

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P3609 SAS No.: P3609 SDG NO.: P3609

EPA Sample No.: SSTDCCC040 Date Analyzed: 08/19/2024

Lab File ID: BM047245.D Time Analyzed: 09:42

Instrument ID: BNA_M GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	274830	7.369	1117660	10.12	774502	14.03
UPPER LIMIT	549660	7.869	2235320	10.622	1549000	14.527
LOWER LIMIT	137415	6.869	558830	9.622	387251	13.527
EPA SAMPLE NO.						
01 915-J-WS-081424	196935	7.36	740406	10.12	509042	14.03
02 920-J-WS-081424	161055	7.36	602261	10.12	400185	14.02
03 PB162757BS	200715	7.36	840423	10.12	601718	14.03
04 PB162757BSD	188275	7.36	791375	10.12	557332	14.03

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.