

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P3609 SAS No.: P3609 SDG NO.: P3609

EPA Sample No.: SSTDCCC040 Date Analyzed: 08/22/2024

Lab File ID: BM047308.D Time Analyzed: 09:14

Instrument ID: BNA\_M GC Column: ZB-GR ID: 0.25 (mm)

|                | IS1 (DCB)<br>AREA # | RT #  | IS2 (NPT)<br>AREA # | RT #  | IS3 (ANT)<br>AREA # | RT #   |
|----------------|---------------------|-------|---------------------|-------|---------------------|--------|
| 12 HOUR STD    | 315372              | 7.351 | 1211800             | 10.11 | 822573              | 14.02  |
| UPPER LIMIT    | 630744              | 7.851 | 2423600             | 10.61 | 1645150             | 14.516 |
| LOWER LIMIT    | 157686              | 6.851 | 605900              | 9.61  | 411287              | 13.516 |
| EPA SAMPLE NO. |                     |       |                     |       |                     |        |
| 01 PB162757BL  | 292730              | 7.35  | 1111490             | 10.10 | 771992              | 14.01  |

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.

\* Values outside of QC limits.