

Cover Page

Order ID : P3609

Project ID : Former Schlumberger Site Princeton NJ

Client : JACOBS Engineering Group, Inc.

Lab Sample Number

P3609-01
P3609-02
P3609-03

Client Sample Number

915-J-WS-081424
920-J-WS-081424
TB-01-081424

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : _____

Date: 10/15/2024

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

JACOBS Engineering Group, Inc.

Project Name: Former Schlumberger Site Princeton NJ

Project # N/A

Chemtech Project # P3609

Test Name: SVOCMS Group6

A. Number of Samples and Date of Receipt:

3 Water samples were received on 08/14/2024.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: Hexavalent Chromium, Mercury, Metals Group4, SVOCMS Group3, SVOCMS Group6 and VOCMS Group6. This data package contains results for SVOCMS Group6.

C. Analytical Techniques:

The samples were analyzed on instrument BNA_M using GC Column ZB-SemiVolatiles Guardian which is 30 meters, 0.25 mm ID, 0.5 um df, Catalog # 7HG-G027-17-GGAThe analysis of SVOCMS Group6 was based on method 8270E and extraction was done based on method 3510.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria except for 920-J-WS-081424 [2,4 and6-Tribromophenol - 117%, this compound did not meet the NJDKQP criteria but met the in-house criteria, Therefore no corrective action was required.

The Internal Standards Areas met the acceptable requirements.

The Retention Times were acceptable for all samples.

The RPD met criteria .

The Blank Spike met requirements for all samples .

The Blank Spike Duplicate met requirements for all samples .

The Blank analysis did not indicate the presence of lab contamination.

The % RSD is greater than 15% in the Initial Calibration (Method 8270-BM081024.M) for Benzaldehyde, this compound is passing on Quadratic regression

The Continuous Calibration met the requirements .

The Tuning criteria met requirements.



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

E. Additional Comments:

This data package has been revised due to parameter list changed.

The Form 6 is not included in the data package because the Initial Calibration was performed using 8 points.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <15% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 15% for the Initial Calibration curve for SW-846 analysis.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following “ Results Qualifiers” are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. “10 U”. This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as “12 B”.
E	Indicates the analyte ‘s concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a “P”.
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

ALLIANCE 284 Sheffield Street, Mountainside New Jersey 07092

NEW JERSEY LAB ID#: 20012; NEW YORK LAB ID#: 11376

GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NUMBER: P3609

MATRIX: Water

METHOD: 8270E/3510

		NA	NO	YES
1.	Chromatograms Labeled/Compounds Identified. (Field samples and Method Blanks)			✓
2.	GC/MS Tuning Specifications. DFTPP Meet Criteria. (NOTE THAT THERE ARE DIFFERENT CRITERIA FOR NY ASP CLP, CLP AND NJ)			✓
3.	GC/MS Tuning Frequency - Performed every 24 hours for 600 series and 12 hours for 8000 Series.			✓
4.	GC/MS Calibration - Initial Calibration performed within 30 days before sample analysis and continuing calibration performed within 24 hours of sample analysis for 600 series and 12 hours for 8000 series.			✓
5.	GC/MS Calibration Requirements.			✓
	The % RSD is greater than 15% in the Initial Calibration (Method 8270-BM081024.M) for Benzaldehyde, this compound is passing on Quadratic regression			
	The Continuous Calibration met the requirements .			
6.	Blank Contamination - If yes, list compounds and concentrations in each blank:		✓	
7.	Surrogate Recoveries Meet Criteria		✓	
	If not met, list those compounds and their recoveries which fall outside the acceptable ranges.			
	The Surrogate recoveries met the acceptable criteria except for 920-J-WS-081424 [2,4 and6-Tribromophenol - 117%, this compound did not meet the NJDKQP criteria but met the in-house criteria, Therefor no corrective action was required.			
8.	Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria			✓
	If not met, list those compounds and their recoveries which fall outside the acceptable range.			
	The Blank Spike met requirements for all samples . The Blank Spike Duplicate met requirements for all samples .			
9.	Internal Standard Area/Retention Time Shift Meet Criteria			✓
	Comments:			

ALLIANCE 284 Sheffield Street, Mountainside New Jersey 07092

NEW JERSEY LAB ID#: 20012; NEW YORK LAB ID#: 11376

GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

(CONTINUED)

		NA	NO	YES
10.	Extraction Holding Time Met			✓
	If not met, list number of days exceeded for each sample:			
11.	Analysis Holding Time Met			✓
	If not met, list number of days exceeded for each sample:			

ADDITIONAL COMMENTS:

This data package has been revised due to parameter list changed.

The Form 6 is not included in the data package because the Initial Calibration was performed using 8 points.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <15% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 15% for the Initial Calibration curve for SW-846 analysis.

QA REVIEW

Date

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: P3609

Completed

For thorough review, the report must have the following:

GENERAL:

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Custody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

ANALYTICAL:

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: SOHIL JODHANI

Date: 10/15/2024

LAB CHRONICLE

OrderID:	P3609	OrderDate:	8/14/2024 12:48:00 PM
Client:	JACOBS Engineering Group, Inc.	Project:	Former Schlumberger Site Princeton NJ
Contact:	Mary I. Murphy	Location:	D21,VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
P3609-01	915-J-WS-081424	Water			08/14/24			08/14/24
			SVOCMS Group3	8270-Modified		08/15/24	08/17/24	
			SVOCMS Group6	8270E		08/15/24	08/19/24	
P3609-02	920-J-WS-081424	Water			08/14/24			08/14/24
			SVOCMS Group3	8270-Modified		08/15/24	08/17/24	
			SVOCMS Group6	8270E		08/15/24	08/19/24	



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: P3609
Client: JACOBS Engineering Group, Inc.

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID :				0.000				
Total Svoc :					0.00			
Total Concentration:					0.00			



QC SUMMARY

Surrogate Summary

SW-846

SDG No.: P3609

Client: JACOBS Engineering Group, Inc.

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
P3609-01	915-J-WS-081424	2-Fluorophenol	150	55.5	37		15 (10)	110 (139)
		Phenol-d6	150	34.7	23		15 (10)	110 (134)
		Nitrobenzene-d5	100	68.0	68		30 (49)	130 (133)
		2-Fluorobiphenyl	100	70.3	70		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	155	103		15 (44)	110 (137)
		Terphenyl-d14	100	86.4	86		30 (48)	130 (125)
P3609-02	920-J-WS-081424	2-Fluorophenol	150	57.3	38		15 (10)	110 (139)
		Phenol-d6	150	34.2	23		15 (10)	110 (134)
		Nitrobenzene-d5	100	77.0	77		30 (49)	130 (133)
		2-Fluorobiphenyl	100	80.0	80		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	175	117	*	15 (44)	110 (137)
		Terphenyl-d14	100	92.1	92		30 (48)	130 (125)
PB162757BL	PB162757BL	2-Fluorophenol	150	120	80		15 (10)	110 (139)
		Phenol-d6	150	111	74		15 (10)	110 (134)
		Nitrobenzene-d5	100	75.0	75		30 (49)	130 (133)
		2-Fluorobiphenyl	100	79.1	79		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	131	88		15 (44)	110 (137)
		Terphenyl-d14	100	96.4	96		30 (48)	130 (125)
PB162757BS	PB162757BS	2-Fluorophenol	150	134	89		15 (10)	110 (139)
		Phenol-d6	150	131	87		15 (10)	110 (134)
		Nitrobenzene-d5	100	78.7	79		30 (49)	130 (133)
		2-Fluorobiphenyl	100	80.7	81		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	141	94		15 (44)	110 (137)
		Terphenyl-d14	100	95.4	95		30 (48)	130 (125)
PB162757BSD	PB162757BSD	2-Fluorophenol	150	135	90		15 (10)	110 (139)
		Phenol-d6	150	131	87		15 (10)	110 (134)
		Nitrobenzene-d5	100	78.4	78		30 (49)	130 (133)
		2-Fluorobiphenyl	100	81.7	82		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	142	95		15 (44)	110 (137)
		Terphenyl-d14	100	94.7	95		30 (48)	130 (125)

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P3609

Client: JACOBS Engineering Group, Inc.

Analytical Method: 8270E

DataFile: BM047250.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD		Limits		RPD
								Qual	Qual	Low	High	
PB162757BS	Pyridine	50	29.0	ug/L	58					20 (29)	160 (97)	
	Benzaldehyde	50	16.9	ug/L	34					20 (10)	160 (162)	
	2-Methylphenol	50	49.9	ug/L	100					70 (69)	130 (109)	
	3+4-Methylphenols	50	49.6	ug/L	99					20 (67)	160 (106)	
	Hexachloroethane	50	41.4	ug/L	83					20 (76)	160 (118)	
	Nitrobenzene	50	41.1	ug/L	82					70 (58)	130 (106)	
	Naphthalene	50	43.6	ug/L	87					70 (64)	130 (107)	
	Hexachlorobutadiene	50	44.7	ug/L	89					70 (69)	130 (101)	
	2-Methylnaphthalene	50	45.4	ug/L	91					70 (64)	130 (107)	
	2,4,6-Trichlorophenol	50	49.9	ug/L	100					70 (61)	130 (110)	
	2,4,5-Trichlorophenol	50	47.2	ug/L	94					70 (70)	130 (106)	
	Acenaphthylene	50	47.6	ug/L	95					70 (79)	130 (103)	
	Acenaphthene	50	43.1	ug/L	86					70 (59)	130 (113)	
	Dibenzofuran	50	45.4	ug/L	91					70 (65)	130 (106)	
	2,4-Dinitrotoluene	50	47.7	ug/L	95					70 (60)	130 (115)	
	Fluorene	50	45.1	ug/L	90					70 (64)	130 (107)	
	Hexachlorobenzene	50	44.7	ug/L	89					70 (73)	130 (106)	
	Pentachlorophenol	100	91.2	ug/L	91					20 (47)	160 (114)	
	Phenanthrene	50	45.8	ug/L	92					70 (62)	130 (109)	
	Anthracene	50	47.2	ug/L	94					70 (65)	130 (110)	
	Carbazole	50	45.9	ug/L	92					70 (62)	130 (106)	
	Di-n-butylphthalate	50	46.3	ug/L	93					70 (64)	130 (106)	
	Fluoranthene	50	47.7	ug/L	95					70 (64)	130 (110)	
	Pyrene	50	44.0	ug/L	88					70 (71)	130 (103)	
	Benzo(a)anthracene	50	47.3	ug/L	95					70 (62)	130 (107)	
	Chrysene	50	46.2	ug/L	92					70 (61)	130 (108)	
	bis(2-Ethylhexyl)phthalate	50	45.3	ug/L	91					70 (59)	130 (110)	
	Benzo(b)fluoranthene	50	47.8	ug/L	96					70 (77)	130 (113)	
	Benzo(k)fluoranthene	50	47.9	ug/L	96					70 (77)	130 (105)	
	Benzo(a)pyrene	50	52.1	ug/L	104					70 (72)	130 (131)	
	Indeno(1,2,3-cd)pyrene	50	52.5	ug/L	105					70 (72)	130 (105)	
	Dibenz(a,h)anthracene	50	52.5	ug/L	105					70 (78)	130 (115)	
	Benzo(g,h,i)perylene	50	48.7	ug/L	97					70 (75)	130 (118)	
	1,4-Dioxane	50	26.1	ug/L	52					20 (38)	160 (125)	
	1-Methylnaphthalene	50	42.9	ug/L	86					70 (51)	130 (114)	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P3609

Client: JACOBS Engineering Group, Inc.

Analytical Method: 8270E

DataFile: BM047251.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	RPD		Limits		
							Qual	Qual	Low	High	RPD
PB162757BSD	Pyridine	50	29.0	ug/L	58	0			20 (29)	160 (97)	20 (20)
	Benzaldehyde	50	17.2	ug/L	34	2			20 (10)	160 (162)	20 (20)
	2-Methylphenol	50	50.1	ug/L	100	0			70 (69)	130 (109)	20 (20)
	3+4-Methylphenols	50	49.6	ug/L	99	0			20 (67)	160 (106)	20 (20)
	Hexachloroethane	50	41.6	ug/L	83	0			20 (76)	160 (118)	20 (20)
	Nitrobenzene	50	41.2	ug/L	82	0			70 (58)	130 (106)	20 (20)
	Naphthalene	50	43.8	ug/L	88	0			70 (64)	130 (107)	20 (20)
	Hexachlorobutadiene	50	45.1	ug/L	90	1			70 (69)	130 (101)	20 (20)
	2-Methylnaphthalene	50	45.3	ug/L	91	0			70 (64)	130 (107)	20 (20)
	2,4,6-Trichlorophenol	50	50.1	ug/L	100	0			70 (61)	130 (110)	20 (20)
	2,4,5-Trichlorophenol	50	47.3	ug/L	95	0			70 (70)	130 (106)	20 (20)
	Acenaphthylene	50	48.2	ug/L	96	1			70 (79)	130 (103)	20 (20)
	Acenaphthene	50	44.0	ug/L	88	2			70 (59)	130 (113)	20 (20)
	Dibenzofuran	50	45.6	ug/L	91	0			70 (65)	130 (106)	20 (20)
	2,4-Dinitrotoluene	50	47.3	ug/L	95	1			70 (60)	130 (115)	20 (20)
	Fluorene	50	45.5	ug/L	91	1			70 (64)	130 (107)	20 (20)
	Hexachlorobenzene	50	45.1	ug/L	90	1			70 (73)	130 (106)	20 (20)
	Pentachlorophenol	100	91.8	ug/L	92	1			20 (47)	160 (114)	20 (20)
	Phenanthrene	50	46.0	ug/L	92	0			70 (62)	130 (109)	20 (20)
	Anthracene	50	47.6	ug/L	95	1			70 (65)	130 (110)	20 (20)
	Carbazole	50	46.6	ug/L	93	2			70 (62)	130 (106)	20 (20)
	Di-n-butylphthalate	50	46.4	ug/L	93	0			70 (64)	130 (106)	20 (20)
	Fluoranthene	50	47.9	ug/L	96	0			70 (64)	130 (110)	20 (20)
	Pyrene	50	43.5	ug/L	87	1			70 (71)	130 (103)	20 (20)
	Benzo(a)anthracene	50	47.2	ug/L	94	0			70 (62)	130 (107)	20 (20)
	Chrysene	50	46.7	ug/L	93	1			70 (61)	130 (108)	20 (20)
	bis(2-Ethylhexyl)phthalate	50	44.8	ug/L	90	1			70 (59)	130 (110)	20 (20)
	Benzo(b)fluoranthene	50	47.2	ug/L	94	1			70 (77)	130 (113)	20 (20)
	Benzo(k)fluoranthene	50	47.5	ug/L	95	1			70 (77)	130 (105)	20 (20)
	Benzo(a)pyrene	50	51.8	ug/L	104	1			70 (72)	130 (131)	20 (20)
	Indeno(1,2,3-cd)pyrene	50	53.2	ug/L	106	1			70 (72)	130 (105)	20 (20)
	Dibenz(a,h)anthracene	50	53.0	ug/L	106	1			70 (78)	130 (115)	20 (20)
	Benzo(g,h,i)perylene	50	49.4	ug/L	99	1			70 (75)	130 (118)	20 (20)
	1,4-Dioxane	50	26.0	ug/L	52	0			20 (38)	160 (125)	20 (20)
	1-Methylnaphthalene	50	42.8	ug/L	86	0			70 (51)	130 (114)	20 (20)

4B

SEMIVOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

PB162757BL

Lab Name: CHEMTECH

Contract: JAC005

Lab Code: CHEM Case No.: P3609

SAS No.: P3609 SDG NO.: P3609

Lab File ID: BM047313.D

Lab Sample ID: PB162757BL

Instrument ID: BNA_M

Date Extracted: 08/15/2024

Matrix: (soil/water) Water

Date Analyzed: 08/22/2024

Level: (low/med) LOW

Time Analyzed: 12:32

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
915-J-WS-081424	P3609-01	BM047256.D	08/19/2024
920-J-WS-081424	P3609-02	BM047257.D	08/19/2024
PB162757BS	PB162757BS	BM047250.D	08/19/2024
PB162757BSD	PB162757BSD	BM047251.D	08/19/2024

COMMENTS: _____



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECHContract: JAC005Lab Code: CHEMSAS No.: P3609SDG NO.: P3609Lab File ID: BM047139.DDFTPP Injection Date: 08/10/2024Instrument ID: BNA_MDFTPP Injection Time: 12:26

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	35.7
68	Less than 2.0% of mass 69	0.6 (1.3) 1
69	Mass 69 relative abundance	44.5
70	Less than 2.0% of mass 69	0.2 (0.5) 1
127	10.0 - 80.0% of mass 198	49.9
197	Less than 2.0% of mass 198	0.4
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	7
275	10.0 - 60.0% of mass 198	25.8
365	Greater than 1% of mass 198	3.6
441	Present, but less than mass 443	9.2
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	11.4 (19.7) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC2.5	SSTDICC2.5	BM047140.D	08/10/2024	13:12
SSTDICC005	SSTDICC005	BM047141.D	08/10/2024	13:52
SSTDICC010	SSTDICC010	BM047142.D	08/10/2024	14:31
SSTDICC020	SSTDICC020	BM047143.D	08/10/2024	15:11
SSTDICCC040	SSTDICCC040	BM047144.D	08/10/2024	15:51
SSTDICC050	SSTDICC050	BM047145.D	08/10/2024	16:31
SSTDICC060	SSTDICC060	BM047146.D	08/10/2024	17:11
SSTDICC080	SSTDICC080	BM047147.D	08/10/2024	17:51



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECHContract: JAC005Lab Code: CHEMSAS No.: P3609 SDG NO.: P3609Lab File ID: BM047244.DDFTPP Injection Date: 08/19/2024Instrument ID: BNA_MDFTPP Injection Time: 09:03

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	28.3
68	Less than 2.0% of mass 69	0.5 (1.4) 1
69	Mass 69 relative abundance	35.6
70	Less than 2.0% of mass 69	0.1 (0.3) 1
127	10.0 - 80.0% of mass 198	42.3
197	Less than 2.0% of mass 198	0.4
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.8
275	10.0 - 60.0% of mass 198	25.6
365	Greater than 1% of mass 198	3.4
441	Present, but less than mass 443	9.1
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	11.7 (19.8) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BM047245.D	08/19/2024	09:42
PB162757BS	PB162757BS	BM047250.D	08/19/2024	13:01
PB162757BSD	PB162757BSD	BM047251.D	08/19/2024	13:42
915-J-WS-081424	P3609-01	BM047256.D	08/19/2024	17:28
920-J-WS-081424	P3609-02	BM047257.D	08/19/2024	18:08



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECHContract: JAC005Lab Code: CHEMSAS No.: P3609 SDG NO.: P3609Lab File ID: BM047307.DDFTPP Injection Date: 08/22/2024Instrument ID: BNA_MDFTPP Injection Time: 08:35

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	25.4
68	Less than 2.0% of mass 69	0.5 (1.5) 1
69	Mass 69 relative abundance	32.2
70	Less than 2.0% of mass 69	0.1 (0.4) 1
127	10.0 - 80.0% of mass 198	39.1
197	Less than 2.0% of mass 198	0.3
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.9
275	10.0 - 60.0% of mass 198	26.2
365	Greater than 1% of mass 198	3.2
441	Present, but less than mass 443	9.3
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	11.5 (19.2) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BM047308.D	08/22/2024	09:14
PB162757BL	PB162757BL	BM047313.D	08/22/2024	12:32

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P3609 SAS No.: P3609 SDG NO.: P3609

EPA Sample No.: SSTDCCC040 Date Analyzed: 08/19/2024

Lab File ID: BM047245.D Time Analyzed: 09:42

Instrument ID: BNA_M GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	274830	7.369	1117660	10.12	774502	14.03
UPPER LIMIT	549660	7.869	2235320	10.622	1549000	14.527
LOWER LIMIT	137415	6.869	558830	9.622	387251	13.527
EPA SAMPLE NO.						
01 915-J-WS-081424	196935	7.36	740406	10.12	509042	14.03
02 920-J-WS-081424	161055	7.36	602261	10.12	400185	14.02
03 PB162757BS	200715	7.36	840423	10.12	601718	14.03
04 PB162757BSD	188275	7.36	791375	10.12	557332	14.03

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P3609 SAS No.: P3609 SDG NO.: P3609

EPA Sample No.: SSTDCCC040 Date Analyzed: 08/19/2024

Lab File ID: BM047245.D Time Analyzed: 09:42

Instrument ID: BNA_M GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	1666060	16.792	1393630	21.039	1655410	23.744
UPPER LIMIT	3332120	17.292	2787260	21.539	3310820	24.244
LOWER LIMIT	833030	16.292	696815	20.539	827705	23.244
EPA SAMPLE NO.						
01 915-J-WS-081424	1167960	16.79	1274470	21.03	1251150	23.73
02 920-J-WS-081424	911557	16.79	1070130	21.03	1179970	23.73
03 PB162757BS	1358930	16.79	1331980	21.04	1613230	23.74
04 PB162757BSD	1252430	16.79	1245980	21.03	1572180	23.74

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P3609 SAS No.: P3609 SDG NO.: P3609

EPA Sample No.: SSTDCCC040 Date Analyzed: 08/22/2024

Lab File ID: BM047308.D Time Analyzed: 09:14

Instrument ID: BNA_M GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	315372	7.351	1211800	10.11	822573	14.02
UPPER LIMIT	630744	7.851	2423600	10.61	1645150	14.516
LOWER LIMIT	157686	6.851	605900	9.61	411287	13.516
EPA SAMPLE NO.						
01 PB162757BL	292730	7.35	1111490	10.10	771992	14.01

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P3609 SAS No.: P3609 SDG NO.: P3609

EPA Sample No.: SSTDCCC040 Date Analyzed: 08/22/2024

Lab File ID: BM047308.D Time Analyzed: 09:14

Instrument ID: BNA_M GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	1768060	16.78	1535450	21.021	1629250	23.709
UPPER LIMIT	3536120	17.28	3070900	21.521	3258500	24.209
LOWER LIMIT	884030	16.28	767725	20.521	814625	23.209
EPA SAMPLE NO.						
01 PB162757BL	1724660	16.77	1583440	21.02	1534870	23.71

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.



SAMPLE DATA

Report of Analysis

Client:	JACOBS Engineering Group, Inc.	Date Collected:	08/14/24
Project:	Former Schlumberger Site Princeton NJ	Date Received:	08/14/24
Client Sample ID:	915-J-WS-081424	SDG No.:	P3609
Lab Sample ID:	P3609-01	Matrix:	Water
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	960 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group6
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BM047256.D	1	08/15/24 08:23	08/19/24 17:28	PB162757

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	1.60	U	1.60	5.20	ug/L
100-52-7	Benzaldehyde	4.20	U	4.20	10.4	ug/L
95-48-7	2-Methylphenol	1.20	U	1.20	5.20	ug/L
65794-96-9	3+4-Methylphenols	1.20	U	1.20	10.4	ug/L
67-72-1	Hexachloroethane	1.10	U	1.10	5.20	ug/L
98-95-3	Nitrobenzene	1.30	U	1.30	5.20	ug/L
91-20-3	Naphthalene	1.10	U	1.10	5.20	ug/L
87-68-3	Hexachlorobutadiene	1.30	U	1.30	5.20	ug/L
91-57-6	2-Methylnaphthalene	1.20	U	1.20	5.20	ug/L
88-06-2	2,4,6-Trichlorophenol	0.93	U	0.93	5.20	ug/L
95-95-4	2,4,5-Trichlorophenol	1.10	U	1.10	5.20	ug/L
208-96-8	Acenaphthylene	1.10	U	1.10	5.20	ug/L
83-32-9	Acenaphthene	0.84	U	0.84	5.20	ug/L
132-64-9	Dibenzofuran	0.97	U	0.97	5.20	ug/L
121-14-2	2,4-Dinitrotoluene	1.60	U	1.60	5.20	ug/L
86-73-7	Fluorene	1.00	U	1.00	5.20	ug/L
118-74-1	Hexachlorobenzene	1.20	U	1.20	5.20	ug/L
87-86-5	Pentachlorophenol	1.90	U	1.90	10.4	ug/L
85-01-8	Phenanthrene	0.93	U	0.93	5.20	ug/L
120-12-7	Anthracene	1.10	U	1.10	5.20	ug/L
86-74-8	Carbazole	1.20	U	1.20	5.20	ug/L
84-74-2	Di-n-butylphthalate	1.50	U	1.50	5.20	ug/L
206-44-0	Fluoranthene	1.30	U	1.30	5.20	ug/L
129-00-0	Pyrene	1.10	U	1.10	5.20	ug/L
56-55-3	Benzo(a)anthracene	0.98	U	0.98	5.20	ug/L
218-01-9	Chrysene	0.90	U	0.90	5.20	ug/L
117-81-7	Bis(2-ethylhexyl)phthalate	2.00	U	2.00	5.20	ug/L
205-99-2	Benzo(b)fluoranthene	1.20	U	1.20	5.20	ug/L
207-08-9	Benzo(k)fluoranthene	1.20	U	1.20	5.20	ug/L

Report of Analysis

Client:	JACOBS Engineering Group, Inc.	Date Collected:	08/14/24
Project:	Former Schlumberger Site Princeton NJ	Date Received:	08/14/24
Client Sample ID:	915-J-WS-081424	SDG No.:	P3609
Lab Sample ID:	P3609-01	Matrix:	Water
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	960 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group6
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BM047256.D	1	08/15/24 08:23	08/19/24 17:28	PB162757

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
50-32-8	Benzo(a)pyrene	1.70	U	1.70	5.20	ug/L
193-39-5	Indeno(1,2,3-cd)pyrene	1.10	U	1.10	5.20	ug/L
53-70-3	Dibenzo(a,h)anthracene	1.20	U	1.20	5.20	ug/L
191-24-2	Benzo(g,h,i)perylene	1.20	U	1.20	5.20	ug/L
123-91-1	1,4-Dioxane	1.30	U	1.30	5.20	ug/L
90-12-0	1-Methylnaphthalene	0.90	U	0.90	5.20	ug/L

SURROGATES

367-12-4	2-Fluorophenol	55.5		15 (10) - 110 (139)	37%	SPK: 150
13127-88-3	Phenol-d6	34.7		15 (10) - 110 (134)	23%	SPK: 150
4165-60-0	Nitrobenzene-d5	68.0		30 (49) - 130 (133)	68%	SPK: 100
321-60-8	2-Fluorobiphenyl	70.3		30 (52) - 130 (132)	70%	SPK: 100
118-79-6	2,4,6-Tribromophenol	155		15 (44) - 110 (137)	103%	SPK: 150
1718-51-0	Terphenyl-d14	86.4		30 (48) - 130 (125)	86%	SPK: 100

INTERNAL STANDARDS

3855-82-1	1,4-Dichlorobenzene-d4	197000	7.363
1146-65-2	Naphthalene-d8	740000	10.122
15067-26-2	Acenaphthene-d10	509000	14.027
1517-22-2	Phenanthrene-d10	1170000	16.786
1719-03-5	Chrysene-d12	1270000	21.027
1520-96-3	Perylene-d12	1250000	23.733

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081924\
Data File : BM047256.D
Acq On : 19 Aug 2024 17:28
Operator : RC/JU
Sample : P3609-01
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
915-J-WS-081424

Quant Time: Aug 19 18:03:21 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sun Aug 11 23:47:58 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.363	152	196935	20.000	ng	-0.01
21) Naphthalene-d8	10.122	136	740406	20.000	ng	-0.01
39) Acenaphthene-d10	14.027	164	509042	20.000	ng	0.00
64) Phenanthrene-d10	16.786	188	1167962	20.000	ng	0.00
76) Chrysene-d12	21.027	240	1274470	20.000	ng	0.00
86) Perylene-d12	23.733	264	1251147	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.016	112	611708	55.502	ng	0.00
7) Phenol-d6	6.575	99	527665	34.746	ng	-0.01
23) Nitrobenzene-d5	8.510	82	999874	67.999	ng	-0.01
42) 2,4,6-Tribromophenol	15.533	330	916245	155.232	ng	0.00
45) 2-Fluorobiphenyl	12.633	172	2319353	70.282	ng	-0.01
79) Terphenyl-d14	19.456	244	5138867	86.400	ng	0.00


Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081924\
Data File : BM047256.D
Acq On : 19 Aug 2024 17:28
Operator : RC/JU
Sample : P3609-01
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
915-J-WS-081424


Quant Time: Aug 19 18:03:21 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sun Aug 11 23:47:58 2024
Response via : Initial Calibration




Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081924\
Data File : BM047256.D
Acq On : 19 Aug 2024 17:28
Operator : RC/JU
Sample : P3609-01
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
915-J-WS-081424

Quant Time: Aug 19 18:03:21 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sun Aug 11 23:47:58 2024
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.363 min Scan# 70
Delta R.T. -0.012 min
Lab File: BM047256.D
Acq: 19 Aug 2024 17:28




Instrument :
BNA_M
ClientSampleId :
915-J-WS-081424


Tgt Ion:152 Resp: 196935
Ion Ratio Lower Upper
152 100
150 157.0 120.5 180.7
115 54.6 45.5 68.3



#5
2-Fluorophenol
Concen: 55.502 ng
RT: 5.016 min Scan# 362
Delta R.T. -0.006 min
Lab File: BM047256.D
Acq: 19 Aug 2024 17:28


Tgt Ion:112 Resp: 611708
Ion Ratio Lower Upper
112 100
64 56.4 46.7 70.1
63 30.0 24.8 37.2








#7
Phenol-d6
Concen: 34.746 ng
RT: 6.575 min Scan# 61
Delta R.T. -0.012 min
Lab File: BM047256.D
Acq: 19 Aug 2024 17:28

Instrument :
BNA_M
ClientSampleId :
915-J-WS-081424







Tgt Ion: 99 Resp: 527665
Ion Ratio Lower Upper
99 100
42 16.7 14.1 21.1
71 34.5 28.5 42.7



#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.122 min Scan# 1230
Delta R.T. -0.012 min
Lab File: BM047256.D
Acq: 19 Aug 2024 17:28

Tgt Ion: 136 Resp: 740406
Ion Ratio Lower Upper
136 100
137 10.8 8.8 13.2
54 7.0 6.0 9.0
68 5.7 4.8 7.2









#23
Nitrobenzene-d5
Concen: 67.999 ng
RT: 8.510 min Scan# 91
Delta R.T. -0.012 min
Lab File: BM047256.D
Acq: 19 Aug 2024 17:28




Instrument :
BNA_M
ClientSampleId :
915-J-WS-081424


Tgt Ion: 82 Resp: 999874
Ion Ratio Lower Upper
82 100
128 42.4 32.3 48.5
54 43.6 37.4 56.2



#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.027 min Scan# 1894
Delta R.T. -0.006 min
Lab File: BM047256.D
Acq: 19 Aug 2024 17:28

Tgt Ion: 164 Resp: 509042
Ion Ratio Lower Upper
164 100
162 99.0 80.7 121.1
160 44.0 36.0 54.0





#42

2,4,6-Tribromophenol

Concen: 155.232 ng

RT: 15.533 min Scan# 21

Delta R.T. -0.006 min

Lab File: BM047256.D

Acq: 19 Aug 2024 17:28

Instrument :

BNA_M

ClientSampleId :

915-J-WS-081424





Tgt Ion:330 Resp: 916245

Ion Ratio Lower Upper

330 100

332 95.9 77.4 116.2

141 35.4 31.6 47.4



#45

2-Fluorobiphenyl

Concen: 70.282 ng

RT: 12.633 min Scan# 1657

Delta R.T. -0.012 min

Lab File: BM047256.D

Acq: 19 Aug 2024 17:28




Tgt Ion:172 Resp: 2319353


Ion Ratio Lower Upper

172 100

171 34.4 27.8 41.6

170 23.0 18.7 28.1









#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 16.786 min Scan# 21
Delta R.T. -0.006 min
Lab File: BM047256.D
Acq: 19 Aug 2024 17:28




Instrument :
BNA_M
ClientSampleId :
915-J-WS-081424


Tgt Ion:188 Resp: 1167962
Ion Ratio Lower Upper
188 100
94 8.2 7.0 10.6
80 9.2 8.0 12.0



#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.027 min Scan# 3084
Delta R.T. -0.006 min
Lab File: BM047256.D
Acq: 19 Aug 2024 17:28

Tgt Ion:240 Resp: 1274470
Ion Ratio Lower Upper
240 100
120 10.4 8.2 12.2
236 26.1 20.6 30.8









#79
Terphenyl-d14
Concen: 86.400 ng
RT: 19.456 min Scan# 21
Delta R.T. -0.006 min
Lab File: BM047256.D
Acq: 19 Aug 2024 17:28




Instrument :
BNA_M
ClientSampleId :
915-J-WS-081424

Tgt Ion:244 Resp: 5138867
Ion Ratio Lower Upper
244 100
212 7.5 6.2 9.4
122 8.9 7.1 10.7



#86
Perylene-d12
Concen: 20.000 ng
RT: 23.733 min Scan# 3544
Delta R.T. 0.006 min
Lab File: BM047256.D
Acq: 19 Aug 2024 17:28

Tgt Ion:264 Resp: 1251147
Ion Ratio Lower Upper
264 100
260 23.4 18.5 27.7
265 22.6 17.3 25.9



Report of Analysis

Client:	JACOBS Engineering Group, Inc.	Date Collected:	08/14/24
Project:	Former Schlumberger Site Princeton NJ	Date Received:	08/14/24
Client Sample ID:	920-J-WS-081424	SDG No.:	P3609
Lab Sample ID:	P3609-02	Matrix:	Water
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group6
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BM047257.D	1	08/15/24 08:23	08/19/24 18:08	PB162757

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	1.60	U	1.60	5.00	ug/L
100-52-7	Benzaldehyde	4.00	U	4.00	10.0	ug/L
95-48-7	2-Methylphenol	1.10	U	1.10	5.00	ug/L
65794-96-9	3+4-Methylphenols	1.20	U	1.20	10.0	ug/L
67-72-1	Hexachloroethane	1.00	U	1.00	5.00	ug/L
98-95-3	Nitrobenzene	1.30	U	1.30	5.00	ug/L
91-20-3	Naphthalene	1.00	U	1.00	5.00	ug/L
87-68-3	Hexachlorobutadiene	1.30	U	1.30	5.00	ug/L
91-57-6	2-Methylnaphthalene	1.10	U	1.10	5.00	ug/L
88-06-2	2,4,6-Trichlorophenol	0.89	U	0.89	5.00	ug/L
95-95-4	2,4,5-Trichlorophenol	1.00	U	1.00	5.00	ug/L
208-96-8	Acenaphthylene	1.00	U	1.00	5.00	ug/L
83-32-9	Acenaphthene	0.81	U	0.81	5.00	ug/L
132-64-9	Dibenzofuran	0.93	U	0.93	5.00	ug/L
121-14-2	2,4-Dinitrotoluene	1.50	U	1.50	5.00	ug/L
86-73-7	Fluorene	0.96	U	0.96	5.00	ug/L
118-74-1	Hexachlorobenzene	1.10	U	1.10	5.00	ug/L
87-86-5	Pentachlorophenol	1.90	U	1.90	10.0	ug/L
85-01-8	Phenanthrene	0.89	U	0.89	5.00	ug/L
120-12-7	Anthracene	1.10	U	1.10	5.00	ug/L
86-74-8	Carbazole	1.20	U	1.20	5.00	ug/L
84-74-2	Di-n-butylphthalate	1.50	U	1.50	5.00	ug/L
206-44-0	Fluoranthene	1.30	U	1.30	5.00	ug/L
129-00-0	Pyrene	1.10	U	1.10	5.00	ug/L
56-55-3	Benzo(a)anthracene	0.94	U	0.94	5.00	ug/L
218-01-9	Chrysene	0.86	U	0.86	5.00	ug/L
117-81-7	Bis(2-ethylhexyl)phthalate	1.90	U	1.90	5.00	ug/L
205-99-2	Benzo(b)fluoranthene	1.10	U	1.10	5.00	ug/L
207-08-9	Benzo(k)fluoranthene	1.20	U	1.20	5.00	ug/L

Report of Analysis

Client:	JACOBS Engineering Group, Inc.	Date Collected:	08/14/24
Project:	Former Schlumberger Site Princeton NJ	Date Received:	08/14/24
Client Sample ID:	920-J-WS-081424	SDG No.:	P3609
Lab Sample ID:	P3609-02	Matrix:	Water
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group6
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BM047257.D	1	08/15/24 08:23	08/19/24 18:08	PB162757

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
50-32-8	Benzo(a)pyrene	1.70	U	1.70	5.00	ug/L
193-39-5	Indeno(1,2,3-cd)pyrene	1.00	U	1.00	5.00	ug/L
53-70-3	Dibenzo(a,h)anthracene	1.20	U	1.20	5.00	ug/L
191-24-2	Benzo(g,h,i)perylene	1.20	U	1.20	5.00	ug/L
123-91-1	1,4-Dioxane	1.30	U	1.30	5.00	ug/L
90-12-0	1-Methylnaphthalene	0.86	U	0.86	5.00	ug/L

SURROGATES

367-12-4	2-Fluorophenol	57.3		15 (10) - 110 (139)	38%	SPK: 150
13127-88-3	Phenol-d6	34.2		15 (10) - 110 (134)	23%	SPK: 150
4165-60-0	Nitrobenzene-d5	77.0		30 (49) - 130 (133)	77%	SPK: 100
321-60-8	2-Fluorobiphenyl	80.0		30 (52) - 130 (132)	80%	SPK: 100
118-79-6	2,4,6-Tribromophenol	175	*	15 (44) - 110 (137)	117%	SPK: 150
1718-51-0	Terphenyl-d14	92.1		30 (48) - 130 (125)	92%	SPK: 100

INTERNAL STANDARDS

3855-82-1	1,4-Dichlorobenzene-d4	161000	7.363
1146-65-2	Naphthalene-d8	602000	10.116
15067-26-2	Acenaphthene-d10	400000	14.021
1517-22-2	Phenanthrene-d10	912000	16.786
1719-03-5	Chrysene-d12	1070000	21.027
1520-96-3	Perylene-d12	1180000	23.732

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081924\
Data File : BM047257.D
Acq On : 19 Aug 2024 18:08
Operator : RC/JU
Sample : P3609-02
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
920-J-WS-081424

Quant Time: Aug 19 18:50:48 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sun Aug 11 23:47:58 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.363	152	161055	20.000	ng	-0.01
21) Naphthalene-d8	10.116	136	602261	20.000	ng	-0.02
39) Acenaphthene-d10	14.021	164	400185	20.000	ng	-0.01
64) Phenanthrene-d10	16.786	188	911557	20.000	ng	0.00
76) Chrysene-d12	21.027	240	1070133	20.000	ng	0.00
86) Perylene-d12	23.732	264	1179966	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.016	112	516759	57.332	ng	0.00
7) Phenol-d6	6.575	99	424276	34.162	ng	-0.01
23) Nitrobenzene-d5	8.510	82	920439	76.955	ng	-0.01
42) 2,4,6-Tribromophenol	15.533	330	812875	175.181	ng	0.00
45) 2-Fluorobiphenyl	12.633	172	2074211	79.951	ng	-0.01
79) Terphenyl-d14	19.456	244	4600226	92.112	ng	0.00


Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081924\
Data File : BM047257.D
Acq On : 19 Aug 2024 18:08
Operator : RC/JU
Sample : P3609-02
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
920-J-WS-081424


Quant Time: Aug 19 18:50:48 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sun Aug 11 23:47:58 2024
Response via : Initial Calibration




Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081924\
Data File : BM047257.D
Acq On : 19 Aug 2024 18:08
Operator : RC/JU
Sample : P3609-02
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
920-J-WS-081424

Quant Time: Aug 19 18:50:48 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sun Aug 11 23:47:58 2024
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.363 min Scan# 763
Delta R.T. -0.012 min
Lab File: BM047257.D
Acq: 19 Aug 2024 18:08




Instrument :
BNA_M
ClientSampleId :
920-J-WS-081424


Tgt Ion:152 Resp: 161055
Ion Ratio Lower Upper
152 100
150 156.8 120.5 180.7
115 53.8 45.5 68.3



#5
2-Fluorophenol
Concen: 57.332 ng
RT: 5.016 min Scan# 362
Delta R.T. -0.006 min
Lab File: BM047257.D
Acq: 19 Aug 2024 18:08


Tgt Ion:112 Resp: 516759
Ion Ratio Lower Upper
112 100
64 56.2 46.7 70.1
63 29.9 24.8 37.2








#7
Phenol-d6
Concen: 34.162 ng
RT: 6.575 min Scan# 61
Delta R.T. -0.012 min
Lab File: BM047257.D
Acq: 19 Aug 2024 18:08

Instrument :
BNA_M
ClientSampleId :
920-J-WS-081424







Tgt Ion: 99 Resp: 424276
Ion Ratio Lower Upper
99 100
42 16.1 14.1 21.1
71 33.6 28.5 42.7



#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.116 min Scan# 1229
Delta R.T. -0.018 min
Lab File: BM047257.D
Acq: 19 Aug 2024 18:08

Tgt Ion: 136 Resp: 602261
Ion Ratio Lower Upper
136 100
137 10.4 8.8 13.2
54 6.9 6.0 9.0
68 5.6 4.8 7.2









#23
Nitrobenzene-d5
Concen: 76.955 ng
RT: 8.510 min Scan# 91
Delta R.T. -0.012 min
Lab File: BM047257.D
Acq: 19 Aug 2024 18:08




Instrument :
BNA_M
ClientSampleId :
920-J-WS-081424


Tgt Ion: 82 Resp: 920439
Ion Ratio Lower Upper
82 100
128 43.3 32.3 48.5
54 45.1 37.4 56.2



#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.021 min Scan# 1893
Delta R.T. -0.012 min
Lab File: BM047257.D
Acq: 19 Aug 2024 18:08

Tgt Ion: 164 Resp: 400185
Ion Ratio Lower Upper
164 100
162 100.4 80.7 121.1
160 44.8 36.0 54.0









#42
2,4,6-Tribromophenol
Concen: 175.181 ng
RT: 15.533 min Scan# 2151
Delta R.T. -0.006 min
Lab File: BM047257.D
Acq: 19 Aug 2024 18:08




Instrument :
BNA_M
ClientSampleId :
920-J-WS-081424


Tgt Ion:330 Resp: 812875
Ion Ratio Lower Upper
330 100
332 96.2 77.4 116.2
141 35.2 31.6 47.4



#45
2-Fluorobiphenyl
Concen: 79.951 ng
RT: 12.633 min Scan# 1657
Delta R.T. -0.012 min
Lab File: BM047257.D
Acq: 19 Aug 2024 18:08

Tgt Ion:172 Resp: 2074211
Ion Ratio Lower Upper
172 100
171 35.3 27.8 41.6
170 23.4 18.7 28.1









#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 16.786 min Scan# 21
Delta R.T. -0.006 min
Lab File: BM047257.D
Acq: 19 Aug 2024 18:08




Instrument :
BNA_M
ClientSampleId :
920-J-WS-081424


Tgt Ion	Ratio	Lower	Upper
188	100		
94	8.6	7.0	10.6
80	9.3	8.0	12.0



#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.027 min Scan# 3084
Delta R.T. -0.006 min
Lab File: BM047257.D
Acq: 19 Aug 2024 18:08

Tgt Ion	Ratio	Lower	Upper
240	100		
120	9.8	8.2	12.2
236	25.5	20.6	30.8









#79
Terphenyl-d14
Concen: 92.112 ng
RT: 19.456 min Scan# 21
Delta R.T. -0.006 min
Lab File: BM047257.D
Acq: 19 Aug 2024 18:08




Instrument :
BNA_M
ClientSampleId :
920-J-WS-081424

Tgt Ion:244 Resp: 4600226
Ion Ratio Lower Upper
244 100
212 7.5 6.2 9.4
122 9.2 7.1 10.7



#86
Perylene-d12
Concen: 20.000 ng
RT: 23.732 min Scan# 3544
Delta R.T. 0.006 min
Lab File: BM047257.D
Acq: 19 Aug 2024 18:08

Tgt Ion:264 Resp: 1179966
Ion Ratio Lower Upper
264 100
260 24.0 18.5 27.7
265 22.4 17.3 25.9





CALIBRATION SUMMARY

Method Path : Z:\svoasrv\HPCHEM1\BNA_M\Methods\
 Method File : 8270-BM081024.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Sun Aug 11 23:47:58 2024
 Response Via : Initial Calibration

Calibration Files

2.5 =BM047140.D 5 =BM047141.D 10 =BM047142.D 20 =BM047143.D 40 =BM047144.D 50 =BM047145.D 60 =BM047146.D 80 =BM047147.D

Compound		2.5	5	10	20	40	50	60	80	Avg	%RSD

1) I	1,4-Dichlorobenzen...	-----ISTD-----									
2)	1,4-Dioxane	0.519	0.490	0.495	0.438	0.436	0.442	0.426	0.464		7.88
3)	Pyridine	1.448	1.390	1.458	1.302	1.299	1.322	1.301	1.360		5.22
4)	n-Nitrosodimet...	0.643	0.612	0.629	0.570	0.564	0.583	0.558	0.594		5.71
5) S	2-Fluorophenol	1.215	1.184	1.197	1.071	1.050	1.071	1.046	1.119		6.73
6)	Aniline	1.546	1.516	1.583	1.428	1.401	1.395	1.334	1.458		6.27
7) S	Phenol-d6	1.643	1.599	1.652	1.478	1.470	1.494	1.462	1.542		5.52
8)	2-Chlorophenol	1.306	1.271	1.324	1.185	1.167	1.166	1.127	1.221		6.38
9)	Benzaldehyde	1.187	1.048	1.024	0.776	0.750	0.641		0.904		23.43
10) C	Phenol	1.695	1.600	1.645	1.453	1.450	1.477	1.453	1.539		6.80
11)	bis(2-Chloroet...	1.431	1.361	1.407	1.283	1.267	1.266	1.211	1.318		6.24
12)	1,3-Dichlorobe...	1.567	1.506	1.537	1.408	1.394	1.393	1.350	1.451		5.81
13) C	1,4-Dichlorobe...	1.581	1.532	1.562	1.429	1.405	1.412	1.360	1.469		5.95
14)	1,2-Dichlorobe...	1.526	1.468	1.514	1.314	1.308	1.309	1.261	1.386		8.10
15)	Benzyl Alcohol	1.083	1.074	1.189	1.098	1.076	1.101	1.063	1.098		3.88
16)	2,2'-oxybis(1-...	1.904	1.777	1.859	1.659	1.616	1.547	1.477	1.691		9.48
17)	2-Methylphenol	1.064	1.055	1.115	1.003	0.998	0.993	0.959	1.027		5.20
18)	Hexachloroethane	0.609	0.577	0.600	0.551	0.553	0.560	0.533	0.569		4.84
19) P	n-Nitroso-di-n...	1.087	1.169	1.095	1.138	0.982	0.965	0.959	0.916	1.039	9.12
20)	3+4-Methylphenols	1.462	1.442	1.570	1.411	1.405	1.405	1.352	1.435		4.77
21) I	Naphthalene-d8	-----ISTD-----									
22)	Acetophenone	0.547	0.517	0.510	0.462	0.450	0.452	0.441	0.483		8.55
23) S	Nitrobenzene-d5	0.423	0.417	0.420	0.390	0.378	0.382	0.372	0.397		5.53
24)	Nitrobenzene	0.425	0.427	0.428	0.393	0.380	0.378	0.369	0.400		6.44
25)	Isophorone	0.712	0.706	0.732	0.688	0.668	0.679	0.656	0.692		3.84
26) C	2-Nitrophenol	0.156	0.166	0.180	0.182	0.182	0.189	0.185	0.177		6.56
27)	2,4-Dimethylph...	0.200	0.209	0.217	0.209	0.206	0.209	0.204	0.207		2.64
28)	bis(2-Chloroet...	0.461	0.461	0.458	0.435	0.421	0.414	0.396	0.435		5.97
29) C	2,4-Dichloroph...	0.289	0.300	0.318	0.312	0.310	0.317	0.308	0.308		3.28
30)	1,2,4-Trichlor...	0.357	0.352	0.359	0.346	0.340	0.347	0.337	0.348		2.34
31)	Naphthalene	1.075	1.041	1.042	0.966	0.949	0.964	0.938	0.996		5.48
32)	Benzoic acid	0.182	0.212	0.240	0.251	0.262	0.273	0.268	0.241		13.75
33)	4-Chloroaniline	0.382	0.393	0.409	0.385	0.374	0.375	0.364	0.383		3.81
34) C	Hexachlorobuta...	0.199	0.200	0.207	0.203	0.203	0.211	0.206	0.204		2.03
35)	Caprolactam	0.098	0.103	0.112	0.109	0.107	0.110	0.109	0.107		4.45
36) C	4-Chloro-3-met...	0.320	0.324	0.347	0.328	0.323	0.330	0.324	0.328		2.73
37)	2-Methylnaphth...	0.741	0.731	0.742	0.689	0.681	0.698	0.685	0.709		3.85
38)	1-Methylnaphth...	0.737	0.726	0.735	0.681	0.668	0.690	0.671	0.701		4.34

Method Path : Z:\svoasrv\HPCHEM1\BNA_M\Methods\
 Method File : 8270-BM081024.M

39) I	Acenaphthene-d10	-----ISTD-----									
40)	1,2,4,5-Tetrac...	0.539	0.531	0.554	0.543	0.546	0.559	0.546	0.546	1.70	
41) P	Hexachlorocycl...	0.160	0.180	0.191	0.200	0.199	0.209	0.195	0.190	8.53	
42) S	2,4,6-Tribromo...	0.222	0.233	0.248	0.232	0.228	0.234	0.226	0.232	3.60	
43) C	2,4,6-Trichlor...	0.355	0.372	0.394	0.399	0.394	0.404	0.396	0.388	4.52	
44)	2,4,5-Trichlor...	0.410	0.410	0.438	0.437	0.436	0.447	0.435	0.430	3.37	
45) S	2-Fluorobiphenyl	1.406	1.341	1.336	1.265	1.247	1.261	1.220	1.297	5.09	
46)	1,1'-Biphenyl	1.538	1.482	1.479	1.388	1.361	1.387	1.332	1.424	5.32	
47)	2-Chloronaphth...	1.197	1.169	1.176	1.071	1.060	1.074	1.043	1.113	5.84	
48)	2-Nitroaniline	0.323	0.346	0.367	0.357	0.352	0.354	0.343	0.349	3.94	
49)	Acenaphthylene	1.719	1.699	1.732	1.602	1.575	1.602	1.554	1.640	4.50	
50)	Dimethylphthalate	1.472	1.433	1.468	1.379	1.343	1.365	1.318	1.397	4.38	
51)	2,6-Dinitrotol...	0.311	0.325	0.337	0.328	0.326	0.335	0.323	0.327	2.62	
52) C	Acenaphthene	1.115	1.079	1.086	1.008	0.991	1.017	0.988	1.041	4.94	
53)	3-Nitroaniline	0.310	0.317	0.344	0.331	0.328	0.337	0.325	0.327	3.53	
54) P	2,4-Dinitrophenol		0.152	0.185	0.204	0.204	0.218	0.212	0.196	12.27	
55)	Dibenzofuran	1.790	1.731	1.728	1.579	1.559	1.584	1.535	1.644	6.22	
56) P	4-Nitrophenol	0.247	0.268	0.293	0.293	0.291	0.296	0.287	0.282	6.42	
57)	2,4-Dinitrotol...	0.400	0.435	0.460	0.441	0.440	0.445	0.430	0.436	4.19	
58)	Fluorene	1.469	1.428	1.426	1.309	1.290	1.306	1.265	1.356	6.03	
59)	2,3,4,6-Tetrac...	0.333	0.345	0.369	0.355	0.354	0.366	0.355	0.354	3.44	
60)	Diethylphthalate	1.505	1.486	1.506	1.398	1.365	1.388	1.337	1.426	4.95	
61)	4-Chlorophenyl...	0.681	0.663	0.663	0.621	0.619	0.645	0.627	0.646	3.76	
62)	4-Nitroaniline	0.300	0.324	0.346	0.319	0.317	0.329	0.321	0.322	4.24	
63)	Azobenzene	1.455	1.445	1.457	1.314	1.280	1.285	1.232	1.352	7.14	
64) I	Phenanthrene-d10	-----ISTD-----									
65)	4,6-Dinitro-2-...	0.092	0.107	0.120	0.125	0.126	0.128	0.122	0.117	11.31	
66) c	n-Nitrosodiphe...	0.570	0.561	0.573	0.536	0.526	0.532	0.515	0.545	4.22	
67)	4-Bromophenyl-...	0.190	0.190	0.198	0.191	0.191	0.198	0.191	0.193	1.90	
68)	Hexachlorobenzene	0.236	0.234	0.240	0.229	0.226	0.229	0.221	0.231	2.73	
69)	Atrazine	0.183	0.186	0.182	0.170	0.152	0.145	0.129	0.164	13.50	
70) C	Pentachlorophenol	0.141	0.152	0.165	0.168	0.166	0.174	0.166	0.162	6.91	
71)	Phenanthrene	1.090	1.035	1.036	0.976	0.967	0.974	0.949	1.004	5.05	
72)	Anthracene	1.024	1.000	1.026	0.953	0.946	0.961	0.926	0.977	4.10	
73)	Carbazole	1.071	1.041	1.052	0.971	0.954	0.959	0.933	0.997	5.55	
74)	Di-n-butylphth...	1.175	1.187	1.223	1.165	1.134	1.140	1.098	1.160	3.50	
75) C	Fluoranthene	1.294	1.239	1.255	1.199	1.179	1.207	1.172	1.221	3.62	
76) I	Chrysene-d12	-----ISTD-----									
77)	Benzidine	0.226	0.190	0.373	0.489	0.370	0.389	0.337	0.339	29.96	
78)	Pyrene	1.456	1.398	1.443	1.425	1.414	1.446	1.414	1.428	1.46	
79) S	Terphenyl-d14	0.924	0.916	0.958	0.937	0.928	0.948	0.922	0.933	1.62	
80)	Butylbenzylpht...	0.511	0.552	0.600	0.608	0.599	0.616	0.595	0.583	6.48	
81)	Benzo(a)anthra...	1.390	1.331	1.364	1.300	1.291	1.324	1.293	1.328	2.85	
82)	3,3'-Dichlorob...	0.373	0.390	0.439	0.443	0.427	0.426	0.409	0.415	6.23	
83)	Chrysene	1.350	1.283	1.301	1.224	1.214	1.234	1.211	1.260	4.21	
84)	Bis(2-ethylhex...	0.771	0.808	0.867	0.845	0.832	0.846	0.816	0.826	3.81	
85) c	Di-n-octyl pht...	1.212	1.316	1.430	1.457	1.450	1.503	1.466	1.405	7.33	


Method Path : Z:\svoasrv\HPCHEM1\BNA_M\Methods\
Method File : 8270-BM081024.M

86) I	Perylene-d12	-----ISTD-----								
87)	Indeno(1,2,3-c...	1.326	1.280	1.366	1.279	1.278	1.227	1.294	1.293	3.38
88)	Benzo(b)fluora...	1.182	1.200	1.222	1.161	1.179	1.207	1.164	1.188	1.91
89)	Benzo(k)fluora...	1.192	1.134	1.175	1.097	1.059	1.116	1.080	1.122	4.35
90) C	Benzo(a)pyrene	1.014	1.015	1.066	1.024	1.012	1.046	1.025	1.029	1.93
91)	Dibenzo(a,h)an...	1.096	1.051	1.119	1.021	1.021	0.986	1.029	1.046	4.44
92)	Benzo(g,h,i)pe...	1.119	1.082	1.162	1.076	1.082	1.005	1.077	1.086	4.41

(#) = Out of Range

Benzaldehyde

Response Ratio



$R = -1.583e-001 A^2 + 1.112e+000 A + 2.780e-002$
Coef of Det (r^2) = 0.995802 Curve Fit: Quadratic
Method Name: Z:\svoasrv\HPCHEM1\BNA M\Methods\8270-BM081024.M
Calibration Table Last Updated: Sun Aug 11 23:47:58 2024

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081024\
 Data File : BM047140.D
 Acq On : 10 Aug 2024 13:12
 Operator : RC/JU
 Sample : SSTDICC2.5
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC2.5

Quant Time: Aug 11 23:30:47 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Aug 11 23:24:37 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.375	152	233543	20.000	ng	0.00
21) Naphthalene-d8	10.128	136	912085	20.000	ng	0.00
39) Acenaphthene-d10	14.034	164	589187	20.000	ng	0.00
64) Phenanthrene-d10	16.792	188	1231431	20.000	ng	0.00
76) Chrysene-d12	21.027	240	1169603	20.000	ng	0.00
86) Perylene-d12	23.727	264	1431343	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	0.000	112	0d	0.000	ng	
7) Phenol-d6	0.000	99	0d	0.000	ng	
23) Nitrobenzene-d5	0.000	82	0d	0.000	ng	
42) 2,4,6-Tribromophenol	0.000	330	0d	0.000	ng	
45) 2-Fluorobiphenyl	0.000	172	0d	0.000	ng	
79) Terphenyl-d14	0.000	244	0d	0.000	ng	
Target Compounds						
19) n-Nitroso-di-n-propyla...	8.181	70	31744	2.617	ng	Qvalue # 98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081024\
Data File : BM047140.D
Acq On : 10 Aug 2024 13:12
Operator : RC/JU
Sample : SSTDICC2.5
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTDICC2.5

Quant Time: Aug 11 23:30:47 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sun Aug 11 23:24:37 2024
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 20.000 ng

RT: 7.375 min Scan# 7

Delta R.T. 0.000 min

Lab File: BM047140.D

Acq: 10 Aug 2024 13:12

Instrument :

BNA_M

ClientSampleId :

SSTDICC2.5





Tgt Ion:152 Resp: 233543

Ion	Ratio	Lower	Upper

152	100		

150	154.9	120.5	180.7

115	57.3	45.5	68.3



#5

2-Fluorophenol

Concen: 0.000 ng

Expected RT: 5.02 min

Lab File: BM047140.D

Acq: 10 Aug 2024 13:12


Tgt Ion: 112


Sig	Exp Ratio

112	100

64	58.4

63	31.0





#7

Phenol-d6

Concen: 0.000 ng

Expected RT: 6.59 min

Lab File: BM047140.D

Acq: 10 Aug 2024 13:12

Instrument :

BNA_M

ClientSampleId :

SSTDICC2.5



Tgt Ion: 99

Sig Exp Ratio

99 100

42 17.6

71 35.6



#19

n-Nitroso-di-n-propylamine

Concen: 2.617 ng

RT: 8.181 min Scan# 900

Delta R.T. -0.006 min

Lab File: BM047140.D

Acq: 10 Aug 2024 13:12

Tgt Ion: 70 Resp: 31744




Ion Ratio Lower Upper


70 100

42 53.2 42.0 63.0

101 6.7 7.2 10.8

130 17.7 15.5 23.3





#21

Naphthalene-d8

Concen: 20.000 ng

RT: 10.128 min Scan# 1

Delta R.T. -0.006 min

Lab File: BM047140.D


Acq: 10 Aug 2024 13:12

Instrument :

BNA_M

ClientSampleId :

SSTDICC2.5



Tgt Ion:136 Resp: 912085




Ion Ratio Lower Upper

136 100

137 11.1 8.8 13.2

54 7.7 6.0 9.0

68 6.1 4.8 7.2



#23

Nitrobenzene-d5

Concen: 0.000 ng

Expected RT: 8.52 min

Lab File: BM047140.D

Acq: 10 Aug 2024 13:12


Tgt Ion: 82


Sig Exp Ratio

82 100

128 40.4

54 46.8





#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 14.034 min Scan# 1

Delta R.T. 0.000 min

Lab File: BM047140.D


Acq: 10 Aug 2024 13:12

Instrument :

BNA_M

ClientSampleId :

SSTDICC2.5





Tgt Ion:164 Resp: 589187

Ion Ratio Lower Upper

164 100

162 99.6 80.7 121.1

160 45.5 36.0 54.0



#42

2,4,6-Tribromophenol

Concen: 0.000 ng

Expected RT: 15.54 min

Lab File: BM047140.D

Acq: 10 Aug 2024 13:12


Tgt Ion: 330


Sig Exp Ratio

330 100

332 96.8

141 39.5







#45
2-Fluorobiphenyl
Concen: 0.000 ng
Expected RT: 12.65 min
Lab File: BM047140.D
Acq: 10 Aug 2024 13:12




Instrument :
BNA_M
ClientSampleId :
SSTDICC2.5


Tgt Ion: 172
Sig Exp Ratio
172 100
171 34.7
170 23.4



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 16.792 min Scan# 2364
Delta R.T. 0.000 min
Lab File: BM047140.D
Acq: 10 Aug 2024 13:12

Tgt Ion: 188 Resp: 1231431
Ion Ratio Lower Upper
188 100
94 9.0 7.0 10.6
80 10.4 8.0 12.0









#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.027 min Scan# 30
Delta R.T. -0.006 min
Lab File: BM047140.D
Acq: 10 Aug 2024 13:12

Instrument :
BNA_M
ClientSampleId :
SSTDICC2.5


Tgt Ion:240 Resp: 1169603
Ion Ratio Lower Upper
240 100
120 11.0 8.2 12.2
236 26.1 20.6 30.8



#79
Terphenyl-d14
Concen: 0.000 ng
Expected RT: 19.46 min


Lab File: BM047140.D
Acq: 10 Aug 2024 13:12

Tgt Ion: 244
Sig Exp Ratio
244 100
212 7.8
122 8.9



Ion 212.00 (211.70 to 212.70): BM047140.D\data.ms

Ion 122.00 (121.70 to 122.70): BM047140.D\data.ms



#86

Perylene-d12

Concen: 20.000 ng

RT: 23.727 min Scan# 3

Delta R.T. 0.000 min

Lab File: BM047140.D


Acq: 10 Aug 2024 13:12

Instrument :

BNA_M

ClientSampleId :

SSTDICC2.5





Tgt Ion:264 Resp: 1431343

Ion Ratio Lower Upper

264 100

260 22.7 18.5 27.7

265 21.7 17.3 25.9



Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081024\
 Data File : BM047141.D
 Acq On : 10 Aug 2024 13:52
 Operator : RC/JU
 Sample : SSTDICC005
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC005

Quant Time: Aug 11 23:31:31 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Aug 11 23:24:37 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.375	152	231587	20.000	ng	0.00
21) Naphthalene-d8	10.128	136	953129	20.000	ng	0.00
39) Acenaphthene-d10	14.027	164	629995	20.000	ng	0.00
64) Phenanthrene-d10	16.792	188	1336137	20.000	ng	0.00
76) Chrysene-d12	21.027	240	1301329	20.000	ng	0.00
86) Perylene-d12	23.727	264	1576050	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.022	112	140661	10.853	ng	0.00
7) Phenol-d6	6.581	99	190264	10.654	ng	0.00
23) Nitrobenzene-d5	8.516	82	201738	10.658	ng	0.00
42) 2,4,6-Tribromophenol	15.533	330	70075	9.593	ng	0.00
45) 2-Fluorobiphenyl	12.639	172	442999	10.847	ng	0.00
79) Terphenyl-d14	19.456	244	601320	9.901	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.040	88	30031	5.592	ng	99
3) Pyridine	3.416	79	83824	5.327	ng	96
4) n-Nitrosodimethylamine	3.334	42	37222	5.410	ng	# 94
6) Aniline	6.722	93	89509	5.303	ng	99
8) 2-Chlorophenol	6.945	128	75622	5.350	ng	95
9) Benzaldehyde	6.540	77	68748	5.019	ng	96
10) Phenol	6.604	94	98125	5.506	ng	99
11) bis(2-Chloroethyl)ether	6.822	93	82870	5.430	ng	95
12) 1,3-Dichlorobenzene	7.263	146	90719	5.401	ng	97
13) 1,4-Dichlorobenzene	7.410	146	91512	5.381	ng	98
14) 1,2-Dichlorobenzene	7.716	146	88355	5.506	ng	96
15) Benzyl Alcohol	7.622	79	62692	4.932	ng	98
16) 2,2'-oxybis(1-Chloropr...	7.904	45	110215	5.622	ng	98
17) 2-Methylphenol	7.834	107	61615	5.183	ng	95
18) Hexachloroethane	8.428	117	35243	5.350	ng	89
19) n-Nitroso-di-n-propyla...	8.181	70	67704	5.628	ng	100
20) 3+4-Methylphenols	8.157	107	84662	5.094	ng	99
22) Acetophenone	8.192	105	130329	5.665	ng	98
24) Nitrobenzene	8.557	77	101289	5.314	ng	97
25) Isophorone	9.086	82	169727	5.150	ng	100
26) 2-Nitrophenol	9.263	139	37222	4.410	ng	# 88
27) 2,4-Dimethylphenol	9.339	122	47540	4.808	ng	99
28) bis(2-Chloroethoxy)met...	9.575	93	109875	5.298	ng	99
29) 2,4-Dichlorophenol	9.798	162	68962	4.700	ng	99
30) 1,2,4-Trichlorobenzene	9.998	180	85165	5.129	ng	96
31) Naphthalene	10.181	128	256168	5.395	ng	99
32) Benzoic acid	9.439	122	43353m	3.753	ng	
33) 4-Chloroaniline	10.304	127	90992	4.983	ng	97
34) Hexachlorobutadiene	10.469	225	47407	4.878	ng	97
35) Caprolactam	11.086	113	23316	4.576	ng	95
36) 4-Chloro-3-methylphenol	11.451	107	76286	4.882	ng	95
37) 2-Methylnaphthalene	11.804	142	176480	5.220	ng	99
38) 1-Methylnaphthalene	12.027	142	175570	5.255	ng	97
40) 1,2,4,5-Tetrachloroben...	12.186	216	84909	4.941	ng	99
41) Hexachlorocyclopentadiene	12.163	237	25169	4.195	ng	97
43) 2,4,6-Trichlorophenol	12.445	196	55962	4.582	ng	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081024\
 Data File : BM047141.D
 Acq On : 10 Aug 2024 13:52
 Operator : RC/JU
 Sample : SSTDICC005
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC005


Quant Time: Aug 11 23:31:31 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Aug 11 23:24:37 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	12.522	196	64621	4.765	ng	97
46) 1,1'-Biphenyl	12.851	154	242201	5.400	ng	98
47) 2-Chloronaphthalene	12.886	162	188590	5.380	ng	98
48) 2-Nitroaniline	13.110	65	50873	4.632	ng	91
49) Acenaphthylene	13.745	152	270677	5.238	ng	99
50) Dimethylphthalate	13.504	163	231786	5.268	ng	100
51) 2,6-Dinitrotoluene	13.621	165	49002	4.764	ng	96
52) Acenaphthene	14.092	154	175534	5.355	ng	99
53) 3-Nitroaniline	13.951	138	48787	4.734	ng	99
55) Dibenzofuran	14.433	168	281866	5.444	ng	99
56) 4-Nitrophenol	14.292	139	38873	4.374	ng	93
57) 2,4-Dinitrotoluene	14.421	165	63007	4.589	ng	88
58) Fluorene	15.092	166	231431	5.418	ng	100
59) 2,3,4,6-Tetrachlorophenol	14.674	232	52488	4.710	ng	94
60) Diethylphthalate	14.892	149	236973	5.274	ng	99
61) 4-Chlorophenyl-phenyle...	15.098	204	107266	5.274	ng	98
62) 4-Nitroaniline	15.127	138	47275m	4.656	ng	
63) Azobenzene	15.392	77	229119	5.378	ng	99
65) 4,6-Dinitro-2-methylph...	15.192	198	30637	3.916	ng	93
66) n-Nitrosodiphenylamine	15.315	169	190365	5.231	ng	99
67) 4-Bromophenyl-phenylether	15.992	248	63559	4.937	ng	94
68) Hexachlorobenzene	16.098	284	78852	5.114	ng	98
69) Atrazine	16.286	200	60979	5.568	ng	98
70) Pentachlorophenol	16.451	266	47145	4.362	ng	97
71) Phenanthrene	16.833	178	363970	5.428	ng	99
72) Anthracene	16.921	178	342081	5.243	ng	99
73) Carbazole	17.204	167	357631	5.368	ng	99
74) Di-n-butylphthalate	17.792	149	392569	5.064	ng	98
75) Fluoranthene	18.868	202	432338	5.301	ng	98
77) Benzidine	19.074	184	73589	3.335	ng	96
78) Pyrene	19.233	202	473702	5.098	ng	99
80) Butylbenzylphthalate	20.174	149	166134	4.381	ng	97
81) Benzo(a)anthracene	21.009	228	452360	5.236	ng	98
82) 3,3'-Dichlorobenzidine	20.950	252	121454	4.493	ng	100
83) Chrysene	21.062	228	439298	5.360	ng	98
84) Bis(2-ethylhexyl)phtha...	20.974	149	250857	4.665	ng	99
85) Di-n-octyl phthalate	22.003	149	394399	4.315	ng	97
87) Indeno(1,2,3-cd)pyrene	26.715	276	522557	5.130	ng	99
88) Benzo(b)fluoranthene	22.880	252	465640	4.975	ng	99
89) Benzo(k)fluoranthene	22.933	252	469757	5.314	ng	99
90) Benzo(a)pyrene	23.597	252	399615	4.929	ng	98
91) Dibenzo(a,h)anthracene	26.762	278	431713	5.237	ng	98
92) Benzo(g,h,i)perylene	27.632	276	440920	5.151	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
BNA_M
ClientSampleId :
SSTDICC005





[illegible]



#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.375 min Scan# 70
Delta R.T. -0.000 min
Lab File: BM047141.D
Acq: 10 Aug 2024 13:52




Instrument :
BNA_M
ClientSampleId :
SSTDICC005


Tgt Ion:152 Resp: 231587
Ion Ratio Lower Upper
152 100
150 155.8 120.5 180.7
115 59.1 45.5 68.3



#2
1,4-Dioxane
Concen: 5.592 ng
RT: 3.040 min Scan# 26
Delta R.T. 0.006 min
Lab File: BM047141.D
Acq: 10 Aug 2024 13:52

Tgt Ion: 88 Resp: 30031
Ion Ratio Lower Upper
88 100
58 73.4 58.2 87.4
43 28.6 22.5 33.7









#3
Pyridine
Concen: 5.327 ng
RT: 3.416 min Scan# 90
Delta R.T. 0.006 min
Lab File: BM047141.D
Acq: 10 Aug 2024 13:52




Instrument :
BNA_M
ClientSampleId :
SSTDICC005


Tgt Ion: 79 Resp: 83824
Ion Ratio Lower Upper
79 100
52 60.5 51.2 76.8
51 34.9 26.1 39.1



#4
n-Nitrosodimethylamine
Concen: 5.410 ng
RT: 3.334 min Scan# 76
Delta R.T. 0.006 min
Lab File: BM047141.D
Acq: 10 Aug 2024 13:52

Tgt Ion: 42 Resp: 37222
Ion Ratio Lower Upper
42 100
74 134.9 112.9 169.3
44 17.0 5.3 7.9









#5
2-Fluorophenol
Concen: 10.853 ng
RT: 5.022 min Scan# 363
Delta R.T. -0.000 min
Lab File: BM047141.D
Acq: 10 Aug 2024 13:52




Instrument :
BNA_M
ClientSampleId :
SSTDICC005


Tgt Ion:112 Resp: 140661
Ion Ratio Lower Upper
112 100
64 59.1 46.7 70.1
63 31.5 24.8 37.2



#6
Aniline
Concen: 5.303 ng
RT: 6.722 min Scan# 652
Delta R.T. -0.006 min
Lab File: BM047141.D
Acq: 10 Aug 2024 13:52

Tgt Ion: 93 Resp: 89509
Ion Ratio Lower Upper
93 100
66 38.6 31.4 47.2
65 18.5 15.6 23.4









#7
Phenol-d6
Concen: 10.654 ng
RT: 6.581 min Scan# 61
Delta R.T. -0.006 min
Lab File: BM047141.D
Acq: 10 Aug 2024 13:52




Instrument :
BNA_M
ClientSampleId :
SSTDICC005


Tgt Ion: 99 Resp: 190264
Ion Ratio Lower Upper
99 100
42 19.1 14.1 21.1
71 34.9 28.5 42.7



#8
2-Chlorophenol
Concen: 5.350 ng
RT: 6.945 min Scan# 690
Delta R.T. -0.006 min
Lab File: BM047141.D
Acq: 10 Aug 2024 13:52

Tgt Ion: 128 Resp: 75622
Ion Ratio Lower Upper
128 100
130 31.8 12.1 52.1
64 52.1 26.6 66.6









#9
Benzaldehyde
Concen: 5.019 ng
RT: 6.540 min Scan# 61
Delta R.T. -0.000 min
Lab File: BM047141.D
Acq: 10 Aug 2024 13:52




Instrument :
BNA_M
ClientSampleId :
SSTDICC005


Tgt Ion: 77 Resp: 68748
Ion Ratio Lower Upper
77 100
105 91.4 74.5 114.5
106 87.4 72.3 112.3



#10
Phenol
Concen: 5.506 ng
RT: 6.604 min Scan# 632
Delta R.T. -0.006 min
Lab File: BM047141.D
Acq: 10 Aug 2024 13:52

Tgt Ion: 94 Resp: 98125
Ion Ratio Lower Upper
94 100
65 28.1 9.0 49.0
66 38.3 18.9 58.9









#11
bis(2-Chloroethyl)ether
Concen: 5.430 ng
RT: 6.822 min Scan# 669
Delta R.T. -0.006 min
Lab File: BM047141.D
Acq: 10 Aug 2024 13:52




Instrument :
BNA_M
ClientSampleId :
SSTDICC005


Tgt Ion: 93 Resp: 82870
Ion Ratio Lower Upper
93 100
63 80.3 55.8 95.8
95 30.0 12.7 52.7



#12
1,3-Dichlorobenzene
Concen: 5.401 ng
RT: 7.263 min Scan# 744
Delta R.T. -0.000 min
Lab File: BM047141.D
Acq: 10 Aug 2024 13:52

Tgt Ion: 146 Resp: 90719
Ion Ratio Lower Upper
146 100
148 62.9 52.3 78.5
75 33.0 25.8 38.6





#13

1,4-Dichlorobenzene

Concen: 5.381 ng

RT: 7.410 min Scan# 70

Delta R.T. -0.000 min

Lab File: BM047141.D

Acq: 10 Aug 2024 13:52

Instrument :

BNA_M

ClientSampleId :

SSTDICC005





Tgt Ion:146 Resp: 91512

Ion Ratio Lower Upper

146 100

148 62.3 51.4 77.2

111 38.3 31.5 47.3



#14

1,2-Dichlorobenzene

Concen: 5.506 ng

RT: 7.716 min Scan# 821

Delta R.T. -0.000 min

Lab File: BM047141.D

Acq: 10 Aug 2024 13:52




Tgt Ion:146 Resp: 88355


Ion Ratio Lower Upper

146 100

148 67.9 51.8 77.8

111 43.7 33.3 49.9









#15
Benzyl Alcohol
Concen: 4.932 ng
RT: 7.622 min Scan# 805
Delta R.T. -0.006 min
Lab File: BM047141.D
Acq: 10 Aug 2024 13:52




Instrument :
BNA_M
ClientSampleId :
SSTDICC005


Tgt Ion: 79 Resp: 62692
Ion Ratio Lower Upper
79 100
108 73.7 59.7 89.5
77 67.9 52.9 79.3



#16
2,2'-oxybis(1-chloropropane)
Concen: 5.622 ng
RT: 7.904 min Scan# 853
Delta R.T. -0.012 min
Lab File: BM047141.D
Acq: 10 Aug 2024 13:52

Tgt Ion: 45 Resp: 110215
Ion Ratio Lower Upper
45 100
77 16.3 0.0 35.6
79 11.6 0.0 32.4





#17

2-Methylphenol

Concen: 5.183 ng

RT: 7.834 min Scan# 841

Delta R.T. -0.000 min

Lab File: BM047141.D

Acq: 10 Aug 2024 13:52

Instrument :

BNA_M

ClientSampleId :

SSTDICC005

Tgt Ion:107 Resp: 61615





Ion Ratio Lower Upper

107 100

108 120.5 90.6 135.8

77 51.6 39.3 58.9

79 49.0 38.6 58.0



#18

Hexachloroethane

Concen: 5.350 ng

RT: 8.428 min Scan# 942

Delta R.T. -0.000 min

Lab File: BM047141.D

Acq: 10 Aug 2024 13:52




Tgt Ion:117 Resp: 35243


Ion Ratio Lower Upper

117 100

119 88.0 77.0 115.4

201 88.6 81.3 121.9









#19
n-Nitroso-di-n-propylamine
Concen: 5.628 ng
RT: 8.181 min Scan# 900
Delta R.T. -0.006 min
Lab File: BM047141.D
Acq: 10 Aug 2024 13:52




Instrument :
BNA_M
ClientSampleId :
SSTDICC005


Tgt Ion: 70	Resp: 67704
Ion Ratio	Lower Upper
70	100
42	52.7 42.0 63.0
101	9.4 7.2 10.8
130	19.1 15.5 23.3



#20
3+4-Methylphenols
Concen: 5.094 ng
RT: 8.157 min Scan# 896
Delta R.T. -0.006 min
Lab File: BM047141.D
Acq: 10 Aug 2024 13:52

Tgt Ion: 107	Resp: 84662
Ion Ratio	Lower Upper
107	100
108	90.4 70.7 110.7
77	40.6 17.4 57.4
79	30.9 10.9 50.9









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.128 min Scan# 1231
Delta R.T. -0.006 min
Lab File: BM047141.D
Acq: 10 Aug 2024 13:52




Instrument :
BNA_M
ClientSampleId :
SSTDICC005


Tgt Ion:136 Resp: 953129
Ion Ratio Lower Upper
136 100
137 10.9 8.8 13.2
54 7.7 6.0 9.0
68 6.0 4.8 7.2



#22
Acetophenone
Concen: 5.665 ng
RT: 8.192 min Scan# 902
Delta R.T. -0.006 min
Lab File: BM047141.D
Acq: 10 Aug 2024 13:52

Tgt Ion:105 Resp: 130329
Ion Ratio Lower Upper
105 100
71 4.6 4.5 6.7
51 26.0 20.1 30.1
120 21.0 17.8 26.6





#23

Nitrobenzene-d5

Concen: 10.658 ng

RT: 8.516 min Scan# 91

Delta R.T. -0.006 min

Lab File: BM047141.D

Acq: 10 Aug 2024 13:52

Instrument :

BNA_M

ClientSampleId :

SSTDICC005





Tgt Ion: 82 Resp: 201738

Ion Ratio Lower Upper

82 100

128 37.9 32.3 48.5

54 46.5 37.4 56.2



#24

Nitrobenzene

Concen: 5.314 ng

RT: 8.557 min Scan# 964

Delta R.T. -0.006 min

Lab File: BM047141.D

Acq: 10 Aug 2024 13:52




Tgt Ion: 77 Resp: 101289


Ion Ratio Lower Upper

77 100

123 40.2 34.0 51.0

65 13.0 10.5 15.7





#25

Isophorone

Concen: 5.150 ng

RT: 9.086 min Scan# 1054

Delta R.T. -0.006 min

Lab File: BM047141.D

Acq: 10 Aug 2024 13:52

Instrument :

BNA_M

ClientSampleId :

SSTDICC005





Tgt Ion: 82 Resp: 169727

Ion Ratio Lower Upper

82 100

95 6.8 5.5 8.3

138 17.0 13.5 20.3



#26

2-Nitrophenol

Concen: 4.410 ng

RT: 9.263 min Scan# 1084

Delta R.T. -0.000 min

Lab File: BM047141.D

Acq: 10 Aug 2024 13:52




Tgt Ion:139 Resp: 37222


Ion Ratio Lower Upper

139 100

109 32.1 24.4 36.6

65 65.8 43.0 64.6#





#27

2,4-Dimethylphenol

Concen: 4.808 ng

RT: 9.339 min Scan# 1098

Delta R.T. -0.006 min

Lab File: BM047141.D

Acq: 10 Aug 2024 13:52

Instrument :

BNA_M

ClientSampleId :

SSTDICC005





Tgt Ion:122 Resp: 47540

Ion Ratio Lower Upper

122 100

107 116.9 92.8 139.2

121 57.0 46.1 69.1



#28

bis(2-Chloroethoxy)methane

Concen: 5.298 ng

RT: 9.575 min Scan# 1137

Delta R.T. -0.000 min

Lab File: BM047141.D

Acq: 10 Aug 2024 13:52




Tgt Ion: 93 Resp: 109875


Ion Ratio Lower Upper

93 100

95 32.1 25.8 38.8

123 10.5 9.0 13.4





#29

2,4-Dichlorophenol

Concen: 4.700 ng

RT: 9.798 min Scan# 1175

Delta R.T. -0.000 min

Lab File: BM047141.D

Acq: 10 Aug 2024 13:52

Instrument :

BNA_M

ClientSampleId :

SSTDICC005





Tgt Ion:162 Resp: 68962

Ion Ratio Lower Upper

162 100

164 65.6 44.7 84.7

98 35.9 14.9 54.9



#30

1,2,4-Trichlorobenzene

Concen: 5.129 ng

RT: 9.998 min Scan# 1209

Delta R.T. -0.000 min

Lab File: BM047141.D

Acq: 10 Aug 2024 13:52




Tgt Ion:180 Resp: 85165


Ion Ratio Lower Upper

180 100

182 91.0 76.6 115.0

145 28.4 23.4 35.2





#31

Naphthalene

Concen: 5.395 ng

RT: 10.181 min Scan# 1114

Delta R.T. -0.000 min

Lab File: BM047141.D

Acq: 10 Aug 2024 13:52

Instrument :

BNA_M

ClientSampleId :

SSTDICC005





Tgt Ion:128 Resp: 256168

Ion Ratio Lower Upper

128 100

129 11.1 8.6 13.0

127 13.2 10.2 15.4



#32

Benzoic acid

Concen: 3.753 ng m

RT: 9.439 min Scan# 1114

Delta R.T. -0.094 min

Lab File: BM047141.D

Acq: 10 Aug 2024 13:52




Tgt Ion:122 Resp: 43353


Ion Ratio Lower Upper

122 100

105 127.0 107.1 147.1

77 99.5 82.1 122.1









#33
4-Chloroaniline
Concen: 4.983 ng
RT: 10.304 min Scan# 1261
Delta R.T. -0.006 min
Lab File: BM047141.D
Acq: 10 Aug 2024 13:52




Instrument :
BNA_M
ClientSampleId :
SSTDICC005


Tgt Ion:	127	Resp:	90992
Ion	Ratio	Lower	Upper
127	100		
129	32.0	25.8	38.8
65	34.1	24.9	37.3
92	19.8	15.1	22.7



#34
Hexachlorobutadiene
Concen: 4.878 ng
RT: 10.469 min Scan# 1289
Delta R.T. -0.000 min
Lab File: BM047141.D
Acq: 10 Aug 2024 13:52

Tgt Ion:	225	Resp:	47407
Ion	Ratio	Lower	Upper
225	100		
223	65.1	50.2	75.2
227	65.9	51.4	77.0









#35
Caprolactam
Concen: 4.576 ng
RT: 11.086 min Scan# 113
Delta R.T. -0.035 min
Lab File: BM047141.D
Acq: 10 Aug 2024 13:52




Instrument :
BNA_M
ClientSampleId :
SSTDICC005


Tgt Ion	Ratio	Lower	Upper
113	100		
55	166.8	140.5	180.5
56	130.0	103.6	143.6



#36
4-Chloro-3-methylphenol
Concen: 4.882 ng
RT: 11.451 min Scan# 1456
Delta R.T. -0.000 min
Lab File: BM047141.D
Acq: 10 Aug 2024 13:52

Tgt Ion	Ratio	Lower	Upper
107	100		
144	27.6	21.8	32.6
142	77.3	66.8	100.2









#37
2-Methylnaphthalene
Concen: 5.220 ng
RT: 11.804 min Scan# 1516
Delta R.T. -0.000 min
Lab File: BM047141.D
Acq: 10 Aug 2024 13:52




Instrument :
BNA_M
ClientSampleId :
SSTDICC005


Tgt Ion	Ratio	Lower	Upper
142	100		
141	87.2	70.9	106.3
115	36.1	28.7	43.1



#38
1-Methylnaphthalene
Concen: 5.255 ng
RT: 12.027 min Scan# 1554
Delta R.T. -0.006 min
Lab File: BM047141.D
Acq: 10 Aug 2024 13:52

Tgt Ion	Ratio	Lower	Upper
142	100		
141	90.6	74.6	111.8
116	4.0	3.1	4.7





#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 14.027 min Scan# 1894

Delta R.T. -0.006 min

Lab File: BM047141.D

Acq: 10 Aug 2024 13:52

Instrument :

BNA_M

ClientSampleId :

SSTDICC005





Tgt Ion:164 Resp: 629995

Ion Ratio Lower Upper

164 100

162 101.0 80.7 121.1

160 46.1 36.0 54.0



#40

1,2,4,5-Tetrachlorobenzene

Concen: 4.941 ng

RT: 12.186 min Scan# 1581

Delta R.T. -0.000 min

Lab File: BM047141.D

Acq: 10 Aug 2024 13:52

Tgt Ion:216 Resp: 84909




Ion Ratio Lower Upper


216 100

214 79.0 63.4 95.2

179 22.6 17.6 26.4

108 20.7 15.1 22.7









#41
Hexachlorocyclopentadiene
Concen: 4.195 ng
RT: 12.163 min Scan# 11
Delta R.T. -0.000 min
Lab File: BM047141.D
Acq: 10 Aug 2024 13:52




Instrument :
BNA_M
ClientSampleId :
SSTDICC005


Tgt Ion	Ratio	Lower	Upper
237	100		
235	62.0	44.8	84.8
272	13.3	0.0	33.5



#42
2,4,6-Tribromophenol
Concen: 9.593 ng
RT: 15.533 min Scan# 2150
Delta R.T. -0.006 min
Lab File: BM047141.D
Acq: 10 Aug 2024 13:52

Tgt Ion	Ratio	Lower	Upper
330	100		
332	98.9	77.4	116.2
141	42.2	31.6	47.4









#43
2,4,6-Trichlorophenol
Concen: 4.582 ng
RT: 12.445 min Scan# 1625
Delta R.T. -0.000 min
Lab File: BM047141.D
Acq: 10 Aug 2024 13:52




Instrument :
BNA_M
ClientSampleId :
SSTDICC005


Tgt Ion	196	Resp	55962
Ion Ratio	Lower	Upper	
196	100		
198	94.4	75.6	113.4
200	29.9	24.6	37.0



#44
2,4,5-Trichlorophenol
Concen: 4.765 ng
RT: 12.522 min Scan# 1638
Delta R.T. -0.000 min
Lab File: BM047141.D
Acq: 10 Aug 2024 13:52

Tgt Ion	196	Resp	64621
Ion Ratio	Lower	Upper	
196	100		
198	94.4	77.7	116.5
97	52.8	39.8	59.8
132	26.7	20.7	31.1









#45
2-Fluorobiphenyl
Concen: 10.847 ng
RT: 12.639 min Scan# 1658
Delta R.T. -0.006 min
Lab File: BM047141.D
Acq: 10 Aug 2024 13:52




Instrument :
BNA_M
ClientSampleId :
SSTDICC005


Tgt Ion:172 Resp: 442999
Ion Ratio Lower Upper
172 100
171 34.4 27.8 41.6
170 23.1 18.7 28.1



#46
1,1'-Biphenyl
Concen: 5.400 ng
RT: 12.851 min Scan# 1694
Delta R.T. -0.006 min
Lab File: BM047141.D
Acq: 10 Aug 2024 13:52

Tgt Ion:154 Resp: 242201
Ion Ratio Lower Upper
154 100
153 40.2 19.0 59.0
76 12.8 0.0 31.9





#47

2-Chloronaphthalene

Concen: 5.380 ng

RT: 12.886 min Scan# 1

Delta R.T. -0.000 min

Lab File: BM047141.D

Acq: 10 Aug 2024 13:52

Instrument :

BNA_M

ClientSampleId :

SSTDICC005





Tgt Ion:162 Resp: 188590

Ion Ratio Lower Upper

162 100

127 36.8 30.0 45.0

164 34.2 26.3 39.5



#48

2-Nitroaniline

Concen: 4.632 ng

RT: 13.110 min Scan# 1738

Delta R.T. -0.006 min

Lab File: BM047141.D

Acq: 10 Aug 2024 13:52




Tgt Ion: 65 Resp: 50873


Ion Ratio Lower Upper

65 100

92 64.6 55.1 82.7

138 96.3 85.9 128.9









#49
Acenaphthylene
Concen: 5.238 ng
RT: 13.745 min Scan# 1846
Delta R.T. -0.006 min
Lab File: BM047141.D
Acq: 10 Aug 2024 13:52




Instrument :
BNA_M
ClientSampleId :
SSTDICC005


Tgt Ion:152 Resp: 270677
Ion Ratio Lower Upper
152 100
151 20.0 15.8 23.6
153 12.6 10.4 15.6



#50
Dimethylphthalate
Concen: 5.268 ng
RT: 13.504 min Scan# 1805
Delta R.T. -0.012 min
Lab File: BM047141.D
Acq: 10 Aug 2024 13:52

Tgt Ion:163 Resp: 231786
Ion Ratio Lower Upper
163 100
194 3.8 3.0 4.6
164 10.2 8.0 12.0





#51

2,6-Dinitrotoluene

Concen: 4.764 ng

RT: 13.621 min Scan# 1826

Delta R.T. -0.006 min

Lab File: BM047141.D

Acq: 10 Aug 2024 13:52

Instrument :

BNA_M

ClientSampleId :

SSTDICC005





Tgt Ion:165 Resp: 49002

Ion Ratio Lower Upper

165 100

63 60.6 46.5 69.7

89 63.1 47.3 70.9



#52

Acenaphthene

Concen: 5.355 ng

RT: 14.092 min Scan# 1905

Delta R.T. -0.006 min

Lab File: BM047141.D

Acq: 10 Aug 2024 13:52




Tgt Ion:154 Resp: 175534


Ion Ratio Lower Upper

154 100

153 113.5 90.3 135.5

152 54.6 42.9 64.3









#53
3-Nitroaniline
Concen: 4.734 ng
RT: 13.951 min Scan# 1881
Delta R.T. -0.012 min
Lab File: BM047141.D
Acq: 10 Aug 2024 13:52




Instrument :
BNA_M
ClientSampleId :
SSTDICC005


Tgt Ion:138 Resp: 48787
Ion Ratio Lower Upper
138 100
108 11.8 7.9 11.9
92 114.2 92.0 138.0



#55
Dibenzofuran
Concen: 5.444 ng
RT: 14.433 min Scan# 1963
Delta R.T. -0.006 min
Lab File: BM047141.D
Acq: 10 Aug 2024 13:52

Tgt Ion:168 Resp: 281866
Ion Ratio Lower Upper
168 100
139 37.6 29.6 44.4
169 13.3 10.7 16.1









#56
4-Nitrophenol
Concen: 4.374 ng
RT: 14.292 min Scan# 1940
Delta R.T. -0.006 min
Lab File: BM047141.D
Acq: 10 Aug 2024 13:52




Instrument :
BNA_M
ClientSampleId :
SSTDICC005


Tgt Ion:139 Resp: 38873
Ion Ratio Lower Upper
139 100
109 75.4 52.5 92.5
65 104.4 75.4 115.4



#57
2,4-Dinitrotoluene
Concen: 4.589 ng
RT: 14.421 min Scan# 1961
Delta R.T. -0.006 min
Lab File: BM047141.D
Acq: 10 Aug 2024 13:52

Tgt Ion:165 Resp: 63007
Ion Ratio Lower Upper
165 100
63 58.9 40.3 60.5
89 87.0 61.0 91.6
182 2.3 1.9 2.9









#58
Fluorene
Concen: 5.418 ng
RT: 15.092 min Scan# 2075
Delta R.T. -0.000 min
Lab File: BM047141.D
Acq: 10 Aug 2024 13:52




Instrument :
BNA_M
ClientSampleId :
SSTDICC005


Tgt Ion:166 Resp: 231431
Ion Ratio Lower Upper
166 100
165 96.9 77.4 116.0
167 13.3 11.0 16.6



#59
2,3,4,6-Tetrachlorophenol
Concen: 4.710 ng
RT: 14.674 min Scan# 2004
Delta R.T. -0.000 min
Lab File: BM047141.D
Acq: 10 Aug 2024 13:52

Tgt Ion:232 Resp: 52488
Ion Ratio Lower Upper
232 100
131 50.7 36.9 55.3
130 2.6 2.0 3.0
166 35.3 26.4 39.6









#60
Diethylphthalate
Concen: 5.274 ng
RT: 14.892 min Scan# 2041
Delta R.T. -0.006 min
Lab File: BM047141.D
Acq: 10 Aug 2024 13:52




Instrument :
BNA_M
ClientSampleId :
SSTDICC005


Tgt Ion:149 Resp: 236973
Ion Ratio Lower Upper
149 100
177 21.8 17.6 26.4
150 11.6 9.8 14.6



#61
4-Chlorophenyl-phenylether
Concen: 5.274 ng
RT: 15.098 min Scan# 2076
Delta R.T. -0.000 min
Lab File: BM047141.D
Acq: 10 Aug 2024 13:52

Tgt Ion:204 Resp: 107266
Ion Ratio Lower Upper
204 100
206 31.7 26.2 39.4
141 60.0 47.2 70.8





#62

4-Nitroaniline

Concen: 4.656 ng m

RT: 15.127 min Scan# 2083

Delta R.T. -0.012 min

Lab File: BM047141.D

Acq: 10 Aug 2024 13:52

Instrument :

BNA_M

ClientSampleId :

SSTDICC005





Tgt Ion:138 Resp: 47275

Ion Ratio Lower Upper

138 100

92 56.0 31.9 71.9

108 86.6 63.8 103.8



#63

Azobenzene

Concen: 5.378 ng

RT: 15.392 min Scan# 2126

Delta R.T. -0.000 min

Lab File: BM047141.D

Acq: 10 Aug 2024 13:52

Tgt Ion: 77 Resp: 229119




Ion Ratio Lower Upper


77 100

182 22.5 3.8 43.8

105 19.4 0.0 39.5

51 27.2 6.8 46.8









#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 16.792 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM047141.D
Acq: 10 Aug 2024 13:52




Instrument :
BNA_M
ClientSampleId :
SSTDICC005


Tgt Ion:188 Resp: 1336137
Ion Ratio Lower Upper
188 100
94 8.9 7.0 10.6
80 9.9 8.0 12.0



#65
4,6-Dinitro-2-methylphenol
Concen: 3.916 ng
RT: 15.192 min Scan# 2092
Delta R.T. -0.006 min
Lab File: BM047141.D
Acq: 10 Aug 2024 13:52

Tgt Ion:198 Resp: 30637
Ion Ratio Lower Upper
198 100
51 48.3 23.9 63.9
105 48.9 24.3 64.3









#66
n-Nitrosodiphenylamine
Concen: 5.231 ng
RT: 15.315 min Scan# 2113
Delta R.T. -0.006 min
Lab File: BM047141.D
Acq: 10 Aug 2024 13:52




Instrument :
BNA_M
ClientSampleId :
SSTDICC005


Tgt Ion:169 Resp: 190365
Ion Ratio Lower Upper
169 100
168 67.8 53.8 80.8
167 38.0 29.3 43.9



#67
4-Bromophenyl-phenylether
Concen: 4.937 ng
RT: 15.992 min Scan# 2228
Delta R.T. -0.006 min
Lab File: BM047141.D
Acq: 10 Aug 2024 13:52

Tgt Ion:248 Resp: 63559
Ion Ratio Lower Upper
248 100
250 93.5 77.8 116.8
141 80.4 57.7 86.5









#68
Hexachlorobenzene
Concen: 5.114 ng
RT: 16.098 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM047141.D
Acq: 10 Aug 2024 13:52




Instrument :
BNA_M
ClientSampleId :
SSTDICC005


Tgt Ion:284 Resp: 78852
Ion Ratio Lower Upper
284 100
142 39.0 29.4 44.0
249 31.6 25.1 37.7



#69
Atrazine
Concen: 5.568 ng
RT: 16.286 min Scan# 2278
Delta R.T. -0.006 min
Lab File: BM047141.D
Acq: 10 Aug 2024 13:52

Tgt Ion:200 Resp: 60979
Ion Ratio Lower Upper
200 100
173 26.2 6.1 46.1
215 44.9 26.8 66.8





#70

Pentachlorophenol

Concen: 4.362 ng

RT: 16.451 min Scan# 21

Delta R.T. -0.000 min

Lab File: BM047141.D

Acq: 10 Aug 2024 13:52

Instrument :

BNA_M

ClientSampleId :

SSTDICC005





Tgt Ion:266 Resp: 47145

Ion Ratio Lower Upper

266 100

268 65.8 49.8 74.6

264 63.7 49.6 74.4



#71

Phenanthrene

Concen: 5.428 ng

RT: 16.833 min Scan# 2371

Delta R.T. -0.006 min

Lab File: BM047141.D

Acq: 10 Aug 2024 13:52




Tgt Ion:178 Resp: 363970


Ion Ratio Lower Upper

178 100

176 18.6 15.3 22.9

179 15.1 12.4 18.6









#72
Anthracene
Concen: 5.243 ng
RT: 16.921 min Scan# 21
Delta R.T. -0.006 min
Lab File: BM047141.D
Acq: 10 Aug 2024 13:52




Instrument :
BNA_M
ClientSampleId :
SSTDICC005


Tgt Ion:178 Resp: 342081
Ion Ratio Lower Upper
178 100
176 18.1 14.8 22.2
179 14.7 12.4 18.6



#73
Carbazole
Concen: 5.368 ng
RT: 17.204 min Scan# 2434
Delta R.T. -0.006 min
Lab File: BM047141.D
Acq: 10 Aug 2024 13:52

Tgt Ion:167 Resp: 357631
Ion Ratio Lower Upper
167 100
166 20.9 17.1 25.7
139 12.9 10.0 15.0









#74
Di-n-butylphthalate
Concen: 5.064 ng
RT: 17.792 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM047141.D
Acq: 10 Aug 2024 13:52




Instrument :
BNA_M
ClientSampleId :
SSTDICC005


Tgt Ion:149 Resp: 392569
Ion Ratio Lower Upper
149 100
150 8.6 7.5 11.3
104 5.5 4.1 6.1



#75
Fluoranthene
Concen: 5.301 ng
RT: 18.868 min Scan# 2717
Delta R.T. -0.000 min
Lab File: BM047141.D
Acq: 10 Aug 2024 13:52

Tgt Ion:202 Resp: 432338
Ion Ratio Lower Upper
202 100
101 11.8 0.0 30.2
203 17.4 0.0 37.6









#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.027 min Scan# 3084
Delta R.T. -0.006 min
Lab File: BM047141.D
Acq: 10 Aug 2024 13:52




Instrument :
BNA_M
ClientSampleId :
SSTDICC005


Tgt Ion:240 Resp: 1301329
Ion Ratio Lower Upper
240 100
120 10.9 8.2 12.2
236 25.4 20.6 30.8



#77
Benzidine
Concen: 3.335 ng
RT: 19.074 min Scan# 2752
Delta R.T. -0.000 min
Lab File: BM047141.D
Acq: 10 Aug 2024 13:52

Tgt Ion:184 Resp: 73589
Ion Ratio Lower Upper
184 100
185 13.2 11.8 17.6
183 13.0 9.0 13.6









#78
Pyrene
Concen: 5.098 ng
RT: 19.233 min Scan# 21
Delta R.T. -0.006 min
Lab File: BM047141.D
Acq: 10 Aug 2024 13:52




Instrument :
BNA_M
ClientSampleId :
SSTDICC005


Tgt Ion:202 Resp: 473702
Ion Ratio Lower Upper
202 100
200 20.8 17.2 25.8
203 17.1 14.2 21.2



#79
Terphenyl-d14
Concen: 9.901 ng
RT: 19.456 min Scan# 2817
Delta R.T. -0.006 min
Lab File: BM047141.D
Acq: 10 Aug 2024 13:52

Tgt Ion:244 Resp: 601320
Ion Ratio Lower Upper
244 100
212 7.3 6.2 9.4
122 11.3 7.1 10.7#









#80
Butylbenzylphthalate
Concen: 4.381 ng
RT: 20.174 min Scan# 2939
Delta R.T. -0.000 min
Lab File: BM047141.D
Acq: 10 Aug 2024 13:52




Instrument :
BNA_M
ClientSampleId :
SSTDICC005


Tgt Ion:149 Resp: 166134
Ion Ratio Lower Upper
149 100
91 73.9 57.0 85.4
206 18.7 16.4 24.6



#81
Benzo(a)anthracene
Concen: 5.236 ng
RT: 21.009 min Scan# 3081
Delta R.T. -0.006 min
Lab File: BM047141.D
Acq: 10 Aug 2024 13:52

Tgt Ion:228 Resp: 452360
Ion Ratio Lower Upper
228 100
226 26.1 21.8 32.6
229 19.1 15.8 23.8









#82
3,3'-Dichlorobenzidine
Concen: 4.493 ng
RT: 20.950 min Scan# 3071
Delta R.T. -0.006 min
Lab File: BM047141.D
Acq: 10 Aug 2024 13:52




Instrument :
BNA_M
ClientSampleId :
SSTDICC005


Tgt Ion:252 Resp: 121454
Ion Ratio Lower Upper
252 100
254 64.0 51.0 76.6
126 11.3 8.5 12.7



#83
Chrysene
Concen: 5.360 ng
RT: 21.062 min Scan# 3090
Delta R.T. -0.012 min
Lab File: BM047141.D
Acq: 10 Aug 2024 13:52

Tgt Ion:228 Resp: 439298
Ion Ratio Lower Upper
228 100
226 28.7 24.4 36.6
229 19.8 15.7 23.5









#84
Bis(2-ethylhexyl)phthalate
Concen: 4.665 ng
RT: 20.974 min Scan# 3075
Delta R.T. -0.000 min
Lab File: BM047141.D
Acq: 10 Aug 2024 13:52




Instrument :
BNA_M
ClientSampleId :
SSTDICC005


Tgt Ion:149 Resp: 250857
Ion Ratio Lower Upper
149 100
167 27.0 22.2 33.2
279 3.6 3.0 4.4



#85
Di-n-octyl phthalate
Concen: 4.315 ng
RT: 22.003 min Scan# 3250
Delta R.T. -0.006 min
Lab File: BM047141.D
Acq: 10 Aug 2024 13:52

Tgt Ion:149 Resp: 394399
Ion Ratio Lower Upper
149 100
167 1.6 1.3 1.9
43 9.7 6.9 10.3









#86
Perylene-d12
Concen: 20.000 ng
RT: 23.727 min Scan# 31
Delta R.T. -0.000 min
Lab File: BM047141.D
Acq: 10 Aug 2024 13:52




Instrument :
BNA_M
ClientSampleId :
SSTDICC005


Tgt Ion:264 Resp: 1576050
Ion Ratio Lower Upper
264 100
260 22.8 18.5 27.7
265 21.8 17.3 25.9



#87
Indeno(1,2,3-cd)pyrene
Concen: 5.130 ng
RT: 26.715 min Scan# 4051
Delta R.T. -0.012 min
Lab File: BM047141.D
Acq: 10 Aug 2024 13:52

Tgt Ion:276 Resp: 522557
Ion Ratio Lower Upper
276 100
138 25.7 19.4 29.0
277 25.1 20.1 30.1









#88
Benzo(b)fluoranthene
Concen: 4.975 ng
RT: 22.880 min Scan# 31
Delta R.T. -0.006 min
Lab File: BM047141.D
Acq: 10 Aug 2024 13:52




Instrument :
BNA_M
ClientSampleId :
SSTDICC005


Tgt Ion:252 Resp: 465640
Ion Ratio Lower Upper
252 100
253 22.5 17.8 26.6
125 10.1 7.4 11.0



#89
Benzo(k)fluoranthene
Concen: 5.314 ng
RT: 22.933 min Scan# 3408
Delta R.T. -0.012 min
Lab File: BM047141.D
Acq: 10 Aug 2024 13:52

Tgt Ion:252 Resp: 469757
Ion Ratio Lower Upper
252 100
253 21.8 17.7 26.5
125 10.5 7.4 11.2









#90
Benzo(a)pyrene
Concen: 4.929 ng
RT: 23.597 min Scan# 31
Delta R.T. -0.012 min
Lab File: BM047141.D
Acq: 10 Aug 2024 13:52




Instrument :
BNA_M
ClientSampleId :
SSTDICC005


Tgt Ion:252 Resp: 399615
Ion Ratio Lower Upper
252 100
253 21.6 17.1 25.7
125 11.8 8.2 12.2



#91
Dibenzo(a,h)anthracene
Concen: 5.237 ng
RT: 26.762 min Scan# 4059
Delta R.T. -0.018 min
Lab File: BM047141.D
Acq: 10 Aug 2024 13:52

Tgt Ion:278 Resp: 431713
Ion Ratio Lower Upper
278 100
139 16.1 12.0 18.0
279 23.1 19.2 28.8





#92

Benzo(g,h,i)perylene

Concen: 5.151 ng

RT: 27.632 min Scan# 41

Delta R.T. -0.030 min

Lab File: BM047141.D


Acq: 10 Aug 2024 13:52

Instrument :

BNA_M

ClientSampleId :

SSTDICC005





Tgt Ion:276 Resp: 440920

Ion Ratio Lower Upper

276 100

277 24.0 18.8 28.2

138 22.4 16.2 24.4



Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081024\
 Data File : BM047142.D
 Acq On : 10 Aug 2024 14:31
 Operator : RC/JU
 Sample : SSTDICC010
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC010

Quant Time: Aug 11 23:32:14 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Aug 11 23:24:37 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.375	152	292955	20.000	ng	0.00
21) Naphthalene-d8	10.128	136	1186627	20.000	ng	0.00
39) Acenaphthene-d10	14.033	164	788651	20.000	ng	0.00
64) Phenanthrene-d10	16.792	188	1690400	20.000	ng	0.00
76) Chrysene-d12	21.027	240	1593166	20.000	ng	0.00
86) Perylene-d12	23.727	264	1901592	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.022	112	346900	21.159	ng	0.00
7) Phenol-d6	6.581	99	468308	20.730	ng	0.00
23) Nitrobenzene-d5	8.522	82	494443	20.981	ng	0.00
42) 2,4,6-Tribromophenol	15.539	330	183466	20.063	ng	0.00
45) 2-Fluorobiphenyl	12.639	172	1057713	20.688	ng	0.00
79) Terphenyl-d14	19.456	244	1459931	19.636	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.040	88	71740	10.561	ng	99
3) Pyridine	3.416	79	203646	10.230	ng	97
4) n-Nitrosodimethylamine	3.334	42	89658	10.302	ng	# 99
6) Aniline	6.722	93	222003	10.397	ng	99
8) 2-Chlorophenol	6.951	128	186177	10.412	ng	98
9) Benzaldehyde	6.540	77	153438	9.573	ng	99
10) Phenol	6.604	94	234318	10.395	ng	99
11) bis(2-Chloroethyl)ether	6.822	93	199316	10.325	ng	99
12) 1,3-Dichlorobenzene	7.263	146	220661	10.384	ng	99
13) 1,4-Dichlorobenzene	7.410	146	224423	10.431	ng	99
14) 1,2-Dichlorobenzene	7.716	146	215097	10.596	ng	98
15) Benzyl Alcohol	7.628	79	157264	9.781	ng	98
16) 2,2'-oxybis(1-Chloropr...	7.910	45	260329	10.497	ng	98
17) 2-Methylphenol	7.834	107	154501	10.274	ng	99
18) Hexachloroethane	8.428	117	84578	10.149	ng	97
19) n-Nitroso-di-n-propyla...	8.181	70	160456	10.544	ng	97
20) 3+4-Methylphenols	8.157	107	211260	10.048	ng	98
22) Acetophenone	8.192	105	306786	10.711	ng	98
24) Nitrobenzene	8.563	77	253056	10.664	ng	99
25) Isophorone	9.086	82	418965	10.211	ng	100
26) 2-Nitrophenol	9.263	139	98575	9.380	ng	98
27) 2,4-Dimethylphenol	9.345	122	123790	10.056	ng	99
28) bis(2-Chloroethoxy)met...	9.575	93	273307	10.586	ng	99
29) 2,4-Dichlorophenol	9.792	162	178052	9.748	ng	99
30) 1,2,4-Trichlorobenzene	9.998	180	208733	10.098	ng	98
31) Naphthalene	10.181	128	617578	10.447	ng	99
32) Benzoic acid	9.457	122	125871	8.752	ng	96
33) 4-Chloroaniline	10.304	127	233198	10.258	ng	98
34) Hexachlorobutadiene	10.469	225	118512	9.795	ng	99
35) Caprolactam	11.086	113	61384	9.677	ng	98
36) 4-Chloro-3-methylphenol	11.451	107	192243	9.883	ng	94
37) 2-Methylnaphthalene	11.804	142	433416	10.298	ng	99
38) 1-Methylnaphthalene	12.027	142	430690	10.354	ng	100
40) 1,2,4,5-Tetrachloroben...	12.186	216	209468	9.738	ng	99
41) Hexachlorocyclopentadiene	12.163	237	70782	9.423	ng	96
43) 2,4,6-Trichlorophenol	12.445	196	146662	9.593	ng	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081024\
 Data File : BM047142.D
 Acq On : 10 Aug 2024 14:31
 Operator : RC/JU
 Sample : SSTDICC010
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC010

Quant Time: Aug 11 23:32:14 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Aug 11 23:24:37 2024
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	12.522	196	161605	9.520	ng	98
46) 1,1'-Biphenyl	12.851	154	584326	10.407	ng	99
47) 2-Chloronaphthalene	12.886	162	460927	10.503	ng	99
48) 2-Nitroaniline	13.110	65	136444	9.925	ng	95
49) Acenaphthylene	13.745	152	669998	10.358	ng	100
50) Dimethylphthalate	13.510	163	565056	10.259	ng	100
51) 2,6-Dinitrotoluene	13.621	165	128162	9.954	ng	95
52) Acenaphthene	14.098	154	425366	10.367	ng	98
53) 3-Nitroaniline	13.957	138	124808	9.674	ng	96
54) 2,4-Dinitrophenol	14.174	184	60036	7.780	ng	96
55) Dibenzofuran	14.439	168	682447	10.530	ng	99
56) 4-Nitrophenol	14.292	139	105813	9.511	ng	94
57) 2,4-Dinitrotoluene	14.421	165	171542	9.982	ng	92
58) Fluorene	15.092	166	562937	10.527	ng	100
59) 2,3,4,6-Tetrachlorophenol	14.674	232	135880	9.739	ng	96
60) Diethylphthalate	14.892	149	586005	10.418	ng	98
61) 4-Chlorophenyl-phenyle...	15.098	204	261449	10.269	ng	98
62) 4-Nitroaniline	15.127	138	127938m	10.065	ng	
63) Azobenzene	15.392	77	569823	10.685	ng	99
65) 4,6-Dinitro-2-methylph...	15.192	198	90017	9.095	ng	95
66) n-Nitrosodiphenylamine	15.315	169	473749	10.289	ng	98
67) 4-Bromophenyl-phenylether	15.992	248	160370	9.845	ng	95
68) Hexachlorobenzene	16.098	284	198141	10.157	ng	97
69) Atrazine	16.286	200	157170	11.344	ng	98
70) Pentachlorophenol	16.451	266	128691	9.411	ng	99
71) Phenanthrene	16.833	178	874873	10.313	ng	99
72) Anthracene	16.927	178	845520	10.243	ng	99
73) Carbazole	17.204	167	879612	10.437	ng	99
74) Di-n-butylphthalate	17.792	149	1003294	10.231	ng	100
75) Fluoranthene	18.868	202	1047590	10.152	ng	99
77) Benzidine	19.074	184	151348	5.602	ng	100
78) Pyrene	19.233	202	1113925	9.792	ng	99
80) Butylbenzylphthalate	20.174	149	439513	9.467	ng	99
81) Benzo(a)anthracene	21.009	228	1060395	10.026	ng	100
82) 3,3'-Dichlorobenzidine	20.950	252	311004	9.398	ng	99
83) Chrysene	21.068	228	1022315	10.188	ng	99
84) Bis(2-ethylhexyl)phtha...	20.974	149	643719	9.777	ng	100
85) Di-n-octyl phthalate	22.003	149	1048294	9.368	ng	98
87) Indeno(1,2,3-cd)pyrene	26.715	276	1216673	9.899	ng	99
88) Benzo(b)fluoranthene	22.880	252	1140578	10.100	ng	98
89) Benzo(k)fluoranthene	22.933	252	1078126	10.108	ng	99
90) Benzo(a)pyrene	23.597	252	965036	9.865	ng	99
91) Dibenzo(a,h)anthracene	26.756	278	999745	10.051	ng	99
92) Benzo(g,h,i)perylene	27.644	276	1029079	9.964	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081024\
Data File : BM047142.D
Acq On : 10 Aug 2024 14:31
Operator : RC/JU
Sample : SSTDICC010
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTDICC010

Quant Time: Aug 11 23:32:14 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sun Aug 11 23:24:37 2024
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.375 min Scan# 7
Delta R.T. -0.000 min
Lab File: BM047142.D
Acq: 10 Aug 2024 14:31




Instrument :
BNA_M
ClientSampleId :
SSTDICC010


Tgt Ion	Ratio	Lower	Upper
152	100		
150	157.7	120.5	180.7
115	56.5	45.5	68.3



#2
1,4-Dioxane
Concen: 10.561 ng
RT: 3.040 min Scan# 26
Delta R.T. 0.006 min
Lab File: BM047142.D
Acq: 10 Aug 2024 14:31

Tgt Ion	Ratio	Lower	Upper
88	100		
58	73.3	58.2	87.4
43	28.9	22.5	33.7









#3
Pyridine
Concen: 10.230 ng
RT: 3.416 min Scan# 90
Delta R.T. 0.006 min
Lab File: BM047142.D
Acq: 10 Aug 2024 14:31




Instrument :
BNA_M
ClientSampleId :
SSTDICC010


Tgt Ion: 79 Resp: 203646
Ion Ratio Lower Upper
79 100
52 65.3 51.2 76.8
51 34.9 26.1 39.1



#4
n-Nitrosodimethylamine
Concen: 10.302 ng
RT: 3.334 min Scan# 76
Delta R.T. 0.006 min
Lab File: BM047142.D
Acq: 10 Aug 2024 14:31

Tgt Ion: 42 Resp: 89658
Ion Ratio Lower Upper
42 100
74 141.9 112.9 169.3
44 10.2 5.3 7.9#









#5
2-Fluorophenol
Concen: 21.159 ng
RT: 5.022 min Scan# 363
Delta R.T. -0.000 min
Lab File: BM047142.D
Acq: 10 Aug 2024 14:31




Instrument :
BNA_M
ClientSampleId :
SSTDICC010


Tgt Ion:112 Resp: 346900
Ion Ratio Lower Upper
112 100
64 58.7 46.7 70.1
63 31.2 24.8 37.2



#6
Aniline
Concen: 10.397 ng
RT: 6.722 min Scan# 652
Delta R.T. -0.006 min
Lab File: BM047142.D
Acq: 10 Aug 2024 14:31

Tgt Ion: 93 Resp: 222003
Ion Ratio Lower Upper
93 100
66 38.4 31.4 47.2
65 19.7 15.6 23.4









#7
Phenol-d6
Concen: 20.730 ng
RT: 6.581 min Scan# 61
Delta R.T. -0.006 min
Lab File: BM047142.D
Acq: 10 Aug 2024 14:31




Instrument :
BNA_M
ClientSampleId :
SSTDICC010


Tgt Ion: 99 Resp: 468308
Ion Ratio Lower Upper
99 100
42 17.8 14.1 21.1
71 36.0 28.5 42.7



#8
2-Chlorophenol
Concen: 10.412 ng
RT: 6.951 min Scan# 691
Delta R.T. -0.000 min
Lab File: BM047142.D
Acq: 10 Aug 2024 14:31

Tgt Ion: 128 Resp: 186177
Ion Ratio Lower Upper
128 100
130 34.6 12.1 52.1
64 46.6 26.6 66.6









#9
Benzaldehyde
Concen: 9.573 ng
RT: 6.540 min Scan# 61
Delta R.T. -0.000 min
Lab File: BM047142.D
Acq: 10 Aug 2024 14:31




Instrument :
BNA_M
ClientSampleId :
SSTDICC010


Tgt Ion: 77 Resp: 153438
Ion Ratio Lower Upper
77 100
105 95.8 74.5 114.5
106 92.7 72.3 112.3



#10
Phenol
Concen: 10.395 ng
RT: 6.604 min Scan# 632
Delta R.T. -0.006 min
Lab File: BM047142.D
Acq: 10 Aug 2024 14:31

Tgt Ion: 94 Resp: 234318
Ion Ratio Lower Upper
94 100
65 29.6 9.0 49.0
66 38.4 18.9 58.9









#11
bis(2-Chloroethyl)ether
Concen: 10.325 ng
RT: 6.822 min Scan# 669
Delta R.T. -0.006 min
Lab File: BM047142.D
Acq: 10 Aug 2024 14:31




Instrument :
BNA_M
ClientSampleId :
SSTDICC010


Tgt Ion: 93 Resp: 199316
Ion Ratio Lower Upper
93 100
63 77.4 55.8 95.8
95 32.6 12.7 52.7



#12
1,3-Dichlorobenzene
Concen: 10.384 ng
RT: 7.263 min Scan# 744
Delta R.T. -0.000 min
Lab File: BM047142.D
Acq: 10 Aug 2024 14:31

Tgt Ion: 146 Resp: 220661
Ion Ratio Lower Upper
146 100
148 64.6 52.3 78.5
75 32.3 25.8 38.6





#13

1,4-Dichlorobenzene

Concen: 10.431 ng

RT: 7.410 min Scan# 7

Delta R.T. -0.000 min

Lab File: BM047142.D

Acq: 10 Aug 2024 14:31

Instrument :

BNA_M

ClientSampleId :

SSTDICC010





Tgt Ion:146 Resp: 224423

Ion Ratio Lower Upper

146 100

148 64.7 51.4 77.2

111 38.4 31.5 47.3



#14

1,2-Dichlorobenzene

Concen: 10.596 ng

RT: 7.716 min Scan# 821

Delta R.T. -0.000 min

Lab File: BM047142.D

Acq: 10 Aug 2024 14:31




Tgt Ion:146 Resp: 215097


Ion Ratio Lower Upper

146 100

148 65.9 51.8 77.8

111 43.0 33.3 49.9









#15
Benzyl Alcohol
Concen: 9.781 ng
RT: 7.628 min Scan# 806
Delta R.T. -0.000 min
Lab File: BM047142.D
Acq: 10 Aug 2024 14:31




Instrument :
BNA_M
ClientSampleId :
SSTDICC010


Tgt Ion: 79 Resp: 157264
Ion Ratio Lower Upper
79 100
108 76.5 59.7 89.5
77 65.0 52.9 79.3



#16
2,2'-oxybis(1-Chloropropane)
Concen: 10.497 ng
RT: 7.910 min Scan# 854
Delta R.T. -0.006 min
Lab File: BM047142.D
Acq: 10 Aug 2024 14:31

Tgt Ion: 45 Resp: 260329
Ion Ratio Lower Upper
45 100
77 16.2 0.0 35.6
79 11.4 0.0 32.4





#17

2-Methylphenol

Concen: 10.274 ng

RT: 7.834 min Scan# 841

Delta R.T. -0.000 min

Lab File: BM047142.D


Acq: 10 Aug 2024 14:31

Instrument :

BNA_M




ClientSampleId :

SSTDICC010



Tgt Ion:107 Resp: 154501

Ion	Ratio	Lower	Upper
107	100		
108	111.0	90.6	135.8
77	49.4	39.3	58.9
79	48.6	38.6	58.0



#18

Hexachloroethane

Concen: 10.149 ng

RT: 8.428 min Scan# 942




Delta R.T. -0.000 min


Lab File: BM047142.D

Acq: 10 Aug 2024 14:31

Tgt Ion:117 Resp: 84578

Ion	Ratio	Lower	Upper
117	100		
119	101.9	77.0	115.4
201	101.2	81.3	121.9









#19
n-Nitroso-di-n-propylamine
Concen: 10.544 ng
RT: 8.181 min Scan# 901
Delta R.T. -0.006 min
Lab File: BM047142.D
Acq: 10 Aug 2024 14:31




Instrument :
BNA_M
ClientSampleId :
SSTDICC010


Tgt Ion: 70 Resp: 160456
Ion Ratio Lower Upper
70 100
42 54.9 42.0 63.0
101 9.2 7.2 10.8
130 18.8 15.5 23.3



#20
3+4-Methylphenols
Concen: 10.048 ng
RT: 8.157 min Scan# 896
Delta R.T. -0.006 min
Lab File: BM047142.D
Acq: 10 Aug 2024 14:31

Tgt Ion: 107 Resp: 211260
Ion Ratio Lower Upper
107 100
108 92.4 70.7 110.7
77 38.0 17.4 57.4
79 32.3 10.9 50.9









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.128 min Scan# 11
Delta R.T. -0.006 min
Lab File: BM047142.D
Acq: 10 Aug 2024 14:31




Instrument :
BNA_M
ClientSampleId :
SSTDICC010


Tgt Ion:136 Resp: 1186627
Ion Ratio Lower Upper
136 100
137 10.7 8.8 13.2
54 7.5 6.0 9.0
68 6.4 4.8 7.2



#22
Acetophenone
Concen: 10.711 ng
RT: 8.192 min Scan# 902
Delta R.T. -0.006 min
Lab File: BM047142.D
Acq: 10 Aug 2024 14:31

Tgt Ion:105 Resp: 306786
Ion Ratio Lower Upper
105 100
71 5.2 4.5 6.7
51 26.8 20.1 30.1
120 22.4 17.8 26.6









#23
Nitrobenzene-d5
Concen: 20.981 ng
RT: 8.522 min Scan# 91
Delta R.T. -0.000 min
Lab File: BM047142.D
Acq: 10 Aug 2024 14:31




Instrument :
BNA_M
ClientSampleId :
SSTDICC010


Tgt Ion: 82 Resp: 494443
Ion Ratio Lower Upper
82 100
128 40.2 32.3 48.5
54 45.9 37.4 56.2



#24
Nitrobenzene
Concen: 10.664 ng
RT: 8.563 min Scan# 965
Delta R.T. -0.000 min
Lab File: BM047142.D
Acq: 10 Aug 2024 14:31

Tgt Ion: 77 Resp: 253056
Ion Ratio Lower Upper
77 100
123 42.8 34.0 51.0
65 12.7 10.5 15.7









#25
Isophorone
Concen: 10.211 ng
RT: 9.086 min Scan# 1054
Delta R.T. -0.006 min
Lab File: BM047142.D
Acq: 10 Aug 2024 14:31




Instrument :
BNA_M
ClientSampleId :
SSTDICC010


Tgt Ion: 82 Resp: 418965
Ion Ratio Lower Upper
82 100
95 6.8 5.5 8.3
138 16.6 13.5 20.3



#26
2-Nitrophenol
Concen: 9.380 ng
RT: 9.263 min Scan# 1084
Delta R.T. -0.000 min
Lab File: BM047142.D
Acq: 10 Aug 2024 14:31

Tgt Ion:139 Resp: 98575
Ion Ratio Lower Upper
139 100
109 30.5 24.4 36.6
65 56.0 43.0 64.6





#27

2,4-Dimethylphenol

Concen: 10.056 ng

RT: 9.345 min Scan# 1109

Delta R.T. -0.000 min

Lab File: BM047142.D


Acq: 10 Aug 2024 14:31

Instrument :

BNA_M

ClientSampleId :

SSTDICC010






Tgt Ion:122 Resp: 123790

Ion Ratio Lower Upper

122 100

107 116.9 92.8 139.2

121 59.8 46.1 69.1



#28

bis(2-Chloroethoxy)methane

Concen: 10.586 ng

RT: 9.575 min Scan# 1137

Delta R.T. -0.000 min

Lab File: BM047142.D

Acq: 10 Aug 2024 14:31




Tgt Ion: 93 Resp: 273307


Ion Ratio Lower Upper

93 100

95 31.9 25.8 38.8

123 10.9 9.0 13.4





#29

2,4-Dichlorophenol

Concen: 9.748 ng

RT: 9.792 min Scan# 1174

Delta R.T. -0.006 min

Lab File: BM047142.D

Acq: 10 Aug 2024 14:31

Instrument :

BNA_M

ClientSampleId :

SSTDICC010





Tgt Ion:162 Resp: 178052

Ion Ratio Lower Upper

162 100

164 65.3 44.7 84.7

98 36.3 14.9 54.9



#30

1,2,4-Trichlorobenzene

Concen: 10.098 ng

RT: 9.998 min Scan# 1209

Delta R.T. -0.000 min

Lab File: BM047142.D

Acq: 10 Aug 2024 14:31




Tgt Ion:180 Resp: 208733


Ion Ratio Lower Upper

180 100

182 93.3 76.6 115.0

145 28.3 23.4 35.2









#31
Naphthalene
Concen: 10.447 ng
RT: 10.181 min Scan# 1117
Delta R.T. -0.000 min
Lab File: BM047142.D
Acq: 10 Aug 2024 14:31




Instrument :
BNA_M
ClientSampleId :
SSTDICC010


Tgt Ion:128 Resp: 617578
Ion Ratio Lower Upper
128 100
129 10.9 8.6 13.0
127 12.2 10.2 15.4



#32
Benzoic acid
Concen: 8.752 ng
RT: 9.457 min Scan# 1117
Delta R.T. -0.077 min
Lab File: BM047142.D
Acq: 10 Aug 2024 14:31

Tgt Ion:122 Resp: 125871
Ion Ratio Lower Upper
122 100
105 123.1 107.1 147.1
77 96.7 82.1 122.1









#33
4-Chloroaniline
Concen: 10.258 ng
RT: 10.304 min Scan# 1261
Delta R.T. -0.006 min
Lab File: BM047142.D
Acq: 10 Aug 2024 14:31




Instrument :
BNA_M
ClientSampleId :
SSTDICC010


Tgt Ion:	127	Resp:	233198
Ion	Ratio	Lower	Upper
127	100		
129	30.8	25.8	38.8
65	32.0	24.9	37.3
92	18.3	15.1	22.7



#34
Hexachlorobutadiene
Concen: 9.795 ng
RT: 10.469 min Scan# 1289
Delta R.T. -0.000 min
Lab File: BM047142.D
Acq: 10 Aug 2024 14:31

Tgt Ion:	225	Resp:	118512
Ion	Ratio	Lower	Upper
225	100		
223	64.5	50.2	75.2
227	64.6	51.4	77.0









#35
Caprolactam
Concen: 9.677 ng
RT: 11.086 min Scan# 11394
Delta R.T. -0.035 min
Lab File: BM047142.D
Acq: 10 Aug 2024 14:31




Instrument :
BNA_M
ClientSampleId :
SSTDICC010


Tgt Ion:113 Resp: 61384
Ion Ratio Lower Upper
113 100
55 162.9 140.5 180.5
56 121.2 103.6 143.6



#36
4-Chloro-3-methylphenol
Concen: 9.883 ng
RT: 11.451 min Scan# 1456
Delta R.T. -0.000 min
Lab File: BM047142.D
Acq: 10 Aug 2024 14:31

Tgt Ion:107 Resp: 192243
Ion Ratio Lower Upper
107 100
144 26.6 21.8 32.6
142 90.0 66.8 100.2









#37
2-Methylnaphthalene
Concen: 10.298 ng
RT: 11.804 min Scan# 1516
Delta R.T. -0.000 min
Lab File: BM047142.D
Acq: 10 Aug 2024 14:31




Instrument :
BNA_M
ClientSampleId :
SSTDICC010


Tgt Ion:142 Resp: 433416
Ion Ratio Lower Upper
142 100
141 87.3 70.9 106.3
115 35.2 28.7 43.1



#38
1-Methylnaphthalene
Concen: 10.354 ng
RT: 12.027 min Scan# 1554
Delta R.T. -0.006 min
Lab File: BM047142.D
Acq: 10 Aug 2024 14:31

Tgt Ion:142 Resp: 430690
Ion Ratio Lower Upper
142 100
141 93.6 74.6 111.8
116 3.9 3.1 4.7





#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 14.033 min Scan# 1895

Delta R.T. -0.000 min

Lab File: BM047142.D

Acq: 10 Aug 2024 14:31

Instrument :

BNA_M

ClientSampleId :

SSTDICC010





Tgt Ion:164 Resp: 788651

Ion Ratio Lower Upper

164 100

162 100.6 80.7 121.1

160 45.0 36.0 54.0



#40

1,2,4,5-Tetrachlorobenzene

Concen: 9.738 ng

RT: 12.186 min Scan# 1581

Delta R.T. -0.000 min

Lab File: BM047142.D

Acq: 10 Aug 2024 14:31

Tgt Ion:216 Resp: 209468




Ion Ratio Lower Upper


216 100

214 80.0 63.4 95.2

179 23.0 17.6 26.4

108 20.1 15.1 22.7









#41
Hexachlorocyclopentadiene
Concen: 9.423 ng
RT: 12.163 min Scan# 1577
Delta R.T. -0.000 min
Lab File: BM047142.D
Acq: 10 Aug 2024 14:31




Instrument :
BNA_M
ClientSampleId :
SSTDICC010


Tgt Ion	Ratio	Lower	Upper
237	100		
235	61.3	44.8	84.8
272	12.0	0.0	33.5



#42
2,4,6-Tribromophenol
Concen: 20.063 ng
RT: 15.539 min Scan# 2151
Delta R.T. -0.000 min
Lab File: BM047142.D
Acq: 10 Aug 2024 14:31

Tgt Ion	Ratio	Lower	Upper
330	100		
332	97.5	77.4	116.2
141	42.0	31.6	47.4





#43

2,4,6-Trichlorophenol

Concen: 9.593 ng

RT: 12.445 min Scan# 1

Delta R.T. -0.000 min

Lab File: BM047142.D

Acq: 10 Aug 2024 14:31

Instrument :

BNA_M

ClientSampleId :

SSTDICC010





Tgt Ion:196 Resp: 146662

Ion Ratio Lower Upper

196 100

198 94.5 75.6 113.4

200 30.4 24.6 37.0



#44

2,4,5-Trichlorophenol

Concen: 9.520 ng

RT: 12.522 min Scan# 1638

Delta R.T. -0.000 min

Lab File: BM047142.D

Acq: 10 Aug 2024 14:31

Tgt Ion:196 Resp: 161605




Ion Ratio Lower Upper


196 100

198 98.9 77.7 116.5

97 51.0 39.8 59.8

132 25.9 20.7 31.1









#45
2-Fluorobiphenyl
Concen: 20.688 ng
RT: 12.639 min Scan# 1658
Delta R.T. -0.006 min
Lab File: BM047142.D
Acq: 10 Aug 2024 14:31




Instrument :
BNA_M
ClientSampleId :
SSTDICC010


Tgt Ion:172 Resp: 1057713
Ion Ratio Lower Upper
172 100
171 34.6 27.8 41.6
170 23.4 18.7 28.1



#46
1,1'-Biphenyl
Concen: 10.407 ng
RT: 12.851 min Scan# 1694
Delta R.T. -0.006 min
Lab File: BM047142.D
Acq: 10 Aug 2024 14:31

Tgt Ion:154 Resp: 584326
Ion Ratio Lower Upper
154 100
153 39.1 19.0 59.0
76 12.9 0.0 31.9





#47

2-Chloronaphthalene

Concen: 10.503 ng

RT: 12.886 min Scan# 1

Delta R.T. -0.000 min

Lab File: BM047142.D

Acq: 10 Aug 2024 14:31

Instrument :

BNA_M

ClientSampleId :

SSTDICC010





Tgt Ion:162 Resp: 460927

Ion Ratio Lower Upper

162 100

127 36.6 30.0 45.0

164 32.6 26.3 39.5



#48

2-Nitroaniline

Concen: 9.925 ng

RT: 13.110 min Scan# 1738

Delta R.T. -0.006 min

Lab File: BM047142.D

Acq: 10 Aug 2024 14:31




Tgt Ion: 65 Resp: 136444


Ion Ratio Lower Upper

65 100

92 69.5 55.1 82.7

138 98.8 85.9 128.9









#49
Acenaphthylene
Concen: 10.358 ng
RT: 13.745 min Scan# 1846
Delta R.T. -0.006 min
Lab File: BM047142.D
Acq: 10 Aug 2024 14:31




Instrument :
BNA_M
ClientSampleId :
SSTDICC010


Tgt Ion:152 Resp: 669998
Ion Ratio Lower Upper
152 100
151 19.8 15.8 23.6
153 13.0 10.4 15.6



#50
Dimethylphthalate
Concen: 10.259 ng
RT: 13.510 min Scan# 1806
Delta R.T. -0.006 min
Lab File: BM047142.D
Acq: 10 Aug 2024 14:31

Tgt Ion:163 Resp: 565056
Ion Ratio Lower Upper
163 100
194 4.0 3.0 4.6
164 10.0 8.0 12.0









#51
2,6-Dinitrotoluene
Concen: 9.954 ng
RT: 13.621 min Scan# 1825
Delta R.T. -0.006 min
Lab File: BM047142.D
Acq: 10 Aug 2024 14:31




Instrument :
BNA_M
ClientSampleId :
SSTDICC010


Tgt Ion:165 Resp: 128162
Ion Ratio Lower Upper
165 100
63 61.6 46.5 69.7
89 63.5 47.3 70.9



#52
Acenaphthene
Concen: 10.367 ng
RT: 14.098 min Scan# 1906
Delta R.T. -0.000 min
Lab File: BM047142.D
Acq: 10 Aug 2024 14:31

Tgt Ion:154 Resp: 425366
Ion Ratio Lower Upper
154 100
153 115.7 90.3 135.5
152 53.6 42.9 64.3









#53
3-Nitroaniline
Concen: 9.674 ng
RT: 13.957 min Scan# 1882
Delta R.T. -0.006 min
Lab File: BM047142.D
Acq: 10 Aug 2024 14:31




Instrument :
BNA_M
ClientSampleId :
SSTDICC010


Tgt Ion:138 Resp: 124808
Ion Ratio Lower Upper
138 100
108 10.1 7.9 11.9
92 119.5 92.0 138.0



#54
2,4-Dinitrophenol
Concen: 7.780 ng
RT: 14.174 min Scan# 1919
Delta R.T. -0.001 min
Lab File: BM047142.D
Acq: 10 Aug 2024 14:31

Tgt Ion:184 Resp: 60036
Ion Ratio Lower Upper
184 100
63 68.4 52.7 79.1
154 66.6 50.4 75.6





#55

Dibenzenofuran

Concen: 10.530 ng

RT: 14.439 min Scan# 1964

Delta R.T. -0.000 min

Lab File: BM047142.D

Acq: 10 Aug 2024 14:31

Instrument :

BNA_M

ClientSampleId :

SSTDICC010





Tgt Ion:168 Resp: 682447

Ion Ratio Lower Upper

168 100

139 36.5 29.6 44.4

169 13.3 10.7 16.1



#56

4-Nitrophenol

Concen: 9.511 ng

RT: 14.292 min Scan# 1939

Delta R.T. -0.006 min

Lab File: BM047142.D

Acq: 10 Aug 2024 14:31




Tgt Ion:139 Resp: 105813


Ion Ratio Lower Upper

139 100

109 75.3 52.5 92.5

65 103.0 75.4 115.4









#57
2,4-Dinitrotoluene
Concen: 9.982 ng
RT: 14.421 min Scan# 1961
Delta R.T. -0.006 min
Lab File: BM047142.D
Acq: 10 Aug 2024 14:31




Instrument :
BNA_M
ClientSampleId :
SSTDICC010


Tgt Ion:165 Resp: 171542
Ion Ratio Lower Upper
165 100
63 55.8 40.3 60.5
89 83.7 61.0 91.6
182 2.4 1.9 2.9



#58
Fluorene
Concen: 10.527 ng
RT: 15.092 min Scan# 2075
Delta R.T. -0.000 min
Lab File: BM047142.D
Acq: 10 Aug 2024 14:31

Tgt Ion:166 Resp: 562937
Ion Ratio Lower Upper
166 100
165 96.9 77.4 116.0
167 13.8 11.0 16.6









#59
2,3,4,6-Tetrachlorophenol
Concen: 9.739 ng
RT: 14.674 min Scan# 2004
Delta R.T. -0.000 min
Lab File: BM047142.D
Acq: 10 Aug 2024 14:31




Instrument :
BNA_M
ClientSampleId :
SSTDICC010


Tgt Ion:232 Resp: 135880
Ion Ratio Lower Upper
232 100
131 50.0 36.9 55.3
130 2.6 2.0 3.0
166 34.6 26.4 39.6



#60
Diethylphthalate
Concen: 10.418 ng
RT: 14.892 min Scan# 2041
Delta R.T. -0.006 min
Lab File: BM047142.D
Acq: 10 Aug 2024 14:31

Tgt Ion:149 Resp: 586005
Ion Ratio Lower Upper
149 100
177 20.7 17.6 26.4
150 11.6 9.8 14.6









#61
4-Chlorophenyl-phenylether
Concen: 10.269 ng
RT: 15.098 min Scan# 2076
Delta R.T. -0.000 min
Lab File: BM047142.D
Acq: 10 Aug 2024 14:31




Instrument :
BNA_M
ClientSampleId :
SSTDICC010


Tgt Ion:204 Resp: 261449
Ion Ratio Lower Upper
204 100
206 32.6 26.2 39.4
141 61.1 47.2 70.8



#62
4-Nitroaniline
Concen: 10.065 ng m
RT: 15.127 min Scan# 2081
Delta R.T. -0.012 min
Lab File: BM047142.D
Acq: 10 Aug 2024 14:31

Tgt Ion:138 Resp: 127938
Ion Ratio Lower Upper
138 100
92 54.7 31.9 71.9
108 83.4 63.8 103.8









#63
Azobenzene
Concen: 10.685 ng
RT: 15.392 min Scan# 2126
Delta R.T. -0.000 min
Lab File: BM047142.D
Acq: 10 Aug 2024 14:31




Instrument :
BNA_M
ClientSampleId :
SSTDICC010


Tgt Ion: 77 Resp: 569823
Ion Ratio Lower Upper
77 100
182 24.3 3.8 43.8
105 18.9 0.0 39.5
51 26.8 6.8 46.8



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 16.792 min Scan# 2364
Delta R.T. -0.000 min
Lab File: BM047142.D
Acq: 10 Aug 2024 14:31

Tgt Ion: 188 Resp: 1690400
Ion Ratio Lower Upper
188 100
94 8.7 7.0 10.6
80 9.9 8.0 12.0









#65
4,6-Dinitro-2-methylphenol
Concen: 9.095 ng
RT: 15.192 min Scan# 2092
Delta R.T. -0.006 min
Lab File: BM047142.D
Acq: 10 Aug 2024 14:31




Instrument :
BNA_M
ClientSampleId :
SSTDICC010


Tgt Ion	Ratio	Lower	Upper
198	100		
51	47.3	23.9	63.9
105	46.8	24.3	64.3



#66
n-Nitrosodiphenylamine
Concen: 10.289 ng
RT: 15.315 min Scan# 2113
Delta R.T. -0.006 min
Lab File: BM047142.D
Acq: 10 Aug 2024 14:31

Tgt Ion	Ratio	Lower	Upper
169	100		
168	66.1	53.8	80.8
167	35.2	29.3	43.9









#67
4-Bromophenyl-phenylether
Concen: 9.845 ng
RT: 15.992 min Scan# 21
Delta R.T. -0.006 min
Lab File: BM047142.D
Acq: 10 Aug 2024 14:31




Instrument :
BNA_M
ClientSampleId :
SSTDICC010


Tgt Ion:248 Resp: 160370
Ion Ratio Lower Upper
248 100
250 95.9 77.8 116.8
141 79.5 57.7 86.5



#68
Hexachlorobenzene
Concen: 10.157 ng
RT: 16.098 min Scan# 2246
Delta R.T. -0.000 min
Lab File: BM047142.D
Acq: 10 Aug 2024 14:31

Tgt Ion:284 Resp: 198141
Ion Ratio Lower Upper
284 100
142 39.7 29.4 44.0
249 31.4 25.1 37.7









#69
Atrazine
Concen: 11.344 ng
RT: 16.286 min Scan# 21
Delta R.T. -0.006 min
Lab File: BM047142.D
Acq: 10 Aug 2024 14:31




Instrument :
BNA_M
ClientSampleId :
SSTDICC010


Tgt Ion:200 Resp: 157170
Ion Ratio Lower Upper
200 100
173 26.4 6.1 46.1
215 45.4 26.8 66.8



#70
Pentachlorophenol
Concen: 9.411 ng
RT: 16.451 min Scan# 2306
Delta R.T. -0.000 min
Lab File: BM047142.D
Acq: 10 Aug 2024 14:31

Tgt Ion:266 Resp: 128691
Ion Ratio Lower Upper
266 100
268 61.6 49.8 74.6
264 61.0 49.6 74.4









#71
Phenanthrene
Concen: 10.313 ng
RT: 16.833 min Scan# 21
Delta R.T. -0.006 min
Lab File: BM047142.D
Acq: 10 Aug 2024 14:31




Instrument :
BNA_M
ClientSampleId :
SSTDICC010


Tgt Ion:178 Resp: 874873
Ion Ratio Lower Upper
178 100
176 18.9 15.3 22.9
179 15.3 12.4 18.6



#72
Anthracene
Concen: 10.243 ng
RT: 16.927 min Scan# 2387
Delta R.T. -0.000 min
Lab File: BM047142.D
Acq: 10 Aug 2024 14:31

Tgt Ion:178 Resp: 845520
Ion Ratio Lower Upper
178 100
176 18.1 14.8 22.2
179 15.8 12.4 18.6









#73
Carbazole
Concen: 10.437 ng
RT: 17.204 min Scan# 2434
Delta R.T. -0.006 min
Lab File: BM047142.D
Acq: 10 Aug 2024 14:31




Instrument :
BNA_M
ClientSampleId :
SSTDICC010


Tgt Ion:167 Resp: 879612
Ion Ratio Lower Upper
167 100
166 20.6 17.1 25.7
139 12.8 10.0 15.0



#74
Di-n-butylphthalate
Concen: 10.231 ng
RT: 17.792 min Scan# 2534
Delta R.T. -0.000 min
Lab File: BM047142.D
Acq: 10 Aug 2024 14:31

Tgt Ion:149 Resp: 1003294
Ion Ratio Lower Upper
149 100
150 9.3 7.5 11.3
104 5.1 4.1 6.1









#75
Fluoranthene
Concen: 10.152 ng
RT: 18.868 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM047142.D
Acq: 10 Aug 2024 14:31




Instrument :
BNA_M
ClientSampleId :
SSTDICC010


Tgt Ion:202 Resp: 1047590
Ion Ratio Lower Upper
202 100
101 10.9 0.0 30.2
203 17.5 0.0 37.6



#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.027 min Scan# 3084
Delta R.T. -0.006 min
Lab File: BM047142.D
Acq: 10 Aug 2024 14:31

Tgt Ion:240 Resp: 1593166
Ion Ratio Lower Upper
240 100
120 10.6 8.2 12.2
236 25.5 20.6 30.8









#77
Benzidine
Concen: 5.602 ng
RT: 19.074 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM047142.D
Acq: 10 Aug 2024 14:31




Instrument :
BNA_M
ClientSampleId :
SSTDICC010


Tgt Ion:184 Resp: 151348
Ion Ratio Lower Upper
184 100
185 14.7 11.8 17.6
183 11.3 9.0 13.6



#78
Pyrene
Concen: 9.792 ng
RT: 19.233 min Scan# 2779
Delta R.T. -0.006 min
Lab File: BM047142.D
Acq: 10 Aug 2024 14:31

Tgt Ion:202 Resp: 1113925
Ion Ratio Lower Upper
202 100
200 20.6 17.2 25.8
203 17.7 14.2 21.2









#79
Terphenyl-d14
Concen: 19.636 ng
RT: 19.456 min Scan# 2817
Delta R.T. -0.006 min
Lab File: BM047142.D
Acq: 10 Aug 2024 14:31




Instrument :
BNA_M
ClientSampleId :
SSTDICC010


Tgt Ion:244 Resp: 1459931
Ion Ratio Lower Upper
244 100
212 7.3 6.2 9.4
122 10.5 7.1 10.7



#80
Butylbenzylphthalate
Concen: 9.467 ng
RT: 20.174 min Scan# 2939
Delta R.T. -0.000 min
Lab File: BM047142.D
Acq: 10 Aug 2024 14:31

Tgt Ion:149 Resp: 439513
Ion Ratio Lower Upper
149 100
91 71.5 57.0 85.4
206 19.3 16.4 24.6









#81
Benzo(a)anthracene
Concen: 10.026 ng
RT: 21.009 min Scan# 3082
Delta R.T. -0.006 min
Lab File: BM047142.D
Acq: 10 Aug 2024 14:31




Instrument :
BNA_M
ClientSampleId :
SSTDICC010


Tgt Ion:228 Resp: 1060395
Ion Ratio Lower Upper
228 100
226 27.3 21.8 32.6
229 19.6 15.8 23.8



#82
3,3'-Dichlorobenzidine
Concen: 9.398 ng
RT: 20.950 min Scan# 3071
Delta R.T. -0.006 min
Lab File: BM047142.D
Acq: 10 Aug 2024 14:31

Tgt Ion:252 Resp: 311004
Ion Ratio Lower Upper
252 100
254 64.2 51.0 76.6
126 11.7 8.5 12.7





#83

Chrysene

Concen: 10.188 ng

RT: 21.068 min Scan# 3091

Delta R.T. -0.006 min

Lab File: BM047142.D

Acq: 10 Aug 2024 14:31

Instrument :

BNA_M

ClientSampleId :

SSTDICC010





Tgt Ion:228 Resp: 1022315

Ion Ratio Lower Upper

228 100

226 29.6 24.4 36.6

229 19.7 15.7 23.5



#84

Bis(2-ethylhexyl)phthalate

Concen: 9.777 ng

RT: 20.974 min Scan# 3075

Delta R.T. -0.000 min

Lab File: BM047142.D

Acq: 10 Aug 2024 14:31




Tgt Ion:149 Resp: 643719


Ion Ratio Lower Upper

149 100

167 27.6 22.2 33.2

279 3.7 3.0 4.4









#85
Di-n-octyl phthalate
Concen: 9.368 ng
RT: 22.003 min Scan# 31
Delta R.T. -0.006 min
Lab File: BM047142.D
Acq: 10 Aug 2024 14:31




Instrument :
BNA_M
ClientSampleId :
SSTDICC010


Tgt Ion:149 Resp: 1048294
Ion Ratio Lower Upper
149 100
167 1.5 1.3 1.9
43 9.5 6.9 10.3



#86
Perylene-d12
Concen: 20.000 ng
RT: 23.727 min Scan# 3543
Delta R.T. -0.000 min
Lab File: BM047142.D
Acq: 10 Aug 2024 14:31


Tgt Ion:264 Resp: 1901592
Ion Ratio Lower Upper
264 100
260 23.0 18.5 27.7
265 21.2 17.3 25.9







#87
Indeno(1,2,3-cd)pyrene
Concen: 9.899 ng
RT: 26.715 min Scan# 4051
Delta R.T. -0.012 min
Lab File: BM047142.D
Acq: 10 Aug 2024 14:31

Instrument :
BNA_M
ClientSampleId :
SSTDICC010







Tgt Ion:276 Resp: 1216673
Ion Ratio Lower Upper
276 100
138 25.4 19.4 29.0
277 25.0 20.1 30.1



#88
Benzo(b)fluoranthene
Concen: 10.100 ng
RT: 22.880 min Scan# 3399
Delta R.T. -0.006 min
Lab File: BM047142.D
Acq: 10 Aug 2024 14:31

Tgt Ion:252 Resp: 1140578
Ion Ratio Lower Upper
252 100
253 21.1 17.8 26.6
125 9.7 7.4 11.0









#89
Benzo(k)fluoranthene
Concen: 10.108 ng
RT: 22.933 min Scan# 3410
Delta R.T. -0.012 min
Lab File: BM047142.D
Acq: 10 Aug 2024 14:31




Instrument :
BNA_M
ClientSampleId :
SSTDICC010


Tgt Ion:252 Resp: 1078126
Ion Ratio Lower Upper
252 100
253 21.9 17.7 26.5
125 10.0 7.4 11.2



#90
Benzo(a)pyrene
Concen: 9.865 ng
RT: 23.597 min Scan# 3521
Delta R.T. -0.012 min
Lab File: BM047142.D
Acq: 10 Aug 2024 14:31

Tgt Ion:252 Resp: 965036
Ion Ratio Lower Upper
252 100
253 21.4 17.1 25.7
125 11.5 8.2 12.2









#91
Dibenzo(a,h)anthracene
Concen: 10.051 ng
RT: 26.756 min Scan# 4062
Delta R.T. -0.024 min
Lab File: BM047142.D
Acq: 10 Aug 2024 14:31




Instrument :
BNA_M
ClientSampleId :
SSTDICC010

Tgt Ion:278 Resp: 999745
Ion Ratio Lower Upper
278 100
139 16.6 12.0 18.0
279 24.0 19.2 28.8



#92
Benzo(g,h,i)perylene
Concen: 9.964 ng
RT: 27.644 min Scan# 4209
Delta R.T. -0.018 min
Lab File: BM047142.D
Acq: 10 Aug 2024 14:31

Tgt Ion:276 Resp: 1029079
Ion Ratio Lower Upper
276 100
277 22.8 18.8 28.2
138 20.9 16.2 24.4



Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081024\
 Data File : BM047143.D
 Acq On : 10 Aug 2024 15:11
 Operator : RC/JU
 Sample : SSTDICC020
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC020

Quant Time: Aug 11 23:32:57 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Aug 11 23:24:37 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.375	152	264468	20.000	ng	0.00
21) Naphthalene-d8	10.128	136	1115607	20.000	ng	0.00
39) Acenaphthene-d10	14.034	164	747989	20.000	ng	0.00
64) Phenanthrene-d10	16.792	188	1595663	20.000	ng	0.00
76) Chrysene-d12	21.027	240	1449202	20.000	ng	0.00
86) Perylene-d12	23.727	264	1737656	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.022	112	633062	42.772	ng	0.00
7) Phenol-d6	6.581	99	873552	42.833	ng	0.00
23) Nitrobenzene-d5	8.522	82	936280	42.259	ng	0.00
42) 2,4,6-Tribromophenol	15.539	330	371649	42.851	ng	0.00
45) 2-Fluorobiphenyl	12.639	172	1998004	41.204	ng	0.00
79) Terphenyl-d14	19.457	244	2776457	41.052	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.034	88	131017	21.364	ng	99
3) Pyridine	3.411	79	385480	21.451	ng	99
4) n-Nitrosodimethylamine	3.328	42	166470	21.188	ng	# 94
6) Aniline	6.722	93	418736	21.723	ng	99
8) 2-Chlorophenol	6.952	128	350051	21.684	ng	99
9) Benzaldehyde	6.540	77	270923	21.095	ng	98
10) Phenol	6.610	94	434987	21.375	ng	99
11) bis(2-Chloroethyl)ether	6.822	93	372085	21.351	ng	99
12) 1,3-Dichlorobenzene	7.263	146	406387	21.185	ng	100
13) 1,4-Dichlorobenzene	7.410	146	413095	21.269	ng	98
14) 1,2-Dichlorobenzene	7.716	146	400406	21.850	ng	99
15) Benzyl Alcohol	7.622	79	314490	21.667	ng	99
16) 2,2'-oxybis(1-Chloropr...	7.904	45	491561	21.956	ng	99
17) 2-Methylphenol	7.834	107	294855	21.718	ng	99
18) Hexachloroethane	8.428	117	158678	21.091	ng	99
19) n-Nitroso-di-n-propyla...	8.181	70	300944	21.907	ng	97
20) 3+4-Methylphenols	8.157	107	415106	21.871	ng	97
22) Acetophenone	8.193	105	569010	21.130	ng	99
24) Nitrobenzene	8.563	77	477014	21.382	ng	97
25) Isophorone	9.087	82	816407	21.163	ng	99
26) 2-Nitrophenol	9.263	139	201069	20.351	ng	98
27) 2,4-Dimethylphenol	9.346	122	242158	20.924	ng	100
28) bis(2-Chloroethoxy)met...	9.575	93	511416	21.069	ng	99
29) 2,4-Dichlorophenol	9.793	162	354907	20.667	ng	98
30) 1,2,4-Trichlorobenzene	9.998	180	400292	20.598	ng	99
31) Naphthalene	10.181	128	1162115	20.911	ng	99
32) Benzoic acid	9.487	122	267804	19.806	ng	98
33) 4-Chloroaniline	10.304	127	455929	21.333	ng	100
34) Hexachlorobutadiene	10.463	225	230475	20.262	ng	99
35) Caprolactam	11.098	113	124776	20.923	ng	100
36) 4-Chloro-3-methylphenol	11.451	107	386762	21.148	ng	100
37) 2-Methylnaphthalene	11.804	142	828088	20.927	ng	100
38) 1-Methylnaphthalene	12.028	142	819798	20.963	ng	98
40) 1,2,4,5-Tetrachloroben...	12.187	216	414428	20.313	ng	99
41) Hexachlorocyclopentadiene	12.163	237	142951	20.066	ng	100
43) 2,4,6-Trichlorophenol	12.445	196	294494	20.309	ng	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081024\
 Data File : BM047143.D
 Acq On : 10 Aug 2024 15:11
 Operator : RC/JU
 Sample : SSTDICC020
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC020


Quant Time: Aug 11 23:32:57 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Aug 11 23:24:37 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	12.516	196	327981	20.371	ng	98
46) 1,1'-Biphenyl	12.851	154	1106359	20.776	ng	98
47) 2-Chloronaphthalene	12.887	162	879955	21.142	ng	99
48) 2-Nitroaniline	13.110	65	274398	21.044	ng	96
49) Acenaphthylene	13.745	152	1295508	21.117	ng	99
50) Dimethylphthalate	13.510	163	1098249	21.023	ng	100
51) 2,6-Dinitrotoluene	13.628	165	252319	20.662	ng	99
52) Acenaphthene	14.098	154	812631	20.882	ng	99
53) 3-Nitroaniline	13.957	138	257120	21.012	ng	98
54) 2,4-Dinitrophenol	14.175	184	138230	18.886	ng	99
55) Dibenzofuran	14.439	168	1292618	21.028	ng	98
56) 4-Nitrophenol	14.292	139	219007	20.755	ng	97
57) 2,4-Dinitrotoluene	14.422	165	343833	21.094	ng	96
58) Fluorene	15.092	166	1066342	21.024	ng	98
59) 2,3,4,6-Tetrachlorophenol	14.675	232	275845	20.846	ng	98
60) Diethylphthalate	14.892	149	1126281	21.113	ng	98
61) 4-Chlorophenyl-phenyle...	15.098	204	496242	20.551	ng	100
62) 4-Nitroaniline	15.128	138	258531m	21.445	ng	
63) Azobenzene	15.392	77	1090035	21.550	ng	99
65) 4,6-Dinitro-2-methylph...	15.192	198	191622	20.511	ng	97
66) n-Nitrosodiphenylamine	15.316	169	914838	21.049	ng	99
67) 4-Bromophenyl-phenylether	15.998	248	316488	20.583	ng	99
68) Hexachlorobenzene	16.098	284	382395	20.765	ng	99
69) Atrazine	16.286	200	290743	22.231	ng	99
70) Pentachlorophenol	16.451	266	263728	20.430	ng	99
71) Phenanthrene	16.833	178	1652375	20.634	ng	100
72) Anthracene	16.927	178	1637902	21.021	ng	100
73) Carbazole	17.204	167	1678200	21.094	ng	99
74) Di-n-butylphthalate	17.792	149	1951166	21.077	ng	99
75) Fluoranthene	18.868	202	2002438	20.558	ng	99
77) Benzidine	19.074	184	541138	22.019	ng	99
78) Pyrene	19.233	202	2091746	20.215	ng	99
80) Butylbenzylphthalate	20.174	149	869066	20.580	ng	98
81) Benzo(a)anthracene	21.010	228	1977415	20.554	ng	100
82) 3,3'-Dichlorobenzidine	20.957	252	636034	21.130	ng	98
83) Chrysene	21.068	228	1885561	20.657	ng	99
84) Bis(2-ethylhexyl)phtha...	20.974	149	1257178	20.992	ng	99
85) Di-n-octyl phthalate	22.009	149	2072238	20.358	ng	99
87) Indeno(1,2,3-cd)pyrene	26.715	276	2373808	21.135	ng	99
88) Benzo(b)fluoranthene	22.880	252	2123837	20.581	ng	99
89) Benzo(k)fluoranthene	22.939	252	2041826	20.948	ng	99
90) Benzo(a)pyrene	23.598	252	1851508	20.712	ng	99
91) Dibenzo(a,h)anthracene	26.768	278	1944551	21.393	ng	99
92) Benzo(g,h,i)perylene	27.644	276	2020018	21.405	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
BNA_M
ClientSampleId :
SSTDICC020





[illegible]



#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.375 min Scan# 763
Delta R.T. 0.000 min
Lab File: BM047143.D
Acq: 10 Aug 2024 15:11




Instrument :
BNA_M
ClientSampleId :
SSTDICC020


Tgt Ion:152 Resp: 264468
Ion Ratio Lower Upper
152 100
150 158.7 120.5 180.7
115 58.6 45.5 68.3



#2
1,4-Dioxane
Concen: 21.364 ng
RT: 3.034 min Scan# 25
Delta R.T. 0.000 min
Lab File: BM047143.D
Acq: 10 Aug 2024 15:11

Tgt Ion: 88 Resp: 131017
Ion Ratio Lower Upper
88 100
58 72.4 58.2 87.4
43 29.1 22.5 33.7









#3
Pyridine
Concen: 21.451 ng
RT: 3.411 min Scan# 89
Delta R.T. 0.000 min
Lab File: BM047143.D
Acq: 10 Aug 2024 15:11




Instrument :
BNA_M
ClientSampleId :
SSTDICC020


Tgt Ion: 79 Resp: 385480
Ion Ratio Lower Upper
79 100
52 64.5 51.2 76.8
51 33.3 26.1 39.1



#4
n-Nitrosodimethylamine
Concen: 21.188 ng
RT: 3.328 min Scan# 75
Delta R.T. 0.000 min
Lab File: BM047143.D
Acq: 10 Aug 2024 15:11

Tgt Ion: 42 Resp: 166470
Ion Ratio Lower Upper
42 100
74 134.1 112.9 169.3
44 8.5 5.3 7.9









#5
2-Fluorophenol
Concen: 42.772 ng
RT: 5.022 min Scan# 363
Delta R.T. 0.000 min
Lab File: BM047143.D
Acq: 10 Aug 2024 15:11




Instrument :
BNA_M
ClientSampleId :
SSTDICC020


Tgt Ion:112 Resp: 633062
Ion Ratio Lower Upper
112 100
64 58.4 46.7 70.1
63 31.0 24.8 37.2



#6
Aniline
Concen: 21.723 ng
RT: 6.722 min Scan# 652
Delta R.T. -0.006 min
Lab File: BM047143.D
Acq: 10 Aug 2024 15:11

Tgt Ion: 93 Resp: 418736
Ion Ratio Lower Upper
93 100
66 39.8 31.4 47.2
65 20.7 15.6 23.4









#7
Phenol-d6
Concen: 42.833 ng
RT: 6.581 min Scan# 61
Delta R.T. -0.006 min
Lab File: BM047143.D
Acq: 10 Aug 2024 15:11




Instrument :
BNA_M
ClientSampleId :
SSTDICC020


Tgt Ion: 99 Resp: 873552
Ion Ratio Lower Upper
99 100
42 18.0 14.1 21.1
71 35.0 28.5 42.7



#8
2-Chlorophenol
Concen: 21.684 ng
RT: 6.952 min Scan# 691
Delta R.T. 0.000 min
Lab File: BM047143.D
Acq: 10 Aug 2024 15:11

Tgt Ion: 128 Resp: 350051
Ion Ratio Lower Upper
128 100
130 32.7 12.1 52.1
64 47.1 26.6 66.6









#9
Benzaldehyde
Concen: 21.095 ng
RT: 6.540 min Scan# 61
Delta R.T. -0.000 min
Lab File: BM047143.D
Acq: 10 Aug 2024 15:11




Instrument :
BNA_M
ClientSampleId :
SSTDICC020


Tgt Ion: 77 Resp: 270923
Ion Ratio Lower Upper
77 100
105 93.5 74.5 114.5
106 89.7 72.3 112.3



#10
Phenol
Concen: 21.375 ng
RT: 6.610 min Scan# 633
Delta R.T. 0.000 min
Lab File: BM047143.D
Acq: 10 Aug 2024 15:11

Tgt Ion: 94 Resp: 434987
Ion Ratio Lower Upper
94 100
65 29.5 9.0 49.0
66 39.4 18.9 58.9









#11
bis(2-chloroethyl)ether
Concen: 21.351 ng
RT: 6.822 min Scan# 669
Delta R.T. -0.006 min
Lab File: BM047143.D
Acq: 10 Aug 2024 15:11




Instrument :
BNA_M
ClientSampleId :
SSTDICC020


Tgt Ion: 93 Resp: 372085
Ion Ratio Lower Upper
93 100
63 77.4 55.8 95.8
95 32.5 12.7 52.7



#12
1,3-Dichlorobenzene
Concen: 21.185 ng
RT: 7.263 min Scan# 744
Delta R.T. 0.000 min
Lab File: BM047143.D
Acq: 10 Aug 2024 15:11

Tgt Ion: 146 Resp: 406387
Ion Ratio Lower Upper
146 100
148 65.4 52.3 78.5
75 32.2 25.8 38.6





#13

1,4-Dichlorobenzene

Concen: 21.269 ng

RT: 7.410 min Scan# 7

Delta R.T. 0.000 min

Lab File: BM047143.D

Acq: 10 Aug 2024 15:11

Instrument :

BNA_M

ClientSampleId :

SSTDICC020





Tgt Ion:146 Resp: 413095

Ion Ratio Lower Upper

146 100

148 66.8 51.4 77.2

111 40.0 31.5 47.3



#14

1,2-Dichlorobenzene

Concen: 21.850 ng

RT: 7.716 min Scan# 821

Delta R.T. 0.000 min

Lab File: BM047143.D

Acq: 10 Aug 2024 15:11




Tgt Ion:146 Resp: 400406


Ion Ratio Lower Upper

146 100

148 65.5 51.8 77.8

111 40.3 33.3 49.9









#15
Benzyl Alcohol
Concen: 21.667 ng
RT: 7.622 min Scan# 805
Delta R.T. -0.006 min
Lab File: BM047143.D
Acq: 10 Aug 2024 15:11




Instrument :
BNA_M
ClientSampleId :
SSTDICC020


Tgt Ion: 79 Resp: 314490
Ion Ratio Lower Upper
79 100
108 73.2 59.7 89.5
77 66.6 52.9 79.3



#16
2,2'-oxybis(1-chloropropane)
Concen: 21.956 ng
RT: 7.904 min Scan# 853
Delta R.T. -0.012 min
Lab File: BM047143.D
Acq: 10 Aug 2024 15:11

Tgt Ion: 45 Resp: 491561
Ion Ratio Lower Upper
45 100
77 15.4 0.0 35.6
79 11.6 0.0 32.4





#17

2-Methylphenol

Concen: 21.718 ng

RT: 7.834 min Scan# 841

Delta R.T. 0.000 min

Lab File: BM047143.D

Acq: 10 Aug 2024 15:11

Instrument :

BNA_M

ClientSampleId :

SSTDICC020

Tgt Ion:107 Resp: 294855





Ion Ratio Lower Upper

107 100

108 113.3 90.6 135.8

77 50.2 39.3 58.9

79 46.9 38.6 58.0



#18

Hexachloroethane

Concen: 21.091 ng

RT: 8.428 min Scan# 942

Delta R.T. 0.000 min

Lab File: BM047143.D

Acq: 10 Aug 2024 15:11




Tgt Ion:117 Resp: 158678


Ion Ratio Lower Upper

117 100

119 95.2 77.0 115.4

201 99.9 81.3 121.9





#19

n-Nitroso-di-n-propylamine

Concen: 21.907 ng

RT: 8.181 min Scan# 900

Delta R.T. -0.006 min

Lab File: BM047143.D

Acq: 10 Aug 2024 15:11

Instrument :

BNA_M

ClientSampleId :

SSTDICC020

Tgt Ion: 70 Resp: 300944





Ion Ratio Lower Upper

70 100

42 55.0 42.0 63.0

101 9.2 7.2 10.8

130 19.1 15.5 23.3



#20

3+4-Methylphenols

Concen: 21.871 ng

RT: 8.157 min Scan# 896

Delta R.T. -0.006 min

Lab File: BM047143.D

Acq: 10 Aug 2024 15:11

Tgt Ion: 107 Resp: 415106




Ion Ratio Lower Upper


107 100

108 86.6 70.7 110.7

77 36.1 17.4 57.4

79 30.2 10.9 50.9









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.128 min Scan# 1115607
Delta R.T. -0.006 min
Lab File: BM047143.D
Acq: 10 Aug 2024 15:11




Instrument :
BNA_M
ClientSampleId :
SSTDICC020


Tgt Ion:136 Resp: 1115607
Ion Ratio Lower Upper
136 100
137 10.7 8.8 13.2
54 7.8 6.0 9.0
68 6.3 4.8 7.2



#22
Acetophenone
Concen: 21.130 ng
RT: 8.193 min Scan# 902
Delta R.T. -0.006 min
Lab File: BM047143.D
Acq: 10 Aug 2024 15:11

Tgt Ion:105 Resp: 569010
Ion Ratio Lower Upper
105 100
71 5.8 4.5 6.7
51 26.4 20.1 30.1
120 22.3 17.8 26.6





#23

Nitrobenzene-d5

Concen: 42.259 ng

RT: 8.522 min Scan# 958

Delta R.T. 0.000 min

Lab File: BM047143.D

Acq: 10 Aug 2024 15:11

Instrument :

BNA_M

ClientSampleId :

SSTDICC020





Tgt Ion: 82 Resp: 936280

Ion Ratio Lower Upper

82 100

128 41.7 32.3 48.5

54 46.9 37.4 56.2



#24

Nitrobenzene

Concen: 21.382 ng

RT: 8.563 min Scan# 965

Delta R.T. 0.000 min

Lab File: BM047143.D

Acq: 10 Aug 2024 15:11




Tgt Ion: 77 Resp: 477014


Ion Ratio Lower Upper

77 100

123 44.7 34.0 51.0

65 13.9 10.5 15.7









#25
Isophorone
Concen: 21.163 ng
RT: 9.087 min Scan# 1054
Delta R.T. -0.006 min
Lab File: BM047143.D
Acq: 10 Aug 2024 15:11




Instrument :
BNA_M
ClientSampleId :
SSTDICC020


Tgt Ion: 82 Resp: 816407
Ion Ratio Lower Upper
82 100
95 6.9 5.5 8.3
138 17.3 13.5 20.3



#26
2-Nitrophenol
Concen: 20.351 ng
RT: 9.263 min Scan# 1084
Delta R.T. 0.000 min
Lab File: BM047143.D
Acq: 10 Aug 2024 15:11

Tgt Ion:139 Resp: 201069
Ion Ratio Lower Upper
139 100
109 31.3 24.4 36.6
65 52.2 43.0 64.6





#27

2,4-Dimethylphenol

Concen: 20.924 ng

RT: 9.346 min Scan# 1098

Delta R.T. 0.000 min

Lab File: BM047143.D

Acq: 10 Aug 2024 15:11

Instrument :

BNA_M

ClientSampleId :

SSTDICC020





Tgt Ion:122 Resp: 242158

Ion Ratio Lower Upper

122 100

107 116.5 92.8 139.2

121 57.7 46.1 69.1



#28

bis(2-Chloroethoxy)methane

Concen: 21.069 ng

RT: 9.575 min Scan# 1137

Delta R.T. 0.000 min

Lab File: BM047143.D

Acq: 10 Aug 2024 15:11




Tgt Ion: 93 Resp: 511416


Ion Ratio Lower Upper

93 100

95 33.1 25.8 38.8

123 11.3 9.0 13.4





#29

2,4-Dichlorophenol

Concen: 20.667 ng

RT: 9.793 min Scan# 1174

Delta R.T. -0.006 min

Lab File: BM047143.D

Acq: 10 Aug 2024 15:11

Instrument :

BNA_M

ClientSampleId :

SSTDICC020





Tgt Ion:162 Resp: 354907

Ion Ratio Lower Upper

162 100

164 65.7 44.7 84.7

98 36.5 14.9 54.9



#30

1,2,4-Trichlorobenzene

Concen: 20.598 ng

RT: 9.998 min Scan# 1209

Delta R.T. 0.000 min

Lab File: BM047143.D

Acq: 10 Aug 2024 15:11




Tgt Ion:180 Resp: 400292


Ion Ratio Lower Upper

180 100

182 96.8 76.6 115.0

145 29.6 23.4 35.2









#31
Naphthalene
Concen: 20.911 ng
RT: 10.181 min Scan# 1122
Delta R.T. 0.000 min
Lab File: BM047143.D
Acq: 10 Aug 2024 15:11




Instrument :
BNA_M
ClientSampleId :
SSTDICC020


Tgt Ion:128 Resp: 1162115
Ion Ratio Lower Upper
128 100
129 10.7 8.6 13.0
127 13.3 10.2 15.4



#32
Benzoic acid
Concen: 19.806 ng
RT: 9.487 min Scan# 1122
Delta R.T. -0.047 min
Lab File: BM047143.D
Acq: 10 Aug 2024 15:11

Tgt Ion:122 Resp: 267804
Ion Ratio Lower Upper
122 100
105 125.1 107.1 147.1
77 99.1 82.1 122.1









#33
4-Chloroaniline
Concen: 21.333 ng
RT: 10.304 min Scan# 1261
Delta R.T. -0.006 min
Lab File: BM047143.D
Acq: 10 Aug 2024 15:11




Instrument :
BNA_M
ClientSampleId :
SSTDICC020


Tgt Ion:	127	Resp:	455929
Ion	Ratio	Lower	Upper
127	100		
129	32.1	25.8	38.8
65	31.1	24.9	37.3
92	19.0	15.1	22.7



#34
Hexachlorobutadiene
Concen: 20.262 ng
RT: 10.463 min Scan# 1288
Delta R.T. -0.006 min
Lab File: BM047143.D
Acq: 10 Aug 2024 15:11

Tgt Ion:	225	Resp:	230475
Ion	Ratio	Lower	Upper
225	100		
223	62.5	50.2	75.2
227	63.0	51.4	77.0









#35
Caprolactam
Concen: 20.923 ng
RT: 11.098 min Scan# 11
Delta R.T. -0.023 min
Lab File: BM047143.D
Acq: 10 Aug 2024 15:11




Instrument :
BNA_M
ClientSampleId :
SSTDICC020


Tgt Ion:113 Resp: 124776
Ion Ratio Lower Upper
113 100
55 160.8 140.5 180.5
56 123.3 103.6 143.6



#36
4-Chloro-3-methylphenol
Concen: 21.148 ng
RT: 11.451 min Scan# 1456
Delta R.T. 0.000 min
Lab File: BM047143.D
Acq: 10 Aug 2024 15:11

Tgt Ion:107 Resp: 386762
Ion Ratio Lower Upper
107 100
144 26.7 21.8 32.6
142 83.5 66.8 100.2









#37
2-Methylnaphthalene
Concen: 20.927 ng
RT: 11.804 min Scan# 1516
Delta R.T. 0.000 min
Lab File: BM047143.D
Acq: 10 Aug 2024 15:11




Instrument :
BNA_M
ClientSampleId :
SSTDICC020


Tgt Ion:142 Resp: 828088
Ion Ratio Lower Upper
142 100
141 88.6 70.9 106.3
115 35.5 28.7 43.1



#38
1-Methylnaphthalene
Concen: 20.963 ng
RT: 12.028 min Scan# 1554
Delta R.T. -0.006 min
Lab File: BM047143.D
Acq: 10 Aug 2024 15:11

Tgt Ion:142 Resp: 819798
Ion Ratio Lower Upper
142 100
141 91.7 74.6 111.8
116 3.9 3.1 4.7









#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.034 min Scan# 1895
Delta R.T. 0.000 min
Lab File: BM047143.D
Acq: 10 Aug 2024 15:11




Instrument :
BNA_M
ClientSampleId :
SSTDICC020


Tgt Ion:164 Resp: 747989
Ion Ratio Lower Upper
164 100
162 100.6 80.7 121.1
160 44.4 36.0 54.0



#40
1,2,4,5-Tetrachlorobenzene
Concen: 20.313 ng
RT: 12.187 min Scan# 1581
Delta R.T. 0.000 min
Lab File: BM047143.D
Acq: 10 Aug 2024 15:11

Tgt Ion:216 Resp: 414428
Ion Ratio Lower Upper
216 100
214 79.1 63.4 95.2
179 22.2 17.6 26.4
108 19.9 15.1 22.7









#41
Hexachlorocyclopentadiene
Concen: 20.066 ng
RT: 12.163 min Scan# 1577
Delta R.T. 0.000 min
Lab File: BM047143.D
Acq: 10 Aug 2024 15:11




Instrument :
BNA_M
ClientSampleId :
SSTDICC020


Tgt Ion	Ratio	Lower	Upper
237	100		
235	65.1	44.8	84.8
272	13.7	0.0	33.5



#42
2,4,6-Tribromophenol
Concen: 42.851 ng
RT: 15.539 min Scan# 2151
Delta R.T. 0.000 min
Lab File: BM047143.D
Acq: 10 Aug 2024 15:11

Tgt Ion	Ratio	Lower	Upper
330	100		
332	96.5	77.4	116.2
141	40.3	31.6	47.4









#43
2,4,6-Trichlorophenol
Concen: 20.309 ng
RT: 12.445 min Scan# 1625
Delta R.T. 0.000 min
Lab File: BM047143.D
Acq: 10 Aug 2024 15:11




Instrument :
BNA_M
ClientSampleId :
SSTDICC020


Tgt Ion:196 Resp: 294494
Ion Ratio Lower Upper
196 100
198 95.7 75.6 113.4
200 31.2 24.6 37.0



#44
2,4,5-Trichlorophenol
Concen: 20.371 ng
RT: 12.516 min Scan# 1637
Delta R.T. -0.006 min
Lab File: BM047143.D
Acq: 10 Aug 2024 15:11

Tgt Ion:196 Resp: 327981
Ion Ratio Lower Upper
196 100
198 94.6 77.7 116.5
97 50.6 39.8 59.8
132 25.2 20.7 31.1





#45

2-Fluorobiphenyl

Concen: 41.204 ng

RT: 12.639 min Scan# 1658

Delta R.T. -0.006 min

Lab File: BM047143.D

Acq: 10 Aug 2024 15:11

Instrument :

BNA_M

ClientSampleId :

SSTDICC020





Tgt Ion:172 Resp: 1998004

Ion Ratio Lower Upper

172 100

171 34.9 27.8 41.6

170 23.4 18.7 28.1



#46

1,1'-Biphenyl

Concen: 20.776 ng

RT: 12.851 min Scan# 1694

Delta R.T. -0.006 min

Lab File: BM047143.D

Acq: 10 Aug 2024 15:11




Tgt Ion:154 Resp: 1106359


Ion Ratio Lower Upper

154 100

153 40.0 19.0 59.0

76 13.0 0.0 31.9





#47

2-Chloronaphthalene

Concen: 21.142 ng

RT: 12.887 min Scan# 1

Delta R.T. 0.000 min

Lab File: BM047143.D

Acq: 10 Aug 2024 15:11

Instrument :

BNA_M

ClientSampleId :

SSTDICC020





Tgt Ion:162 Resp: 879955

Ion Ratio Lower Upper

162 100

127 36.1 30.0 45.0

164 32.6 26.3 39.5



#48

2-Nitroaniline

Concen: 21.044 ng

RT: 13.110 min Scan# 1738

Delta R.T. -0.006 min

Lab File: BM047143.D

Acq: 10 Aug 2024 15:11




Tgt Ion: 65 Resp: 274398


Ion Ratio Lower Upper

65 100

92 68.0 55.1 82.7

138 102.1 85.9 128.9









#49
Acenaphthylene
Concen: 21.117 ng
RT: 13.745 min Scan# 1846
Delta R.T. -0.006 min
Lab File: BM047143.D
Acq: 10 Aug 2024 15:11




Instrument :
BNA_M
ClientSampleId :
SSTDICC020


Tgt Ion:152 Resp: 1295508
Ion Ratio Lower Upper
152 100
151 20.0 15.8 23.6
153 13.5 10.4 15.6



#50
Dimethylphthalate
Concen: 21.023 ng
RT: 13.510 min Scan# 1806
Delta R.T. -0.006 min
Lab File: BM047143.D
Acq: 10 Aug 2024 15:11

Tgt Ion:163 Resp: 1098249
Ion Ratio Lower Upper
163 100
194 3.6 3.0 4.6
164 10.2 8.0 12.0









#51
2,6-Dinitrotoluene
Concen: 20.662 ng
RT: 13.628 min Scan# 1826
Delta R.T. 0.000 min
Lab File: BM047143.D
Acq: 10 Aug 2024 15:11




Instrument :
BNA_M
ClientSampleId :
SSTDICC020


Tgt Ion:165 Resp: 252319
Ion Ratio Lower Upper
165 100
63 57.9 46.5 69.7
89 60.7 47.3 70.9



#52
Acenaphthene
Concen: 20.882 ng
RT: 14.098 min Scan# 1906
Delta R.T. 0.000 min
Lab File: BM047143.D
Acq: 10 Aug 2024 15:11

Tgt Ion:154 Resp: 812631
Ion Ratio Lower Upper
154 100
153 111.7 90.3 135.5
152 53.5 42.9 64.3









#53
3-Nitroaniline
Concen: 21.012 ng
RT: 13.957 min Scan# 1882
Delta R.T. -0.006 min
Lab File: BM047143.D
Acq: 10 Aug 2024 15:11




Instrument :
BNA_M
ClientSampleId :
SSTDICC020


Tgt Ion:138 Resp: 257120
Ion Ratio Lower Upper
138 100
108 10.0 7.9 11.9
92 117.8 92.0 138.0



#54
2,4-Dinitrophenol
Concen: 18.886 ng
RT: 14.175 min Scan# 1919
Delta R.T. -0.000 min
Lab File: BM047143.D
Acq: 10 Aug 2024 15:11

Tgt Ion:184 Resp: 138230
Ion Ratio Lower Upper
184 100
63 64.9 52.7 79.1
154 63.5 50.4 75.6





#55

Dibenzofuran

Concen: 21.028 ng

RT: 14.439 min Scan# 1964

Delta R.T. 0.000 min

Lab File: BM047143.D

Acq: 10 Aug 2024 15:11

Instrument :

BNA_M

ClientSampleId :

SSTDICC020





Tgt Ion:168 Resp: 1292618

Ion Ratio Lower Upper

168 100

139 35.8 29.6 44.4

169 13.0 10.7 16.1



#56

4-Nitrophenol

Concen: 20.755 ng

RT: 14.292 min Scan# 1939

Delta R.T. -0.006 min

Lab File: BM047143.D

Acq: 10 Aug 2024 15:11




Tgt Ion:139 Resp: 219007


Ion Ratio Lower Upper

139 100

109 74.0 52.5 92.5

65 98.5 75.4 115.4





#57

2,4-Dinitrotoluene

Concen: 21.094 ng

RT: 14.422 min Scan# 1962

Delta R.T. -0.006 min

Lab File: BM047143.D

Acq: 10 Aug 2024 15:11

Instrument :

BNA_M

ClientSampleId :

SSTDICC020

Tgt Ion:165 Resp: 343833





Ion Ratio Lower Upper

165 100

63 52.0 40.3 60.5

89 81.1 61.0 91.6

182 2.7 1.9 2.9



#58

Fluorene

Concen: 21.024 ng

RT: 15.092 min Scan# 2075

Delta R.T. 0.000 min

Lab File: BM047143.D

Acq: 10 Aug 2024 15:11




Tgt Ion:166 Resp: 1066342


Ion Ratio Lower Upper

166 100

165 98.4 77.4 116.0


167 13.4 11.0 16.6








#59
2,3,4,6-Tetrachlorophenol
Concen: 20.846 ng
RT: 14.675 min Scan# 2004
Delta R.T. 0.000 min
Lab File: BM047143.D
Acq: 10 Aug 2024 15:11


Instrument :
BNA_M
ClientSampleId :
SSTDICC020





Tgt Ion:232 Resp: 275845
Ion Ratio Lower Upper
232 100
131 47.9 36.9 55.3
130 2.5 2.0 3.0
166 32.4 26.4 39.6




#60
Diethylphthalate
Concen: 21.113 ng
RT: 14.892 min Scan# 2041
Delta R.T. -0.006 min
Lab File: BM047143.D
Acq: 10 Aug 2024 15:11



Tgt Ion:149 Resp: 1126281
Ion Ratio Lower Upper
149 100
177 21.0 17.6 26.4
150 12.0 9.8 14.6









#61
4-Chlorophenyl-phenylether
Concen: 20.551 ng
RT: 15.098 min Scan# 2076
Delta R.T. 0.000 min
Lab File: BM047143.D
Acq: 10 Aug 2024 15:11




Instrument :
BNA_M
ClientSampleId :
SSTDICC020


Tgt Ion:204 Resp: 496242
Ion Ratio Lower Upper
204 100
206 32.8 26.2 39.4
141 59.5 47.2 70.8



#62
4-Nitroaniline
Concen: 21.445 ng m
RT: 15.128 min Scan# 2081
Delta R.T. -0.012 min
Lab File: BM047143.D
Acq: 10 Aug 2024 15:11

Tgt Ion:138 Resp: 258531
Ion Ratio Lower Upper
138 100
92 54.4 31.9 71.9
108 85.1 63.8 103.8









#63
Azobenzene
Concen: 21.550 ng
RT: 15.392 min Scan# 2126
Delta R.T. 0.000 min
Lab File: BM047143.D
Acq: 10 Aug 2024 15:11

Instrument :
BNA_M
ClientSampleId :
SSTDICC020

Tgt Ion: 77 Resp: 1090035




Ion	Ratio	Lower	Upper
77	100		
182	24.2	3.8	43.8
105	19.2	0.0	39.5
51	27.3	6.8	46.8




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 16.792 min Scan# 2364
Delta R.T. 0.000 min
Lab File: BM047143.D
Acq: 10 Aug 2024 15:11

Tgt Ion: 188 Resp: 1595663

Ion	Ratio	Lower	Upper
188	100		
94	8.8	7.0	10.6
80	9.6	8.0	12.0





#65

4,6-Dinitro-2-methylphenol

Concen: 20.511 ng

RT: 15.192 min Scan# 2092

Delta R.T. -0.006 min

Lab File: BM047143.D

Acq: 10 Aug 2024 15:11

Instrument :

BNA_M

ClientSampleId :

SSTDICC020





Tgt Ion:198 Resp: 191622

Ion Ratio Lower Upper

198 100

51 46.1 23.9 63.9

105 45.5 24.3 64.3



#66

n-Nitrosodiphenylamine

Concen: 21.049 ng

RT: 15.316 min Scan# 2113

Delta R.T. -0.006 min

Lab File: BM047143.D

Acq: 10 Aug 2024 15:11




Tgt Ion:169 Resp: 914838


Ion Ratio Lower Upper

169 100

168 68.0 53.8 80.8

167 36.4 29.3 43.9









#67
4-Bromophenyl-phenylether
Concen: 20.583 ng
RT: 15.998 min Scan# 211
Delta R.T. 0.000 min
Lab File: BM047143.D
Acq: 10 Aug 2024 15:11




Instrument :
BNA_M
ClientSampleId :
SSTDICC020


Tgt Ion:248 Resp: 316488
Ion Ratio Lower Upper
248 100
250 97.0 77.8 116.8
141 70.4 57.7 86.5



#68
Hexachlorobenzene
Concen: 20.765 ng
RT: 16.098 min Scan# 2246
Delta R.T. 0.000 min
Lab File: BM047143.D
Acq: 10 Aug 2024 15:11

Tgt Ion:284 Resp: 382395
Ion Ratio Lower Upper
284 100
142 36.3 29.4 44.0
249 30.4 25.1 37.7





#69

Atrazine

Concen: 22.231 ng

RT: 16.286 min Scan# 21

Delta R.T. -0.006 min

Lab File: BM047143.D

Acq: 10 Aug 2024 15:11

Instrument :

BNA_M

ClientSampleId :

SSTDICC020





Tgt Ion:200 Resp: 290743

Ion Ratio Lower Upper

200 100

173 27.0 6.1 46.1

215 46.1 26.8 66.8



#70

Pentachlorophenol

Concen: 20.430 ng

RT: 16.451 min Scan# 2306

Delta R.T. 0.000 min

Lab File: BM047143.D

Acq: 10 Aug 2024 15:11




Tgt Ion:266 Resp: 263728


Ion Ratio Lower Upper

266 100

268 63.1 49.8 74.6

264 60.8 49.6 74.4









#71
Phenanthrene
Concen: 20.634 ng
RT: 16.833 min Scan# 21
Delta R.T. -0.006 min
Lab File: BM047143.D
Acq: 10 Aug 2024 15:11




Instrument :
BNA_M
ClientSampleId :
SSTDICC020


Tgt Ion:178 Resp: 1652375
Ion Ratio Lower Upper
178 100
176 18.9 15.3 22.9
179 15.2 12.4 18.6



#72
Anthracene
Concen: 21.021 ng
RT: 16.927 min Scan# 2387
Delta R.T. 0.000 min
Lab File: BM047143.D
Acq: 10 Aug 2024 15:11

Tgt Ion:178 Resp: 1637902
Ion Ratio Lower Upper
178 100
176 18.6 14.8 22.2
179 15.1 12.4 18.6









#73
Carbazole
Concen: 21.094 ng
RT: 17.204 min Scan# 2435
Delta R.T. -0.006 min
Lab File: BM047143.D
Acq: 10 Aug 2024 15:11




Instrument :
BNA_M
ClientSampleId :
SSTDICC020


Tgt Ion:167 Resp: 1678200
Ion Ratio Lower Upper
167 100
166 21.0 17.1 25.7
139 12.2 10.0 15.0



#74
Di-n-butylphthalate
Concen: 21.077 ng
RT: 17.792 min Scan# 2534
Delta R.T. 0.000 min
Lab File: BM047143.D
Acq: 10 Aug 2024 15:11

Tgt Ion:149 Resp: 1951166
Ion Ratio Lower Upper
149 100
150 9.1 7.5 11.3
104 5.3 4.1 6.1









#75
Fluoranthene
Concen: 20.558 ng
RT: 18.868 min Scan# 21
Delta R.T. 0.000 min
Lab File: BM047143.D
Acq: 10 Aug 2024 15:11




Instrument :
BNA_M
ClientSampleId :
SSTDICC020


Tgt Ion:202 Resp: 2002438
Ion Ratio Lower Upper
202 100
101 10.4 0.0 30.2
203 17.1 0.0 37.6



#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.027 min Scan# 3084
Delta R.T. -0.006 min
Lab File: BM047143.D
Acq: 10 Aug 2024 15:11

Tgt Ion:240 Resp: 1449202
Ion Ratio Lower Upper
240 100
120 10.7 8.2 12.2
236 25.8 20.6 30.8









#77
Benzidine
Concen: 22.019 ng
RT: 19.074 min Scan# 21
Delta R.T. 0.000 min
Lab File: BM047143.D
Acq: 10 Aug 2024 15:11




Instrument :
BNA_M
ClientSampleId :
SSTDICC020


Tgt Ion:184 Resp: 541138
Ion Ratio Lower Upper
184 100
185 14.9 11.8 17.6
183 10.8 9.0 13.6



#78
Pyrene
Concen: 20.215 ng
RT: 19.233 min Scan# 2779
Delta R.T. -0.006 min
Lab File: BM047143.D
Acq: 10 Aug 2024 15:11

Tgt Ion:202 Resp: 2091746
Ion Ratio Lower Upper
202 100
200 20.7 17.2 25.8
203 17.2 14.2 21.2









#79
Terphenyl-d14
Concen: 41.052 ng
RT: 19.457 min Scan# 2817
Delta R.T. -0.006 min
Lab File: BM047143.D
Acq: 10 Aug 2024 15:11




Instrument :
BNA_M
ClientSampleId :
SSTDICC020


Tgt Ion:244 Resp: 2776457
Ion Ratio Lower Upper
244 100
212 7.5 6.2 9.4
122 10.5 7.1 10.7



#80
Butylbenzylphthalate
Concen: 20.580 ng
RT: 20.174 min Scan# 2939
Delta R.T. 0.000 min
Lab File: BM047143.D
Acq: 10 Aug 2024 15:11

Tgt Ion:149 Resp: 869066
Ion Ratio Lower Upper
149 100
91 69.6 57.0 85.4
206 19.0 16.4 24.6









#81
Benzo(a)anthracene
Concen: 20.554 ng
RT: 21.010 min Scan# 3082
Delta R.T. -0.006 min
Lab File: BM047143.D
Acq: 10 Aug 2024 15:11




Instrument :
BNA_M
ClientSampleId :
SSTDICC020


Tgt Ion:228 Resp: 1977415
Ion Ratio Lower Upper
228 100
226 27.4 21.8 32.6
229 19.8 15.8 23.8



#82
3,3'-Dichlorobenzidine
Concen: 21.130 ng
RT: 20.957 min Scan# 3072
Delta R.T. 0.000 min
Lab File: BM047143.D
Acq: 10 Aug 2024 15:11

Tgt Ion:252 Resp: 636034
Ion Ratio Lower Upper
252 100
254 65.5 51.0 76.6
126 10.4 8.5 12.7









#83
Chrysene
Concen: 20.657 ng
RT: 21.068 min Scan# 3091
Delta R.T. -0.006 min
Lab File: BM047143.D
Acq: 10 Aug 2024 15:11




Instrument :
BNA_M
ClientSampleId :
SSTDICC020


Tgt Ion:228 Resp: 1885561
Ion Ratio Lower Upper
228 100
226 29.6 24.4 36.6
229 19.6 15.7 23.5



#84
Bis(2-ethylhexyl)phthalate
Concen: 20.992 ng
RT: 20.974 min Scan# 3075
Delta R.T. 0.000 min
Lab File: BM047143.D
Acq: 10 Aug 2024 15:11

Tgt Ion:149 Resp: 1257178
Ion Ratio Lower Upper
149 100
167 26.9 22.2 33.2
279 3.8 3.0 4.4









#85
Di-n-octyl phthalate
Concen: 20.358 ng
RT: 22.009 min Scan# 31
Delta R.T. 0.000 min
Lab File: BM047143.D
Acq: 10 Aug 2024 15:11




Instrument :
BNA_M
ClientSampleId :
SSTDICC020


Tgt Ion:149 Resp: 2072238
Ion Ratio Lower Upper
149 100
167 1.6 1.3 1.9
43 9.1 6.9 10.3



#86
Perylene-d12
Concen: 20.000 ng
RT: 23.727 min Scan# 3543
Delta R.T. 0.000 min
Lab File: BM047143.D
Acq: 10 Aug 2024 15:11

Tgt Ion:264 Resp: 1737656
Ion Ratio Lower Upper
264 100
260 23.0 18.5 27.7
265 21.4 17.3 25.9









#87
Indeno(1,2,3-cd)pyrene
Concen: 21.135 ng
RT: 26.715 min Scan# 4051
Delta R.T. -0.012 min
Lab File: BM047143.D
Acq: 10 Aug 2024 15:11




Instrument :
BNA_M
ClientSampleId :
SSTDICC020


Tgt Ion:276 Resp: 2373808
Ion Ratio Lower Upper
276 100
138 24.9 19.4 29.0
277 25.0 20.1 30.1



#88
Benzo(b)fluoranthene
Concen: 20.581 ng
RT: 22.880 min Scan# 3399
Delta R.T. -0.006 min
Lab File: BM047143.D
Acq: 10 Aug 2024 15:11

Tgt Ion:252 Resp: 2123837
Ion Ratio Lower Upper
252 100
253 21.5 17.8 26.6
125 9.5 7.4 11.0





#89

Benzo(k)fluoranthene

Concen: 20.948 ng

RT: 22.939 min Scan# 3410

Delta R.T. -0.006 min

Lab File: BM047143.D

Acq: 10 Aug 2024 15:11

Instrument :

BNA_M

ClientSampleId :

SSTDICC020





Tgt Ion:252 Resp: 2041826

Ion Ratio Lower Upper

252 100

253 21.7 17.7 26.5

125 9.6 7.4 11.2



#90

Benzo(a)pyrene

Concen: 20.712 ng

RT: 23.598 min Scan# 3521

Delta R.T. -0.012 min

Lab File: BM047143.D

Acq: 10 Aug 2024 15:11




Tgt Ion:252 Resp: 1851508


Ion Ratio Lower Upper

252 100

253 21.8 17.1 25.7

125 10.7 8.2 12.2









#91
Dibenzo(a,h)anthracene
Concen: 21.393 ng
RT: 26.768 min Scan# 4062
Delta R.T. -0.012 min
Lab File: BM047143.D
Acq: 10 Aug 2024 15:11




Instrument :
BNA_M
ClientSampleId :
SSTDICC020

Tgt Ion:278 Resp: 1944551
Ion Ratio Lower Upper
278 100
139 15.2 12.0 18.0
279 23.4 19.2 28.8



#92
Benzo(g,h,i)perylene
Concen: 21.405 ng
RT: 27.644 min Scan# 4209
Delta R.T. -0.018 min
Lab File: BM047143.D
Acq: 10 Aug 2024 15:11

Tgt Ion:276 Resp: 2020018
Ion Ratio Lower Upper
276 100
277 23.6 18.8 28.2
138 20.9 16.2 24.4



Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081024\
 Data File : BM047144.D
 Acq On : 10 Aug 2024 15:51
 Operator : RC/JU
 Sample : SSTDICCC040
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICCC040

Quant Time: Aug 11 23:33:42 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Aug 11 23:24:37 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.375	152	343346	20.000	ng	0.00
21) Naphthalene-d8	10.134	136	1387685	20.000	ng	0.00
39) Acenaphthene-d10	14.033	164	915049	20.000	ng	0.00
64) Phenanthrene-d10	16.792	188	1901562	20.000	ng	0.00
76) Chrysene-d12	21.033	240	1674112	20.000	ng	0.00
86) Perylene-d12	23.727	264	2019244	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.022	112	1471191	76.564	ng	0.00
7) Phenol-d6	6.587	99	2029645	76.657	ng	0.00
23) Nitrobenzene-d5	8.522	82	2162113	78.454	ng	0.00
42) 2,4,6-Tribromophenol	15.539	330	847891	79.913	ng	0.00
45) 2-Fluorobiphenyl	12.645	172	4628326	78.021	ng	0.00
79) Terphenyl-d14	19.462	244	6271744	80.275	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.034	88	301045	37.812	ng	100
3) Pyridine	3.411	79	894414m	38.338	ng	
4) n-Nitrosodimethylamine	3.328	42	391342	38.367	ng	100
6) Aniline	6.728	93	980743	39.189	ng	100
8) 2-Chlorophenol	6.952	128	813486	38.816	ng	100
9) Benzaldehyde	6.540	77	532876	37.348	ng	100
10) Phenol	6.610	94	997737	37.765	ng	100
11) bis(2-Chloroethyl)ether	6.828	93	881082	38.943	ng	100
12) 1,3-Dichlorobenzene	7.263	146	966606	38.813	ng	100
13) 1,4-Dichlorobenzene	7.410	146	981161	38.912	ng	100
14) 1,2-Dichlorobenzene	7.716	146	902511	37.935	ng	100
15) Benzyl Alcohol	7.628	79	754177	40.023	ng	100
16) 2,2'-oxybis(1-Chloropr...	7.916	45	1139263	39.195	ng	100
17) 2-Methylphenol	7.834	107	688729	39.076	ng	100
18) Hexachloroethane	8.428	117	378211	38.722	ng	100
19) n-Nitroso-di-n-propyla...	8.187	70	674051	37.795	ng	100
20) 3+4-Methylphenols	8.163	107	969100	39.329	ng	100
22) Acetophenone	8.198	105	1283137	38.307	ng	100
24) Nitrobenzene	8.563	77	1091975	39.350	ng	100
25) Isophorone	9.093	82	1909053	39.785	ng	100
26) 2-Nitrophenol	9.263	139	504016	41.011	ng	100
27) 2,4-Dimethylphenol	9.345	122	579838	40.278	ng	100
28) bis(2-Chloroethoxy)met...	9.575	93	1208106	40.013	ng	100
29) 2,4-Dichlorophenol	9.798	162	865465	40.518	ng	100
30) 1,2,4-Trichlorobenzene	9.998	180	960853	39.749	ng	100
31) Naphthalene	10.181	128	2680978	38.782	ng	100
32) Benzoic acid	9.534	122	695240	41.337	ng	100
33) 4-Chloroaniline	10.310	127	1069700	40.238	ng	100
34) Hexachlorobutadiene	10.469	225	562225	39.737	ng	100
35) Caprolactam	11.122	113	303209	40.875	ng	100
36) 4-Chloro-3-methylphenol	11.451	107	909901	39.998	ng	100
37) 2-Methylnaphthalene	11.804	142	1912342	38.853	ng	100
38) 1-Methylnaphthalene	12.034	142	1890364	38.861	ng	100
40) 1,2,4,5-Tetrachloroben...	12.186	216	994174	39.833	ng	100
41) Hexachlorocyclopentadiene	12.163	237	366656	42.071	ng	100
43) 2,4,6-Trichlorophenol	12.445	196	730450	41.176	ng	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081024\
 Data File : BM047144.D
 Acq On : 10 Aug 2024 15:51
 Operator : RC/JU
 Sample : SSTDICCC040
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICCC040


Quant Time: Aug 11 23:33:42 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Aug 11 23:24:37 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	12.522	196	800233	40.628	ng	100
46) 1,1'-Biphenyl	12.857	154	2539356	38.981	ng	100
47) 2-Chloronaphthalene	12.886	162	1959972	38.494	ng	100
48) 2-Nitroaniline	13.116	65	652655	40.916	ng	100
49) Acenaphthylene	13.751	152	2931534	39.060	ng	100
50) Dimethylphthalate	13.516	163	2522996	39.479	ng	100
51) 2,6-Dinitrotoluene	13.628	165	601084	40.235	ng	100
52) Acenaphthene	14.098	154	1844024	38.733	ng	100
53) 3-Nitroaniline	13.963	138	606076	40.486	ng	100
54) 2,4-Dinitrophenol	14.175	184	373189	41.679	ng	100
55) Dibenzofuran	14.439	168	2890230	38.434	ng	100
56) 4-Nitrophenol	14.298	139	537017	41.602	ng	100
57) 2,4-Dinitrotoluene	14.428	165	806412	40.441	ng	100
58) Fluorene	15.092	166	2395462	38.607	ng	100
59) 2,3,4,6-Tetrachlorophenol	14.675	232	650207	40.167	ng	100
60) Diethylphthalate	14.898	149	2558000	39.196	ng	100
61) 4-Chlorophenyl-phenyle...	15.098	204	1137143	38.496	ng	100
62) 4-Nitroaniline	15.139	138	584150	39.609	ng	100
63) Azobenzene	15.392	77	2403859	38.848	ng	100
65) 4,6-Dinitro-2-methylph...	15.198	198	476893	42.835	ng	100
66) n-Nitrosodiphenylamine	15.322	169	2039493	39.377	ng	100
67) 4-Bromophenyl-phenylether	15.998	248	725772	39.608	ng	100
68) Hexachlorobenzene	16.098	284	870505	39.667	ng	100
69) Atrazine	16.292	200	648319	41.597	ng	100
70) Pentachlorophenol	16.451	266	638108	41.480	ng	100
71) Phenanthrene	16.839	178	3712379	38.901	ng	100
72) Anthracene	16.927	178	3623679	39.024	ng	100
73) Carbazole	17.210	167	3693008	38.952	ng	100
74) Di-n-butylphthalate	17.792	149	4430949	40.165	ng	100
75) Fluoranthene	18.868	202	4559377	39.279	ng	100
77) Benzidine	19.074	184	1636101	57.629	ng	100
78) Pyrene	19.239	202	4770524	39.909	ng	100
80) Butylbenzylphthalate	20.174	149	2035184	41.719	ng	100
81) Benzo(a)anthracene	21.015	228	4353304	39.170	ng	100
82) 3,3'-Dichlorobenzidine	20.956	252	1483651	42.667	ng	100
83) Chrysene	21.074	228	4097336	38.858	ng	100
84) Bis(2-ethylhexyl)phtha...	20.974	149	2828936	40.891	ng	100
85) Di-n-octyl phthalate	22.009	149	4878689	41.489	ng	100
87) Indeno(1,2,3-cd)pyrene	26.727	276	5166805	39.586	ng	100
88) Benzo(b)fluoranthene	22.886	252	4689366	39.106	ng	100
89) Benzo(k)fluoranthene	22.945	252	4431625	39.126	ng	100
90) Benzo(a)pyrene	23.609	252	4136829	39.824	ng	100
91) Dibenzo(a,h)anthracene	26.780	278	4125251	39.055	ng	100
92) Benzo(g,h,i)perylene	27.662	276	4346786	39.636	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
BNA_M
ClientSampleId :
SSTDICCC040





[illegible]



#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.375 min Scan# 763
Delta R.T. 0.000 min
Lab File: BM047144.D
Acq: 10 Aug 2024 15:51




Instrument :
BNA_M
ClientSampleId :
SSTDICCC040


Tgt Ion:152 Resp: 343346
Ion Ratio Lower Upper
152 100
150 150.6 120.5 180.7
115 56.9 45.5 68.3



#2
1,4-Dioxane
Concen: 37.812 ng
RT: 3.034 min Scan# 25
Delta R.T. 0.000 min
Lab File: BM047144.D
Acq: 10 Aug 2024 15:51

Tgt Ion: 88 Resp: 301045
Ion Ratio Lower Upper
88 100
58 72.8 58.2 87.4
43 28.1 22.5 33.7





#3

Pyridine

Concen: 38.338 ng m

RT: 3.411 min Scan# 89

Delta R.T. 0.000 min

Lab File: BM047144.D

Acq: 10 Aug 2024 15:51

Instrument :





BNA_M

ClientSampleId :

SSTDICCC040

Tgt Ion: 79 Resp: 894414

Ion	Ratio	Lower	Upper
79	100		
52	64.0	51.2	76.8
51	32.6	26.1	39.1



#4

n-Nitrosodimethylamine

Concen: 38.367 ng

RT: 3.328 min Scan# 75




Delta R.T. 0.000 min


Lab File: BM047144.D

Acq: 10 Aug 2024 15:51

Tgt Ion: 42 Resp: 391342

Ion	Ratio	Lower	Upper
42	100		
74	141.1	112.9	169.3
44	6.6	5.3	7.9









#5
2-Fluorophenol
Concen: 76.564 ng
RT: 5.022 min Scan# 363
Delta R.T. 0.000 min
Lab File: BM047144.D
Acq: 10 Aug 2024 15:51




Instrument :
BNA_M
ClientSampleId :
SSTDICCC040


Tgt Ion:112 Resp: 1471191
Ion Ratio Lower Upper
112 100
64 58.4 46.7 70.1
63 31.0 24.8 37.2



#6
Aniline
Concen: 39.189 ng
RT: 6.728 min Scan# 653
Delta R.T. 0.000 min
Lab File: BM047144.D
Acq: 10 Aug 2024 15:51

Tgt Ion: 93 Resp: 980743
Ion Ratio Lower Upper
93 100
66 39.3 31.4 47.2
65 19.5 15.6 23.4









#7
Phenol-d6
Concen: 76.657 ng
RT: 6.587 min Scan# 629
Delta R.T. 0.000 min
Lab File: BM047144.D
Acq: 10 Aug 2024 15:51




Instrument :
BNA_M
ClientSampleId :
SSTDICCC040


Tgt Ion: 99 Resp: 2029645
Ion Ratio Lower Upper
99 100
42 17.6 14.1 21.1
71 35.6 28.5 42.7



#8
2-Chlorophenol
Concen: 38.816 ng
RT: 6.952 min Scan# 691
Delta R.T. 0.000 min
Lab File: BM047144.D
Acq: 10 Aug 2024 15:51

Tgt Ion: 128 Resp: 813486
Ion Ratio Lower Upper
128 100
130 32.1 12.1 52.1
64 46.6 26.6 66.6









#9
Benzaldehyde
Concen: 37.348 ng
RT: 6.540 min Scan# 61
Delta R.T. -0.000 min
Lab File: BM047144.D
Acq: 10 Aug 2024 15:51




Instrument :
BNA_M
ClientSampleId :
SSTDICCC040


Tgt Ion: 77 Resp: 532876
Ion Ratio Lower Upper
77 100
105 94.5 74.5 114.5
106 92.3 72.3 112.3



#10
Phenol
Concen: 37.765 ng
RT: 6.610 min Scan# 633
Delta R.T. 0.000 min
Lab File: BM047144.D
Acq: 10 Aug 2024 15:51

Tgt Ion: 94 Resp: 997737
Ion Ratio Lower Upper
94 100
65 29.0 9.0 49.0
66 38.9 18.9 58.9









#11
bis(2-Chloroethyl)ether
Concen: 38.943 ng
RT: 6.828 min Scan# 670
Delta R.T. 0.000 min
Lab File: BM047144.D
Acq: 10 Aug 2024 15:51




Instrument :
BNA_M
ClientSampleId :
SSTDICCC040


Tgt Ion: 93 Resp: 881082
Ion Ratio Lower Upper
93 100
63 75.8 55.8 95.8
95 32.7 12.7 52.7



#12
1,3-Dichlorobenzene
Concen: 38.813 ng
RT: 7.263 min Scan# 744
Delta R.T. 0.000 min
Lab File: BM047144.D
Acq: 10 Aug 2024 15:51

Tgt Ion: 146 Resp: 966606
Ion Ratio Lower Upper
146 100
148 65.4 52.3 78.5
75 32.2 25.8 38.6





#13

1,4-Dichlorobenzene

Concen: 38.912 ng

RT: 7.410 min Scan# 769

Delta R.T. 0.000 min

Lab File: BM047144.D

Acq: 10 Aug 2024 15:51

Instrument :

BNA_M

ClientSampleId :

SSTDICCC040





Tgt Ion:146 Resp: 981161

Ion Ratio Lower Upper

146 100

148 64.3 51.4 77.2

111 39.4 31.5 47.3



#14

1,2-Dichlorobenzene

Concen: 37.935 ng

RT: 7.716 min Scan# 821

Delta R.T. 0.000 min

Lab File: BM047144.D

Acq: 10 Aug 2024 15:51




Tgt Ion:146 Resp: 902511


Ion Ratio Lower Upper

146 100

148 64.8 51.8 77.8

111 41.6 33.3 49.9









#15
Benzyl Alcohol
Concen: 40.023 ng
RT: 7.628 min Scan# 806
Delta R.T. 0.000 min
Lab File: BM047144.D
Acq: 10 Aug 2024 15:51




Instrument :
BNA_M
ClientSampleId :
SSTDICCC040


Tgt Ion: 79 Resp: 754177
Ion Ratio Lower Upper
79 100
108 74.6 59.7 89.5
77 66.1 52.9 79.3



#16
2,2'-oxybis(1-chloropropane)
Concen: 39.195 ng
RT: 7.916 min Scan# 855
Delta R.T. 0.000 min
Lab File: BM047144.D
Acq: 10 Aug 2024 15:51

Tgt Ion: 45 Resp: 1139263
Ion Ratio Lower Upper
45 100
77 15.6 0.0 35.6
79 12.4 0.0 32.4





#17

2-Methylphenol

Concen: 39.076 ng

RT: 7.834 min Scan# 841

Delta R.T. 0.000 min

Lab File: BM047144.D

Acq: 10 Aug 2024 15:51

Instrument :

BNA_M

ClientSampleId :

SSTDICCC040

Tgt Ion:107 Resp: 688729





Ion	Ratio	Lower	Upper

107	100		

108	113.2	90.6	135.8

77	49.1	39.3	58.9

79	48.3	38.6	58.0



#18

Hexachloroethane

Concen: 38.722 ng

RT: 8.428 min Scan# 942

Delta R.T. 0.000 min

Lab File: BM047144.D

Acq: 10 Aug 2024 15:51




Tgt Ion:117 Resp: 378211


Ion	Ratio	Lower	Upper

117	100		

119	96.2	77.0	115.4

201	101.6	81.3	121.9





#19

n-Nitroso-di-n-propylamine

Concen: 37.795 ng

RT: 8.187 min Scan# 901

Delta R.T. 0.000 min

Lab File: BM047144.D

Acq: 10 Aug 2024 15:51

Instrument :

BNA_M

ClientSampleId :

SSTDICCC040

Tgt Ion: 70 Resp: 674051





Ion Ratio Lower Upper

70 100

42 52.5 42.0 63.0

101 9.0 7.2 10.8

130 19.4 15.5 23.3



#20

3+4-Methylphenols

Concen: 39.329 ng

RT: 8.163 min Scan# 897

Delta R.T. 0.000 min

Lab File: BM047144.D

Acq: 10 Aug 2024 15:51

Tgt Ion: 107 Resp: 969100




Ion Ratio Lower Upper


107 100

108 90.7 70.7 110.7

77 37.4 17.4 57.4

79 30.9 10.9 50.9









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.134 min Scan# 1232
Delta R.T. 0.000 min
Lab File: BM047144.D
Acq: 10 Aug 2024 15:51




Instrument :
BNA_M
ClientSampleId :
SSTDICCC040


Tgt Ion:136 Resp: 1387685
Ion Ratio Lower Upper
136 100
137 11.0 8.8 13.2
54 7.5 6.0 9.0
68 6.0 4.8 7.2



#22
Acetophenone
Concen: 38.307 ng
RT: 8.198 min Scan# 903
Delta R.T. 0.000 min
Lab File: BM047144.D
Acq: 10 Aug 2024 15:51

Tgt Ion:105 Resp: 1283137
Ion Ratio Lower Upper
105 100
71 5.6 4.5 6.7
51 25.1 20.1 30.1
120 22.2 17.8 26.6





#23

Nitrobenzene-d5

Concen: 78.454 ng

RT: 8.522 min Scan# 958

Delta R.T. 0.000 min

Lab File: BM047144.D

Acq: 10 Aug 2024 15:51

Instrument :





BNA_M

ClientSampleId :

SSTDICCC040

Tgt Ion: 82 Resp: 2162113

Ion	Ratio	Lower	Upper
82	100		
128	40.4	32.3	48.5
54	46.8	37.4	56.2



#24

Nitrobenzene

Concen: 39.350 ng

RT: 8.563 min Scan# 965




Delta R.T. 0.000 min


Lab File: BM047144.D

Acq: 10 Aug 2024 15:51

Tgt Ion: 77 Resp: 1091975

Ion	Ratio	Lower	Upper
77	100		
123	42.5	34.0	51.0
65	13.1	10.5	15.7









#25
Isophorone
Concen: 39.785 ng
RT: 9.093 min Scan# 1055
Delta R.T. 0.000 min
Lab File: BM047144.D
Acq: 10 Aug 2024 15:51




Instrument :
BNA_M
ClientSampleId :
SSTDICCC040


Tgt Ion: 82 Resp: 1909053
Ion Ratio Lower Upper
82 100
95 6.9 5.5 8.3
138 16.9 13.5 20.3



#26
2-Nitrophenol
Concen: 41.011 ng
RT: 9.263 min Scan# 1084
Delta R.T. 0.000 min
Lab File: BM047144.D
Acq: 10 Aug 2024 15:51

Tgt Ion:139 Resp: 504016
Ion Ratio Lower Upper
139 100
109 30.5 24.4 36.6
65 53.8 43.0 64.6









#27
2,4-Dimethylphenol
Concen: 40.278 ng
RT: 9.345 min Scan# 1098
Delta R.T. 0.000 min
Lab File: BM047144.D
Acq: 10 Aug 2024 15:51




Instrument :
BNA_M
ClientSampleId :
SSTDICCC040


Tgt Ion:122 Resp: 579838
Ion Ratio Lower Upper
122 100
107 116.0 92.8 139.2
121 57.6 46.1 69.1



#28
bis(2-Chloroethoxy)methane
Concen: 40.013 ng
RT: 9.575 min Scan# 1137
Delta R.T. 0.000 min
Lab File: BM047144.D
Acq: 10 Aug 2024 15:51

Tgt Ion: 93 Resp: 1208106
Ion Ratio Lower Upper
93 100
95 32.3 25.8 38.8
123 11.2 9.0 13.4





#29

2,4-Dichlorophenol

Concen: 40.518 ng

RT: 9.798 min Scan# 1175

Delta R.T. 0.000 min

Lab File: BM047144.D

Acq: 10 Aug 2024 15:51

Instrument :

BNA_M

ClientSampleId :

SSTDICCC040





Tgt Ion:162 Resp: 865465

Ion Ratio Lower Upper

162 100

164 64.7 44.7 84.7

98 34.9 14.9 54.9



#30

1,2,4-Trichlorobenzene

Concen: 39.749 ng

RT: 9.998 min Scan# 1209

Delta R.T. 0.000 min

Lab File: BM047144.D

Acq: 10 Aug 2024 15:51




Tgt Ion:180 Resp: 960853


Ion Ratio Lower Upper

180 100

182 95.8 76.6 115.0

145 29.3 23.4 35.2









#31
Naphthalene
Concen: 38.782 ng
RT: 10.181 min Scan# 1240
Delta R.T. 0.000 min
Lab File: BM047144.D
Acq: 10 Aug 2024 15:51




Instrument :
BNA_M
ClientSampleId :
SSTDICCC040


Tgt Ion:128 Resp: 2680978
Ion Ratio Lower Upper
128 100
129 10.8 8.6 13.0
127 12.8 10.2 15.4



#32
Benzoic acid
Concen: 41.337 ng
RT: 9.534 min Scan# 1130
Delta R.T. 0.000 min
Lab File: BM047144.D
Acq: 10 Aug 2024 15:51

Tgt Ion:122 Resp: 695240
Ion Ratio Lower Upper
122 100
105 127.1 107.1 147.1
77 102.1 82.1 122.1









#33
4-Chloroaniline
Concen: 40.238 ng
RT: 10.310 min Scan# 1262
Delta R.T. 0.000 min
Lab File: BM047144.D
Acq: 10 Aug 2024 15:51




Instrument :
BNA_M
ClientSampleId :
SSTDICCC040


Tgt Ion:127 Resp: 1069700
Ion Ratio Lower Upper
127 100
129 32.3 25.8 38.8
65 31.1 24.9 37.3
92 18.9 15.1 22.7



#34
Hexachlorobutadiene
Concen: 39.737 ng
RT: 10.469 min Scan# 1289
Delta R.T. 0.000 min
Lab File: BM047144.D
Acq: 10 Aug 2024 15:51

Tgt Ion:225 Resp: 562225
Ion Ratio Lower Upper
225 100
223 62.7 50.2 75.2
227 64.2 51.4 77.0









#35
Caprolactam
Concen: 40.875 ng
RT: 11.122 min Scan# 1400
Delta R.T. 0.000 min
Lab File: BM047144.D
Acq: 10 Aug 2024 15:51




Instrument :
BNA_M
ClientSampleId :
SSTDICCC040


Tgt Ion:113 Resp: 303209
Ion Ratio Lower Upper
113 100
55 160.5 140.5 180.5
56 123.6 103.6 143.6



#36
4-Chloro-3-methylphenol
Concen: 39.998 ng
RT: 11.451 min Scan# 1456
Delta R.T. 0.000 min
Lab File: BM047144.D
Acq: 10 Aug 2024 15:51

Tgt Ion:107 Resp: 909901
Ion Ratio Lower Upper
107 100
144 27.2 21.8 32.6
142 83.5 66.8 100.2









#37
2-Methylnaphthalene
Concen: 38.853 ng
RT: 11.804 min Scan# 1516
Delta R.T. 0.000 min
Lab File: BM047144.D
Acq: 10 Aug 2024 15:51




Instrument :
BNA_M
ClientSampleId :
SSTDICCC040


Tgt Ion:142 Resp: 1912342
Ion Ratio Lower Upper
142 100
141 88.6 70.9 106.3
115 35.9 28.7 43.1



#38
1-Methylnaphthalene
Concen: 38.861 ng
RT: 12.034 min Scan# 1555
Delta R.T. 0.000 min
Lab File: BM047144.D
Acq: 10 Aug 2024 15:51

Tgt Ion:142 Resp: 1890364
Ion Ratio Lower Upper
142 100
141 93.2 74.6 111.8
116 3.9 3.1 4.7





#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 14.033 min Scan# 1895

Delta R.T. 0.000 min

Lab File: BM047144.D

Acq: 10 Aug 2024 15:51

Instrument :





BNA_M

ClientSampleId :

SSTDICCC040

Tgt Ion:164 Resp: 915049

Ion	Ratio	Lower	Upper
164	100		
162	100.9	80.7	121.1
160	45.0	36.0	54.0



#40

1,2,4,5-Tetrachlorobenzene

Concen: 39.833 ng

RT: 12.186 min Scan# 1581




Delta R.T. 0.000 min


Lab File: BM047144.D

Acq: 10 Aug 2024 15:51

Tgt Ion:216 Resp: 994174

Ion	Ratio	Lower	Upper
216	100		
214	79.3	63.4	95.2
179	22.0	17.6	26.4
108	18.9	15.1	22.7









#41
Hexachlorocyclopentadiene
Concen: 42.071 ng
RT: 12.163 min Scan# 1577
Delta R.T. 0.000 min
Lab File: BM047144.D
Acq: 10 Aug 2024 15:51




Instrument :
BNA_M
ClientSampleId :
SSTDICCC040


Tgt Ion:237 Resp: 366656
Ion Ratio Lower Upper
237 100
235 64.8 44.8 84.8
272 13.5 0.0 33.5



#42
2,4,6-Tribromophenol
Concen: 79.913 ng
RT: 15.539 min Scan# 2151
Delta R.T. 0.000 min
Lab File: BM047144.D
Acq: 10 Aug 2024 15:51

Tgt Ion:330 Resp: 847891
Ion Ratio Lower Upper
330 100
332 96.8 77.4 116.2
141 39.5 31.6 47.4





#43

2,4,6-Trichlorophenol

Concen: 41.176 ng

RT: 12.445 min Scan# 1

Delta R.T. 0.000 min

Lab File: BM047144.D

Acq: 10 Aug 2024 15:51

Instrument :

BNA_M

ClientSampleId :

SSTDICCC040





Tgt Ion:196 Resp: 730450

Ion Ratio Lower Upper

196 100

198 94.5 75.6 113.4

200 30.8 24.6 37.0



#44

2,4,5-Trichlorophenol

Concen: 40.628 ng

RT: 12.522 min Scan# 1638

Delta R.T. 0.000 min

Lab File: BM047144.D

Acq: 10 Aug 2024 15:51

Tgt Ion:196 Resp: 800233




Ion Ratio Lower Upper


196 100

198 97.1 77.7 116.5

97 49.8 39.8 59.8

132 25.9 20.7 31.1









#45
2-Fluorobiphenyl
Concen: 78.021 ng
RT: 12.645 min Scan# 1659
Delta R.T. 0.000 min
Lab File: BM047144.D
Acq: 10 Aug 2024 15:51




Instrument :
BNA_M
ClientSampleId :
SSTDICCC040


Tgt Ion:172 Resp: 4628326
Ion Ratio Lower Upper
172 100
171 34.7 27.8 41.6
170 23.4 18.7 28.1



#46
1,1'-Biphenyl
Concen: 38.981 ng
RT: 12.857 min Scan# 1695
Delta R.T. 0.000 min
Lab File: BM047144.D
Acq: 10 Aug 2024 15:51

Tgt Ion:154 Resp: 2539356
Ion Ratio Lower Upper
154 100
153 39.0 19.0 59.0
76 11.9 0.0 31.9





#47

2-Chloronaphthalene

Concen: 38.494 ng

RT: 12.886 min Scan# 1700

Delta R.T. 0.000 min

Lab File: BM047144.D

Acq: 10 Aug 2024 15:51

Instrument :

BNA_M

ClientSampleId :

SSTDICCC040





Tgt Ion:162 Resp: 1959972

Ion Ratio Lower Upper

162 100

127 37.5 30.0 45.0

164 32.9 26.3 39.5



#48

2-Nitroaniline

Concen: 40.916 ng

RT: 13.116 min Scan# 1739

Delta R.T. 0.000 min

Lab File: BM047144.D

Acq: 10 Aug 2024 15:51




Tgt Ion: 65 Resp: 652655


Ion Ratio Lower Upper

65 100

92 68.9 55.1 82.7

138 107.4 85.9 128.9





#49

Acenaphthylene

Concen: 39.060 ng

RT: 13.751 min Scan# 1847

Delta R.T. 0.000 min

Lab File: BM047144.D

Acq: 10 Aug 2024 15:51

Instrument :





BNA_M

ClientSampleId :

SSTDICCC040

Tgt Ion:152 Resp: 2931534

Ion	Ratio	Lower	Upper
152	100		
151	19.7	15.8	23.6
153	13.0	10.4	15.6



#50

Dimethylphthalate

Concen: 39.479 ng

RT: 13.516 min Scan# 1807




Delta R.T. 0.000 min


Lab File: BM047144.D

Acq: 10 Aug 2024 15:51

Tgt Ion:163 Resp: 2522996

Ion	Ratio	Lower	Upper
163	100		
194	3.8	3.0	4.6
164	10.0	8.0	12.0









#51
2,6-Dinitrotoluene
Concen: 40.235 ng
RT: 13.628 min Scan# 1826
Delta R.T. 0.000 min
Lab File: BM047144.D
Acq: 10 Aug 2024 15:51




Instrument :
BNA_M
ClientSampleId :
SSTDICCC040


Tgt Ion:165 Resp: 601084
Ion Ratio Lower Upper
165 100
63 58.1 46.5 69.7
89 59.1 47.3 70.9



#52
Acenaphthene
Concen: 38.733 ng
RT: 14.098 min Scan# 1906
Delta R.T. 0.000 min
Lab File: BM047144.D
Acq: 10 Aug 2024 15:51

Tgt Ion:154 Resp: 1844024
Ion Ratio Lower Upper
154 100
153 112.9 90.3 135.5
152 53.6 42.9 64.3









#53
3-Nitroaniline
Concen: 40.486 ng
RT: 13.963 min Scan# 1883
Delta R.T. -0.000 min
Lab File: BM047144.D
Acq: 10 Aug 2024 15:51




Instrument :
BNA_M
ClientSampleId :
SSTDICCC040


Tgt Ion:138 Resp: 606076
Ion Ratio Lower Upper
138 100
108 9.9 7.9 11.9
92 115.0 92.0 138.0



#54
2,4-Dinitrophenol
Concen: 41.679 ng
RT: 14.175 min Scan# 1919
Delta R.T. -0.000 min
Lab File: BM047144.D
Acq: 10 Aug 2024 15:51

Tgt Ion:184 Resp: 373189
Ion Ratio Lower Upper
184 100
63 65.9 52.7 79.1
154 63.0 50.4 75.6









#55
Dibenzofuran
Concen: 38.434 ng
RT: 14.439 min Scan# 1964
Delta R.T. 0.000 min
Lab File: BM047144.D
Acq: 10 Aug 2024 15:51




Instrument :
BNA_M
ClientSampleId :
SSTDICCC040


Tgt Ion:168 Resp: 2890230
Ion Ratio Lower Upper
168 100
139 37.0 29.6 44.4
169 13.4 10.7 16.1



#56
4-Nitrophenol
Concen: 41.602 ng
RT: 14.298 min Scan# 1940
Delta R.T. 0.000 min
Lab File: BM047144.D
Acq: 10 Aug 2024 15:51

Tgt Ion:139 Resp: 537017
Ion Ratio Lower Upper
139 100
109 72.5 52.5 92.5
65 95.4 75.4 115.4









#57
2,4-Dinitrotoluene
Concen: 40.441 ng
RT: 14.428 min Scan# 1962
Delta R.T. 0.000 min
Lab File: BM047144.D
Acq: 10 Aug 2024 15:51




Instrument :
BNA_M
ClientSampleId :
SSTDICCC040


Tgt Ion:165 Resp: 806412
Ion Ratio Lower Upper
165 100
63 50.4 40.3 60.5
89 76.3 61.0 91.6
182 2.4 1.9 2.9



#58
Fluorene
Concen: 38.607 ng
RT: 15.092 min Scan# 2075
Delta R.T. 0.000 min
Lab File: BM047144.D
Acq: 10 Aug 2024 15:51

Tgt Ion:166 Resp: 2395462
Ion Ratio Lower Upper
166 100
165 96.7 77.4 116.0
167 13.8 11.0 16.6









#59
2,3,4,6-Tetrachlorophenol
Concen: 40.167 ng
RT: 14.675 min Scan# 2004
Delta R.T. 0.000 min
Lab File: BM047144.D
Acq: 10 Aug 2024 15:51




Instrument :
BNA_M
ClientSampleId :
SSTDICCC040


Tgt Ion:232 Resp: 650207
Ion Ratio Lower Upper
232 100
131 46.1 36.9 55.3
130 2.5 2.0 3.0
166 33.0 26.4 39.6



#60
Diethylphthalate
Concen: 39.196 ng
RT: 14.898 min Scan# 2042
Delta R.T. 0.000 min
Lab File: BM047144.D
Acq: 10 Aug 2024 15:51

Tgt Ion:149 Resp: 2558000
Ion Ratio Lower Upper
149 100
177 22.0 17.6 26.4
150 12.2 9.8 14.6









#61
4-Chlorophenyl-phenylether
Concen: 38.496 ng
RT: 15.098 min Scan# 2076
Delta R.T. 0.000 min
Lab File: BM047144.D
Acq: 10 Aug 2024 15:51




Instrument :
BNA_M
ClientSampleId :
SSTDICCC040


Tgt Ion:204 Resp: 1137143
Ion Ratio Lower Upper
204 100
206 32.8 26.2 39.4
141 59.0 47.2 70.8



#62
4-Nitroaniline
Concen: 39.609 ng
RT: 15.139 min Scan# 2083
Delta R.T. 0.000 min
Lab File: BM047144.D
Acq: 10 Aug 2024 15:51

Tgt Ion:138 Resp: 584150
Ion Ratio Lower Upper
138 100
92 51.9 31.9 71.9
108 83.8 63.8 103.8









#63
Azobenzene
Concen: 38.848 ng
RT: 15.392 min Scan# 2126
Delta R.T. 0.000 min
Lab File: BM047144.D
Acq: 10 Aug 2024 15:51

Instrument :
BNA_M
ClientSampleId :
SSTDICCC040

Tgt Ion: 77 Resp: 2403859




Ion	Ratio	Lower	Upper
77	100		
182	23.8	3.8	43.8
105	19.5	0.0	39.5
51	26.8	6.8	46.8




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 16.792 min Scan# 2364
Delta R.T. 0.000 min
Lab File: BM047144.D
Acq: 10 Aug 2024 15:51

Tgt Ion: 188 Resp: 1901562

Ion	Ratio	Lower	Upper
188	100		
94	8.8	7.0	10.6
80	10.0	8.0	12.0









#65
4,6-Dinitro-2-methylphenol
Concen: 42.835 ng
RT: 15.198 min Scan# 2093
Delta R.T. 0.000 min
Lab File: BM047144.D
Acq: 10 Aug 2024 15:51




Instrument :
BNA_M
ClientSampleId :
SSTDICCC040


Tgt Ion:198 Resp: 476893
Ion Ratio Lower Upper
198 100
51 43.9 23.9 63.9
105 44.3 24.3 64.3



#66
n-Nitrosodiphenylamine
Concen: 39.377 ng
RT: 15.322 min Scan# 2114
Delta R.T. 0.000 min
Lab File: BM047144.D
Acq: 10 Aug 2024 15:51

Tgt Ion:169 Resp: 2039493
Ion Ratio Lower Upper
169 100
168 67.3 53.8 80.8
167 36.6 29.3 43.9









#67
4-Bromophenyl-phenylether
Concen: 39.608 ng
RT: 15.998 min Scan# 21
Delta R.T. 0.000 min
Lab File: BM047144.D
Acq: 10 Aug 2024 15:51




Instrument :
BNA_M
ClientSampleId :
SSTDICCC040


Tgt Ion:248 Resp: 725772
Ion Ratio Lower Upper
248 100
250 97.3 77.8 116.8
141 72.1 57.7 86.5



#68
Hexachlorobenzene
Concen: 39.667 ng
RT: 16.098 min Scan# 2246
Delta R.T. 0.000 min
Lab File: BM047144.D
Acq: 10 Aug 2024 15:51

Tgt Ion:284 Resp: 870505
Ion Ratio Lower Upper
284 100
142 36.7 29.4 44.0
249 31.4 25.1 37.7









#69
Atrazine
Concen: 41.597 ng
RT: 16.292 min Scan# 2279
Delta R.T. 0.000 min
Lab File: BM047144.D
Acq: 10 Aug 2024 15:51




Instrument :
BNA_M
ClientSampleId :
SSTDICCC040


Tgt Ion:200 Resp: 648319
Ion Ratio Lower Upper
200 100
173 26.1 6.1 46.1
215 46.8 26.8 66.8



#70
Pentachlorophenol
Concen: 41.480 ng
RT: 16.451 min Scan# 2306
Delta R.T. 0.000 min
Lab File: BM047144.D
Acq: 10 Aug 2024 15:51

Tgt Ion:266 Resp: 638108
Ion Ratio Lower Upper
266 100
268 62.2 49.8 74.6
264 62.0 49.6 74.4









#71
Phenanthrene
Concen: 38.901 ng
RT: 16.839 min Scan# 21
Delta R.T. 0.000 min
Lab File: BM047144.D
Acq: 10 Aug 2024 15:51




Instrument :
BNA_M
ClientSampleId :
SSTDICCC040


Tgt Ion:178 Resp: 3712379
Ion Ratio Lower Upper
178 100
176 19.1 15.3 22.9
179 15.5 12.4 18.6



#72
Anthracene
Concen: 39.024 ng
RT: 16.927 min Scan# 2387
Delta R.T. 0.000 min
Lab File: BM047144.D
Acq: 10 Aug 2024 15:51

Tgt Ion:178 Resp: 3623679
Ion Ratio Lower Upper
178 100
176 18.5 14.8 22.2
179 15.5 12.4 18.6









#73
Carbazole
Concen: 38.952 ng
RT: 17.210 min Scan# 2435
Delta R.T. 0.000 min
Lab File: BM047144.D
Acq: 10 Aug 2024 15:51




Instrument :
BNA_M
ClientSampleId :
SSTDICCC040


Tgt Ion:167 Resp: 3693008
Ion Ratio Lower Upper
167 100
166 21.4 17.1 25.7
139 12.5 10.0 15.0



#74
Di-n-butylphthalate
Concen: 40.165 ng
RT: 17.792 min Scan# 2534
Delta R.T. 0.000 min
Lab File: BM047144.D
Acq: 10 Aug 2024 15:51

Tgt Ion:149 Resp: 4430949
Ion Ratio Lower Upper
149 100
150 9.4 7.5 11.3
104 5.1 4.1 6.1









#75
Fluoranthene
Concen: 39.279 ng
RT: 18.868 min Scan# 2717
Delta R.T. 0.000 min
Lab File: BM047144.D
Acq: 10 Aug 2024 15:51




Instrument :
BNA_M
ClientSampleId :
SSTDICCC040


Tgt Ion:202 Resp: 4559377
Ion Ratio Lower Upper
202 100
101 10.2 0.0 30.2
203 17.6 0.0 37.6



#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.033 min Scan# 3085
Delta R.T. 0.000 min
Lab File: BM047144.D
Acq: 10 Aug 2024 15:51

Tgt Ion:240 Resp: 1674112
Ion Ratio Lower Upper
240 100
120 10.2 8.2 12.2
236 25.7 20.6 30.8









#77
Benzidine
Concen: 57.629 ng
RT: 19.074 min Scan# 2752
Delta R.T. 0.000 min
Lab File: BM047144.D
Acq: 10 Aug 2024 15:51




Instrument :
BNA_M
ClientSampleId :
SSTDICCC040


Tgt Ion:184 Resp: 1636101
Ion Ratio Lower Upper
184 100
185 14.7 11.8 17.6
183 11.3 9.0 13.6



#78
Pyrene
Concen: 39.909 ng
RT: 19.239 min Scan# 2780
Delta R.T. 0.000 min
Lab File: BM047144.D
Acq: 10 Aug 2024 15:51

Tgt Ion:202 Resp: 4770524
Ion Ratio Lower Upper
202 100
200 21.5 17.2 25.8
203 17.7 14.2 21.2









#79
Terphenyl-d14
Concen: 80.275 ng
RT: 19.462 min Scan# 2818
Delta R.T. 0.000 min
Lab File: BM047144.D
Acq: 10 Aug 2024 15:51




Instrument :
BNA_M
ClientSampleId :
SSTDICCC040


Tgt Ion:244 Resp: 6271744
Ion Ratio Lower Upper
244 100
212 7.8 6.2 9.4
122 8.9 7.1 10.7



#80
Butylbenzylphthalate
Concen: 41.719 ng
RT: 20.174 min Scan# 2939
Delta R.T. 0.000 min
Lab File: BM047144.D
Acq: 10 Aug 2024 15:51

Tgt Ion:149 Resp: 2035184
Ion Ratio Lower Upper
149 100
91 71.2 57.0 85.4
206 20.5 16.4 24.6









#81
Benzo(a)anthracene
Concen: 39.170 ng
RT: 21.015 min Scan# 3082
Delta R.T. 0.000 min
Lab File: BM047144.D
Acq: 10 Aug 2024 15:51




Instrument :
BNA_M
ClientSampleId :
SSTDICCC040


Tgt Ion:228 Resp: 4353304
Ion Ratio Lower Upper
228 100
226 27.2 21.8 32.6
229 19.8 15.8 23.8



#82
3,3'-Dichlorobenzidine
Concen: 42.667 ng
RT: 20.956 min Scan# 3072
Delta R.T. 0.000 min
Lab File: BM047144.D
Acq: 10 Aug 2024 15:51

Tgt Ion:252 Resp: 1483651
Ion Ratio Lower Upper
252 100
254 63.8 51.0 76.6
126 10.6 8.5 12.7





#83

Chrysene

Concen: 38.858 ng

RT: 21.074 min Scan# 3092

Delta R.T. 0.000 min

Lab File: BM047144.D

Acq: 10 Aug 2024 15:51

Instrument :

BNA_M

ClientSampleId :

SSTDICCC040





Tgt Ion:228 Resp: 4097336

Ion Ratio Lower Upper

228 100

226 30.5 24.4 36.6

229 19.6 15.7 23.5



#84

Bis(2-ethylhexyl)phthalate

Concen: 40.891 ng

RT: 20.974 min Scan# 3075

Delta R.T. 0.000 min

Lab File: BM047144.D

Acq: 10 Aug 2024 15:51




Tgt Ion:149 Resp: 2828936


Ion Ratio Lower Upper

149 100

167 27.7 22.2 33.2

279 3.7 3.0 4.4









#85
Di-n-octyl phthalate
Concen: 41.489 ng
RT: 22.009 min Scan# 31
Delta R.T. 0.000 min
Lab File: BM047144.D
Acq: 10 Aug 2024 15:51




Instrument :
BNA_M
ClientSampleId :
SSTDICCC040


Tgt Ion:149 Resp: 4878689
Ion Ratio Lower Upper
149 100
167 1.6 1.3 1.9
43 8.6 6.9 10.3



#86
Perylene-d12
Concen: 20.000 ng
RT: 23.727 min Scan# 3543
Delta R.T. 0.000 min
Lab File: BM047144.D
Acq: 10 Aug 2024 15:51

Tgt Ion:264 Resp: 2019244
Ion Ratio Lower Upper
264 100
260 23.1 18.5 27.7
265 21.6 17.3 25.9









#87
Indeno(1,2,3-cd)pyrene
Concen: 39.586 ng
RT: 26.727 min Scan# 4053
Delta R.T. 0.000 min
Lab File: BM047144.D
Acq: 10 Aug 2024 15:51




Instrument :
BNA_M
ClientSampleId :
SSTDICCC040


Tgt Ion:276 Resp: 5166805
Ion Ratio Lower Upper
276 100
138 24.2 19.4 29.0
277 25.1 20.1 30.1



#88
Benzo(b)fluoranthene
Concen: 39.106 ng
RT: 22.886 min Scan# 3400
Delta R.T. 0.000 min
Lab File: BM047144.D
Acq: 10 Aug 2024 15:51

Tgt Ion:252 Resp: 4689366
Ion Ratio Lower Upper
252 100
253 22.2 17.8 26.6
125 9.2 7.4 11.0





#89

Benzo(k)fluoranthene

Concen: 39.126 ng

RT: 22.945 min Scan# 3410

Delta R.T. 0.000 min

Lab File: BM047144.D

Acq: 10 Aug 2024 15:51

Instrument :

BNA_M

ClientSampleId :

SSTDICCC040





Tgt Ion:252 Resp: 4431625

Ion Ratio Lower Upper

252 100

253 22.1 17.7 26.5

125 9.3 7.4 11.2



#90

Benzo(a)pyrene

Concen: 39.824 ng

RT: 23.609 min Scan# 3523

Delta R.T. 0.000 min

Lab File: BM047144.D

Acq: 10 Aug 2024 15:51




Tgt Ion:252 Resp: 4136829


Ion Ratio Lower Upper

252 100

253 21.4 17.1 25.7

125 10.2 8.2 12.2









#91
Dibenzo(a,h)anthracene
Concen: 39.055 ng
RT: 26.780 min Scan# 4062
Delta R.T. 0.000 min
Lab File: BM047144.D
Acq: 10 Aug 2024 15:51




Instrument :
BNA_M
ClientSampleId :
SSTDICCC040

Tgt Ion:278 Resp: 4125251
Ion Ratio Lower Upper
278 100
139 15.0 12.0 18.0
279 24.0 19.2 28.8



#92
Benzo(g,h,i)perylene
Concen: 39.636 ng
RT: 27.662 min Scan# 4212
Delta R.T. 0.000 min
Lab File: BM047144.D
Acq: 10 Aug 2024 15:51

Tgt Ion:276 Resp: 4346786
Ion Ratio Lower Upper
276 100
277 23.5 18.8 28.2
138 20.3 16.2 24.4



Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081024\
 Data File : BM047145.D
 Acq On : 10 Aug 2024 16:31
 Operator : RC/JU
 Sample : SSTDICC050
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC050

Quant Time: Aug 11 23:34:27 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Aug 11 23:24:37 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.375	152	315091	20.000	ng	0.00
21) Naphthalene-d8	10.133	136	1281047	20.000	ng	0.00
39) Acenaphthene-d10	14.033	164	834055	20.000	ng	0.00
64) Phenanthrene-d10	16.792	188	1737042	20.000	ng	0.00
76) Chrysene-d12	21.033	240	1511494	20.000	ng	0.00
86) Perylene-d12	23.733	264	1841387	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.022	112	1654844	93.844	ng	0.00
7) Phenol-d6	6.587	99	2315558	95.298	ng	0.00
23) Nitrobenzene-d5	8.528	82	2421031	95.162	ng	0.00
42) 2,4,6-Tribromophenol	15.545	330	952366	98.476	ng	0.00
45) 2-Fluorobiphenyl	12.645	172	5200055	96.172	ng	0.00
79) Terphenyl-d14	19.462	244	7014667	99.443	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.034	88	343225	46.976	ng	98
3) Pyridine	3.410	79	1022963m	47.780	ng	
4) n-Nitrosodimethylamine	3.328	42	444430	47.479	ng	99
6) Aniline	6.728	93	1103960	48.069	ng	99
8) 2-Chlorophenol	6.951	128	918928	47.779	ng	99
9) Benzaldehyde	6.540	77	590756	53.942	ng	98
10) Phenol	6.616	94	1142509	47.123	ng	99
11) bis(2-Chloroethyl)ether	6.828	93	997816	48.057	ng	100
12) 1,3-Dichlorobenzene	7.263	146	1098474	48.063	ng	99
13) 1,4-Dichlorobenzene	7.410	146	1107084	47.843	ng	99
14) 1,2-Dichlorobenzene	7.722	146	1030619	47.204	ng	99
15) Benzyl Alcohol	7.628	79	847663	49.018	ng	99
16) 2,2'-oxybis(1-Chloropr...	7.916	45	1272990	47.724	ng	100
17) 2-Methylphenol	7.834	107	786481	48.623	ng	99
18) Hexachloroethane	8.428	117	435333	48.568	ng	100
19) n-Nitroso-di-n-propyla...	8.192	70	759990	46.435	ng	99
20) 3+4-Methylphenols	8.163	107	1106673	48.940	ng	98
22) Acetophenone	8.198	105	1441742	46.625	ng	99
24) Nitrobenzene	8.569	77	1216089	47.470	ng	96
25) Isophorone	9.092	82	2139220	48.292	ng	100
26) 2-Nitrophenol	9.263	139	582500	51.343	ng	98
27) 2,4-Dimethylphenol	9.345	122	658437	49.545	ng	100
28) bis(2-Chloroethoxy)met...	9.575	93	1347358	48.339	ng	100
29) 2,4-Dichlorophenol	9.798	162	993725	50.395	ng	99
30) 1,2,4-Trichlorobenzene	9.998	180	1089421	48.819	ng	99
31) Naphthalene	10.180	128	3038911	47.619	ng	99
32) Benzoic acid	9.539	122	838043	53.976	ng	97
33) 4-Chloroaniline	10.310	127	1197445	48.793	ng	99
34) Hexachlorobutadiene	10.469	225	650107	49.773	ng	99
35) Caprolactam	11.127	113	344173	50.259	ng	99
36) 4-Chloro-3-methylphenol	11.457	107	1033238	49.201	ng	99
37) 2-Methylnaphthalene	11.810	142	2180425	47.987	ng	99
38) 1-Methylnaphthalene	12.033	142	2140830	47.673	ng	97
40) 1,2,4,5-Tetrachloroben...	12.186	216	1138431	50.043	ng	99
41) Hexachlorocyclopentadiene	12.163	237	415149	52.261	ng	100
43) 2,4,6-Trichlorophenol	12.445	196	821910	50.831	ng	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081024\
 Data File : BM047145.D
 Acq On : 10 Aug 2024 16:31
 Operator : RC/JU
 Sample : SSTDICC050
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC050

Quant Time: Aug 11 23:34:27 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Aug 11 23:24:37 2024
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	12.521	196	909210	50.644	ng	99
46) 1,1'-Biphenyl	12.857	154	2837162	47.782	ng	99
47) 2-Chloronaphthalene	12.892	162	2210354	47.627	ng	99
48) 2-Nitroaniline	13.116	65	733636	50.459	ng	98
49) Acenaphthylene	13.751	152	3284244	48.009	ng	100
50) Dimethylphthalate	13.516	163	2800312	48.073	ng	99
51) 2,6-Dinitrotoluene	13.633	165	679884	49.929	ng	95
52) Acenaphthene	14.098	154	2067281	47.640	ng	99
53) 3-Nitroaniline	13.963	138	683158	50.067	ng	100
54) 2,4-Dinitrophenol	14.174	184	424486	52.012	ng	99
55) Dibenzofuran	14.439	168	3249834	47.413	ng	100
56) 4-Nitrophenol	14.298	139	606437	51.542	ng	96
57) 2,4-Dinitrotoluene	14.427	165	917536	50.482	ng	100
58) Fluorene	15.098	166	2690806	47.578	ng	100
59) 2,3,4,6-Tetrachlorophenol	14.674	232	738067	50.022	ng	99
60) Diethylphthalate	14.898	149	2846430	47.851	ng	100
61) 4-Chlorophenyl-phenyle...	15.104	204	1289728	47.901	ng	98
62) 4-Nitroaniline	15.139	138	661759	49.229	ng	99
63) Azobenzene	15.398	77	2669713	47.334	ng	98
65) 4,6-Dinitro-2-methylph...	15.198	198	545293	53.617	ng	99
66) n-Nitrosodiphenylamine	15.321	169	2285297	48.302	ng	100
67) 4-Bromophenyl-phenylether	15.998	248	831353	49.667	ng	99
68) Hexachlorobenzene	16.098	284	983119	49.041	ng	99
69) Atrazine	16.292	200	662204	46.512	ng	99
70) Pentachlorophenol	16.457	266	719505	51.201	ng	98
71) Phenanthrene	16.839	178	4199256	48.170	ng	100
72) Anthracene	16.927	178	4107556	48.425	ng	100
73) Carbazole	17.209	167	4142913	47.836	ng	100
74) Di-n-butylphthalate	17.798	149	4924066	48.863	ng	100
75) Fluoranthene	18.874	202	5120629	48.293	ng	99
77) Benzidine	19.074	184	1399934	54.615	ng	100
78) Pyrene	19.239	202	5342784	49.505	ng	99
80) Butylbenzylphthalate	20.180	149	2261582	51.348	ng	97
81) Benzo(a)anthracene	21.015	228	4879819	48.632	ng	100
82) 3,3'-Dichlorobenzidine	20.956	252	1613787	51.403	ng	100
83) Chrysene	21.074	228	4586625	48.178	ng	99
84) Bis(2-ethylhexyl)phtha...	20.980	149	3143788	50.331	ng	99
85) Di-n-octyl phthalate	22.009	149	5477285	51.591	ng	99
87) Indeno(1,2,3-cd)pyrene	26.732	276	5881591	49.416	ng	100
88) Benzo(b)fluoranthene	22.891	252	5425845	49.618	ng	99
89) Benzo(k)fluoranthene	22.950	252	4875078	47.199	ng	99
90) Benzo(a)pyrene	23.609	252	4657835	49.170	ng	99
91) Dibenzo(a,h)anthracene	26.785	278	4698026	48.774	ng	100
92) Benzo(g,h,i)perylene	27.668	276	4981580	49.812	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081024\
Data File : BM047145.D
Acq On : 10 Aug 2024 16:31
Operator : RC/JU
Sample : SSTDICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTDICC050

Quant Time: Aug 11 23:34:27 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sun Aug 11 23:24:37 2024
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.375 min Scan# 763
Delta R.T. -0.000 min
Lab File: BM047145.D
Acq: 10 Aug 2024 16:31




Instrument :
BNA_M
ClientSampleId :
SSTDICC050


Tgt Ion:152 Resp: 315091
Ion Ratio Lower Upper
152 100
150 154.4 120.5 180.7
115 56.9 45.5 68.3



#2
1,4-Dioxane
Concen: 46.976 ng
RT: 3.034 min Scan# 25
Delta R.T. -0.000 min
Lab File: BM047145.D
Acq: 10 Aug 2024 16:31

Tgt Ion: 88 Resp: 343225
Ion Ratio Lower Upper
88 100
58 70.8 58.2 87.4
43 28.0 22.5 33.7









#3
Pyridine
Concen: 47.780 ng m
RT: 3.410 min Scan# 89
Delta R.T. -0.000 min
Lab File: BM047145.D
Acq: 10 Aug 2024 16:31




Instrument :
BNA_M
ClientSampleId :
SSTDICC050


Tgt Ion: 79 Resp: 1022963
Ion Ratio Lower Upper
79 100
52 61.3 51.2 76.8
51 31.5 26.1 39.1



#4
n-Nitrosodimethylamine
Concen: 47.479 ng
RT: 3.328 min Scan# 75
Delta R.T. -0.000 min
Lab File: BM047145.D
Acq: 10 Aug 2024 16:31

Tgt Ion: 42 Resp: 444430
Ion Ratio Lower Upper
42 100
74 141.7 112.9 169.3
44 6.4 5.3 7.9









#5
2-Fluorophenol
Concen: 93.844 ng
RT: 5.022 min Scan# 363
Delta R.T. -0.000 min
Lab File: BM047145.D
Acq: 10 Aug 2024 16:31




Instrument :
BNA_M
ClientSampleId :
SSTDICC050


Tgt Ion:112 Resp: 1654844
Ion Ratio Lower Upper
112 100
64 58.7 46.7 70.1
63 31.4 24.8 37.2



#6
Aniline
Concen: 48.069 ng
RT: 6.728 min Scan# 653
Delta R.T. -0.000 min
Lab File: BM047145.D
Acq: 10 Aug 2024 16:31

Tgt Ion: 93 Resp: 1103960
Ion Ratio Lower Upper
93 100
66 39.6 31.4 47.2
65 20.3 15.6 23.4









#7
Phenol-d6
Concen: 95.298 ng
RT: 6.587 min Scan# 61
Delta R.T. -0.000 min
Lab File: BM047145.D
Acq: 10 Aug 2024 16:31




Instrument :
BNA_M
ClientSampleId :
SSTDICC050


Tgt Ion: 99 Resp: 2315558
Ion Ratio Lower Upper
99 100
42 17.9 14.1 21.1
71 36.0 28.5 42.7



#8
2-Chlorophenol
Concen: 47.779 ng
RT: 6.951 min Scan# 691
Delta R.T. -0.000 min
Lab File: BM047145.D
Acq: 10 Aug 2024 16:31

Tgt Ion: 128 Resp: 918928
Ion Ratio Lower Upper
128 100
130 32.4 12.1 52.1
64 47.8 26.6 66.6









#9
Benzaldehyde
Concen: 53.942 ng
RT: 6.540 min Scan# 61
Delta R.T. -0.000 min
Lab File: BM047145.D
Acq: 10 Aug 2024 16:31




Instrument :
BNA_M
ClientSampleId :
SSTDICC050


Tgt Ion: 77 Resp: 590756
Ion Ratio Lower Upper
77 100
105 93.2 74.5 114.5
106 89.6 72.3 112.3



#10
Phenol
Concen: 47.123 ng
RT: 6.616 min Scan# 634
Delta R.T. 0.006 min
Lab File: BM047145.D
Acq: 10 Aug 2024 16:31

Tgt Ion: 94 Resp: 1142509
Ion Ratio Lower Upper
94 100
65 29.5 9.0 49.0
66 38.5 18.9 58.9









#11
bis(2-chloroethyl)ether
Concen: 48.057 ng
RT: 6.828 min Scan# 61
Delta R.T. -0.000 min
Lab File: BM047145.D
Acq: 10 Aug 2024 16:31




Instrument :
BNA_M
ClientSampleId :
SSTDICC050


Tgt Ion: 93 Resp: 997816
Ion Ratio Lower Upper
93 100
63 75.5 55.8 95.8
95 32.2 12.7 52.7



#12
1,3-Dichlorobenzene
Concen: 48.063 ng
RT: 7.263 min Scan# 744
Delta R.T. -0.000 min
Lab File: BM047145.D
Acq: 10 Aug 2024 16:31

Tgt Ion: 146 Resp: 1098474
Ion Ratio Lower Upper
146 100
148 64.6 52.3 78.5
75 32.4 25.8 38.6





#13

1,4-Dichlorobenzene

Concen: 47.843 ng

RT: 7.410 min Scan# 7

Delta R.T. -0.000 min

Lab File: BM047145.D

Acq: 10 Aug 2024 16:31

Instrument :

BNA_M

ClientSampleId :

SSTDICC050





Tgt Ion:146 Resp: 1107084

Ion Ratio Lower Upper

146 100

148 64.9 51.4 77.2

111 39.3 31.5 47.3



#14

1,2-Dichlorobenzene

Concen: 47.204 ng

RT: 7.722 min Scan# 822

Delta R.T. 0.006 min

Lab File: BM047145.D

Acq: 10 Aug 2024 16:31




Tgt Ion:146 Resp: 1030619


Ion Ratio Lower Upper

146 100

148 64.4 51.8 77.8

111 41.2 33.3 49.9









#15
Benzyl Alcohol
Concen: 49.018 ng
RT: 7.628 min Scan# 806
Delta R.T. -0.000 min
Lab File: BM047145.D
Acq: 10 Aug 2024 16:31




Instrument :
BNA_M
ClientSampleId :
SSTDICC050


Tgt Ion: 79 Resp: 847663
Ion Ratio Lower Upper
79 100
108 75.2 59.7 89.5
77 66.5 52.9 79.3



#16
2,2'-oxybis(1-Chloropropane)
Concen: 47.724 ng
RT: 7.916 min Scan# 855
Delta R.T. -0.000 min
Lab File: BM047145.D
Acq: 10 Aug 2024 16:31

Tgt Ion: 45 Resp: 1272990
Ion Ratio Lower Upper
45 100
77 15.6 0.0 35.6
79 12.4 0.0 32.4





#17

2-Methylphenol

Concen: 48.623 ng

RT: 7.834 min Scan# 841

Delta R.T. -0.000 min

Lab File: BM047145.D


Acq: 10 Aug 2024 16:31

Instrument :

BNA_M



ClientSampleId :

SSTDIC050



Tgt Ion:107 Resp: 786481

Ion	Ratio	Lower	Upper
107	100		
108	112.8	90.6	135.8
77	49.2	39.3	58.9
79	47.5	38.6	58.0



#18

Hexachloroethane

Concen: 48.568 ng

RT: 8.428 min Scan# 942




Delta R.T. -0.000 min


Lab File: BM047145.D

Acq: 10 Aug 2024 16:31

Tgt Ion:117 Resp: 435333

Ion	Ratio	Lower	Upper
117	100		
119	96.7	77.0	115.4
201	101.4	81.3	121.9









#19
n-Nitroso-di-n-propylamine
Concen: 46.435 ng
RT: 8.192 min Scan# 901
Delta R.T. 0.006 min
Lab File: BM047145.D
Acq: 10 Aug 2024 16:31




Instrument :
BNA_M
ClientSampleId :
SSTDICC050


Tgt Ion: 70 Resp: 759990
Ion Ratio Lower Upper
70 100
42 52.4 42.0 63.0
101 10.0 7.2 10.8
130 21.1 15.5 23.3



#20
3+4-Methylphenols
Concen: 48.940 ng
RT: 8.163 min Scan# 897
Delta R.T. -0.000 min
Lab File: BM047145.D
Acq: 10 Aug 2024 16:31

Tgt Ion: 107 Resp: 1106673
Ion Ratio Lower Upper
107 100
108 89.2 70.7 110.7
77 36.0 17.4 57.4
79 30.0 10.9 50.9









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.133 min Scan# 1232
Delta R.T. -0.000 min
Lab File: BM047145.D
Acq: 10 Aug 2024 16:31




Instrument :
BNA_M
ClientSampleId :
SSTDICC050


Tgt Ion:136 Resp: 1281047
Ion Ratio Lower Upper
136 100
137 10.6 8.8 13.2
54 7.2 6.0 9.0
68 6.2 4.8 7.2



#22
Acetophenone
Concen: 46.625 ng
RT: 8.198 min Scan# 903
Delta R.T. -0.000 min
Lab File: BM047145.D
Acq: 10 Aug 2024 16:31

Tgt Ion:105 Resp: 1441742
Ion Ratio Lower Upper
105 100
71 6.4 4.5 6.7
51 25.2 20.1 30.1
120 23.3 17.8 26.6





#23

Nitrobenzene-d5

Concen: 95.162 ng

RT: 8.528 min Scan# 91

Delta R.T. 0.006 min

Lab File: BM047145.D

Acq: 10 Aug 2024 16:31

Instrument :





BNA_M

ClientSampleId :

SSTDICC050

Tgt Ion: 82 Resp: 2421031

Ion	Ratio	Lower	Upper
82	100		
128	41.4	32.3	48.5
54	45.7	37.4	56.2



#24

Nitrobenzene

Concen: 47.470 ng

RT: 8.569 min Scan# 966




Delta R.T. 0.006 min


Lab File: BM047145.D

Acq: 10 Aug 2024 16:31

Tgt Ion: 77 Resp: 1216089

Ion	Ratio	Lower	Upper
77	100		
123	45.6	34.0	51.0
65	13.4	10.5	15.7









#25
Isophorone
Concen: 48.292 ng
RT: 9.092 min Scan# 1055
Delta R.T. -0.000 min
Lab File: BM047145.D
Acq: 10 Aug 2024 16:31




Instrument :
BNA_M
ClientSampleId :
SSTDICC050


Tgt Ion: 82 Resp: 2139220
Ion Ratio Lower Upper
82 100
95 6.8 5.5 8.3
138 17.0 13.5 20.3



#26
2-Nitrophenol
Concen: 51.343 ng
RT: 9.263 min Scan# 1084
Delta R.T. -0.000 min
Lab File: BM047145.D
Acq: 10 Aug 2024 16:31

Tgt Ion:139 Resp: 582500
Ion Ratio Lower Upper
139 100
109 30.1 24.4 36.6
65 52.0 43.0 64.6





#27

2,4-Dimethylphenol

Concen: 49.545 ng

RT: 9.345 min Scan# 1098

Delta R.T. -0.000 min

Lab File: BM047145.D

Acq: 10 Aug 2024 16:31

Instrument :

BNA_M

ClientSampleId :

SSTDICC050





Tgt Ion:122 Resp: 658437

Ion Ratio Lower Upper

122 100

107 116.1 92.8 139.2

121 58.1 46.1 69.1



#28

bis(2-Chloroethoxy)methane

Concen: 48.339 ng

RT: 9.575 min Scan# 1137

Delta R.T. -0.000 min

Lab File: BM047145.D

Acq: 10 Aug 2024 16:31




Tgt Ion: 93 Resp: 1347358


Ion Ratio Lower Upper

93 100

95 32.5 25.8 38.8

123 11.1 9.0 13.4





#29

2,4-Dichlorophenol

Concen: 50.395 ng

RT: 9.798 min Scan# 1175

Delta R.T. -0.000 min

Lab File: BM047145.D


Acq: 10 Aug 2024 16:31

Instrument :

BNA_M

ClientSampleId :

SSTDICC050






Tgt Ion:162 Resp: 993725

Ion Ratio Lower Upper

162 100

164 65.2 44.7 84.7

98 34.4 14.9 54.9



#30

1,2,4-Trichlorobenzene


Concen: 48.819 ng

RT: 9.998 min Scan# 1209

Delta R.T. -0.000 min

Lab File: BM047145.D

Acq: 10 Aug 2024 16:31





Tgt Ion:180 Resp: 1089421


Ion Ratio Lower Upper

180 100

182 97.1 76.6 115.0

145 29.3 23.4 35.2





#31

Naphthalene

Concen: 47.619 ng

RT: 10.180 min Scan# 1130

Delta R.T. -0.000 min

Lab File: BM047145.D

Acq: 10 Aug 2024 16:31

Instrument :

BNA_M

ClientSampleId :

SSTDICC050





Tgt Ion:128 Resp: 3038911

Ion Ratio Lower Upper

128 100

129 10.9 8.6 13.0

127 13.1 10.2 15.4



#32

Benzoic acid

Concen: 53.976 ng

RT: 9.539 min Scan# 1131

Delta R.T. 0.006 min

Lab File: BM047145.D

Acq: 10 Aug 2024 16:31




Tgt Ion:122 Resp: 838043


Ion Ratio Lower Upper

122 100

105 123.3 107.1 147.1

77 99.9 82.1 122.1





#33

4-Chloroaniline

Concen: 48.793 ng

RT: 10.310 min Scan# 1262

Delta R.T. -0.000 min

Lab File: BM047145.D

Acq: 10 Aug 2024 16:31

Instrument :

BNA_M

ClientSampleId :

SSTDIC050

Tgt Ion:127 Resp: 1197445





Ion	Ratio	Lower	Upper

127	100		

129	32.8	25.8	38.8

65	30.9	24.9	37.3

92	19.0	15.1	22.7



#34

Hexachlorobutadiene

Concen: 49.773 ng

RT: 10.469 min Scan# 1289

Delta R.T. -0.000 min

Lab File: BM047145.D

Acq: 10 Aug 2024 16:31




Tgt Ion:225 Resp: 650107


Ion	Ratio	Lower	Upper

225	100		

223	62.0	50.2	75.2

227	64.4	51.4	77.0









#35
Caprolactam
Concen: 50.259 ng
RT: 11.127 min Scan# 1400
Delta R.T. 0.006 min
Lab File: BM047145.D
Acq: 10 Aug 2024 16:31




Instrument :
BNA_M
ClientSampleId :
SSTDICC050


Tgt Ion:113 Resp: 344173
Ion Ratio Lower Upper
113 100
55 160.6 140.5 180.5
56 120.7 103.6 143.6



#36
4-Chloro-3-methylphenol
Concen: 49.201 ng
RT: 11.457 min Scan# 1457
Delta R.T. 0.006 min
Lab File: BM047145.D
Acq: 10 Aug 2024 16:31

Tgt Ion:107 Resp: 1033238
Ion Ratio Lower Upper
107 100
144 27.8 21.8 32.6
142 83.9 66.8 100.2









#37
2-Methylnaphthalene
Concen: 47.987 ng
RT: 11.810 min Scan# 1516
Delta R.T. 0.006 min
Lab File: BM047145.D
Acq: 10 Aug 2024 16:31




Instrument :
BNA_M
ClientSampleId :
SSTDICC050


Tgt Ion:142 Resp: 2180425
Ion Ratio Lower Upper
142 100
141 87.4 70.9 106.3
115 35.0 28.7 43.1



#38
1-Methylnaphthalene
Concen: 47.673 ng
RT: 12.033 min Scan# 1555
Delta R.T. -0.000 min
Lab File: BM047145.D
Acq: 10 Aug 2024 16:31

Tgt Ion:142 Resp: 2140830
Ion Ratio Lower Upper
142 100
141 90.4 74.6 111.8
116 3.7 3.1 4.7









#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.033 min Scan# 1895
Delta R.T. -0.000 min
Lab File: BM047145.D
Acq: 10 Aug 2024 16:31




Instrument :
BNA_M
ClientSampleId :
SSTDICC050


Tgt Ion:164 Resp: 834055
Ion Ratio Lower Upper
164 100
162 100.7 80.7 121.1
160 45.7 36.0 54.0



#40
1,2,4,5-Tetrachlorobenzene
Concen: 50.043 ng
RT: 12.186 min Scan# 1581
Delta R.T. -0.000 min
Lab File: BM047145.D
Acq: 10 Aug 2024 16:31

Tgt Ion:216 Resp: 1138431
Ion Ratio Lower Upper
216 100
214 78.7 63.4 95.2
179 21.7 17.6 26.4
108 18.7 15.1 22.7









#41
Hexachlorocyclopentadiene
Concen: 52.261 ng
RT: 12.163 min Scan# 1577
Delta R.T. -0.000 min
Lab File: BM047145.D
Acq: 10 Aug 2024 16:31




Instrument :
BNA_M
ClientSampleId :
SSTDICC050


Tgt Ion:237 Resp: 415149
Ion Ratio Lower Upper
237 100
235 64.7 44.8 84.8
272 13.0 0.0 33.5



#42
2,4,6-Tribromophenol
Concen: 98.476 ng
RT: 15.545 min Scan# 2152
Delta R.T. 0.006 min
Lab File: BM047145.D
Acq: 10 Aug 2024 16:31

Tgt Ion:330 Resp: 952366
Ion Ratio Lower Upper
330 100
332 97.0 77.4 116.2
141 39.2 31.6 47.4





#43

2,4,6-Trichlorophenol

Concen: 50.831 ng

RT: 12.445 min Scan# 1

Delta R.T. -0.000 min

Lab File: BM047145.D

Acq: 10 Aug 2024 16:31

Instrument :

BNA_M

ClientSampleId :

SSTDICC050





Tgt Ion:196 Resp: 821910

Ion Ratio Lower Upper

196 100

198 96.2 75.6 113.4

200 30.7 24.6 37.0



#44

2,4,5-Trichlorophenol

Concen: 50.644 ng

RT: 12.521 min Scan# 1638

Delta R.T. -0.000 min

Lab File: BM047145.D

Acq: 10 Aug 2024 16:31

Tgt Ion:196 Resp: 909210




Ion Ratio Lower Upper


196 100

198 96.2 77.7 116.5

97 50.6 39.8 59.8

132 26.3 20.7 31.1





#45

2-Fluorobiphenyl

Concen: 96.172 ng

RT: 12.645 min Scan# 1659

Delta R.T. 0.000 min

Lab File: BM047145.D

Acq: 10 Aug 2024 16:31

Instrument :

BNA_M

ClientSampleId :

SSTDICC050





Tgt Ion:172 Resp: 5200055

Ion Ratio Lower Upper

172 100

171 34.9 27.8 41.6

170 23.4 18.7 28.1



#46

1,1'-Biphenyl

Concen: 47.782 ng

RT: 12.857 min Scan# 1695

Delta R.T. -0.000 min

Lab File: BM047145.D

Acq: 10 Aug 2024 16:31




Tgt Ion:154 Resp: 2837162


Ion Ratio Lower Upper

154 100

153 39.9 19.0 59.0

76 12.2 0.0 31.9









#47
2-Chloronaphthalene
Concen: 47.627 ng
RT: 12.892 min Scan# 1701
Delta R.T. 0.006 min
Lab File: BM047145.D
Acq: 10 Aug 2024 16:31




Instrument :
BNA_M
ClientSampleId :
SSTDICC050


Tgt Ion:162 Resp: 2210354
Ion Ratio Lower Upper
162 100
127 36.7 30.0 45.0
164 33.4 26.3 39.5



#48
2-Nitroaniline
Concen: 50.459 ng
RT: 13.116 min Scan# 1739
Delta R.T. -0.000 min
Lab File: BM047145.D
Acq: 10 Aug 2024 16:31

Tgt Ion: 65 Resp: 733636
Ion Ratio Lower Upper
65 100
92 68.7 55.1 82.7
138 104.8 85.9 128.9









#49
Acenaphthylene
Concen: 48.009 ng
RT: 13.751 min Scan# 1847
Delta R.T. -0.000 min
Lab File: BM047145.D
Acq: 10 Aug 2024 16:31




Instrument :
BNA_M
ClientSampleId :
SSTDICC050


Tgt Ion:152 Resp: 3284244
Ion Ratio Lower Upper
152 100
151 20.0 15.8 23.6
153 13.0 10.4 15.6



#50
Dimethylphthalate
Concen: 48.073 ng
RT: 13.516 min Scan# 1807
Delta R.T. -0.000 min
Lab File: BM047145.D
Acq: 10 Aug 2024 16:31

Tgt Ion:163 Resp: 2800312
Ion Ratio Lower Upper
163 100
194 4.0 3.0 4.6
164 10.3 8.0 12.0









#51
2,6-Dinitrotoluene
Concen: 49.929 ng
RT: 13.633 min Scan# 1827
Delta R.T. 0.006 min
Lab File: BM047145.D
Acq: 10 Aug 2024 16:31




Instrument :
BNA_M
ClientSampleId :
SSTDICC050


Tgt Ion:165 Resp: 679884
Ion Ratio Lower Upper
165 100
63 53.9 46.5 69.7
89 55.3 47.3 70.9



#52
Acenaphthene
Concen: 47.640 ng
RT: 14.098 min Scan# 1906
Delta R.T. -0.000 min
Lab File: BM047145.D
Acq: 10 Aug 2024 16:31

Tgt Ion:154 Resp: 2067281
Ion Ratio Lower Upper
154 100
153 111.8 90.3 135.5
152 53.2 42.9 64.3









#53
3-Nitroaniline
Concen: 50.067 ng
RT: 13.963 min Scan# 1883
Delta R.T. -0.000 min
Lab File: BM047145.D
Acq: 10 Aug 2024 16:31




Instrument :
BNA_M
ClientSampleId :
SSTDICC050


Tgt Ion:138 Resp: 683158
Ion Ratio Lower Upper
138 100
108 9.3 7.9 11.9
92 115.1 92.0 138.0



#54
2,4-Dinitrophenol
Concen: 52.012 ng
RT: 14.174 min Scan# 1919
Delta R.T. -0.001 min
Lab File: BM047145.D
Acq: 10 Aug 2024 16:31

Tgt Ion:184 Resp: 424486
Ion Ratio Lower Upper
184 100
63 64.2 52.7 79.1
154 62.6 50.4 75.6





#55

Dibenzenofuran

Concen: 47.413 ng

RT: 14.439 min Scan# 1964

Delta R.T. -0.000 min

Lab File: BM047145.D

Acq: 10 Aug 2024 16:31

Instrument :





BNA_M

ClientSampleId :

SSTDICC050

Tgt Ion:168 Resp: 3249834

Ion	Ratio	Lower	Upper
168	100		
139	36.9	29.6	44.4
169	13.3	10.7	16.1



#56

4-Nitrophenol

Concen: 51.542 ng

RT: 14.298 min Scan# 1940




Delta R.T. -0.000 min


Lab File: BM047145.D

Acq: 10 Aug 2024 16:31

Tgt Ion:139 Resp: 606437

Ion	Ratio	Lower	Upper
139	100		
109	75.7	52.5	92.5
65	99.9	75.4	115.4









#57
2,4-Dinitrotoluene
Concen: 50.482 ng
RT: 14.427 min Scan# 1962
Delta R.T. -0.000 min
Lab File: BM047145.D
Acq: 10 Aug 2024 16:31




Instrument :
BNA_M
ClientSampleId :
SSTDICC050


Tgt Ion:165 Resp: 917536
Ion Ratio Lower Upper
165 100
63 50.4 40.3 60.5
89 76.1 61.0 91.6
182 2.5 1.9 2.9



#58
Fluorene
Concen: 47.578 ng
RT: 15.098 min Scan# 2076
Delta R.T. 0.006 min
Lab File: BM047145.D
Acq: 10 Aug 2024 16:31

Tgt Ion:166 Resp: 2690806
Ion Ratio Lower Upper
166 100
165 96.6 77.4 116.0
167 13.4 11.0 16.6









#59
2,3,4,6-Tetrachlorophenol
Concen: 50.022 ng
RT: 14.674 min Scan# 2004
Delta R.T. -0.000 min
Lab File: BM047145.D
Acq: 10 Aug 2024 16:31




Instrument :
BNA_M
ClientSampleId :
SSTDICC050


Tgt Ion:232 Resp: 738067
Ion Ratio Lower Upper
232 100
131 45.1 36.9 55.3
130 2.5 2.0 3.0
166 32.9 26.4 39.6



#60
Diethylphthalate
Concen: 47.851 ng
RT: 14.898 min Scan# 2042
Delta R.T. -0.000 min
Lab File: BM047145.D
Acq: 10 Aug 2024 16:31

Tgt Ion:149 Resp: 2846430
Ion Ratio Lower Upper
149 100
177 22.2 17.6 26.4
150 12.4 9.8 14.6









#61
4-Chlorophenyl-phenylether
Concen: 47.901 ng
RT: 15.104 min Scan# 2076
Delta R.T. 0.006 min
Lab File: BM047145.D
Acq: 10 Aug 2024 16:31




Instrument :
BNA_M
ClientSampleId :
SSTDICC050


Tgt Ion:204 Resp: 1289728
Ion Ratio Lower Upper
204 100
206 32.4 26.2 39.4
141 56.5 47.2 70.8



#62
4-Nitroaniline
Concen: 49.229 ng
RT: 15.139 min Scan# 2083
Delta R.T. -0.000 min
Lab File: BM047145.D
Acq: 10 Aug 2024 16:31

Tgt Ion:138 Resp: 661759
Ion Ratio Lower Upper
138 100
92 52.0 31.9 71.9
108 85.1 63.8 103.8









#63
Azobenzene
Concen: 47.334 ng
RT: 15.398 min Scan# 2126
Delta R.T. 0.006 min
Lab File: BM047145.D
Acq: 10 Aug 2024 16:31




Instrument :
BNA_M
ClientSampleId :
SSTDICC050


Tgt Ion: 77 Resp: 2669713
Ion Ratio Lower Upper
77 100
182 25.8 3.8 43.8
105 20.2 0.0 39.5
51 26.2 6.8 46.8



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 16.792 min Scan# 2364
Delta R.T. -0.000 min
Lab File: BM047145.D
Acq: 10 Aug 2024 16:31

Tgt Ion: 188 Resp: 1737042
Ion Ratio Lower Upper
188 100
94 9.3 7.0 10.6
80 10.4 8.0 12.0









#65
4,6-Dinitro-2-methylphenol
Concen: 53.617 ng
RT: 15.198 min Scan# 2093
Delta R.T. -0.000 min
Lab File: BM047145.D
Acq: 10 Aug 2024 16:31




Instrument :
BNA_M
ClientSampleId :
SSTDICC050


Tgt Ion:198 Resp: 545293
Ion Ratio Lower Upper
198 100
51 43.6 23.9 63.9
105 43.6 24.3 64.3



#66
n-Nitrosodiphenylamine
Concen: 48.302 ng
RT: 15.321 min Scan# 2114
Delta R.T. -0.000 min
Lab File: BM047145.D
Acq: 10 Aug 2024 16:31

Tgt Ion:169 Resp: 2285297
Ion Ratio Lower Upper
169 100
168 67.6 53.8 80.8
167 36.5 29.3 43.9









#67
4-Bromophenyl-phenylether
Concen: 49.667 ng
RT: 15.998 min Scan# 211
Delta R.T. -0.000 min
Lab File: BM047145.D
Acq: 10 Aug 2024 16:31




Instrument :
BNA_M
ClientSampleId :
SSTDICC050


Tgt Ion:	248	Resp:	831353
Ion Ratio	Lower	Upper	
248	100		
250	99.3	77.8	116.8
141	71.6	57.7	86.5



#68
Hexachlorobenzene
Concen: 49.041 ng
RT: 16.098 min Scan# 2246
Delta R.T. -0.000 min
Lab File: BM047145.D
Acq: 10 Aug 2024 16:31

Tgt Ion:	284	Resp:	983119
Ion Ratio	Lower	Upper	
284	100		
142	37.0	29.4	44.0
249	30.9	25.1	37.7









#69
Atrazine
Concen: 46.512 ng
RT: 16.292 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM047145.D
Acq: 10 Aug 2024 16:31




Instrument :
BNA_M
ClientSampleId :
SSTDICC050


Tgt Ion:200 Resp: 662204
Ion Ratio Lower Upper
200 100
173 26.1 6.1 46.1
215 47.5 26.8 66.8



#70
Pentachlorophenol
Concen: 51.201 ng
RT: 16.457 min Scan# 2307
Delta R.T. 0.006 min
Lab File: BM047145.D
Acq: 10 Aug 2024 16:31

Tgt Ion:266 Resp: 719505
Ion Ratio Lower Upper
266 100
268 64.9 49.8 74.6
264 62.7 49.6 74.4









#71
Phenanthrene
Concen: 48.170 ng
RT: 16.839 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM047145.D
Acq: 10 Aug 2024 16:31




Instrument :
BNA_M
ClientSampleId :
SSTDICC050


Tgt Ion:178 Resp: 4199256
Ion Ratio Lower Upper
178 100
176 19.0 15.3 22.9
179 15.4 12.4 18.6



#72
Anthracene
Concen: 48.425 ng
RT: 16.927 min Scan# 2387
Delta R.T. -0.000 min
Lab File: BM047145.D
Acq: 10 Aug 2024 16:31

Tgt Ion:178 Resp: 4107556
Ion Ratio Lower Upper
178 100
176 18.8 14.8 22.2
179 15.5 12.4 18.6









#73
Carbazole
Concen: 47.836 ng
RT: 17.209 min Scan# 2435
Delta R.T. -0.000 min
Lab File: BM047145.D
Acq: 10 Aug 2024 16:31




Instrument :
BNA_M
ClientSampleId :
SSTDICC050


Tgt Ion:167 Resp: 4142913
Ion Ratio Lower Upper
167 100
166 21.4 17.1 25.7
139 12.5 10.0 15.0



#74
Di-n-butylphthalate
Concen: 48.863 ng
RT: 17.798 min Scan# 2535
Delta R.T. 0.006 min
Lab File: BM047145.D
Acq: 10 Aug 2024 16:31

Tgt Ion:149 Resp: 4924066
Ion Ratio Lower Upper
149 100
150 9.3 7.5 11.3
104 4.9 4.1 6.1









#75
Fluoranthene
Concen: 48.293 ng
RT: 18.874 min Scan# 21
Delta R.T. 0.006 min
Lab File: BM047145.D
Acq: 10 Aug 2024 16:31




Instrument :
BNA_M
ClientSampleId :
SSTDIC050


Tgt Ion:202 Resp: 5120629
Ion Ratio Lower Upper
202 100
101 9.6 0.0 30.2
203 17.6 0.0 37.6



#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.033 min Scan# 3085
Delta R.T. -0.000 min
Lab File: BM047145.D
Acq: 10 Aug 2024 16:31

Tgt Ion:240 Resp: 1511494
Ion Ratio Lower Upper
240 100
120 10.3 8.2 12.2
236 25.8 20.6 30.8









#77
Benzidine
Concen: 54.615 ng
RT: 19.074 min Scan# 2752
Delta R.T. -0.000 min
Lab File: BM047145.D
Acq: 10 Aug 2024 16:31




Instrument :
BNA_M
ClientSampleId :
SSTDICC050


Tgt Ion:184 Resp: 1399934
Ion Ratio Lower Upper
184 100
185 14.5 11.8 17.6
183 11.3 9.0 13.6



#78
Pyrene
Concen: 49.505 ng
RT: 19.239 min Scan# 2780
Delta R.T. -0.000 min
Lab File: BM047145.D
Acq: 10 Aug 2024 16:31

Tgt Ion:202 Resp: 5342784
Ion Ratio Lower Upper
202 100
200 21.1 17.2 25.8
203 17.6 14.2 21.2









#79
Terphenyl-d14
Concen: 99.443 ng
RT: 19.462 min Scan# 2818
Delta R.T. -0.000 min
Lab File: BM047145.D
Acq: 10 Aug 2024 16:31




Instrument :
BNA_M
ClientSampleId :
SSTDICC050


Tgt Ion:244 Resp: 7014667
Ion Ratio Lower Upper
244 100
212 7.9 6.2 9.4
122 8.9 7.1 10.7



#80
Butylbenzylphthalate
Concen: 51.348 ng
RT: 20.180 min Scan# 2940
Delta R.T. 0.006 min
Lab File: BM047145.D
Acq: 10 Aug 2024 16:31

Tgt Ion:149 Resp: 2261582
Ion Ratio Lower Upper
149 100
91 68.1 57.0 85.4
206 21.2 16.4 24.6









#81
Benzo(a)anthracene
Concen: 48.632 ng
RT: 21.015 min Scan# 3082
Delta R.T. -0.000 min
Lab File: BM047145.D
Acq: 10 Aug 2024 16:31




Instrument :
BNA_M
ClientSampleId :
SSTDICC050


Tgt Ion:228 Resp: 4879819
Ion Ratio Lower Upper
228 100
226 27.4 21.8 32.6
229 19.9 15.8 23.8



#82
3,3'-Dichlorobenzidine
Concen: 51.403 ng
RT: 20.956 min Scan# 3072
Delta R.T. -0.000 min
Lab File: BM047145.D
Acq: 10 Aug 2024 16:31

Tgt Ion:252 Resp: 1613787
Ion Ratio Lower Upper
252 100
254 64.0 51.0 76.6
126 10.9 8.5 12.7









#83
Chrysene
Concen: 48.178 ng
RT: 21.074 min Scan# 3092
Delta R.T. -0.000 min
Lab File: BM047145.D
Acq: 10 Aug 2024 16:31




Instrument :
BNA_M
ClientSampleId :
SSTDICC050


Tgt Ion:228 Resp: 4586625
Ion Ratio Lower Upper
228 100
226 30.7 24.4 36.6
229 20.0 15.7 23.5



#84
Bis(2-ethylhexyl)phthalate
Concen: 50.331 ng
RT: 20.980 min Scan# 3076
Delta R.T. 0.006 min
Lab File: BM047145.D
Acq: 10 Aug 2024 16:31

Tgt Ion:149 Resp: 3143788
Ion Ratio Lower Upper
149 100
167 27.9 22.2 33.2
279 4.3 3.0 4.4









#85
Di-n-octyl phthalate
Concen: 51.591 ng
RT: 22.009 min Scan# 31
Delta R.T. -0.000 min
Lab File: BM047145.D
Acq: 10 Aug 2024 16:31




Instrument :
BNA_M
ClientSampleId :
SSTDICC050


Tgt Ion:149 Resp: 5477285
Ion Ratio Lower Upper
149 100
167 1.6 1.3 1.9
43 8.4 6.9 10.3



#86
Perylene-d12
Concen: 20.000 ng
RT: 23.733 min Scan# 3544
Delta R.T. 0.006 min
Lab File: BM047145.D
Acq: 10 Aug 2024 16:31

Tgt Ion:264 Resp: 1841387
Ion Ratio Lower Upper
264 100
260 23.2 18.5 27.7
265 21.8 17.3 25.9









#87
Indeno(1,2,3-cd)pyrene
Concen: 49.416 ng
RT: 26.732 min Scan# 4054
Delta R.T. 0.006 min
Lab File: BM047145.D
Acq: 10 Aug 2024 16:31




Instrument :
BNA_M
ClientSampleId :
SSTDICC050


Tgt Ion:276 Resp: 5881591
Ion Ratio Lower Upper
276 100
138 23.9 19.4 29.0
277 25.0 20.1 30.1



#88
Benzo(b)fluoranthene
Concen: 49.618 ng
RT: 22.891 min Scan# 3401
Delta R.T. 0.006 min
Lab File: BM047145.D
Acq: 10 Aug 2024 16:31

Tgt Ion:252 Resp: 5425845
Ion Ratio Lower Upper
252 100
253 22.5 17.8 26.6
125 9.0 7.4 11.0





#89

Benzo(k)fluoranthene

Concen: 47.199 ng

RT: 22.950 min Scan# 3411

Delta R.T. 0.006 min

Lab File: BM047145.D

Acq: 10 Aug 2024 16:31

Instrument :

BNA_M

ClientSampleId :

SSTDIC050





Tgt Ion:252 Resp: 4875078

Ion Ratio Lower Upper

252 100

253 21.9 17.7 26.5

125 8.7 7.4 11.2



#90

Benzo(a)pyrene

Concen: 49.170 ng

RT: 23.609 min Scan# 3523

Delta R.T. -0.000 min

Lab File: BM047145.D

Acq: 10 Aug 2024 16:31




Tgt Ion:252 Resp: 4657835


Ion Ratio Lower Upper

252 100

253 21.8 17.1 25.7

125 10.2 8.2 12.2









#91
Dibenzo(a,h)anthracene
Concen: 48.774 ng
RT: 26.785 min Scan# 4063
Delta R.T. 0.006 min
Lab File: BM047145.D
Acq: 10 Aug 2024 16:31




Instrument :
BNA_M
ClientSampleId :
SSTDICC050

Tgt Ion:278 Resp: 4698026
Ion Ratio Lower Upper
278 100
139 15.2 12.0 18.0
279 23.9 19.2 28.8



#92
Benzo(g,h,i)perylene
Concen: 49.812 ng
RT: 27.668 min Scan# 4213
Delta R.T. 0.006 min
Lab File: BM047145.D
Acq: 10 Aug 2024 16:31

Tgt Ion:276 Resp: 4981580
Ion Ratio Lower Upper
276 100
277 23.8 18.8 28.2
138 20.7 16.2 24.4



Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081024\
 Data File : BM047146.D
 Acq On : 10 Aug 2024 17:11
 Operator : RC/JU
 Sample : SSTDICC060
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC060

Quant Time: Aug 11 23:35:12 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Aug 11 23:24:37 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.375	152	381037	20.000	ng	0.00
21) Naphthalene-d8	10.133	136	1526728	20.000	ng	0.00
39) Acenaphthene-d10	14.033	164	991306	20.000	ng	0.00
64) Phenanthrene-d10	16.798	188	2082135	20.000	ng	0.00
76) Chrysene-d12	21.033	240	1812683	20.000	ng	0.00
86) Perylene-d12	23.733	264	2151546	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.028	112	2449346	114.860	ng	0.00
7) Phenol-d6	6.593	99	3414785	116.214	ng	0.00
23) Nitrobenzene-d5	8.528	82	3495384	115.282	ng	0.00
42) 2,4,6-Tribromophenol	15.545	330	1391826	121.088	ng	0.00
45) 2-Fluorobiphenyl	12.645	172	7500034	116.705	ng	0.00
79) Terphenyl-d14	19.462	244	10315735	121.942	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.034	88	505618	57.225	ng	98
3) Pyridine	3.410	79	1511433m	58.377	ng	
4) n-Nitrosodimethylamine	3.334	42	666445	58.875	ng	99
6) Aniline	6.728	93	1595030	57.431	ng	100
8) 2-Chlorophenol	6.957	128	1333099	57.317	ng	98
9) Benzaldehyde	6.540	77	733143	58.359	ng	100
10) Phenol	6.616	94	1688399	57.586	ng	99
11) bis(2-Chloroethyl)ether	6.834	93	1446822	57.623	ng	100
12) 1,3-Dichlorobenzene	7.263	146	1592469	57.619	ng	98
13) 1,4-Dichlorobenzene	7.410	146	1614359	57.691	ng	99
14) 1,2-Dichlorobenzene	7.722	146	1495805	56.653	ng	99
15) Benzyl Alcohol	7.634	79	1258590	60.184	ng	99
16) 2,2'-oxybis(1-Chloropr...	7.916	45	1768027	54.811	ng	99
17) 2-Methylphenol	7.839	107	1134669	58.008	ng	99
18) Hexachloroethane	8.428	117	639660	59.012	ng	100
19) n-Nitroso-di-n-propyla...	8.192	70	1096068	55.379	ng	99
20) 3+4-Methylphenols	8.169	107	1605895	58.726	ng	99
22) Acetophenone	8.198	105	2070558	56.185	ng	# 98
24) Nitrobenzene	8.569	77	1733451	56.777	ng	100
25) Isophorone	9.098	82	3108196	58.876	ng	98
26) 2-Nitrophenol	9.269	139	865172	63.987	ng	97
27) 2,4-Dimethylphenol	9.345	122	957294	60.441	ng	98
28) bis(2-Chloroethoxy)met...	9.581	93	1895700	57.068	ng	99
29) 2,4-Dichlorophenol	9.798	162	1451562	61.767	ng	100
30) 1,2,4-Trichlorobenzene	10.004	180	1589392	59.763	ng	99
31) Naphthalene	10.186	128	4415000	58.049	ng	100
32) Benzoic acid	9.575	122	1248981	67.498	ng	99
33) 4-Chloroaniline	10.316	127	1718710	58.764	ng	100
34) Hexachlorobutadiene	10.469	225	965378	62.017	ng	100
35) Caprolactam	11.151	113	503257m	61.664	ng	
36) 4-Chloro-3-methylphenol	11.463	107	1512201	60.420	ng	100
37) 2-Methylnaphthalene	11.810	142	3194966	59.000	ng	99
38) 1-Methylnaphthalene	12.033	142	3159405	59.033	ng	97
40) 1,2,4,5-Tetrachloroben...	12.192	216	1663474	61.523	ng	99
41) Hexachlorocyclopentadiene	12.169	237	620612	65.733	ng	97
43) 2,4,6-Trichlorophenol	12.451	196	1202229	62.557	ng	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081024\
 Data File : BM047146.D
 Acq On : 10 Aug 2024 17:11
 Operator : RC/JU
 Sample : SSTDICC060
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC060


Quant Time: Aug 11 23:35:12 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Aug 11 23:24:37 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	12.522	196	1329069	62.287	ng	98
46) 1,1'-Biphenyl	12.857	154	4126067	58.466	ng	99
47) 2-Chloronaphthalene	12.892	162	3192787	57.882	ng	99
48) 2-Nitroaniline	13.121	65	1051464	60.847	ng	100
49) Acenaphthylene	13.751	152	4765674	58.613	ng	100
50) Dimethylphthalate	13.521	163	4059126	58.629	ng	99
51) 2,6-Dinitrotoluene	13.633	165	995456	61.508	ng	96
52) Acenaphthene	14.104	154	3025820	58.668	ng	99
53) 3-Nitroaniline	13.969	138	1000744	61.708	ng	100
54) 2,4-Dinitrophenol	14.180	184	647477	66.750	ng	97
55) Dibenzofuran	14.445	168	4709496	57.809	ng	99
56) 4-Nitrophenol	14.304	139	879989	62.927	ng	96
57) 2,4-Dinitrotoluene	14.433	165	1323234	61.255	ng	98
58) Fluorene	15.098	166	3885304	57.801	ng	100
59) 2,3,4,6-Tetrachlorophenol	14.680	232	1088974	62.097	ng	98
60) Diethylphthalate	14.904	149	4128072	58.389	ng	99
61) 4-Chlorophenyl-phenyle...	15.104	204	1916971	59.903	ng	96
62) 4-Nitroaniline	15.145	138	978393	61.238	ng	99
63) Azobenzene	15.398	77	3820106	56.986	ng	98
65) 4,6-Dinitro-2-methylph...	15.204	198	799367	65.573	ng	96
66) n-Nitrosodiphenylamine	15.327	169	3326078	58.648	ng	99
67) 4-Bromophenyl-phenylether	15.998	248	1234868	61.547	ng	98
68) Hexachlorobenzene	16.104	284	1431249	59.562	ng	97
69) Atrazine	16.298	200	906491	53.118	ng	99
70) Pentachlorophenol	16.457	266	1086504	64.503	ng	99
71) Phenanthrene	16.839	178	6082429	58.208	ng	100
72) Anthracene	16.933	178	6001835	59.030	ng	99
73) Carbazole	17.209	167	5990202	57.702	ng	100
74) Di-n-butylphthalate	17.798	149	7121063	58.952	ng	100
75) Fluoranthene	18.874	202	7540518	59.328	ng	99
77) Benzidine	19.074	184	2113957	68.768	ng	100
78) Pyrene	19.239	202	7862143	60.744	ng	100
80) Butylbenzylphthalate	20.180	149	3347760	63.380	ng	96
81) Benzo(a)anthracene	21.015	228	7197782	59.814	ng	99
82) 3,3'-Dichlorobenzidine	20.956	252	2319044	61.593	ng	99
83) Chrysene	21.074	228	6712104	58.790	ng	99
84) Bis(2-ethylhexyl)phtha...	20.980	149	4599112	61.396	ng	# 99
85) Di-n-octyl phthalate	22.009	149	8171642	64.180	ng	99
87) Indeno(1,2,3-cd)pyrene	26.744	276	7917132	56.929	ng	99
88) Benzo(b)fluoranthene	22.897	252	7789248	60.962	ng	99
89) Benzo(k)fluoranthene	22.950	252	7202578	59.680	ng	99
90) Benzo(a)pyrene	23.615	252	6751397	60.997	ng	99
91) Dibenzo(a,h)anthracene	26.791	278	6367453	56.576	ng	100
92) Benzo(g,h,i)perylene	27.679	276	6483852	55.488	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
BNA_M
ClientSampleId :
SSTDICC060





[illegible]



#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.375 min Scan# 763
Delta R.T. -0.000 min
Lab File: BM047146.D
Acq: 10 Aug 2024 17:11




Instrument :
BNA_M
ClientSampleId :
SSTDICC060


Tgt Ion:152 Resp: 381037
Ion Ratio Lower Upper
152 100
150 155.6 120.5 180.7
115 55.7 45.5 68.3



#2
1,4-Dioxane
Concen: 57.225 ng
RT: 3.034 min Scan# 25
Delta R.T. -0.000 min
Lab File: BM047146.D
Acq: 10 Aug 2024 17:11

Tgt Ion: 88 Resp: 505618
Ion Ratio Lower Upper
88 100
58 71.2 58.2 87.4
43 27.9 22.5 33.7









#3
Pyridine
Concen: 58.377 ng m
RT: 3.410 min Scan# 89
Delta R.T. -0.000 min
Lab File: BM047146.D
Acq: 10 Aug 2024 17:11




Instrument :
BNA_M
ClientSampleId :
SSTDICC060


Tgt Ion: 79 Resp: 1511433
Ion Ratio Lower Upper
79 100
52 63.6 51.2 76.8
51 32.5 26.1 39.1



#4
n-Nitrosodimethylamine
Concen: 58.875 ng
RT: 3.334 min Scan# 76
Delta R.T. 0.006 min
Lab File: BM047146.D
Acq: 10 Aug 2024 17:11

Tgt Ion: 42 Resp: 666445
Ion Ratio Lower Upper
42 100
74 142.7 112.9 169.3
44 6.1 5.3 7.9





#5

2-Fluorophenol

Concen: 114.860 ng

RT: 5.028 min Scan# 3

Delta R.T. 0.006 min

Lab File: BM047146.D

Acq: 10 Aug 2024 17:11

Instrument :

BNA_M

ClientSampleId :

SSTDICC060





Tgt Ion:112 Resp: 2449346

Ion Ratio Lower Upper

112 100

64 57.4 46.7 70.1

63 31.3 24.8 37.2



#6

Aniline

Concen: 57.431 ng

RT: 6.728 min Scan# 653

Delta R.T. -0.000 min

Lab File: BM047146.D

Acq: 10 Aug 2024 17:11




Tgt Ion: 93 Resp: 1595030


Ion Ratio Lower Upper

93 100

66 39.2 31.4 47.2

65 19.9 15.6 23.4









#7
Phenol-d6
Concen: 116.214 ng
RT: 6.593 min Scan# 61
Delta R.T. 0.006 min
Lab File: BM047146.D
Acq: 10 Aug 2024 17:11




Instrument :
BNA_M
ClientSampleId :
SSTDICC060


Tgt Ion: 99 Resp: 3414785
Ion Ratio Lower Upper
99 100
42 17.6 14.1 21.1
71 36.6 28.5 42.7



#8
2-Chlorophenol
Concen: 57.317 ng
RT: 6.957 min Scan# 692
Delta R.T. 0.006 min
Lab File: BM047146.D
Acq: 10 Aug 2024 17:11

Tgt Ion: 128 Resp: 1333099
Ion Ratio Lower Upper
128 100
130 32.5 12.1 52.1
64 44.9 26.6 66.6









#9
Benzaldehyde
Concen: 58.359 ng
RT: 6.540 min Scan# 61
Delta R.T. -0.000 min
Lab File: BM047146.D
Acq: 10 Aug 2024 17:11




Instrument :
BNA_M
ClientSampleId :
SSTDICC060


Tgt Ion: 77 Resp: 733143
Ion Ratio Lower Upper
77 100
105 94.6 74.5 114.5
106 92.1 72.3 112.3



#10
Phenol
Concen: 57.586 ng
RT: 6.616 min Scan# 634
Delta R.T. 0.006 min
Lab File: BM047146.D
Acq: 10 Aug 2024 17:11

Tgt Ion: 94 Resp: 1688399
Ion Ratio Lower Upper
94 100
65 29.6 9.0 49.0
66 39.8 18.9 58.9





#11

bis(2-Chloroethyl)ether

Concen: 57.623 ng

RT: 6.834 min Scan# 6

Delta R.T. 0.006 min

Lab File: BM047146.D

Acq: 10 Aug 2024 17:11

Instrument :





BNA_M

ClientSampleId :

SSTDICC060

Tgt Ion: 93 Resp: 1446822

Ion	Ratio	Lower	Upper
93	100		
63	75.7	55.8	95.8
95	32.7	12.7	52.7



#12

1,3-Dichlorobenzene

Concen: 57.619 ng

RT: 7.263 min Scan# 744




Delta R.T. -0.000 min


Lab File: BM047146.D

Acq: 10 Aug 2024 17:11

Tgt Ion:146 Resp: 1592469

Ion	Ratio	Lower	Upper
146	100		
148	63.4	52.3	78.5
75	32.5	25.8	38.6





#13

1,4-Dichlorobenzene

Concen: 57.691 ng

RT: 7.410 min Scan# 70

Delta R.T. -0.000 min

Lab File: BM047146.D

Acq: 10 Aug 2024 17:11

Instrument :

BNA_M

ClientSampleId :

SSTDICC060





Tgt Ion:146 Resp: 1614359

Ion Ratio Lower Upper

146 100

148 64.7 51.4 77.2

111 38.9 31.5 47.3



#14

1,2-Dichlorobenzene

Concen: 56.653 ng

RT: 7.722 min Scan# 822

Delta R.T. 0.006 min

Lab File: BM047146.D

Acq: 10 Aug 2024 17:11




Tgt Ion:146 Resp: 1495805


Ion Ratio Lower Upper

146 100

148 65.4 51.8 77.8

111 41.7 33.3 49.9









#15
Benzyl Alcohol
Concen: 60.184 ng
RT: 7.634 min Scan# 806
Delta R.T. 0.006 min
Lab File: BM047146.D
Acq: 10 Aug 2024 17:11




Instrument :
BNA_M
ClientSampleId :
SSTDICC060


Tgt Ion: 79 Resp: 1258590
Ion Ratio Lower Upper
79 100
108 74.8 59.7 89.5
77 65.3 52.9 79.3



#16
2,2'-oxybis(1-chloropropane)
Concen: 54.811 ng
RT: 7.916 min Scan# 855
Delta R.T. -0.000 min
Lab File: BM047146.D
Acq: 10 Aug 2024 17:11

Tgt Ion: 45 Resp: 1768027
Ion Ratio Lower Upper
45 100
77 16.2 0.0 35.6
79 13.0 0.0 32.4





#17

2-Methylphenol

Concen: 58.008 ng

RT: 7.839 min Scan# 841

Delta R.T. 0.006 min

Lab File: BM047146.D


Acq: 10 Aug 2024 17:11

Instrument :

BNA_M



ClientSampleId :

SSTDICC060



Tgt Ion:107 Resp: 1134669

Ion	Ratio	Lower	Upper
107	100		
108	112.2	90.6	135.8
77	49.2	39.3	58.9
79	46.8	38.6	58.0



#18

Hexachloroethane

Concen: 59.012 ng

RT: 8.428 min Scan# 942




Delta R.T. -0.000 min


Lab File: BM047146.D

Acq: 10 Aug 2024 17:11

Tgt Ion:117 Resp: 639660

Ion	Ratio	Lower	Upper
117	100		
119	95.6	77.0	115.4
201	101.2	81.3	121.9





#19

n-Nitroso-di-n-propylamine

Concen: 55.379 ng

RT: 8.192 min Scan# 901

Delta R.T. 0.006 min

Lab File: BM047146.D

Acq: 10 Aug 2024 17:11

Instrument :

BNA_M

ClientSampleId :

SSTDICC060

Tgt Ion: 70 Resp: 1096068





Ion Ratio Lower Upper

70 100

42 53.3 42.0 63.0

101 9.0 7.2 10.8

130 19.7 15.5 23.3



#20

3+4-Methylphenols

Concen: 58.726 ng

RT: 8.169 min Scan# 898

Delta R.T. 0.006 min

Lab File: BM047146.D

Acq: 10 Aug 2024 17:11

Tgt Ion: 107 Resp: 1605895




Ion Ratio Lower Upper


107 100

108 89.7 70.7 110.7

77 37.8 17.4 57.4

79 30.3 10.9 50.9









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.133 min Scan# 1232
Delta R.T. -0.000 min
Lab File: BM047146.D
Acq: 10 Aug 2024 17:11




Instrument :
BNA_M
ClientSampleId :
SSTDICC060


Tgt Ion:136 Resp: 1526728
Ion Ratio Lower Upper
136 100
137 10.8 8.8 13.2
54 7.1 6.0 9.0
68 5.9 4.8 7.2



#22
Acetophenone
Concen: 56.185 ng
RT: 8.198 min Scan# 903
Delta R.T. -0.000 min
Lab File: BM047146.D
Acq: 10 Aug 2024 17:11

Tgt Ion:105 Resp: 2070558
Ion Ratio Lower Upper
105 100
71 8.2 4.5 6.7#
51 26.0 20.1 30.1
120 22.4 17.8 26.6





#23

Nitrobenzene-d5

Concen: 115.282 ng

RT: 8.528 min Scan# 91

Delta R.T. 0.006 min

Lab File: BM047146.D

Acq: 10 Aug 2024 17:11

Instrument :





BNA_M

ClientSampleId :

SSTDICC060

Tgt Ion: 82 Resp: 3495384

Ion	Ratio	Lower	Upper
82	100		
128	41.3	32.3	48.5
54	46.0	37.4	56.2



#24

Nitrobenzene

Concen: 56.777 ng

RT: 8.569 min Scan# 966




Delta R.T. 0.006 min


Lab File: BM047146.D

Acq: 10 Aug 2024 17:11

Tgt Ion: 77 Resp: 1733451


Ion	Ratio	Lower	Upper
77	100		
123	42.4	34.0	51.0
65	13.6	10.5	15.7








#25
Isophorone
Concen: 58.876 ng
RT: 9.098 min Scan# 1056
Delta R.T. 0.006 min
Lab File: BM047146.D
Acq: 10 Aug 2024 17:11

Instrument :
BNA_M
ClientSampleId :
SSTDICC060







Tgt Ion: 82 Resp: 3108196
Ion Ratio Lower Upper
82 100
95 7.3 5.5 8.3
138 17.7 13.5 20.3



#26
2-Nitrophenol
Concen: 63.987 ng
RT: 9.269 min Scan# 1085
Delta R.T. 0.006 min
Lab File: BM047146.D
Acq: 10 Aug 2024 17:11

Tgt Ion:139 Resp: 865172
Ion Ratio Lower Upper
139 100
109 30.4 24.4 36.6
65 50.9 43.0 64.6





#27

2,4-Dimethylphenol

Concen: 60.441 ng

RT: 9.345 min Scan# 1098

Delta R.T. -0.000 min

Lab File: BM047146.D

Acq: 10 Aug 2024 17:11

Instrument :

BNA_M

ClientSampleId :

SSTDICC060





Tgt Ion:122 Resp: 957294

Ion Ratio Lower Upper

122 100

107 118.6 92.8 139.2

121 58.3 46.1 69.1



#28

bis(2-Chloroethoxy)methane

Concen: 57.068 ng

RT: 9.581 min Scan# 1138

Delta R.T. 0.006 min

Lab File: BM047146.D

Acq: 10 Aug 2024 17:11




Tgt Ion: 93 Resp: 1895700


Ion Ratio Lower Upper

93 100

95 32.4 25.8 38.8

123 12.6 9.0 13.4





#29

2,4-Dichlorophenol

Concen: 61.767 ng

RT: 9.798 min Scan# 1175

Delta R.T. -0.000 min

Lab File: BM047146.D

Acq: 10 Aug 2024 17:11

Instrument :

BNA_M

ClientSampleId :

SSTDICC060





Tgt Ion:162 Resp: 1451562

Ion Ratio Lower Upper

162 100

164 64.4 44.7 84.7

98 35.0 14.9 54.9



#30

1,2,4-Trichlorobenzene

Concen: 59.763 ng

RT: 10.004 min Scan# 1210

Delta R.T. 0.006 min

Lab File: BM047146.D

Acq: 10 Aug 2024 17:11




Tgt Ion:180 Resp: 1589392


Ion Ratio Lower Upper

180 100

182 96.9 76.6 115.0

145 29.1 23.4 35.2









#31
Naphthalene
Concen: 58.049 ng
RT: 10.186 min Scan# 1137
Delta R.T. 0.006 min
Lab File: BM047146.D
Acq: 10 Aug 2024 17:11




Instrument :
BNA_M
ClientSampleId :
SSTDICC060


Tgt Ion:128 Resp: 4415000
Ion Ratio Lower Upper
128 100
129 10.9 8.6 13.0
127 13.0 10.2 15.4



#32
Benzoic acid
Concen: 67.498 ng
RT: 9.575 min Scan# 1137
Delta R.T. 0.041 min
Lab File: BM047146.D
Acq: 10 Aug 2024 17:11

Tgt Ion:122 Resp: 1248981
Ion Ratio Lower Upper
122 100
105 127.7 107.1 147.1
77 100.2 82.1 122.1









#33
4-Chloroaniline
Concen: 58.764 ng
RT: 10.316 min Scan# 1262
Delta R.T. 0.006 min
Lab File: BM047146.D
Acq: 10 Aug 2024 17:11

Instrument :
BNA_M
ClientSampleId :
SSTDICC060

Tgt Ion:127 Resp: 1718710




Ion	Ratio	Lower	Upper
127	100		
129	32.4	25.8	38.8
65	31.3	24.9	37.3
92	19.2	15.1	22.7




#34
Hexachlorobutadiene
Concen: 62.017 ng
RT: 10.469 min Scan# 1289
Delta R.T. -0.000 min
Lab File: BM047146.D
Acq: 10 Aug 2024 17:11

Tgt Ion:225 Resp: 965378

Ion	Ratio	Lower	Upper
225	100		
223	62.7	50.2	75.2
227	63.7	51.4	77.0









#35
Caprolactam
Concen: 61.664 ng m
RT: 11.151 min Scan# 1405
Delta R.T. 0.029 min
Lab File: BM047146.D
Acq: 10 Aug 2024 17:11




Instrument :
BNA_M
ClientSampleId :
SSTDICC060


Tgt Ion:113 Resp: 503257
Ion Ratio Lower Upper
113 100
55 156.7 140.5 180.5
56 119.4 103.6 143.6



#36
4-Chloro-3-methylphenol
Concen: 60.420 ng
RT: 11.463 min Scan# 1458
Delta R.T. 0.012 min
Lab File: BM047146.D
Acq: 10 Aug 2024 17:11

Tgt Ion:107 Resp: 1512201
Ion Ratio Lower Upper
107 100
144 27.4 21.8 32.6
142 83.4 66.8 100.2









#37
2-Methylnaphthalene
Concen: 59.000 ng
RT: 11.810 min Scan# 1516
Delta R.T. 0.006 min
Lab File: BM047146.D
Acq: 10 Aug 2024 17:11




Instrument :
BNA_M
ClientSampleId :
SSTDICC060


Tgt Ion:142 Resp: 3194966
Ion Ratio Lower Upper
142 100
141 87.3 70.9 106.3
115 35.3 28.7 43.1



#38
1-Methylnaphthalene
Concen: 59.033 ng
RT: 12.033 min Scan# 1555
Delta R.T. -0.000 min
Lab File: BM047146.D
Acq: 10 Aug 2024 17:11

Tgt Ion:142 Resp: 3159405
Ion Ratio Lower Upper
142 100
141 90.5 74.6 111.8
116 3.9 3.1 4.7









#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.033 min Scan# 11
Delta R.T. -0.000 min
Lab File: BM047146.D
Acq: 10 Aug 2024 17:11




Instrument :
BNA_M
ClientSampleId :
SSTDICC060


Tgt Ion:164 Resp: 991306
Ion Ratio Lower Upper
164 100
162 100.1 80.7 121.1
160 44.9 36.0 54.0



#40
1,2,4,5-Tetrachlorobenzene
Concen: 61.523 ng
RT: 12.192 min Scan# 1582
Delta R.T. 0.006 min
Lab File: BM047146.D
Acq: 10 Aug 2024 17:11

Tgt Ion:216 Resp: 1663474
Ion Ratio Lower Upper
216 100
214 78.7 63.4 95.2
179 21.7 17.6 26.4
108 18.3 15.1 22.7









#41
Hexachlorocyclopentadiene
Concen: 65.733 ng
RT: 12.169 min Scan# 1577
Delta R.T. 0.006 min
Lab File: BM047146.D
Acq: 10 Aug 2024 17:11




Instrument :
BNA_M
ClientSampleId :
SSTDICC060


Tgt Ion:237 Resp: 620612
Ion Ratio Lower Upper
237 100
235 61.8 44.8 84.8
272 13.2 0.0 33.5



#42
2,4,6-Tribromophenol
Concen: 121.088 ng
RT: 15.545 min Scan# 2152
Delta R.T. 0.006 min
Lab File: BM047146.D
Acq: 10 Aug 2024 17:11

Tgt Ion:330 Resp: 1391826
Ion Ratio Lower Upper
330 100
332 96.5 77.4 116.2
141 38.4 31.6 47.4





#43

2,4,6-Trichlorophenol

Concen: 62.557 ng

RT: 12.451 min Scan# 1

Delta R.T. 0.006 min

Lab File: BM047146.D

Acq: 10 Aug 2024 17:11

Instrument :

BNA_M

ClientSampleId :

SSTDICC060





Tgt Ion:196 Resp: 1202229

Ion Ratio Lower Upper

196 100

198 96.9 75.6 113.4

200 31.3 24.6 37.0



#44

2,4,5-Trichlorophenol

Concen: 62.287 ng

RT: 12.522 min Scan# 1638

Delta R.T. -0.000 min

Lab File: BM047146.D

Acq: 10 Aug 2024 17:11

Tgt Ion:196 Resp: 1329069




Ion Ratio Lower Upper


196 100

198 94.3 77.7 116.5

97 49.4 39.8 59.8

132 25.8 20.7 31.1









#45
2-Fluorobiphenyl
Concen: 116.705 ng
RT: 12.645 min Scan# 1659
Delta R.T. 0.000 min
Lab File: BM047146.D
Acq: 10 Aug 2024 17:11




Instrument :
BNA_M
ClientSampleId :
SSTDICC060


Tgt Ion:172 Resp: 7500034
Ion Ratio Lower Upper
172 100
171 35.0 27.8 41.6
170 23.5 18.7 28.1



#46
1,1'-Biphenyl
Concen: 58.466 ng
RT: 12.857 min Scan# 1695
Delta R.T. -0.000 min
Lab File: BM047146.D
Acq: 10 Aug 2024 17:11


Tgt Ion:154 Resp: 4126067
Ion Ratio Lower Upper
154 100
153 39.9 19.0 59.0
76 12.1 0.0 31.9








#47
2-Chloronaphthalene
Concen: 57.882 ng
RT: 12.892 min Scan# 1701
Delta R.T. 0.006 min
Lab File: BM047146.D
Acq: 10 Aug 2024 17:11


Instrument :
BNA_M
ClientSampleId :
SSTDICC060





Tgt Ion:162 Resp: 3192787
Ion Ratio Lower Upper
162 100
127 37.0 30.0 45.0
164 33.1 26.3 39.5




#48
2-Nitroaniline
Concen: 60.847 ng
RT: 13.121 min Scan# 1740
Delta R.T. 0.006 min
Lab File: BM047146.D
Acq: 10 Aug 2024 17:11



Tgt Ion: 65 Resp: 1051464
Ion Ratio Lower Upper
65 100
92 68.9 55.1 82.7
138 107.0 85.9 128.9









#49
Acenaphthylene
Concen: 58.613 ng
RT: 13.751 min Scan# 1847
Delta R.T. -0.000 min
Lab File: BM047146.D
Acq: 10 Aug 2024 17:11




Instrument :
BNA_M
ClientSampleId :
SSTDICC060


Tgt Ion:152 Resp: 4765674
Ion Ratio Lower Upper
152 100
151 19.7 15.8 23.6
153 13.3 10.4 15.6



#50
Dimethylphthalate
Concen: 58.629 ng
RT: 13.521 min Scan# 1808
Delta R.T. 0.006 min
Lab File: BM047146.D
Acq: 10 Aug 2024 17:11

Tgt Ion:163 Resp: 4059126
Ion Ratio Lower Upper
163 100
194 4.0 3.0 4.6
164 10.4 8.0 12.0









#51
2,6-Dinitrotoluene
Concen: 61.508 ng
RT: 13.633 min Scan# 1827
Delta R.T. 0.006 min
Lab File: BM047146.D
Acq: 10 Aug 2024 17:11




Instrument :
BNA_M
ClientSampleId :
SSTDICC060


Tgt Ion:165 Resp: 995456
Ion Ratio Lower Upper
165 100
63 54.7 46.5 69.7
89 56.8 47.3 70.9



#52
Acenaphthene
Concen: 58.668 ng
RT: 14.104 min Scan# 1907
Delta R.T. 0.006 min
Lab File: BM047146.D
Acq: 10 Aug 2024 17:11

Tgt Ion:154 Resp: 3025820
Ion Ratio Lower Upper
154 100
153 112.1 90.3 135.5
152 53.3 42.9 64.3









#53
3-Nitroaniline
Concen: 61.708 ng
RT: 13.969 min Scan# 1883
Delta R.T. 0.006 min
Lab File: BM047146.D
Acq: 10 Aug 2024 17:11




Instrument :
BNA_M
ClientSampleId :
SSTDICC060


Tgt Ion:138 Resp: 1000744
Ion Ratio Lower Upper
138 100
108 9.6 7.9 11.9
92 114.9 92.0 138.0



#54
2,4-Dinitrophenol
Concen: 66.750 ng
RT: 14.180 min Scan# 1920
Delta R.T. 0.005 min
Lab File: BM047146.D
Acq: 10 Aug 2024 17:11

Tgt Ion:184 Resp: 647477
Ion Ratio Lower Upper
184 100
63 62.4 52.7 79.1
154 63.7 50.4 75.6









#55
Dibenzenofuran
Concen: 57.809 ng
RT: 14.445 min Scan# 1965
Delta R.T. 0.006 min
Lab File: BM047146.D
Acq: 10 Aug 2024 17:11




Instrument :
BNA_M
ClientSampleId :
SSTDICC060


Tgt Ion:168 Resp: 4709496
Ion Ratio Lower Upper
168 100
139 36.5 29.6 44.4
169 13.2 10.7 16.1



#56
4-Nitrophenol
Concen: 62.927 ng
RT: 14.304 min Scan# 1941
Delta R.T. 0.006 min
Lab File: BM047146.D
Acq: 10 Aug 2024 17:11


Tgt Ion:139 Resp: 879989
Ion Ratio Lower Upper
139 100
109 77.5 52.5 92.5
65 98.0 75.4 115.4








#57
2,4-Dinitrotoluene
Concen: 61.255 ng
RT: 14.433 min Scan# 1963
Delta R.T. 0.006 min
Lab File: BM047146.D
Acq: 10 Aug 2024 17:11

Instrument :
BNA_M
ClientSampleId :
SSTDICC060







Tgt Ion:165 Resp: 1323234
Ion Ratio Lower Upper
165 100
63 53.0 40.3 60.5
89 77.0 61.0 91.6
182 2.7 1.9 2.9



#58
Fluorene
Concen: 57.801 ng
RT: 15.098 min Scan# 2076
Delta R.T. 0.006 min
Lab File: BM047146.D
Acq: 10 Aug 2024 17:11

Tgt Ion:166 Resp: 3885304
Ion Ratio Lower Upper
166 100
165 96.6 77.4 116.0
167 13.5 11.0 16.6









#59
2,3,4,6-Tetrachlorophenol
Concen: 62.097 ng
RT: 14.680 min Scan# 2004
Delta R.T. 0.006 min
Lab File: BM047146.D
Acq: 10 Aug 2024 17:11




Instrument :
BNA_M
ClientSampleId :
SSTDICC060


Tgt Ion:232 Resp: 1088974
Ion Ratio Lower Upper
232 100
131 44.4 36.9 55.3
130 2.4 2.0 3.0
166 32.1 26.4 39.6



#60
Diethylphthalate
Concen: 58.389 ng
RT: 14.904 min Scan# 2043
Delta R.T. 0.006 min
Lab File: BM047146.D
Acq: 10 Aug 2024 17:11

Tgt Ion:149 Resp: 4128072
Ion Ratio Lower Upper
149 100
177 22.9 17.6 26.4
150 12.4 9.8 14.6









#61
4-Chlorophenyl-phenylether
Concen: 59.903 ng
RT: 15.104 min Scan# 2076
Delta R.T. 0.006 min
Lab File: BM047146.D
Acq: 10 Aug 2024 17:11




Instrument :
BNA_M
ClientSampleId :
SSTDICC060


Tgt Ion:204 Resp: 1916971
Ion Ratio Lower Upper
204 100
206 33.1 26.2 39.4
141 54.2 47.2 70.8



#62
4-Nitroaniline
Concen: 61.238 ng
RT: 15.145 min Scan# 2084
Delta R.T. 0.006 min
Lab File: BM047146.D
Acq: 10 Aug 2024 17:11

Tgt Ion:138 Resp: 978393
Ion Ratio Lower Upper
138 100
92 52.0 31.9 71.9
108 85.0 63.8 103.8









#63
Azobenzene
Concen: 56.986 ng
RT: 15.398 min Scan# 2126
Delta R.T. 0.006 min
Lab File: BM047146.D
Acq: 10 Aug 2024 17:11




Instrument :
BNA_M
ClientSampleId :
SSTDICC060


Tgt Ion: 77 Resp: 3820106
Ion Ratio Lower Upper
77 100
182 25.7 3.8 43.8
105 19.8 0.0 39.5
51 26.5 6.8 46.8



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 16.798 min Scan# 2365
Delta R.T. 0.006 min
Lab File: BM047146.D
Acq: 10 Aug 2024 17:11

Tgt Ion: 188 Resp: 2082135
Ion Ratio Lower Upper
188 100
94 8.5 7.0 10.6
80 9.5 8.0 12.0





#65

4,6-Dinitro-2-methylphenol

Concen: 65.573 ng

RT: 15.204 min Scan# 2094

Delta R.T. 0.006 min

Lab File: BM047146.D

Acq: 10 Aug 2024 17:11

Instrument :

BNA_M

ClientSampleId :

SSTDICC060





Tgt Ion:198 Resp: 799367

Ion Ratio Lower Upper

198 100

51 40.9 23.9 63.9

105 42.1 24.3 64.3



#66

n-Nitrosodiphenylamine

Concen: 58.648 ng

RT: 15.327 min Scan# 2115

Delta R.T. 0.006 min

Lab File: BM047146.D

Acq: 10 Aug 2024 17:11




Tgt Ion:169 Resp: 3326078


Ion Ratio Lower Upper

169 100

168 66.6 53.8 80.8

167 36.1 29.3 43.9





#67

4-Bromophenyl-phenylether

Concen: 61.547 ng

RT: 15.998 min Scan# 21

Delta R.T. -0.000 min

Lab File: BM047146.D

Acq: 10 Aug 2024 17:11

Instrument :

BNA_M

ClientSampleId :

SSTDICC060





Tgt Ion:248 Resp: 1234868

Ion Ratio Lower Upper

248 100

250 96.2 77.8 116.8

141 69.4 57.7 86.5



#68

Hexachlorobenzene

Concen: 59.562 ng

RT: 16.104 min Scan# 2247

Delta R.T. 0.006 min

Lab File: BM047146.D

Acq: 10 Aug 2024 17:11




Tgt Ion:284 Resp: 1431249


Ion Ratio Lower Upper

284 100

142 33.7 29.4 44.0

249 31.8 25.1 37.7









#69
Atrazine
Concen: 53.118 ng
RT: 16.298 min Scan# 21
Delta R.T. 0.006 min
Lab File: BM047146.D
Acq: 10 Aug 2024 17:11




Instrument :
BNA_M
ClientSampleId :
SSTDICC060


Tgt Ion:200 Resp: 906491
Ion Ratio Lower Upper
200 100
173 25.5 6.1 46.1
215 46.6 26.8 66.8



#70
Pentachlorophenol
Concen: 64.503 ng
RT: 16.457 min Scan# 2307
Delta R.T. 0.006 min
Lab File: BM047146.D
Acq: 10 Aug 2024 17:11

Tgt Ion:266 Resp: 1086504
Ion Ratio Lower Upper
266 100
268 63.5 49.8 74.6
264 62.9 49.6 74.4









#71
Phenanthrene
Concen: 58.208 ng
RT: 16.839 min Scan# 2372
Delta R.T. -0.000 min
Lab File: BM047146.D
Acq: 10 Aug 2024 17:11




Instrument :
BNA_M
ClientSampleId :
SSTDICC060


Tgt Ion:178 Resp: 6082429
Ion Ratio Lower Upper
178 100
176 19.3 15.3 22.9
179 15.7 12.4 18.6



#72
Anthracene
Concen: 59.030 ng
RT: 16.933 min Scan# 2388
Delta R.T. 0.006 min
Lab File: BM047146.D
Acq: 10 Aug 2024 17:11

Tgt Ion:178 Resp: 6001835
Ion Ratio Lower Upper
178 100
176 18.8 14.8 22.2
179 15.4 12.4 18.6









#73
Carbazole
Concen: 57.702 ng
RT: 17.209 min Scan# 2435
Delta R.T. -0.000 min
Lab File: BM047146.D
Acq: 10 Aug 2024 17:11




Instrument :
BNA_M
ClientSampleId :
SSTDICC060


Tgt Ion:167 Resp: 5990202
Ion Ratio Lower Upper
167 100
166 21.4 17.1 25.7
139 12.6 10.0 15.0



#74
Di-n-butylphthalate
Concen: 58.952 ng
RT: 17.798 min Scan# 2535
Delta R.T. 0.006 min
Lab File: BM047146.D
Acq: 10 Aug 2024 17:11

Tgt Ion:149 Resp: 7121063
Ion Ratio Lower Upper
149 100
150 9.3 7.5 11.3
104 5.1 4.1 6.1









#75
Fluoranthene
Concen: 59.328 ng
RT: 18.874 min Scan# 21
Delta R.T. 0.006 min
Lab File: BM047146.D
Acq: 10 Aug 2024 17:11




Instrument :
BNA_M
ClientSampleId :
SSTDICC060


Tgt Ion:202 Resp: 7540518
Ion Ratio Lower Upper
202 100
101 9.2 0.0 30.2
203 17.5 0.0 37.6



#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.033 min Scan# 3085
Delta R.T. -0.000 min
Lab File: BM047146.D
Acq: 10 Aug 2024 17:11

Tgt Ion:240 Resp: 1812683
Ion Ratio Lower Upper
240 100
120 10.2 8.2 12.2
236 25.8 20.6 30.8









#77
Benzidine
Concen: 68.768 ng
RT: 19.074 min Scan# 2752
Delta R.T. -0.000 min
Lab File: BM047146.D
Acq: 10 Aug 2024 17:11




Instrument :
BNA_M
ClientSampleId :
SSTDICC060


Tgt Ion:184 Resp: 2113957
Ion Ratio Lower Upper
184 100
185 14.7 11.8 17.6
183 11.6 9.0 13.6



#78
Pyrene
Concen: 60.744 ng
RT: 19.239 min Scan# 2780
Delta R.T. -0.000 min
Lab File: BM047146.D
Acq: 10 Aug 2024 17:11


Tgt Ion:202 Resp: 7862143
Ion Ratio Lower Upper
202 100
200 21.6 17.2 25.8
203 17.7 14.2 21.2








#79
Terphenyl-d14
Concen: 121.942 ng
RT: 19.462 min Scan# 2818
Delta R.T. -0.000 min
Lab File: BM047146.D
Acq: 10 Aug 2024 17:11


Instrument :
BNA_M
ClientSampleId :
SSTDICC060





Tgt Ion:244 Resp:10315735
Ion Ratio Lower Upper
244 100
212 8.1 6.2 9.4
122 8.3 7.1 10.7




#80
Butylbenzylphthalate
Concen: 63.380 ng
RT: 20.180 min Scan# 2940
Delta R.T. 0.006 min
Lab File: BM047146.D
Acq: 10 Aug 2024 17:11



Tgt Ion:149 Resp: 3347760
Ion Ratio Lower Upper
149 100
91 67.5 57.0 85.4
206 22.0 16.4 24.6









#81
Benzo(a)anthracene
Concen: 59.814 ng
RT: 21.015 min Scan# 3082
Delta R.T. -0.000 min
Lab File: BM047146.D
Acq: 10 Aug 2024 17:11




Instrument :
BNA_M
ClientSampleId :
SSTDICC060


Tgt Ion:228 Resp: 7197782
Ion Ratio Lower Upper
228 100
226 28.1 21.8 32.6
229 20.3 15.8 23.8



#82
3,3'-Dichlorobenzidine
Concen: 61.593 ng
RT: 20.956 min Scan# 3072
Delta R.T. -0.000 min
Lab File: BM047146.D
Acq: 10 Aug 2024 17:11

Tgt Ion:252 Resp: 2319044
Ion Ratio Lower Upper
252 100
254 64.6 51.0 76.6
126 10.4 8.5 12.7





#83

Chrysene

Concen: 58.790 ng

RT: 21.074 min Scan# 3092

Delta R.T. -0.000 min

Lab File: BM047146.D

Acq: 10 Aug 2024 17:11

Instrument :

BNA_M

ClientSampleId :

SSTDICC060





Tgt Ion:228 Resp: 6712104

Ion Ratio Lower Upper

228 100

226 31.1 24.4 36.6

229 20.1 15.7 23.5



#84

Bis(2-ethylhexyl)phthalate

Concen: 61.396 ng

RT: 20.980 min Scan# 3076

Delta R.T. 0.006 min

Lab File: BM047146.D

Acq: 10 Aug 2024 17:11




Tgt Ion:149 Resp: 4599112


Ion Ratio Lower Upper

149 100

167 28.2 22.2 33.2

279 4.6 3.0 4.4#









#85
Di-n-octyl phthalate
Concen: 64.180 ng
RT: 22.009 min Scan# 31
Delta R.T. -0.000 min
Lab File: BM047146.D
Acq: 10 Aug 2024 17:11




Instrument :
BNA_M
ClientSampleId :
SSTDICC060


Tgt Ion:149 Resp: 8171642
Ion Ratio Lower Upper
149 100
167 1.6 1.3 1.9
43 8.2 6.9 10.3



#86
Perylene-d12
Concen: 20.000 ng
RT: 23.733 min Scan# 3544
Delta R.T. 0.006 min
Lab File: BM047146.D
Acq: 10 Aug 2024 17:11

Tgt Ion:264 Resp: 2151546
Ion Ratio Lower Upper
264 100
260 22.9 18.5 27.7
265 21.5 17.3 25.9









#87
Indeno(1,2,3-cd)pyrene
Concen: 56.929 ng
RT: 26.744 min Scan# 4053
Delta R.T. 0.017 min
Lab File: BM047146.D
Acq: 10 Aug 2024 17:11




Instrument :
BNA_M
ClientSampleId :
SSTDICC060


Tgt Ion:276 Resp: 7917132
Ion Ratio Lower Upper
276 100
138 23.4 19.4 29.0
277 25.1 20.1 30.1



#88
Benzo(b)fluoranthene
Concen: 60.962 ng
RT: 22.897 min Scan# 3402
Delta R.T. 0.012 min
Lab File: BM047146.D
Acq: 10 Aug 2024 17:11

Tgt Ion:252 Resp: 7789248
Ion Ratio Lower Upper
252 100
253 22.3 17.8 26.6
125 8.6 7.4 11.0





#89

Benzo(k)fluoranthene

Concen: 59.680 ng

RT: 22.950 min Scan# 3410

Delta R.T. 0.006 min

Lab File: BM047146.D

Acq: 10 Aug 2024 17:11

Instrument :

BNA_M

ClientSampleId :

SSTDICC060





Tgt Ion:252 Resp: 7202578

Ion Ratio Lower Upper

252 100

253 22.0 17.7 26.5

125 8.4 7.4 11.2



#90

Benzo(a)pyrene

Concen: 60.997 ng

RT: 23.615 min Scan# 3524

Delta R.T. 0.006 min

Lab File: BM047146.D

Acq: 10 Aug 2024 17:11




Tgt Ion:252 Resp: 6751397


Ion Ratio Lower Upper

252 100

253 21.9 17.1 25.7

125 9.8 8.2 12.2









#91
Dibenzo(a,h)anthracene
Concen: 56.576 ng
RT: 26.791 min Scan# 4064
Delta R.T. 0.012 min
Lab File: BM047146.D
Acq: 10 Aug 2024 17:11




Instrument :
BNA_M
ClientSampleId :
SSTDICC060

Tgt Ion:278 Resp: 6367453
Ion Ratio Lower Upper
278 100
139 14.6 12.0 18.0
279 24.0 19.2 28.8



#92
Benzo(g,h,i)perylene
Concen: 55.488 ng
RT: 27.679 min Scan# 4215
Delta R.T. 0.017 min
Lab File: BM047146.D
Acq: 10 Aug 2024 17:11

Tgt Ion:276 Resp: 6483852
Ion Ratio Lower Upper
276 100
277 24.1 18.8 28.2
138 20.1 16.2 24.4



Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081024\
 Data File : BM047147.D
 Acq On : 10 Aug 2024 17:51
 Operator : RC/JU
 Sample : SSTDICC080
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC080

Quant Time: Aug 11 23:35:57 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Aug 11 23:24:37 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.375	152	304450	20.000	ng	0.00
21) Naphthalene-d8	10.134	136	1204959	20.000	ng	0.00
39) Acenaphthene-d10	14.033	164	795761	20.000	ng	0.00
64) Phenanthrene-d10	16.798	188	1674812	20.000	ng	0.00
76) Chrysene-d12	21.033	240	1443265	20.000	ng	0.00
86) Perylene-d12	23.733	264	1789187	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.028	112	2548471	149.572	ng	0.00
7) Phenol-d6	6.593	99	3559773	151.625	ng	0.00
23) Nitrobenzene-d5	8.528	82	3582678	149.714	ng	0.00
42) 2,4,6-Tribromophenol	15.545	330	1437262	155.767	ng	0.00
45) 2-Fluorobiphenyl	12.645	172	7769782	150.612	ng	0.00
79) Terphenyl-d14	19.462	244	10644639	158.037	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.034	88	518861	73.496	ng	99
3) Pyridine	3.410	79	1583783m	76.560	ng	
4) n-Nitrosodimethylamine	3.328	42	678926	75.066	ng	96
6) Aniline	6.734	93	1624949	73.227	ng	99
8) 2-Chlorophenol	6.951	128	1372864	73.876	ng	99
10) Phenol	6.616	94	1769382	75.529	ng	99
11) bis(2-Chloroethyl)ether	6.834	93	1474612	73.503	ng	99
12) 1,3-Dichlorobenzene	7.263	146	1643503	74.424	ng	99
13) 1,4-Dichlorobenzene	7.410	146	1656531	74.089	ng	100
14) 1,2-Dichlorobenzene	7.722	146	1535913	72.806	ng	99
15) Benzyl Alcohol	7.634	79	1293991	77.443	ng	99
16) 2,2'-oxybis(1-Chloropr...	7.916	45	1798177	69.769	ng	100
17) 2-Methylphenol	7.834	107	1167820	74.722	ng	99
18) Hexachloroethane	8.428	117	649682	75.015	ng	99
19) n-Nitroso-di-n-propyla...	8.198	70	1114969	70.505	ng	98
20) 3+4-Methylphenols	8.169	107	1646698	75.366	ng	99
22) Acetophenone	8.198	105	2124216	73.033	ng	# 98
24) Nitrobenzene	8.569	77	1777843	73.781	ng	99
25) Isophorone	9.098	82	3163607	75.927	ng	99
26) 2-Nitrophenol	9.263	139	891212	83.514	ng	99
27) 2,4-Dimethylphenol	9.345	122	981455	78.514	ng	99
28) bis(2-Chloroethoxy)met...	9.581	93	1908847	72.808	ng	98
29) 2,4-Dichlorophenol	9.798	162	1485940	80.115	ng	100
30) 1,2,4-Trichlorobenzene	10.004	180	1625842	77.458	ng	100
31) Naphthalene	10.181	128	4520274	75.304	ng	99
32) Benzoic acid	9.575	122	1293498	88.571	ng	99
33) 4-Chloroaniline	10.316	127	1753660	75.970	ng	100
34) Hexachlorobutadiene	10.469	225	991890	80.736	ng	99
35) Caprolactam	11.151	113	523592m	81.288	ng	
36) 4-Chloro-3-methylphenol	11.463	107	1559534	78.951	ng	99
37) 2-Methylnaphthalene	11.810	142	3301008	77.236	ng	99
38) 1-Methylnaphthalene	12.033	142	3232647	76.532	ng	99
40) 1,2,4,5-Tetrachloroben...	12.192	216	1736758	80.017	ng	99
41) Hexachlorocyclopentadiene	12.169	237	620195	81.830	ng	97
43) 2,4,6-Trichlorophenol	12.451	196	1259185	81.622	ng	99
44) 2,4,5-Trichlorophenol	12.522	196	1383585	80.776	ng	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081024\
 Data File : BM047147.D
 Acq On : 10 Aug 2024 17:51
 Operator : RC/JU
 Sample : SSTDICC080
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC080


Quant Time: Aug 11 23:35:57 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Aug 11 23:24:37 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) 1,1'-Biphenyl	12.857	154	4241252	74.866	ng	98
47) 2-Chloronaphthalene	12.892	162	3319138	74.960	ng	99
48) 2-Nitroaniline	13.122	65	1090521	78.614	ng	99
49) Acenaphthylene	13.751	152	4945706	75.775	ng	99
50) Dimethylphthalate	13.522	163	4196226	75.503	ng	99
51) 2,6-Dinitrotoluene	13.639	165	1028020	79.128	ng	91
52) Acenaphthene	14.104	154	3144083	75.941	ng	99
53) 3-Nitroaniline	13.969	138	1034545	79.468	ng	100
54) 2,4-Dinitrophenol	14.180	184	674624	86.639	ng	96
55) Dibenzofuran	14.445	168	4887362	74.735	ng	99
56) 4-Nitrophenol	14.304	139	913035	81.334	ng	96
57) 2,4-Dinitrotoluene	14.433	165	1370129	79.012	ng	97
58) Fluorene	15.098	166	4025459	74.602	ng	99
59) 2,3,4,6-Tetrachlorophenol	14.680	232	1128723	80.180	ng	98
60) Diethylphthalate	14.904	149	4257104	75.010	ng	98
61) 4-Chlorophenyl-phenyle...	15.104	204	1997247	77.748	ng	96
62) 4-Nitroaniline	15.145	138	1020427	79.564	ng	99
63) Azobenzene	15.398	77	3920766	72.861	ng	97
65) 4,6-Dinitro-2-methylph...	15.204	198	820182	83.643	ng	97
66) n-Nitrosodiphenylamine	15.327	169	3446873	75.560	ng	99
67) 4-Bromophenyl-phenylether	15.998	248	1277742	79.172	ng	99
68) Hexachlorobenzene	16.104	284	1481615	76.654	ng	98
69) Atrazine	16.298	200	862020	62.797	ng	99
70) Pentachlorophenol	16.457	266	1115403	82.323	ng	98
71) Phenanthrene	16.839	178	6356974	75.631	ng	100
72) Anthracene	16.933	178	6203287	75.850	ng	100
73) Carbazole	17.210	167	6250976	74.858	ng	100
74) Di-n-butylphthalate	17.798	149	7355197	75.700	ng	100
75) Fluoranthene	18.874	202	7852169	76.805	ng	99
77) Benzidine	19.074	184	1943953	79.424	ng	100
78) Pyrene	19.239	202	8163599	79.218	ng	100
80) Butylbenzylphthalate	20.174	149	3437732	81.742	ng	100
81) Benzo(a)anthracene	21.015	228	7463024	77.892	ng	99
82) 3,3'-Dichlorobenzidine	20.956	252	2359199	78.698	ng	99
83) Chrysene	21.074	228	6992860	76.926	ng	99
84) Bis(2-ethylhexyl)phtha...	20.974	149	4711826	79.001	ng	99
85) Di-n-octyl phthalate	22.009	149	8464049	83.492	ng	99
87) Indeno(1,2,3-cd)pyrene	26.750	276	9258642	80.058	ng	99
88) Benzo(b)fluoranthene	22.897	252	8329223	78.391	ng	99
89) Benzo(k)fluoranthene	22.950	252	7726234	76.985	ng	99
90) Benzo(a)pyrene	23.615	252	7338002	79.723	ng	98
91) Dibenzo(a,h)anthracene	26.803	278	7361474	78.655	ng	99
92) Benzo(g,h,i)perylene	27.685	276	7705035	79.293	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
BNA_M
ClientSampleId :
SSTDICC080





[illegible]



#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.375 min Scan# 763
Delta R.T. -0.000 min
Lab File: BM047147.D
Acq: 10 Aug 2024 17:51




Instrument :
BNA_M
ClientSampleId :
SSTDICC080


Tgt Ion:152 Resp: 304450
Ion Ratio Lower Upper
152 100
150 156.0 120.5 180.7
115 56.2 45.5 68.3



#2
1,4-Dioxane
Concen: 73.496 ng
RT: 3.034 min Scan# 25
Delta R.T. -0.000 min
Lab File: BM047147.D
Acq: 10 Aug 2024 17:51

Tgt Ion: 88 Resp: 518861
Ion Ratio Lower Upper
88 100
58 71.3 58.2 87.4
43 27.8 22.5 33.7









#3
Pyridine
Concen: 76.560 ng m
RT: 3.410 min Scan# 89
Delta R.T. -0.000 min
Lab File: BM047147.D
Acq: 10 Aug 2024 17:51




Instrument :
BNA_M
ClientSampleId :
SSTDICC080


Tgt Ion: 79 Resp: 1583783
Ion Ratio Lower Upper
79 100
52 64.1 51.2 76.8
51 32.4 26.1 39.1



#4
n-Nitrosodimethylamine
Concen: 75.066 ng
RT: 3.328 min Scan# 75
Delta R.T. -0.000 min
Lab File: BM047147.D
Acq: 10 Aug 2024 17:51

Tgt Ion: 42 Resp: 678926
Ion Ratio Lower Upper
42 100
74 145.6 112.9 169.3
44 6.3 5.3 7.9









#5
2-Fluorophenol
Concen: 149.572 ng
RT: 5.028 min Scan# 363
Delta R.T. 0.006 min
Lab File: BM047147.D
Acq: 10 Aug 2024 17:51




Instrument :
BNA_M
ClientSampleId :
SSTDICC080


Tgt Ion:112 Resp: 2548471
Ion Ratio Lower Upper
112 100
64 57.1 46.7 70.1
63 30.5 24.8 37.2



#6
Aniline
Concen: 73.227 ng
RT: 6.734 min Scan# 654
Delta R.T. 0.006 min
Lab File: BM047147.D
Acq: 10 Aug 2024 17:51

Tgt Ion: 93 Resp: 1624949
Ion Ratio Lower Upper
93 100
66 38.6 31.4 47.2
65 19.8 15.6 23.4









#7
Phenol-d6
Concen: 151.625 ng
RT: 6.593 min Scan# 61
Delta R.T. 0.006 min
Lab File: BM047147.D
Acq: 10 Aug 2024 17:51




Instrument :
BNA_M
ClientSampleId :
SSTDICC080


Tgt Ion: 99 Resp: 3559773
Ion Ratio Lower Upper
99 100
42 17.2 14.1 21.1
71 36.0 28.5 42.7



#8
2-Chlorophenol
Concen: 73.876 ng
RT: 6.951 min Scan# 691
Delta R.T. -0.000 min
Lab File: BM047147.D
Acq: 10 Aug 2024 17:51

Tgt Ion: 128 Resp: 1372864
Ion Ratio Lower Upper
128 100
130 31.6 12.1 52.1
64 47.3 26.6 66.6









#10
Phenol
Concen: 75.529 ng
RT: 6.616 min Scan# 61
Delta R.T. 0.006 min
Lab File: BM047147.D
Acq: 10 Aug 2024 17:51

Instrument :
BNA_M
ClientSampleId :
SSTDICC080

Tgt Ion: 94 Resp: 1769382




Ion	Ratio	Lower	Upper
94	100		
65	29.1	9.0	49.0
66	39.8	18.9	58.9




#11
bis(2-Chloroethyl)ether
Concen: 73.503 ng
RT: 6.834 min Scan# 671
Delta R.T. 0.006 min
Lab File: BM047147.D
Acq: 10 Aug 2024 17:51

Tgt Ion: 93 Resp: 1474612

Ion	Ratio	Lower	Upper
93	100		
63	74.5	55.8	95.8
95	32.6	12.7	52.7





#12

1,3-Dichlorobenzene

Concen: 74.424 ng

RT: 7.263 min Scan# 744

Delta R.T. -0.000 min

Lab File: BM047147.D

Acq: 10 Aug 2024 17:51

Instrument :

BNA_M

ClientSampleId :

SSTDICC080





Tgt Ion:146 Resp: 1643503

Ion Ratio Lower Upper

146 100

148 64.5 52.3 78.5

75 31.7 25.8 38.6



#13

1,4-Dichlorobenzene

Concen: 74.089 ng

RT: 7.410 min Scan# 769

Delta R.T. -0.000 min

Lab File: BM047147.D

Acq: 10 Aug 2024 17:51




Tgt Ion:146 Resp: 1656531


Ion Ratio Lower Upper

146 100

148 64.6 51.4 77.2

111 39.1 31.5 47.3









#14
1,2-Dichlorobenzene
Concen: 72.806 ng
RT: 7.722 min Scan# 81
Delta R.T. 0.006 min
Lab File: BM047147.D
Acq: 10 Aug 2024 17:51




Instrument :
BNA_M
ClientSampleId :
SSTDICC080


Tgt Ion:146 Resp: 1535913
Ion Ratio Lower Upper
146 100
148 63.8 51.8 77.8
111 41.0 33.3 49.9



#15
Benzyl Alcohol
Concen: 77.443 ng
RT: 7.634 min Scan# 807
Delta R.T. 0.006 min
Lab File: BM047147.D
Acq: 10 Aug 2024 17:51

Tgt Ion: 79 Resp: 1293991
Ion Ratio Lower Upper
79 100
108 74.5 59.7 89.5
77 64.4 52.9 79.3









#16
2,2'-oxybis(1-Chloropropane)
Concen: 69.769 ng
RT: 7.916 min Scan# 855
Delta R.T. -0.000 min
Lab File: BM047147.D
Acq: 10 Aug 2024 17:51




Instrument :
BNA_M
ClientSampleId :
SSTDICC080


Tgt Ion: 45 Resp: 1798177
Ion Ratio Lower Upper
45 100
77 15.7 0.0 35.6
79 12.7 0.0 32.4



#17
2-Methylphenol
Concen: 74.722 ng
RT: 7.834 min Scan# 841
Delta R.T. -0.000 min
Lab File: BM047147.D
Acq: 10 Aug 2024 17:51


Tgt Ion: 107 Resp: 1167820
Ion Ratio Lower Upper
107 100
108 112.9 90.6 135.8
77 50.7 39.3 58.9
79 49.2 38.6 58.0








#18
Hexachloroethane
Concen: 75.015 ng
RT: 8.428 min Scan# 941
Delta R.T. -0.000 min
Lab File: BM047147.D
Acq: 10 Aug 2024 17:51

Instrument :
BNA_M
ClientSampleId :
SSTDICC080







Tgt Ion:117 Resp: 649682
Ion Ratio Lower Upper
117 100
119 96.2 77.0 115.4
201 104.1 81.3 121.9



#19
n-Nitroso-di-n-propylamine
Concen: 70.505 ng
RT: 8.198 min Scan# 903
Delta R.T. 0.012 min
Lab File: BM047147.D
Acq: 10 Aug 2024 17:51


Tgt Ion: 70 Resp: 1114969
Ion Ratio Lower Upper
70 100
42 51.9 42.0 63.0
101 9.6 7.2 10.8
130 21.0 15.5 23.3








#20
3+4-Methylphenols
Concen: 75.366 ng
RT: 8.169 min Scan# 897
Delta R.T. 0.006 min
Lab File: BM047147.D
Acq: 10 Aug 2024 17:51

Instrument :
BNA_M
ClientSampleId :
SSTDICC080



Tgt Ion:107 Resp: 1646698




Ion	Ratio	Lower	Upper
107	100		
108	89.3	70.7	110.7
77	38.0	17.4	57.4
79	29.9	10.9	50.9




#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.134 min Scan# 1232
Delta R.T. -0.000 min
Lab File: BM047147.D
Acq: 10 Aug 2024 17:51

Tgt Ion:136 Resp: 1204959


Ion	Ratio	Lower	Upper
136	100		
137	11.1	8.8	13.2
54	7.2	6.0	9.0
68	5.9	4.8	7.2








#22
Acetophenone
Concen: 73.033 ng
RT: 8.198 min Scan# 903
Delta R.T. -0.000 min
Lab File: BM047147.D
Acq: 10 Aug 2024 17:51

Instrument :
BNA_M
ClientSampleId :
SSTDICC080







Tgt Ion:105 Resp: 2124216
Ion Ratio Lower Upper
105 100
71 8.4 4.5 6.7#
51 25.6 20.1 30.1
120 22.7 17.8 26.6



#23
Nitrobenzene-d5
Concen: 149.714 ng
RT: 8.528 min Scan# 959
Delta R.T. 0.006 min
Lab File: BM047147.D
Acq: 10 Aug 2024 17:51

Tgt Ion: 82 Resp: 3582678
Ion Ratio Lower Upper
82 100
128 41.4 32.3 48.5
54 46.1 37.4 56.2









#24
Nitrobenzene
Concen: 73.781 ng
RT: 8.569 min Scan# 965
Delta R.T. 0.006 min
Lab File: BM047147.D
Acq: 10 Aug 2024 17:51




Instrument :
BNA_M
ClientSampleId :
SSTDICC080


Tgt Ion: 77 Resp: 1777843
Ion Ratio Lower Upper
77 100
123 43.4 34.0 51.0
65 13.7 10.5 15.7



#25
Isophorone
Concen: 75.927 ng
RT: 9.098 min Scan# 1056
Delta R.T. 0.006 min
Lab File: BM047147.D
Acq: 10 Aug 2024 17:51

Tgt Ion: 82 Resp: 3163607
Ion Ratio Lower Upper
82 100
95 7.3 5.5 8.3
138 17.5 13.5 20.3









#26
2-Nitrophenol
Concen: 83.514 ng
RT: 9.263 min Scan# 1084
Delta R.T. -0.000 min
Lab File: BM047147.D
Acq: 10 Aug 2024 17:51




Instrument :
BNA_M
ClientSampleId :
SSTDICC080


Tgt Ion	Ratio	Lower	Upper
139	100		
109	30.5	24.4	36.6
65	52.3	43.0	64.6



#27
2,4-Dimethylphenol
Concen: 78.514 ng
RT: 9.345 min Scan# 1098
Delta R.T. -0.000 min
Lab File: BM047147.D
Acq: 10 Aug 2024 17:51

Tgt Ion	Ratio	Lower	Upper
122	100		
107	117.1	92.8	139.2
121	57.7	46.1	69.1









#28
bis(2-Chloroethoxy)methane
Concen: 72.808 ng
RT: 9.581 min Scan# 1138
Delta R.T. 0.006 min
Lab File: BM047147.D
Acq: 10 Aug 2024 17:51




Instrument :
BNA_M
ClientSampleId :
SSTDICC080


Tgt Ion: 93 Resp: 1908847
Ion Ratio Lower Upper
93 100
95 32.8 25.8 38.8
123 12.6 9.0 13.4



#29
2,4-Dichlorophenol
Concen: 80.115 ng
RT: 9.798 min Scan# 1175
Delta R.T. -0.000 min
Lab File: BM047147.D
Acq: 10 Aug 2024 17:51

Tgt Ion: 162 Resp: 1485940
Ion Ratio Lower Upper
162 100
164 64.8 44.7 84.7
98 34.7 14.9 54.9





#30

1,2,4-Trichlorobenzene

Concen: 77.458 ng

RT: 10.004 min Scan# 11

Delta R.T. 0.006 min

Lab File: BM047147.D

Acq: 10 Aug 2024 17:51

Instrument :

BNA_M

ClientSampleId :

SSTDICC080





Tgt Ion:180 Resp: 1625842

Ion Ratio Lower Upper

180 100

182 96.0 76.6 115.0

145 28.7 23.4 35.2



#31

Naphthalene

Concen: 75.304 ng

RT: 10.181 min Scan# 1240

Delta R.T. -0.000 min

Lab File: BM047147.D

Acq: 10 Aug 2024 17:51




Tgt Ion:128 Resp: 4520274


Ion Ratio Lower Upper

128 100

129 11.2 8.6 13.0

127 13.0 10.2 15.4









#32
Benzoic acid
Concen: 88.571 ng
RT: 9.575 min Scan# 1137
Delta R.T. 0.041 min
Lab File: BM047147.D
Acq: 10 Aug 2024 17:51




Instrument :
BNA_M
ClientSampleId :
SSTDICC080


Tgt Ion:122 Resp: 1293498
Ion Ratio Lower Upper
122 100
105 126.0 107.1 147.1
77 100.9 82.1 122.1



#33
4-Chloroaniline
Concen: 75.970 ng
RT: 10.316 min Scan# 1263
Delta R.T. 0.006 min
Lab File: BM047147.D
Acq: 10 Aug 2024 17:51

Tgt Ion:127 Resp: 1753660
Ion Ratio Lower Upper
127 100
129 32.4 25.8 38.8
65 31.4 24.9 37.3
92 19.2 15.1 22.7









#34
Hexachlorobutadiene
Concen: 80.736 ng
RT: 10.469 min Scan# 11
Delta R.T. -0.000 min
Lab File: BM047147.D
Acq: 10 Aug 2024 17:51




Instrument :
BNA_M
ClientSampleId :
SSTDICC080


Tgt Ion:225 Resp: 991890
Ion Ratio Lower Upper
225 100
223 64.1 50.2 75.2
227 64.1 51.4 77.0



#35
Caprolactam
Concen: 81.288 ng m
RT: 11.151 min Scan# 1405
Delta R.T. 0.029 min
Lab File: BM047147.D
Acq: 10 Aug 2024 17:51

Tgt Ion:113 Resp: 523592
Ion Ratio Lower Upper
113 100
55 155.2 140.5 180.5
56 118.1 103.6 143.6









#36
4-Chloro-3-methylphenol
Concen: 78.951 ng
RT: 11.463 min Scan# 1458
Delta R.T. 0.012 min
Lab File: BM047147.D
Acq: 10 Aug 2024 17:51




Instrument :
BNA_M
ClientSampleId :
SSTDICC080


Tgt Ion:107 Resp: 1559534
Ion Ratio Lower Upper
107 100
144 26.7 21.8 32.6
142 82.3 66.8 100.2



#37
2-Methylnaphthalene
Concen: 77.236 ng
RT: 11.810 min Scan# 1517
Delta R.T. 0.006 min
Lab File: BM047147.D
Acq: 10 Aug 2024 17:51

Tgt Ion:142 Resp: 3301008
Ion Ratio Lower Upper
142 100
141 88.0 70.9 106.3
115 35.1 28.7 43.1





#38

1-Methylnaphthalene

Concen: 76.532 ng

RT: 12.033 min Scan# 1

Delta R.T. -0.000 min

Lab File: BM047147.D

Acq: 10 Aug 2024 17:51

Instrument :

BNA_M

ClientSampleId :

SSTDICC080





Tgt Ion:142 Resp: 3232647

Ion Ratio Lower Upper

142 100

141 92.1 74.6 111.8

116 3.9 3.1 4.7



#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 14.033 min Scan# 1895

Delta R.T. -0.000 min

Lab File: BM047147.D

Acq: 10 Aug 2024 17:51




Tgt Ion:164 Resp: 795761


Ion Ratio Lower Upper

164 100

162 99.7 80.7 121.1

160 44.4 36.0 54.0





#40

1,2,4,5-Tetrachlorobenzene

Concen: 80.017 ng

RT: 12.192 min Scan# 1

Delta R.T. 0.006 min

Lab File: BM047147.D

Acq: 10 Aug 2024 17:51

Instrument :





BNA_M

ClientSampleId :

SSTDICC080

Tgt Ion:216 Resp: 1736758

Ion	Ratio	Lower	Upper
216	100		
214	78.9	63.4	95.2
179	21.7	17.6	26.4
108	17.7	15.1	22.7



#41

Hexachlorocyclopentadiene

Concen: 81.830 ng

RT: 12.169 min Scan# 1578




Delta R.T. 0.006 min


Lab File: BM047147.D

Acq: 10 Aug 2024 17:51

Tgt Ion:237 Resp: 620195

Ion	Ratio	Lower	Upper
237	100		
235	62.3	44.8	84.8
272	12.6	0.0	33.5









#42
2,4,6-Tribromophenol
Concen: 155.767 ng
RT: 15.545 min Scan# 2151
Delta R.T. 0.006 min
Lab File: BM047147.D
Acq: 10 Aug 2024 17:51




Instrument :
BNA_M
ClientSampleId :
SSTDICC080


Tgt Ion:330 Resp: 1437262
Ion Ratio Lower Upper
330 100
332 96.9 77.4 116.2
141 37.8 31.6 47.4



#43
2,4,6-Trichlorophenol
Concen: 81.622 ng
RT: 12.451 min Scan# 1626
Delta R.T. 0.006 min
Lab File: BM047147.D
Acq: 10 Aug 2024 17:51

Tgt Ion:196 Resp: 1259185
Ion Ratio Lower Upper
196 100
198 95.2 75.6 113.4
200 30.7 24.6 37.0









#44
2,4,5-Trichlorophenol
Concen: 80.776 ng
RT: 12.522 min Scan# 1638
Delta R.T. -0.000 min
Lab File: BM047147.D
Acq: 10 Aug 2024 17:51




Instrument :
BNA_M
ClientSampleId :
SSTDICC080


Tgt Ion:196 Resp: 1383585
Ion Ratio Lower Upper
196 100
198 95.4 77.7 116.5
97 51.0 39.8 59.8
132 27.2 20.7 31.1



#45
2-Fluorobiphenyl
Concen: 150.612 ng
RT: 12.645 min Scan# 1659
Delta R.T. 0.000 min
Lab File: BM047147.D
Acq: 10 Aug 2024 17:51

Tgt Ion:172 Resp: 7769782
Ion Ratio Lower Upper
172 100
171 35.3 27.8 41.6
170 23.4 18.7 28.1









#46
1,1'-Biphenyl
Concen: 74.866 ng
RT: 12.857 min Scan# 1695
Delta R.T. -0.000 min
Lab File: BM047147.D
Acq: 10 Aug 2024 17:51




Instrument :
BNA_M
ClientSampleId :
SSTDICC080


Tgt Ion:154 Resp: 4241252
Ion Ratio Lower Upper
154 100
153 40.2 19.0 59.0
76 12.2 0.0 31.9



#47
2-Chloronaphthalene
Concen: 74.960 ng
RT: 12.892 min Scan# 1701
Delta R.T. 0.006 min
Lab File: BM047147.D
Acq: 10 Aug 2024 17:51

Tgt Ion:162 Resp: 3319138
Ion Ratio Lower Upper
162 100
127 36.8 30.0 45.0
164 33.4 26.3 39.5









#48
2-Nitroaniline
Concen: 78.614 ng
RT: 13.122 min Scan# 1739
Delta R.T. 0.006 min
Lab File: BM047147.D
Acq: 10 Aug 2024 17:51




Instrument :
BNA_M
ClientSampleId :
SSTDICC080


Tgt Ion: 65 Resp: 1090521
Ion Ratio Lower Upper
65 100
92 69.4 55.1 82.7
138 109.5 85.9 128.9



#49
Acenaphthylene
Concen: 75.775 ng
RT: 13.751 min Scan# 1847
Delta R.T. -0.000 min
Lab File: BM047147.D
Acq: 10 Aug 2024 17:51

Tgt Ion: 152 Resp: 4945706
Ion Ratio Lower Upper
152 100
151 20.0 15.8 23.6
153 13.4 10.4 15.6









#50
Dimethylphthalate
Concen: 75.503 ng
RT: 13.522 min Scan# 1808
Delta R.T. 0.006 min
Lab File: BM047147.D
Acq: 10 Aug 2024 17:51




Instrument :
BNA_M
ClientSampleId :
SSTDICC080


Tgt Ion:163 Resp: 4196226
Ion Ratio Lower Upper
163 100
194 4.1 3.0 4.6
164 10.4 8.0 12.0



#51
2,6-Dinitrotoluene
Concen: 79.128 ng
RT: 13.639 min Scan# 1828
Delta R.T. 0.012 min
Lab File: BM047147.D
Acq: 10 Aug 2024 17:51

Tgt Ion:165 Resp: 1028020
Ion Ratio Lower Upper
165 100
63 50.5 46.5 69.7
89 52.9 47.3 70.9









#52
Acenaphthene
Concen: 75.941 ng
RT: 14.104 min Scan# 1906
Delta R.T. 0.006 min
Lab File: BM047147.D
Acq: 10 Aug 2024 17:51




Instrument :
BNA_M
ClientSampleId :
SSTDICC080


Tgt Ion:154 Resp: 3144083
Ion Ratio Lower Upper
154 100
153 112.4 90.3 135.5
152 52.9 42.9 64.3



#53
3-Nitroaniline
Concen: 79.468 ng
RT: 13.969 min Scan# 1884
Delta R.T. 0.006 min
Lab File: BM047147.D
Acq: 10 Aug 2024 17:51

Tgt Ion:138 Resp: 1034545
Ion Ratio Lower Upper
138 100
108 9.9 7.9 11.9
92 115.5 92.0 138.0









#54
2,4-Dinitrophenol
Concen: 86.639 ng
RT: 14.180 min Scan# 1919
Delta R.T. 0.005 min
Lab File: BM047147.D
Acq: 10 Aug 2024 17:51




Instrument :
BNA_M
ClientSampleId :
SSTDICC080


Tgt Ion:184 Resp: 674624
Ion Ratio Lower Upper
184 100
63 60.8 52.7 79.1
154 63.4 50.4 75.6



#55
Dibenzofuran
Concen: 74.735 ng
RT: 14.445 min Scan# 1965
Delta R.T. 0.006 min
Lab File: BM047147.D
Acq: 10 Aug 2024 17:51

Tgt Ion:168 Resp: 4887362
Ion Ratio Lower Upper
168 100
139 36.3 29.6 44.4
169 13.1 10.7 16.1









#56
4-Nitrophenol
Concen: 81.334 ng
RT: 14.304 min Scan# 1941
Delta R.T. 0.006 min
Lab File: BM047147.D
Acq: 10 Aug 2024 17:51




Instrument :
BNA_M
ClientSampleId :
SSTDICC080


Tgt Ion:139 Resp: 913035
Ion Ratio Lower Upper
139 100
109 76.5 52.5 92.5
65 99.0 75.4 115.4



#57
2,4-Dinitrotoluene
Concen: 79.012 ng
RT: 14.433 min Scan# 1963
Delta R.T. 0.006 min
Lab File: BM047147.D
Acq: 10 Aug 2024 17:51

Tgt Ion:165 Resp: 1370129
Ion Ratio Lower Upper
165 100
63 54.0 40.3 60.5
89 78.0 61.0 91.6
182 2.7 1.9 2.9









#58
Fluorene
Concen: 74.602 ng
RT: 15.098 min Scan# 2075
Delta R.T. 0.006 min
Lab File: BM047147.D
Acq: 10 Aug 2024 17:51




Instrument :
BNA_M
ClientSampleId :
SSTDICC080


Tgt Ion:166 Resp: 4025459
Ion Ratio Lower Upper
166 100
165 97.2 77.4 116.0
167 13.7 11.0 16.6



#59
2,3,4,6-Tetrachlorophenol
Concen: 80.180 ng
RT: 14.680 min Scan# 2005
Delta R.T. 0.006 min
Lab File: BM047147.D
Acq: 10 Aug 2024 17:51

Tgt Ion:232 Resp: 1128723
Ion Ratio Lower Upper
232 100
131 44.8 36.9 55.3
130 2.3 2.0 3.0
166 32.0 26.4 39.6









#60
Diethylphthalate
Concen: 75.010 ng
RT: 14.904 min Scan# 2043
Delta R.T. 0.006 min
Lab File: BM047147.D
Acq: 10 Aug 2024 17:51




Instrument :
BNA_M
ClientSampleId :
SSTDICC080


Tgt Ion:149 Resp: 4257104
Ion Ratio Lower Upper
149 100
177 23.0 17.6 26.4
150 12.8 9.8 14.6



#61
4-Chlorophenyl-phenylether
Concen: 77.748 ng
RT: 15.104 min Scan# 2077
Delta R.T. 0.006 min
Lab File: BM047147.D
Acq: 10 Aug 2024 17:51

Tgt Ion:204 Resp: 1997247
Ion Ratio Lower Upper
204 100
206 32.7 26.2 39.4
141 54.7 47.2 70.8









#62
4-Nitroaniline
Concen: 79.564 ng
RT: 15.145 min Scan# 2084
Delta R.T. 0.006 min
Lab File: BM047147.D
Acq: 10 Aug 2024 17:51




Instrument :
BNA_M
ClientSampleId :
SSTDICC080


Tgt Ion:138 Resp: 1020427
Ion Ratio Lower Upper
138 100
92 51.9 31.9 71.9
108 84.8 63.8 103.8



#63
Azobenzene
Concen: 72.861 ng
RT: 15.398 min Scan# 2127
Delta R.T. 0.006 min
Lab File: BM047147.D
Acq: 10 Aug 2024 17:51

Tgt Ion: 77 Resp: 3920766
Ion Ratio Lower Upper
77 100
182 26.7 3.8 43.8
105 20.2 0.0 39.5
51 26.4 6.8 46.8









#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 16.798 min Scan# 21
Delta R.T. 0.006 min
Lab File: BM047147.D
Acq: 10 Aug 2024 17:51




Instrument :
BNA_M
ClientSampleId :
SSTDICC080


Tgt Ion:188 Resp: 1674812
Ion Ratio Lower Upper
188 100
94 8.5 7.0 10.6
80 9.3 8.0 12.0



#65
4,6-Dinitro-2-methylphenol
Concen: 83.643 ng
RT: 15.204 min Scan# 2094
Delta R.T. 0.006 min
Lab File: BM047147.D
Acq: 10 Aug 2024 17:51

Tgt Ion:198 Resp: 820182
Ion Ratio Lower Upper
198 100
51 41.8 23.9 63.9
105 43.0 24.3 64.3









#66
n-Nitrosodiphenylamine
Concen: 75.560 ng
RT: 15.327 min Scan# 2114
Delta R.T. 0.006 min
Lab File: BM047147.D
Acq: 10 Aug 2024 17:51




Instrument :
BNA_M
ClientSampleId :
SSTDICC080


Tgt Ion:169 Resp: 3446873
Ion Ratio Lower Upper
169 100
168 66.8 53.8 80.8
167 37.0 29.3 43.9



#67
4-Bromophenyl-phenylether
Concen: 79.172 ng
RT: 15.998 min Scan# 2229
Delta R.T. -0.000 min
Lab File: BM047147.D
Acq: 10 Aug 2024 17:51

Tgt Ion:248 Resp: 1277742
Ion Ratio Lower Upper
248 100
250 97.9 77.8 116.8
141 71.0 57.7 86.5









#68
Hexachlorobenzene
Concen: 76.654 ng
RT: 16.104 min Scan# 21
Delta R.T. 0.006 min
Lab File: BM047147.D
Acq: 10 Aug 2024 17:51




Instrument :
BNA_M
ClientSampleId :
SSTDICC080


Tgt Ion:284 Resp: 1481615
Ion Ratio Lower Upper
284 100
142 34.2 29.4 44.0
249 31.6 25.1 37.7



#69
Atrazine
Concen: 62.797 ng
RT: 16.298 min Scan# 2280
Delta R.T. 0.006 min
Lab File: BM047147.D
Acq: 10 Aug 2024 17:51

Tgt Ion:200 Resp: 862020
Ion Ratio Lower Upper
200 100
173 25.4 6.1 46.1
215 47.7 26.8 66.8





#70

Pentachlorophenol

Concen: 82.323 ng

RT: 16.457 min Scan# 21

Delta R.T. 0.006 min

Lab File: BM047147.D

Acq: 10 Aug 2024 17:51

Instrument :

BNA_M

ClientSampleId :

SSTDICC080





Tgt Ion:266 Resp: 1115403

Ion Ratio Lower Upper

266 100

268 64.5 49.8 74.6

264 62.8 49.6 74.4



#71

Phenanthrene

Concen: 75.631 ng

RT: 16.839 min Scan# 2372

Delta R.T. -0.000 min

Lab File: BM047147.D

Acq: 10 Aug 2024 17:51




Tgt Ion:178 Resp: 6356974


Ion Ratio Lower Upper

178 100

176 19.3 15.3 22.9

179 15.6 12.4 18.6









#72
Anthracene
Concen: 75.850 ng
RT: 16.933 min Scan# 21
Delta R.T. 0.006 min
Lab File: BM047147.D
Acq: 10 Aug 2024 17:51




Instrument :
BNA_M
ClientSampleId :
SSTDICC080


Tgt Ion:178 Resp: 6203287
Ion Ratio Lower Upper
178 100
176 18.8 14.8 22.2
179 15.5 12.4 18.6



#73
Carbazole
Concen: 74.858 ng
RT: 17.210 min Scan# 2435
Delta R.T. -0.000 min
Lab File: BM047147.D
Acq: 10 Aug 2024 17:51

Tgt Ion:167 Resp: 6250976
Ion Ratio Lower Upper
167 100
166 21.4 17.1 25.7
139 12.4 10.0 15.0









#74
Di-n-butylphthalate
Concen: 75.700 ng
RT: 17.798 min Scan# 21
Delta R.T. 0.006 min
Lab File: BM047147.D
Acq: 10 Aug 2024 17:51




Instrument :
BNA_M
ClientSampleId :
SSTDICC080


Tgt Ion:149 Resp: 7355197
Ion Ratio Lower Upper
149 100
150 9.4 7.5 11.3
104 5.1 4.1 6.1



#75
Fluoranthene
Concen: 76.805 ng
RT: 18.874 min Scan# 2718
Delta R.T. 0.006 min
Lab File: BM047147.D
Acq: 10 Aug 2024 17:51

Tgt Ion:202 Resp: 7852169
Ion Ratio Lower Upper
202 100
101 9.2 0.0 30.2
203 17.8 0.0 37.6









#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.033 min Scan# 3085
Delta R.T. -0.000 min
Lab File: BM047147.D
Acq: 10 Aug 2024 17:51




Instrument :
BNA_M
ClientSampleId :
SSTDICC080


Tgt Ion:240 Resp: 1443265
Ion Ratio Lower Upper
240 100
120 10.2 8.2 12.2
236 25.4 20.6 30.8



#77
Benzidine
Concen: 79.424 ng
RT: 19.074 min Scan# 2752
Delta R.T. -0.000 min
Lab File: BM047147.D
Acq: 10 Aug 2024 17:51

Tgt Ion:184 Resp: 1943953
Ion Ratio Lower Upper
184 100
185 14.6 11.8 17.6
183 11.5 9.0 13.6









#78
Pyrene
Concen: 79.218 ng
RT: 19.239 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM047147.D
Acq: 10 Aug 2024 17:51




Instrument :
BNA_M
ClientSampleId :
SSTDICC080


Tgt Ion:202 Resp: 8163599
Ion Ratio Lower Upper
202 100
200 21.4 17.2 25.8
203 17.9 14.2 21.2



#79
Terphenyl-d14
Concen: 158.037 ng
RT: 19.462 min Scan# 2818
Delta R.T. -0.000 min
Lab File: BM047147.D
Acq: 10 Aug 2024 17:51

Tgt Ion:244 Resp:10644639
Ion Ratio Lower Upper
244 100
212 8.3 6.2 9.4
122 8.6 7.1 10.7









#80
Butylbenzylphthalate
Concen: 81.742 ng
RT: 20.174 min Scan# 2939
Delta R.T. -0.000 min
Lab File: BM047147.D
Acq: 10 Aug 2024 17:51




Instrument :
BNA_M
ClientSampleId :
SSTDICC080


Tgt Ion:149 Resp: 3437732
Ion Ratio Lower Upper
149 100
91 71.5 57.0 85.4
206 20.8 16.4 24.6



#81
Benzo(a)anthracene
Concen: 77.892 ng
RT: 21.015 min Scan# 3082
Delta R.T. -0.000 min
Lab File: BM047147.D
Acq: 10 Aug 2024 17:51

Tgt Ion:228 Resp: 7463024
Ion Ratio Lower Upper
228 100
226 28.0 21.8 32.6
229 20.1 15.8 23.8









#82
3,3'-Dichlorobenzidine
Concen: 78.698 ng
RT: 20.956 min Scan# 3072
Delta R.T. -0.000 min
Lab File: BM047147.D
Acq: 10 Aug 2024 17:51




Instrument :
BNA_M
ClientSampleId :
SSTDICC080


Tgt Ion:252 Resp: 2359199
Ion Ratio Lower Upper
252 100
254 64.8 51.0 76.6
126 10.4 8.5 12.7



#83
Chrysene
Concen: 76.926 ng
RT: 21.074 min Scan# 3092
Delta R.T. -0.000 min
Lab File: BM047147.D
Acq: 10 Aug 2024 17:51

Tgt Ion:228 Resp: 6992860
Ion Ratio Lower Upper
228 100
226 31.1 24.4 36.6
229 20.2 15.7 23.5









#84
Bis(2-ethylhexyl)phthalate
Concen: 79.001 ng
RT: 20.974 min Scan# 3075
Delta R.T. -0.000 min
Lab File: BM047147.D
Acq: 10 Aug 2024 17:51




Instrument :
BNA_M
ClientSampleId :
SSTDICC080


Tgt Ion:149 Resp: 4711826
Ion Ratio Lower Upper
149 100
167 28.1 22.2 33.2
279 4.1 3.0 4.4



#85
Di-n-octyl phthalate
Concen: 83.492 ng
RT: 22.009 min Scan# 3251
Delta R.T. -0.000 min
Lab File: BM047147.D
Acq: 10 Aug 2024 17:51

Tgt Ion:149 Resp: 8464049
Ion Ratio Lower Upper
149 100
167 1.6 1.3 1.9
43 8.1 6.9 10.3









#86
Perylene-d12
Concen: 20.000 ng
RT: 23.733 min Scan# 31
Delta R.T. 0.006 min
Lab File: BM047147.D
Acq: 10 Aug 2024 17:51




Instrument :
BNA_M
ClientSampleId :
SSTDICC080


Tgt Ion:264 Resp: 1789187
Ion Ratio Lower Upper
264 100
260 23.2 18.5 27.7
265 21.6 17.3 25.9



#87
Indeno(1,2,3-cd)pyrene
Concen: 80.058 ng
RT: 26.750 min Scan# 4057
Delta R.T. 0.023 min
Lab File: BM047147.D
Acq: 10 Aug 2024 17:51

Tgt Ion:276 Resp: 9258642
Ion Ratio Lower Upper
276 100
138 23.1 19.4 29.0
277 25.1 20.1 30.1









#88
Benzo(b)fluoranthene
Concen: 78.391 ng
RT: 22.897 min Scan# 3400
Delta R.T. 0.012 min
Lab File: BM047147.D
Acq: 10 Aug 2024 17:51




Instrument :
BNA_M
ClientSampleId :
SSTDICC080


Tgt Ion:252 Resp: 8329223
Ion Ratio Lower Upper
252 100
253 22.1 17.8 26.6
125 8.6 7.4 11.0



#89
Benzo(k)fluoranthene
Concen: 76.985 ng
RT: 22.950 min Scan# 3411
Delta R.T. 0.006 min
Lab File: BM047147.D
Acq: 10 Aug 2024 17:51

Tgt Ion:252 Resp: 7726234
Ion Ratio Lower Upper
252 100
253 22.3 17.7 26.5
125 8.7 7.4 11.2









#90
Benzo(a)pyrene
Concen: 79.723 ng
RT: 23.615 min Scan# 31
Delta R.T. 0.006 min
Lab File: BM047147.D
Acq: 10 Aug 2024 17:51




Instrument :
BNA_M
ClientSampleId :
SSTDICC080


Tgt Ion:252 Resp: 7338002
Ion Ratio Lower Upper
252 100
253 22.0 17.1 25.7
125 9.3 8.2 12.2



#91
Dibenzo(a,h)anthracene
Concen: 78.655 ng
RT: 26.803 min Scan# 4066
Delta R.T. 0.023 min
Lab File: BM047147.D
Acq: 10 Aug 2024 17:51

Tgt Ion:278 Resp: 7361474
Ion Ratio Lower Upper
278 100
139 14.3 12.0 18.0
279 23.6 19.2 28.8





#92

Benzo(g,h,i)perylene

Concen: 79.293 ng

RT: 27.685 min Scan# 41

Delta R.T. 0.023 min

Lab File: BM047147.D


Acq: 10 Aug 2024 17:51

Instrument :

BNA_M

ClientSampleId :

SSTDICC080





Tgt Ion:276 Resp: 7705035

Ion Ratio Lower Upper

276 100

277 23.6 18.8 28.2

138 19.4 16.2 24.4



Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081024\
 Data File : BM047148.D
 Acq On : 10 Aug 2024 19:10
 Operator : RC/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 ICVBM081024

Quant Time: Aug 11 23:52:13 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Aug 11 23:47:58 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.375	152	296367	20.000	ng	0.00
21) Naphthalene-d8	10.128	136	1206414	20.000	ng	0.00
39) Acenaphthene-d10	14.033	164	800129	20.000	ng	0.00
64) Phenanthrene-d10	16.792	188	1673175	20.000	ng	0.00
76) Chrysene-d12	21.027	240	1477990	20.000	ng	0.00
86) Perylene-d12	23.727	264	1812426	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.022	112	1247183	75.195	ng	0.00
7) Phenol-d6	6.587	99	1751460	76.636	ng	0.00
23) Nitrobenzene-d5	8.522	82	1846961	77.088	ng	0.00
42) 2,4,6-Tribromophenol	15.539	330	757124	81.608	ng	0.00
45) 2-Fluorobiphenyl	12.639	172	4083420	78.722	ng	0.00
79) Terphenyl-d14	19.456	244	5496459	79.687	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.034	88	253056	36.823	ng	98
3) Pyridine	3.410	79	765159m	37.969	ng	
4) n-Nitrosodimethylamine	3.328	42	328593	37.322	ng	100
6) Aniline	6.722	93	844703	39.104	ng	99
8) 2-Chlorophenol	6.951	128	703988	38.916	ng	98
9) Benzaldehyde	6.540	77	476271	39.535	ng	98
10) Phenol	6.610	94	856285	37.549	ng	99
11) bis(2-Chloroethyl)ether	6.828	93	741565	37.972	ng	99
12) 1,3-Dichlorobenzene	7.263	146	836221	38.900	ng	99
13) 1,4-Dichlorobenzene	7.410	146	845433	38.844	ng	98
14) 1,2-Dichlorobenzene	7.716	146	795642	38.744	ng	99
15) Benzyl Alcohol	7.628	79	643803	39.581	ng	97
16) 2,2'-oxybis(1-Chloropr...	7.916	45	947270	37.800	ng	99
17) 2-Methylphenol	7.834	107	595894	39.168	ng	99
18) Hexachloroethane	8.428	117	324138	38.447	ng	99
19) n-Nitroso-di-n-propyla...	8.187	70	582895	37.864	ng	99
20) 3+4-Methylphenols	8.163	107	844879	39.723	ng	99
22) Acetophenone	8.192	105	1109726	38.108	ng	# 98
24) Nitrobenzene	8.563	77	940366	38.978	ng	96
25) Isophorone	9.086	82	1650061	39.554	ng	100
26) 2-Nitrophenol	9.263	139	440983	41.274	ng	97
27) 2,4-Dimethylphenol	9.345	122	494953	39.547	ng	98
28) bis(2-Chloroethoxy)met...	9.575	93	1024303	39.022	ng	100
29) 2,4-Dichlorophenol	9.792	162	761049	40.983	ng	100
30) 1,2,4-Trichlorobenzene	9.998	180	846099	40.261	ng	100
31) Naphthalene	10.181	128	2317673	38.564	ng	100
32) Benzoic acid	9.522	122	626290	43.072	ng	99
33) 4-Chloroaniline	10.304	127	930580	40.265	ng	98
34) Hexachlorobutadiene	10.469	225	495141	40.254	ng	99
35) Caprolactam	11.116	113	261463	40.543	ng	98
36) 4-Chloro-3-methylphenol	11.451	107	787338	39.811	ng	100
37) 2-Methylnaphthalene	11.804	142	1667488	38.968	ng	100
38) 1-Methylnaphthalene	12.033	142	1644135	38.877	ng	99
40) 1,2,4,5-Tetrachloroben...	12.186	216	885065	40.555	ng	99
41) Hexachlorocyclopentadiene	12.163	237	312767	41.042	ng	98
43) 2,4,6-Trichlorophenol	12.445	196	632529	40.777	ng	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081024\
 Data File : BM047148.D
 Acq On : 10 Aug 2024 19:10
 Operator : RC/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 ICVBM081024

Quant Time: Aug 11 23:52:13 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Aug 11 23:47:58 2024
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	12.516	196	707267	41.066	ng	99
46) 1,1'-Biphenyl	12.851	154	2199985	38.622	ng	98
47) 2-Chloronaphthalene	12.886	162	1712308	38.460	ng	100
48) 2-Nitroaniline	13.116	65	561121	40.230	ng	100
49) Acenaphthylene	13.751	152	2577143	39.270	ng	100
50) Dimethylphthalate	13.516	163	2199717	39.364	ng	100
51) 2,6-Dinitrotoluene	13.627	165	526940	40.338	ng	96
52) Acenaphthene	14.098	154	1616361	38.828	ng	100
53) 3-Nitroaniline	13.957	138	529385	40.443	ng	98
54) 2,4-Dinitrophenol	14.174	184	315633	40.314	ng	97
55) Dibenzofuran	14.439	168	2550249	38.784	ng	99
56) 4-Nitrophenol	14.292	139	466721	41.349	ng	98
57) 2,4-Dinitrotoluene	14.427	165	708519	40.635	ng	98
58) Fluorene	15.092	166	2108277	38.859	ng	100
59) 2,3,4,6-Tetrachlorophenol	14.674	232	577329	40.787	ng	99
60) Diethylphthalate	14.898	149	2232502	39.122	ng	99
61) 4-Chlorophenyl-phenyle...	15.098	204	1001595	38.777	ng	99
62) 4-Nitroaniline	15.133	138	517467	40.127	ng	98
63) Azobenzene	15.392	77	2048152	37.854	ng	98
65) 4,6-Dinitro-2-methylph...	15.198	198	416340	42.500	ng	94
66) n-Nitrosodiphenylamine	15.321	169	1790716	39.293	ng	100
67) 4-Bromophenyl-phenylether	15.998	248	647415	40.155	ng	97
68) Hexachlorobenzene	16.098	284	774906	40.130	ng	98
69) Atrazine	16.292	200	566536	41.312	ng	100
70) Pentachlorophenol	16.451	266	561329	41.470	ng	98
71) Phenanthrene	16.833	178	3272056	38.967	ng	100
72) Anthracene	16.927	178	3237131	39.620	ng	100
73) Carbazole	17.209	167	3262489	39.108	ng	99
74) Di-n-butylphthalate	17.792	149	3851216	39.675	ng	100
75) Fluoranthene	18.868	202	3981412	38.982	ng	99
77) Benzidine	19.074	184	1486859	59.321	ng	100
78) Pyrene	19.233	202	4150742	39.331	ng	99
80) Butylbenzylphthalate	20.174	149	1778335	41.291	ng	98
81) Benzo(a)anthracene	21.015	228	3835235	39.088	ng	100
82) 3,3'-Dichlorobenzidine	20.956	252	1322453	43.078	ng	98
83) Chrysene	21.068	228	3598062	38.651	ng	99
84) Bis(2-ethylhexyl)phtha...	20.974	149	2464399	40.349	ng	100
85) Di-n-octyl phthalate	22.003	149	4287924	41.304	ng	99
87) Indeno(1,2,3-cd)pyrene	26.726	276	4748737	40.535	ng	99
88) Benzo(b)fluoranthene	22.886	252	4137733	38.443	ng	100
89) Benzo(k)fluoranthene	22.944	252	4000536	39.350	ng	99
90) Benzo(a)pyrene	23.603	252	3707577	39.764	ng	99
91) Dibenzo(a,h)anthracene	26.774	278	3803894	40.122	ng	99
92) Benzo(g,h,i)perylene	27.656	276	3997872	40.615	ng	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081024\
 Data File : BM047148.D
 Acq On : 10 Aug 2024 19:10
 Operator : RC/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 ICVBM081024

Quant Time: Aug 11 23:52:13 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Aug 11 23:47:58 2024
 Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.375 min Scan# 763
Delta R.T. -0.000 min
Lab File: BM047148.D
Acq: 10 Aug 2024 19:10




Instrument :
BNA_M
ClientSampleId :
ICVBM081024


Tgt Ion:152 Resp: 296367
Ion Ratio Lower Upper
152 100
150 151.6 120.5 180.7
115 54.8 45.5 68.3



#2
1,4-Dioxane
Concen: 36.823 ng
RT: 3.034 min Scan# 25
Delta R.T. -0.000 min
Lab File: BM047148.D
Acq: 10 Aug 2024 19:10

Tgt Ion: 88 Resp: 253056
Ion Ratio Lower Upper
88 100
58 70.9 58.2 87.4
43 27.5 22.5 33.7









#3
Pyridine
Concen: 37.969 ng m
RT: 3.410 min Scan# 89
Delta R.T. -0.000 min
Lab File: BM047148.D
Acq: 10 Aug 2024 19:10




Instrument :
BNA_M
ClientSampleId :
ICVBM081024


Tgt Ion: 79 Resp: 765159
Ion Ratio Lower Upper
79 100
52 62.2 51.2 76.8
51 31.2 26.1 39.1



#4
n-Nitrosodimethylamine
Concen: 37.322 ng
RT: 3.328 min Scan# 75
Delta R.T. -0.000 min
Lab File: BM047148.D
Acq: 10 Aug 2024 19:10

Tgt Ion: 42 Resp: 328593
Ion Ratio Lower Upper
42 100
74 141.2 112.9 169.3
44 7.2 5.3 7.9









#5
2-Fluorophenol
Concen: 75.195 ng
RT: 5.022 min Scan# 363
Delta R.T. -0.000 min
Lab File: BM047148.D
Acq: 10 Aug 2024 19:10




Instrument :
BNA_M
ClientSampleId :
ICVBM081024


Tgt Ion:112 Resp: 1247183
Ion Ratio Lower Upper
112 100
64 57.9 46.7 70.1
63 31.1 24.8 37.2



#6
Aniline
Concen: 39.104 ng
RT: 6.722 min Scan# 652
Delta R.T. -0.006 min
Lab File: BM047148.D
Acq: 10 Aug 2024 19:10

Tgt Ion: 93 Resp: 844703
Ion Ratio Lower Upper
93 100
66 38.2 31.4 47.2
65 19.6 15.6 23.4









#7
Phenol-d6
Concen: 76.636 ng
RT: 6.587 min Scan# 61
Delta R.T. -0.000 min
Lab File: BM047148.D
Acq: 10 Aug 2024 19:10




Instrument :
BNA_M
ClientSampleId :
ICVBM081024


Tgt Ion: 99 Resp: 1751460
Ion Ratio Lower Upper
99 100
42 17.0 14.1 21.1
71 35.5 28.5 42.7



#8
2-Chlorophenol
Concen: 38.916 ng
RT: 6.951 min Scan# 691
Delta R.T. -0.000 min
Lab File: BM047148.D
Acq: 10 Aug 2024 19:10

Tgt Ion: 128 Resp: 703988
Ion Ratio Lower Upper
128 100
130 32.0 12.1 52.1
64 44.6 26.6 66.6









#9
Benzaldehyde
Concen: 39.535 ng
RT: 6.540 min Scan# 61
Delta R.T. -0.000 min
Lab File: BM047148.D
Acq: 10 Aug 2024 19:10




Instrument :
BNA_M
ClientSampleId :
ICVBM081024


Tgt Ion: 77 Resp: 476271
Ion Ratio Lower Upper
77 100
105 97.6 74.5 114.5
106 91.7 72.3 112.3



#10
Phenol
Concen: 37.549 ng
RT: 6.610 min Scan# 633
Delta R.T. -0.000 min
Lab File: BM047148.D
Acq: 10 Aug 2024 19:10

Tgt Ion: 94 Resp: 856285
Ion Ratio Lower Upper
94 100
65 29.0 9.0 49.0
66 39.5 18.9 58.9









#11
bis(2-Chloroethyl)ether
Concen: 37.972 ng
RT: 6.828 min Scan# 611
Delta R.T. -0.000 min
Lab File: BM047148.D
Acq: 10 Aug 2024 19:10




Instrument :
BNA_M
ClientSampleId :
ICVBM081024


Tgt Ion: 93 Resp: 741565
Ion Ratio Lower Upper
93 100
63 74.5 55.8 95.8
95 32.9 12.7 52.7



#12
1,3-Dichlorobenzene
Concen: 38.900 ng
RT: 7.263 min Scan# 744
Delta R.T. -0.000 min
Lab File: BM047148.D
Acq: 10 Aug 2024 19:10

Tgt Ion: 146 Resp: 836221
Ion Ratio Lower Upper
146 100
148 64.3 52.3 78.5
75 31.6 25.8 38.6





#13

1,4-Dichlorobenzene

Concen: 38.844 ng

RT: 7.410 min Scan# 7

Delta R.T. -0.000 min

Lab File: BM047148.D

Acq: 10 Aug 2024 19:10

Instrument :

BNA_M

ClientSampleId :

ICVBM081024





Tgt Ion:146 Resp: 845433

Ion Ratio Lower Upper

146 100

148 65.9 51.4 77.2

111 38.5 31.5 47.3



#14

1,2-Dichlorobenzene

Concen: 38.744 ng

RT: 7.716 min Scan# 821

Delta R.T. -0.000 min

Lab File: BM047148.D

Acq: 10 Aug 2024 19:10




Tgt Ion:146 Resp: 795642


Ion Ratio Lower Upper

146 100

148 63.6 51.8 77.8

111 41.1 33.3 49.9









#15
Benzyl Alcohol
Concen: 39.581 ng
RT: 7.628 min Scan# 806
Delta R.T. -0.000 min
Lab File: BM047148.D
Acq: 10 Aug 2024 19:10




Instrument :
BNA_M
ClientSampleId :
ICVBM081024


Tgt Ion: 79 Resp: 643803
Ion Ratio Lower Upper
79 100
108 76.0 59.7 89.5
77 63.1 52.9 79.3



#16
2,2'-oxybis(1-chloropropane)
Concen: 37.800 ng
RT: 7.916 min Scan# 855
Delta R.T. -0.000 min
Lab File: BM047148.D
Acq: 10 Aug 2024 19:10

Tgt Ion: 45 Resp: 947270
Ion Ratio Lower Upper
45 100
77 15.5 0.0 35.6
79 12.8 0.0 32.4





#17

2-Methylphenol

Concen: 39.168 ng

RT: 7.834 min Scan# 841

Delta R.T. -0.000 min

Lab File: BM047148.D


Acq: 10 Aug 2024 19:10

Instrument :

BNA_M



ClientSampleId :

ICVBM081024



Tgt Ion:107 Resp: 595894

Ion	Ratio	Lower	Upper
107	100		
108	113.5	90.6	135.8
77	49.5	39.3	58.9
79	47.6	38.6	58.0



#18

Hexachloroethane

Concen: 38.447 ng

RT: 8.428 min Scan# 942




Delta R.T. -0.000 min


Lab File: BM047148.D

Acq: 10 Aug 2024 19:10

Tgt Ion:117 Resp: 324138

Ion	Ratio	Lower	Upper
117	100		
119	96.9	77.0	115.4
201	101.3	81.3	121.9





#19

n-Nitroso-di-n-propylamine

Concen: 37.864 ng

RT: 8.187 min Scan# 901

Delta R.T. -0.000 min

Lab File: BM047148.D

Acq: 10 Aug 2024 19:10

Instrument :

BNA_M

ClientSampleId :

ICVBM081024

Tgt Ion: 70 Resp: 582895





Ion Ratio Lower Upper

70 100

42 51.7 42.0 63.0

101 9.7 7.2 10.8

130 19.8 15.5 23.3



#20

3+4-Methylphenols

Concen: 39.723 ng

RT: 8.163 min Scan# 897

Delta R.T. -0.000 min

Lab File: BM047148.D

Acq: 10 Aug 2024 19:10

Tgt Ion: 107 Resp: 844879




Ion Ratio Lower Upper


107 100

108 90.3 70.7 110.7

77 36.8 17.4 57.4

79 30.6 10.9 50.9









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.128 min Scan# 1231
Delta R.T. -0.006 min
Lab File: BM047148.D
Acq: 10 Aug 2024 19:10




Instrument :
BNA_M
ClientSampleId :
ICVBM081024


Tgt Ion:136 Resp: 1206414
Ion Ratio Lower Upper
136 100
137 10.9 8.8 13.2
54 7.6 6.0 9.0
68 6.2 4.8 7.2



#22
Acetophenone
Concen: 38.108 ng
RT: 8.192 min Scan# 902
Delta R.T. -0.006 min
Lab File: BM047148.D
Acq: 10 Aug 2024 19:10

Tgt Ion:105 Resp: 1109726
Ion Ratio Lower Upper
105 100
71 7.6 4.5 6.7#
51 25.9 20.1 30.1
120 23.3 17.8 26.6





#23

Nitrobenzene-d5

Concen: 77.088 ng

RT: 8.522 min Scan# 91

Delta R.T. -0.000 min

Lab File: BM047148.D

Acq: 10 Aug 2024 19:10

Instrument :





BNA_M

ClientSampleId :

ICVBM081024

Tgt Ion: 82 Resp: 1846961

Ion	Ratio	Lower	Upper
82	100		
128	41.6	32.3	48.5
54	46.5	37.4	56.2



#24

Nitrobenzene

Concen: 38.978 ng

RT: 8.563 min Scan# 965




Delta R.T. -0.000 min


Lab File: BM047148.D

Acq: 10 Aug 2024 19:10

Tgt Ion: 77 Resp: 940366

Ion	Ratio	Lower	Upper
77	100		
123	45.3	34.0	51.0
65	13.4	10.5	15.7





#25

Isophorone

Concen: 39.554 ng

RT: 9.086 min Scan# 1054

Delta R.T. -0.006 min

Lab File: BM047148.D

Acq: 10 Aug 2024 19:10

Instrument :





BNA_M

ClientSampleId :

ICVBM081024

Tgt Ion: 82 Resp: 1650061

Ion	Ratio	Lower	Upper
82	100		
95	7.1	5.5	8.3
138	17.0	13.5	20.3



#26

2-Nitrophenol

Concen: 41.274 ng

RT: 9.263 min Scan# 1084




Delta R.T. -0.000 min


Lab File: BM047148.D

Acq: 10 Aug 2024 19:10

Tgt Ion:139 Resp: 440983

Ion	Ratio	Lower	Upper
139	100		
109	30.4	24.4	36.6
65	50.6	43.0	64.6





#27

2,4-Dimethylphenol

Concen: 39.547 ng

RT: 9.345 min Scan# 1098

Delta R.T. -0.000 min

Lab File: BM047148.D


Acq: 10 Aug 2024 19:10

Instrument :

BNA_M




ClientSampleId :

ICVBM081024



Tgt Ion:122 Resp: 494953

Ion	Ratio	Lower	Upper
122	100		
107	117.9	92.8	139.2
121	58.4	46.1	69.1



#28

bis(2-Chloroethoxy)methane

Concen: 39.022 ng

RT: 9.575 min Scan# 1137




Delta R.T. -0.000 min


Lab File: BM047148.D

Acq: 10 Aug 2024 19:10

Tgt Ion: 93 Resp: 1024303

Ion	Ratio	Lower	Upper
93	100		
95	32.5	25.8	38.8
123	11.4	9.0	13.4





#29

2,4-Dichlorophenol

Concen: 40.983 ng

RT: 9.792 min Scan# 1174

Delta R.T. -0.006 min

Lab File: BM047148.D


Acq: 10 Aug 2024 19:10

Instrument :

BNA_M

ClientSampleId :

ICVBM081024






Tgt Ion:162 Resp: 761049

Ion Ratio Lower Upper

162 100

164 64.3 44.7 84.7

98 35.1 14.9 54.9



#30

1,2,4-Trichlorobenzene

Concen: 40.261 ng

RT: 9.998 min Scan# 1209

Delta R.T. -0.000 min

Lab File: BM047148.D

Acq: 10 Aug 2024 19:10




Tgt Ion:180 Resp: 846099


Ion Ratio Lower Upper

180 100

182 95.8 76.6 115.0

145 28.8 23.4 35.2









#31
Naphthalene
Concen: 38.564 ng
RT: 10.181 min Scan# 1128
Delta R.T. -0.000 min
Lab File: BM047148.D
Acq: 10 Aug 2024 19:10




Instrument :
BNA_M
ClientSampleId :
ICVBM081024


Tgt Ion:128 Resp: 2317673
Ion Ratio Lower Upper
128 100
129 10.9 8.6 13.0
127 12.7 10.2 15.4



#32
Benzoic acid
Concen: 43.072 ng
RT: 9.522 min Scan# 1128
Delta R.T. -0.012 min
Lab File: BM047148.D
Acq: 10 Aug 2024 19:10

Tgt Ion:122 Resp: 626290
Ion Ratio Lower Upper
122 100
105 126.1 107.1 147.1
77 100.0 82.1 122.1









#33
4-Chloroaniline
Concen: 40.265 ng
RT: 10.304 min Scan# 1261
Delta R.T. -0.006 min
Lab File: BM047148.D
Acq: 10 Aug 2024 19:10




Instrument :
BNA_M
ClientSampleId :
ICVBM081024


Tgt Ion:	127	Resp:	930580
Ion	Ratio	Lower	Upper
127	100		
129	31.5	25.8	38.8
65	32.1	24.9	37.3
92	19.6	15.1	22.7



#34
Hexachlorobutadiene
Concen: 40.254 ng
RT: 10.469 min Scan# 1289
Delta R.T. -0.000 min
Lab File: BM047148.D
Acq: 10 Aug 2024 19:10

Tgt Ion:	225	Resp:	495141
Ion	Ratio	Lower	Upper
225	100		
223	63.0	50.2	75.2
227	65.9	51.4	77.0









#35
Caprolactam
Concen: 40.543 ng
RT: 11.116 min Scan# 1116
Delta R.T. -0.006 min
Lab File: BM047148.D
Acq: 10 Aug 2024 19:10




Instrument :
BNA_M
ClientSampleId :
ICVBM081024


Tgt Ion:113 Resp: 261463
Ion Ratio Lower Upper
113 100
55 157.5 140.5 180.5
56 121.0 103.6 143.6



#36
4-Chloro-3-methylphenol
Concen: 39.811 ng
RT: 11.451 min Scan# 1456
Delta R.T. -0.000 min
Lab File: BM047148.D
Acq: 10 Aug 2024 19:10

Tgt Ion:107 Resp: 787338
Ion Ratio Lower Upper
107 100
144 27.0 21.8 32.6
142 83.5 66.8 100.2









#37
2-Methylnaphthalene
Concen: 38.968 ng
RT: 11.804 min Scan# 1516
Delta R.T. -0.000 min
Lab File: BM047148.D
Acq: 10 Aug 2024 19:10




Instrument :
BNA_M
ClientSampleId :
ICVBM081024


Tgt Ion:142 Resp: 1667488
Ion Ratio Lower Upper
142 100
141 88.3 70.9 106.3
115 35.5 28.7 43.1



#38
1-Methylnaphthalene
Concen: 38.877 ng
RT: 12.033 min Scan# 1555
Delta R.T. -0.000 min
Lab File: BM047148.D
Acq: 10 Aug 2024 19:10

Tgt Ion:142 Resp: 1644135
Ion Ratio Lower Upper
142 100
141 92.2 74.6 111.8
116 3.7 3.1 4.7





#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 14.033 min Scan# 1895

Delta R.T. -0.000 min

Lab File: BM047148.D

Acq: 10 Aug 2024 19:10

Instrument :

BNA_M

ClientSampleId :

ICVBM081024





Tgt Ion:164 Resp: 800129

Ion Ratio Lower Upper

164 100

162 99.7 80.7 121.1

160 44.3 36.0 54.0



#40

1,2,4,5-Tetrachlorobenzene

Concen: 40.555 ng

RT: 12.186 min Scan# 1581

Delta R.T. -0.000 min

Lab File: BM047148.D

Acq: 10 Aug 2024 19:10

Tgt Ion:216 Resp: 885065




Ion Ratio Lower Upper


216 100

214 78.6 63.4 95.2

179 21.5 17.6 26.4

108 18.4 15.1 22.7









#41
Hexachlorocyclopentadiene
Concen: 41.042 ng
RT: 12.163 min Scan# 1577
Delta R.T. -0.000 min
Lab File: BM047148.D
Acq: 10 Aug 2024 19:10




Instrument :
BNA_M
ClientSampleId :
ICVBM081024


Tgt Ion:237 Resp: 312767
Ion Ratio Lower Upper
237 100
235 63.3 44.8 84.8
272 13.6 0.0 33.5



#42
2,4,6-Tribromophenol
Concen: 81.608 ng
RT: 15.539 min Scan# 2151
Delta R.T. -0.000 min
Lab File: BM047148.D
Acq: 10 Aug 2024 19:10

Tgt Ion:330 Resp: 757124
Ion Ratio Lower Upper
330 100
332 97.0 77.4 116.2
141 38.1 31.6 47.4









#43
2,4,6-Trichlorophenol
Concen: 40.777 ng
RT: 12.445 min Scan# 1625
Delta R.T. -0.000 min
Lab File: BM047148.D
Acq: 10 Aug 2024 19:10




Instrument :
BNA_M
ClientSampleId :
ICVBM081024


Tgt Ion:196 Resp: 632529
Ion Ratio Lower Upper
196 100
198 97.3 75.6 113.4
200 30.8 24.6 37.0



#44
2,4,5-Trichlorophenol
Concen: 41.066 ng
RT: 12.516 min Scan# 1637
Delta R.T. -0.006 min
Lab File: BM047148.D
Acq: 10 Aug 2024 19:10

Tgt Ion:196 Resp: 707267
Ion Ratio Lower Upper
196 100
198 95.4 77.7 116.5
97 50.5 39.8 59.8
132 25.6 20.7 31.1





#45

2-Fluorobiphenyl

Concen: 78.722 ng

RT: 12.639 min Scan# 1658

Delta R.T. -0.006 min

Lab File: BM047148.D

Acq: 10 Aug 2024 19:10

Instrument :

BNA_M

ClientSampleId :

ICVBM081024





Tgt Ion:172 Resp: 4083420

Ion Ratio Lower Upper

172 100

171 34.6 27.8 41.6

170 23.2 18.7 28.1



#46

1,1'-Biphenyl

Concen: 38.622 ng

RT: 12.851 min Scan# 1694

Delta R.T. -0.006 min

Lab File: BM047148.D

Acq: 10 Aug 2024 19:10




Tgt Ion:154 Resp: 2199985


Ion Ratio Lower Upper

154 100

153 40.1 19.0 59.0

76 12.3 0.0 31.9









#47
2-Chloronaphthalene
Concen: 38.460 ng
RT: 12.886 min Scan# 1700
Delta R.T. -0.000 min
Lab File: BM047148.D
Acq: 10 Aug 2024 19:10




Instrument :
BNA_M
ClientSampleId :
ICVBM081024


Tgt Ion:162 Resp: 1712308
Ion Ratio Lower Upper
162 100
127 37.1 30.0 45.0
164 33.0 26.3 39.5



#48
2-Nitroaniline
Concen: 40.230 ng
RT: 13.116 min Scan# 1739
Delta R.T. -0.000 min
Lab File: BM047148.D
Acq: 10 Aug 2024 19:10

Tgt Ion: 65 Resp: 561121
Ion Ratio Lower Upper
65 100
92 69.3 55.1 82.7
138 107.6 85.9 128.9





#49

Acenaphthylene

Concen: 39.270 ng

RT: 13.751 min Scan# 1847

Delta R.T. -0.000 min

Lab File: BM047148.D


Acq: 10 Aug 2024 19:10

Instrument :

BNA_M



ClientSampleId :

ICVBM081024



Tgt Ion:152 Resp: 2577143

Ion	Ratio	Lower	Upper
152	100		
151	19.6	15.8	23.6
153	12.9	10.4	15.6



#50

Dimethylphthalate

Concen: 39.364 ng

RT: 13.516 min Scan# 1807




Delta R.T. -0.000 min


Lab File: BM047148.D

Acq: 10 Aug 2024 19:10

Tgt Ion:163 Resp: 2199717

Ion	Ratio	Lower	Upper
163	100		
194	4.0	3.0	4.6
164	9.8	8.0	12.0









#51
2,6-Dinitrotoluene
Concen: 40.338 ng
RT: 13.627 min Scan# 1826
Delta R.T. -0.000 min
Lab File: BM047148.D
Acq: 10 Aug 2024 19:10




Instrument :
BNA_M
ClientSampleId :
ICVBM081024


Tgt Ion:165 Resp: 526940
Ion Ratio Lower Upper
165 100
63 54.3 46.5 69.7
89 57.2 47.3 70.9



#52
Acenaphthene
Concen: 38.828 ng
RT: 14.098 min Scan# 1906
Delta R.T. -0.000 min
Lab File: BM047148.D
Acq: 10 Aug 2024 19:10

Tgt Ion:154 Resp: 1616361
Ion Ratio Lower Upper
154 100
153 112.9 90.3 135.5
152 53.5 42.9 64.3









#53
3-Nitroaniline
Concen: 40.443 ng
RT: 13.957 min Scan# 1882
Delta R.T. -0.006 min
Lab File: BM047148.D
Acq: 10 Aug 2024 19:10




Instrument :
BNA_M
ClientSampleId :
ICVBM081024


Tgt Ion:138 Resp: 529385
Ion Ratio Lower Upper
138 100
108 9.9 7.9 11.9
92 117.5 92.0 138.0



#54
2,4-Dinitrophenol
Concen: 40.314 ng
RT: 14.174 min Scan# 1919
Delta R.T. -0.001 min
Lab File: BM047148.D
Acq: 10 Aug 2024 19:10

Tgt Ion:184 Resp: 315633
Ion Ratio Lower Upper
184 100
63 60.8 52.7 79.1
154 62.7 50.4 75.6





#55

Dibenzofuran

Concen: 38.784 ng

RT: 14.439 min Scan# 1964

Delta R.T. -0.000 min

Lab File: BM047148.D

Acq: 10 Aug 2024 19:10

Instrument :

BNA_M

ClientSampleId :

ICVBM081024





Tgt Ion:168 Resp: 2550249

Ion Ratio Lower Upper

168 100

139 36.4 29.6 44.4

169 13.3 10.7 16.1



#56

4-Nitrophenol

Concen: 41.349 ng

RT: 14.292 min Scan# 1939

Delta R.T. -0.006 min

Lab File: BM047148.D

Acq: 10 Aug 2024 19:10




Tgt Ion:139 Resp: 466721


Ion Ratio Lower Upper

139 100

109 73.3 52.5 92.5


65 98.5 75.4 115.4








#57
2,4-Dinitrotoluene
Concen: 40.635 ng
RT: 14.427 min Scan# 1962
Delta R.T. -0.000 min
Lab File: BM047148.D
Acq: 10 Aug 2024 19:10

Instrument :
BNA_M
ClientSampleId :
ICVBM081024







Tgt Ion:165 Resp: 708519
Ion Ratio Lower Upper
165 100
63 51.0 40.3 60.5
89 74.1 61.0 91.6
182 2.7 1.9 2.9



#58
Fluorene
Concen: 38.859 ng
RT: 15.092 min Scan# 2075
Delta R.T. -0.000 min
Lab File: BM047148.D
Acq: 10 Aug 2024 19:10

Tgt Ion:166 Resp: 2108277
Ion Ratio Lower Upper
166 100
165 96.6 77.4 116.0
167 13.6 11.0 16.6









#59
2,3,4,6-Tetrachlorophenol
Concen: 40.787 ng
RT: 14.674 min Scan# 2004
Delta R.T. -0.000 min
Lab File: BM047148.D
Acq: 10 Aug 2024 19:10




Instrument :
BNA_M
ClientSampleId :
ICVBM081024


Tgt Ion:232 Resp: 577329
Ion Ratio Lower Upper
232 100
131 45.3 36.9 55.3
130 2.4 2.0 3.0
166 32.4 26.4 39.6



#60
Diethylphthalate
Concen: 39.122 ng
RT: 14.898 min Scan# 2042
Delta R.T. -0.000 min
Lab File: BM047148.D
Acq: 10 Aug 2024 19:10

Tgt Ion:149 Resp: 2232502
Ion Ratio Lower Upper
149 100
177 22.5 17.6 26.4
150 12.1 9.8 14.6









#61
4-Chlorophenyl-phenylether
Concen: 38.777 ng
RT: 15.098 min Scan# 2076
Delta R.T. -0.000 min
Lab File: BM047148.D
Acq: 10 Aug 2024 19:10




Instrument :
BNA_M
ClientSampleId :
ICVBM081024


Tgt Ion:204 Resp: 1001595
Ion Ratio Lower Upper
204 100
206 32.5 26.2 39.4
141 57.9 47.2 70.8



#62
4-Nitroaniline
Concen: 40.127 ng
RT: 15.133 min Scan# 2082
Delta R.T. -0.006 min
Lab File: BM047148.D
Acq: 10 Aug 2024 19:10

Tgt Ion:138 Resp: 517467
Ion Ratio Lower Upper
138 100
92 52.9 31.9 71.9
108 82.2 63.8 103.8









#63
Azobenzene
Concen: 37.854 ng
RT: 15.392 min Scan# 2126
Delta R.T. -0.000 min
Lab File: BM047148.D
Acq: 10 Aug 2024 19:10

Instrument :
BNA_M
ClientSampleId :
ICVBM081024

Tgt Ion: 77 Resp: 2048152




Ion	Ratio	Lower	Upper
77	100		
182	26.1	3.8	43.8
105	19.6	0.0	39.5
51	26.6	6.8	46.8




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 16.792 min Scan# 2364
Delta R.T. -0.000 min
Lab File: BM047148.D
Acq: 10 Aug 2024 19:10

Tgt Ion: 188 Resp: 1673175

Ion	Ratio	Lower	Upper
188	100		
94	9.0	7.0	10.6
80	10.2	8.0	12.0









#65
4,6-Dinitro-2-methylphenol
Concen: 42.500 ng
RT: 15.198 min Scan# 2093
Delta R.T. -0.000 min
Lab File: BM047148.D
Acq: 10 Aug 2024 19:10




Instrument :
BNA_M
ClientSampleId :
ICVBM081024


Tgt Ion:198 Resp: 416340
Ion Ratio Lower Upper
198 100
51 39.7 23.9 63.9
105 41.0 24.3 64.3



#66
n-Nitrosodiphenylamine
Concen: 39.293 ng
RT: 15.321 min Scan# 2114
Delta R.T. -0.000 min
Lab File: BM047148.D
Acq: 10 Aug 2024 19:10

Tgt Ion:169 Resp: 1790716
Ion Ratio Lower Upper
169 100
168 67.0 53.8 80.8
167 36.3 29.3 43.9









#67
4-Bromophenyl-phenylether
Concen: 40.155 ng
RT: 15.998 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM047148.D
Acq: 10 Aug 2024 19:10




Instrument :
BNA_M
ClientSampleId :
ICVBM081024


Tgt Ion:248 Resp: 647415
Ion Ratio Lower Upper
248 100
250 95.4 77.8 116.8
141 68.4 57.7 86.5



#68
Hexachlorobenzene
Concen: 40.130 ng
RT: 16.098 min Scan# 2246
Delta R.T. -0.000 min
Lab File: BM047148.D
Acq: 10 Aug 2024 19:10

Tgt Ion:284 Resp: 774906
Ion Ratio Lower Upper
284 100
142 34.5 29.4 44.0
249 31.1 25.1 37.7









#69
Atrazine
Concen: 41.312 ng
RT: 16.292 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM047148.D
Acq: 10 Aug 2024 19:10




Instrument :
BNA_M
ClientSampleId :
ICVBM081024


Tgt Ion:200 Resp: 566536
Ion Ratio Lower Upper
200 100
173 25.8 6.1 46.1
215 46.7 26.8 66.8



#70
Pentachlorophenol
Concen: 41.470 ng
RT: 16.451 min Scan# 2306
Delta R.T. -0.000 min
Lab File: BM047148.D
Acq: 10 Aug 2024 19:10

Tgt Ion:266 Resp: 561329
Ion Ratio Lower Upper
266 100
268 64.6 49.8 74.6
264 63.0 49.6 74.4









#71
Phenanthrene
Concen: 38.967 ng
RT: 16.833 min Scan# 21
Delta R.T. -0.006 min
Lab File: BM047148.D
Acq: 10 Aug 2024 19:10




Instrument :
BNA_M
ClientSampleId :
ICVBM081024


Tgt Ion:178 Resp: 3272056
Ion Ratio Lower Upper
178 100
176 19.1 15.3 22.9
179 15.4 12.4 18.6



#72
Anthracene
Concen: 39.620 ng
RT: 16.927 min Scan# 2387
Delta R.T. -0.000 min
Lab File: BM047148.D
Acq: 10 Aug 2024 19:10

Tgt Ion:178 Resp: 3237131
Ion Ratio Lower Upper
178 100
176 18.6 14.8 22.2
179 15.3 12.4 18.6









#73
Carbazole
Concen: 39.108 ng
RT: 17.209 min Scan# 2435
Delta R.T. -0.000 min
Lab File: BM047148.D
Acq: 10 Aug 2024 19:10




Instrument :
BNA_M
ClientSampleId :
ICVBM081024


Tgt Ion:167 Resp: 3262489
Ion Ratio Lower Upper
167 100
166 21.2 17.1 25.7
139 12.0 10.0 15.0



#74
Di-n-butylphthalate
Concen: 39.675 ng
RT: 17.792 min Scan# 2534
Delta R.T. -0.000 min
Lab File: BM047148.D
Acq: 10 Aug 2024 19:10

Tgt Ion:149 Resp: 3851216
Ion Ratio Lower Upper
149 100
150 9.3 7.5 11.3
104 5.0 4.1 6.1









#75
Fluoranthene
Concen: 38.982 ng
RT: 18.868 min Scan# 2717
Delta R.T. -0.000 min
Lab File: BM047148.D
Acq: 10 Aug 2024 19:10




Instrument :
BNA_M
ClientSampleId :
ICVBM081024


Tgt Ion:202 Resp: 3981412
Ion Ratio Lower Upper
202 100
101 9.6 0.0 30.2
203 17.5 0.0 37.6



#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.027 min Scan# 3084
Delta R.T. -0.006 min
Lab File: BM047148.D
Acq: 10 Aug 2024 19:10

Tgt Ion:240 Resp: 1477990
Ion Ratio Lower Upper
240 100
120 10.0 8.2 12.2
236 25.8 20.6 30.8









#77
Benzidine
Concen: 59.321 ng
RT: 19.074 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM047148.D
Acq: 10 Aug 2024 19:10




Instrument :
BNA_M
ClientSampleId :
ICVBM081024


Tgt Ion:184 Resp: 1486859
Ion Ratio Lower Upper
184 100
185 14.8 11.8 17.6
183 11.2 9.0 13.6



#78
Pyrene
Concen: 39.331 ng
RT: 19.233 min Scan# 2779
Delta R.T. -0.006 min
Lab File: BM047148.D
Acq: 10 Aug 2024 19:10

Tgt Ion:202 Resp: 4150742
Ion Ratio Lower Upper
202 100
200 20.9 17.2 25.8
203 17.6 14.2 21.2









#79
Terphenyl-d14
Concen: 79.687 ng
RT: 19.456 min Scan# 2817
Delta R.T. -0.006 min
Lab File: BM047148.D
Acq: 10 Aug 2024 19:10




Instrument :
BNA_M
ClientSampleId :
ICVBM081024


Tgt Ion:244 Resp: 5496459
Ion Ratio Lower Upper
244 100
212 7.7 6.2 9.4
122 9.7 7.1 10.7



#80
Butylbenzylphthalate
Concen: 41.291 ng
RT: 20.174 min Scan# 2939
Delta R.T. -0.000 min
Lab File: BM047148.D
Acq: 10 Aug 2024 19:10

Tgt Ion:149 Resp: 1778335
Ion Ratio Lower Upper
149 100
91 68.6 57.0 85.4
206 20.5 16.4 24.6









#81
Benzo(a)anthracene
Concen: 39.088 ng
RT: 21.015 min Scan# 3082
Delta R.T. -0.000 min
Lab File: BM047148.D
Acq: 10 Aug 2024 19:10




Instrument :
BNA_M
ClientSampleId :
ICVBM081024


Tgt Ion:228 Resp: 3835235
Ion Ratio Lower Upper
228 100
226 27.4 21.8 32.6
229 19.6 15.8 23.8



#82
3,3'-Dichlorobenzidine
Concen: 43.078 ng
RT: 20.956 min Scan# 3072
Delta R.T. -0.000 min
Lab File: BM047148.D
Acq: 10 Aug 2024 19:10

Tgt Ion:252 Resp: 1322453
Ion Ratio Lower Upper
252 100
254 65.5 51.0 76.6
126 9.7 8.5 12.7









#83
Chrysene
Concen: 38.651 ng
RT: 21.068 min Scan# 3092
Delta R.T. -0.006 min
Lab File: BM047148.D
Acq: 10 Aug 2024 19:10




Instrument :
BNA_M
ClientSampleId :
ICVBM081024


Tgt Ion:228 Resp: 3598062
Ion Ratio Lower Upper
228 100
226 30.3 24.4 36.6
229 20.1 15.7 23.5



#84
Bis(2-ethylhexyl)phthalate
Concen: 40.349 ng
RT: 20.974 min Scan# 3075
Delta R.T. -0.000 min
Lab File: BM047148.D
Acq: 10 Aug 2024 19:10

Tgt Ion:149 Resp: 2464399
Ion Ratio Lower Upper
149 100
167 27.8 22.2 33.2
279 4.1 3.0 4.4









#85
Di-n-octyl phthalate
Concen: 41.304 ng
RT: 22.003 min Scan# 31
Delta R.T. -0.006 min
Lab File: BM047148.D
Acq: 10 Aug 2024 19:10




Instrument :
BNA_M
ClientSampleId :
ICVBM081024


Tgt Ion:149 Resp: 4287924
Ion Ratio Lower Upper
149 100
167 1.5 1.3 1.9
43 8.3 6.9 10.3



#86
Perylene-d12
Concen: 20.000 ng
RT: 23.727 min Scan# 3543
Delta R.T. -0.000 min
Lab File: BM047148.D
Acq: 10 Aug 2024 19:10

Tgt Ion:264 Resp: 1812426
Ion Ratio Lower Upper
264 100
260 23.3 18.5 27.7
265 21.8 17.3 25.9









#87
Indeno(1,2,3-cd)pyrene
Concen: 40.535 ng
RT: 26.726 min Scan# 4053
Delta R.T. -0.000 min
Lab File: BM047148.D
Acq: 10 Aug 2024 19:10




Instrument :
BNA_M
ClientSampleId :
ICVBM081024


Tgt Ion:276 Resp: 4748737
Ion Ratio Lower Upper
276 100
138 23.6 19.4 29.0
277 25.0 20.1 30.1



#88
Benzo(b)fluoranthene
Concen: 38.443 ng
RT: 22.886 min Scan# 3400
Delta R.T. -0.000 min
Lab File: BM047148.D
Acq: 10 Aug 2024 19:10

Tgt Ion:252 Resp: 4137733
Ion Ratio Lower Upper
252 100
253 21.9 17.8 26.6
125 9.1 7.4 11.0





#89

Benzo(k)fluoranthene

Concen: 39.350 ng

RT: 22.944 min Scan# 3410

Delta R.T. -0.000 min

Lab File: BM047148.D

Acq: 10 Aug 2024 19:10

Instrument :

BNA_M

ClientSampleId :

ICVBM081024





Tgt Ion:252 Resp: 4000536

Ion Ratio Lower Upper

252 100

253 21.7 17.7 26.5

125 8.9 7.4 11.2



#90

Benzo(a)pyrene

Concen: 39.764 ng

RT: 23.603 min Scan# 3522

Delta R.T. -0.006 min

Lab File: BM047148.D

Acq: 10 Aug 2024 19:10




Tgt Ion:252 Resp: 3707577


Ion Ratio Lower Upper

252 100

253 21.8 17.1 25.7

125 10.1 8.2 12.2









#91
Dibenzo(a,h)anthracene
Concen: 40.122 ng
RT: 26.774 min Scan# 4061
Delta R.T. -0.006 min
Lab File: BM047148.D
Acq: 10 Aug 2024 19:10




Instrument :
BNA_M
ClientSampleId :
ICVBM081024

Tgt Ion:278 Resp: 3803894
Ion Ratio Lower Upper
278 100
139 14.8 12.0 18.0
279 23.7 19.2 28.8



#92
Benzo(g,h,i)perylene
Concen: 40.615 ng
RT: 27.656 min Scan# 4211
Delta R.T. -0.006 min
Lab File: BM047148.D
Acq: 10 Aug 2024 19:10

Tgt Ion:276 Resp: 3997872
Ion Ratio Lower Upper
276 100
277 23.6 18.8 28.2
138 20.2 16.2 24.4



Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081024\
 Data File : BM047148.D
 Acq On : 10 Aug 2024 19:10
 Operator : RC/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 ICVBM081024

Quant Time: Aug 11 23:52:13 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Aug 11 23:47:58 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	86	0.00
2	1,4-Dioxane	0.464	0.427	8.0	84	0.00
3	Pyridine	1.360	1.291	5.1	86	0.00
4	n-Nitrosodimethylamine	0.594	0.554	6.7	84	0.00
5 S	2-Fluorophenol	1.119	1.052	6.0	85	0.00
6	Aniline	1.458	1.425	2.3	86	0.00
7 S	Phenol-d6	1.542	1.477	4.2	86	0.00
8	2-Chlorophenol	1.221	1.188	2.7	87	0.00
9	Benzaldehyde	0.904	0.804	11.1	89	0.00
10 C	Phenol	1.539	1.445	6.1	86	0.00
11	bis(2-Chloroethyl)ether	1.318	1.251	5.1	84	0.00
12	1,3-Dichlorobenzene	1.451	1.411	2.8	87	0.00
13 C	1,4-Dichlorobenzene	1.469	1.426	2.9	86	0.00
14	1,2-Dichlorobenzene	1.386	1.342	3.2	88	0.00
15	Benzyl Alcohol	1.098	1.086	1.1	85	0.00
16	2,2'-oxybis(1-Chloropropane	1.691	1.598	5.5	83	0.00
17	2-Methylphenol	1.027	1.005	2.1	87	0.00
18	Hexachloroethane	0.569	0.547	3.9	86	0.00
19 P	n-Nitroso-di-n-propylamine	1.039	0.983	5.4	86	0.00
20	3+4-Methylphenols	1.435	1.425	0.7	87	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	87	0.00
22	Acetophenone	0.483	0.460	4.8	86	0.00
23 S	Nitrobenzene-d5	0.397	0.383	3.5	85	0.00
24	Nitrobenzene	0.400	0.390	2.5	86	0.00
25	Isophorone	0.692	0.684	1.2	86	0.00
26 C	2-Nitrophenol	0.177	0.183	-3.4	87	0.00
27	2,4-Dimethylphenol	0.207	0.205	1.0	85	0.00
28	bis(2-Chloroethoxy)methane	0.435	0.425	2.3	85	0.00
29 C	2,4-Dichlorophenol	0.308	0.315	-2.3	88	0.00
30	1,2,4-Trichlorobenzene	0.348	0.351	-0.9	88	0.00
31	Naphthalene	0.996	0.961	3.5	86	0.00
32	Benzoic acid	0.241	0.260	-7.9	90	-0.01
33	4-Chloroaniline	0.383	0.386	-0.8	87	0.00
34 C	Hexachlorobutadiene	0.204	0.205	-0.5	88	0.00
35	Caprolactam	0.107	0.108	-0.9	86	0.00
36 C	4-Chloro-3-methylphenol	0.328	0.326	0.6	87	0.00
37	2-Methylnaphthalene	0.709	0.691	2.5	87	0.00
38	1-Methylnaphthalene	0.701	0.681	2.9	87	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	87	0.00
40	1,2,4,5-Tetrachlorobenzene	0.546	0.553	-1.3	89	0.00
41 P	Hexachlorocyclopentadiene	0.190	0.195	-2.6	85	0.00
42 S	2,4,6-Tribromophenol	0.232	0.237	-2.2	89	0.00
43 C	2,4,6-Trichlorophenol	0.388	0.395	-1.8	87	0.00
44	2,4,5-Trichlorophenol	0.430	0.442	-2.8	88	0.00
45 S	2-Fluorobiphenyl	1.297	1.276	1.6	88	0.00
46	1,1'-Biphenyl	1.424	1.375	3.4	87	0.00
47	2-Chloronaphthalene	1.113	1.070	3.9	87	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081024\
 Data File : BM047148.D
 Acq On : 10 Aug 2024 19:10
 Operator : RC/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 ICVBM081024

Quant Time: Aug 11 23:52:13 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Aug 11 23:47:58 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.349	0.351	-0.6	86	0.00
49	Acenaphthylene	1.640	1.610	1.8	88	0.00
50	Dimethylphthalate	1.397	1.375	1.6	87	0.00
51	2,6-Dinitrotoluene	0.327	0.329	-0.6	88	0.00
52 C	Acenaphthene	1.041	1.010	3.0	88	0.00
53	3-Nitroaniline	0.327	0.331	-1.2	87	0.00
54 P	2,4-Dinitrophenol	0.196	0.197	-0.5	85	0.00
55	Dibenzofuran	1.644	1.594	3.0	88	0.00
56 P	4-Nitrophenol	0.282	0.292	-3.5	87	0.00
57	2,4-Dinitrotoluene	0.436	0.443	-1.6	88	0.00
58	Fluorene	1.356	1.317	2.9	88	0.00
59	2,3,4,6-Tetrachlorophenol	0.354	0.361	-2.0	89	0.00
60	Diethylphthalate	1.426	1.395	2.2	87	0.00
61	4-Chlorophenyl-phenylether	0.646	0.626	3.1	88	0.00
62	4-Nitroaniline	0.322	0.323	-0.3	89	0.00
63	Azobenzene	1.352	1.280	5.3	85	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	88	0.00
65	4,6-Dinitro-2-methylphenol	0.117	0.124	-6.0	87	0.00
66 c	n-Nitrosodiphenylamine	0.545	0.535	1.8	88	0.00
67	4-Bromophenyl-phenylether	0.193	0.193	0.0	89	0.00
68	Hexachlorobenzene	0.231	0.232	-0.4	89	0.00
69	Atrazine	0.164	0.169	-3.0	87	0.00
70 C	Pentachlorophenol	0.162	0.168	-3.7	88	0.00
71	Phenanthrene	1.004	0.978	2.6	88	0.00
72	Anthracene	0.977	0.967	1.0	89	0.00
73	Carbazole	0.997	0.975	2.2	88	0.00
74	Di-n-butylphthalate	1.160	1.151	0.8	87	0.00
75 C	Fluoranthene	1.221	1.190	2.5	87	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	88	0.00
77	Benzydine	0.339	0.503	-48.4#	91	0.00
78	Pyrene	1.428	1.404	1.7	87	0.00
79 S	Terphenyl-d14	0.933	0.930	0.3	88	0.00
80	Butylbenzylphthalate	0.583	0.602	-3.3	87	0.00
81	Benzo(a)anthracene	1.328	1.297	2.3	88	0.00
82	3,3'-Dichlorobenzidine	0.415	0.447	-7.7	89	0.00
83	Chrysene	1.260	1.217	3.4	88	0.00
84	Bis(2-ethylhexyl)phthalate	0.826	0.834	-1.0	87	0.00
85 c	Di-n-octyl phthalate	1.405	1.451	-3.3	88	0.00
86 I	Perylene-d12	1.000	1.000	0.0	90	0.00
87	Indeno(1,2,3-cd)pyrene	1.293	1.310	-1.3	92	0.00
88	Benzo(b)fluoranthene	1.188	1.141	4.0	88	0.00
89	Benzo(k)fluoranthene	1.122	1.104	1.6	90	0.00
90 C	Benzo(a)pyrene	1.029	1.023	0.6	90	0.00
91	Dibenzo(a,h)anthracene	1.046	1.049	-0.3	92	0.00
92	Benzo(g,h,i)perylene	1.086	1.103	-1.6	92	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081024\
Data File : BM047148.D
Acq On : 10 Aug 2024 19:10
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
ICVBM081024

Quant Time: Aug 11 23:52:13 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sun Aug 11 23:47:58 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081024\
 Data File : BM047148.D
 Acq On : 10 Aug 2024 19:10
 Operator : RC/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 ICVBM081024

Quant Time: Aug 11 23:52:13 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Aug 11 23:47:58 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	86	0.00
2	1,4-Dioxane	40.000	36.823	7.9	84	0.00
3	Pyridine	40.000	37.969	5.1	86	0.00
4	n-Nitrosodimethylamine	40.000	37.322	6.7	84	0.00
5 S	2-Fluorophenol	80.000	75.195	6.0	85	0.00
6	Aniline	40.000	39.104	2.2	86	0.00
7 S	Phenol-d6	80.000	76.636	4.2	86	0.00
8	2-Chlorophenol	40.000	38.916	2.7	87	0.00
9	Benzaldehyde	40.000	39.535	1.2	89	0.00
10 C	Phenol	40.000	37.549	6.1	86	0.00
11	bis(2-Chloroethyl)ether	40.000	37.972	5.1	84	0.00
12	1,3-Dichlorobenzene	40.000	38.900	2.8	87	0.00
13 C	1,4-Dichlorobenzene	40.000	38.844	2.9	86	0.00
14	1,2-Dichlorobenzene	40.000	38.744	3.1	88	0.00
15	Benzyl Alcohol	40.000	39.581	1.0	85	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	37.800	5.5	83	0.00
17	2-Methylphenol	40.000	39.168	2.1	87	0.00
18	Hexachloroethane	40.000	38.447	3.9	86	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	37.864	5.3	86	0.00
20	3+4-Methylphenols	40.000	39.723	0.7	87	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	87	0.00
22	Acetophenone	40.000	38.108	4.7	86	0.00
23 S	Nitrobenzene-d5	80.000	77.088	3.6	85	0.00
24	Nitrobenzene	40.000	38.978	2.6	86	0.00
25	Isophorone	40.000	39.554	1.1	86	0.00
26 C	2-Nitrophenol	40.000	41.274	-3.2	87	0.00
27	2,4-Dimethylphenol	40.000	39.547	1.1	85	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.022	2.4	85	0.00
29 C	2,4-Dichlorophenol	40.000	40.983	-2.5	88	0.00
30	1,2,4-Trichlorobenzene	40.000	40.261	-0.7	88	0.00
31	Naphthalene	40.000	38.564	3.6	86	0.00
32	Benzoic acid	40.000	43.072	-7.7	90	-0.01
33	4-Chloroaniline	40.000	40.265	-0.7	87	0.00
34 C	Hexachlorobutadiene	40.000	40.254	-0.6	88	0.00
35	Caprolactam	40.000	40.543	-1.4	86	0.00
36 C	4-Chloro-3-methylphenol	40.000	39.811	0.5	87	0.00
37	2-Methylnaphthalene	40.000	38.968	2.6	87	0.00
38	1-Methylnaphthalene	40.000	38.877	2.8	87	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	87	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	40.555	-1.4	89	0.00
41 P	Hexachlorocyclopentadiene	40.000	41.042	-2.6	85	0.00
42 S	2,4,6-Tribromophenol	80.000	81.608	-2.0	89	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.777	-1.9	87	0.00
44	2,4,5-Trichlorophenol	40.000	41.066	-2.7	88	0.00
45 S	2-Fluorobiphenyl	80.000	78.722	1.6	88	0.00
46	1,1'-Biphenyl	40.000	38.622	3.4	87	0.00
47	2-Chloronaphthalene	40.000	38.460	3.8	87	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081024\
 Data File : BM047148.D
 Acq On : 10 Aug 2024 19:10
 Operator : RC/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 ICVBM081024

Quant Time: Aug 11 23:52:13 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Aug 11 23:47:58 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	40.230	-0.6	86	0.00
49	Acenaphthylene	40.000	39.270	1.8	88	0.00
50	Dimethylphthalate	40.000	39.364	1.6	87	0.00
51	2,6-Dinitrotoluene	40.000	40.338	-0.8	88	0.00
52 C	Acenaphthene	40.000	38.828	2.9	88	0.00
53	3-Nitroaniline	40.000	40.443	-1.1	87	0.00
54 P	2,4-Dinitrophenol	40.000	40.314	-0.8	85	0.00
55	Dibenzofuran	40.000	38.784	3.0	88	0.00
56 P	4-Nitrophenol	40.000	41.349	-3.4	87	0.00
57	2,4-Dinitrotoluene	40.000	40.635	-1.6	88	0.00
58	Fluorene	40.000	38.859	2.9	88	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	40.787	-2.0	89	0.00
60	Diethylphthalate	40.000	39.122	2.2	87	0.00
61	4-Chlorophenyl-phenylether	40.000	38.777	3.1	88	0.00
62	4-Nitroaniline	40.000	40.127	-0.3	89	0.00
63	Azobenzene	40.000	37.854	5.4	85	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	88	0.00
65	4,6-Dinitro-2-methylphenol	40.000	42.500	-6.3	87	0.00
66 c	n-Nitrosodiphenylamine	40.000	39.293	1.8	88	0.00
67	4-Bromophenyl-phenylether	40.000	40.155	-0.4	89	0.00
68	Hexachlorobenzene	40.000	40.130	-0.3	89	0.00
69	Atrazine	40.000	41.312	-3.3	87	0.00
70 C	Pentachlorophenol	40.000	41.470	-3.7	88	0.00
71	Phenanthrene	40.000	38.967	2.6	88	0.00
72	Anthracene	40.000	39.620	1.0	89	0.00
73	Carbazole	40.000	39.108	2.2	88	0.00
74	Di-n-butylphthalate	40.000	39.675	0.8	87	0.00
75 C	Fluoranthene	40.000	38.982	2.5	87	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	88	0.00
77	Benzydine	40.000	59.321	-48.3#	91	0.00
78	Pyrene	40.000	39.331	1.7	87	0.00
79 S	Terphenyl-d14	80.000	79.687	0.4	88	0.00
80	Butylbenzylphthalate	40.000	41.291	-3.2	87	0.00
81	Benzo(a)anthracene	40.000	39.088	2.3	88	0.00
82	3,3'-Dichlorobenzidine	40.000	43.078	-7.7	89	0.00
83	Chrysene	40.000	38.651	3.4	88	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	40.349	-0.9	87	0.00
85 c	Di-n-octyl phthalate	40.000	41.304	-3.3	88	0.00
86 I	Perylene-d12	20.000	20.000	0.0	90	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	40.535	-1.3	92	0.00
88	Benzo(b)fluoranthene	40.000	38.443	3.9	88	0.00
89	Benzo(k)fluoranthene	40.000	39.350	1.6	90	0.00
90 C	Benzo(a)pyrene	40.000	39.764	0.6	90	0.00
91	Dibenzo(a,h)anthracene	40.000	40.122	-0.3	92	0.00
92	Benzo(g,h,i)perylene	40.000	40.615	-1.5	92	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081024\
Data File : BM047148.D
Acq On : 10 Aug 2024 19:10
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
ICVBM081024

Quant Time: Aug 11 23:52:13 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sun Aug 11 23:47:58 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range SPCC's out = 0 CCC's out = 0



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: JAC005

Lab Code: CHEM Case No.: P3609 SAS No.: P3609 SDG No.: P3609

Instrument ID: BNA_M Calibration Date/Time: 08/19/2024 09:42

Lab File ID: BM047245.D Init. Calib. Date(s): 08/10/2024 08/10/2024

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 13:12 17:51

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
Pyridine	1.360	1.126		-17.2	
2-Fluorophenol	1.119	0.985		-12.0	
Benzaldehyde	0.904	0.709		-21.6	
Phenol-d6	1.542	1.370		-11.2	
2-Methylphenol	1.027	0.951		-7.4	
3+4-Methylphenols	1.435	1.329		-7.4	
Nitrobenzene-d5	0.397	0.345		-13.1	
Hexachloroethane	0.569	0.514		-9.7	
Nitrobenzene	0.400	0.352		-12.0	
Naphthalene	0.996	0.908		-8.8	
Hexachlorobutadiene	0.204	0.213		4.4	20.0
2-Methylnaphthalene	0.709	0.667		-5.9	
2,4,6-Trichlorophenol	0.388	0.387		-0.3	20.0
2-Fluorobiphenyl	1.297	1.203		-7.2	
2,4,5-Trichlorophenol	0.430	0.431		0.2	
Acenaphthylene	1.640	1.482		-9.6	
Acenaphthene	1.041	0.938		-9.9	20.0
Dibenzofuran	1.644	1.483		-9.8	
2,4-Dinitrotoluene	0.436	0.410		-6.0	
Fluorene	1.356	1.232		-9.1	
2,4,6-Tribromophenol	0.232	0.233		0.4	
Hexachlorobenzene	0.231	0.217		-6.1	
Pentachlorophenol	0.162	0.158		-2.5	20.0
Phenanthrene	1.004	0.896		-10.8	
Anthracene	0.977	0.888		-9.1	
Carbazole	0.997	0.892		-10.5	
Di-n-butylphthalate	1.160	1.070		-7.8	
Fluoranthene	1.221	1.121		-8.2	20.0
Pyrene	1.428	1.389		-2.7	
Terphenyl-d14	0.933	0.922		-1.2	
Benzo (a) anthracene	1.328	1.210		-8.9	
Chrysene	1.260	1.127		-10.6	
Bis (2-ethylhexyl) phthalate	0.826	0.796		-3.6	
Benzo (b) fluoranthene	1.188	1.040		-12.5	
Benzo (k) fluoranthene	1.122	0.981		-12.6	
Benzo (a) pyrene	1.029	0.906		-12.0	20.0
Indeno (1,2,3-cd) pyrene	1.293	1.055		-18.4	
Dibenzo (a,h) anthracene	1.046	0.868		-17.0	
Benzo (g,h,i) perylene	1.086	0.869		-20.0	



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: JAC005
Lab Code: CHEM Case No.: P3609 SAS No.: P3609 SDG No.: P3609
Instrument ID: BNA_M Calibration Date/Time: 08/19/2024 09:42
Lab File ID: BM047245.D Init. Calib. Date(s): 08/10/2024 08/10/2024
EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 13:12 17:51
GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	<u>RRF</u>	RRF040	MIN RRF	%D	MAX%D
1,4-Dioxane	0.464	0.372		-19.8	20.0
1-Methylnaphthalene	0.701	0.654		-6.7	

All other compounds must meet a minimum RRF of 0.010.

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081924\
 Data File : BM047245.D
 Acq On : 19 Aug 2024 09:42
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040

Manual Integrations**APPROVED**

Reviewed By :Jagrut Upadhyay 08/20/2024
 Supervised By :mohammad ahmed 08/21/2024

Quant Time: Aug 19 11:18:18 2024

Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081024.M

Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

QLast Update : Sun Aug 11 23:47:58 2024

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.369	152	274830	20.000	ng	0.00
21) Naphthalene-d8	10.122	136	1117659	20.000	ng	-0.01
39) Acenaphthene-d10	14.027	164	774502	20.000	ng	0.00
64) Phenanthrene-d10	16.792	188	1666055	20.000	ng	0.00
76) Chrysene-d12	21.039	240	1393632	20.000	ng	0.00
86) Perylene-d12	23.744	264	1655413	20.000	ng	0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.016	112	1082412	70.374	ng	0.00
7) Phenol-d6	6.581	99	1506202	71.069	ng	0.00
23) Nitrobenzene-d5	8.516	82	1544533	69.585	ng	0.00
42) 2,4,6-Tribromophenol	15.539	330	722218	80.421	ng	0.00
45) 2-Fluorobiphenyl	12.633	172	3725750	74.203	ng	-0.01
79) Terphenyl-d14	19.462	244	5138818	79.011	ng	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.028	88	204283	32.055	ng	96
3) Pyridine	3.405	79	618664	33.106	ng	97
4) n-Nitrosodimethylamine	3.322	42	249562	30.567	ng	96
6) Aniline	6.716	93	726743	36.280	ng	98
8) 2-Chlorophenol	6.946	128	613579	36.576	ng	97
9) Benzaldehyde	6.528	77	389532	32.528	ng	100
10) Phenol	6.610	94	743089	35.139	ng	98
11) bis(2-Chloroethyl)ether	6.816	93	624007	34.456	ng	98
12) 1,3-Dichlorobenzene	7.257	146	739956	37.119	ng	98
13) 1,4-Dichlorobenzene	7.399	146	750115	37.165	ng	99
14) 1,2-Dichlorobenzene	7.710	146	709235	37.243	ng	99
15) Benzyl Alcohol	7.616	79	595977m	39.512	ng	
16) 2,2'-oxybis(1-Chloropr...	7.904	45	770805m	33.169	ng	
17) 2-Methylphenol	7.828	107	522725	37.051	ng	98
18) Hexachloroethane	8.416	117	282269	36.104	ng	94
19) n-Nitroso-di-n-propyla...	8.175	70	490524	34.361	ng	98
20) 3+4-Methylphenols	8.157	107	730719	37.048	ng	99
22) Acetophenone	8.187	105	958453	35.527	ng	99
24) Nitrobenzene	8.557	77	785862	35.161	ng	96
25) Isophorone	9.081	82	1420787	36.763	ng	99
26) 2-Nitrophenol	9.257	139	391755	39.578	ng	95
27) 2,4-Dimethylphenol	9.340	122	445047	38.384	ng	99
28) bis(2-Chloroethoxy)met...	9.563	93	871685	35.845	ng	99
29) 2,4-Dichlorophenol	9.792	162	692233	40.237	ng	98
30) 1,2,4-Trichlorobenzene	9.992	180	768282	39.461	ng	99
31) Naphthalene	10.175	128	2030688	36.472	ng	100
32) Benzoic acid	9.534	122	560880	41.637	ng	97
33) 4-Chloroaniline	10.298	127	808887	37.779	ng	99
34) Hexachlorobutadiene	10.457	225	475986	41.770	ng	99
35) Caprolactam	11.110	113	235829	39.472	ng	92
36) 4-Chloro-3-methylphenol	11.451	107	706688	38.570	ng	99
37) 2-Methylnaphthalene	11.798	142	1491295	37.618	ng	99
38) 1-Methylnaphthalene	12.028	142	1462706	37.334	ng	98
40) 1,2,4,5-Tetrachloroben...	12.186	216	841763	39.847	ng	98
41) Hexachlorocyclopentadiene	12.157	237	239288	32.439	ng	95
43) 2,4,6-Trichlorophenol	12.445	196	599628	39.935	ng	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081924\
 Data File : BM047245.D
 Acq On : 19 Aug 2024 09:42
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040

Manual Integrations**APPROVED**

Reviewed By :Jagrut Upadhyay 08/20/2024
 Supervised By :mohammad ahmed 08/21/2024

Quant Time: Aug 19 11:18:18 2024

Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081024.M

Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

QLast Update : Sun Aug 11 23:47:58 2024

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	12.522	196	668381	40.092	ng	97
46) 1,1'-Biphenyl	12.851	154	1967045	35.675	ng	99
47) 2-Chloronaphthalene	12.886	162	1538296	35.695	ng	98
48) 2-Nitroaniline	13.110	65	470455	34.845	ng	96
49) Acenaphthylene	13.745	152	2296157	36.146	ng	100
50) Dimethylphthalate	13.510	163	1999459	36.964	ng	100
51) 2,6-Dinitrotoluene	13.628	165	481854	38.107	ng	88
52) Acenaphthene	14.092	154	1452604	36.049	ng	98
53) 3-Nitroaniline	13.957	138	473408	37.363	ng	97
54) 2,4-Dinitrophenol	14.180	184	274199	36.181	ng	92
55) Dibenzofuran	14.433	168	2296458	36.080	ng	98
56) 4-Nitrophenol	14.310	139	400919	36.695	ng	97
57) 2,4-Dinitrotoluene	14.433	165	635042	37.626	ng	97
58) Fluorene	15.092	166	1908826	36.346	ng	100
59) 2,3,4,6-Tetrachlorophenol	14.680	232	550495	40.178	ng	93
60) Diethylphthalate	14.892	149	2006026	36.316	ng	98
61) 4-Chlorophenyl-phenyle...	15.098	204	943737	37.746	ng	96
62) 4-Nitroaniline	15.133	138	462456	37.048	ng	96
63) Azobenzene	15.392	77	1742713	33.274	ng	95
65) 4,6-Dinitro-2-methylph...	15.204	198	382355	39.198	ng	90
66) n-Nitrosodiphenylamine	15.316	169	1636928	36.072	ng	99
67) 4-Bromophenyl-phenylether	15.992	248	621686	38.724	ng	98
68) Hexachlorobenzene	16.104	284	723291	37.617	ng	95
69) Atrazine	16.292	200	548707	40.182	ng	99
70) Pentachlorophenol	16.457	266	524917	38.946	ng	98
71) Phenanthrene	16.833	178	2985541	35.707	ng	100
72) Anthracene	16.927	178	2959126	36.372	ng	100
73) Carbazole	17.210	167	2971451	35.772	ng	99
74) Di-n-butylphthalate	17.792	149	3566712	36.901	ng	100
75) Fluoranthene	18.874	202	3736572	36.741	ng	99
77) Benzidine	19.080	184	1751945	74.129	ng	99
78) Pyrene	19.239	202	3870284	38.894	ng	99
80) Butylbenzylphthalate	20.174	149	1617756	39.837	ng	97
81) Benzo(a)anthracene	21.021	228	3372910	36.457	ng	99
82) 3,3'-Dichlorobenzidine	20.962	252	1158555	40.023	ng	99
83) Chrysene	21.080	228	3142558	35.801	ng	100
84) Bis(2-ethylhexyl)phtha...	20.974	149	2219045	38.531	ng	99
85) Di-n-octyl phthalate	22.009	149	3946297	40.314	ng	98
87) Indeno(1,2,3-cd)pyrene	26.756	276	3492007	32.635	ng	98
88) Benzo(b)fluoranthene	22.898	252	3442805	35.020	ng	100
89) Benzo(k)fluoranthene	22.956	252	3246337	34.961	ng	99
90) Benzo(a)pyrene	23.621	252	2998995	35.215	ng	98
91) Dibenzo(a,h)anthracene	26.803	278	2874097	33.190	ng	99
92) Benzo(g,h,i)perylene	27.685	276	2876281	31.992	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081924\
Data File : BM047245.D
Acq On : 19 Aug 2024 09:42
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040

Manual Integrations**APPROVED**

Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024


Quant Time: Aug 19 11:18:18 2024

Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081024.M

Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

QLast Update : Sun Aug 11 23:47:58 2024

Response via : Initial Calibration




Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081924\
Data File : BM047245.D
Acq On : 19 Aug 2024 09:42
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1


Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024


Quant Time: Aug 19 11:18:18 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sun Aug 11 23:47:58 2024
Response via : Initial Calibration





#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.369 min Scan# 7
Delta R.T. -0.006 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion: 152 Resp: 274830
Ion Ratio Lower Upper
152 100
150 152.7 120.5 180.7
115 52.7 45.5 68.3




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#2
1,4-Dioxane
Concen: 32.055 ng
RT: 3.028 min Scan# 24
Delta R.T. -0.006 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42


Tgt Ion: 88 Resp: 204283
Ion Ratio Lower Upper
88 100
58 69.1 58.2 87.4
43 27.2 22.5 33.7





#3
Pyridine
Concen: 33.106 ng
RT: 3.405 min Scan# 89
Delta R.T. -0.006 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion: 79 Resp: 618664
Ion Ratio Lower Upper
79 100
52 61.5 51.2 76.8
51 31.9 26.1 39.1




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#4
n-Nitrosodimethylamine
Concen: 30.567 ng
RT: 3.322 min Scan# 74
Delta R.T. -0.006 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42


Tgt Ion: 42 Resp: 249562
Ion Ratio Lower Upper
42 100
74 145.5 112.9 169.3
44 7.3 5.3 7.9





#5
2-Fluorophenol
Concen: 70.374 ng
RT: 5.016 min Scan# 362
Delta R.T. -0.006 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:112 Resp: 108241
Ion Ratio Lower Upper
112 100
64 57.0 46.7 70.1
63 30.0 24.8 37.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#6
Aniline
Concen: 36.280 ng
RT: 6.716 min Scan# 651
Delta R.T. -0.012 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42


Tgt Ion: 93 Resp: 726743
Ion Ratio Lower Upper
93 100
66 37.6 31.4 47.2
65 19.3 15.6 23.4





#7
Phenol-d6
Concen: 71.069 ng
RT: 6.581 min Scan# 61
Delta R.T. -0.006 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040




Tgt Ion: 99 Resp: 150620
Ion Ratio Lower Upper
99 100
42 16.9 14.1 21.1
71 35.0 28.5 42.7

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024




#8
2-Chlorophenol
Concen: 36.576 ng
RT: 6.946 min Scan# 690
Delta R.T. -0.006 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42




Tgt Ion:128 Resp: 613579
Ion Ratio Lower Upper
128 100
130 32.3 12.1 52.1
64 43.4 26.6 66.6





#9
Benzaldehyde
Concen: 32.528 ng
RT: 6.528 min Scan# 619
Delta R.T. -0.012 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040




Tgt Ion: 77 Resp: 389532
Ion Ratio Lower Upper
77 100
105 94.4 74.5 114.5
106 92.0 72.3 112.3

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024




#10
Phenol
Concen: 35.139 ng
RT: 6.610 min Scan# 633
Delta R.T. -0.000 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42



Tgt Ion: 94 Resp: 743089
Ion Ratio Lower Upper
94 100
65 28.2 9.0 49.0
66 37.3 18.9 58.9






#11
bis(2-Chloroethyl)ether
Concen: 34.456 ng
RT: 6.816 min Scan# 609
Delta R.T. -0.012 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040






Tgt Ion: 93 Resp: 62400
Ion Ratio Lower Upper
93 100
63 73.9 55.8 95.8
95 31.5 12.7 52.7


Manual Integrations

APPROVED



Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024




#12
1,3-Dichlorobenzene
Concen: 37.119 ng
RT: 7.257 min Scan# 743
Delta R.T. -0.006 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42




Tgt Ion:146 Resp: 739956
Ion Ratio Lower Upper
146 100
148 65.5 52.3 78.5
75 28.5 25.8 38.6





#13
1,4-Dichlorobenzene
Concen: 37.165 ng
RT: 7.399 min Scan# 70
Delta R.T. -0.012 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:146 Resp: 75011
Ion Ratio Lower Upper
146 100
148 64.1 51.4 77.2
111 38.4 31.5 47.3




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#14
1,2-Dichlorobenzene
Concen: 37.243 ng
RT: 7.710 min Scan# 820
Delta R.T. -0.006 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42


Tgt Ion:146 Resp: 709235
Ion Ratio Lower Upper
146 100
148 64.6 51.8 77.8
111 40.1 33.3 49.9





#15
Benzyl Alcohol
Concen: 39.512 ng m
RT: 7.616 min Scan# 804
Delta R.T. -0.012 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion: 79 Resp: 59597
Ion Ratio Lower Upper
79 100
108 75.6 59.7 89.5
77 63.6 52.9 79.3




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#16
2,2'-oxybis(1-Chloropropane)
Concen: 33.169 ng m
RT: 7.904 min Scan# 853
Delta R.T. -0.012 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42


Tgt Ion: 45 Resp: 770805
Ion Ratio Lower Upper
45 100
77 16.4 0.0 35.6
79 13.6 0.0 32.4





#17
2-Methylphenol
Concen: 37.051 ng
RT: 7.828 min Scan# 841
Delta R.T. -0.006 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:107 Resp: 522725
Ion Ratio Lower Upper
107 100
108 115.3 90.6 135.8
77 48.2 39.3 58.9
79 46.3 38.6 58.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#18
Hexachloroethane
Concen: 36.104 ng
RT: 8.416 min Scan# 940
Delta R.T. -0.012 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42


Tgt Ion:117 Resp: 282269
Ion Ratio Lower Upper
117 100
119 97.0 77.0 115.4
201 111.9 81.3 121.9





#19
n-Nitroso-di-n-propylamine
Concen: 34.361 ng
RT: 8.175 min Scan# 899
Delta R.T. -0.012 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion: 70 Resp: 490524
Ion Ratio Lower Upper
70 100
42 51.5 42.0 63.0
101 9.6 7.2 10.8
130 21.0 15.5 23.3




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#20
3+4-Methylphenols
Concen: 37.048 ng
RT: 8.157 min Scan# 896
Delta R.T. -0.006 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42


Tgt Ion:107 Resp: 730719
Ion Ratio Lower Upper
107 100
108 90.2 70.7 110.7
77 36.3 17.4 57.4
79 29.4 10.9 50.9





#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.122 min Scan# 1117659
Delta R.T. -0.012 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:136 Resp: 1117659
Ion Ratio Lower Upper
136 100
137 10.9 8.8 13.2
54 6.7 6.0 9.0
68 6.1 4.8 7.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#22
Acetophenone
Concen: 35.527 ng
RT: 8.187 min Scan# 901
Delta R.T. -0.012 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42

Tgt Ion:105 Resp: 958453
Ion Ratio Lower Upper
105 100
71 5.5 4.5 6.7
51 24.0 20.1 30.1
120 22.5 17.8 26.6






#23
Nitrobenzene-d5
Concen: 69.585 ng
RT: 8.516 min Scan# 91
Delta R.T. -0.006 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040






Tgt Ion: 82 Resp: 154453
Ion Ratio Lower Upper
82 100
128 43.6 32.3 48.5
54 44.5 37.4 56.2


Manual Integrations

APPROVED



Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024




#24
Nitrobenzene
Concen: 35.161 ng
RT: 8.557 min Scan# 964
Delta R.T. -0.006 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42




Tgt Ion: 77 Resp: 785862
Ion Ratio Lower Upper
77 100
123 45.6 34.0 51.0
65 13.3 10.5 15.7





#25
Isophorone
Concen: 36.763 ng
RT: 9.081 min Scan# 1053
Delta R.T. -0.012 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion: 82 Resp: 142078
Ion Ratio Lower Upper
82 100
95 6.8 5.5 8.3
138 17.5 13.5 20.3




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#26
2-Nitrophenol
Concen: 39.578 ng
RT: 9.257 min Scan# 1083
Delta R.T. -0.006 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42


Tgt Ion:139 Resp: 391755
Ion Ratio Lower Upper
139 100
109 28.8 24.4 36.6
65 49.0 43.0 64.6





#27
2,4-Dimethylphenol
Concen: 38.384 ng
RT: 9.340 min Scan# 1098
Delta R.T. -0.006 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:122 Resp: 44504
Ion Ratio Lower Upper
122 100
107 116.6 92.8 139.2
121 58.7 46.1 69.1




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#28
bis(2-Chloroethoxy)methane
Concen: 35.845 ng
RT: 9.563 min Scan# 1135
Delta R.T. -0.012 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42


Tgt Ion: 93 Resp: 871685
Ion Ratio Lower Upper
93 100
95 33.0 25.8 38.8
123 11.1 9.0 13.4





#29
2,4-Dichlorophenol
Concen: 40.237 ng
RT: 9.792 min Scan# 1174
Delta R.T. -0.006 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:162 Resp: 69223
Ion Ratio Lower Upper
162 100
164 66.5 44.7 84.7
98 33.8 14.9 54.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#30
1,2,4-Trichlorobenzene
Concen: 39.461 ng
RT: 9.992 min Scan# 1208
Delta R.T. -0.006 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42


Tgt Ion:180 Resp: 768282
Ion Ratio Lower Upper
180 100
182 96.5 76.6 115.0
145 28.4 23.4 35.2





#31
Naphthalene
Concen: 36.472 ng
RT: 10.175 min Scan# 1239
Delta R.T. -0.006 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:128 Resp: 2030688
Ion Ratio Lower Upper
128 100
129 10.8 8.6 13.0
127 12.9 10.2 15.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#32
Benzoic acid
Concen: 41.637 ng
RT: 9.534 min Scan# 1130
Delta R.T. -0.000 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42


Tgt Ion:122 Resp: 560880
Ion Ratio Lower Upper
122 100
105 125.6 107.1 147.1
77 97.2 82.1 122.1





#33
4-Chloroaniline
Concen: 37.779 ng
RT: 10.298 min Scan# 1260
Delta R.T. -0.012 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion	Ratio	Lower	Upper
127	100		
129	31.8	25.8	38.8
65	30.7	24.9	37.3
92	18.6	15.1	22.7




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#34
Hexachlorobutadiene
Concen: 41.770 ng
RT: 10.457 min Scan# 1287
Delta R.T. -0.012 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42


Tgt Ion	Ratio	Lower	Upper
225	100		
223	62.4	50.2	75.2
227	65.9	51.4	77.0





#35
Caprolactam
Concen: 39.472 ng
RT: 11.110 min Scan# 1110
Delta R.T. -0.012 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:113 Resp: 235829
Ion Ratio Lower Upper
113 100
55 149.9 140.5 180.5
56 114.4 103.6 143.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#36
4-Chloro-3-methylphenol
Concen: 38.570 ng
RT: 11.451 min Scan# 1456
Delta R.T. -0.000 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42


Tgt Ion:107 Resp: 706688
Ion Ratio Lower Upper
107 100
144 27.8 21.8 32.6
142 84.6 66.8 100.2





#37
2-Methylnaphthalene
Concen: 37.618 ng
RT: 11.798 min Scan# 1515
Delta R.T. -0.006 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:142 Resp: 1491295
Ion Ratio Lower Upper
142 100
141 88.0 70.9 106.3
115 34.2 28.7 43.1




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#38
1-Methylnaphthalene
Concen: 37.334 ng
RT: 12.028 min Scan# 1554
Delta R.T. -0.006 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42


Tgt Ion:142 Resp: 1462706
Ion Ratio Lower Upper
142 100
141 91.1 74.6 111.8
116 3.7 3.1 4.7





#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.027 min Scan# 1895
Delta R.T. -0.006 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:164 Resp: 774502
Ion Ratio Lower Upper
164 100
162 99.9 80.7 121.1
160 44.2 36.0 54.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#40
1,2,4,5-Tetrachlorobenzene
Concen: 39.847 ng
RT: 12.186 min Scan# 1581
Delta R.T. -0.000 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42


Tgt Ion:216 Resp: 841763
Ion Ratio Lower Upper
216 100
214 78.0 63.4 95.2
179 20.9 17.6 26.4
108 16.1 15.1 22.7





#41
Hexachlorocyclopentadiene
Concen: 32.439 ng
RT: 12.157 min Scan# 1576
Delta R.T. -0.006 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:237 Resp: 239288
Ion Ratio Lower Upper
237 100
235 60.7 44.8 84.8
272 12.1 0.0 33.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#42
2,4,6-Tribromophenol
Concen: 80.421 ng
RT: 15.539 min Scan# 2151
Delta R.T. -0.000 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42


Tgt Ion:330 Resp: 722218
Ion Ratio Lower Upper
330 100
332 95.9 77.4 116.2
141 34.9 31.6 47.4





#43
2,4,6-Trichlorophenol
Concen: 39.935 ng
RT: 12.445 min Scan# 1625
Delta R.T. -0.000 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040




Tgt Ion:196 Resp: 599628
Ion Ratio Lower Upper
196 100
198 97.0 75.6 113.4
200 30.5 24.6 37.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024




#44
2,4,5-Trichlorophenol
Concen: 40.092 ng
RT: 12.522 min Scan# 1638
Delta R.T. -0.000 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42




Tgt Ion:196 Resp: 668381
Ion Ratio Lower Upper
196 100
198 96.1 77.7 116.5
97 44.6 39.8 59.8
132 24.0 20.7 31.1





#45
2-Fluorobiphenyl
Concen: 74.203 ng
RT: 12.633 min Scan# 1657
Delta R.T. -0.012 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:172 Resp: 3725750
Ion Ratio Lower Upper
172 100
171 34.9 27.8 41.6
170 23.0 18.7 28.1




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#46
1,1'-Biphenyl
Concen: 35.675 ng
RT: 12.851 min Scan# 1694
Delta R.T. -0.006 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42


Tgt Ion:154 Resp: 1967045
Ion Ratio Lower Upper
154 100
153 39.8 19.0 59.0
76 11.4 0.0 31.9





#47
2-Chloronaphthalene
Concen: 35.695 ng
RT: 12.886 min Scan# 1700
Delta R.T. -0.000 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:162 Resp: 1538290
Ion Ratio Lower Upper
162 100
127 35.3 30.0 45.0
164 32.8 26.3 39.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#48
2-Nitroaniline
Concen: 34.845 ng
RT: 13.110 min Scan# 1738
Delta R.T. -0.006 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42


Tgt Ion: 65 Resp: 470455
Ion Ratio Lower Upper
65 100
92 69.9 55.1 82.7
138 113.1 85.9 128.9





#49
Acenaphthylene
Concen: 36.146 ng
RT: 13.745 min Scan# 1846
Delta R.T. -0.006 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:152 Resp: 229615
Ion Ratio Lower Upper
152 100
151 19.5 15.8 23.6
153 12.9 10.4 15.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#50
Dimethylphthalate
Concen: 36.964 ng
RT: 13.510 min Scan# 1806
Delta R.T. -0.006 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42


Tgt Ion:163 Resp: 1999459
Ion Ratio Lower Upper
163 100
194 4.2 3.0 4.6
164 10.0 8.0 12.0





#51
2,6-Dinitrotoluene
Concen: 38.107 ng
RT: 13.628 min Scan# 11
Delta R.T. -0.000 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:165 Resp: 481854
Ion Ratio Lower Upper
165 100
63 48.2 46.5 69.7
89 51.3 47.3 70.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#52
Acenaphthene
Concen: 36.049 ng
RT: 14.092 min Scan# 1905
Delta R.T. -0.006 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42


Tgt Ion:154 Resp: 1452604
Ion Ratio Lower Upper
154 100
153 111.0 90.3 135.5
152 52.3 42.9 64.3





#53
3-Nitroaniline
Concen: 37.363 ng
RT: 13.957 min Scan# 1882
Delta R.T. -0.006 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:138 Resp: 473408
Ion Ratio Lower Upper
138 100
108 10.0 7.9 11.9
92 111.3 92.0 138.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#54
2,4-Dinitrophenol
Concen: 36.181 ng
RT: 14.180 min Scan# 1920
Delta R.T. 0.005 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42


Tgt Ion:184 Resp: 274199
Ion Ratio Lower Upper
184 100
63 58.3 52.7 79.1
154 57.8 50.4 75.6





#55
Dibenzenofuran
Concen: 36.080 ng
RT: 14.433 min Scan# 1963
Delta R.T. -0.006 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:168 Resp: 2296458
Ion Ratio Lower Upper
168 100
139 35.1 29.6 44.4
169 13.5 10.7 16.1




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#56
4-Nitrophenol
Concen: 36.695 ng
RT: 14.310 min Scan# 1942
Delta R.T. 0.012 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42


Tgt Ion:139 Resp: 400919
Ion Ratio Lower Upper
139 100
109 71.5 52.5 92.5
65 90.9 75.4 115.4





#57
2,4-Dinitrotoluene
Concen: 37.626 ng
RT: 14.433 min Scan# 1963
Delta R.T. 0.006 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:165 Resp: 63504
Ion Ratio Lower Upper
165 100
63 51.8 40.3 60.5
89 73.2 61.0 91.6
182 2.7 1.9 2.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#58
Fluorene
Concen: 36.346 ng
RT: 15.092 min Scan# 2075
Delta R.T. -0.000 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42


Tgt Ion:166 Resp: 1908826
Ion Ratio Lower Upper
166 100
165 96.8 77.4 116.0
167 13.6 11.0 16.6





#59
2,3,4,6-Tetrachlorophenol
Concen: 40.178 ng
RT: 14.680 min Scan# 2005
Delta R.T. 0.006 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:232 Resp: 550491
Ion Ratio Lower Upper
232 100
131 40.6 36.9 55.3
130 2.2 2.0 3.0
166 30.3 26.4 39.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#60
Diethylphthalate
Concen: 36.316 ng
RT: 14.892 min Scan# 2041
Delta R.T. -0.006 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42


Tgt Ion:149 Resp: 2006026
Ion Ratio Lower Upper
149 100
177 23.1 17.6 26.4
150 12.3 9.8 14.6





#61
4-Chlorophenyl-phenylether
Concen: 37.746 ng
RT: 15.098 min Scan# 2076
Delta R.T. -0.000 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:204 Resp: 94373
Ion Ratio Lower Upper
204 100
206 33.2 26.2 39.4
141 54.5 47.2 70.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#62
4-Nitroaniline
Concen: 37.048 ng
RT: 15.133 min Scan# 2082
Delta R.T. -0.006 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42


Tgt Ion:138 Resp: 462456
Ion Ratio Lower Upper
138 100
92 50.3 31.9 71.9
108 79.8 63.8 103.8





#63
Azobenzene
Concen: 33.274 ng
RT: 15.392 min Scan# 2126
Delta R.T. -0.000 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion: 77 Resp: 1742711
Ion Ratio Lower Upper
77 100
182 27.8 3.8 43.8
105 20.8 0.0 39.5
51 25.3 6.8 46.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 16.792 min Scan# 2364
Delta R.T. -0.000 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42


Tgt Ion:188 Resp: 1666055
Ion Ratio Lower Upper
188 100
94 8.7 7.0 10.6
80 9.3 8.0 12.0





#65
4,6-Dinitro-2-methylphenol
Concen: 39.198 ng
RT: 15.204 min Scan# 2094
Delta R.T. 0.006 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:198 Resp: 38235
Ion Ratio Lower Upper
198 100
51 37.4 23.9 63.9
105 38.2 24.3 64.3




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#66
n-Nitrosodiphenylamine
Concen: 36.072 ng
RT: 15.316 min Scan# 2113
Delta R.T. -0.006 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42


Tgt Ion:169 Resp: 1636928
Ion Ratio Lower Upper
169 100
168 66.4 53.8 80.8
167 35.7 29.3 43.9





#67
4-Bromophenyl-phenylether
Concen: 38.724 ng
RT: 15.992 min Scan# 211
Delta R.T. -0.006 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:248 Resp: 621680
Ion Ratio Lower Upper
248 100
250 96.9 77.8 116.8
141 68.6 57.7 86.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#68
Hexachlorobenzene
Concen: 37.617 ng
RT: 16.104 min Scan# 2247
Delta R.T. 0.006 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42


Tgt Ion:284 Resp: 723291
Ion Ratio Lower Upper
284 100
142 31.6 29.4 44.0
249 30.6 25.1 37.7





#69
Atrazine
Concen: 40.182 ng
RT: 16.292 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:200 Resp: 54870
Ion Ratio Lower Upper
200 100
173 25.2 6.1 46.1
215 47.4 26.8 66.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#70
Pentachlorophenol
Concen: 38.946 ng
RT: 16.457 min Scan# 2307
Delta R.T. 0.006 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42


Tgt Ion:266 Resp: 524917
Ion Ratio Lower Upper
266 100
268 64.4 49.8 74.6
264 63.3 49.6 74.4





#71
Phenanthrene
Concen: 35.707 ng
RT: 16.833 min Scan# 21
Delta R.T. -0.006 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:178 Resp: 298554
Ion Ratio Lower Upper
178 100
176 19.3 15.3 22.9
179 15.7 12.4 18.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#72
Anthracene
Concen: 36.372 ng
RT: 16.927 min Scan# 2387
Delta R.T. -0.000 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42


Tgt Ion:178 Resp: 2959126
Ion Ratio Lower Upper
178 100
176 18.5 14.8 22.2
179 15.3 12.4 18.6





#73
Carbazole
Concen: 35.772 ng
RT: 17.210 min Scan# 2435
Delta R.T. -0.000 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:167 Resp: 297145
Ion Ratio Lower Upper
167 100
166 21.1 17.1 25.7
139 11.9 10.0 15.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#74
Di-n-butylphthalate
Concen: 36.901 ng
RT: 17.792 min Scan# 2534
Delta R.T. -0.000 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42


Tgt Ion:149 Resp: 3566712
Ion Ratio Lower Upper
149 100
150 9.4 7.5 11.3
104 4.7 4.1 6.1





#75
Fluoranthene
Concen: 36.741 ng
RT: 18.874 min Scan# 21
Delta R.T. 0.006 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:202 Resp: 3736572
Ion Ratio Lower Upper
202 100
101 9.1 0.0 30.2
203 17.6 0.0 37.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.039 min Scan# 3086
Delta R.T. 0.006 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42


Tgt Ion:240 Resp: 1393632
Ion Ratio Lower Upper
240 100
120 9.0 8.2 12.2
236 25.3 20.6 30.8





#77
Benzidine
Concen: 74.129 ng
RT: 19.080 min Scan# 21
Delta R.T. 0.006 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:184 Resp: 175194
Ion Ratio Lower Upper
184 100
185 13.9 11.8 17.6
183 11.4 9.0 13.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#78
Pyrene
Concen: 38.894 ng
RT: 19.239 min Scan# 2780
Delta R.T. -0.000 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42

Tgt Ion:202 Resp: 3870284
Ion Ratio Lower Upper
202 100
200 20.9 17.2 25.8
203 17.5 14.2 21.2






#79
Terphenyl-d14
Concen: 79.011 ng
RT: 19.462 min Scan# 2818
Delta R.T. -0.000 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040






Tgt Ion:244 Resp: 5138818
Ion Ratio Lower Upper
244 100
212 7.3 6.2 9.4
122 8.5 7.1 10.7

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#80
Butylbenzylphthalate
Concen: 39.837 ng
RT: 20.174 min Scan# 2939
Delta R.T. -0.000 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42

Tgt Ion:149 Resp: 1617756
Ion Ratio Lower Upper
149 100
91 69.1 57.0 85.4
206 21.8 16.4 24.6






#81
Benzo(a)anthracene
Concen: 36.457 ng
RT: 21.021 min Scan# 3082
Delta R.T. 0.006 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040






Tgt Ion:228 Resp: 3372910
Ion Ratio Lower Upper
228 100
226 27.6 21.8 32.6
229 19.7 15.8 23.8


Manual Integrations

APPROVED



Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024




#82
3,3'-Dichlorobenzidine
Concen: 40.023 ng
RT: 20.962 min Scan# 3073
Delta R.T. 0.006 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42




Tgt Ion:252 Resp: 1158555
Ion Ratio Lower Upper
252 100
254 64.6 51.0 76.6
126 9.3 8.5 12.7





#83
Chrysene
Concen: 35.801 ng
RT: 21.080 min Scan# 3092
Delta R.T. 0.006 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:228 Resp: 3142558
Ion Ratio Lower Upper
228 100
226 30.3 24.4 36.6
229 20.0 15.7 23.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#84
Bis(2-ethylhexyl)phthalate
Concen: 38.531 ng
RT: 20.974 min Scan# 3075
Delta R.T. -0.000 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42


Tgt Ion:149 Resp: 2219045
Ion Ratio Lower Upper
149 100
167 28.2 22.2 33.2
279 4.2 3.0 4.4





#85
Di-n-octyl phthalate
Concen: 40.314 ng
RT: 22.009 min Scan# 31
Delta R.T. -0.000 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:149 Resp: 394629
Ion Ratio Lower Upper
149 100
167 1.6 1.3 1.9
43 7.6 6.9 10.3




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#86
Perylene-d12
Concen: 20.000 ng
RT: 23.744 min Scan# 3546
Delta R.T. 0.018 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42


Tgt Ion:264 Resp: 1655413
Ion Ratio Lower Upper
264 100
260 23.8 18.5 27.7
265 21.5 17.3 25.9





#87
Indeno(1,2,3-cd)pyrene
Concen: 32.635 ng
RT: 26.756 min Scan# 4053
Delta R.T. 0.029 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion: 276 Resp: 349200
Ion Ratio Lower Upper
276 100
138 22.5 19.4 29.0
277 25.2 20.1 30.1




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#88
Benzo(b)fluoranthene
Concen: 35.020 ng
RT: 22.898 min Scan# 3402
Delta R.T. 0.012 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42


Tgt Ion: 252 Resp: 3442805
Ion Ratio Lower Upper
252 100
253 22.2 17.8 26.6
125 8.7 7.4 11.0





#89
Benzo(k)fluoranthene
Concen: 34.961 ng
RT: 22.956 min Scan# 3410
Delta R.T. 0.012 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:252 Resp: 324633
Ion Ratio Lower Upper
252 100
253 21.6 17.7 26.5
125 8.6 7.4 11.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#90
Benzo(a)pyrene
Concen: 35.215 ng
RT: 23.621 min Scan# 3525
Delta R.T. 0.012 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42

Tgt Ion:252 Resp: 2998995
Ion Ratio Lower Upper
252 100
253 22.4 17.1 25.7
125 9.2 8.2 12.2





#91
Dibenzo(a,h)anthracene
Concen: 33.190 ng
RT: 26.803 min Scan# 4062
Delta R.T. 0.023 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42

Instrument :

BNA_M





ClientSampleId :

SSTDCCC040

Tgt Ion:278 Resp: 287409
Ion Ratio Lower Upper
278 100
139 14.7 12.0 18.0
279 23.3 19.2 28.8




Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#92
Benzo(g,h,i)perylene
Concen: 31.992 ng
RT: 27.685 min Scan# 4216
Delta R.T. 0.023 min
Lab File: BM047245.D
Acq: 19 Aug 2024 09:42

Tgt Ion:276 Resp: 2876281
Ion Ratio Lower Upper
276 100
277 23.9 18.8 28.2
138 19.1 16.2 24.4



Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081924\
 Data File : BM047245.D
 Acq On : 19 Aug 2024 09:42
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 19 11:18:18 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Aug 11 23:47:58 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	80	0.00
2	1,4-Dioxane	0.464	0.372	19.8	68	0.00
3	Pyridine	1.360	1.126	17.2	69	0.00
4	n-Nitrosodimethylamine	0.594	0.454	23.6	64	0.00
5 S	2-Fluorophenol	1.119	0.985	12.0	74	0.00
6	Aniline	1.458	1.322	9.3	74	-0.01
7 S	Phenol-d6	1.542	1.370	11.2	74	0.00
8	2-Chlorophenol	1.221	1.116	8.6	75	0.00
9	Benzaldehyde	0.904	0.709	21.6	73	-0.01
10 C	Phenol	1.539	1.352	12.2	74	0.00
11	bis(2-Chloroethyl)ether	1.318	1.135	13.9	71	-0.01
12	1,3-Dichlorobenzene	1.451	1.346	7.2	77	0.00
13 C	1,4-Dichlorobenzene	1.469	1.365	7.1	76	-0.01
14	1,2-Dichlorobenzene	1.386	1.290	6.9	79	0.00
15	Benzyl Alcohol	1.098	1.084	1.3	79	-0.01
16	2,2'-oxybis(1-Chloropropane	1.691	1.402	17.1	68	-0.01
17	2-Methylphenol	1.027	0.951	7.4	76	0.00
18	Hexachloroethane	0.569	0.514	9.7	75	-0.01
19 P	n-Nitroso-di-n-propylamine	1.039	0.892	14.1	73	-0.01
20	3+4-Methylphenols	1.435	1.329	7.4	75	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	81	-0.01
22	Acetophenone	0.483	0.429	11.2	75	-0.01
23 S	Nitrobenzene-d5	0.397	0.345	13.1	71	0.00
24	Nitrobenzene	0.400	0.352	12.0	72	0.00
25	Isophorone	0.692	0.636	8.1	74	-0.01
26 C	2-Nitrophenol	0.177	0.175	1.1	78	0.00
27	2,4-Dimethylphenol	0.207	0.199	3.9	77	0.00
28	bis(2-Chloroethoxy)methane	0.435	0.390	10.3	72	-0.01
29 C	2,4-Dichlorophenol	0.308	0.310	-0.6	80	0.00
30	1,2,4-Trichlorobenzene	0.348	0.344	1.1	80	0.00
31	Naphthalene	0.996	0.908	8.8	76	0.00
32	Benzoic acid	0.241	0.251	-4.1	81	0.00
33	4-Chloroaniline	0.383	0.362	5.5	76	-0.01
34 C	Hexachlorobutadiene	0.204	0.213	-4.4	85	-0.01
35	Caprolactam	0.107	0.106	0.9	78	-0.01
36 C	4-Chloro-3-methylphenol	0.328	0.316	3.7	78	0.00
37	2-Methylnaphthalene	0.709	0.667	5.9	78	0.00
38	1-Methylnaphthalene	0.701	0.654	6.7	77	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	85	0.00
40	1,2,4,5-Tetrachlorobenzene	0.546	0.543	0.5	85	0.00
41 P	Hexachlorocyclopentadiene	0.190	0.154	18.9	65	0.00
42 S	2,4,6-Tribromophenol	0.232	0.233	-0.4	85	0.00
43 C	2,4,6-Trichlorophenol	0.388	0.387	0.3	82	0.00
44	2,4,5-Trichlorophenol	0.430	0.431	-0.2	84	0.00
45 S	2-Fluorobiphenyl	1.297	1.203	7.2	80	-0.01
46	1,1'-Biphenyl	1.424	1.270	10.8	77	0.00
47	2-Chloronaphthalene	1.113	0.993	10.8	78	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081924\
 Data File : BM047245.D
 Acq On : 19 Aug 2024 09:42
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 19 11:18:18 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Aug 11 23:47:58 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.349	0.304	12.9	72	0.00
49	Acenaphthylene	1.640	1.482	9.6	78	0.00
50	Dimethylphthalate	1.397	1.291	7.6	79	0.00
51	2,6-Dinitrotoluene	0.327	0.311	4.9	80	0.00
52 C	Acenaphthene	1.041	0.938	9.9	79	0.00
53	3-Nitroaniline	0.327	0.306	6.4	78	0.00
54 P	2,4-Dinitrophenol	0.196	0.177	9.7	73	0.00
55	Dibenzofuran	1.644	1.483	9.8	79	0.00
56 P	4-Nitrophenol	0.282	0.259	8.2	75	0.01
57	2,4-Dinitrotoluene	0.436	0.410	6.0	79	0.00
58	Fluorene	1.356	1.232	9.1	80	0.00
59	2,3,4,6-Tetrachlorophenol	0.354	0.355	-0.3	85	0.00
60	Diethylphthalate	1.426	1.295	9.2	78	0.00
61	4-Chlorophenyl-phenylether	0.646	0.609	5.7	83	0.00
62	4-Nitroaniline	0.322	0.299	7.1	79	0.00
63	Azobenzene	1.352	1.125	16.8	72	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	88	0.00
65	4,6-Dinitro-2-methylphenol	0.117	0.115	1.7	80	0.00
66 c	n-Nitrosodiphenylamine	0.545	0.491	9.9	80	0.00
67	4-Bromophenyl-phenylether	0.193	0.187	3.1	86	0.00
68	Hexachlorobenzene	0.231	0.217	6.1	83	0.00
69	Atrazine	0.164	0.165	-0.6	85	0.00
70 C	Pentachlorophenol	0.162	0.158	2.5	82	0.00
71	Phenanthrene	1.004	0.896	10.8	80	0.00
72	Anthracene	0.977	0.888	9.1	82	0.00
73	Carbazole	0.997	0.892	10.5	80	0.00
74	Di-n-butylphthalate	1.160	1.070	7.8	80	0.00
75 C	Fluoranthene	1.221	1.121	8.2	82	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	83	0.00
77	Benzidine	0.339	0.629	-85.5#	107	0.00
78	Pyrene	1.428	1.389	2.7	81	0.00
79 S	Terphenyl-d14	0.933	0.922	1.2	82	0.00
80	Butylbenzylphthalate	0.583	0.580	0.5	79	0.00
81	Benzo(a)anthracene	1.328	1.210	8.9	77	0.00
82	3,3'-Dichlorobenzidine	0.415	0.416	-0.2	78	0.00
83	Chrysene	1.260	1.127	10.6	77	0.00
84	Bis(2-ethylhexyl)phthalate	0.826	0.796	3.6	78	0.00
85 c	Di-n-octyl phthalate	1.405	1.416	-0.8	81	0.00
86 I	Perylene-d12	1.000	1.000	0.0	82	0.02
87	Indeno(1,2,3-cd)pyrene	1.293	1.055	18.4	68	0.03
88	Benzo(b)fluoranthene	1.188	1.040	12.5	73	0.01
89	Benzo(k)fluoranthene	1.122	0.981	12.6	73	0.01
90 C	Benzo(a)pyrene	1.029	0.906	12.0	72	0.01
91	Dibenzo(a,h)anthracene	1.046	0.868	17.0	70	0.02
92	Benzo(g,h,i)perylene	1.086	0.869	20.0	66	0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081924\
Data File : BM047245.D
Acq On : 19 Aug 2024 09:42
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC040

Quant Time: Aug 19 11:18:18 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sun Aug 11 23:47:58 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081924\
 Data File : BM047245.D
 Acq On : 19 Aug 2024 09:42
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 19 11:18:18 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Aug 11 23:47:58 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	80	0.00
2	1,4-Dioxane	40.000	32.055	19.9	68	0.00
3	Pyridine	40.000	33.106	17.2	69	0.00
4	n-Nitrosodimethylamine	40.000	30.567	23.6	64	0.00
5 S	2-Fluorophenol	80.000	70.374	12.0	74	0.00
6	Aniline	40.000	36.280	9.3	74	-0.01
7 S	Phenol-d6	80.000	71.069	11.2	74	0.00
8	2-Chlorophenol	40.000	36.576	8.6	75	0.00
9	Benzaldehyde	40.000	32.528	18.7	73	-0.01
10 C	Phenol	40.000	35.139	12.2	74	0.00
11	bis(2-Chloroethyl)ether	40.000	34.456	13.9	71	-0.01
12	1,3-Dichlorobenzene	40.000	37.119	7.2	77	0.00
13 C	1,4-Dichlorobenzene	40.000	37.165	7.1	76	-0.01
14	1,2-Dichlorobenzene	40.000	37.243	6.9	79	0.00
15	Benzyl Alcohol	40.000	39.512	1.2	79	-0.01
16	2,2'-oxybis(1-Chloropropane	40.000	33.169	17.1	68	-0.01
17	2-Methylphenol	40.000	37.051	7.4	76	0.00
18	Hexachloroethane	40.000	36.104	9.7	75	-0.01
19 P	n-Nitroso-di-n-propylamine	40.000	34.361	14.1	73	-0.01
20	3+4-Methylphenols	40.000	37.048	7.4	75	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	81	-0.01
22	Acetophenone	40.000	35.527	11.2	75	-0.01
23 S	Nitrobenzene-d5	80.000	69.585	13.0	71	0.00
24	Nitrobenzene	40.000	35.161	12.1	72	0.00
25	Isophorone	40.000	36.763	8.1	74	-0.01
26 C	2-Nitrophenol	40.000	39.578	1.1	78	0.00
27	2,4-Dimethylphenol	40.000	38.384	4.0	77	0.00
28	bis(2-Chloroethoxy)methane	40.000	35.845	10.4	72	-0.01
29 C	2,4-Dichlorophenol	40.000	40.237	-0.6	80	0.00
30	1,2,4-Trichlorobenzene	40.000	39.461	1.3	80	0.00
31	Naphthalene	40.000	36.472	8.8	76	0.00
32	Benzoic acid	40.000	41.637	-4.1	81	0.00
33	4-Chloroaniline	40.000	37.779	5.6	76	-0.01
34 C	Hexachlorobutadiene	40.000	41.770	-4.4	85	-0.01
35	Caprolactam	40.000	39.472	1.3	78	-0.01
36 C	4-Chloro-3-methylphenol	40.000	38.570	3.6	78	0.00
37	2-Methylnaphthalene	40.000	37.618	6.0	78	0.00
38	1-Methylnaphthalene	40.000	37.334	6.7	77	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	85	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.847	0.4	85	0.00
41 P	Hexachlorocyclopentadiene	40.000	32.439	18.9	65	0.00
42 S	2,4,6-Tribromophenol	80.000	80.421	-0.5	85	0.00
43 C	2,4,6-Trichlorophenol	40.000	39.935	0.2	82	0.00
44	2,4,5-Trichlorophenol	40.000	40.092	-0.2	84	0.00
45 S	2-Fluorobiphenyl	80.000	74.203	7.2	80	-0.01
46	1,1'-Biphenyl	40.000	35.675	10.8	77	0.00
47	2-Chloronaphthalene	40.000	35.695	10.8	78	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081924\
 Data File : BM047245.D
 Acq On : 19 Aug 2024 09:42
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 19 11:18:18 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Aug 11 23:47:58 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	34.845	12.9	72	0.00
49	Acenaphthylene	40.000	36.146	9.6	78	0.00
50	Dimethylphthalate	40.000	36.964	7.6	79	0.00
51	2,6-Dinitrotoluene	40.000	38.107	4.7	80	0.00
52 C	Acenaphthene	40.000	36.049	9.9	79	0.00
53	3-Nitroaniline	40.000	37.363	6.6	78	0.00
54 P	2,4-Dinitrophenol	40.000	36.181	9.5	73	0.00
55	Dibenzofuran	40.000	36.080	9.8	79	0.00
56 P	4-Nitrophenol	40.000	36.695	8.3	75	0.01
57	2,4-Dinitrotoluene	40.000	37.626	5.9	79	0.00
58	Fluorene	40.000	36.346	9.1	80	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	40.178	-0.4	85	0.00
60	Diethylphthalate	40.000	36.316	9.2	78	0.00
61	4-Chlorophenyl-phenylether	40.000	37.746	5.6	83	0.00
62	4-Nitroaniline	40.000	37.048	7.4	79	0.00
63	Azobenzene	40.000	33.274	16.8	72	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	88	0.00
65	4,6-Dinitro-2-methylphenol	40.000	39.198	2.0	80	0.00
66 c	n-Nitrosodiphenylamine	40.000	36.072	9.8	80	0.00
67	4-Bromophenyl-phenylether	40.000	38.724	3.2	86	0.00
68	Hexachlorobenzene	40.000	37.617	6.0	83	0.00
69	Atrazine	40.000	40.182	-0.5	85	0.00
70 C	Pentachlorophenol	40.000	38.946	2.6	82	0.00
71	Phenanthrene	40.000	35.707	10.7	80	0.00
72	Anthracene	40.000	36.372	9.1	82	0.00
73	Carbazole	40.000	35.772	10.6	80	0.00
74	Di-n-butylphthalate	40.000	36.901	7.7	80	0.00
75 C	Fluoranthene	40.000	36.741	8.1	82	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	83	0.00
77	Benzidine	40.000	74.129	-85.3#	107	0.00
78	Pyrene	40.000	38.894	2.8	81	0.00
79 S	Terphenyl-d14	80.000	79.011	1.2	82	0.00
80	Butylbenzylphthalate	40.000	39.837	0.4	79	0.00
81	Benzo(a)anthracene	40.000	36.457	8.9	77	0.00
82	3,3'-Dichlorobenzidine	40.000	40.023	-0.1	78	0.00
83	Chrysene	40.000	35.801	10.5	77	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	38.531	3.7	78	0.00
85 c	Di-n-octyl phthalate	40.000	40.314	-0.8	81	0.00
86 I	Perylene-d12	20.000	20.000	0.0	82	0.02
87	Indeno(1,2,3-cd)pyrene	40.000	32.635	18.4	68	0.03
88	Benzo(b)fluoranthene	40.000	35.020	12.4	73	0.01
89	Benzo(k)fluoranthene	40.000	34.961	12.6	73	0.01
90 C	Benzo(a)pyrene	40.000	35.215	12.0	72	0.01
91	Dibenzo(a,h)anthracene	40.000	33.190	17.0	70	0.02
92	Benzo(g,h,i)perylene	40.000	31.992	20.0	66	0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081924\
Data File : BM047245.D
Acq On : 19 Aug 2024 09:42
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC040

Quant Time: Aug 19 11:18:18 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sun Aug 11 23:47:58 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range SPCC's out = 0 CCC's out = 0



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: JAC005

Lab Code: CHEM Case No.: P3609 SAS No.: P3609 SDG No.: P3609

Instrument ID: BNA_M Calibration Date/Time: 08/22/2024 09:14

Lab File ID: BM047308.D Init. Calib. Date(s): 08/10/2024 08/10/2024

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 13:12 17:51

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
Pyridine	1.360	1.164		-14.4	
2-Fluorophenol	1.119	1.034		-7.6	
Benzaldehyde	0.904	0.830		-8.2	
Phenol-d6	1.542	1.396		-9.5	
2-Methylphenol	1.027	0.949		-7.6	
3+4-Methylphenols	1.435	1.298		-9.5	
Nitrobenzene-d5	0.397	0.355		-10.6	
Hexachloroethane	0.569	0.525		-7.7	
Nitrobenzene	0.400	0.365		-8.8	
Naphthalene	0.996	0.950		-4.6	
Hexachlorobutadiene	0.204	0.229		12.3	20.0
2-Methylnaphthalene	0.709	0.696		-1.8	
2,4,6-Trichlorophenol	0.388	0.405		4.4	20.0
2-Fluorobiphenyl	1.297	1.252		-3.5	
2,4,5-Trichlorophenol	0.430	0.449		4.4	
Acenaphthylene	1.640	1.557		-5.1	
Acenaphthene	1.041	0.985		-5.4	20.0
Dibenzofuran	1.644	1.575		-4.2	
2,4-Dinitrotoluene	0.436	0.431		-1.1	
Fluorene	1.356	1.297		-4.4	
2,4,6-Tribromophenol	0.232	0.243		4.7	
Hexachlorobenzene	0.231	0.232		0.4	
Pentachlorophenol	0.162	0.164		1.2	20.0
Phenanthrene	1.004	0.946		-5.8	
Anthracene	0.977	0.929		-4.9	
Carbazole	0.997	0.928		-6.9	
Di-n-butylphthalate	1.160	1.134		-2.2	
Fluoranthene	1.221	1.198		-1.9	20.0
Pyrene	1.428	1.416		-0.8	
Terphenyl-d14	0.933	1.039		11.4	
Benzo (a) anthracene	1.328	1.267		-4.6	
Chrysene	1.260	1.185		-6.0	
Bis (2-ethylhexyl) phthalate	0.826	0.838		1.5	
Benzo (b) fluoranthene	1.188	1.174		-1.2	
Benzo (k) fluoranthene	1.122	1.122		0.0	
Benzo (a) pyrene	1.029	0.998		-3.0	20.0
Indeno (1,2,3-cd) pyrene	1.293	1.113		-13.9	
Dibenzo (a,h) anthracene	1.046	0.901		-13.9	
Benzo (g,h,i) perylene	1.086	0.915		-15.7	



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: JAC005
Lab Code: CHEM Case No.: P3609 SAS No.: P3609 SDG No.: P3609
Instrument ID: BNA_M Calibration Date/Time: 08/22/2024 09:14
Lab File ID: BM047308.D Init. Calib. Date(s): 08/10/2024 08/10/2024
EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 13:12 17:51
GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	<u>RRF</u>	RRF040	MIN RRF	%D	MAX%D
1,4-Dioxane	0.464	0.417		-10.1	20.0
1-Methylnaphthalene	0.701	0.678		-3.3	

All other compounds must meet a minimum RRF of 0.010.

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM082224\
 Data File : BM047308.D
 Acq On : 22 Aug 2024 09:14
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC040

Quant Time: Aug 22 10:59:43 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Aug 11 23:47:58 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.351	152	315372	20.000	ng	-0.02
21) Naphthalene-d8	10.110	136	1211803	20.000	ng	-0.02
39) Acenaphthene-d10	14.016	164	822573	20.000	ng	-0.02
64) Phenanthrene-d10	16.780	188	1768064	20.000	ng	-0.01
76) Chrysene-d12	21.021	240	1535454	20.000	ng	-0.01
86) Perylene-d12	23.709	264	1629252	20.000	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.010	112	1303996	73.882	ng	-0.01
7) Phenol-d6	6.569	99	1761491	72.430	ng	-0.02
23) Nitrobenzene-d5	8.504	82	1721029	71.513	ng	-0.02
42) 2,4,6-Tribromophenol	15.527	330	800401	83.918	ng	-0.01
45) 2-Fluorobiphenyl	12.622	172	4120450	77.269	ng	-0.02
79) Terphenyl-d14	19.445	244	6380197	89.037	ng	-0.02
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.028	88	263249	35.998	ng	96
3) Pyridine	3.399	79	734306	34.243	ng	98
4) n-Nitrosodimethylamine	3.322	42	306437	32.708	ng	96
6) Aniline	6.704	93	809995	35.238	ng	98
8) 2-Chlorophenol	6.928	128	714965	37.141	ng	99
9) Benzaldehyde	6.516	77	523783	41.837	ng	99
10) Phenol	6.593	94	863321	35.576	ng	98
11) bis(2-Chloroethyl)ether	6.804	93	742990	35.752	ng	97
12) 1,3-Dichlorobenzene	7.240	146	879161	38.433	ng	97
13) 1,4-Dichlorobenzene	7.387	146	886468	38.275	ng	98
14) 1,2-Dichlorobenzene	7.698	146	825943	37.796	ng	98
15) Benzyl Alcohol	7.604	79	619139	35.771	ng	98
16) 2,2'-oxybis(1-Chloropr...	7.892	45	880322	33.012	ng	98
17) 2-Methylphenol	7.816	107	598616	36.976	ng	99
18) Hexachloroethane	8.404	117	331304	36.929	ng	94
19) n-Nitroso-di-n-propyla...	8.169	70	561893	34.301	ng	96
20) 3+4-Methylphenols	8.145	107	818840	36.179	ng	100
22) Acetophenone	8.175	105	1094687	37.424	ng	98
24) Nitrobenzene	8.545	77	884972	36.519	ng	97
25) Isophorone	9.069	82	1591106	37.971	ng	98
26) 2-Nitrophenol	9.245	139	449274	41.863	ng	94
27) 2,4-Dimethylphenol	9.328	122	496910	39.527	ng	99
28) bis(2-Chloroethoxy)met...	9.551	93	993074	37.664	ng	100
29) 2,4-Dichlorophenol	9.775	162	782273	41.938	ng	99
30) 1,2,4-Trichlorobenzene	9.975	180	881137	41.742	ng	100
31) Naphthalene	10.157	128	2303393	38.156	ng	99
32) Benzoic acid	9.522	122	603943	41.350	ng	99
33) 4-Chloroaniline	10.286	127	873816	37.641	ng	99
34) Hexachlorobutadiene	10.439	225	554015	44.840	ng	100
35) Caprolactam	11.104	113	259311	40.031	ng	91
36) 4-Chloro-3-methylphenol	11.439	107	785639	39.548	ng	100
37) 2-Methylnaphthalene	11.786	142	1687918	39.270	ng	98
38) 1-Methylnaphthalene	12.010	142	1643957	38.700	ng	99
40) 1,2,4,5-Tetrachloroben...	12.169	216	957174	42.662	ng	98
41) Hexachlorocyclopentadiene	12.139	237	284615	36.329	ng	98
43) 2,4,6-Trichlorophenol	12.427	196	665686	41.744	ng	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM082224\
 Data File : BM047308.D
 Acq On : 22 Aug 2024 09:14
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC040

Quant Time: Aug 22 10:59:43 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Aug 11 23:47:58 2024
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	12.510	196	738243	41.695	ng	96
46) 1,1'-Biphenyl	12.833	154	2220122	37.912	ng	98
47) 2-Chloronaphthalene	12.869	162	1733753	37.879	ng	99
48) 2-Nitroaniline	13.098	65	519350	36.219	ng	96
49) Acenaphthylene	13.727	152	2561019	37.959	ng	100
50) Dimethylphthalate	13.498	163	2232553	38.861	ng	100
51) 2,6-Dinitrotoluene	13.616	165	536890	39.978	ng	88
52) Acenaphthene	14.080	154	1620780	37.872	ng	100
53) 3-Nitroaniline	13.945	138	510904	37.966	ng	100
54) 2,4-Dinitrophenol	14.169	184	303462	37.702	ng	93
55) Dibenzofuran	14.421	168	2590656	38.323	ng	99
56) 4-Nitrophenol	14.298	139	415593	35.815	ng	98
57) 2,4-Dinitrotoluene	14.416	165	708560	39.529	ng	# 97
58) Fluorene	15.074	166	2133336	38.248	ng	100
59) 2,3,4,6-Tetrachlorophenol	14.663	232	609310	41.872	ng	93
60) Diethylphthalate	14.880	149	2237417	38.138	ng	98
61) 4-Chlorophenyl-phenyle...	15.080	204	1071908	40.367	ng	95
62) 4-Nitroaniline	15.127	138	500966	37.788	ng	95
63) Azobenzene	15.374	77	1953128	35.112	ng	95
65) 4,6-Dinitro-2-methylph...	15.186	198	425285	41.084	ng	92
66) n-Nitrosodiphenylamine	15.304	169	1831258	38.026	ng	99
67) 4-Bromophenyl-phenylether	15.980	248	705879	41.431	ng	96
68) Hexachlorobenzene	16.086	284	821372	40.254	ng	94
69) Atrazine	16.274	200	511524	35.298	ng	99
70) Pentachlorophenol	16.439	266	580674	40.597	ng	98
71) Phenanthrene	16.821	178	3343698	37.683	ng	100
72) Anthracene	16.910	178	3284636	38.044	ng	100
73) Carbazole	17.198	167	3281091	37.220	ng	99
74) Di-n-butylphthalate	17.774	149	4010258	39.097	ng	99
75) Fluoranthene	18.856	202	4234649	39.236	ng	99
77) Benzidine	19.062	184	1551050	59.567	ng	100
78) Pyrene	19.221	202	4349870	39.676	ng	100
80) Butylbenzylphthalate	20.162	149	1852458	41.403	ng	93
81) Benzo(a)anthracene	21.003	228	3891993	38.182	ng	100
82) 3,3'-Dichlorobenzidine	20.945	252	1389102	43.555	ng	99
83) Chrysene	21.062	228	3637684	37.614	ng	100
84) Bis(2-ethylhexyl)phtha...	20.956	149	2572281	40.539	ng	98
85) Di-n-octyl phthalate	21.986	149	4548357	42.173	ng	98
87) Indeno(1,2,3-cd)pyrene	26.703	276	3627893	34.449	ng	98
88) Benzo(b)fluoranthene	22.874	252	3826072	39.544	ng	98
89) Benzo(k)fluoranthene	22.927	252	3657262	40.018	ng	99
90) Benzo(a)pyrene	23.586	252	3250485	38.781	ng	98
91) Dibenzo(a,h)anthracene	26.750	278	2934810	34.436	ng	98
92) Benzo(g,h,i)perylene	27.638	276	2982909	33.711	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM082224\
 Data File : BM047308.D
 Acq On : 22 Aug 2024 09:14
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC040


Quant Time: Aug 22 10:59:43 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Aug 11 23:47:58 2024
 Response via : Initial Calibration




Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM082224\
Data File : BM047308.D
Acq On : 22 Aug 2024 09:14
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Quant Time: Aug 22 10:59:43 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sun Aug 11 23:47:58 2024
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.351 min Scan# 71
Delta R.T. -0.024 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:152 Resp: 315372
Ion Ratio Lower Upper
152 100
150 154.8 120.5 180.7
115 54.0 45.5 68.3



#2
1,4-Dioxane
Concen: 35.998 ng
RT: 3.028 min Scan# 24
Delta R.T. -0.006 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14

Tgt Ion: 88 Resp: 263249
Ion Ratio Lower Upper
88 100
58 69.3 58.2 87.4
43 26.7 22.5 33.7









#3
Pyridine
Concen: 34.243 ng
RT: 3.399 min Scan# 81
Delta R.T. -0.012 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion: 79 Resp: 734306
Ion Ratio Lower Upper
79 100
52 62.1 51.2 76.8
51 31.3 26.1 39.1



#4
n-Nitrosodimethylamine
Concen: 32.708 ng
RT: 3.322 min Scan# 74
Delta R.T. -0.006 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14

Tgt Ion: 42 Resp: 306437
Ion Ratio Lower Upper
42 100
74 146.2 112.9 169.3
44 6.6 5.3 7.9









#5
2-Fluorophenol
Concen: 73.882 ng
RT: 5.010 min Scan# 30
Delta R.T. -0.012 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:112 Resp: 1303996
Ion Ratio Lower Upper
112 100
64 55.8 46.7 70.1
63 29.5 24.8 37.2



#6
Aniline
Concen: 35.238 ng
RT: 6.704 min Scan# 649
Delta R.T. -0.024 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14


Tgt Ion: 93 Resp: 809995
Ion Ratio Lower Upper
93 100
66 37.8 31.4 47.2
65 19.5 15.6 23.4








#7
Phenol-d6
Concen: 72.430 ng
RT: 6.569 min Scan# 61
Delta R.T. -0.018 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion: 99 Resp: 1761491




Ion	Ratio	Lower	Upper
99	100		
42	16.2	14.1	21.1
71	34.7	28.5	42.7




#8
2-Chlorophenol
Concen: 37.141 ng
RT: 6.928 min Scan# 687
Delta R.T. -0.024 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14

Tgt Ion: 128 Resp: 714965

Ion	Ratio	Lower	Upper
128	100		
130	32.1	12.1	52.1
64	44.9	26.6	66.6









#9
Benzaldehyde
Concen: 41.837 ng
RT: 6.516 min Scan# 617
Delta R.T. -0.024 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion: 77 Resp: 523783
Ion Ratio Lower Upper
77 100
105 93.8 74.5 114.5
106 93.0 72.3 112.3



#10
Phenol
Concen: 35.576 ng
RT: 6.593 min Scan# 630
Delta R.T. -0.018 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14

Tgt Ion: 94 Resp: 863321
Ion Ratio Lower Upper
94 100
65 27.9 9.0 49.0
66 38.1 18.9 58.9









#11
bis(2-Chloroethyl)ether
Concen: 35.752 ng
RT: 6.804 min Scan# 666
Delta R.T. -0.024 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion: 93 Resp: 742990
Ion Ratio Lower Upper
93 100
63 72.6 55.8 95.8
95 31.8 12.7 52.7



#12
1,3-Dichlorobenzene
Concen: 38.433 ng
RT: 7.240 min Scan# 740
Delta R.T. -0.024 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14

Tgt Ion: 146 Resp: 879161
Ion Ratio Lower Upper
146 100
148 64.0 52.3 78.5
75 29.3 25.8 38.6









#13
1,4-Dichlorobenzene
Concen: 38.275 ng
RT: 7.387 min Scan# 701
Delta R.T. -0.024 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:146 Resp: 886468
Ion Ratio Lower Upper
146 100
148 65.2 51.4 77.2
111 37.6 31.5 47.3



#14
1,2-Dichlorobenzene
Concen: 37.796 ng
RT: 7.698 min Scan# 818
Delta R.T. -0.018 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14

Tgt Ion:146 Resp: 825943
Ion Ratio Lower Upper
146 100
148 64.3 51.8 77.8
111 39.3 33.3 49.9









#15
Benzyl Alcohol
Concen: 35.771 ng
RT: 7.604 min Scan# 806
Delta R.T. -0.024 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion: 79 Resp: 619139
Ion Ratio Lower Upper
79 100
108 75.9 59.7 89.5
77 63.9 52.9 79.3



#16
2,2'-oxybis(1-Chloropropane)
Concen: 33.012 ng
RT: 7.892 min Scan# 851
Delta R.T. -0.024 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14

Tgt Ion: 45 Resp: 880322
Ion Ratio Lower Upper
45 100
77 16.7 0.0 35.6
79 13.1 0.0 32.4









#17
2-Methylphenol
Concen: 36.976 ng
RT: 7.816 min Scan# 81
Delta R.T. -0.018 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:107 Resp: 598616
Ion Ratio Lower Upper
107 100
108 112.4 90.6 135.8
77 47.6 39.3 58.9
79 46.5 38.6 58.0



#18
Hexachloroethane
Concen: 36.929 ng
RT: 8.404 min Scan# 938
Delta R.T. -0.024 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14

Tgt Ion:117 Resp: 331304
Ion Ratio Lower Upper
117 100
119 96.8 77.0 115.4
201 113.5 81.3 121.9









#19
n-Nitroso-di-n-propylamine
Concen: 34.301 ng
RT: 8.169 min Scan# 898
Delta R.T. -0.018 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion: 70 Resp: 561893
Ion Ratio Lower Upper
70 100
42 50.1 42.0 63.0
101 10.1 7.2 10.8
130 22.0 15.5 23.3



#20
3+4-Methylphenols
Concen: 36.179 ng
RT: 8.145 min Scan# 894
Delta R.T. -0.018 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14

Tgt Ion: 107 Resp: 818840
Ion Ratio Lower Upper
107 100
108 90.6 70.7 110.7
77 37.2 17.4 57.4
79 30.1 10.9 50.9









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.110 min Scan# 1232
Delta R.T. -0.024 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:136 Resp: 1211803
Ion Ratio Lower Upper
136 100
137 10.5 8.8 13.2
54 6.6 6.0 9.0
68 5.6 4.8 7.2



#22
Acetophenone
Concen: 37.424 ng
RT: 8.175 min Scan# 899
Delta R.T. -0.024 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14

Tgt Ion:105 Resp: 1094687
Ion Ratio Lower Upper
105 100
71 5.6 4.5 6.7
51 23.5 20.1 30.1
120 22.9 17.8 26.6









#23
Nitrobenzene-d5
Concen: 71.513 ng
RT: 8.504 min Scan# 91
Delta R.T. -0.018 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion: 82 Resp: 1721029
Ion Ratio Lower Upper
82 100
128 43.7 32.3 48.5
54 44.2 37.4 56.2



#24
Nitrobenzene
Concen: 36.519 ng
RT: 8.545 min Scan# 962
Delta R.T. -0.018 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14


Tgt Ion: 77 Resp: 884972
Ion Ratio Lower Upper
77 100
123 45.2 34.0 51.0
65 13.1 10.5 15.7








#25
Isophorone
Concen: 37.971 ng
RT: 9.069 min Scan# 1051
Delta R.T. -0.024 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040







Tgt Ion: 82 Resp: 1591106
Ion Ratio Lower Upper
82 100
95 7.1 5.5 8.3
138 17.8 13.5 20.3



#26
2-Nitrophenol
Concen: 41.863 ng
RT: 9.245 min Scan# 1081
Delta R.T. -0.018 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14

Tgt Ion: 139 Resp: 449274
Ion Ratio Lower Upper
139 100
109 30.2 24.4 36.6
65 47.7 43.0 64.6





#27

2,4-Dimethylphenol

Concen: 39.527 ng

RT: 9.328 min Scan# 1095

Delta R.T. -0.018 min

Lab File: BM047308.D

Acq: 22 Aug 2024 09:14

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040





Tgt Ion:122 Resp: 496910

Ion Ratio Lower Upper

122 100

107 115.2 92.8 139.2

121 58.2 46.1 69.1



#28

bis(2-Chloroethoxy)methane

Concen: 37.664 ng

RT: 9.551 min Scan# 1133

Delta R.T. -0.024 min

Lab File: BM047308.D

Acq: 22 Aug 2024 09:14




Tgt Ion: 93 Resp: 993074


Ion Ratio Lower Upper

93 100

95 32.1 25.8 38.8

123 11.4 9.0 13.4









#29
2,4-Dichlorophenol
Concen: 41.938 ng
RT: 9.775 min Scan# 1175
Delta R.T. -0.024 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:162 Resp: 782273
Ion Ratio Lower Upper
162 100
164 65.0 44.7 84.7
98 32.8 14.9 54.9



#30
1,2,4-Trichlorobenzene
Concen: 41.742 ng
RT: 9.975 min Scan# 1205
Delta R.T. -0.024 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14

Tgt Ion:180 Resp: 881137
Ion Ratio Lower Upper
180 100
182 95.7 76.6 115.0
145 28.6 23.4 35.2









#31
Naphthalene
Concen: 38.156 ng
RT: 10.157 min Scan# 1128
Delta R.T. -0.024 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:128 Resp: 2303393
Ion Ratio Lower Upper
128 100
129 11.0 8.6 13.0
127 13.2 10.2 15.4



#32
Benzoic acid
Concen: 41.350 ng
RT: 9.522 min Scan# 1128
Delta R.T. -0.012 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14

Tgt Ion:122 Resp: 603943
Ion Ratio Lower Upper
122 100
105 125.1 107.1 147.1
77 101.2 82.1 122.1









#33
4-Chloroaniline
Concen: 37.641 ng
RT: 10.286 min Scan# 1258
Delta R.T. -0.024 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:	127	Resp:	873816
Ion	Ratio	Lower	Upper
127	100		
129	32.3	25.8	38.8
65	29.6	24.9	37.3
92	18.8	15.1	22.7



#34
Hexachlorobutadiene
Concen: 44.840 ng
RT: 10.439 min Scan# 1284
Delta R.T. -0.030 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14

Tgt Ion:	225	Resp:	554015
Ion	Ratio	Lower	Upper
225	100		
223	63.1	50.2	75.2
227	64.0	51.4	77.0









#35
Caprolactam
Concen: 40.031 ng
RT: 11.104 min Scan# 11
Delta R.T. -0.018 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:113 Resp: 259311
Ion Ratio Lower Upper
113 100
55 148.1 140.5 180.5
56 113.4 103.6 143.6



#36
4-Chloro-3-methylphenol
Concen: 39.548 ng
RT: 11.439 min Scan# 1454
Delta R.T. -0.012 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14

Tgt Ion:107 Resp: 785639
Ion Ratio Lower Upper
107 100
144 26.8 21.8 32.6
142 83.9 66.8 100.2









#37
2-Methylnaphthalene
Concen: 39.270 ng
RT: 11.786 min Scan# 1513
Delta R.T. -0.018 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:142 Resp: 1687918
Ion Ratio Lower Upper
142 100
141 87.9 70.9 106.3
115 33.5 28.7 43.1



#38
1-Methylnaphthalene
Concen: 38.700 ng
RT: 12.010 min Scan# 1551
Delta R.T. -0.024 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14

Tgt Ion:142 Resp: 1643957
Ion Ratio Lower Upper
142 100
141 91.8 74.6 111.8
116 3.8 3.1 4.7





#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 14.016 min Scan# 11

Delta R.T. -0.018 min

Lab File: BM047308.D

Acq: 22 Aug 2024 09:14

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040





Tgt Ion:164 Resp: 822573

Ion Ratio Lower Upper

164 100

162 99.3 80.7 121.1

160 44.2 36.0 54.0



#40

1,2,4,5-Tetrachlorobenzene

Concen: 42.662 ng

RT: 12.169 min Scan# 1578

Delta R.T. -0.018 min

Lab File: BM047308.D

Acq: 22 Aug 2024 09:14

Tgt Ion:216 Resp: 957174




Ion Ratio Lower Upper


216 100

214 78.6 63.4 95.2

179 20.6 17.6 26.4

108 16.0 15.1 22.7









#41
Hexachlorocyclopentadiene
Concen: 36.329 ng
RT: 12.139 min Scan# 11
Delta R.T. -0.024 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:237 Resp: 284615
Ion Ratio Lower Upper
237 100
235 63.0 44.8 84.8
272 13.1 0.0 33.5



#42
2,4,6-Tribromophenol
Concen: 83.918 ng
RT: 15.527 min Scan# 2149
Delta R.T. -0.012 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14

Tgt Ion:330 Resp: 800401
Ion Ratio Lower Upper
330 100
332 96.8 77.4 116.2
141 34.8 31.6 47.4









#43
2,4,6-Trichlorophenol
Concen: 41.744 ng
RT: 12.427 min Scan# 1622
Delta R.T. -0.018 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion	196	Resp	665686
Ion Ratio	Lower	Upper	
196	100		
198	94.0	75.6	113.4
200	30.6	24.6	37.0



#44
2,4,5-Trichlorophenol
Concen: 41.695 ng
RT: 12.510 min Scan# 1636
Delta R.T. -0.012 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14

Tgt Ion	196	Resp	738243
Ion Ratio	Lower	Upper	
196	100		
198	95.8	77.7	116.5
97	44.0	39.8	59.8
132	24.2	20.7	31.1









#45
2-Fluorobiphenyl
Concen: 77.269 ng
RT: 12.622 min Scan# 1655
Delta R.T. -0.023 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:172 Resp: 4120450
Ion Ratio Lower Upper
172 100
171 34.8 27.8 41.6
170 23.2 18.7 28.1



#46
1,1'-Biphenyl
Concen: 37.912 ng
RT: 12.833 min Scan# 1691
Delta R.T. -0.024 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14

Tgt Ion:154 Resp: 2220122
Ion Ratio Lower Upper
154 100
153 40.0 19.0 59.0
76 11.4 0.0 31.9









#47
2-Chloronaphthalene
Concen: 37.879 ng
RT: 12.869 min Scan# 1
Delta R.T. -0.018 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:162 Resp: 1733753
Ion Ratio Lower Upper
162 100
127 36.1 30.0 45.0
164 32.9 26.3 39.5



#48
2-Nitroaniline
Concen: 36.219 ng
RT: 13.098 min Scan# 1736
Delta R.T. -0.018 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14

Tgt Ion: 65 Resp: 519350
Ion Ratio Lower Upper
65 100
92 69.7 55.1 82.7
138 113.2 85.9 128.9









#49
Acenaphthylene
Concen: 37.959 ng
RT: 13.727 min Scan# 1843
Delta R.T. -0.024 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:152 Resp: 2561019
Ion Ratio Lower Upper
152 100
151 19.7 15.8 23.6
153 13.2 10.4 15.6



#50
Dimethylphthalate
Concen: 38.861 ng
RT: 13.498 min Scan# 1804
Delta R.T. -0.018 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14

Tgt Ion:163 Resp: 2232553
Ion Ratio Lower Upper
163 100
194 4.2 3.0 4.6
164 10.0 8.0 12.0









#51
2,6-Dinitrotoluene
Concen: 39.978 ng
RT: 13.616 min Scan# 1824
Delta R.T. -0.012 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:165 Resp: 536890
Ion Ratio Lower Upper
165 100
63 48.8 46.5 69.7
89 50.6 47.3 70.9



#52
Acenaphthene
Concen: 37.872 ng
RT: 14.080 min Scan# 1903
Delta R.T. -0.018 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14

Tgt Ion:154 Resp: 1620780
Ion Ratio Lower Upper
154 100
153 113.1 90.3 135.5
152 53.3 42.9 64.3









#53
3-Nitroaniline
Concen: 37.966 ng
RT: 13.945 min Scan# 1880
Delta R.T. -0.018 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:138 Resp: 510904
Ion Ratio Lower Upper
138 100
108 9.9 7.9 11.9
92 114.5 92.0 138.0



#54
2,4-Dinitrophenol
Concen: 37.702 ng
RT: 14.169 min Scan# 1918
Delta R.T. -0.006 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14

Tgt Ion:184 Resp: 303462
Ion Ratio Lower Upper
184 100
63 57.0 52.7 79.1
154 60.8 50.4 75.6





#55

Dibenzofuran

Concen: 38.323 ng

RT: 14.421 min Scan# 1964

Delta R.T. -0.018 min

Lab File: BM047308.D

Acq: 22 Aug 2024 09:14

Instrument :





BNA_M

ClientSampleId :

SSTDCCC040

Tgt Ion:168 Resp: 2590656

Ion	Ratio	Lower	Upper
168	100		
139	36.0	29.6	44.4
169	13.2	10.7	16.1



#56

4-Nitrophenol

Concen: 35.815 ng

RT: 14.298 min Scan# 1940




Delta R.T. -0.000 min


Lab File: BM047308.D

Acq: 22 Aug 2024 09:14

Tgt Ion:139 Resp: 415593


Ion	Ratio	Lower	Upper
139	100		
109	71.6	52.5	92.5
65	92.4	75.4	115.4








#57
2,4-Dinitrotoluene
Concen: 39.529 ng
RT: 14.416 min Scan# 1960
Delta R.T. -0.012 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040







Tgt Ion:165 Resp: 708560
Ion Ratio Lower Upper
165 100
63 51.4 40.3 60.5
89 73.0 61.0 91.6
182 2.9 1.9 2.9#



#58
Fluorene
Concen: 38.248 ng
RT: 15.074 min Scan# 2072
Delta R.T. -0.018 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14

Tgt Ion:166 Resp: 2133336
Ion Ratio Lower Upper
166 100
165 96.3 77.4 116.0
167 13.8 11.0 16.6









#59
2,3,4,6-Tetrachlorophenol
Concen: 41.872 ng
RT: 14.663 min Scan# 2002
Delta R.T. -0.012 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:232 Resp: 609310
Ion Ratio Lower Upper
232 100
131 40.6 36.9 55.3
130 2.1 2.0 3.0
166 30.5 26.4 39.6



#60
Diethylphthalate
Concen: 38.138 ng
RT: 14.880 min Scan# 2039
Delta R.T. -0.018 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14

Tgt Ion:149 Resp: 2237417
Ion Ratio Lower Upper
149 100
177 23.5 17.6 26.4
150 12.3 9.8 14.6









#61
4-Chlorophenyl-phenylether
Concen: 40.367 ng
RT: 15.080 min Scan# 2076
Delta R.T. -0.018 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:204 Resp: 1071908
Ion Ratio Lower Upper
204 100
206 33.2 26.2 39.4
141 54.0 47.2 70.8



#62
4-Nitroaniline
Concen: 37.788 ng
RT: 15.127 min Scan# 2081
Delta R.T. -0.012 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14

Tgt Ion:138 Resp: 500966
Ion Ratio Lower Upper
138 100
92 49.6 31.9 71.9
108 78.2 63.8 103.8









#63
Azobenzene
Concen: 35.112 ng
RT: 15.374 min Scan# 2126
Delta R.T. -0.018 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion: 77 Resp: 1953128
Ion Ratio Lower Upper
77 100
182 27.7 3.8 43.8
105 20.4 0.0 39.5
51 25.1 6.8 46.8



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 16.780 min Scan# 2362
Delta R.T. -0.012 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14

Tgt Ion:188 Resp: 1768064
Ion Ratio Lower Upper
188 100
94 8.2 7.0 10.6
80 8.8 8.0 12.0









#65
4,6-Dinitro-2-methylphenol
Concen: 41.084 ng
RT: 15.186 min Scan# 2091
Delta R.T. -0.012 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:198 Resp: 425285
Ion Ratio Lower Upper
198 100
51 37.2 23.9 63.9
105 40.2 24.3 64.3



#66
n-Nitrosodiphenylamine
Concen: 38.026 ng
RT: 15.304 min Scan# 2111
Delta R.T. -0.018 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14

Tgt Ion:169 Resp: 1831258
Ion Ratio Lower Upper
169 100
168 66.2 53.8 80.8
167 36.4 29.3 43.9









#67
4-Bromophenyl-phenylether
Concen: 41.431 ng
RT: 15.980 min Scan# 21
Delta R.T. -0.018 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:248 Resp: 705879
Ion Ratio Lower Upper
248 100
250 96.8 77.8 116.8
141 65.3 57.7 86.5



#68
Hexachlorobenzene
Concen: 40.254 ng
RT: 16.086 min Scan# 2244
Delta R.T. -0.012 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14

Tgt Ion:284 Resp: 821372
Ion Ratio Lower Upper
284 100
142 30.8 29.4 44.0
249 30.8 25.1 37.7









#69
Atrazine
Concen: 35.298 ng
RT: 16.274 min Scan# 21
Delta R.T. -0.018 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:200 Resp: 511524
Ion Ratio Lower Upper
200 100
173 26.0 6.1 46.1
215 47.3 26.8 66.8



#70
Pentachlorophenol
Concen: 40.597 ng
RT: 16.439 min Scan# 2304
Delta R.T. -0.012 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14

Tgt Ion:266 Resp: 580674
Ion Ratio Lower Upper
266 100
268 63.7 49.8 74.6
264 63.2 49.6 74.4









#71
Phenanthrene
Concen: 37.683 ng
RT: 16.821 min Scan# 21
Delta R.T. -0.018 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:178 Resp: 3343698
Ion Ratio Lower Upper
178 100
176 19.0 15.3 22.9
179 15.6 12.4 18.6



#72
Anthracene
Concen: 38.044 ng
RT: 16.910 min Scan# 2384
Delta R.T. -0.018 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14

Tgt Ion:178 Resp: 3284636
Ion Ratio Lower Upper
178 100
176 18.5 14.8 22.2
179 15.5 12.4 18.6









#73
Carbazole
Concen: 37.220 ng
RT: 17.198 min Scan# 2435
Delta R.T. -0.012 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:167 Resp: 3281091
Ion Ratio Lower Upper
167 100
166 21.1 17.1 25.7
139 12.1 10.0 15.0



#74
Di-n-butylphthalate
Concen: 39.097 ng
RT: 17.774 min Scan# 2531
Delta R.T. -0.018 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14

Tgt Ion:149 Resp: 4010258
Ion Ratio Lower Upper
149 100
150 9.3 7.5 11.3
104 4.7 4.1 6.1









#75
Fluoranthene
Concen: 39.236 ng
RT: 18.856 min Scan# 21
Delta R.T. -0.012 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:202 Resp: 4234649
Ion Ratio Lower Upper
202 100
101 8.9 0.0 30.2
203 17.7 0.0 37.6



#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.021 min Scan# 3083
Delta R.T. -0.012 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14

Tgt Ion:240 Resp: 1535454
Ion Ratio Lower Upper
240 100
120 9.2 8.2 12.2
236 25.9 20.6 30.8









#77
Benzidine
Concen: 59.567 ng
RT: 19.062 min Scan# 21
Delta R.T. -0.012 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:184 Resp: 1551050
Ion Ratio Lower Upper
184 100
185 14.5 11.8 17.6
183 11.2 9.0 13.6



#78
Pyrene
Concen: 39.676 ng
RT: 19.221 min Scan# 2777
Delta R.T. -0.018 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14

Tgt Ion:202 Resp: 4349870
Ion Ratio Lower Upper
202 100
200 21.1 17.2 25.8
203 17.7 14.2 21.2









#79
Terphenyl-d14
Concen: 89.037 ng
RT: 19.445 min Scan# 21
Delta R.T. -0.018 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:244 Resp: 6380197
Ion Ratio Lower Upper
244 100
212 7.7 6.2 9.4
122 8.7 7.1 10.7



#80
Butylbenzylphthalate
Concen: 41.403 ng
RT: 20.162 min Scan# 2937
Delta R.T. -0.012 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14

Tgt Ion:149 Resp: 1852458
Ion Ratio Lower Upper
149 100
91 64.1 57.0 85.4
206 22.2 16.4 24.6









#81
Benzo(a)anthracene
Concen: 38.182 ng
RT: 21.003 min Scan# 3080
Delta R.T. -0.012 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:228 Resp: 3891993
Ion Ratio Lower Upper
228 100
226 27.2 21.8 32.6
229 19.7 15.8 23.8



#82
3,3'-Dichlorobenzidine
Concen: 43.555 ng
RT: 20.945 min Scan# 3070
Delta R.T. -0.012 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14

Tgt Ion:252 Resp: 1389102
Ion Ratio Lower Upper
252 100
254 64.5 51.0 76.6
126 9.5 8.5 12.7









#83
Chrysene
Concen: 37.614 ng
RT: 21.062 min Scan# 3092
Delta R.T. -0.012 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:228 Resp: 3637684
Ion Ratio Lower Upper
228 100
226 30.2 24.4 36.6
229 19.7 15.7 23.5



#84
Bis(2-ethylhexyl)phthalate
Concen: 40.539 ng
RT: 20.956 min Scan# 3072
Delta R.T. -0.018 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14

Tgt Ion:149 Resp: 2572281
Ion Ratio Lower Upper
149 100
167 29.0 22.2 33.2
279 4.3 3.0 4.4









#85
Di-n-octyl phthalate
Concen: 42.173 ng
RT: 21.986 min Scan# 31
Delta R.T. -0.024 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:149 Resp: 4548357
Ion Ratio Lower Upper
149 100
167 1.6 1.3 1.9
43 7.6 6.9 10.3



#86
Perylene-d12
Concen: 20.000 ng
RT: 23.709 min Scan# 3540
Delta R.T. -0.018 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14

Tgt Ion:264 Resp: 1629252
Ion Ratio Lower Upper
264 100
260 23.3 18.5 27.7
265 21.9 17.3 25.9









#87
Indeno(1,2,3-cd)pyrene
Concen: 34.449 ng
RT: 26.703 min Scan# 4053
Delta R.T. -0.024 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:276 Resp: 3627893
Ion Ratio Lower Upper
276 100
138 22.1 19.4 29.0
277 25.0 20.1 30.1



#88
Benzo(b)fluoranthene
Concen: 39.544 ng
RT: 22.874 min Scan# 3398
Delta R.T. -0.012 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14

Tgt Ion:252 Resp: 3826072
Ion Ratio Lower Upper
252 100
253 21.6 17.8 26.6
125 8.2 7.4 11.0





#89

Benzo(k)fluoranthene

Concen: 40.018 ng

RT: 22.927 min Scan# 34

Delta R.T. -0.018 min

Lab File: BM047308.D

Acq: 22 Aug 2024 09:14

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040





Tgt Ion:252 Resp: 3657262

Ion Ratio Lower Upper

252 100

253 22.1 17.7 26.5

125 8.4 7.4 11.2



#90

Benzo(a)pyrene

Concen: 38.781 ng

RT: 23.586 min Scan# 3519

Delta R.T. -0.024 min

Lab File: BM047308.D

Acq: 22 Aug 2024 09:14




Tgt Ion:252 Resp: 3250485


Ion Ratio Lower Upper

252 100

253 22.2 17.1 25.7

125 9.6 8.2 12.2









#91
Dibenzo(a,h)anthracene
Concen: 34.436 ng
RT: 26.750 min Scan# 4062
Delta R.T. -0.030 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion:278 Resp: 2934810
Ion Ratio Lower Upper
278 100
139 13.8 12.0 18.0
279 23.7 19.2 28.8



#92
Benzo(g,h,i)perylene
Concen: 33.711 ng
RT: 27.638 min Scan# 4208
Delta R.T. -0.024 min
Lab File: BM047308.D
Acq: 22 Aug 2024 09:14

Tgt Ion:276 Resp: 2982909
Ion Ratio Lower Upper
276 100
277 23.1 18.8 28.2
138 18.8 16.2 24.4



Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM082224\
 Data File : BM047308.D
 Acq On : 22 Aug 2024 09:14
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 22 10:59:43 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Aug 11 23:47:58 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	92	-0.02
2	1,4-Dioxane	0.464	0.417	10.1	87	0.00
3	Pyridine	1.360	1.164	14.4	82	-0.01
4	n-Nitrosodimethylamine	0.594	0.486	18.2	78	0.00
5 S	2-Fluorophenol	1.119	1.034	7.6	89	-0.01
6	Aniline	1.458	1.284	11.9	83	-0.02
7 S	Phenol-d6	1.542	1.396	9.5	87	-0.02
8	2-Chlorophenol	1.221	1.134	7.1	88	-0.02
9	Benzaldehyde	0.904	0.830	8.2	98	-0.02
10 C	Phenol	1.539	1.369	11.0	87	-0.02
11	bis(2-Chloroethyl)ether	1.318	1.178	10.6	84	-0.02
12	1,3-Dichlorobenzene	1.451	1.394	3.9	91	-0.02
13 C	1,4-Dichlorobenzene	1.469	1.405	4.4	90	-0.02
14	1,2-Dichlorobenzene	1.386	1.309	5.6	92	-0.02
15	Benzyl Alcohol	1.098	0.982	10.6	82	-0.02
16	2,2'-oxybis(1-Chloropropane	1.691	1.396	17.4	77	-0.02
17	2-Methylphenol	1.027	0.949	7.6	87	-0.02
18	Hexachloroethane	0.569	0.525	7.7	88	-0.02
19 P	n-Nitroso-di-n-propylamine	1.039	0.891	14.2	83	-0.02
20	3+4-Methylphenols	1.435	1.298	9.5	84	-0.02
21 I	Naphthalene-d8	1.000	1.000	0.0	87	-0.02
22	Acetophenone	0.483	0.452	6.4	85	-0.02
23 S	Nitrobenzene-d5	0.397	0.355	10.6	80	-0.02
24	Nitrobenzene	0.400	0.365	8.8	81	-0.02
25	Isophorone	0.692	0.657	5.1	83	-0.02
26 C	2-Nitrophenol	0.177	0.185	-4.5	89	-0.02
27	2,4-Dimethylphenol	0.207	0.205	1.0	86	-0.02
28	bis(2-Chloroethoxy)methane	0.435	0.410	5.7	82	-0.02
29 C	2,4-Dichlorophenol	0.308	0.323	-4.9	90	-0.02
30	1,2,4-Trichlorobenzene	0.348	0.364	-4.6	92	-0.02
31	Naphthalene	0.996	0.950	4.6	86	-0.02
32	Benzoic acid	0.241	0.249	-3.3	87	-0.01
33	4-Chloroaniline	0.383	0.361	5.7	82	-0.02
34 C	Hexachlorobutadiene	0.204	0.229	-12.3	99	-0.03
35	Caprolactam	0.107	0.107	0.0	86	-0.02
36 C	4-Chloro-3-methylphenol	0.328	0.324	1.2	86	-0.01
37	2-Methylnaphthalene	0.709	0.696	1.8	88	-0.02
38	1-Methylnaphthalene	0.701	0.678	3.3	87	-0.02
39 I	Acenaphthene-d10	1.000	1.000	0.0	90	-0.02
40	1,2,4,5-Tetrachlorobenzene	0.546	0.582	-6.6	96	-0.02
41 P	Hexachlorocyclopentadiene	0.190	0.173	8.9	78	-0.02
42 S	2,4,6-Tribromophenol	0.232	0.243	-4.7	94	-0.01
43 C	2,4,6-Trichlorophenol	0.388	0.405	-4.4	91	-0.02
44	2,4,5-Trichlorophenol	0.430	0.449	-4.4	92	-0.01
45 S	2-Fluorobiphenyl	1.297	1.252	3.5	89	-0.02
46	1,1'-Biphenyl	1.424	1.349	5.3	87	-0.02
47	2-Chloronaphthalene	1.113	1.054	5.3	88	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM082224\
 Data File : BM047308.D
 Acq On : 22 Aug 2024 09:14
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 22 10:59:43 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Aug 11 23:47:58 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.349	0.316	9.5	80	-0.02
49	Acenaphthylene	1.640	1.557	5.1	87	-0.02
50	Dimethylphthalate	1.397	1.357	2.9	88	-0.02
51	2,6-Dinitrotoluene	0.327	0.326	0.3	89	-0.01
52 C	Acenaphthene	1.041	0.985	5.4	88	-0.02
53	3-Nitroaniline	0.327	0.311	4.9	84	-0.02
54 P	2,4-Dinitrophenol	0.196	0.184	6.1	81	0.00
55	Dibenzofuran	1.644	1.575	4.2	90	-0.02
56 P	4-Nitrophenol	0.282	0.253	10.3	77	0.00
57	2,4-Dinitrotoluene	0.436	0.431	1.1	88	-0.01
58	Fluorene	1.356	1.297	4.4	89	-0.02
59	2,3,4,6-Tetrachlorophenol	0.354	0.370	-4.5	94	-0.01
60	Diethylphthalate	1.426	1.360	4.6	87	-0.02
61	4-Chlorophenyl-phenylether	0.646	0.652	-0.9	94	-0.02
62	4-Nitroaniline	0.322	0.305	5.3	86	-0.01
63	Azobenzene	1.352	1.187	12.2	81	-0.02
64 I	Phenanthrene-d10	1.000	1.000	0.0	93	-0.01
65	4,6-Dinitro-2-methylphenol	0.117	0.120	-2.6	89	-0.01
66 c	n-Nitrosodiphenylamine	0.545	0.518	5.0	90	-0.02
67	4-Bromophenyl-phenylether	0.193	0.200	-3.6	97	-0.02
68	Hexachlorobenzene	0.231	0.232	-0.4	94	-0.01
69	Atrazine	0.164	0.145	11.6	79	-0.02
70 C	Pentachlorophenol	0.162	0.164	-1.2	91	-0.01
71	Phenanthrene	1.004	0.946	5.8	90	-0.02
72	Anthracene	0.977	0.929	4.9	91	-0.02
73	Carbazole	0.997	0.928	6.9	89	-0.01
74	Di-n-butylphthalate	1.160	1.134	2.2	91	-0.02
75 C	Fluoranthene	1.221	1.198	1.9	93	-0.01
76 I	Chrysene-d12	1.000	1.000	0.0	92	-0.01
77	Benzydine	0.339	0.505	-49.0#	95	-0.01
78	Pyrene	1.428	1.416	0.8	91	-0.02
79 S	Terphenyl-d14	0.933	1.039	-11.4	102	-0.02
80	Butylbenzylphthalate	0.583	0.603	-3.4	91	-0.01
81	Benzo(a)anthracene	1.328	1.267	4.6	89	-0.01
82	3,3'-Dichlorobenzidine	0.415	0.452	-8.9	94	-0.01
83	Chrysene	1.260	1.185	6.0	89	-0.01
84	Bis(2-ethylhexyl)phthalate	0.826	0.838	-1.5	91	-0.02
85 c	Di-n-octyl phthalate	1.405	1.481	-5.4	93	-0.02
86 I	Perylene-d12	1.000	1.000	0.0	81	-0.02
87	Indeno(1,2,3-cd)pyrene	1.293	1.113	13.9	70	-0.02
88	Benzo(b)fluoranthene	1.188	1.174	1.2	82	-0.01
89	Benzo(k)fluoranthene	1.122	1.122	0.0	83	-0.02
90 C	Benzo(a)pyrene	1.029	0.998	3.0	79	-0.02
91	Dibenzo(a,h)anthracene	1.046	0.901	13.9	71	-0.03
92	Benzo(g,h,i)perylene	1.086	0.915	15.7	69	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM082224\
Data File : BM047308.D
Acq On : 22 Aug 2024 09:14
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC040

Quant Time: Aug 22 10:59:43 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sun Aug 11 23:47:58 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM082224\
 Data File : BM047308.D
 Acq On : 22 Aug 2024 09:14
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 22 10:59:43 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Aug 11 23:47:58 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	92	-0.02
2	1,4-Dioxane	40.000	35.998	10.0	87	0.00
3	Pyridine	40.000	34.243	14.4	82	-0.01
4	n-Nitrosodimethylamine	40.000	32.708	18.2	78	0.00
5 S	2-Fluorophenol	80.000	73.882	7.6	89	-0.01
6	Aniline	40.000	35.238	11.9	83	-0.02
7 S	Phenol-d6	80.000	72.430	9.5	87	-0.02
8	2-Chlorophenol	40.000	37.141	7.1	88	-0.02
9	Benzaldehyde	40.000	41.837	-4.6	98	-0.02
10 C	Phenol	40.000	35.576	11.1	87	-0.02
11	bis(2-Chloroethyl)ether	40.000	35.752	10.6	84	-0.02
12	1,3-Dichlorobenzene	40.000	38.433	3.9	91	-0.02
13 C	1,4-Dichlorobenzene	40.000	38.275	4.3	90	-0.02
14	1,2-Dichlorobenzene	40.000	37.796	5.5	92	-0.02
15	Benzyl Alcohol	40.000	35.771	10.6	82	-0.02
16	2,2'-oxybis(1-Chloropropane	40.000	33.012	17.5	77	-0.02
17	2-Methylphenol	40.000	36.976	7.6	87	-0.02
18	Hexachloroethane	40.000	36.929	7.7	88	-0.02
19 P	n-Nitroso-di-n-propylamine	40.000	34.301	14.2	83	-0.02
20	3+4-Methylphenols	40.000	36.179	9.6	84	-0.02
21 I	Naphthalene-d8	20.000	20.000	0.0	87	-0.02
22	Acetophenone	40.000	37.424	6.4	85	-0.02
23 S	Nitrobenzene-d5	80.000	71.513	10.6	80	-0.02
24	Nitrobenzene	40.000	36.519	8.7	81	-0.02
25	Isophorone	40.000	37.971	5.1	83	-0.02
26 C	2-Nitrophenol	40.000	41.863	-4.7	89	-0.02
27	2,4-Dimethylphenol	40.000	39.527	1.2	86	-0.02
28	bis(2-Chloroethoxy)methane	40.000	37.664	5.8	82	-0.02
29 C	2,4-Dichlorophenol	40.000	41.938	-4.8	90	-0.02
30	1,2,4-Trichlorobenzene	40.000	41.742	-4.4	92	-0.02
31	Naphthalene	40.000	38.156	4.6	86	-0.02
32	Benzoic acid	40.000	41.350	-3.4	87	-0.01
33	4-Chloroaniline	40.000	37.641	5.9	82	-0.02
34 C	Hexachlorobutadiene	40.000	44.840	-12.1	99	-0.03
35	Caprolactam	40.000	40.031	-0.1	86	-0.02
36 C	4-Chloro-3-methylphenol	40.000	39.548	1.1	86	-0.01
37	2-Methylnaphthalene	40.000	39.270	1.8	88	-0.02
38	1-Methylnaphthalene	40.000	38.700	3.2	87	-0.02
39 I	Acenaphthene-d10	20.000	20.000	0.0	90	-0.02
40	1,2,4,5-Tetrachlorobenzene	40.000	42.662	-6.7	96	-0.02
41 P	Hexachlorocyclopentadiene	40.000	36.329	9.2	78	-0.02
42 S	2,4,6-Tribromophenol	80.000	83.918	-4.9	94	-0.01
43 C	2,4,6-Trichlorophenol	40.000	41.744	-4.4	91	-0.02
44	2,4,5-Trichlorophenol	40.000	41.695	-4.2	92	-0.01
45 S	2-Fluorobiphenyl	80.000	77.269	3.4	89	-0.02
46	1,1'-Biphenyl	40.000	37.912	5.2	87	-0.02
47	2-Chloronaphthalene	40.000	37.879	5.3	88	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM082224\
 Data File : BM047308.D
 Acq On : 22 Aug 2024 09:14
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 22 10:59:43 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Aug 11 23:47:58 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	36.219	9.5	80	-0.02
49	Acenaphthylene	40.000	37.959	5.1	87	-0.02
50	Dimethylphthalate	40.000	38.861	2.8	88	-0.02
51	2,6-Dinitrotoluene	40.000	39.978	0.1	89	-0.01
52 C	Acenaphthene	40.000	37.872	5.3	88	-0.02
53	3-Nitroaniline	40.000	37.966	5.1	84	-0.02
54 P	2,4-Dinitrophenol	40.000	37.702	5.7	81	0.00
55	Dibenzofuran	40.000	38.323	4.2	90	-0.02
56 P	4-Nitrophenol	40.000	35.815	10.5	77	0.00
57	2,4-Dinitrotoluene	40.000	39.529	1.2	88	-0.01
58	Fluorene	40.000	38.248	4.4	89	-0.02
59	2,3,4,6-Tetrachlorophenol	40.000	41.872	-4.7	94	-0.01
60	Diethylphthalate	40.000	38.138	4.7	87	-0.02
61	4-Chlorophenyl-phenylether	40.000	40.367	-0.9	94	-0.02
62	4-Nitroaniline	40.000	37.788	5.5	86	-0.01
63	Azobenzene	40.000	35.112	12.2	81	-0.02
64 I	Phenanthrene-d10	20.000	20.000	0.0	93	-0.01
65	4,6-Dinitro-2-methylphenol	40.000	41.084	-2.7	89	-0.01
66 c	n-Nitrosodiphenylamine	40.000	38.026	4.9	90	-0.02
67	4-Bromophenyl-phenylether	40.000	41.431	-3.6	97	-0.02
68	Hexachlorobenzene	40.000	40.254	-0.6	94	-0.01
69	Atrazine	40.000	35.298	11.8	79	-0.02
70 C	Pentachlorophenol	40.000	40.597	-1.5	91	-0.01
71	Phenanthrene	40.000	37.683	5.8	90	-0.02
72	Anthracene	40.000	38.044	4.9	91	-0.02
73	Carbazole	40.000	37.220	7.0	89	-0.01
74	Di-n-butylphthalate	40.000	39.097	2.3	91	-0.02
75 C	Fluoranthene	40.000	39.236	1.9	93	-0.01
76 I	Chrysene-d12	20.000	20.000	0.0	92	-0.01
77	Benzidine	40.000	59.567	-48.9#	95	-0.01
78	Pyrene	40.000	39.676	0.8	91	-0.02
79 S	Terphenyl-d14	80.000	89.037	-11.3	102	-0.02
80	Butylbenzylphthalate	40.000	41.403	-3.5	91	-0.01
81	Benzo(a)anthracene	40.000	38.182	4.5	89	-0.01
82	3,3'-Dichlorobenzidine	40.000	43.555	-8.9	94	-0.01
83	Chrysene	40.000	37.614	6.0	89	-0.01
84	Bis(2-ethylhexyl)phthalate	40.000	40.539	-1.3	91	-0.02
85 c	Di-n-octyl phthalate	40.000	42.173	-5.4	93	-0.02
86 I	Perylene-d12	20.000	20.000	0.0	81	-0.02
87	Indeno(1,2,3-cd)pyrene	40.000	34.449	13.9	70	-0.02
88	Benzo(b)fluoranthene	40.000	39.544	1.1	82	-0.01
89	Benzo(k)fluoranthene	40.000	40.018	-0.0	83	-0.02
90 C	Benzo(a)pyrene	40.000	38.781	3.0	79	-0.02
91	Dibenzo(a,h)anthracene	40.000	34.436	13.9	71	-0.03
92	Benzo(g,h,i)perylene	40.000	33.711	15.7	69	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM082224\
Data File : BM047308.D
Acq On : 22 Aug 2024 09:14
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC040

Quant Time: Aug 22 10:59:43 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sun Aug 11 23:47:58 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range SPCC's out = 0 CCC's out = 0





QC SAMPLE DATA

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081024\
Data File : BM047139.D
Acq On : 10 Aug 2024 12:26
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081024.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Sun Aug 11 23:47:58 2024



AutoFind: Scans 2444, 2445, 2446; Background Corrected with Scan 2437

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	35.7	125256	PASS
68	69	0.00	2	1.3	2089	PASS
69	198	0.00	100	44.5	156224	PASS
70	69	0.00	2	0.5	803	PASS
127	198	10	80	49.9	174976	PASS
197	198	0.00	2	0.4	1460	PASS
198	198	100	100	100.0	350933	PASS
199	198	5	9	7.0	24517	PASS
275	198	10	60	25.8	90373	PASS
365	198	1	100	3.6	12789	PASS
441	198	0.01	100	9.2	32133	PASS
442	442	50	100	100.0	203891	PASS
443	442	15	24	19.7	40160	PASS

Instrument :

BNA_M

ClientSampleId :

DFTPP




DDT Breakdown

Date	Instrument Name	DFTPP Data File
8/10/2024	BNA_M	<u>BM047139.D</u>
Compound Name	Response	Retention Time
DDT	1324125	20.321
DDD	17903	19.927
DDE	740	19.362
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
18643	1342768	1.39

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081024\
Data File : BM047139.D
Acq On : 10 Aug 2024 12:26
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP

Quant Time: Aug 12 01:52:09 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sun Aug 11 23:47:58 2024
Response via : Initial Calibration



TIC: BM047139.D\data.ms

(70) Pentachlorophenol (C)

16.451min (-0.000) 14982.60 ng




response 465922

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	65.50
264.00	62.00	63.60
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081024\
Data File : BM047139.D
Acq On : 10 Aug 2024 12:26
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP

Quant Time: Aug 12 01:52:09 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sun Aug 11 23:47:58 2024
Response via : Initial Calibration



TIC: BM047139.D\data.ms

(77) Benzidine

19.074min (-0.000) 6495.62 ng

response 2296747



Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.70	14.34
183.00	11.30	11.36
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081924\
Data File : BM047244.D
Acq On : 19 Aug 2024 09:03
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081024.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Sun Aug 11 23:47:58 2024



AutoFind: Scans 2443, 2444, 2445; Background Corrected with Scan 2436

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	28.3	100766	PASS
68	69	0.00	2	1.4	1739	PASS
69	198	0.00	100	35.6	126854	PASS
70	69	0.00	2	0.3	398	PASS
127	198	10	80	42.3	150684	PASS
197	198	0.00	2	0.4	1439	PASS
198	198	100	100	100.0	356095	PASS
199	198	5	9	6.8	24112	PASS
275	198	10	60	25.6	91270	PASS
365	198	1	100	3.4	12027	PASS
441	198	0.01	100	9.1	32381	PASS
442	442	50	100	100.0	209880	PASS
443	442	15	24	19.8	41587	PASS

Instrument :
BNA_M
ClientSampleId :
DFTPP




DDT Breakdown

Date	Instrument Name	DFTPP Data File
8/19/2024	BNA_M	<u>BM047244.D</u>
Compound Name	Response	Retention Time
DDT	1140644	20.321
DDD	107172	19.88
DDE	3211	19.368
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
110383	1251027	8.82

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081924\
Data File : BM047244.D
Acq On : 19 Aug 2024 09:03
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP

Quant Time: Aug 19 11:15:50 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sun Aug 11 23:47:58 2024
Response via : Initial Calibration



TIC: BM047244.D\data.ms

(70) Pentachlorophenol (C)

16.457min (+ 0.006) 216859.29 ng




response 499995

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	64.77
264.00	62.00	62.88
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081924\
Data File : BM047244.D
Acq On : 19 Aug 2024 09:03
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP

Quant Time: Aug 19 11:15:50 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sun Aug 11 23:47:58 2024
Response via : Initial Calibration



TIC: BM047244.D\data.ms

(77) Benzidine

19.074min (-0.000) 40695.55 ng m

response 2114570



Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.70	14.26
183.00	11.30	11.34
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM082224\
Data File : BM047307.D
Acq On : 22 Aug 2024 08:35
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081024.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Sun Aug 11 23:47:58 2024



AutoFind: Scans 2441, 2442, 2443; Background Corrected with Scan 2434

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	25.4	193181	PASS
68	69	0.00	2	1.5	3779	PASS
69	198	0.00	100	32.2	244634	PASS
70	69	0.00	2	0.4	1038	PASS
127	198	10	80	39.1	297621	PASS
197	198	0.00	2	0.3	2079	PASS
198	198	100	100	100.0	760427	PASS
199	198	5	9	6.9	52152	PASS
275	198	10	60	26.2	199595	PASS
365	198	1	100	3.2	24227	PASS
441	198	0.01	100	9.3	70976	PASS
442	442	50	100	100.0	454208	PASS
443	442	15	24	19.2	87264	PASS

Instrument :

BNA_M

ClientSampleId :

DFTPP




DDT Breakdown

Date	Instrument Name	DFTPP Data File
8/22/2024	BNA_M	<u>BM047307.D</u>
Compound Name	Response	Retention Time
DDT	2599731	20.309
DDD	150494	19.868
DDE	4792	19.351
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
155286	2755017	5.64

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM082224\
Data File : BM047307.D
Acq On : 22 Aug 2024 08:35
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP

Quant Time: Aug 22 10:58:16 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sun Aug 11 23:47:58 2024
Response via : Initial Calibration



TIC: BM047307.D\data.ms

(70) Pentachlorophenol (C)

16.439min (-0.012) 305697.00 ng




response 1016426

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	63.44
264.00	62.00	63.25
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM082224\
Data File : BM047307.D
Acq On : 22 Aug 2024 08:35
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP

Quant Time: Aug 22 10:58:16 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sun Aug 11 23:47:58 2024
Response via : Initial Calibration



TIC: BM047307.D\data.ms

(77) Benzidine**19.068min (-0.006) 142115.28 ng****response 5191261**

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.70	14.05
183.00	11.30	11.27
0.00	0.00	0.00

Report of Analysis

Client:	JACOBS Engineering Group, Inc.	Date Collected:	
Project:	Former Schlumberger Site Princeton NJ	Date Received:	
Client Sample ID:	PB162757BL	SDG No.:	P3609
Lab Sample ID:	PB162757BL	Matrix:	Water
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group6
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BM047313.D	1	08/15/24 08:23	08/22/24 12:32	PB162757

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	1.60	U	1.60	5.00	ug/L
100-52-7	Benzaldehyde	4.00	U	4.00	10.0	ug/L
95-48-7	2-Methylphenol	1.10	U	1.10	5.00	ug/L
65794-96-9	3+4-Methylphenols	1.20	U	1.20	10.0	ug/L
67-72-1	Hexachloroethane	1.00	U	1.00	5.00	ug/L
98-95-3	Nitrobenzene	1.30	U	1.30	5.00	ug/L
91-20-3	Naphthalene	1.00	U	1.00	5.00	ug/L
87-68-3	Hexachlorobutadiene	1.30	U	1.30	5.00	ug/L
91-57-6	2-Methylnaphthalene	1.10	U	1.10	5.00	ug/L
88-06-2	2,4,6-Trichlorophenol	0.89	U	0.89	5.00	ug/L
95-95-4	2,4,5-Trichlorophenol	1.00	U	1.00	5.00	ug/L
208-96-8	Acenaphthylene	1.00	U	1.00	5.00	ug/L
83-32-9	Acenaphthene	0.81	U	0.81	5.00	ug/L
132-64-9	Dibenzofuran	0.93	U	0.93	5.00	ug/L
121-14-2	2,4-Dinitrotoluene	1.50	U	1.50	5.00	ug/L
86-73-7	Fluorene	0.96	U	0.96	5.00	ug/L
118-74-1	Hexachlorobenzene	1.10	U	1.10	5.00	ug/L
87-86-5	Pentachlorophenol	1.90	U	1.90	10.0	ug/L
85-01-8	Phenanthrene	0.89	U	0.89	5.00	ug/L
120-12-7	Anthracene	1.10	U	1.10	5.00	ug/L
86-74-8	Carbazole	1.20	U	1.20	5.00	ug/L
84-74-2	Di-n-butylphthalate	1.50	U	1.50	5.00	ug/L
206-44-0	Fluoranthene	1.30	U	1.30	5.00	ug/L
129-00-0	Pyrene	1.10	U	1.10	5.00	ug/L
56-55-3	Benzo(a)anthracene	0.94	U	0.94	5.00	ug/L
218-01-9	Chrysene	0.86	U	0.86	5.00	ug/L
117-81-7	Bis(2-ethylhexyl)phthalate	1.90	U	1.90	5.00	ug/L
205-99-2	Benzo(b)fluoranthene	1.10	U	1.10	5.00	ug/L
207-08-9	Benzo(k)fluoranthene	1.20	U	1.20	5.00	ug/L

Report of Analysis

Client:	JACOBS Engineering Group, Inc.	Date Collected:	
Project:	Former Schlumberger Site Princeton NJ	Date Received:	
Client Sample ID:	PB162757BL	SDG No.:	P3609
Lab Sample ID:	PB162757BL	Matrix:	Water
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group6
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BM047313.D	1	08/15/24 08:23	08/22/24 12:32	PB162757

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
50-32-8	Benzo(a)pyrene	1.70	U	1.70	5.00	ug/L
193-39-5	Indeno(1,2,3-cd)pyrene	1.00	U	1.00	5.00	ug/L
53-70-3	Dibenzo(a,h)anthracene	1.20	U	1.20	5.00	ug/L
191-24-2	Benzo(g,h,i)perylene	1.20	U	1.20	5.00	ug/L
123-91-1	1,4-Dioxane	1.30	U	1.30	5.00	ug/L
90-12-0	1-Methylnaphthalene	0.86	U	0.86	5.00	ug/L

SURROGATES

367-12-4	2-Fluorophenol	120		15 (10) - 110 (139)	80%	SPK: 150
13127-88-3	Phenol-d6	111		15 (10) - 110 (134)	74%	SPK: 150
4165-60-0	Nitrobenzene-d5	75.0		30 (49) - 130 (133)	75%	SPK: 100
321-60-8	2-Fluorobiphenyl	79.1		30 (52) - 130 (132)	79%	SPK: 100
118-79-6	2,4,6-Tribromophenol	131		15 (44) - 110 (137)	88%	SPK: 150
1718-51-0	Terphenyl-d14	96.4		30 (48) - 130 (125)	96%	SPK: 100

INTERNAL STANDARDS

3855-82-1	1,4-Dichlorobenzene-d4	293000	7.351
1146-65-2	Naphthalene-d8	1110000	10.104
15067-26-2	Acenaphthene-d10	772000	14.01
1517-22-2	Phenanthrene-d10	1720000	16.774
1719-03-5	Chrysene-d12	1580000	21.015
1520-96-3	Perylene-d12	1530000	23.709

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM082224\
Data File : BM047313.D
Acq On : 22 Aug 2024 12:32
Operator : RC/JU
Sample : PB162757BL
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
PB162757BL

Quant Time: Aug 22 13:38:33 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sun Aug 11 23:47:58 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.351	152	292730	20.000	ng	-0.02
21) Naphthalene-d8	10.104	136	1111486	20.000	ng	-0.03
39) Acenaphthene-d10	14.010	164	771992	20.000	ng	-0.02
64) Phenanthrene-d10	16.774	188	1724658	20.000	ng	-0.02
76) Chrysene-d12	21.015	240	1583438	20.000	ng	-0.02
86) Perylene-d12	23.709	264	1534865	20.000	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.010	112	1962459	119.790	ng	-0.01
7) Phenol-d6	6.569	99	2501470	110.813	ng	-0.02
23) Nitrobenzene-d5	8.504	82	1655941	75.019	ng	-0.02
42) 2,4,6-Tribromophenol	15.527	330	1174894	131.253	ng	-0.01
45) 2-Fluorobiphenyl	12.621	172	3958679	79.099	ng	-0.02
79) Terphenyl-d14	19.445	244	7120475	96.357	ng	-0.02


Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM082224\
Data File : BM047313.D
Acq On : 22 Aug 2024 12:32
Operator : RC/JU
Sample : PB162757BL
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
PB162757BL


Quant Time: Aug 22 13:38:33 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sun Aug 11 23:47:58 2024
Response via : Initial Calibration




Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM082224\
Data File : BM047313.D
Acq On : 22 Aug 2024 12:32
Operator : RC/JU
Sample : PB162757BL
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
PB162757BL

Quant Time: Aug 22 13:38:33 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sun Aug 11 23:47:58 2024
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.351 min Scan# 71
Delta R.T. -0.024 min
Lab File: BM047313.D
Acq: 22 Aug 2024 12:32




Instrument :
BNA_M
ClientSampleId :
PB162757BL


Tgt Ion:152 Resp: 292730
Ion Ratio Lower Upper
152 100
150 157.4 120.5 180.7
115 54.1 45.5 68.3



#5
2-Fluorophenol
Concen: 119.790 ng
RT: 5.010 min Scan# 361
Delta R.T. -0.012 min
Lab File: BM047313.D
Acq: 22 Aug 2024 12:32


Tgt Ion:112 Resp: 1962459
Ion Ratio Lower Upper
112 100
64 56.7 46.7 70.1
63 29.8 24.8 37.2








#7
Phenol-d6
Concen: 110.813 ng
RT: 6.569 min Scan# 61
Delta R.T. -0.018 min
Lab File: BM047313.D
Acq: 22 Aug 2024 12:32

Instrument :
BNA_M
ClientSampleId :
PB162757BL



Tgt Ion: 99 Resp: 2501470




Ion	Ratio	Lower	Upper
99	100		
42	16.4	14.1	21.1
71	35.2	28.5	42.7




#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.104 min Scan# 1227
Delta R.T. -0.030 min
Lab File: BM047313.D
Acq: 22 Aug 2024 12:32

Tgt Ion: 136 Resp: 1111486

Ion	Ratio	Lower	Upper
136	100		
137	10.6	8.8	13.2
54	6.9	6.0	9.0
68	5.8	4.8	7.2









#23
Nitrobenzene-d5
Concen: 75.019 ng
RT: 8.504 min Scan# 91
Delta R.T. -0.018 min
Lab File: BM047313.D
Acq: 22 Aug 2024 12:32




Instrument :
BNA_M
ClientSampleId :
PB162757BL


Tgt Ion: 82 Resp: 1655941
Ion Ratio Lower Upper
82 100
128 43.4 32.3 48.5
54 43.5 37.4 56.2



#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.010 min Scan# 1891
Delta R.T. -0.024 min
Lab File: BM047313.D
Acq: 22 Aug 2024 12:32

Tgt Ion: 164 Resp: 771992
Ion Ratio Lower Upper
164 100
162 100.3 80.7 121.1
160 44.8 36.0 54.0









#42
2,4,6-Tribromophenol
Concen: 131.253 ng
RT: 15.527 min Scan# 2151
Delta R.T. -0.012 min
Lab File: BM047313.D
Acq: 22 Aug 2024 12:32




Instrument :
BNA_M
ClientSampleId :
PB162757BL


Tgt Ion:330 Resp: 1174894
Ion Ratio Lower Upper
330 100
332 95.8 77.4 116.2
141 35.7 31.6 47.4



#45
2-Fluorobiphenyl
Concen: 79.099 ng
RT: 12.621 min Scan# 1655
Delta R.T. -0.024 min
Lab File: BM047313.D
Acq: 22 Aug 2024 12:32

Tgt Ion:172 Resp: 3958679
Ion Ratio Lower Upper
172 100
171 34.9 27.8 41.6
170 23.3 18.7 28.1









#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 16.774 min Scan# 21
Delta R.T. -0.018 min
Lab File: BM047313.D
Acq: 22 Aug 2024 12:32




Instrument :
BNA_M
ClientSampleId :
PB162757BL


Tgt Ion:188 Resp: 1724658
Ion Ratio Lower Upper
188 100
94 8.4 7.0 10.6
80 9.4 8.0 12.0



#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.015 min Scan# 3082
Delta R.T. -0.018 min
Lab File: BM047313.D
Acq: 22 Aug 2024 12:32

Tgt Ion:240 Resp: 1583438
Ion Ratio Lower Upper
240 100
120 9.9 8.2 12.2
236 25.5 20.6 30.8





#79

Terphenyl-d14

Concen: 96.357 ng

RT: 19.445 min Scan# 216

Delta R.T. -0.018 min

Lab File: BM047313.D

Acq: 22 Aug 2024 12:32

Instrument :

BNA_M

ClientSampleId :

PB162757BL





Tgt Ion:244 Resp: 7120475

Ion Ratio Lower Upper

244 100

212 7.8 6.2 9.4

122 8.7 7.1 10.7



#86

Perylene-d12

Concen: 20.000 ng

RT: 23.709 min Scan# 3540

Delta R.T. -0.018 min

Lab File: BM047313.D

Acq: 22 Aug 2024 12:32




Tgt Ion:264 Resp: 1534865

Ion Ratio Lower Upper

264 100

260 23.0 18.5 27.7

265 22.0 17.3 25.9



Report of Analysis

Client:	JACOBS Engineering Group, Inc.	Date Collected:	
Project:	Former Schlumberger Site Princeton NJ	Date Received:	
Client Sample ID:	PB162757BS	SDG No.:	P3609
Lab Sample ID:	PB162757BS	Matrix:	Water
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group6
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BM047250.D	1	08/15/24 08:23	08/19/24 13:01	PB162757

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	29.0		1.60	5.00	ug/L
100-52-7	Benzaldehyde	16.9		4.00	10.0	ug/L
95-48-7	2-Methylphenol	49.9		1.10	5.00	ug/L
65794-96-9	3+4-Methylphenols	49.6		1.20	10.0	ug/L
67-72-1	Hexachloroethane	41.4		1.00	5.00	ug/L
98-95-3	Nitrobenzene	41.1		1.30	5.00	ug/L
91-20-3	Naphthalene	43.6		1.00	5.00	ug/L
87-68-3	Hexachlorobutadiene	44.7		1.30	5.00	ug/L
91-57-6	2-Methylnaphthalene	45.4		1.10	5.00	ug/L
88-06-2	2,4,6-Trichlorophenol	49.9		0.89	5.00	ug/L
95-95-4	2,4,5-Trichlorophenol	47.2		1.00	5.00	ug/L
208-96-8	Acenaphthylene	47.6		1.00	5.00	ug/L
83-32-9	Acenaphthene	43.1		0.81	5.00	ug/L
132-64-9	Dibenzofuran	45.4		0.93	5.00	ug/L
121-14-2	2,4-Dinitrotoluene	47.7		1.50	5.00	ug/L
86-73-7	Fluorene	45.1		0.96	5.00	ug/L
118-74-1	Hexachlorobenzene	44.7		1.10	5.00	ug/L
87-86-5	Pentachlorophenol	91.2	E	1.90	10.0	ug/L
85-01-8	Phenanthrene	45.8		0.89	5.00	ug/L
120-12-7	Anthracene	47.2		1.10	5.00	ug/L
86-74-8	Carbazole	45.9		1.20	5.00	ug/L
84-74-2	Di-n-butylphthalate	46.3		1.50	5.00	ug/L
206-44-0	Fluoranthene	47.7		1.30	5.00	ug/L
129-00-0	Pyrene	44.0		1.10	5.00	ug/L
56-55-3	Benzo(a)anthracene	47.3		0.94	5.00	ug/L
218-01-9	Chrysene	46.2		0.86	5.00	ug/L
117-81-7	Bis(2-ethylhexyl)phthalate	45.3		1.90	5.00	ug/L
205-99-2	Benzo(b)fluoranthene	47.8		1.10	5.00	ug/L
207-08-9	Benzo(k)fluoranthene	47.9		1.20	5.00	ug/L

Report of Analysis

Client:	JACOBS Engineering Group, Inc.	Date Collected:	
Project:	Former Schlumberger Site Princeton NJ	Date Received:	
Client Sample ID:	PB162757BS	SDG No.:	P3609
Lab Sample ID:	PB162757BS	Matrix:	Water
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group6
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BM047250.D	1	08/15/24 08:23	08/19/24 13:01	PB162757

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
50-32-8	Benzo(a)pyrene	52.1		1.70	5.00	ug/L
193-39-5	Indeno(1,2,3-cd)pyrene	52.5		1.00	5.00	ug/L
53-70-3	Dibenzo(a,h)anthracene	52.5		1.20	5.00	ug/L
191-24-2	Benzo(g,h,i)perylene	48.7		1.20	5.00	ug/L
123-91-1	1,4-Dioxane	26.1		1.30	5.00	ug/L
90-12-0	1-Methylnaphthalene	42.9		0.86	5.00	ug/L

SURROGATES

367-12-4	2-Fluorophenol	134		15 (10) - 110 (139)	89%	SPK: 150
13127-88-3	Phenol-d6	131		15 (10) - 110 (134)	87%	SPK: 150
4165-60-0	Nitrobenzene-d5	78.7		30 (49) - 130 (133)	79%	SPK: 100
321-60-8	2-Fluorobiphenyl	80.7		30 (52) - 130 (132)	81%	SPK: 100
118-79-6	2,4,6-Tribromophenol	141		15 (44) - 110 (137)	94%	SPK: 150
1718-51-0	Terphenyl-d14	95.4		30 (48) - 130 (125)	95%	SPK: 100

INTERNAL STANDARDS

3855-82-1	1,4-Dichlorobenzene-d4	201000	7.363
1146-65-2	Naphthalene-d8	840000	10.122
15067-26-2	Acenaphthene-d10	602000	14.027
1517-22-2	Phenanthrene-d10	1360000	16.792
1719-03-5	Chrysene-d12	1330000	21.039
1520-96-3	Perylene-d12	1610000	23.738

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081924\
 Data File : BM047250.D
 Acq On : 19 Aug 2024 13:01
 Operator : RC/JU
 Sample : PB162757BS
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :

BNA_M

ClientSampleId :

PB162757BS

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 08/20/2024

Supervised By :mohammad ahmed 08/21/2024

Quant Time: Aug 19 13:38:34 2024

Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081024.M

Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

QLast Update : Sun Aug 11 23:47:58 2024

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.363	152	200715	20.000	ng	-0.01
21) Naphthalene-d8	10.122	136	840423	20.000	ng	-0.01
39) Acenaphthene-d10	14.027	164	601718	20.000	ng	0.00
64) Phenanthrene-d10	16.792	188	1358929	20.000	ng	0.00
76) Chrysene-d12	21.039	240	1331978	20.000	ng	0.00
86) Perylene-d12	23.738	264	1613228	20.000	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.022	112	1500748	133.602	ng	0.00
7) Phenol-d6	6.587	99	2024398	130.792	ng	0.00
23) Nitrobenzene-d5	8.516	82	1313241	78.682	ng	0.00
42) 2,4,6-Tribromophenol	15.539	330	983835	141.011	ng	0.00
45) 2-Fluorobiphenyl	12.633	172	3148181	80.705	ng	-0.01
79) Terphenyl-d14	19.462	244	5929827	95.394	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.022	88	121620	26.131	ng	99
3) Pyridine	3.399	79	396398	29.045	ng	97
4) n-Nitrosodimethylamine	3.316	42	216424	36.296	ng	99
6) Aniline	6.710	93	406471	27.784	ng	98
8) 2-Chlorophenol	6.945	128	628413	51.293	ng	96
9) Benzaldehyde	6.528	77	171778	16.936	ng	100
10) Phenol	6.610	94	748552	48.467	ng	97
11) bis(2-Chloroethyl)ether	6.816	93	560447	42.374	ng	98
12) 1,3-Dichlorobenzene	7.257	146	618674	42.495	ng	97
13) 1,4-Dichlorobenzene	7.398	146	636233	43.163	ng	98
14) 1,2-Dichlorobenzene	7.704	146	629748	45.279	ng	98
15) Benzyl Alcohol	7.616	79	565608	51.345	ng	98
16) 2,2'-oxybis(1-Chloropr...	7.886	45	712313	41.970	ng	99
17) 2-Methylphenol	7.828	107	514304	49.915	ng	99
18) Hexachloroethane	8.416	117	236430	41.408	ng	94
19) n-Nitroso-di-n-propyla...	8.175	70	449654	43.129	ng	99
20) 3+4-Methylphenols	8.157	107	714481	49.601	ng	99
22) Acetophenone	8.186	105	796460	39.261	ng	98
24) Nitrobenzene	8.557	77	691067	41.119	ng	97
25) Isophorone	9.081	82	1305301	44.916	ng	98
26) 2-Nitrophenol	9.257	139	364187	48.930	ng	94
27) 2,4-Dimethylphenol	9.339	122	472416	54.185	ng	99
28) bis(2-Chloroethoxy)met...	9.563	93	802839	43.905	ng	99
29) 2,4-Dichlorophenol	9.792	162	686506	53.068	ng	98
30) 1,2,4-Trichlorobenzene	9.992	180	665098	45.431	ng	99
31) Naphthalene	10.175	128	1823534	43.556	ng	100
32) Benzoic acid	9.528	122	514032	50.747	ng	97
33) 4-Chloroaniline	10.298	127	437665	27.184	ng	99
34) Hexachlorobutadiene	10.457	225	383339	44.736	ng	99
35) Caprolactam	11.104	113	221498m	49.303	ng	
36) 4-Chloro-3-methylphenol	11.451	107	689653	50.057	ng	98
37) 2-Methylnaphthalene	11.798	142	1352299	45.365	ng	98
38) 1-Methylnaphthalene	12.022	142	1262961	42.869	ng	98
40) 1,2,4,5-Tetrachloroben...	12.180	216	704034	42.897	ng	99
41) Hexachlorocyclopentadiene	12.157	237	514477	89.772	ng	97
43) 2,4,6-Trichlorophenol	12.445	196	581555	49.854	ng	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081924\
 Data File : BM047250.D
 Acq On : 19 Aug 2024 13:01
 Operator : RC/JU
 Sample : PB162757BS
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :

BNA_M

ClientSampleId :

PB162757BS

Manual Integrations**APPROVED**

Reviewed By :Jagrut Upadhyay 08/20/2024
 Supervised By :mohammad ahmed 08/21/2024

Quant Time: Aug 19 13:38:34 2024

Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081024.M

Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

QLast Update : Sun Aug 11 23:47:58 2024

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	12.527	196	610765	47.156	ng	97
46) 1,1'-Biphenyl	12.845	154	1692324	39.506	ng	99
47) 2-Chloronaphthalene	12.886	162	1420731	42.433	ng	97
48) 2-Nitroaniline	13.110	65	481315	45.887	ng	98
49) Acenaphthylene	13.745	152	2350545	47.627	ng	100
50) Dimethylphthalate	13.510	163	1955524	46.533	ng	99
51) 2,6-Dinitrotoluene	13.627	165	447612	45.564	ng	89
52) Acenaphthene	14.092	154	1350366	43.134	ng	99
53) 3-Nitroaniline	13.957	138	385947	39.207	ng	98
54) 2,4-Dinitrophenol	14.180	184	595111	101.074	ng	93
55) Dibenzofuran	14.433	168	2242954	45.358	ng	98
56) 4-Nitrophenol	14.310	139	830321	97.818	ng	97
57) 2,4-Dinitrotoluene	14.427	165	625506	47.703	ng	97
58) Fluorene	15.092	166	1840031	45.097	ng	99
59) 2,3,4,6-Tetrachlorophenol	14.674	232	563384	52.926	ng	94
60) Diethylphthalate	14.892	149	1969145	45.885	ng	98
61) 4-Chlorophenyl-phenyle...	15.092	204	888294	45.730	ng	97
62) 4-Nitroaniline	15.133	138	465668	48.018	ng	96
63) Azobenzene	15.392	77	1740175	42.767	ng	95
65) 4,6-Dinitro-2-methylph...	15.204	198	406204	51.054	ng	90
66) n-Nitrosodiphenylamine	15.315	169	1666769	45.031	ng	98
67) 4-Bromophenyl-phenylether	15.992	248	605167	46.214	ng	98
68) Hexachlorobenzene	16.098	284	701756	44.746	ng	96
69) Atrazine	16.292	200	559997	50.278	ng	99
70) Pentachlorophenol	16.457	266	1002308	91.172	ng	98
71) Phenanthrene	16.833	178	3126126	45.838	ng	100
72) Anthracene	16.927	178	3130614	47.177	ng	100
73) Carbazole	17.209	167	3107281	45.861	ng	99
74) Di-n-butylphthalate	17.792	149	3652536	46.330	ng	99
75) Fluoranthene	18.868	202	3957445	47.707	ng	99
77) Benzidine	19.074	184	973849	43.113	ng	99
78) Pyrene	19.239	202	4183940	43.992	ng	99
80) Butylbenzylphthalate	20.174	149	1835254	47.284	ng	95
81) Benzo(a)anthracene	21.021	228	4186583	47.346	ng	100
82) 3,3'-Dichlorobenzidine	20.956	252	1316983	47.602	ng	98
83) Chrysene	21.074	228	3874920	46.188	ng	100
84) Bis(2-ethylhexyl)phtha...	20.974	149	2493859	45.307	ng #	98
85) Di-n-octyl phthalate	22.009	149	4621455	49.397	ng	98
87) Indeno(1,2,3-cd)pyrene	26.762	276	5478272	52.537	ng	98
88) Benzo(b)fluoranthene	22.897	252	4575044	47.755	ng	99
89) Benzo(k)fluoranthene	22.956	252	4333239	47.886	ng	99
90) Benzo(a)pyrene	23.621	252	4321255	52.069	ng	98
91) Dibenzo(a,h)anthracene	26.809	278	4427346	52.465	ng	99
92) Benzo(g,h,i)perylene	27.691	276	4267391	48.706	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081924\
 Data File : BM047250.D
 Acq On : 19 Aug 2024 13:01
 Operator : RC/JU
 Sample : PB162757BS
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :

BNA_M


ClientSampleId :

PB162757BS

Manual Integrations**APPROVED**

Reviewed By :Jagrut Upadhyay 08/20/2024
 Supervised By :mohammad ahmed 08/21/2024

Quant Time: Aug 19 13:38:34 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Aug 11 23:47:58 2024
 Response via : Initial Calibration




Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081924\
Data File : BM047250.D
Acq On : 19 Aug 2024 13:01
Operator : RC/JU
Sample : PB162757BS
Misc :
ALS Vial : 7 Sample Multiplier: 1


Instrument :
BNA_M
ClientSampleId :
PB162757BS

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024


Quant Time: Aug 19 13:38:34 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sun Aug 11 23:47:58 2024
Response via : Initial Calibration





#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.363 min Scan# 70
Delta R.T. -0.012 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01




Instrument :
BNA_M
ClientSampleId :
PB162757BS



Tgt Ion:152 Resp: 200715
Ion Ratio Lower Upper
152 100
150 151.0 120.5 180.7
115 53.9 45.5 68.3




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#2
1,4-Dioxane
Concen: 26.131 ng
RT: 3.022 min Scan# 23
Delta R.T. -0.012 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01


Tgt Ion: 88 Resp: 121620
Ion Ratio Lower Upper
88 100
58 72.4 58.2 87.4
43 29.5 22.5 33.7





#3
Pyridine
Concen: 29.045 ng
RT: 3.399 min Scan# 81
Delta R.T. -0.012 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01




Instrument :
BNA_M
ClientSampleId :
PB162757BS



Tgt Ion: 79 Resp: 396398
Ion Ratio Lower Upper
79 100
52 61.6 51.2 76.8
51 31.8 26.1 39.1




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#4
n-Nitrosodimethylamine
Concen: 36.296 ng
RT: 3.316 min Scan# 73
Delta R.T. -0.012 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01


Tgt Ion: 42 Resp: 216424
Ion Ratio Lower Upper
42 100
74 140.6 112.9 169.3
44 7.3 5.3 7.9





#5
2-Fluorophenol
Concen: 133.602 ng
RT: 5.022 min Scan# 30
Delta R.T. -0.000 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01




Instrument :
BNA_M
ClientSampleId :
PB162757BS



Tgt Ion:112 Resp: 150074
Ion Ratio Lower Upper
112 100
64 56.5 46.7 70.1
63 29.3 24.8 37.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#6
Aniline
Concen: 27.784 ng
RT: 6.710 min Scan# 650
Delta R.T. -0.018 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01


Tgt Ion: 93 Resp: 406471
Ion Ratio Lower Upper
93 100
66 38.3 31.4 47.2
65 18.8 15.6 23.4





#7
Phenol-d6
Concen: 130.792 ng
RT: 6.587 min Scan# 61
Delta R.T. -0.000 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01




Instrument :
BNA_M
ClientSampleId :
PB162757BS




Tgt Ion: 99 Resp: 2024398
Ion Ratio Lower Upper
99 100
42 16.0 14.1 21.1
71 34.3 28.5 42.7

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024




#8
2-Chlorophenol
Concen: 51.293 ng
RT: 6.945 min Scan# 690
Delta R.T. -0.006 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01




Tgt Ion:128 Resp: 628413
Ion Ratio Lower Upper
128 100
130 33.7 12.1 52.1
64 43.9 26.6 66.6





#9
Benzaldehyde
Concen: 16.936 ng
RT: 6.528 min Scan# 619
Delta R.T. -0.012 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01




Instrument :
BNA_M
ClientSampleId :
PB162757BS



Tgt Ion: 77 Resp: 171778
Ion Ratio Lower Upper
77 100
105 95.1 74.5 114.5
106 92.1 72.3 112.3




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#10
Phenol
Concen: 48.467 ng
RT: 6.610 min Scan# 633
Delta R.T. -0.000 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01

Tgt Ion: 94 Resp: 748552
Ion Ratio Lower Upper
94 100
65 27.9 9.0 49.0
66 37.1 18.9 58.9









#11
bis(2-Chloroethyl)ether
Concen: 42.374 ng
RT: 6.816 min Scan# 60
Delta R.T. -0.012 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01

Instrument :
BNA_M
ClientSampleId :
PB162757BS

Tgt Ion: 93 Resp: 56044
Ion Ratio Lower Upper
93 100
63 73.3 55.8 95.8
95 32.6 12.7 52.7




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#12
1,3-Dichlorobenzene
Concen: 42.495 ng
RT: 7.257 min Scan# 743
Delta R.T. -0.006 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01

Tgt Ion:146 Resp: 618674
Ion Ratio Lower Upper
146 100
148 64.4 52.3 78.5
75 28.8 25.8 38.6





#13

1,4-Dichlorobenzene

Concen: 43.163 ng

RT: 7.398 min Scan# 70

Delta R.T. -0.012 min

Lab File: BM047250.D

Acq: 19 Aug 2024 13:01

Instrument :

BNA_M

ClientSampleId :

PB162757BS

Tgt Ion:146 Resp: 63623

Ion Ratio Lower Upper





146 100

148 63.5 51.4 77.2

111 37.3 31.5 47.3

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024

#14

1,2-Dichlorobenzene

Concen: 45.279 ng

RT: 7.704 min Scan# 819

Delta R.T. -0.012 min

Lab File: BM047250.D

Acq: 19 Aug 2024 13:01




Tgt Ion:146 Resp: 629748


Ion Ratio Lower Upper

146 100

148 63.4 51.8 77.8


111 40.5 33.3 49.9





#15
Benzyl Alcohol
Concen: 51.345 ng
RT: 7.616 min Scan# 804
Delta R.T. -0.012 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01




Instrument :
BNA_M
ClientSampleId :
PB162757BS




Tgt Ion: 79 Resp: 565608
Ion Ratio Lower Upper
79 100
108 75.9 59.7 89.5
77 64.1 52.9 79.3

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024




#16
2,2'-oxybis(1-Chloropropane)
Concen: 41.970 ng
RT: 7.886 min Scan# 850
Delta R.T. -0.030 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01




Tgt Ion: 45 Resp: 712313
Ion Ratio Lower Upper
45 100
77 16.0 0.0 35.6
79 12.2 0.0 32.4





#17
2-Methylphenol
Concen: 49.915 ng
RT: 7.828 min Scan# 841
Delta R.T. -0.006 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01




Instrument :
BNA_M
ClientSampleId :
PB162757BS



Tgt Ion:107 Resp: 514304
Ion Ratio Lower Upper
107 100
108 114.2 90.6 135.8
77 48.3 39.3 58.9
79 47.1 38.6 58.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#18
Hexachloroethane
Concen: 41.408 ng
RT: 8.416 min Scan# 940
Delta R.T. -0.012 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01


Tgt Ion:117 Resp: 236430
Ion Ratio Lower Upper
117 100
119 95.6 77.0 115.4
201 113.0 81.3 121.9





#19
n-Nitroso-di-n-propylamine
Concen: 43.129 ng
RT: 8.175 min Scan# 899
Delta R.T. -0.012 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01




Instrument :
BNA_M
ClientSampleId :
PB162757BS



Tgt Ion: 70 Resp: 449654
Ion Ratio Lower Upper
70 100
42 52.1 42.0 63.0
101 10.2 7.2 10.8
130 20.8 15.5 23.3




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#20
3+4-Methylphenols
Concen: 49.601 ng
RT: 8.157 min Scan# 896
Delta R.T. -0.006 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01


Tgt Ion:107 Resp: 714481
Ion Ratio Lower Upper
107 100
108 90.4 70.7 110.7
77 37.3 17.4 57.4
79 29.7 10.9 50.9





#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.122 min Scan# 1
Delta R.T. -0.012 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01




Instrument :
BNA_M
ClientSampleId :
PB162757BS



Tgt Ion:136 Resp: 84042
Ion Ratio Lower Upper
136 100
137 10.6 8.8 13.2
54 7.1 6.0 9.0
68 5.8 4.8 7.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#22
Acetophenone
Concen: 39.261 ng
RT: 8.186 min Scan# 901
Delta R.T. -0.012 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01


Tgt Ion:105 Resp: 796460
Ion Ratio Lower Upper
105 100
71 4.9 4.5 6.7
51 24.5 20.1 30.1
120 23.4 17.8 26.6





#23
Nitrobenzene-d5
Concen: 78.682 ng
RT: 8.516 min Scan# 91
Delta R.T. -0.006 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01




Instrument :
BNA_M
ClientSampleId :
PB162757BS




Tgt Ion: 82 Resp: 131324
Ion Ratio Lower Upper
82 100
128 43.2 32.3 48.5
54 43.8 37.4 56.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024




#24
Nitrobenzene
Concen: 41.119 ng
RT: 8.557 min Scan# 964
Delta R.T. -0.006 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01




Tgt Ion: 77 Resp: 691067
Ion Ratio Lower Upper
77 100
123 44.5 34.0 51.0
65 12.8 10.5 15.7





#25
Isophorone
Concen: 44.916 ng
RT: 9.081 min Scan# 1053
Delta R.T. -0.012 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01




Instrument :
BNA_M
ClientSampleId :
PB162757BS



Tgt Ion: 82 Resp: 130530
Ion Ratio Lower Upper
82 100
95 6.9 5.5 8.3
138 17.9 13.5 20.3




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#26
2-Nitrophenol
Concen: 48.930 ng
RT: 9.257 min Scan# 1083
Delta R.T. -0.006 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01


Tgt Ion:139 Resp: 364187
Ion Ratio Lower Upper
139 100
109 28.8 24.4 36.6
65 48.0 43.0 64.6





#27
2,4-Dimethylphenol
Concen: 54.185 ng
RT: 9.339 min Scan# 1097
Delta R.T. -0.006 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01




Instrument :
BNA_M
ClientSampleId :
PB162757BS



Tgt Ion:122 Resp: 472410
Ion Ratio Lower Upper
122 100
107 115.5 92.8 139.2
121 59.2 46.1 69.1




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#28
bis(2-Chloroethoxy)methane
Concen: 43.905 ng
RT: 9.563 min Scan# 1135
Delta R.T. -0.012 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01

Tgt Ion: 93 Resp: 802839
Ion Ratio Lower Upper
93 100
95 33.3 25.8 38.8
123 11.4 9.0 13.4





#29

2,4-Dichlorophenol

Concen: 53.068 ng

RT: 9.792 min Scan# 1175

Delta R.T. -0.006 min

Lab File: BM047250.D


Acq: 19 Aug 2024 13:01

Instrument :

BNA_M

ClientSampleId :

PB162757BS



Tgt Ion:162 Resp: 686500

Ion Ratio Lower Upper

162 100

164 63.9 44.7 84.7




98 32.9 14.9 54.9

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 08/20/2024

Supervised By :mohammad ahmed 08/21/2024



#30

1,2,4-Trichlorobenzene


Concen: 45.431 ng

RT: 9.992 min Scan# 1208

Delta R.T. -0.006 min

Lab File: BM047250.D

Acq: 19 Aug 2024 13:01





Tgt Ion:180 Resp: 665098


Ion Ratio Lower Upper

180 100

182 95.7 76.6 115.0


145 28.1 23.4 35.2





#31
Naphthalene
Concen: 43.556 ng
RT: 10.175 min Scan# 1129
Delta R.T. -0.006 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01




Instrument :
BNA_M
ClientSampleId :
PB162757BS



Tgt Ion:128 Resp: 1823534
Ion Ratio Lower Upper
128 100
129 10.9 8.6 13.0
127 12.9 10.2 15.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#32
Benzoic acid
Concen: 50.747 ng
RT: 9.528 min Scan# 1129
Delta R.T. -0.006 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01


Tgt Ion:122 Resp: 514032
Ion Ratio Lower Upper
122 100
105 125.5 107.1 147.1
77 97.3 82.1 122.1





#33
4-Chloroaniline
Concen: 27.184 ng
RT: 10.298 min Scan# 1260
Delta R.T. -0.012 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01




Instrument :
BNA_M
ClientSampleId :
PB162757BS




Tgt Ion:127 Resp: 43766
Ion Ratio Lower Upper
127 100
129 32.3 25.8 38.8
65 31.6 24.9 37.3
92 19.4 15.1 22.7

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024




#34
Hexachlorobutadiene
Concen: 44.736 ng
RT: 10.457 min Scan# 1287
Delta R.T. -0.012 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01




Tgt Ion:225 Resp: 383339
Ion Ratio Lower Upper
225 100
223 63.6 50.2 75.2
227 65.2 51.4 77.0





#35
Caprolactam
Concen: 49.303 ng m
RT: 11.104 min Scan# 113
Delta R.T. -0.018 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01




Instrument :
BNA_M
ClientSampleId :
PB162757BS



Tgt Ion:113 Resp: 221498
Ion Ratio Lower Upper
113 100
55 151.6 140.5 180.5
56 116.3 103.6 143.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#36
4-Chloro-3-methylphenol
Concen: 50.057 ng
RT: 11.451 min Scan# 1456
Delta R.T. -0.000 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01


Tgt Ion:107 Resp: 689653
Ion Ratio Lower Upper
107 100
144 27.2 21.8 32.6
142 80.9 66.8 100.2





#37
2-Methylnaphthalene
Concen: 45.365 ng
RT: 11.798 min Scan# 1515
Delta R.T. -0.006 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01




Instrument :
BNA_M
ClientSampleId :
PB162757BS



Tgt Ion:142 Resp: 1352299
Ion Ratio Lower Upper
142 100
141 87.5 70.9 106.3
115 34.0 28.7 43.1




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#38
1-Methylnaphthalene
Concen: 42.869 ng
RT: 12.022 min Scan# 1553
Delta R.T. -0.012 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01


Tgt Ion:142 Resp: 1262961
Ion Ratio Lower Upper
142 100
141 91.5 74.6 111.8
116 3.8 3.1 4.7





#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.027 min Scan# 1895
Delta R.T. -0.006 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01




Instrument :
BNA_M
ClientSampleId :
PB162757BS



Tgt Ion:164 Resp: 601718
Ion Ratio Lower Upper
164 100
162 98.9 80.7 121.1
160 44.0 36.0 54.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#40
1,2,4,5-Tetrachlorobenzene
Concen: 42.897 ng
RT: 12.180 min Scan# 1580
Delta R.T. -0.006 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01


Tgt Ion:216 Resp: 704034
Ion Ratio Lower Upper
216 100
214 78.9 63.4 95.2
179 20.9 17.6 26.4
108 17.8 15.1 22.7





#41
Hexachlorocyclopentadiene
Concen: 89.772 ng
RT: 12.157 min Scan# 1576
Delta R.T. -0.006 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01




Instrument :
BNA_M
ClientSampleId :
PB162757BS



Tgt Ion:237 Resp: 51447
Ion Ratio Lower Upper
237 100
235 62.0 44.8 84.8
272 13.0 0.0 33.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#42
2,4,6-Tribromophenol
Concen: 141.011 ng
RT: 15.539 min Scan# 2151
Delta R.T. -0.000 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01


Tgt Ion:330 Resp: 983835
Ion Ratio Lower Upper
330 100
332 96.5 77.4 116.2
141 35.4 31.6 47.4





#43
2,4,6-Trichlorophenol
Concen: 49.854 ng
RT: 12.445 min Scan# 1625
Delta R.T. -0.000 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01




Instrument :
BNA_M
ClientSampleId :
PB162757BS




Tgt Ion:196 Resp: 58155
Ion Ratio Lower Upper
196 100
198 96.8 75.6 113.4
200 30.5 24.6 37.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024




#44
2,4,5-Trichlorophenol
Concen: 47.156 ng
RT: 12.527 min Scan# 1639
Delta R.T. 0.006 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01




Tgt Ion:196 Resp: 610765
Ion Ratio Lower Upper
196 100
198 95.4 77.7 116.5
97 44.9 39.8 59.8
132 25.9 20.7 31.1





#45
2-Fluorobiphenyl
Concen: 80.705 ng
RT: 12.633 min Scan# 1657
Delta R.T. -0.012 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01




Instrument :
BNA_M
ClientSampleId :
PB162757BS



Tgt Ion:172 Resp: 314818
Ion Ratio Lower Upper
172 100
171 34.8 27.8 41.6
170 23.3 18.7 28.1




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#46
1,1'-Biphenyl
Concen: 39.506 ng
RT: 12.845 min Scan# 1693
Delta R.T. -0.012 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01


Tgt Ion:154 Resp: 1692324
Ion Ratio Lower Upper
154 100
153 39.6 19.0 59.0
76 12.1 0.0 31.9





#47
2-Chloronaphthalene
Concen: 42.433 ng
RT: 12.886 min Scan# 1700
Delta R.T. -0.000 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01




Instrument :
BNA_M
ClientSampleId :
PB162757BS



Tgt Ion:162 Resp: 142073
Ion Ratio Lower Upper
162 100
127 34.5 30.0 45.0
164 33.1 26.3 39.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#48
2-Nitroaniline
Concen: 45.887 ng
RT: 13.110 min Scan# 1738
Delta R.T. -0.006 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01


Tgt Ion: 65 Resp: 481315
Ion Ratio Lower Upper
65 100
92 68.7 55.1 82.7
138 109.9 85.9 128.9





#49
Acenaphthylene
Concen: 47.627 ng
RT: 13.745 min Scan# 1846
Delta R.T. -0.006 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01




Instrument :
BNA_M
ClientSampleId :
PB162757BS



Tgt Ion:152 Resp: 2350549
Ion Ratio Lower Upper
152 100
151 19.7 15.8 23.6
153 13.3 10.4 15.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#50
Dimethylphthalate
Concen: 46.533 ng
RT: 13.510 min Scan# 1806
Delta R.T. -0.006 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01


Tgt Ion:163 Resp: 1955524
Ion Ratio Lower Upper
163 100
194 4.3 3.0 4.6
164 10.1 8.0 12.0





#51
2,6-Dinitrotoluene
Concen: 45.564 ng
RT: 13.627 min Scan# 1826
Delta R.T. -0.000 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01




Instrument :
BNA_M
ClientSampleId :
PB162757BS




Tgt Ion:165 Resp: 447612
Ion Ratio Lower Upper
165 100
63 49.0 46.5 69.7
89 51.7 47.3 70.9

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024




#52
Acenaphthene
Concen: 43.134 ng
RT: 14.092 min Scan# 1905
Delta R.T. -0.006 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01




Tgt Ion:154 Resp: 1350366
Ion Ratio Lower Upper
154 100
153 112.2 90.3 135.5
152 53.0 42.9 64.3





#53
3-Nitroaniline
Concen: 39.207 ng
RT: 13.957 min Scan# 11
Delta R.T. -0.006 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01




Instrument :
BNA_M
ClientSampleId :
PB162757BS




Tgt Ion:138 Resp: 38594
Ion Ratio Lower Upper
138 100
108 10.1 7.9 11.9
92 113.1 92.0 138.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024




#54
2,4-Dinitrophenol
Concen: 101.074 ng
RT: 14.180 min Scan# 1920
Delta R.T. 0.005 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01




Tgt Ion:184 Resp: 595111
Ion Ratio Lower Upper
184 100
63 56.9 52.7 79.1
154 60.5 50.4 75.6





#55
Dibenzenofuran
Concen: 45.358 ng
RT: 14.433 min Scan# 1964
Delta R.T. -0.006 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01




Instrument :
BNA_M
ClientSampleId :
PB162757BS



Tgt Ion:168 Resp: 2242954
Ion Ratio Lower Upper
168 100
139 35.9 29.6 44.4
169 13.7 10.7 16.1




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#56
4-Nitrophenol
Concen: 97.818 ng
RT: 14.310 min Scan# 1942
Delta R.T. 0.012 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01


Tgt Ion:139 Resp: 830321
Ion Ratio Lower Upper
139 100
109 73.6 52.5 92.5
65 91.6 75.4 115.4





#57
2,4-Dinitrotoluene
Concen: 47.703 ng
RT: 14.427 min Scan# 1962
Delta R.T. -0.000 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01




Instrument :
BNA_M
ClientSampleId :
PB162757BS



Tgt Ion:165 Resp: 625500
Ion Ratio Lower Upper
165 100
63 49.1 40.3 60.5
89 72.9 61.0 91.6
182 2.6 1.9 2.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#58
Fluorene
Concen: 45.097 ng
RT: 15.092 min Scan# 2075
Delta R.T. -0.000 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01


Tgt Ion:166 Resp: 1840031
Ion Ratio Lower Upper
166 100
165 95.9 77.4 116.0
167 13.4 11.0 16.6





#59
2,3,4,6-Tetrachlorophenol
Concen: 52.926 ng
RT: 14.674 min Scan# 2004
Delta R.T. -0.000 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01




Instrument :
BNA_M
ClientSampleId :
PB162757BS




Tgt Ion:232 Resp: 563384
Ion Ratio Lower Upper
232 100
131 40.5 36.9 55.3
130 2.2 2.0 3.0
166 30.9 26.4 39.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024




#60
Diethylphthalate
Concen: 45.885 ng
RT: 14.892 min Scan# 2041
Delta R.T. -0.006 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01




Tgt Ion:149 Resp: 1969145
Ion Ratio Lower Upper
149 100
177 23.2 17.6 26.4
150 12.5 9.8 14.6





#61
4-Chlorophenyl-phenylether
Concen: 45.730 ng
RT: 15.092 min Scan# 2076
Delta R.T. -0.006 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01




Instrument :
BNA_M
ClientSampleId :
PB162757BS



Tgt Ion:204 Resp: 888294
Ion Ratio Lower Upper
204 100
206 32.8 26.2 39.4
141 55.3 47.2 70.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#62
4-Nitroaniline
Concen: 48.018 ng
RT: 15.133 min Scan# 2082
Delta R.T. -0.006 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01


Tgt Ion:138 Resp: 465668
Ion Ratio Lower Upper
138 100
92 50.5 31.9 71.9
108 79.2 63.8 103.8





#63
Azobenzene
Concen: 42.767 ng
RT: 15.392 min Scan# 2126
Delta R.T. -0.000 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01




Instrument :
BNA_M
ClientSampleId :
PB162757BS



Tgt Ion: 77 Resp: 174017
Ion Ratio Lower Upper
77 100
182 28.6 3.8 43.8
105 20.5 0.0 39.5
51 25.2 6.8 46.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 16.792 min Scan# 2364
Delta R.T. -0.000 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01


Tgt Ion:188 Resp: 1358929
Ion Ratio Lower Upper
188 100
94 8.4 7.0 10.6
80 9.2 8.0 12.0





#65
4,6-Dinitro-2-methylphenol
Concen: 51.054 ng
RT: 15.204 min Scan# 2094
Delta R.T. 0.006 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01




Instrument :
BNA_M
ClientSampleId :
PB162757BS



Tgt Ion:198 Resp: 406204
Ion Ratio Lower Upper
198 100
51 37.2 23.9 63.9
105 37.8 24.3 64.3




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#66
n-Nitrosodiphenylamine
Concen: 45.031 ng
RT: 15.315 min Scan# 2113
Delta R.T. -0.006 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01


Tgt Ion:169 Resp: 1666769
Ion Ratio Lower Upper
169 100
168 65.5 53.8 80.8
167 35.9 29.3 43.9





#67
4-Bromophenyl-phenylether
Concen: 46.214 ng
RT: 15.992 min Scan# 211
Delta R.T. -0.006 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01




Instrument :
BNA_M
ClientSampleId :
PB162757BS



Tgt Ion:248 Resp: 60516
Ion Ratio Lower Upper
248 100
250 98.9 77.8 116.8
141 69.7 57.7 86.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#68
Hexachlorobenzene
Concen: 44.746 ng
RT: 16.098 min Scan# 2246
Delta R.T. -0.000 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01


Tgt Ion:284 Resp: 701756
Ion Ratio Lower Upper
284 100
142 33.0 29.4 44.0
249 31.2 25.1 37.7





#69
Atrazine
Concen: 50.278 ng
RT: 16.292 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01




Instrument :
BNA_M
ClientSampleId :
PB162757BS



Tgt Ion:200 Resp: 55999
Ion Ratio Lower Upper
200 100
173 25.8 6.1 46.1
215 47.7 26.8 66.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#70
Pentachlorophenol
Concen: 91.172 ng
RT: 16.457 min Scan# 2307
Delta R.T. 0.006 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01


Tgt Ion:266 Resp: 1002308
Ion Ratio Lower Upper
266 100
268 64.2 49.8 74.6
264 63.1 49.6 74.4





#71
Phenanthrene
Concen: 45.838 ng
RT: 16.833 min Scan# 2371
Delta R.T. -0.006 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01




Instrument :
BNA_M
ClientSampleId :
PB162757BS



Tgt Ion:178 Resp: 3126120
Ion Ratio Lower Upper
178 100
176 19.1 15.3 22.9
179 15.6 12.4 18.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#72
Anthracene
Concen: 47.177 ng
RT: 16.927 min Scan# 2387
Delta R.T. -0.000 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01


Tgt Ion:178 Resp: 3130614
Ion Ratio Lower Upper
178 100
176 18.6 14.8 22.2
179 15.3 12.4 18.6





#73
Carbazole
Concen: 45.861 ng
RT: 17.209 min Scan# 2435
Delta R.T. -0.000 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01




Instrument :
BNA_M
ClientSampleId :
PB162757BS



Tgt Ion:167 Resp: 310728
Ion Ratio Lower Upper
167 100
166 21.6 17.1 25.7
139 12.0 10.0 15.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#74
Di-n-butylphthalate
Concen: 46.330 ng
RT: 17.792 min Scan# 2534
Delta R.T. -0.000 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01


Tgt Ion:149 Resp: 3652536
Ion Ratio Lower Upper
149 100
150 9.1 7.5 11.3
104 4.7 4.1 6.1





#75
Fluoranthene
Concen: 47.707 ng
RT: 18.868 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01




Instrument :
BNA_M
ClientSampleId :
PB162757BS



Tgt Ion:202 Resp: 395744
Ion Ratio Lower Upper
202 100
101 9.7 0.0 30.2
203 17.9 0.0 37.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.039 min Scan# 3086
Delta R.T. 0.006 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01


Tgt Ion:240 Resp: 1331978
Ion Ratio Lower Upper
240 100
120 8.9 8.2 12.2
236 25.5 20.6 30.8





#77
Benzidine
Concen: 43.113 ng
RT: 19.074 min Scan# 2752
Delta R.T. -0.000 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01




Instrument :
BNA_M
ClientSampleId :
PB162757BS



Tgt Ion:184 Resp: 973849
Ion Ratio Lower Upper
184 100
185 14.0 11.8 17.6
183 11.5 9.0 13.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#78
Pyrene
Concen: 43.992 ng
RT: 19.239 min Scan# 2780
Delta R.T. -0.000 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01


Tgt Ion:202 Resp: 4183940
Ion Ratio Lower Upper
202 100
200 20.7 17.2 25.8
203 17.5 14.2 21.2





#79
Terphenyl-d14
Concen: 95.394 ng
RT: 19.462 min Scan# 2818
Delta R.T. -0.000 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01




Instrument :
BNA_M
ClientSampleId :
PB162757BS



Tgt Ion:244 Resp: 592982
Ion Ratio Lower Upper
244 100
212 7.5 6.2 9.4
122 8.5 7.1 10.7




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#80
Butylbenzylphthalate
Concen: 47.284 ng
RT: 20.174 min Scan# 2939
Delta R.T. -0.000 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01


Tgt Ion:149 Resp: 1835254
Ion Ratio Lower Upper
149 100
91 66.2 57.0 85.4
206 21.6 16.4 24.6





#81
Benzo(a)anthracene
Concen: 47.346 ng
RT: 21.021 min Scan# 3082
Delta R.T. 0.006 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01




Instrument :
BNA_M
ClientSampleId :
PB162757BS



Tgt Ion:228 Resp: 418658
Ion Ratio Lower Upper
228 100
226 27.4 21.8 32.6
229 19.7 15.8 23.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#82
3,3'-Dichlorobenzidine
Concen: 47.602 ng
RT: 20.956 min Scan# 3072
Delta R.T. -0.000 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01


Tgt Ion:252 Resp: 1316983
Ion Ratio Lower Upper
252 100
254 65.0 51.0 76.6
126 10.1 8.5 12.7





#83
Chrysene
Concen: 46.188 ng
RT: 21.074 min Scan# 3092
Delta R.T. -0.000 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01




Instrument :
BNA_M
ClientSampleId :
PB162757BS



Tgt Ion:228 Resp: 3874920
Ion Ratio Lower Upper
228 100
226 30.3 24.4 36.6
229 19.7 15.7 23.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#84
Bis(2-ethylhexyl)phthalate
Concen: 45.307 ng
RT: 20.974 min Scan# 3075
Delta R.T. -0.000 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01


Tgt Ion:149 Resp: 2493859
Ion Ratio Lower Upper
149 100
167 28.7 22.2 33.2
279 4.5 3.0 4.4#





#85
Di-n-octyl phthalate
Concen: 49.397 ng
RT: 22.009 min Scan# 31
Delta R.T. -0.000 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01




Instrument :
BNA_M
ClientSampleId :
PB162757BS



Tgt Ion:149 Resp: 462145
Ion Ratio Lower Upper
149 100
167 1.6 1.3 1.9
43 7.8 6.9 10.3




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#86
Perylene-d12
Concen: 20.000 ng
RT: 23.738 min Scan# 3545
Delta R.T. 0.011 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01


Tgt Ion:264 Resp: 1613228
Ion Ratio Lower Upper
264 100
260 23.3 18.5 27.7
265 22.0 17.3 25.9





#87
Indeno(1,2,3-cd)pyrene
Concen: 52.537 ng
RT: 26.762 min Scan# 4059
Delta R.T. 0.035 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01




Instrument :
BNA_M
ClientSampleId :
PB162757BS




Tgt Ion: 276 Resp: 547827
Ion Ratio Lower Upper
276 100
138 22.3 19.4 29.0
277 25.2 20.1 30.1

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024




#88
Benzo(b)fluoranthene
Concen: 47.755 ng
RT: 22.897 min Scan# 3402
Delta R.T. 0.011 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01




Tgt Ion: 252 Resp: 4575044
Ion Ratio Lower Upper
252 100
253 22.2 17.8 26.6
125 8.3 7.4 11.0





#89
Benzo(k)fluoranthene
Concen: 47.886 ng
RT: 22.956 min Scan# 3410
Delta R.T. 0.011 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01




Instrument :
BNA_M
ClientSampleId :
PB162757BS



Tgt Ion:252 Resp: 4333239
Ion Ratio Lower Upper
252 100
253 21.8 17.7 26.5
125 8.3 7.4 11.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#90
Benzo(a)pyrene
Concen: 52.069 ng
RT: 23.621 min Scan# 3525
Delta R.T. 0.011 min
Lab File: BM047250.D
Acq: 19 Aug 2024 13:01

Tgt Ion:252 Resp: 4321255
Ion Ratio Lower Upper
252 100
253 22.1 17.1 25.7
125 9.6 8.2 12.2





#91

Dibenzo(a,h)anthracene

Concen: 52.465 ng

RT: 26.809 min Scan# 4062

Delta R.T. 0.029 min

Lab File: BM047250.D

Acq: 19 Aug 2024 13:01

Instrument :

BNA_M

ClientSampleId :

PB162757BS

Tgt Ion:278 Resp: 4427340

Ion Ratio Lower Upper





278 100

139 14.3 12.0 18.0

279 24.4 19.2 28.8

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024

#92

Benzo(g,h,i)perylene

Concen: 48.706 ng

RT: 27.691 min Scan# 4217

Delta R.T. 0.029 min

Lab File: BM047250.D

Acq: 19 Aug 2024 13:01




Tgt Ion:276 Resp: 4267391

Ion Ratio Lower Upper

276 100

277 23.8 18.8 28.2

138 19.4 16.2 24.4



Report of Analysis

Client:	JACOBS Engineering Group, Inc.	Date Collected:	
Project:	Former Schlumberger Site Princeton NJ	Date Received:	
Client Sample ID:	PB162757BSD	SDG No.:	P3609
Lab Sample ID:	PB162757BSD	Matrix:	Water
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group6
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BM047251.D	1	08/15/24 08:23	08/19/24 13:42	PB162757

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	29.0		1.60	5.00	ug/L
100-52-7	Benzaldehyde	17.2		4.00	10.0	ug/L
95-48-7	2-Methylphenol	50.1		1.10	5.00	ug/L
65794-96-9	3+4-Methylphenols	49.6		1.20	10.0	ug/L
67-72-1	Hexachloroethane	41.6		1.00	5.00	ug/L
98-95-3	Nitrobenzene	41.2		1.30	5.00	ug/L
91-20-3	Naphthalene	43.8		1.00	5.00	ug/L
87-68-3	Hexachlorobutadiene	45.1		1.30	5.00	ug/L
91-57-6	2-Methylnaphthalene	45.3		1.10	5.00	ug/L
88-06-2	2,4,6-Trichlorophenol	50.1		0.89	5.00	ug/L
95-95-4	2,4,5-Trichlorophenol	47.3		1.00	5.00	ug/L
208-96-8	Acenaphthylene	48.2		1.00	5.00	ug/L
83-32-9	Acenaphthene	44.0		0.81	5.00	ug/L
132-64-9	Dibenzofuran	45.6		0.93	5.00	ug/L
121-14-2	2,4-Dinitrotoluene	47.3		1.50	5.00	ug/L
86-73-7	Fluorene	45.5		0.96	5.00	ug/L
118-74-1	Hexachlorobenzene	45.1		1.10	5.00	ug/L
87-86-5	Pentachlorophenol	91.8	E	1.90	10.0	ug/L
85-01-8	Phenanthrene	46.0		0.89	5.00	ug/L
120-12-7	Anthracene	47.6		1.10	5.00	ug/L
86-74-8	Carbazole	46.6		1.20	5.00	ug/L
84-74-2	Di-n-butylphthalate	46.4		1.50	5.00	ug/L
206-44-0	Fluoranthene	47.9		1.30	5.00	ug/L
129-00-0	Pyrene	43.5		1.10	5.00	ug/L
56-55-3	Benzo(a)anthracene	47.2		0.94	5.00	ug/L
218-01-9	Chrysene	46.7		0.86	5.00	ug/L
117-81-7	Bis(2-ethylhexyl)phthalate	44.8		1.90	5.00	ug/L
205-99-2	Benzo(b)fluoranthene	47.2		1.10	5.00	ug/L
207-08-9	Benzo(k)fluoranthene	47.5		1.20	5.00	ug/L

Report of Analysis

Client:	JACOBS Engineering Group, Inc.	Date Collected:	
Project:	Former Schlumberger Site Princeton NJ	Date Received:	
Client Sample ID:	PB162757BSD	SDG No.:	P3609
Lab Sample ID:	PB162757BSD	Matrix:	Water
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group6
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BM047251.D	1	08/15/24 08:23	08/19/24 13:42	PB162757

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
50-32-8	Benzo(a)pyrene	51.8		1.70	5.00	ug/L
193-39-5	Indeno(1,2,3-cd)pyrene	53.2		1.00	5.00	ug/L
53-70-3	Dibenzo(a,h)anthracene	53.0		1.20	5.00	ug/L
191-24-2	Benzo(g,h,i)perylene	49.4		1.20	5.00	ug/L
123-91-1	1,4-Dioxane	26.0		1.30	5.00	ug/L
90-12-0	1-Methylnaphthalene	42.8		0.86	5.00	ug/L

SURROGATES

367-12-4	2-Fluorophenol	135		15 (10) - 110 (139)	90%	SPK: 150
13127-88-3	Phenol-d6	131		15 (10) - 110 (134)	87%	SPK: 150
4165-60-0	Nitrobenzene-d5	78.4		30 (49) - 130 (133)	78%	SPK: 100
321-60-8	2-Fluorobiphenyl	81.7		30 (52) - 130 (132)	82%	SPK: 100
118-79-6	2,4,6-Tribromophenol	142		15 (44) - 110 (137)	95%	SPK: 150
1718-51-0	Terphenyl-d14	94.7		30 (48) - 130 (125)	95%	SPK: 100

INTERNAL STANDARDS

3855-82-1	1,4-Dichlorobenzene-d4	188000	7.363
1146-65-2	Naphthalene-d8	791000	10.122
15067-26-2	Acenaphthene-d10	557000	14.027
1517-22-2	Phenanthrene-d10	1250000	16.792
1719-03-5	Chrysene-d12	1250000	21.033
1520-96-3	Perylene-d12	1570000	23.738

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081924\
 Data File : BM047251.D
 Acq On : 19 Aug 2024 13:42
 Operator : RC/JU
 Sample : PB162757BSD
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 PB162757BSD

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 08/20/2024
 Supervised By :mohammad ahmed 08/21/2024

Quant Time: Aug 19 14:49:49 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Aug 11 23:47:58 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.363	152	188275	20.000	ng	-0.01
21) Naphthalene-d8	10.122	136	791375	20.000	ng	-0.01
39) Acenaphthene-d10	14.027	164	557332	20.000	ng	0.00
64) Phenanthrene-d10	16.792	188	1252425	20.000	ng	0.00
76) Chrysene-d12	21.033	240	1245983	20.000	ng	0.00
86) Perylene-d12	23.738	264	1572179	20.000	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.022	112	1424206	135.166	ng	0.00
7) Phenol-d6	6.581	99	1895913	130.584	ng	0.00
23) Nitrobenzene-d5	8.516	82	1231431	78.353	ng	0.00
42) 2,4,6-Tribromophenol	15.539	330	920269	142.405	ng	0.00
45) 2-Fluorobiphenyl	12.633	172	2950276	81.655	ng	-0.01
79) Terphenyl-d14	19.462	244	5506473	94.697	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.022	88	113428	25.981	ng	96
3) Pyridine	3.399	79	370799	28.964	ng	97
4) n-Nitrosodimethylamine	3.316	42	205628	36.764	ng	99
6) Aniline	6.710	93	383691	27.960	ng	99
8) 2-Chlorophenol	6.945	128	598876	52.112	ng	97
9) Benzaldehyde	6.528	77	163403	17.223	ng	99
10) Phenol	6.610	94	702697	48.505	ng	98
11) bis(2-Chloroethyl)ether	6.816	93	529915	42.713	ng	95
12) 1,3-Dichlorobenzene	7.251	146	593110	43.431	ng	97
13) 1,4-Dichlorobenzene	7.398	146	605106	43.763	ng	98
14) 1,2-Dichlorobenzene	7.710	146	595731	45.664	ng	99
15) Benzyl Alcohol	7.616	79	535220	51.797	ng	99
16) 2,2'-oxybis(1-Chloropr...	7.887	45	672110	42.218	ng	99
17) 2-Methylphenol	7.828	107	484486	50.128	ng	99
18) Hexachloroethane	8.416	117	222857	41.610	ng	94
19) n-Nitroso-di-n-propyla...	8.175	70	421681	43.118	ng	98
20) 3+4-Methylphenols	8.157	107	670847	49.649	ng	98
22) Acetophenone	8.187	105	752379	39.386	ng	98
24) Nitrobenzene	8.557	77	652013	41.200	ng	96
25) Isophorone	9.081	82	1219558	44.567	ng	98
26) 2-Nitrophenol	9.257	139	342308	48.841	ng	96
27) 2,4-Dimethylphenol	9.339	122	441628	53.793	ng	100
28) bis(2-Chloroethoxy)met...	9.563	93	748328	43.460	ng	100
29) 2,4-Dichlorophenol	9.792	162	640100	52.547	ng	99
30) 1,2,4-Trichlorobenzene	9.992	180	623817	45.252	ng	99
31) Naphthalene	10.169	128	1725748	43.775	ng	99
32) Benzoic acid	9.522	122	486844	51.041	ng	96
33) 4-Chloroaniline	10.298	127	412674	27.220	ng	98
34) Hexachlorobutadiene	10.457	225	363591	45.061	ng	100
35) Caprolactam	11.104	113	203814m	48.179	ng	
36) 4-Chloro-3-methylphenol	11.451	107	637315	49.126	ng	99
37) 2-Methylnaphthalene	11.798	142	1272582	45.337	ng	98
38) 1-Methylnaphthalene	12.022	142	1186326	42.764	ng	98
40) 1,2,4,5-Tetrachloroben...	12.186	216	663901	43.673	ng	98
41) Hexachlorocyclopentadiene	12.157	237	473446	89.192	ng	98
43) 2,4,6-Trichlorophenol	12.445	196	541460	50.113	ng	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081924\
 Data File : BM047251.D
 Acq On : 19 Aug 2024 13:42
 Operator : RC/JU
 Sample : PB162757BSD
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 PB162757BSD

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 08/20/2024
 Supervised By :mohammad ahmed 08/21/2024

Quant Time: Aug 19 14:49:49 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Aug 11 23:47:58 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	12.527	196	567752	47.326	ng	96
46) 1,1'-Biphenyl	12.851	154	1586071	39.974	ng	100
47) 2-Chloronaphthalene	12.886	162	1342415	43.287	ng	98
48) 2-Nitroaniline	13.110	65	443955	45.696	ng	98
49) Acenaphthylene	13.745	152	2204793	48.232	ng	100
50) Dimethylphthalate	13.510	163	1805626	46.388	ng	100
51) 2,6-Dinitrotoluene	13.627	165	416728	45.799	ng	87
52) Acenaphthene	14.092	154	1275241	43.979	ng	98
53) 3-Nitroaniline	13.957	138	357076	39.163	ng	96
54) 2,4-Dinitrophenol	14.180	184	550923	101.021	ng	92
55) Dibenzofuran	14.433	168	2087079	45.567	ng	98
56) 4-Nitrophenol	14.304	139	774486	98.507	ng	99
57) 2,4-Dinitrotoluene	14.427	165	573934	47.256	ng	97
58) Fluorene	15.092	166	1718229	45.466	ng	99
59) 2,3,4,6-Tetrachlorophenol	14.674	232	529330	53.688	ng	93
60) Diethylphthalate	14.892	149	1818603	45.752	ng	98
61) 4-Chlorophenyl-phenyle...	15.092	204	834673	46.392	ng	96
62) 4-Nitroaniline	15.133	138	430471	47.923	ng	94
63) Azobenzene	15.392	77	1610976	42.744	ng	94
65) 4,6-Dinitro-2-methylph...	15.204	198	372358	50.780	ng	91
66) n-Nitrosodiphenylamine	15.315	169	1539154	45.119	ng	100
67) 4-Bromophenyl-phenylether	15.992	248	562507	46.609	ng	96
68) Hexachlorobenzene	16.098	284	651895	45.101	ng	98
69) Atrazine	16.292	200	516465	50.312	ng	98
70) Pentachlorophenol	16.457	266	930264	91.815	ng	99
71) Phenanthrene	16.833	178	2889868	45.977	ng	100
72) Anthracene	16.927	178	2909916	47.580	ng	100
73) Carbazole	17.209	167	2910966	46.617	ng	99
74) Di-n-butylphthalate	17.792	149	3370294	46.385	ng	99
75) Fluoranthene	18.868	202	3662035	47.900	ng	99
77) Benzidine	19.074	184	972955	46.046	ng	99
78) Pyrene	19.239	202	3872962	43.533	ng	99
80) Butylbenzylphthalate	20.174	149	1685588	46.426	ng	95
81) Benzo(a)anthracene	21.021	228	3904446	47.203	ng	100
82) 3,3'-Dichlorobenzidine	20.956	252	1267661	48.982	ng	99
83) Chrysene	21.080	228	3667613	46.734	ng	100
84) Bis(2-ethylhexyl)phtha...	20.974	149	2306020	44.786	ng #	99
85) Di-n-octyl phthalate	22.009	149	4339330	49.582	ng	98
87) Indeno(1,2,3-cd)pyrene	26.756	276	5410859	53.245	ng	98
88) Benzo(b)fluoranthene	22.897	252	4408520	47.218	ng	99
89) Benzo(k)fluoranthene	22.956	252	4184991	47.455	ng	99
90) Benzo(a)pyrene	23.621	252	4186037	51.757	ng	98
91) Dibenzo(a,h)anthracene	26.809	278	4361715	53.036	ng	99
92) Benzo(g,h,i)perylene	27.691	276	4213855	49.351	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081924\
 Data File : BM047251.D
 Acq On : 19 Aug 2024 13:42
 Operator : RC/JU
 Sample : PB162757BSD
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :

BNA_M


ClientSampleId :

PB162757BSD

Manual Integrations**APPROVED**

Reviewed By :Jagrut Upadhyay 08/20/2024
 Supervised By :mohammad ahmed 08/21/2024

Quant Time: Aug 19 14:49:49 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Aug 11 23:47:58 2024
 Response via : Initial Calibration




Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081924\
Data File : BM047251.D
Acq On : 19 Aug 2024 13:42
Operator : RC/JU
Sample : PB162757BSD
Misc :
ALS Vial : 8 Sample Multiplier: 1


Instrument :
BNA_M
ClientSampleId :
PB162757BSD

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024


Quant Time: Aug 19 14:49:49 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sun Aug 11 23:47:58 2024
Response via : Initial Calibration





#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.363 min Scan# 70
Delta R.T. -0.012 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42




Instrument :
BNA_M
ClientSampleId :
PB162757BSD



Tgt Ion:152 Resp: 188275
Ion Ratio Lower Upper
152 100
150 158.9 120.5 180.7
115 53.3 45.5 68.3




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#2
1,4-Dioxane
Concen: 25.981 ng
RT: 3.022 min Scan# 23
Delta R.T. -0.012 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42


Tgt Ion: 88 Resp: 113428
Ion Ratio Lower Upper
88 100
58 68.5 58.2 87.4
43 28.2 22.5 33.7





#3
Pyridine
Concen: 28.964 ng
RT: 3.399 min Scan# 81
Delta R.T. -0.012 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42




Instrument :
BNA_M
ClientSampleId :
PB162757BSD



Tgt Ion: 79 Resp: 370799
Ion Ratio Lower Upper
79 100
52 60.8 51.2 76.8
51 31.7 26.1 39.1




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#4
n-Nitrosodimethylamine
Concen: 36.764 ng
RT: 3.316 min Scan# 73
Delta R.T. -0.012 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42


Tgt Ion: 42 Resp: 205628
Ion Ratio Lower Upper
42 100
74 142.2 112.9 169.3
44 7.1 5.3 7.9





#5
2-Fluorophenol
Concen: 135.166 ng
RT: 5.022 min Scan# 30
Delta R.T. -0.000 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42




Instrument :
BNA_M
ClientSampleId :
PB162757BSD



Tgt Ion:112 Resp: 1424200
Ion Ratio Lower Upper
112 100
64 55.4 46.7 70.1
63 29.3 24.8 37.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#6
Aniline
Concen: 27.960 ng
RT: 6.710 min Scan# 650
Delta R.T. -0.018 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42


Tgt Ion: 93 Resp: 383691
Ion Ratio Lower Upper
93 100
66 38.7 31.4 47.2
65 19.2 15.6 23.4





#7
Phenol-d6
Concen: 130.584 ng
RT: 6.581 min Scan# 61
Delta R.T. -0.006 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42




Instrument :
BNA_M
ClientSampleId :
PB162757BSD



Tgt Ion: 99 Resp: 189591
Ion Ratio Lower Upper
99 100
42 16.7 14.1 21.1
71 35.0 28.5 42.7




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#8
2-Chlorophenol
Concen: 52.112 ng
RT: 6.945 min Scan# 690
Delta R.T. -0.006 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42


Tgt Ion:128 Resp: 598876
Ion Ratio Lower Upper
128 100
130 32.7 12.1 52.1
64 43.5 26.6 66.6





#9
Benzaldehyde
Concen: 17.223 ng
RT: 6.528 min Scan# 619
Delta R.T. -0.012 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42




Instrument :
BNA_M
ClientSampleId :
PB162757BSD




Tgt Ion: 77 Resp: 16340
Ion Ratio Lower Upper
77 100
105 95.0 74.5 114.5
106 90.8 72.3 112.3

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024




#10
Phenol
Concen: 48.505 ng
RT: 6.610 min Scan# 633
Delta R.T. -0.000 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42



Tgt Ion: 94 Resp: 702697
Ion Ratio Lower Upper
94 100
65 28.1 9.0 49.0
66 37.6 18.9 58.9









#11
bis(2-Chloroethyl)ether
Concen: 42.713 ng
RT: 6.816 min Scan# 608
Delta R.T. -0.012 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42

Instrument :
BNA_M
ClientSampleId :
PB162757BSD

Tgt Ion: 93 Resp: 529915
Ion Ratio Lower Upper
93 100
63 70.5 55.8 95.8
95 31.8 12.7 52.7




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#12
1,3-Dichlorobenzene
Concen: 43.431 ng
RT: 7.251 min Scan# 742
Delta R.T. -0.012 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42


Tgt Ion:146 Resp: 593110
Ion Ratio Lower Upper
146 100
148 63.5 52.3 78.5
75 29.5 25.8 38.6





#13
1,4-Dichlorobenzene
Concen: 43.763 ng
RT: 7.398 min Scan# 70
Delta R.T. -0.012 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42




Instrument :
BNA_M
ClientSampleId :
PB162757BSD




Tgt Ion:146 Resp: 605100
Ion Ratio Lower Upper
146 100
148 63.2 51.4 77.2
111 37.9 31.5 47.3

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024




#14
1,2-Dichlorobenzene
Concen: 45.664 ng
RT: 7.710 min Scan# 820
Delta R.T. -0.006 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42




Tgt Ion:146 Resp: 595731
Ion Ratio Lower Upper
146 100
148 65.4 51.8 77.8
111 40.2 33.3 49.9





#15
Benzyl Alcohol
Concen: 51.797 ng
RT: 7.616 min Scan# 804
Delta R.T. -0.012 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42




Instrument :
BNA_M
ClientSampleId :
PB162757BSD




Tgt Ion: 79 Resp: 535220
Ion Ratio Lower Upper
79 100
108 74.5 59.7 89.5
77 64.4 52.9 79.3

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024




#16
2,2'-oxybis(1-Chloropropane)
Concen: 42.218 ng
RT: 7.887 min Scan# 850
Delta R.T. -0.030 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42



Tgt Ion: 45 Resp: 672110
Ion Ratio Lower Upper
45 100
77 16.4 0.0 35.6
79 12.1 0.0 32.4





#17

2-Methylphenol

Concen: 50.128 ng

RT: 7.828 min Scan# 841

Delta R.T. -0.006 min

Lab File: BM047251.D

Acq: 19 Aug 2024 13:42

Instrument :

BNA_M

ClientSampleId :

PB162757BSD

Tgt Ion:107 Resp: 484480

Ion Ratio Lower Upper

107 100

108 113.4 90.6 135.8

77 48.2 39.3 58.9





79 47.5 38.6 58.0

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 08/20/2024

Supervised By :mohammad ahmed 08/21/2024



#18

Hexachloroethane

Concen: 41.610 ng

RT: 8.416 min Scan# 940

Delta R.T. -0.012 min

Lab File: BM047251.D

Acq: 19 Aug 2024 13:42




Tgt Ion:117 Resp: 222857


Ion Ratio Lower Upper

117 100

119 96.2 77.0 115.4


201 113.5 81.3 121.9





#19
n-Nitroso-di-n-propylamine
Concen: 43.118 ng
RT: 8.175 min Scan# 899
Delta R.T. -0.012 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42




Instrument :
BNA_M
ClientSampleId :
PB162757BSD



Tgt Ion: 70 Resp: 42168
Ion Ratio Lower Upper
70 100
42 51.5 42.0 63.0
101 10.2 7.2 10.8
130 21.4 15.5 23.3




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#20
3+4-Methylphenols
Concen: 49.649 ng
RT: 8.157 min Scan# 896
Delta R.T. -0.006 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42


Tgt Ion:107 Resp: 670847
Ion Ratio Lower Upper
107 100
108 88.5 70.7 110.7
77 37.2 17.4 57.4
79 30.0 10.9 50.9





#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.122 min Scan# 1232
Delta R.T. -0.012 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42




Instrument :
BNA_M
ClientSampleId :
PB162757BSD



Tgt Ion:136 Resp: 791379
Ion Ratio Lower Upper
136 100
137 10.9 8.8 13.2
54 6.8 6.0 9.0
68 5.5 4.8 7.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#22
Acetophenone
Concen: 39.386 ng
RT: 8.187 min Scan# 901
Delta R.T. -0.012 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42


Tgt Ion:105 Resp: 752379
Ion Ratio Lower Upper
105 100
71 4.6 4.5 6.7
51 24.0 20.1 30.1
120 22.9 17.8 26.6





#23
Nitrobenzene-d5
Concen: 78.353 ng
RT: 8.516 min Scan# 91
Delta R.T. -0.006 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42




Instrument :
BNA_M
ClientSampleId :
PB162757BSD



Tgt Ion: 82 Resp: 123143
Ion Ratio Lower Upper
82 100
128 43.9 32.3 48.5
54 44.7 37.4 56.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#24
Nitrobenzene
Concen: 41.200 ng
RT: 8.557 min Scan# 964
Delta R.T. -0.006 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42


Tgt Ion: 77 Resp: 652013
Ion Ratio Lower Upper
77 100
123 45.6 34.0 51.0
65 13.3 10.5 15.7





#25
Isophorone
Concen: 44.567 ng
RT: 9.081 min Scan# 1053
Delta R.T. -0.012 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42




Instrument :
BNA_M
ClientSampleId :
PB162757BSD



Tgt Ion: 82 Resp: 1219558
Ion Ratio Lower Upper
82 100
95 6.9 5.5 8.3
138 18.1 13.5 20.3




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#26
2-Nitrophenol
Concen: 48.841 ng
RT: 9.257 min Scan# 1083
Delta R.T. -0.006 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42


Tgt Ion:139 Resp: 342308
Ion Ratio Lower Upper
139 100
109 29.3 24.4 36.6
65 50.0 43.0 64.6





#27
2,4-Dimethylphenol
Concen: 53.793 ng
RT: 9.339 min Scan# 1097
Delta R.T. -0.006 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42




Instrument :
BNA_M
ClientSampleId :
PB162757BSD



Tgt Ion:122 Resp: 441628
Ion Ratio Lower Upper
122 100
107 116.1 92.8 139.2
121 58.5 46.1 69.1




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#28
bis(2-Chloroethoxy)methane
Concen: 43.460 ng
RT: 9.563 min Scan# 1135
Delta R.T. -0.012 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42

Tgt Ion: 93 Resp: 748328
Ion Ratio Lower Upper
93 100
95 32.1 25.8 38.8
123 11.1 9.0 13.4





#29

2,4-Dichlorophenol

Concen: 52.547 ng

RT: 9.792 min Scan# 1175

Delta R.T. -0.006 min

Lab File: BM047251.D

Acq: 19 Aug 2024 13:42

Instrument :

BNA_M

ClientSampleId :

PB162757BSD

Tgt Ion:162 Resp: 640100

Ion Ratio Lower Upper





162 100

164 64.7 44.7 84.7

98 33.6 14.9 54.9

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024

#30

1,2,4-Trichlorobenzene

Concen: 45.252 ng

RT: 9.992 min Scan# 1208

Delta R.T. -0.006 min

Lab File: BM047251.D

Acq: 19 Aug 2024 13:42




Tgt Ion:180 Resp: 623817


Ion Ratio Lower Upper

180 100

182 95.8 76.6 115.0


145 27.8 23.4 35.2





#31
Naphthalene
Concen: 43.775 ng
RT: 10.169 min Scan# 1128
Delta R.T. -0.012 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42




Instrument :
BNA_M
ClientSampleId :
PB162757BSD



Tgt Ion:128 Resp: 1725748
Ion Ratio Lower Upper
128 100
129 11.1 8.6 13.0
127 13.0 10.2 15.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#32
Benzoic acid
Concen: 51.041 ng
RT: 9.522 min Scan# 1128
Delta R.T. -0.012 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42


Tgt Ion:122 Resp: 486844
Ion Ratio Lower Upper
122 100
105 124.5 107.1 147.1
77 96.6 82.1 122.1





#33
4-Chloroaniline
Concen: 27.220 ng
RT: 10.298 min Scan# 1260
Delta R.T. -0.012 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42




Instrument :
BNA_M
ClientSampleId :
PB162757BSD




Tgt Ion:127 Resp: 412674
Ion Ratio Lower Upper
127 100
129 31.9 25.8 38.8
65 29.4 24.9 37.3
92 18.6 15.1 22.7

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024




#34
Hexachlorobutadiene
Concen: 45.061 ng
RT: 10.457 min Scan# 1287
Delta R.T. -0.012 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42




Tgt Ion:225 Resp: 363591
Ion Ratio Lower Upper
225 100
223 62.8 50.2 75.2
227 64.0 51.4 77.0





#35
Caprolactam
Concen: 48.179 ng m
RT: 11.104 min Scan# 11397
Delta R.T. -0.018 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42




Instrument :
BNA_M
ClientSampleId :
PB162757BSD



Tgt Ion:113 Resp: 203814
Ion Ratio Lower Upper
113 100
55 151.8 140.5 180.5
56 118.2 103.6 143.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#36
4-Chloro-3-methylphenol
Concen: 49.126 ng
RT: 11.451 min Scan# 1456
Delta R.T. -0.000 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42


Tgt Ion:107 Resp: 637315
Ion Ratio Lower Upper
107 100
144 27.7 21.8 32.6
142 84.2 66.8 100.2





#37
2-Methylnaphthalene
Concen: 45.337 ng
RT: 11.798 min Scan# 1515
Delta R.T. -0.006 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42




Instrument :
BNA_M
ClientSampleId :
PB162757BSD




Tgt Ion:142 Resp: 127258
Ion Ratio Lower Upper
142 100
141 87.4 70.9 106.3
115 33.5 28.7 43.1

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024




#38
1-Methylnaphthalene
Concen: 42.764 ng
RT: 12.022 min Scan# 1553
Delta R.T. -0.012 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42




Tgt Ion:142 Resp: 1186326
Ion Ratio Lower Upper
142 100
141 91.6 74.6 111.8
116 3.9 3.1 4.7





#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.027 min Scan# 1894
Delta R.T. -0.006 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42




Instrument :
BNA_M
ClientSampleId :
PB162757BSD



Tgt Ion:164 Resp: 55733
Ion Ratio Lower Upper
164 100
162 100.8 80.7 121.1
160 44.6 36.0 54.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#40
1,2,4,5-Tetrachlorobenzene
Concen: 43.673 ng
RT: 12.186 min Scan# 1581
Delta R.T. -0.000 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42


Tgt Ion:216 Resp: 663901
Ion Ratio Lower Upper
216 100
214 78.3 63.4 95.2
179 21.1 17.6 26.4
108 17.5 15.1 22.7





#41
Hexachlorocyclopentadiene
Concen: 89.192 ng
RT: 12.157 min Scan# 1576
Delta R.T. -0.006 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42




Instrument :
BNA_M
ClientSampleId :
PB162757BSD



Tgt Ion:237 Resp: 473440
Ion Ratio Lower Upper
237 100
235 63.1 44.8 84.8
272 12.8 0.0 33.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#42
2,4,6-Tribromophenol
Concen: 142.405 ng
RT: 15.539 min Scan# 2151
Delta R.T. -0.000 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42


Tgt Ion:330 Resp: 920269
Ion Ratio Lower Upper
330 100
332 96.5 77.4 116.2
141 35.0 31.6 47.4





#43
2,4,6-Trichlorophenol
Concen: 50.113 ng
RT: 12.445 min Scan# 1625
Delta R.T. -0.000 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42




Instrument :
BNA_M
ClientSampleId :
PB162757BSD




Tgt Ion:196 Resp: 541460
Ion Ratio Lower Upper
196 100
198 96.6 75.6 113.4
200 30.9 24.6 37.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024




#44
2,4,5-Trichlorophenol
Concen: 47.326 ng
RT: 12.527 min Scan# 1639
Delta R.T. 0.006 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42




Tgt Ion:196 Resp: 567752
Ion Ratio Lower Upper
196 100
198 94.6 77.7 116.5
97 45.0 39.8 59.8
132 25.3 20.7 31.1





#45
2-Fluorobiphenyl
Concen: 81.655 ng
RT: 12.633 min Scan# 1657
Delta R.T. -0.012 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42




Instrument :
BNA_M
ClientSampleId :
PB162757BSD



Tgt Ion:172 Resp: 2950270
Ion Ratio Lower Upper
172 100
171 35.4 27.8 41.6
170 23.5 18.7 28.1




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#46
1,1'-Biphenyl
Concen: 39.974 ng
RT: 12.851 min Scan# 1694
Delta R.T. -0.006 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42


Tgt Ion:154 Resp: 1586071
Ion Ratio Lower Upper
154 100
153 39.1 19.0 59.0
76 11.6 0.0 31.9





#47
2-Chloronaphthalene
Concen: 43.287 ng
RT: 12.886 min Scan# 1700
Delta R.T. -0.000 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42




Instrument :
BNA_M
ClientSampleId :
PB162757BSD




Tgt Ion:162 Resp: 1342415
Ion Ratio Lower Upper
162 100
127 35.8 30.0 45.0
164 33.7 26.3 39.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024




#48
2-Nitroaniline
Concen: 45.696 ng
RT: 13.110 min Scan# 1738
Delta R.T. -0.006 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42




Tgt Ion: 65 Resp: 443955
Ion Ratio Lower Upper
65 100
92 70.3 55.1 82.7
138 110.0 85.9 128.9





#49
Acenaphthylene
Concen: 48.232 ng
RT: 13.745 min Scan# 1846
Delta R.T. -0.006 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42




Instrument :
BNA_M
ClientSampleId :
PB162757BSD



Tgt Ion:152 Resp: 220479
Ion Ratio Lower Upper
152 100
151 19.5 15.8 23.6
153 13.0 10.4 15.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#50
Dimethylphthalate
Concen: 46.388 ng
RT: 13.510 min Scan# 1806
Delta R.T. -0.006 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42


Tgt Ion:163 Resp: 1805626
Ion Ratio Lower Upper
163 100
194 4.3 3.0 4.6
164 10.0 8.0 12.0





#51
2,6-Dinitrotoluene
Concen: 45.799 ng
RT: 13.627 min Scan# 11
Delta R.T. -0.000 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42




Instrument :
BNA_M
ClientSampleId :
PB162757BSD



Tgt Ion:165 Resp: 416728
Ion Ratio Lower Upper
165 100
63 47.7 46.5 69.7
89 50.8 47.3 70.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#52
Acenaphthene
Concen: 43.979 ng
RT: 14.092 min Scan# 1905
Delta R.T. -0.006 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42


Tgt Ion:154 Resp: 1275241
Ion Ratio Lower Upper
154 100
153 110.6 90.3 135.5
152 51.9 42.9 64.3





#53
3-Nitroaniline
Concen: 39.163 ng
RT: 13.957 min Scan# 11
Delta R.T. -0.006 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42




Instrument :
BNA_M
ClientSampleId :
PB162757BSD



Tgt Ion:138 Resp: 357070
Ion Ratio Lower Upper
138 100
108 9.9 7.9 11.9
92 110.5 92.0 138.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#54
2,4-Dinitrophenol
Concen: 101.021 ng
RT: 14.180 min Scan# 1920
Delta R.T. 0.005 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42


Tgt Ion:184 Resp: 550923
Ion Ratio Lower Upper
184 100
63 56.9 52.7 79.1
154 59.5 50.4 75.6





#55
Dibenzenofuran
Concen: 45.567 ng
RT: 14.433 min Scan# 1963
Delta R.T. -0.006 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42




Instrument :
BNA_M
ClientSampleId :
PB162757BSD



Tgt Ion:168 Resp: 2087079
Ion Ratio Lower Upper
168 100
139 35.7 29.6 44.4
169 13.5 10.7 16.1




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#56
4-Nitrophenol
Concen: 98.507 ng
RT: 14.304 min Scan# 1941
Delta R.T. 0.006 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42


Tgt Ion:139 Resp: 774486
Ion Ratio Lower Upper
139 100
109 74.1 52.5 92.5
65 94.5 75.4 115.4





#57
2,4-Dinitrotoluene
Concen: 47.256 ng
RT: 14.427 min Scan# 1962
Delta R.T. -0.000 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42




Instrument :
BNA_M
ClientSampleId :
PB162757BSD




Tgt Ion:165 Resp: 573934
Ion Ratio Lower Upper
165 100
63 48.9 40.3 60.5
89 73.1 61.0 91.6
182 2.7 1.9 2.9

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024




#58
Fluorene
Concen: 45.466 ng
RT: 15.092 min Scan# 2075
Delta R.T. -0.000 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42




Tgt Ion:166 Resp: 1718229
Ion Ratio Lower Upper
166 100
165 98.0 77.4 116.0
167 13.6 11.0 16.6





#59
2,3,4,6-Tetrachlorophenol
Concen: 53.688 ng
RT: 14.674 min Scan# 2004
Delta R.T. -0.000 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42




Instrument :
BNA_M
ClientSampleId :
PB162757BSD




Tgt Ion:232 Resp: 529330
Ion Ratio Lower Upper
232 100
131 39.9 36.9 55.3
130 2.1 2.0 3.0
166 30.3 26.4 39.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024




#60
Diethylphthalate
Concen: 45.752 ng
RT: 14.892 min Scan# 2041
Delta R.T. -0.006 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42




Tgt Ion:149 Resp: 1818603
Ion Ratio Lower Upper
149 100
177 23.5 17.6 26.4
150 12.3 9.8 14.6





#61
4-Chlorophenyl-phenylether
Concen: 46.392 ng
RT: 15.092 min Scan# 2076
Delta R.T. -0.006 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42




Instrument :
BNA_M
ClientSampleId :
PB162757BSD




Tgt Ion:204 Resp: 83467
Ion Ratio Lower Upper
204 100
206 32.1 26.2 39.4
141 54.9 47.2 70.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024




#62
4-Nitroaniline
Concen: 47.923 ng
RT: 15.133 min Scan# 2082
Delta R.T. -0.006 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42




Tgt Ion:138 Resp: 430471
Ion Ratio Lower Upper
138 100
92 49.2 31.9 71.9
108 77.7 63.8 103.8





#63
Azobenzene
Concen: 42.744 ng
RT: 15.392 min Scan# 2126
Delta R.T. -0.000 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42




Instrument :
BNA_M
ClientSampleId :
PB162757BSD



Tgt Ion: 77 Resp: 1610970
Ion Ratio Lower Upper
77 100
182 28.8 3.8 43.8
105 21.5 0.0 39.5
51 25.3 6.8 46.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 16.792 min Scan# 2364
Delta R.T. -0.000 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42


Tgt Ion:188 Resp: 1252425
Ion Ratio Lower Upper
188 100
94 8.5 7.0 10.6
80 9.2 8.0 12.0





#65
4,6-Dinitro-2-methylphenol
Concen: 50.780 ng
RT: 15.204 min Scan# 2094
Delta R.T. 0.006 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42




Instrument :
BNA_M
ClientSampleId :
PB162757BSD




Tgt Ion:198 Resp: 372358
Ion Ratio Lower Upper
198 100
51 36.5 23.9 63.9
105 39.7 24.3 64.3

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024




#66
n-Nitrosodiphenylamine
Concen: 45.119 ng
RT: 15.315 min Scan# 2113
Delta R.T. -0.006 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42




Tgt Ion:169 Resp: 1539154
Ion Ratio Lower Upper
169 100
168 67.2 53.8 80.8
167 36.2 29.3 43.9





#67
4-Bromophenyl-phenylether
Concen: 46.609 ng
RT: 15.992 min Scan# 211
Delta R.T. -0.006 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42




Instrument :
BNA_M
ClientSampleId :
PB162757BSD




Tgt Ion:248 Resp: 56250
Ion Ratio Lower Upper
248 100
250 98.8 77.8 116.8
141 66.9 57.7 86.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024




#68
Hexachlorobenzene
Concen: 45.101 ng
RT: 16.098 min Scan# 2246
Delta R.T. -0.000 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42



Tgt Ion:284 Resp: 651895
Ion Ratio Lower Upper
284 100
142 34.7 29.4 44.0
249 31.3 25.1 37.7





#69

Atrazine

Concen: 50.312 ng

RT: 16.292 min Scan# 21

Delta R.T. -0.000 min

Lab File: BM047251.D


Acq: 19 Aug 2024 13:42

Instrument :

BNA_M

ClientSampleId :

PB162757BSD






Tgt Ion:	200	Resp:	51646
Ion Ratio	Lower	Upper	
200	100		
173	24.8	6.1	46.1
215	47.5	26.8	66.8

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#70

Pentachlorophenol


Concen: 91.815 ng

RT: 16.457 min Scan# 2307



Delta R.T. 0.006 min


Lab File: BM047251.D

Acq: 19 Aug 2024 13:42




Tgt Ion:	266	Resp:	930264
Ion Ratio	Lower	Upper	
266	100		
268	63.2	49.8	74.6
264	62.5	49.6	74.4





#71
Phenanthrene
Concen: 45.977 ng
RT: 16.833 min Scan# 21
Delta R.T. -0.006 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42




Instrument :
BNA_M
ClientSampleId :
PB162757BSD



Tgt Ion:178 Resp: 288986
Ion Ratio Lower Upper
178 100
176 18.8 15.3 22.9
179 15.7 12.4 18.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#72
Anthracene
Concen: 47.580 ng
RT: 16.927 min Scan# 2387
Delta R.T. -0.000 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42


Tgt Ion:178 Resp: 2909916
Ion Ratio Lower Upper
178 100
176 18.5 14.8 22.2
179 15.5 12.4 18.6





#73
Carbazole
Concen: 46.617 ng
RT: 17.209 min Scan# 2435
Delta R.T. -0.000 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42




Instrument :
BNA_M
ClientSampleId :
PB162757BSD



Tgt Ion:167 Resp: 2910960
Ion Ratio Lower Upper
167 100
166 21.2 17.1 25.7
139 12.0 10.0 15.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#74
Di-n-butylphthalate
Concen: 46.385 ng
RT: 17.792 min Scan# 2534
Delta R.T. -0.000 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42


Tgt Ion:149 Resp: 3370294
Ion Ratio Lower Upper
149 100
150 8.9 7.5 11.3
104 4.6 4.1 6.1





#75
Fluoranthene
Concen: 47.900 ng
RT: 18.868 min Scan# 2717
Delta R.T. -0.000 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42




Instrument :
BNA_M
ClientSampleId :
PB162757BSD



Tgt Ion:202 Resp: 3662035
Ion Ratio Lower Upper
202 100
101 9.6 0.0 30.2
203 17.8 0.0 37.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.033 min Scan# 3085
Delta R.T. -0.000 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42


Tgt Ion:240 Resp: 1245983
Ion Ratio Lower Upper
240 100
120 10.5 8.2 12.2
236 25.6 20.6 30.8





#77
Benzidine
Concen: 46.046 ng
RT: 19.074 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42




Instrument :
BNA_M
ClientSampleId :
PB162757BSD



Tgt Ion:184 Resp: 97295
Ion Ratio Lower Upper
184 100
185 14.3 11.8 17.6
183 11.1 9.0 13.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#78
Pyrene
Concen: 43.533 ng
RT: 19.239 min Scan# 2780
Delta R.T. -0.000 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42


Tgt Ion:202 Resp: 3872962
Ion Ratio Lower Upper
202 100
200 21.1 17.2 25.8
203 17.6 14.2 21.2





#79
Terphenyl-d14
Concen: 94.697 ng
RT: 19.462 min Scan# 2818
Delta R.T. -0.000 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42




Instrument :
BNA_M
ClientSampleId :
PB162757BSD



Tgt Ion:244 Resp: 550647
Ion Ratio Lower Upper
244 100
212 7.3 6.2 9.4
122 8.4 7.1 10.7




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#80
Butylbenzylphthalate
Concen: 46.426 ng
RT: 20.174 min Scan# 2939
Delta R.T. -0.000 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42


Tgt Ion:149 Resp: 1685588
Ion Ratio Lower Upper
149 100
91 66.6 57.0 85.4
206 21.9 16.4 24.6





#81
Benzo(a)anthracene
Concen: 47.203 ng
RT: 21.021 min Scan# 3082
Delta R.T. 0.006 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42




Instrument :
BNA_M
ClientSampleId :
PB162757BSD



Tgt Ion:228 Resp: 3904440
Ion Ratio Lower Upper
228 100
226 27.3 21.8 32.6
229 19.5 15.8 23.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#82
3,3'-Dichlorobenzidine
Concen: 48.982 ng
RT: 20.956 min Scan# 3072
Delta R.T. -0.000 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42


Tgt Ion:252 Resp: 1267661
Ion Ratio Lower Upper
252 100
254 63.4 51.0 76.6
126 10.0 8.5 12.7





#83
Chrysene
Concen: 46.734 ng
RT: 21.080 min Scan# 3092
Delta R.T. 0.006 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42




Instrument :
BNA_M
ClientSampleId :
PB162757BSD



Tgt Ion:228 Resp: 366761
Ion Ratio Lower Upper
228 100
226 30.3 24.4 36.6
229 19.4 15.7 23.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#84
Bis(2-ethylhexyl)phthalate
Concen: 44.786 ng
RT: 20.974 min Scan# 3075
Delta R.T. -0.000 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42


Tgt Ion:149 Resp: 2306020
Ion Ratio Lower Upper
149 100
167 28.3 22.2 33.2
279 4.6 3.0 4.4#





#85
Di-n-octyl phthalate
Concen: 49.582 ng
RT: 22.009 min Scan# 31
Delta R.T. -0.000 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42




Instrument :
BNA_M
ClientSampleId :
PB162757BSD



Tgt Ion:149 Resp: 4339330
Ion Ratio Lower Upper
149 100
167 1.6 1.3 1.9
43 7.7 6.9 10.3




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#86
Perylene-d12
Concen: 20.000 ng
RT: 23.738 min Scan# 3545
Delta R.T. 0.012 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42


Tgt Ion:264 Resp: 1572179
Ion Ratio Lower Upper
264 100
260 23.1 18.5 27.7
265 21.8 17.3 25.9





#87
Indeno(1,2,3-cd)pyrene
Concen: 53.245 ng
RT: 26.756 min Scan# 4053
Delta R.T. 0.029 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42




Instrument :
BNA_M
ClientSampleId :
PB162757BSD



Tgt Ion:276 Resp: 5410859
Ion Ratio Lower Upper
276 100
138 22.2 19.4 29.0
277 25.0 20.1 30.1




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#88
Benzo(b)fluoranthene
Concen: 47.218 ng
RT: 22.897 min Scan# 3402
Delta R.T. 0.012 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42


Tgt Ion:252 Resp: 4408520
Ion Ratio Lower Upper
252 100
253 21.9 17.8 26.6
125 8.4 7.4 11.0





#89
Benzo(k)fluoranthene
Concen: 47.455 ng
RT: 22.956 min Scan# 3412
Delta R.T. 0.012 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42




Instrument :
BNA_M
ClientSampleId :
PB162757BSD



Tgt Ion:252 Resp: 418499
Ion Ratio Lower Upper
252 100
253 22.2 17.7 26.5
125 8.5 7.4 11.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#90
Benzo(a)pyrene
Concen: 51.757 ng
RT: 23.621 min Scan# 3525
Delta R.T. 0.012 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42


Tgt Ion:252 Resp: 4186037
Ion Ratio Lower Upper
252 100
253 22.0 17.1 25.7
125 9.1 8.2 12.2





#91
Dibenzo(a,h)anthracene
Concen: 53.036 ng
RT: 26.809 min Scan# 4062
Delta R.T. 0.029 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42




Instrument :
BNA_M
ClientSampleId :
PB162757BSD




Tgt Ion:278 Resp: 436171
Ion Ratio Lower Upper
278 100
139 13.9 12.0 18.0
279 23.6 19.2 28.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 08/20/2024
Supervised By :mohammad ahmed 08/21/2024



#92
Benzo(g,h,i)perylene
Concen: 49.351 ng
RT: 27.691 min Scan# 4217
Delta R.T. 0.029 min
Lab File: BM047251.D
Acq: 19 Aug 2024 13:42



Tgt Ion:276 Resp: 4213855
Ion Ratio Lower Upper
276 100
277 23.6 18.8 28.2
138 19.2 16.2 24.4



Manual Integration Report

Sequence:	BM081024	Instrument	BNA_m

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
SSTDICC005	BM047141.D	4-Nitroaniline	Jagrut	8/12/2024 5:01:33 PM	mohammad	8/13/2024 5:17:30 AM	Peak Integrated by Software
SSTDICC005	BM047141.D	Benzoic acid	Jagrut	8/12/2024 5:01:33 PM	mohammad	8/13/2024 5:17:30 AM	Peak Integrated by Software
SSTDICC010	BM047142.D	4-Nitroaniline	Jagrut	8/12/2024 5:01:36 PM	mohammad	8/13/2024 5:17:30 AM	Peak Integrated by Software
SSTDICC020	BM047143.D	4-Nitroaniline	Jagrut	8/12/2024 5:01:39 PM	mohammad	8/13/2024 5:17:30 AM	Peak Integrated by Software
SSTDICCC040	BM047144.D	Pyridine	Jagrut	8/12/2024 5:01:41 PM	mohammad	8/13/2024 5:17:30 AM	Peak Integrated by Software
SSTDICC050	BM047145.D	Pyridine	Jagrut	8/12/2024 5:01:44 PM	mohammad	8/13/2024 5:17:30 AM	Peak Integrated by Software
SSTDICC060	BM047146.D	Caprolactam	Jagrut	8/12/2024 5:01:46 PM	mohammad	8/13/2024 5:17:30 AM	Peak Integrated by Software
SSTDICC060	BM047146.D	Pyridine	Jagrut	8/12/2024 5:01:46 PM	mohammad	8/13/2024 5:17:30 AM	Peak Integrated by Software
SSTDICC080	BM047147.D	Caprolactam	Jagrut	8/12/2024 5:01:48 PM	mohammad	8/13/2024 5:17:30 AM	Peak Integrated by Software
SSTDICC080	BM047147.D	Pyridine	Jagrut	8/12/2024 5:01:48 PM	mohammad	8/13/2024 5:17:30 AM	Peak Integrated by Software
SSTDICV040	BM047148.D	Pyridine	Jagrut	8/12/2024 5:01:51 PM	mohammad	8/13/2024 5:17:30 AM	Peak Integrated by Software

Manual Integration Report

Sequence:

bm081924

Instrument

BNA_m

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
SSTDCCC040	BM047245.D	2,2"-oxybis(1-Chloropropane)	Jagrut	8/20/2024 2:27:05 PM	mohammad	8/21/2024 1:24:18 AM	Peak Integrated by Software
SSTDCCC040	BM047245.D	Benzyl Alcohol	Jagrut	8/20/2024 2:27:05 PM	mohammad	8/21/2024 1:24:18 AM	Peak Integrated by Software
PB162757BS	BM047250.D	Caprolactam	Jagrut	8/20/2024 2:27:12 PM	mohammad	8/21/2024 1:24:18 AM	Peak Integrated by Software
PB162757BSD	BM047251.D	Caprolactam	Jagrut	8/20/2024 2:27:15 PM	mohammad	8/21/2024 1:24:18 AM	Peak Integrated by Software



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:

bm082224

Instrument

BNA_m

Sample ID

File ID

Parameter

Review By

Review On

Supervised
By

Supervised On

Reason

Instrument ID: **BNA_M**

Daily Analysis Runlog For Sequence/QCBatch ID # BM081024

Review By	Jagrut	Review On	8/12/2024 5:03:24 PM
Supervise By	mohammad	Supervise On	8/13/2024 5:17:30 AM
SubDirectory	BM081024	HP Acquire Method	BNA_M
		HP Processing Method	bm081024
STD. NAME	STD REF.#		
Tune/Reschk	SP6573		
Initial Calibration Stds	SP6550,SP6551,SP6552,SP6553,SP6554,SP6555,SP6556,SP6557		
CCC	SP6553		
Internal Standard/PEM	S12039,10ul/1000ul sample		
ICV/I.BLK	SP6559		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BM047139.D	10 Aug 2024 12:26	RC/JU	Ok
2	SSTDICC2.5	BM047140.D	10 Aug 2024 13:12	RC/JU	Ok
3	SSTDICC005	BM047141.D	10 Aug 2024 13:52	RC/JU	Ok,M
4	SSTDICC010	BM047142.D	10 Aug 2024 14:31	RC/JU	Ok,M
5	SSTDICC020	BM047143.D	10 Aug 2024 15:11	RC/JU	Ok,M
6	SSTDICCC040	BM047144.D	10 Aug 2024 15:51	RC/JU	Ok,M
7	SSTDICC050	BM047145.D	10 Aug 2024 16:31	RC/JU	Ok,M
8	SSTDICC060	BM047146.D	10 Aug 2024 17:11	RC/JU	Ok,M
9	SSTDICC080	BM047147.D	10 Aug 2024 17:51	RC/JU	Ok,M
10	SSTDICV040	BM047148.D	10 Aug 2024 19:10	RC/JU	Ok,M
11	PB162577TB	BM047149.D	10 Aug 2024 19:50	RC/JU	Ok

M : Manual Integration

Instrument ID: **BNA_M**

Daily Analysis Runlog For Sequence/QC Batch ID # BM081924

Review By	Jagrut	Review On	8/20/2024 2:27:36 PM
Supervise By	mohammad	Supervise On	8/21/2024 1:24:18 AM
SubDirectory	BM081924	HP Acquire Method	BNA_M
		HP Processing Method	bm081024
STD. NAME	STD REF.#		
Tune/Reschk	SP6573		
Initial Calibration Stds	SP6550,SP6551,SP6552,SP6553,SP6554,SP6555,SP6556,SP6557		
CCC	SP6553		
Internal Standard/PEM	S12041,10ul/1000ul sample		
ICV/I.BLK	SP6559		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BM047244.D	19 Aug 2024 09:03	RC/JU	Ok
2	SSTDCCC040	BM047245.D	19 Aug 2024 09:42	RC/JU	Ok,M
3	PB162722BL	BM047246.D	19 Aug 2024 10:22	RC/JU	Ok
4	PB162722BS	BM047247.D	19 Aug 2024 11:02	RC/JU	Ok,M
5	PB162722BSD	BM047248.D	19 Aug 2024 11:41	RC/JU	Ok,M
6	PB162757BL	BM047249.D	19 Aug 2024 12:21	RC/JU	Not Ok
7	PB162757BS	BM047250.D	19 Aug 2024 13:01	RC/JU	Ok,M
8	PB162757BSD	BM047251.D	19 Aug 2024 13:42	RC/JU	Ok,M
9	P3596-01	BM047252.D	19 Aug 2024 14:48	RC/JU	ReRun
10	P3596-02	BM047253.D	19 Aug 2024 15:28	RC/JU	ReRun
11	P3580-01	BM047254.D	19 Aug 2024 16:08	RC/JU	Not Ok
12	P3580-02	BM047255.D	19 Aug 2024 16:48	RC/JU	ReRun
13	P3609-01	BM047256.D	19 Aug 2024 17:28	RC/JU	Ok
14	P3609-02	BM047257.D	19 Aug 2024 18:08	RC/JU	Ok
15	P3619-01	BM047258.D	19 Aug 2024 18:49	RC/JU	Not Ok
16	P3643-06	BM047259.D	19 Aug 2024 19:29	RC/JU	Ok,M
17	P3637-01	BM047260.D	19 Aug 2024 20:09	RC/JU	Ok
18	P3643-07	BM047261.D	19 Aug 2024 20:48	RC/JU	Ok,M

M : Manual Integration

Instrument ID: **BNA_M**

Daily Analysis Runlog For Sequence/QC Batch ID # BM082224

Review By	yogesh	Review On	8/23/2024 6:46:01 AM
Supervise By	mohammad	Supervise On	8/23/2024 6:56:43 AM
SubDirectory	BM082224	HP Acquire Method	BNA_M
		HP Processing Method	bm081024
STD. NAME	STD REF.#		
Tune/Reschk	SP6573		
Initial Calibration Stds	SP6550,SP6551,SP6552,SP6553,SP6554,SP6555,SP6556,SP6557		
CCC	SP6553		
Internal Standard/PEM	S12041,10ul/1000ul sample		
ICV/I.BLK	SP6559		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BM047307.D	22 Aug 2024 08:35	RC/JU	Ok
2	SSTDCCC040	BM047308.D	22 Aug 2024 09:14	RC/JU	Ok
3	PB162876BL	BM047309.D	22 Aug 2024 09:54	RC/JU	Ok
4	PB162876BS	BM047310.D	22 Aug 2024 10:33	RC/JU	Ok
5	PB162574BL	BM047311.D	22 Aug 2024 11:13	RC/JU	Ok
6	PB162574BS	BM047312.D	22 Aug 2024 11:52	RC/JU	Not Ok
7	PB162757BL	BM047313.D	22 Aug 2024 12:32	RC/JU	Ok
8	P3671-14	BM047314.D	22 Aug 2024 13:11	RC/JU	Ok
9	P3671-09	BM047315.D	22 Aug 2024 13:51	RC/JU	Ok
10	P3671-13	BM047316.D	22 Aug 2024 14:31	RC/JU	Ok
11	P3671-10	BM047317.D	22 Aug 2024 15:10	RC/JU	Ok
12	P3671-10MS	BM047318.D	22 Aug 2024 15:50	RC/JU	Ok
13	P3671-10MSD	BM047319.D	22 Aug 2024 16:30	RC/JU	Ok
14	P3671-03	BM047320.D	22 Aug 2024 17:10	RC/JU	Ok
15	P3671-04	BM047321.D	22 Aug 2024 17:50	RC/JU	Ok
16	P3671-08	BM047322.D	22 Aug 2024 18:29	RC/JU	Ok,M
17	P3685-01	BM047323.D	22 Aug 2024 19:09	RC/JU	Ok
18	P3685-03	BM047324.D	22 Aug 2024 19:49	RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_M

Daily Analysis Runlog For Sequence/QC Batch ID # BM081024

Review By	Jagrut	Review On	8/12/2024 5:03:24 PM
Supervise By	mohammad	Supervise On	8/13/2024 5:17:30 AM
SubDirectory	BM081024	HP Acquire Method	BNA_M
		HP Processing Method	bm081024

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6550,SP6551,SP6552,SP6553,SP6554,SP6555,SP6556,SP6557
CCC	SP6553
Internal Standard/PEM	S12039,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BM047139.D	10 Aug 2024 12:26		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BM047140.D	10 Aug 2024 13:12		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BM047141.D	10 Aug 2024 13:52	Compound #54 removed from 5ppm	RC/JU	Ok,M
4	SSTDICC010	SSTDICC010	BM047142.D	10 Aug 2024 14:31	Com#77 failing in calibration for both 8270DOD and 8270E	RC/JU	Ok,M
5	SSTDICC020	SSTDICC020	BM047143.D	10 Aug 2024 15:11	Compound #9 Kept on QR	RC/JU	Ok,M
6	SSTDICCC040	SSTDICCC040	BM047144.D	10 Aug 2024 15:51	The Calibration is Good For 8270 DOD and also good for 625.1 Method except com#77	RC/JU	Ok,M
7	SSTDICC050	SSTDICC050	BM047145.D	10 Aug 2024 16:31		RC/JU	Ok,M
8	SSTDICC060	SSTDICC060	BM047146.D	10 Aug 2024 17:11		RC/JU	Ok,M
9	SSTDICC080	SSTDICC080	BM047147.D	10 Aug 2024 17:51	Compound #9 removed from 80ppm	RC/JU	Ok,M
10	SSTDICV040	ICVBM081024	BM047148.D	10 Aug 2024 19:10	Com#77 failing in ICV for Both 8270 DOD and 8270 E and also in 625.1.	RC/JU	Ok,M
11	PB162577TB	PB162577TB	BM047149.D	10 Aug 2024 19:50		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_M

Daily Analysis Runlog For Sequence/QC Batch ID # BM081924

Review By	Jagrut	Review On	8/20/2024 2:27:36 PM
Supervise By	mohammad	Supervise On	8/21/2024 1:24:18 AM
SubDirectory	BM081924	HP Acquire Method	BNA_M
		HP Processing Method	bm081024

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6550,SP6551,SP6552,SP6553,SP6554,SP6555,SP6556,SP6557
CCC	SP6553
Internal Standard/PEM	S12041,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BM047244.D	19 Aug 2024 09:03		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BM047245.D	19 Aug 2024 09:42		RC/JU	Ok,M
3	PB162722BL	PB162722BL	BM047246.D	19 Aug 2024 10:22		RC/JU	Ok
4	PB162722BS	PB162722BS	BM047247.D	19 Aug 2024 11:02		RC/JU	Ok,M
5	PB162722BSD	PB162722BSD	BM047248.D	19 Aug 2024 11:41		RC/JU	Ok,M
6	PB162757BL	PB162757BL	BM047249.D	19 Aug 2024 12:21	Internal Standard Fail	RC/JU	Not Ok
7	PB162757BS	PB162757BS	BM047250.D	19 Aug 2024 13:01		RC/JU	Ok,M
8	PB162757BSD	PB162757BSD	BM047251.D	19 Aug 2024 13:42		RC/JU	Ok,M
9	P3596-01	918-J-WS-081324	BM047252.D	19 Aug 2024 14:48	Internal Standard Fail	RC/JU	ReRun
10	P3596-02	918-J-WS-081324-FD	BM047253.D	19 Aug 2024 15:28	Internal Standard Fail	RC/JU	ReRun
11	P3580-01	926-K1-SD-0-0.5-08122	BM047254.D	19 Aug 2024 16:08	Not match with Sim Analysis - Re-extraction	RC/JU	Not Ok
12	P3580-02	926-K1-SD-0.5-1-08122	BM047255.D	19 Aug 2024 16:48	Internal Standard Fail	RC/JU	ReRun
13	P3609-01	915-J-WS-081424	BM047256.D	19 Aug 2024 17:28		RC/JU	Ok
14	P3609-02	920-J-WS-081424	BM047257.D	19 Aug 2024 18:08		RC/JU	Ok
15	P3619-01	917-J-WS-081424	BM047258.D	19 Aug 2024 18:49	Sample ID Removed From Login Page	RC/JU	Not Ok
16	P3643-06	917-J-SD-0.5-1-081524	BM047259.D	19 Aug 2024 19:29		RC/JU	Ok,M
17	P3637-01	ETGI - 370	BM047260.D	19 Aug 2024 20:09		RC/JU	Ok

Instrument ID: BNA_M

Daily Analysis Runlog For Sequence/QC Batch ID # BM081924

Review By	Jagrut	Review On	8/20/2024 2:27:36 PM		
Supervise By	mohammad	Supervise On	8/21/2024 1:24:18 AM		
SubDirectory	BM081924	HP Acquire Method	BNA_M	HP Processing Method	bm081024

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6550,SP6551,SP6552,SP6553,SP6554,SP6555,SP6556,SP6557
CCC	SP6553
Internal Standard/PEM	S12041,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

18	P3643-07	918-J-SD-0-0.5-081524	BM047261.D	19 Aug 2024 20:48		RC/JU	Ok,M

M : Manual Integration

Instrument ID: BNA_M

Daily Analysis Runlog For Sequence/QC Batch ID # BM082224

Review By	yogesh	Review On	8/23/2024 6:46:01 AM				
Supervise By	mohammad	Supervise On	8/23/2024 6:56:43 AM				
SubDirectory	BM082224	HP Acquire Method	BNA_M	HP Processing Method	bm081024		
STD. NAME		STD REF.#					
Tune/Reschk		SP6573					
Initial Calibration Stds		SP6550,SP6551,SP6552,SP6553,SP6554,SP6555,SP6556,SP6557					
CCC		SP6553					
Internal Standard/PEM		S12041,10ul/1000ul sample					
ICV/I.BLK		SP6559					
Surrogate Standard							
MS/MSD Standard							
LCS Standard							

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BM047307.D	22 Aug 2024 08:35		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BM047308.D	22 Aug 2024 09:14		RC/JU	Ok
3	PB162876BL	PB162876BL	BM047309.D	22 Aug 2024 09:54		RC/JU	Ok
4	PB162876BS	PB162876BS	BM047310.D	22 Aug 2024 10:33		RC/JU	Ok
5	PB162574BL	PB162574BL	BM047311.D	22 Aug 2024 11:13		RC/JU	Ok
6	PB162574BS	PB162574BS	BM047312.D	22 Aug 2024 11:52	Surrogate Fail and Com#9 Recovery is zero	RC/JU	Not Ok
7	PB162757BL	PB162757BL	BM047313.D	22 Aug 2024 12:32		RC/JU	Ok
8	P3671-14	S-912-J-SO-0.5-1-0819	BM047314.D	22 Aug 2024 13:11		RC/JU	Ok
9	P3671-09	S-928-K1-SO-0.5-1-0819	BM047315.D	22 Aug 2024 13:51		RC/JU	Ok
10	P3671-13	S-912-J-SO-0.5-1-0819	BM047316.D	22 Aug 2024 14:31		RC/JU	Ok
11	P3671-10	S-928-K1-SO-0.5-1-0819	BM047317.D	22 Aug 2024 15:10		RC/JU	Ok
12	P3671-10MS	S-928-K1-SO-0.5-1-0819	BM047318.D	22 Aug 2024 15:50		RC/JU	Ok
13	P3671-10MSD	S-928-K1-SO-0.5-1-0819	BM047319.D	22 Aug 2024 16:30		RC/JU	Ok
14	P3671-03	S-911-K1-SO-0.5-1-0819	BM047320.D	22 Aug 2024 17:10		RC/JU	Ok
15	P3671-04	S-911-K1-SO-0.5-1-0819	BM047321.D	22 Aug 2024 17:50		RC/JU	Ok
16	P3671-08	S-905-J-SO-0.5-1.0-0819	BM047322.D	22 Aug 2024 18:29	SIM Analysis not required	RC/JU	Ok,M
17	P3685-01	Stock Pile-1	BM047323.D	22 Aug 2024 19:09		RC/JU	Ok

Instrument ID: BNA_M

Daily Analysis Runlog For Sequence/QC Batch ID # BM082224

Review By	yogesh	Review On	8/23/2024 6:46:01 AM		
Supervise By	mohammad	Supervise On	8/23/2024 6:56:43 AM		
SubDirectory	BM082224	HP Acquire Method	BNA_M	HP Processing Method	bm081024

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6550,SP6551,SP6552,SP6553,SP6554,SP6555,SP6556,SP6557
CCC	SP6553
Internal Standard/PEM	S12041,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

18	P3685-03	Stock Pile-2	BM047324.D	22 Aug 2024 19:49		RC/JU	Ok

M : Manual Integration

SOP ID: M3510C,3580A-Extraction SVOC-20			
Clean Up SOP #:	N/A	Extraction Start Date :	08/15/2024
Matrix :	Water	Extraction Start Time :	08:23
Welgh By:	N/A	Extraction End Date :	08/15/2024
Balance check:	N/A	Filter By:	RS
Balance ID:	N/A	Extraction End Time :	13:20
pH Strip Lot#:	E3574	pH Meter ID:	N/A
		Concentration By:	EH
		Supervisor By :	rajesh
Extraction Method: ☒ Separatory Funne		☐ Continous Liquid/Liquid	
		☐ Sonication	
		☐ Waste Dilution	
		☐ Soxhlet	

Standarded Name	MLS USED	Concentration ug/mL	STD REF. # FROM LOG
Spike Sol 1	1.0ML	50/100 PPM	SP6604
Surrogate	1.0ML	100/150 PPM	SP6524
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
Methylene Chloride	N/A	E3787
Baked Na2SO4	N/A	EP2521
10N NaoH	N/A	EP2523
H2SO4 1:1	N/A	EP2524
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

1.5 ML Vial lot# 2210673. pH Adjusted<2 with 1:1 H2SO4 &>11 with 10 N NaOH.P3619-01 Added in batch at 10:20.

KD Bath ID:	Water bath -01	Envap ID:	NEVAP-02
KD Bath Temperature:	60 °C	Envap Temperature:	40 °C

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
08/15/24	RP (Ext. Lab)	RC/SVOC
13:25	Preparation Group	Analysis Group

Analytical Method: M3510C,3580A-Extraction SVOC-20

Concentration Date: 08/15/2024

Sample ID	Client Sample ID	Test	g / <u>g</u> mL	PH	Surr/Spike By:		Final Vol.(mL)	JarID	Comments	Prep Pos
					AddedBy	VerifiedBy				
PB162757BL	SBLK757	SVOCMS Group6	1000	6	RUPESH	rajesh	1			SEP-1
PB162757BS	SLCS757	SVOCMS Group6	1000	6	RUPESH	rajesh	1			2
PB162757BS D	SLCSD757	SVOCMS Group6	1000	6	RUPESH	rajesh	1			3
P3609-01	915-J-WS-081424	SVOCMS Group6	960	6	RUPESH	rajesh	1	E		4
P3609-02	920-J-WS-081424	SVOCMS Group6	1000	6	RUPESH	rajesh	1	E		5
P3619-01	917-J-WS-081424	SVOCMS Group6	980	6	RUPESH	rajesh	1	E		6

* Extracts relinquished on the same date as received.

8/15/24

162797
8-23 Bw

WORKLIST(Hardcopy Internal Chain)

WorkList Name : P3609

WorkList ID : 182678

Department : Extraction

Date : 08-15-2024 08:06:34

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3609-01	915-J-WS-081424	Water	SVOCMS Group3	Cool 4 deg C	JACO05	D21	08/14/2024	8270-Modified
P3609-01	915-J-WS-081424	Water	SVOCMS Group6	Cool 4 deg C	JACO05	D21	08/14/2024	8270E
P3609-02	920-J-WS-081424	Water	SVOCMS Group3	Cool 4 deg C	JACO05	D21	08/14/2024	8270-Modified
P3609-02	920-J-WS-081424	Water	SVOCMS Group6	Cool 4 deg C	JACO05	D21	08/14/2024	8270E

Date/Time 08/15/24 8:20
Raw Sample Received by: Rs (See Log)
Raw Sample Relinquished by: Rm Sm

Date/Time 08/15/24 8:05
Raw Sample Received by: Rm Sm
Raw Sample Relinquished by: Rs (See Log)

02
168787
10/20

WORKLIST(Hardcopy Internal Chain)

WorkList Name : p3619

WorkList ID : 182686

Department : Extraction

Date : 08-15-2024 10:20:14

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3619-01	917-J-WS-081424	Water	SVOCMS Group3	Cool 4 deg C	JACO05	G11	08/14/2024	8270-Modified
P3619-01	917-J-WS-081424	Water	SVOCMS Group6	Cool 4 deg C	JACO05	G11	08/14/2024	8270E
P3620-01	915-J-WPO-0.25-081424	Water	SVOC-SIMGroup1	Cool 4 deg C	JACO05	G11	08/14/2024	8270-Modified
P3620-02	920-J-WPO-0.25-081424	Water	SVOC-SIMGroup1	Cool 4 deg C	JACO05	G11	08/14/2024	8270-Modified

Date/Time 08/15/24 10:20

Raw Sample Received by: RS (Serling)

Raw Sample Relinquished by: JDCSM

Date/Time 08/15/24 10:30

Raw Sample Received by: JDCSM

Raw Sample Relinquished by: RS (Serling)

Prep Standard - Chemical Standard Summary

Order ID : P3609

Test : SVOCMS Group6

Prepbatch ID : PB162757,

Sequence ID/Qc Batch ID: BM081924,BM082224,

Standard ID :

EP2521,EP2523,EP2524,SP6524,SP6549,SP6550,SP6551,SP6552,SP6553,SP6554,SP6555,SP6556,SP6557,SP6558,SP6559,SP6573,SP6604,

Chemical ID :

10ul/1000ul

sample,E3551,E3657,E3744,E3746,E3768,E3780,E3787,M5037,S10102,S10247,S10398,S10583,S10972,S10973,S10974,S10975,S10976,S10977,S10996,S10997,S10998,S10999,S11000,S11001,S11002,S11003,S11012,S11092,S11096,S11102,S11138,S11141,S11148,S11434,S11555,S11561,S11566,S11567,S11766,S11767,S11768,S11769,S11770,S12033,S12041,S12076,S12097,S12098,S12099,S12100,S12101,S12102,S12103,S12104,S12105,S12112,S12117,S12118,S12119,S12120,S12121,S12122,S12123,S12124,S12125,S12126,S12449,S12450,S12451,S12452,S12453,W3112,



<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3923	Baked Sodium Sulfate	EP2521	08/10/2024	01/03/2025	Rajesh Parikh	Extraction_SC ALE_2 (EX-SC-2)	None	RUPESHKUMAR SHAH 08/10/2024
<u>FROM</u> 4000.00000gram of E3551 = Final Quantity: 4000.000 gram								

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1874	10 N SODIUM HYDROXIDE SOLN	EP2523	08/14/2024	02/14/2025	Rajesh Parikh	Extraction_SCALE_2 (EX-SC-2)	None	RUPESHKUMAR SHAH 08/14/2024
<u>FROM</u>	1000.00000ml of W3112 + 400.00000gram of E3657 = Final Quantity: 1000.000 ml							

Extractions STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
314	1.1 H2SO4 SOLN	EP2524	08/14/2024	12/15/2024	Rajesh Parikh	None	None	RUPESHKUMAR SHAH 08/14/2024

FROM 1000.00000ml of M5037 + 1000.00000ml of W3112 = Final Quantity: 2000.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
19	8270/CLP Surrogate Solution, 100 PPM BN/150 PPM ACID	SP6524	05/31/2024	11/29/2024	Jagrut Upadhyay	None	None	Yogesh Patel 06/13/2024

FROM 1930.00000ml of E3744 + 2.90000ml of S11003 + 3.00000ml of S10977 + 5.30000ml of S10996 + 5.30000ml of S10997 + 5.30000ml of S10998 + 5.30000ml of S10999 + 5.30000ml of S11000 + 5.30000ml of S11001 + 5.30000ml of S11002 + 5.40000ml of S10972 + 5.40000ml of S10973 + 5.40000ml of S10974 + 5.40000ml of S10975 + 5.40000ml of S10976 = Final Quantity: 2000.000 ml

[illegible]

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
413	80 ng BNA ICC, 80 PPM	SP6550	07/09/2024	08/26/2024	Jagrut Upadhyay	None	None	mohammad ahmed 07/09/2024
<u>FROM</u> 0.01000ml of S12033 + 0.20000ml of E3746 + 0.80000ml of SP6549 = Final Quantity: 1.010 ml								

SVOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
412	60 ng BNA ICC, 60 PPM	SP6551	07/09/2024	08/26/2024	Jagrut Upadhyay	None	None	mohammad ahmed
								07/09/2024

FROM 0.01000ml of S12033 + 0.40000ml of E3746 + 0.60000ml of SP6549 = Final Quantity: 1.010 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
411	50 ng BNA ICC, 50 PPM	SP6552	07/09/2024	08/26/2024	Jagrut Upadhyay	None	None	mohammad ahmed
								07/09/2024

FROM 0.01000ml of S12033 + 0.50000ml of E3746 + 0.50000ml of SP6549 = Final Quantity: 1.010 ml

SVOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
410	40 ng BNA ICC, 40 PPM	SP6553	07/09/2024	08/26/2024	Jagrut Upadhyay	None	None	mohammad ahmed
								07/09/2024

FROM 0.01000ml of S12033 + 0.60000ml of E3746 + 0.40000ml of SP6549 = Final Quantity: 1.010 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3678	20 ng BNA ICC, 20 PPM	SP6554	07/09/2024	08/26/2024	Jagrut Upadhyay	None	None	mohammad ahmed
								07/09/2024

FROM 0.01000ml of S12033 + 0.80000ml of E3746 + 0.20000ml of SP6549 = Final Quantity: 1.010 ml

SVOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
408	10 ng BNA ICC, 10 PPM	SP6555	07/09/2024	08/26/2024	Jagrut Upadhyay	None	None	mohammad ahmed
								07/09/2024

FROM 0.01000ml of S12033 + 0.90000ml of E3746 + 0.10000ml of SP6549 = Final Quantity: 1.010 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
407	5 ng BNA ICC, 5 PPM	SP6556	07/09/2024	08/26/2024	Jagrut Upadhyay	None	None	mohammad ahmed
								07/09/2024

FROM 0.01000ml of S12033 + 0.95000ml of E3746 + 0.05000ml of SP6549 = Final Quantity: 1.010 ml

SVOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
175	2.5 ng BNA ICC, 2.5 PPM	SP6557	07/09/2024	08/26/2024	Jagrut Upadhyay	None	None	mohammad ahmed
								07/09/2024

FROM 0.01000ml of S12033 + 0.50000ml of E3746 + 0.50000ml of SP6556 = Final Quantity: 1.010 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
18	Second Source Calibration Stock Standard, 100 PPM,	SP6558	07/09/2024	11/30/2024	Jagrut Upadhyay	None	None	mohammad ahmed
	(8270/625/CLP)							07/11/2024

FROM 0.04000ml of S10977 + 0.08000ml of S11003 + 0.10000ml of S11766 + 0.20000ml of S11566 + 0.20000ml of S12097 +
0.20000ml of S12117 + 1.18000ml of E3768 = Final Quantity: 2.000 ml

SVOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
416	40 ng BNA ICV, 40 PPM	SP6559	07/09/2024	11/30/2024	Jagrut Upadhyay	None	None	mohammad ahmed
								07/11/2024

FROM 0.01000ml of S12033 + 0.60000ml of E3768 + 0.40000ml of SP6558 = Final Quantity: 1.010 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3895	50 ug/ml DFTPP 8270E	SP6573	07/15/2024	01/08/2025	Rahul Chavli	None	None	Yogesh Patel
								07/17/2024

FROM 1.00000ml of S10247 + 19.00000ml of E3768 = Final Quantity: 20.000 ml



<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
171	8270/625 Spike Solution, 50/100 PPM	SP6604	08/12/2024	10/26/2024	Rahul Chavli	None	None	Yogesh Patel 08/21/2024

FROM 0.20000ml of S12126 + 0.40000ml of S10398 + 0.40000ml of S10583 + 0.40000ml of S11012 + 0.40000ml of S11138 + 0.40000ml of S11141 + 0.50000ml of S12105 + 0.50000ml of S12453 + 0.70000ml of S12097 + 0.80000ml of S11566 + 0.80000ml of S12118 + 1.20000ml of S11555 + 1.20000ml of S11561 + 1.20000ml of S11767 + 1.20000ml of S11769 + 1.20000ml of S12099 + 1.20000ml of S12101 + 1.20000ml of S12104 + 1.20000ml of S12119 + 1.20000ml of S12121 + 1.20000ml of S12449 + 1.20000ml of S12451 + 1.30000ml of S11567 + 1.30000ml of S11768 + 1.30000ml of S11770 + 1.30000ml of S12098 + 1.30000ml of S12100 + 1.30000ml of S12102 + 1.30000ml of S12103 + 1.30000ml of S12120 + 1.30000ml of S12122 + 1.30000ml of S12123 + 1.30000ml of S12124 + 1.30000ml of S12450 + 1.30000ml of S12452 + 1.40000ml of S12125 + 163.00000ml of E3780 = Final Quantity: 200.000 ml

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	PC19631-100 / SODIUM SULFATE, ANHYDROUS, PEST GRADE, 1	313201	01/03/2025	01/03/2024 / Rajesh	07/20/2023 / Rajesh	E3551

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	PC19510-5 / Sodium Hydroxide Pellets 2.5 Kg, Pk of 4	23B1556310	12/31/2025	12/04/2023 / Rajesh	12/01/2023 / Rajesh	E3657

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9254-03 / Acetone, Ultra Resi (cs/4x4L)	23H14626005	11/29/2024	05/29/2024 / Rajesh	05/23/2024 / Rajesh	E3744

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9644-A4 / Methylene Chloride,U-Resi, Cycle-Tainer (215L)	24C0162011	11/25/2024	05/25/2024 / Rajesh	05/08/2024 / Rajesh	E3746

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9644-A4 / Methylene Chloride,U-Resi, Cycle-Tainer (215L)	24E2462004	01/08/2025	07/08/2024 / Rajesh	06/21/2024 / Rajesh	E3768

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9254-03 / Acetone, Ultra Resi (cs/4x4L)	23H1462005	02/10/2025	08/10/2024 / Rajesh	07/24/2024 / Rajesh	E3780

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9644-A4 / Methylene Chloride,U-Resi, Cycle-Tainer (215L)	24G0862022	02/13/2025	08/13/2024 / Rajesh	08/07/2024 / Rajesh	E3787

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9673-33 / Sulfuric Acid, Instra-Analyzed (cs/6c2.5L)	0000250349	12/15/2024	01/06/2022 / mohan	09/18/2021 / mohan	M5037

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CPI International	Z-112090-04 / CLP Acid Surrogate Solution, 7500 mg/L, 1ml	440246	09/29/2024	03/29/2024 / Jagrut	12/09/2021 / Christian	S10102

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31615 / SV Mixture, GC/MS Tuning Mixture, CH ₂ Cl ₂ , 1mL,	A0182667	01/15/2025	07/15/2024 / Rahul	03/18/2022 / Christian	S10247

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555871 / Custom Standard, 4-nitrophenol Std [CS 5238-4]	A0185300	10/26/2024	04/26/2024 / Rahul	05/18/2022 / Christian	S10398

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555868 / Custom Standard, Benzidine Std [CS 5328-1]	A0186373	02/12/2025	08/12/2024 / Rahul	07/05/2022 / Christian	S10583

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31087 / Acid Surrogate 10,000ug/ml,methanol,5ml/ ampul	A0188108	08/31/2030	05/31/2024 / Jagrut	12/28/2022 / Christian	S10972

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31087 / Acid Surrogate 10,000ug/ml,methanol,5ml/ ampul	A0188108	11/30/2024	05/31/2024 / Jagrut	12/28/2022 / Christian	S10973

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31087 / Acid Surrogate 10,000ug/ml,methanol,5ml/ ampul	A0188108	11/30/2024	05/31/2024 / Jagrut	12/28/2022 / Christian	S10974

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31087 / Acid Surrogate 10,000ug/ml,methanol,5ml/ ampul	A0188108	11/30/2024	05/31/2024 / Jagrut	12/28/2022 / Christian	S10975

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31087 / Acid Surrogate 10,000ug/ml,methanol,5ml/ ampul	A0188108	11/30/2024	05/31/2024 / Jagrut	12/28/2022 / Christian	S10976

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31087 / Acid Surrogate 10,000ug/ml,methanol,5ml/ ampul	A0188108	11/30/2024	05/31/2024 / Jagrut	12/28/2022 / Christian	S10977

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31086 / Base Neutral Surrogate 5000ug/ml,CH ₂ Cl ₂ ,5ml	A0189418	08/31/2028	05/31/2024 / Jagrut	12/28/2022 / Christian	S10996

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31086 / Base Neutral Surrogate 5000ug/ml,CH ₂ Cl ₂ ,5ml	A0189418	11/30/2024	05/31/2024 / Jagrut	12/28/2022 / Christian	S10997

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31086 / Base Neutral Surrogate 5000ug/ml,CH ₂ Cl ₂ ,5ml	A0189418	11/30/2024	05/31/2024 / Jagrut	12/28/2022 / Christian	S10998

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31086 / Base Neutral Surrogate 5000ug/ml,CH ₂ Cl ₂ ,5ml	A0189418	11/30/2024	05/31/2024 / Jagrut	12/28/2022 / Christian	S10999

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31086 / Base Neutral Surrogate 5000ug/ml,CH ₂ Cl ₂ ,5ml	A0189418	11/30/2024	05/31/2024 / Jagrut	12/28/2022 / Christian	S11000

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31086 / Base Neutral Surrogate 5000ug/ml,CH ₂ Cl ₂ ,5ml	A0189418	11/30/2024	05/31/2024 / Jagrut	12/28/2022 / Christian	S11001

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31086 / Base Neutral Surrogate 5000ug/ml,CH ₂ Cl ₂ ,5ml	A0189418	11/30/2024	05/31/2024 / Jagrut	12/28/2022 / Christian	S11002

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31086 / Base Neutral Surrogate 5000ug/ml,CH ₂ Cl ₂ ,5ml	A0189418	11/30/2024	05/31/2024 / Jagrut	12/28/2022 / Christian	S11003

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555872 / Custom Standard, pentachlorophenol Std [CS 5328-5]	A0193449	10/26/2024	04/26/2024 / Rahul	01/13/2023 / Christian	S11012

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CPI International	Z-110817-01 / Custom 8270 Mix, 4-55, 1000 mg/L, 1 ml, (Maximum Expiration: 90 Days)	414125	01/09/2025	07/09/2024 / Jagrut	02/07/2023 / Christian	S11092

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CPI International	z-110381-01 / 8270 Calibration Solution, 76-1, 500 & 1,000 mg/L, 1ml	495831	01/09/2025	07/09/2024 / Jagrut	02/07/2023 / Christian	S11096

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CPI International	Z-010442-07 / Benzaldehyde Solution, 1000 mg/L, 1.3 ml, (Maximum Expiration: 90 Days)	495833	01/09/2025	07/09/2024 / Jagrut	02/07/2023 / Christian	S11102

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555870 / Custom Standard, 2,4-dinitrophenol Std [CS 5328-3]	A0194698	02/12/2025	08/12/2024 / Rahul	02/20/2023 / Christian	S11138

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555869 / Custom Standard, hexachlorocyclopentadiene Std [CS 5328-2]	A0194702	02/12/2025	08/12/2024 / Rahul	02/20/2023 / Christian	S11141

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CPI International	Z-010074-07 / 3,3'-Dichlorobenzidine Solution, 1,000 mg/L, 1 ml, (Maximum Expiration: 180 days)	406703	01/09/2025	07/09/2024 / Jagrut	03/06/2023 / Christian	S11148

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CPI International	Z-110094-02 / CLP Base/Neutral Surrogate Solution, 5000 mg/L, 1ml	503442	08/26/2024	07/09/2024 / Jagrut	07/26/2023 / yogesh	S11434

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555223 / Custom 8270 Plus Std #1 [2nd lot at \$100 per ampul if requested - contact ARM with Request] [CS 4978-1]	A0201940	02/12/2025	08/12/2024 / Rahul	09/18/2023 / Kiran	S11555

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555223 / Custom 8270 Plus Std #1 [2nd lot at \$100 per ampul if requested - contact ARM with Request] [CS 4978-1]	A0201940	02/12/2025	08/12/2024 / Rahul	09/18/2023 / Kiran	S11561

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555223 / Custom 8270 Plus Std #1 [2nd lot at \$100 per ampul if requested - contact ARM with Request]	A0201940	12/05/2024	06/05/2024 / Rahul	09/18/2023 / Kiran	S11566

[CS 4978-1]

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555223 / Custom 8270 Plus Std #1 [2nd lot at \$100 per ampul if requested - contact ARM with Request]	A0201940	02/12/2025	08/12/2024 / Rahul	09/18/2023 / Kiran	S11567

[CS 4978-1]

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31853 / 1,4-Dioxane, 2000 ug/ml , Solvent: Methylene Chloride	A0196453	12/14/2024	06/14/2024 / Rahul	11/21/2023 / Rahul	S11766

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31853 / 1,4-Dioxane, 2000 ug/ml , Solvent: Methylene Chloride	A0196453	01/12/2025	07/12/2024 / Rahul	11/21/2023 / Rahul	S11767

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31853 / 1,4-Dioxane, 2000 ug/ml , Solvent: Methylene Chloride	A0196453	02/12/2025	08/12/2024 / Rahul	11/21/2023 / Rahul	S11768

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31853 / 1,4-Dioxane, 2000 ug/ml , Solvent: Methylene Chloride	A0196453	02/12/2025	08/12/2024 / Rahul	11/21/2023 / Rahul	S11769

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31853 / 1,4-Dioxane, 2000 ug/ml , Solvent: Methylene Chloride	A0196453	02/12/2025	08/12/2024 / Rahul	11/21/2023 / Rahul	S11770

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31206 / SV Mix, CLP method, Internal Std, 2000ug/mL, CH2Cl2, 1mL	A0201320	01/01/2025	07/01/2024 / Rahul	12/21/2023 / Rahul	S12033

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31206 / SV Mix, CLP method, Internal Std, 2000ug/mL, CH2Cl2, 1mL	A0201320	02/13/2025	08/13/2024 / anahy	12/21/2023 / Rahul	S12041

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CPI International	Z-110816-01 / Custom 8270 Mix, 4-79, 1000 mg/L, 1 mL, (Maximum Expiration: 180 Days)	414127	01/09/2025	07/09/2024 / Jagrut	01/31/2024 / Rahul	S12076

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555224 / Custom 8270 Plus Std #2 [2nd lot at \$85 per ampul if requested - contact ARM with Request]	A0207706	01/09/2025	07/09/2024 / Jagrut	02/05/2024 / Rahul	S12097

[CS 4978-2]

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555224 / Custom 8270 Plus Std #2 [2nd lot at \$85 per ampul if requested - contact ARM with Request]	A0207706	02/12/2025	08/12/2024 / Rahul	02/05/2024 / Rahul	S12098

[CS 4978-2]

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555224 / Custom 8270 Plus Std #2 [2nd lot at \$85 per ampul if requested - contact ARM with Request]	A0207706	02/12/2025	08/12/2024 / Rahul	02/05/2024 / Rahul	S12099

[CS 4978-2]

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555224 / Custom 8270 Plus Std #2 [2nd lot at \$85 per ampul if requested - contact ARM with Request]	A0207706	02/12/2025	08/12/2024 / Rahul	02/05/2024 / Rahul	S12100

[CS 4978-2]

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555224 / Custom 8270 Plus Std #2 [2nd lot at \$85 per ampul if requested - contact ARM with Request]	A0207706	02/12/2025	08/12/2024 / Rahul	02/05/2024 / Rahul	S12101

[CS 4978-2]

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555224 / Custom 8270 Plus Std #2 [2nd lot at \$85 per ampul if requested - contact ARM with Request]	A0207706	02/12/2025	08/12/2024 / Rahul	02/05/2024 / Rahul	S12102

[CS 4978-2]

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555224 / Custom 8270 Plus Std #2 [2nd lot at \$85 per ampul if requested - contact ARM with Request]	A0207706	02/12/2025	08/12/2024 / Rahul	02/05/2024 / Rahul	S12103

[CS 4978-2]

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555224 / Custom 8270 Plus Std #2 [2nd lot at \$85 per ampul if requested - contact ARM with Request]	A0207706	02/12/2025	08/12/2024 / Rahul	02/05/2024 / Rahul	S12104

[CS 4978-2]

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555224 / Custom 8270 Plus Std #2 [2nd lot at \$85 per ampul if requested - contact ARM with Request]	A0207706	02/12/2025	08/12/2024 / Rahul	02/05/2024 / Rahul	S12105

[CS 4978-2]

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CPI International	z-010223-01 / 1,4-Dioxane Solution, 2,000mg/L, 1ml	454157	01/09/2025	07/09/2024 / Jagrut	03/08/2024 / Rahul	S12112

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH ₂ Cl ₂ [New Solvent 100% CH ₂ Cl ₂]	A0203726	12/05/2024	06/05/2024 / Rahul	03/15/2024 / Rahul	S12117

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH ₂ Cl ₂ [New Solvent 100% CH ₂ Cl ₂]	A0203726	04/30/2025	01/10/2025 / Jagrut	03/15/2024 / Rahul	S12118

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH ₂ Cl ₂ [New Solvent 100% CH ₂ Cl ₂]	A0203726	02/12/2025	08/12/2024 / Rahul	03/15/2024 / Rahul	S12119

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH ₂ Cl ₂ [New Solvent 100% CH ₂ Cl ₂]	A0203726	02/12/2025	08/12/2024 / Rahul	03/15/2024 / Rahul	S12120

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH ₂ Cl ₂ [New Solvent 100% CH ₂ Cl ₂]	A0203726	02/12/2025	08/12/2024 / Rahul	03/15/2024 / Rahul	S12121

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH ₂ Cl ₂ [New Solvent 100% CH ₂ Cl ₂]	A0203726	02/12/2025	08/12/2024 / Rahul	03/15/2024 / Rahul	S12122

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH ₂ Cl ₂ [New Solvent 100% CH ₂ Cl ₂]	A0203726	02/12/2025	08/12/2024 / Rahul	03/15/2024 / Rahul	S12123

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH ₂ Cl ₂ [New Solvent 100% CH ₂ Cl ₂]	A0203726	02/12/2025	08/12/2024 / Rahul	03/15/2024 / Rahul	S12124

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH ₂ Cl ₂ [New Solvent 100% CH ₂ Cl ₂]	A0203726	02/12/2025	08/12/2024 / Rahul	03/15/2024 / Rahul	S12125

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH ₂ Cl ₂ [New Solvent 100% CH ₂ Cl ₂]	A0203726	02/12/2025	08/12/2024 / Rahul	03/15/2024 / Rahul	S12126

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555223 / Custom 8270 Plus Std #1 [2nd lot at \$100 per ampul if requested - contact ARM with Request]	A0214021	02/12/2025	08/12/2024 / Rahul	07/23/2024 / RAHUL	S12449

[CS 4978-1]

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555223 / Custom 8270 Plus Std #1 [2nd lot at \$100 per ampul if requested - contact ARM with Request]	A0214021	02/12/2025	08/12/2024 / Rahul	07/23/2024 / RAHUL	S12450

[CS 4978-1]

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555223 / Custom 8270 Plus Std #1 [2nd lot at \$100 per ampul if requested - contact ARM with Request]	A0214021	02/12/2025	08/12/2024 / Rahul	07/23/2024 / RAHUL	S12451

[CS 4978-1]

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555223 / Custom 8270 Plus Std #1 [2nd lot at \$100 per ampul if requested - contact ARM with Request]	A0214021	02/12/2025	08/12/2024 / Rahul	07/23/2024 / RAHUL	S12452

[CS 4978-1]

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555223 / Custom 8270 Plus Std #1 [2nd lot at \$100 per ampul if requested - contact ARM with Request]	A0214021	02/12/2025	08/12/2024 / Rahul	07/23/2024 / RAHUL	S12453

[CS 4978-1]

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	DIW / DI Water	Daily Lab-Certified	07/03/2029	07/03/2024 / lwona	07/03/2024 / lwona	W3112



5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received: _____

Certificate of Analysis

Rev 0

Page 1 of 1

Catalog No.:	Lot No.:	Storage:	Solvent:	Exp. Date:	Description:
Z-010074-07	406703	≤ -10 °C	Methylene Chloride	3/30/2025	3,3'-Dichlorobenzidine Solution, 1,000 mg/L, 1 mL

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
3,3'-dichlorobenzidine	91-94-1	99.5	74.3.26P	989 ± 7.53

Received on
02/07/23
by
CG
S11084
to
S11088

*Not a certified value

Certified By: _____

Jacob Mulloy
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values
listed are determined gravimetrically.



5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received: _____

Certificate of Analysis

Rev 0

Page 1 of 1

Catalog No.:	Lot No.:	Storage:	Solvent:	Exp. Date:	Description:
Z-110817-01	414125	≤ -10 °C	Methylene Chloride	6/21/2025	Custom 8270 Mix, 4-55, 1000 mg/L, 1 mL

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
acetophenone	98-86-2	99.2	85.8.1P	998 ± 11.5
benzoic acid	65-85-0	100	123.7.1P	1010 ± 5.88
biphenyl	92-52-4	99.9	366.29.1P	999 ± 5.82
1,2,4,5-tetrachlorobenzene	95-94-3	99.7	53.7.2P	993 ± 5.79

Received on
02/07/23
by
CG
S11089
to
S11093

*Not a certified value

Manufactured by o2si smart solutions, Accredited to ISO 9001:2008 by NSF and ISO/IEC 17025:2005 (Certification No. 3031.01) and ISO Guide 34:2009 (Certification No. 3031.02) by A2LA

Certified By: _____

Shane Overcash
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetrically.



5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received: _____

Certificate of Analysis

Rev 0

Page 1 of 1

Catalog No.: Lot No.: **Storage:** **Solvent:** **Exp. Date:** **Description:**
Z-112090 440246 $\leq -10^{\circ}\text{C}$ Methylene Chloride 2/16/2026 CLP Acid Surrogate Solution, 7,500 mg/L, 1 mL
-04

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
2-chlorophenol-d ₄	93951-73-6	99.3	248.12.7P	7487 $\pm$ 17.2
2-fluorophenol	367-12-4	99.8	10.7.3.3P	7513 $\pm$ 17.26
phenol-d ₆	13127-88-3	99.9	949.120.8P	7481 $\pm$ 17.19
2,4,6-tribromophenol	118-79-6	99.8	12.1.6P	7469 $\pm$ 17.17

Received on

02/25/21

by
CG

S9236
to

S9240

*Not a certified value

Manufactured by o2si smart solutions, Accredited to ISO 9001:2008 by NSF and ISO/IEC 17025:2005 (Certification No. 3031.01) and ISO Guide 34:2009 (Certification No. 3031.02) by A2LA

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetrically.

Certified By: _____

Erica Castiglione
Chemist



5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Receivalon
02/07/23 by CG

Sheet S11096
to
S11099

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received: _____

Certificate of Analysis

Rev 0

Page 1 of 4

Catalog No.:	Lot No.:	Storage:	Solvent:	Exp. Date:	Description:
Z-110381-01	495831	≤ -10 °C	Methylene Chloride	10/30/2027	Method 8270 Calibration Solution, 76-1, 500 & 1,000 mg/L, 1 mL

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
acenaphthene	83-32-9	99.9	13.1.5P	1003 ± 17.27
acenaphthylene	208-96-8	97.6	14.290.1P	999.8 ± 17.22
aniline	62-53-3	99.9	64.7.1P	995 ± 17.13
anthracene	120-12-7	99.5	15.7.1P	1001 ± 17.24
azobenzene	103-33-3	98.1	252.7.2P	999.1 ± 17.21
benzo[a]anthracene	56-55-3	100	16.7.3P	1001 ± 17.24
benzo[b]fluoranthene	205-99-2	99.8	17.421.3P	1001 ± 19.91
benzo[k]fluoranthene	207-08-9	98.9	18.421.4P	1001 ± 17.92
benzo[ghi]perylene	191-24-2	93	19.286.4P	999.6 ± 19.88
benzo[a]pyrene	50-32-8	97	20.286.2P	999.1 ± 26.35
benzyl alcohol	100-51-6	99.9	65.18.1P	1001 ± 17.24
bis(2-chloroethoxy)methane	111-91-1	99.1	31.3.15P	999.7 ± 17.89
bis(2-chloroethyl)ether	111-44-4	99.8	32.7.1P	1001 ± 17.23
bis(2-chloro-1-methylethyl) ether	108-60-1	99.5	34.3.13P	999.5 ± 17.89
bis(2-ethylhexyl)adipate	103-23-1	99.5	874.7.1P	999.5 ± 17.21
bis(2-ethylhexyl)phthalate	117-81-7	99.4	33.29.1P	998.8 ± 19.86
4-bromophenyl phenyl ether	101-55-3	99.4	35.7.1P	999.1 ± 17.2
butyl benzyl phthalate	85-68-7	98.4	36.1.6P	984.7 ± 19.58
carbazole	86-74-8	99.4	239.7.2P	1000 ± 17.22

*Not a certified value

Manufactured by o2si smart solutions, Accredited to ISO 9001:2008 by NSF and ISO/IEC 17025:2005 (Certification No. 3031.01) and ISO Guide 34:2009 (Certification No. 3031.02) by A2LA

Brianna Smith

Certified By: _____

Brianna Smith
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetrically.

Certificate of Analysis

Page 4 of 4

Catalog No.: Z-110381-01

Lot No.: 495831

Expiration Date: 10/30/2027

<u>Compound</u>	<u>CAS No.</u>	<u>Purity (%)</u>	<u>Compound Lot No.</u>	<u>Concentration, mg/L</u>
1,2,4-trichlorobenzene	120-82-1	99.6	54.29.1P	1000 ± 17.22
2,4,5-trichlorophenol	95-95-4	96.5	121.7.1.1P	1000 ± 17.22
2,4,6-trichlorophenol	88-06-2	99.6	113.7.1P	1002 ± 17.25

*Not a certified value

Manufactured by o2si smart solutions, Accredited to ISO 9001:2008 by NSF and ISO/IEC 17025:2005 (Certification No. 3031.01) and ISO Guide 34:2009 (Certification No. 3031.02) by A2LA

Certified By: _____



Briana Smith
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetrically.



5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received: _____

Certificate of Analysis

Rev 0

Page 1 of 1

Catalog No.:	Lot No.:	Storage:	Solvent:	Exp. Date:	Description:
Z-010442-07	495833	≤ -10 °C	Methylene Chloride	1/16/2028	Benzaldehyde Solution, 1000 mg/L, 1.3 mL

<u>Compound</u>	<u>CAS No.</u>	<u>Purity (%)</u>	<u>Compound Lot No.</u>	<u>Concentration, mg/L</u>
benzaldehyde	100-52-7	98.3	442.421.1P	996.8 ± 11.49

Received on
02/07/23
by CG

S11101
to
S11103

*Not a certified value

Certified By: _____

S. Hunter

Scott Hunter
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values
listed are determined gravimetrically.



CERTIFIED REFERENCE MATERIAL

110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

Certificate of Analysis



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Received on
03/10/22
by
CG
S10242
to
S10247

Catalog No. : 31615 **Lot No.:** A0182667
Description : GC/MS Tuning Mixture
GC/MS Tuning Mixture 1,000µg/mL, Methylene Chloride, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : March 31, 2025 **Storage:** 10°C or colder
Handling: Contains carcinogen/reproductive toxin. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)	
1	Pentachlorophenol CAS # 87-86-5 (Lot 211229RSR) Purity 99%	1,003.6 µg/mL	+/- 5.8897 µg/mL +/- 45.7132 µg/mL +/- 66.0037 µg/mL	Gravimetric Unstressed Stressed
2	DFTPP (Decafluorotriphenylphosphine) CAS # 5074-71-5 (Lot Q117-147) Purity 95%	1,006.6 µg/mL	+/- 5.9074 µg/mL +/- 45.8508 µg/mL +/- 66.2023 µg/mL	Gravimetric Unstressed Stressed
3	Benzidine CAS # 92-87-5 (Lot 211228JLM) Purity 99%	1,008.4 µg/mL	+/- 5.9179 µg/mL +/- 45.9318 µg/mL +/- 66.3193 µg/mL	Gravimetric Unstressed Stressed
4	4,4'-DDT CAS # 50-29-3 (Lot 210916JLM) Purity 99%	1,007.6 µg/mL	+/- 5.9132 µg/mL +/- 45.8954 µg/mL +/- 66.2667 µg/mL	Gravimetric Unstressed Stressed
Solvent:	Methylene chloride CAS # 75-09-2 Purity 99%			

Column:

30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant pressure 10 psi.

Temp. Program:

75°C (hold 1 min.) to 330°C
@ 20°C/min. (hold 10 min.)

Inj. Temp:


250°C

Det. Temp:

330°C

Det. Type:

FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Morgan Craighead - Mix Technician

Date Mixed: 08-Mar-2022

Balance: B345965662

Marlene Cowan - Operations Tech I

Date Passed: 10-Mar-2022

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



CERTIFIED REFERENCE MATERIAL

110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

Gravimetric Certificate



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555871 Lot No.: A0185300
Description : Custom 4-Nitrophenol Standard
Custom 4-Nitrophenol Standard 25,000µg/mL, Methanol, 1mL/ampul
Container Size : 2 mL Pkg Amt: > 1 mL
Expiration Date : May 31, 2025 Storage: 10°C or colder
Ship: Ambient

Received by
CG ON
05/18/22
S10393
+0
S10402

CERTIFIED VALUES

Component #	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)
1	4-Nitrophenol CAS # 100-02-7 Purity 99% (Lot MKCN1089)	25,060.0 µg/mL	+/- 231.9100 µg/mL Gravimetric +/- 753.2622 µg/mL Unstressed +/- 905.6020 µg/mL Stressed

Solvent: Methanol
CAS # 67-56-1
Purity 99%

Katelyn McGinnis - Operations Tech I

Date Mixed: 16-May-2022 Balance: 1128342314

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified combined stressed uncertainty value (includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ stressed} = k \sqrt{U_{gravimetric}^2 + U_{homogeneity}^2 + U_{storage\ stability}^2 + U_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- It is important to note that the shipping stability uncertainty was obtained under temperature extremes for specific time intervals; therefore, the certified combined stressed uncertainty value should only be applied to the product if it was stored at non-standard temperature conditions up to and including 7 days. Contact Restek Technical Service at www.restek.com/Contact-Us for use recommendations if your shipment was in-transit for more than 7 days at non-standard temperature conditions.
- Apply the certified combined unstressed uncertainty value if the product was received under standard shipping conditions. Apply the certified combined stressed uncertainty value if the product was received under non-standard conditions as specified below.

Label Conditions	Standard Conditions	Non-Standard Conditions
25°C Nominal (Room Temperature)	< 60°C	≥ 60°C up to 7 days
10°C or colder (Refrigerate)	< 40°C	≥ 40°C up to 7 days
0°C or colder (Freezer) -20°C or colder (Deep Freezer)	< 25°C	≥ 25°C up to 7 days

- Separate (not combined) uncertainty values for gravimetric uncertainty are also displayed on the certificate, if needed, separate homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty values are available by contacting Restek Technical Service at www.restek.com/Contact-Us.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampules are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampules. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Gravimetric Certificate



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Received by
CG
on
07/05/22
S 10583
to
S 10592

Catalog No. : 555868 **Lot No.:** A0186373

Description : Custom Benzidine Standard
Custom Benzidine Standard 25,000µg/mL, Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : June 30, 2025 **Storage:** 10°C or colder

Handling: Contains carcinogen/reproductive toxin. **Ship:** Ambient

CERTIFIED VALUES

Component #	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)
1	Benzidine	25,200.0 µg/mL	+/- 233.2055 µg/mL Gravimetric
	CAS # 92-87-5 (Lot 220511RSR)		+/- 351.6606 µg/mL Unstressed
	Purity 99%		+/- 512.6054 µg/mL Stressed

Solvent: Methanol
CAS # 67-56-1
Purity 99%

Tom Suckal - Mix Technician

Date Mixed: 16-Jun-2022 Balance: 1122030677

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified combined stressed uncertainty value (includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ stressed} = k \sqrt{U_{gravimetric}^2 + U_{homogeneity}^2 + U_{storage\ stability}^2 + U_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- It is important to note that the shipping stability uncertainty was obtained under temperature extremes for specific time intervals; therefore, the certified combined stressed uncertainty value should only be applied to the product if it was stored at non-standard temperature conditions up to and including 7 days. Contact Restek Technical Service at www.restek.com/Contact-Us for use recommendations if your shipment was in-transit for more than 7 days at non-standard temperature conditions.
- Apply the certified combined unstressed uncertainty value if the product was received under standard shipping conditions. Apply the certified combined stressed uncertainty value if the product was received under non-standard conditions as specified below.

Label Conditions	Standard Conditions	Non-Standard Conditions
25°C Nominal (Room Temperature)	< 60°C	≥ 60°C up to 7 days
10°C or colder (Refrigerate)	< 40°C	≥ 40°C up to 7 days
0°C or colder (Freezer) -20°C or colder (Deep Freezer)	< 25°C	≥ 25°C up to 7 days

- Separate (not combined) uncertainty values for gravimetric uncertainty are also displayed on the certificate, if needed, separate homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty values are available by contacting Restek Technical Service at www.restek.com/Contact-Us.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampules are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampules. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31087 **Lot No.:** A0188108
Description : Acid Surrogate Mix (4/89 SOW)
Acid Surrogate 10,000µg/mL, Methanol, 5mL/ampul
Container Size : 5 mL **Pkg Amt:** > 5 mL
Expiration Date : August 31, 2030 **Storage:** 10°C or colder
Ship: Ambient

Received by
CG on
12/28/22
S10951
to
S10980

CERTIFIED VALUES

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)
1	2-Fluorophenol CAS # 367-12-4 Purity 99% (Lot STBF3761V)	10,088.5 µg/mL	+/- 58.6554 µg/mL Gravimetric +/- 294.4162 µg/mL Unstressed +/- 357.2628 µg/mL Stressed
2	Phenol-d6 CAS # 13127-88-3 Purity 99% (Lot PR-31262)	10,043.3 µg/mL	+/- 58.3923 µg/mL Gravimetric +/- 293.0957 µg/mL Unstressed +/- 355.6603 µg/mL Stressed
3	2,4,6-Tribromophenol CAS # 118-79-6 Purity 99% (Lot MKCJ7664)	10,010.0 µg/mL	+/- 58.1990 µg/mL Gravimetric +/- 292.1253 µg/mL Unstressed +/- 354.4829 µg/mL Stressed

Solvent: Methanol
CAS # 67-56-1
Purity 99%

Column:

30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant pressure 10 psi.

Temp. Program:

40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

Inj. Temp:


250°C

Det. Temp:

330°C

Det. Type:

FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Morgan Craighead - Mix Technician

Date Mixed: 02-Aug-2022

Balance: 1127510105

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 05-Aug-2022

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis



ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01



ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31087 **Lot No.:** A0188108
Description : Acid Surrogate Mix (4/89 SOW)
Acid Surrogate 10,000µg/mL, Methanol, 5mL/ampul
Container Size : 5 mL **Pkg Amt:** > 5 mL
Expiration Date : August 31, 2030 **Storage:** 10°C or colder
Ship: Ambient

Received by
CG on
12/28/22
S10951
FO
S10980

CERTIFIED VALUES

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)
1	2-Fluorophenol CAS # 367-12-4 Purity 99% (Lot STBF3761V)	10,088.5 µg/mL	+/- 58.6554 µg/mL Gravimetric +/- 294.4162 µg/mL Unstressed +/- 357.2628 µg/mL Stressed
2	Phenol-d6 CAS # 13127-88-3 Purity 99% (Lot PR-31262)	10,043.3 µg/mL	+/- 58.3923 µg/mL Gravimetric +/- 293.0957 µg/mL Unstressed +/- 355.6603 µg/mL Stressed
3	2,4,6-Tribromophenol CAS # 118-79-6 Purity 99% (Lot MKCJ7664)	10,010.0 µg/mL	+/- 58.1990 µg/mL Gravimetric +/- 292.1253 µg/mL Unstressed +/- 354.4829 µg/mL Stressed

Solvent: Methanol
CAS # 67-56-1
Purity 99%

Column:

30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant pressure 10 psi.

Temp. Program:

40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

Inj. Temp:


250°C

Det. Temp:

330°C

Det. Type:

FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Morgan Craighead - Mix Technician

Date Mixed: 02-Aug-2022

Balance: 1127510105

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 05-Aug-2022

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis



ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01



ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31087 **Lot No.:** A0188108

Description : Acid Surrogate Mix (4/89 SOW)
Acid Surrogate 10,000µg/mL, Methanol, 5mL/ampul

Container Size : 5 mL **Pkg Amt:** > 5 mL

Expiration Date : August 31, 2030 **Storage:** 10°C or colder
Ship: Ambient

Received by
CG on
12/28/22
S10951
FO
S10980

CERTIFIED VALUES

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)
1	2-Fluorophenol CAS # 367-12-4 Purity 99% (Lot STBF3761V)	10,088.5 µg/mL	+/- 58.6554 µg/mL Gravimetric +/- 294.4162 µg/mL Unstressed +/- 357.2628 µg/mL Stressed
2	Phenol-d6 CAS # 13127-88-3 Purity 99% (Lot PR-31262)	10,043.3 µg/mL	+/- 58.3923 µg/mL Gravimetric +/- 293.0957 µg/mL Unstressed +/- 355.6603 µg/mL Stressed
3	2,4,6-Tribromophenol CAS # 118-79-6 Purity 99% (Lot MKCJ7664)	10,010.0 µg/mL	+/- 58.1990 µg/mL Gravimetric +/- 292.1253 µg/mL Unstressed +/- 354.4829 µg/mL Stressed

Solvent: Methanol
CAS # 67-56-1
Purity 99%

Column:

30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant pressure 10 psi.

Temp. Program:

40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

Inj. Temp:


250°C

Det. Temp:

330°C

Det. Type:

FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Morgan Craighead - Mix Technician

Date Mixed: 02-Aug-2022 Balance: 1127510105


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 05-Aug-2022

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis



ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01



ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31087 **Lot No.:** A0188108

Description : Acid Surrogate Mix (4/89 SOW)

Acid Surrogate 10,000µg/mL, Methanol, 5mL/ampul

Container Size : 5 mL **Pkg Amt:** > 5 mL

Expiration Date : August 31, 2030 **Storage:** 10°C or colder

Ship: Ambient

Received by
CG on
12/28/22
S10951
FO
S10980

CERTIFIED VALUES

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)
1	2-Fluorophenol CAS # 367-12-4 Purity 99% (Lot STBF3761V)	10,088.5 µg/mL	+/- 58.6554 µg/mL Gravimetric +/- 294.4162 µg/mL Unstressed +/- 357.2628 µg/mL Stressed
2	Phenol-d6 CAS # 13127-88-3 Purity 99% (Lot PR-31262)	10,043.3 µg/mL	+/- 58.3923 µg/mL Gravimetric +/- 293.0957 µg/mL Unstressed +/- 355.6603 µg/mL Stressed
3	2,4,6-Tribromophenol CAS # 118-79-6 Purity 99% (Lot MKCJ7664)	10,010.0 µg/mL	+/- 58.1990 µg/mL Gravimetric +/- 292.1253 µg/mL Unstressed +/- 354.4829 µg/mL Stressed

Solvent: Methanol
CAS # 67-56-1
Purity 99%

Column:

30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant pressure 10 psi.

Temp. Program:

40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

Inj. Temp:


250°C

Det. Temp:

330°C

Det. Type:

FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Morgan Craighead - Mix Technician

Date Mixed: 02-Aug-2022 Balance: 1127510105


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 05-Aug-2022

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis



ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01



ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31087 **Lot No.:** A0188108

Description : Acid Surrogate Mix (4/89 SOW)

Acid Surrogate 10,000µg/mL, Methanol, 5mL/ampul

Container Size : 5 mL **Pkg Amt:** > 5 mL

Expiration Date : August 31, 2030 **Storage:** 10°C or colder

Ship: Ambient

Received by
CG on
12/28/22
S10951
FO
S10980

CERTIFIED VALUES

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)
1	2-Fluorophenol CAS # 367-12-4 Purity 99% (Lot STBF3761V)	10,088.5 µg/mL	+/- 58.6554 µg/mL Gravimetric +/- 294.4162 µg/mL Unstressed +/- 357.2628 µg/mL Stressed
2	Phenol-d6 CAS # 13127-88-3 Purity 99% (Lot PR-31262)	10,043.3 µg/mL	+/- 58.3923 µg/mL Gravimetric +/- 293.0957 µg/mL Unstressed +/- 355.6603 µg/mL Stressed
3	2,4,6-Tribromophenol CAS # 118-79-6 Purity 99% (Lot MKCJ7664)	10,010.0 µg/mL	+/- 58.1990 µg/mL Gravimetric +/- 292.1253 µg/mL Unstressed +/- 354.4829 µg/mL Stressed

Solvent: Methanol
CAS # 67-56-1
Purity 99%

Column:

30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant pressure 10 psi.

Temp. Program:

40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

Inj. Temp:


250°C

Det. Temp:

330°C

Det. Type:

FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Morgan Craighead - Mix Technician

Date Mixed: 02-Aug-2022

Balance: 1127510105

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 05-Aug-2022

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis



ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01



ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31087 **Lot No.:** A0188108

Description : Acid Surrogate Mix (4/89 SOW)

Acid Surrogate 10,000µg/mL, Methanol, 5mL/ampul

Container Size : 5 mL **Pkg Amt:** > 5 mL

Expiration Date : August 31, 2030 **Storage:** 10°C or colder

Ship: Ambient

Received by
CG on
12/28/22
S10951
FO
S10980

CERTIFIED VALUES

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)
1	2-Fluorophenol CAS # 367-12-4 Purity 99% (Lot STBF3761V)	10,088.5 µg/mL	+/- 58.6554 µg/mL Gravimetric +/- 294.4162 µg/mL Unstressed +/- 357.2628 µg/mL Stressed
2	Phenol-d6 CAS # 13127-88-3 Purity 99% (Lot PR-31262)	10,043.3 µg/mL	+/- 58.3923 µg/mL Gravimetric +/- 293.0957 µg/mL Unstressed +/- 355.6603 µg/mL Stressed
3	2,4,6-Tribromophenol CAS # 118-79-6 Purity 99% (Lot MKCJ7664)	10,010.0 µg/mL	+/- 58.1990 µg/mL Gravimetric +/- 292.1253 µg/mL Unstressed +/- 354.4829 µg/mL Stressed

Solvent: Methanol
CAS # 67-56-1
Purity 99%

Column:

30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant pressure 10 psi.

Temp. Program:

40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

Inj. Temp:


250°C

Det. Temp:

330°C

Det. Type:

FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Morgan Craighead - Mix Technician

Date Mixed: 02-Aug-2022

Balance: 1127510105

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 05-Aug-2022

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31086 **Lot No.:** A0189418
Description : B/N Surrogate Mix (4/89 SOW)
Base Neutral Surrogate 5000µg/mL, Methylene Chloride, 5mL/ampul
Container Size : 5 mL **Pkg Amt:** > 5 mL
Expiration Date : August 31, 2028 **Storage:** 10°C or colder
Handling: Sonicate prior to use. **Ship:** Ambient

Received by
CG on
12/28/22
\$10981
to
S11010

CERTIFIED VALUES

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)
1	Nitrobenzene-d5 CAS # 4165-60-0 (Lot PR-29940A) Purity 99%	5,009.8 µg/mL	+/- 29.1271 µg/mL Gravimetric +/- 225.6421 µg/mL Unstressed +/- 250.3778 µg/mL Stressed
2	2-Fluorobiphenyl CAS # 321-60-8 (Lot 00021384) Purity 99%	5,026.6 µg/mL	+/- 29.2250 µg/mL Gravimetric +/- 226.4003 µg/mL Unstressed +/- 251.2191 µg/mL Stressed
3	p-Terphenyl-d14 CAS # 1718-51-0 (Lot PR-30504) Purity 99%	5,027.3 µg/mL	+/- 29.2289 µg/mL Gravimetric +/- 226.4304 µg/mL Unstressed +/- 251.2524 µg/mL Stressed

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Tech Tips:

Due to the limited solubility of p-terphenyl-d14 in methanol, we do not recommend that this mixture be diluted in methanol.

Column:

30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant pressure 10 psi.

Temp. Program:

40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

Inj. Temp:


250°C

Det. Temp:


330°C

Det. Type:

FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


John Friedline - Operations Technician I

Date Mixed: 09-Sep-2022

Balance: 1128353505


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 13-Sep-2022

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31086 **Lot No.:** A0189418
Description : B/N Surrogate Mix (4/89 SOW)
Base Neutral Surrogate 5000µg/mL, Methylene Chloride, 5mL/ampul
Container Size : 5 mL **Pkg Amt:** > 5 mL
Expiration Date : August 31, 2028 **Storage:** 10°C or colder
Handling: Sonicate prior to use. **Ship:** Ambient

Received by
CG on
12/28/22
\$10981
to
S11010

CERTIFIED VALUES

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)
1	Nitrobenzene-d5 CAS # 4165-60-0 (Lot PR-29940A) Purity 99%	5,009.8 µg/mL	+/- 29.1271 µg/mL Gravimetric +/- 225.6421 µg/mL Unstressed +/- 250.3778 µg/mL Stressed
2	2-Fluorobiphenyl CAS # 321-60-8 (Lot 00021384) Purity 99%	5,026.6 µg/mL	+/- 29.2250 µg/mL Gravimetric +/- 226.4003 µg/mL Unstressed +/- 251.2191 µg/mL Stressed
3	p-Terphenyl-d14 CAS # 1718-51-0 (Lot PR-30504) Purity 99%	5,027.3 µg/mL	+/- 29.2289 µg/mL Gravimetric +/- 226.4304 µg/mL Unstressed +/- 251.2524 µg/mL Stressed

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Tech Tips:

Due to the limited solubility of p-terphenyl-d14 in methanol, we do not recommend that this mixture be diluted in methanol.

Column:

30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant pressure 10 psi.

Temp. Program:

40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

Inj. Temp:


250°C

Det. Temp:


330°C

Det. Type:

FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


John Friedline - Operations Technician I

Date Mixed: 09-Sep-2022

Balance: 1128353505


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 13-Sep-2022

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Received by
CG on
12/28/22
\$10981
to
S11010

Catalog No. : 31086 **Lot No.:** A0189418
Description : B/N Surrogate Mix (4/89 SOW)
Base Neutral Surrogate 5000µg/mL, Methylene Chloride, 5mL/ampul
Container Size : 5 mL **Pkg Amt:** > 5 mL
Expiration Date : August 31, 2028 **Storage:** 10°C or colder
Handling: Sonicate prior to use. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)
1	Nitrobenzene-d5 CAS # 4165-60-0 (Lot PR-29940A) Purity 99%	5,009.8 µg/mL	+/- 29.1271 µg/mL Gravimetric +/- 225.6421 µg/mL Unstressed +/- 250.3778 µg/mL Stressed
2	2-Fluorobiphenyl CAS # 321-60-8 (Lot 00021384) Purity 99%	5,026.6 µg/mL	+/- 29.2250 µg/mL Gravimetric +/- 226.4003 µg/mL Unstressed +/- 251.2191 µg/mL Stressed
3	p-Terphenyl-d14 CAS # 1718-51-0 (Lot PR-30504) Purity 99%	5,027.3 µg/mL	+/- 29.2289 µg/mL Gravimetric +/- 226.4304 µg/mL Unstressed +/- 251.2524 µg/mL Stressed

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Tech Tips:

Due to the limited solubility of p-terphenyl-d14 in methanol, we do not recommend that this mixture be diluted in methanol.

Column:

30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant pressure 10 psi.

Temp. Program:

40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

Inj. Temp:


250°C

Det. Temp:


330°C

Det. Type:

FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


John Friedline - Operations Technician I

Date Mixed: 09-Sep-2022

Balance: 1128353505


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 13-Sep-2022

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Received by
CG on
12/28/22
\$10981
to
\$11010

Catalog No. : 31086 **Lot No.:** A0189418
Description : B/N Surrogate Mix (4/89 SOW)
Base Neutral Surrogate 5000µg/mL, Methylene Chloride, 5mL/ampul
Container Size : 5 mL **Pkg Amt:** > 5 mL
Expiration Date : August 31, 2028 **Storage:** 10°C or colder
Handling: Sonicate prior to use. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)
1	Nitrobenzene-d5 CAS # 4165-60-0 (Lot PR-29940A) Purity 99%	5,009.8 µg/mL	+/- 29.1271 µg/mL Gravimetric +/- 225.6421 µg/mL Unstressed +/- 250.3778 µg/mL Stressed
2	2-Fluorobiphenyl CAS # 321-60-8 (Lot 00021384) Purity 99%	5,026.6 µg/mL	+/- 29.2250 µg/mL Gravimetric +/- 226.4003 µg/mL Unstressed +/- 251.2191 µg/mL Stressed
3	p-Terphenyl-d14 CAS # 1718-51-0 (Lot PR-30504) Purity 99%	5,027.3 µg/mL	+/- 29.2289 µg/mL Gravimetric +/- 226.4304 µg/mL Unstressed +/- 251.2524 µg/mL Stressed

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Tech Tips:

Due to the limited solubility of p-terphenyl-d14 in methanol, we do not recommend that this mixture be diluted in methanol.

Column:

30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant pressure 10 psi.

Temp. Program:

40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

Inj. Temp:


250°C

Det. Temp:


330°C

Det. Type:

FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


John Friedline - Operations Technician I

Date Mixed: 09-Sep-2022

Balance: 1128353505


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 13-Sep-2022

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31086 **Lot No.:** A0189418
Description : B/N Surrogate Mix (4/89 SOW)
Base Neutral Surrogate 5000µg/mL, Methylene Chloride, 5mL/ampul
Container Size : 5 mL **Pkg Amt:** > 5 mL
Expiration Date : August 31, 2028 **Storage:** 10°C or colder
Handling: Sonicate prior to use. **Ship:** Ambient

Received by
CG on
12/28/22
\$10981
to
S11010

CERTIFIED VALUES

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)
1	Nitrobenzene-d5 CAS # 4165-60-0 (Lot PR-29940A) Purity 99%	5,009.8 µg/mL	+/- 29.1271 µg/mL Gravimetric +/- 225.6421 µg/mL Unstressed +/- 250.3778 µg/mL Stressed
2	2-Fluorobiphenyl CAS # 321-60-8 (Lot 00021384) Purity 99%	5,026.6 µg/mL	+/- 29.2250 µg/mL Gravimetric +/- 226.4003 µg/mL Unstressed +/- 251.2191 µg/mL Stressed
3	p-Terphenyl-d14 CAS # 1718-51-0 (Lot PR-30504) Purity 99%	5,027.3 µg/mL	+/- 29.2289 µg/mL Gravimetric +/- 226.4304 µg/mL Unstressed +/- 251.2524 µg/mL Stressed

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Tech Tips:

Due to the limited solubility of p-terphenyl-d14 in methanol, we do not recommend that this mixture be diluted in methanol.

Column:

30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant pressure 10 psi.

Temp. Program:

40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

Inj. Temp:


250°C

Det. Temp:


330°C

Det. Type:

FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


John Friedline - Operations Technician I

Date Mixed: 09-Sep-2022

Balance: 1128353505


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 13-Sep-2022

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Received by
CG on
12/28/22
\$10981
to
S11010

Catalog No. : 31086 **Lot No.:** A0189418
Description : B/N Surrogate Mix (4/89 SOW)
Base Neutral Surrogate 5000µg/mL, Methylene Chloride, 5mL/ampul
Container Size : 5 mL **Pkg Amt:** > 5 mL
Expiration Date : August 31, 2028 **Storage:** 10°C or colder
Handling: Sonicate prior to use. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)
1	Nitrobenzene-d5 CAS # 4165-60-0 (Lot PR-29940A) Purity 99%	5,009.8 µg/mL	+/- 29.1271 µg/mL Gravimetric +/- 225.6421 µg/mL Unstressed +/- 250.3778 µg/mL Stressed
2	2-Fluorobiphenyl CAS # 321-60-8 (Lot 00021384) Purity 99%	5,026.6 µg/mL	+/- 29.2250 µg/mL Gravimetric +/- 226.4003 µg/mL Unstressed +/- 251.2191 µg/mL Stressed
3	p-Terphenyl-d14 CAS # 1718-51-0 (Lot PR-30504) Purity 99%	5,027.3 µg/mL	+/- 29.2289 µg/mL Gravimetric +/- 226.4304 µg/mL Unstressed +/- 251.2524 µg/mL Stressed

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Tech Tips:

Due to the limited solubility of p-terphenyl-d14 in methanol, we do not recommend that this mixture be diluted in methanol.

Column:

30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant pressure 10 psi.

Temp. Program:

40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

Inj. Temp:


250°C

Det. Temp:


330°C

Det. Type:

FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


John Friedline - Operations Technician I

Date Mixed: 09-Sep-2022

Balance: 1128353505


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 13-Sep-2022

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31086 **Lot No.:** A0189418
Description : B/N Surrogate Mix (4/89 SOW)
Base Neutral Surrogate 5000µg/mL, Methylene Chloride, 5mL/ampul
Container Size : 5 mL **Pkg Amt:** > 5 mL
Expiration Date : August 31, 2028 **Storage:** 10°C or colder
Handling: Sonicate prior to use. **Ship:** Ambient

Received by
CG on
12/28/22
\$10981
to
S11010

CERTIFIED VALUES

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)
1	Nitrobenzene-d5 CAS # 4165-60-0 (Lot PR-29940A) Purity 99%	5,009.8 µg/mL	+/- 29.1271 µg/mL Gravimetric +/- 225.6421 µg/mL Unstressed +/- 250.3778 µg/mL Stressed
2	2-Fluorobiphenyl CAS # 321-60-8 (Lot 00021384) Purity 99%	5,026.6 µg/mL	+/- 29.2250 µg/mL Gravimetric +/- 226.4003 µg/mL Unstressed +/- 251.2191 µg/mL Stressed
3	p-Terphenyl-d14 CAS # 1718-51-0 (Lot PR-30504) Purity 99%	5,027.3 µg/mL	+/- 29.2289 µg/mL Gravimetric +/- 226.4304 µg/mL Unstressed +/- 251.2524 µg/mL Stressed

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Tech Tips:

Due to the limited solubility of p-terphenyl-d14 in methanol, we do not recommend that this mixture be diluted in methanol.

Column:

30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant pressure 10 psi.

Temp. Program:

40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

Inj. Temp:


250°C

Det. Temp:


330°C

Det. Type:

FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


John Friedline - Operations Technician I

Date Mixed: 09-Sep-2022

Balance: 1128353505


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 13-Sep-2022

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31086 **Lot No.:** A0189418
Description : B/N Surrogate Mix (4/89 SOW)
Base Neutral Surrogate 5000µg/mL, Methylene Chloride, 5mL/ampul
Container Size : 5 mL **Pkg Amt:** > 5 mL
Expiration Date : August 31, 2028 **Storage:** 10°C or colder
Handling: Sonicate prior to use. **Ship:** Ambient

Received by
CG on
12/28/22
\$10981
to
S11010

CERTIFIED VALUES

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)
1	Nitrobenzene-d5 CAS # 4165-60-0 (Lot PR-29940A) Purity 99%	5,009.8 µg/mL	+/- 29.1271 µg/mL Gravimetric +/- 225.6421 µg/mL Unstressed +/- 250.3778 µg/mL Stressed
2	2-Fluorobiphenyl CAS # 321-60-8 (Lot 00021384) Purity 99%	5,026.6 µg/mL	+/- 29.2250 µg/mL Gravimetric +/- 226.4003 µg/mL Unstressed +/- 251.2191 µg/mL Stressed
3	p-Terphenyl-d14 CAS # 1718-51-0 (Lot PR-30504) Purity 99%	5,027.3 µg/mL	+/- 29.2289 µg/mL Gravimetric +/- 226.4304 µg/mL Unstressed +/- 251.2524 µg/mL Stressed

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Tech Tips:

Due to the limited solubility of p-terphenyl-d14 in methanol, we do not recommend that this mixture be diluted in methanol.

Column:

30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant pressure 10 psi.

Temp. Program:

40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

Inj. Temp:


250°C

Det. Temp:


330°C

Det. Type:

FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


John Friedline - Operations Technician I

Date Mixed: 09-Sep-2022

Balance: 1128353505


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 13-Sep-2022

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555872 **Lot No.:** A0193449
Description : Custom Pentachlorophenol Standard
Custom Pentachlorophenol Standard 25,000µg/mL, Methanol,
1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : January 31, 2026 **Storage:** 10°C or colder
Ship: Ambient

Received on
01/13/23
by
CG
S11011
+0
S11015

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Pentachlorophenol	87-86-5	RP221012	99%	25,050.0 µg/mL	+/- 778.6378

Solvent: Methanol
CAS # 67-56-1
Purity 99%

Russ Bookhamer

Russ Bookhamer - Operations Technician I

Date Mixed: 11-Jan-2023

Balance: B442140311

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555870 Lot No.: A0194698
Description : Custom 2,4-Dinitrophenol Standard
Custom 2,4-Dinitrophenol Standard 25,000µg/mL, Methanol, 1mL/ampul
Container Size : 2 mL Pkg Amt: > 1 mL
Expiration Date : February 28, 2026 Storage: 10°C or colder
Ship: Ambient

CERTIFIED

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)
1	2,4-Dinitrophenol	51-28-5	DR221221RSR	99%	25,195.0 µg/mL

Solvent: Methanol
CAS # 67-56-1
Purity 99%

Russ Bookhamer - Operations Technician I

Date Mixed: 15-Feb-2023

Balance: B442140311

Manufactured under Restek®
Registered Quality
Certificate #FM 8

tified Reference Material Notes

es:

n date valid for unopened ampul stored in compliance with the recommended conditions.
nty, concentration, and expiration of the CRM are based on the unopened product being stored according to the
ended condition found in the storage field.

d/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD,
LC/MS, RI, and/or melting point.

nds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A
n factor is used to calculate the amount of compound necessary to achieve the desired concentration of the
mpound in solution.

isomeric compounds is reported as the sum of the isomers.

lues are rounded to the nearest whole number.

rtainty Value Notes:

rtainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded
ty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability
ty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

erage factor of 2, which gives a level of confidence of approximately 95%.

ged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure
imum packaged amount can be sufficiently transferred.

Notes:

tion is based upon gravimetric preparation using either a balance whose calibration has been verified daily
Γ traceable weights, and/or dilutions with Class A glassware.

:

the unopened product, when stored in compliance with the recommended conditions, is guaranteed through
ion displayed on the product label and certificate. Contact Restek for additional opened product stability
n, with the knowledge/understanding that open product stability is subject to the specific handling and
ntal conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with
tards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom
rm. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861,
ides complete instructions.

ssolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555869 **Lot No.:** A0194702

Description : Custom Hexachlorocyclopentadiene Standard

Custom Hexachlorocyclopentadiene Standard 25,000µg/mL, Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL


Expiration Date : February 28, 2026 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)
1	Hexachlorocyclopentadiene	77-47-4	0012019	99%	25,008.0 µg/mL

Solvent: Methanol
CAS # 67-56-1
Purity 99%


Russ Bookhamer - Operations Technician I

Date Mixed: 15-Feb-2023

Balance: B442140311

Manufactured under Restek
Registered Quality
Certificate #FM1

ified Reference Material Notes

es:

n date valid for unopened ampul stored in compliance with the recommended conditions.
ty, concentration, and expiration of the CRM are based on the unopened product being stored according to the
ended condition found in the storage field.

d/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD,
LC/MS, RI, and/or melting point.
nds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A
n factor is used to calculate the amount of compound necessary to achieve the desired concentration of the
mpound in solution.
isomeric compounds is reported as the sum of the isomers.
ues are rounded to the nearest whole number.

rtainty Value Notes:

rtainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded
ty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability
ty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

age factor of 2, which gives a level of confidence of approximately 95%.
ged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure
nimum packaged amount can be sufficiently transferred.

Notes:

tion is based upon gravimetric preparation using either a balance whose calibration has been verified daily
traceable weights, and/or dilutions with Class A glassware.

the unopened product, when stored in compliance with the recommended conditions, is guaranteed through
ion displayed on the product label and certificate. Contact Restek for additional opened product stability
i, with the knowledge/understanding that open product stability is subject to the specific handling and
ntal conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with
ards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom
m. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861,
des complete instructions.

ssolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely



**PRODUCTOS
QUÍMICOS
MONTERREY, S.A. DE C.V.**

MIRADOR 201, COL. MIRADOR
MONTERREY, N.L. MEXICO
CP 64070
TEL +52 81 13 52 57 57
www.pqm.com.mx

CERTIFICATE OF ANALYSIS

PRODUCT :	SODIUM SULFATE CRYSTALS ANHYDROUS		
QUALITY :	ACS (CODE RMB3375)	FORMULA :	Na ₂ SO ₄
SPECIFICATION NUMBER :	6399	RELEASE DATE:	ABR/21/2023
LOT NUMBER :	313201		

TEST	SPECIFICATIONS	LOT VALUES
Assay (Na ₂ SO ₄)	Min. 99.0%	99.7 %
pH of a 5% solution at 25°C	5.2 - 9.2	6.1
Insoluble matter	Max. 0.01%	0.005 %
Loss on ignition	Max. 0.5%	0.1 %
Chloride (Cl)	Max. 0.001%	<0.001 %
Nitrogen compounds (as N)	Max. 5 ppm	<5 ppm
Phosphate (PO ₄)	Max. 0.001%	<0.001 %
Heavy metals (as Pb)	Max. 5 ppm	<5 ppm
Iron (Fe)	Max. 0.001%	<0.001 %
Calcium (Ca)	Max. 0.01%	0.002 %
Magnesium (Mg)	Max. 0.005%	0.001 %
Potassium (K)	Max. 0.008%	0.003 %
Extraction-concentration suitability	Passes test	Passes test
Appearance	Passes test	Passes test
Identification	Passes test	Passes test
Solubility and foreign matter	Passes test	Passes test
Retained on US Standard No. 10 sieve	Max. 1%	0.1 %
Retained on US Standard No. 60 sieve	Min. 94%	97.3 %
Through US Standard No. 60 sieve	Max. 5%	2.5 %
Through US Standard No. 100 sieve	Max. 10%	0.1 %

COMMENTS

QC: PhC Irma Belmares

If you need further details, please call our factory or contact our local distributor.

Recd. by R3 on 7/24/23 E 3551

RC-02-01, Ed. 3



Certificate of Analysis

Sodium Hydroxide (Pellets)

Material: 0583
Grade: ACS GRADE
Batch Number: 23B1556310

Chemical Formula: NaOH
Molecular Weight: 40
CAS #: 1310-73-2
Appearance:

Manufacture Date: 12/14/2022
Expiration Date: 12/31/2025

Storage: Room Temperature

Pellets

TEST	SPECIFICATION	ANALYSIS	DISPOSITION
Calcium	$\leq 0.005 \%$	$< 0.005 \%$	PASS
Chloride	$\leq 0.005 \%$	0.002 %	PASS
Heavy Metals	$\leq 0.002 \%$	$< 0.002 \%$	PASS
Iron	$\leq 0.001 \%$	$< 0.001 \%$	PASS
Magnesium	$\leq 0.002 \%$	$< 0.002 \%$	PASS
Mercury	$\leq 0.1 \text{ ppm}$	$< 0.1 \text{ ppm}$	PASS
Nickel	$\leq 0.001 \%$	$< 0.001 \%$	PASS
Nitrogen Compounds	$\leq 0.001 \%$	$< 0.001 \%$	PASS
Phosphate	$\leq 0.001 \%$	$< 0.001 \%$	PASS
Potassium	$\leq 0.02 \%$	$< 0.02 \%$	PASS
Purity	$\geq 97.0 \%$	99.2 %	PASS
Sodium Carbonate	$\leq 1.0 \%$	0.5 %	PASS
Sulfate	$\leq 0.003 \%$	$< 0.003 \%$	PASS

Internal ID #: 710

Signature

We certify that this batch conforms to the specifications listed.


This document has been electronically produced and is valid without a signature.

Leona Edwardson, Quality Control Sr. Manager - Solon
VWR Chemicals, LLC.
28600 Fountain Parkway, Solon OH 44139 USA

Additional Information

Analysis may have been rounded to significant digits in specification limits.

Product meets analytical specifications of the grades listed.



Acetone
BAKER RESI-ANALYZED® Reagent
For Organic Residue Analysis



Material No.: 9254-03
Batch No.: 23H1462005
Manufactured Date: 2023-07-26
Expiration Date: 2026-07-25
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ((CH ₃) ₂ CO) (by GC, corrected for water)	≥ 99.4 %	99.7 %
Color (APHA)	≤ 10	5
Residue after Evaporation	≤ 1.0 ppm	0.3 ppm
Substances Reducing Permanganate	Passes Test	Passes Test
Titration Acid (μeq/g)	≤ 0.3	0.1
Titration Base (μeq/g)	≤ 0.6	< 0.1
Water (H ₂ O)	≤ 0.5 %	0.3 %
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	< 1
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	≤ 10	1

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

Rec'd by RP on 5/23/24

E 3744

Ken Koehnlein
Sr. Manager, Quality Assurance

Methylene Chloride
ULTRA RESI-ANALYZED
For Organic Residue Analysis
(dichloromethane)



Material No.: 9266-A4
Batch No.: 24C0162011
Manufactured Date: 2024-01-04
Expiration Date: 2025-04-04
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	< 1
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	≤ 10	2
Assay (CH ₂ Cl ₂) (by GC, exclusive of preservative, corrected for water)	$\geq 99.8 \%$	100.0 %
Color (APHA)	≤ 10	10
Residue after Evaporation	$\leq 1.0 \text{ ppm}$	0.2 ppm
Titration Acid ($\mu\text{eq/g}$)	≤ 0.3	< 0.1
Chloride (Cl)	$\leq 10 \text{ ppm}$	< 5 ppm
Water (by KF, coulometric)	$\leq 0.02 \%$	< 0.01 %

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC
Manufacturer source batch: MG24A04224

E 3746

Ken Koehnlein
Sr. Manager, Quality Assurance

Methylene Chloride
ULTRA RESI-ANALYZED
For Organic Residue Analysis
(dichloromethane)



Material No.: 9266-A4
Batch No.: 24E2462004
Manufactured Date: 2024-04-10
Expiration Date: 2025-07-10
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	3
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	≤ 10	3
Assay (CH_2Cl_2) (by GC, exclusive of preservative, corrected for water)	$\geq 99.8 \%$	100.0 %
Color (APHA)	≤ 10	5
Residue after Evaporation	$\leq 1.0 \text{ ppm}$	0.1 ppm
Titration Acid ($\mu\text{eq/g}$)	≤ 0.3	< 0.1
Chloride (Cl)	$\leq 10 \text{ ppm}$	5 ppm
Water (by KF, coulometric)	$\leq 0.02 \%$	< 0.01 %

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC
Manufacturer source batch: MG24D10725

E 3768

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials, LLC

100 Matsonford Rd, Suite 200, Radnor, PA 19087, U.S.A. Phone 610.386.1700

Page 1 of 1

Acetone
BAKER RESI-ANALYZED® Reagent
For Organic Residue Analysis



Material No.: 9254-03
Batch No.: 23H1462005
Manufactured Date: 2023-07-26
Expiration Date: 2026-07-25
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ((CH ₃) ₂ CO) (by GC, corrected for water)	≥ 99.4 %	99.7 %
Color (APHA)	≤ 10	5
Residue after Evaporation	≤ 1.0 ppm	0.3 ppm
Substances Reducing Permanganate	Passes Test	Passes Test
Titration Acid (μeq/g)	≤ 0.3	0.1
Titration Base (μeq/g)	≤ 0.6	< 0.1
Water (H ₂ O)	≤ 0.5 %	0.3 %
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	< 1
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	≤ 10	1

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

Recd by RP on 7/24/24

E 3780

Ken Koehnlein
Sr. Manager, Quality Assurance

Methylene Chloride
ULTRA RESI-ANALYZED
For Organic Residue Analysis
(dichloromethane)



Material No.: 9266-A4
Batch No.: 24G0862022
Manufactured Date: 2024-06-05
Expiration Date: 2025-09-04
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	3
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	≤ 10	4
Assay (CH_2Cl_2) (by GC, exclusive of preservative, corrected for water)	$\geq 99.8 \%$	100.0 %
Color (APHA)	≤ 10	5
Residue after Evaporation	$\leq 1.0 \text{ ppm}$	0.3 ppm
Titration Acid ($\mu\text{eq/g}$)	≤ 0.3	< 0.1
Chloride (Cl)	$\leq 10 \text{ ppm}$	< 5 ppm
Water (by KF, coulometric)	$\leq 0.02 \%$	< 0.01 %

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC
Manufacturer source batch: MG24F05012

E 3787

Jamie Croak
Director Quality Operations, Bioscience Production

Sulfuric Acid

BAKER INSTRA-ANALYZED® Reagent

For Trace Metal Analysis

Low Selenium

avantorsm



Material No.: 9673-33

Batch No.: 0000250349

Manufactured Date: 2019/12/17

Retest Date: 2024/12/15

Revision No: 1

Certificate of Analysis

Test	Specification	Result
ACS - Assay (H ₂ SO ₄)	95.0 - 98.0 %	96.5
Appearance	Passes Test	PT
ACS - Color (APHA)	<= 10	5
ACS - Residue after Ignition	<= 3 ppm	1
ACS - Substances Reducing Permanganate (as SO ₂)	<= 2 ppm	< 2
Ammonium (NH ₄)	<= 1 ppm	< 1
Chloride (Cl)	<= 0.1 ppm	< 0.1
Nitrate (NO ₃)	<= 0.2 ppm	< 0.1
Phosphate (PO ₄)	<= 0.5 ppm	< 0.1
Trace Impurities - Aluminum (Al)	<= 30.0 ppb	0.2
Arsenic and Antimony (as As)	<= 4 ppb	< 2
Trace Impurities - Barium (Ba)	<= 10.0 ppb	< 1.0
Trace Impurities - Beryllium (Be)	<= 10.0 ppb	< 1.0
Trace Impurities - Bismuth (Bi)	<= 10.0 ppb	< 1.0
Trace Impurities - Boron (B)	<= 10.0 ppb	< 5.0
Trace Impurities - Cadmium (Cd)	<= 2.0 ppb	< 0.3
Trace Impurities - Calcium (Ca)	<= 50.0 ppb	2.9
Trace Impurities - Chromium (Cr)	<= 6.0 ppb	< 0.4
Trace Impurities - Cobalt (Co)	<= 0.5 ppb	< 0.3
Trace Impurities - Copper (Cu)	<= 1.0 ppb	< 0.1
Trace Impurities - Gallium (Ga)	<= 10.0 ppb	< 1.0
Trace Impurities - Germanium (Ge)	<= 10.0 ppb	< 10.0
Trace Impurities - Gold (Au)	<= 10.0 ppb	< 0.2
Heavy Metals (as Pb)	<= 500 ppb	< 100

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials, LLC


100 Matsonford Rd, Suite 200, Radnor, PA 19087. U.S.A. Phone: 610.386.1700

Material No.: 9673-33
Batch No.: 0000250349

Test	Specification	Result
Trace Impurities - Iron (Fe)	<= 50.0 ppb	4.1
Trace Impurities - Lead (Pb)	<= 0.5 ppb	< 0.5
Trace Impurities - Lithium (Li)	<= 10.0 ppb	< 1.0
Trace Impurities - Magnesium (Mg)	<= 7.0 ppb	0.4
Trace Impurities - Manganese (Mn)	<= 1.0 ppb	< 0.4
Trace Impurities - Mercury (Hg)	<= 0.5 ppb	< 0.1
Trace Impurities - Molybdenum (Mo)	<= 10.0 ppb	< 5.0
Trace Impurities - Nickel (Ni)	<= 2.0 ppb	< 0.3
Trace Impurities - Niobium (Nb)	<= 10.0 ppb	< 1.0
Trace Impurities - Potassium (K)	<= 500.0 ppb	< 2.0
Trace Impurities - Selenium (Se)	<= 50.0 ppb	22.9
Trace Impurities - Silicon (Si)	<= 100.0 ppb	< 10.0
Trace Impurities - Silver (Ag)	<= 1.0 ppb	< 0.3
Trace Impurities - Sodium (Na)	<= 500.0 ppb	2.7
Trace Impurities - Strontium (Sr)	<= 5.0 ppb	< 0.2
Trace Impurities - Tantalum (Ta)	<= 10.0 ppb	< 5.0
Trace Impurities - Thallium (Tl)	<= 20.0 ppb	< 5.0
Trace Impurities - Tin (Sn)	<= 5.0 ppb	< 0.8
Trace Impurities - Titanium (Ti)	<= 10.0 ppb	< 1.0
Trace Impurities - Vanadium (V)	<= 10.0 ppb	< 1.0
Trace Impurities - Zinc (Zn)	<= 5.0 ppb	0.3
Trace Impurities - Zirconium (Zr)	<= 10.0 ppb	< 1.0

For Laboratory, Research or Manufacturing Use

Country of Origin: US
Packaging Site: Phillipsburg Mfg Ctr & DC


Jamie Ethier
Vice President Global Quality

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials, LLC
100 Matsonford Rd, Suite 200, Radnor, PA 19087. U.S.A. Phone: 610.386.1700



5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

511434 } Y.P.
07/20/23
Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received: _____

Certificate of Analysis

Rev 0

Page 1 of 1

Catalog No.:	Lot No.:	Storage:	Solvent:	Exp. Date:	Description:
Z-110094-02	503442	≤ -10 °C	Methylene Chloride	8/26/2024	CLP Base/Neutral Surrogate Solution, 5,000 mg/L, 1 ml

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
1,2-dichlorobenzene-d ₄	2199-69-1	99.7	247.29.3P	5052 ± 122.61
2-fluorobiphenyl	321-60-8	99.7	8.7.1.1P	5005 ± 121.47
nitrobenzene-d ₅	4165-60-0	100	7.9.2P	5040 ± 122.21
p-terphenyl-d ₁₄	1718-51-0	99.6	9.12.9P	5027 ± 122

*Not a certified value

Certified By: _____

Joanna Radu
Joanna Radu
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values
listed are determined gravimetrically.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No.: 555223 Lot No.: A0201940
Description: Custom 8270 Plus Standard #1
Custom 8270 Plus Standard #1 1,000µg/mL, Methylene Chloride, 1mL/ampul
Container Size: 2 mL
Expiration Date: September 30, 2025
Handling: This product is photosensitive.
Pkg Amt: > 1 mL
Storage: 10°C or colder
Ship: Ambient

511539
↓
51568
} 7.P.
09/19/

CERTIFIED VALUES						
Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	3,3'-Dichlorobenzidine	91-94-1	S230321RSR	99%	1,001.0 µg/mL	+/- 22.9799
2	Atrazine	1912-24-9	5FYWL	99%	1,010.0 µg/mL	+/- 23.1865
3	Benzidine	92-87-5	S221205RSR	99%	1,008.0 µg/mL	+/- 23.1406
4	epsilon-Caprolactam	105-60-2	I16X016	99%	1,008.0 µg/mL	+/- 23.1406
Solvent: Methylene chloride						
CAS # 75-09-2						
Purity 99%						

REVIEWED

By Jennifer Pollino at 7:10 am, Sep 13, 2023

Sam Moodler - Operations Tech I

Date Mixed: 13-Sep-2023

Balance: B345965662

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No.: 555223 Lot No.: A0201940
Description: Custom 8270 Plus Standard #1
Custom 8270 Plus Standard #1 1,000µg/mL, Methylene Chloride, 1mL/ampul
Container Size: 2 mL
Expiration Date: September 30, 2025
Handling: This product is photosensitive.
Pkg Amt: > 1 mL
Storage: 10°C or colder
Ship: Ambient

511539
↓
51568
} 7.P.
09/19/

CERTIFIED VALUES						
Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	3,3'-Dichlorobenzidine	91-94-1	S230321RSR	99%	1,001.0 µg/mL	+/- 22.9799
2	Atrazine	1912-24-9	5FYWL	99%	1,010.0 µg/mL	+/- 23.1865
3	Benzidine	92-87-5	S221205RSR	99%	1,008.0 µg/mL	+/- 23.1406
4	epsilon-Caprolactam	105-60-2	I16X016	99%	1,008.0 µg/mL	+/- 23.1406
Solvent: Methylene chloride						
CAS # 75-09-2						
Purity 99%						

REVIEWED

By Jennifer Pollino at 7:10 am, Sep 13, 2023

Sam Moodler - Operations Tech I

Date Mixed: 13-Sep-2023

Balance: B345965662

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No.: 555223 Lot No.: A0201940
Description: Custom 8270 Plus Standard #1
Custom 8270 Plus Standard #1 1,000µg/mL, Methylene Chloride, 1mL/ampul
Container Size: 2 mL
Expiration Date: September 30, 2025
Handling: This product is photosensitive.
Pkg Amt: > 1 mL
Storage: 10°C or colder
Ship: Ambient

511539
↓
51568
} 7.P.
09/19/

CERTIFIED VALUES						
Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	3,3'-Dichlorobenzidine	91-94-1	S230321RSR	99%	1,001.0 µg/mL	+/- 22.9799
2	Atrazine	1912-24-9	5FYWL	99%	1,010.0 µg/mL	+/- 23.1865
3	Benzidine	92-87-5	S221205RSR	99%	1,008.0 µg/mL	+/- 23.1406
4	epsilon-Caprolactam	105-60-2	I16X016	99%	1,008.0 µg/mL	+/- 23.1406
Solvent: Methylene chloride						
CAS # 75-09-2						
Purity 99%						

REVIEWED

By Jennifer Pollino at 7:10 am, Sep 13, 2023

Sam Moodler - Operations Tech I

Date Mixed: 13-Sep-2023

Balance: B345965662

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No.: 555223 Lot No.: A0201940
Description: Custom 8270 Plus Standard #1
Custom 8270 Plus Standard #1 1,000µg/mL, Methylene Chloride, 1mL/ampul
Container Size: 2 mL
Expiration Date: September 30, 2025
Handling: This product is photosensitive.
Pkg Amt: > 1 mL
Storage: 10°C or colder
Ship: Ambient

511539
↓
51568
} 7.P.
09/19/

CERTIFIED VALUES						
Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	3,3'-Dichlorobenzidine	91-94-1	S230321RSR	99%	1,001.0 µg/mL	+/- 22.9799
2	Atrazine	1912-24-9	5FYWL	99%	1,010.0 µg/mL	+/- 23.1865
3	Benzidine	92-87-5	S221205RSR	99%	1,008.0 µg/mL	+/- 23.1406
4	epsilon-Caprolactam	105-60-2	I16X016	99%	1,008.0 µg/mL	+/- 23.1406
Solvent: Methylene chloride						
CAS # 75-09-2						
Purity 99%						

REVIEWED

By Jennifer Pollino at 7:10 am, Sep 13, 2023

Sam Moodler - Operations Tech I

Date Mixed: 13-Sep-2023

Balance: B345965662

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31853 **Lot No.:** A0196453
Description : 1,4-dioxane
1,4-Dioxane 2,000µg/mL, Methylene Chloride, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : March 31, 2028 **Storage:** 0°C or colder
Ship: Ambient

S11749
↓
S11794 } RC / 11/30/23

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,4-Dioxane	123-91-1	SHBN3770	99%	2,013.0 µg/mL	+/- 25.0521

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Quality Confirmation Test

Column:

30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant flow 1.8 mL/min.

Temp. Program:

80°C (hold 0.1 min.) to 330°C
@ 9.6°C/min. (hold 2.86 min.)

Inj. Temp:

250°C

Det. Temp:

340°C

Det. Type:


FID

Split Vent:

100 mL/min.

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Sam Moodler
Sam Moodler - Operations Tech I

Date Mixed: 30-Mar-2023

Balance Serial # B707717271

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 31-Mar-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31853 **Lot No.:** A0196453
Description : 1,4-dioxane
1,4-Dioxane 2,000µg/mL, Methylene Chloride, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : March 31, 2028 **Storage:** 0°C or colder
Ship: Ambient

S11749
↓
S11794 } RC / 11/30/23

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,4-Dioxane	123-91-1	SHBN3770	99%	2,013.0 µg/mL	+/- 25.0521

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Quality Confirmation Test

Column:

30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant flow 1.8 mL/min.

Temp. Program:

80°C (hold 0.1 min.) to 330°C
@ 9.6°C/min. (hold 2.86 min.)

Inj. Temp:

250°C

Det. Temp:

340°C

Det. Type:


FID

Split Vent:

100 ml/min.

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Sam Moodler
Sam Moodler - Operations Tech I

Date Mixed: 30-Mar-2023

Balance Serial # B707717271

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 31-Mar-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31853 **Lot No.:** A0196453
Description : 1,4-dioxane
1,4-Dioxane 2,000µg/mL, Methylene Chloride, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : March 31, 2028 **Storage:** 0°C or colder
Ship: Ambient

S11749
↓
S11794 } RC / 11/30/23

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,4-Dioxane	123-91-1	SHBN3770	99%	2,013.0 µg/mL	+/- 25.0521

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Quality Confirmation Test

Column:

30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant flow 1.8 mL/min.

Temp. Program:

80°C (hold 0.1 min.) to 330°C
@ 9.6°C/min. (hold 2.86 min.)

Inj. Temp:

250°C

Det. Temp:

340°C

Det. Type:


FID

Split Vent:

100 ml/min.

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Sam Moodler
Sam Moodler - Operations Tech I

Date Mixed: 30-Mar-2023

Balance Serial # B707717271

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 31-Mar-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31853 **Lot No.:** A0196453
Description : 1,4-dioxane
1,4-Dioxane 2,000µg/mL, Methylene Chloride, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : March 31, 2028 **Storage:** 0°C or colder
Ship: Ambient

S11749
↓
S11794 } RC / 11/30/23

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,4-Dioxane	123-91-1	SHBN3770	99%	2,013.0 µg/mL	+/- 25.0521

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Quality Confirmation Test

Column:

30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant flow 1.8 mL/min.

Temp. Program:

80°C (hold 0.1 min.) to 330°C
@ 9.6°C/min. (hold 2.86 min.)

Inj. Temp:

250°C

Det. Temp:

340°C

Det. Type:


FID

Split Vent:

100 ml/min.

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Sam Moodler
Sam Moodler - Operations Tech I

Date Mixed: 30-Mar-2023

Balance Serial # B707717271

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 31-Mar-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31853 **Lot No.:** A0196453
Description : 1,4-dioxane
1,4-Dioxane 2,000µg/mL, Methylene Chloride, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : March 31, 2028 **Storage:** 0°C or colder
Ship: Ambient

S11749
↓
S11794 } RC / 11/30/23

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,4-Dioxane	123-91-1	SHBN3770	99%	2,013.0 µg/mL	+/- 25.0521

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Quality Confirmation Test

Column:

30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant flow 1.8 mL/min.

Temp. Program:

80°C (hold 0.1 min.) to 330°C
@ 9.6°C/min. (hold 2.86 min.)

Inj. Temp:

250°C

Det. Temp:

340°C

Det. Type:


FID

Split Vent:

100 mL/min.

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Sam Moodler
Sam Moodler - Operations Tech I

Date Mixed: 30-Mar-2023

Balance Serial # B707717271

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 31-Mar-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31206 **Lot No.:** A0201320

Description : SV Internal Standard Mix 2mg/ml
SV Internal Standard Mix 2mg/ml 2000 µg/ml, Methylene Chloride,
1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2029 **Storage:** 10°C or colder

Handling: Sonication required. Mix is photosensitive. **Ship:** Ambient

S12013
↓
S12042 } RC
12/26/23

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,4-Dichlorobenzene-d4	3855-82-1	PR-30447	99%	2,017.0 µg/mL	+/- 90.8469
2	Naphthalene-d8	1146-65-2	M-2180	99%	2,011.3 µg/mL	+/- 90.5917
3	Acenaphthene-d10	15067-26-2	PR-33507	99%	2,008.6 µg/mL	+/- 90.4685
4	Phenanthrene-d10	1517-22-2	PR-32303	99%	2,019.4 µg/mL	+/- 90.9550
5	Chrysene-d12	1719-03-5	PR-32210	99%	2,013.7 µg/mL	+/- 90.6968
6	Perylene-d12	1520-96-3	PR-33205	99%	2,012.7 µg/mL	+/- 90.6517

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Quality Confirmation Test

Column:

30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant pressure 10 psi.

Temp. Program:

75°C (hold 1 min.) to 330°C
@ 20°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

330°C

Det. Type:


FID

Split Vent:

10 ml/min.

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Peter Robbins - Operations Technician I

Date Mixed: 23-Aug-2023

Balance Serial # B345965662

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 25-Aug-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31206 **Lot No.:** A0201320

Description : SV Internal Standard Mix 2mg/ml
SV Internal Standard Mix 2mg/ml 2000 µg/ml, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2029 **Storage:** 10°C or colder

Handling: Sonication required. Mix is photosensitive. **Ship:** Ambient

S12013
↓
S12042 } RC
12/26/23

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,4-Dichlorobenzene-d4	3855-82-1	PR-30447	99%	2,017.0 µg/mL	+/- 90.8469
2	Naphthalene-d8	1146-65-2	M-2180	99%	2,011.3 µg/mL	+/- 90.5917
3	Acenaphthene-d10	15067-26-2	PR-33507	99%	2,008.6 µg/mL	+/- 90.4685
4	Phenanthrene-d10	1517-22-2	PR-32303	99%	2,019.4 µg/mL	+/- 90.9550
5	Chrysene-d12	1719-03-5	PR-32210	99%	2,013.7 µg/mL	+/- 90.6968
6	Perylene-d12	1520-96-3	PR-33205	99%	2,012.7 µg/mL	+/- 90.6517

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Quality Confirmation Test

Column:

30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant pressure 10 psi.

Temp. Program:

75°C (hold 1 min.) to 330°C
@ 20°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

330°C

Det. Type:


FID

Split Vent:

10 ml/min.

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Peter Robbins - Operations Technician I

Date Mixed: 23-Aug-2023

Balance Serial # B345965662

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 25-Aug-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received: _____

Certificate of Analysis

Rev 0

Page 1 of 1

Catalog No.:	Lot No.:	Storage:	Solvent:	Exp. Date:	Description:
Z-110816-01	414127	≤ -10 °C	Methylene Chloride	6/21/2025	Custom 8270 Mix, 4-79, 1000 mg/L, 1 mL

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
atrazine	1912-24-9	99.5	337.7.3P	997 ± 5.81
benzidine	92-87-5	99.9	124.18.6.2P	991.8 ± 5.77
caprolactam	105-60-2	99.9	271.1.6P	999 ± 5.82

912075 } RC
↓
912079 } 02/01/24

*Not a certified value

Manufactured by o2si smart solutions, Accredited to ISO 9001:2008 by NSF and ISO/IEC 17025:2005 (Certification No. 3031.01) and ISO Guide 34:2009 (Certification No. 3031.02) by A2LA

Certified By: _____

Shane Overcash
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetrically.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555224 **Lot No.:** A0207706

Description : Custom 8270 Plus Standard #2

Custom 8270 Plus Standard #2 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : February 28, 2026 **Storage:** 10°C or colder

Ship: Ambient

512082 } RC/
↓
512111 } 02/22/24

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,2,4,5-Tetrachlorobenzene	95-94-3	MKCT9480	99%	1,001.0 µg/mL	+/- 29.424320
2	Acetophenone	98-86-2	STBH8205	99%	1,004.0 µg/mL	+/- 29.512504
3	Benzaldehyde	100-52-7	RD231129RSRA	99%	1,005.0 µg/mL	+/- 29.541899
4	Benzoic acid	65-85-0	MKCR2694	99%	1,003.0 µg/mL	+/- 29.483110
5	Biphenyl	92-52-4	MKCL6515	99%	1,006.0 µg/mL	+/- 29.571294

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

John Friedline - Operations Technician I

Date Mixed: 12-Feb-2024

Balance: B345965662



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555224 **Lot No.:** A0207706

Description : Custom 8270 Plus Standard #2

Custom 8270 Plus Standard #2 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : February 28, 2026 **Storage:** 10°C or colder

Ship: Ambient

512082 } RC/
↓
512111 } 02/22/24

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,2,4,5-Tetrachlorobenzene	95-94-3	MKCT9480	99%	1,001.0 µg/mL	+/- 29.424320
2	Acetophenone	98-86-2	STBH8205	99%	1,004.0 µg/mL	+/- 29.512504
3	Benzaldehyde	100-52-7	RD231129RSRA	99%	1,005.0 µg/mL	+/- 29.541899
4	Benzoic acid	65-85-0	MKCR2694	99%	1,003.0 µg/mL	+/- 29.483110
5	Biphenyl	92-52-4	MKCL6515	99%	1,006.0 µg/mL	+/- 29.571294

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

John Friedline - Operations Technician I

Date Mixed: 12-Feb-2024

Balance: B345965662



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555224 **Lot No.:** A0207706

Description : Custom 8270 Plus Standard #2

Custom 8270 Plus Standard #2 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : February 28, 2026 **Storage:** 10°C or colder

Ship: Ambient

512082 } RC/
↓
512111 } 02/22/24

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,2,4,5-Tetrachlorobenzene	95-94-3	MKCT9480	99%	1,001.0 µg/mL	+/- 29.424320
2	Acetophenone	98-86-2	STBH8205	99%	1,004.0 µg/mL	+/- 29.512504
3	Benzaldehyde	100-52-7	RD231129RSRA	99%	1,005.0 µg/mL	+/- 29.541899
4	Benzoic acid	65-85-0	MKCR2694	99%	1,003.0 µg/mL	+/- 29.483110
5	Biphenyl	92-52-4	MKCL6515	99%	1,006.0 µg/mL	+/- 29.571294

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

John Friedline - Operations Technician I

Date Mixed: 12-Feb-2024

Balance: B345965662



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555224 **Lot No.:** A0207706

Description : Custom 8270 Plus Standard #2

Custom 8270 Plus Standard #2 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : February 28, 2026 **Storage:** 10°C or colder

Ship: Ambient

512082 } RC/
↓
512111 } 02/22/24

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,2,4,5-Tetrachlorobenzene	95-94-3	MKCT9480	99%	1,001.0 µg/mL	+/- 29.424320
2	Acetophenone	98-86-2	STBH8205	99%	1,004.0 µg/mL	+/- 29.512504
3	Benzaldehyde	100-52-7	RD231129RSRA	99%	1,005.0 µg/mL	+/- 29.541899
4	Benzoic acid	65-85-0	MKCR2694	99%	1,003.0 µg/mL	+/- 29.483110
5	Biphenyl	92-52-4	MKCL6515	99%	1,006.0 µg/mL	+/- 29.571294

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

John Friedline - Operations Technician I

Date Mixed: 12-Feb-2024

Balance: B345965662



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555224 **Lot No.:** A0207706

Description : Custom 8270 Plus Standard #2

Custom 8270 Plus Standard #2 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : February 28, 2026 **Storage:** 10°C or colder

Ship: Ambient

512082 } RC/
↓
512111 } 02/22/24

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,2,4,5-Tetrachlorobenzene	95-94-3	MKCT9480	99%	1,001.0 µg/mL	+/- 29.424320
2	Acetophenone	98-86-2	STBH8205	99%	1,004.0 µg/mL	+/- 29.512504
3	Benzaldehyde	100-52-7	RD231129RSRA	99%	1,005.0 µg/mL	+/- 29.541899
4	Benzoic acid	65-85-0	MKCR2694	99%	1,003.0 µg/mL	+/- 29.483110
5	Biphenyl	92-52-4	MKCL6515	99%	1,006.0 µg/mL	+/- 29.571294

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

John Friedline - Operations Technician I

Date Mixed: 12-Feb-2024

Balance: B345965662



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555224 **Lot No.:** A0207706

Description : Custom 8270 Plus Standard #2

Custom 8270 Plus Standard #2 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : February 28, 2026 **Storage:** 10°C or colder

Ship: Ambient

512082 } RC/
↓
512111 } 02/22/24

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,2,4,5-Tetrachlorobenzene	95-94-3	MKCT9480	99%	1,001.0 µg/mL	+/- 29.424320
2	Acetophenone	98-86-2	STBH8205	99%	1,004.0 µg/mL	+/- 29.512504
3	Benzaldehyde	100-52-7	RD231129RSRA	99%	1,005.0 µg/mL	+/- 29.541899
4	Benzoic acid	65-85-0	MKCR2694	99%	1,003.0 µg/mL	+/- 29.483110
5	Biphenyl	92-52-4	MKCL6515	99%	1,006.0 µg/mL	+/- 29.571294

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

John Friedline - Operations Technician I

Date Mixed: 12-Feb-2024

Balance: B345965662



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555224 **Lot No.:** A0207706

Description : Custom 8270 Plus Standard #2

Custom 8270 Plus Standard #2 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : February 28, 2026 **Storage:** 10°C or colder

Ship: Ambient

512082 } RC/
↓
512111 } 02/22/24

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,2,4,5-Tetrachlorobenzene	95-94-3	MKCT9480	99%	1,001.0 µg/mL	+/- 29.424320
2	Acetophenone	98-86-2	STBH8205	99%	1,004.0 µg/mL	+/- 29.512504
3	Benzaldehyde	100-52-7	RD231129RSRA	99%	1,005.0 µg/mL	+/- 29.541899
4	Benzoic acid	65-85-0	MKCR2694	99%	1,003.0 µg/mL	+/- 29.483110
5	Biphenyl	92-52-4	MKCL6515	99%	1,006.0 µg/mL	+/- 29.571294

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

John Friedline - Operations Technician I

Date Mixed: 12-Feb-2024

Balance: B345965662



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555224 **Lot No.:** A0207706

Description : Custom 8270 Plus Standard #2

Custom 8270 Plus Standard #2 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : February 28, 2026 **Storage:** 10°C or colder

Ship: Ambient

512082 } RC/
↓
512111 } 02/22/24

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,2,4,5-Tetrachlorobenzene	95-94-3	MKCT9480	99%	1,001.0 µg/mL	+/- 29.424320
2	Acetophenone	98-86-2	STBH8205	99%	1,004.0 µg/mL	+/- 29.512504
3	Benzaldehyde	100-52-7	RD231129RSRA	99%	1,005.0 µg/mL	+/- 29.541899
4	Benzoic acid	65-85-0	MKCR2694	99%	1,003.0 µg/mL	+/- 29.483110
5	Biphenyl	92-52-4	MKCL6515	99%	1,006.0 µg/mL	+/- 29.571294

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

John Friedline - Operations Technician I

Date Mixed: 12-Feb-2024

Balance: B345965662



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555224 **Lot No.:** A0207706

Description : Custom 8270 Plus Standard #2

Custom 8270 Plus Standard #2 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : February 28, 2026 **Storage:** 10°C or colder

Ship: Ambient

512082 } RC/
↓
512111 } 02/22/24

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,2,4,5-Tetrachlorobenzene	95-94-3	MKCT9480	99%	1,001.0 µg/mL	+/- 29.424320
2	Acetophenone	98-86-2	STBH8205	99%	1,004.0 µg/mL	+/- 29.512504
3	Benzaldehyde	100-52-7	RD231129RSRA	99%	1,005.0 µg/mL	+/- 29.541899
4	Benzoic acid	65-85-0	MKCR2694	99%	1,003.0 µg/mL	+/- 29.483110
5	Biphenyl	92-52-4	MKCL6515	99%	1,006.0 µg/mL	+/- 29.571294

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

John Friedline - Operations Technician I

Date Mixed: 12-Feb-2024

Balance: B345965662



5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received: _____

Certificate of Analysis

Rev 0

Page 1 of 1

Catalog No.:	Lot No.:	Storage:	Solvent:	Exp. Date:	Description:
Z-020223-01	454157	≤ -10 °C	P/T Methanol	6/10/2026	1,4-Dioxane Solution, 2000 mg/L, 1 mL

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
1,4-dioxane	123-91-1	100	223.1.3P	1997 ± 57.08

512112 } RC/
↓
912116 } 03/08/24

*Not a certified value

Certified By: _____

Melissa Workoff
Melissa Workoff
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values
listed are determined gravimetrically.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31850 **Lot No.:** A0203726

Description : 8270 MegaMix®
8270 MegaMix® 500-1000 µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2025 **Storage:** 0°C or colder

Handling: Sonication required. Mix is photosensitive. **Ship:** Ambient

512117 } RC/
↓ 03/18/24
512146 }

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Pyridine	110-86-1	SHBP6240	99%	1,001.6 µg/mL	+/- 36.4412
2	N-Nitrosodimethylamine	62-75-9	230209JLM	99%	1,005.9 µg/mL	+/- 36.5968
3	Phenol	108-95-2	MKCK1120	99%	1,003.3 µg/mL	+/- 36.5038
4	Aniline	62-53-3	X22F726	99%	1,005.8 µg/mL	+/- 36.5928
5	Bis(2-chloroethyl)ether	111-44-4	SHBL6942	99%	1,008.1 µg/mL	+/- 36.6776
6	2-Chlorophenol	95-57-8	STBJ3909	99%	1,001.8 µg/mL	+/- 36.4492
7	1,3-Dichlorobenzene	541-73-1	BCCD5315	99%	1,002.3 µg/mL	+/- 36.4654
8	1,4-Dichlorobenzene	106-46-7	MKBS7929V	99%	1,003.7 µg/mL	+/- 36.5159
9	Benzyl alcohol	100-51-6	SHBK5469	99%	1,008.7 µg/mL	+/- 36.6979
10	1,2-Dichlorobenzene	95-50-1	SHBN3835	99%	1,000.3 µg/mL	+/- 36.3926
11	2-Methylphenol (o-cresol)	95-48-7	SHBN7598	99%	1,003.5 µg/mL	+/- 36.5099
12	2,2'-oxybis(1-chloropropane)	108-60-1	29-MAR-45-5	99%	1,007.3 µg/mL	+/- 36.6493
13	3-Methylphenol (m-cresol)	108-39-4	STBJ0710	99%	504.3 µg/mL	+/- 18.3500
14	4-Methylphenol (p-cresol)	106-44-5	SHBN3411	99%	503.6 µg/mL	+/- 18.3237
15	N-Nitroso-di-n-propylamine	621-64-7	N63MG	99%	1,008.3 µg/mL	+/- 36.6857
16	Hexachloroethane	67-72-1	QTORH	99%	1,007.5 µg/mL	+/- 36.6554
17	Nitrobenzene	98-95-3	10224044	99%	1,008.6 µg/mL	+/- 36.6938

18	Isophorone	78-59-1	MKCC9506	99%	1,005.9	µg/mL	+/- 36.5988
19	2-Nitrophenol	88-75-5	RP230710	99%	1,003.2	µg/mL	+/- 36.4998
20	2,4-Dimethylphenol	105-67-9	XW5GK	99%	1,003.8	µg/mL	+/- 36.5200
21	Bis(2-chloroethoxy)methane	111-91-1	13670200	99%	1,002.1	µg/mL	+/- 36.4573
22	2,4-Dichlorophenol	120-83-2	BCBZ6787	99%	1,003.7	µg/mL	+/- 36.5180
23	1,2,4-Trichlorobenzene	120-82-1	SHBP5900	99%	1,007.6	µg/mL	+/- 36.6574
24	Naphthalene	91-20-3	STBL1057	99%	1,008.3	µg/mL	+/- 36.6837
25	4-Chloroaniline	106-47-8	BCCJ3217	99%	1,001.3	µg/mL	+/- 36.4290
26	Hexachlorobutadiene	87-68-3	RP230823RSR	98%	1,008.3	µg/mL	+/- 36.6829
27	4-Chloro-3-methylphenol	59-50-7	BCCD4461	99%	1,003.1	µg/mL	+/- 36.4937
28	2-Methylnaphthalene	91-57-6	STBK0259	96%	1,001.9	µg/mL	+/- 36.4505
29	1-Methylnaphthalene	90-12-0	5234.00-8	98%	1,000.0	µg/mL	+/- 36.3838
30	Hexachlorocyclopentadiene	77-47-4	099063I14L	98%	1,008.5	µg/mL	+/- 36.6909
31	2,4,6-Trichlorophenol	88-06-2	STBJ5914	99%	1,004.4	µg/mL	+/- 36.5442
32	2,4,5-Trichlorophenol	95-95-4	FHN01	98%	1,001.9	µg/mL	+/- 36.4512
33	2-Chloronaphthalene	91-58-7	RPN7O	99%	1,001.1	µg/mL	+/- 36.4230
34	2-Nitroaniline	88-74-4	RP230531	99%	1,002.9	µg/mL	+/- 36.4876
35	1,4-Dinitrobenzene	100-25-4	RP230816	99%	1,005.7	µg/mL	+/- 36.5887
36	Acenaphthylene	208-96-8	p06V	98%	1,009.5	µg/mL	+/- 36.7265
37	1,3-Dinitrobenzene	99-65-0	1-DXX-24-1	99%	1,004.4	µg/mL	+/- 36.5422
38	Dimethylphthalate	131-11-3	358221L17K	99%	1,005.9	µg/mL	+/- 36.5968
39	2,6-Dinitrotoluene	606-20-2	BCCG1833	99%	1,003.2	µg/mL	+/- 36.4998
40	1,2-Dinitrobenzene	528-29-0	RP230428	99%	1,002.2	µg/mL	+/- 36.4634
41	Acenaphthene	83-32-9	MKCR7169	99%	1,009.3	µg/mL	+/- 36.7221
42	3-Nitroaniline	99-09-2	RP230822RSR	99%	1,003.9	µg/mL	+/- 36.5240
43	2,4-Dinitrophenol	51-28-5	DR230417RSR	99%	1,002.0	µg/mL	+/- 36.4553
44	Dibenzofuran	132-64-9	MKCD9952	99%	1,006.7	µg/mL	+/- 36.6251
45	2,4-Dinitrotoluene	121-14-2	MKAA0690V	99%	1,003.8	µg/mL	+/- 36.5220
46	4-Nitrophenol	100-02-7	RP230627	99%	1,002.3	µg/mL	+/- 36.4674
47	2,3,4,6-Tetrachlorophenol	58-90-2	PR-30126	99%	1,008.7	µg/mL	+/- 36.6979
48	2,3,5,6-Tetrachlorophenol	935-95-5	RP230919	99%	1,006.3	µg/mL	+/- 36.6130
49	Fluorene	86-73-7	10241100	99%	1,008.3	µg/mL	+/- 36.6857
50	4-Chlorophenyl phenyl ether	7005-72-3	MKCT7248	99%	1,003.8	µg/mL	+/- 36.5220
51	Diethylphthalate	84-66-2	MKCD2547	99%	1,008.6	µg/mL	+/- 36.6958
52	4-Nitroaniline	100-01-6	RP230111	99%	1,001.1	µg/mL	+/- 36.4230
53	4,6-Dinitro-2-methylphenol (Dinitro-o-cresol)	534-52-1	230718JLM	99%	1,002.0	µg/mL	+/- 36.4553

54	Diphenylamine	122-39-4	MKCH1042	99%	1,002.3	µg/mL	+/- 36.4674
55	Azobenzene	103-33-3	BCKK0887	99%	1,005.8	µg/mL	+/- 36.5928
56	4-Bromophenyl phenyl ether	101-55-3	STBH6361	99%	1,003.0	µg/mL	+/- 36.4917
57	Hexachlorobenzene	118-74-1	14821700	99%	1,007.5	µg/mL	+/- 36.6554
58	Pentachlorophenol	87-86-5	RP230530RSR	99%	1,008.8	µg/mL	+/- 36.7019
59	Phenanthrene	85-01-8	MKCQ8876	99%	1,008.4	µg/mL	+/- 36.6877
60	Anthracene	120-12-7	MKCR0570	99%	1,009.0	µg/mL	+/- 36.7100
61	Carbazole	86-74-8	14351100	99%	1,000.9	µg/mL	+/- 36.4149
62	Di-n-butylphthalate	84-74-2	MKCN4337	99%	1,007.6	µg/mL	+/- 36.6595
63	Fluoranthene	206-44-0	MKCQ4728	99%	1,009.6	µg/mL	+/- 36.7302
64	Pyrene	129-00-0	BCCG8479	98%	1,007.2	µg/mL	+/- 36.6453
65	Benzyl butyl phthalate	85-68-7	X12I018	99%	1,002.1	µg/mL	+/- 36.4573
66	Bis(2-ethylhexyl)adipate	103-23-1	MKCM1988	99%	1,005.2	µg/mL	+/- 36.5705
67	Benz(a)anthracene	56-55-3	I220012022BAA	99%	1,002.2	µg/mL	+/- 36.4614
68	Chrysene	218-01-9	RP230601	99%	1,008.3	µg/mL	+/- 36.6837
69	Bis(2-ethylhexyl)phthalate	117-81-7	MKCQ3468	99%	1,001.8	µg/mL	+/- 36.4472
70	Di-n-octyl phthalate	117-84-0	14382700	99%	1,006.0	µg/mL	+/- 36.6008
71	Benzo(b)fluoranthene	205-99-2	012013B	99%	1,002.8	µg/mL	+/- 36.4836
72	Benzo(k)fluoranthene	207-08-9	012022K	99%	1,003.0	µg/mL	+/- 36.4917
73	Benzo(a)pyrene	50-32-8	P54915-0703	99%	1,002.3	µg/mL	+/- 36.4674
74	Indeno(1,2,3-cd)pyrene	193-39-5	12-JKL-118-9	97%	1,009.4	µg/mL	+/- 36.7243
75	Dibenz(a,h)anthracene	53-70-3	2-ASA-59-1	99%	1,007.6	µg/mL	+/- 36.6595
76	Benzo(g,h,i)perylene	191-24-2	RP231003RSR	99%	1,002.9	µg/mL	+/- 36.4876

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31850 **Lot No.:** A0203726

Description : 8270 MegaMix®
8270 MegaMix® 500-1000 µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2025 **Storage:** 0°C or colder

Handling: Sonication required. Mix is photosensitive. **Ship:** Ambient

512117 } RC/
↓ 03/18/24
512146 }

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Pyridine	110-86-1	SHBP6240	99%	1,001.6 µg/mL	+/- 36.4412
2	N-Nitrosodimethylamine	62-75-9	230209JLM	99%	1,005.9 µg/mL	+/- 36.5968
3	Phenol	108-95-2	MKCK1120	99%	1,003.3 µg/mL	+/- 36.5038
4	Aniline	62-53-3	X22F726	99%	1,005.8 µg/mL	+/- 36.5928
5	Bis(2-chloroethyl)ether	111-44-4	SHBL6942	99%	1,008.1 µg/mL	+/- 36.6776
6	2-Chlorophenol	95-57-8	STBJ3909	99%	1,001.8 µg/mL	+/- 36.4492
7	1,3-Dichlorobenzene	541-73-1	BCCD5315	99%	1,002.3 µg/mL	+/- 36.4654
8	1,4-Dichlorobenzene	106-46-7	MKBS7929V	99%	1,003.7 µg/mL	+/- 36.5159
9	Benzyl alcohol	100-51-6	SHBK5469	99%	1,008.7 µg/mL	+/- 36.6979
10	1,2-Dichlorobenzene	95-50-1	SHBN3835	99%	1,000.3 µg/mL	+/- 36.3926
11	2-Methylphenol (o-cresol)	95-48-7	SHBN7598	99%	1,003.5 µg/mL	+/- 36.5099
12	2,2'-oxybis(1-chloropropane)	108-60-1	29-MAR-45-5	99%	1,007.3 µg/mL	+/- 36.6493
13	3-Methylphenol (m-cresol)	108-39-4	STBJ0710	99%	504.3 µg/mL	+/- 18.3500
14	4-Methylphenol (p-cresol)	106-44-5	SHBN3411	99%	503.6 µg/mL	+/- 18.3237
15	N-Nitroso-di-n-propylamine	621-64-7	N63MG	99%	1,008.3 µg/mL	+/- 36.6857
16	Hexachloroethane	67-72-1	QTORH	99%	1,007.5 µg/mL	+/- 36.6554
17	Nitrobenzene	98-95-3	10224044	99%	1,008.6 µg/mL	+/- 36.6938

18	Isophorone	78-59-1	MKCC9506	99%	1,005.9	µg/mL	+/- 36.5988
19	2-Nitrophenol	88-75-5	RP230710	99%	1,003.2	µg/mL	+/- 36.4998
20	2,4-Dimethylphenol	105-67-9	XW5GK	99%	1,003.8	µg/mL	+/- 36.5200
21	Bis(2-chloroethoxy)methane	111-91-1	13670200	99%	1,002.1	µg/mL	+/- 36.4573
22	2,4-Dichlorophenol	120-83-2	BCBZ6787	99%	1,003.7	µg/mL	+/- 36.5180
23	1,2,4-Trichlorobenzene	120-82-1	SHBP5900	99%	1,007.6	µg/mL	+/- 36.6574
24	Naphthalene	91-20-3	STBL1057	99%	1,008.3	µg/mL	+/- 36.6837
25	4-Chloroaniline	106-47-8	BCCJ3217	99%	1,001.3	µg/mL	+/- 36.4290
26	Hexachlorobutadiene	87-68-3	RP230823RSR	98%	1,008.3	µg/mL	+/- 36.6829
27	4-Chloro-3-methylphenol	59-50-7	BCCD4461	99%	1,003.1	µg/mL	+/- 36.4937
28	2-Methylnaphthalene	91-57-6	STBK0259	96%	1,001.9	µg/mL	+/- 36.4505
29	1-Methylnaphthalene	90-12-0	5234.00-8	98%	1,000.0	µg/mL	+/- 36.3838
30	Hexachlorocyclopentadiene	77-47-4	099063I14L	98%	1,008.5	µg/mL	+/- 36.6909
31	2,4,6-Trichlorophenol	88-06-2	STBJ5914	99%	1,004.4	µg/mL	+/- 36.5442
32	2,4,5-Trichlorophenol	95-95-4	FHN01	98%	1,001.9	µg/mL	+/- 36.4512
33	2-Chloronaphthalene	91-58-7	RPN7O	99%	1,001.1	µg/mL	+/- 36.4230
34	2-Nitroaniline	88-74-4	RP230531	99%	1,002.9	µg/mL	+/- 36.4876
35	1,4-Dinitrobenzene	100-25-4	RP230816	99%	1,005.7	µg/mL	+/- 36.5887
36	Acenaphthylene	208-96-8	p06V	98%	1,009.5	µg/mL	+/- 36.7265
37	1,3-Dinitrobenzene	99-65-0	1-DXX-24-1	99%	1,004.4	µg/mL	+/- 36.5422
38	Dimethylphthalate	131-11-3	358221L17K	99%	1,005.9	µg/mL	+/- 36.5968
39	2,6-Dinitrotoluene	606-20-2	BCCG1833	99%	1,003.2	µg/mL	+/- 36.4998
40	1,2-Dinitrobenzene	528-29-0	RP230428	99%	1,002.2	µg/mL	+/- 36.4634
41	Acenaphthene	83-32-9	MKCR7169	99%	1,009.3	µg/mL	+/- 36.7221
42	3-Nitroaniline	99-09-2	RP230822RSR	99%	1,003.9	µg/mL	+/- 36.5240
43	2,4-Dinitrophenol	51-28-5	DR230417RSR	99%	1,002.0	µg/mL	+/- 36.4553
44	Dibenzofuran	132-64-9	MKCD9952	99%	1,006.7	µg/mL	+/- 36.6251
45	2,4-Dinitrotoluene	121-14-2	MKAA0690V	99%	1,003.8	µg/mL	+/- 36.5220
46	4-Nitrophenol	100-02-7	RP230627	99%	1,002.3	µg/mL	+/- 36.4674
47	2,3,4,6-Tetrachlorophenol	58-90-2	PR-30126	99%	1,008.7	µg/mL	+/- 36.6979
48	2,3,5,6-Tetrachlorophenol	935-95-5	RP230919	99%	1,006.3	µg/mL	+/- 36.6130
49	Fluorene	86-73-7	10241100	99%	1,008.3	µg/mL	+/- 36.6857
50	4-Chlorophenyl phenyl ether	7005-72-3	MKCT7248	99%	1,003.8	µg/mL	+/- 36.5220
51	Diethylphthalate	84-66-2	MKCD2547	99%	1,008.6	µg/mL	+/- 36.6958
52	4-Nitroaniline	100-01-6	RP230111	99%	1,001.1	µg/mL	+/- 36.4230
53	4,6-Dinitro-2-methylphenol (Dinitro-o-cresol)	534-52-1	230718JLM	99%	1,002.0	µg/mL	+/- 36.4553

54	Diphenylamine	122-39-4	MKCH1042	99%	1,002.3	µg/mL	+/- 36.4674
55	Azobenzene	103-33-3	BCKK0887	99%	1,005.8	µg/mL	+/- 36.5928
56	4-Bromophenyl phenyl ether	101-55-3	STBH6361	99%	1,003.0	µg/mL	+/- 36.4917
57	Hexachlorobenzene	118-74-1	14821700	99%	1,007.5	µg/mL	+/- 36.6554
58	Pentachlorophenol	87-86-5	RP230530RSR	99%	1,008.8	µg/mL	+/- 36.7019
59	Phenanthrene	85-01-8	MKCQ8876	99%	1,008.4	µg/mL	+/- 36.6877
60	Anthracene	120-12-7	MKCR0570	99%	1,009.0	µg/mL	+/- 36.7100
61	Carbazole	86-74-8	14351100	99%	1,000.9	µg/mL	+/- 36.4149
62	Di-n-butylphthalate	84-74-2	MKCN4337	99%	1,007.6	µg/mL	+/- 36.6595
63	Fluoranthene	206-44-0	MKCQ4728	99%	1,009.6	µg/mL	+/- 36.7302
64	Pyrene	129-00-0	BCCG8479	98%	1,007.2	µg/mL	+/- 36.6453
65	Benzyl butyl phthalate	85-68-7	X12I018	99%	1,002.1	µg/mL	+/- 36.4573
66	Bis(2-ethylhexyl)adipate	103-23-1	MKCM1988	99%	1,005.2	µg/mL	+/- 36.5705
67	Benz(a)anthracene	56-55-3	I220012022BAA	99%	1,002.2	µg/mL	+/- 36.4614
68	Chrysene	218-01-9	RP230601	99%	1,008.3	µg/mL	+/- 36.6837
69	Bis(2-ethylhexyl)phthalate	117-81-7	MKCQ3468	99%	1,001.8	µg/mL	+/- 36.4472
70	Di-n-octyl phthalate	117-84-0	14382700	99%	1,006.0	µg/mL	+/- 36.6008
71	Benzo(b)fluoranthene	205-99-2	012013B	99%	1,002.8	µg/mL	+/- 36.4836
72	Benzo(k)fluoranthene	207-08-9	012022K	99%	1,003.0	µg/mL	+/- 36.4917
73	Benzo(a)pyrene	50-32-8	P54915-0703	99%	1,002.3	µg/mL	+/- 36.4674
74	Indeno(1,2,3-cd)pyrene	193-39-5	12-JKL-118-9	97%	1,009.4	µg/mL	+/- 36.7243
75	Dibenz(a,h)anthracene	53-70-3	2-ASA-59-1	99%	1,007.6	µg/mL	+/- 36.6595
76	Benzo(g,h,i)perylene	191-24-2	RP231003RSR	99%	1,002.9	µg/mL	+/- 36.4876

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31850 **Lot No.:** A0203726

Description : 8270 MegaMix®
8270 MegaMix® 500-1000 µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2025 **Storage:** 0°C or colder

Handling: Sonication required. Mix is photosensitive. **Ship:** Ambient

512117 } RC/
↓ 03/18/24
512146 }

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Pyridine	110-86-1	SHBP6240	99%	1,001.6 µg/mL	+/- 36.4412
2	N-Nitrosodimethylamine	62-75-9	230209JLM	99%	1,005.9 µg/mL	+/- 36.5968
3	Phenol	108-95-2	MKCK1120	99%	1,003.3 µg/mL	+/- 36.5038
4	Aniline	62-53-3	X22F726	99%	1,005.8 µg/mL	+/- 36.5928
5	Bis(2-chloroethyl)ether	111-44-4	SHBL6942	99%	1,008.1 µg/mL	+/- 36.6776
6	2-Chlorophenol	95-57-8	STBJ3909	99%	1,001.8 µg/mL	+/- 36.4492
7	1,3-Dichlorobenzene	541-73-1	BCCD5315	99%	1,002.3 µg/mL	+/- 36.4654
8	1,4-Dichlorobenzene	106-46-7	MKBS7929V	99%	1,003.7 µg/mL	+/- 36.5159
9	Benzyl alcohol	100-51-6	SHBK5469	99%	1,008.7 µg/mL	+/- 36.6979
10	1,2-Dichlorobenzene	95-50-1	SHBN3835	99%	1,000.3 µg/mL	+/- 36.3926
11	2-Methylphenol (o-cresol)	95-48-7	SHBN7598	99%	1,003.5 µg/mL	+/- 36.5099
12	2,2'-oxybis(1-chloropropane)	108-60-1	29-MAR-45-5	99%	1,007.3 µg/mL	+/- 36.6493
13	3-Methylphenol (m-cresol)	108-39-4	STBJ0710	99%	504.3 µg/mL	+/- 18.3500
14	4-Methylphenol (p-cresol)	106-44-5	SHBN3411	99%	503.6 µg/mL	+/- 18.3237
15	N-Nitroso-di-n-propylamine	621-64-7	N63MG	99%	1,008.3 µg/mL	+/- 36.6857
16	Hexachloroethane	67-72-1	QTORH	99%	1,007.5 µg/mL	+/- 36.6554
17	Nitrobenzene	98-95-3	10224044	99%	1,008.6 µg/mL	+/- 36.6938

18	Isophorone	78-59-1	MKCC9506	99%	1,005.9	µg/mL	+/- 36.5988
19	2-Nitrophenol	88-75-5	RP230710	99%	1,003.2	µg/mL	+/- 36.4998
20	2,4-Dimethylphenol	105-67-9	XW5GK	99%	1,003.8	µg/mL	+/- 36.5200
21	Bis(2-chloroethoxy)methane	111-91-1	13670200	99%	1,002.1	µg/mL	+/- 36.4573
22	2,4-Dichlorophenol	120-83-2	BCBZ6787	99%	1,003.7	µg/mL	+/- 36.5180
23	1,2,4-Trichlorobenzene	120-82-1	SHBP5900	99%	1,007.6	µg/mL	+/- 36.6574
24	Naphthalene	91-20-3	STBL1057	99%	1,008.3	µg/mL	+/- 36.6837
25	4-Chloroaniline	106-47-8	BCCJ3217	99%	1,001.3	µg/mL	+/- 36.4290
26	Hexachlorobutadiene	87-68-3	RP230823RSR	98%	1,008.3	µg/mL	+/- 36.6829
27	4-Chloro-3-methylphenol	59-50-7	BCCD4461	99%	1,003.1	µg/mL	+/- 36.4937
28	2-Methylnaphthalene	91-57-6	STBK0259	96%	1,001.9	µg/mL	+/- 36.4505
29	1-Methylnaphthalene	90-12-0	5234.00-8	98%	1,000.0	µg/mL	+/- 36.3838
30	Hexachlorocyclopentadiene	77-47-4	099063I14L	98%	1,008.5	µg/mL	+/- 36.6909
31	2,4,6-Trichlorophenol	88-06-2	STBJ5914	99%	1,004.4	µg/mL	+/- 36.5442
32	2,4,5-Trichlorophenol	95-95-4	FHN01	98%	1,001.9	µg/mL	+/- 36.4512
33	2-Chloronaphthalene	91-58-7	RPN7O	99%	1,001.1	µg/mL	+/- 36.4230
34	2-Nitroaniline	88-74-4	RP230531	99%	1,002.9	µg/mL	+/- 36.4876
35	1,4-Dinitrobenzene	100-25-4	RP230816	99%	1,005.7	µg/mL	+/- 36.5887
36	Acenaphthylene	208-96-8	p06V	98%	1,009.5	µg/mL	+/- 36.7265
37	1,3-Dinitrobenzene	99-65-0	1-DXX-24-1	99%	1,004.4	µg/mL	+/- 36.5422
38	Dimethylphthalate	131-11-3	358221L17K	99%	1,005.9	µg/mL	+/- 36.5968
39	2,6-Dinitrotoluene	606-20-2	BCCG1833	99%	1,003.2	µg/mL	+/- 36.4998
40	1,2-Dinitrobenzene	528-29-0	RP230428	99%	1,002.2	µg/mL	+/- 36.4634
41	Acenaphthene	83-32-9	MKCR7169	99%	1,009.3	µg/mL	+/- 36.7221
42	3-Nitroaniline	99-09-2	RP230822RSR	99%	1,003.9	µg/mL	+/- 36.5240
43	2,4-Dinitrophenol	51-28-5	DR230417RSR	99%	1,002.0	µg/mL	+/- 36.4553
44	Dibenzofuran	132-64-9	MKCD9952	99%	1,006.7	µg/mL	+/- 36.6251
45	2,4-Dinitrotoluene	121-14-2	MKAA0690V	99%	1,003.8	µg/mL	+/- 36.5220
46	4-Nitrophenol	100-02-7	RP230627	99%	1,002.3	µg/mL	+/- 36.4674
47	2,3,4,6-Tetrachlorophenol	58-90-2	PR-30126	99%	1,008.7	µg/mL	+/- 36.6979
48	2,3,5,6-Tetrachlorophenol	935-95-5	RP230919	99%	1,006.3	µg/mL	+/- 36.6130
49	Fluorene	86-73-7	10241100	99%	1,008.3	µg/mL	+/- 36.6857
50	4-Chlorophenyl phenyl ether	7005-72-3	MKCT7248	99%	1,003.8	µg/mL	+/- 36.5220
51	Diethylphthalate	84-66-2	MKCD2547	99%	1,008.6	µg/mL	+/- 36.6958
52	4-Nitroaniline	100-01-6	RP230111	99%	1,001.1	µg/mL	+/- 36.4230
53	4,6-Dinitro-2-methylphenol (Dinitro-o-cresol)	534-52-1	230718JLM	99%	1,002.0	µg/mL	+/- 36.4553

54	Diphenylamine	122-39-4	MKCH1042	99%	1,002.3	µg/mL	+/- 36.4674
55	Azobenzene	103-33-3	BCKK0887	99%	1,005.8	µg/mL	+/- 36.5928
56	4-Bromophenyl phenyl ether	101-55-3	STBH6361	99%	1,003.0	µg/mL	+/- 36.4917
57	Hexachlorobenzene	118-74-1	14821700	99%	1,007.5	µg/mL	+/- 36.6554
58	Pentachlorophenol	87-86-5	RP230530RSR	99%	1,008.8	µg/mL	+/- 36.7019
59	Phenanthrene	85-01-8	MKCQ8876	99%	1,008.4	µg/mL	+/- 36.6877
60	Anthracene	120-12-7	MKCR0570	99%	1,009.0	µg/mL	+/- 36.7100
61	Carbazole	86-74-8	14351100	99%	1,000.9	µg/mL	+/- 36.4149
62	Di-n-butylphthalate	84-74-2	MKCN4337	99%	1,007.6	µg/mL	+/- 36.6595
63	Fluoranthene	206-44-0	MKCQ4728	99%	1,009.6	µg/mL	+/- 36.7302
64	Pyrene	129-00-0	BCCG8479	98%	1,007.2	µg/mL	+/- 36.6453
65	Benzyl butyl phthalate	85-68-7	X12I018	99%	1,002.1	µg/mL	+/- 36.4573
66	Bis(2-ethylhexyl)adipate	103-23-1	MKCM1988	99%	1,005.2	µg/mL	+/- 36.5705
67	Benz(a)anthracene	56-55-3	I220012022BAA	99%	1,002.2	µg/mL	+/- 36.4614
68	Chrysene	218-01-9	RP230601	99%	1,008.3	µg/mL	+/- 36.6837
69	Bis(2-ethylhexyl)phthalate	117-81-7	MKCQ3468	99%	1,001.8	µg/mL	+/- 36.4472
70	Di-n-octyl phthalate	117-84-0	14382700	99%	1,006.0	µg/mL	+/- 36.6008
71	Benzo(b)fluoranthene	205-99-2	012013B	99%	1,002.8	µg/mL	+/- 36.4836
72	Benzo(k)fluoranthene	207-08-9	012022K	99%	1,003.0	µg/mL	+/- 36.4917
73	Benzo(a)pyrene	50-32-8	P54915-0703	99%	1,002.3	µg/mL	+/- 36.4674
74	Indeno(1,2,3-cd)pyrene	193-39-5	12-JKL-118-9	97%	1,009.4	µg/mL	+/- 36.7243
75	Dibenz(a,h)anthracene	53-70-3	2-ASA-59-1	99%	1,007.6	µg/mL	+/- 36.6595
76	Benzo(g,h,i)perylene	191-24-2	RP231003RSR	99%	1,002.9	µg/mL	+/- 36.4876

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31850 **Lot No.:** A0203726

Description : 8270 MegaMix®
8270 MegaMix® 500-1000 µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2025 **Storage:** 0°C or colder

Handling: Sonication required. Mix is photosensitive. **Ship:** Ambient

512117 } RC/
↓ 03/18/24
512146 }

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Pyridine	110-86-1	SHBP6240	99%	1,001.6 µg/mL	+/- 36.4412
2	N-Nitrosodimethylamine	62-75-9	230209JLM	99%	1,005.9 µg/mL	+/- 36.5968
3	Phenol	108-95-2	MKCK1120	99%	1,003.3 µg/mL	+/- 36.5038
4	Aniline	62-53-3	X22F726	99%	1,005.8 µg/mL	+/- 36.5928
5	Bis(2-chloroethyl)ether	111-44-4	SHBL6942	99%	1,008.1 µg/mL	+/- 36.6776
6	2-Chlorophenol	95-57-8	STBJ3909	99%	1,001.8 µg/mL	+/- 36.4492
7	1,3-Dichlorobenzene	541-73-1	BCCD5315	99%	1,002.3 µg/mL	+/- 36.4654
8	1,4-Dichlorobenzene	106-46-7	MKBS7929V	99%	1,003.7 µg/mL	+/- 36.5159
9	Benzyl alcohol	100-51-6	SHBK5469	99%	1,008.7 µg/mL	+/- 36.6979
10	1,2-Dichlorobenzene	95-50-1	SHBN3835	99%	1,000.3 µg/mL	+/- 36.3926
11	2-Methylphenol (o-cresol)	95-48-7	SHBN7598	99%	1,003.5 µg/mL	+/- 36.5099
12	2,2'-oxybis(1-chloropropane)	108-60-1	29-MAR-45-5	99%	1,007.3 µg/mL	+/- 36.6493
13	3-Methylphenol (m-cresol)	108-39-4	STBJ0710	99%	504.3 µg/mL	+/- 18.3500
14	4-Methylphenol (p-cresol)	106-44-5	SHBN3411	99%	503.6 µg/mL	+/- 18.3237
15	N-Nitroso-di-n-propylamine	621-64-7	N63MG	99%	1,008.3 µg/mL	+/- 36.6857
16	Hexachloroethane	67-72-1	QTORH	99%	1,007.5 µg/mL	+/- 36.6554
17	Nitrobenzene	98-95-3	10224044	99%	1,008.6 µg/mL	+/- 36.6938

18	Isophorone	78-59-1	MKCC9506	99%	1,005.9	µg/mL	+/- 36.5988
19	2-Nitrophenol	88-75-5	RP230710	99%	1,003.2	µg/mL	+/- 36.4998
20	2,4-Dimethylphenol	105-67-9	XW5GK	99%	1,003.8	µg/mL	+/- 36.5200
21	Bis(2-chloroethoxy)methane	111-91-1	13670200	99%	1,002.1	µg/mL	+/- 36.4573
22	2,4-Dichlorophenol	120-83-2	BCBZ6787	99%	1,003.7	µg/mL	+/- 36.5180
23	1,2,4-Trichlorobenzene	120-82-1	SHBP5900	99%	1,007.6	µg/mL	+/- 36.6574
24	Naphthalene	91-20-3	STBL1057	99%	1,008.3	µg/mL	+/- 36.6837
25	4-Chloroaniline	106-47-8	BCCJ3217	99%	1,001.3	µg/mL	+/- 36.4290
26	Hexachlorobutadiene	87-68-3	RP230823RSR	98%	1,008.3	µg/mL	+/- 36.6829
27	4-Chloro-3-methylphenol	59-50-7	BCCD4461	99%	1,003.1	µg/mL	+/- 36.4937
28	2-Methylnaphthalene	91-57-6	STBK0259	96%	1,001.9	µg/mL	+/- 36.4505
29	1-Methylnaphthalene	90-12-0	5234.00-8	98%	1,000.0	µg/mL	+/- 36.3838
30	Hexachlorocyclopentadiene	77-47-4	099063I14L	98%	1,008.5	µg/mL	+/- 36.6909
31	2,4,6-Trichlorophenol	88-06-2	STBJ5914	99%	1,004.4	µg/mL	+/- 36.5442
32	2,4,5-Trichlorophenol	95-95-4	FHN01	98%	1,001.9	µg/mL	+/- 36.4512
33	2-Chloronaphthalene	91-58-7	RPN7O	99%	1,001.1	µg/mL	+/- 36.4230
34	2-Nitroaniline	88-74-4	RP230531	99%	1,002.9	µg/mL	+/- 36.4876
35	1,4-Dinitrobenzene	100-25-4	RP230816	99%	1,005.7	µg/mL	+/- 36.5887
36	Acenaphthylene	208-96-8	p06V	98%	1,009.5	µg/mL	+/- 36.7265
37	1,3-Dinitrobenzene	99-65-0	1-DXX-24-1	99%	1,004.4	µg/mL	+/- 36.5422
38	Dimethylphthalate	131-11-3	358221L17K	99%	1,005.9	µg/mL	+/- 36.5968
39	2,6-Dinitrotoluene	606-20-2	BCCG1833	99%	1,003.2	µg/mL	+/- 36.4998
40	1,2-Dinitrobenzene	528-29-0	RP230428	99%	1,002.2	µg/mL	+/- 36.4634
41	Acenaphthene	83-32-9	MKCR7169	99%	1,009.3	µg/mL	+/- 36.7221
42	3-Nitroaniline	99-09-2	RP230822RSR	99%	1,003.9	µg/mL	+/- 36.5240
43	2,4-Dinitrophenol	51-28-5	DR230417RSR	99%	1,002.0	µg/mL	+/- 36.4553
44	Dibenzofuran	132-64-9	MKCD9952	99%	1,006.7	µg/mL	+/- 36.6251
45	2,4-Dinitrotoluene	121-14-2	MKAA0690V	99%	1,003.8	µg/mL	+/- 36.5220
46	4-Nitrophenol	100-02-7	RP230627	99%	1,002.3	µg/mL	+/- 36.4674
47	2,3,4,6-Tetrachlorophenol	58-90-2	PR-30126	99%	1,008.7	µg/mL	+/- 36.6979
48	2,3,5,6-Tetrachlorophenol	935-95-5	RP230919	99%	1,006.3	µg/mL	+/- 36.6130
49	Fluorene	86-73-7	10241100	99%	1,008.3	µg/mL	+/- 36.6857
50	4-Chlorophenyl phenyl ether	7005-72-3	MKCT7248	99%	1,003.8	µg/mL	+/- 36.5220
51	Diethylphthalate	84-66-2	MKCD2547	99%	1,008.6	µg/mL	+/- 36.6958
52	4-Nitroaniline	100-01-6	RP230111	99%	1,001.1	µg/mL	+/- 36.4230
53	4,6-Dinitro-2-methylphenol (Dinitro-o-cresol)	534-52-1	230718JLM	99%	1,002.0	µg/mL	+/- 36.4553

54	Diphenylamine	122-39-4	MKCH1042	99%	1,002.3	µg/mL	+/- 36.4674
55	Azobenzene	103-33-3	BCKK0887	99%	1,005.8	µg/mL	+/- 36.5928
56	4-Bromophenyl phenyl ether	101-55-3	STBH6361	99%	1,003.0	µg/mL	+/- 36.4917
57	Hexachlorobenzene	118-74-1	14821700	99%	1,007.5	µg/mL	+/- 36.6554
58	Pentachlorophenol	87-86-5	RP230530RSR	99%	1,008.8	µg/mL	+/- 36.7019
59	Phenanthrene	85-01-8	MKCQ8876	99%	1,008.4	µg/mL	+/- 36.6877
60	Anthracene	120-12-7	MKCR0570	99%	1,009.0	µg/mL	+/- 36.7100
61	Carbazole	86-74-8	14351100	99%	1,000.9	µg/mL	+/- 36.4149
62	Di-n-butylphthalate	84-74-2	MKCN4337	99%	1,007.6	µg/mL	+/- 36.6595
63	Fluoranthene	206-44-0	MKCQ4728	99%	1,009.6	µg/mL	+/- 36.7302
64	Pyrene	129-00-0	BCCG8479	98%	1,007.2	µg/mL	+/- 36.6453
65	Benzyl butyl phthalate	85-68-7	X12I018	99%	1,002.1	µg/mL	+/- 36.4573
66	Bis(2-ethylhexyl)adipate	103-23-1	MKCM1988	99%	1,005.2	µg/mL	+/- 36.5705
67	Benz(a)anthracene	56-55-3	I220012022BAA	99%	1,002.2	µg/mL	+/- 36.4614
68	Chrysene	218-01-9	RP230601	99%	1,008.3	µg/mL	+/- 36.6837
69	Bis(2-ethylhexyl)phthalate	117-81-7	MKCQ3468	99%	1,001.8	µg/mL	+/- 36.4472
70	Di-n-octyl phthalate	117-84-0	14382700	99%	1,006.0	µg/mL	+/- 36.6008
71	Benzo(b)fluoranthene	205-99-2	012013B	99%	1,002.8	µg/mL	+/- 36.4836
72	Benzo(k)fluoranthene	207-08-9	012022K	99%	1,003.0	µg/mL	+/- 36.4917
73	Benzo(a)pyrene	50-32-8	P54915-0703	99%	1,002.3	µg/mL	+/- 36.4674
74	Indeno(1,2,3-cd)pyrene	193-39-5	12-JKL-118-9	97%	1,009.4	µg/mL	+/- 36.7243
75	Dibenz(a,h)anthracene	53-70-3	2-ASA-59-1	99%	1,007.6	µg/mL	+/- 36.6595
76	Benzo(g,h,i)perylene	191-24-2	RP231003RSR	99%	1,002.9	µg/mL	+/- 36.4876

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31850 **Lot No.:** A0203726

Description : 8270 MegaMix®
8270 MegaMix® 500-1000 µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2025 **Storage:** 0°C or colder

Handling: Sonication required. Mix is photosensitive. **Ship:** Ambient

512117 } RC/
↓ 03/18/24
512146 }

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Pyridine	110-86-1	SHBP6240	99%	1,001.6 µg/mL	+/- 36.4412
2	N-Nitrosodimethylamine	62-75-9	230209JLM	99%	1,005.9 µg/mL	+/- 36.5968
3	Phenol	108-95-2	MKCK1120	99%	1,003.3 µg/mL	+/- 36.5038
4	Aniline	62-53-3	X22F726	99%	1,005.8 µg/mL	+/- 36.5928
5	Bis(2-chloroethyl)ether	111-44-4	SHBL6942	99%	1,008.1 µg/mL	+/- 36.6776
6	2-Chlorophenol	95-57-8	STBJ3909	99%	1,001.8 µg/mL	+/- 36.4492
7	1,3-Dichlorobenzene	541-73-1	BCCD5315	99%	1,002.3 µg/mL	+/- 36.4654
8	1,4-Dichlorobenzene	106-46-7	MKBS7929V	99%	1,003.7 µg/mL	+/- 36.5159
9	Benzyl alcohol	100-51-6	SHBK5469	99%	1,008.7 µg/mL	+/- 36.6979
10	1,2-Dichlorobenzene	95-50-1	SHBN3835	99%	1,000.3 µg/mL	+/- 36.3926
11	2-Methylphenol (o-cresol)	95-48-7	SHBN7598	99%	1,003.5 µg/mL	+/- 36.5099
12	2,2'-oxybis(1-chloropropane)	108-60-1	29-MAR-45-5	99%	1,007.3 µg/mL	+/- 36.6493
13	3-Methylphenol (m-cresol)	108-39-4	STBJ0710	99%	504.3 µg/mL	+/- 18.3500
14	4-Methylphenol (p-cresol)	106-44-5	SHBN3411	99%	503.6 µg/mL	+/- 18.3237
15	N-Nitroso-di-n-propylamine	621-64-7	N63MG	99%	1,008.3 µg/mL	+/- 36.6857
16	Hexachloroethane	67-72-1	QTORH	99%	1,007.5 µg/mL	+/- 36.6554
17	Nitrobenzene	98-95-3	10224044	99%	1,008.6 µg/mL	+/- 36.6938

18	Isophorone	78-59-1	MKCC9506	99%	1,005.9	µg/mL	+/- 36.5988
19	2-Nitrophenol	88-75-5	RP230710	99%	1,003.2	µg/mL	+/- 36.4998
20	2,4-Dimethylphenol	105-67-9	XW5GK	99%	1,003.8	µg/mL	+/- 36.5200
21	Bis(2-chloroethoxy)methane	111-91-1	13670200	99%	1,002.1	µg/mL	+/- 36.4573
22	2,4-Dichlorophenol	120-83-2	BCBZ6787	99%	1,003.7	µg/mL	+/- 36.5180
23	1,2,4-Trichlorobenzene	120-82-1	SHBP5900	99%	1,007.6	µg/mL	+/- 36.6574
24	Naphthalene	91-20-3	STBL1057	99%	1,008.3	µg/mL	+/- 36.6837
25	4-Chloroaniline	106-47-8	BCCJ3217	99%	1,001.3	µg/mL	+/- 36.4290
26	Hexachlorobutadiene	87-68-3	RP230823RSR	98%	1,008.3	µg/mL	+/- 36.6829
27	4-Chloro-3-methylphenol	59-50-7	BCCD4461	99%	1,003.1	µg/mL	+/- 36.4937
28	2-Methylnaphthalene	91-57-6	STBK0259	96%	1,001.9	µg/mL	+/- 36.4505
29	1-Methylnaphthalene	90-12-0	5234.00-8	98%	1,000.0	µg/mL	+/- 36.3838
30	Hexachlorocyclopentadiene	77-47-4	099063I14L	98%	1,008.5	µg/mL	+/- 36.6909
31	2,4,6-Trichlorophenol	88-06-2	STBJ5914	99%	1,004.4	µg/mL	+/- 36.5442
32	2,4,5-Trichlorophenol	95-95-4	FHN01	98%	1,001.9	µg/mL	+/- 36.4512
33	2-Chloronaphthalene	91-58-7	RPN7O	99%	1,001.1	µg/mL	+/- 36.4230
34	2-Nitroaniline	88-74-4	RP230531	99%	1,002.9	µg/mL	+/- 36.4876
35	1,4-Dinitrobenzene	100-25-4	RP230816	99%	1,005.7	µg/mL	+/- 36.5887
36	Acenaphthylene	208-96-8	p06V	98%	1,009.5	µg/mL	+/- 36.7265
37	1,3-Dinitrobenzene	99-65-0	1-DXX-24-1	99%	1,004.4	µg/mL	+/- 36.5422
38	Dimethylphthalate	131-11-3	358221L17K	99%	1,005.9	µg/mL	+/- 36.5968
39	2,6-Dinitrotoluene	606-20-2	BCCG1833	99%	1,003.2	µg/mL	+/- 36.4998
40	1,2-Dinitrobenzene	528-29-0	RP230428	99%	1,002.2	µg/mL	+/- 36.4634
41	Acenaphthene	83-32-9	MKCR7169	99%	1,009.3	µg/mL	+/- 36.7221
42	3-Nitroaniline	99-09-2	RP230822RSR	99%	1,003.9	µg/mL	+/- 36.5240
43	2,4-Dinitrophenol	51-28-5	DR230417RSR	99%	1,002.0	µg/mL	+/- 36.4553
44	Dibenzofuran	132-64-9	MKCD9952	99%	1,006.7	µg/mL	+/- 36.6251
45	2,4-Dinitrotoluene	121-14-2	MKAA0690V	99%	1,003.8	µg/mL	+/- 36.5220
46	4-Nitrophenol	100-02-7	RP230627	99%	1,002.3	µg/mL	+/- 36.4674
47	2,3,4,6-Tetrachlorophenol	58-90-2	PR-30126	99%	1,008.7	µg/mL	+/- 36.6979
48	2,3,5,6-Tetrachlorophenol	935-95-5	RP230919	99%	1,006.3	µg/mL	+/- 36.6130
49	Fluorene	86-73-7	10241100	99%	1,008.3	µg/mL	+/- 36.6857
50	4-Chlorophenyl phenyl ether	7005-72-3	MKCT7248	99%	1,003.8	µg/mL	+/- 36.5220
51	Diethylphthalate	84-66-2	MKCD2547	99%	1,008.6	µg/mL	+/- 36.6958
52	4-Nitroaniline	100-01-6	RP230111	99%	1,001.1	µg/mL	+/- 36.4230
53	4,6-Dinitro-2-methylphenol (Dinitro-o-cresol)	534-52-1	230718JLM	99%	1,002.0	µg/mL	+/- 36.4553

54	Diphenylamine	122-39-4	MKCH1042	99%	1,002.3	µg/mL	+/- 36.4674
55	Azobenzene	103-33-3	BCKK0887	99%	1,005.8	µg/mL	+/- 36.5928
56	4-Bromophenyl phenyl ether	101-55-3	STBH6361	99%	1,003.0	µg/mL	+/- 36.4917
57	Hexachlorobenzene	118-74-1	14821700	99%	1,007.5	µg/mL	+/- 36.6554
58	Pentachlorophenol	87-86-5	RP230530RSR	99%	1,008.8	µg/mL	+/- 36.7019
59	Phenanthrene	85-01-8	MKCQ8876	99%	1,008.4	µg/mL	+/- 36.6877
60	Anthracene	120-12-7	MKCR0570	99%	1,009.0	µg/mL	+/- 36.7100
61	Carbazole	86-74-8	14351100	99%	1,000.9	µg/mL	+/- 36.4149
62	Di-n-butylphthalate	84-74-2	MKCN4337	99%	1,007.6	µg/mL	+/- 36.6595
63	Fluoranthene	206-44-0	MKCQ4728	99%	1,009.6	µg/mL	+/- 36.7302
64	Pyrene	129-00-0	BCCG8479	98%	1,007.2	µg/mL	+/- 36.6453
65	Benzyl butyl phthalate	85-68-7	X12I018	99%	1,002.1	µg/mL	+/- 36.4573
66	Bis(2-ethylhexyl)adipate	103-23-1	MKCM1988	99%	1,005.2	µg/mL	+/- 36.5705
67	Benz(a)anthracene	56-55-3	I220012022BAA	99%	1,002.2	µg/mL	+/- 36.4614
68	Chrysene	218-01-9	RP230601	99%	1,008.3	µg/mL	+/- 36.6837
69	Bis(2-ethylhexyl)phthalate	117-81-7	MKCQ3468	99%	1,001.8	µg/mL	+/- 36.4472
70	Di-n-octyl phthalate	117-84-0	14382700	99%	1,006.0	µg/mL	+/- 36.6008
71	Benzo(b)fluoranthene	205-99-2	012013B	99%	1,002.8	µg/mL	+/- 36.4836
72	Benzo(k)fluoranthene	207-08-9	012022K	99%	1,003.0	µg/mL	+/- 36.4917
73	Benzo(a)pyrene	50-32-8	P54915-0703	99%	1,002.3	µg/mL	+/- 36.4674
74	Indeno(1,2,3-cd)pyrene	193-39-5	12-JKL-118-9	97%	1,009.4	µg/mL	+/- 36.7243
75	Dibenz(a,h)anthracene	53-70-3	2-ASA-59-1	99%	1,007.6	µg/mL	+/- 36.6595
76	Benzo(g,h,i)perylene	191-24-2	RP231003RSR	99%	1,002.9	µg/mL	+/- 36.4876

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31850 **Lot No.:** A0203726

Description : 8270 MegaMix®
8270 MegaMix® 500-1000 µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2025 **Storage:** 0°C or colder

Handling: Sonication required. Mix is photosensitive. **Ship:** Ambient

512117 } RC/
↓ 03/18/24
512146 }

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Pyridine	110-86-1	SHBP6240	99%	1,001.6 µg/mL	+/- 36.4412
2	N-Nitrosodimethylamine	62-75-9	230209JLM	99%	1,005.9 µg/mL	+/- 36.5968
3	Phenol	108-95-2	MKCK1120	99%	1,003.3 µg/mL	+/- 36.5038
4	Aniline	62-53-3	X22F726	99%	1,005.8 µg/mL	+/- 36.5928
5	Bis(2-chloroethyl)ether	111-44-4	SHBL6942	99%	1,008.1 µg/mL	+/- 36.6776
6	2-Chlorophenol	95-57-8	STBJ3909	99%	1,001.8 µg/mL	+/- 36.4492
7	1,3-Dichlorobenzene	541-73-1	BCCD5315	99%	1,002.3 µg/mL	+/- 36.4654
8	1,4-Dichlorobenzene	106-46-7	MKBS7929V	99%	1,003.7 µg/mL	+/- 36.5159
9	Benzyl alcohol	100-51-6	SHBK5469	99%	1,008.7 µg/mL	+/- 36.6979
10	1,2-Dichlorobenzene	95-50-1	SHBN3835	99%	1,000.3 µg/mL	+/- 36.3926
11	2-Methylphenol (o-cresol)	95-48-7	SHBN7598	99%	1,003.5 µg/mL	+/- 36.5099
12	2,2'-oxybis(1-chloropropane)	108-60-1	29-MAR-45-5	99%	1,007.3 µg/mL	+/- 36.6493
13	3-Methylphenol (m-cresol)	108-39-4	STBJ0710	99%	504.3 µg/mL	+/- 18.3500
14	4-Methylphenol (p-cresol)	106-44-5	SHBN3411	99%	503.6 µg/mL	+/- 18.3237
15	N-Nitroso-di-n-propylamine	621-64-7	N63MG	99%	1,008.3 µg/mL	+/- 36.6857
16	Hexachloroethane	67-72-1	QTORH	99%	1,007.5 µg/mL	+/- 36.6554
17	Nitrobenzene	98-95-3	10224044	99%	1,008.6 µg/mL	+/- 36.6938

18	Isophorone	78-59-1	MKCC9506	99%	1,005.9	µg/mL	+/- 36.5988
19	2-Nitrophenol	88-75-5	RP230710	99%	1,003.2	µg/mL	+/- 36.4998
20	2,4-Dimethylphenol	105-67-9	XW5GK	99%	1,003.8	µg/mL	+/- 36.5200
21	Bis(2-chloroethoxy)methane	111-91-1	13670200	99%	1,002.1	µg/mL	+/- 36.4573
22	2,4-Dichlorophenol	120-83-2	BCBZ6787	99%	1,003.7	µg/mL	+/- 36.5180
23	1,2,4-Trichlorobenzene	120-82-1	SHBP5900	99%	1,007.6	µg/mL	+/- 36.6574
24	Naphthalene	91-20-3	STBL1057	99%	1,008.3	µg/mL	+/- 36.6837
25	4-Chloroaniline	106-47-8	BCCJ3217	99%	1,001.3	µg/mL	+/- 36.4290
26	Hexachlorobutadiene	87-68-3	RP230823RSR	98%	1,008.3	µg/mL	+/- 36.6829
27	4-Chloro-3-methylphenol	59-50-7	BCCD4461	99%	1,003.1	µg/mL	+/- 36.4937
28	2-Methylnaphthalene	91-57-6	STBK0259	96%	1,001.9	µg/mL	+/- 36.4505
29	1-Methylnaphthalene	90-12-0	5234.00-8	98%	1,000.0	µg/mL	+/- 36.3838
30	Hexachlorocyclopentadiene	77-47-4	099063I14L	98%	1,008.5	µg/mL	+/- 36.6909
31	2,4,6-Trichlorophenol	88-06-2	STBJ5914	99%	1,004.4	µg/mL	+/- 36.5442
32	2,4,5-Trichlorophenol	95-95-4	FHN01	98%	1,001.9	µg/mL	+/- 36.4512
33	2-Chloronaphthalene	91-58-7	RPN7O	99%	1,001.1	µg/mL	+/- 36.4230
34	2-Nitroaniline	88-74-4	RP230531	99%	1,002.9	µg/mL	+/- 36.4876
35	1,4-Dinitrobenzene	100-25-4	RP230816	99%	1,005.7	µg/mL	+/- 36.5887
36	Acenaphthylene	208-96-8	p06V	98%	1,009.5	µg/mL	+/- 36.7265
37	1,3-Dinitrobenzene	99-65-0	1-DXX-24-1	99%	1,004.4	µg/mL	+/- 36.5422
38	Dimethylphthalate	131-11-3	358221L17K	99%	1,005.9	µg/mL	+/- 36.5968
39	2,6-Dinitrotoluene	606-20-2	BCCG1833	99%	1,003.2	µg/mL	+/- 36.4998
40	1,2-Dinitrobenzene	528-29-0	RP230428	99%	1,002.2	µg/mL	+/- 36.4634
41	Acenaphthene	83-32-9	MKCR7169	99%	1,009.3	µg/mL	+/- 36.7221
42	3-Nitroaniline	99-09-2	RP230822RSR	99%	1,003.9	µg/mL	+/- 36.5240
43	2,4-Dinitrophenol	51-28-5	DR230417RSR	99%	1,002.0	µg/mL	+/- 36.4553
44	Dibenzofuran	132-64-9	MKCD9952	99%	1,006.7	µg/mL	+/- 36.6251
45	2,4-Dinitrotoluene	121-14-2	MKAA0690V	99%	1,003.8	µg/mL	+/- 36.5220
46	4-Nitrophenol	100-02-7	RP230627	99%	1,002.3	µg/mL	+/- 36.4674
47	2,3,4,6-Tetrachlorophenol	58-90-2	PR-30126	99%	1,008.7	µg/mL	+/- 36.6979
48	2,3,5,6-Tetrachlorophenol	935-95-5	RP230919	99%	1,006.3	µg/mL	+/- 36.6130
49	Fluorene	86-73-7	10241100	99%	1,008.3	µg/mL	+/- 36.6857
50	4-Chlorophenyl phenyl ether	7005-72-3	MKCT7248	99%	1,003.8	µg/mL	+/- 36.5220
51	Diethylphthalate	84-66-2	MKCD2547	99%	1,008.6	µg/mL	+/- 36.6958
52	4-Nitroaniline	100-01-6	RP230111	99%	1,001.1	µg/mL	+/- 36.4230
53	4,6-Dinitro-2-methylphenol (Dinitro-o-cresol)	534-52-1	230718JLM	99%	1,002.0	µg/mL	+/- 36.4553

54	Diphenylamine	122-39-4	MKCH1042	99%	1,002.3	µg/mL	+/- 36.4674
55	Azobenzene	103-33-3	BCKK0887	99%	1,005.8	µg/mL	+/- 36.5928
56	4-Bromophenyl phenyl ether	101-55-3	STBH6361	99%	1,003.0	µg/mL	+/- 36.4917
57	Hexachlorobenzene	118-74-1	14821700	99%	1,007.5	µg/mL	+/- 36.6554
58	Pentachlorophenol	87-86-5	RP230530RSR	99%	1,008.8	µg/mL	+/- 36.7019
59	Phenanthrene	85-01-8	MKCQ8876	99%	1,008.4	µg/mL	+/- 36.6877
60	Anthracene	120-12-7	MKCR0570	99%	1,009.0	µg/mL	+/- 36.7100
61	Carbazole	86-74-8	14351100	99%	1,000.9	µg/mL	+/- 36.4149
62	Di-n-butylphthalate	84-74-2	MKCN4337	99%	1,007.6	µg/mL	+/- 36.6595
63	Fluoranthene	206-44-0	MKCQ4728	99%	1,009.6	µg/mL	+/- 36.7302
64	Pyrene	129-00-0	BCCG8479	98%	1,007.2	µg/mL	+/- 36.6453
65	Benzyl butyl phthalate	85-68-7	X12I018	99%	1,002.1	µg/mL	+/- 36.4573
66	Bis(2-ethylhexyl)adipate	103-23-1	MKCM1988	99%	1,005.2	µg/mL	+/- 36.5705
67	Benz(a)anthracene	56-55-3	I220012022BAA	99%	1,002.2	µg/mL	+/- 36.4614
68	Chrysene	218-01-9	RP230601	99%	1,008.3	µg/mL	+/- 36.6837
69	Bis(2-ethylhexyl)phthalate	117-81-7	MKCQ3468	99%	1,001.8	µg/mL	+/- 36.4472
70	Di-n-octyl phthalate	117-84-0	14382700	99%	1,006.0	µg/mL	+/- 36.6008
71	Benzo(b)fluoranthene	205-99-2	012013B	99%	1,002.8	µg/mL	+/- 36.4836
72	Benzo(k)fluoranthene	207-08-9	012022K	99%	1,003.0	µg/mL	+/- 36.4917
73	Benzo(a)pyrene	50-32-8	P54915-0703	99%	1,002.3	µg/mL	+/- 36.4674
74	Indeno(1,2,3-cd)pyrene	193-39-5	12-JKL-118-9	97%	1,009.4	µg/mL	+/- 36.7243
75	Dibenz(a,h)anthracene	53-70-3	2-ASA-59-1	99%	1,007.6	µg/mL	+/- 36.6595
76	Benzo(g,h,i)perylene	191-24-2	RP231003RSR	99%	1,002.9	µg/mL	+/- 36.4876

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31850 **Lot No.:** A0203726

Description : 8270 MegaMix®
8270 MegaMix® 500-1000 µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2025 **Storage:** 0°C or colder

Handling: Sonication required. Mix is photosensitive. **Ship:** Ambient

512117 } RC/
↓ 03/18/24
512146 }

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Pyridine	110-86-1	SHBP6240	99%	1,001.6 µg/mL	+/- 36.4412
2	N-Nitrosodimethylamine	62-75-9	230209JLM	99%	1,005.9 µg/mL	+/- 36.5968
3	Phenol	108-95-2	MKCK1120	99%	1,003.3 µg/mL	+/- 36.5038
4	Aniline	62-53-3	X22F726	99%	1,005.8 µg/mL	+/- 36.5928
5	Bis(2-chloroethyl)ether	111-44-4	SHBL6942	99%	1,008.1 µg/mL	+/- 36.6776
6	2-Chlorophenol	95-57-8	STBJ3909	99%	1,001.8 µg/mL	+/- 36.4492
7	1,3-Dichlorobenzene	541-73-1	BCCD5315	99%	1,002.3 µg/mL	+/- 36.4654
8	1,4-Dichlorobenzene	106-46-7	MKBS7929V	99%	1,003.7 µg/mL	+/- 36.5159
9	Benzyl alcohol	100-51-6	SHBK5469	99%	1,008.7 µg/mL	+/- 36.6979
10	1,2-Dichlorobenzene	95-50-1	SHBN3835	99%	1,000.3 µg/mL	+/- 36.3926
11	2-Methylphenol (o-cresol)	95-48-7	SHBN7598	99%	1,003.5 µg/mL	+/- 36.5099
12	2,2'-oxybis(1-chloropropane)	108-60-1	29-MAR-45-5	99%	1,007.3 µg/mL	+/- 36.6493
13	3-Methylphenol (m-cresol)	108-39-4	STBJ0710	99%	504.3 µg/mL	+/- 18.3500
14	4-Methylphenol (p-cresol)	106-44-5	SHBN3411	99%	503.6 µg/mL	+/- 18.3237
15	N-Nitroso-di-n-propylamine	621-64-7	N63MG	99%	1,008.3 µg/mL	+/- 36.6857
16	Hexachloroethane	67-72-1	QTORH	99%	1,007.5 µg/mL	+/- 36.6554
17	Nitrobenzene	98-95-3	10224044	99%	1,008.6 µg/mL	+/- 36.6938

18	Isophorone	78-59-1	MKCC9506	99%	1,005.9	µg/mL	+/- 36.5988
19	2-Nitrophenol	88-75-5	RP230710	99%	1,003.2	µg/mL	+/- 36.4998
20	2,4-Dimethylphenol	105-67-9	XW5GK	99%	1,003.8	µg/mL	+/- 36.5200
21	Bis(2-chloroethoxy)methane	111-91-1	13670200	99%	1,002.1	µg/mL	+/- 36.4573
22	2,4-Dichlorophenol	120-83-2	BCBZ6787	99%	1,003.7	µg/mL	+/- 36.5180
23	1,2,4-Trichlorobenzene	120-82-1	SHBP5900	99%	1,007.6	µg/mL	+/- 36.6574
24	Naphthalene	91-20-3	STBL1057	99%	1,008.3	µg/mL	+/- 36.6837
25	4-Chloroaniline	106-47-8	BCCJ3217	99%	1,001.3	µg/mL	+/- 36.4290
26	Hexachlorobutadiene	87-68-3	RP230823RSR	98%	1,008.3	µg/mL	+/- 36.6829
27	4-Chloro-3-methylphenol	59-50-7	BCCD4461	99%	1,003.1	µg/mL	+/- 36.4937
28	2-Methylnaphthalene	91-57-6	STBK0259	96%	1,001.9	µg/mL	+/- 36.4505
29	1-Methylnaphthalene	90-12-0	5234.00-8	98%	1,000.0	µg/mL	+/- 36.3838
30	Hexachlorocyclopentadiene	77-47-4	099063I14L	98%	1,008.5	µg/mL	+/- 36.6909
31	2,4,6-Trichlorophenol	88-06-2	STBJ5914	99%	1,004.4	µg/mL	+/- 36.5442
32	2,4,5-Trichlorophenol	95-95-4	FHN01	98%	1,001.9	µg/mL	+/- 36.4512
33	2-Chloronaphthalene	91-58-7	RPN7O	99%	1,001.1	µg/mL	+/- 36.4230
34	2-Nitroaniline	88-74-4	RP230531	99%	1,002.9	µg/mL	+/- 36.4876
35	1,4-Dinitrobenzene	100-25-4	RP230816	99%	1,005.7	µg/mL	+/- 36.5887
36	Acenaphthylene	208-96-8	p06V	98%	1,009.5	µg/mL	+/- 36.7265
37	1,3-Dinitrobenzene	99-65-0	1-DXX-24-1	99%	1,004.4	µg/mL	+/- 36.5422
38	Dimethylphthalate	131-11-3	358221L17K	99%	1,005.9	µg/mL	+/- 36.5968
39	2,6-Dinitrotoluene	606-20-2	BCCG1833	99%	1,003.2	µg/mL	+/- 36.4998
40	1,2-Dinitrobenzene	528-29-0	RP230428	99%	1,002.2	µg/mL	+/- 36.4634
41	Acenaphthene	83-32-9	MKCR7169	99%	1,009.3	µg/mL	+/- 36.7221
42	3-Nitroaniline	99-09-2	RP230822RSR	99%	1,003.9	µg/mL	+/- 36.5240
43	2,4-Dinitrophenol	51-28-5	DR230417RSR	99%	1,002.0	µg/mL	+/- 36.4553
44	Dibenzofuran	132-64-9	MKCD9952	99%	1,006.7	µg/mL	+/- 36.6251
45	2,4-Dinitrotoluene	121-14-2	MKAA0690V	99%	1,003.8	µg/mL	+/- 36.5220
46	4-Nitrophenol	100-02-7	RP230627	99%	1,002.3	µg/mL	+/- 36.4674
47	2,3,4,6-Tetrachlorophenol	58-90-2	PR-30126	99%	1,008.7	µg/mL	+/- 36.6979
48	2,3,5,6-Tetrachlorophenol	935-95-5	RP230919	99%	1,006.3	µg/mL	+/- 36.6130
49	Fluorene	86-73-7	10241100	99%	1,008.3	µg/mL	+/- 36.6857
50	4-Chlorophenyl phenyl ether	7005-72-3	MKCT7248	99%	1,003.8	µg/mL	+/- 36.5220
51	Diethylphthalate	84-66-2	MKCD2547	99%	1,008.6	µg/mL	+/- 36.6958
52	4-Nitroaniline	100-01-6	RP230111	99%	1,001.1	µg/mL	+/- 36.4230
53	4,6-Dinitro-2-methylphenol (Dinitro-o-cresol)	534-52-1	230718JLM	99%	1,002.0	µg/mL	+/- 36.4553

54	Diphenylamine	122-39-4	MKCH1042	99%	1,002.3	µg/mL	+/- 36.4674
55	Azobenzene	103-33-3	BCKK0887	99%	1,005.8	µg/mL	+/- 36.5928
56	4-Bromophenyl phenyl ether	101-55-3	STBH6361	99%	1,003.0	µg/mL	+/- 36.4917
57	Hexachlorobenzene	118-74-1	14821700	99%	1,007.5	µg/mL	+/- 36.6554
58	Pentachlorophenol	87-86-5	RP230530RSR	99%	1,008.8	µg/mL	+/- 36.7019
59	Phenanthrene	85-01-8	MKCQ8876	99%	1,008.4	µg/mL	+/- 36.6877
60	Anthracene	120-12-7	MKCR0570	99%	1,009.0	µg/mL	+/- 36.7100
61	Carbazole	86-74-8	14351100	99%	1,000.9	µg/mL	+/- 36.4149
62	Di-n-butylphthalate	84-74-2	MKCN4337	99%	1,007.6	µg/mL	+/- 36.6595
63	Fluoranthene	206-44-0	MKCQ4728	99%	1,009.6	µg/mL	+/- 36.7302
64	Pyrene	129-00-0	BCCG8479	98%	1,007.2	µg/mL	+/- 36.6453
65	Benzyl butyl phthalate	85-68-7	X12I018	99%	1,002.1	µg/mL	+/- 36.4573
66	Bis(2-ethylhexyl)adipate	103-23-1	MKCM1988	99%	1,005.2	µg/mL	+/- 36.5705
67	Benz(a)anthracene	56-55-3	I220012022BAA	99%	1,002.2	µg/mL	+/- 36.4614
68	Chrysene	218-01-9	RP230601	99%	1,008.3	µg/mL	+/- 36.6837
69	Bis(2-ethylhexyl)phthalate	117-81-7	MKCQ3468	99%	1,001.8	µg/mL	+/- 36.4472
70	Di-n-octyl phthalate	117-84-0	14382700	99%	1,006.0	µg/mL	+/- 36.6008
71	Benzo(b)fluoranthene	205-99-2	012013B	99%	1,002.8	µg/mL	+/- 36.4836
72	Benzo(k)fluoranthene	207-08-9	012022K	99%	1,003.0	µg/mL	+/- 36.4917
73	Benzo(a)pyrene	50-32-8	P54915-0703	99%	1,002.3	µg/mL	+/- 36.4674
74	Indeno(1,2,3-cd)pyrene	193-39-5	12-JKL-118-9	97%	1,009.4	µg/mL	+/- 36.7243
75	Dibenz(a,h)anthracene	53-70-3	2-ASA-59-1	99%	1,007.6	µg/mL	+/- 36.6595
76	Benzo(g,h,i)perylene	191-24-2	RP231003RSR	99%	1,002.9	µg/mL	+/- 36.4876

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31850 **Lot No.:** A0203726

Description : 8270 MegaMix®
8270 MegaMix® 500-1000 µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2025 **Storage:** 0°C or colder

Handling: Sonication required. Mix is photosensitive. **Ship:** Ambient

512117 } RC/
↓ 03/18/24
512146 }

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Pyridine	110-86-1	SHBP6240	99%	1,001.6 µg/mL	+/- 36.4412
2	N-Nitrosodimethylamine	62-75-9	230209JLM	99%	1,005.9 µg/mL	+/- 36.5968
3	Phenol	108-95-2	MKCK1120	99%	1,003.3 µg/mL	+/- 36.5038
4	Aniline	62-53-3	X22F726	99%	1,005.8 µg/mL	+/- 36.5928
5	Bis(2-chloroethyl)ether	111-44-4	SHBL6942	99%	1,008.1 µg/mL	+/- 36.6776
6	2-Chlorophenol	95-57-8	STBJ3909	99%	1,001.8 µg/mL	+/- 36.4492
7	1,3-Dichlorobenzene	541-73-1	BCCD5315	99%	1,002.3 µg/mL	+/- 36.4654
8	1,4-Dichlorobenzene	106-46-7	MKBS7929V	99%	1,003.7 µg/mL	+/- 36.5159
9	Benzyl alcohol	100-51-6	SHBK5469	99%	1,008.7 µg/mL	+/- 36.6979
10	1,2-Dichlorobenzene	95-50-1	SHBN3835	99%	1,000.3 µg/mL	+/- 36.3926
11	2-Methylphenol (o-cresol)	95-48-7	SHBN7598	99%	1,003.5 µg/mL	+/- 36.5099
12	2,2'-oxybis(1-chloropropane)	108-60-1	29-MAR-45-5	99%	1,007.3 µg/mL	+/- 36.6493
13	3-Methylphenol (m-cresol)	108-39-4	STBJ0710	99%	504.3 µg/mL	+/- 18.3500
14	4-Methylphenol (p-cresol)	106-44-5	SHBN3411	99%	503.6 µg/mL	+/- 18.3237
15	N-Nitroso-di-n-propylamine	621-64-7	N63MG	99%	1,008.3 µg/mL	+/- 36.6857
16	Hexachloroethane	67-72-1	QTORH	99%	1,007.5 µg/mL	+/- 36.6554
17	Nitrobenzene	98-95-3	10224044	99%	1,008.6 µg/mL	+/- 36.6938

18	Isophorone	78-59-1	MKCC9506	99%	1,005.9	µg/mL	+/- 36.5988
19	2-Nitrophenol	88-75-5	RP230710	99%	1,003.2	µg/mL	+/- 36.4998
20	2,4-Dimethylphenol	105-67-9	XW5GK	99%	1,003.8	µg/mL	+/- 36.5200
21	Bis(2-chloroethoxy)methane	111-91-1	13670200	99%	1,002.1	µg/mL	+/- 36.4573
22	2,4-Dichlorophenol	120-83-2	BCBZ6787	99%	1,003.7	µg/mL	+/- 36.5180
23	1,2,4-Trichlorobenzene	120-82-1	SHBP5900	99%	1,007.6	µg/mL	+/- 36.6574
24	Naphthalene	91-20-3	STBL1057	99%	1,008.3	µg/mL	+/- 36.6837
25	4-Chloroaniline	106-47-8	BCCJ3217	99%	1,001.3	µg/mL	+/- 36.4290
26	Hexachlorobutadiene	87-68-3	RP230823RSR	98%	1,008.3	µg/mL	+/- 36.6829
27	4-Chloro-3-methylphenol	59-50-7	BCCD4461	99%	1,003.1	µg/mL	+/- 36.4937
28	2-Methylnaphthalene	91-57-6	STBK0259	96%	1,001.9	µg/mL	+/- 36.4505
29	1-Methylnaphthalene	90-12-0	5234.00-8	98%	1,000.0	µg/mL	+/- 36.3838
30	Hexachlorocyclopentadiene	77-47-4	099063I14L	98%	1,008.5	µg/mL	+/- 36.6909
31	2,4,6-Trichlorophenol	88-06-2	STBJ5914	99%	1,004.4	µg/mL	+/- 36.5442
32	2,4,5-Trichlorophenol	95-95-4	FHN01	98%	1,001.9	µg/mL	+/- 36.4512
33	2-Chloronaphthalene	91-58-7	RPN7O	99%	1,001.1	µg/mL	+/- 36.4230
34	2-Nitroaniline	88-74-4	RP230531	99%	1,002.9	µg/mL	+/- 36.4876
35	1,4-Dinitrobenzene	100-25-4	RP230816	99%	1,005.7	µg/mL	+/- 36.5887
36	Acenaphthylene	208-96-8	p06V	98%	1,009.5	µg/mL	+/- 36.7265
37	1,3-Dinitrobenzene	99-65-0	1-DXX-24-1	99%	1,004.4	µg/mL	+/- 36.5422
38	Dimethylphthalate	131-11-3	358221L17K	99%	1,005.9	µg/mL	+/- 36.5968
39	2,6-Dinitrotoluene	606-20-2	BCCG1833	99%	1,003.2	µg/mL	+/- 36.4998
40	1,2-Dinitrobenzene	528-29-0	RP230428	99%	1,002.2	µg/mL	+/- 36.4634
41	Acenaphthene	83-32-9	MKCR7169	99%	1,009.3	µg/mL	+/- 36.7221
42	3-Nitroaniline	99-09-2	RP230822RSR	99%	1,003.9	µg/mL	+/- 36.5240
43	2,4-Dinitrophenol	51-28-5	DR230417RSR	99%	1,002.0	µg/mL	+/- 36.4553
44	Dibenzofuran	132-64-9	MKCD9952	99%	1,006.7	µg/mL	+/- 36.6251
45	2,4-Dinitrotoluene	121-14-2	MKAA0690V	99%	1,003.8	µg/mL	+/- 36.5220
46	4-Nitrophenol	100-02-7	RP230627	99%	1,002.3	µg/mL	+/- 36.4674
47	2,3,4,6-Tetrachlorophenol	58-90-2	PR-30126	99%	1,008.7	µg/mL	+/- 36.6979
48	2,3,5,6-Tetrachlorophenol	935-95-5	RP230919	99%	1,006.3	µg/mL	+/- 36.6130
49	Fluorene	86-73-7	10241100	99%	1,008.3	µg/mL	+/- 36.6857
50	4-Chlorophenyl phenyl ether	7005-72-3	MKCT7248	99%	1,003.8	µg/mL	+/- 36.5220
51	Diethylphthalate	84-66-2	MKCD2547	99%	1,008.6	µg/mL	+/- 36.6958
52	4-Nitroaniline	100-01-6	RP230111	99%	1,001.1	µg/mL	+/- 36.4230
53	4,6-Dinitro-2-methylphenol (Dinitro-o-cresol)	534-52-1	230718JLM	99%	1,002.0	µg/mL	+/- 36.4553

54	Diphenylamine	122-39-4	MKCH1042	99%	1,002.3	µg/mL	+/- 36.4674
55	Azobenzene	103-33-3	BCKK0887	99%	1,005.8	µg/mL	+/- 36.5928
56	4-Bromophenyl phenyl ether	101-55-3	STBH6361	99%	1,003.0	µg/mL	+/- 36.4917
57	Hexachlorobenzene	118-74-1	14821700	99%	1,007.5	µg/mL	+/- 36.6554
58	Pentachlorophenol	87-86-5	RP230530RSR	99%	1,008.8	µg/mL	+/- 36.7019
59	Phenanthrene	85-01-8	MKCQ8876	99%	1,008.4	µg/mL	+/- 36.6877
60	Anthracene	120-12-7	MKCR0570	99%	1,009.0	µg/mL	+/- 36.7100
61	Carbazole	86-74-8	14351100	99%	1,000.9	µg/mL	+/- 36.4149
62	Di-n-butylphthalate	84-74-2	MKCN4337	99%	1,007.6	µg/mL	+/- 36.6595
63	Fluoranthene	206-44-0	MKCQ4728	99%	1,009.6	µg/mL	+/- 36.7302
64	Pyrene	129-00-0	BCCG8479	98%	1,007.2	µg/mL	+/- 36.6453
65	Benzyl butyl phthalate	85-68-7	X12I018	99%	1,002.1	µg/mL	+/- 36.4573
66	Bis(2-ethylhexyl)adipate	103-23-1	MKCM1988	99%	1,005.2	µg/mL	+/- 36.5705
67	Benz(a)anthracene	56-55-3	I220012022BAA	99%	1,002.2	µg/mL	+/- 36.4614
68	Chrysene	218-01-9	RP230601	99%	1,008.3	µg/mL	+/- 36.6837
69	Bis(2-ethylhexyl)phthalate	117-81-7	MKCQ3468	99%	1,001.8	µg/mL	+/- 36.4472
70	Di-n-octyl phthalate	117-84-0	14382700	99%	1,006.0	µg/mL	+/- 36.6008
71	Benzo(b)fluoranthene	205-99-2	012013B	99%	1,002.8	µg/mL	+/- 36.4836
72	Benzo(k)fluoranthene	207-08-9	012022K	99%	1,003.0	µg/mL	+/- 36.4917
73	Benzo(a)pyrene	50-32-8	P54915-0703	99%	1,002.3	µg/mL	+/- 36.4674
74	Indeno(1,2,3-cd)pyrene	193-39-5	12-JKL-118-9	97%	1,009.4	µg/mL	+/- 36.7243
75	Dibenz(a,h)anthracene	53-70-3	2-ASA-59-1	99%	1,007.6	µg/mL	+/- 36.6595
76	Benzo(g,h,i)perylene	191-24-2	RP231003RSR	99%	1,002.9	µg/mL	+/- 36.4876

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31850 **Lot No.:** A0203726

Description : 8270 MegaMix®
8270 MegaMix® 500-1000 µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2025 **Storage:** 0°C or colder

Handling: Sonication required. Mix is photosensitive. **Ship:** Ambient

512117 } RC/
↓ 03/18/24
512146 }

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Pyridine	110-86-1	SHBP6240	99%	1,001.6 µg/mL	+/- 36.4412
2	N-Nitrosodimethylamine	62-75-9	230209JLM	99%	1,005.9 µg/mL	+/- 36.5968
3	Phenol	108-95-2	MKCK1120	99%	1,003.3 µg/mL	+/- 36.5038
4	Aniline	62-53-3	X22F726	99%	1,005.8 µg/mL	+/- 36.5928
5	Bis(2-chloroethyl)ether	111-44-4	SHBL6942	99%	1,008.1 µg/mL	+/- 36.6776
6	2-Chlorophenol	95-57-8	STBJ3909	99%	1,001.8 µg/mL	+/- 36.4492
7	1,3-Dichlorobenzene	541-73-1	BCCD5315	99%	1,002.3 µg/mL	+/- 36.4654
8	1,4-Dichlorobenzene	106-46-7	MKBS7929V	99%	1,003.7 µg/mL	+/- 36.5159
9	Benzyl alcohol	100-51-6	SHBK5469	99%	1,008.7 µg/mL	+/- 36.6979
10	1,2-Dichlorobenzene	95-50-1	SHBN3835	99%	1,000.3 µg/mL	+/- 36.3926
11	2-Methylphenol (o-cresol)	95-48-7	SHBN7598	99%	1,003.5 µg/mL	+/- 36.5099
12	2,2'-oxybis(1-chloropropane)	108-60-1	29-MAR-45-5	99%	1,007.3 µg/mL	+/- 36.6493
13	3-Methylphenol (m-cresol)	108-39-4	STBJ0710	99%	504.3 µg/mL	+/- 18.3500
14	4-Methylphenol (p-cresol)	106-44-5	SHBN3411	99%	503.6 µg/mL	+/- 18.3237
15	N-Nitroso-di-n-propylamine	621-64-7	N63MG	99%	1,008.3 µg/mL	+/- 36.6857
16	Hexachloroethane	67-72-1	QTORH	99%	1,007.5 µg/mL	+/- 36.6554
17	Nitrobenzene	98-95-3	10224044	99%	1,008.6 µg/mL	+/- 36.6938

18	Isophorone	78-59-1	MKCC9506	99%	1,005.9	µg/mL	+/- 36.5988
19	2-Nitrophenol	88-75-5	RP230710	99%	1,003.2	µg/mL	+/- 36.4998
20	2,4-Dimethylphenol	105-67-9	XW5GK	99%	1,003.8	µg/mL	+/- 36.5200
21	Bis(2-chloroethoxy)methane	111-91-1	13670200	99%	1,002.1	µg/mL	+/- 36.4573
22	2,4-Dichlorophenol	120-83-2	BCBZ6787	99%	1,003.7	µg/mL	+/- 36.5180
23	1,2,4-Trichlorobenzene	120-82-1	SHBP5900	99%	1,007.6	µg/mL	+/- 36.6574
24	Naphthalene	91-20-3	STBL1057	99%	1,008.3	µg/mL	+/- 36.6837
25	4-Chloroaniline	106-47-8	BCCJ3217	99%	1,001.3	µg/mL	+/- 36.4290
26	Hexachlorobutadiene	87-68-3	RP230823RSR	98%	1,008.3	µg/mL	+/- 36.6829
27	4-Chloro-3-methylphenol	59-50-7	BCCD4461	99%	1,003.1	µg/mL	+/- 36.4937
28	2-Methylnaphthalene	91-57-6	STBK0259	96%	1,001.9	µg/mL	+/- 36.4505
29	1-Methylnaphthalene	90-12-0	5234.00-8	98%	1,000.0	µg/mL	+/- 36.3838
30	Hexachlorocyclopentadiene	77-47-4	099063I14L	98%	1,008.5	µg/mL	+/- 36.6909
31	2,4,6-Trichlorophenol	88-06-2	STBJ5914	99%	1,004.4	µg/mL	+/- 36.5442
32	2,4,5-Trichlorophenol	95-95-4	FHN01	98%	1,001.9	µg/mL	+/- 36.4512
33	2-Chloronaphthalene	91-58-7	RPN7O	99%	1,001.1	µg/mL	+/- 36.4230
34	2-Nitroaniline	88-74-4	RP230531	99%	1,002.9	µg/mL	+/- 36.4876
35	1,4-Dinitrobenzene	100-25-4	RP230816	99%	1,005.7	µg/mL	+/- 36.5887
36	Acenaphthylene	208-96-8	p06V	98%	1,009.5	µg/mL	+/- 36.7265
37	1,3-Dinitrobenzene	99-65-0	1-DXX-24-1	99%	1,004.4	µg/mL	+/- 36.5422
38	Dimethylphthalate	131-11-3	358221L17K	99%	1,005.9	µg/mL	+/- 36.5968
39	2,6-Dinitrotoluene	606-20-2	BCCG1833	99%	1,003.2	µg/mL	+/- 36.4998
40	1,2-Dinitrobenzene	528-29-0	RP230428	99%	1,002.2	µg/mL	+/- 36.4634
41	Acenaphthene	83-32-9	MKCR7169	99%	1,009.3	µg/mL	+/- 36.7221
42	3-Nitroaniline	99-09-2	RP230822RSR	99%	1,003.9	µg/mL	+/- 36.5240
43	2,4-Dinitrophenol	51-28-5	DR230417RSR	99%	1,002.0	µg/mL	+/- 36.4553
44	Dibenzofuran	132-64-9	MKCD9952	99%	1,006.7	µg/mL	+/- 36.6251
45	2,4-Dinitrotoluene	121-14-2	MKAA0690V	99%	1,003.8	µg/mL	+/- 36.5220
46	4-Nitrophenol	100-02-7	RP230627	99%	1,002.3	µg/mL	+/- 36.4674
47	2,3,4,6-Tetrachlorophenol	58-90-2	PR-30126	99%	1,008.7	µg/mL	+/- 36.6979
48	2,3,5,6-Tetrachlorophenol	935-95-5	RP230919	99%	1,006.3	µg/mL	+/- 36.6130
49	Fluorene	86-73-7	10241100	99%	1,008.3	µg/mL	+/- 36.6857
50	4-Chlorophenyl phenyl ether	7005-72-3	MKCT7248	99%	1,003.8	µg/mL	+/- 36.5220
51	Diethylphthalate	84-66-2	MKCD2547	99%	1,008.6	µg/mL	+/- 36.6958
52	4-Nitroaniline	100-01-6	RP230111	99%	1,001.1	µg/mL	+/- 36.4230
53	4,6-Dinitro-2-methylphenol (Dinitro-o-cresol)	534-52-1	230718JLM	99%	1,002.0	µg/mL	+/- 36.4553

54	Diphenylamine	122-39-4	MKCH1042	99%	1,002.3	µg/mL	+/- 36.4674
55	Azobenzene	103-33-3	BCKK0887	99%	1,005.8	µg/mL	+/- 36.5928
56	4-Bromophenyl phenyl ether	101-55-3	STBH6361	99%	1,003.0	µg/mL	+/- 36.4917
57	Hexachlorobenzene	118-74-1	14821700	99%	1,007.5	µg/mL	+/- 36.6554
58	Pentachlorophenol	87-86-5	RP230530RSR	99%	1,008.8	µg/mL	+/- 36.7019
59	Phenanthrene	85-01-8	MKCQ8876	99%	1,008.4	µg/mL	+/- 36.6877
60	Anthracene	120-12-7	MKCR0570	99%	1,009.0	µg/mL	+/- 36.7100
61	Carbazole	86-74-8	14351100	99%	1,000.9	µg/mL	+/- 36.4149
62	Di-n-butylphthalate	84-74-2	MKCN4337	99%	1,007.6	µg/mL	+/- 36.6595
63	Fluoranthene	206-44-0	MKCQ4728	99%	1,009.6	µg/mL	+/- 36.7302
64	Pyrene	129-00-0	BCCG8479	98%	1,007.2	µg/mL	+/- 36.6453
65	Benzyl butyl phthalate	85-68-7	X12I018	99%	1,002.1	µg/mL	+/- 36.4573
66	Bis(2-ethylhexyl)adipate	103-23-1	MKCM1988	99%	1,005.2	µg/mL	+/- 36.5705
67	Benz(a)anthracene	56-55-3	I220012022BAA	99%	1,002.2	µg/mL	+/- 36.4614
68	Chrysene	218-01-9	RP230601	99%	1,008.3	µg/mL	+/- 36.6837
69	Bis(2-ethylhexyl)phthalate	117-81-7	MKCQ3468	99%	1,001.8	µg/mL	+/- 36.4472
70	Di-n-octyl phthalate	117-84-0	14382700	99%	1,006.0	µg/mL	+/- 36.6008
71	Benzo(b)fluoranthene	205-99-2	012013B	99%	1,002.8	µg/mL	+/- 36.4836
72	Benzo(k)fluoranthene	207-08-9	012022K	99%	1,003.0	µg/mL	+/- 36.4917
73	Benzo(a)pyrene	50-32-8	P54915-0703	99%	1,002.3	µg/mL	+/- 36.4674
74	Indeno(1,2,3-cd)pyrene	193-39-5	12-JKL-118-9	97%	1,009.4	µg/mL	+/- 36.7243
75	Dibenz(a,h)anthracene	53-70-3	2-ASA-59-1	99%	1,007.6	µg/mL	+/- 36.6595
76	Benzo(g,h,i)perylene	191-24-2	RP231003RSR	99%	1,002.9	µg/mL	+/- 36.4876

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31850 **Lot No.:** A0203726

Description : 8270 MegaMix®
8270 MegaMix® 500-1000 µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2025 **Storage:** 0°C or colder

Handling: Sonication required. Mix is photosensitive. **Ship:** Ambient

512117 } RC/
↓ 03/18/24
512146 }

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Pyridine	110-86-1	SHBP6240	99%	1,001.6 µg/mL	+/- 36.4412
2	N-Nitrosodimethylamine	62-75-9	230209JLM	99%	1,005.9 µg/mL	+/- 36.5968
3	Phenol	108-95-2	MKCK1120	99%	1,003.3 µg/mL	+/- 36.5038
4	Aniline	62-53-3	X22F726	99%	1,005.8 µg/mL	+/- 36.5928
5	Bis(2-chloroethyl)ether	111-44-4	SHBL6942	99%	1,008.1 µg/mL	+/- 36.6776
6	2-Chlorophenol	95-57-8	STBJ3909	99%	1,001.8 µg/mL	+/- 36.4492
7	1,3-Dichlorobenzene	541-73-1	BCCD5315	99%	1,002.3 µg/mL	+/- 36.4654
8	1,4-Dichlorobenzene	106-46-7	MKBS7929V	99%	1,003.7 µg/mL	+/- 36.5159
9	Benzyl alcohol	100-51-6	SHBK5469	99%	1,008.7 µg/mL	+/- 36.6979
10	1,2-Dichlorobenzene	95-50-1	SHBN3835	99%	1,000.3 µg/mL	+/- 36.3926
11	2-Methylphenol (o-cresol)	95-48-7	SHBN7598	99%	1,003.5 µg/mL	+/- 36.5099
12	2,2'-oxybis(1-chloropropane)	108-60-1	29-MAR-45-5	99%	1,007.3 µg/mL	+/- 36.6493
13	3-Methylphenol (m-cresol)	108-39-4	STBJ0710	99%	504.3 µg/mL	+/- 18.3500
14	4-Methylphenol (p-cresol)	106-44-5	SHBN3411	99%	503.6 µg/mL	+/- 18.3237
15	N-Nitroso-di-n-propylamine	621-64-7	N63MG	99%	1,008.3 µg/mL	+/- 36.6857
16	Hexachloroethane	67-72-1	QTORH	99%	1,007.5 µg/mL	+/- 36.6554
17	Nitrobenzene	98-95-3	10224044	99%	1,008.6 µg/mL	+/- 36.6938

18	Isophorone	78-59-1	MKCC9506	99%	1,005.9	µg/mL	+/- 36.5988
19	2-Nitrophenol	88-75-5	RP230710	99%	1,003.2	µg/mL	+/- 36.4998
20	2,4-Dimethylphenol	105-67-9	XW5GK	99%	1,003.8	µg/mL	+/- 36.5200
21	Bis(2-chloroethoxy)methane	111-91-1	13670200	99%	1,002.1	µg/mL	+/- 36.4573
22	2,4-Dichlorophenol	120-83-2	BCBZ6787	99%	1,003.7	µg/mL	+/- 36.5180
23	1,2,4-Trichlorobenzene	120-82-1	SHBP5900	99%	1,007.6	µg/mL	+/- 36.6574
24	Naphthalene	91-20-3	STBL1057	99%	1,008.3	µg/mL	+/- 36.6837
25	4-Chloroaniline	106-47-8	BCCJ3217	99%	1,001.3	µg/mL	+/- 36.4290
26	Hexachlorobutadiene	87-68-3	RP230823RSR	98%	1,008.3	µg/mL	+/- 36.6829
27	4-Chloro-3-methylphenol	59-50-7	BCCD4461	99%	1,003.1	µg/mL	+/- 36.4937
28	2-Methylnaphthalene	91-57-6	STBK0259	96%	1,001.9	µg/mL	+/- 36.4505
29	1-Methylnaphthalene	90-12-0	5234.00-8	98%	1,000.0	µg/mL	+/- 36.3838
30	Hexachlorocyclopentadiene	77-47-4	099063I14L	98%	1,008.5	µg/mL	+/- 36.6909
31	2,4,6-Trichlorophenol	88-06-2	STBJ5914	99%	1,004.4	µg/mL	+/- 36.5442
32	2,4,5-Trichlorophenol	95-95-4	FHN01	98%	1,001.9	µg/mL	+/- 36.4512
33	2-Chloronaphthalene	91-58-7	RPN7O	99%	1,001.1	µg/mL	+/- 36.4230
34	2-Nitroaniline	88-74-4	RP230531	99%	1,002.9	µg/mL	+/- 36.4876
35	1,4-Dinitrobenzene	100-25-4	RP230816	99%	1,005.7	µg/mL	+/- 36.5887
36	Acenaphthylene	208-96-8	p06V	98%	1,009.5	µg/mL	+/- 36.7265
37	1,3-Dinitrobenzene	99-65-0	1-DXX-24-1	99%	1,004.4	µg/mL	+/- 36.5422
38	Dimethylphthalate	131-11-3	358221L17K	99%	1,005.9	µg/mL	+/- 36.5968
39	2,6-Dinitrotoluene	606-20-2	BCCG1833	99%	1,003.2	µg/mL	+/- 36.4998
40	1,2-Dinitrobenzene	528-29-0	RP230428	99%	1,002.2	µg/mL	+/- 36.4634
41	Acenaphthene	83-32-9	MKCR7169	99%	1,009.3	µg/mL	+/- 36.7221
42	3-Nitroaniline	99-09-2	RP230822RSR	99%	1,003.9	µg/mL	+/- 36.5240
43	2,4-Dinitrophenol	51-28-5	DR230417RSR	99%	1,002.0	µg/mL	+/- 36.4553
44	Dibenzofuran	132-64-9	MKCD9952	99%	1,006.7	µg/mL	+/- 36.6251
45	2,4-Dinitrotoluene	121-14-2	MKAA0690V	99%	1,003.8	µg/mL	+/- 36.5220
46	4-Nitrophenol	100-02-7	RP230627	99%	1,002.3	µg/mL	+/- 36.4674
47	2,3,4,6-Tetrachlorophenol	58-90-2	PR-30126	99%	1,008.7	µg/mL	+/- 36.6979
48	2,3,5,6-Tetrachlorophenol	935-95-5	RP230919	99%	1,006.3	µg/mL	+/- 36.6130
49	Fluorene	86-73-7	10241100	99%	1,008.3	µg/mL	+/- 36.6857
50	4-Chlorophenyl phenyl ether	7005-72-3	MKCT7248	99%	1,003.8	µg/mL	+/- 36.5220
51	Diethylphthalate	84-66-2	MKCD2547	99%	1,008.6	µg/mL	+/- 36.6958
52	4-Nitroaniline	100-01-6	RP230111	99%	1,001.1	µg/mL	+/- 36.4230
53	4,6-Dinitro-2-methylphenol (Dinitro-o-cresol)	534-52-1	230718JLM	99%	1,002.0	µg/mL	+/- 36.4553

54	Diphenylamine	122-39-4	MKCH1042	99%	1,002.3	µg/mL	+/- 36.4674
55	Azobenzene	103-33-3	BCKK0887	99%	1,005.8	µg/mL	+/- 36.5928
56	4-Bromophenyl phenyl ether	101-55-3	STBH6361	99%	1,003.0	µg/mL	+/- 36.4917
57	Hexachlorobenzene	118-74-1	14821700	99%	1,007.5	µg/mL	+/- 36.6554
58	Pentachlorophenol	87-86-5	RP230530RSR	99%	1,008.8	µg/mL	+/- 36.7019
59	Phenanthrene	85-01-8	MKCQ8876	99%	1,008.4	µg/mL	+/- 36.6877
60	Anthracene	120-12-7	MKCR0570	99%	1,009.0	µg/mL	+/- 36.7100
61	Carbazole	86-74-8	14351100	99%	1,000.9	µg/mL	+/- 36.4149
62	Di-n-butylphthalate	84-74-2	MKCN4337	99%	1,007.6	µg/mL	+/- 36.6595
63	Fluoranthene	206-44-0	MKCQ4728	99%	1,009.6	µg/mL	+/- 36.7302
64	Pyrene	129-00-0	BCCG8479	98%	1,007.2	µg/mL	+/- 36.6453
65	Benzyl butyl phthalate	85-68-7	X12I018	99%	1,002.1	µg/mL	+/- 36.4573
66	Bis(2-ethylhexyl)adipate	103-23-1	MKCM1988	99%	1,005.2	µg/mL	+/- 36.5705
67	Benz(a)anthracene	56-55-3	I220012022BAA	99%	1,002.2	µg/mL	+/- 36.4614
68	Chrysene	218-01-9	RP230601	99%	1,008.3	µg/mL	+/- 36.6837
69	Bis(2-ethylhexyl)phthalate	117-81-7	MKCQ3468	99%	1,001.8	µg/mL	+/- 36.4472
70	Di-n-octyl phthalate	117-84-0	14382700	99%	1,006.0	µg/mL	+/- 36.6008
71	Benzo(b)fluoranthene	205-99-2	012013B	99%	1,002.8	µg/mL	+/- 36.4836
72	Benzo(k)fluoranthene	207-08-9	012022K	99%	1,003.0	µg/mL	+/- 36.4917
73	Benzo(a)pyrene	50-32-8	P54915-0703	99%	1,002.3	µg/mL	+/- 36.4674
74	Indeno(1,2,3-cd)pyrene	193-39-5	12-JKL-118-9	97%	1,009.4	µg/mL	+/- 36.7243
75	Dibenz(a,h)anthracene	53-70-3	2-ASA-59-1	99%	1,007.6	µg/mL	+/- 36.6595
76	Benzo(g,h,i)perylene	191-24-2	RP231003RSR	99%	1,002.9	µg/mL	+/- 36.4876

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555223 **Lot No.:** A0214021

Description : Custom 8270 Plus Standard #1

Custom 8270 Plus Standard #1 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2026 **Storage:** 10°C or colder

Handling: This product is photosensitive. **Ship:** Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	3,3'-Dichlorobenzidine	91-94-1	S240326RSR	99%	1,004.0 µg/mL	+/- 23.0487
2	Atrazine	1912-24-9	5FYWL	99%	1,005.0 µg/mL	+/- 23.0717
3	Benzidine	92-87-5	S240430RSR	99%	1,006.0 µg/mL	+/- 23.0947
4	epsilon-Caprolactam	105-60-2	Y16H012	99%	1,000.0 µg/mL	+/- 22.9569

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

S12449 } RC/
↓
S12508 } 7/24/24

Rebecca Gingerich - Operations Tech II

Date Mixed: 18-Jul-2024

Balance: 1128353505

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555223 **Lot No.:** A0214021

Description : Custom 8270 Plus Standard #1

Custom 8270 Plus Standard #1 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2026 **Storage:** 10°C or colder

Handling: This product is photosensitive. **Ship:** Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	3,3'-Dichlorobenzidine	91-94-1	S240326RSR	99%	1,004.0 µg/mL	+/- 23.0487
2	Atrazine	1912-24-9	5FYWL	99%	1,005.0 µg/mL	+/- 23.0717
3	Benzidine	92-87-5	S240430RSR	99%	1,006.0 µg/mL	+/- 23.0947
4	epsilon-Caprolactam	105-60-2	Y16H012	99%	1,000.0 µg/mL	+/- 22.9569

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

S12449 } RC/
↓
S12508 } 7/24/24

Rebecca Gingerich - Operations Tech II

Date Mixed: 18-Jul-2024

Balance: 1128353505

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555223 **Lot No.:** A0214021

Description : Custom 8270 Plus Standard #1

Custom 8270 Plus Standard #1 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2026 **Storage:** 10°C or colder

Handling: This product is photosensitive. **Ship:** Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	3,3'-Dichlorobenzidine	91-94-1	S240326RSR	99%	1,004.0 µg/mL	+/- 23.0487
2	Atrazine	1912-24-9	5FYWL	99%	1,005.0 µg/mL	+/- 23.0717
3	Benzidine	92-87-5	S240430RSR	99%	1,006.0 µg/mL	+/- 23.0947
4	epsilon-Caprolactam	105-60-2	Y16H012	99%	1,000.0 µg/mL	+/- 22.9569

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

S12449 } RC/
↓
S12508 } 7/24/24

Rebecca Gingerich - Operations Tech II

Date Mixed: 18-Jul-2024

Balance: 1128353505

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/ μ ECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555223 **Lot No.:** A0214021

Description : Custom 8270 Plus Standard #1

Custom 8270 Plus Standard #1 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2026 **Storage:** 10°C or colder

Handling: This product is photosensitive. **Ship:** Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	3,3'-Dichlorobenzidine	91-94-1	S240326RSR	99%	1,004.0 µg/mL	+/- 23.0487
2	Atrazine	1912-24-9	5FYWL	99%	1,005.0 µg/mL	+/- 23.0717
3	Benzidine	92-87-5	S240430RSR	99%	1,006.0 µg/mL	+/- 23.0947
4	epsilon-Caprolactam	105-60-2	Y16H012	99%	1,000.0 µg/mL	+/- 22.9569

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

S12449 } RC/
↓
S12508 } 7/24/24

Rebecca Gingerich - Operations Tech II

Date Mixed: 18-Jul-2024

Balance: 1128353505

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555223 **Lot No.:** A0214021

Description : Custom 8270 Plus Standard #1

Custom 8270 Plus Standard #1 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2026 **Storage:** 10°C or colder

Handling: This product is photosensitive. **Ship:** Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	3,3'-Dichlorobenzidine	91-94-1	S240326RSR	99%	1,004.0 µg/mL	+/- 23.0487
2	Atrazine	1912-24-9	5FYWL	99%	1,005.0 µg/mL	+/- 23.0717
3	Benzidine	92-87-5	S240430RSR	99%	1,006.0 µg/mL	+/- 23.0947
4	epsilon-Caprolactam	105-60-2	Y16H012	99%	1,000.0 µg/mL	+/- 22.9569

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

S12449 } RC/
↓ 7/24/24
S12508 }

Rebecca Gingerich - Operations Tech II

Date Mixed: 18-Jul-2024

Balance: 1128353505

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



SHIPPING DOCUMENTS

Copyright © 2023



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Laboratory Certification

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (L-A-B)	L2219
Maine	2024021
Maryland	296
New Hampshire	255423
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	T104704488

LOGIN REPORT/SAMPLE TRANSFER

Order ID : P3609 JACO05

Order Date : 8/14/2024 12:48:00 PM

Project Mgr : Yazmeen

Client Name : JACOBS Engineering Grou

Project Name : Former Schlumberger Site I

Report Type : Level 4

Client Contact : Mary I. Murphy

Receive DateTime : 8/14/2024 2:15:00 PM

EDD Type : CH2MHILL

Invoice Name : JACOBS Engineering Grou

Purchase Order :

Hard Copy Date :

Invoice Contact : Mary I. Murphy

Date Signoff : 8/14/2024 4:40:21 PM

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
P3609-01	915-J-WS-081424	Water	08/14/2024	10:25					
					VOCMS Group6		8260-Low	10 Bus. Days	
P3609-02	920-J-WS-081424	Water	08/14/2024	11:20					
					VOCMS Group6		8260-Low	10 Bus. Days	
P3609-03	TB-01-081424	Water	08/14/2024	11:25					
					VOCMS Group6		8260-Low	10 Bus. Days	

Relinquished By :

Date / Time :


8/14/24

Received By :

Date / Time :


8/14/24

Storage Area : VOA Refridgerator Room