

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P3609 SAS No.: P3609 SDG NO.: P3609

EPA Sample No.: SSTDCCC040 Date Analyzed: 08/19/2024

Lab File ID: BM047245.D Time Analyzed: 09:42

Instrument ID: BNA_M GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	1666060	16.792	1393630	21.039	1655410	23.744
UPPER LIMIT	3332120	17.292	2787260	21.539	3310820	24.244
LOWER LIMIT	833030	16.292	696815	20.539	827705	23.244
EPA SAMPLE NO.						
01 915-J-WS-081424	1167960	16.79	1274470	21.03	1251150	23.73
02 920-J-WS-081424	911557	16.79	1070130	21.03	1179970	23.73
03 PB162757BS	1358930	16.79	1331980	21.04	1613230	23.74
04 PB162757BSD	1252430	16.79	1245980	21.03	1572180	23.74

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.