

284 Sheffield Street, Mountainside, New Jersey 07092, Phone: 908 789 8900,

Fax: 908 789 8922

51

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH	Contract: JACO05	
Lab Code: CHEM	SAS No.: P3609	SDG NO.: P3609
Lab File ID: BM047307.D	DFTPP Injection Date:	08/22/2024
Instrument ID: BNA_M	DFTPP Injection Time:	08:35

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE	
51	10.0 - 80.0% of mass 198	25.4	
68	Less than 2.0% of mass 69	0.5 (1.5)	1
69	Mass 69 relative abundance	32.2	
70	Less than 2.0% of mass 69	0.1 (0.4)	1
127	10.0 - 80.0% of mass 198	39.1	
197	Less than 2.0% of mass 198	0.3	
198	Base Peak, 100% relative abundance	100	
199	5.0 to 9.0% of mass 198	6.9	
275	10.0 - 60.0% of mass 198	26.2	
365	Greater than 1% of mass 198	3.2	
441	Present, but less than mass 443	9.3	
442	Greater than 50% of mass 198	100	
443	15.0 - 24.0% of mass 442	11.5 (19.2)	2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BM047308.D	08/22/2024	09:14
PB162757BL	PB162757BL	BM047313.D	08/22/2024	12:32