

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: JACO05  
Lab Code: CHEM Case No.: P3609 SAS No.: P3609 SDG NO.: P3609  
Lab File ID: VN083295.D Date Analyzed: 08/14/2024  
Instrument ID: MSVOA\_N Time Analyzed: 12:41  
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

|                 | IS1<br>AREA # | RT #  | IS2<br>AREA # | RT # | IS3<br>AREA # | RT #   |
|-----------------|---------------|-------|---------------|------|---------------|--------|
| 12 HOUR STD     | 165534        | 8.22  | 282606        | 9.10 | 250996        | 11.87  |
| UPPER LIMIT     | 331068        | 8.724 | 565212        | 9.6  | 501992        | 12.365 |
| LOWER LIMIT     | 82767         | 7.724 | 141303        | 8.6  | 125498        | 11.365 |
| EPA SAMPLE NO.  |               |       |               |      |               |        |
| 915-J-WS-081424 | 123251        | 8.22  | 241969        | 9.10 | 252623        | 11.87  |
| 920-J-WS-081424 | 120820        | 8.22  | 238487        | 9.11 | 244841        | 11.87  |
| TB-01-081424    | 130232        | 8.22  | 253989        | 9.10 | 263608        | 11.87  |
| VN0814WBL01     | 145129        | 8.23  | 276647        | 9.10 | 286871        | 11.87  |
| VN0814WBS01     | 164632        | 8.22  | 284346        | 9.10 | 250581        | 11.87  |
| VN0814WBSD01    | 148957        | 8.22  | 263108        | 9.10 | 232246        | 11.87  |

IS1 = Pentafluorobenzene  
IS2 = 1,4-Difluorobenzene  
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area  
AREA LOWER LIMIT = -50% of internal standard area  
RT UPPER LIMIT = +0.50 minutes of internal standard RT  
RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.  
\* Values outside of QC limits.