

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P3645

Client: JACOBS Engineering Group, Inc.

Analytical Method: 8270-Modified DataFile: BN033494.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	RPD		Limits		
							Qual	Qual	Low	High	RPD
PB162787BSD	Naphthalene	0.4	0.38	ug/L	95	5			70 (67)	130 (120)	20 (20)
	2-Methylnaphthalene	0.4	0.37	ug/L	93	3			70 (50)	130 (122)	20 (20)
	Acenaphthylene	0.4	0.37	ug/L	93	3			70 (60)	130 (119)	20 (20)
	Acenaphthene	0.4	0.38	ug/L	95	5			70 (65)	130 (119)	20 (20)
	Fluorene	0.4	0.35	ug/L	88	3			70 (58)	130 (122)	20 (20)
	Phenanthrene	0.4	0.39	ug/L	98	3			70 (65)	130 (117)	20 (20)
	Anthracene	0.4	0.38	ug/L	95	5			70 (57)	130 (118)	20 (20)
	Fluoranthene	0.4	0.35	ug/L	88	6			70 (53)	130 (126)	20 (20)
	Pyrene	0.4	0.39	ug/L	98	3			70 (54)	130 (124)	20 (20)
	Benzo(a)anthracene	0.4	0.42	ug/L	105	10			70 (54)	130 (130)	20 (20)
	Chrysene	0.4	0.46	ug/L	115	14			70 (64)	130 (126)	20 (20)
	Benzo(b)fluoranthene	0.4	0.44	ug/L	110	15			70 (65)	130 (121)	20 (20)
	Benzo(k)fluoranthene	0.4	0.43	ug/L	108	10			70 (72)	130 (119)	20 (20)
	Benzo(a)pyrene	0.4	0.46	ug/L	115	14			70 (68)	130 (120)	20 (20)
	Indeno(1,2,3-cd)pyrene	0.4	0.49	ug/L	123	11			70 (70)	130 (127)	20 (20)
	Dibenz(a,h)anthracene	0.4	0.48	ug/L	120	11			70 (65)	130 (121)	20 (20)
	Benzo(g,h,i)perylene	0.4	0.47	ug/L	117	14			70 (76)	130 (117)	20 (20)
	1,4-Dioxane	0.4	0.31	ug/L	78	7			20 (42)	160 (127)	20 (20)