

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P3645 SAS No.: P3645 SDG NO.: P3645

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 08/20/2024

Lab File ID: BN033489.D Time Analyzed: 04:44

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	9356	7.552	25244	10.31	13324	14.19
UPPER LIMIT	18712	8.052	50488	10.814	26648	14.689
LOWER LIMIT	4678	7.052	12622	9.814	6662	13.689
EPA SAMPLE NO.						
01 PB162787BS	7475	7.55	18804	10.31	8660	14.19
02 PB162787BSD	6965	7.55	17427	10.31	7934	14.19

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.