

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P3645 SAS No.: P3645 SDG NO.: P3645

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 08/20/2024

Lab File ID: BN033507.D Time Analyzed: 15:51

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	8554	7.552	22897	10.31	11496	14.19
UPPER LIMIT	17108	8.052	45794	10.814	22992	14.689
LOWER LIMIT	4277	7.052	11448.5	9.814	5748	13.689
EPA SAMPLE NO.						
01 914-J-WS-081524	7581	7.55	20168	10.31	10160	14.19
02 916-J-WS-081524	7720	7.55	19614	10.31	9099	14.19
03 PB162787BL	8441	7.55	21567	10.31	9670	14.19

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.