

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P3645 SAS No.: P3645 SDG NO.: P3645

EPA Sample No.: SSTDCCC040 Date Analyzed: 08/20/2024

Lab File ID: BM047273.D Time Analyzed: 16:55

Instrument ID: BNA_M GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	329969	7.357	1193760	10.12	731367	14.02
UPPER LIMIT	659938	7.857	2387520	10.616	1462730	14.522
LOWER LIMIT	164985	6.857	596880	9.616	365684	13.522
EPA SAMPLE NO.						
01 916-J-WS-081524	312772	7.36	1142660	10.11	745797	14.02
02 914-J-WS-081524	298308	7.36	1085870	10.11	684460	14.02

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.