

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P3645 SAS No.: P3645 SDG NO.: P3645

EPA Sample No.: SSTDCCC040 Date Analyzed: 08/20/2024

Lab File ID: BM047273.D Time Analyzed: 16:55

Instrument ID: BNA_M GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	1421240	16.786	1062420	21.027	1267880	23.721
UPPER LIMIT	2842480	17.286	2124840	21.527	2535760	24.221
LOWER LIMIT	710620	16.286	531210	20.527	633940	23.221
EPA SAMPLE NO.						
01 916-J-WS-081524	1530970	16.78	1309540	21.02	1291410	23.72
02 914-J-WS-081524	1321030	16.78	1060410	21.02	1115040	23.72

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.