

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: JAC005

Lab Code: CHEM

SAS No.: P3645

SDG NO.: P3645

Lab File ID: BM047139.D

DFTPP Injection Date: 08/10/2024

Instrument ID: BNA_M

DFTPP Injection Time: 12:26

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	35.7
68	Less than 2.0% of mass 69	0.6 (1.3) 1
69	Mass 69 relative abundance	44.5
70	Less than 2.0% of mass 69	0.2 (0.5) 1
127	10.0 - 80.0% of mass 198	49.9
197	Less than 2.0% of mass 198	0.4
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	7
275	10.0 - 60.0% of mass 198	25.8
365	Greater than 1% of mass 198	3.6
441	Present, but less than mass 443	9.2
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	11.4 (19.7) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC2.5	SSTDICC2.5	BM047140.D	08/10/2024	13:12
SSTDICC005	SSTDICC005	BM047141.D	08/10/2024	13:52
SSTDICC010	SSTDICC010	BM047142.D	08/10/2024	14:31
SSTDICC020	SSTDICC020	BM047143.D	08/10/2024	15:11
SSTDICCC040	SSTDICCC040	BM047144.D	08/10/2024	15:51
SSTDICC050	SSTDICC050	BM047145.D	08/10/2024	16:31
SSTDICC060	SSTDICC060	BM047146.D	08/10/2024	17:11
SSTDICC080	SSTDICC080	BM047147.D	08/10/2024	17:51