

Instrument ID: BNA_M

Daily Analysis Runlog For Sequence/QC Batch ID # BM081024

Review By	Jagrut	Review On	8/12/2024 5:03:24 PM
Supervise By	mohammad	Supervise On	8/13/2024 5:17:30 AM
SubDirectory	BM081024	HP Acquire Method	BNA_M
		HP Processing Method	bm081024

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6550,SP6551,SP6552,SP6553,SP6554,SP6555,SP6556,SP6557
CCC	SP6553
Internal Standard/PEM	S12039,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BM047139.D	10 Aug 2024 12:26		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BM047140.D	10 Aug 2024 13:12		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BM047141.D	10 Aug 2024 13:52	Compound #54 removed from 5ppm	RC/JU	Ok,M
4	SSTDICC010	SSTDICC010	BM047142.D	10 Aug 2024 14:31	Com#77 failing in calibration for both 8270DOD and 8270E	RC/JU	Ok,M
5	SSTDICC020	SSTDICC020	BM047143.D	10 Aug 2024 15:11	Compound #9 Kept on QR	RC/JU	Ok,M
6	SSTDICCC040	SSTDICCC040	BM047144.D	10 Aug 2024 15:51	The Calibration is Good For 8270 DOD and also good for 625.1 Method except com#77	RC/JU	Ok,M
7	SSTDICC050	SSTDICC050	BM047145.D	10 Aug 2024 16:31		RC/JU	Ok,M
8	SSTDICC060	SSTDICC060	BM047146.D	10 Aug 2024 17:11		RC/JU	Ok,M
9	SSTDICC080	SSTDICC080	BM047147.D	10 Aug 2024 17:51	Compound #9 removed from 80ppm	RC/JU	Ok,M
10	SSTDICV040	ICVBM081024	BM047148.D	10 Aug 2024 19:10	Com#77 failing in ICV for Both 8270 DOD and 8270 E and also in 625.1.	RC/JU	Ok,M
11	PB162577TB	PB162577TB	BM047149.D	10 Aug 2024 19:50		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_M

Daily Analysis Runlog For Sequence/QC Batch ID # BM082024

Review By	yogesh	Review On	8/21/2024 1:29:58 AM
Supervise By	mohammad	Supervise On	8/22/2024 7:51:34 AM
SubDirectory	BM082024	HP Acquire Method	BNA_M
		HP Processing Method	bm081024

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6550,SP6551,SP6552,SP6553,SP6554,SP6555,SP6556,SP6557
CCC	SP6553
Internal Standard/PEM	S12041,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BM047262.D	20 Aug 2024 08:32		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BM047263.D	20 Aug 2024 09:11		RC/JU	Ok
3	PB162798BL	PB162798BL	BM047264.D	20 Aug 2024 09:51		RC/JU	Ok
4	PB162798BS	PB162798BS	BM047265.D	20 Aug 2024 10:30		RC/JU	Ok,M
5	P3643-01	932-K1-SD-0-0.5-08152	BM047266.D	20 Aug 2024 11:28		RC/JU	Ok
6	P3643-02MS	932-K1-SD-0-0.5-08152	BM047267.D	20 Aug 2024 12:07		RC/JU	Ok,M
7	P3643-03MSD	932-K1-SD-0-0.5-08152	BM047268.D	20 Aug 2024 12:47		RC/JU	Ok
8	P3643-04	932-K1-SD-0.5-1-08152	BM047269.D	20 Aug 2024 13:27		RC/JU	Ok
9	P3643-08	918-J-SD-0.5-1-081524	BM047270.D	20 Aug 2024 14:06		RC/JU	Ok
10	P3643-09	924-K1-SD-0-0.5-08152	BM047271.D	20 Aug 2024 14:46		RC/JU	Ok
11	DFTPP	DFTPP	BM047272.D	20 Aug 2024 15:35		RC/JU	Ok
12	SSTDCCC040	SSTDCCC040	BM047273.D	20 Aug 2024 16:55		RC/JU	Ok
13	PB162806TB	PB162806TB	BM047274.D	20 Aug 2024 17:36		RC/JU	Ok
14	PB162861BL	PB162861BL	BM047275.D	20 Aug 2024 18:16		RC/JU	Ok
15	PB162861BS	PB162861BS	BM047276.D	20 Aug 2024 18:56		RC/JU	Ok
16	P3637-02	ETGI - 370	BM047277.D	20 Aug 2024 19:35		RC/JU	Ok
17	P3637-02MS	ETGI - 370MS	BM047278.D	20 Aug 2024 20:15		RC/JU	Ok
18	P3637-02MSD	ETGI - 370MSD	BM047279.D	20 Aug 2024 20:55		RC/JU	Ok

Instrument ID: BNA_M

Daily Analysis Runlog For Sequence/QC Batch ID # BM082024

Review By	yogesh	Review On	8/21/2024 1:29:58 AM		
Supervise By	mohammad	Supervise On	8/22/2024 7:51:34 AM		
SubDirectory	BM082024	HP Acquire Method	BNA_M	HP Processing Method	bm081024
STD. NAME		STD REF.#			
Tune/Reschk		SP6573			
Initial Calibration Stds		SP6550,SP6551,SP6552,SP6553,SP6554,SP6555,SP6556,SP6557			
CCC		SP6553			
Internal Standard/PEM		S12041,10ul/1000ul sample			
ICV/I.BLK		SP6559			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	P3645-01	914-J-WS-081524	BM047280.D	20 Aug 2024 21:34		RC/JU	Ok
20	P3645-02	916-J-WS-081524	BM047281.D	20 Aug 2024 22:14		RC/JU	Ok
21	P3652-01	LIVONIA-SOIL	BM047282.D	20 Aug 2024 22:54		RC/JU	Ok,M
22	P3652-01MS	LIVONIA-SOILMS	BM047283.D	20 Aug 2024 23:34	CCC fail low side for compound #19,35,56,62	RC/JU	Not Ok
23	P3652-01MSD	LIVONIA-SOILMSD	BM047284.D	21 Aug 2024 00:13	CCC fail low side for compound #19,35,56,62	RC/JU	Not Ok
24	P3643-13	921-J-SD-0-0.5-081524	BM047285.D	21 Aug 2024 00:53	FD	RC/JU	Ok
25	P3643-05	917-J-SD-0-0.5-081524	BM047286.D	21 Aug 2024 01:33		RC/JU	Ok
26	P3643-11	921-J-SD-0-0.5-081524	BM047287.D	21 Aug 2024 02:12		RC/JU	Ok,M
27	P3580-01	926-K1-SD-0-0.5-08122	BM047288.D	21 Aug 2024 02:52		RC/JU	Ok,M

M : Manual Integration

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP081324

Review By	yogesh	Review On	8/14/2024 5:49:15 AM
Supervise By	mohammad	Supervise On	8/14/2024 6:23:14 AM
SubDirectory	BP081324	HP Acquire Method	BNA_P
		HP Processing Method	BP081324

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6595,SP6594,SP6593,SP6592,SP6591,SP6590,SP6589,SP6588
CCC	SP6591
Internal Standard/PEM	S12041,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BP021480.D	13 Aug 2024 11:49		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BP021481.D	13 Aug 2024 12:32		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BP021482.D	13 Aug 2024 13:15	Compound #32,54,56,65,70,77,85 removed from 5ppm	RC/JU	Ok,M
4	SSTDICC010	SSTDICC010	BP021483.D	13 Aug 2024 13:59		RC/JU	Ok,M
5	SSTDICC020	SSTDICC020	BP021484.D	13 Aug 2024 14:42	Compound #48,51,53,56,57,62,80,85 Kept on LR and Compound #26,32,54,65,77 Kept on QR	RC/JU	Ok,M
6	SSTDICCC040	SSTDICCC040	BP021485.D	13 Aug 2024 15:26	The Calibration is Good For 8270 DOD except com#32,77 and good for 8270E	RC/JU	Ok,M
7	SSTDICC050	SSTDICC050	BP021486.D	13 Aug 2024 16:09		RC/JU	Ok,M
8	SSTDICC060	SSTDICC060	BP021487.D	13 Aug 2024 16:53		RC/JU	Ok,M
9	SSTDICC080	SSTDICC080	BP021488.D	13 Aug 2024 17:36	Compound #9 removed from 80ppm	RC/JU	Ok,M
10	SSTDICV040	ICVBP081324	BP021489.D	13 Aug 2024 19:03	Com#32 and 77 failed in ICV for 8270 DOD Criteria	RC/JU	Ok,M
11	PB162612TB	PB162612TB	BP021490.D	13 Aug 2024 19:46	Already analyzed with ok status	RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP082124

Review By	yogesh	Review On	8/22/2024 2:06:51 AM
Supervise By	mohammad	Supervise On	8/22/2024 7:52:20 AM
SubDirectory	BP082124	HP Acquire Method	BNA_P
		HP Processing Method	BP081324

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6595,SP6594,SP6593,SP6592,SP6591,SP6590,SP6589,SP6588
CCC	SP6591
Internal Standard/PEM	S12041,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BP021591.D	21 Aug 2024 09:28		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BP021592.D	21 Aug 2024 10:10		RC/JU	Ok,M
3	PB162788BL	PB162788BL	BP021593.D	21 Aug 2024 10:52		RC/JU	Ok
4	PB162788BS	PB162788BS	BP021594.D	21 Aug 2024 12:19		RC/JU	Ok,M
5	PB162788BSD	PB162788BSD	BP021595.D	21 Aug 2024 13:08		RC/JU	Ok,M
6	PB162717BS	PB162717BS	BP021596.D	21 Aug 2024 13:52		RC/JU	Ok,M
7	PB162631BS	PB162631BS	BP021597.D	21 Aug 2024 14:35		RC/JU	Ok,M
8	PB162878BL	PB162878BL	BP021598.D	21 Aug 2024 15:17		RC/JU	Ok
9	PB162878BS	PB162878BS	BP021599.D	21 Aug 2024 15:59		RC/JU	Ok,M
10	PB162878BSD	PB162878BSD	BP021600.D	21 Aug 2024 16:42		RC/JU	Ok,M
11	P3666-03	SP-303-20240816	BP021601.D	21 Aug 2024 17:25		RC/JU	Ok
12	P3617-02	Field Blank	BP021602.D	21 Aug 2024 18:08		RC/JU	Ok
13	P3617-03	VCPS-B2-081424-6W	BP021603.D	21 Aug 2024 18:51		RC/JU	Ok
14	P3617-04	VCPS-B2-081424-6W-I	BP021604.D	21 Aug 2024 19:35		RC/JU	Ok,M
15	SSTDCCC040	SSTDCCC040EC	BP021605.D	21 Aug 2024 20:18		RC/JU	Ok,M

M : Manual Integration