

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: JAC005

Lab Code: CHEM

SAS No.: P3645 SDG NO.: P3645

Lab File ID: BM047272.D

DFTPP Injection Date: 08/20/2024

Instrument ID: BNA_M

DFTPP Injection Time: 15:35

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	27.7
68	Less than 2.0% of mass 69	0.4 (1.3) 1
69	Mass 69 relative abundance	35.2
70	Less than 2.0% of mass 69	0.1 (0.4) 1
127	10.0 - 80.0% of mass 198	41.5
197	Less than 2.0% of mass 198	0.4
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.7
275	10.0 - 60.0% of mass 198	25.7
365	Greater than 1% of mass 198	3.3
441	Present, but less than mass 443	9.2
442	Greater than 50% of mass 198	58.4
443	15.0 - 24.0% of mass 442	11.6 (19.8) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BM047273.D	08/20/2024	16:55
914-J-WS-081524	P3645-01	BM047280.D	08/20/2024	21:34
916-J-WS-081524	P3645-02	BM047281.D	08/20/2024	22:14