

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: JAC005

Lab Code: CHEM

SAS No.: P3645 SDG NO.: P3645

Lab File ID: BP021480.D

DFTPP Injection Date: 08/13/2024

Instrument ID: BNA_P

DFTPP Injection Time: 11:49

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	30.1
68	Less than 2.0% of mass 69	0.6 (1.5) 1
69	Mass 69 relative abundance	40.6
70	Less than 2.0% of mass 69	0.2 (0.4) 1
127	10.0 - 80.0% of mass 198	43.9
197	Less than 2.0% of mass 198	0.5
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.6
275	10.0 - 60.0% of mass 198	28.5
365	Greater than 1% of mass 198	3.5
441	Present, but less than mass 443	15
442	Greater than 50% of mass 198	97.2
443	15.0 - 24.0% of mass 442	18.7 (19.3) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC2.5	SSTDICC2.5	BP021481.D	08/13/2024	12:32
SSTDICC005	SSTDICC005	BP021482.D	08/13/2024	13:15
SSTDICC010	SSTDICC010	BP021483.D	08/13/2024	13:59
SSTDICC020	SSTDICC020	BP021484.D	08/13/2024	14:42
SSTDICCC040	SSTDICCC040	BP021485.D	08/13/2024	15:26
SSTDICC050	SSTDICC050	BP021486.D	08/13/2024	16:09
SSTDICC060	SSTDICC060	BP021487.D	08/13/2024	16:53
SSTDICC080	SSTDICC080	BP021488.D	08/13/2024	17:36