

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: JAC005

Lab Code: CHEM

SAS No.: P3645

SDG NO.: P3645

Lab File ID: BP021591.D

DFTPP Injection Date: 08/21/2024

Instrument ID: BNA_P

DFTPP Injection Time: 09:28

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	28.9
68	Less than 2.0% of mass 69	0.6 (1.7) 1
69	Mass 69 relative abundance	36.5
70	Less than 2.0% of mass 69	0.2 (0.6) 1
127	10.0 - 80.0% of mass 198	43.3
197	Less than 2.0% of mass 198	0.3
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.7
275	10.0 - 60.0% of mass 198	29.2
365	Greater than 1% of mass 198	3.7
441	Present, but less than mass 443	14.9
442	Greater than 50% of mass 198	94.9
443	15.0 - 24.0% of mass 442	18.7 (19.7) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BP021592.D	08/21/2024	10:10
PB162788BL	PB162788BL	BP021593.D	08/21/2024	10:52
PB162788BS	PB162788BS	BP021594.D	08/21/2024	12:19
PB162788BSD	PB162788BSD	BP021595.D	08/21/2024	13:08