

Manual Integration Report

Sequence:

vn080724

Instrument

MSVOA_n

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDICC100	VN083135.D	1,2,3-Trichloropropane	JOHN	8/8/2024 9:35:12 AM	MMDadoda	8/9/2024 6:15:44 AM	Peak Integrated by Software incorrectly
VSTDICCC050	VN083136.D	1,2,3-Trichloropropane	JOHN	8/8/2024 9:35:17 AM	MMDadoda	8/9/2024 6:15:45 AM	Peak Integrated by Software incorrectly
VSTDICCC050	VN083136.D	Vinyl Acetate	JOHN	8/8/2024 9:35:17 AM	MMDadoda	8/9/2024 6:15:45 AM	Peak Integrated by Software incorrectly
VSTDICC020	VN083137.D	1,2,3-Trichloropropane	JOHN	8/8/2024 9:35:21 AM	MMDadoda	8/9/2024 6:15:46 AM	Peak Integrated by Software incorrectly
VSTDICC020	VN083137.D	Vinyl Acetate	JOHN	8/8/2024 9:35:21 AM	MMDadoda	8/9/2024 6:15:46 AM	Peak Integrated by Software incorrectly
VSTDICC010	VN083138.D	1,2,3-Trichloropropane	JOHN	8/8/2024 9:35:26 AM	MMDadoda	8/9/2024 6:15:48 AM	Peak Integrated by Software incorrectly
VSTDICC005	VN083139.D	1,2,3-Trichloropropane	JOHN	8/8/2024 9:35:30 AM	MMDadoda	8/9/2024 6:15:49 AM	Peak Integrated by Software incorrectly
VSTDICC005	VN083139.D	Vinyl Acetate	JOHN	8/8/2024 9:35:30 AM	MMDadoda	8/9/2024 6:15:49 AM	Peak Integrated by Software incorrectly
VSTDICC001	VN083140.D	1,2,3-Trichloropropane	JOHN	8/8/2024 9:35:34 AM	MMDadoda	8/9/2024 6:15:51 AM	Peak Integrated by Software incorrectly
VSTDICC001	VN083140.D	1,4-Dichlorobenzene	JOHN	8/8/2024 9:35:34 AM	MMDadoda	8/9/2024 6:15:51 AM	Peak Integrated by Software incorrectly
VSTDICV050	VN083142.D	1,2,3-Trichloropropane	JOHN	8/8/2024 9:35:41 AM	MMDadoda	8/9/2024 6:15:52 AM	Peak Integrated by Software incorrectly

Manual Integration Report

Sequence:	vn081924	Instrument	MSVOA_n
-----------	----------	------------	---------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDCCC050	VN083367.D	1,2,3-Trichloropropane	SAM	8/20/2024 1:15:47 PM	MMDadoda	8/20/2024 4:36:00 PM	Peak Integrated by Software incorrectly
VSTDCCC050	VN083367.D	Vinyl Acetate	SAM	8/20/2024 1:15:47 PM	MMDadoda	8/20/2024 4:36:00 PM	Peak Integrated by Software incorrectly
VN0819WBS01	VN083370.D	1,2,3-Trichloropropane	SAM	8/20/2024 1:15:52 PM	MMDadoda	8/20/2024 4:36:01 PM	Peak Integrated by Software incorrectly
VN0819WBS01	VN083370.D	Vinyl Acetate	SAM	8/20/2024 1:15:52 PM	MMDadoda	8/20/2024 4:36:01 PM	Peak Integrated by Software incorrectly
VN0819WBSD0 1	VN083371.D	1,2,3-Trichloropropane	SAM	8/20/2024 1:16:01 PM	MMDadoda	8/20/2024 4:36:02 PM	Peak Integrated by Software incorrectly
VN0819WBSD0 1	VN083371.D	Vinyl Acetate	SAM	8/20/2024 1:16:01 PM	MMDadoda	8/20/2024 4:36:02 PM	Peak Integrated by Software incorrectly
VSTDCCC050	VN083389.D	1,2,3-Trichloropropane	SAM	8/20/2024 1:42:21 PM	MMDadoda	8/20/2024 4:36:04 PM	Peak Integrated by Software incorrectly
VSTDCCC050	VN083389.D	Vinyl Acetate	SAM	8/20/2024 1:42:21 PM	MMDadoda	8/20/2024 4:36:04 PM	Peak Integrated by Software incorrectly

Manual Integration Report

Sequence:

vn082224

Instrument

MSVOA_n

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDCCC050	VN083424.D	1,2,3-Trichloropropane	SAM	8/23/2024 11:47:13 AM	MMDadoda	8/23/2024 4:54:15 PM	Peak Integrated by Software incorrectly
VN0822WBS01	VN083427.D	1,2,3-Trichloropropane	SAM	8/23/2024 11:47:17 AM	MMDadoda	8/23/2024 4:54:16 PM	Peak Integrated by Software incorrectly
VSTDCCC050	VN083445.D	1,2,3-Trichloropropane	SAM	8/23/2024 11:58:14 AM	MMDadoda	8/23/2024 4:54:19 PM	Peak Integrated by Software incorrectly