

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: JACO05
Lab Code: CHEM Case No.: P3645 SAS No.: P3645 SDG NO.: P3645
Lab File ID: VN083424.D Date Analyzed: 08/22/2024
Instrument ID: MSVOA_N Time Analyzed: 08:46
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	157462	8.22	268222	9.10	234428	11.87
UPPER LIMIT	314924	8.724	536444	9.6	468856	12.365
LOWER LIMIT	78731	7.724	134111	8.6	117214	11.365
EPA SAMPLE NO.						
TB-02-081524	125922	8.22	257527	9.10	277904	11.87
VN0822WBL01	142322	8.22	288957	9.10	306835	11.87
VN0822WBS01	155463	8.22	274386	9.10	237980	11.87

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.