

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: JACO05
 Lab Code: CHEM Case No.: P3645 SAS No.: P3645 SDG NO.: P3645
 Lab File ID: VN083367.D Date Analyzed: 08/19/2024
 Instrument ID: MSVOA_N Time Analyzed: 11:43
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	128940	13.794				
UPPER LIMIT	257880	14.294				
LOWER LIMIT	64470	13.294				
EPA SAMPLE NO.						
914-J-WS-081524	121097	13.79				
916-J-WS-081524	115723	13.79				
VN0819WBL01	134032	13.79				
VN0819WBS01	116728	13.79				
VN0819WBSD01	112767	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.