

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081924\
 Data File : BN033489.D
 Acq On : 20 Aug 2024 04:44
 Operator : MA/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 20 05:11:58 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN081924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 19 23:32:18 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	73	0.00
2	1,4-Dioxane	0.460	0.396	13.9	87	0.00
3	n-Nitrosodimethylamine	0.535	0.463	13.5	74	0.00
4 S	2-Fluorophenol	1.271	1.003	21.1	57	0.00
5 S	Phenol-d6	1.512	1.274	15.7	68	0.00
6	bis(2-Chloroethyl)ether	1.072	1.015	5.3	98	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	104	0.00
8 S	Nitrobenzene-d5	0.332	0.284	14.5	97	0.00
9	Naphthalene	1.069	0.986	7.8	108	0.00
10	Hexachlorobutadiene	0.213	0.196	8.0	105	0.00
11 SURR	2-Methylnaphthalene-d10	0.572	0.523	8.6	110	0.00
12	2-Methylnaphthalene	0.677	0.626	7.5	111	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	111	0.00
14 S	2,4,6-Tribromophenol	0.215	0.163	24.2	98	0.00
15 S	2-Fluorobiphenyl	1.634	1.511	7.5	113	0.00
16	Acenaphthylene	1.754	1.526	13.0	110	-0.01
17	Acenaphthene	1.234	1.122	9.1	113	0.00
18	Fluorene	1.555	1.402	9.8	113	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	112	0.00
20	4,6-Dinitro-2-methylphenol	0.062	0.044	29.0#	96	0.00
21	4-Bromophenyl-phenylether	0.243	0.219	9.9	110	0.00
22	Hexachlorobenzene	0.268	0.255	4.9	114	0.00
23	Atrazine	0.194	0.158	18.6	102	0.00
24	Pentachlorophenol	0.116	0.084	27.6#	97	0.00
25	Phenanthrene	1.113	1.038	6.7	112	0.00
26	Anthracene	0.985	0.858	12.9	109	0.00
27 SURR	Fluoranthene-d10	0.961	0.802	16.5	103	0.00
28	Fluoranthene	1.230	1.056	14.1	106	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	99	0.00
30	Pyrene	1.785	1.802	-1.0	105	0.00
31 S	Terphenyl-d14	0.909	0.886	2.5	102	0.00
32	Benzo(a)anthracene	1.446	1.315	9.1	99	0.00
33	Chrysene	1.437	1.353	5.8	101	0.00
34	Bis(2-ethylhexyl)phthalate	0.915	0.740	19.1	90	0.00
35 I	Perylene-d12	1.000	1.000	0.0	97	0.00
36	Indeno(1,2,3-cd)pyrene	1.661	1.557	6.3	98	-0.01
37	Benzo(b)fluoranthene	1.494	1.423	4.8	103	0.00
38	Benzo(k)fluoranthene	1.470	1.396	5.0	102	0.00
39 C	Benzo(a)pyrene	1.236	1.083	12.4	96	0.00
40	Dibenzo(a,h)anthracene	1.328	1.239	6.7	98	0.00
41	Benzo(g,h,i)perylene	1.420	1.309	7.8	98	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4