

Surrogate Summary

SW-846

SDG No.: P3657

Client: JACOBS Engineering Group, Inc.

Analytical Method: 8270-Modified

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
P3657-01	917-J-WS-081624	2-Methylnaphthalene-d10	0.4	0.31	77		30 (20)	150 (139)
		Fluoranthene-d10	0.4	0.37	93		30 (30)	150 (150)
		Nitrobenzene-d5	0.4	0.29	73		30 (27)	130 (123)
		2-Fluorobiphenyl	0.4	0.31	78		30 (34)	130 (132)
		Terphenyl-d14	0.4	0.46	115		30 (35)	130 (157)
PB162821BL	PB162821BL	2-Methylnaphthalene-d10	0.4	0.32	80		30 (20)	150 (139)
		Fluoranthene-d10	0.4	0.31	77		30 (30)	150 (150)
		Nitrobenzene-d5	0.4	0.33	81		30 (27)	130 (123)
		2-Fluorobiphenyl	0.4	0.36	90		30 (34)	130 (132)
		Terphenyl-d14	0.4	0.36	90		30 (35)	130 (157)
PB162821BS	PB162821BS	2-Methylnaphthalene-d10	0.4	0.45	112		30 (20)	150 (139)
		Fluoranthene-d10	0.4	0.32	79		30 (30)	150 (150)
		Nitrobenzene-d5	0.4	0.34	84		30 (27)	130 (123)
		2-Fluorobiphenyl	0.4	0.37	93		30 (34)	130 (132)
		Terphenyl-d14	0.4	0.38	94		30 (35)	130 (157)
PB162821BSD	PB162821BSD	2-Methylnaphthalene-d10	0.4	0.45	113		30 (20)	150 (139)
		Fluoranthene-d10	0.4	0.35	87		30 (30)	150 (150)
		Nitrobenzene-d5	0.4	0.36	90		30 (27)	130 (123)
		2-Fluorobiphenyl	0.4	0.39	98		30 (34)	130 (132)
		Terphenyl-d14	0.4	0.36	90		30 (35)	130 (157)