

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: JAC005

Lab Code: CHEM

SAS No.: P3657 SDG NO.: P3657

Lab File ID: BN033488.D

DFTPP Injection Date: 08/20/2024

Instrument ID: BNA_N

DFTPP Injection Time: 04:04

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	36.7
68	Less than 2.0% of mass 69	0.6 (1.7) 1
69	Mass 69 relative abundance	38
70	Less than 2.0% of mass 69	0.2 (0.5) 1
127	10.0 - 80.0% of mass 198	48.9
197	Less than 2.0% of mass 198	0.5
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.8
275	10.0 - 60.0% of mass 198	26.9
365	Greater than 1% of mass 198	3.8
441	Present, but less than mass 443	10.6
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	12.7 (17.9) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC0.4	SSTDCCC0.4	BN033489.D	08/20/2024	04:44
PB162821BL	PB162821BL	BN033490.D	08/20/2024	05:20
PB162821BS	PB162821BS	BN033491.D	08/20/2024	05:56
PB162821BSD	PB162821BSD	BN033492.D	08/20/2024	06:32
917-J-WS-081624	P3657-01	BN033501.D	08/20/2024	12:11