

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP082024\
 Data File : BP021575.D
 Acq On : 20 Aug 2024 16:40
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040

Quant Time: Aug 20 17:15:31 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 14 00:00:30 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.805	152	372530	20.000	ng	0.00
21) Naphthalene-d8	10.593	136	1563751	20.000	ng	0.00
39) Acenaphthene-d10	14.463	164	1068846	20.000	ng	0.00
64) Phenanthrene-d10	17.275	188	2403929	20.000	ng	0.02
76) Chrysene-d12	21.739	240	2426350	20.000	ng	0.04
86) Perylene-d12	25.180	264	2940308	20.000	ng	0.08
System Monitoring Compounds						
5) 2-Fluorophenol	5.399	112	2044869	81.506	ng	0.00
7) Phenol-d6	6.969	99	2928471	80.016	ng	0.00
23) Nitrobenzene-d5	8.952	82	2633490	86.924	ng	0.00
42) 2,4,6-Tribromophenol	15.981	330	1133914	95.320	ng	0.01
45) 2-Fluorobiphenyl	13.063	172	5181271	73.981	ng	0.00
79) Terphenyl-d14	19.998	244	9297974	74.447	ng	0.02
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.317	88	424462	35.681	ng	99
3) Pyridine	3.717	79	1208987	36.406	ng	97
4) n-Nitrosodimethylamine	3.622	42	371373	37.206	ng	95
6) Aniline	7.134	93	1348749	36.863	ng	96
8) 2-Chlorophenol	7.375	128	979003	42.226	ng	95
9) Benzaldehyde	6.946	77	1037198	44.262	ng	93
10) Phenol	6.999	94	1481572	39.093	ng	96
11) bis(2-Chloroethyl)ether	7.228	93	1172312	36.557	ng	96
12) 1,3-Dichlorobenzene	7.699	146	1084084	39.364	ng	97
13) 1,4-Dichlorobenzene	7.840	146	1102920	39.661	ng	98
14) 1,2-Dichlorobenzene	8.158	146	1043542	38.940	ng	99
15) Benzyl Alcohol	8.040	79	1015548	38.673	ng	96
16) 2,2'-oxybis(1-Chloropr...	8.334	45	1061205	35.934	ng	97
17) 2-Methylphenol	8.240	107	976782	40.769	ng	97
18) Hexachloroethane	8.887	117	454819	40.786	ng	98
19) n-Nitroso-di-n-propyla...	8.605	70	850374	33.637	ng	95
20) 3+4-Methylphenols	8.563	107	1356088	41.976	ng	99
22) Acetophenone	8.616	105	1802202	37.514	ng	99
24) Nitrobenzene	8.993	77	1345015	40.876	ng	93
25) Isophorone	9.522	82	2538499	35.467	ng	98
26) 2-Nitrophenol	9.704	139	508327	56.363	ng	92
27) 2,4-Dimethylphenol	9.769	122	711129	40.260	ng	98
28) bis(2-Chloroethoxy)met...	9.999	93	1593547	36.474	ng	99
29) 2,4-Dichlorophenol	10.240	162	947435	45.397	ng	98
30) 1,2,4-Trichlorobenzene	10.457	180	1037019	38.758	ng	98
31) Naphthalene	10.646	128	3220205	39.419	ng	100
32) Benzoic acid	9.905	122	875031	65.645	ng	92
33) 4-Chloroaniline	10.746	127	1251607	40.788	ng	97
34) Hexachlorobutadiene	10.940	225	645550	38.319	ng	98
35) Caprolactam	11.534	113	416030	47.892	ng	88
36) 4-Chloro-3-methylphenol	11.881	107	1284015	44.353	ng	96
37) 2-Methylnaphthalene	12.263	142	2186594	39.219	ng	100
38) 1-Methylnaphthalene	12.481	142	2166454	39.399	ng	99
40) 1,2,4,5-Tetrachloroben...	12.634	216	1175676	37.541	ng	100
41) Hexachlorocyclopentadiene	12.616	237	464856	46.834	ng	99
43) 2,4,6-Trichlorophenol	12.875	196	805218	46.287	ng	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP082024\
 Data File : BP021575.D
 Acq On : 20 Aug 2024 16:40
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040


Quant Time: Aug 20 17:15:31 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 14 00:00:30 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	12.946	196	901983	46.667	ng	97
46) 1,1'-Biphenyl	13.281	154	2870734	37.962	ng	98
47) 2-Chloronaphthalene	13.322	162	2258356	38.359	ng	99
48) 2-Nitroaniline	13.516	65	775883	47.126	ng	88
49) Acenaphthylene	14.181	152	3412612	39.430	ng	100
50) Dimethylphthalate	13.916	163	2903067	40.557	ng	100
51) 2,6-Dinitrotoluene	14.022	165	668456	46.431	ng	91
52) Acenaphthene	14.534	154	2178301	38.359	ng	99
53) 3-Nitroaniline	14.351	138	704846	49.513	ng	# 90
54) 2,4-Dinitrophenol	14.569	184	301078	66.049	ng	90
55) Dibenzofuran	14.869	168	3469479	39.716	ng	99
56) 4-Nitrophenol	14.663	139	623596	54.273	ng	88
57) 2,4-Dinitrotoluene	14.828	165	940265	50.859	ng	# 85
58) Fluorene	15.534	166	2801684	39.039	ng	100
59) 2,3,4,6-Tetrachlorophenol	15.092	232	787234	47.986	ng	98
60) Diethylphthalate	15.304	149	3034612	41.319	ng	100
61) 4-Chlorophenyl-phenyle...	15.528	204	1408811	37.977	ng	99
62) 4-Nitroaniline	15.539	138	744682	50.846	ng	92
63) Azobenzene	15.822	77	3217680	35.641	ng	97
65) 4,6-Dinitro-2-methylph...	15.604	198	455554	58.431	ng	99
66) n-Nitrosodiphenylamine	15.739	169	2431907	37.750	ng	100
67) 4-Bromophenyl-phenylether	16.439	248	941570	37.008	ng	99
68) Hexachlorobenzene	16.563	284	1154799	38.408	ng	99
69) Atrazine	16.722	200	704532	31.853	ng	100
70) Pentachlorophenol	16.916	266	808216	49.250	ng	98
71) Phenanthrene	17.322	178	4562430	39.464	ng	100
72) Anthracene	17.416	178	4446404	39.216	ng	100
73) Carbazole	17.686	167	4532118	41.477	ng	99
74) Di-n-butylphthalate	18.298	149	5580792	41.587	ng	100
75) Fluoranthene	19.404	202	5912433	41.680	ng	100
77) Benzidine	19.592	184	2026298	35.722	ng	100
78) Pyrene	19.780	202	6147221	38.762	ng	99
80) Butylbenzylphthalate	20.722	149	2641829	44.200	ng	97
81) Benzo(a)anthracene	21.716	228	6085084	39.950	ng	100
82) 3,3'-Dichlorobenzidine	21.627	252	2216157	45.379	ng	100
83) Chrysene	21.786	228	5688074	39.553	ng	99
84) Bis(2-ethylhexyl)phtha...	21.663	149	4058627	45.083	ng	99
85) Di-n-octyl phthalate	22.992	149	7075138	45.068	ng	98
87) Indeno(1,2,3-cd)pyrene	29.109	276	6485659	34.029	ng	93
88) Benzo(b)fluoranthene	24.092	252	6649815	39.379	ng	99
89) Benzo(k)fluoranthene	24.168	252	6261875	38.012	ng	100
90) Benzo(a)pyrene	25.021	252	5826473	39.759	ng	99
91) Dibenzo(a,h)anthracene	29.221	278	6298888	39.775	ng	100
92) Benzo(g,h,i)perylene	30.303	276	6370124	40.133	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Aug 20 17:15:31 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 14 00:00:30 2024
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 20.000 ng
 RT: 7.805 min Scan# 81
 Delta R.T. -0.006 min
 Lab File: BP021575.D
 Acq: 20 Aug 2024 16:40




Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040


Tgt Ion:152 Resp: 372530
 Ion Ratio Lower Upper
 152 100
 150 164.1 122.4 183.6
 115 70.5 54.7 82.1



#2
 1,4-Dioxane
 Concen: 35.681 ng
 RT: 3.317 min Scan# 56
 Delta R.T. -0.000 min
 Lab File: BP021575.D
 Acq: 20 Aug 2024 16:40

Tgt Ion: 88 Resp: 424462
 Ion Ratio Lower Upper
 88 100
 58 61.5 48.7 73.1
 43 20.1 15.4 23.2









#3
Pyridine
Concen: 36.406 ng
RT: 3.717 min Scan# 11
Delta R.T. 0.012 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion: 79 Resp: 1208987
Ion Ratio Lower Upper
79 100
52 51.3 39.0 58.6
51 25.2 19.2 28.8



#4
n-Nitrosodimethylamine
Concen: 37.206 ng
RT: 3.622 min Scan# 108
Delta R.T. 0.006 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40

Tgt Ion: 42 Resp: 371373
Ion Ratio Lower Upper
42 100
74 185.7 154.3 231.5
44 6.5 6.2 9.4









#5
2-Fluorophenol
Concen: 81.506 ng
RT: 5.399 min Scan# 410
Delta R.T. 0.006 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:112 Resp: 2044869
Ion Ratio Lower Upper
112 100
64 54.4 49.0 73.6
63 27.9 24.6 36.8



#6
Aniline
Concen: 36.863 ng
RT: 7.134 min Scan# 705
Delta R.T. -0.000 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40

Tgt Ion: 93 Resp: 1348749
Ion Ratio Lower Upper
93 100
66 35.1 30.2 45.2
65 17.5 14.7 22.1









#7
Phenol-d6
Concen: 80.016 ng
RT: 6.969 min Scan# 61
Delta R.T. 0.006 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion: 99 Resp: 2928471
Ion Ratio Lower Upper
99 100
42 12.4 10.2 15.4
71 30.3 25.8 38.8



#8
2-Chlorophenol
Concen: 42.226 ng
RT: 7.375 min Scan# 746
Delta R.T. -0.000 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40

Tgt Ion: 128 Resp: 979003
Ion Ratio Lower Upper
128 100
130 31.6 11.6 51.6
64 49.4 35.3 75.3









#9
Benzaldehyde
Concen: 44.262 ng
RT: 6.946 min Scan# 61
Delta R.T. -0.000 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion: 77 Resp: 1037198
Ion Ratio Lower Upper
77 100
105 98.7 72.1 112.1
106 95.4 67.9 107.9



#10
Phenol
Concen: 39.093 ng
RT: 6.999 min Scan# 682
Delta R.T. 0.012 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40

Tgt Ion: 94 Resp: 1481572
Ion Ratio Lower Upper
94 100
65 24.1 6.5 46.5
66 32.5 14.7 54.7









#11
bis(2-Chloroethyl)ether
Concen: 36.557 ng
RT: 7.228 min Scan# 721
Delta R.T. -0.000 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion: 93 Resp: 1172312




Ion	Ratio	Lower	Upper
93	100		
63	66.1	51.0	91.0
95	32.6	12.9	52.9




#12
1,3-Dichlorobenzene
Concen: 39.364 ng
RT: 7.699 min Scan# 801
Delta R.T. -0.000 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40

Tgt Ion: 146 Resp: 1084084

Ion	Ratio	Lower	Upper
146	100		
148	63.8	51.8	77.6
75	34.8	31.3	46.9









#13
1,4-Dichlorobenzene
Concen: 39.661 ng
RT: 7.840 min Scan# 81
Delta R.T. -0.000 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:146 Resp: 1102920
Ion Ratio Lower Upper
146 100
148 64.2 51.5 77.3
111 48.0 40.3 60.5



#14
1,2-Dichlorobenzene
Concen: 38.940 ng
RT: 8.158 min Scan# 879
Delta R.T. -0.000 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40

Tgt Ion:146 Resp: 1043542
Ion Ratio Lower Upper
146 100
148 65.0 51.5 77.3
111 50.9 41.1 61.7









#15
Benzyl Alcohol
Concen: 38.673 ng
RT: 8.040 min Scan# 858
Delta R.T. 0.006 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion: 79 Resp: 1015548
Ion Ratio Lower Upper
79 100
108 78.3 57.9 86.9
77 63.6 51.8 77.8



#16
2,2'-oxybis(1-Chloropropane)
Concen: 35.934 ng
RT: 8.334 min Scan# 909
Delta R.T. -0.000 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40

Tgt Ion: 45 Resp: 1061205
Ion Ratio Lower Upper
45 100
77 20.0 1.9 41.9
79 15.8 0.0 36.6









#17
2-Methylphenol
Concen: 40.769 ng
RT: 8.240 min Scan# 892
Delta R.T. 0.006 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:107 Resp: 976782




Ion	Ratio	Lower	Upper
107	100		
108	109.8	88.8	133.2
77	44.7	38.7	58.1
79	43.9	37.4	56.2




#18
Hexachloroethane
Concen: 40.786 ng
RT: 8.887 min Scan# 1003
Delta R.T. -0.006 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40

Tgt Ion:117 Resp: 454819

Ion	Ratio	Lower	Upper
117	100		
119	94.3	73.9	110.9
201	95.3	74.9	112.3









#19
n-Nitroso-di-n-propylamine
Concen: 33.637 ng
RT: 8.605 min Scan# 91
Delta R.T. -0.000 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion: 70 Resp: 850374




Ion	Ratio	Lower	Upper
70	100		
42	43.9	33.1	49.7
101	10.7	7.4	11.2
130	19.6	13.3	19.9




#20
3+4-Methylphenols
Concen: 41.976 ng
RT: 8.563 min Scan# 948
Delta R.T. 0.006 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40

Tgt Ion: 107 Resp: 1356088

Ion	Ratio	Lower	Upper
107	100		
108	87.5	67.6	107.6
77	32.4	15.0	55.0
79	28.2	8.9	48.9









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.593 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:136 Resp: 1563751
Ion Ratio Lower Upper
136 100
137 10.9 8.8 13.2
54 7.9 6.3 9.5
68 7.5 6.8 10.2



#22
Acetophenone
Concen: 37.514 ng
RT: 8.616 min Scan# 957
Delta R.T. -0.000 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40

Tgt Ion:105 Resp: 1802202
Ion Ratio Lower Upper
105 100
71 5.0 4.6 7.0
51 19.7 16.3 24.5
120 20.7 16.2 24.2





#23

Nitrobenzene-d5

Concen: 86.924 ng

RT: 8.952 min Scan# 1014

Delta R.T. -0.000 min

Lab File: BP021575.D

Acq: 20 Aug 2024 16:40

Instrument :

BNA_P

ClientSampleId :

SSTDCCC040





Tgt Ion: 82 Resp: 2633490

Ion Ratio Lower Upper

82 100

128 37.8 26.5 39.7

54 39.1 30.3 45.5



#24

Nitrobenzene

Concen: 40.876 ng

RT: 8.993 min Scan# 1021

Delta R.T. -0.000 min

Lab File: BP021575.D

Acq: 20 Aug 2024 16:40




Tgt Ion: 77 Resp: 1345015


Ion Ratio Lower Upper

77 100

123 42.6 29.5 44.3

65 12.3 9.9 14.9









#25
Isophorone
Concen: 35.467 ng
RT: 9.522 min Scan# 1111
Delta R.T. 0.000 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion: 82 Resp: 2538499
Ion Ratio Lower Upper
82 100
95 6.8 5.4 8.2
138 14.7 10.6 16.0



#26
2-Nitrophenol
Concen: 56.363 ng
RT: 9.704 min Scan# 1142
Delta R.T. 0.005 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40

Tgt Ion: 139 Resp: 508327
Ion Ratio Lower Upper
139 100
109 34.3 28.9 43.3
65 53.4 49.0 73.6









#27
2,4-Dimethylphenol
Concen: 40.260 ng
RT: 9.769 min Scan# 1151
Delta R.T. 0.012 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:122 Resp: 711129
Ion Ratio Lower Upper
122 100
107 129.7 106.2 159.4
121 59.4 48.8 73.2



#28
bis(2-Chloroethoxy)methane
Concen: 36.474 ng
RT: 9.999 min Scan# 1192
Delta R.T. -0.006 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40

Tgt Ion: 93 Resp: 1593547
Ion Ratio Lower Upper
93 100
95 32.2 26.2 39.4
123 9.7 7.6 11.4









#29
2,4-Dichlorophenol
Concen: 45.397 ng
RT: 10.240 min Scan# 1233
Delta R.T. -0.000 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:162 Resp: 947435
Ion Ratio Lower Upper
162 100
164 63.8 45.9 85.9
98 44.8 26.3 66.3



#30
1,2,4-Trichlorobenzene
Concen: 38.758 ng
RT: 10.457 min Scan# 1270
Delta R.T. -0.000 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40

Tgt Ion:180 Resp: 1037019
Ion Ratio Lower Upper
180 100
182 95.4 78.3 117.5
145 30.7 24.0 36.0









#31
Naphthalene
Concen: 39.419 ng
RT: 10.646 min Scan# 1117
Delta R.T. -0.000 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:128 Resp: 3220205
Ion Ratio Lower Upper
128 100
129 10.6 8.6 12.8
127 13.0 10.6 15.8



#32
Benzoic acid
Concen: 65.645 ng
RT: 9.905 min Scan# 1176
Delta R.T. 0.030 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40

Tgt Ion:122 Resp: 875031
Ion Ratio Lower Upper
122 100
105 137.7 124.6 164.6
77 103.6 96.2 136.2





#33

4-Chloroaniline

Concen: 40.788 ng

RT: 10.746 min Scan# 1318

Delta R.T. 0.006 min

Lab File: BP021575.D

Acq: 20 Aug 2024 16:40

Instrument :

BNA_P

ClientSampleId :

SSTDCCC040

Tgt Ion:127 Resp: 1251607





Ion Ratio Lower Upper

127 100

129 31.8 24.5 36.7

65 33.2 28.9 43.3

92 23.4 19.3 28.9



#34

Hexachlorobutadiene

Concen: 38.319 ng

RT: 10.940 min Scan# 1352

Delta R.T. -0.000 min

Lab File: BP021575.D

Acq: 20 Aug 2024 16:40




Tgt Ion:225 Resp: 645550


Ion Ratio Lower Upper

225 100

223 62.5 49.6 74.4

227 64.6 50.1 75.1









#35
 Caprolactam
 Concen: 47.892 ng
 RT: 11.534 min Scan# 1447
 Delta R.T. 0.035 min
 Lab File: BP021575.D
 Acq: 20 Aug 2024 16:40




Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040


Tgt Ion:113 Resp: 416030
 Ion Ratio Lower Upper
 113 100
 55 133.6 130.6 170.6
 56 102.6 94.6 134.6



#36
 4-Chloro-3-methylphenol
 Concen: 44.353 ng
 RT: 11.881 min Scan# 1512
 Delta R.T. 0.018 min
 Lab File: BP021575.D
 Acq: 20 Aug 2024 16:40

Tgt Ion:107 Resp: 1284015
 Ion Ratio Lower Upper
 107 100
 144 21.5 16.6 24.8
 142 67.2 51.0 76.4









#37
2-Methylnaphthalene
Concen: 39.219 ng
RT: 12.263 min Scan# 1577
Delta R.T. -0.000 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:142 Resp: 2186594
Ion Ratio Lower Upper
142 100
141 90.0 71.7 107.5
115 39.1 31.4 47.2



#38
1-Methylnaphthalene
Concen: 39.399 ng
RT: 12.481 min Scan# 1614
Delta R.T. -0.000 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40

Tgt Ion:142 Resp: 2166454
Ion Ratio Lower Upper
142 100
141 92.7 73.5 110.3
116 4.5 3.4 5.2









#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.463 min Scan# 1950
Delta R.T. 0.006 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:164 Resp: 1068846
Ion Ratio Lower Upper
164 100
162 100.1 81.5 122.3
160 43.4 35.2 52.8



#40
1,2,4,5-Tetrachlorobenzene
Concen: 37.541 ng
RT: 12.634 min Scan# 1640
Delta R.T. 0.006 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40

Tgt Ion:216 Resp: 1175676
Ion Ratio Lower Upper
216 100
214 78.2 62.5 93.7
179 20.6 16.4 24.6
108 20.7 16.7 25.1





#41

Hexachlorocyclopentadiene

Concen: 46.834 ng

RT: 12.616 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP021575.D

Acq: 20 Aug 2024 16:40

Instrument :

BNA_P

ClientSampleId :

SSTDCCC040





Tgt Ion:237 Resp: 464856

Ion Ratio Lower Upper

237 100

235 63.4 43.9 83.9

272 12.7 0.0 33.3



#42

2,4,6-Tribromophenol

Concen: 95.320 ng

RT: 15.981 min Scan# 2209

Delta R.T. 0.012 min

Lab File: BP021575.D

Acq: 20 Aug 2024 16:40




Tgt Ion:330 Resp: 1133914


Ion Ratio Lower Upper

330 100

332 97.2 77.4 116.0

141 31.0 24.5 36.7









#43
2,4,6-Trichlorophenol
Concen: 46.287 ng
RT: 12.875 min Scan# 1680
Delta R.T. 0.006 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:196 Resp: 805218
Ion Ratio Lower Upper
196 100
198 96.3 76.6 114.8
200 30.9 24.5 36.7



#44
2,4,5-Trichlorophenol
Concen: 46.667 ng
RT: 12.946 min Scan# 1693
Delta R.T. 0.012 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40

Tgt Ion:196 Resp: 901983
Ion Ratio Lower Upper
196 100
198 98.0 78.9 118.3
97 56.1 49.4 74.0
132 26.1 21.1 31.7









#45
2-Fluorobiphenyl
Concen: 73.981 ng
RT: 13.063 min Scan# 1713
Delta R.T. -0.000 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:172 Resp: 5181271
Ion Ratio Lower Upper
172 100
171 35.1 28.2 42.4
170 23.2 18.4 27.6



#46
1,1'-Biphenyl
Concen: 37.962 ng
RT: 13.281 min Scan# 1750
Delta R.T. -0.000 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40

Tgt Ion:154 Resp: 2870734
Ion Ratio Lower Upper
154 100
153 40.5 20.7 60.7
76 15.7 0.0 37.7





#47

2-Chloronaphthalene

Concen: 38.359 ng

RT: 13.322 min Scan# 1756

Delta R.T. 0.006 min

Lab File: BP021575.D

Acq: 20 Aug 2024 16:40

Instrument :

BNA_P

ClientSampleId :

SSTDCCC040





Tgt Ion:162 Resp: 2258356

Ion Ratio Lower Upper

162 100

127 39.9 31.9 47.9

164 32.4 26.8 40.2



#48

2-Nitroaniline

Concen: 47.126 ng

RT: 13.516 min Scan# 1790

Delta R.T. 0.012 min

Lab File: BP021575.D

Acq: 20 Aug 2024 16:40




Tgt Ion: 65 Resp: 775883


Ion Ratio Lower Upper

65 100

92 78.4 58.1 87.1

138 97.3 65.5 98.3









#49
Acenaphthylene
Concen: 39.430 ng
RT: 14.181 min Scan# 1903
Delta R.T. 0.012 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:152 Resp: 3412612
Ion Ratio Lower Upper
152 100
151 19.8 15.9 23.9
153 13.1 10.6 16.0



#50
Dimethylphthalate
Concen: 40.557 ng
RT: 13.916 min Scan# 1858
Delta R.T. 0.006 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40

Tgt Ion:163 Resp: 2903067
Ion Ratio Lower Upper
163 100
194 4.2 3.3 4.9
164 10.3 8.3 12.5





#51

2,6-Dinitrotoluene

Concen: 46.431 ng

RT: 14.022 min Scan# 1876

Delta R.T. -0.000 min

Lab File: BP021575.D

Acq: 20 Aug 2024 16:40

Instrument :

BNA_P

ClientSampleId :

SSTDCCC040





Tgt Ion:165 Resp: 668456

Ion Ratio Lower Upper

165 100

63 57.2 51.5 77.3

89 64.1 57.8 86.8



#52

Acenaphthene

Concen: 38.359 ng

RT: 14.534 min Scan# 1963

Delta R.T. 0.012 min

Lab File: BP021575.D

Acq: 20 Aug 2024 16:40




Tgt Ion:154 Resp: 2178301


Ion Ratio Lower Upper

154 100

153 112.8 91.1 136.7

152 53.0 43.0 64.4









#53
3-Nitroaniline
Concen: 49.513 ng
RT: 14.351 min Scan# 1932
Delta R.T. 0.000 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:138 Resp: 704846
Ion Ratio Lower Upper
138 100
108 13.1 16.2 24.4#
92 141.5 122.2 183.4



#54
2,4-Dinitrophenol
Concen: 66.049 ng
RT: 14.569 min Scan# 1969
Delta R.T. 0.018 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40

Tgt Ion:184 Resp: 301078
Ion Ratio Lower Upper
184 100
63 67.1 64.9 97.3
154 63.6 52.1 78.1









#55
Dibenzofuran
Concen: 39.716 ng
RT: 14.869 min Scan# 2018
Delta R.T. 0.012 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:168 Resp: 3469479
Ion Ratio Lower Upper
168 100
139 36.6 28.6 43.0
169 13.4 10.6 16.0



#56
4-Nitrophenol
Concen: 54.273 ng
RT: 14.663 min Scan# 1985
Delta R.T. 0.012 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40

Tgt Ion:139 Resp: 623596
Ion Ratio Lower Upper
139 100
109 82.5 70.2 110.2
65 104.4 101.8 141.8









#57
2,4-Dinitrotoluene
Concen: 50.859 ng
RT: 14.828 min Scan# 2010
Delta R.T. 0.018 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:165 Resp: 940265
Ion Ratio Lower Upper
165 100
63 42.2 42.5 63.7#
89 81.5 77.1 115.7
182 2.7 2.0 3.0



#58
Fluorene
Concen: 39.039 ng
RT: 15.534 min Scan# 2133
Delta R.T. 0.018 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40

Tgt Ion:166 Resp: 2801684
Ion Ratio Lower Upper
166 100
165 97.4 78.1 117.1
167 13.5 11.0 16.4









#59
2,3,4,6-Tetrachlorophenol
Concen: 47.986 ng
RT: 15.092 min Scan# 2058
Delta R.T. 0.006 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:232 Resp: 787234
Ion Ratio Lower Upper
232 100
131 42.3 33.1 49.7
130 2.3 1.7 2.5
166 29.8 23.0 34.4



#60
Diethylphthalate
Concen: 41.319 ng
RT: 15.304 min Scan# 2094
Delta R.T. -0.000 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40

Tgt Ion:149 Resp: 3034612
Ion Ratio Lower Upper
149 100
177 22.1 17.8 26.8
150 12.7 10.1 15.1









#61
4-Chlorophenyl-phenylether
Concen: 37.977 ng
RT: 15.528 min Scan# 2131
Delta R.T. 0.006 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:204 Resp: 1408811
Ion Ratio Lower Upper
204 100
206 33.3 26.4 39.6
141 58.9 46.1 69.1



#62
4-Nitroaniline
Concen: 50.846 ng
RT: 15.539 min Scan# 2134
Delta R.T. 0.011 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40

Tgt Ion:138 Resp: 744682
Ion Ratio Lower Upper
138 100
92 59.5 46.4 86.4
108 86.0 73.2 113.2









#63
Azobenzene
Concen: 35.641 ng
RT: 15.822 min Scan# 2182
Delta R.T. -0.000 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion: 77 Resp: 3217680




Ion	Ratio	Lower	Upper
77	100		
182	21.9	0.0	39.9
105	22.3	0.9	40.9
51	22.5	1.4	41.4




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.275 min Scan# 2429
Delta R.T. 0.023 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40

Tgt Ion: 188 Resp: 2403929

Ion	Ratio	Lower	Upper
188	100		
94	12.2	10.2	15.4
80	13.1	11.6	17.4









#65
4,6-Dinitro-2-methylphenol
Concen: 58.431 ng
RT: 15.604 min Scan# 2145
Delta R.T. 0.017 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:198 Resp: 455554
Ion Ratio Lower Upper
198 100
51 40.2 21.5 61.5
105 52.2 32.9 72.9



#66
n-Nitrosodiphenylamine
Concen: 37.750 ng
RT: 15.739 min Scan# 2168
Delta R.T. -0.000 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40

Tgt Ion:169 Resp: 2431907
Ion Ratio Lower Upper
169 100
168 66.8 53.4 80.2
167 36.3 28.7 43.1









#67
4-Bromophenyl-phenylether
Concen: 37.008 ng
RT: 16.439 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:248 Resp: 941570
Ion Ratio Lower Upper
248 100
250 97.9 78.6 117.8
141 69.2 53.9 80.9



#68
Hexachlorobenzene
Concen: 38.408 ng
RT: 16.563 min Scan# 2308
Delta R.T. 0.012 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40

Tgt Ion:284 Resp: 1154799
Ion Ratio Lower Upper
284 100
142 30.3 23.4 35.2
249 30.6 24.1 36.1









#69
 Atrazine
 Concen: 31.853 ng
 RT: 16.722 min Scan# 21
 Delta R.T. 0.006 min
 Lab File: BP021575.D
 Acq: 20 Aug 2024 16:40




Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040


Tgt Ion:200 Resp: 704532
 Ion Ratio Lower Upper
 200 100
 173 26.0 6.2 46.2
 215 48.4 28.4 68.4



#70
 Pentachlorophenol
 Concen: 49.250 ng
 RT: 16.916 min Scan# 2368
 Delta R.T. 0.024 min
 Lab File: BP021575.D
 Acq: 20 Aug 2024 16:40

Tgt Ion:266 Resp: 808216
 Ion Ratio Lower Upper
 266 100
 268 65.9 51.8 77.6
 264 65.3 50.5 75.7





#71

Phenanthrene

Concen: 39.464 ng

RT: 17.322 min Scan# 2433

Delta R.T. 0.023 min

Lab File: BP021575.D

Acq: 20 Aug 2024 16:40

Instrument :

BNA_P

ClientSampleId :

SSTDCCC040





Tgt Ion:178 Resp: 4562430

Ion Ratio Lower Upper

178 100

176 19.3 15.4 23.2

179 15.7 12.4 18.6



#72

Anthracene

Concen: 39.216 ng

RT: 17.416 min Scan# 2453

Delta R.T. 0.018 min

Lab File: BP021575.D

Acq: 20 Aug 2024 16:40




Tgt Ion:178 Resp: 4446404


Ion Ratio Lower Upper

178 100

176 18.5 14.9 22.3

179 15.3 12.5 18.7









#73
Carbazole
Concen: 41.477 ng
RT: 17.686 min Scan# 2496
Delta R.T. 0.018 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:167 Resp: 4532118
Ion Ratio Lower Upper
167 100
166 21.7 17.0 25.4
139 13.0 9.9 14.9



#74
Di-n-butylphthalate
Concen: 41.587 ng
RT: 18.298 min Scan# 2603
Delta R.T. 0.023 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40

Tgt Ion:149 Resp: 5580792
Ion Ratio Lower Upper
149 100
150 9.3 7.5 11.3
104 5.9 4.9 7.3





#75

Fluoranthene

Concen: 41.680 ng

RT: 19.404 min Scan# 21

Delta R.T. 0.023 min

Lab File: BP021575.D

Acq: 20 Aug 2024 16:40

Instrument :

BNA_P

ClientSampleId :

SSTDCCC040





Tgt Ion:202 Resp: 5912433

Ion Ratio Lower Upper

202 100

101 14.0 0.0 34.0

203 17.6 0.0 37.5



#76

Chrysene-d12

Concen: 20.000 ng

RT: 21.739 min Scan# 3188

Delta R.T. 0.041 min

Lab File: BP021575.D

Acq: 20 Aug 2024 16:40




Tgt Ion:240 Resp: 2426350


Ion Ratio Lower Upper

240 100

120 11.0 9.9 14.9

236 25.9 20.6 30.8





#77

Benzidine

Concen: 35.722 ng

RT: 19.592 min Scan# 2823

Delta R.T. 0.017 min

Lab File: BP021575.D

Acq: 20 Aug 2024 16:40

Instrument :

BNA_P

ClientSampleId :

SSTDCCC040





Tgt Ion:184 Resp: 2026298

Ion Ratio Lower Upper

184 100

185 14.1 11.4 17.0

183 11.7 9.3 13.9



#78

Pyrene

Concen: 38.762 ng

RT: 19.780 min Scan# 2855

Delta R.T. 0.029 min

Lab File: BP021575.D

Acq: 20 Aug 2024 16:40




Tgt Ion:202 Resp: 6147221


Ion Ratio Lower Upper

202 100

200 20.6 16.7 25.1

203 17.4 14.1 21.1





#79

Terphenyl-d14

Concen: 74.447 ng

RT: 19.998 min Scan# 2892

Delta R.T. 0.023 min

Lab File: BP021575.D

Acq: 20 Aug 2024 16:40

Instrument :

BNA_P

ClientSampleId :

SSTDCCC040





Tgt Ion:244 Resp: 9297974

Ion Ratio Lower Upper

244 100

212 7.0 5.6 8.4

122 11.3 9.4 14.2



#80

Butylbenzylphthalate

Concen: 44.200 ng

RT: 20.722 min Scan# 3015

Delta R.T. 0.012 min

Lab File: BP021575.D

Acq: 20 Aug 2024 16:40




Tgt Ion:149 Resp: 2641829


Ion Ratio Lower Upper

149 100

91 88.4 73.3 109.9

206 21.8 17.1 25.7









#81
Benzo(a)anthracene
Concen: 39.950 ng
RT: 21.716 min Scan# 3178
Delta R.T. 0.035 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:228 Resp: 6085084
Ion Ratio Lower Upper
228 100
226 27.2 21.4 32.2
229 19.6 15.8 23.6



#82
3,3'-Dichlorobenzidine
Concen: 45.379 ng
RT: 21.627 min Scan# 3169
Delta R.T. 0.018 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40

Tgt Ion:252 Resp: 2216157
Ion Ratio Lower Upper
252 100
254 64.5 51.7 77.5
126 11.1 8.7 13.1





#83

Chrysene

Concen: 39.553 ng

RT: 21.786 min Scan# 3196

Delta R.T. 0.035 min

Lab File: BP021575.D

Acq: 20 Aug 2024 16:40

Instrument :

BNA_P

ClientSampleId :

SSTDCCC040





Tgt Ion:228 Resp: 5688074

Ion Ratio Lower Upper

228 100

226 29.6 23.9 35.9

229 19.4 15.8 23.6



#84

Bis(2-ethylhexyl)phthalate

Concen: 45.083 ng

RT: 21.663 min Scan# 3175

Delta R.T. 0.018 min

Lab File: BP021575.D

Acq: 20 Aug 2024 16:40




Tgt Ion:149 Resp: 4058627


Ion Ratio Lower Upper

149 100

167 27.5 22.3 33.5

279 4.6 3.8 5.6





#85

Di-n-octyl phthalate

Concen: 45.068 ng

RT: 22.992 min Scan# 34

Delta R.T. 0.029 min

Lab File: BP021575.D

Acq: 20 Aug 2024 16:40

Instrument :

BNA_P

ClientSampleId :

SSTDCCC040





Tgt Ion:149 Resp: 7075138

Ion Ratio Lower Upper

149 100

167 1.6 1.3 1.9

43 8.1 7.3 10.9



#86

Perylene-d12

Concen: 20.000 ng

RT: 25.180 min Scan# 3773

Delta R.T. 0.076 min

Lab File: BP021575.D

Acq: 20 Aug 2024 16:40




Tgt Ion:264 Resp: 2940308


Ion Ratio Lower Upper

264 100

260 23.1 18.6 28.0

265 22.0 17.1 25.7





#87

Indeno(1,2,3-cd)pyrene

Concen: 34.029 ng

RT: 29.109 min Scan# 44

Delta R.T. 0.118 min

Lab File: BP021575.D

Acq: 20 Aug 2024 16:40

Instrument :

BNA_P

ClientSampleId :

SSTDCCC040





Tgt Ion:276 Resp: 6485659

Ion Ratio Lower Upper

276 100

138 18.7 12.8 19.2

277 24.2 16.5 24.7



#88

Benzo(b)fluoranthene

Concen: 39.379 ng

RT: 24.092 min Scan# 3588

Delta R.T. 0.065 min

Lab File: BP021575.D

Acq: 20 Aug 2024 16:40




Tgt Ion:252 Resp: 6649815


Ion Ratio Lower Upper

252 100

253 21.8 17.6 26.4

125 9.7 8.6 13.0









#89
Benzo(k)fluoranthene
Concen: 38.012 ng
RT: 24.168 min Scan# 30
Delta R.T. 0.059 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:252 Resp: 6261875
Ion Ratio Lower Upper
252 100
253 22.0 17.5 26.3
125 9.7 7.8 11.8



#90
Benzo(a)pyrene
Concen: 39.759 ng
RT: 25.021 min Scan# 3746
Delta R.T. 0.071 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40

Tgt Ion:252 Resp: 5826473
Ion Ratio Lower Upper
252 100
253 22.2 17.7 26.5
125 12.0 8.7 13.1





#91

Dibenzo(a,h)anthracene

Concen: 39.775 ng

RT: 29.221 min Scan# 4437

Delta R.T. 0.135 min

Lab File: BP021575.D

Acq: 20 Aug 2024 16:40

Instrument :

BNA_P

ClientSampleId :

SSTDCCC040





Tgt Ion:278 Resp: 6298888

Ion Ratio Lower Upper

278 100

139 15.1 12.2 18.4

279 24.1 19.1 28.7



#92

Benzo(g,h,i)perylene

Concen: 40.133 ng

RT: 30.303 min Scan# 4644

Delta R.T. 0.129 min

Lab File: BP021575.D

Acq: 20 Aug 2024 16:40

Tgt Ion:276 Resp: 6370124

Ion Ratio Lower Upper

276 100

277 23.9 19.0 28.6

138 19.0 16.0 24.0

