

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P3657 SAS No.: P3657 SDG NO.: P3657

EPA Sample No.: SSTDCCC040 Date Analyzed: 08/20/2024

Lab File ID: BP021566.D Time Analyzed: 10:07

Instrument ID: BNA_P GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	357434	7.805	1454010	10.59	971251	14.45
UPPER LIMIT	714868	8.305	2908020	11.093	1942500	14.951
LOWER LIMIT	178717	7.305	727005	10.093	485626	13.951
EPA SAMPLE NO.						
01 PB162822BL	385209	7.81	1560540	10.59	1021930	14.45
02 PB162822BS	494651	7.81	2073180	10.60	1389380	14.45
03 PB162822BSD	491039	7.81	2018200	10.59	1318740	14.46

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.