

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P3657 SAS No.: P3657 SDG NO.: P3657

EPA Sample No.: SSTDCCC040 Date Analyzed: 08/20/2024

Lab File ID: BP021575.D Time Analyzed: 16:40

Instrument ID: BNA_P GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	2403930	17.275	2426350	21.739	2940310	25.18
UPPER LIMIT	4807860	17.775	4852700	22.239	5880620	25.68
LOWER LIMIT	1201970	16.775	1213180	21.239	1470160	24.68
EPA SAMPLE NO.						
01 917-J-WS-081624	2099850	17.27	1963860	21.73	2261180	25.17

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.