

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: JACO05
Lab Code: CHEM Case No.: P3657 SAS No.: P3657 SDG No.: P3657
Instrument ID: BNA_P Calibration Date(s): 08/13/2024 08/13/2024
Calibration Time(s): 12:32 17:36

LAB FILE ID:		RRF2.5 = BP021481.D			RRF005 = BP021482.D		RRF010 = BP021483.D	
		RRF020 = BP021484.D			RRF040 = BP021485.D		RRF050 = BP021486.D	
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
Pyridine		1.744	1.720	1.915	1.802	1.741	1.783	4.0
2-Fluorophenol		1.332	1.286	1.405	1.348	1.324	1.347	3.2
Benzaldehyde		1.364	1.447	1.409	1.198	1.073	1.258	13.7
Phenol-d6		1.897	1.960	2.062	1.978	1.953	1.965	2.9
2-Methylphenol		1.257	1.313	1.350	1.290	1.277	1.286	3.0
3+4-Methylphenols		1.595	1.820	1.795	1.758	1.734	1.734	4.5
Nitrobenzene-d5		0.300	0.339	0.386	0.410	0.412	0.387	13.1
Hexachloroethane		0.599	0.590	0.617	0.594	0.587	0.599	2.0
Nitrobenzene		0.338	0.385	0.430	0.443	0.435	0.421	10.4
Naphthalene		1.084	1.041	1.074	1.041	1.002	1.045	2.6
Hexachlorobutadiene		0.217	0.213	0.218	0.215	0.210	0.215	1.5
2-Methylnaphthalene		0.747	0.742	0.730	0.705	0.683	0.713	3.8
2,4,6-Trichlorophenol		0.261	0.284	0.318	0.343	0.339	0.326	12.5
2-Fluorobiphenyl		1.371	1.314	1.335	1.327	1.264	1.310	3.1
2,4,5-Trichlorophenol		0.284	0.316	0.360	0.382	0.378	0.362	12.7
Acenaphthylene		1.660	1.610	1.657	1.645	1.561	1.619	2.5
Acenaphthene		1.094	1.058	1.074	1.073	1.023	1.063	2.2
Dibenzofuran		1.735	1.637	1.677	1.649	1.563	1.635	3.7
2,4-Dinitrotoluene		0.140	0.214	0.262	0.320	0.328	0.284	29.5
Fluorene		1.431	1.400	1.408	1.354	1.260	1.343	5.7
2,4,6-Tribromophenol		0.179	0.204	0.222	0.236	0.231	0.223	10.5
Hexachlorobenzene		0.255	0.249	0.250	0.252	0.245	0.250	1.3
Pentachlorophenol			0.108	0.125	0.137	0.143	0.137	13.1
Phenanthrene		0.990	0.963	0.998	0.965	0.932	0.962	2.7
Anthracene		0.954	0.955	0.970	0.950	0.915	0.943	2.2
Carbazole		0.908	0.906	0.941	0.930	0.873	0.909	2.5
Di-n-butylphthalate		1.049	1.065	1.149	1.149	1.133	1.116	3.7
Fluoranthene		1.215	1.153	1.233	1.200	1.137	1.180	3.2
Pyrene		1.283	1.470	1.295	1.290	1.262	1.307	5.6
Terphenyl-d14		1.069	1.196	1.034	1.006	0.973	1.029	8.2
Benzo (a) anthracene		1.271	1.247	1.291	1.268	1.224	1.256	1.8
Chrysene		1.213	1.180	1.205	1.197	1.151	1.185	1.8
Bis (2-ethylhexyl) phthalate		0.607	0.635	0.741	0.781	0.804	0.742	11.7
Benzo (b) fluoranthene		1.106	1.152	1.158	1.155	1.119	1.149	2.4

All other compounds must meet a minimum RRF of 0.010.

Form VI SV-1

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: JAC005
Lab Code: CHEM Case No.: P3657 SAS No.: P3657 SDG No.: P3657
Instrument ID: BNA_P Calibration Date(s): 08/13/2024 08/13/2024
Calibration Time(s): 12:32 17:36

LAB FILE ID:		RRF2.5 = BP021481.D			RRF005 = BP021482.D		RRF010 = BP021483.D	
		RRF020 = BP021484.D			RRF040 = BP021485.D		RRF050 = BP021486.D	
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
Benzo(k) fluoranthene		1.115	1.101	1.124	1.141	1.103	1.121	1.7
Benzo(a)pyrene		0.953	0.953	1.003	1.012	0.996	0.997	3.3
Indeno(1,2,3-cd)pyrene		1.244	1.156	1.317	1.327	1.297	1.296	5.8
Dibenzo(a,h)anthracene		1.034	0.965	1.106	1.102	1.080	1.077	5.5
Benzo(g,h,i)perylene		1.052	0.955	1.096	1.102	1.074	1.080	5.9
1,4-Dioxane		0.685	0.614	0.681	0.642	0.599	0.639	5.4
1-Methylnaphthalene		0.742	0.735	0.725	0.697	0.673	0.703	4.4