

**DATA PACKAGE
SEMI-VOLATILE ORGANICS**

PROJECT NAME : FORMER SCHLUMBERGER SITE PRINCETON NJ

JACOBS ENGINEERING GROUP, INC.

412 Mt. Kemble Ave

Downtown Building

Morristown, NJ - 07960

Phone No: 9732670555

ORDER ID : P3657

ATTENTION : Mary I. Murphy



Laboratory Certification ID # 20012

| | | |
|--|------------|--|
| 1) SEMI-VOLATILE DATA | 2 | |
| 2) Signature Page | 4 | |
| 3) Case Narrative | 5 | |
| 4) Qualifier Page | 7 | |
| 5) Conformance/Non Conformance | 8 | |
| 6) QA Checklist | 10 | |
| 7) Chronicle | 11 | |
| 8) Hit Summary | 12 | |
| 9) QC Data Summary For SVOCMS Group6 | 13 | |
| 9.1) Deuterated Monitoring Compound Summary | 14 | |
| 9.2) LCS/LCSD Summary | 15 | |
| 9.3) Method Blank Summary | 17 | |
| 9.4) GS/MS Tune Summary | 18 | |
| 9.5) Internal Standard Area and RT Summary | 21 | |
| 10) Sample Data | 25 | |
| 10.1) 917-J-WS-081624 | 26 | |
| 11) Calibration Data Summary | 36 | |
| 11.1) Initial Calibration Data | 37 | |
| 11.1.1) BP081324 | 37 | |
| 11.2) Continued Calibration Data | 444 | |
| 11.2.1) BP021566.D | 444 | |
| 11.2.2) BP021575.D | 501 | |
| 12) QC Sample Data | 558 | |
| 12.1) Tune Raw Data | 559 | |
| 12.2) Method Blank Data | 571 | |
| 12.3) LCS Data | 581 | |
| 12.4) LCSD Data | 632 | |
| 13) Manual Integration | 683 | |
| 14) Analytical Runlogs | 687 | |
| 15) Extraction Logs | 693 | |
| 15.1) PB162822.pdf | 693 | |
| 15.2) PB162822IC.pdf | 695 | |
| 16) Standard Prep Logs | 696 | |
| 17) Shipping Document | 855 | |
| 17.1) Chain Of Custody | 856 | |

Table Of Contents for P3657

| | | |
|---------------------------|-----|----|
| | 861 | 1 |
| | 862 | 2 |
| | | 3 |
| | | 4 |
| | | 5 |
| | | 6 |
| | | 7 |
| | | 8 |
| | | 9 |
| | | 10 |
| | | 11 |
| | | 12 |
| | | 13 |
| | | 14 |
| | | 15 |
| | | 16 |
| | | 17 |
| 17.3) Internal COC | | |

Cover Page

Order ID : P3657

Project ID : Former Schlumberger Site Princeton NJ

Client : JACOBS Engineering Group, Inc.

Lab Sample Number

P3657-01
P3657-02

Client Sample Number

917-J-WS-081624
TB-01-081624

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : _____

Date: 10/15/2024

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012

CASE NARRATIVE

JACOBS Engineering Group, Inc.

Project Name: Former Schlumberger Site Princeton NJ

Project # N/A

Chemtech Project # P3657

Test Name: SVOCMS Group6

A. Number of Samples and Date of Receipt:

2 Water samples were received on 08/16/2024.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: Hexavalent Chromium, Mercury, Metals Group4, SVOCMS Group3, SVOCMS Group6 and VOCMS Group6. This data package contains results for SVOCMS Group6.

C. Analytical Techniques:

The samples were analyzed on instrument BNA_P using GC Column ZB-SemiVolatile Guardian which is 30 meters, 0.25 mm ID, 0.5 um df, Catalog # 7HG-G027-17-GGAThe analysis of SVOCMS Group6 was based on method 8270E and extraction was done based on method 3510.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria except for 917-J-WS-081624 [2,4,6-Tribromophenol - 119%], this compound did not meet the NJDKQP criteria but met the in-house criteria.

The Internal Standards Areas met the acceptable requirements.

The Retention Times were acceptable for all samples.

The RPD for {PB162822BSD} with File ID: BP021569.D met criteria except for Benzaldehyde[24%], this compound did not meet the NJDKQP criteria and in-house criteria due to difference in results of BS and BSD.

The Blank Spike met requirements for all samples .

The Blank Spike Duplicate met requirements for all samples .

The Blank analysis did not indicate the presence of lab contamination.

The % RSD is greater than 15% in the Initial Calibration (8270-BP081324.M) for 2,4-Dinitrotoluene, this compound is passing on Linear Regression.

The Continuous Calibration File ID BP021566.D met the requirements except for 2,4-Dinitrotoluene and Pentachlorophenol but associated QC within limits therefore no corrective action taken.

The Continuous Calibration File ID BP021575.D met the requirements except for 2,4-Dinitrotoluene and Pentachlorophenol but associated sample having no positive hit therefore no corrective action taken.

The Tuning criteria met requirements.

E. Additional Comments:

This data package has been revised due to parameter list changed.

The Form 6 is not included in the data package because the Initial Calibration was performed using 8 points.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <15% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 15% for the Initial Calibration curve for SW-846 analysis.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature _____

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following "Results Qualifiers" are used:

- Value** If the result is a value greater than or equal to the detection limit, report the value
- U** Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. "10 U". This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
- ND** Indicates the analyte was analyzed for, but not detected
- J** Indicates an estimated value. This flag is used:
(1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.)
(2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
- B** Indicates the analyte was found in the blank as well as the sample report as "12 B".
- E** Indicates the analyte's concentration exceeds the calibrated range of the instrument for that specific analysis.
- D** This flag identifies all compounds identified in an analysis at a secondary dilution factor.
- P** This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a "P".
- N** This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
- A** This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
- Q** Indicates the LCS did not meet the control limits requirements

ALLIANCE 284 Sheffield Street, Mountainside New Jersey 07092

NEW JERSEY LAB ID#: 20012: NEW YORK LAB ID#: 11376

GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NUMBER: P3657

MATRIX: Water

METHOD: 8270E/3510

| | NA | NO | YES |
|---|----|----|-----|
| 1. Chromatograms Labeled/Compounds Identified. (Field samples and Method Blanks) | | | ✓ |
| 2. GC/MS Tuning Specifications. DFTPP Meet Criteria. (NOTE THAT THERE ARE DIFFERENT CRITERIA FOR NY ASP CLP, CLP AND NJ) | | | ✓ |
| 3. GC/MS Tuning Frequency - Performed every 24 hours for 600 series and 12 hours for 8000 Series. | | | ✓ |
| 4. GC/MS Calibration - Initial Calibration performed within 30 days before sample analysis and continuing calibration performed within 24 hours of sample analysis for 600 series and 12 hours for 8000 series. | | | ✓ |
| 5. GC/MS Calibration Requirements. | | | ✓ |

The % RSD is greater than 15% in the Initial Calibration (8270-BP081324.M) for 2,4-Dinitrotoluene, this compound is passing on Linear Regression.

The Continuous Calibration File ID BP021566.D met the requirements except for 2,4-Dinitrotoluene and Pentachlorophenol but associated QC within limits therefore no corrective action taken.

The Continuous Calibration File ID BP021575.D met the requirements except for 2,4-Dinitrotoluene and Pentachlorophenol but associated sample having no positive hit therefore no corrective action taken.

| | |
|---|---|
| 6. Blank Contamination - If yes, list compounds and concentrations in each blank: | ✓ |
| 7. Surrogate Recoveries Meet Criteria | ✓ |

If not met, list those compounds and their recoveries which fall outside the acceptable ranges.

The Surrogate recoveries met the acceptable criteria except for 917-J-WS-081624 [2,4 and 6-Tribromophenol - 119%], this compound did not meet the NJDKQP criteria but met the in-house criteria.

| | |
|--|---|
| 8. Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria | ✓ |
| If not met, list those compounds and their recoveries which fall outside the acceptable range. | |

The Blank Spike met requirements for all samples .
The Blank Spike Duplicate met requirements for all samples .

ALLIANCE 284 Sheffield Street, Mountainside New Jersey 07092

NEW JERSEY LAB ID#: 20012: NEW YORK LAB ID#: 11376

GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY**(CONTINUED)**

| | | NA | NO | YES |
|-----|---|----|----|-----|
| 9. | Internal Standard Area/Retention Time Shift Meet Criteria | | | ✓ |
| | Comments: | | | |
| 10. | Extraction Holding Time Met | | | ✓ |
| | If not met, list number of days exceeded for each sample: | | | |
| 11. | Analysis Holding Time Met | | | ✓ |
| | If not met, list number of days exceeded for each sample: | | | |

ADDITIONAL COMMENTS:

This data package has been revised due to parameter list changed.

The Form 6 is not included in the data package because the Initial Calibration was performed using 8 points.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <15% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 15% for the Initial Calibration curve for SW-846 analysis.

QA REVIEW

Date

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: P3657

Completed

For thorough review, the report must have the following:

GENERAL:

Are all original paperwork present (chain of custody, record of communication, airbill, sample management lab chronicle, login page) ✓

Check chain-of-custody for proper relinquish/return of samples ✓

Is the chain of custody signed and complete ✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts ✓

Collect information for each project id from server. Were all requirements followed ✓

COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page ✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody ✓

CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results ✓

Do requested analyses on Chain of Custody agree with the log-in page ✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Castody ✓

Were the samples received within hold time ✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle ✓

ANALYTICAL:

Was method requirement followed? ✓

Was client requirement followed? ✓

Does the case narrative summarize all QC failure? ✓

All runlogs and manual integration are reviewed for requirements ✓

All manual calculations and /or hand notations verified ✓

QA Review Signature: SOHIL JODHANI

Date: 10/15/2024

LAB CHRONICLE

| | | | |
|-----------------|--------------------------------|-------------------|---------------------------------------|
| OrderID: | P3657 | OrderDate: | 8/16/2024 2:45:00 PM |
| Client: | JACOBS Engineering Group, Inc. | Project: | Former Schlumberger Site Princeton NJ |
| Contact: | Mary I. Murphy | Location: | G11,VOA Ref. #3 Water |

| LabID | ClientID | Matrix | Test | Method | Sample Date | Prep Date | Anal Date | Received |
|-----------------|------------------------|--------------|---------------|---------------|-----------------|-----------|-----------|-----------------|
| P3657-01 | 917-J-WS-081624 | Water | | | 08/16/24 | | | 08/16/24 |
| | | | SVOCMS Group3 | 8270-Modified | 08/19/24 | 08/20/24 | | |
| | | | SVOCMS Group6 | 8270E | 08/19/24 | 08/21/24 | | |



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet

SW-846

SDG No.: P3657

Client: JACOBS Engineering Group, Inc.



QC SUMMARY

Surrogate Summary

SW-846

SDG No.: P3657

Client: JACOBS Engineering Group, Inc.

Analytical Method: 8270E

| Lab Sample ID | Client ID | Parameter | Spike (PPM) | Result (PPM) | Recovery (%) | Qual | Limits (%) | |
|---------------|-----------------|----------------------|-------------|--------------|--------------|------|------------|-----------|
| | | | | | | | Low | High |
| P3657-01 | 917-J-WS-081624 | 2-Fluorophenol | 150 | 52.1 | 35 | | 15 (10) | 110 (139) |
| | | Phenol-d6 | 150 | 32.9 | 22 | | 15 (10) | 110 (134) |
| | | Nitrobenzene-d5 | 100 | 82.0 | 82 | | 30 (49) | 130 (133) |
| | | 2-Fluorobiphenyl | 100 | 76.3 | 76 | | 30 (52) | 130 (132) |
| | | 2,4,6-Tribromophenol | 150 | 179 | 119 | * | 15 (44) | 110 (137) |
| | | Terphenyl-d14 | 100 | 96.0 | 96 | | 30 (48) | 130 (125) |
| PB162822BL | PB162822BL | 2-Fluorophenol | 150 | 130 | 87 | | 15 (10) | 110 (139) |
| | | Phenol-d6 | 150 | 120 | 80 | | 15 (10) | 110 (134) |
| | | Nitrobenzene-d5 | 100 | 90.5 | 91 | | 30 (49) | 130 (133) |
| | | 2-Fluorobiphenyl | 100 | 78.7 | 79 | | 30 (52) | 130 (132) |
| | | 2,4,6-Tribromophenol | 150 | 158 | 105 | | 15 (44) | 110 (137) |
| | | Terphenyl-d14 | 100 | 82.4 | 82 | | 30 (48) | 130 (125) |
| PB162822BS | PB162822BS | 2-Fluorophenol | 150 | 139 | 93 | | 15 (10) | 110 (139) |
| | | Phenol-d6 | 150 | 131 | 87 | | 15 (10) | 110 (134) |
| | | Nitrobenzene-d5 | 100 | 93.7 | 94 | | 30 (49) | 130 (133) |
| | | 2-Fluorobiphenyl | 100 | 79.6 | 80 | | 30 (52) | 130 (132) |
| | | 2,4,6-Tribromophenol | 150 | 157 | 105 | | 15 (44) | 110 (137) |
| | | Terphenyl-d14 | 100 | 85.2 | 85 | | 30 (48) | 130 (125) |
| PB162822BSD | PB162822BSD | 2-Fluorophenol | 150 | 137 | 91 | | 15 (10) | 110 (139) |
| | | Phenol-d6 | 150 | 128 | 85 | | 15 (10) | 110 (134) |
| | | Nitrobenzene-d5 | 100 | 93.0 | 93 | | 30 (49) | 130 (133) |
| | | 2-Fluorobiphenyl | 100 | 80.2 | 80 | | 30 (52) | 130 (132) |
| | | 2,4,6-Tribromophenol | 150 | 154 | 103 | | 15 (44) | 110 (137) |
| | | Terphenyl-d14 | 100 | 88.7 | 89 | | 30 (48) | 130 (125) |

() = LABORATORY INHOUSE LIMIT

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P3657

Client: JACOBS Engineering Group, Inc.

Analytical Method: 8270E

DataFile: BP021568.D

| Lab Sample ID | Parameter | Spike | Result | Unit | Rec | RPD | Qual | Qual | Limits | | |
|---------------|----------------------------|-------|--------|------|-----|-----|------|------|---------|-----------|--|
| | | | | | | | | | Low | High | |
| PB162822BS | Pyridine | 50 | 28.7 | ug/L | 57 | | | | 20 (29) | 160 (97) | |
| | Benzaldehyde | 50 | 16.8 | ug/L | 34 | | | | 20 (10) | 160 (162) | |
| | 2-Methylphenol | 50 | 49.7 | ug/L | 99 | | | | 70 (69) | 130 (109) | |
| | 3+4-Methylphenols | 50 | 51.5 | ug/L | 103 | | | | 20 (67) | 160 (106) | |
| | Hexachloroethane | 50 | 45.4 | ug/L | 91 | | | | 20 (76) | 160 (118) | |
| | Nitrobenzene | 50 | 44.7 | ug/L | 89 | | | | 70 (58) | 130 (106) | |
| | Naphthalene | 50 | 43.9 | ug/L | 88 | | | | 70 (64) | 130 (107) | |
| | Hexachlorobutadiene | 50 | 41.5 | ug/L | 83 | | | | 70 (69) | 130 (101) | |
| | 2-Methylnaphthalene | 50 | 44.8 | ug/L | 90 | | | | 70 (64) | 130 (107) | |
| | 2,4,6-Trichlorophenol | 50 | 57.1 | ug/L | 114 | | | | 70 (61) | 130 (110) | |
| | 2,4,5-Trichlorophenol | 50 | 55.3 | ug/L | 111 | | | | 70 (70) | 130 (106) | |
| | Acenaphthylene | 50 | 48.9 | ug/L | 98 | | | | 70 (79) | 130 (103) | |
| | Acenaphthene | 50 | 51.6 | ug/L | 103 | | | | 70 (59) | 130 (113) | |
| | Dibenzofuran | 50 | 45.7 | ug/L | 91 | | | | 70 (65) | 130 (106) | |
| | 2,4-Dinitrotoluene | 50 | 58.1 | ug/L | 116 | | | | 70 (60) | 130 (115) | |
| | Fluorene | 50 | 43.7 | ug/L | 87 | | | | 70 (64) | 130 (107) | |
| | Hexachlorobenzene | 50 | 44.5 | ug/L | 89 | | | | 70 (73) | 130 (106) | |
| | Pentachlorophenol | 100 | 110 | ug/L | 110 | | | | 20 (47) | 160 (114) | |
| | Phenanthrene | 50 | 46.6 | ug/L | 93 | | | | 70 (62) | 130 (109) | |
| | Anthracene | 50 | 47.6 | ug/L | 95 | | | | 70 (65) | 130 (110) | |
| | Carbazole | 50 | 47.8 | ug/L | 96 | | | | 70 (62) | 130 (106) | |
| | Di-n-butylphthalate | 50 | 48.4 | ug/L | 97 | | | | 70 (64) | 130 (106) | |
| | Fluoranthene | 50 | 46.1 | ug/L | 92 | | | | 70 (64) | 130 (110) | |
| | Pyrene | 50 | 48.5 | ug/L | 97 | | | | 70 (71) | 130 (103) | |
| | Benzo(a)anthracene | 50 | 47.9 | ug/L | 96 | | | | 70 (62) | 130 (107) | |
| | Chrysene | 50 | 48.0 | ug/L | 96 | | | | 70 (61) | 130 (108) | |
| | bis(2-Ethylhexyl)phthalate | 50 | 54.2 | ug/L | 108 | | | | 70 (59) | 130 (110) | |
| | Benzo(b)fluoranthene | 50 | 51.7 | ug/L | 103 | | | | 70 (77) | 130 (113) | |
| | Benzo(k)fluoranthene | 50 | 50.4 | ug/L | 101 | | | | 70 (77) | 130 (105) | |
| | Benzo(a)pyrene | 50 | 54.3 | ug/L | 109 | | | | 70 (72) | 130 (131) | |
| | Indeno(1,2,3-cd)pyrene | 50 | 54.2 | ug/L | 108 | | | | 70 (72) | 130 (105) | |
| | Dibenz(a,h)anthracene | 50 | 53.3 | ug/L | 107 | | | | 70 (78) | 130 (115) | |
| | Benzo(g,h,i)perylene | 50 | 51.4 | ug/L | 103 | | | | 70 (75) | 130 (118) | |
| | 1,4-Dioxane | 50 | 28.5 | ug/L | 57 | | | | 20 (38) | 160 (125) | |
| | 1-Methylnaphthalene | 50 | 42.7 | ug/L | 85 | | | | 70 (51) | 130 (114) | |

() = LABORATORY INHOUSE LIMIT

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P3657

Client: JACOBS Engineering Group, Inc.

Analytical Method: 8270E

DataFile: BP021569.D

| Lab Sample ID | Parameter | Spike | Result | Unit | Rec | RPD | Qual | RPD | Limits | | |
|---------------|----------------------------|-------|--------|------|-----|-----|------|-----|---------|-----------|---------|
| | | | | | | | | | Low | High | RPD |
| PB162822BSD | Pyridine | 50 | 28.8 | ug/L | 58 | 0 | * | | 20 (29) | 160 (97) | 20 (20) |
| | Benzaldehyde | 50 | 21.4 | ug/L | 43 | 24 | | | 20 (10) | 160 (162) | 20 (20) |
| | 2-Methylphenol | 50 | 48.7 | ug/L | 97 | 2 | | | 70 (69) | 130 (109) | 20 (20) |
| | 3+4-Methylphenols | 50 | 50.1 | ug/L | 100 | 3 | | | 20 (67) | 160 (106) | 20 (20) |
| | Hexachloroethane | 50 | 44.9 | ug/L | 90 | 1 | | | 20 (76) | 160 (118) | 20 (20) |
| | Nitrobenzene | 50 | 44.7 | ug/L | 89 | 0 | | | 70 (58) | 130 (106) | 20 (20) |
| | Naphthalene | 50 | 43.7 | ug/L | 87 | 0 | | | 70 (64) | 130 (107) | 20 (20) |
| | Hexachlorobutadiene | 50 | 41.6 | ug/L | 83 | 0 | | | 70 (69) | 130 (101) | 20 (20) |
| | 2-Methylnaphthalene | 50 | 44.2 | ug/L | 88 | 1 | | | 70 (64) | 130 (107) | 20 (20) |
| | 2,4,6-Trichlorophenol | 50 | 57.4 | ug/L | 115 | 1 | | | 70 (61) | 130 (110) | 20 (20) |
| | 2,4,5-Trichlorophenol | 50 | 55.2 | ug/L | 110 | 0 | | | 70 (70) | 130 (106) | 20 (20) |
| | Acenaphthylene | 50 | 49.4 | ug/L | 99 | 1 | | | 70 (79) | 130 (103) | 20 (20) |
| | Acenaphthene | 50 | 51.4 | ug/L | 103 | 0 | | | 70 (59) | 130 (113) | 20 (20) |
| | Dibenzofuran | 50 | 46.3 | ug/L | 93 | 1 | | | 70 (65) | 130 (106) | 20 (20) |
| | 2,4-Dinitrotoluene | 50 | 57.8 | ug/L | 116 | 1 | | | 70 (60) | 130 (115) | 20 (20) |
| | Fluorene | 50 | 43.6 | ug/L | 87 | 0 | | | 70 (64) | 130 (107) | 20 (20) |
| | Hexachlorobenzene | 50 | 45.0 | ug/L | 90 | 1 | | | 70 (73) | 130 (106) | 20 (20) |
| | Pentachlorophenol | 100 | 110 | ug/L | 110 | 0 | | | 20 (47) | 160 (114) | 20 (20) |
| | Phenanthrene | 50 | 47.6 | ug/L | 95 | 2 | | | 70 (62) | 130 (109) | 20 (20) |
| | Anthracene | 50 | 48.7 | ug/L | 97 | 2 | | | 70 (65) | 130 (110) | 20 (20) |
| | Carbazole | 50 | 48.9 | ug/L | 98 | 2 | | | 70 (62) | 130 (106) | 20 (20) |
| | Di-n-butylphthalate | 50 | 48.1 | ug/L | 96 | 1 | | | 70 (64) | 130 (106) | 20 (20) |
| | Fluoranthene | 50 | 46.4 | ug/L | 93 | 1 | | | 70 (64) | 130 (110) | 20 (20) |
| | Pyrene | 50 | 48.0 | ug/L | 96 | 1 | | | 70 (71) | 130 (103) | 20 (20) |
| | Benzo(a)anthracene | 50 | 48.7 | ug/L | 97 | 2 | | | 70 (62) | 130 (107) | 20 (20) |
| | Chrysene | 50 | 48.8 | ug/L | 98 | 2 | | | 70 (61) | 130 (108) | 20 (20) |
| | bis(2-Ethylhexyl)phthalate | 50 | 53.5 | ug/L | 107 | 1 | | | 70 (59) | 130 (110) | 20 (20) |
| | Benzo(b)fluoranthene | 50 | 49.6 | ug/L | 99 | 4 | | | 70 (77) | 130 (113) | 20 (20) |
| | Benzo(k)fluoranthene | 50 | 51.0 | ug/L | 102 | 1 | | | 70 (77) | 130 (105) | 20 (20) |
| | Benzo(a)pyrene | 50 | 54.0 | ug/L | 108 | 1 | | | 70 (72) | 130 (131) | 20 (20) |
| | Indeno(1,2,3-cd)pyrene | 50 | 53.9 | ug/L | 108 | 1 | | | 70 (72) | 130 (105) | 20 (20) |
| | Dibenz(a,h)anthracene | 50 | 53.3 | ug/L | 107 | 0 | | | 70 (78) | 130 (115) | 20 (20) |
| | Benzo(g,h,i)perylene | 50 | 50.9 | ug/L | 102 | 1 | | | 70 (75) | 130 (118) | 20 (20) |
| | 1,4-Dioxane | 50 | 28.8 | ug/L | 58 | 1 | | | 20 (38) | 160 (125) | 20 (20) |
| | 1-Methylnaphthalene | 50 | 42.0 | ug/L | 84 | 2 | | | 70 (51) | 130 (114) | 20 (20) |

() = LABORATORY INHOUSE LIMIT

4B

SEMIVOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

PB162822BL

Lab Name: CHEMTECH

Contract: JACO05

Lab Code: CHEM Case No.: P3657

SAS No.: P3657 SDG NO.: P3657

Lab File ID: BP021567.D

Lab Sample ID: PB162822BL

Instrument ID: BNA_P

Date Extracted: 08/19/2024

Matrix: (soil/water) Water

Date Analyzed: 08/20/2024

Level: (low/med) LOW

Time Analyzed: 10:49

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

| EPA SAMPLE NO. | LAB SAMPLE ID | LAB FILE ID | DATE ANALYZED |
|-------------------|------------------|----------------|------------------|
| PB162822BS | PB162822BS | BP021568.D | 08/20/2024 |
| PB162822BSD | PB162822BSD | BP021569.D | 08/20/2024 |
| 917-J-WS-081624 | P3657-01 | BP021589.D | 08/21/2024 |

COMMENTS:



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECHContract: JAC005Lab Code: CHEMSAS No.: P3657 SDG NO.: P3657Lab File ID: BP021480.DDFTPP Injection Date: 08/13/2024Instrument ID: BNA_PDFTPP Injection Time: 11:49

| m/e | ION ABUNDANCE CRITERIA | % RELATIVE ABUNDANCE |
|-----|------------------------------------|----------------------|
| 51 | 10.0 - 80.0% of mass 198 | 30.1 |
| 68 | Less than 2.0% of mass 69 | 0.6 (1.5) 1 |
| 69 | Mass 69 relative abundance | 40.6 |
| 70 | Less than 2.0% of mass 69 | 0.2 (0.4) 1 |
| 127 | 10.0 - 80.0% of mass 198 | 43.9 |
| 197 | Less than 2.0% of mass 198 | 0.5 |
| 198 | Base Peak, 100% relative abundance | 100 |
| 199 | 5.0 to 9.0% of mass 198 | 6.6 |
| 275 | 10.0 - 60.0% of mass 198 | 28.5 |
| 365 | Greater than 1% of mass 198 | 3.5 |
| 441 | Present, but less than mass 443 | 15 |
| 442 | Greater than 50% of mass 198 | 100 |
| 443 | 15.0 - 24.0% of mass 442 | 18.7 (19.3) 2 |

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

| EPA SAMPLE NO. | LAB SAMPLE ID | LAB FILE ID | DATE ANALYZED | TIME ANALYZED |
|-------------------|------------------|----------------|------------------|------------------|
| SSTDICC2.5 | SSTDICC2.5 | BP021481.D | 08/13/2024 | 12:32 |
| SSTDICC005 | SSTDICC005 | BP021482.D | 08/13/2024 | 13:15 |
| SSTDICC010 | SSTDICC010 | BP021483.D | 08/13/2024 | 13:59 |
| SSTDICC020 | SSTDICC020 | BP021484.D | 08/13/2024 | 14:42 |
| SSTDICCC040 | SSTDICCC040 | BP021485.D | 08/13/2024 | 15:26 |
| SSTDICC050 | SSTDICC050 | BP021486.D | 08/13/2024 | 16:09 |
| SSTDICC060 | SSTDICC060 | BP021487.D | 08/13/2024 | 16:53 |
| SSTDICC080 | SSTDICC080 | BP021488.D | 08/13/2024 | 17:36 |



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: JAC005

Lab Code: CHEM

SAS No.: P3657 SDG NO.: P3657

Lab File ID: BP021565.D

DFTPP Injection Date: 08/20/2024

Instrument ID: BNA_P

DFTPP Injection Time: 09:25

| m/e | ION ABUNDANCE CRITERIA | % RELATIVE ABUNDANCE |
|-----|------------------------------------|----------------------|
| 51 | 10.0 - 80.0% of mass 198 | 29.2 |
| 68 | Less than 2.0% of mass 69 | 0.6 (1.6) 1 |
| 69 | Mass 69 relative abundance | 36.7 |
| 70 | Less than 2.0% of mass 69 | 0.2 (0.5) 1 |
| 127 | 10.0 - 80.0% of mass 198 | 44.9 |
| 197 | Less than 2.0% of mass 198 | 0.5 |
| 198 | Base Peak, 100% relative abundance | 100 |
| 199 | 5.0 to 9.0% of mass 198 | 6.7 |
| 275 | 10.0 - 60.0% of mass 198 | 28.9 |
| 365 | Greater than 1% of mass 198 | 3.8 |
| 441 | Present, but less than mass 443 | 14.4 |
| 442 | Greater than 50% of mass 198 | 100 |
| 443 | 15.0 - 24.0% of mass 442 | 18.2 (19.1) 2 |

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

| EPA SAMPLE NO. | LAB SAMPLE ID | LAB FILE ID | DATE ANALYZED | TIME ANALYZED |
|-------------------|------------------|----------------|------------------|------------------|
| SSTDCCC040 | SSTDCCC040 | BP021566.D | 08/20/2024 | 10:07 |
| PB162822BL | PB162822BL | BP021567.D | 08/20/2024 | 10:49 |
| PB162822BS | PB162822BS | BP021568.D | 08/20/2024 | 11:31 |
| PB162822BSD | PB162822BSD | BP021569.D | 08/20/2024 | 12:13 |



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

5B

SEMICVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: JAC005

Lab Code: CHEM

SAS No.: P3657 SDG NO.: P3657

Lab File ID: BP021574.D

DFTPP Injection Date: 08/20/2024

Instrument ID: BNA_P

DFTPP Injection Time: 15:57

| m/e | ION ABUNDANCE CRITERIA | % RELATIVE ABUNDANCE |
|-----|------------------------------------|----------------------|
| 51 | 10.0 - 80.0% of mass 198 | 28.5 |
| 68 | Less than 2.0% of mass 69 | 0.6 (1.6) 1 |
| 69 | Mass 69 relative abundance | 36.5 |
| 70 | Less than 2.0% of mass 69 | 0.1 (0.4) 1 |
| 127 | 10.0 - 80.0% of mass 198 | 44.3 |
| 197 | Less than 2.0% of mass 198 | 0.3 |
| 198 | Base Peak, 100% relative abundance | 100 |
| 199 | 5.0 to 9.0% of mass 198 | 6.8 |
| 275 | 10.0 - 60.0% of mass 198 | 30.1 |
| 365 | Greater than 1% of mass 198 | 4 |
| 441 | Present, but less than mass 443 | 15.2 |
| 442 | Greater than 50% of mass 198 | 100 |
| 443 | 15.0 - 24.0% of mass 442 | 18.8 (19.1) 2 |

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

| EPA SAMPLE NO. | LAB SAMPLE ID | LAB FILE ID | DATE ANALYZED | TIME ANALYZED |
|-------------------|------------------|----------------|------------------|------------------|
| SSTDCCC040 | SSTDCCC040 | BP021575.D | 08/20/2024 | 16:40 |
| 917-J-WS-081624 | P3657-01 | BP021589.D | 08/21/2024 | 02:44 |



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

8B

SEMOVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH
Lab Code: CHEM Case No.: P3657 SAS No.: P3657 SDG No.: P3657
EPA Sample No.: SSTDCCC040 Date Analyzed: 08/20/2024
Lab File ID: BP021566.D Time Analyzed: 10:07
Instrument ID: BNA_P GC Column: ZB-GR ID: 0.25 (mm)

| | IS1 (DCB) AREA # | RT # | IS2 (NPT) AREA # | RT # | IS3 (ANT) AREA # | RT # |
|----------------|---------------------|-------|---------------------|--------|---------------------|--------|
| 12 HOUR STD | 357434 | 7.805 | 1454010 | 10.59 | 971251 | 14.45 |
| UPPER LIMIT | 714868 | 8.305 | 2908020 | 11.093 | 1942500 | 14.951 |
| LOWER LIMIT | 178717 | 7.305 | 727005 | 10.093 | 485626 | 13.951 |
| EPA SAMPLE NO. | | | | | | |
| 01 PB162822BL | 385209 | 7.81 | 1560540 | 10.59 | 1021930 | 14.45 |
| 02 PB162822BS | 494651 | 7.81 | 2073180 | 10.60 | 1389380 | 14.45 |
| 03 PB162822BSD | 491039 | 7.81 | 2018200 | 10.59 | 1318740 | 14.46 |

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

| | | | | | | | |
|-----------------|------------|-----------|----------------|------------|-------|-----------|-------|
| Lab Name: | CHEMTECH | | | | | | |
| Lab Code: | CHEM | Case No.: | P3657 | SAS No.: | P3657 | SDG NO.: | P3657 |
| EPA Sample No.: | SSTDCCC040 | | Date Analyzed: | 08/20/2024 | | | |
| Lab File ID: | BP021566.D | | Time Analyzed: | 10:07 | | | |
| Instrument ID: | BNA_P | | GC Column: | ZB-GR | ID: | 0.25 (mm) | |

| | IS4 (PHN) AREA # | RT # | IS5 (CRY) AREA # | RT # | IS6 (PRY) AREA # | RT # |
|----------------|---------------------|--------|---------------------|--------|---------------------|--------|
| 12 HOUR STD | 2102940 | 17.263 | 2067040 | 21.727 | 2510090 | 25.174 |
| | 4205880 | 17.763 | 4134080 | 22.227 | 5020180 | 25.674 |
| | 1051470 | 16.763 | 1033520 | 21.227 | 1255050 | 24.674 |
| EPA SAMPLE NO. | | | | | | |
| 01 PB162822BL | 2408800 | 17.26 | 2576180 | 21.72 | 2892860 | 25.18 |
| 02 PB162822BS | 3038380 | 17.26 | 2684400 | 21.73 | 2880700 | 25.17 |
| 03 PB162822BSD | 2796130 | 17.27 | 2515500 | 21.74 | 2791670 | 25.19 |

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

8B

SEMOVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH
Lab Code: CHEM Case No.: P3657 SAS No.: P3657 SDG NO.: P3657
EPA Sample No.: SSTDCCC040 Date Analyzed: 08/20/2024
Lab File ID: BP021575.D Time Analyzed: 16:40
Instrument ID: BNA_P GC Column: ZB-GR ID: 0.25 (mm)

| | IS1 (DCB) AREA # | RT # | IS2 (NPT) AREA # | RT # | IS3 (ANT) AREA # | RT # |
|----------------|---------------------|--------|---------------------|---------|---------------------|--------|
| 12 HOUR STD | 372530 | 7.805 | 1563750 | 10.59 | 1068850 | 14.46 |
| UPPER LIMIT | 745060 | 8.305 | 3127500 | 11.093 | 2137700 | 14.963 |
| LOWER LIMIT | 186265 | 7.305 | 781875 | 10.093 | 534425 | 13.963 |
| EPA SAMPLE NO. | | | | | | |
| 01 | 917-J-WS-081624 | 352980 | 7.81 | 1458860 | 10.60 | 956855 |
| | | | | | | 14.46 |

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

| | | | | | | | |
|-----------------|------------|-----------|----------------|------------|-------|-----------|-------|
| Lab Name: | CHEMTECH | | | | | | |
| Lab Code: | CHEM | Case No.: | P3657 | SAS No.: | P3657 | SDG NO.: | P3657 |
| EPA Sample No.: | SSTDCCC040 | | Date Analyzed: | 08/20/2024 | | | |
| Lab File ID: | BP021575.D | | Time Analyzed: | 16:40 | | | |
| Instrument ID: | BNA_P | | GC Column: | ZB-GR | ID: | 0.25 (mm) | |

| | IS4 (PHN) AREA # | RT # | IS5 (CRY) AREA # | RT # | IS6 (PRY) AREA # | RT # |
|--------------------|---------------------|--------|---------------------|--------|---------------------|-------|
| 12 HOUR STD | 2403930 | 17.275 | 2426350 | 21.739 | 2940310 | 25.18 |
| | 4807860 | 17.775 | 4852700 | 22.239 | 5880620 | 25.68 |
| | 1201970 | 16.775 | 1213180 | 21.239 | 1470160 | 24.68 |
| EPA SAMPLE NO. | | | | | | |
| 01 917-J-WS-081624 | 2099850 | 17.27 | 1963860 | 21.73 | 2261180 | 25.17 |

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.



SAMPLE

DATA



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

| | | | | | | |
|--------------------|---------------------------------------|--------------|-----|-----------------|---------------|------|
| Client: | JACOBS Engineering Group, Inc. | | | Date Collected: | 08/16/24 | |
| Project: | Former Schlumberger Site Princeton NJ | | | Date Received: | 08/16/24 | |
| Client Sample ID: | 917-J-WS-081624 | | | SDG No.: | P3657 | |
| Lab Sample ID: | P3657-01 | | | Matrix: | Water | |
| Analytical Method: | SW8270 | | | % Solid: | 0 | |
| Sample Wt/Vol: | 960 | Units: | mL | Final Vol: | 1000 | uL |
| Soil Aliquot Vol: | | | uL | Test: | SVOCMS Group6 | |
| Extraction Type : | | Decanted : | N | Level : | LOW | |
| Injection Volume : | | GPC Factor : | 1.0 | GPC Cleanup : | N | PH : |
| Prep Method : | SW3510C | | | | | |

| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed | Prep Batch ID |
|-------------------|-----------|----------------|----------------|---------------|
| BP021589.D | 1 | 08/19/24 09:40 | 08/21/24 02:44 | PB162822 |

| CAS Number | Parameter | Conc. | Qualifier | MDL | LOQ / CRQL | Units |
|----------------|----------------------------|-------|-----------|------|------------|-------|
| TARGETS | | | | | | |
| 110-86-1 | Pyridine | 1.60 | U | 1.60 | 5.20 | ug/L |
| 100-52-7 | Benzaldehyde | 4.20 | U | 4.20 | 10.4 | ug/L |
| 95-48-7 | 2-Methylphenol | 1.20 | U | 1.20 | 5.20 | ug/L |
| 65794-96-9 | 3+4-Methylphenols | 1.20 | U | 1.20 | 10.4 | ug/L |
| 67-72-1 | Hexachloroethane | 1.10 | U | 1.10 | 5.20 | ug/L |
| 98-95-3 | Nitrobenzene | 1.30 | U | 1.30 | 5.20 | ug/L |
| 91-20-3 | Naphthalene | 1.10 | U | 1.10 | 5.20 | ug/L |
| 87-68-3 | Hexachlorobutadiene | 1.30 | U | 1.30 | 5.20 | ug/L |
| 91-57-6 | 2-Methylnaphthalene | 1.20 | U | 1.20 | 5.20 | ug/L |
| 88-06-2 | 2,4,6-Trichlorophenol | 0.93 | U | 0.93 | 5.20 | ug/L |
| 95-95-4 | 2,4,5-Trichlorophenol | 1.10 | U | 1.10 | 5.20 | ug/L |
| 208-96-8 | Acenaphthylene | 1.10 | U | 1.10 | 5.20 | ug/L |
| 83-32-9 | Acenaphthene | 0.84 | U | 0.84 | 5.20 | ug/L |
| 132-64-9 | Dibenzofuran | 0.97 | U | 0.97 | 5.20 | ug/L |
| 121-14-2 | 2,4-Dinitrotoluene | 1.60 | U | 1.60 | 5.20 | ug/L |
| 86-73-7 | Fluorene | 1.00 | U | 1.00 | 5.20 | ug/L |
| 118-74-1 | Hexachlorobenzene | 1.20 | U | 1.20 | 5.20 | ug/L |
| 87-86-5 | Pentachlorophenol | 1.90 | U | 1.90 | 10.4 | ug/L |
| 85-01-8 | Phenanthrene | 0.93 | U | 0.93 | 5.20 | ug/L |
| 120-12-7 | Anthracene | 1.10 | U | 1.10 | 5.20 | ug/L |
| 86-74-8 | Carbazole | 1.20 | U | 1.20 | 5.20 | ug/L |
| 84-74-2 | Di-n-butylphthalate | 1.50 | U | 1.50 | 5.20 | ug/L |
| 206-44-0 | Fluoranthene | 1.30 | U | 1.30 | 5.20 | ug/L |
| 129-00-0 | Pyrene | 1.10 | U | 1.10 | 5.20 | ug/L |
| 56-55-3 | Benzo(a)anthracene | 0.98 | U | 0.98 | 5.20 | ug/L |
| 218-01-9 | Chrysene | 0.90 | U | 0.90 | 5.20 | ug/L |
| 117-81-7 | Bis(2-ethylhexyl)phthalate | 2.00 | U | 2.00 | 5.20 | ug/L |
| 205-99-2 | Benzo(b)fluoranthene | 1.20 | U | 1.20 | 5.20 | ug/L |
| 207-08-9 | Benzo(k)fluoranthene | 1.20 | U | 1.20 | 5.20 | ug/L |



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

| | | | | | | |
|--------------------|---------------------------------------|--------|----|-----------------|---------------|------|
| Client: | JACOBS Engineering Group, Inc. | | | Date Collected: | 08/16/24 | |
| Project: | Former Schlumberger Site Princeton NJ | | | Date Received: | 08/16/24 | |
| Client Sample ID: | 917-J-WS-081624 | | | SDG No.: | P3657 | |
| Lab Sample ID: | P3657-01 | | | Matrix: | Water | |
| Analytical Method: | SW8270 | | | % Solid: | 0 | |
| Sample Wt/Vol: | 960 | Units: | mL | Final Vol: | 1000 | uL |
| Soil Aliquot Vol: | uL | | | Test: | SVOCMS Group6 | |
| Extraction Type : | Decanted : N | | | Level : | LOW | |
| Injection Volume : | GPC Factor : 1.0 | | | GPC Cleanup : | N | PH : |
| Prep Method : | SW3510C | | | | | |

| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed | Prep Batch ID |
|-------------------|-----------|----------------|----------------|---------------|
| BP021589.D | 1 | 08/19/24 09:40 | 08/21/24 02:44 | PB162822 |

| CAS Number | Parameter | Conc. | Qualifier | MDL | LOQ / CRQL | Units |
|-------------------|------------------------|-------|-----------|---------------------|------------|----------|
| 50-32-8 | Benzo(a)pyrene | 1.70 | U | 1.70 | 5.20 | ug/L |
| 193-39-5 | Indeno(1,2,3-cd)pyrene | 1.10 | U | 1.10 | 5.20 | ug/L |
| 53-70-3 | Dibenzo(a,h)anthracene | 1.20 | U | 1.20 | 5.20 | ug/L |
| 191-24-2 | Benzo(g,h,i)perylene | 1.20 | U | 1.20 | 5.20 | ug/L |
| 123-91-1 | 1,4-Dioxane | 1.30 | U | 1.30 | 5.20 | ug/L |
| 90-12-0 | 1-Methylnaphthalene | 0.90 | U | 0.90 | 5.20 | ug/L |
| SURROGATES | | | | | | |
| 367-12-4 | 2-Fluorophenol | 52.1 | | 15 (10) - 110 (139) | 35% | SPK: 150 |
| 13127-88-3 | Phenol-d6 | 32.9 | | 15 (10) - 110 (134) | 22% | SPK: 150 |
| 4165-60-0 | Nitrobenzene-d5 | 82.0 | | 30 (49) - 130 (133) | 82% | SPK: 100 |
| 321-60-8 | 2-Fluorobiphenyl | 76.3 | | 30 (52) - 130 (132) | 76% | SPK: 100 |
| 118-79-6 | 2,4,6-Tribromophenol | 179 | * | 15 (44) - 110 (137) | 119% | SPK: 150 |
| 1718-51-0 | Terphenyl-d14 | 96.0 | | 30 (48) - 130 (125) | 96% | SPK: 100 |

INTERNAL STANDARDS

| | | | |
|------------|------------------------|---------|--------|
| 3855-82-1 | 1,4-Dichlorobenzene-d4 | 353000 | 7.805 |
| 1146-65-2 | Naphthalene-d8 | 1460000 | 10.599 |
| 15067-26-2 | Acenaphthene-d10 | 957000 | 14.457 |
| 1517-22-2 | Phenanthrene-d10 | 2100000 | 17.269 |
| 1719-03-5 | Chrysene-d12 | 1960000 | 21.727 |
| 1520-96-3 | Perylene-d12 | 2260000 | 25.168 |

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP082024\
 Data File : BP021589.D
 Acq On : 21 Aug 2024 02:44
 Operator : RC/JU
 Sample : P3657-01
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 917-J-WS-081624

Quant Time: Aug 21 05:40:51 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 14 00:00:30 2024
 Response via : Initial Calibration


| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|------------------------------------|--------|------|----------|---------|-------|----------|
| Internal Standards | | | | | | |
| 1) 1,4-Dichlorobenzene-d4 | 7.805 | 152 | 352980 | 20.000 | ng | 0.00 |
| 21) Naphthalene-d8 | 10.599 | 136 | 1458861 | 20.000 | ng | 0.00 |
| 39) Acenaphthene-d10 | 14.457 | 164 | 956855 | 20.000 | ng | 0.00 |
| 64) Phenanthrene-d10 | 17.269 | 188 | 2099848 | 20.000 | ng | 0.02 |
| 76) Chrysene-d12 | 21.727 | 240 | 1963856 | 20.000 | ng | 0.03 |
| 86) Perylene-d12 | 25.168 | 264 | 2261178 | 20.000 | ng | 0.06 |
| System Monitoring Compounds | | | | | | |
| 5) 2-Fluorophenol | 5.405 | 112 | 1238318 | 52.092 | ng | 0.01 |
| 7) Phenol-d6 | 6.969 | 99 | 1142521 | 32.947 | ng | 0.00 |
| 23) Nitrobenzene-d5 | 8.952 | 82 | 2318497 | 82.029 | ng | 0.00 |
| 42) 2,4,6-Tribromophenol | 15.981 | 330 | 1906699 | 179.043 | ng | 0.01 |
| 45) 2-Fluorobiphenyl | 13.069 | 172 | 4784620 | 76.313 | ng | 0.00 |
| 79) Terphenyl-d14 | 19.992 | 244 | 9705536 | 96.011 | ng | 0.02 |


| Target Compounds | Qvalue |
|--|--------|
| (#= qualifier out of range (m) = manual integration (+) = signals summed | |

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP082024\
 Data File : BP021589.D
 Acq On : 21 Aug 2024 02:44
 Operator : RC/JU
 Sample : P3657-01
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 917-J-WS-081624

Quant Time: Aug 21 05:40:51 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 14 00:00:30 2024
 Response via : Initial Calibration







#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.805 min Scan# 8
Delta R.T. -0.006 min
Lab File: BP021589.D
Acq: 21 Aug 2024 02:44


Instrument : BNA_P
ClientSampleId : 917-J-WS-081624


Tgt Ion:152 Resp: 352980
Ion Ratio Lower Upper
152 100
150 154.7 122.4 183.6
115 68.6 54.7 82.1



#5
2-Fluorophenol
Concen: 52.092 ng
RT: 5.405 min Scan# 411
Delta R.T. 0.012 min
Lab File: BP021589.D
Acq: 21 Aug 2024 02:44

Tgt Ion:112 Resp: 1238318
Ion Ratio Lower Upper
112 100
64 54.3 49.0 73.6
63 27.8 24.6 36.8







#7
 Phenol-d6
 Concen: 32.947 ng
 RT: 6.969 min Scan# 6
 Delta R.T. 0.006 min
 Lab File: BP021589.D
 Acq: 21 Aug 2024 02:44


Instrument : BNA_P
 ClientSampleId : 917-J-WS-081624


Tgt Ion: 99 Resp: 1142521
 Ion Ratio Lower Upper
 99 100
 42 12.7 10.2 15.4
 71 30.9 25.8 38.8



#21
 Naphthalene-d8
 Concen: 20.000 ng
 RT: 10.599 min Scan# 1294
 Delta R.T. -0.000 min
 Lab File: BP021589.D
 Acq: 21 Aug 2024 02:44

Tgt Ion:136 Resp: 1458861
 Ion Ratio Lower Upper
 136 100
 137 10.8 8.8 13.2
 54 7.6 6.3 9.5
 68 7.0 6.8 10.2







#23
 Nitrobenzene-d5
 Concen: 82.029 ng
 RT: 8.952 min Scan# 1
 Delta R.T. -0.000 min
 Lab File: BP021589.D
 Acq: 21 Aug 2024 02:44


Instrument : BNA_P
 ClientSampleId : 917-J-WS-081624


Tgt Ion: 82 Resp: 2318497
 Ion Ratio Lower Upper
 82 100
 128 37.0 26.5 39.7
 54 39.5 30.3 45.5



#39
 Acenaphthene-d10
 Concen: 20.000 ng
 RT: 14.457 min Scan# 1950
 Delta R.T. -0.000 min
 Lab File: BP021589.D
 Acq: 21 Aug 2024 02:44

Tgt Ion: 164 Resp: 956855
 Ion Ratio Lower Upper
 164 100
 162 101.4 81.5 122.3
 160 43.6 35.2 52.8





#42

2,4,6-Tribromophenol

Concen: 179.043 ng

RT: 15.981 min Scan# 2

Delta R.T. 0.012 min

Lab File: BP021589.D

Acq: 21 Aug 2024 02:44

Instrument :

BNA_P

ClientSampleId :

917-J-WS-081624

Tgt Ion:330 Resp: 1906699

Ion Ratio Lower Upper

330 100

332 96.7 77.4 116.0

141 29.8 24.5 36.7

Abundance



15.981

1000000

500000

0

Time-->



#45

2-Fluorobiphenyl

Concen: 76.313 ng

RT: 13.069 min Scan# 1714

Delta R.T. 0.006 min

Lab File: BP021589.D

Acq: 21 Aug 2024 02:44

Tgt Ion:172 Resp: 4784620

Ion Ratio Lower Upper

172 100

171 35.1 28.2 42.4

170 23.5 18.4 27.6

Abundance

13.069


3000000


2000000

1000000

0

Time-->







#64
 Phenanthrene-d10
 Concen: 20.000 ng
 RT: 17.269 min Scan# 2
 Delta R.T. 0.018 min
 Lab File: BP021589.D
 Acq: 21 Aug 2024 02:44


Instrument : BNA_P
 ClientSampleId : 917-J-WS-081624


Tgt Ion:188 Resp: 2099848
 Ion Ratio Lower Upper
 188 100
 94 11.5 10.2 15.4
 80 13.1 11.6 17.4



#76
 Chrysene-d12
 Concen: 20.000 ng
 RT: 21.727 min Scan# 3186
 Delta R.T. 0.029 min
 Lab File: BP021589.D
 Acq: 21 Aug 2024 02:44

Tgt Ion:240 Resp: 1963856
 Ion Ratio Lower Upper
 240 100
 120 11.5 9.9 14.9
 236 25.8 20.6 30.8







#79
Terphenyl-d14
Concen: 96.011 ng
RT: 19.992 min Scan# 2
Delta R.T. 0.017 min
Lab File: BP021589.D
Acq: 21 Aug 2024 02:44


Instrument : BNA_P
ClientSampleId : 917-J-WS-081624

Tgt Ion:244 Resp: 9705536
Ion Ratio Lower Upper
244 100
212 7.4 5.6 8.4
122 10.8 9.4 14.2



#86
Perylene-d12
Concen: 20.000 ng
RT: 25.168 min Scan# 3771
Delta R.T. 0.065 min
Lab File: BP021589.D
Acq: 21 Aug 2024 02:44

Tgt Ion:264 Resp: 2261178
Ion Ratio Lower Upper
264 100
260 23.2 18.6 28.0
265 21.8 17.1 25.7





CALIBRATION

SUMMARY

Method Path : Z:\svoasrv\HPCHEM1\BNA_P\Methods\
 Method File : 8270E-BP081324.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Aug 14 00:00:30 2024
 Response Via : Initial Calibration

Calibration Files
 2.5 =BP021481.D 5 =BP021482.D 10 =BP021483.D 20 =BP021484.D 40 =BP021485.D 50 =BP021486.D 60 =BP021487.D 80 =BP021488.D

| | Compound | 2.5 | 5 | 10 | 20 | 40 | 50 | 60 | 80 | Avg | %RSD |
|-------|------------------------------|-------|-------|-------|-------|-------|-------|-------|-------|--------|------|
| <hr/> | | | | | | | | | | | |
| 1) I | 1,4-Dichlorobenzene | | | | | | | | | | |
| 2) | 1,4-Dioxane | 0.685 | 0.614 | 0.681 | 0.642 | 0.599 | 0.645 | 0.605 | 0.639 | 5.45 | |
| 3) | Pyridine | 1.744 | 1.720 | 1.915 | 1.802 | 1.741 | 1.834 | 1.725 | 1.783 | 4.02 | |
| 4) | n-Nitrosodimethylamine | 0.545 | 0.525 | 0.568 | 0.532 | 0.516 | 0.546 | 0.519 | 0.536 | 3.41 | |
| 5) S | 2-Fluorophenol | 1.332 | 1.286 | 1.405 | 1.348 | 1.324 | 1.401 | 1.333 | 1.347 | 3.18 | |
| 6) | Aniline | 1.964 | 2.035 | 2.053 | 1.978 | 1.925 | 1.954 | 1.841 | 1.964 | 3.60 | |
| 7) S | Phenol-d6 | 1.897 | 1.960 | 2.062 | 1.978 | 1.953 | 1.998 | 1.905 | 1.965 | 2.87 | |
| 8) | 2-Chlorophenol | 1.199 | 1.221 | 1.291 | 1.250 | 1.236 | 1.285 | 1.232 | 1.245 | 2.68 | |
| 9) | Benzaldehyde | 1.364 | 1.447 | 1.409 | 1.198 | 1.073 | 1.057 | | 1.258 | 13.68 | |
| 10) C | Phenol | 1.968 | 2.023 | 2.138 | 2.045 | 2.020 | 2.077 | 1.971 | 2.035 | 2.94 | |
| 11) | bis(2-Chloroethyl)ether | 1.757 | 1.800 | 1.787 | 1.702 | 1.680 | 1.690 | 1.636 | 1.722 | 3.53 | |
| 12) | 1,3-Dichlorobenzene | 1.579 | 1.479 | 1.546 | 1.452 | 1.422 | 1.467 | 1.404 | 1.479 | 4.29 | |
| 13) C | 1,4-Dichlorobenzene | 1.571 | 1.514 | 1.553 | 1.475 | 1.435 | 1.491 | 1.413 | 1.493 | 3.89 | |
| 14) | 1,2-Dichlorobenzene | 1.523 | 1.475 | 1.513 | 1.410 | 1.377 | 1.416 | 1.356 | 1.439 | 4.57 | |
| 15) | Benzyl Alcohol | 1.306 | 1.454 | 1.474 | 1.415 | 1.417 | 1.437 | 1.365 | 1.410 | 4.07 | |
| 16) | 2,2'-oxybis(1-chloropropane) | 1.639 | 1.716 | 1.642 | 1.546 | 1.549 | 1.532 | 1.474 | 1.585 | 5.23 | |
| 17) | 2-Methylphenol | 1.257 | 1.313 | 1.350 | 1.290 | 1.277 | 1.288 | 1.230 | 1.286 | 2.99 | |
| 18) | Hexachloroethane | 0.599 | 0.590 | 0.617 | 0.594 | 0.587 | 0.614 | 0.590 | 0.599 | 2.04 | |
| 19) P | n-Nitroso-di-n-butylamine | 1.329 | 1.436 | 1.555 | 1.433 | 1.340 | 1.308 | 1.271 | 1.185 | 1.357 | 8.44 |
| 20) | 3+4-Methylphenols | 1.595 | 1.820 | 1.795 | 1.758 | 1.734 | 1.769 | 1.670 | 1.734 | 4.49 | |
| 21) I | Naphthalene-d8 | | | | | | | | | | |
| 22) | Acetophenone | 0.636 | 0.621 | 0.643 | 0.612 | 0.591 | 0.609 | 0.590 | 0.614 | 3.32 | |
| 23) S | Nitrobenzene-d5 | 0.300 | 0.339 | 0.386 | 0.410 | 0.412 | 0.436 | 0.430 | 0.387 | 13.07 | |
| 24) | Nitrobenzene | 0.338 | 0.385 | 0.430 | 0.443 | 0.435 | 0.460 | 0.454 | 0.421 | 10.43 | |
| 25) | Isophorone | 0.947 | 0.970 | 0.951 | 0.913 | 0.877 | 0.886 | 0.864 | 0.915 | 4.51 | |
| 26) C | 2-Nitrophenol | 0.054 | 0.063 | 0.076 | 0.094 | 0.107 | 0.120 | | 0.086 | 30.30# | |
| 27) | 2,4-Dimethylphenol | 0.218 | 0.225 | 0.232 | 0.225 | 0.221 | 0.233 | 0.227 | 0.226 | 2.34 | |
| 28) | bis(2-Chloroethyl)ether | 0.577 | 0.574 | 0.572 | 0.549 | 0.542 | 0.554 | 0.543 | 0.559 | 2.72 | |
| 29) C | 2,4-Dichlorophenol | 0.226 | 0.249 | 0.270 | 0.273 | 0.275 | 0.288 | 0.288 | 0.267 | 8.43 | |
| 30) | 1,2,4-Trichlorobenzene | 0.355 | 0.332 | 0.347 | 0.340 | 0.331 | 0.349 | 0.343 | 0.342 | 2.61 | |
| 31) | Naphthalene | 1.084 | 1.041 | 1.074 | 1.041 | 1.002 | 1.046 | 1.026 | 1.045 | 2.65 | |
| 32) | Benzoic acid | | 0.082 | 0.104 | 0.129 | 0.151 | 0.168 | 0.187 | 0.137 | 29.06 | |
| 33) | 4-Chloroaniline | 0.394 | 0.403 | 0.406 | 0.395 | 0.379 | 0.389 | 0.381 | 0.392 | 2.58 | |
| 34) C | Hexachlorobutane | 0.217 | 0.213 | 0.218 | 0.215 | 0.210 | 0.219 | 0.217 | 0.215 | 1.50 | |
| 35) | Caprolactam | 0.099 | 0.115 | 0.118 | 0.114 | 0.109 | 0.113 | 0.110 | 0.111 | 5.44 | |
| 36) C | 4-Chloro-3-methylphenol | 0.332 | 0.375 | 0.382 | 0.383 | 0.370 | 0.378 | 0.372 | 0.370 | 4.74 | |
| 37) | 2-Methylnaphthalene | 0.747 | 0.742 | 0.730 | 0.705 | 0.683 | 0.701 | 0.683 | 0.713 | 3.75 | |
| 38) | 1-Methylnaphthalene | 0.742 | 0.735 | 0.725 | 0.697 | 0.673 | 0.685 | 0.666 | 0.703 | 4.37 | |

Method Path : Z:\svoasrv\HPCHEM1\BNA_P\Methods\
 Method File : 8270E-BP081324.M

| | | | | |
|-------|-------------------|---|-------|--|
| 39) I | Acenaphthene-d10 | -----ISTD----- | | |
| 40) | 1,2,4,5-Tetrac... | 0.574 0.543 0.592 0.595 0.578 0.616 0.604 0.586 | 4.05 | |
| 41) P | Hexachlorocycl... | 0.159 0.153 0.174 0.192 0.195 0.214 0.214 0.186 | 13.17 | |
| 42) S | 2,4,6-Tribromo... | 0.179 0.204 0.222 0.236 0.231 0.245 0.240 0.223 | 10.47 | |
| 43) C | 2,4,6-Trichlor... | 0.261 0.284 0.318 0.343 0.339 0.367 0.367 0.326 | 12.49 | |
| 44) | 2,4,5-Trichlor... | 0.284 0.316 0.360 0.382 0.378 0.408 0.404 0.362 | 12.72 | |
| 45) S | 2-Fluorobiphenyl | 1.371 1.314 1.335 1.327 1.264 1.309 1.253 1.310 | 3.12 | |
| 46) | 1,1'-Biphenyl | 1.462 1.400 1.434 1.428 1.362 1.433 1.386 1.415 | 2.40 | |
| 47) | 2-Chloronaphth... | 1.121 1.075 1.117 1.116 1.065 1.129 1.088 1.102 | 2.30 | |
| 48) | 2-Nitroaniline | 0.146 0.181 0.231 0.285 0.294 0.329 0.328 0.256 | 28.08 | |
| 49) | Acenaphthylene | 1.660 1.610 1.657 1.645 1.561 1.633 1.570 1.619 | 2.51 | |
| 50) | Dimethylphthalate | 1.387 1.365 1.369 1.352 1.276 1.340 1.287 1.339 | 3.17 | |
| 51) | 2,6-Dinitrotol... | 0.129 0.187 0.227 0.255 0.261 0.282 0.283 0.232 | 24.32 | |
| 52) C | Acenaphthene | 1.094 1.058 1.074 1.073 1.023 1.072 1.042 1.063 | 2.23 | |
| 53) | 3-Nitroaniline | 0.127 0.185 0.221 0.254 0.254 0.279 0.277 0.228 | 24.33 | |
| 54) P | 2,4-Dinitrophenol | 0.045 0.052 0.066 0.073 0.085 0.093 0.069 | 27.01 | |
| 55) | Dibenzofuran | 1.735 1.637 1.677 1.649 1.563 1.607 1.575 1.635 | 3.67 | |
| 56) P | 4-Nitrophenol | 0.141 0.177 0.202 0.201 0.225 0.223 0.195 | 16.26 | |
| 57) | 2,4-Dinitrotol... | 0.140 0.214 0.262 0.320 0.328 0.362 0.365 0.284 | 29.47 | |
| 58) | Fluorene | 1.431 1.400 1.408 1.354 1.260 1.315 1.233 1.343 | 5.70 | |
| 59) | 2,3,4,6-Tetrac... | 0.250 0.278 0.302 0.328 0.317 0.340 0.333 0.307 | 10.67 | |
| 60) | Diethylphthalate | 1.405 1.395 1.427 1.397 1.318 1.368 1.310 1.374 | 3.26 | |
| 61) | 4-Chlorophenyl... | 0.741 0.724 0.718 0.695 0.656 0.680 0.644 0.694 | 5.19 | |
| 62) | 4-Nitroaniline | 0.148 0.198 0.242 0.273 0.262 0.288 0.278 0.241 | 21.07 | |
| 63) | Azobenzene | 1.797 1.779 1.778 1.714 1.555 1.648 1.555 1.689 | 6.21 | |
| 64) I | Phenanthrene-d10 | -----ISTD----- | | |
| 65) | 4,6-Dinitro-2.... | 0.031 0.039 0.052 0.060 0.069 0.075 0.054 | 30.94 | |
| 66) c | n-Nitrosodiphe... | 0.534 0.541 0.549 0.538 0.523 0.534 0.531 0.536 | 1.52 | |
| 67) | 4-Bromophenyl.... | 0.211 0.210 0.212 0.207 0.210 0.215 0.216 0.212 | 1.39 | |
| 68) | Hexachlorobenzene | 0.255 0.249 0.250 0.252 0.245 0.252 0.248 0.250 | 1.32 | |
| 69) | Atrazine | 0.181 0.186 0.196 0.191 0.179 0.182 0.174 0.184 | 4.00 | |
| 70) C | Pentachlorophenol | 0.108 0.125 0.137 0.143 0.151 0.156 0.137 | 13.07 | |
| 71) | Phenanthrene | 0.990 0.963 0.998 0.965 0.932 0.952 0.932 0.962 | 2.71 | |
| 72) | Anthracene | 0.954 0.955 0.970 0.950 0.915 0.943 0.915 0.943 | 2.21 | |
| 73) | Carbazole | 0.908 0.906 0.941 0.930 0.873 0.913 0.892 0.909 | 2.49 | |
| 74) | Di-n-butylphth... | 1.049 1.065 1.149 1.149 1.133 1.144 1.126 1.116 | 3.75 | |
| 75) C | Fluoranthene | 1.215 1.153 1.233 1.200 1.137 1.179 1.144 1.180 | 3.17 | |
| 76) I | Chrysene-d12 | -----ISTD----- | | |
| 77) | Benzidine | 0.297 0.451 0.477 0.476 0.408 0.409 0.420 | 16.09 | |
| 78) | Pyrene | 1.283 1.470 1.295 1.290 1.262 1.270 1.282 1.307 | 5.55 | |
| 79) S | Terphenyl-d14 | 1.069 1.196 1.034 1.006 0.973 0.994 0.935 1.029 | 8.24 | |
| 80) | Butylbenzylpht... | 0.298 0.365 0.411 0.470 0.489 0.508 0.519 0.437 | 18.83 | |
| 81) | Benzo(a)anthra... | 1.271 1.247 1.291 1.268 1.224 1.244 1.242 1.256 | 1.79 | |
| 82) | 3,3'-Dichlorob... | 0.349 0.358 0.413 0.428 0.419 0.429 0.422 0.403 | 8.47 | |
| 83) | Chrysene | 1.213 1.180 1.205 1.197 1.151 1.182 1.169 1.185 | 1.81 | |
| 84) | Bis(2-ethylhex... | 0.607 0.635 0.741 0.781 0.804 0.800 0.826 0.742 | 11.70 | |
| 85) c | Di-n-octyl pht... | 0.839 1.115 1.247 1.285 1.332 1.376 1.199 | 16.51 | |

Method Path : Z:\svoasrv\HPCHEM1\BNA_P\Methods\


Method File : 8270E-BP081324.M

| | | | | | | | | | | | | | | | |
|-----|---|--------------------|---|------|--|--|--|--|--|--|--|--|--|--|--|
| 86) | I | Perylene-d12 | -----ISTD----- | | | | | | | | | | | | |
| 87) | | Indeno(1,2,3-c...) | 1.244 1.156 1.317 1.327 1.297 1.384 1.350 1.296 | 5.84 | | | | | | | | | | | |
| 88) | | Benzo(b)fluora... | 1.106 1.152 1.158 1.155 1.119 1.189 1.163 1.149 | 2.42 | | | | | | | | | | | |
| 89) | | Benzo(k)fluora... | 1.115 1.101 1.124 1.141 1.103 1.150 1.110 1.121 | 1.69 | | | | | | | | | | | |
| 90) | C | Benzo(a)pyrene | 0.953 0.953 1.003 1.012 0.996 1.040 1.019 0.997 | 3.28 | | | | | | | | | | | |
| 91) | | Dibenzo(a,h)an... | 1.034 0.965 1.106 1.102 1.080 1.144 1.109 1.077 | 5.54 | | | | | | | | | | | |
| 92) | | Benzo(g,h,i)pe... | 1.052 0.955 1.096 1.102 1.074 1.151 1.128 1.080 | 5.91 | | | | | | | | | | | |

(#) = Out of Range

2,4-Dinitrotoluene

Response Ratio



Response = 3.838e-001 * Amt - 9.639e-002
Coef of Det (r^2) = 0.996361 Curve Fit: Linear
P:\3657\NANO\CMS\Groups\srsv\HPCHEM1\BNA P\Methods\8270E-BP081324.M
Calibration Table Last Updated: Wed Aug 14 00:00:30 2024

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081324\
 Data File : BP021481.D
 Acq On : 13 Aug 2024 12:32
 Operator : RC/JU
 Sample : SSTDICC2.5
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICC2.5

Quant Time: Aug 13 23:06:00 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 13 23:03:48 2024
 Response via : Initial Calibration


| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|------------------------------------|--------|------|----------|--------|-------|----------|
| Internal Standards | | | | | | |
| 1) 1,4-Dichlorobenzene-d4 | 7.805 | 152 | 291697 | 20.000 | ng | 0.00 |
| 21) Naphthalene-d8 | 10.587 | 136 | 1211879 | 20.000 | ng | -0.01 |
| 39) Acenaphthene-d10 | 14.445 | 164 | 840945 | 20.000 | ng | -0.01 |
| 64) Phenanthrene-d10 | 17.239 | 188 | 1860621 | 20.000 | ng | -0.01 |
| 76) Chrysene-d12 | 21.704 | 240 | 1845246 | 20.000 | ng | 0.00 |
| 86) Perylene-d12 | 25.115 | 264 | 2154249 | 20.000 | ng | 0.01 |
| System Monitoring Compounds | | | | | | |
| 5) 2-Fluorophenol | 0.000 | 112 | 0d | 0.000 | ng | |
| 7) Phenol-d6 | 0.000 | 99 | 0d | 0.000 | ng | |
| 23) Nitrobenzene-d5 | 0.000 | 82 | 0d | 0.000 | ng | |
| 42) 2,4,6-Tribromophenol | 0.000 | 330 | 0d | 0.000 | ng | |
| 45) 2-Fluorobiphenyl | 0.000 | 172 | 0d | 0.000 | ng | |
| 79) Terphenyl-d14 | 0.000 | 244 | 0d | 0.000 | ng | |
| Target Compounds | | | | | | |
| 19) n-Nitroso-di-n-propyla... | 8.599 | 70 | 48460 | 2.448 | ng | 99 |


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081324\
 Data File : BP021481.D
 Acq On : 13 Aug 2024 12:32
 Operator : RC/JU
 Sample : SSTDICC2.5
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC2.5

Quant Time: Aug 13 23:06:00 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 13 23:03:48 2024
 Response via : Initial Calibration







#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.805 min Scan# 8
Delta R.T. -0.006 min
Lab File: BP021481.D
Acq: 13 Aug 2024 12:32

Instrument : BNA_P
ClientSampleId : SSTDICC2.5






Tgt Ion:152 Resp: 291697
Ion Ratio Lower Upper
152 100
150 154.6 122.4 183.6
115 68.5 54.7 82.1



#5
2-Fluorophenol
Concen: 0.000 ng
Expected RT: 5.39 min

Lab File: BP021481.D
Acq: 13 Aug 2024 12:32

Tgt Ion: 112
Sig Exp Ratio
112 100
64 61.3
63 30.7




#7
Phenol-d6
Concen: 0.000 ng
Expected RT: 6.96 min


Instrument :
BNA_P
ClientSampleId :
SSTDICC2.5

Tgt Ion: 99
Sig Exp Ratio
99 100
42 12.8
71 32.3

#19
n-Nitroso-di-n-propylamine
Concen: 2.448 ng
RT: 8.599 min Scan# 954
Delta R.T. -0.006 min
Lab File: BP021481.D
Acq: 13 Aug 2024 12:32

Tgt Ion: 70 Resp: 48460
Ion Ratio Lower Upper
70 100
42 41.7 33.1 49.7
101 7.8 7.4 11.2
130 15.6 13.3 19.9







#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.587 min Scan# 1
Delta R.T. -0.012 min
Lab File: BP021481.D
Acq: 13 Aug 2024 12:32

Instrument : BNA_P
ClientSampleId : SSTDICC2.5

Tgt Ion:136 Resp: 1211879

Ion Ratio Lower Upper


| | | | |
|-----|------|-----|------|
| 136 | 100 | | |
| 137 | 11.4 | 8.8 | 13.2 |
| 54 | 8.4 | 6.3 | 9.5 |
| 68 | 8.2 | 6.8 | 10.2 |



#23
Nitrobenzene-d5
Concen: 0.000 ng
Expected RT: 8.95 min

Lab File: BP021481.D
Acq: 13 Aug 2024 12:32



Tgt Ion: 82
Sig Exp Ratio
82 100
128 33.1
54 37.9



#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.445 min Scan# 1
Delta R.T. -0.012 min
Lab File: BP021481.D
Acq: 13 Aug 2024 12:32

Instrument : BNA_P
ClientSampleId : SSTDICC2.5






Tgt Ion:164 Resp: 840945
Ion Ratio Lower Upper
164 100
162 101.1 81.5 122.3
160 44.1 35.2 52.8



#42
2,4,6-Tribromophenol
Concen: 0.000 ng
Expected RT: 15.97 min

Lab File: BP021481.D
Acq: 13 Aug 2024 12:32

Tgt Ion: 330
Sig Exp Ratio
330 100
332 96.7
141 30.6




#45
2-Fluorobiphenyl
Concen: 0.000 ng
Expected RT: 13.06 min


Instrument :
BNA_P
ClientSampleId :
SSTDICC2.5

Tgt Ion: 172
Sig Exp Ratio
172 100
171 35.3
170 23.0

#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.239 min Scan# 2423
Delta R.T. -0.012 min
Lab File: BP021481.D
Acq: 13 Aug 2024 12:32

Tgt Ion:188 Resp: 1860621
Ion Ratio Lower Upper
188 100
94 13.0 10.2 15.4
80 14.6 11.6 17.4







#76
 Chrysene-d12
 Concen: 20.000 ng
 RT: 21.704 min Scan# 3
 Delta R.T. 0.006 min
 Lab File: BP021481.D
 Acq: 13 Aug 2024 12:32

Instrument : BNA_P
 ClientSampleId : SSTDICC2.5

Tgt Ion:240 Resp: 1845246
 Ion Ratio Lower Upper
 240 100
 120 12.7 9.9 14.9
 236 25.4 20.6 30.8



#79
 Terphenyl-d14
 Concen: 0.000 ng
 Expected RT: 19.97 min
 Lab File: BP021481.D
 Acq: 13 Aug 2024 12:32

Tgt Ion: 244
 Sig Exp Ratio
 244 100
 212 7.0
 122 11.8

#86

Perylene-d₁₂

Concen: 20.000 ng

RT: 25.115 min Scan# 3 Instrument :

Delta R.T. 0.012 min BNA_P

Lab File: BP021481.D ClientSampleId :

Acq: 13 Aug 2024 12:32 SSTDICC2.5





Tgt Ion:264 Resp: 2154249

Ion Ratio Lower Upper

264 100

260 23.9 18.6 28.0

265 21.8 17.1 25.7



Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081324\

Data File : BP021482.D

Acq On : 13 Aug 2024 13:15

Operator : RC/JU

Sample : SSTDICC005

Misc :

ALS Vial : 3 Sample Multiplier: 1

Quant Time: Aug 13 23:48:54 2024

Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M

Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

QLast Update : Tue Aug 13 23:44:24 2024

Response via : Initial Calibration

Instrument :

BNA_P

ClientSampleId :

SSTDICC005

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|----------|------|------|----------|------|-------|----------|
|----------|------|------|----------|------|-------|----------|

Internal Standards

| | | | | | | |
|---------------------------|--------|-----|---------|--------|----|-------|
| 1) 1,4-Dichlorobenzene-d4 | 7.804 | 152 | 312360 | 20.000 | ng | 0.00 |
| 21) Naphthalene-d8 | 10.587 | 136 | 1334707 | 20.000 | ng | -0.01 |
| 39) Acenaphthene-d10 | 14.451 | 164 | 933864 | 20.000 | ng | 0.00 |
| 64) Phenanthrene-d10 | 17.257 | 188 | 2070828 | 20.000 | ng | 0.00 |
| 76) Chrysene-d12 | 21.698 | 240 | 2040393 | 20.000 | ng | 0.00 |
| 86) Perylene-d12 | 25.115 | 264 | 2299535 | 20.000 | ng | 0.01 |

System Monitoring Compounds

| | | | | | | |
|--------------------------|--------|-----|---------|--------|----|------|
| 5) 2-Fluorophenol | 5.393 | 112 | 208031 | 9.889 | ng | 0.00 |
| 7) Phenol-d6 | 6.957 | 99 | 296244 | 9.654 | ng | 0.00 |
| 23) Nitrobenzene-d5 | 8.946 | 82 | 199944 | 7.732 | ng | 0.00 |
| 42) 2,4,6-Tribromophenol | 15.963 | 330 | 83806 | 8.063 | ng | 0.00 |
| 45) 2-Fluorobiphenyl | 13.063 | 172 | 640316 | 10.464 | ng | 0.00 |
| 79) Terphenyl-d14 | 19.980 | 244 | 1090116 | 10.379 | ng | 0.00 |

Target Compounds

| | | | | Qvalue | |
|-------------------------------|--------|-----|--------|--------|----|
| 2) 1,4-Dioxane | 3.316 | 88 | 53510 | 5.365 | ng |
| 3) Pyridine | 3.711 | 79 | 136152 | 4.890 | ng |
| 4) n-Nitrosodimethylamine | 3.616 | 42 | 42551 | 5.084 | ng |
| 6) Aniline | 7.128 | 93 | 153398 | 5.000 | ng |
| 8) 2-Chlorophenol | 7.369 | 128 | 93591 | 4.814 | ng |
| 9) Benzaldehyde | 6.946 | 77 | 106515 | 5.421 | ng |
| 10) Phenol | 6.981 | 94 | 153713 | 4.837 | ng |
| 11) bis(2-Chloroethyl)ether | 7.228 | 93 | 137203 | 5.103 | ng |
| 12) 1,3-Dichlorobenzene | 7.699 | 146 | 123281 | 5.339 | ng |
| 13) 1,4-Dichlorobenzene | 7.840 | 146 | 122689 | 5.262 | ng |
| 14) 1,2-Dichlorobenzene | 8.157 | 146 | 118950 | 5.294 | ng |
| 15) Benzyl Alcohol | 8.028 | 79 | 101967 | 4.631 | ng |
| 16) 2,2'-oxybis(1-Chloropr... | 8.334 | 45 | 127954 | 5.167 | ng |
| 17) 2-Methylphenol | 8.228 | 107 | 98168 | 4.887 | ng |
| 18) Hexachloroethane | 8.887 | 117 | 46773 | 5.002 | ng |
| 19) n-Nitroso-di-n-propyla... | 8.599 | 70 | 112153 | 5.291 | ng |
| 20) 3+4-Methylphenols | 8.551 | 107 | 124528 | 4.597 | ng |
| 22) Acetophenone | 8.610 | 105 | 212108 | 5.173 | ng |
| 24) Nitrobenzene | 8.987 | 77 | 112817 | 4.017 | ng |
| 25) Isophorone | 9.516 | 82 | 315900 | 5.171 | ng |
| 26) 2-Nitrophenol | 9.698 | 139 | 18035 | 4.396 | ng |
| 27) 2,4-Dimethylphenol | 9.757 | 122 | 72762 | 4.826 | ng |
| 28) bis(2-Chloroethoxy)met... | 9.993 | 93 | 192662 | 5.167 | ng |
| 29) 2,4-Dichlorophenol | 10.228 | 162 | 75312 | 4.228 | ng |
| 30) 1,2,4-Trichlorobenzene | 10.451 | 180 | 118446 | 5.186 | ng |
| 31) Naphthalene | 10.640 | 128 | 361804 | 5.189 | ng |
| 33) 4-Chloroaniline | 10.740 | 127 | 131360 | 5.015 | ng |
| 34) Hexachlorobutadiene | 10.940 | 225 | 72436 | 5.038 | ng |
| 35) Caprolactam | 11.475 | 113 | 33090 | 4.463 | ng |
| 36) 4-Chloro-3-methylphenol | 11.857 | 107 | 110736 | 4.481 | ng |
| 37) 2-Methylnaphthalene | 12.251 | 142 | 249268 | 5.238 | ng |
| 38) 1-Methylnaphthalene | 12.475 | 142 | 247615 | 5.276 | ng |
| 40) 1,2,4,5-Tetrachloroben... | 12.622 | 216 | 133916 | 4.894 | ng |
| 41) Hexachlorocyclopentadiene | 12.616 | 237 | 37190 | 4.288 | ng |
| 43) 2,4,6-Trichlorophenol | 12.857 | 196 | 60845 | 4.003 | ng |
| 44) 2,4,5-Trichlorophenol | 12.928 | 196 | 66306 | 3.926 | ng |

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081324\
 Data File : BP021482.D
 Acq On : 13 Aug 2024 13:15
 Operator : RC/JU
 Sample : SSTDICC005
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC005

Quant Time: Aug 13 23:48:54 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 13 23:44:24 2024
 Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024
 Supervised By :mohammad ahmed 08/14/2024

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|-------------------------------|--------|------|----------|-------|-------|----------|
| 46) 1,1'-Biphenyl | 13.275 | 154 | 341430 | 5.168 | ng | 99 |
| 47) 2-Chloronaphthalene | 13.310 | 162 | 261761 | 5.089 | ng | 99 |
| 48) 2-Nitroaniline | 13.498 | 65 | 34186 | 7.264 | ng | 97 |
| 49) Acenaphthylene | 14.163 | 152 | 387614 | 5.126 | ng | 99 |
| 50) Dimethylphthalate | 13.898 | 163 | 323804 | 5.178 | ng | 99 |
| 51) 2,6-Dinitrotoluene | 14.010 | 165 | 30050 | 6.169 | ng | # 81 |
| 52) Acenaphthene | 14.516 | 154 | 255510 | 5.150 | ng | 99 |
| 53) 3-Nitroaniline | 14.334 | 138 | 29621 | 6.160 | ng | # 79 |
| 55) Dibenzofuran | 14.851 | 168 | 405023 | 5.307 | ng | 99 |
| 57) 2,4-Dinitrotoluene | 14.804 | 165 | 32636 | 6.843 | ng | 90 |
| 58) Fluorene | 15.516 | 166 | 333975 | 5.326 | ng | 99 |
| 59) 2,3,4,6-Tetrachlorophenol | 15.086 | 232 | 58426 | 4.076 | ng | # 98 |
| 60) Diethylphthalate | 15.281 | 149 | 328116 | 5.113 | ng | 99 |
| 61) 4-Chlorophenyl-phenyle... | 15.510 | 204 | 173085 | 5.340 | ng | 100 |
| 62) 4-Nitroaniline | 15.510 | 138 | 34574 | 5.481 | ng | # 67 |
| 63) Azobenzene | 15.810 | 77 | 419466 | 5.318 | ng | 98 |
| 66) n-Nitrosodiphenylamine | 15.728 | 169 | 276698 | 4.986 | ng | 99 |
| 67) 4-Bromophenyl-phenylether | 16.428 | 248 | 109250 | 4.985 | ng | 97 |
| 68) Hexachlorobenzene | 16.545 | 284 | 132184 | 5.104 | ng | 97 |
| 69) Atrazine | 16.692 | 200 | 93767 | 4.921 | ng | 99 |
| 71) Phenanthrene | 17.292 | 178 | 512757 | 5.149 | ng | 99 |
| 72) Anthracene | 17.386 | 178 | 493987 | 5.058 | ng | 99 |
| 73) Carbazole | 17.669 | 167 | 470253 | 4.996 | ng | 99 |
| 74) Di-n-butylphthalate | 18.269 | 149 | 542975 | 4.697 | ng | 99 |
| 75) Fluoranthene | 19.380 | 202 | 629080 | 5.148 | ng | 99 |
| 78) Pyrene | 19.757 | 202 | 654333 | 4.906 | ng | 100 |
| 80) Butylbenzylphthalate | 20.710 | 149 | 152079 | 6.382 | ng | 95 |
| 81) Benzo(a)anthracene | 21.680 | 228 | 648521 | 5.063 | ng | 100 |
| 82) 3,3'-Dichlorobenzidine | 21.592 | 252 | 178137 | 4.338 | ng | 98 |
| 83) Chrysene | 21.745 | 228 | 618621 | 5.115 | ng | 99 |
| 84) Bis(2-ethylhexyl)phtha... | 21.651 | 149 | 309738 | 4.091 | ng | 100 |
| 87) Indeno(1,2,3-cd)pyrene | 28.962 | 276 | 715393m | 4.800 | ng | |
| 88) Benzo(b)fluoranthene | 24.015 | 252 | 636027 | 4.816 | ng | 99 |
| 89) Benzo(k)fluoranthene | 24.092 | 252 | 640982 | 4.975 | ng | 99 |
| 90) Benzo(a)pyrene | 24.939 | 252 | 548144 | 4.783 | ng | 100 |
| 91) Dibenzo(a,h)anthracene | 29.068 | 278 | 594702 | 4.802 | ng | 98 |
| 92) Benzo(g,h,i)perylene | 30.162 | 276 | 604771 | 4.872 | ng | 96 |

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081324\
 Data File : BP021482.D
 Acq On : 13 Aug 2024 13:15
 Operator : RC/JU
 Sample : SSTDICC005
 Misc :
 ALS Vial : 3 Sample Multiplier: 1


Quant Time: Aug 13 23:48:54 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 13 23:44:24 2024
 Response via : Initial Calibration

Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC005

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024
 Supervised By :mohammad ahmed 08/14/2024







#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.804 min Scan# 8
Delta R.T. -0.007 min
Lab File: BP021482.D
Acq: 13 Aug 2024 13:15

Instrument : BNA_P
ClientSampleId : SSTDICC005


Manual Integrations
APPROVED







Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#2
1,4-Dioxane
Concen: 5.365 ng
RT: 3.316 min Scan# 56
Delta R.T. -0.000 min
Lab File: BP021482.D
Acq: 13 Aug 2024 13:15

Tgt Ion: 88 Resp: 53510
Ion Ratio Lower Upper
88 100
58 62.2 48.7 73.1
43 20.5 15.4 23.2





#3


Pyridine
Concen: 4.890 ng
RT: 3.711 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP021482.D
Acq: 13 Aug 2024 13:15

Instrument : BNA_P
ClientSampleId : SSTDICC005

Tgt Ion: 79 Resp: 13615
Ion Ratio Lower Upper
79 100
52 48.5 39.0 58.6
51 25.2 19.2 28.8

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024




#4

n-Nitrosodimethylamine
Concen: 5.084 ng
RT: 3.616 min Scan# 107
Delta R.T. -0.000 min
Lab File: BP021482.D
Acq: 13 Aug 2024 13:15

Tgt Ion: 42 Resp: 42551
Ion Ratio Lower Upper
42 100
74 182.7 154.3 231.5
44 21.2 6.2 9.4#







#5
2-Fluorophenol
Concen: 9.889 ng
RT: 5.393 min Scan# 409
Delta R.T. -0.000 min
Lab File: BP021482.D
Acq: 13 Aug 2024 13:15

Instrument : BNA_P
ClientSampleId : SSTDICC005


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#6
Aniline
Concen: 5.000 ng
RT: 7.128 min Scan# 704
Delta R.T. -0.006 min
Lab File: BP021482.D
Acq: 13 Aug 2024 13:15

Tgt Ion: 93 Resp: 153398
Ion Ratio Lower Upper
93 100
66 35.4 30.2 45.2
65 18.0 14.7 22.1







#7
 Phenol-d6
 Concen: 9.654 ng
 RT: 6.957 min Scan# 6
 Delta R.T. -0.006 min
 Lab File: BP021482.D
 Acq: 13 Aug 2024 13:15

Instrument : BNA_P
 ClientSampleId : SSTDICC005


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
 Supervised By :mohammad ahmed 08/14/2024



#8
 2-Chlorophenol
 Concen: 4.814 ng
 RT: 7.369 min Scan# 745
 Delta R.T. -0.006 min
 Lab File: BP021482.D
 Acq: 13 Aug 2024 13:15


Tgt Ion:128 Resp: 93591
 Ion Ratio Lower Upper
 128 100
 130 30.8 11.6 51.6
 64 56.2 35.3 75.3





#9
Benzaldehyde
Concen: 5.421 ng
RT: 6.946 min Scan# 6
Delta R.T. -0.000 min
Lab File: BP021482.D
Acq: 13 Aug 2024 13:15




Instrument : BNA_P
ClientSampleId : SSTDICC005




Tgt Ion: 77 Resp: 10651
Ion Ratio Lower Upper
77 100
105 91.9 72.1 112.1
106 86.6 67.9 107.9

Manual Integrations APPROVED



Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024






#10
Phenol
Concen: 4.837 ng
RT: 6.981 min Scan# 679
Delta R.T. -0.006 min
Lab File: BP021482.D
Acq: 13 Aug 2024 13:15



Tgt Ion: 94 Resp: 153713
Ion Ratio Lower Upper
94 100
65 25.6 6.5 46.5
66 33.2 14.7 54.7









#11
bis(2-Chloroethyl)ether
Concen: 5.103 ng
RT: 7.228 min Scan# 7
Delta R.T. -0.000 min
Lab File: BP021482.D
Acq: 13 Aug 2024 13:15

Instrument : BNA_P
ClientSampleId : SSTDICC005


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024





#12
1,3-Dichlorobenzene
Concen: 5.339 ng
RT: 7.699 min Scan# 801
Delta R.T. -0.000 min
Lab File: BP021482.D
Acq: 13 Aug 2024 13:15

Tgt Ion:146 Resp: 123281
Ion Ratio Lower Upper
146 100
148 64.9 51.8 77.6
75 39.9 31.3 46.9







Abundance Scan 825 (7.840 min): BP021482.D\data.ms



Abundance Scan 879 (8.157 min): BP021482.D\data.ms



Abundance Scan 879 (8.157 min): BP021482.D\data.ms (-82)



#13

1,4-Dichlorobenzene

Concen: 5.262 ng

RT: 7.840 min Scan# 8

Delta R.T. -0.000 min

Lab File: BP021482.D

Acq: 13 Aug 2024 13:15

Instrument :

BNA_P

ClientSampleId :

SSTDICC005

Tgt Ion:146 Resp: 122689

Ion Ratio Lower Upper

146 100

148 64.0 51.5 77.3


111 46.8 40.3 60.5

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#14

1,2-Dichlorobenzene

Concen: 5.294 ng

RT: 8.157 min Scan# 879

Delta R.T. -0.000 min

Lab File: BP021482.D

Acq: 13 Aug 2024 13:15


Tgt Ion:146 Resp: 118950




Ion Ratio Lower Upper

146 100

148 62.9 51.5 77.3

111 52.0 41.1 61.7





#15

Benzyl Alcohol

Concen: 4.631 ng

RT: 8.028 min Scan# 8

Delta R.T. -0.006 min

Lab File: BP021482.D

Acq: 13 Aug 2024 13:15

Instrument :

BNA_P





ClientSampleId :

SSTDICC005

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#16

2,2'-oxybis(1-Chloropropane)

Concen: 5.167 ng

RT: 8.334 min Scan# 909

Delta R.T. -0.000 min

Lab File: BP021482.D

Acq: 13 Aug 2024 13:15


Tgt Ion: 45 Resp: 127954




Ion Ratio Lower Upper

45 100

77 23.2 1.9 41.9

79 16.6 0.0 36.6





#17

2-Methylphenol

Concen: 4.887 ng

RT: 8.228 min Scan# 8

Delta R.T. -0.006 min

Lab File: BP021482.D

Acq: 13 Aug 2024 13:15

Instrument :

BNA_P




ClientSampleId :

SSTDICC005

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#18

Hexachloroethane

Concen: 5.002 ng

RT: 8.887 min Scan# 1003

Delta R.T. -0.006 min

Lab File: BP021482.D

Acq: 13 Aug 2024 13:15


Tgt Ion:117 Resp: 46773







Ion Ratio Lower Upper

117 100

119 96.2 73.9 110.9

201 90.9 74.9 112.3






#19

n-Nitroso-di-n-propylamine
Concen: 5.291 ng
RT: 8.599 min Scan# 9
Delta R.T. -0.006 min
Lab File: BP021482.D
Acq: 13 Aug 2024 13:15

Instrument :
BNA_P
ClientSampleId :
SSTDICC005


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#20
3+4-Methylphenols
Concen: 4.597 ng
RT: 8.551 min Scan# 946
Delta R.T. -0.006 min
Lab File: BP021482.D
Acq: 13 Aug 2024 13:15

Tgt Ion:107 Resp: 124528
Ion Ratio Lower Upper
107 100
108 89.0 67.6 107.6
77 36.6 15.0 55.0
79 31.8 8.9 48.9





#21

Naphthalene-d8

Concen: 20.000 ng

RT: 10.587 min Scan# 1

Delta R.T. -0.012 min

Lab File: BP021482.D

Acq: 13 Aug 2024 13:15

Instrument :

BNA_P



ClientSampleId :

SSTDICC005

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#22

Acetophenone

Concen: 5.173 ng

RT: 8.610 min Scan# 956

Delta R.T. -0.006 min

Lab File: BP021482.D

Acq: 13 Aug 2024 13:15

Tgt Ion:105 Resp: 212108


Ion Ratio Lower Upper


105 100

71 4.2 4.6 7.0#

51 19.9 16.3 24.5

120 19.0 16.2 24.2







#23
 Nitrobenzene-d5
 Concen: 7.732 ng
 RT: 8.946 min Scan# 1
 Delta R.T. -0.006 min
 Lab File: BP021482.D
 Acq: 13 Aug 2024 13:15

Instrument : BNA_P
 ClientSampleId : SSTDICC005


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
 Supervised By :mohammad ahmed 08/14/2024



#24
 Nitrobenzene
 Concen: 4.017 ng
 RT: 8.987 min Scan# 1020
 Delta R.T. -0.006 min
 Lab File: BP021482.D
 Acq: 13 Aug 2024 13:15


Tgt Ion: 77 Resp: 112817
 Ion Ratio Lower Upper
 77 100
 123 35.0 29.5 44.3
 65 13.6 9.9 14.9





#25
Isophorone
Concen: 5.171 ng
RT: 9.516 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP021482.D
Acq: 13 Aug 2024 13:15



Instrument : BNA_P
ClientSampleId : SSTDICC005



Tgt Ion: 82 Resp: 315900
Ion Ratio Lower Upper
82 100
95 7.4 5.4 8.2
138 12.9 10.6 16.0


Manual Integrations
APPROVED




Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#26
2-Nitrophenol
Concen: 4.396 ng
RT: 9.698 min Scan# 1141
Delta R.T. -0.001 min
Lab File: BP021482.D
Acq: 13 Aug 2024 13:15

Tgt Ion:139 Resp: 18035
Ion Ratio Lower Upper
139 100
109 38.1 28.9 43.3
65 63.3 49.0 73.6





#27

2,4-Dimethylphenol

Concen: 4.826 ng

RT: 9.757 min Scan# 1

Delta R.T. -0.000 min

Lab File: BP021482.D

Acq: 13 Aug 2024 13:15

Instrument :

BNA_P





ClientSampleId :

SSTDICC005

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#28

bis(2-Chloroethoxy)methane

Concen: 5.167 ng


RT: 9.993 min Scan# 1191


Delta R.T. -0.012 min

Lab File: BP021482.D

Acq: 13 Aug 2024 13:15

| Tgt | Ion: | 93 | Resp: | 192662 |
|-----|-------|-------|-------|--------|
| Ion | Ratio | Lower | Upper | |
| 93 | 100 | | | |
| 95 | 32.5 | 26.2 | 39.4 | |
| 123 | 9.5 | 7.6 | 11.4 | |






#29
2,4-Dichlorophenol
Concen: 4.228 ng
RT: 10.228 min Scan# 1
Delta R.T. -0.012 min
Lab File: BP021482.D
Acq: 13 Aug 2024 13:15

Instrument : BNA_P
ClientSampleId : SSTDICC005


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#30
1,2,4-Trichlorobenzene
Concen: 5.186 ng
RT: 10.451 min Scan# 1269
Delta R.T. -0.006 min
Lab File: BP021482.D
Acq: 13 Aug 2024 13:15

Tgt Ion:180 Resp: 118446
Ion Ratio Lower Upper
180 100
182 92.2 78.3 117.5
145 28.6 24.0 36.0







#31
Naphthalene
Concen: 5.189 ng
RT: 10.640 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP021482.D
Acq: 13 Aug 2024 13:15

Instrument : BNA_P
ClientSampleId : SSTDICC005


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#33
4-Chloroaniline
Concen: 5.015 ng
RT: 10.740 min Scan# 1318
Delta R.T. -0.000 min
Lab File: BP021482.D
Acq: 13 Aug 2024 13:15

Tgt Ion:127 Resp: 131360
Ion Ratio Lower Upper
127 100
129 32.3 24.5 36.7
65 38.1 28.9 43.3
92 25.1 19.3 28.9



#34

Hexachlorobutadiene

Concen: 5.038 ng

RT: 10.940 min Scan# 1

Delta R.T. -0.000 min

Lab File: BP021482.D




Acq: 13 Aug 2024 13:15

Instrument :

BNA_P

ClientSampleId :





SSTDICC005



| Tgt | Ion:225 | Resp: | 72430 |
|-----|---------|-------|-------|
| Ion | Ratio | Lower | Upper |
| 225 | 100 | | |
| 223 | 61.7 | 49.6 | 74.4 |
| 227 | 66.6 | 50.1 | 75.1 |


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#35
Caprolactam
Concen: 4.463 ng
RT: 11.475 min Scan# 1443
Delta R.T. -0.024 min
Lab File: BP021482.D
Acq: 13 Aug 2024 13:15

| Tgt | Ion:113 | Resp: | 33090 |
|-----|---------|-------|-------|
| Ion | Ratio | Lower | Upper |
| 113 | 100 | | |
| 55 | 153.8 | 130.6 | 170.6 |
| 56 | 115.8 | 94.6 | 134.6 |







#36
4-Chloro-3-methylphenol
Concen: 4.481 ng
RT: 11.857 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP021482.D
Acq: 13 Aug 2024 13:15

Instrument : BNA_P
ClientSampleId : SSTDICC005


1 Manual Integrations
2 APPROVED


3 Reviewed By :Yogesh Patel 08/14/2024
4 Supervised By :mohammad ahmed 08/14/2024



#37
2-Methylnaphthalene
Concen: 5.238 ng
RT: 12.251 min Scan# 1575
Delta R.T. -0.012 min
Lab File: BP021482.D
Acq: 13 Aug 2024 13:15

Tgt Ion:142 Resp: 249268
Ion Ratio Lower Upper
142 100
141 87.0 71.7 107.5
115 39.7 31.4 47.2







#38
1-Methylnaphthalene
Concen: 5.276 ng
RT: 12.475 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP021482.D
Acq: 13 Aug 2024 13:15

Instrument : BNA_P
ClientSampleId : SSTDICC005


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.451 min Scan# 1949
Delta R.T. -0.006 min
Lab File: BP021482.D
Acq: 13 Aug 2024 13:15

Tgt Ion:164 Resp: 933864
Ion Ratio Lower Upper
164 100
162 100.3 81.5 122.3
160 44.3 35.2 52.8







#40
1,2,4,5-Tetrachlorobenzene
Concen: 4.894 ng
RT: 12.622 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP021482.D
Acq: 13 Aug 2024 13:15

Instrument :
BNA_P
ClientSampleId :
SSTDICC005


Manual Integrations APPROVED




Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#41
Hexachlorocyclopentadiene
Concen: 4.288 ng
RT: 12.616 min Scan# 1637
Delta R.T. -0.006 min
Lab File: BP021482.D
Acq: 13 Aug 2024 13:15

Tgt Ion:237 Resp: 37190
Ion Ratio Lower Upper
237 100
235 60.5 43.9 83.9
272 14.3 0.0 33.3





#42

2,4,6-Tribromophenol

Concen: 8.063 ng

RT: 15.963 min Scan# 2

Delta R.T. -0.006 min

Lab File: BP021482.D

Acq: 13 Aug 2024 13:15

Instrument :

BNA_P




ClientSampleId :

SSTDICC005

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#43

2,4,6-Trichlorophenol

Concen: 4.003 ng


RT: 12.857 min Scan# 1678

Delta R.T. -0.012 min

Lab File: BP021482.D

Acq: 13 Aug 2024 13:15

| Tgt | Ion:196 | Resp: | 60845 |
|-----|---------|-------|-------|
| Ion | Ratio | Lower | Upper |
| 196 | 100 | | |
| 198 | 94.2 | 76.6 | 114.8 |
| 200 | 32.3 | 24.5 | 36.7 |



#44

2,4,5-Trichlorophenol

Concen: 3.926 ng

RT: 12.928 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP021482.D




Acq: 13 Aug 2024 13:15

Instrument :

BNA_P

ClientSampleId :

SSTDICC005



Tgt Ion:196 Resp: 66300

Ion Ratio Lower Upper

196 100

198 98.3 78.9 118.3





97 63.1 49.4 74.0

132 27.6 21.1 31.7

Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#45

2-Fluorobiphenyl

Concen: 10.464 ng

RT: 13.063 min Scan# 1713

Delta R.T. -0.000 min

Lab File: BP021482.D

Acq: 13 Aug 2024 13:15


Tgt Ion:172 Resp: 640316


Ion Ratio Lower Upper

172 100

171 35.2 28.2 42.4

170 23.9 18.4 27.6







#46
1,1'-Biphenyl
Concen: 5.168 ng
RT: 13.275 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP021482.D
Acq: 13 Aug 2024 13:15

Instrument : BNA_P
ClientSampleId : SSTDICC005


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#47
2-Chloronaphthalene
Concen: 5.089 ng
RT: 13.310 min Scan# 1755
Delta R.T. -0.006 min
Lab File: BP021482.D
Acq: 13 Aug 2024 13:15

Tgt Ion:162 Resp: 261761
Ion Ratio Lower Upper
162 100
127 40.1 31.9 47.9
164 31.9 26.8 40.2







#48
2-Nitroaniline
Concen: 7.264 ng
RT: 13.498 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP021482.D
Acq: 13 Aug 2024 13:15

Instrument : BNA_P
ClientSampleId : SSTDICC005


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024




#49
Acenaphthylene
Concen: 5.126 ng
RT: 14.163 min Scan# 1900
Delta R.T. -0.006 min
Lab File: BP021482.D
Acq: 13 Aug 2024 13:15

Tgt Ion:152 Resp: 387614
Ion Ratio Lower Upper
152 100
151 19.2 15.9 23.9
153 12.9 10.6 16.0





Abundance Scan 1855 (13.898 min): BP021482.D\data.ms (-)



#50

Dimethylphthalate

Concen: 5.178 ng

RT: 13.898 min Scan# 1

Delta R.T. -0.012 min

Lab File: BP021482.D

Acq: 13 Aug 2024 13:15

Instrument :

BNA_P



ClientSampleId :

SSTDICC005


**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



Abundance Scan 1874 (14.010 min): BP021482.D\data.ms (-)



#51

2,6-Dinitrotoluene

Concen: 6.169 ng

RT: 14.010 min Scan# 1874

Delta R.T. -0.012 min

Lab File: BP021482.D

Acq: 13 Aug 2024 13:15


Tgt Ion:165 Resp: 30050


Ion Ratio Lower Upper

165 100


63 73.6 51.5 77.3

89 93.9 57.8 86.8#





Abundance Scan 1960 (14.516 min): BP021482.D\data.ms (-)



#52

Acenaphthene

Concen: 5.150 ng

RT: 14.516 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP021482.D

Acq: 13 Aug 2024 13:15

Instrument :

BNA_P


ClientSampleId :

SSTDICC005

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#53

3-Nitroaniline

Concen: 6.160 ng

RT: 14.334 min Scan# 1929

Delta R.T. -0.017 min

Lab File: BP021482.D

Acq: 13 Aug 2024 13:15


Tgt Ion:138 Resp: 29621


Ion Ratio Lower Upper

138 100


108 63.9 16.2 24.4#

92 167.4 122.2 183.4






Abundance Scan 2017 (14.851 min): BP021482.D\data.ms (-)



Abundance Scan 2017 (14.851 min): BP021482.D\data.ms (-)



#55

Dibenzofuran

Concen: 5.307 ng

RT: 14.851 min Scan# 2

Delta R.T. -0.006 min

Lab File: BP021482.D

Acq: 13 Aug 2024 13:15

Instrument :

BNA_P


ClientSampleId :

SSTDICC005


**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/14/2024


Supervised By :mohammad ahmed 08/14/2024



Abundance Scan 2009 (14.804 min): BP021482.D\data.ms (-)



Abundance Scan 2009 (14.804 min): BP021482.D\data.ms (-)



#57

2,4-Dinitrotoluene

Concen: 6.843 ng

RT: 14.804 min Scan# 2009

Delta R.T. -0.006 min

Lab File: BP021482.D

Acq: 13 Aug 2024 13:15

Tgt Ion:165 Resp: 32636


Ion Ratio Lower Upper


165 100

63 58.2 42.5 63.7

89 107.5 77.1 115.7

182 2.7 2.0 3.0







#58
Fluorene
Concen: 5.326 ng
RT: 15.516 min Scan# 2
Delta R.T. -0.000 min
Lab File: BP021482.D
Acq: 13 Aug 2024 13:15

Instrument : BNA_P
ClientSampleId : SSTDICC005


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#59
2,3,4,6-Tetrachlorophenol
Concen: 4.076 ng
RT: 15.086 min Scan# 2057
Delta R.T. -0.000 min
Lab File: BP021482.D
Acq: 13 Aug 2024 13:15

Tgt Ion:232 Resp: 58426
Ion Ratio Lower Upper
232 100
131 40.4 33.1 49.7
130 2.7 1.7 2.5#
166 29.6 23.0 34.4







#60
Diethylphthalate
Concen: 5.113 ng
RT: 15.281 min Scan# 2129
Delta R.T. -0.024 min
Lab File: BP021482.D
Acq: 13 Aug 2024 13:15

Instrument : BNA_P
ClientSampleId : SSTDICC005


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#61
4-Chlorophenyl-phenylether
Concen: 5.340 ng
RT: 15.510 min Scan# 2129
Delta R.T. -0.012 min
Lab File: BP021482.D
Acq: 13 Aug 2024 13:15

Tgt Ion:204 Resp: 173085
Ion Ratio Lower Upper
204 100
206 32.7 26.4 39.6
141 57.6 46.1 69.1



#62

4-Nitroaniline

Concen: 5.481 ng

RT: 15.510 min Scan# 2180

Delta R.T. -0.018 min

Lab File: BP021482.D

Acq: 13 Aug 2024 13:15

Instrument :

BNA_P

ClientSampleId :

SSTDICC005

Tgt Ion:138 Resp: 34574

Ion Ratio Lower Upper

138 100

92 76.5 46.4 86.4

108 138.4 73.2 113.2

Manual Integrations

APPROVED


Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024

Abundance

Scan 2129 (15.510 min): BP021482.D\data.ms (-)


166.1



Abundance

Scan 2129 (15.510 min): BP021482.D\data.ms (-)

166.1




Time--> 15.40 15.50

Abundance

Scan 2129 (15.510 min): BP021482.D\data.ms (-)

166.1



Sub


50

m/z--> 50 100 150 200 250 300 350 400 450 500

Abundance

Scan 2182 (15.822 min): BP021485.D\data.ms (-)

77.1



Sub


50

m/z--> 50 100 150 200 250 300 350 400 450 500

Abundance

Scan 2180 (15.810 min): BP021482.D\data.ms (-)

77.1



Sub

50

m/z--> 50 100 150 200 250 300 350 400 450 500

Abundance

Scan 2180 (15.810 min): BP021482.D\data.ms (-)

77.1



Time--> 15.80


Abundance

Scan 2180 (15.810 min): BP021482.D\data.ms (-)

77.1



Time--> 15.80



Tgt Ion:138 Resp: 34574

Ion Ratio Lower Upper

138 100

92 76.5 46.4 86.4

108 138.4 73.2 113.2

Manual Integrations

APPROVED


Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024

Abundance

Scan 2129 (15.510 min): BP021482.D\data.ms (-)


166.1



Abundance

Scan 2129 (15.510 min): BP021482.D\data.ms (-)

166.1




Time--> 15.40 15.50

Abundance

Scan 2129 (15.510 min): BP021482.D\data.ms (-)

166.1



Sub


50

m/z--> 50 100 150 200 250 300 350 400 450 500

Abundance

Scan 2182 (15.822 min): BP021485.D\data.ms (-)

77.1



Sub


50

m/z--> 50 100 150 200 250 300 350 400 450 500

Abundance

Scan 2180 (15.810 min): BP021482.D\data.ms (-)

77.1



Sub

50

m/z--> 50 100 150 200 250 300 350 400 450 500

Abundance

Scan 2180 (15.810 min): BP021482.D\data.ms (-)

77.1



Time--> 15.80


Abundance

Scan 2180 (15.810 min): BP021482.D\data.ms (-)


77.1



Time--> 15.80



Abundance Scan 2426 (17.257 min): BP021482.D\data.ms (-)



#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 17.257 min Scan# 2

Delta R.T. 0.006 min

Lab File: BP021482.D

Acq: 13 Aug 2024 13:15

Instrument :

BNA_P

ClientSampleId :

SSTDICC005

Tgt Ion:188 Resp: 207082

Ion Ratio Lower Upper

188 100

94 11.9 10.2 15.4


80 13.3 11.6 17.4

Manual Integrations


APPROVED

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



Abundance Scan 2166 (15.728 min): BP021482.D\data.ms (-)



#66

n-Nitrosodiphenylamine

Concen: 4.986 ng

RT: 15.728 min Scan# 2166

Delta R.T. -0.012 min

Lab File: BP021482.D

Acq: 13 Aug 2024 13:15


Tgt Ion:169 Resp: 276698



Ion Ratio Lower Upper

169 100

168 65.8 53.4 80.2

167 35.2 28.7 43.1






#67
4-Bromophenyl-phenylether
Concen: 4.985 ng
RT: 16.428 min Scan# 21
Delta R.T. -0.012 min
Lab File: BP021482.D
Acq: 13 Aug 2024 13:15

Instrument : BNA_P
ClientSampleId : SSTDICC005


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#68
Hexachlorobenzene
Concen: 5.104 ng
RT: 16.545 min Scan# 2305
Delta R.T. -0.006 min
Lab File: BP021482.D
Acq: 13 Aug 2024 13:15

Tgt Ion:284 Resp: 132184
Ion Ratio Lower Upper
284 100
142 31.5 23.4 35.2
249 30.8 24.1 36.1







#69
Atrazine
Concen: 4.921 ng
RT: 16.692 min Scan# 2
Delta R.T. -0.024 min
Lab File: BP021482.D
Acq: 13 Aug 2024 13:15

Instrument : BNA_P
ClientSampleId : SSTDICC005


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#71
Phenanthrene
Concen: 5.149 ng
RT: 17.292 min Scan# 2432
Delta R.T. -0.006 min
Lab File: BP021482.D
Acq: 13 Aug 2024 13:15

Tgt Ion:178 Resp: 512757
Ion Ratio Lower Upper
178 100
176 19.0 15.4 23.2
179 15.2 12.4 18.6







#72
Anthracene
Concen: 5.058 ng
RT: 17.386 min Scan# 2
Delta R.T. -0.012 min
Lab File: BP021482.D
Acq: 13 Aug 2024 13:15

Instrument : BNA_P
ClientSampleId : SSTDICC005


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#73
Carbazole
Concen: 4.996 ng
RT: 17.669 min Scan# 2496
Delta R.T. -0.000 min
Lab File: BP021482.D
Acq: 13 Aug 2024 13:15

Tgt Ion:167 Resp: 470253
Ion Ratio Lower Upper
167 100
166 20.9 17.0 25.4
139 12.3 9.9 14.9







#74
Di-n-butylphthalate
Concen: 4.697 ng
RT: 18.269 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP021482.D
Acq: 13 Aug 2024 13:15

Instrument : BNA_P
ClientSampleId : SSTDICC005


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#75
Fluoranthene
Concen: 5.148 ng
RT: 19.380 min Scan# 2787
Delta R.T. -0.000 min
Lab File: BP021482.D
Acq: 13 Aug 2024 13:15

Tgt Ion:202 Resp: 629080
Ion Ratio Lower Upper
202 100
101 14.8 0.0 34.0
203 17.8 0.0 37.5







#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.698 min Scan# 3
Delta R.T. -0.000 min
Lab File: BP021482.D
Acq: 13 Aug 2024 13:15

Instrument : BNA_P
ClientSampleId : SSTDICC005


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#78
Pyrene
Concen: 4.906 ng
RT: 19.757 min Scan# 2851
Delta R.T. 0.006 min
Lab File: BP021482.D
Acq: 13 Aug 2024 13:15

Tgt Ion:202 Resp: 654333
Ion Ratio Lower Upper
202 100
200 20.9 16.7 25.1
203 17.3 14.1 21.1







#79
Terphenyl-d14
Concen: 10.379 ng
RT: 19.980 min Scan# 2
Delta R.T. 0.005 min
Lab File: BP021482.D
Acq: 13 Aug 2024 13:15

Instrument : BNA_P
ClientSampleId : SSTDICC005


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#80
Butylbenzylphthalate
Concen: 6.382 ng
RT: 20.710 min Scan# 3013
Delta R.T. -0.000 min
Lab File: BP021482.D
Acq: 13 Aug 2024 13:15

Tgt Ion:149 Resp: 152079
Ion Ratio Lower Upper
149 100
91 97.1 73.3 109.9
206 21.7 17.1 25.7





#81

Benzo(a)anthracene

Concen: 5.063 ng

RT: 21.680 min Scan# 3178

Delta R.T. -0.000 min

Lab File: BP021482.D

Acq: 13 Aug 2024 13:15



Instrument : BNA_P

ClientSampleId : SSTDICC005

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#82

3,3'-Dichlorobenzidine

Concen: 4.338 ng


RT: 21.592 min Scan# 3163


Delta R.T. -0.018 min

Lab File: BP021482.D

Acq: 13 Aug 2024 13:15

| Tgt | Ion:252 | Resp: | 178137 |
|-----------|---------|-------|--------|
| Ion Ratio | Lower | Upper | |
| 252 | 100 | | |
| 254 | 63.3 | 51.7 | 77.5 |
| 126 | 11.2 | 8.7 | 13.1 |







#83
Chrysene
Concen: 5.115 ng
RT: 21.745 min Scan# 3
Delta R.T. -0.006 min
Lab File: BP021482.D
Acq: 13 Aug 2024 13:15

Instrument : BNA_P
ClientSampleId : SSTDICC005


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024





#84
Bis(2-ethylhexyl)phthalate
Concen: 4.091 ng
RT: 21.651 min Scan# 3173
Delta R.T. 0.006 min
Lab File: BP021482.D
Acq: 13 Aug 2024 13:15

Tgt Ion:149 Resp: 309738
Ion Ratio Lower Upper
149 100
167 27.8 22.3 33.5
279 5.2 3.8 5.6







Abundance Scan 3762 (25.115 min): BP021482.D\data.ms (-)



Abundance Scan 4416 (28.962 min): BP021482.D\data.ms (-)



Abundance Scan 4416 (28.962 min): BP021482.D\data.ms (-)



#86

Perylene-d₁₂

Concen: 20.000 ng

RT: 25.115 min Scan# 3

Delta R.T. 0.012 min

Lab File: BP021482.D

Acq: 13 Aug 2024 13:15

Instrument :

BNA_P



ClientSampleId :

SSTDICC005


Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel 08/14/2024


Supervised By :mohammad ahmed 08/14/2024



Abundance Scan 4416 (28.962 min): BP021482.D\data.ms (-)



Abundance Scan 4416 (28.962 min): BP021482.D\data.ms (-)



#87

Indeno(1,2,3-cd)pyrene

Concen: 4.800 ng m

RT: 28.962 min Scan# 4416

Delta R.T. -0.030 min

Lab File: BP021482.D

Acq: 13 Aug 2024 13:15


Tgt Ion:276 Resp: 715393

Ion Ratio Lower Upper

276 100

138 15.3 12.8 19.2

277 19.6 16.5 24.7



#88

Benzo(b)fluoranthene

Concen: 4.816 ng

RT: 24.015 min Scan# 3

Delta R.T. -0.012 min

Lab File: BP021482.D



Acq: 13 Aug 2024 13:15

Instrument :

BNA_P

ClientSampleId :





SSTDICC005



| Tgt | Ion:252 | Resp: | 63602 |
|-----|---------|-------|-------|
| Ion | Ratio | Lower | Upper |
| 252 | 100 | | |
| 253 | 21.7 | 17.6 | 26.4 |
| 125 | 10.5 | 8.6 | 13.0 |


Manual Integrations APPROVED



Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#89
Benzo(k)fluoranthene
Concen: 4.975 ng
RT: 24.092 min Scan# 3588
Delta R.T. -0.018 min
Lab File: BP021482.D
Acq: 13 Aug 2024 13:15

| Tgt | Ion:252 | Resp: | 640982 |
|-----|---------|-------|--------|
| Ion | Ratio | Lower | Upper |
| 252 | 100 | | |
| 253 | 21.8 | 17.5 | 26.3 |
| 125 | 10.9 | 7.8 | 11.8 |





#90

Benzo(a)pyrene

Concen: 4.783 ng

RT: 24.939 min Scan# 3

Delta R.T. -0.012 min

Lab File: BP021482.D

Acq: 13 Aug 2024 13:15

Instrument :

BNA_P


ClientSampleId :

SSTDICC005

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#91

Dibenzo(a,h)anthracene

Concen: 4.802 ng

RT: 29.068 min Scan# 4434

Delta R.T. -0.018 min

Lab File: BP021482.D

Acq: 13 Aug 2024 13:15


Tgt Ion:278 Resp: 594702


Ion Ratio Lower Upper

278 100

139 14.9 12.2 18.4

279 22.6 19.1 28.7





#92

Benzo(g,h,i)perylene

Concen: 4.872 ng

RT: 30.162 min Scan# 4

Delta R.T. -0.012 min

Lab File: BP021482.D

Acq: 13 Aug 2024 13:15

Instrument :

BNA_P


ClientSampleId :

SSTDICC005

Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081324\
 Data File : BP021483.D
 Acq On : 13 Aug 2024 13:59
 Operator : RC/JU
 Sample : SSTDICC010
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC010

Quant Time: Aug 13 23:50:33 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 13 23:44:24 2024
 Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024
 Supervised By :mohammad ahmed 08/14/2024

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|------------------------------------|--------|------|----------|--------|-------|----------|
| Internal Standards | | | | | | |
| 1) 1,4-Dichlorobenzene-d4 | 7.805 | 152 | 385438 | 20.000 | ng | 0.00 |
| 21) Naphthalene-d8 | 10.599 | 136 | 1772285 | 20.000 | ng | 0.00 |
| 39) Acenaphthene-d10 | 14.451 | 164 | 1303990 | 20.000 | ng | 0.00 |
| 64) Phenanthrene-d10 | 17.251 | 188 | 2799153 | 20.000 | ng | 0.00 |
| 76) Chrysene-d12 | 21.710 | 240 | 2268651 | 20.000 | ng | 0.01 |
| 86) Perylene-d12 | 25.110 | 264 | 2172035 | 20.000 | ng | 0.00 |
| System Monitoring Compounds | | | | | | |
| 5) 2-Fluorophenol | 5.393 | 112 | 495524 | 19.090 | ng | 0.00 |
| 7) Phenol-d6 | 6.958 | 99 | 755554 | 19.953 | ng | 0.00 |
| 23) Nitrobenzene-d5 | 8.952 | 82 | 600158 | 17.479 | ng | 0.00 |
| 42) 2,4,6-Tribromophenol | 15.963 | 330 | 266074 | 18.334 | ng | 0.00 |
| 45) 2-Fluorobiphenyl | 13.057 | 172 | 1713239 | 20.051 | ng | 0.00 |
| 79) Terphenyl-d14 | 19.963 | 244 | 2713318 | 23.235 | ng | -0.01 |
| Target Compounds | | | | | | |
| | | | | Qvalue | | |
| 2) 1,4-Dioxane | 3.317 | 88 | 118289 | 9.611 | ng | 97 |
| 3) Pyridine | 3.705 | 79 | 331408 | 9.645 | ng | 98 |
| 4) n-Nitrosodimethylamine | 3.617 | 42 | 101144 | 9.794 | ng | # 99 |
| 6) Aniline | 7.128 | 93 | 392151 | 10.359 | ng | 99 |
| 8) 2-Chlorophenol | 7.369 | 128 | 235350 | 9.811 | ng | 97 |
| 9) Benzaldehyde | 6.946 | 77 | 278892 | 11.503 | ng | 99 |
| 10) Phenol | 6.987 | 94 | 389958 | 9.945 | ng | 99 |
| 11) bis(2-Chloroethyl)ether | 7.228 | 93 | 346942 | 10.456 | ng | 99 |
| 12) 1,3-Dichlorobenzene | 7.699 | 146 | 285122 | 10.006 | ng | 98 |
| 13) 1,4-Dichlorobenzene | 7.840 | 146 | 291753 | 10.140 | ng | 99 |
| 14) 1,2-Dichlorobenzene | 8.158 | 146 | 284225 | 10.251 | ng | 99 |
| 15) Benzyl Alcohol | 8.034 | 79 | 280251 | 10.315 | ng | 98 |
| 16) 2,2'-oxybis(1-Chloropr... | 8.334 | 45 | 330688m | 10.822 | ng | |
| 17) 2-Methylphenol | 8.228 | 107 | 252971 | 10.205 | ng | 98 |
| 18) Hexachloroethane | 8.893 | 117 | 113709 | 9.855 | ng | 99 |
| 19) n-Nitroso-di-n-propyla... | 8.599 | 70 | 299730 | 11.459 | ng | 99 |
| 20) 3+4-Methylphenols | 8.558 | 107 | 350826 | 10.496 | ng | 98 |
| 22) Acetophenone | 8.616 | 105 | 550125 | 10.104 | ng | # 98 |
| 24) Nitrobenzene | 8.993 | 77 | 341460 | 9.156 | ng | 96 |
| 25) Isophorone | 9.516 | 82 | 859297 | 10.593 | ng | 99 |
| 26) 2-Nitrophenol | 9.705 | 139 | 55468 | 10.066 | ng | 94 |
| 27) 2,4-Dimethylphenol | 9.757 | 122 | 199430 | 9.962 | ng | 98 |
| 28) bis(2-Chloroethoxy)met... | 10.005 | 93 | 508512 | 10.270 | ng | 99 |
| 29) 2,4-Dichlorophenol | 10.240 | 162 | 220504 | 9.322 | ng | 98 |
| 30) 1,2,4-Trichlorobenzene | 10.457 | 180 | 293816 | 9.689 | ng | 96 |
| 31) Naphthalene | 10.646 | 128 | 922302 | 9.962 | ng | 99 |
| 32) Benzoic acid | 9.828 | 122 | 72405 | 10.529 | ng | 89 |
| 33) 4-Chloroaniline | 10.740 | 127 | 357126 | 10.269 | ng | 98 |
| 34) Hexachlorobutadiene | 10.946 | 225 | 188389 | 9.867 | ng | 99 |
| 35) Caprolactam | 11.481 | 113 | 101824 | 10.344 | ng | 98 |
| 36) 4-Chloro-3-methylphenol | 11.863 | 107 | 332115 | 10.122 | ng | 98 |
| 37) 2-Methylnaphthalene | 12.257 | 142 | 657592 | 10.407 | ng | 99 |
| 38) 1-Methylnaphthalene | 12.475 | 142 | 651444 | 10.453 | ng | 99 |
| 40) 1,2,4,5-Tetrachloroben... | 12.628 | 216 | 354329 | 9.274 | ng | 100 |
| 41) Hexachlorocyclopentadiene | 12.622 | 237 | 99734 | 8.236 | ng | 96 |
| 43) 2,4,6-Trichlorophenol | 12.863 | 196 | 185031 | 8.718 | ng | 96 |

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081324\
 Data File : BP021483.D
 Acq On : 13 Aug 2024 13:59
 Operator : RC/JU
 Sample : SSTDICC010
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC010

Quant Time: Aug 13 23:50:33 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 13 23:44:24 2024
 Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024
 Supervised By :mohammad ahmed 08/14/2024

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|-------------------------------|--------|------|----------|--------|-------|----------|
| 44) 2,4,5-Trichlorophenol | 12.934 | 196 | 206114 | 8.741 | ng | 99 |
| 46) 1,1'-Biphenyl | 13.275 | 154 | 912835 | 9.894 | ng | 99 |
| 47) 2-Chloronaphthalene | 13.310 | 162 | 700953 | 9.759 | ng | 99 |
| 48) 2-Nitroaniline | 13.504 | 65 | 117865 | 10.374 | ng | 97 |
| 49) Acenaphthylene | 14.169 | 152 | 1049611 | 9.941 | ng | 100 |
| 50) Dimethylphthalate | 13.893 | 163 | 889818 | 10.190 | ng | 100 |
| 51) 2,6-Dinitrotoluene | 14.004 | 165 | 122236 | 10.348 | ng | 90 |
| 52) Acenaphthene | 14.516 | 154 | 690129 | 9.961 | ng | 100 |
| 53) 3-Nitroaniline | 14.328 | 138 | 120754 | 10.365 | ng | # 95 |
| 54) 2,4-Dinitrophenol | 14.540 | 184 | 29452 | 10.721 | ng | # 52 |
| 55) Dibenzofuran | 14.851 | 168 | 1067432 | 10.016 | ng | 100 |
| 56) 4-Nitrophenol | 14.628 | 139 | 91997 | 11.027 | ng | 92 |
| 57) 2,4-Dinitrotoluene | 14.804 | 165 | 139246 | 10.586 | ng | 94 |
| 58) Fluorene | 15.522 | 166 | 912792 | 10.425 | ng | 99 |
| 59) 2,3,4,6-Tetrachlorophenol | 15.075 | 232 | 181235 | 9.055 | ng | 99 |
| 60) Diethylphthalate | 15.292 | 149 | 909304 | 10.148 | ng | 98 |
| 61) 4-Chlorophenyl-phenyle... | 15.516 | 204 | 471863 | 10.426 | ng | 98 |
| 62) 4-Nitroaniline | 15.516 | 138 | 129100 | 9.743 | ng | 87 |
| 63) Azobenzene | 15.810 | 77 | 1159642 | 10.529 | ng | 99 |
| 65) 4,6-Dinitro-2-methylph... | 15.581 | 198 | 44079 | 11.022 | ng | 91 |
| 66) n-Nitrosodiphenylamine | 15.722 | 169 | 757847 | 10.103 | ng | 100 |
| 67) 4-Bromophenyl-phenylether | 16.422 | 248 | 293464 | 9.906 | ng | 98 |
| 68) Hexachlorobenzene | 16.545 | 284 | 347933 | 9.938 | ng | 96 |
| 69) Atrazine | 16.698 | 200 | 260277 | 10.106 | ng | 99 |
| 70) Pentachlorophenol | 16.892 | 266 | 150520 | 7.877 | ng | 99 |
| 71) Phenanthrene | 17.298 | 178 | 1348415 | 10.017 | ng | 99 |
| 72) Anthracene | 17.392 | 178 | 1336422 | 10.123 | ng | 99 |
| 73) Carbazole | 17.657 | 167 | 1267960 | 9.966 | ng | 99 |
| 74) Di-n-butylphthalate | 18.275 | 149 | 1490471 | 9.539 | ng | 99 |
| 75) Fluoranthene | 19.375 | 202 | 1613504 | 9.769 | ng | 100 |
| 77) Benzidine | 19.569 | 184 | 336598 | 9.509 | ng | 99 |
| 78) Pyrene | 19.751 | 202 | 1667213 | 11.243 | ng | 99 |
| 80) Butylbenzylphthalate | 20.704 | 149 | 414136 | 10.409 | ng | 96 |
| 81) Benzo(a)anthracene | 21.680 | 228 | 1415007 | 9.936 | ng | 100 |
| 82) 3,3'-Dichlorobenzidine | 21.598 | 252 | 405660 | 8.884 | ng | 99 |
| 83) Chrysene | 21.751 | 228 | 1338613 | 9.955 | ng | 99 |
| 84) Bis(2-ethylhexyl)phtha... | 21.639 | 149 | 720802 | 8.563 | ng | 100 |
| 85) Di-n-octyl phthalate | 22.951 | 149 | 951183 | 10.621 | ng | 99 |
| 87) Indeno(1,2,3-cd)pyrene | 28.974 | 276 | 1255326 | 8.917 | ng | 98 |
| 88) Benzo(b)fluoranthene | 24.021 | 252 | 1250809 | 10.027 | ng | 99 |
| 89) Benzo(k)fluoranthene | 24.092 | 252 | 1195420 | 9.824 | ng | 99 |
| 90) Benzo(a)pyrene | 24.927 | 252 | 1035462 | 9.565 | ng | 99 |
| 91) Dibenzo(a,h)anthracene | 29.051 | 278 | 1047693 | 8.956 | ng | 99 |
| 92) Benzo(g,h,i)perylene | 30.168 | 276 | 1037272 | 8.847 | ng | 99 |

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081324\
 Data File : BP021483.D
 Acq On : 13 Aug 2024 13:59
 Operator : RC/JU
 Sample : SSTDICC010
 Misc :
 ALS Vial : 4 Sample Multiplier: 1


Quant Time: Aug 13 23:50:33 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 13 23:44:24 2024
 Response via : Initial Calibration

Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC010

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024
 Supervised By :mohammad ahmed 08/14/2024







#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.805 min Scan# 8
Delta R.T. -0.006 min
Lab File: BP021483.D
Acq: 13 Aug 2024 13:59

Instrument : BNA_P
ClientSampleId : SSTDICC010


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#2
1,4-Dioxane
Concen: 9.611 ng
RT: 3.317 min Scan# 56
Delta R.T. -0.000 min
Lab File: BP021483.D
Acq: 13 Aug 2024 13:59

Tgt Ion: 88 Resp: 118289
Ion Ratio Lower Upper
88 100
58 63.3 48.7 73.1
43 19.9 15.4 23.2







#3
Pyridine
Concen: 9.645 ng
RT: 3.705 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP021483.D
Acq: 13 Aug 2024 13:59

Instrument : BNA_P
ClientSampleId : SSTDICC010


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#4
n-Nitrosodimethylamine
Concen: 9.794 ng
RT: 3.617 min Scan# 107
Delta R.T. -0.000 min
Lab File: BP021483.D
Acq: 13 Aug 2024 13:59

Tgt Ion: 42 Resp: 101144
Ion Ratio Lower Upper
42 100
74 194.5 154.3 231.5
44 12.0 6.2 9.4#






#5
2-Fluorophenol
Concen: 19.090 ng
RT: 5.393 min Scan# 409
Delta R.T. -0.000 min
Lab File: BP021483.D
Acq: 13 Aug 2024 13:59

Instrument : BNA_P
ClientSampleId : SSTDICC010


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024




#6
Aniline
Concen: 10.359 ng
RT: 7.128 min Scan# 704
Delta R.T. -0.006 min
Lab File: BP021483.D
Acq: 13 Aug 2024 13:59


Tgt Ion: 93 Resp: 392151
Ion Ratio Lower Upper
93 100
66 37.6 30.2 45.2
65 19.3 14.7 22.1




Abundance Scan 676 (6.964 min): BP021485.D\data.ms (-66)



Abundance Scan 675 (6.958 min): BP021483.D\data.ms



Abundance Scan 675 (6.958 min): BP021483.D\data.ms (-62)



#7

Phenol-d6

Concen: 19.953 ng

RT: 6.958 min Scan# 6

Delta R.T. -0.006 min

Lab File: BP021483.D

Acq: 13 Aug 2024 13:59

Instrument :

BNA_P

ClientSampleId :

SSTDICC010


**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/14/2024


Supervised By :mohammad ahmed 08/14/2024

Abundance


6.958




Abundance Scan 746 (7.375 min): BP021485.D\data.ms (-73)



Abundance Scan 745 (7.369 min): BP021483.D\data.ms



Abundance Scan 745 (7.369 min): BP021483.D\data.ms (-69)



#8

2-Chlorophenol

Concen: 9.811 ng

RT: 7.369 min Scan# 745

Delta R.T. -0.006 min

Lab File: BP021483.D

Acq: 13 Aug 2024 13:59

Tgt Ion:128 Resp: 235350

Ion Ratio Lower Upper


128 100

130 31.6 11.6 51.6


64 58.6 35.3 75.3

Abundance

7.369



Abundance Scan 673 (6.946 min): BP021485.D\data.ms (-66)



#9

Benzaldehyde

Concen: 11.503 ng

RT: 6.946 min Scan# 6

Delta R.T. -0.000 min

Lab File: BP021483.D

Acq: 13 Aug 2024 13:59


Instrument :

BNA_P


ClientSampleId :

SSTDICC010

m/z--> Scan 673 (6.946 min): BP021483.D\data.ms




Abundance Scan 673 (6.946 min): BP021483.D\data.ms (-62)




Abundance

6.946



Abundance Scan 680 (6.987 min): BP021485.D\data.ms (-67)



#10

Phenol

Concen: 9.945 ng


RT: 6.987 min Scan# 680

Delta R.T. -0.000 min

Lab File: BP021483.D


Acq: 13 Aug 2024 13:59

Abundance Scan 680 (6.987 min): BP021483.D\data.ms







Abundance

6.987



Abundance Scan 680 (6.987 min): BP021483.D\data.ms (-62)









#11
bis(2-Chloroethyl)ether
Concen: 10.456 ng
RT: 7.228 min Scan# 7
Delta R.T. -0.000 min
Lab File: BP021483.D
Acq: 13 Aug 2024 13:59

Instrument : BNA_P
ClientSampleId : SSTDICC010


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#12
1,3-Dichlorobenzene
Concen: 10.006 ng
RT: 7.699 min Scan# 801
Delta R.T. -0.000 min
Lab File: BP021483.D
Acq: 13 Aug 2024 13:59

Tgt Ion:146 Resp: 285122
Ion Ratio Lower Upper
146 100
148 63.6 51.8 77.6
75 41.0 31.3 46.9





#13

1,4-Dichlorobenzene

Concen: 10.140 ng

RT: 7.840 min Scan# 8

Delta R.T. -0.000 min

Lab File: BP021483.D

Acq: 13 Aug 2024 13:59

Instrument :

BNA_P

ClientSampleId :

SSTDICC010

Tgt

Ion:146

Ion Ratio

146

100

148

63.4

111

49.6

Resp:

29175

Lower

51.5

77.3

40.3



60.5

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#14

1,2-Dichlorobenzene

Concen: 10.251 ng

RT: 8.158 min Scan# 879

Delta R.T. -0.000 min

Lab File: BP021483.D

Acq: 13 Aug 2024 13:59

Tgt

Ion:146

Ion Ratio

146

100

148

64.2

111

Resp:

284225


Lower


51.5

77.3


41.1

61.7





Ref Scan 858 (8.034 min): BP021483.D\data.ms



Sub Scan 858 (8.034 min): BP021483.D\data.ms (-80)

Time-->

#15

Benzyl Alcohol

Concen: 10.315 ng

RT: 8.034 min Scan# 8

Delta R.T. -0.000 min

Lab File: BP021483.D

Acq: 13 Aug 2024 13:59

Instrument :

BNA_P

ClientSampleId :

SSTDICC010

Tgt Ion: 79 Resp: 280253

Ion Ratio Lower Upper

79 100

108 72.4 57.9 86.9


77 67.5 51.8 77.8

Manual Integrations

APPROVED


Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



Ref Scan 909 (8.334 min): BP021483.D\data.ms

Time-->



Sub Scan 909 (8.334 min): BP021483.D\data.ms (-85)

Time-->

#16

2,2'-oxybis(1-Chloropropane)

Concen: 10.822 ng

RT: 8.334 min Scan# 909

Delta R.T. -0.000 min

Lab File: BP021483.D

Acq: 13 Aug 2024 13:59


Tgt Ion: 45 Resp: 330688

Ion Ratio Lower Upper

45 100

77 21.4 1.9 41.9

79 16.1 0.0 36.6



Sub Scan 909 (8.334 min): BP021483.D\data.ms (-85)

Time-->

Abundance




8.334

100000

50000

0

8.30 8.334 8.40



#17

2-Methylphenol

Concen: 10.205 ng

RT: 8.228 min Scan# 8

Delta R.T. -0.006 min

Lab File: BP021483.D

Acq: 13 Aug 2024 13:59

Instrument :

BNA_P




ClientSampleId :

SSTDICC010

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#18

Hexachloroethane

Concen: 9.855 ng

RT: 8.893 min Scan# 1004

Delta R.T. -0.000 min

Lab File: BP021483.D

Acq: 13 Aug 2024 13:59


Tgt Ion:117 Resp: 113709


Ion Ratio Lower Upper

117 100

119 93.5 73.9 110.9

201 92.1 74.9 112.3







#19
n-Nitroso-di-n-propylamine
Concen: 11.459 ng
RT: 8.599 min Scan# 9
Delta R.T. -0.006 min
Lab File: BP021483.D
Acq: 13 Aug 2024 13:59

Instrument :
BNA_P
ClientSampleId :
SSTDICC010


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#20
3+4-Methylphenols
Concen: 10.496 ng
RT: 8.558 min Scan# 947
Delta R.T. -0.000 min
Lab File: BP021483.D
Acq: 13 Aug 2024 13:59

Tgt Ion:107 Resp: 350826
Ion Ratio Lower Upper
107 100
108 90.2 67.6 107.6
77 35.1 15.0 55.0
79 30.9 8.9 48.9





#21

Naphthalene-d8

Concen: 20.000 ng

RT: 10.599 min Scan# 1

Delta R.T. -0.000 min

Lab File: BP021483.D

Acq: 13 Aug 2024 13:59

Instrument :

BNA_P



ClientSampleId :

SSTDICC010

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#22

Acetophenone

Concen: 10.104 ng

RT: 8.616 min Scan# 957

Delta R.T. -0.000 min

Lab File: BP021483.D

Acq: 13 Aug 2024 13:59

Tgt Ion:105 Resp: 550125


Ion Ratio Lower Upper




105 100

71 3.7 4.6 7.0#

51 19.6 16.3 24.5

120 19.6 16.2 24.2





#23

Nitrobenzene-d5

Concen: 17.479 ng

RT: 8.952 min Scan# 1

Delta R.T. -0.000 min

Lab File: BP021483.D

Acq: 13 Aug 2024 13:59

Instrument :

BNA_P

ClientSampleId :

SSTDICC010




**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024

Abundance

8.952



#24

Nitrobenzene

Concen: 9.156 ng

RT: 8.993 min Scan# 1021

Delta R.T. -0.000 min

Lab File: BP021483.D

Acq: 13 Aug 2024 13:59

Tgt Ion: 77 Resp: 341460

Ion Ratio Lower Upper


77 100


123 34.3 29.5 44.3

65 12.9 9.9 14.9

Abundance

8.993







#25
 Isophorone
 Concen: 10.593 ng
 RT: 9.516 min Scan# 1111
 Delta R.T. -0.006 min
 Lab File: BP021483.D
 Acq: 13 Aug 2024 13:59

Instrument : BNA_P
 ClientSampleId : SSTDICC010


Manual Integrations
APPROVED




Reviewed By :Yogesh Patel 08/14/2024
 Supervised By :mohammad ahmed 08/14/2024



#26
 2-Nitrophenol
 Concen: 10.066 ng
 RT: 9.705 min Scan# 1142
 Delta R.T. 0.006 min
 Lab File: BP021483.D
 Acq: 13 Aug 2024 13:59

Tgt Ion:139 Resp: 55468
 Ion Ratio Lower Upper
 139 100
 109 41.2 28.9 43.3
 65 64.4 49.0 73.6





#27

2,4-Dimethylphenol

Concen: 9.962 ng

RT: 9.757 min Scan# 1

Delta R.T. -0.000 min

Lab File: BP021483.D

Acq: 13 Aug 2024 13:59

Instrument :

BNA_P





ClientSampleId :

SSTDICC010

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#28

bis(2-Chloroethoxy)methane

Concen: 10.270 ng

RT: 10.005 min Scan# 1193

Delta R.T. -0.000 min

Lab File: BP021483.D

Acq: 13 Aug 2024 13:59


Tgt Ion: 93 Resp: 508512




Ion Ratio Lower Upper

93 100

95 32.3 26.2 39.4

123 9.4 7.6 11.4





#29

2,4-Dichlorophenol

Concen: 9.322 ng

RT: 10.240 min Scan# 1

Delta R.T. -0.000 min

Lab File: BP021483.D

Acq: 13 Aug 2024 13:59

Instrument :

BNA_P





ClientSampleId :

SSTDICC010

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#30

1,2,4-Trichlorobenzene

Concen: 9.689 ng

RT: 10.457 min Scan# 1270

Delta R.T. -0.000 min

Lab File: BP021483.D

Acq: 13 Aug 2024 13:59


Tgt Ion:180 Resp: 293816


Ion Ratio Lower Upper

180 100

182 93.0 78.3 117.5

145 29.5 24.0 36.0





#31

Naphthalene

Concen: 9.962 ng

RT: 10.646 min Scan# 1

Delta R.T. -0.000 min

Lab File: BP021483.D

Acq: 13 Aug 2024 13:59

Instrument :

BNA_P

ClientSampleId :

SSTDICC010

Tgt Ion:128 Resp: 92230

Ion Ratio Lower Upper

128 100

129 11.2 8.6 12.8

127 13.4 10.6 15.8


Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024

Abundance



#32

Benzoic acid

Concen: 10.529 ng

RT: 9.828 min Scan# 1163

Delta R.T. -0.047 min

Lab File: BP021483.D

Acq: 13 Aug 2024 13:59

Tgt Ion:122 Resp: 72405


Ion Ratio Lower Upper

122 100

105 129.6 124.6 164.6

77 106.7 96.2 136.2

Abundance



#33

4-Chloroaniline

Concen: 10.269 ng

RT: 10.740 min Scan# 1

Delta R.T. -0.000 min

Lab File: BP021483.D




Acq: 13 Aug 2024 13:59

Instrument :

BNA_P

ClientSampleId :

SSTDICC010



Tgt Ion:127 Resp: 357120

Ion Ratio Lower Upper

127 100

129 32.1 24.5 36.7





65 36.5 28.9 43.3

92 24.9 19.3 28.9

Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#34

Hexachlorobutadiene

Concen: 9.867 ng

RT: 10.946 min Scan# 1353

Delta R.T. 0.006 min

Lab File: BP021483.D

Acq: 13 Aug 2024 13:59


Tgt Ion:225 Resp: 188389


Ion Ratio Lower Upper

225 100

223 62.1 49.6 74.4

227 61.7 50.1 75.1







#35
Caprolactam
Concen: 10.344 ng
RT: 11.481 min Scan# 1447
Delta R.T. -0.018 min
Lab File: BP021483.D
Acq: 13 Aug 2024 13:59

Instrument : BNA_P
ClientSampleId : SSTDICC010


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#36
4-Chloro-3-methylphenol
Concen: 10.122 ng
RT: 11.863 min Scan# 1509
Delta R.T. -0.000 min
Lab File: BP021483.D
Acq: 13 Aug 2024 13:59

Tgt Ion:107 Resp: 332115
Ion Ratio Lower Upper
107 100
144 21.2 16.6 24.8
142 65.3 51.0 76.4





#37

2-Methylnaphthalene

Concen: 10.407 ng

RT: 12.257 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP021483.D

Acq: 13 Aug 2024 13:59

Instrument :

BNA_P



ClientSampleId :

SSTDICC010

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#38

1-Methylnaphthalene

Concen: 10.453 ng

RT: 12.475 min Scan# 1613

Delta R.T. -0.006 min

Lab File: BP021483.D

Acq: 13 Aug 2024 13:59


Tgt Ion:142 Resp: 651444


Ion Ratio Lower Upper

142 100


141 91.1 73.5 110.3

116 4.5 3.4 5.2






Abundance Scan 1949 (14.451 min): BP021483.D\data.ms (-)



Abundance Scan 1949 (14.451 min): BP021483.D\data.ms (-)



#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 14.451 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP021483.D

Acq: 13 Aug 2024 13:59

Instrument :

BNA_P


ClientSampleId :

SSTDICC010


**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/14/2024


Supervised By :mohammad ahmed 08/14/2024



Abundance Scan 1639 (12.628 min): BP021483.D\data.ms (-)



Abundance Scan 1639 (12.628 min): BP021483.D\data.ms (-)



#40

1,2,4,5-Tetrachlorobenzene

Concen: 9.274 ng

RT: 12.628 min Scan# 1639

Delta R.T. -0.000 min

Lab File: BP021483.D

Acq: 13 Aug 2024 13:59

Tgt Ion:216 Resp: 354329


Ion Ratio Lower Upper


216 100

214 78.1 62.5 93.7

179 20.4 16.4 24.6

108 21.1 16.7 25.1







#41
Hexachlorocyclopentadiene
Concen: 8.236 ng
RT: 12.622 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP021483.D
Acq: 13 Aug 2024 13:59

Instrument :
BNA_P
ClientSampleId :
SSTDICC010


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#42
2,4,6-Tribromophenol
Concen: 18.334 ng
RT: 15.963 min Scan# 2206
Delta R.T. -0.006 min
Lab File: BP021483.D
Acq: 13 Aug 2024 13:59

Tgt Ion:330 Resp: 266074
Ion Ratio Lower Upper
330 100
332 96.6 77.4 116.0
141 31.7 24.5 36.7



#43

2,4,6-Trichlorophenol

Concen: 8.718 ng

RT: 12.863 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP021483.D

Acq: 13 Aug 2024 13:59

Instrument :

BNA_P

ClientSampleId :

SSTDICC010

Tgt Ion:196 Resp: 18503

Ion Ratio Lower Upper

196 100

198 100.1 76.6 114.8


200 32.5 24.5 36.7

Manual Integrations

APPROVED



Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024




#44
2,4,5-Trichlorophenol
Concen: 8.741 ng
RT: 12.934 min Scan# 1691
Delta R.T. -0.000 min
Lab File: BP021483.D
Acq: 13 Aug 2024 13:59




Tgt Ion:196 Resp: 206114
Ion Ratio Lower Upper
196 100
198 97.3 78.9 118.3
97 61.0 49.4 74.0
132 25.7 21.1 31.7



#44
2,4,5-Trichlorophenol
Concen: 8.741 ng
RT: 12.934 min Scan# 1691
Delta R.T. -0.000 min
Lab File: BP021483.D
Acq: 13 Aug 2024 13:59

Tgt Ion:196 Resp: 206114
Ion Ratio Lower Upper
196 100
198 97.3 78.9 118.3
97 61.0 49.4 74.0
132 25.7 21.1 31.7





#45

2-Fluorobiphenyl

Concen: 20.051 ng

RT: 13.057 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP021483.D

Acq: 13 Aug 2024 13:59

Instrument :

BNA_P




ClientSampleId :

SSTDICC010

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#46

1,1'-Biphenyl

Concen: 9.894 ng

RT: 13.275 min Scan# 1749

Delta R.T. -0.006 min

Lab File: BP021483.D

Acq: 13 Aug 2024 13:59



Tgt Ion:154 Resp: 912835


Ion Ratio Lower Upper

154 100

153 39.6 20.7 60.7

76 17.6 0.0 37.7







#47
2-Chloronaphthalene
Concen: 9.759 ng
RT: 13.310 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP021483.D
Acq: 13 Aug 2024 13:59

Instrument : BNA_P
ClientSampleId : SSTDICC010


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#48
2-Nitroaniline
Concen: 10.374 ng
RT: 13.504 min Scan# 1788
Delta R.T. 0.000 min
Lab File: BP021483.D
Acq: 13 Aug 2024 13:59

Tgt Ion: 65 Resp: 117865
Ion Ratio Lower Upper
65 100
92 74.7 58.1 87.1
138 84.7 65.5 98.3



#49

Acenaphthylene

Concen: 9.941 ng

RT: 14.169 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP021483.D




Acq: 13 Aug 2024 13:59

Instrument :

BNA_P

ClientSampleId :

SSTDICC010



Tgt Ion:152 Resp: 104961:

Ion Ratio Lower Upper

152 100

151 20.0 15.9 23.9





153 13.0 10.6 16.0

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#50

Dimethylphthalate

Concen: 10.190 ng

RT: 13.893 min Scan# 1854

Delta R.T. -0.017 min

Lab File: BP021483.D

Acq: 13 Aug 2024 13:59


Tgt Ion:163 Resp: 889818


Ion Ratio Lower Upper

163 100

194 3.9 3.3 4.9

164 10.4 8.3 12.5







#51
2,6-Dinitrotoluene
Concen: 10.348 ng
RT: 14.004 min Scan# 1
Delta R.T. -0.018 min
Lab File: BP021483.D
Acq: 13 Aug 2024 13:59

Instrument : BNA_P
ClientSampleId : SSTDICC010


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#52
Acenaphthene
Concen: 9.961 ng
RT: 14.516 min Scan# 1960
Delta R.T. -0.006 min
Lab File: BP021483.D
Acq: 13 Aug 2024 13:59

Tgt Ion:154 Resp: 690129
Ion Ratio Lower Upper
154 100
153 113.8 91.1 136.7
152 53.5 43.0 64.4







#53
3-Nitroaniline
Concen: 10.365 ng
RT: 14.328 min Scan# 1
Delta R.T. -0.023 min
Lab File: BP021483.D
Acq: 13 Aug 2024 13:59

Instrument : BNA_P
ClientSampleId : SSTDICC010


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#54
2,4-Dinitrophenol
Concen: 10.721 ng
RT: 14.540 min Scan# 1964
Delta R.T. -0.011 min
Lab File: BP021483.D
Acq: 13 Aug 2024 13:59

Tgt Ion:184 Resp: 29452
Ion Ratio Lower Upper
184 100
63 81.7 64.9 97.3
154 149.0 52.1 78.1#







#55
Dibenzofuran
Concen: 10.016 ng
RT: 14.851 min Scan# 2
Instrument : BNA_P
Delta R.T. -0.006 min
Lab File: BP021483.D
Acq: 13 Aug 2024 13:59
ClientSampleId : SSTDICC010


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#56
4-Nitrophenol
Concen: 11.027 ng
RT: 14.628 min Scan# 1979
Delta R.T. -0.023 min
Lab File: BP021483.D
Acq: 13 Aug 2024 13:59

Tgt Ion:139 Resp: 91997
Ion Ratio Lower Upper
139 100
109 99.2 70.2 110.2
65 129.6 101.8 141.8







#57
2,4-Dinitrotoluene
Concen: 10.586 ng
RT: 14.804 min Scan# 2130
Delta R.T. -0.006 min
Lab File: BP021483.D
Acq: 13 Aug 2024 13:59

Instrument : BNA_P
ClientSampleId : SSTDICC010


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#58
Fluorene
Concen: 10.425 ng
RT: 15.522 min Scan# 2131
Delta R.T. 0.006 min
Lab File: BP021483.D
Acq: 13 Aug 2024 13:59

Tgt Ion:166 Resp: 912792
Ion Ratio Lower Upper
166 100
165 98.2 78.1 117.1
167 13.9 11.0 16.4



#59

2,3,4,6-Tetrachlorophenol

Concen: 9.055 ng

RT: 15.075 min Scan# 2

Delta R.T. -0.012 min

Lab File: BP021483.D




Acq: 13 Aug 2024 13:59

Instrument :

BNA_P

ClientSampleId :

SSTDICC010



Tgt Ion:232 Resp: 181239

Ion Ratio Lower Upper

232 100

131 40.9 33.1 49.7





130 2.3 1.7 2.5

166 30.0 23.0 34.4

Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024

#60
Diethylphthalate
Concen: 10.148 ng
RT: 15.292 min Scan# 2092
Delta R.T. -0.012 min
Lab File: BP021483.D
Acq: 13 Aug 2024 13:59


Tgt Ion:149 Resp: 909304


Ion Ratio Lower Upper

149 100


177 23.4 17.8 26.8

150 12.6 10.1 15.1





Abundance Scan 2130 (15.516 min): BP021483.D\data.ms



#61

4-Chlorophenyl-phenylether

Concen: 10.426 ng

RT: 15.516 min Scan# 2130

Delta R.T. -0.006 min

Lab File: BP021483.D

Acq: 13 Aug 2024 13:59

Instrument : BNA_P

ClientSampleId : SSTDICC010

Tgt Ion:204 Resp: 47186

Ion Ratio Lower Upper

204 100

206 33.0 26.4 39.6



141 55.2 46.1 69.1

Manual Integrations


APPROVED

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



Abundance Scan 2130 (15.516 min): BP021483.D\data.ms



#62

4-Nitroaniline

Concen: 9.743 ng

RT: 15.516 min Scan# 2130

Delta R.T. -0.012 min

Lab File: BP021483.D

Acq: 13 Aug 2024 13:59


Tgt Ion:138 Resp: 129100







Ion Ratio Lower Upper

138 100

92 71.5 46.4 86.4

108 110.9 73.2 113.2





#63

Azobenzene

Concen: 10.529 ng

RT: 15.810 min Scan# 2

Delta R.T. -0.012 min

Lab File: BP021483.D

Acq: 13 Aug 2024 13:59

Instrument :

BNA_P

ClientSampleId :


SSTDICC010

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 17.251 min Scan# 2425

Delta R.T. -0.000 min

Lab File: BP021483.D

Acq: 13 Aug 2024 13:59


Tgt Ion:188 Resp: 2799153


Ion Ratio Lower Upper

188 100

94 12.4 10.2 15.4

80 13.9 11.6 17.4







#65
4,6-Dinitro-2-methylphenol
Concen: 11.022 ng
RT: 15.581 min Scan# 2142
Delta R.T. -0.006 min
Lab File: BP021483.D
Acq: 13 Aug 2024 13:59

Instrument : BNA_P
ClientSampleId : SSTDICC010


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#66
n-Nitrosodiphenylamine
Concen: 10.103 ng
RT: 15.722 min Scan# 2165
Delta R.T. -0.018 min
Lab File: BP021483.D
Acq: 13 Aug 2024 13:59

Tgt Ion:169 Resp: 757847
Ion Ratio Lower Upper
169 100
168 66.8 53.4 80.2
167 36.2 28.7 43.1







#67
4-Bromophenyl-phenylether
Concen: 9.906 ng
RT: 16.422 min Scan# 2
Delta R.T. -0.018 min
Lab File: BP021483.D
Acq: 13 Aug 2024 13:59

Instrument : BNA_P
ClientSampleId : SSTDICC010


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#68
Hexachlorobenzene
Concen: 9.938 ng
RT: 16.545 min Scan# 2305
Delta R.T. -0.006 min
Lab File: BP021483.D
Acq: 13 Aug 2024 13:59

Tgt Ion:284 Resp: 347933
Ion Ratio Lower Upper
284 100
142 32.8 23.4 35.2
249 29.7 24.1 36.1







#69
Atrazine
Concen: 10.106 ng
RT: 16.698 min Scan# 21
Delta R.T. -0.018 min
Lab File: BP021483.D
Acq: 13 Aug 2024 13:59

Instrument : BNA_P
ClientSampleId : SSTDICC010


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024




#70
Pentachlorophenol
Concen: 7.877 ng
RT: 16.892 min Scan# 2364
Delta R.T. 0.000 min
Lab File: BP021483.D
Acq: 13 Aug 2024 13:59

Tgt Ion:266 Resp: 150520
Ion Ratio Lower Upper
266 100
268 66.3 51.8 77.6
264 63.0 50.5 75.7






Abundance Scan 2433 (17.298 min): BP021483.D\data.ms (-)



Abundance Scan 2433 (17.298 min): BP021483.D\data.ms (-)



#71

Phenanthrene

Concen: 10.017 ng

RT: 17.298 min Scan# 2433

Delta R.T. -0.000 min

Lab File: BP021483.D

Acq: 13 Aug 2024 13:59

Instrument :

BNA_P


ClientSampleId :

SSTDICC010


**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/14/2024


Supervised By :mohammad ahmed 08/14/2024



Abundance Scan 2449 (17.392 min): BP021483.D\data.ms (-)



Abundance Scan 2449 (17.392 min): BP021483.D\data.ms (-)



#72

Anthracene

Concen: 10.123 ng

RT: 17.392 min Scan# 2449

Delta R.T. -0.006 min

Lab File: BP021483.D

Acq: 13 Aug 2024 13:59


Tgt Ion:178 Resp: 1336422

Ion Ratio Lower Upper

178 100

176 18.2 14.9 22.3


179 15.0 12.5 18.7




Abundance

Scan 2449 (17.392 min): BP021483.D\data.ms (-)

Time--> 17.30 17.40 17.50







#73
Carbazole
Concen: 9.966 ng
RT: 17.657 min Scan# 2
Delta R.T. -0.012 min
Lab File: BP021483.D
Acq: 13 Aug 2024 13:59

Instrument : BNA_P
ClientSampleId : SSTDICC010


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#74
Di-n-butylphthalate
Concen: 9.539 ng
RT: 18.275 min Scan# 2599
Delta R.T. -0.000 min
Lab File: BP021483.D
Acq: 13 Aug 2024 13:59

Tgt Ion:149 Resp: 1490471
Ion Ratio Lower Upper
149 100
150 9.0 7.5 11.3
104 6.0 4.9 7.3







#75
Fluoranthene
Concen: 9.769 ng
RT: 19.375 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP021483.D
Acq: 13 Aug 2024 13:59

Instrument : BNA_P
ClientSampleId : SSTDICC010


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#76
Chrysene-d₁₂
Concen: 20.000 ng
RT: 21.710 min Scan# 3183
Delta R.T. 0.012 min
Lab File: BP021483.D
Acq: 13 Aug 2024 13:59

Tgt Ion:240 Resp: 2268651
Ion Ratio Lower Upper
240 100
120 11.7 9.9 14.9
236 26.2 20.6 30.8







#77
Benzidine
Concen: 9.509 ng
RT: 19.569 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP021483.D
Acq: 13 Aug 2024 13:59

Instrument : BNA_P
ClientSampleId : SSTDICC010


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#78
Pyrene
Concen: 11.243 ng
RT: 19.751 min Scan# 2850
Delta R.T. -0.000 min
Lab File: BP021483.D
Acq: 13 Aug 2024 13:59

Tgt Ion:202 Resp: 1667213
Ion Ratio Lower Upper
202 100
200 20.4 16.7 25.1
203 17.1 14.1 21.1







#79
Terphenyl-d14
Concen: 23.235 ng
RT: 19.963 min Scan# 21
Delta R.T. -0.012 min
Lab File: BP021483.D
Acq: 13 Aug 2024 13:59

Instrument : BNA_P
ClientSampleId : SSTDICC010


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024




#80
Butylbenzylphthalate
Concen: 10.409 ng
RT: 20.704 min Scan# 3012
Delta R.T. -0.006 min
Lab File: BP021483.D
Acq: 13 Aug 2024 13:59

Tgt Ion:149 Resp: 414136
Ion Ratio Lower Upper
149 100
91 95.2 73.3 109.9
206 22.4 17.1 25.7






Abundance Scan 3178 (21.680 min): BP021483.D\data.ms (-)



Abundance Scan 3178 (21.680 min): BP021483.D\data.ms (-)



#81

Benzo(a)anthracene

Concen: 9.936 ng

RT: 21.680 min Scan# 3

Instrument :

Delta R.T. -0.000 min

BNA_P

Lab File: BP021483.D

ClientSampleId :



Acq: 13 Aug 2024 13:59

SSTDICC010


**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/14/2024


Supervised By :mohammad ahmed 08/14/2024



Abundance Scan 3164 (21.598 min): BP021483.D\data.ms (-)



Abundance Scan 3164 (21.598 min): BP021483.D\data.ms (-)



#82

3,3'-Dichlorobenzidine

Concen: 8.884 ng

RT: 21.598 min Scan# 3164

Delta R.T. -0.012 min

Lab File: BP021483.D

Acq: 13 Aug 2024 13:59


Tgt Ion:252 Resp: 405660


Ion Ratio Lower Upper

252 100

254 63.9 51.7 77.5

126 10.7 8.7 13.1







#83

Chrysene
Concen: 9.955 ng
RT: 21.751 min Scan# 3
Delta R.T. -0.000 min
Lab File: BP021483.D
Acq: 13 Aug 2024 13:59

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024




#84

Bis(2-ethylhexyl)phthalate
Concen: 8.563 ng
RT: 21.639 min Scan# 3171
Delta R.T. -0.006 min
Lab File: BP021483.D
Acq: 13 Aug 2024 13:59

Tgt Ion:149 Resp: 720802
Ion Ratio Lower Upper
149 100
167 27.8 22.3 33.5
279 4.7 3.8 5.6







#85
Di-n-octyl phthalate
Concen: 10.621 ng
RT: 22.951 min Scan# 3
Delta R.T. -0.012 min
Lab File: BP021483.D
Acq: 13 Aug 2024 13:59

Instrument : BNA_P
ClientSampleId : SSTDICC010


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#86
Perylene-d₁₂
Concen: 20.000 ng
RT: 25.110 min Scan# 3761
Delta R.T. 0.006 min
Lab File: BP021483.D
Acq: 13 Aug 2024 13:59

Tgt Ion:264 Resp: 2172035
Ion Ratio Lower Upper
264 100
260 23.9 18.6 28.0
265 21.3 17.1 25.7



#87

Indeno(1,2,3-cd)pyrene

Concen: 8.917 ng

RT: 28.974 min Scan# 4

Delta R.T. -0.018 min

Lab File: BP021483.D

Acq: 13 Aug 2024 13:59

Instrument : BNA_P

ClientSampleId : SSTDICC010

Tgt Ion:276 Resp: 1255320

Ion Ratio Lower Upper

276 100


138 16.6 12.8 19.2

277 21.5 16.5 24.7

Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#88

Benzo(b)fluoranthene

Concen: 10.027 ng

RT: 24.021 min Scan# 3576

Delta R.T. -0.006 min

Lab File: BP021483.D

Acq: 13 Aug 2024 13:59



Tgt Ion:252 Resp: 1250809

Ion Ratio Lower Upper

252 100

253 22.0 17.6 26.4

125 10.1 8.6 13.0



#89

Benzo(k)fluoranthene

Concen: 9.824 ng

RT: 24.092 min Scan# 3

Delta R.T. -0.018 min

Lab File: BP021483.D




Acq: 13 Aug 2024 13:59

Instrument :

BNA_P

ClientSampleId :

SSTDICC010



Tgt Ion:252 Resp: 1195420

Ion Ratio Lower Upper

252 100





253 22.3 17.5 26.3

125 10.3 7.8 11.8

Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024

#90
Benzo(a)pyrene
Concen: 9.565 ng
RT: 24.927 min Scan# 3730
Delta R.T. -0.024 min
Lab File: BP021483.D
Acq: 13 Aug 2024 13:59


Tgt Ion:252 Resp: 1035462




Ion Ratio Lower Upper

252 100

253 21.8 17.7 26.5

125 11.2 8.7 13.1









#91
Dibenzo(a,h)anthracene
Concen: 8.956 ng
RT: 29.051 min Scan# 4431
Delta R.T. -0.035 min
Lab File: BP021483.D
Acq: 13 Aug 2024 13:59

Instrument :
BNA_P
ClientSampleId :
SSTDICC010


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#92
Benzo(g,h,i)perylene
Concen: 8.847 ng
RT: 30.168 min Scan# 4621
Delta R.T. -0.006 min
Lab File: BP021483.D
Acq: 13 Aug 2024 13:59

Tgt Ion:276 Resp: 1037272
Ion Ratio Lower Upper
276 100
277 24.5 19.0 28.6
138 19.4 16.0 24.0



Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081324\
 Data File : BP021484.D
 Acq On : 13 Aug 2024 14:42
 Operator : RC/JU
 Sample : SSTDICC020
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC020

Quant Time: Aug 13 23:51:48 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 13 23:44:24 2024
 Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024
 Supervised By :mohammad ahmed 08/14/2024

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|------------------------------------|--------|------|----------|--------|-------|----------|
| Internal Standards | | | | | | |
| 1) 1,4-Dichlorobenzene-d4 | 7.805 | 152 | 325874 | 20.000 | ng | 0.00 |
| 21) Naphthalene-d8 | 10.593 | 136 | 1390587 | 20.000 | ng | 0.00 |
| 39) Acenaphthene-d10 | 14.451 | 164 | 947946 | 20.000 | ng | 0.00 |
| 64) Phenanthrene-d10 | 17.251 | 188 | 2036378 | 20.000 | ng | 0.00 |
| 76) Chrysene-d12 | 21.704 | 240 | 2051361 | 20.000 | ng | 0.00 |
| 86) Perylene-d12 | 25.110 | 264 | 2332049 | 20.000 | ng | 0.00 |
| System Monitoring Compounds | | | | | | |
| 5) 2-Fluorophenol | 5.393 | 112 | 915778 | 41.728 | ng | 0.00 |
| 7) Phenol-d6 | 6.958 | 99 | 1343954 | 41.979 | ng | 0.00 |
| 23) Nitrobenzene-d5 | 8.952 | 82 | 1073821 | 39.857 | ng | 0.00 |
| 42) 2,4,6-Tribromophenol | 15.963 | 330 | 421534 | 39.955 | ng | 0.00 |
| 45) 2-Fluorobiphenyl | 13.063 | 172 | 2531601 | 40.758 | ng | 0.00 |
| 79) Terphenyl-d14 | 19.975 | 244 | 4240552 | 40.160 | ng | 0.00 |
| Target Compounds | | | | | | |
| | | | | Qvalue | | |
| 2) 1,4-Dioxane | 3.317 | 88 | 221781 | 21.313 | ng | 98 |
| 3) Pyridine | 3.705 | 79 | 623899 | 21.477 | ng | 99 |
| 4) n-Nitrosodimethylamine | 3.617 | 42 | 185073 | 21.196 | ng | 96 |
| 6) Aniline | 7.128 | 93 | 669123 | 20.906 | ng | 99 |
| 8) 2-Chlorophenol | 7.369 | 128 | 420713 | 20.744 | ng | 98 |
| 9) Benzaldehyde | 6.946 | 77 | 459162 | 22.400 | ng | 100 |
| 10) Phenol | 6.987 | 94 | 696832 | 21.019 | ng | 99 |
| 11) bis(2-Chloroethyl)ether | 7.228 | 93 | 582174 | 20.753 | ng | 98 |
| 12) 1,3-Dichlorobenzene | 7.699 | 146 | 503772 | 20.912 | ng | 99 |
| 13) 1,4-Dichlorobenzene | 7.840 | 146 | 506071 | 20.804 | ng | 99 |
| 14) 1,2-Dichlorobenzene | 8.158 | 146 | 493188 | 21.038 | ng | 99 |
| 15) Benzyl Alcohol | 8.034 | 79 | 480369 | 20.912 | ng | 99 |
| 16) 2,2'-oxybis(1-Chloropr... | 8.340 | 45 | 535190 | 20.717 | ng | 99 |
| 17) 2-Methylphenol | 8.234 | 107 | 439840 | 20.986 | ng | 100 |
| 18) Hexachloroethane | 8.893 | 117 | 201027 | 20.608 | ng | 98 |
| 19) n-Nitroso-di-n-propyla... | 8.605 | 70 | 467087 | 21.121 | ng | 100 |
| 20) 3+4-Methylphenols | 8.558 | 107 | 584839 | 20.695 | ng | 99 |
| 22) Acetophenone | 8.616 | 105 | 893676 | 20.919 | ng | # 99 |
| 24) Nitrobenzene | 8.993 | 77 | 597794 | 20.430 | ng | 99 |
| 25) Isophorone | 9.516 | 82 | 1322653 | 20.781 | ng | 100 |
| 26) 2-Nitrophenol | 9.705 | 139 | 105219 | 20.765 | ng | 97 |
| 27) 2,4-Dimethylphenol | 9.758 | 122 | 322640 | 20.541 | ng | 98 |
| 28) bis(2-Chloroethoxy)met... | 9.999 | 93 | 795422 | 20.473 | ng | 99 |
| 29) 2,4-Dichlorophenol | 10.234 | 162 | 375130 | 20.213 | ng | 98 |
| 30) 1,2,4-Trichlorobenzene | 10.458 | 180 | 482172 | 20.265 | ng | 98 |
| 31) Naphthalene | 10.646 | 128 | 1492878 | 20.550 | ng | 100 |
| 32) Benzoic acid | 9.840 | 122 | 144329 | 19.969 | ng | 95 |
| 33) 4-Chloroaniline | 10.734 | 127 | 564170 | 20.675 | ng | 99 |
| 34) Hexachlorobutadiene | 10.946 | 225 | 302459 | 20.189 | ng | 98 |
| 35) Caprolactam | 11.481 | 113 | 163665 | 21.189 | ng | 98 |
| 36) 4-Chloro-3-methylphenol | 11.857 | 107 | 531060 | 20.628 | ng | 99 |
| 37) 2-Methylnaphthalene | 12.257 | 142 | 1015274 | 20.478 | ng | 100 |
| 38) 1-Methylnaphthalene | 12.475 | 142 | 1008222 | 20.619 | ng | 100 |
| 40) 1,2,4,5-Tetrachloroben... | 12.628 | 216 | 561317 | 20.209 | ng | 99 |
| 41) Hexachlorocyclopentadiene | 12.616 | 237 | 164693 | 18.709 | ng | 99 |
| 43) 2,4,6-Trichlorophenol | 12.863 | 196 | 301820 | 19.562 | ng | 99 |

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081324\
 Data File : BP021484.D
 Acq On : 13 Aug 2024 14:42
 Operator : RC/JU
 Sample : SSTDICC020
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC020

Quant Time: Aug 13 23:51:48 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 13 23:44:24 2024
 Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024
 Supervised By :mohammad ahmed 08/14/2024

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|-------------------------------|--------|------|----------|--------|-------|----------|
| 44) 2,4,5-Trichlorophenol | 12.934 | 196 | 341066 | 19.897 | ng | 99 |
| 46) 1,1'-Biphenyl | 13.281 | 154 | 1359024 | 20.263 | ng | 99 |
| 47) 2-Chloronaphthalene | 13.322 | 162 | 1058999 | 20.282 | ng | 99 |
| 48) 2-Nitroaniline | 13.504 | 65 | 218570 | 18.481 | ng | 97 |
| 49) Acenaphthylene | 14.169 | 152 | 1571109 | 20.468 | ng | 100 |
| 50) Dimethylphthalate | 13.904 | 163 | 1297733 | 20.442 | ng | 99 |
| 51) 2,6-Dinitrotoluene | 14.016 | 165 | 215611 | 19.423 | ng | 93 |
| 52) Acenaphthene | 14.516 | 154 | 1018510 | 20.223 | ng | 100 |
| 53) 3-Nitroaniline | 14.346 | 138 | 209499 | 19.232 | ng | # 91 |
| 54) 2,4-Dinitrophenol | 14.540 | 184 | 49171 | 19.714 | ng | # 74 |
| 55) Dibenzofuran | 14.851 | 168 | 1589415 | 20.515 | ng | 99 |
| 56) 4-Nitrophenol | 14.634 | 139 | 167932 | 20.016 | ng | 97 |
| 57) 2,4-Dinitrotoluene | 14.804 | 165 | 248400 | 18.676 | ng | 92 |
| 58) Fluorene | 15.516 | 166 | 1334998 | 20.975 | ng | 100 |
| 59) 2,3,4,6-Tetrachlorophenol | 15.081 | 232 | 286326 | 19.679 | ng | 98 |
| 60) Diethylphthalate | 15.293 | 149 | 1353063 | 20.773 | ng | 100 |
| 61) 4-Chlorophenyl-phenyle... | 15.522 | 204 | 680988 | 20.699 | ng | 97 |
| 62) 4-Nitroaniline | 15.522 | 138 | 229769 | 19.603 | ng | 92 |
| 63) Azobenzene | 15.816 | 77 | 1685601 | 21.052 | ng | 99 |
| 65) 4,6-Dinitro-2-methylph... | 15.581 | 198 | 79397 | 19.364 | ng | 96 |
| 66) n-Nitrosodiphenylamine | 15.734 | 169 | 1118083 | 20.488 | ng | 100 |
| 67) 4-Bromophenyl-phenylether | 16.434 | 248 | 432165 | 20.052 | ng | 99 |
| 68) Hexachlorobenzene | 16.545 | 284 | 509441 | 20.002 | ng | 98 |
| 69) Atrazine | 16.710 | 200 | 398442 | 21.266 | ng | 99 |
| 70) Pentachlorophenol | 16.892 | 266 | 255526 | 18.381 | ng | 99 |
| 71) Phenanthrene | 17.292 | 178 | 2033070 | 20.760 | ng | 99 |
| 72) Anthracene | 17.398 | 178 | 1976030 | 20.574 | ng | 99 |
| 73) Carbazole | 17.663 | 167 | 1916837 | 20.709 | ng | 99 |
| 74) Di-n-butylphthalate | 18.269 | 149 | 2340289 | 20.587 | ng | 99 |
| 75) Fluoranthene | 19.375 | 202 | 2511684 | 20.902 | ng | 100 |
| 77) Benzidine | 19.575 | 184 | 924193 | 20.258 | ng | 99 |
| 78) Pyrene | 19.751 | 202 | 2656705 | 19.814 | ng | 100 |
| 80) Butylbenzylphthalate | 20.710 | 149 | 843482 | 18.934 | ng | 98 |
| 81) Benzo(a)anthracene | 21.686 | 228 | 2648478 | 20.566 | ng | 100 |
| 82) 3,3'-Dichlorobenzidine | 21.604 | 252 | 847490 | 20.526 | ng | 99 |
| 83) Chrysene | 21.751 | 228 | 2472528 | 20.336 | ng | 100 |
| 84) Bis(2-ethylhexyl)phtha... | 21.651 | 149 | 1520788 | 19.981 | ng | 99 |
| 85) Di-n-octyl phthalate | 22.963 | 149 | 2286827 | 20.217 | ng | 99 |
| 87) Indeno(1,2,3-cd)pyrene | 28.980 | 276 | 3071376m | 20.320 | ng | |
| 88) Benzo(b)fluoranthene | 24.033 | 252 | 2700298 | 20.161 | ng | 98 |
| 89) Benzo(k)fluoranthene | 24.104 | 252 | 2620830 | 20.059 | ng | 99 |
| 90) Benzo(a)pyrene | 24.951 | 252 | 2339364 | 20.127 | ng | 98 |
| 91) Dibenzo(a,h)anthracene | 29.080 | 278 | 2579763 | 20.539 | ng | 99 |
| 92) Benzo(g,h,i)perylene | 30.168 | 276 | 2555517 | 20.300 | ng | 100 |

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081324\
 Data File : BP021484.D
 Acq On : 13 Aug 2024 14:42
 Operator : RC/JU
 Sample : SSTDICC020
 Misc :
 ALS Vial : 5 Sample Multiplier: 1


Quant Time: Aug 13 23:51:48 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 13 23:44:24 2024
 Response via : Initial Calibration

Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC020

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024
 Supervised By :mohammad ahmed 08/14/2024






#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.805 min Scan# 8
Delta R.T. -0.006 min
Lab File: BP021484.D
Acq: 13 Aug 2024 14:42

Instrument : BNA_P
ClientSampleId : SSTDICC020


Manual Integrations
APPROVED







Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#2
1,4-Dioxane
Concen: 21.313 ng
RT: 3.317 min Scan# 56
Delta R.T. 0.000 min
Lab File: BP021484.D
Acq: 13 Aug 2024 14:42

Tgt Ion: 88 Resp: 221781
Ion Ratio Lower Upper
88 100
58 62.4 48.7 73.1
43 19.4 15.4 23.2






#3

Pyridine
Concen: 21.477 ng
RT: 3.705 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP021484.D
Acq: 13 Aug 2024 14:42

Instrument : BNA_P
ClientSampleId : SSTDICC020

Manual Integrations APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024




#4

n-Nitrosodimethylamine
Concen: 21.196 ng
RT: 3.617 min Scan# 107
Delta R.T. 0.000 min
Lab File: BP021484.D
Acq: 13 Aug 2024 14:42

Tgt Ion: 42 Resp: 185073
Ion Ratio Lower Upper
42 100
74 186.5 154.3 231.5
44 9.2 6.2 9.4







#5
2-Fluorophenol
Concen: 41.728 ng
RT: 5.393 min Scan# 409
Delta R.T. 0.000 min
Lab File: BP021484.D
Acq: 13 Aug 2024 14:42

Instrument : BNA_P
ClientSampleId : SSTDICC020


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024




#6
Aniline
Concen: 20.906 ng
RT: 7.128 min Scan# 704
Delta R.T. -0.006 min
Lab File: BP021484.D
Acq: 13 Aug 2024 14:42


Tgt Ion: 93 Resp: 669123
Ion Ratio Lower Upper
93 100
66 37.6 30.2 45.2
65 19.0 14.7 22.1




Abundance Scan 676 (6.964 min): BP021485.D\data.ms (-66)



Abundance Scan 675 (6.958 min): BP021484.D\data.ms



Abundance Scan 675 (6.958 min): BP021484.D\data.ms (-62)



#7

Phenol-d6

Concen: 41.979 ng

RT: 6.958 min Scan# 6

Delta R.T. -0.006 min

Lab File: BP021484.D

Acq: 13 Aug 2024 14:42

Instrument :

BNA_P


ClientSampleId :

SSTDICC020


**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/14/2024


Supervised By :mohammad ahmed 08/14/2024




Abundance Scan 746 (7.375 min): BP021485.D\data.ms (-73)



Abundance Scan 745 (7.369 min): BP021484.D\data.ms



Abundance Scan 745 (7.369 min): BP021484.D\data.ms (-69)



#8

2-Chlorophenol

Concen: 20.744 ng

RT: 7.369 min Scan# 745

Delta R.T. -0.006 min

Lab File: BP021484.D

Acq: 13 Aug 2024 14:42


Tgt Ion:128 Resp: 420713

Ion Ratio Lower Upper




128 100

130 30.5 11.6 51.6

64 56.6 35.3 75.3



Abundance Scan 673 (6.946 min): BP021485.D\data.ms (-66)



#9

Benzaldehyde

Concen: 22.400 ng

RT: 6.946 min Scan# 6

Delta R.T. -0.000 min

Lab File: BP021484.D

Acq: 13 Aug 2024 14:42

Instrument :

BNA_P





ClientSampleId :

SSTDICC020

Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#10

Phenol

Concen: 21.019 ng


RT: 6.987 min Scan# 680




Delta R.T. 0.000 min

Lab File: BP021484.D

Acq: 13 Aug 2024 14:42

| Tgt | Ion | Resp: | | |
|-----|------|--------|-------|-------|
| | | | Lower | Upper |
| | 94 | 696832 | | |
| | 100 | | | |
| 65 | 25.7 | | 6.5 | 46.5 |
| 66 | 34.1 | | 14.7 | 54.7 |









#11
bis(2-Chloroethyl)ether
Concen: 20.753 ng
RT: 7.228 min Scan# 7
Delta R.T. 0.000 min
Lab File: BP021484.D
Acq: 13 Aug 2024 14:42

Instrument : BNA_P
ClientSampleId : SSTDICC020


Manual Integrations APPROVED







Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#12
1,3-Dichlorobenzene
Concen: 20.912 ng
RT: 7.699 min Scan# 801
Delta R.T. 0.000 min
Lab File: BP021484.D
Acq: 13 Aug 2024 14:42

Tgt Ion:146 Resp: 503772
Ion Ratio Lower Upper
146 100
148 63.8 51.8 77.6
75 40.2 31.3 46.9





#13

1,4-Dichlorobenzene

Concen: 20.804 ng

RT: 7.840 min Scan# 8

Delta R.T. 0.000 min

Lab File: BP021484.D

Acq: 13 Aug 2024 14:42

Instrument :

BNA_P


ClientSampleId :

SSTDICC020

Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#14

1,2-Dichlorobenzene

Concen: 21.038 ng

RT: 8.158 min Scan# 879

Delta R.T. 0.000 min

Lab File: BP021484.D

Acq: 13 Aug 2024 14:42


Tgt Ion:146 Resp: 493188




Ion Ratio Lower Upper

146 100

148 63.9 51.5 77.3

111 50.1 41.1 61.7





#15

Benzyl Alcohol

Concen: 20.912 ng

RT: 8.034 min Scan# 8

Delta R.T. 0.000 min

Lab File: BP021484.D

Acq: 13 Aug 2024 14:42

Instrument :

BNA_P





ClientSampleId :

SSTDICC020

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#16

2,2'-oxybis(1-Chloropropane)

Concen: 20.717 ng

RT: 8.340 min Scan# 910

Delta R.T. 0.006 min

Lab File: BP021484.D

Acq: 13 Aug 2024 14:42


Tgt Ion: 45 Resp: 535190




Ion Ratio Lower Upper

45 100

77 21.5 1.9 41.9

79 16.7 0.0 36.6





#17

2-Methylphenol

Concen: 20.986 ng

RT: 8.234 min Scan# 8

Delta R.T. 0.000 min

Lab File: BP021484.D

Acq: 13 Aug 2024 14:42




Instrument : BNA_P

ClientSampleId : SSTDICC020

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#18

Hexachloroethane

Concen: 20.608 ng

RT: 8.893 min Scan# 1004

Delta R.T. 0.000 min

Lab File: BP021484.D

Acq: 13 Aug 2024 14:42


Tgt Ion:117 Resp: 201027


Ion Ratio Lower Upper

117 100

119 93.0 73.9 110.9

201 90.6 74.9 112.3







#19
n-Nitroso-di-n-propylamine
Concen: 21.121 ng
RT: 8.605 min Scan# 9
Delta R.T. 0.000 min
Lab File: BP021484.D
Acq: 13 Aug 2024 14:42

Instrument : BNA_P
ClientSampleId : SSTDICC020


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#20
3+4-Methylphenols
Concen: 20.695 ng
RT: 8.558 min Scan# 947
Delta R.T. 0.000 min
Lab File: BP021484.D
Acq: 13 Aug 2024 14:42

Tgt Ion:107 Resp: 584839
Ion Ratio Lower Upper
107 100
108 86.3 67.6 107.6
77 34.7 15.0 55.0
79 29.2 8.9 48.9





#21

Naphthalene-d8

Concen: 20.000 ng

RT: 10.593 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP021484.D

Acq: 13 Aug 2024 14:42

Instrument :

BNA_P



ClientSampleId :

SSTDICC020

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#22

Acetophenone

Concen: 20.919 ng


RT: 8.616 min Scan# 957


Delta R.T. 0.000 min

Lab File: BP021484.D

Acq: 13 Aug 2024 14:42

| Tgt | Ion:105 | Resp: | 893676 |
|-----|---------|-------|--------|
| Ion | Ratio | Lower | Upper |
| 105 | 100 | | |
| 71 | 4.6 | 4.6 | 7.0# |
| 51 | 20.1 | 16.3 | 24.5 |
| 120 | 19.6 | 16.2 | 24.2 |







#23
 Nitrobenzene-d5
 Concen: 39.857 ng
 RT: 8.952 min Scan# 1
 Delta R.T. 0.000 min
 Lab File: BP021484.D
 Acq: 13 Aug 2024 14:42

Instrument : BNA_P
 ClientSampleId : SSTDICC020


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
 Supervised By :mohammad ahmed 08/14/2024



#24
 Nitrobenzene
 Concen: 20.430 ng
 RT: 8.993 min Scan# 1021
 Delta R.T. 0.000 min
 Lab File: BP021484.D
 Acq: 13 Aug 2024 14:42

Tgt Ion: 77 Resp: 597794
 Ion Ratio Lower Upper
 77 100
 123 36.4 29.5 44.3
 65 12.6 9.9 14.9







#25
Isophorone
Concen: 20.781 ng
RT: 9.516 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP021484.D
Acq: 13 Aug 2024 14:42

Instrument : BNA_P
ClientSampleId : SSTDICC020


Manual Integrations
APPROVED




Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#26
2-Nitrophenol
Concen: 20.765 ng
RT: 9.705 min Scan# 1142
Delta R.T. 0.006 min
Lab File: BP021484.D
Acq: 13 Aug 2024 14:42

Tgt Ion:139 Resp: 105219
Ion Ratio Lower Upper
139 100
109 35.2 28.9 43.3
65 64.6 49.0 73.6





#27

2,4-Dimethylphenol

Concen: 20.541 ng

RT: 9.758 min Scan# 1151

Delta R.T. 0.000 min

Lab File: BP021484.D

Acq: 13 Aug 2024 14:42

Instrument :

BNA_P





ClientSampleId :

SSTDICC020

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#28

bis(2-Chloroethoxy)methane

Concen: 20.473 ng

RT: 9.999 min Scan# 1192

Delta R.T. -0.006 min

Lab File: BP021484.D

Acq: 13 Aug 2024 14:42


Tgt Ion: 93 Resp: 795422


Ion Ratio Lower Upper

93 100

95 32.2 26.2 39.4

123 9.2 7.6 11.4







#29
2,4-Dichlorophenol
Concen: 20.213 ng
RT: 10.234 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP021484.D
Acq: 13 Aug 2024 14:42

Instrument : BNA_P
ClientSampleId : SSTDICC020


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#30
1,2,4-Trichlorobenzene
Concen: 20.265 ng
RT: 10.458 min Scan# 1270
Delta R.T. 0.000 min
Lab File: BP021484.D
Acq: 13 Aug 2024 14:42

Tgt Ion:180 Resp: 482172
Ion Ratio Lower Upper
180 100
182 96.0 78.3 117.5
145 28.6 24.0 36.0





#31

Naphthalene

Concen: 20.550 ng

RT: 10.646 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP021484.D

Acq: 13 Aug 2024 14:42

Instrument :

BNA_P

ClientSampleId :

SSTDICC020

Tgt Ion:128 Resp: 149287

Ion Ratio Lower Upper

128 100

129 10.5 8.6 12.8



127 13.3 10.6 15.8

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#32

Benzoic acid

Concen: 19.969 ng

RT: 9.840 min Scan# 1165

Delta R.T. -0.035 min

Lab File: BP021484.D

Acq: 13 Aug 2024 14:42


Tgt Ion:122 Resp: 144329


Ion Ratio Lower Upper

122 100

105 151.1 124.6 164.6

77 121.0 96.2 136.2







#33
4-Chloroaniline
Concen: 20.675 ng
RT: 10.734 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP021484.D
Acq: 13 Aug 2024 14:42

Instrument : BNA_P
ClientSampleId : SSTDICC020


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#34
Hexachlorobutadiene
Concen: 20.189 ng
RT: 10.946 min Scan# 1353
Delta R.T. 0.006 min
Lab File: BP021484.D
Acq: 13 Aug 2024 14:42

Tgt Ion:225 Resp: 302459
Ion Ratio Lower Upper
225 100
223 64.1 49.6 74.4
227 63.4 50.1 75.1







#35
Caprolactam
Concen: 21.189 ng
RT: 11.481 min Scan# 1447
Delta R.T. -0.018 min
Lab File: BP021484.D
Acq: 13 Aug 2024 14:42

Instrument : BNA_P
ClientSampleId : SSTDICC020


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#36
4-Chloro-3-methylphenol
Concen: 20.628 ng
RT: 11.857 min Scan# 1508
Delta R.T. -0.006 min
Lab File: BP021484.D
Acq: 13 Aug 2024 14:42

Tgt Ion:107 Resp: 531060
Ion Ratio Lower Upper
107 100
144 20.6 16.6 24.8
142 63.2 51.0 76.4



#37

2-Methylnaphthalene

Concen: 20.478 ng

RT: 12.257 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP021484.D


Acq: 13 Aug 2024 14:42

Instrument :


BNA_P

ClientSampleId :


SSTDICC020



Abundance Scan 1576 (12.257 min): BP021484.D\data.ms (-)



Abundance Scan 1576 (12.257 min): BP021484.D\data.ms (-)



Tgt Ion:142 Resp: 101527

Ion Ratio Lower Upper

142 100

141 89.1 71.7 107.5



115 39.3 31.4 47.2

Manual Integrations


APPROVED

Reviewed By :Yogesh Patel 08/14/2024


Supervised By :mohammad ahmed 08/14/2024



Abundance Scan 1613 (12.475 min): BP021484.D\data.ms (-)



Abundance Scan 1613 (12.475 min): BP021484.D\data.ms (-)



#38

1-Methylnaphthalene

Concen: 20.619 ng

RT: 12.475 min Scan# 1613

Delta R.T. -0.006 min

Lab File: BP021484.D

Acq: 13 Aug 2024 14:42


Tgt Ion:142 Resp: 1008222


Ion Ratio Lower Upper

142 100

141 91.5 73.5 110.3

116 4.2 3.4 5.2





#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 14.451 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP021484.D

Acq: 13 Aug 2024 14:42

Instrument :

BNA_P



ClientSampleId :

SSTDICC020

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#40

1,2,4,5-Tetrachlorobenzene


Concen: 20.209 ng


RT: 12.628 min Scan# 1639

Delta R.T. 0.000 min

Lab File: BP021484.D

Acq: 13 Aug 2024 14:42







#41
Hexachlorocyclopentadiene
Concen: 18.709 ng
RT: 12.616 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP021484.D
Acq: 13 Aug 2024 14:42

Instrument : BNA_P
ClientSampleId : SSTDICC020


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#42
2,4,6-Tribromophenol
Concen: 39.955 ng
RT: 15.963 min Scan# 2206
Delta R.T. -0.006 min
Lab File: BP021484.D
Acq: 13 Aug 2024 14:42

Tgt Ion:330 Resp: 421534
Ion Ratio Lower Upper
330 100
332 96.4 77.4 116.0
141 31.2 24.5 36.7



#43

2,4,6-Trichlorophenol

Concen: 19.562 ng

RT: 12.863 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP021484.D


Acq: 13 Aug 2024 14:42

Instrument :


BNA_P

ClientSampleId :


SSTDICC020



Abundance Scan 1679 (12.863 min): BP021484.D\data.ms



Abundance Scan 1679 (12.863 min): BP021484.D\data.ms (-)



Tgt Ion:196 Resp: 301820

Ion Ratio Lower Upper

196 100

198 96.8 76.6 114.8



200 30.8 24.5 36.7

Manual Integrations


APPROVED

Reviewed By :Yogesh Patel 08/14/2024


Supervised By :mohammad ahmed 08/14/2024



Abundance Scan 1691 (12.934 min): BP021484.D\data.ms



Abundance Scan 1691 (12.934 min): BP021484.D\data.ms (-)



#44

2,4,5-Trichlorophenol

Concen: 19.897 ng

RT: 12.934 min Scan# 1691

Delta R.T. 0.000 min

Lab File: BP021484.D

Acq: 13 Aug 2024 14:42

Tgt Ion:196 Resp: 341066


Ion Ratio Lower Upper


196 100

198 96.7 78.9 118.3


97 60.8 49.4 74.0

132 26.4 21.1 31.7






Abundance Scan 1713 (13.063 min): BP021484.D\data.ms (-)



Abundance Scan 1713 (13.063 min): BP021484.D\data.ms (-)



#45

2-Fluorobiphenyl

Concen: 40.758 ng

RT: 13.063 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP021484.D

Acq: 13 Aug 2024 14:42

Instrument :

BNA_P


ClientSampleId :

SSTDICC020


**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/14/2024


Supervised By :mohammad ahmed 08/14/2024



Abundance Scan 1750 (13.281 min): BP021484.D\data.ms (-)



Abundance Scan 1750 (13.281 min): BP021484.D\data.ms (-)



#46

1,1'-Biphenyl

Concen: 20.263 ng

RT: 13.281 min Scan# 1750

Delta R.T. -0.000 min

Lab File: BP021484.D

Acq: 13 Aug 2024 14:42


Tgt Ion:154 Resp: 1359024


Ion Ratio Lower Upper

154 100

153 40.1 20.7 60.7

76 17.4 0.0 37.7







#47
2-Chloronaphthalene
Concen: 20.282 ng
RT: 13.322 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP021484.D
Acq: 13 Aug 2024 14:42

Instrument : BNA_P
ClientSampleId : SSTDICC020


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#48
2-Nitroaniline
Concen: 18.481 ng
RT: 13.504 min Scan# 1788
Delta R.T. 0.000 min
Lab File: BP021484.D
Acq: 13 Aug 2024 14:42

Tgt Ion: 65 Resp: 218570
Ion Ratio Lower Upper
65 100
92 74.3 58.1 87.1
138 85.1 65.5 98.3



#49

Acenaphthylene

Concen: 20.468 ng

RT: 14.169 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP021484.D




Acq: 13 Aug 2024 14:42

Instrument :

BNA_P

ClientSampleId :

SSTDICC020



Tgt Ion:152 Resp: 1571109

Ion Ratio Lower Upper

152 100





151 19.8 15.9 23.9

153 13.0 10.6 16.0

Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#50

Dimethylphthalate

Concen: 20.442 ng

RT: 13.904 min Scan# 1856

Delta R.T. -0.006 min

Lab File: BP021484.D

Acq: 13 Aug 2024 14:42


Tgt Ion:163 Resp: 1297733


Ion Ratio Lower Upper

163 100

194 4.3 3.3 4.9

164 9.7 8.3 12.5







#51
2,6-Dinitrotoluene
Concen: 19.423 ng
RT: 14.016 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP021484.D
Acq: 13 Aug 2024 14:42

Instrument : BNA_P
ClientSampleId : SSTDICC020


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#52
Acenaphthene
Concen: 20.223 ng
RT: 14.516 min Scan# 1960
Delta R.T. -0.006 min
Lab File: BP021484.D
Acq: 13 Aug 2024 14:42

Tgt Ion:154 Resp: 1018510
Ion Ratio Lower Upper
154 100
153 113.8 91.1 136.7
152 54.0 43.0 64.4







#53
3-Nitroaniline
Concen: 19.232 ng
RT: 14.346 min Scan# 1
Delta R.T. -0.005 min
Lab File: BP021484.D
Acq: 13 Aug 2024 14:42

Instrument : BNA_P
ClientSampleId : SSTDICC020


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#54
2,4-Dinitrophenol
Concen: 19.714 ng
RT: 14.540 min Scan# 1964
Delta R.T. -0.011 min
Lab File: BP021484.D
Acq: 13 Aug 2024 14:42

Tgt Ion:184 Resp: 49171
Ion Ratio Lower Upper
184 100
63 85.8 64.9 97.3
154 105.7 52.1 78.1#







#55
Dibenzofuran
Concen: 20.515 ng
RT: 14.851 min Scan# 2
Instrument : BNA_P
Delta R.T. -0.006 min
Lab File: BP021484.D
Acq: 13 Aug 2024 14:42
ClientSampleId : SSTDICC020


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#56
4-Nitrophenol
Concen: 20.016 ng
RT: 14.634 min Scan# 1980
Delta R.T. -0.017 min
Lab File: BP021484.D
Acq: 13 Aug 2024 14:42

Tgt Ion:139 Resp: 167932
Ion Ratio Lower Upper
139 100
109 93.8 70.2 110.2
65 124.4 101.8 141.8







#57
2,4-Dinitrotoluene
Concen: 18.676 ng
RT: 14.804 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP021484.D
Acq: 13 Aug 2024 14:42

Instrument : BNA_P
ClientSampleId : SSTDICC020


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#58
Fluorene
Concen: 20.975 ng
RT: 15.516 min Scan# 2130
Delta R.T. 0.000 min
Lab File: BP021484.D
Acq: 13 Aug 2024 14:42

Tgt Ion:166 Resp: 1334998
Ion Ratio Lower Upper
166 100
165 97.3 78.1 117.1
167 13.3 11.0 16.4



#59

2,3,4,6-Tetrachlorophenol

Concen: 19.679 ng

RT: 15.081 min Scan# 2

Delta R.T. -0.006 min

Lab File: BP021484.D


Acq: 13 Aug 2024 14:42

Instrument :



BNA_P

ClientSampleId :

SSTDICC020



Abundance Scan 2056 (15.081 min): BP021484.D\data.ms



Tgt Ion:232 Resp: 286320



Ion Ratio Lower Upper

| | |
|-----|------|
| 232 | 100 |
| 131 | 42.6 |
| 130 | 2.3 |
| 166 | 29.6 |
| | 23.0 |
| | 34.4 |




Manual Integrations**APPROVED**


Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024




#60
Diethylphthalate
Concen: 20.773 ng
RT: 15.293 min Scan# 2092
Delta R.T. -0.012 min
Lab File: BP021484.D
Acq: 13 Aug 2024 14:42


Tgt Ion:149 Resp: 1353063
Ion Ratio Lower Upper
149 100
177 22.2 17.8 26.8
150 12.4 10.1 15.1



Abundance Scan 2131 (15.522 min): BP021484.D\data.ms



Abundance Scan 2131 (15.522 min): BP021484.D\data.ms (-)



#61

4-Chlorophenyl-phenylether

Concen: 20.699 ng

RT: 15.522 min Scan# 2131

Delta R.T. 0.000 min

Lab File: BP021484.D

Acq: 13 Aug 2024 14:42


Instrument : BNA_P

ClientSampleId : SSTDICC020


**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/14/2024


Supervised By :mohammad ahmed 08/14/2024



Abundance Scan 2131 (15.522 min): BP021484.D\data.ms



Abundance Scan 2131 (15.522 min): BP021484.D\data.ms (-)



#62

4-Nitroaniline

Concen: 19.603 ng

RT: 15.522 min Scan# 2131

Delta R.T. -0.006 min

Lab File: BP021484.D

Acq: 13 Aug 2024 14:42


Tgt Ion:138 Resp: 229769

Ion Ratio Lower Upper

138 100

92 70.1 46.4 86.4

108 103.7 73.2 113.2



#63

Azobenzene

Concen: 21.052 ng

RT: 15.816 min Scan# 2

Delta R.T. -0.006 min

Lab File: BP021484.D




Acq: 13 Aug 2024 14:42

Instrument :

BNA_P

ClientSampleId :

SSTDICC020



Tgt Ion: 77 Resp: 168560





Ion Ratio Lower Upper

| | | | |
|-----|------|-----|------|
| 77 | 100 | | |
| 182 | 19.3 | 0.0 | 39.9 |
| 105 | 20.3 | 0.9 | 40.9 |
| 51 | 21.2 | 1.4 | 41.4 |

Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 17.251 min Scan# 2425

Delta R.T. 0.000 min


Lab File: BP021484.D


Acq: 13 Aug 2024 14:42

Tgt Ion:188 Resp: 2036378

Ion Ratio Lower Upper

| | | | |
|-----|------|------|------|
| 188 | 100 | | |
| 94 | 12.6 | 10.2 | 15.4 |
| 80 | 14.4 | 11.6 | 17.4 |







#65
4,6-Dinitro-2-methylphenol
Concen: 19.364 ng
RT: 15.581 min Scan# 2142
Delta R.T. -0.006 min
Lab File: BP021484.D
Acq: 13 Aug 2024 14:42

Instrument : BNA_P
ClientSampleId : SSTDICC020


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#66
n-Nitrosodiphenylamine
Concen: 20.488 ng
RT: 15.734 min Scan# 2167
Delta R.T. -0.006 min
Lab File: BP021484.D
Acq: 13 Aug 2024 14:42

Tgt Ion:169 Resp: 1118083
Ion Ratio Lower Upper
169 100
168 66.8 53.4 80.2
167 36.7 28.7 43.1







#67
4-Bromophenyl-phenylether
Concen: 20.052 ng
RT: 16.434 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP021484.D
Acq: 13 Aug 2024 14:42

Instrument : BNA_P
ClientSampleId : SSTDICC020


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#68
Hexachlorobenzene
Concen: 20.002 ng
RT: 16.545 min Scan# 2305
Delta R.T. -0.006 min
Lab File: BP021484.D
Acq: 13 Aug 2024 14:42

Tgt Ion:284 Resp: 509441
Ion Ratio Lower Upper
284 100
142 31.5 23.4 35.2
249 30.3 24.1 36.1







#69
Atrazine
Concen: 21.266 ng
RT: 16.710 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP021484.D
Acq: 13 Aug 2024 14:42

Instrument : BNA_P
ClientSampleId : SSTDICC020


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024




#70
Pentachlorophenol
Concen: 18.381 ng
RT: 16.892 min Scan# 2364
Delta R.T. 0.000 min
Lab File: BP021484.D
Acq: 13 Aug 2024 14:42

Tgt Ion:266 Resp: 255526
Ion Ratio Lower Upper
266 100
268 64.4 51.8 77.6
264 62.0 50.5 75.7





Abundance Scan 2432 (17.292 min): BP021484.D\data.ms (-)



#71

Phenanthrene

Concen: 20.760 ng

RT: 17.292 min Scan# 2

Delta R.T. -0.006 min

Lab File: BP021484.D

Acq: 13 Aug 2024 14:42

Instrument :

BNA_P


ClientSampleId :

SSTDICC020

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#72

Anthracene

Concen: 20.574 ng

RT: 17.398 min Scan# 2450

Delta R.T. 0.000 min

Lab File: BP021484.D

Acq: 13 Aug 2024 14:42


Tgt Ion:178 Resp: 1976030

Ion Ratio Lower Upper

178 100

176 18.0 14.9 22.3

179 15.4 12.5 18.7






Tgt Ion:178 Resp: 1976030

Ion Ratio Lower Upper

178 100

176 18.0 14.9 22.3

179 15.4 12.5 18.7



#73

Carbazole

Concen: 20.709 ng

RT: 17.663 min Scan# 2

Delta R.T. -0.006 min

Lab File: BP021484.D

Acq: 13 Aug 2024 14:42

Instrument :

BNA_P





ClientSampleId :

SSTDICC020

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#74

Di-n-butylphthalate

Concen: 20.587 ng

RT: 18.269 min Scan# 2598

Delta R.T. -0.006 min

Lab File: BP021484.D

Acq: 13 Aug 2024 14:42


Tgt Ion:149 Resp: 2340289


Ion Ratio Lower Upper

149 100

150 8.9 7.5 11.3

104 6.1 4.9 7.3







#75
Fluoranthene
Concen: 20.902 ng
RT: 19.375 min Scan# 2
Instrument : BNA_P
Delta R.T. -0.006 min
Lab File: BP021484.D
Acq: 13 Aug 2024 14:42
ClientSampleId : SSTDICC020


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#76
Chrysene-d₁₂
Concen: 20.000 ng
RT: 21.704 min Scan# 3182
Delta R.T. 0.006 min
Lab File: BP021484.D
Acq: 13 Aug 2024 14:42

Tgt Ion:240 Resp: 2051361
Ion Ratio Lower Upper
240 100
120 11.3 9.9 14.9
236 25.6 20.6 30.8







#77
Benzidine
Concen: 20.258 ng
RT: 19.575 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP021484.D
Acq: 13 Aug 2024 14:42

Instrument : BNA_P
ClientSampleId : SSTDICC020


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#78
Pyrene
Concen: 19.814 ng
RT: 19.751 min Scan# 2850
Delta R.T. 0.000 min
Lab File: BP021484.D
Acq: 13 Aug 2024 14:42

Tgt Ion:202 Resp: 2656705
Ion Ratio Lower Upper
202 100
200 21.0 16.7 25.1
203 17.5 14.1 21.1







#79
Terphenyl-d14
Concen: 40.160 ng
RT: 19.975 min Scan# 2
Delta R.T. -0.000 min
Lab File: BP021484.D
Acq: 13 Aug 2024 14:42

Instrument : BNA_P
ClientSampleId : SSTDICC020


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#80
Butylbenzylphthalate
Concen: 18.934 ng
RT: 20.710 min Scan# 3013
Delta R.T. -0.000 min
Lab File: BP021484.D
Acq: 13 Aug 2024 14:42

Tgt Ion:149 Resp: 843482
Ion Ratio Lower Upper
149 100
91 93.7 73.3 109.9
206 22.9 17.1 25.7





#81

Benzo(a)anthracene

Concen: 20.566 ng

RT: 21.686 min Scan# 3178

Delta R.T. 0.006 min

Lab File: BP021484.D

Acq: 13 Aug 2024 14:42

Instrument :

BNA_P



ClientSampleId :

SSTDICC020

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#82

3,3'-Dichlorobenzidine

Concen: 20.526 ng

RT: 21.604 min Scan# 3165

Delta R.T. -0.006 min

Lab File: BP021484.D

Acq: 13 Aug 2024 14:42


Tgt Ion:252 Resp: 847490


Ion Ratio Lower Upper

252 100

254 65.3 51.7 77.5

126 10.9 8.7 13.1







#83

Chrysene
Concen: 20.336 ng
RT: 21.751 min Scan# 3
Delta R.T. 0.000 min
Lab File: BP021484.D
Acq: 13 Aug 2024 14:42

Instrument : BNA_P
ClientSampleId : SSTDICC020

Manual Integrations APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024




#84

Bis(2-ethylhexyl)phthalate
Concen: 19.981 ng
RT: 21.651 min Scan# 3173
Delta R.T. 0.006 min
Lab File: BP021484.D
Acq: 13 Aug 2024 14:42

Tgt Ion:149 Resp: 1520788
Ion Ratio Lower Upper
149 100
167 27.6 22.3 33.5
279 4.3 3.8 5.6







#85
Di-n-octyl phthalate
Concen: 20.217 ng
RT: 22.963 min Scan# 3
Instrument : BNA_P
Delta R.T. -0.000 min
Lab File: BP021484.D
Acq: 13 Aug 2024 14:42
ClientSampleId : SSTDICC020


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#86
Perylene-d₁₂
Concen: 20.000 ng
RT: 25.110 min Scan# 3761
Delta R.T. 0.006 min
Lab File: BP021484.D
Acq: 13 Aug 2024 14:42

Tgt Ion:264 Resp: 2332049
Ion Ratio Lower Upper
264 100
260 23.9 18.6 28.0
265 21.7 17.1 25.7



#87

Indeno(1,2,3-cd)pyrene

Concen: 20.320 ng m

RT: 28.980 min Scan# 4

Delta R.T. -0.012 min

Lab File: BP021484.D

Acq: 13 Aug 2024 14:42

Instrument :

BNA_P

ClientSampleId :

SSTDICC020

Tgt Ion:276 Resp: 3071370

Ion Ratio Lower Upper

276 100

138 16.1 12.8 19.2

277 20.7 16.5 24.7

Manual Integrations

APPROVED



Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024

Abundance

Scan 4419 (28.980 min): BP021484.D\data.ms (-)


276.1



Abundance

Scan 4419 (28.980 min): BP021484.D\data.ms (-)



276.1



Abundance

Scan 3577 (24.027 min): BP021485.D\data.ms (-)



252.1



Abundance

Scan 3578 (24.033 min): BP021484.D\data.ms (-)





252.1



Abundance

Scan 3578 (24.033 min): BP021484.D\data.ms (-)

252.1



#89

Benzo(k)fluoranthene

Concen: 20.059 ng

RT: 24.104 min Scan# 3

Delta R.T. -0.006 min

Lab File: BP021484.D




Acq: 13 Aug 2024 14:42

Instrument :

BNA_P





ClientSampleId :

SSTDICC020


**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#90

Benzo(a)pyrene

Concen: 20.127 ng


RT: 24.951 min Scan# 3734




Delta R.T. 0.000 min

Lab File: BP021484.D

Acq: 13 Aug 2024 14:42

| Tgt | Ion:252 | Resp: | 2339364 |
|-----|---------|-------|---------|
| Ion | Ratio | Lower | Upper |
| 252 | 100 | | |
| 253 | 21.2 | 17.7 | 26.5 |
| 125 | 11.3 | 8.7 | 13.1 |









#91
Dibenzo(a,h)anthracene
Concen: 20.539 ng
RT: 29.080 min Scan# 4437
Delta R.T. -0.006 min
Lab File: BP021484.D
Acq: 13 Aug 2024 14:42

Instrument :
BNA_P
ClientSampleId :
SSTDICC020


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#92
Benzo(g,h,i)perylene
Concen: 20.300 ng
RT: 30.168 min Scan# 4621
Delta R.T. -0.006 min
Lab File: BP021484.D
Acq: 13 Aug 2024 14:42

Tgt Ion:276 Resp: 2555517
Ion Ratio Lower Upper
276 100
277 23.8 19.0 28.6
138 19.9 16.0 24.0



Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081324\
 Data File : BP021485.D
 Acq On : 13 Aug 2024 15:26
 Operator : RC/JU
 Sample : SSTDICCC040
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDICCC040

Quant Time: Aug 13 23:53:03 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 13 23:44:24 2024
 Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024
 Supervised By :mohammad ahmed 08/14/2024

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|------------------------------------|--------|------|----------|--------|-------|----------|
| Internal Standards | | | | | | |
| 1) 1,4-Dichlorobenzene-d4 | 7.811 | 152 | 334672 | 20.000 | ng | 0.00 |
| 21) Naphthalene-d8 | 10.599 | 136 | 1379126 | 20.000 | ng | 0.00 |
| 39) Acenaphthene-d10 | 14.457 | 164 | 890240 | 20.000 | ng | 0.00 |
| 64) Phenanthrene-d10 | 17.251 | 188 | 1905279 | 20.000 | ng | 0.00 |
| 76) Chrysene-d12 | 21.698 | 240 | 1820074 | 20.000 | ng | 0.00 |
| 86) Perylene-d12 | 25.104 | 264 | 2085332 | 20.000 | ng | 0.00 |
| System Monitoring Compounds | | | | | | |
| 5) 2-Fluorophenol | 5.393 | 112 | 1804372 | 80.056 | ng | 0.00 |
| 7) Phenol-d6 | 6.964 | 99 | 2647910 | 80.535 | ng | 0.00 |
| 23) Nitrobenzene-d5 | 8.952 | 82 | 2262476 | 84.675 | ng | 0.00 |
| 42) 2,4,6-Tribromophenol | 15.969 | 330 | 840745 | 84.855 | ng | 0.00 |
| 45) 2-Fluorobiphenyl | 13.063 | 172 | 4724987 | 81.001 | ng | 0.00 |
| 79) Terphenyl-d14 | 19.975 | 244 | 7322059 | 78.155 | ng | 0.00 |
| Target Compounds | | | | | | |
| | | | | Qvalue | | |
| 2) 1,4-Dioxane | 3.317 | 88 | 429612 | 40.199 | ng | 100 |
| 3) Pyridine | 3.705 | 79 | 1205871m | 40.419 | ng | |
| 4) n-Nitrosodimethylamine | 3.617 | 42 | 356290 | 39.732 | ng | 100 |
| 6) Aniline | 7.134 | 93 | 1324158 | 40.285 | ng | 100 |
| 8) 2-Chlorophenol | 7.375 | 128 | 836602 | 40.166 | ng | 100 |
| 9) Benzaldehyde | 6.946 | 77 | 801968 | 38.095 | ng | 100 |
| 10) Phenol | 6.987 | 94 | 1368534 | 40.195 | ng | 100 |
| 11) bis(2-Chloroethyl)ether | 7.228 | 93 | 1139337 | 39.547 | ng | 100 |
| 12) 1,3-Dichlorobenzene | 7.699 | 146 | 972133 | 39.292 | ng | 100 |
| 13) 1,4-Dichlorobenzene | 7.840 | 146 | 987045 | 39.509 | ng | 100 |
| 14) 1,2-Dichlorobenzene | 8.158 | 146 | 943886 | 39.206 | ng | 100 |
| 15) Benzyl Alcohol | 8.034 | 79 | 947378 | 40.158 | ng | 100 |
| 16) 2,2'-oxybis(1-Chloropr... | 8.334 | 45 | 1035140 | 39.016 | ng | 100 |
| 17) 2-Methylphenol | 8.234 | 107 | 863176 | 40.102 | ng | 100 |
| 18) Hexachloroethane | 8.893 | 117 | 397406 | 39.668 | ng | 100 |
| 19) n-Nitroso-di-n-propyla... | 8.605 | 70 | 897052 | 39.497 | ng | 100 |
| 20) 3+4-Methylphenols | 8.558 | 107 | 1176757 | 40.546 | ng | 100 |
| 22) Acetophenone | 8.616 | 105 | 1688225 | 39.846 | ng | 100 |
| 24) Nitrobenzene | 8.993 | 77 | 1221296 | 42.085 | ng | 100 |
| 25) Isophorone | 9.522 | 82 | 2519531 | 39.915 | ng | 100 |
| 26) 2-Nitrophenol | 9.699 | 139 | 259209 | 39.575 | ng | 100 |
| 27) 2,4-Dimethylphenol | 9.758 | 122 | 621559 | 39.900 | ng | 100 |
| 28) bis(2-Chloroethoxy)met... | 10.005 | 93 | 1515653 | 39.336 | ng | 100 |
| 29) 2,4-Dichlorophenol | 10.240 | 162 | 751870 | 40.849 | ng | 100 |
| 30) 1,2,4-Trichlorobenzene | 10.457 | 180 | 938177 | 39.757 | ng | 100 |
| 31) Naphthalene | 10.646 | 128 | 2870666 | 39.845 | ng | 100 |
| 32) Benzoic acid | 9.875 | 122 | 356582 | 38.509 | ng | 100 |
| 33) 4-Chloroaniline | 10.740 | 127 | 1090714 | 40.303 | ng | 100 |
| 34) Hexachlorobutadiene | 10.940 | 225 | 594148 | 39.989 | ng | 100 |
| 35) Caprolactam | 11.499 | 113 | 314109 | 41.005 | ng | 100 |
| 36) 4-Chloro-3-methylphenol | 11.863 | 107 | 1055054 | 41.323 | ng | 100 |
| 37) 2-Methylnaphthalene | 12.263 | 142 | 1944719 | 39.550 | ng | 100 |
| 38) 1-Methylnaphthalene | 12.481 | 142 | 1921454 | 39.621 | ng | 100 |
| 40) 1,2,4,5-Tetrachloroben... | 12.628 | 216 | 1058707 | 40.588 | ng | 100 |
| 41) Hexachlorocyclopentadiene | 12.622 | 237 | 341369 | 41.293 | ng | 100 |
| 43) 2,4,6-Trichlorophenol | 12.869 | 196 | 610003 | 42.100 | ng | 100 |

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081324\
 Data File : BP021485.D
 Acq On : 13 Aug 2024 15:26
 Operator : RC/JU
 Sample : SSTDICCC040
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDICCC040

Quant Time: Aug 13 23:53:03 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 13 23:44:24 2024
 Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024
 Supervised By :mohammad ahmed 08/14/2024

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|-------------------------------|--------|------|----------|--------|-------|----------|
| 44) 2,4,5-Trichlorophenol | 12.934 | 196 | 680956 | 42.300 | ng | 100 |
| 46) 1,1'-Biphenyl | 13.281 | 154 | 2542708 | 40.370 | ng | 100 |
| 47) 2-Chloronaphthalene | 13.316 | 162 | 1987834 | 40.538 | ng | 100 |
| 48) 2-Nitroaniline | 13.504 | 65 | 506883 | 38.074 | ng | 100 |
| 49) Acenaphthylene | 14.169 | 152 | 2928814 | 40.630 | ng | 100 |
| 50) Dimethylphthalate | 13.910 | 163 | 2408078 | 40.392 | ng | 100 |
| 51) 2,6-Dinitrotoluene | 14.022 | 165 | 454687 | 38.650 | ng | 100 |
| 52) Acenaphthene | 14.522 | 154 | 1911119 | 40.406 | ng | 100 |
| 53) 3-Nitroaniline | 14.351 | 138 | 452937 | 39.107 | ng | 100 |
| 54) 2,4-Dinitrophenol | 14.551 | 184 | 116797 | 38.943 | ng | 100 |
| 55) Dibenzofuran | 14.857 | 168 | 2936517 | 40.359 | ng | 100 |
| 56) 4-Nitrophenol | 14.651 | 139 | 358912 | 39.073 | ng | 100 |
| 57) 2,4-Dinitrotoluene | 14.810 | 165 | 570222 | 38.397 | ng | 100 |
| 58) Fluorene | 15.516 | 166 | 2410461 | 40.326 | ng | 100 |
| 59) 2,3,4,6-Tetrachlorophenol | 15.087 | 232 | 584357 | 42.766 | ng | 100 |
| 60) Diethylphthalate | 15.304 | 149 | 2486968 | 40.656 | ng | 100 |
| 61) 4-Chlorophenyl-phenyle... | 15.522 | 204 | 1236572 | 40.022 | ng | 100 |
| 62) 4-Nitroaniline | 15.528 | 138 | 485672 | 40.451 | ng | 100 |
| 63) Azobenzene | 15.822 | 77 | 3052445 | 40.594 | ng | 100 |
| 65) 4,6-Dinitro-2-methylph... | 15.587 | 198 | 199156 | 38.603 | ng | 100 |
| 66) n-Nitrosodiphenylamine | 15.740 | 169 | 2051265 | 40.175 | ng | 100 |
| 67) 4-Bromophenyl-phenylether | 16.439 | 248 | 790673 | 39.211 | ng | 100 |
| 68) Hexachlorobenzene | 16.551 | 284 | 958921 | 40.240 | ng | 100 |
| 69) Atrazine | 16.716 | 200 | 726072 | 41.419 | ng | 100 |
| 70) Pentachlorophenol | 16.892 | 266 | 520400 | 40.011 | ng | 100 |
| 71) Phenanthrene | 17.298 | 178 | 3678868 | 40.150 | ng | 100 |
| 72) Anthracene | 17.398 | 178 | 3619950 | 40.283 | ng | 100 |
| 73) Carbazole | 17.669 | 167 | 3543089 | 40.912 | ng | 100 |
| 74) Di-n-butylphthalate | 18.275 | 149 | 4378720 | 41.170 | ng | 100 |
| 75) Fluoranthene | 19.381 | 202 | 4572881 | 40.674 | ng | 100 |
| 77) Benzidine | 19.575 | 184 | 1735103 | 41.023 | ng | 100 |
| 78) Pyrene | 19.751 | 202 | 4694393 | 39.461 | ng | 100 |
| 80) Butylbenzylphthalate | 20.710 | 149 | 1712106 | 38.676 | ng | 100 |
| 81) Benzo(a)anthracene | 21.680 | 228 | 4615389 | 40.395 | ng | 100 |
| 82) 3,3'-Dichlorobenzidine | 21.610 | 252 | 1559563 | 42.572 | ng | 100 |
| 83) Chrysene | 21.751 | 228 | 4358695 | 40.405 | ng | 100 |
| 84) Bis(2-ethylhexyl)phtha... | 21.645 | 149 | 2841806 | 42.082 | ng | 100 |
| 85) Di-n-octyl phthalate | 22.963 | 149 | 4538217 | 39.238 | ng | 100 |
| 87) Indeno(1,2,3-cd)pyrene | 28.992 | 276 | 5533549m | 40.941 | ng | |
| 88) Benzo(b)fluoranthene | 24.027 | 252 | 4815875 | 40.211 | ng | 100 |
| 89) Benzo(k)fluoranthene | 24.110 | 252 | 4757000 | 40.717 | ng | 100 |
| 90) Benzo(a)pyrene | 24.951 | 252 | 4221348 | 40.616 | ng | 100 |
| 91) Dibenzo(a,h)anthracene | 29.086 | 278 | 4594554 | 40.908 | ng | 100 |
| 92) Benzo(g,h,i)perylene | 30.174 | 276 | 4595577 | 40.824 | ng | 100 |

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081324\
 Data File : BP021485.D
 Acq On : 13 Aug 2024 15:26
 Operator : RC/JU
 Sample : SSTDICCC040
 Misc :
 ALS Vial : 6 Sample Multiplier: 1


Quant Time: Aug 13 23:53:03 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 13 23:44:24 2024
 Response via : Initial Calibration

Instrument :
 BNA_P
 ClientSampleId :
 SSTDICCC040

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024
 Supervised By :mohammad ahmed 08/14/2024







#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.811 min Scan# 8
Instrument : BNA_P
Delta R.T. -0.000 min
Lab File: BP021485.D
ClientSampleId : SSTDICCC040
Acq: 13 Aug 2024 15:26


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#2
1,4-Dioxane
Concen: 40.199 ng
RT: 3.317 min Scan# 56
Delta R.T. 0.000 min
Lab File: BP021485.D
Acq: 13 Aug 2024 15:26

Tgt Ion: 88 Resp: 429612
Ion Ratio Lower Upper
88 100
58 60.9 48.7 73.1
43 19.3 15.4 23.2







#3
Pyridine
Concen: 40.419 ng m
RT: 3.705 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP021485.D
Acq: 13 Aug 2024 15:26

Instrument : BNA_P
ClientSampleId : SSTDICCC040


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#4
n-Nitrosodimethylamine
Concen: 39.732 ng
RT: 3.617 min Scan# 107
Delta R.T. 0.000 min
Lab File: BP021485.D
Acq: 13 Aug 2024 15:26

Tgt Ion: 42 Resp: 356290
Ion Ratio Lower Upper
42 100
74 192.9 154.3 231.5
44 7.8 6.2 9.4







#5
2-Fluorophenol
Concen: 80.056 ng
RT: 5.393 min Scan# 409
Delta R.T. 0.000 min
Lab File: BP021485.D
Acq: 13 Aug 2024 15:26

Instrument : BNA_P
ClientSampleId : SSTDICCC040


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024






#6
Aniline
Concen: 40.285 ng
RT: 7.134 min Scan# 705
Delta R.T. 0.000 min
Lab File: BP021485.D
Acq: 13 Aug 2024 15:26

Tgt Ion: 93 Resp: 1324158
Ion Ratio Lower Upper
93 100
66 37.7 30.2 45.2
65 18.4 14.7 22.1



Abundance Scan 676 (6.964 min): BP021485.D\data.ms (-66)



#7

Phenol-d6

Concen: 80.535 ng

RT: 6.964 min Scan# 6

Delta R.T. 0.000 min

Lab File: BP021485.D

Acq: 13 Aug 2024 15:26

Instrument :

BNA_P





ClientSampleId :

SSTDICCC040

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#8

2-Chlorophenol

Concen: 40.166 ng

RT: 7.375 min Scan# 746

Delta R.T. 0.000 min

Lab File: BP021485.D

Acq: 13 Aug 2024 15:26


Tgt Ion:128 Resp: 836602

Ion Ratio Lower Upper


128 100

130 31.6 11.6 51.6

64 55.3 35.3 75.3



Abundance Scan 673 (6.946 min): BP021485.D\data.ms (-66)



#9

Benzaldehyde

Concen: 38.095 ng

RT: 6.946 min Scan# 6

Delta R.T. -0.000 min

Lab File: BP021485.D

Acq: 13 Aug 2024 15:26

Instrument :

BNA_P

ClientSampleId :

SSTDICCC040

Tgt Ion: 77 Resp: 801963

Ion Ratio Lower Upper

77 100

105 92.1 72.1 112.1



106 87.9 67.9 107.9

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#10

Phenol

Concen: 40.195 ng

RT: 6.987 min Scan# 680

Delta R.T. 0.000 min

Lab File: BP021485.D

Acq: 13 Aug 2024 15:26



Tgt Ion: 94 Resp: 1368534




Ion Ratio Lower Upper

94 100

65 26.5 6.5 46.5

66 34.7 14.7 54.7









#11
 bis(2-Chloroethyl)ether
 Concen: 39.547 ng
 RT: 7.228 min Scan# 7
 Delta R.T. 0.000 min
 Lab File: BP021485.D
 Acq: 13 Aug 2024 15:26

Instrument : BNA_P
 ClientSampleId : SSTDICCC040


Manual Integrations
APPROVED







Reviewed By :Yogesh Patel 08/14/2024
 Supervised By :mohammad ahmed 08/14/2024



#12
 1,3-Dichlorobenzene
 Concen: 39.292 ng
 RT: 7.699 min Scan# 801
 Delta R.T. 0.000 min
 Lab File: BP021485.D
 Acq: 13 Aug 2024 15:26

Tgt Ion:146 Resp: 972133
 Ion Ratio Lower Upper
 146 100
 148 64.7 51.8 77.6
 75 39.1 31.3 46.9





#13

1,4-Dichlorobenzene

Concen: 39.509 ng

RT: 7.840 min Scan# 8

Delta R.T. 0.000 min

Lab File: BP021485.D

Acq: 13 Aug 2024 15:26

Instrument :

BNA_P


ClientSampleId :

SSTDICCC040

Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#14

1,2-Dichlorobenzene

Concen: 39.206 ng

RT: 8.158 min Scan# 879

Delta R.T. 0.000 min

Lab File: BP021485.D

Acq: 13 Aug 2024 15:26


Tgt Ion:146 Resp: 943886




Ion Ratio Lower Upper

146 100

148 64.4 51.5 77.3

111 51.4 41.1 61.7





#15

Benzyl Alcohol

Concen: 40.158 ng

RT: 8.034 min Scan# 8

Delta R.T. 0.000 min

Lab File: BP021485.D

Acq: 13 Aug 2024 15:26

Instrument :

BNA_P





ClientSampleId :

SSTDICCC040

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#16

2,2'-oxybis(1-Chloropropane)

Concen: 39.016 ng

RT: 8.334 min Scan# 909

Delta R.T. 0.000 min

Lab File: BP021485.D

Acq: 13 Aug 2024 15:26


Tgt Ion: 45 Resp: 1035140


Ion Ratio Lower Upper

45 100

77 21.9 1.9 41.9

79 16.6 0.0 36.6





#17

2-Methylphenol

Concen: 40.102 ng

RT: 8.234 min Scan# 8

Delta R.T. 0.000 min

Lab File: BP021485.D

Acq: 13 Aug 2024 15:26

Instrument :

BNA_P



ClientSampleId :

SSTDICCC040

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#18

Hexachloroethane

Concen: 39.668 ng


RT: 8.893 min Scan# 1004




Delta R.T. 0.000 min

Lab File: BP021485.D

Acq: 13 Aug 2024 15:26

| Tgt | Ion:117 | Resp: | 397406 |
|-----|---------|-------|--------|
| Ion | Ratio | Lower | Upper |
| 117 | 100 | | |
| 119 | 92.4 | 73.9 | 110.9 |
| 201 | 93.6 | 74.9 | 112.3 |





#19





n-Nitroso-di-n-propylamine
Concen: 39.497 ng
RT: 8.605 min Scan# 9

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion: 70 Resp: 89705
Ion Ratio Lower Upper
70 100
42 41.4 33.1 49.7
101 9.3 7.4 11.2
130 16.6 13.3 19.9

Manual Integrations APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024




#20

3+4-Methylphenols
Concen: 40.546 ng
RT: 8.558 min Scan# 947
Delta R.T. 0.000 min
Lab File: BP021485.D
Acq: 13 Aug 2024 15:26

Tgt Ion:107 Resp: 1176757
Ion Ratio Lower Upper
107 100
108 87.6 67.6 107.6
77 35.0 15.0 55.0
79 28.9 8.9 48.9







#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.599 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP021485.D
Acq: 13 Aug 2024 15:26

Instrument : BNA_P
ClientSampleId : SSTDICCC040


Manual Integrations APPROVED




Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#22
Acetophenone
Concen: 39.846 ng
RT: 8.616 min Scan# 957
Delta R.T. 0.000 min
Lab File: BP021485.D
Acq: 13 Aug 2024 15:26

Tgt Ion:105 Resp: 1688225
Ion Ratio Lower Upper
105 100
71 5.8 4.6 7.0
51 20.4 16.3 24.5
120 20.2 16.2 24.2





#23

Nitrobenzene-d5

Concen: 84.675 ng

RT: 8.952 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP021485.D

Acq: 13 Aug 2024 15:26

Instrument :

BNA_P

ClientSampleId :

SSTDICCC040





**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024

Abundance

8.952



#24

Nitrobenzene

Concen: 42.085 ng

RT: 8.993 min Scan# 1021

Delta R.T. 0.000 min

Lab File: BP021485.D

Acq: 13 Aug 2024 15:26

Tgt Ion: 77 Resp: 1221296

Ion Ratio Lower Upper


77 100


123 36.9 29.5 44.3

65 12.4 9.9 14.9

Abundance

8.993







#25
Isophorone
Concen: 39.915 ng
RT: 9.522 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP021485.D
Acq: 13 Aug 2024 15:26

Instrument : BNA_P
ClientSampleId : SSTDICCC040


Manual Integrations
APPROVED




Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#26
2-Nitrophenol
Concen: 39.575 ng
RT: 9.699 min Scan# 1141
Delta R.T. -0.000 min
Lab File: BP021485.D
Acq: 13 Aug 2024 15:26

Tgt Ion:139 Resp: 259209
Ion Ratio Lower Upper
139 100
109 36.1 28.9 43.3
65 61.3 49.0 73.6





#27

2,4-Dimethylphenol

Concen: 39.900 ng

RT: 9.758 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP021485.D

Acq: 13 Aug 2024 15:26

Instrument :

BNA_P





ClientSampleId :

SSTDICCC040

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#28

bis(2-Chloroethoxy)methane

Concen: 39.336 ng

RT: 10.005 min Scan# 1193

Delta R.T. 0.000 min

Lab File: BP021485.D

Acq: 13 Aug 2024 15:26


Tgt Ion: 93 Resp: 1515653

Ion Ratio Lower Upper

93 100

95 32.8 26.2 39.4

123 9.5 7.6 11.4



#29

2,4-Dichlorophenol

Concen: 40.849 ng

RT: 10.240 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP021485.D




Acq: 13 Aug 2024 15:26

Instrument :

BNA_P

ClientSampleId :

SSTDICCC040



Tgt Ion:162 Resp: 751870

Ion Ratio Lower Upper

162 100





164 65.9 45.9 85.9

98 46.3 26.3 66.3

Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#30
1,2,4-Trichlorobenzene
Concen: 39.757 ng
RT: 10.457 min Scan# 1270
Delta R.T. 0.000 min
Lab File: BP021485.D
Acq: 13 Aug 2024 15:26


Tgt Ion:180 Resp: 938177


Ion Ratio Lower Upper

180 100

182 97.9 78.3 117.5

145 30.0 24.0 36.0





#31

Naphthalene

Concen: 39.845 ng

RT: 10.646 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP021485.D

Acq: 13 Aug 2024 15:26

Instrument :

BNA_P



ClientSampleId :

SSTDICCC040

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#32

Benzoic acid

Concen: 38.509 ng

RT: 9.875 min Scan# 1171


Delta R.T. 0.000 min



Lab File: BP021485.D

Acq: 13 Aug 2024 15:26

Tgt Ion:122 Resp: 356582

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 122 | 100 | | |
| 105 | 144.6 | 124.6 | 164.6 |
| 77 | 116.2 | 96.2 | 136.2 |






#33
4-Chloroaniline
Concen: 40.303 ng
RT: 10.740 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP021485.D
Acq: 13 Aug 2024 15:26

Instrument : BNA_P
ClientSampleId : SSTDICCC040


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#34
Hexachlorobutadiene
Concen: 39.989 ng
RT: 10.940 min Scan# 1352
Delta R.T. 0.000 min
Lab File: BP021485.D
Acq: 13 Aug 2024 15:26

Tgt Ion:225 Resp: 594148
Ion Ratio Lower Upper
225 100
223 62.0 49.6 74.4
227 62.6 50.1 75.1





#35

Caprolactam

Concen: 41.005 ng

RT: 11.499 min Scan# 1447

Delta R.T. 0.000 min

Lab File: BP021485.D

Acq: 13 Aug 2024 15:26

Instrument :

BNA_P



ClientSampleId :

SSTDICCC040

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#36

4-Chloro-3-methylphenol

Concen: 41.323 ng

RT: 11.863 min Scan# 1509

Delta R.T. 0.000 min

Lab File: BP021485.D

Acq: 13 Aug 2024 15:26


Tgt Ion:107 Resp: 1055054

Ion Ratio Lower Upper

107 100

144 20.7 16.6 24.8

142 63.7 51.0 76.4



#37

2-Methylnaphthalene

Concen: 39.550 ng

RT: 12.263 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP021485.D


Acq: 13 Aug 2024 15:26

Instrument :


BNA_P

ClientSampleId :


SSTDICCC040



Abundance Scan 1577 (12.263 min): BP021485.D\data.ms (-)



Abundance Scan 1577 (12.263 min): BP021485.D\data.ms (-)



Tgt Ion:142 Resp: 1944719

Ion Ratio Lower Upper

142 100

141 89.6 71.7 107.5



115 39.3 31.4 47.2

Manual Integrations


APPROVED

Reviewed By :Yogesh Patel 08/14/2024


Supervised By :mohammad ahmed 08/14/2024



Abundance Scan 1614 (12.481 min): BP021485.D\data.ms (-)



Abundance Scan 1614 (12.481 min): BP021485.D\data.ms (-)



#38

1-Methylnaphthalene

Concen: 39.621 ng

RT: 12.481 min Scan# 1614

Delta R.T. 0.000 min

Lab File: BP021485.D

Acq: 13 Aug 2024 15:26


Tgt Ion:142 Resp: 1921454

Ion Ratio Lower Upper

142 100

141 91.9 73.5 110.3

116 4.3 3.4 5.2



#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 14.457 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP021485.D


Acq: 13 Aug 2024 15:26

Instrument :


BNA_P

ClientSampleId :


SSTDICCC040



Abundance Scan 1950 (14.457 min): BP021485.D\data.ms (-)



Abundance Scan 1950 (14.457 min): BP021485.D\data.ms (-)



Tgt Ion:164 Resp: 890240

Ion Ratio Lower Upper

164 100

162 101.9 81.5 122.3



160 44.0 35.2 52.8

Manual Integrations


APPROVED

Reviewed By :Yogesh Patel 08/14/2024


Supervised By :mohammad ahmed 08/14/2024



Abundance Scan 1639 (12.628 min): BP021485.D\data.ms (-)



Abundance Scan 1639 (12.628 min): BP021485.D\data.ms (-)



#40

1,2,4,5-Tetrachlorobenzene

Concen: 40.588 ng

RT: 12.628 min Scan# 1639

Delta R.T. 0.000 min

Lab File: BP021485.D

Acq: 13 Aug 2024 15:26

Tgt Ion:216 Resp: 1058707


Ion Ratio Lower Upper


216 100

214 78.1 62.5 93.7

179 20.5 16.4 24.6

108 20.9 16.7 25.1







#41
Hexachlorocyclopentadiene
Concen: 41.293 ng
RT: 12.622 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP021485.D
Acq: 13 Aug 2024 15:26

Instrument : BNA_P
ClientSampleId : SSTDICCC040


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#42
2,4,6-Tribromophenol
Concen: 84.855 ng
RT: 15.969 min Scan# 2207
Delta R.T. 0.000 min
Lab File: BP021485.D
Acq: 13 Aug 2024 15:26

Tgt Ion:330 Resp: 840745
Ion Ratio Lower Upper
330 100
332 96.7 77.4 116.0
141 30.6 24.5 36.7



#43

2,4,6-Trichlorophenol

Concen: 42.100 ng

RT: 12.869 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP021485.D


Acq: 13 Aug 2024 15:26

Instrument :


BNA_P

ClientSampleId :


SSTDICCC040



Abundance Scan 1680 (12.869 min): BP021485.D\data.ms (-)



Abundance Scan 1680 (12.869 min): BP021485.D\data.ms (-)



Tgt Ion:196 Resp: 61000

Ion Ratio Lower Upper

196 100

198 95.7 76.6 114.8



200 30.6 24.5 36.7

Manual Integrations


APPROVED

Reviewed By :Yogesh Patel 08/14/2024


Supervised By :mohammad ahmed 08/14/2024



Abundance Scan 1691 (12.934 min): BP021485.D\data.ms (-)



Abundance Scan 1691 (12.934 min): BP021485.D\data.ms (-)



#44

2,4,5-Trichlorophenol

Concen: 42.300 ng

RT: 12.934 min Scan# 1691

Delta R.T. 0.000 min

Lab File: BP021485.D

Acq: 13 Aug 2024 15:26

Tgt Ion:196 Resp: 680956


Ion Ratio Lower Upper




196 100

198 98.6 78.9 118.3

97 61.7 49.4 74.0

132 26.4 21.1 31.7





#45

2-Fluorobiphenyl

Concen: 81.001 ng

RT: 13.063 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP021485.D

Acq: 13 Aug 2024 15:26

Instrument :

BNA_P




ClientSampleId :

SSTDICCC040

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#46

1,1'-Biphenyl

Concen: 40.370 ng

RT: 13.281 min Scan# 1750

Delta R.T. -0.000 min

Lab File: BP021485.D

Acq: 13 Aug 2024 15:26


Tgt Ion:154 Resp: 2542708


Ion Ratio Lower Upper

154 100

153 40.7 20.7 60.7

76 17.7 0.0 37.7





#47

2-Chloronaphthalene

Concen: 40.538 ng

RT: 13.316 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP021485.D

Acq: 13 Aug 2024 15:26

Instrument :

BNA_P



ClientSampleId :

SSTDICCC040

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#48

2-Nitroaniline

Concen: 38.074 ng

RT: 13.504 min Scan# 1788

Delta R.T. 0.000 min

Lab File: BP021485.D

Acq: 13 Aug 2024 15:26


Tgt Ion: 65 Resp: 506883


Ion Ratio Lower Upper

65 100

92 72.6 58.1 87.1

138 81.9 65.5 98.3





#49

Acenaphthylene

Concen: 40.630 ng

RT: 14.169 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP021485.D

Acq: 13 Aug 2024 15:26

Instrument :

BNA_P



ClientSampleId :

SSTDICCC040

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#50

Dimethylphthalate

Concen: 40.392 ng

RT: 13.910 min Scan# 1857

Delta R.T. 0.000 min

Lab File: BP021485.D

Acq: 13 Aug 2024 15:26


Tgt Ion:163 Resp: 2408078


Ion Ratio Lower Upper

163 100

194 4.1 3.3 4.9

164 10.4 8.3 12.5







#51
2,6-Dinitrotoluene
Concen: 38.650 ng
RT: 14.022 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP021485.D
Acq: 13 Aug 2024 15:26

Instrument : BNA_P
ClientSampleId : SSTDICCC040


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#52
Acenaphthene
Concen: 40.406 ng
RT: 14.522 min Scan# 1961
Delta R.T. -0.000 min
Lab File: BP021485.D
Acq: 13 Aug 2024 15:26

Tgt Ion:154 Resp: 1911119
Ion Ratio Lower Upper
154 100
153 113.9 91.1 136.7
152 53.7 43.0 64.4







#53
3-Nitroaniline
Concen: 39.107 ng
RT: 14.351 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP021485.D
Acq: 13 Aug 2024 15:26

Instrument : BNA_P
ClientSampleId : SSTDICCC040


Manual Integrations APPROVED




Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#54
2,4-Dinitrophenol
Concen: 38.943 ng
RT: 14.551 min Scan# 1966
Delta R.T. 0.000 min
Lab File: BP021485.D
Acq: 13 Aug 2024 15:26

Tgt Ion:184 Resp: 116797
Ion Ratio Lower Upper
184 100
63 81.1 64.9 97.3
154 65.1 52.1 78.1





#55

Dibenzofuran

Concen: 40.359 ng

RT: 14.857 min Scan# 2

Delta R.T. 0.000 min

Lab File: BP021485.D

Acq: 13 Aug 2024 15:26

Instrument :

BNA_P





ClientSampleId :

SSTDICCC040

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#56

4-Nitrophenol

Concen: 39.073 ng

RT: 14.651 min Scan# 1983

Delta R.T. 0.000 min

Lab File: BP021485.D

Acq: 13 Aug 2024 15:26


Tgt Ion:139 Resp: 358912


Ion Ratio Lower Upper

139 100

109 90.2 70.2 110.2

65 121.8 101.8 141.8







#57
2,4-Dinitrotoluene
Concen: 38.397 ng
RT: 14.810 min Scan# 2130
Delta R.T. 0.000 min
Lab File: BP021485.D
Acq: 13 Aug 2024 15:26

Instrument : BNA_P
ClientSampleId : SSTDICCC040


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#58
Fluorene
Concen: 40.326 ng
RT: 15.516 min Scan# 2130
Delta R.T. 0.000 min
Lab File: BP021485.D
Acq: 13 Aug 2024 15:26

Tgt Ion:166 Resp: 2410461
Ion Ratio Lower Upper
166 100
165 97.6 78.1 117.1
167 13.7 11.0 16.4







#59
2,3,4,6-Tetrachlorophenol
Concen: 42.766 ng
RT: 15.087 min Scan# 2
Delta R.T. 0.000 min
Lab File: BP021485.D
Acq: 13 Aug 2024 15:26

Instrument : BNA_P
ClientSampleId : SSTDICCC040


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024




#60
Diethylphthalate
Concen: 40.656 ng
RT: 15.304 min Scan# 2094
Delta R.T. 0.000 min
Lab File: BP021485.D
Acq: 13 Aug 2024 15:26

Tgt Ion:149 Resp: 2486968
Ion Ratio Lower Upper
149 100
177 22.3 17.8 26.8
150 12.6 10.1 15.1





Abundance Scan 2131 (15.522 min): BP021485.D\data.ms



#61

4-Chlorophenyl-phenylether

Concen: 40.022 ng

RT: 15.522 min Scan# 2131

Delta R.T. 0.000 min

Lab File: BP021485.D

Acq: 13 Aug 2024 15:26

Instrument : BNA_P


ClientSampleId :

SSTDICCC040


**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



Abundance Scan 2132 (15.528 min): BP021485.D\data.ms



#62

4-Nitroaniline

Concen: 40.451 ng

RT: 15.528 min Scan# 2132

Delta R.T. -0.000 min

Lab File: BP021485.D

Acq: 13 Aug 2024 15:26


Tgt Ion:138 Resp: 485672


Ion Ratio Lower Upper

138 100

92 66.4 46.4 86.4

108 93.2 73.2 113.2





#63

Azobenzene

Concen: 40.594 ng

RT: 15.822 min Scan# 2182

Delta R.T. 0.000 min

Lab File: BP021485.D

Acq: 13 Aug 2024 15:26

Instrument :

BNA_P



ClientSampleId :

SSTDICCC040

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 17.251 min Scan# 2425

Delta R.T. 0.000 min

Lab File: BP021485.D

Acq: 13 Aug 2024 15:26


Tgt Ion:188 Resp: 1905279


Ion Ratio Lower Upper

188 100

94 12.8 10.2 15.4

80 14.5 11.6 17.4







#65
4,6-Dinitro-2-methylphenol
Concen: 38.603 ng
RT: 15.587 min Scan# 2142
Delta R.T. -0.000 min
Lab File: BP021485.D
Acq: 13 Aug 2024 15:26

Instrument : BNA_P
ClientSampleId : SSTDICCC040


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#66
n-Nitrosodiphenylamine
Concen: 40.175 ng
RT: 15.740 min Scan# 2168
Delta R.T. 0.000 min
Lab File: BP021485.D
Acq: 13 Aug 2024 15:26

Tgt Ion:169 Resp: 2051265
Ion Ratio Lower Upper
169 100
168 66.8 53.4 80.2
167 35.9 28.7 43.1






#67
4-Bromophenyl-phenylether
Concen: 39.211 ng
RT: 16.439 min Scan# 2
Delta R.T. 0.000 min
Lab File: BP021485.D
Acq: 13 Aug 2024 15:26

Instrument : BNA_P
ClientSampleId : SSTDICCC040


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#68
Hexachlorobenzene
Concen: 40.240 ng
RT: 16.551 min Scan# 2306
Delta R.T. 0.000 min
Lab File: BP021485.D
Acq: 13 Aug 2024 15:26

Tgt Ion:284 Resp: 958921
Ion Ratio Lower Upper
284 100
142 29.3 23.4 35.2
249 30.1 24.1 36.1







#69
Atrazine
Concen: 41.419 ng
RT: 16.716 min Scan# 2
Delta R.T. 0.000 min
Lab File: BP021485.D
Acq: 13 Aug 2024 15:26

Instrument : BNA_P
ClientSampleId : SSTDICCC040


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#70
Pentachlorophenol
Concen: 40.011 ng
RT: 16.892 min Scan# 2364
Delta R.T. 0.000 min
Lab File: BP021485.D
Acq: 13 Aug 2024 15:26

Tgt Ion:266 Resp: 520400
Ion Ratio Lower Upper
266 100
268 64.7 51.8 77.6
264 63.1 50.5 75.7







#71
Phenanthrene
Concen: 40.150 ng
RT: 17.298 min Scan# 2433
Delta R.T. 0.000 min
Lab File: BP021485.D
Acq: 13 Aug 2024 15:26

Instrument : BNA_P
ClientSampleId : SSTDICCC040


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#72
Anthracene
Concen: 40.283 ng
RT: 17.398 min Scan# 2450
Delta R.T. 0.000 min
Lab File: BP021485.D
Acq: 13 Aug 2024 15:26

Tgt Ion:178 Resp: 3619950
Ion Ratio Lower Upper
178 100
176 18.6 14.9 22.3
179 15.6 12.5 18.7







#73
Carbazole
Concen: 40.912 ng
RT: 17.669 min Scan# 2
Delta R.T. 0.000 min
Lab File: BP021485.D
Acq: 13 Aug 2024 15:26

Instrument : BNA_P
ClientSampleId : SSTDICCC040


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#74
Di-n-butylphthalate
Concen: 41.170 ng
RT: 18.275 min Scan# 2599
Delta R.T. 0.000 min
Lab File: BP021485.D
Acq: 13 Aug 2024 15:26

Tgt Ion:149 Resp: 4378720
Ion Ratio Lower Upper
149 100
150 9.4 7.5 11.3
104 6.1 4.9 7.3



#75

Fluoranthene

Concen: 40.674 ng

RT: 19.381 min Scan# 2

Delta R.T. 0.000 min

Lab File: BP021485.D




Acq: 13 Aug 2024 15:26

Instrument :

BNA_P

ClientSampleId :

SSTDICCC040



Tgt Ion:202 Resp: 457288

Ion Ratio Lower Upper

202 100

101 14.0 0.0 34.0





203 17.5 0.0 37.5

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#76

Chrysene-d₁₂

Concen: 20.000 ng

RT: 21.698 min Scan# 3181

Delta R.T. 0.000 min

Lab File: BP021485.D

Acq: 13 Aug 2024 15:26


Tgt Ion:240 Resp: 1820074


Ion Ratio Lower Upper

240 100

120 12.4 9.9 14.9

236 25.7 20.6 30.8







#77
Benzidine
Concen: 41.023 ng
RT: 19.575 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP021485.D
Acq: 13 Aug 2024 15:26

Instrument : BNA_P
ClientSampleId : SSTDICCC040


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#78
Pyrene
Concen: 39.461 ng
RT: 19.751 min Scan# 2850
Delta R.T. 0.000 min
Lab File: BP021485.D
Acq: 13 Aug 2024 15:26

Tgt Ion:202 Resp: 4694393
Ion Ratio Lower Upper
202 100
200 20.9 16.7 25.1
203 17.6 14.1 21.1







#79
Terphenyl-d14
Concen: 78.155 ng
RT: 19.975 min Scan# 2
Delta R.T. -0.000 min
Lab File: BP021485.D
Acq: 13 Aug 2024 15:26

Instrument : BNA_P
ClientSampleId : SSTDICCC040


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#80
Butylbenzylphthalate
Concen: 38.676 ng
RT: 20.710 min Scan# 3013
Delta R.T. -0.000 min
Lab File: BP021485.D
Acq: 13 Aug 2024 15:26

Tgt Ion:149 Resp: 1712106
Ion Ratio Lower Upper
149 100
91 91.6 73.3 109.9
206 21.4 17.1 25.7





#81

Benzo(a)anthracene

Concen: 40.395 ng

RT: 21.680 min Scan# 3178

Delta R.T. 0.000 min

Lab File: BP021485.D

Acq: 13 Aug 2024 15:26

Instrument :

BNA_P



ClientSampleId :

SSTDICCC040

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#82

3,3'-Dichlorobenzidine

Concen: 42.572 ng


RT: 21.610 min Scan# 3166


Delta R.T. 0.000 min

Lab File: BP021485.D

Acq: 13 Aug 2024 15:26

| Tgt | Ion: | 252 | Resp: | 1559563 |
|-----|-------|-------|-------|---------|
| Ion | Ratio | Lower | Upper | |
| 252 | 100 | | | |
| 254 | 64.6 | 51.7 | 77.5 | |
| 126 | 10.9 | 8.7 | 13.1 | |







#83
Chrysene
Concen: 40.405 ng
RT: 21.751 min Scan# 3190
Delta R.T. 0.000 min
Lab File: BP021485.D
Acq: 13 Aug 2024 15:26

Instrument : BNA_P
ClientSampleId : SSTDICCC040


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#84
Bis(2-ethylhexyl)phthalate
Concen: 42.082 ng
RT: 21.645 min Scan# 3172
Delta R.T. 0.000 min
Lab File: BP021485.D
Acq: 13 Aug 2024 15:26

Tgt Ion:149 Resp: 2841806
Ion Ratio Lower Upper
149 100
167 27.9 22.3 33.5
279 4.7 3.8 5.6







#85
Di-n-octyl phthalate
Concen: 39.238 ng
RT: 22.963 min Scan# 3
Delta R.T. -0.000 min
Lab File: BP021485.D
Acq: 13 Aug 2024 15:26

Instrument : BNA_P
ClientSampleId : SSTDICCC040


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#86
Perylene-d₁₂
Concen: 20.000 ng
RT: 25.104 min Scan# 3760
Delta R.T. 0.000 min
Lab File: BP021485.D
Acq: 13 Aug 2024 15:26

Tgt Ion:264 Resp: 2085332
Ion Ratio Lower Upper
264 100
260 23.3 18.6 28.0
265 21.4 17.1 25.7






#87

Indeno(1,2,3-cd)pyrene
Concen: 40.941 ng/m
RT: 28.992 min Scan# 4
Delta R.T. 0.000 min
Lab File: BP021485.D
Acq: 13 Aug 2024 15:26

Instrument : BNA_P
ClientSampleId : SSTDICCC040

Manual Integrations APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#88

Benzo(b)fluoranthene
Concen: 40.211 ng
RT: 24.027 min Scan# 3577
Delta R.T. 0.000 min
Lab File: BP021485.D
Acq: 13 Aug 2024 15:26

Tgt Ion:252 Resp: 4815875
Ion Ratio Lower Upper
252 100
253 22.0 17.6 26.4
125 10.8 8.6 13.0



#89

Benzo(k)fluoranthene

Concen: 40.717 ng

RT: 24.110 min Scan# 3

Delta R.T. 0.000 min

Lab File: BP021485.D




Acq: 13 Aug 2024 15:26

Instrument :

BNA_P

ClientSampleId :

SSTDICCC040



Tgt Ion:252 Resp: 4757000

Ion Ratio Lower Upper

252 100





253 21.9 17.5 26.3

125 9.8 7.8 11.8

Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#90
Benzo(a)pyrene
Concen: 40.616 ng
RT: 24.951 min Scan# 3734
Delta R.T. 0.000 min
Lab File: BP021485.D
Acq: 13 Aug 2024 15:26


Tgt Ion:252 Resp: 4221348

Ion Ratio Lower Upper

252 100

253 22.1 17.7 26.5

125 10.9 8.7 13.1



#91

Dibenzo(a,h)anthracene

Concen: 40.908 ng

RT: 29.086 min Scan# 4

Delta R.T. 0.000 min

Lab File: BP021485.D

Acq: 13 Aug 2024 15:26

Instrument :

BNA_P

ClientSampleId :

SSTDICCC040

Tgt Ion:278 Resp: 4594554

Ion Ratio Lower Upper

278 100

139 15.3 12.2 18.4

279 23.9 19.1 28.7

Manual Integrations


APPROVED

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024

Abundance

29.086



#92

Benzo(g,h,i)perylene

Concen: 40.824 ng

RT: 30.174 min Scan# 4622

Delta R.T. 0.000 min

Lab File: BP021485.D

Acq: 13 Aug 2024 15:26

Tgt Ion:276 Resp: 4595577

Ion Ratio Lower Upper




276 100

277 23.8 19.0 28.6

138 20.0 16.0 24.0

Abundance

30.174



Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081324\

Data File : BP021486.D

Acq On : 13 Aug 2024 16:09

Operator : RC/JU

Sample : SSTDICC050

Misc :

ALS Vial : 7 Sample Multiplier: 1

Quant Time: Aug 13 23:54:25 2024

Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M

Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

QLast Update : Tue Aug 13 23:44:24 2024

Response via : Initial Calibration

Instrument :

BNA_P

ClientSampleId :

SSTDICC050

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|-------------------------------|--------|------|----------|---------|-------|----------|
| Internal Standards | | | | | | |
| 1) 1,4-Dichlorobenzene-d4 | 7.805 | 152 | 410033 | 20.000 | ng | 0.00 |
| 21) Naphthalene-d8 | 10.593 | 136 | 1728756 | 20.000 | ng | 0.00 |
| 39) Acenaphthene-d10 | 14.446 | 164 | 1119203 | 20.000 | ng | -0.01 |
| 64) Phenanthrene-d10 | 17.257 | 188 | 2278560 | 20.000 | ng | 0.00 |
| 76) Chrysene-d12 | 21.710 | 240 | 2093381 | 20.000 | ng | 0.01 |
| 86) Perylene-d12 | 25.121 | 264 | 2377136 | 20.000 | ng | 0.02 |
| System Monitoring Compounds | | | | | | |
| 5) 2-Fluorophenol | 5.393 | 112 | 2713912 | 98.280 | ng | 0.00 |
| 7) Phenol-d6 | 6.964 | 99 | 4004371 | 99.407 | ng | 0.00 |
| 23) Nitrobenzene-d5 | 8.958 | 82 | 3559213 | 106.266 | ng | 0.00 |
| 42) 2,4,6-Tribromophenol | 15.969 | 330 | 1295133 | 103.974 | ng | 0.00 |
| 45) 2-Fluorobiphenyl | 13.063 | 172 | 7073510 | 96.455 | ng | 0.00 |
| 79) Terphenyl-d14 | 19.981 | 244 | 10188627 | 94.554 | ng | 0.00 |
| Target Compounds | | | | Qvalue | | |
| 2) 1,4-Dioxane | 3.317 | 88 | 614292 | 46.916 | ng | 99 |
| 3) Pyridine | 3.705 | 79 | 1785033m | 48.836 | ng | |
| 4) n-Nitrosodimethylamine | 3.617 | 42 | 529008 | 48.151 | ng | 99 |
| 6) Aniline | 7.134 | 93 | 1973018 | 48.993 | ng | 100 |
| 8) 2-Chlorophenol | 7.369 | 128 | 1266906 | 49.646 | ng | 98 |
| 9) Benzaldehyde | 6.946 | 77 | 1099572 | 42.632 | ng | 99 |
| 10) Phenol | 6.987 | 94 | 2070173 | 49.628 | ng | 99 |
| 11) bis(2-Chloroethyl)ether | 7.228 | 93 | 1722098 | 48.789 | ng | 100 |
| 12) 1,3-Dichlorobenzene | 7.699 | 146 | 1457386 | 48.079 | ng | 99 |
| 13) 1,4-Dichlorobenzene | 7.840 | 146 | 1470548 | 48.044 | ng | 99 |
| 14) 1,2-Dichlorobenzene | 8.158 | 146 | 1411170 | 47.842 | ng | 99 |
| 15) Benzyl Alcohol | 8.034 | 79 | 1452516 | 50.255 | ng | 99 |
| 16) 2,2'-oxybis(1-Chloropr... | 8.334 | 45 | 1588042 | 48.854 | ng | 99 |
| 17) 2-Methylphenol | 8.234 | 107 | 1308851 | 49.632 | ng | 99 |
| 18) Hexachloroethane | 8.887 | 117 | 601480 | 49.004 | ng | 97 |
| 19) n-Nitroso-di-n-propyla... | 8.611 | 70 | 1340426 | 48.171 | ng | 99 |
| 20) 3+4-Methylphenols | 8.564 | 107 | 1777289 | 49.982 | ng | 98 |
| 22) Acetophenone | 8.616 | 105 | 2552368 | 48.058 | ng | 99 |
| 24) Nitrobenzene | 8.999 | 77 | 1881606 | 51.725 | ng | 98 |
| 25) Isophorone | 9.522 | 82 | 3789274 | 47.889 | ng | 99 |
| 26) 2-Nitrophenol | 9.705 | 139 | 464597 | 49.970 | ng | 98 |
| 27) 2,4-Dimethylphenol | 9.763 | 122 | 956577 | 48.987 | ng | 98 |
| 28) bis(2-Chloroethoxy)met... | 10.005 | 93 | 2342749 | 48.504 | ng | 100 |
| 29) 2,4-Dichlorophenol | 10.234 | 162 | 1189252 | 51.545 | ng | 98 |
| 30) 1,2,4-Trichlorobenzene | 10.458 | 180 | 1428575 | 48.296 | ng | 99 |
| 31) Naphthalene | 10.646 | 128 | 4330348 | 47.949 | ng | 100 |
| 32) Benzoic acid | 9.899 | 122 | 652668 | 50.213 | ng | 99 |
| 33) 4-Chloroaniline | 10.746 | 127 | 1640026 | 48.345 | ng | 99 |
| 34) Hexachlorobutadiene | 10.946 | 225 | 906501 | 48.672 | ng | 100 |
| 35) Caprolactam | 11.510 | 113 | 469071 | 48.850 | ng | 98 |
| 36) 4-Chloro-3-methylphenol | 11.863 | 107 | 1600490 | 50.008 | ng | 99 |
| 37) 2-Methylnaphthalene | 12.263 | 142 | 2950679 | 47.872 | ng | 100 |
| 38) 1-Methylnaphthalene | 12.481 | 142 | 2908087 | 47.838 | ng | 100 |
| 40) 1,2,4,5-Tetrachloroben... | 12.628 | 216 | 1617735 | 49.332 | ng | 100 |
| 41) Hexachlorocyclopentadiene | 12.616 | 237 | 546269 | 52.560 | ng | 99 |
| 43) 2,4,6-Trichlorophenol | 12.863 | 196 | 949480 | 52.124 | ng | 99 |

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081324\
 Data File : BP021486.D
 Acq On : 13 Aug 2024 16:09
 Operator : RC/JU
 Sample : SSTDICC050
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024
 Supervised By :mohammad ahmed 08/14/2024

Quant Time: Aug 13 23:54:25 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 13 23:44:24 2024
 Response via : Initial Calibration

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|-------------------------------|--------|------|----------|--------|-------|----------|
| 44) 2,4,5-Trichlorophenol | 12.934 | 196 | 1056277 | 52.191 | ng | 98 |
| 46) 1,1'-Biphenyl | 13.275 | 154 | 3811862 | 48.139 | ng | 100 |
| 47) 2-Chloronaphthalene | 13.310 | 162 | 2979553 | 48.332 | ng | 98 |
| 48) 2-Nitroaniline | 13.504 | 65 | 821676 | 47.603 | ng | 98 |
| 49) Acenaphthylene | 14.169 | 152 | 4367223 | 48.190 | ng | 100 |
| 50) Dimethylphthalate | 13.910 | 163 | 3569162 | 47.620 | ng | 99 |
| 51) 2,6-Dinitrotoluene | 14.010 | 165 | 731267 | 48.330 | ng | 99 |
| 52) Acenaphthene | 14.516 | 154 | 2861701 | 48.126 | ng | 100 |
| 53) 3-Nitroaniline | 14.340 | 138 | 710103 | 47.788 | ng | 97 |
| 54) 2,4-Dinitrophenol | 14.551 | 184 | 205628 | 49.369 | ng | 97 |
| 55) Dibenzofuran | 14.857 | 168 | 4373407 | 47.811 | ng | 100 |
| 56) 4-Nitrophenol | 14.646 | 139 | 562138 | 47.429 | ng | 96 |
| 57) 2,4-Dinitrotoluene | 14.810 | 165 | 918371 | 47.777 | ng | 98 |
| 58) Fluorene | 15.522 | 166 | 3525015 | 46.908 | ng | 99 |
| 59) 2,3,4,6-Tetrachlorophenol | 15.081 | 232 | 885682 | 51.558 | ng | 99 |
| 60) Diethylphthalate | 15.304 | 149 | 3687157 | 47.945 | ng | 99 |
| 61) 4-Chlorophenyl-phenyle... | 15.516 | 204 | 1836235 | 47.272 | ng | 97 |
| 62) 4-Nitroaniline | 15.522 | 138 | 733231 | 47.987 | ng | 94 |
| 63) Azobenzene | 15.810 | 77 | 4349860 | 46.014 | ng | 99 |
| 65) 4,6-Dinitro-2-methylph... | 15.587 | 198 | 342052 | 49.749 | ng | 98 |
| 66) n-Nitrosodiphenylamine | 15.734 | 169 | 2980276 | 48.807 | ng | 100 |
| 67) 4-Bromophenyl-phenylether | 16.434 | 248 | 1198687 | 49.707 | ng | 98 |
| 68) Hexachlorobenzene | 16.545 | 284 | 1395323 | 48.961 | ng | 99 |
| 69) Atrazine | 16.716 | 200 | 1019051 | 48.608 | ng | 98 |
| 70) Pentachlorophenol | 16.892 | 266 | 812046 | 52.206 | ng | 99 |
| 71) Phenanthrene | 17.304 | 178 | 5308898 | 48.447 | ng | 100 |
| 72) Anthracene | 17.392 | 178 | 5212543 | 48.503 | ng | 100 |
| 73) Carbazole | 17.669 | 167 | 4974161 | 48.027 | ng | 99 |
| 74) Di-n-butylphthalate | 18.287 | 149 | 6454560 | 50.745 | ng | 100 |
| 75) Fluoranthene | 19.386 | 202 | 6476130 | 48.166 | ng | 99 |
| 77) Benzidine | 19.575 | 184 | 2493655 | 52.808 | ng | 100 |
| 78) Pyrene | 19.757 | 202 | 6603640 | 48.263 | ng | 99 |
| 80) Butylbenzylphthalate | 20.710 | 149 | 2559896 | 49.198 | ng | 98 |
| 81) Benzo(a)anthracene | 21.686 | 228 | 6405297 | 48.741 | ng | 99 |
| 82) 3,3'-Dichlorobenzidine | 21.610 | 252 | 2191535 | 52.013 | ng | 99 |
| 83) Chrysene | 21.763 | 228 | 6024463 | 48.555 | ng | 100 |
| 84) Bis(2-ethylhexyl)phtha... | 21.645 | 149 | 4205342 | 54.143 | ng | 100 |
| 85) Di-n-octyl phthalate | 22.969 | 149 | 6722548 | 49.144 | ng | 100 |
| 87) Indeno(1,2,3-cd)pyrene | 29.004 | 276 | 7707464m | 50.025 | ng | |
| 88) Benzo(b)fluoranthene | 24.039 | 252 | 6647186 | 48.689 | ng | 100 |
| 89) Benzo(k)fluoranthene | 24.116 | 252 | 6552987 | 49.204 | ng | 99 |
| 90) Benzo(a)pyrene | 24.957 | 252 | 5918922 | 49.959 | ng | 99 |
| 91) Dibenzo(a,h)anthracene | 29.092 | 278 | 6420896 | 50.151 | ng | 99 |
| 92) Benzo(g,h,i)perylene | 30.221 | 276 | 6383870 | 49.748 | ng | 99 |

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081324\
Data File : BP021486.D
Acq On : 13 Aug 2024 16:09
Operator : RC/JU
Sample : SSTDICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1


Quant Time: Aug 13 23:54:25 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 13 23:44:24 2024
Response via : Initial Calibration

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024







#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.805 min Scan# 81
Delta R.T. -0.006 min
Lab File: BP021486.D
Acq: 13 Aug 2024 16:09

Instrument : BNA_P
ClientSampleId : SSTDICC050


Manual Integrations
APPROVED




Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#2
1,4-Dioxane
Concen: 46.916 ng
RT: 3.317 min Scan# 56
Delta R.T. 0.000 min
Lab File: BP021486.D
Acq: 13 Aug 2024 16:09

Tgt Ion: 88 Resp: 614292
Ion Ratio Lower Upper
88 100
58 61.9 48.7 73.1
43 19.4 15.4 23.2









#3

Pyridine
Concen: 48.836 ng m
RT: 3.705 min Scan# 11
Delta R.T. 0.000 min
Lab File: BP021486.D
Acq: 13 Aug 2024 16:09

Instrument : BNA_P
ClientSampleId : SSTDICC050

Manual Integrations APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024




#4

n-Nitrosodimethylamine
Concen: 48.151 ng
RT: 3.617 min Scan# 107
Delta R.T. 0.000 min
Lab File: BP021486.D
Acq: 13 Aug 2024 16:09

Tgt Ion: 42 Resp: 529008
Ion Ratio Lower Upper
42 100
74 191.6 154.3 231.5
44 7.1 6.2 9.4







#5
2-Fluorophenol
Concen: 98.280 ng
RT: 5.393 min Scan# 409
Delta R.T. 0.000 min
Lab File: BP021486.D
Acq: 13 Aug 2024 16:09

Instrument : BNA_P
ClientSampleId : SSTDICC050


Manual Integrations
APPROVED




Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#6
Aniline
Concen: 48.993 ng
RT: 7.134 min Scan# 705
Delta R.T. 0.000 min
Lab File: BP021486.D
Acq: 13 Aug 2024 16:09

Tgt Ion: 93 Resp: 1973018
Ion Ratio Lower Upper
93 100
66 37.6 30.2 45.2
65 18.8 14.7 22.1





#7

Phenol-d6

Concen: 99.407 ng

RT: 6.964 min Scan# 6

Delta R.T. 0.000 min

Lab File: BP021486.D

Acq: 13 Aug 2024 16:09

Instrument :

BNA_P

ClientSampleId :

SSTDICC050





**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024

Abundance

6.964



#8

2-Chlorophenol

Concen: 49.646 ng

RT: 7.369 min Scan# 745

Delta R.T. -0.006 min

Lab File: BP021486.D

Acq: 13 Aug 2024 16:09

Tgt Ion:128 Resp: 1266906

Ion Ratio Lower Upper


128 100

130 31.8 11.6 51.6




64 57.0 35.3 75.3

Abundance

7.369



Abundance Scan 673 (6.946 min): BP021485.D\data.ms (-66)



#9

Benzaldehyde

Concen: 42.632 ng

RT: 6.946 min Scan# 6

Delta R.T. -0.000 min

Lab File: BP021486.D

Acq: 13 Aug 2024 16:09

Instrument :

BNA_P



ClientSampleId :

SSTDICC050


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024



Supervised By :mohammad ahmed 08/14/2024






#10
Phenol
Concen: 49.628 ng
RT: 6.987 min Scan# 680
Delta R.T. 0.000 min
Lab File: BP021486.D
Acq: 13 Aug 2024 16:09



Tgt Ion: 94 Resp: 2070173
Ion Ratio Lower Upper
94 100
65 25.8 6.5 46.5
66 34.5 14.7 54.7









#11
bis(2-Chloroethyl)ether
Concen: 48.789 ng
RT: 7.228 min Scan# 7
Delta R.T. 0.000 min
Lab File: BP021486.D
Acq: 13 Aug 2024 16:09

Instrument : BNA_P
ClientSampleId : SSTDICC050


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#12
1,3-Dichlorobenzene
Concen: 48.079 ng
RT: 7.699 min Scan# 801
Delta R.T. 0.000 min
Lab File: BP021486.D
Acq: 13 Aug 2024 16:09

Tgt Ion:146 Resp: 1457386
Ion Ratio Lower Upper
146 100
148 65.4 51.8 77.6
75 39.0 31.3 46.9







#13
1,4-Dichlorobenzene
Concen: 48.044 ng
RT: 7.840 min Scan# 81
Delta R.T. 0.000 min
Lab File: BP021486.D
Acq: 13 Aug 2024 16:09

Instrument : BNA_P
ClientSampleId : SSTDICC050


Manual Integrations APPROVED




Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#14
1,2-Dichlorobenzene
Concen: 47.842 ng
RT: 8.158 min Scan# 879
Delta R.T. 0.000 min
Lab File: BP021486.D
Acq: 13 Aug 2024 16:09

Tgt Ion:146 Resp: 1411170
Ion Ratio Lower Upper
146 100
148 65.1 51.5 77.3
111 50.9 41.1 61.7





#15

Benzyl Alcohol

Concen: 50.255 ng

RT: 8.034 min Scan# 8

Delta R.T. 0.000 min

Lab File: BP021486.D

Acq: 13 Aug 2024 16:09

Instrument :

BNA_P





ClientSampleId :

SSTDICC050

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#16

2,2'-oxybis(1-Chloropropane)

Concen: 48.854 ng


RT: 8.334 min Scan# 909




Delta R.T. 0.000 min

Lab File: BP021486.D

Acq: 13 Aug 2024 16:09

| Tgt | Ion | Resp: | 1588042 |
|-----|-----------|-------|---------|
| | Ion Ratio | Lower | Upper |
| 45 | 100 | | |
| 77 | 21.0 | 1.9 | 41.9 |
| 79 | 16.5 | 0.0 | 36.6 |





#17

2-Methylphenol

Concen: 49.632 ng

RT: 8.234 min Scan# 8

Delta R.T. 0.000 min

Lab File: BP021486.D

Acq: 13 Aug 2024 16:09

Instrument :

BNA_P




ClientSampleId :

SSTDICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#18

Hexachloroethane

Concen: 49.004 ng

RT: 8.887 min Scan# 1003

Delta R.T. -0.006 min

Lab File: BP021486.D

Acq: 13 Aug 2024 16:09


Tgt Ion:117 Resp: 601480




Ion Ratio Lower Upper

117 100

119 94.4 73.9 110.9

201 90.5 74.9 112.3





#19

n-Nitroso-di-n-propylamine
Concen: 48.171 ng
RT: 8.611 min Scan# 9




Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion: 70 Resp: 1340420

| | Ion Ratio | Lower | Upper |
|-----|-----------|-------|-------|
| 70 | 100 | | |
| 42 | 41.2 | 33.1 | 49.7 |
| 101 | 9.6 | 7.4 | 11.2 |
| 130 | 17.1 | 13.3 | 19.9 |

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024





#20

3+4-Methylphenols
Concen: 49.982 ng
RT: 8.564 min Scan# 948
Delta R.T. 0.006 min
Lab File: BP021486.D
Acq: 13 Aug 2024 16:09

Tgt Ion:107 Resp: 1777289

| | Ion Ratio | Lower | Upper |
|-----|-----------|-------|-------|
| 107 | 100 | | |
| 108 | 89.5 | 67.6 | 107.6 |
| 77 | 34.2 | 15.0 | 55.0 |
| 79 | 30.6 | 8.9 | 48.9 |







#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.593 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP021486.D
Acq: 13 Aug 2024 16:09

Instrument : BNA_P
ClientSampleId : SSTDICC050


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#22
Acetophenone
Concen: 48.058 ng
RT: 8.616 min Scan# 957
Delta R.T. 0.000 min
Lab File: BP021486.D
Acq: 13 Aug 2024 16:09

Tgt Ion:105 Resp: 2552368
Ion Ratio Lower Upper
105 100
71 6.4 4.6 7.0
51 19.5 16.3 24.5
120 20.2 16.2 24.2







#23
Nitrobenzene-d5
Concen: 106.266 ng
RT: 8.958 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP021486.D
Acq: 13 Aug 2024 16:09

Instrument : BNA_P
ClientSampleId : SSTDICC050


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#24
Nitrobenzene
Concen: 51.725 ng
RT: 8.999 min Scan# 1022
Delta R.T. 0.006 min
Lab File: BP021486.D
Acq: 13 Aug 2024 16:09

Tgt Ion: 77 Resp: 1881606
Ion Ratio Lower Upper
77 100
123 38.3 29.5 44.3
65 12.4 9.9 14.9







#25
 Isophorone
 Concen: 47.889 ng
 RT: 9.522 min Scan# 1
 Delta R.T. 0.000 min
 Lab File: BP021486.D
 Acq: 13 Aug 2024 16:09

Instrument : BNA_P
 ClientSampleId : SSTDICC050


Manual Integrations
APPROVED







Reviewed By :Yogesh Patel 08/14/2024
 Supervised By :mohammad ahmed 08/14/2024



#26
 2-Nitrophenol
 Concen: 49.970 ng
 RT: 9.705 min Scan# 1142
 Delta R.T. 0.006 min
 Lab File: BP021486.D
 Acq: 13 Aug 2024 16:09

Tgt Ion:139 Resp: 464597
 Ion Ratio Lower Upper
 139 100
 109 36.5 28.9 43.3
 65 63.5 49.0 73.6





#27

2,4-Dimethylphenol

Concen: 48.987 ng

RT: 9.763 min Scan# 1

Delta R.T. 0.006 min

Lab File: BP021486.D

Acq: 13 Aug 2024 16:09

Instrument :

BNA_P


ClientSampleId :

SSTDICC050

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#28

bis(2-Chloroethoxy)methane

Concen: 48.504 ng

RT: 10.005 min Scan# 1193

Delta R.T. 0.000 min

Lab File: BP021486.D

Acq: 13 Aug 2024 16:09


Tgt Ion: 93 Resp: 2342749


Ion Ratio Lower Upper

93 100

95 32.8 26.2 39.4

123 9.5 7.6 11.4





#29

2,4-Dichlorophenol

Concen: 51.545 ng

RT: 10.234 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP021486.D

Acq: 13 Aug 2024 16:09

Instrument :

BNA_P



ClientSampleId :

SSTDICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#30

1,2,4-Trichlorobenzene

Concen: 48.296 ng

RT: 10.458 min Scan# 1270

Delta R.T. 0.000 min

Lab File: BP021486.D

Acq: 13 Aug 2024 16:09


Tgt Ion:180 Resp: 1428575


Ion Ratio Lower Upper

180 100

182 96.7 78.3 117.5

145 30.5 24.0 36.0





#31

Naphthalene

Concen: 47.949 ng

RT: 10.646 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP021486.D

Acq: 13 Aug 2024 16:09

Instrument :

BNA_P

ClientSampleId :

SSTDICC050

Tgt Ion:128 Resp: 433034

Ion Ratio Lower Upper

128 100

129 10.7 8.6 12.8



127 13.1 10.6 15.8

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#32

Benzoic acid

Concen: 50.213 ng

RT: 9.899 min Scan# 1175

Delta R.T. 0.024 min

Lab File: BP021486.D

Acq: 13 Aug 2024 16:09


Tgt Ion:122 Resp: 652668


Ion Ratio Lower Upper

122 100

105 142.9 124.6 164.6

77 116.7 96.2 136.2







#33
4-Chloroaniline
Concen: 48.345 ng
RT: 10.746 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP021486.D
Acq: 13 Aug 2024 16:09

Instrument : BNA_P
ClientSampleId : SSTDICC050


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#34
Hexachlorobutadiene
Concen: 48.672 ng
RT: 10.946 min Scan# 1353
Delta R.T. 0.006 min
Lab File: BP021486.D
Acq: 13 Aug 2024 16:09

Tgt Ion:225 Resp: 906501
Ion Ratio Lower Upper
225 100
223 62.2 49.6 74.4
227 63.1 50.1 75.1





#35

Caprolactam

Concen: 48.850 ng

RT: 11.510 min Scan# 1449

Delta R.T. 0.012 min

Lab File: BP021486.D

Acq: 13 Aug 2024 16:09

Instrument :

BNA_P



ClientSampleId :

SSTDICC050

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#36

4-Chloro-3-methylphenol

Concen: 50.008 ng

RT: 11.863 min Scan# 1509


Delta R.T. 0.000 min


Lab File: BP021486.D

Acq: 13 Aug 2024 16:09

Tgt Ion:107 Resp: 1600490

| | Ion | Ratio | Lower | Upper |
|-----|------|-------|-------|-------|
| 107 | 100 | | | |
| 144 | 21.0 | 16.6 | 24.8 | |
| 142 | 64.3 | 51.0 | 76.4 | |





#37

2-Methylnaphthalene

Concen: 47.872 ng

RT: 12.263 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP021486.D

Acq: 13 Aug 2024 16:09

Instrument :

BNA_P

ClientSampleId :

SSTDICC050

Tgt Ion:142 Resp: 2950679

Ion Ratio Lower Upper

142 100

141 90.0 71.7 107.5



115 39.7 31.4 47.2

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#38

1-Methylnaphthalene

Concen: 47.838 ng

RT: 12.481 min Scan# 1614

Delta R.T. 0.000 min

Lab File: BP021486.D

Acq: 13 Aug 2024 16:09


Tgt Ion:142 Resp: 2908087


Ion Ratio Lower Upper

142 100

141 92.0 73.5 110.3

116 4.2 3.4 5.2





#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 14.446 min Scan# 1

Delta R.T. -0.012 min

Lab File: BP021486.D

Acq: 13 Aug 2024 16:09

Instrument :

BNA_P



ClientSampleId :

SSTDICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#40

1,2,4,5-Tetrachlorobenzene

Concen: 49.332 ng

RT: 12.628 min Scan# 1639

Delta R.T. 0.000 min

Lab File: BP021486.D

Acq: 13 Aug 2024 16:09

Tgt Ion:216 Resp: 1617735


Ion Ratio Lower Upper


216 100

214 78.1 62.5 93.7

179 20.3 16.4 24.6

108 20.5 16.7 25.1







#41
Hexachlorocyclopentadiene
Concen: 52.560 ng
RT: 12.616 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP021486.D
Acq: 13 Aug 2024 16:09

Instrument : BNA_P
ClientSampleId : SSTDICC050


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#42
2,4,6-Tribromophenol
Concen: 103.974 ng
RT: 15.969 min Scan# 2207
Delta R.T. 0.000 min
Lab File: BP021486.D
Acq: 13 Aug 2024 16:09

Tgt Ion:330 Resp: 1295133
Ion Ratio Lower Upper
330 100
332 96.5 77.4 116.0
141 29.8 24.5 36.7



#43

2,4,6-Trichlorophenol

Concen: 52.124 ng

RT: 12.863 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP021486.D


Acq: 13 Aug 2024 16:09

Instrument :


BNA_P

ClientSampleId :


SSTDICC050



Abundance Scan 1679 (12.863 min): BP021486.D\data.ms



Abundance Scan 1679 (12.863 min): BP021486.D\data.ms (-)



Tgt Ion:196 Resp: 949480

Ion Ratio Lower Upper

196 100

198 94.2 76.6 114.8



200 30.0 24.5 36.7

Manual Integrations


APPROVED

Reviewed By :Yogesh Patel 08/14/2024


Supervised By :mohammad ahmed 08/14/2024



Abundance Scan 1691 (12.934 min): BP021486.D\data.ms



Abundance Scan 1691 (12.934 min): BP021486.D\data.ms (-)



#44

2,4,5-Trichlorophenol

Concen: 52.191 ng

RT: 12.934 min Scan# 1691

Delta R.T. 0.000 min

Lab File: BP021486.D

Acq: 13 Aug 2024 16:09

Tgt Ion:196 Resp: 1056277


Ion Ratio Lower Upper




196 100

198 97.0 78.9 118.3

97 58.2 49.4 74.0

132 26.3 21.1 31.7





#45

2-Fluorobiphenyl

Concen: 96.455 ng

RT: 13.063 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP021486.D

Acq: 13 Aug 2024 16:09

Instrument :

BNA_P




ClientSampleId :

SSTDICC050

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#46

1,1'-Biphenyl

Concen: 48.139 ng

RT: 13.275 min Scan# 1749

Delta R.T. -0.006 min

Lab File: BP021486.D

Acq: 13 Aug 2024 16:09


Tgt Ion:154 Resp: 3811862


Ion Ratio Lower Upper

154 100

153 40.3 20.7 60.7

76 17.6 0.0 37.7







#47
2-Chloronaphthalene
Concen: 48.332 ng
RT: 13.310 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP021486.D
Acq: 13 Aug 2024 16:09

Instrument : BNA_P
ClientSampleId : SSTDICC050


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#48
2-Nitroaniline
Concen: 47.603 ng
RT: 13.504 min Scan# 1788
Delta R.T. 0.000 min
Lab File: BP021486.D
Acq: 13 Aug 2024 16:09

Tgt Ion: 65 Resp: 821676
Ion Ratio Lower Upper
65 100
92 73.6 58.1 87.1
138 84.3 65.5 98.3





#49

Acenaphthylene

Concen: 48.190 ng

RT: 14.169 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP021486.D

Acq: 13 Aug 2024 16:09

Instrument :

BNA_P



ClientSampleId :

SSTDICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#50

Dimethylphthalate

Concen: 47.620 ng

RT: 13.910 min Scan# 1857

Delta R.T. 0.000 min

Lab File: BP021486.D

Acq: 13 Aug 2024 16:09


Tgt Ion:163 Resp: 3569162


Ion Ratio Lower Upper

163 100

194 4.2 3.3 4.9

164 9.9 8.3 12.5







#51
2,6-Dinitrotoluene
Concen: 48.330 ng
RT: 14.010 min Scan# 1
Delta R.T. -0.012 min
Lab File: BP021486.D
Acq: 13 Aug 2024 16:09

Instrument : BNA_P
ClientSampleId : SSTDICC050


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#52
Acenaphthene
Concen: 48.126 ng
RT: 14.516 min Scan# 1960
Delta R.T. -0.006 min
Lab File: BP021486.D
Acq: 13 Aug 2024 16:09

Tgt Ion:154 Resp: 2861701
Ion Ratio Lower Upper
154 100
153 113.8 91.1 136.7
152 53.6 43.0 64.4







#53
 3-Nitroaniline
 Concen: 47.788 ng
 RT: 14.340 min Scan# 1
 Delta R.T. -0.011 min
 Lab File: BP021486.D
 Acq: 13 Aug 2024 16:09

Instrument : BNA_P
 ClientSampleId : SSTDICC050


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
 Supervised By :mohammad ahmed 08/14/2024



#54
 2,4-Dinitrophenol
 Concen: 49.369 ng
 RT: 14.551 min Scan# 1966
 Delta R.T. 0.000 min
 Lab File: BP021486.D
 Acq: 13 Aug 2024 16:09

Tgt Ion:184 Resp: 205628
 Ion Ratio Lower Upper
 184 100
 63 77.6 64.9 97.3
 154 64.3 52.1 78.1







#55
Dibenzofuran
Concen: 47.811 ng
RT: 14.857 min Scan# 2
Instrument : BNA_P
Delta R.T. 0.000 min
Lab File: BP021486.D
Acq: 13 Aug 2024 16:09
ClientSampleId : SSTDICC050


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#56
4-Nitrophenol
Concen: 47.429 ng
RT: 14.646 min Scan# 1982
Delta R.T. -0.005 min
Lab File: BP021486.D
Acq: 13 Aug 2024 16:09

Tgt Ion:139 Resp: 562138
Ion Ratio Lower Upper
139 100
109 86.5 70.2 110.2
65 117.6 101.8 141.8







#57
2,4-Dinitrotoluene
Concen: 47.777 ng
RT: 14.810 min Scan# 211837
Delta R.T. 0.000 min
Lab File: BP021486.D
Acq: 13 Aug 2024 16:09

Instrument : BNA_P
ClientSampleId : SSTDICC050


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#58
Fluorene
Concen: 46.908 ng
RT: 15.522 min Scan# 2131
Delta R.T. 0.006 min
Lab File: BP021486.D
Acq: 13 Aug 2024 16:09

Tgt Ion:166 Resp: 3525015
Ion Ratio Lower Upper
166 100
165 98.2 78.1 117.1
167 13.5 11.0 16.4







#59
2,3,4,6-Tetrachlorophenol
Concen: 51.558 ng
RT: 15.081 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP021486.D
Acq: 13 Aug 2024 16:09

Instrument : BNA_P
ClientSampleId : SSTDICC050


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024




#60
Diethylphthalate
Concen: 47.945 ng
RT: 15.304 min Scan# 2094
Delta R.T. 0.000 min
Lab File: BP021486.D
Acq: 13 Aug 2024 16:09

Tgt Ion:149 Resp: 3687157
Ion Ratio Lower Upper
149 100
177 23.2 17.8 26.8
150 12.4 10.1 15.1





Abundance Scan 2130 (15.516 min): BP021486.D\data.ms



#61

4-Chlorophenyl-phenylether

Concen: 47.272 ng

RT: 15.516 min Scan# 2131

Delta R.T. -0.006 min

Lab File: BP021486.D

Acq: 13 Aug 2024 16:09

Instrument : BNA_P


ClientSampleId :

SSTDICC050


**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



Abundance Scan 2131 (15.522 min): BP021486.D\data.ms



#62

4-Nitroaniline

Concen: 47.987 ng

RT: 15.522 min Scan# 2131

Delta R.T. -0.006 min

Lab File: BP021486.D

Acq: 13 Aug 2024 16:09


Tgt Ion:138 Resp: 733231


Ion Ratio Lower Upper

138 100

92 62.4 46.4 86.4

108 87.4 73.2 113.2






#63
Azobenzene
Concen: 46.014 ng
RT: 15.810 min Scan# 2
Delta R.T. -0.012 min
Lab File: BP021486.D
Acq: 13 Aug 2024 16:09

Instrument : BNA_P
ClientSampleId : SSTDICC050


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.257 min Scan# 2426
Delta R.T. 0.006 min
Lab File: BP021486.D
Acq: 13 Aug 2024 16:09

Tgt Ion:188 Resp: 2278560
Ion Ratio Lower Upper
188 100
94 12.2 10.2 15.4
80 14.1 11.6 17.4







#65
4,6-Dinitro-2-methylphenol
Concen: 49.749 ng
RT: 15.587 min Scan# 2142
Delta R.T. -0.000 min
Lab File: BP021486.D
Acq: 13 Aug 2024 16:09

Instrument : BNA_P
ClientSampleId : SSTDICC050


Manual Integrations
APPROVED




Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#66
n-Nitrosodiphenylamine
Concen: 48.807 ng
RT: 15.734 min Scan# 2167
Delta R.T. -0.006 min
Lab File: BP021486.D
Acq: 13 Aug 2024 16:09

Tgt Ion:169 Resp: 2980276
Ion Ratio Lower Upper
169 100
168 66.6 53.4 80.2
167 36.1 28.7 43.1









#67
4-Bromophenyl-phenylether
Concen: 49.707 ng
RT: 16.434 min Scan# 219868
Delta R.T. -0.006 min
Lab File: BP021486.D
Acq: 13 Aug 2024 16:09

Instrument : BNA_P
ClientSampleId : SSTDICC050


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#68
Hexachlorobenzene
Concen: 48.961 ng
RT: 16.545 min Scan# 2305
Delta R.T. -0.006 min
Lab File: BP021486.D
Acq: 13 Aug 2024 16:09

Tgt Ion:284 Resp: 1395323
Ion Ratio Lower Upper
284 100
142 30.0 23.4 35.2
249 30.4 24.1 36.1







#69
Atrazine
Concen: 48.608 ng
RT: 16.716 min Scan# 2
Delta R.T. 0.000 min
Lab File: BP021486.D
Acq: 13 Aug 2024 16:09

Instrument : BNA_P
ClientSampleId : SSTDICC050


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024




#70
Pentachlorophenol
Concen: 52.206 ng
RT: 16.892 min Scan# 2364
Delta R.T. 0.000 min
Lab File: BP021486.D
Acq: 13 Aug 2024 16:09

Tgt Ion:266 Resp: 812046
Ion Ratio Lower Upper
266 100
268 63.1 51.8 77.6
264 63.1 50.5 75.7






Abundance Scan 2434 (17.304 min): BP021486.D\data.ms



Abundance Scan 2434 (17.304 min): BP021486.D\data.ms (-)



#71

Phenanthrene

Concen: 48.447 ng

RT: 17.304 min Scan# 2449

Delta R.T. 0.006 min

Lab File: BP021486.D

Acq: 13 Aug 2024 16:09

Instrument :

BNA_P



ClientSampleId :

SSTDICC050


**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/14/2024


Supervised By :mohammad ahmed 08/14/2024



Abundance Scan 2449 (17.392 min): BP021486.D\data.ms



Abundance Scan 2449 (17.392 min): BP021486.D\data.ms (-)



#72

Anthracene

Concen: 48.503 ng

RT: 17.392 min Scan# 2449

Delta R.T. -0.006 min

Lab File: BP021486.D

Acq: 13 Aug 2024 16:09


Tgt Ion:178 Resp: 5212543


Ion Ratio Lower Upper

178 100

176 18.8 14.9 22.3

179 15.8 12.5 18.7







#73
Carbazole
Concen: 48.027 ng
RT: 17.669 min Scan# 2
Delta R.T. 0.000 min
Lab File: BP021486.D
Acq: 13 Aug 2024 16:09

Instrument : BNA_P
ClientSampleId : SSTDICC050


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#74
Di-n-butylphthalate
Concen: 50.745 ng
RT: 18.287 min Scan# 2601
Delta R.T. 0.012 min
Lab File: BP021486.D
Acq: 13 Aug 2024 16:09

Tgt Ion:149 Resp: 6454560
Ion Ratio Lower Upper
149 100
150 9.4 7.5 11.3
104 6.0 4.9 7.3







#75
Fluoranthene
Concen: 48.166 ng
RT: 19.386 min Scan# 2
Delta R.T. 0.006 min
Lab File: BP021486.D
Acq: 13 Aug 2024 16:09

Instrument : BNA_P
ClientSampleId : SSTDICC050


Manual Integrations
APPROVED




Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#76
Chrysene-d₁₂
Concen: 20.000 ng
RT: 21.710 min Scan# 3183
Delta R.T. 0.012 min
Lab File: BP021486.D
Acq: 13 Aug 2024 16:09

Tgt Ion:240 Resp: 2093381
Ion Ratio Lower Upper
240 100
120 11.9 9.9 14.9
236 25.7 20.6 30.8









#77
Benzidine
Concen: 52.808 ng
RT: 19.575 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP021486.D
Acq: 13 Aug 2024 16:09

Instrument : BNA_P
ClientSampleId : SSTDICC050


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#78
Pyrene
Concen: 48.263 ng
RT: 19.757 min Scan# 2851
Delta R.T. 0.006 min
Lab File: BP021486.D
Acq: 13 Aug 2024 16:09

Tgt Ion:202 Resp: 6603640
Ion Ratio Lower Upper
202 100
200 21.3 16.7 25.1
203 17.9 14.1 21.1



#79

Terphenyl-d14

Concen: 94.554 ng

RT: 19.981 min Scan# 2

Delta R.T. 0.006 min

Lab File: BP021486.D




Acq: 13 Aug 2024 16:09

Instrument :

BNA_P

ClientSampleId :

SSTDICC050



Tgt Ion:244 Resp:1018862

Ion Ratio Lower Upper

244 100





212 7.0 5.6 8.4

122 10.8 9.4 14.2

Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024

#80
Butylbenzylphthalate
Concen: 49.198 ng
RT: 20.710 min Scan# 3013
Delta R.T. -0.000 min
Lab File: BP021486.D
Acq: 13 Aug 2024 16:09


Tgt Ion:149 Resp: 2559896


Ion Ratio Lower Upper

149 100

91 93.4 73.3 109.9

206 21.8 17.1 25.7





#81

Benzo(a)anthracene

Concen: 48.741 ng

RT: 21.686 min Scan# 3178

Delta R.T. 0.006 min

Lab File: BP021486.D

Acq: 13 Aug 2024 16:09

Instrument :

BNA_P



ClientSampleId :

SSTDICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#82

3,3'-Dichlorobenzidine

Concen: 52.013 ng


RT: 21.610 min Scan# 3166


Delta R.T. 0.000 min

Lab File: BP021486.D

Acq: 13 Aug 2024 16:09

| Tgt | Ion:252 | Resp: 2191535 | |
|-----|---------|---------------|-------|
| Ion | Ratio | Lower | Upper |
| 252 | 100 | | |
| 254 | 64.3 | 51.7 | 77.5 |
| 126 | 10.3 | 8.7 | 13.1 |







#83

Chrysene
Concen: 48.555 ng
RT: 21.763 min Scan# 3
Delta R.T. 0.012 min
Lab File: BP021486.D
Acq: 13 Aug 2024 16:09

Instrument : BNA_P
ClientSampleId : SSTDICC050

Manual Integrations APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024




#84

Bis(2-ethylhexyl)phthalate
Concen: 54.143 ng
RT: 21.645 min Scan# 3172
Delta R.T. 0.000 min
Lab File: BP021486.D
Acq: 13 Aug 2024 16:09

Tgt Ion:149 Resp: 4205342
Ion Ratio Lower Upper
149 100
167 27.9 22.3 33.5
279 4.8 3.8 5.6





#85

Di-n-octyl phthalate

Concen: 49.144 ng

RT: 22.969 min Scan# 3

Delta R.T. 0.006 min

Lab File: BP021486.D

Acq: 13 Aug 2024 16:09

Instrument :

BNA_P



ClientSampleId :

SSTDICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#86

Perylene-d₁₂

Concen: 20.000 ng

RT: 25.121 min Scan# 3763

Delta R.T. 0.018 min

Lab File: BP021486.D

Acq: 13 Aug 2024 16:09


Tgt Ion:264 Resp: 2377136

Ion Ratio Lower Upper

264 100

260 23.5 18.6 28.0

265 21.2 17.1 25.7



#87

Indeno(1,2,3-cd)pyrene

Concen: 50.025 ng m

RT: 29.004 min Scan# 4

Delta R.T. 0.012 min

Lab File: BP021486.D



Acq: 13 Aug 2024 16:09

Instrument :

BNA_P

ClientSampleId :

SSTDICC050



Tgt Ion:276 Resp: 770746

Ion Ratio Lower Upper

276 100





138 15.5 12.8 19.2

277 20.6 16.5 24.7

Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#88

Benzo(b)fluoranthene

Concen: 48.689 ng

RT: 24.039 min Scan# 3579

Delta R.T. 0.012 min

Lab File: BP021486.D

Acq: 13 Aug 2024 16:09


Tgt Ion:252 Resp: 6647186

Ion Ratio Lower Upper

252 100

253 22.1 17.6 26.4

125 10.6 8.6 13.0



#89

Benzo(k)fluoranthene

Concen: 49.204 ng

RT: 24.116 min Scan# 3

Delta R.T. 0.006 min

Lab File: BP021486.D




Acq: 13 Aug 2024 16:09

Instrument :

BNA_P

ClientSampleId :

SSTDICC050



Tgt Ion:252 Resp: 655298

Ion Ratio Lower Upper

252 100





253 22.3 17.5 26.3

125 10.0 7.8 11.8

Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#90

Benzo(a)pyrene

Concen: 49.959 ng

RT: 24.957 min Scan# 3735

Delta R.T. 0.006 min

Lab File: BP021486.D

Acq: 13 Aug 2024 16:09


Tgt Ion:252 Resp: 5918922







Ion Ratio Lower Upper

252 100

253 21.8 17.7 26.5

125 11.2 8.7 13.1





#91

Dibenzo(a,h)anthracene

Concen: 50.151 ng

RT: 29.092 min Scan# 4438

Delta R.T. 0.006 min

Lab File: BP021486.D

Acq: 13 Aug 2024 16:09

Instrument :

BNA_P


ClientSampleId :

SSTDICC050

Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#92

Benzo(g,h,i)perylene

Concen: 49.748 ng

RT: 30.221 min Scan# 4630

Delta R.T. 0.047 min


Lab File: BP021486.D

Acq: 13 Aug 2024 16:09

Tgt Ion:276 Resp: 6383870

Ion Ratio Lower Upper

| | 100 | | |
|-----|------|------|------|
| 276 | 100 | | |
| 277 | 23.5 | 19.0 | 28.6 |
| 138 | 20.7 | 16.0 | 24.0 |



Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081324\
 Data File : BP021487.D
 Acq On : 13 Aug 2024 16:53
 Operator : RC/JU
 Sample : SSTDICC060
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC060

Quant Time: Aug 13 23:48:28 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 13 23:44:24 2024
 Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024
 Supervised By :mohammad ahmed 08/14/2024

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|------------------------------------|--------|------|----------|---------|-------|----------|
| Internal Standards | | | | | | |
| 1) 1,4-Dichlorobenzene-d4 | 7.810 | 152 | 382208 | 20.000 | ng | 0.00 |
| 21) Naphthalene-d8 | 10.599 | 136 | 1539217 | 20.000 | ng | 0.00 |
| 39) Acenaphthene-d10 | 14.457 | 164 | 953193 | 20.000 | ng | 0.00 |
| 64) Phenanthrene-d10 | 17.257 | 188 | 1991489 | 20.000 | ng | 0.00 |
| 76) Chrysene-d12 | 21.710 | 240 | 1907005 | 20.000 | ng | 0.01 |
| 86) Perylene-d12 | 25.121 | 264 | 2191159 | 20.000 | ng | 0.02 |
| System Monitoring Compounds | | | | | | |
| 5) 2-Fluorophenol | 5.393 | 112 | 3212821 | 124.817 | ng | 0.00 |
| 7) Phenol-d6 | 6.963 | 99 | 4582988 | 122.053 | ng | 0.00 |
| 23) Nitrobenzene-d5 | 8.957 | 82 | 4024825 | 134.966 | ng | 0.00 |
| 42) 2,4,6-Tribromophenol | 15.969 | 330 | 1401053 | 132.067 | ng | 0.00 |
| 45) 2-Fluorobiphenyl | 13.069 | 172 | 7484438 | 119.833 | ng | 0.00 |
| 79) Terphenyl-d14 | 19.986 | 244 | 11374626 | 115.877 | ng | 0.01 |
| Target Compounds | | | | | | |
| | | | | Qvalue | | |
| 2) 1,4-Dioxane | 3.322 | 88 | 739240 | 60.569 | ng | 99 |
| 3) Pyridine | 3.705 | 79 | 2102990m | 61.723 | ng | |
| 4) n-Nitrosodimethylamine | 3.622 | 42 | 625716 | 61.099 | ng | 99 |
| 6) Aniline | 7.134 | 93 | 2239942 | 59.670 | ng | 100 |
| 8) 2-Chlorophenol | 7.375 | 128 | 1472845 | 61.917 | ng | 98 |
| 9) Benzaldehyde | 6.946 | 77 | 1212390 | 50.428 | ng | 98 |
| 10) Phenol | 6.993 | 94 | 2381743 | 61.254 | ng | 98 |
| 11) bis(2-Chloroethyl)ether | 7.234 | 93 | 1938063 | 58.905 | ng | 99 |
| 12) 1,3-Dichlorobenzene | 7.699 | 146 | 1682261 | 59.538 | ng | 100 |
| 13) 1,4-Dichlorobenzene | 7.846 | 146 | 1709124 | 59.904 | ng | 99 |
| 14) 1,2-Dichlorobenzene | 8.157 | 146 | 1624118 | 59.070 | ng | 100 |
| 15) Benzyl Alcohol | 8.034 | 79 | 1647924 | 61.166 | ng | 99 |
| 16) 2,2'-oxybis(1-Chloropr... | 8.334 | 45 | 1756837 | 57.982 | ng | 100 |
| 17) 2-Methylphenol | 8.234 | 107 | 1477294 | 60.098 | ng | 99 |
| 18) Hexachloroethane | 8.893 | 117 | 704431 | 61.570 | ng | 99 |
| 19) n-Nitroso-di-n-propyla... | 8.610 | 70 | 1457549 | 56.194 | ng | 99 |
| 20) 3+4-Methylphenols | 8.563 | 107 | 2028575 | 61.202 | ng | 99 |
| 22) Acetophenone | 8.616 | 105 | 2813557 | 59.500 | ng | # 99 |
| 24) Nitrobenzene | 8.999 | 77 | 2126322 | 65.650 | ng | 98 |
| 25) Isophorone | 9.522 | 82 | 4092218 | 58.086 | ng | 99 |
| 26) 2-Nitrophenol | 9.704 | 139 | 554873 | 60.117 | ng | 99 |
| 27) 2,4-Dimethylphenol | 9.763 | 122 | 1074335 | 61.792 | ng | 97 |
| 28) bis(2-Chloroethoxy)met... | 10.004 | 93 | 2556281 | 59.443 | ng | 100 |
| 29) 2,4-Dichlorophenol | 10.234 | 162 | 1331749 | 64.829 | ng | 98 |
| 30) 1,2,4-Trichlorobenzene | 10.457 | 180 | 1609924 | 61.128 | ng | 98 |
| 31) Naphthalene | 10.646 | 128 | 4829277 | 60.059 | ng | 99 |
| 32) Benzoic acid | 9.904 | 122 | 775935 | 61.192 | ng | 99 |
| 33) 4-Chloroaniline | 10.746 | 127 | 1798121 | 59.532 | ng | 99 |
| 34) Hexachlorobutadiene | 10.940 | 225 | 1011679 | 61.008 | ng | 99 |
| 35) Caprolactam | 11.510 | 113 | 522232 | 61.084 | ng | 98 |
| 36) 4-Chloro-3-methylphenol | 11.863 | 107 | 1746318 | 61.283 | ng | 98 |
| 37) 2-Methylnaphthalene | 12.257 | 142 | 3238044 | 59.003 | ng | 99 |
| 38) 1-Methylnaphthalene | 12.475 | 142 | 3163697 | 58.451 | ng | 99 |
| 40) 1,2,4,5-Tetrachloroben... | 12.634 | 216 | 1762492 | 63.107 | ng | 100 |
| 41) Hexachlorocyclopentadiene | 12.622 | 237 | 610821 | 69.006 | ng | 99 |
| 43) 2,4,6-Trichlorophenol | 12.857 | 196 | 1050099 | 67.687 | ng | 99 |

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081324\
 Data File : BP021487.D
 Acq On : 13 Aug 2024 16:53
 Operator : RC/JU
 Sample : SSTDICC060
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC060

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024
 Supervised By :mohammad ahmed 08/14/2024

Quant Time: Aug 13 23:48:28 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 13 23:44:24 2024
 Response via : Initial Calibration

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|-------------------------------|--------|------|----------|--------|-------|----------|
| 44) 2,4,5-Trichlorophenol | 12.934 | 196 | 1165351 | 67.609 | ng | 98 |
| 46) 1,1'-Biphenyl | 13.275 | 154 | 4097218 | 60.754 | ng | 100 |
| 47) 2-Chloronaphthalene | 13.316 | 162 | 3228869 | 61.498 | ng | 99 |
| 48) 2-Nitroaniline | 13.504 | 65 | 940456 | 62.204 | ng | 99 |
| 49) Acenaphthylene | 14.175 | 152 | 4669645 | 60.501 | ng | 100 |
| 50) Dimethylphthalate | 13.904 | 163 | 3832687 | 60.041 | ng | 100 |
| 51) 2,6-Dinitrotoluene | 14.016 | 165 | 805457 | 61.336 | ng | 99 |
| 52) Acenaphthene | 14.522 | 154 | 3066208 | 60.546 | ng | 98 |
| 53) 3-Nitroaniline | 14.345 | 138 | 797905 | 61.781 | ng | 99 |
| 54) 2,4-Dinitrophenol | 14.545 | 184 | 242951 | 61.764 | ng | 95 |
| 55) Dibenzofuran | 14.857 | 168 | 4594089 | 58.970 | ng | 99 |
| 56) 4-Nitrophenol | 14.639 | 139 | 644221 | 62.066 | ng | 93 |
| 57) 2,4-Dinitrotoluene | 14.816 | 165 | 1034061 | 61.548 | ng | # 94 |
| 58) Fluorene | 15.522 | 166 | 3759913 | 58.748 | ng | 100 |
| 59) 2,3,4,6-Tetrachlorophenol | 15.086 | 232 | 973647 | 66.550 | ng | 99 |
| 60) Diethylphthalate | 15.292 | 149 | 3911032 | 59.713 | ng | 100 |
| 61) 4-Chlorophenyl-phenyle... | 15.528 | 204 | 1945714 | 58.814 | ng | 97 |
| 62) 4-Nitroaniline | 15.534 | 138 | 824354 | 62.407 | ng | 98 |
| 63) Azobenzene | 15.822 | 77 | 4711824 | 58.524 | ng | 98 |
| 65) 4,6-Dinitro-2-methylph... | 15.586 | 198 | 410096 | 61.846 | ng | 98 |
| 66) n-Nitrosodiphenylamine | 15.734 | 169 | 3192615 | 59.821 | ng | 100 |
| 67) 4-Bromophenyl-phenylether | 16.433 | 248 | 1286048 | 61.016 | ng | 97 |
| 68) Hexachlorobenzene | 16.551 | 284 | 1505847 | 60.456 | ng | 100 |
| 69) Atrazine | 16.710 | 200 | 1088343 | 59.397 | ng | 99 |
| 70) Pentachlorophenol | 16.892 | 266 | 899879 | 66.192 | ng | 99 |
| 71) Phenanthrene | 17.292 | 178 | 5685784 | 59.366 | ng | 100 |
| 72) Anthracene | 17.392 | 178 | 5636907 | 60.013 | ng | 100 |
| 73) Carbazole | 17.675 | 167 | 5454823 | 60.260 | ng | 99 |
| 74) Di-n-butylphthalate | 18.269 | 149 | 6834417 | 61.477 | ng | 100 |
| 75) Fluoranthene | 19.380 | 202 | 7044903 | 59.949 | ng | 99 |
| 77) Benzidine | 19.569 | 184 | 2334836 | 54.599 | ng | 100 |
| 78) Pyrene | 19.757 | 202 | 7263907 | 58.277 | ng | 100 |
| 80) Butylbenzylphthalate | 20.710 | 149 | 2906885 | 60.438 | ng | 98 |
| 81) Benzo(a)anthracene | 21.692 | 228 | 7119658 | 59.472 | ng | 99 |
| 82) 3,3'-Dichlorobenzidine | 21.598 | 252 | 2455338 | 63.969 | ng | 100 |
| 83) Chrysene | 21.757 | 228 | 6762569 | 59.831 | ng | 99 |
| 84) Bis(2-ethylhexyl)phtha... | 21.645 | 149 | 4576797 | 64.685 | ng | 100 |
| 85) Di-n-octyl phthalate | 22.968 | 149 | 7622203 | 59.983 | ng | 99 |
| 87) Indeno(1,2,3-cd)pyrene | 28.998 | 276 | 9100343m | 64.078 | ng | |
| 88) Benzo(b)fluoranthene | 24.045 | 252 | 7812709 | 62.083 | ng | 99 |
| 89) Benzo(k)fluoranthene | 24.121 | 252 | 7561966 | 61.599 | ng | 99 |
| 90) Benzo(a)pyrene | 24.974 | 252 | 6839630 | 62.630 | ng | 99 |
| 91) Dibenzo(a,h)anthracene | 29.103 | 278 | 7520005 | 63.721 | ng | 100 |
| 92) Benzo(g,h,i)perylene | 30.203 | 276 | 7563028 | 63.940 | ng | 100 |

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081324\
 Data File : BP021487.D
 Acq On : 13 Aug 2024 16:53
 Operator : RC/JU
 Sample : SSTDICC060
 Misc :
 ALS Vial : 8 Sample Multiplier: 1


Quant Time: Aug 13 23:48:28 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 13 23:44:24 2024
 Response via : Initial Calibration

Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC060

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024
 Supervised By :mohammad ahmed 08/14/2024







#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.810 min Scan# 8
Delta R.T. -0.001 min
Lab File: BP021487.D
Acq: 13 Aug 2024 16:53

Instrument : BNA_P
ClientSampleId : SSTDICC060


Manual Integrations
APPROVED




Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#2
1,4-Dioxane
Concen: 60.569 ng
RT: 3.322 min Scan# 57
Delta R.T. 0.006 min
Lab File: BP021487.D
Acq: 13 Aug 2024 16:53

Tgt Ion: 88 Resp: 739240
Ion Ratio Lower Upper
88 100
58 61.4 48.7 73.1
43 19.1 15.4 23.2









#3
Pyridine
Concen: 61.723 ng m
RT: 3.705 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP021487.D
Acq: 13 Aug 2024 16:53

Instrument : BNA_P
ClientSampleId : SSTDICC060


Manual Integrations
APPROVED




Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#4
n-Nitrosodimethylamine
Concen: 61.099 ng
RT: 3.622 min Scan# 108
Delta R.T. 0.006 min
Lab File: BP021487.D
Acq: 13 Aug 2024 16:53

Tgt Ion: 42 Resp: 625716
Ion Ratio Lower Upper
42 100
74 192.1 154.3 231.5
44 7.0 6.2 9.4









#5
2-Fluorophenol
Concen: 124.817 ng
RT: 5.393 min Scan# 409
Delta R.T. -0.000 min
Lab File: BP021487.D
Acq: 13 Aug 2024 16:53

Instrument :
BNA_P
ClientSampleId :
SSTDICC060


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024






#6
Aniline
Concen: 59.670 ng
RT: 7.134 min Scan# 705
Delta R.T. -0.000 min
Lab File: BP021487.D
Acq: 13 Aug 2024 16:53

Tgt Ion: 93 Resp: 2239942
Ion Ratio Lower Upper
93 100
66 38.1 30.2 45.2
65 18.4 14.7 22.1



Abundance Scan 676 (6.964 min): BP021485.D\data.ms (-66)



#7

Phenol-d6

Concen: 122.053 ng

RT: 6.963 min Scan# 6

Delta R.T. -0.000 min

Lab File: BP021487.D

Acq: 13 Aug 2024 16:53

Instrument :

BNA_P

ClientSampleId :

SSTDICC060

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024

Abundance

2500000

2000000

1500000

1000000




500000

0

6.963

Time-->

6.90 7.00 7.10



#8

2-Chlorophenol

Concen: 61.917 ng

RT: 7.375 min Scan# 746

Delta R.T. -0.000 min

Lab File: BP021487.D

Acq: 13 Aug 2024 16:53

Tgt Ion:128 Resp: 1472845

Ion Ratio Lower Upper

128 100

130 32.0 11.6 51.6

64 53.5 35.3 75.3

Abundance

800000

600000

400000




200000

0

7.375

Time-->

7.30 7.40 7.50



#9

Benzaldehyde

Concen: 50.428 ng

RT: 6.946 min Scan# 6

Delta R.T. -0.000 min

Lab File: BP021487.D

Acq: 13 Aug 2024 16:53

Instrument :

BNA_P





ClientSampleId :

SSTDICC060

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#10

Phenol

Concen: 61.254 ng


RT: 6.993 min Scan# 681




Delta R.T. 0.006 min

Lab File: BP021487.D

Acq: 13 Aug 2024 16:53

| Tgt | Ion | Resp: | | |
|-----|------|---------|-------|-------|
| | 94 | 2381743 | Lower | Upper |
| | 100 | | | |
| 94 | 100 | | | |
| 65 | 25.8 | | 6.5 | 46.5 |
| 66 | 33.6 | | 14.7 | 54.7 |









#11
bis(2-Chloroethyl)ether
Concen: 58.905 ng
RT: 7.234 min Scan# 7
Delta R.T. 0.006 min
Lab File: BP021487.D
Acq: 13 Aug 2024 16:53

Instrument : BNA_P
ClientSampleId : SSTDICC060


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#12
1,3-Dichlorobenzene
Concen: 59.538 ng
RT: 7.699 min Scan# 801
Delta R.T. -0.000 min
Lab File: BP021487.D
Acq: 13 Aug 2024 16:53

Tgt Ion:146 Resp: 1682261
Ion Ratio Lower Upper
146 100
148 64.2 51.8 77.6
75 39.1 31.3 46.9





#13

1,4-Dichlorobenzene

Concen: 59.904 ng

RT: 7.846 min Scan# 8

Delta R.T. 0.006 min

Lab File: BP021487.D

Acq: 13 Aug 2024 16:53

Instrument :

BNA_P



ClientSampleId :

SSTDICC060

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#14

1,2-Dichlorobenzene

Concen: 59.070 ng


RT: 8.157 min Scan# 879




Delta R.T. -0.000 min

Lab File: BP021487.D

Acq: 13 Aug 2024 16:53

| Tgt | Ion:146 | Resp: | 1624118 |
|-----|---------|-------|---------|
| Ion | Ratio | Lower | Upper |
| 146 | 100 | | |
| 148 | 64.8 | 51.5 | 77.3 |
| 111 | 51.6 | 41.1 | 61.7 |





#15

Benzyl Alcohol

Concen: 61.166 ng

RT: 8.034 min Scan# 8

Delta R.T. -0.000 min

Lab File: BP021487.D

Acq: 13 Aug 2024 16:53

Instrument :

BNA_P




ClientSampleId :

SSTDICC060

Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#16

2,2'-oxybis(1-Chloropropane)

Concen: 57.982 ng

RT: 8.334 min Scan# 909

Delta R.T. -0.000 min

Lab File: BP021487.D

Acq: 13 Aug 2024 16:53



Tgt Ion: 45 Resp: 1756837

Ion Ratio Lower Upper

45 100

77 21.7 1.9 41.9


79 16.6 0.0 36.6




#17

Abundance

Time--> 8.00 8.10






#17
2-Methylphenol
Concen: 60.098 ng
RT: 8.234 min Scan# 84
Delta R.T. -0.000 min
Lab File: BP021487.D
Acq: 13 Aug 2024 16:53

Instrument : BNA_P
ClientSampleId : SSTDICC060


Manual Integrations
APPROVED







Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#18
Hexachloroethane
Concen: 61.570 ng
RT: 8.893 min Scan# 1004
Delta R.T. -0.000 min
Lab File: BP021487.D
Acq: 13 Aug 2024 16:53

Tgt Ion:117 Resp: 704431
Ion Ratio Lower Upper
117 100
119 93.9 73.9 110.9
201 93.8 74.9 112.3






#19

n-Nitroso-di-n-propylamine
Concen: 56.194 ng
RT: 8.610 min Scan# 9
Delta R.T. 0.006 min
Lab File: BP021487.D
Acq: 13 Aug 2024 16:53

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Manual Integrations APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024




#20

3+4-Methylphenols
Concen: 61.202 ng
RT: 8.563 min Scan# 948
Delta R.T. 0.006 min
Lab File: BP021487.D
Acq: 13 Aug 2024 16:53

Tgt Ion:107 Resp: 2028575
Ion Ratio Lower Upper
107 100
108 88.4 67.6 107.6
77 34.6 15.0 55.0
79 29.6 8.9 48.9





#21

Naphthalene-d8

Concen: 20.000 ng

RT: 10.599 min Scan# 1

Delta R.T. -0.001 min

Lab File: BP021487.D

Acq: 13 Aug 2024 16:53

Instrument :

BNA_P



ClientSampleId :

SSTDICC060

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#22

Acetophenone

Concen: 59.500 ng


RT: 8.616 min Scan# 957


Delta R.T. -0.000 min

Lab File: BP021487.D

Acq: 13 Aug 2024 16:53

| Tgt | Ion:105 | Resp: | 2813557 |
|-----|---------|-------|---------|
| Ion | Ratio | Lower | Upper |
| 105 | 100 | | |
| 71 | 8.0 | 4.6 | 7.0# |
| 51 | 19.9 | 16.3 | 24.5 |
| 120 | 20.2 | 16.2 | 24.2 |







#23
Nitrobenzene-d5
Concen: 134.966 ng
RT: 8.957 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP021487.D
Acq: 13 Aug 2024 16:53

Instrument : BNA_P
ClientSampleId : SSTDICC060


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024




#24
Nitrobenzene
Concen: 65.650 ng
RT: 8.999 min Scan# 1022
Delta R.T. 0.006 min
Lab File: BP021487.D
Acq: 13 Aug 2024 16:53

Tgt Ion: 77 Resp: 2126322
Ion Ratio Lower Upper
77 100
123 38.7 29.5 44.3
65 12.6 9.9 14.9





Abundance Scan 1111 (9.522 min): BP021487.D\data.ms



#25

Isophorone

Concen: 58.086 ng

RT: 9.522 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP021487.D

Acq: 13 Aug 2024 16:53

Instrument :

BNA_P



ClientSampleId :

SSTDICC060


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



Abundance Scan 1142 (9.704 min): BP021487.D\data.ms



#26

2-Nitrophenol

Concen: 60.117 ng

RT: 9.704 min Scan# 1142

Delta R.T. 0.005 min

Lab File: BP021487.D

Acq: 13 Aug 2024 16:53


Tgt Ion:139 Resp: 554873







Ion Ratio Lower Upper

139 100

109 34.7 28.9 43.3

65 61.6 49.0 73.6





#27

2,4-Dimethylphenol

Concen: 61.792 ng

RT: 9.763 min Scan# 1

Delta R.T. 0.006 min

Lab File: BP021487.D

Acq: 13 Aug 2024 16:53

Instrument :

BNA_P


ClientSampleId :

SSTDICC060

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#28

bis(2-Chloroethoxy)methane

Concen: 59.443 ng

RT: 10.004 min Scan# 1193

Delta R.T. -0.000 min

Lab File: BP021487.D

Acq: 13 Aug 2024 16:53


Tgt Ion: 93 Resp: 2556281

Ion Ratio Lower Upper

93 100

95 32.6 26.2 39.4

123 9.4 7.6 11.4



#29

2,4-Dichlorophenol

Concen: 64.829 ng

RT: 10.234 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP021487.D




Acq: 13 Aug 2024 16:53

Instrument :

BNA_P

ClientSampleId :

SSTDICC060



Tgt Ion:162 Resp: 1331749

Ion Ratio Lower Upper

162 100





164 64.4 45.9 85.9

98 47.7 26.3 66.3

Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#30
1,2,4-Trichlorobenzene
Concen: 61.128 ng
RT: 10.457 min Scan# 1270
Delta R.T. -0.000 min
Lab File: BP021487.D
Acq: 13 Aug 2024 16:53


Tgt Ion:180 Resp: 1609924







Ion Ratio Lower Upper

180 100

182 95.5 78.3 117.5

145 29.5 24.0 36.0





#31

Naphthalene

Concen: 60.059 ng

RT: 10.646 min Scan# 1

Delta R.T. -0.000 min

Lab File: BP021487.D

Acq: 13 Aug 2024 16:53

Instrument :

BNA_P


ClientSampleId :

SSTDICC060

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#32

Benzoic acid

Concen: 61.192 ng

RT: 9.904 min Scan# 1176

Delta R.T. 0.029 min

Lab File: BP021487.D

Acq: 13 Aug 2024 16:53


Tgt Ion:122 Resp: 775935


Ion Ratio Lower Upper

122 100

105 144.5 124.6 164.6

77 114.7 96.2 136.2







#33
4-Chloroaniline
Concen: 59.532 ng
RT: 10.746 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP021487.D
Acq: 13 Aug 2024 16:53

Instrument : BNA_P
ClientSampleId : SSTDICC060


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#34
Hexachlorobutadiene
Concen: 61.008 ng
RT: 10.940 min Scan# 1352
Delta R.T. -0.000 min
Lab File: BP021487.D
Acq: 13 Aug 2024 16:53

Tgt Ion:225 Resp: 1011679
Ion Ratio Lower Upper
225 100
223 62.3 49.6 74.4
227 63.7 50.1 75.1





#35

Caprolactam

Concen: 61.084 ng

RT: 11.510 min Scan# 1

Delta R.T. 0.012 min

Lab File: BP021487.D

Acq: 13 Aug 2024 16:53

Instrument :

BNA_P



ClientSampleId :

SSTDICC060

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#36

4-Chloro-3-methylphenol

Concen: 61.283 ng


RT: 11.863 min Scan# 1509




Delta R.T. -0.000 min

Lab File: BP021487.D

Acq: 13 Aug 2024 16:53

| Tgt | Ion:107 | Resp: | 1746318 |
|-----|---------|-------|---------|
| Ion | Ratio | Lower | Upper |
| 107 | 100 | | |
| 144 | 21.1 | 16.6 | 24.8 |
| 142 | 65.4 | 51.0 | 76.4 |





#37

2-Methylnaphthalene

Concen: 59.003 ng

RT: 12.257 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP021487.D

Acq: 13 Aug 2024 16:53





Instrument : BNA_P

ClientSampleId : SSTDICC060

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#38

1-Methylnaphthalene

Concen: 58.451 ng

RT: 12.475 min Scan# 1613

Delta R.T. -0.006 min

Lab File: BP021487.D

Acq: 13 Aug 2024 16:53


Tgt Ion:142 Resp: 3163697


Ion Ratio Lower Upper

142 100

141 90.7 73.5 110.3

116 4.4 3.4 5.2





#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 14.457 min Scan# 1

Delta R.T. -0.000 min

Lab File: BP021487.D

Acq: 13 Aug 2024 16:53

Instrument :

BNA_P



ClientSampleId :

SSTDICC060

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#40

1,2,4,5-Tetrachlorobenzene

Concen: 63.107 ng


RT: 12.634 min Scan# 1640


Delta R.T. 0.006 min

Lab File: BP021487.D

Acq: 13 Aug 2024 16:53

| Tgt | Ion:216 | Resp: 1762492 | |
|-----|---------|---------------|-------|
| Ion | Ratio | Lower | Upper |
| 216 | 100 | | |
| 214 | 78.4 | 62.5 | 93.7 |
| 179 | 20.4 | 16.4 | 24.6 |
| 108 | 20.5 | 16.7 | 25.1 |







#41
Hexachlorocyclopentadiene
Concen: 69.006 ng
RT: 12.622 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP021487.D
Acq: 13 Aug 2024 16:53

Instrument : BNA_P
ClientSampleId : SSTDICC060


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#42
2,4,6-Tribromophenol
Concen: 132.067 ng
RT: 15.969 min Scan# 2207
Delta R.T. -0.000 min
Lab File: BP021487.D
Acq: 13 Aug 2024 16:53

Tgt Ion:330 Resp: 1401053
Ion Ratio Lower Upper
330 100
332 96.5 77.4 116.0
141 29.9 24.5 36.7







#43
2,4,6-Trichlorophenol
Concen: 67.687 ng
RT: 12.857 min Scan# 1
Delta R.T. -0.012 min
Lab File: BP021487.D
Acq: 13 Aug 2024 16:53

Instrument : BNA_P
ClientSampleId : SSTDICC060


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024




#44
2,4,5-Trichlorophenol
Concen: 67.609 ng
RT: 12.934 min Scan# 1691
Delta R.T. -0.000 min
Lab File: BP021487.D
Acq: 13 Aug 2024 16:53

Tgt Ion:196 Resp: 1165351
Ion Ratio Lower Upper
196 100
198 96.1 78.9 118.3
97 60.0 49.4 74.0
132 26.3 21.1 31.7





Abundance Scan 1714 (13.069 min): BP021487.D\data.ms



#45

2-Fluorobiphenyl

Concen: 119.833 ng

RT: 13.069 min Scan# 1

Delta R.T. 0.006 min

Lab File: BP021487.D

Acq: 13 Aug 2024 16:53

Instrument :

BNA_P

ClientSampleId :

SSTDICC060

Tgt Ion:172 Resp: 748443

Ion Ratio Lower Upper

172 100

171 35.6 28.2 42.4


170 23.5 18.4 27.6

Manual Integrations


APPROVED

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



Abundance Scan 1749 (13.275 min): BP021487.D\data.ms



#46

1,1'-Biphenyl

Concen: 60.754 ng

RT: 13.275 min Scan# 1749

Delta R.T. -0.006 min

Lab File: BP021487.D

Acq: 13 Aug 2024 16:53

Tgt Ion:154 Resp: 4097218


Ion Ratio Lower Upper




154 100

153 40.6 20.7 60.7

76 17.7 0.0 37.7

Abundance Scan 1749 (13.275 min): BP021487.D\data.ms (-)





#47

2-Chloronaphthalene

Concen: 61.498 ng

RT: 13.316 min Scan# 1

Delta R.T. -0.000 min

Lab File: BP021487.D

Acq: 13 Aug 2024 16:53

Instrument :

BNA_P





ClientSampleId :

SSTDICC060

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#48

2-Nitroaniline

Concen: 62.204 ng

RT: 13.504 min Scan# 1788

Delta R.T. 0.000 min

Lab File: BP021487.D

Acq: 13 Aug 2024 16:53


Tgt Ion: 65 Resp: 940456

Ion Ratio Lower Upper

65 100

92 72.4 58.1 87.1

138 84.1 65.5 98.3



#49

Acenaphthylene

Concen: 60.501 ng

RT: 14.175 min Scan# 1

Delta R.T. 0.006 min

Lab File: BP021487.D




Acq: 13 Aug 2024 16:53

Instrument :

BNA_P

ClientSampleId :

SSTDICC060



Tgt Ion:152 Resp: 466964

Ion Ratio Lower Upper

152 100





151 20.1 15.9 23.9

153 13.1 10.6 16.0

Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#50

Dimethylphthalate

Concen: 60.041 ng

RT: 13.904 min Scan# 1856

Delta R.T. -0.006 min

Lab File: BP021487.D

Acq: 13 Aug 2024 16:53


Tgt Ion:163 Resp: 3832687


Ion Ratio Lower Upper

163 100

194 4.3 3.3 4.9

164 10.2 8.3 12.5







#51
2,6-Dinitrotoluene
Concen: 61.336 ng
RT: 14.016 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP021487.D
Acq: 13 Aug 2024 16:53

Instrument : BNA_P
ClientSampleId : SSTDICC060


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#52
Acenaphthene
Concen: 60.546 ng
RT: 14.522 min Scan# 1961
Delta R.T. -0.000 min
Lab File: BP021487.D
Acq: 13 Aug 2024 16:53

Tgt Ion:154 Resp: 3066208
Ion Ratio Lower Upper
154 100
153 111.6 91.1 136.7
152 53.1 43.0 64.4







#53
3-Nitroaniline
Concen: 61.781 ng
RT: 14.345 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP021487.D
Acq: 13 Aug 2024 16:53

Instrument : BNA_P
ClientSampleId : SSTDICC060


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#54
2,4-Dinitrophenol
Concen: 61.764 ng
RT: 14.545 min Scan# 1965
Delta R.T. -0.006 min
Lab File: BP021487.D
Acq: 13 Aug 2024 16:53

Tgt Ion:184 Resp: 242951
Ion Ratio Lower Upper
184 100
63 77.9 64.9 97.3
154 70.2 52.1 78.1







#55
Dibenzofuran
Concen: 58.970 ng
RT: 14.857 min Scan# 2
Delta R.T. 0.000 min
Lab File: BP021487.D
Acq: 13 Aug 2024 16:53

Instrument : BNA_P
ClientSampleId : SSTDICC060


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#56
4-Nitrophenol
Concen: 62.066 ng
RT: 14.639 min Scan# 1981
Delta R.T. -0.012 min
Lab File: BP021487.D
Acq: 13 Aug 2024 16:53

Tgt Ion:139 Resp: 644221
Ion Ratio Lower Upper
139 100
109 83.8 70.2 110.2
65 114.0 101.8 141.8







#57
2,4-Dinitrotoluene
Concen: 61.548 ng
RT: 14.816 min Scan# 2130
Delta R.T. 0.006 min
Lab File: BP021487.D
Acq: 13 Aug 2024 16:53

Instrument : BNA_P
ClientSampleId : SSTDICC060


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#58
Fluorene
Concen: 58.748 ng
RT: 15.522 min Scan# 2131
Delta R.T. 0.006 min
Lab File: BP021487.D
Acq: 13 Aug 2024 16:53

Tgt Ion:166 Resp: 3759913
Ion Ratio Lower Upper
166 100
165 97.4 78.1 117.1
167 13.6 11.0 16.4






#59
2,3,4,6-Tetrachlorophenol
Concen: 66.550 ng
RT: 15.086 min Scan# 2
Delta R.T. -0.000 min
Lab File: BP021487.D
Acq: 13 Aug 2024 16:53

Instrument : BNA_P
ClientSampleId : SSTDICC060


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024




#60
Diethylphthalate
Concen: 59.713 ng
RT: 15.292 min Scan# 2092
Delta R.T. -0.012 min
Lab File: BP021487.D
Acq: 13 Aug 2024 16:53

Tgt Ion:149 Resp: 3911032
Ion Ratio Lower Upper
149 100
177 22.6 17.8 26.8
150 12.6 10.1 15.1





Abundance Scan 2132 (15.528 min): BP021487.D\data.ms



#61

4-Chlorophenyl-phenylether

Concen: 58.814 ng

RT: 15.528 min Scan# 2132

Delta R.T. 0.006 min

Lab File: BP021487.D

Acq: 13 Aug 2024 16:53

Instrument :

BNA_P



ClientSampleId :

SSTDICC060


**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



Abundance Scan 2133 (15.534 min): BP021487.D\data.ms



#62

4-Nitroaniline

Concen: 62.407 ng

RT: 15.534 min Scan# 2133

Delta R.T. 0.006 min

Lab File: BP021487.D

Acq: 13 Aug 2024 16:53


Tgt Ion:138 Resp: 824354




Ion Ratio Lower Upper

138 100

92 65.4 46.4 86.4

108 90.0 73.2 113.2





#63

Azobenzene

Concen: 58.524 ng

RT: 15.822 min Scan# 2

Delta R.T. -0.000 min

Lab File: BP021487.D

Acq: 13 Aug 2024 16:53

Instrument :

BNA_P




ClientSampleId :

SSTDICC060

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 17.257 min Scan# 2426

Delta R.T. 0.006 min

Lab File: BP021487.D

Acq: 13 Aug 2024 16:53


Tgt Ion:188 Resp: 1991489


Ion Ratio Lower Upper

188 100

94 12.1 10.2 15.4

80 13.4 11.6 17.4







#65
4,6-Dinitro-2-methylphenol
Concen: 61.846 ng
RT: 15.586 min Scan# 2142
Delta R.T. -0.001 min
Lab File: BP021487.D
Acq: 13 Aug 2024 16:53

Instrument : BNA_P
ClientSampleId : SSTDICC060


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#66
n-Nitrosodiphenylamine
Concen: 59.821 ng
RT: 15.734 min Scan# 2167
Delta R.T. -0.006 min
Lab File: BP021487.D
Acq: 13 Aug 2024 16:53

Tgt Ion:169 Resp: 3192615
Ion Ratio Lower Upper
169 100
168 66.5 53.4 80.2
167 36.0 28.7 43.1







#67
4-Bromophenyl-phenylether
Concen: 61.016 ng
RT: 16.433 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP021487.D
Acq: 13 Aug 2024 16:53

Instrument : BNA_P
ClientSampleId : SSTDICC060


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#68
Hexachlorobenzene
Concen: 60.456 ng
RT: 16.551 min Scan# 2306
Delta R.T. -0.000 min
Lab File: BP021487.D
Acq: 13 Aug 2024 16:53

Tgt Ion:284 Resp: 1505847
Ion Ratio Lower Upper
284 100
142 29.3 23.4 35.2
249 30.1 24.1 36.1







#69
Atrazine
Concen: 59.397 ng
RT: 16.710 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP021487.D
Acq: 13 Aug 2024 16:53

Instrument : BNA_P
ClientSampleId : SSTDICC060


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024





#70
Pentachlorophenol
Concen: 66.192 ng
RT: 16.892 min Scan# 2364
Delta R.T. 0.000 min
Lab File: BP021487.D
Acq: 13 Aug 2024 16:53

Tgt Ion:266 Resp: 899879
Ion Ratio Lower Upper
266 100
268 63.9 51.8 77.6
264 63.1 50.5 75.7






Abundance Scan 2432 (17.292 min): BP021487.D\data.ms



Abundance Scan 2449 (17.392 min): BP021487.D\data.ms



#71

Phenanthrene

Concen: 59.366 ng

RT: 17.292 min Scan# 2449

Delta R.T. -0.006 min

Lab File: BP021487.D

Acq: 13 Aug 2024 16:53

Instrument :

BNA_P


ClientSampleId :

SSTDICC060

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#72

Anthracene

Concen: 60.013 ng

RT: 17.392 min Scan# 2449

Delta R.T. -0.006 min

Lab File: BP021487.D

Acq: 13 Aug 2024 16:53


Tgt Ion:178 Resp: 5636907


Ion Ratio Lower Upper

178 100

176 18.4 14.9 22.3

179 15.6 12.5 18.7







#73
Carbazole
Concen: 60.260 ng
RT: 17.675 min Scan# 2
Delta R.T. 0.006 min
Lab File: BP021487.D
Acq: 13 Aug 2024 16:53

Instrument : BNA_P
ClientSampleId : SSTDICC060


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#74
Di-n-butylphthalate
Concen: 61.477 ng
RT: 18.269 min Scan# 2598
Delta R.T. -0.006 min
Lab File: BP021487.D
Acq: 13 Aug 2024 16:53

Tgt Ion:149 Resp: 6834417
Ion Ratio Lower Upper
149 100
150 9.4 7.5 11.3
104 6.4 4.9 7.3







#75
Fluoranthene
Concen: 59.949 ng
RT: 19.380 min Scan# 2
Delta R.T. -0.000 min
Lab File: BP021487.D
Acq: 13 Aug 2024 16:53

Instrument : BNA_P
ClientSampleId : SSTDICC060


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#76
Chrysene-d₁₂
Concen: 20.000 ng
RT: 21.710 min Scan# 3183
Delta R.T. 0.012 min
Lab File: BP021487.D
Acq: 13 Aug 2024 16:53

Tgt Ion:240 Resp: 1907005
Ion Ratio Lower Upper
240 100
120 11.3 9.9 14.9
236 26.0 20.6 30.8







#77
Benzidine
Concen: 54.599 ng
RT: 19.569 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP021487.D
Acq: 13 Aug 2024 16:53

Instrument : BNA_P
ClientSampleId : SSTDICC060


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#78
Pyrene
Concen: 58.277 ng
RT: 19.757 min Scan# 2851
Delta R.T. 0.006 min
Lab File: BP021487.D
Acq: 13 Aug 2024 16:53

Tgt Ion:202 Resp: 7263907
Ion Ratio Lower Upper
202 100
200 21.1 16.7 25.1
203 17.7 14.1 21.1





#79

Terphenyl-d14

Concen: 115.877 ng

RT: 19.986 min Scan# 2

Delta R.T. 0.011 min

Lab File: BP021487.D

Acq: 13 Aug 2024 16:53

Instrument :

BNA_P



ClientSampleId :

SSTDICC060

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#80

Butylbenzylphthalate

Concen: 60.438 ng

RT: 20.710 min Scan# 3013

Delta R.T. -0.000 min

Lab File: BP021487.D

Acq: 13 Aug 2024 16:53


Tgt Ion:149 Resp: 2906885


Ion Ratio Lower Upper

149 100

91 90.1 73.3 109.9

206 22.3 17.1 25.7





#81

Benzo(a)anthracene

Concen: 59.472 ng

RT: 21.692 min Scan# 3180

Delta R.T. 0.012 min

Lab File: BP021487.D

Acq: 13 Aug 2024 16:53

Instrument :

BNA_P



ClientSampleId :

SSTDICC060

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#82

3,3'-Dichlorobenzidine

Concen: 63.969 ng

RT: 21.598 min Scan# 3164

Delta R.T. -0.012 min

Lab File: BP021487.D

Acq: 13 Aug 2024 16:53


Tgt Ion:252 Resp: 2455338


Ion Ratio Lower Upper

252 100

254 64.7 51.7 77.5

126 10.8 8.7 13.1





#83

Chrysene

Concen: 59.831 ng

RT: 21.757 min Scan# 3

Delta R.T. 0.006 min

Lab File: BP021487.D

Acq: 13 Aug 2024 16:53

Instrument :

BNA_P



ClientSampleId :

SSTDICC060

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#84

Bis(2-ethylhexyl)phthalate

Concen: 64.685 ng

RT: 21.645 min Scan# 3172

Delta R.T. -0.000 min

Lab File: BP021487.D

Acq: 13 Aug 2024 16:53


Tgt Ion:149 Resp: 4576797



Ion Ratio Lower Upper

149 100

167 27.7 22.3 33.5

279 4.6 3.8 5.6





#85

Di-n-octyl phthalate

Concen: 59.983 ng

RT: 22.968 min Scan# 3

Delta R.T. 0.005 min

Lab File: BP021487.D

Acq: 13 Aug 2024 16:53

Instrument :

BNA_P


ClientSampleId :

SSTDICC060

Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#86

Perylene-d₁₂

Concen: 20.000 ng

RT: 25.121 min Scan# 3763

Delta R.T. 0.018 min

Lab File: BP021487.D

Acq: 13 Aug 2024 16:53


Tgt Ion:264 Resp: 2191159


Ion Ratio Lower Upper

264 100

260 23.5 18.6 28.0

265 21.1 17.1 25.7







#87
Indeno(1,2,3-cd)pyrene
Concen: 64.078 ng/m
RT: 28.998 min Scan# 4
Instrument : BNA_P
Delta R.T. 0.006 min
Lab File: BP021487.D
Acq: 13 Aug 2024 16:53
ClientSampleId : SSTDICC060


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#88
Benzo(b)fluoranthene
Concen: 62.083 ng
RT: 24.045 min Scan# 3580
Delta R.T. 0.018 min
Lab File: BP021487.D
Acq: 13 Aug 2024 16:53

Tgt Ion:252 Resp: 7812709
Ion Ratio Lower Upper
252 100
253 22.0 17.6 26.4
125 9.8 8.6 13.0



#89

Benzo(k)fluoranthene

Concen: 61.599 ng

RT: 24.121 min Scan# 3

Delta R.T. 0.012 min

Lab File: BP021487.D




Acq: 13 Aug 2024 16:53

Instrument :

BNA_P

ClientSampleId :

SSTDICC060



Tgt Ion:252 Resp: 7561960

Ion Ratio Lower Upper

252 100

253 22.3 17.5 26.3





125 10.0 7.8 11.8

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#90

Benzo(a)pyrene

Concen: 62.630 ng

RT: 24.974 min Scan# 3738

Delta R.T. 0.023 min

Lab File: BP021487.D

Acq: 13 Aug 2024 16:53


Tgt Ion:252 Resp: 6839630



Ion Ratio Lower Upper

252 100

253 21.6 17.7 26.5

125 11.5 8.7 13.1









#91
Dibenzo(a,h)anthracene
Concen: 63.721 ng
RT: 29.103 min Scan# 4
Delta R.T. 0.018 min
Lab File: BP021487.D
Acq: 13 Aug 2024 16:53

Instrument : BNA_P
ClientSampleId : SSTDICC060


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#92
Benzo(g,h,i)perylene
Concen: 63.940 ng
RT: 30.203 min Scan# 4627
Delta R.T. 0.029 min
Lab File: BP021487.D
Acq: 13 Aug 2024 16:53

Tgt Ion:276 Resp: 7563028
Ion Ratio Lower Upper
276 100
277 24.0 19.0 28.6
138 19.8 16.0 24.0



Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081324\
 Data File : BP021488.D
 Acq On : 13 Aug 2024 17:36
 Operator : RC/JU
 Sample : SSTDICC080
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC080

Quant Time: Aug 13 23:48:09 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 13 23:44:24 2024
 Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024
 Supervised By :mohammad ahmed 08/14/2024

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|------------------------------------|--------|------|----------|---------|-------|----------|
| Internal Standards | | | | | | |
| 1) 1,4-Dichlorobenzene-d4 | 7.805 | 152 | 405636 | 20.000 | ng | 0.00 |
| 21) Naphthalene-d8 | 10.599 | 136 | 1590621 | 20.000 | ng | 0.00 |
| 39) Acenaphthene-d10 | 14.451 | 164 | 991938 | 20.000 | ng | 0.00 |
| 64) Phenanthrene-d10 | 17.239 | 188 | 2007991 | 20.000 | ng | -0.01 |
| 76) Chrysene-d12 | 21.698 | 240 | 1834396 | 20.000 | ng | 0.00 |
| 86) Perylene-d12 | 25.110 | 264 | 2148798 | 20.000 | ng | 0.00 |
| System Monitoring Compounds | | | | | | |
| 5) 2-Fluorophenol | 5.393 | 112 | 4326213 | 158.364 | ng | 0.00 |
| 7) Phenol-d6 | 6.969 | 99 | 6182538 | 155.142 | ng | 0.00 |
| 23) Nitrobenzene-d5 | 8.957 | 82 | 5476064 | 177.696 | ng | 0.00 |
| 42) 2,4,6-Tribromophenol | 15.951 | 330 | 1902622 | 172.341 | ng | -0.02 |
| 45) 2-Fluorobiphenyl | 13.063 | 172 | 9945684 | 153.020 | ng | 0.00 |
| 79) Terphenyl-d14 | 19.969 | 244 | 13720759 | 145.311 | ng | 0.00 |
| Target Compounds | | | | | | |
| | | | | Qvalue | | |
| 2) 1,4-Dioxane | 3.317 | 88 | 981936 | 75.807 | ng | 99 |
| 3) Pyridine | 3.705 | 79 | 2799665m | 77.425 | ng | |
| 4) n-Nitrosodimethylamine | 3.617 | 42 | 842879 | 77.551 | ng | 100 |
| 6) Aniline | 7.140 | 93 | 2987211 | 74.981 | ng | 99 |
| 8) 2-Chlorophenol | 7.375 | 128 | 1999068 | 79.186 | ng | 100 |
| 10) Phenol | 6.993 | 94 | 3198266 | 77.502 | ng | 98 |
| 11) bis(2-Chloroethyl)ether | 7.234 | 93 | 2653675 | 75.997 | ng | 98 |
| 12) 1,3-Dichlorobenzene | 7.699 | 146 | 2278631 | 75.987 | ng | 99 |
| 13) 1,4-Dichlorobenzene | 7.840 | 146 | 2292726 | 75.717 | ng | 99 |
| 14) 1,2-Dichlorobenzene | 8.158 | 146 | 2200854 | 75.423 | ng | 100 |
| 15) Benzyl Alcohol | 8.040 | 79 | 2214742 | 77.457 | ng | 99 |
| 16) 2,2'-oxybis(1-Chloropr... | 8.334 | 45 | 2391178 | 74.359 | ng | 99 |
| 17) 2-Methylphenol | 8.234 | 107 | 1995331 | 76.484 | ng | 100 |
| 18) Hexachloroethane | 8.887 | 117 | 957451 | 78.852 | ng | 98 |
| 19) n-Nitroso-di-n-propyla... | 8.610 | 70 | 1923242 | 69.865 | ng | 99 |
| 20) 3+4-Methylphenols | 8.563 | 107 | 2709790 | 77.033 | ng | 99 |
| 22) Acetophenone | 8.622 | 105 | 3753281 | 76.808 | ng | 99 |
| 24) Nitrobenzene | 8.999 | 77 | 2888411 | 86.298 | ng | 97 |
| 25) Isophorone | 9.522 | 82 | 5496234 | 75.494 | ng | 100 |
| 27) 2,4-Dimethylphenol | 9.763 | 122 | 1443721 | 80.354 | ng | 99 |
| 28) bis(2-Chloroethoxy)met... | 10.004 | 93 | 3455210 | 77.749 | ng | 99 |
| 29) 2,4-Dichlorophenol | 10.240 | 162 | 1832311 | 86.313 | ng | 100 |
| 30) 1,2,4-Trichlorobenzene | 10.457 | 180 | 2181467 | 80.153 | ng | 99 |
| 31) Naphthalene | 10.646 | 128 | 6530682 | 78.593 | ng | 99 |
| 32) Benzoic acid | 9.922 | 122 | 1191515 | 79.618 | ng | 99 |
| 33) 4-Chloroaniline | 10.746 | 127 | 2421138 | 77.568 | ng | 99 |
| 34) Hexachlorobutadiene | 10.946 | 225 | 1379696 | 80.513 | ng | 97 |
| 35) Caprolactam | 11.522 | 113 | 702638m | 79.529 | ng | |
| 36) 4-Chloro-3-methylphenol | 11.869 | 107 | 2368621 | 80.436 | ng | 98 |
| 37) 2-Methylnaphthalene | 12.257 | 142 | 4347439 | 76.659 | ng | 100 |
| 38) 1-Methylnaphthalene | 12.481 | 142 | 4237982 | 75.769 | ng | 99 |
| 40) 1,2,4,5-Tetrachloroben... | 12.628 | 216 | 2395337 | 82.416 | ng | 100 |
| 41) Hexachlorocyclopentadiene | 12.616 | 237 | 847184 | 91.971 | ng | 98 |
| 43) 2,4,6-Trichlorophenol | 12.863 | 196 | 1454749 | 90.107 | ng | 100 |
| 44) 2,4,5-Trichlorophenol | 12.928 | 196 | 1603816 | 89.412 | ng | 97 |
| 46) 1,1'-Biphenyl | 13.275 | 154 | 5498161 | 78.343 | ng | 99 |

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081324\
 Data File : BP021488.D
 Acq On : 13 Aug 2024 17:36
 Operator : RC/JU
 Sample : SSTDICC080
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC080

Quant Time: Aug 13 23:48:09 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 13 23:44:24 2024
 Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024
 Supervised By :mohammad ahmed 08/14/2024

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|-------------------------------|--------|------|-----------|--------|-------|----------|
| 47) 2-Chloronaphthalene | 13.310 | 162 | 4315309 | 78.980 | ng | 99 |
| 48) 2-Nitroaniline | 13.504 | 65 | 1301093 | 81.000 | ng | 93 |
| 49) Acenaphthylene | 14.169 | 152 | 6229370 | 77.556 | ng | 100 |
| 50) Dimethylphthalate | 13.904 | 163 | 5104657 | 76.844 | ng | 99 |
| 51) 2,6-Dinitrotoluene | 14.010 | 165 | 1121870 | 80.745 | ng | 97 |
| 52) Acenaphthene | 14.510 | 154 | 4136001 | 78.480 | ng | 99 |
| 53) 3-Nitroaniline | 14.340 | 138 | 1100154 | 80.567 | ng | # 95 |
| 54) 2,4-Dinitrophenol | 14.545 | 184 | 370353 | 79.528 | ng | 92 |
| 55) Dibenzofuran | 14.851 | 168 | 6247950 | 77.067 | ng | 99 |
| 56) 4-Nitrophenol | 14.640 | 139 | 885944 | 80.388 | ng | 97 |
| 57) 2,4-Dinitrotoluene | 14.804 | 165 | 1449693 | 81.172 | ng | 96 |
| 58) Fluorene | 15.510 | 166 | 4891299 | 73.441 | ng | 100 |
| 59) 2,3,4,6-Tetrachlorophenol | 15.075 | 232 | 1322486 | 86.863 | ng | 100 |
| 60) Diethylphthalate | 15.292 | 149 | 5198353 | 76.268 | ng | 99 |
| 61) 4-Chlorophenyl-phenyle... | 15.510 | 204 | 2556419 | 74.256 | ng | 98 |
| 62) 4-Nitroaniline | 15.522 | 138 | 1101954 | 79.329 | ng | 94 |
| 63) Azobenzene | 15.804 | 77 | 6169252 | 73.633 | ng | 97 |
| 65) 4,6-Dinitro-2-methylph... | 15.575 | 198 | 600088 | 79.442 | ng | 100 |
| 66) n-Nitrosodiphenylamine | 15.722 | 169 | 4264380 | 79.247 | ng | 99 |
| 67) 4-Bromophenyl-phenylether | 16.416 | 248 | 1731598 | 81.480 | ng | 98 |
| 68) Hexachlorobenzene | 16.534 | 284 | 1994136 | 79.402 | ng | 99 |
| 69) Atrazine | 16.698 | 200 | 1395641 | 75.542 | ng | 98 |
| 70) Pentachlorophenol | 16.881 | 266 | 1256277 | 91.649 | ng | 100 |
| 71) Phenanthrene | 17.286 | 178 | 7482235 | 77.481 | ng | 99 |
| 72) Anthracene | 17.381 | 178 | 7350512 | 77.613 | ng | 100 |
| 73) Carbazole | 17.651 | 167 | 7163944 | 78.491 | ng | 99 |
| 74) Di-n-butylphthalate | 18.257 | 149 | 9044596 | 80.689 | ng | 100 |
| 75) Fluoranthene | 19.369 | 202 | 9185836 | 77.525 | ng | 99 |
| 77) Benzidine | 19.557 | 184 | 3000610 | 82.290 | ng | 99 |
| 78) Pyrene | 19.745 | 202 | 9406096 | 78.450 | ng | 99 |
| 80) Butylbenzylphthalate | 20.704 | 149 | 3806046 | 80.964 | ng | 98 |
| 81) Benzo(a)anthracene | 21.680 | 228 | 9116699 | 79.168 | ng | 98 |
| 82) 3,3'-Dichlorobenzidine | 21.598 | 252 | 3092890 | 83.768 | ng | 100 |
| 83) Chrysene | 21.751 | 228 | 8577622 | 78.894 | ng | 98 |
| 84) Bis(2-ethylhexyl)phtha... | 21.639 | 149 | 6062099 | 89.068 | ng | 100 |
| 85) Di-n-octyl phthalate | 22.957 | 149 | 10099312 | 80.797 | ng | 99 |
| 87) Indeno(1,2,3-cd)pyrene | 29.009 | 276 | 11599453m | 83.285 | ng | |
| 88) Benzo(b)fluoranthene | 24.045 | 252 | 9994213 | 80.984 | ng | 100 |
| 89) Benzo(k)fluoranthene | 24.121 | 252 | 9544162 | 79.279 | ng | 99 |
| 90) Benzo(a)pyrene | 24.962 | 252 | 8757562 | 81.773 | ng | 100 |
| 91) Dibenzo(a,h)anthracene | 29.098 | 278 | 9530458 | 82.349 | ng | 100 |
| 92) Benzo(g,h,i)perylene | 30.209 | 276 | 9695156 | 83.581 | ng | 100 |

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081324\
 Data File : BP021488.D
 Acq On : 13 Aug 2024 17:36
 Operator : RC/JU
 Sample : SSTDICC080
 Misc :
 ALS Vial : 9 Sample Multiplier: 1


Quant Time: Aug 13 23:48:09 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 13 23:44:24 2024
 Response via : Initial Calibration

Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC080

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024
 Supervised By :mohammad ahmed 08/14/2024







#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.805 min Scan# 8
Delta R.T. -0.006 min
Lab File: BP021488.D
Acq: 13 Aug 2024 17:36

Instrument : BNA_P
ClientSampleId : SSTDICC080


Manual Integrations
APPROVED







Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#2
1,4-Dioxane
Concen: 75.807 ng
RT: 3.317 min Scan# 56
Delta R.T. -0.000 min
Lab File: BP021488.D
Acq: 13 Aug 2024 17:36

Tgt Ion: 88 Resp: 981936
Ion Ratio Lower Upper
88 100
58 61.6 48.7 73.1
43 19.6 15.4 23.2





#3

Pyridine

Concen: 77.425 ng m

RT: 3.705 min Scan# 1

Delta R.T. -0.000 min

Lab File: BP021488.D

Acq: 13 Aug 2024 17:36

Instrument :

BNA_P


ClientSampleId :

SSTDICC080

Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#4

n-Nitrosodimethylamine

Concen: 77.551 ng

RT: 3.617 min Scan# 107

Delta R.T. -0.000 min

Lab File: BP021488.D

Acq: 13 Aug 2024 17:36


Tgt Ion: 42 Resp: 842879


Ion Ratio Lower Upper

42 100

74 192.8 154.3 231.5

44 6.8 6.2 9.4







#5
2-Fluorophenol
Concen: 158.364 ng
RT: 5.393 min Scan# 409
Delta R.T. -0.000 min
Lab File: BP021488.D
Acq: 13 Aug 2024 17:36

Instrument : BNA_P
ClientSampleId : SSTDICC080


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024




#6
Aniline
Concen: 74.981 ng
RT: 7.140 min Scan# 706
Delta R.T. 0.006 min
Lab File: BP021488.D
Acq: 13 Aug 2024 17:36


Tgt Ion: 93 Resp: 2987211
Ion Ratio Lower Upper
93 100
66 37.3 30.2 45.2
65 18.4 14.7 22.1




Abundance Scan 676 (6.964 min): BP021485.D\data.ms (-66)



Abundance Scan 677 (6.969 min): BP021488.D\data.ms



Abundance Scan 677 (6.969 min): BP021488.D\data.ms (-62)



#7

Phenol-d6

Concen: 155.142 ng

RT: 6.969 min Scan# 6

Delta R.T. 0.006 min

Lab File: BP021488.D

Acq: 13 Aug 2024 17:36

Instrument :

BNA_P

ClientSampleId :

SSTDICC080

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024

Abundance

3000000

6.969

2000000

1000000

0

Time-->

6.90 7.00 7.10

#8

2-Chlorophenol

Concen: 79.186 ng

RT: 7.375 min Scan# 746

Delta R.T. -0.000 min

Lab File: BP021488.D

Acq: 13 Aug 2024 17:36

Tgt Ion:128 Resp: 1999068

Ion Ratio Lower Upper

128 100

130 31.7 11.6 51.6

64 55.2 35.3 75.3

Abundance

1000000

7.375

500000

0


Time-->

7.30 7.40 7.50

#9

BP021488.D 8270E-BP081324.M

P3657-SVOCMS Group6



#10

Phenol

Concen: 77.502 ng

RT: 6.993 min Scan# 6

Delta R.T. 0.006 min

Lab File: BP021488.D

Acq: 13 Aug 2024 17:36

Instrument :

BNA_P



ClientSampleId :

SSTDICC080

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#11

bis(2-Chloroethyl)ether

Concen: 75.997 ng

RT: 7.234 min Scan# 722


Delta R.T. 0.006 min




Lab File: BP021488.D

Acq: 13 Aug 2024 17:36

Tgt Ion: 93 Resp: 2653675

| | Ion Ratio | Lower | Upper |
|----|-----------|-------|-------|
| 93 | 100 | | |
| 63 | 69.6 | 51.0 | 91.0 |
| 95 | 32.2 | 12.9 | 52.9 |





#12

1,3-Dichlorobenzene

Concen: 75.987 ng

RT: 7.699 min Scan# 8

Delta R.T. -0.000 min

Lab File: BP021488.D

Acq: 13 Aug 2024 17:36

Instrument :

BNA_P





ClientSampleId :

SSTDICC080

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#13

1,4-Dichlorobenzene

Concen: 75.717 ng

RT: 7.840 min Scan# 825

Delta R.T. -0.000 min

Lab File: BP021488.D

Acq: 13 Aug 2024 17:36


Tgt Ion:146 Resp: 2292726




Ion Ratio Lower Upper

146 100

148 64.4 51.5 77.3

111 48.2 40.3 60.5





#14

1,2-Dichlorobenzene

Concen: 75.423 ng

RT: 8.158 min Scan# 8

Delta R.T. -0.000 min

Lab File: BP021488.D

Acq: 13 Aug 2024 17:36

Instrument :

BNA_P




ClientSampleId :

SSTDICC080

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#15

Benzyl Alcohol

Concen: 77.457 ng

RT: 8.040 min Scan# 859

Delta R.T. 0.006 min

Lab File: BP021488.D

Acq: 13 Aug 2024 17:36

Tgt Ion: 79 Resp: 2214742

Ion Ratio Lower Upper

79 100

108 73.1 57.9 86.9

77 63.3 51.8 77.8

Time--> 8.10 8.15 8.20

Abundance


8.158

Time--> 8.10 8.15 8.20

Abundance

8.040

Time--> 8.00 8.05 8.10





#16
2,2'-oxybis(1-Chloropropane)
Concen: 74.359 ng
RT: 8.334 min Scan# 909
Delta R.T. -0.000 min
Lab File: BP021488.D
Acq: 13 Aug 2024 17:36

Instrument :
BNA_P
ClientSampleId :
SSTDICC080


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#17
2-Methylphenol
Concen: 76.484 ng
RT: 8.234 min Scan# 892
Delta R.T. -0.000 min
Lab File: BP021488.D
Acq: 13 Aug 2024 17:36

Tgt Ion:107 Resp: 1995331
Ion Ratio Lower Upper
107 100
108 110.5 88.8 133.2
77 47.9 38.7 58.1
79 46.6 37.4 56.2





#18

Hexachloroethane

Concen: 78.852 ng

RT: 8.887 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP021488.D

Acq: 13 Aug 2024 17:36

Instrument :

BNA_P



ClientSampleId :

SSTDICC080

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#19

n-Nitroso-di-n-propylamine

Concen: 69.865 ng


RT: 8.610 min Scan# 956


Delta R.T. 0.006 min

Lab File: BP021488.D

Acq: 13 Aug 2024 17:36

| Tgt | Ion: | 70 | Resp: | 1923242 |
|-----|-----------|-------|-------|---------|
| | Ion Ratio | Lower | Upper | |
| | 70 | 100 | | |
| | 42 | 41.9 | 33.1 | 49.7 |
| | 101 | 9.9 | 7.4 | 11.2 |
| | 130 | 17.4 | 13.3 | 19.9 |






#20
3+4-Methylphenols
Concen: 77.033 ng
RT: 8.563 min Scan# 9
Instrument : BNA_P
Delta R.T. 0.006 min
Lab File: BP021488.D
Acq: 13 Aug 2024 17:36
ClientSampleId : SSTDICC080


Manual Integrations APPROVED




Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.599 min Scan# 1294
Delta R.T. -0.000 min
Lab File: BP021488.D
Acq: 13 Aug 2024 17:36

Tgt Ion:136 Resp: 1590621
Ion Ratio Lower Upper
136 100
137 10.9 8.8 13.2
54 8.2 6.3 9.5
68 8.5 6.8 10.2





#22

Acetophenone

Concen: 76.808 ng

RT: 8.622 min Scan# 9

Delta R.T. 0.006 min

Lab File: BP021488.D

Acq: 13 Aug 2024 17:36

Instrument :

BNA_P





ClientSampleId :

SSTDICC080

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#23

Nitrobenzene-d5

Concen: 177.696 ng

RT: 8.957 min Scan# 1015

Delta R.T. 0.006 min

Lab File: BP021488.D

Acq: 13 Aug 2024 17:36


Tgt Ion: 82 Resp: 5476064


Ion Ratio Lower Upper

82 100



128 35.4 26.5 39.7

54 38.7 30.3 45.5






Abundance Scan 1022 (8.999 min): BP021488.D\data.ms



Abundance Scan 1111 (9.522 min): BP021488.D\data.ms



#24

Nitrobenzene

Concen: 86.298 ng

RT: 8.999 min Scan# 1

Delta R.T. 0.006 min

Lab File: BP021488.D

Acq: 13 Aug 2024 17:36

Instrument :

BNA_P


ClientSampleId :

SSTDICC080

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#25

Isophorone

Concen: 75.494 ng

RT: 9.522 min Scan# 1111

Delta R.T. 0.000 min

Lab File: BP021488.D

Acq: 13 Aug 2024 17:36


Tgt Ion: 82 Resp: 5496234







Ion Ratio Lower Upper

82 100

95 6.7 5.4 8.2

138 13.3 10.6 16.0





#27

2,4-Dimethylphenol

Concen: 80.354 ng

RT: 9.763 min Scan# 1

Delta R.T. 0.006 min

Lab File: BP021488.D

Acq: 13 Aug 2024 17:36

Instrument :

BNA_P


ClientSampleId :

SSTDICC080

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#28

bis(2-Chloroethoxy)methane

Concen: 77.749 ng

RT: 10.004 min Scan# 1193

Delta R.T. -0.000 min

Lab File: BP021488.D

Acq: 13 Aug 2024 17:36


Tgt Ion: 93 Resp: 3455210


Ion Ratio Lower Upper

93 100

95 32.2 26.2 39.4

123 9.7 7.6 11.4







#29
2,4-Dichlorophenol
Concen: 86.313 ng
RT: 10.240 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP021488.D
Acq: 13 Aug 2024 17:36

Instrument : BNA_P
ClientSampleId : SSTDICC080


Manual Integrations
APPROVED




Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#30
1,2,4-Trichlorobenzene
Concen: 80.153 ng
RT: 10.457 min Scan# 1270
Delta R.T. -0.000 min
Lab File: BP021488.D
Acq: 13 Aug 2024 17:36

Tgt Ion:180 Resp: 2181467
Ion Ratio Lower Upper
180 100
182 96.4 78.3 117.5
145 29.5 24.0 36.0





#31

Naphthalene

Concen: 78.593 ng

RT: 10.646 min Scan# 1

Delta R.T. -0.000 min

Lab File: BP021488.D

Acq: 13 Aug 2024 17:36

Instrument :

BNA_P




ClientSampleId :

SSTDICC080


**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/14/2024


Supervised By :mohammad ahmed 08/14/2024



#32
Benzoic acid
Concen: 79.618 ng
RT: 9.922 min Scan# 1179
Delta R.T. 0.047 min
Lab File: BP021488.D
Acq: 13 Aug 2024 17:36



Tgt Ion:122 Resp: 1191515
Ion Ratio Lower Upper
122 100
105 143.7 124.6 164.6
77 114.4 96.2 136.2



#33

4-Chloroaniline

Concen: 77.568 ng

RT: 10.746 min Scan# 1

Delta R.T. 0.006 min

Lab File: BP021488.D




Acq: 13 Aug 2024 17:36

Instrument :

BNA_P

ClientSampleId :

SSTDICC080



Tgt Ion:127 Resp: 242113





Ion Ratio Lower Upper

| | | | |
|-----|------|------|------|
| 127 | 100 | | |
| 129 | 31.4 | 24.5 | 36.7 |
| 65 | 35.8 | 28.9 | 43.3 |
| 92 | 24.4 | 19.3 | 28.9 |

Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#34

Hexachlorobutadiene

Concen: 80.513 ng

RT: 10.946 min Scan# 1353

Delta R.T. 0.006 min


Lab File: BP021488.D


Acq: 13 Aug 2024 17:36

Tgt Ion:225 Resp: 1379696

Ion Ratio Lower Upper

| | | | |
|-----|------|------|------|
| 225 | 100 | | |
| 223 | 63.8 | 49.6 | 74.4 |
| 227 | 65.3 | 50.1 | 75.1 |





#35

Caprolactam

Concen: 79.529 ng m

RT: 11.522 min Scan# 1

Delta R.T. 0.023 min

Lab File: BP021488.D

Acq: 13 Aug 2024 17:36

Instrument :

BNA_P



ClientSampleId :

SSTDICC080

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#36

4-Chloro-3-methylphenol

Concen: 80.436 ng


RT: 11.869 min Scan# 1510


Delta R.T. 0.006 min

Lab File: BP021488.D


Acq: 13 Aug 2024 17:36

| Tgt | Ion:107 | Resp: | 2368621 |
|-----|---------|-------|---------|
| Ion | Ratio | Lower | Upper |
| 107 | 100 | | |
| 144 | 21.4 | 16.6 | 24.8 |
| 142 | 65.8 | 51.0 | 76.4 |






Abundance Scan 1576 (12.257 min): BP021488.D\data.ms (-)



Abundance Scan 1576 (12.257 min): BP021488.D\data.ms (-)



#37

2-Methylnaphthalene

Concen: 76.659 ng

RT: 12.257 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP021488.D

Acq: 13 Aug 2024 17:36

Instrument :

BNA_P


ClientSampleId :

SSTDICC080


**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/14/2024


Supervised By :mohammad ahmed 08/14/2024



Abundance Scan 1614 (12.481 min): BP021488.D\data.ms (-)



Abundance Scan 1614 (12.481 min): BP021488.D\data.ms (-)



#38

1-Methylnaphthalene

Concen: 75.769 ng

RT: 12.481 min Scan# 1614

Delta R.T. -0.000 min

Lab File: BP021488.D

Acq: 13 Aug 2024 17:36


Tgt Ion:142 Resp: 4237982


Ion Ratio Lower Upper

142 100


141 92.6 73.5 110.3

116 4.3 3.4 5.2






Abundance Scan 1949 (14.451 min): BP021488.D\data.ms (-)



Abundance Scan 1949 (14.451 min): BP021488.D\data.ms (-)



#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 14.451 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP021488.D

Acq: 13 Aug 2024 17:36

Instrument :

BNA_P



ClientSampleId :

SSTDICC080


**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/14/2024


Supervised By :mohammad ahmed 08/14/2024



Abundance Scan 1639 (12.628 min): BP021488.D\data.ms (-)



Abundance Scan 1639 (12.628 min): BP021488.D\data.ms (-)



#40

1,2,4,5-Tetrachlorobenzene

Concen: 82.416 ng

RT: 12.628 min Scan# 1639

Delta R.T. -0.000 min

Lab File: BP021488.D

Acq: 13 Aug 2024 17:36

Tgt Ion:216 Resp: 2395337


Ion Ratio Lower Upper


216 100

214 78.3 62.5 93.7

179 20.3 16.4 24.6

108 20.4 16.7 25.1







#41
Hexachlorocyclopentadiene
Concen: 91.971 ng
RT: 12.616 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP021488.D
Acq: 13 Aug 2024 17:36

Instrument : BNA_P
ClientSampleId : SSTDICC080


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#42
2,4,6-Tribromophenol
Concen: 172.341 ng
RT: 15.951 min Scan# 2204
Delta R.T. -0.018 min
Lab File: BP021488.D
Acq: 13 Aug 2024 17:36

Tgt Ion:330 Resp: 1902622
Ion Ratio Lower Upper
330 100
332 96.5 77.4 116.0
141 29.3 24.5 36.7



#43

2,4,6-Trichlorophenol

Concen: 90.107 ng

RT: 12.863 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP021488.D


Acq: 13 Aug 2024 17:36

Instrument :


BNA_P

ClientSampleId :


SSTDICC080



Abundance Scan 1679 (12.863 min): BP021488.D\data.ms (-)



Abundance Scan 1679 (12.863 min): BP021488.D\data.ms (-)



Tgt Ion:196 Resp: 1454749

Ion Ratio Lower Upper

196 100



198 95.8 76.6 114.8

200 30.9 24.5 36.7


Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel 08/14/2024


Supervised By :mohammad ahmed 08/14/2024



Abundance Scan 1690 (12.928 min): BP021488.D\data.ms (-)



Abundance Scan 1690 (12.928 min): BP021488.D\data.ms (-)



#44

2,4,5-Trichlorophenol

Concen: 89.412 ng

RT: 12.928 min Scan# 1690

Delta R.T. -0.006 min

Lab File: BP021488.D

Acq: 13 Aug 2024 17:36

Tgt Ion:196 Resp: 1603816


Ion Ratio Lower Upper




196 100

198 94.7 78.9 118.3

97 59.3 49.4 74.0

132 26.3 21.1 31.7





#45

2-Fluorobiphenyl

Concen: 153.020 ng

RT: 13.063 min Scan# 1

Delta R.T. -0.000 min

Lab File: BP021488.D

Acq: 13 Aug 2024 17:36

Instrument :

BNA_P




ClientSampleId :

SSTDICC080

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#46

1,1'-Biphenyl

Concen: 78.343 ng

RT: 13.275 min Scan# 1749

Delta R.T. -0.006 min

Lab File: BP021488.D

Acq: 13 Aug 2024 17:36


Tgt Ion:154 Resp: 5498161


Ion Ratio Lower Upper

154 100

153 40.7 20.7 60.7

76 16.6 0.0 37.7







#47
2-Chloronaphthalene
Concen: 78.980 ng
RT: 13.310 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP021488.D
Acq: 13 Aug 2024 17:36

Instrument : BNA_P
ClientSampleId : SSTDICC080


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#48
2-Nitroaniline
Concen: 81.000 ng
RT: 13.504 min Scan# 1788
Delta R.T. 0.000 min
Lab File: BP021488.D
Acq: 13 Aug 2024 17:36

Tgt Ion: 65 Resp: 1301093
Ion Ratio Lower Upper
65 100
92 76.7 58.1 87.1
138 89.8 65.5 98.3





#49

Acenaphthylene

Concen: 77.556 ng

RT: 14.169 min Scan# 1

Delta R.T. -0.000 min

Lab File: BP021488.D

Acq: 13 Aug 2024 17:36

Instrument :

BNA_P



ClientSampleId :

SSTDICC080

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#50

Dimethylphthalate

Concen: 76.844 ng

RT: 13.904 min Scan# 1856

Delta R.T. -0.006 min

Lab File: BP021488.D

Acq: 13 Aug 2024 17:36


Tgt Ion:163 Resp: 5104657


Ion Ratio Lower Upper

163 100

194 4.2 3.3 4.9

164 10.2 8.3 12.5







#51
2,6-Dinitrotoluene
Concen: 80.745 ng
RT: 14.010 min Scan# 1
Delta R.T. -0.012 min
Lab File: BP021488.D
Acq: 13 Aug 2024 17:36

Instrument : BNA_P
ClientSampleId : SSTDICC080


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#52
Acenaphthene
Concen: 78.480 ng
RT: 14.510 min Scan# 1959
Delta R.T. -0.012 min
Lab File: BP021488.D
Acq: 13 Aug 2024 17:36

Tgt Ion:154 Resp: 4136001
Ion Ratio Lower Upper
154 100
153 112.6 91.1 136.7
152 52.9 43.0 64.4







#53
3-Nitroaniline
Concen: 80.567 ng
RT: 14.340 min Scan# 1
Delta R.T. -0.011 min
Lab File: BP021488.D
Acq: 13 Aug 2024 17:36

Instrument : BNA_P
ClientSampleId : SSTDICC080


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#54
2,4-Dinitrophenol
Concen: 79.528 ng
RT: 14.545 min Scan# 1965
Delta R.T. -0.006 min
Lab File: BP021488.D
Acq: 13 Aug 2024 17:36

Tgt Ion:184 Resp: 370353
Ion Ratio Lower Upper
184 100
63 74.3 64.9 97.3
154 59.3 52.1 78.1







#55
Dibenzofuran
Concen: 77.067 ng
RT: 14.851 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP021488.D
Acq: 13 Aug 2024 17:36

Instrument : BNA_P
ClientSampleId : SSTDICC080


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#56
4-Nitrophenol
Concen: 80.388 ng
RT: 14.640 min Scan# 1981
Delta R.T. -0.011 min
Lab File: BP021488.D
Acq: 13 Aug 2024 17:36

Tgt Ion:139 Resp: 885944
Ion Ratio Lower Upper
139 100
109 87.2 70.2 110.2
65 117.9 101.8 141.8







#57
2,4-Dinitrotoluene
Concen: 81.172 ng
RT: 14.804 min Scan# 2129
Delta R.T. -0.006 min
Lab File: BP021488.D
Acq: 13 Aug 2024 17:36

Instrument : BNA_P
ClientSampleId : SSTDICC080


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#58
Fluorene
Concen: 73.441 ng
RT: 15.510 min Scan# 2129
Delta R.T. -0.006 min
Lab File: BP021488.D
Acq: 13 Aug 2024 17:36

Tgt Ion:166 Resp: 4891299
Ion Ratio Lower Upper
166 100
165 97.9 78.1 117.1
167 13.6 11.0 16.4







#59
2,3,4,6-Tetrachlorophenol
Concen: 86.863 ng
RT: 15.075 min Scan# 2
Delta R.T. -0.012 min
Lab File: BP021488.D
Acq: 13 Aug 2024 17:36

Instrument : BNA_P
ClientSampleId : SSTDICC080


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#60
Diethylphthalate
Concen: 76.268 ng
RT: 15.292 min Scan# 2092
Delta R.T. -0.012 min
Lab File: BP021488.D
Acq: 13 Aug 2024 17:36

Tgt Ion:149 Resp: 5198353
Ion Ratio Lower Upper
149 100
177 23.3 17.8 26.8
150 12.7 10.1 15.1







#61
4-Chlorophenyl-phenylether
Concen: 74.256 ng
RT: 15.510 min Scan# 2131
Delta R.T. -0.012 min
Lab File: BP021488.D
Acq: 13 Aug 2024 17:36

Instrument : BNA_P
ClientSampleId : SSTDICC080


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#62
4-Nitroaniline
Concen: 79.329 ng
RT: 15.522 min Scan# 2131
Delta R.T. -0.006 min
Lab File: BP021488.D
Acq: 13 Aug 2024 17:36

Tgt Ion:138 Resp: 1101954
Ion Ratio Lower Upper
138 100
92 61.5 46.4 86.4
108 87.6 73.2 113.2





#63

Azobenzene

Concen: 73.633 ng

RT: 15.804 min Scan# 2

Delta R.T. -0.018 min

Lab File: BP021488.D

Acq: 13 Aug 2024 17:36

Instrument :

BNA_P



ClientSampleId :

SSTDICC080

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 17.239 min Scan# 2423

Delta R.T. -0.012 min

Lab File: BP021488.D

Acq: 13 Aug 2024 17:36


Tgt Ion:188 Resp: 2007991



Ion Ratio Lower Upper

188 100

94 12.5 10.2 15.4

80 14.4 11.6 17.4






#65
4,6-Dinitro-2-methylphenol
Concen: 79.442 ng
RT: 15.575 min Scan# 2165
Delta R.T. -0.012 min
Lab File: BP021488.D
Acq: 13 Aug 2024 17:36

Instrument : BNA_P
ClientSampleId : SSTDICC080


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#66
n-Nitrosodiphenylamine
Concen: 79.247 ng
RT: 15.722 min Scan# 2165
Delta R.T. -0.018 min
Lab File: BP021488.D
Acq: 13 Aug 2024 17:36

Tgt Ion:169 Resp: 4264380
Ion Ratio Lower Upper
169 100
168 65.7 53.4 80.2
167 35.8 28.7 43.1







#67
4-Bromophenyl-phenylether
Concen: 81.480 ng
RT: 16.416 min Scan# 2
Delta R.T. -0.024 min
Lab File: BP021488.D
Acq: 13 Aug 2024 17:36

Instrument : BNA_P
ClientSampleId : SSTDICC080


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#68
Hexachlorobenzene
Concen: 79.402 ng
RT: 16.534 min Scan# 2303
Delta R.T. -0.018 min
Lab File: BP021488.D
Acq: 13 Aug 2024 17:36

Tgt Ion:284 Resp: 1994136
Ion Ratio Lower Upper
284 100
142 29.7 23.4 35.2
249 30.9 24.1 36.1







#69
Atrazine
Concen: 75.542 ng
RT: 16.698 min Scan# 2
Delta R.T. -0.018 min
Lab File: BP021488.D
Acq: 13 Aug 2024 17:36

Instrument : BNA_P
ClientSampleId : SSTDICC080


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#70
Pentachlorophenol
Concen: 91.649 ng
RT: 16.881 min Scan# 2362
Delta R.T. -0.011 min
Lab File: BP021488.D
Acq: 13 Aug 2024 17:36

Tgt Ion:266 Resp: 1256277
Ion Ratio Lower Upper
266 100
268 64.2 51.8 77.6
264 63.3 50.5 75.7







#71
Phenanthrene
Concen: 77.481 ng
RT: 17.286 min Scan# 2431
Delta R.T. -0.012 min
Lab File: BP021488.D
Acq: 13 Aug 2024 17:36

Instrument : BNA_P
ClientSampleId : SSTDICC080


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024




#72
Anthracene
Concen: 77.613 ng
RT: 17.381 min Scan# 2447
Delta R.T. -0.018 min
Lab File: BP021488.D
Acq: 13 Aug 2024 17:36

Tgt Ion:178 Resp: 7350512
Ion Ratio Lower Upper
178 100
176 18.8 14.9 22.3
179 15.8 12.5 18.7





Abundance Scan 2493 (17.651 min): BP021488.D\data.ms





#73
Carbazole
Concen: 78.491 ng
RT: 17.651 min Scan# 2
Delta R.T. -0.018 min
Lab File: BP021488.D
Acq: 13 Aug 2024 17:36


Instrument : BNA_P
ClientSampleId : SSTDICC080

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024






Abundance Scan 2596 (18.257 min): BP021488.D\data.ms




#74
Di-n-butylphthalate
Concen: 80.689 ng
RT: 18.257 min Scan# 2596
Delta R.T. -0.018 min
Lab File: BP021488.D
Acq: 13 Aug 2024 17:36

Tgt Ion:149 Resp: 9044596
Ion Ratio Lower Upper
149 100
150 9.6 7.5 11.3
104 6.1 4.9 7.3









#77
Benzidine
Concen: 82.290 ng
RT: 19.557 min Scan# 21
Delta R.T. -0.018 min
Lab File: BP021488.D
Acq: 13 Aug 2024 17:36

Instrument : BNA_P
ClientSampleId : SSTDICC080


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#78
Pyrene
Concen: 78.450 ng
RT: 19.745 min Scan# 2849
Delta R.T. -0.006 min
Lab File: BP021488.D
Acq: 13 Aug 2024 17:36

Tgt Ion:202 Resp: 9406096
Ion Ratio Lower Upper
202 100
200 21.6 16.7 25.1
203 18.1 14.1 21.1





#79

Terphenyl-d14

Concen: 145.311 ng

RT: 19.969 min Scan# 2

Delta R.T. -0.006 min

Lab File: BP021488.D

Acq: 13 Aug 2024 17:36

Instrument :

BNA_P



ClientSampleId :

SSTDICC080

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#80

Butylbenzylphthalate

Concen: 80.964 ng

RT: 20.704 min Scan# 3012

Delta R.T. -0.006 min

Lab File: BP021488.D

Acq: 13 Aug 2024 17:36


Tgt Ion:149 Resp: 3806046

Ion Ratio Lower Upper

149 100

91 89.4 73.3 109.9

206 22.3 17.1 25.7



#81

Benzo(a)anthracene

Concen: 79.168 ng

RT: 21.680 min Scan# 3

Delta R.T. -0.000 min

Lab File: BP021488.D




Acq: 13 Aug 2024 17:36

Instrument :

BNA_P

ClientSampleId :

SSTDICC080



Tgt Ion:228 Resp: 911669

Ion Ratio Lower Upper

228 100





226 27.8 21.4 32.2

229 20.2 15.8 23.6

Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#82

3,3'-Dichlorobenzidine

Concen: 83.768 ng

RT: 21.598 min Scan# 3164

Delta R.T. -0.012 min

Lab File: BP021488.D

Acq: 13 Aug 2024 17:36


Tgt Ion:252 Resp: 3092890


Ion Ratio Lower Upper

252 100

254 64.5 51.7 77.5

126 10.4 8.7 13.1







#83

Chrysene
Concen: 78.894 ng
RT: 21.751 min Scan# 3
Delta R.T. -0.000 min
Lab File: BP021488.D
Acq: 13 Aug 2024 17:36

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024




#84

Bis(2-ethylhexyl)phthalate
Concen: 89.068 ng
RT: 21.639 min Scan# 3171
Delta R.T. -0.006 min
Lab File: BP021488.D
Acq: 13 Aug 2024 17:36

Tgt Ion:149 Resp: 6062099
Ion Ratio Lower Upper
149 100
167 28.2 22.3 33.5
279 4.7 3.8 5.6







#85
Di-n-octyl phthalate
Concen: 80.797 ng
RT: 22.957 min Scan# 3
Delta R.T. -0.006 min
Lab File: BP021488.D
Acq: 13 Aug 2024 17:36

Instrument : BNA_P
ClientSampleId : SSTDICC080


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#86
Perylene-d12
Concen: 20.000 ng
RT: 25.110 min Scan# 3761
Delta R.T. 0.006 min
Lab File: BP021488.D
Acq: 13 Aug 2024 17:36

Tgt Ion:264 Resp: 2148798
Ion Ratio Lower Upper
264 100
260 23.0 18.6 28.0
265 21.4 17.1 25.7



#89

Benzo(k)fluoranthene

Concen: 79.279 ng

RT: 24.121 min Scan# 3

Delta R.T. 0.012 min

Lab File: BP021488.D




Acq: 13 Aug 2024 17:36

Instrument :

BNA_P

ClientSampleId :

SSTDICC080



Tgt Ion:252 Resp: 954416

Ion Ratio Lower Upper

252 100





253 22.1 17.5 26.3

125 10.2 7.8 11.8

Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#90
Benzo(a)pyrene
Concen: 81.773 ng
RT: 24.962 min Scan# 3736
Delta R.T. 0.012 min
Lab File: BP021488.D
Acq: 13 Aug 2024 17:36


Tgt Ion:252 Resp: 8757562



Ion Ratio Lower Upper

252 100

253 22.2 17.7 26.5

125 11.2 8.7 13.1








#91
Dibenzo(a,h)anthracene
Concen: 82.349 ng
RT: 29.098 min Scan# 4
Delta R.T. 0.012 min
Lab File: BP021488.D
Acq: 13 Aug 2024 17:36

Instrument : BNA_P
ClientSampleId : SSTDICC080


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#92
Benzo(g,h,i)perylene
Concen: 83.581 ng
RT: 30.209 min Scan# 4628
Delta R.T. 0.035 min
Lab File: BP021488.D
Acq: 13 Aug 2024 17:36

Tgt Ion:276 Resp: 9695156
Ion Ratio Lower Upper
276 100
277 23.9 19.0 28.6
138 20.1 16.0 24.0



Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081324\
 Data File : BP021489.D
 Acq On : 13 Aug 2024 19:03
 Operator : RC/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 ICSVBP081324

Quant Time: Aug 14 00:10:09 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 14 00:00:30 2024
 Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024
 Supervised By :mohammad ahmed 08/14/2024

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|------------------------------------|--------|------|----------|--------|-------|----------|
| Internal Standards | | | | | | |
| 1) 1,4-Dichlorobenzene-d4 | 7.805 | 152 | 443421 | 20.000 | ng | 0.00 |
| 21) Naphthalene-d8 | 10.599 | 136 | 2000407 | 20.000 | ng | 0.00 |
| 39) Acenaphthene-d10 | 14.463 | 164 | 1334240 | 20.000 | ng | 0.00 |
| 64) Phenanthrene-d10 | 17.257 | 188 | 2538103 | 20.000 | ng | 0.00 |
| 76) Chrysene-d12 | 21.716 | 240 | 1932921 | 20.000 | ng | 0.02 |
| 86) Perylene-d12 | 25.121 | 264 | 2127555 | 20.000 | ng | 0.02 |
| System Monitoring Compounds | | | | | | |
| 5) 2-Fluorophenol | 5.393 | 112 | 2413201 | 80.810 | ng | 0.00 |
| 7) Phenol-d6 | 6.964 | 99 | 3740158 | 85.856 | ng | 0.00 |
| 23) Nitrobenzene-d5 | 8.958 | 82 | 3463320 | 89.361 | ng | 0.00 |
| 42) 2,4,6-Tribromophenol | 15.975 | 330 | 1239599 | 83.477 | ng | 0.00 |
| 45) 2-Fluorobiphenyl | 13.069 | 172 | 6875867 | 78.649 | ng | 0.00 |
| 79) Terphenyl-d14 | 19.986 | 244 | 8578102 | 86.216 | ng | 0.01 |
| Target Compounds | | | | | | |
| | | | | Qvalue | | |
| 2) 1,4-Dioxane | 3.317 | 88 | 537338 | 37.948 | ng | 100 |
| 3) Pyridine | 3.705 | 79 | 1598732m | 40.445 | ng | |
| 4) n-Nitrosodimethylamine | 3.617 | 42 | 478013 | 40.233 | ng | 100 |
| 6) Aniline | 7.128 | 93 | 1811514 | 41.596 | ng | 100 |
| 8) 2-Chlorophenol | 7.375 | 128 | 1167341 | 42.300 | ng | 99 |
| 9) Benzaldehyde | 6.946 | 77 | 1127798 | 40.434 | ng | 99 |
| 10) Phenol | 6.993 | 94 | 1922019 | 42.607 | ng | 97 |
| 11) bis(2-Chloroethyl)ether | 7.228 | 93 | 1585750 | 41.543 | ng | 100 |
| 12) 1,3-Dichlorobenzene | 7.699 | 146 | 1300530 | 39.674 | ng | 99 |
| 13) 1,4-Dichlorobenzene | 7.840 | 146 | 1309606 | 39.564 | ng | 99 |
| 14) 1,2-Dichlorobenzene | 8.158 | 146 | 1271884 | 39.873 | ng | 99 |
| 15) Benzyl Alcohol | 8.034 | 79 | 1390052 | 44.472 | ng | 99 |
| 16) 2,2'-oxybis(1-Chloropr... | 8.340 | 45 | 1471975 | 41.875 | ng | 99 |
| 17) 2-Methylphenol | 8.234 | 107 | 1245871 | 43.687 | ng | 100 |
| 18) Hexachloroethane | 8.893 | 117 | 540967 | 40.755 | ng | 99 |
| 19) n-Nitroso-di-n-propyla... | 8.611 | 70 | 1303123 | 43.304 | ng | 99 |
| 20) 3+4-Methylphenols | 8.563 | 107 | 1724643 | 44.850 | ng | 98 |
| 22) Acetophenone | 8.622 | 105 | 2454249 | 39.936 | ng | 98 |
| 24) Nitrobenzene | 8.999 | 77 | 1820293 | 43.244 | ng | 98 |
| 25) Isophorone | 9.516 | 82 | 3726496 | 40.700 | ng | 100 |
| 26) 2-Nitrophenol | 9.705 | 139 | 501343 | 47.780 | ng | 99 |
| 27) 2,4-Dimethylphenol | 9.758 | 122 | 930478 | 41.179 | ng | 99 |
| 28) bis(2-Chloroethoxy)met... | 10.005 | 93 | 2260063 | 40.438 | ng | 99 |
| 29) 2,4-Dichlorophenol | 10.240 | 162 | 1159801 | 43.442 | ng | 99 |
| 30) 1,2,4-Trichlorobenzene | 10.458 | 180 | 1339874 | 39.146 | ng | 97 |
| 31) Naphthalene | 10.646 | 128 | 4141293 | 39.629 | ng | 99 |
| 32) Benzoic acid | 9.899 | 122 | 747114 | 49.839 | ng | 98 |
| 33) 4-Chloroaniline | 10.746 | 127 | 1594388 | 40.617 | ng | 99 |
| 34) Hexachlorobutadiene | 10.946 | 225 | 842702 | 39.102 | ng | 99 |
| 35) Caprolactam | 11.505 | 113 | 464342 | 41.786 | ng | 97 |
| 36) 4-Chloro-3-methylphenol | 11.857 | 107 | 1591956 | 42.987 | ng | 99 |
| 37) 2-Methylnaphthalene | 12.263 | 142 | 2882235 | 40.412 | ng | 99 |
| 38) 1-Methylnaphthalene | 12.475 | 142 | 2825877 | 40.173 | ng | 100 |
| 40) 1,2,4,5-Tetrachloroben... | 12.640 | 216 | 1573947 | 40.261 | ng | 100 |
| 41) Hexachlorocyclopentadiene | 12.622 | 237 | 535050 | 43.183 | ng | 99 |
| 43) 2,4,6-Trichlorophenol | 12.869 | 196 | 960914 | 44.249 | ng | 99 |

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081324\
 Data File : BP021489.D
 Acq On : 13 Aug 2024 19:03
 Operator : RC/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 ICBP081324

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024
 Supervised By :mohammad ahmed 08/14/2024

Quant Time: Aug 14 00:10:09 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 14 00:00:30 2024
 Response via : Initial Calibration

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|-------------------------------|--------|------|----------|--------|-------|----------|
| 44) 2,4,5-Trichlorophenol | 12.940 | 196 | 1071161 | 44.396 | ng | 98 |
| 46) 1,1'-Biphenyl | 13.281 | 154 | 3732699 | 39.542 | ng | 99 |
| 47) 2-Chloronaphthalene | 13.322 | 162 | 2917301 | 39.695 | ng | 99 |
| 48) 2-Nitroaniline | 13.510 | 65 | 874355 | 43.044 | ng | 98 |
| 49) Acenaphthylene | 14.181 | 152 | 4270895 | 39.531 | ng | 100 |
| 50) Dimethylphthalate | 13.910 | 163 | 3496130 | 39.127 | ng | 99 |
| 51) 2,6-Dinitrotoluene | 14.022 | 165 | 735994 | 41.424 | ng | 97 |
| 52) Acenaphthene | 14.528 | 154 | 2790584 | 39.366 | ng | 99 |
| 53) 3-Nitroaniline | 14.351 | 138 | 705420 | 40.483 | ng | 97 |
| 54) 2,4-Dinitrophenol | 14.546 | 184 | 215120 | 45.065 | ng | # 56 |
| 55) Dibenzofuran | 14.863 | 168 | 4202990 | 38.543 | ng | 99 |
| 56) 4-Nitrophenol | 14.640 | 139 | 545969 | 39.582 | ng | 98 |
| 57) 2,4-Dinitrotoluene | 14.810 | 165 | 917790 | 40.864 | ng | 93 |
| 58) Fluorene | 15.522 | 166 | 3409197 | 38.055 | ng | 100 |
| 59) 2,3,4,6-Tetrachlorophenol | 15.093 | 232 | 866459 | 42.310 | ng | 99 |
| 60) Diethylphthalate | 15.293 | 149 | 3461071 | 37.752 | ng | 99 |
| 61) 4-Chlorophenyl-phenyle... | 15.528 | 204 | 1795669 | 38.777 | ng | 98 |
| 62) 4-Nitroaniline | 15.534 | 138 | 685273 | 38.254 | ng | 99 |
| 63) Azobenzene | 15.828 | 77 | 4263599 | 37.833 | ng | 99 |
| 65) 4,6-Dinitro-2-methylph... | 15.593 | 198 | 343500 | 46.281 | ng | 99 |
| 66) n-Nitrosodiphenylamine | 15.734 | 169 | 2863059 | 42.093 | ng | 100 |
| 67) 4-Bromophenyl-phenylether | 16.440 | 248 | 1135965 | 42.289 | ng | 99 |
| 68) Hexachlorobenzene | 16.551 | 284 | 1306627 | 41.161 | ng | 97 |
| 69) Atrazine | 16.710 | 200 | 907263 | 38.851 | ng | 99 |
| 70) Pentachlorophenol | 16.898 | 266 | 770079 | 44.446 | ng | 100 |
| 71) Phenanthrene | 17.298 | 178 | 4826486 | 39.541 | ng | 100 |
| 72) Anthracene | 17.398 | 178 | 4745227 | 39.639 | ng | 100 |
| 73) Carbazole | 17.675 | 167 | 4321248 | 37.457 | ng | 99 |
| 74) Di-n-butylphthalate | 18.281 | 149 | 5456093 | 38.509 | ng | 100 |
| 75) Fluoranthene | 19.386 | 202 | 5434874 | 36.288 | ng | 99 |
| 77) Benzidine | 19.569 | 184 | 1367550 | 30.320 | ng | 99 |
| 78) Pyrene | 19.763 | 202 | 5438173 | 43.044 | ng | 100 |
| 80) Butylbenzylphthalate | 20.710 | 149 | 1972270 | 41.647 | ng | 97 |
| 81) Benzo(a)anthracene | 21.692 | 228 | 4874495 | 40.172 | ng | 100 |
| 82) 3,3'-Dichlorobenzidine | 21.604 | 252 | 1592617 | 40.936 | ng | 100 |
| 83) Chrysene | 21.763 | 228 | 4554886 | 39.759 | ng | 100 |
| 84) Bis(2-ethylhexyl)phtha... | 21.657 | 149 | 3111178 | 43.381 | ng | 100 |
| 85) Di-n-octyl phthalate | 22.974 | 149 | 4908258 | 39.871 | ng | 100 |
| 87) Indeno(1,2,3-cd)pyrene | 28.986 | 276 | 5668848m | 41.105 | ng | |
| 88) Benzo(b)fluoranthene | 24.039 | 252 | 4934155 | 40.381 | ng | 99 |
| 89) Benzo(k)fluoranthene | 24.121 | 252 | 4765945 | 39.984 | ng | 99 |
| 90) Benzo(a)pyrene | 24.968 | 252 | 4311183 | 40.657 | ng | 100 |
| 91) Dibenzo(a,h)anthracene | 29.104 | 278 | 4717815 | 41.172 | ng | 99 |
| 92) Benzo(g,h,i)perylene | 30.203 | 276 | 4769954 | 41.532 | ng | 99 |

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081324\
 Data File : BP021489.D
 Acq On : 13 Aug 2024 19:03
 Operator : RC/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 11 Sample Multiplier: 1


Quant Time: Aug 14 00:10:09 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 14 00:00:30 2024
 Response via : Initial Calibration

Instrument :
 BNA_P
 ClientSampleId :
 ICSVBP081324

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 08/14/2024
 Supervised By :mohammad ahmed 08/14/2024







#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.805 min Scan# 81
Delta R.T. -0.006 min
Lab File: BP021489.D
Acq: 13 Aug 2024 19:03

Instrument : BNA_P
ClientSampleId : ICVBP081324


Manual Integrations
APPROVED




Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#2
1,4-Dioxane
Concen: 37.948 ng
RT: 3.317 min Scan# 56
Delta R.T. 0.000 min
Lab File: BP021489.D
Acq: 13 Aug 2024 19:03

Tgt Ion: 88 Resp: 537338
Ion Ratio Lower Upper
88 100
58 61.0 48.7 73.1
43 19.3 15.4 23.2





#3

Pyridine

Concen: 40.445 ng m

RT: 3.705 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP021489.D

Acq: 13 Aug 2024 19:03

Instrument :

BNA_P





ClientSampleId :

ICVBP081324

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#4

n-Nitrosodimethylamine

Concen: 40.233 ng

RT: 3.617 min Scan# 107

Delta R.T. 0.000 min

Lab File: BP021489.D

Acq: 13 Aug 2024 19:03


Tgt Ion: 42 Resp: 478013


Ion Ratio Lower Upper

42 100

74 192.8 154.3 231.5

44 7.0 6.2 9.4







#5
2-Fluorophenol
Concen: 80.810 ng
RT: 5.393 min Scan# 409
Delta R.T. 0.000 min
Lab File: BP021489.D
Acq: 13 Aug 2024 19:03

Instrument : BNA_P
ClientSampleId : ICVBP081324


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024




#6
Aniline
Concen: 41.596 ng
RT: 7.128 min Scan# 704
Delta R.T. -0.006 min
Lab File: BP021489.D
Acq: 13 Aug 2024 19:03


Tgt Ion: 93 Resp: 1811514
Ion Ratio Lower Upper
93 100
66 37.7 30.2 45.2
65 18.6 14.7 22.1




Abundance Scan 676 (6.964 min): BP021485.D\data.ms (-66)



Abundance Scan 676 (6.964 min): BP021489.D\data.ms



Abundance Scan 676 (6.964 min): BP021489.D\data.ms (-62)



#7

Phenol-d6

Concen: 85.856 ng

RT: 6.964 min Scan# 6

Delta R.T. 0.000 min

Lab File: BP021489.D

Acq: 13 Aug 2024 19:03

Instrument :

BNA_P

ClientSampleId :

ICVBP081324


**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/14/2024


Supervised By :mohammad ahmed 08/14/2024

Abundance


6.964




Abundance Scan 746 (7.375 min): BP021485.D\data.ms (-73)



Abundance Scan 746 (7.375 min): BP021489.D\data.ms



Abundance Scan 746 (7.375 min): BP021489.D\data.ms (-69)



#8

2-Chlorophenol

Concen: 42.300 ng

RT: 7.375 min Scan# 746

Delta R.T. 0.000 min

Lab File: BP021489.D

Acq: 13 Aug 2024 19:03

Tgt Ion:128 Resp: 1167341

Ion Ratio Lower Upper


128 100

130 31.7 11.6 51.6


64 54.4 35.3 75.3

Abundance


7.375




Abundance Scan 673 (6.946 min): BP021485.D\data.ms (-66)



Abundance Scan 673 (6.946 min): BP021489.D\data.ms



Abundance Scan 673 (6.946 min): BP021489.D\data.ms (-62)



#9

Benzaldehyde

Concen: 40.434 ng

RT: 6.946 min Scan# 6

Delta R.T. -0.000 min

Lab File: BP021489.D

Acq: 13 Aug 2024 19:03

Instrument :

BNA_P


ClientSampleId :

ICVBP081324


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024


Supervised By :mohammad ahmed 08/14/2024




Abundance Scan 680 (6.987 min): BP021485.D\data.ms (-67)



Abundance Scan 681 (6.993 min): BP021489.D\data.ms



Abundance Scan 681 (6.993 min): BP021489.D\data.ms (-62)



#10

Phenol

Concen: 42.607 ng

RT: 6.993 min Scan# 681

Delta R.T. 0.006 min

Lab File: BP021489.D

Acq: 13 Aug 2024 19:03


Tgt Ion: 94 Resp: 1922019

Ion Ratio Lower Upper

94 100

65 25.5 6.5 46.5

66 32.9 14.7 54.7




#11
 bis(2-Chloroethyl)ether
 Concen: 41.543 ng
 RT: 7.228 min Scan# 7
 Delta R.T. 0.000 min
 Lab File: BP021489.D
 Acq: 13 Aug 2024 19:03

Instrument : BNA_P
 ClientSampleId : ICVBP081324

Tgt Ion: 93 Resp: 1585750
 Ion Ratio Lower Upper
 93 100
 63 71.1 51.0 91.0
 95 32.5 12.9 52.9






Manual Integrations APPROVED


Reviewed By :Yogesh Patel 08/14/2024
 Supervised By :mohammad ahmed 08/14/2024



#12
 1,3-Dichlorobenzene
 Concen: 39.674 ng
 RT: 7.699 min Scan# 801
 Delta R.T. 0.000 min
 Lab File: BP021489.D
 Acq: 13 Aug 2024 19:03

Tgt Ion:146 Resp: 1300530
 Ion Ratio Lower Upper
 146 100
 148 64.9 51.8 77.6
 75 38.5 31.3 46.9





#13

1,4-Dichlorobenzene

Concen: 39.564 ng

RT: 7.840 min Scan# 81

Delta R.T. 0.000 min

Lab File: BP021489.D

Acq: 13 Aug 2024 19:03

Instrument :

BNA_P



ClientSampleId :

ICVBP081324

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#14

1,2-Dichlorobenzene

Concen: 39.873 ng

RT: 8.158 min Scan# 879

Delta R.T. 0.000 min

Lab File: BP021489.D

Acq: 13 Aug 2024 19:03


Tgt Ion:146 Resp: 1271884




Ion Ratio Lower Upper

146 100

148 64.1 51.5 77.3

111 50.9 41.1 61.7





#15

Benzyl Alcohol

Concen: 44.472 ng

RT: 8.034 min Scan# 8

Delta R.T. 0.000 min

Lab File: BP021489.D

Acq: 13 Aug 2024 19:03

Instrument :

BNA_P





ClientSampleId :

ICVBP081324

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#16

2,2'-oxybis(1-Chloropropane)

Concen: 41.875 ng

RT: 8.340 min Scan# 910

Delta R.T. 0.006 min

Lab File: BP021489.D

Acq: 13 Aug 2024 19:03


Tgt Ion: 45 Resp: 1471975







Ion Ratio Lower Upper

45 100

77 21.1 1.9 41.9

79 16.5 0.0 36.6





#17

2-Methylphenol

Concen: 43.687 ng

RT: 8.234 min Scan# 8

Delta R.T. 0.000 min

Lab File: BP021489.D

Acq: 13 Aug 2024 19:03

Instrument :

BNA_P


ClientSampleId :

ICVBP081324

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#18

Hexachloroethane

Concen: 40.755 ng

RT: 8.893 min Scan# 1004

Delta R.T. 0.000 min

Lab File: BP021489.D

Acq: 13 Aug 2024 19:03


Tgt Ion:117 Resp: 540967

Ion Ratio Lower Upper




117 100

119 92.9 73.9 110.9

201 91.9 74.9 112.3



Abundance Scan 955 (8.605 min): BP021485.D\data.ms (-94)



#19

n-Nitroso-di-n-propylamine

Concen: 43.304 ng

RT: 8.611 min Scan# 9

Delta R.T. 0.006 min

Lab File: BP021489.D

Acq: 13 Aug 2024 19:03

Instrument :

BNA_P




ClientSampleId :

ICVBP081324

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#20

3+4-Methylphenols

Concen: 44.850 ng

RT: 8.563 min Scan# 948

Delta R.T. 0.006 min

Lab File: BP021489.D

Acq: 13 Aug 2024 19:03

Tgt Ion:107 Resp: 1724643


Ion Ratio Lower Upper


107 100

108 89.2 67.6 107.6

77 34.0 15.0 55.0

79 29.9 8.9 48.9





#21

Naphthalene-d8

Concen: 20.000 ng

RT: 10.599 min Scan# 1

Delta R.T. -0.000 min

Lab File: BP021489.D

Acq: 13 Aug 2024 19:03

Instrument :

BNA_P



ClientSampleId :

ICVBP081324

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#22

Acetophenone

Concen: 39.936 ng

RT: 8.622 min Scan# 958

Delta R.T. 0.006 min

Lab File: BP021489.D

Acq: 13 Aug 2024 19:03

Tgt Ion:105 Resp: 2454249


Ion Ratio Lower Upper


105 100

71 5.3 4.6 7.0

51 18.9 16.3 24.5

120 19.9 16.2 24.2







#23
Nitrobenzene-d5
Concen: 89.361 ng
RT: 8.958 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP021489.D
Acq: 13 Aug 2024 19:03

Instrument : BNA_P
ClientSampleId : ICVBP081324


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#24
Nitrobenzene
Concen: 43.244 ng
RT: 8.999 min Scan# 1022
Delta R.T. 0.006 min
Lab File: BP021489.D
Acq: 13 Aug 2024 19:03

Tgt Ion: 77 Resp: 1820293
Ion Ratio Lower Upper
77 100
123 38.4 29.5 44.3
65 12.3 9.9 14.9







#25
Isophorone
Concen: 40.700 ng
RT: 9.516 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP021489.D
Acq: 13 Aug 2024 19:03

Instrument : BNA_P
ClientSampleId : ICVBP081324


Manual Integrations
APPROVED




Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#26
2-Nitrophenol
Concen: 47.780 ng
RT: 9.705 min Scan# 1142
Delta R.T. 0.006 min
Lab File: BP021489.D
Acq: 13 Aug 2024 19:03

Tgt Ion:139 Resp: 501343
Ion Ratio Lower Upper
139 100
109 35.9 28.9 43.3
65 62.2 49.0 73.6





#27

2,4-Dimethylphenol

Concen: 41.179 ng

RT: 9.758 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP021489.D

Acq: 13 Aug 2024 19:03

Instrument :

BNA_P





ClientSampleId :

ICVBP081324

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#28

bis(2-Chloroethoxy)methane

Concen: 40.438 ng

RT: 10.005 min Scan# 1193

Delta R.T. 0.000 min

Lab File: BP021489.D

Acq: 13 Aug 2024 19:03


Tgt Ion: 93 Resp: 2260063

Ion Ratio Lower Upper

93 100

95 32.2 26.2 39.4

123 8.9 7.6 11.4



#29

2,4-Dichlorophenol

Concen: 43.442 ng

RT: 10.240 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP021489.D

Acq: 13 Aug 2024 19:03

Instrument :

BNA_P

ClientSampleId :

ICVBP081324

Tgt Ion:162 Resp: 115980

Ion Ratio Lower Upper

162 100

164 65.1 45.9 85.9

98 46.3 26.3 66.3

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024

Abundance

10.240

600000

400000

200000

0

Time-->


10.20 10.40

10.240

600000

400000

200000



#31

Naphthalene

Concen: 39.629 ng

RT: 10.646 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP021489.D

Acq: 13 Aug 2024 19:03

Instrument :

BNA_P

ClientSampleId :

ICVBP081324

Tgt Ion:128 Resp: 414129

Ion Ratio Lower Upper

128 100

129 11.2 8.6 12.8



127 13.3 10.6 15.8

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#32

Benzoic acid

Concen: 49.839 ng

RT: 9.899 min Scan# 1175

Delta R.T. 0.024 min

Lab File: BP021489.D

Acq: 13 Aug 2024 19:03


Tgt Ion:122 Resp: 747114


Ion Ratio Lower Upper

122 100

105 141.1 124.6 164.6

77 115.3 96.2 136.2







#33
4-Chloroaniline
Concen: 40.617 ng
RT: 10.746 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP021489.D
Acq: 13 Aug 2024 19:03

Instrument : BNA_P
ClientSampleId : ICVBP081324


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#34
Hexachlorobutadiene
Concen: 39.102 ng
RT: 10.946 min Scan# 1353
Delta R.T. 0.006 min
Lab File: BP021489.D
Acq: 13 Aug 2024 19:03

Tgt Ion:225 Resp: 842702
Ion Ratio Lower Upper
225 100
223 63.0 49.6 74.4
227 62.9 50.1 75.1







#35
Caprolactam
Concen: 41.786 ng
RT: 11.505 min Scan# 14489.D
Delta R.T. 0.006 min
Lab File: BP021489.D
Acq: 13 Aug 2024 19:03

Instrument : BNA_P
ClientSampleId : ICVBP081324


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#36
4-Chloro-3-methylphenol
Concen: 42.987 ng
RT: 11.857 min Scan# 1508
Delta R.T. -0.006 min
Lab File: BP021489.D
Acq: 13 Aug 2024 19:03

Tgt Ion:107 Resp: 1591956
Ion Ratio Lower Upper
107 100
144 20.9 16.6 24.8
142 64.9 51.0 76.4







#37
2-Methylnaphthalene
Concen: 40.412 ng
RT: 12.263 min Scan# 1
Instrument : BNA_P
Delta R.T. 0.000 min
Lab File: BP021489.D
Acq: 13 Aug 2024 19:03
ClientSampleId : ICVBP081324


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#38
1-Methylnaphthalene
Concen: 40.173 ng
RT: 12.475 min Scan# 1613
Delta R.T. -0.006 min
Lab File: BP021489.D
Acq: 13 Aug 2024 19:03

Tgt Ion:142 Resp: 2825877
Ion Ratio Lower Upper
142 100
141 92.3 73.5 110.3
116 4.4 3.4 5.2



#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 14.463 min Scan# 1

Delta R.T. 0.006 min

Lab File: BP021489.D

Acq: 13 Aug 2024 19:03

Instrument :

BNA_P

ClientSampleId :

ICVBP081324

Tgt Ion:164 Resp: 1334240

Ion Ratio Lower Upper

164 100

162 102.1 81.5 122.3


160 44.9 35.2 52.8

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#40

1,2,4,5-Tetrachlorobenzene

Concen: 40.261 ng

RT: 12.640 min Scan# 1641

Delta R.T. 0.012 min

Lab File: BP021489.D

Acq: 13 Aug 2024 19:03

Tgt Ion:216 Resp: 1573947



Ion Ratio Lower Upper


216 100

214 77.9 62.5 93.7

179 20.3 16.4 24.6

108 20.5 16.7 25.1







#41
Hexachlorocyclopentadiene
Concen: 43.183 ng
RT: 12.622 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP021489.D
Acq: 13 Aug 2024 19:03

Instrument :
BNA_P
ClientSampleId :
ICVBP081324


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#42
2,4,6-Tribromophenol
Concen: 83.477 ng
RT: 15.975 min Scan# 2208
Delta R.T. 0.006 min
Lab File: BP021489.D
Acq: 13 Aug 2024 19:03

Tgt Ion:330 Resp: 1239599
Ion Ratio Lower Upper
330 100
332 96.3 77.4 116.0
141 30.2 24.5 36.7



#43

2,4,6-Trichlorophenol

Concen: 44.249 ng

RT: 12.869 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP021489.D

Acq: 13 Aug 2024 19:03

Instrument :

BNA_P

ClientSampleId :

ICVBP081324

Tgt Ion:196 Resp: 96091

Ion Ratio Lower Upper

196 100

198 97.1 76.6 114.8


200 31.1 24.5 36.7

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#44

2,4,5-Trichlorophenol

Concen: 44.396 ng

RT: 12.940 min Scan# 1692

Delta R.T. 0.006 min

Lab File: BP021489.D

Acq: 13 Aug 2024 19:03

Tgt Ion:196 Resp: 1071161


Ion Ratio Lower Upper




196 100

198 96.9 78.9 118.3

97 58.5 49.4 74.0

132 25.9 21.1 31.7





#45

2-Fluorobiphenyl

Concen: 78.649 ng

RT: 13.069 min Scan# 1

Delta R.T. 0.006 min

Lab File: BP021489.D

Acq: 13 Aug 2024 19:03

Instrument :

BNA_P





ClientSampleId :

ICVBP081324

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#46

1,1'-Biphenyl

Concen: 39.542 ng

RT: 13.281 min Scan# 1750

Delta R.T. -0.000 min

Lab File: BP021489.D

Acq: 13 Aug 2024 19:03


Tgt Ion:154 Resp: 3732699


Ion Ratio Lower Upper

154 100

153 40.3 20.7 60.7

76 16.9 0.0 37.7





#47

2-Chloronaphthalene

Concen: 39.695 ng

RT: 13.322 min Scan# 1

Delta R.T. 0.006 min

Lab File: BP021489.D

Acq: 13 Aug 2024 19:03

Instrument :

BNA_P



ClientSampleId :

ICVBP081324

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#48

2-Nitroaniline

Concen: 43.044 ng

RT: 13.510 min Scan# 1789

Delta R.T. 0.006 min

Lab File: BP021489.D

Acq: 13 Aug 2024 19:03


Tgt Ion: 65 Resp: 874355

Ion Ratio Lower Upper

65 100

92 74.0 58.1 87.1

138 84.4 65.5 98.3



#49

Acenaphthylene

Concen: 39.531 ng

RT: 14.181 min Scan# 1

Delta R.T. 0.012 min

Lab File: BP021489.D




Acq: 13 Aug 2024 19:03

Instrument :

BNA_P

ClientSampleId :

ICVBP081324



Tgt Ion:152 Resp: 427089

Ion Ratio Lower Upper

152 100





151 19.9 15.9 23.9

153 13.0 10.6 16.0

Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#50

Dimethylphthalate

Concen: 39.127 ng

RT: 13.910 min Scan# 1857

Delta R.T. 0.000 min

Lab File: BP021489.D

Acq: 13 Aug 2024 19:03


Tgt Ion:163 Resp: 3496130


Ion Ratio Lower Upper

163 100

194 4.1 3.3 4.9

164 9.7 8.3 12.5







#51
2,6-Dinitrotoluene
Concen: 41.424 ng
RT: 14.022 min Scan# 1
Instrument : BNA_P
Delta R.T. 0.000 min
Lab File: BP021489.D
Acq: 13 Aug 2024 19:03
ClientSampleId : ICVBP081324


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#52
Acenaphthene
Concen: 39.366 ng
RT: 14.528 min Scan# 1962
Delta R.T. 0.006 min
Lab File: BP021489.D
Acq: 13 Aug 2024 19:03

Tgt Ion:154 Resp: 2790584
Ion Ratio Lower Upper
154 100
153 112.8 91.1 136.7
152 54.0 43.0 64.4







#53
3-Nitroaniline
Concen: 40.483 ng
RT: 14.351 min Scan# 1
Instrument : BNA_P
Delta R.T. 0.000 min
Lab File: BP021489.D
Acq: 13 Aug 2024 19:03
ClientSampleId : ICVBP081324


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#54
2,4-Dinitrophenol
Concen: 45.065 ng
RT: 14.546 min Scan# 1965
Delta R.T. -0.005 min
Lab File: BP021489.D
Acq: 13 Aug 2024 19:03

Tgt Ion:184 Resp: 215120
Ion Ratio Lower Upper
184 100
63 79.1 64.9 97.3
154 140.5 52.1 78.1#







#55
Dibenzofuran
Concen: 38.543 ng
RT: 14.863 min Scan# 2
Instrument : BNA_P
Delta R.T. 0.006 min
Lab File: BP021489.D
Acq: 13 Aug 2024 19:03
ClientSampleId : ICVBP081324


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#56
4-Nitrophenol
Concen: 39.582 ng
RT: 14.640 min Scan# 1981
Delta R.T. -0.011 min
Lab File: BP021489.D
Acq: 13 Aug 2024 19:03

Tgt Ion:139 Resp: 545969
Ion Ratio Lower Upper
139 100
109 87.4 70.2 110.2
65 120.5 101.8 141.8







#57
2,4-Dinitrotoluene
Concen: 40.864 ng
RT: 14.810 min Scan# 2130
Delta R.T. 0.000 min
Lab File: BP021489.D
Acq: 13 Aug 2024 19:03

Instrument : BNA_P
ClientSampleId : ICVBP081324


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#58
Fluorene
Concen: 38.055 ng
RT: 15.522 min Scan# 2131
Delta R.T. 0.006 min
Lab File: BP021489.D
Acq: 13 Aug 2024 19:03

Tgt Ion:166 Resp: 3409197
Ion Ratio Lower Upper
166 100
165 97.4 78.1 117.1
167 13.5 11.0 16.4



#59

2,3,4,6-Tetrachlorophenol

Concen: 42.310 ng

RT: 15.093 min Scan# 2

Delta R.T. 0.006 min

Lab File: BP021489.D

Acq: 13 Aug 2024 19:03

Instrument :

BNA_P

ClientSampleId :

ICVBP081324

Tgt Ion:232 Resp: 866459

Ion Ratio Lower Upper

232 100

131 42.1 33.1 49.7

130 2.2 1.7 2.5

166 29.6 23.0 34.4

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024

Abundance

600000 15.093

400000

200000

0

Time--> 15.00 15.10 15.20

#60

Diethylphthalate

Concen: 37.752 ng

RT: 15.293 min Scan# 2092

Delta R.T. -0.012 min

Lab File: BP021489.D

Acq: 13 Aug 2024 19:03

Tgt Ion:149 Resp: 3461071

Ion Ratio Lower Upper

149 100

177 22.8 17.8 26.8

150 12.7 10.1 15.1

Abundance

2000000 15.293



1500000


1000000

500000


0

Time--> 15.20 15.30 15.40





Abundance Scan 2132 (15.528 min): BP021489.D\data.ms



#61

4-Chlorophenyl-phenylether

Concen: 38.777 ng

RT: 15.528 min Scan# 2131

Delta R.T. 0.006 min

Lab File: BP021489.D

Acq: 13 Aug 2024 19:03


Instrument : BNA_P

ClientSampleId : ICVBP081324

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#62

4-Nitroaniline

Concen: 38.254 ng

RT: 15.534 min Scan# 2133

Delta R.T. 0.006 min

Lab File: BP021489.D

Acq: 13 Aug 2024 19:03


Tgt Ion:138 Resp: 685273


Ion Ratio Lower Upper

138 100

92 67.4 46.4 86.4

108 93.0 73.2 113.2





#63

Azobenzene

Concen: 37.833 ng

RT: 15.828 min Scan# 2

Delta R.T. 0.006 min

Lab File: BP021489.D

Acq: 13 Aug 2024 19:03

Instrument :

BNA_P


ClientSampleId :

ICVBP081324

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 17.257 min Scan# 2426

Delta R.T. 0.006 min

Lab File: BP021489.D

Acq: 13 Aug 2024 19:03


Tgt Ion:188 Resp: 2538103

Ion Ratio Lower Upper

188 100

94 13.2 10.2 15.4

80 15.0 11.6 17.4



#65

4,6-Dinitro-2-methylphenol

Concen: 46.281 ng

RT: 15.593 min Scan# 2142

Delta R.T. 0.006 min

Lab File: BP021489.D ClientSampleId :

Acq: 13 Aug 2024 19:03 ICSVBP081324

Instrument : BNA_P

ClientSampleId : ICSVBP081324

Tgt Ion:198 Resp: 343500

Ion Ratio Lower Upper

198 100

51 39.6 21.5 61.5

105 53.0 32.9 72.9

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024

Abundance

300000 15.593

200000

100000

0

Time--> 15.50 15.60 15.70

200000

100000

0

Time--> 15.50 15.60 15.70

Abundance

300000 15.593

200000

100000

0

Time--> 15.50 15.60 15.70

200000

100000

0

Time--> 15.50 15.60 15.70

Abundance

300000 15.593

200000

100000

0

Time--> 15.50 15.60 15.70

200000

100000

0

Time--> 15.50 15.60 15.70

Abundance

1500000 15.734

1000000

500000

0

Time--> 15.60 15.70 15.80

1500000

1000000

500000

0

Time--> 15.60 15.70 15.80

Abundance

1500000 15.734

1000000

500000

0

Time--> 15.60 15.70 15.80

1500000

1000000

500000

#67

4-Bromophenyl-phenylether

Concen: 42.289 ng

RT: 16.440 min Scan# 2

Delta R.T. 0.000 min

Lab File: BP021489.D

Acq: 13 Aug 2024 19:03

Instrument :

BNA_P

ClientSampleId :

ICVBP081324

Tgt Ion:248 Resp: 113596

Ion Ratio Lower Upper

248 100

250 98.2 78.6 117.8

141 66.3 53.9 80.9

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024

Abundance

Scan 2287 (16.440 min): BP021489.D\data.ms (-)

248.0

313.4378.9444.8 518.9

51.1 141.1

50 100 150 200 250 300 350 400 450 500

m/z-->

Sub

Scan 2287 (16.440 min): BP021489.D\data.ms (-)

248.0

335.6402.6469.3534.7

51.0 141.1

50 100 150 200 250 300 350 400 450 500

m/z-->

Time-->

#68

Hexachlorobenzene

Concen: 41.161 ng

RT: 16.551 min Scan# 2306

Delta R.T. 0.000 min

Lab File: BP021489.D

Acq: 13 Aug 2024 19:03

Tgt Ion:284 Resp: 1306627

Ion Ratio Lower Upper

284 100

142 31.8 23.4 35.2

249 30.4 24.1 36.1

Abundance

Scan 2306 (16.551 min): BP021489.D\data.ms (-)

283.9

350.5418.1484.1

71.0 142.0

50 100 150 200 250 300 350 400 450 500

m/z-->

Raw

Scan 2306 (16.551 min): BP021489.D\data.ms (-)

283.9

351.9418.8 532.4

71.0 142.0

50 100 150 200 250 300 350 400 450 500

m/z-->

Sub

Scan 2306 (16.551 min): BP021489.D\data.ms (-)

283.9


349.5 420.0486.1

71.0 142.0

50 100 150 200 250 300 350 400 450 500

m/z-->

Time-->



Tgt Ion:248 Resp: 113596

Ion Ratio Lower Upper

248 100

250 98.2 78.6 117.8

141 66.3 53.9 80.9

Abundance

Scan 2287 (16.440 min): BP021489.D\data.ms (-)

248.0

313.4378.9444.8 518.9

51.1 141.1

50 100 150 200 250 300 350 400 450 500

m/z-->

Sub

Scan 2287 (16.440 min): BP021489.D\data.ms (-)

248.0

335.6402.6469.3534.7

51.0 141.1

50 100 150 200 250 300 350 400 450 500

m/z-->

Time-->

#68

Hexachlorobenzene

Concen: 41.161 ng

RT: 16.551 min Scan# 2306

Delta R.T. 0.000 min

Lab File: BP021489.D

Acq: 13 Aug 2024 19:03

Tgt Ion:284 Resp: 1306627

Ion Ratio Lower Upper

284 100

142 31.8 23.4 35.2

249 30.4 24.1 36.1

Abundance

Scan 2306 (16.551 min): BP021489.D\data.ms (-)

283.9

350.5418.1484.1

71.0 142.0

50 100 150 200 250 300 350 400 450 500

m/z-->

Raw

Scan 2306 (16.551 min): BP021489.D\data.ms (-)

283.9

351.9418.8 532.4

71.0 142.0

50 100 150 200 250 300 350 400 450 500

m/z-->

Sub

Scan 2306 (16.551 min): BP021489.D\data.ms (-)

283.9


349.5 420.0486.1

71.0 142.0

50 100 150 200 250 300 350 400 450 500

m/z-->

Time-->





#69
Atrazine
Concen: 38.851 ng
RT: 16.710 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP021489.D
Acq: 13 Aug 2024 19:03

Instrument : BNA_P
ClientSampleId : ICVBP081324


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#70
Pentachlorophenol
Concen: 44.446 ng
RT: 16.898 min Scan# 2365
Delta R.T. 0.006 min
Lab File: BP021489.D
Acq: 13 Aug 2024 19:03

Tgt Ion:266 Resp: 770079
Ion Ratio Lower Upper
266 100
268 64.2 51.8 77.6
264 63.0 50.5 75.7







#71
Phenanthrene
Concen: 39.541 ng
RT: 17.298 min Scan# 2433
Delta R.T. 0.000 min
Lab File: BP021489.D
Acq: 13 Aug 2024 19:03

Instrument : BNA_P
ClientSampleId : ICVBP081324


Manual Integrations
APPROVED




Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#72
Anthracene
Concen: 39.639 ng
RT: 17.398 min Scan# 2450
Delta R.T. 0.000 min
Lab File: BP021489.D
Acq: 13 Aug 2024 19:03

Tgt Ion:178 Resp: 4745227
Ion Ratio Lower Upper
178 100
176 18.4 14.9 22.3
179 15.4 12.5 18.7








#73
Carbazole
Concen: 37.457 ng
RT: 17.675 min Scan# 2
Instrument : BNA_P
Delta R.T. 0.006 min
Lab File: BP021489.D
Acq: 13 Aug 2024 19:03
ClientSampleId : ICVBP081324

Tgt Ion:167 Resp: 4321243
Ion Ratio Lower Upper
167 100
166 21.6 17.0 25.4
139 12.8 9.9 14.9



Manual Integrations APPROVED

Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#74
Di-n-butylphthalate
Concen: 38.509 ng
RT: 18.281 min Scan# 2600
Delta R.T. 0.006 min
Lab File: BP021489.D
Acq: 13 Aug 2024 19:03

Tgt Ion:149 Resp: 5456093
Ion Ratio Lower Upper
149 100
150 9.5 7.5 11.3
104 6.2 4.9 7.3



#75

Fluoranthene

Concen: 36.288 ng

RT: 19.386 min Scan# 2

Delta R.T. 0.006 min

Lab File: BP021489.D



Acq: 13 Aug 2024 19:03

Instrument :

BNA_P

ClientSampleId :

ICVBP081324



Tgt Ion:202 Resp: 543487

Ion Ratio Lower Upper

202 100




101 14.6 0.0 34.0

203 17.7 0.0 37.5

Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#76

Chrysene-d₁₂

Concen: 20.000 ng

RT: 21.716 min Scan# 3184

Delta R.T. 0.018 min

Lab File: BP021489.D

Acq: 13 Aug 2024 19:03


Tgt Ion:240 Resp: 1932921




Ion Ratio Lower Upper

240 100

120 11.7 9.9 14.9

236 25.8 20.6 30.8









#77
Benzidine
Concen: 30.320 ng
RT: 19.569 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP021489.D
Acq: 13 Aug 2024 19:03

Instrument : BNA_P
ClientSampleId : ICVBP081324


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#78
Pyrene
Concen: 43.044 ng
RT: 19.763 min Scan# 2852
Delta R.T. 0.012 min
Lab File: BP021489.D
Acq: 13 Aug 2024 19:03

Tgt Ion:202 Resp: 5438173
Ion Ratio Lower Upper
202 100
200 21.2 16.7 25.1
203 17.7 14.1 21.1



#79

Terphenyl-d14

Concen: 86.216 ng

RT: 19.986 min Scan# 2

Delta R.T. 0.011 min

Lab File: BP021489.D

Acq: 13 Aug 2024 19:03

Instrument :

BNA_P

ClientSampleId :

ICVBP081324

Tgt Ion:244 Resp: 857810

Ion Ratio Lower Upper

244 100

212 7.0 5.6 8.4

122 11.8 9.4 14.2

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024

Abundance

5000000

4000000

3000000

2000000

1000000

0

19.986

Time-->

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00

20.00


20.00

20.00

20.00

20.00

20.00



#81

Benzo(a)anthracene

Concen: 40.172 ng

RT: 21.692 min Scan# 3

Delta R.T. 0.012 min

Lab File: BP021489.D

Acq: 13 Aug 2024 19:03

Instrument :

BNA_P



ClientSampleId :

ICVBP081324

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024



#82

3,3'-Dichlorobenzidine

Concen: 40.936 ng

RT: 21.604 min Scan# 3165

Delta R.T. -0.006 min

Lab File: BP021489.D

Acq: 13 Aug 2024 19:03


Tgt Ion:252 Resp: 1592617


Ion Ratio Lower Upper

252 100

254 64.6 51.7 77.5

126 11.7 8.7 13.1







#83
Chrysene
Concen: 39.759 ng
RT: 21.763 min Scan# 3190
Delta R.T. 0.012 min
Lab File: BP021489.D
Acq: 13 Aug 2024 19:03

Instrument : BNA_P
ClientSampleId : ICVBP081324


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#84
Bis(2-ethylhexyl)phthalate
Concen: 43.381 ng
RT: 21.657 min Scan# 3174
Delta R.T. 0.012 min
Lab File: BP021489.D
Acq: 13 Aug 2024 19:03

Tgt Ion:149 Resp: 3111178
Ion Ratio Lower Upper
149 100
167 28.0 22.3 33.5
279 4.5 3.8 5.6







#85
Di-n-octyl phthalate
Concen: 39.871 ng
RT: 22.974 min Scan# 3
Delta R.T. 0.011 min
Lab File: BP021489.D
Acq: 13 Aug 2024 19:03

Instrument : BNA_P
ClientSampleId : ICVBP081324


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#86
Perylene-d₁₂
Concen: 20.000 ng
RT: 25.121 min Scan# 3763
Delta R.T. 0.018 min
Lab File: BP021489.D
Acq: 13 Aug 2024 19:03

Tgt Ion:264 Resp: 2127555
Ion Ratio Lower Upper
264 100
260 23.3 18.6 28.0
265 21.3 17.1 25.7



#87

Indeno(1,2,3-cd)pyrene

Concen: 41.105 ng m

RT: 28.986 min Scan# 4

Delta R.T. -0.006 min

Lab File: BP021489.D

Acq: 13 Aug 2024 19:03

Instrument :

BNA_P

ClientSampleId :

ICVBP081324

Tgt Ion:276 Resp: 566884

Ion Ratio Lower Upper

276 100

138 15.6 12.8 19.2

277 19.5 16.5 24.7

Manual Integrations

APPROVED



Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024

Abundance

Scan 4420 (28.986 min): BP021489.D\data.ms (-)



276.1



Abundance

Scan 4420 (28.986 min): BP021489.D\data.ms (-)



276.1



Abundance

Scan 3577 (24.027 min): BP021485.D\data.ms (-)



252.1



Abundance

Scan 3579 (24.039 min): BP021489.D\data.ms (-)





252.1



Abundance

Scan 3579 (24.039 min): BP021489.D\data.ms (-)

252.1



#89

Benzo(k)fluoranthene

Concen: 39.984 ng

RT: 24.121 min Scan# 3

Delta R.T. 0.012 min

Lab File: BP021489.D

Acq: 13 Aug 2024 19:03

Instrument :

BNA_P

ClientSampleId :

ICVBP081324

Tgt Ion:252 Resp: 476594

Ion Ratio Lower Upper

252 100

253 21.8 17.5 26.3

125 10.2 7.8 11.8

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :mohammad ahmed 08/14/2024

Abundance

Scan 3593 (24.121 min): BP021489.D\data.ms (-)

252.1

Raw

Abundance

Scan 3593 (24.121 min): BP021489.D\data.ms (-)

252.1

Sub

Abundance

Scan 3593 (24.121 min): BP021489.D\data.ms (-)

252.1

m/z-->

50 100 150 200 250 300 350 400 450 500

44.0 126.1 320.6 386.9 453.7 540.7

m/z-->

50 100 150 200 250 300 350 400 450 500




52.0 126.1 337.3 412.3 478.6 544.1

m/z-->

Time-->

Time-->

24.00 24.121 24.20







#91
Dibenzo(a,h)anthracene
Concen: 41.172 ng
RT: 29.104 min Scan# 4
Delta R.T. 0.018 min
Lab File: BP021489.D
Acq: 13 Aug 2024 19:03

Instrument : BNA_P
ClientSampleId : ICVBP081324


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 08/14/2024
Supervised By :mohammad ahmed 08/14/2024



#92
Benzo(g,h,i)perylene
Concen: 41.532 ng
RT: 30.203 min Scan# 4627
Delta R.T. 0.029 min
Lab File: BP021489.D
Acq: 13 Aug 2024 19:03

Tgt Ion:276 Resp: 4769954
Ion Ratio Lower Upper
276 100
277 23.7 19.0 28.6
138 20.4 16.0 24.0



Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081324\
 Data File : BP021489.D
 Acq On : 13 Aug 2024 19:03
 Operator : RC/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 ICVBP081324

Quant Time: Aug 14 00:10:09 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 14 00:00:30 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

| | Compound | AvgRF | CCRF | %Dev | Area% | Dev(min) |
|------|-----------------------------|-------|-------|--------|-------|----------|
| 1 I | 1,4-Dichlorobenzene-d4 | 1.000 | 1.000 | 0.0 | 132 | 0.00 |
| 2 | 1,4-Dioxane | 0.639 | 0.606 | 5.2 | 125 | 0.00 |
| 3 | Pyridine | 1.783 | 1.803 | -1.1 | 133 | 0.00 |
| 4 | n-Nitrosodimethylamine | 0.536 | 0.539 | -0.6 | 134 | 0.00 |
| 5 S | 2-Fluorophenol | 1.347 | 1.361 | -1.0 | 134 | 0.00 |
| 6 | Aniline | 1.964 | 2.043 | -4.0 | 137 | 0.00 |
| 7 S | Phenol-d6 | 1.965 | 2.109 | -7.3 | 141 | 0.00 |
| 8 | 2-Chlorophenol | 1.245 | 1.316 | -5.7 | 140 | 0.00 |
| 9 | Benzaldehyde | 1.258 | 1.272 | -1.1 | 141 | 0.00 |
| 10 C | Phenol | 2.035 | 2.167 | -6.5 | 140 | 0.00 |
| 11 | bis(2-Chloroethyl)ether | 1.722 | 1.788 | -3.8 | 139 | 0.00 |
| 12 | 1,3-Dichlorobenzene | 1.479 | 1.466 | 0.9 | 134 | 0.00 |
| 13 C | 1,4-Dichlorobenzene | 1.493 | 1.477 | 1.1 | 133 | 0.00 |
| 14 | 1,2-Dichlorobenzene | 1.439 | 1.434 | 0.3 | 135 | 0.00 |
| 15 | Benzyl Alcohol | 1.410 | 1.567 | -11.1 | 147 | 0.00 |
| 16 | 2,2'-oxybis(1-Chloropropane | 1.585 | 1.660 | -4.7 | 142 | 0.00 |
| 17 | 2-Methylphenol | 1.286 | 1.405 | -9.3 | 144 | 0.00 |
| 18 | Hexachloroethane | 0.599 | 0.610 | -1.8 | 136 | 0.00 |
| 19 P | n-Nitroso-di-n-propylamine | 1.357 | 1.469 | -8.3 | 145 | 0.00 |
| 20 | 3+4-Methylphenols | 1.734 | 1.945 | -12.2 | 147 | 0.00 |
| 21 I | Naphthalene-d8 | 1.000 | 1.000 | 0.0 | 145 | 0.00 |
| 22 | Acetophenone | 0.614 | 0.613 | 0.2 | 145 | 0.00 |
| 23 S | Nitrobenzene-d5 | 0.387 | 0.433 | -11.9 | 153# | 0.00 |
| 24 | Nitrobenzene | 0.421 | 0.455 | -8.1 | 149 | 0.00 |
| 25 | Isophorone | 0.915 | 0.931 | -1.7 | 148 | 0.00 |
| 26 C | 2-Nitrophenol | 0.086 | 0.125 | -45.3# | 193# | 0.00 |
| 27 | 2,4-Dimethylphenol | 0.226 | 0.233 | -3.1 | 150 | 0.00 |
| 28 | bis(2-Chloroethoxy)methane | 0.559 | 0.565 | -1.1 | 149 | 0.00 |
| 29 C | 2,4-Dichlorophenol | 0.267 | 0.290 | -8.6 | 154# | 0.00 |
| 30 | 1,2,4-Trichlorobenzene | 0.342 | 0.335 | 2.0 | 143 | 0.00 |
| 31 | Naphthalene | 1.045 | 1.035 | 1.0 | 144 | 0.00 |
| 32 | Benzoic acid | 0.137 | 0.187 | -36.5# | 210# | 0.02 |
| 33 | 4-Chloroaniline | 0.392 | 0.399 | -1.8 | 146 | 0.00 |
| 34 C | Hexachlorobutadiene | 0.215 | 0.211 | 1.9 | 142 | 0.00 |
| 35 | Caprolactam | 0.111 | 0.116 | -4.5 | 148 | 0.00 |
| 36 C | 4-Chloro-3-methylphenol | 0.370 | 0.398 | -7.6 | 151# | 0.00 |
| 37 | 2-Methylnaphthalene | 0.713 | 0.720 | -1.0 | 148 | 0.00 |
| 38 | 1-Methylnaphthalene | 0.703 | 0.706 | -0.4 | 147 | 0.00 |
| 39 I | Acenaphthene-d10 | 1.000 | 1.000 | 0.0 | 150 | 0.00 |
| 40 | 1,2,4,5-Tetrachlorobenzene | 0.586 | 0.590 | -0.7 | 149 | 0.01 |
| 41 P | Hexachlorocyclopentadiene | 0.186 | 0.201 | -8.1 | 157# | 0.00 |
| 42 S | 2,4,6-Tribromophenol | 0.223 | 0.232 | -4.0 | 147 | 0.00 |
| 43 C | 2,4,6-Trichlorophenol | 0.326 | 0.360 | -10.4 | 158# | 0.00 |
| 44 | 2,4,5-Trichlorophenol | 0.362 | 0.401 | -10.8 | 157# | 0.00 |
| 45 S | 2-Fluorobiphenyl | 1.310 | 1.288 | 1.7 | 146 | 0.00 |
| 46 | 1,1'-Biphenyl | 1.415 | 1.399 | 1.1 | 147 | 0.00 |
| 47 | 2-Chloronaphthalene | 1.102 | 1.093 | 0.8 | 147 | 0.00 |

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081324\
 Data File : BP021489.D
 Acq On : 13 Aug 2024 19:03
 Operator : RC/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 ICVBP081324

Quant Time: Aug 14 00:10:09 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 14 00:00:30 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

| | Compound | AvgRF | CCRF | %Dev | Area% | Dev(min) |
|------|----------------------------|-------|-------|--------|-------|----------|
| 48 | 2-Nitroaniline | 0.256 | 0.328 | -28.1# | 172# | 0.00 |
| 49 | Acenaphthylene | 1.619 | 1.600 | 1.2 | 146 | 0.01 |
| 50 | Dimethylphthalate | 1.339 | 1.310 | 2.2 | 145 | 0.00 |
| 51 | 2,6-Dinitrotoluene | 0.232 | 0.276 | -19.0 | 162# | 0.00 |
| 52 C | Acenaphthene | 1.063 | 1.046 | 1.6 | 146 | 0.00 |
| 53 | 3-Nitroaniline | 0.228 | 0.264 | -15.8 | 156# | 0.00 |
| 54 P | 2,4-Dinitrophenol | 0.069 | 0.081 | -17.4 | 184# | 0.00 |
| 55 | Dibenzofuran | 1.635 | 1.575 | 3.7 | 143 | 0.00 |
| 56 P | 4-Nitrophenol | 0.195 | 0.205 | -5.1 | 152# | -0.01 |
| 57 | 2,4-Dinitrotoluene | 0.284 | 0.344 | -21.1 | 161# | 0.00 |
| 58 | Fluorene | 1.343 | 1.278 | 4.8 | 141 | 0.00 |
| 59 | 2,3,4,6-Tetrachlorophenol | 0.307 | 0.325 | -5.9 | 148 | 0.00 |
| 60 | Diethylphthalate | 1.374 | 1.297 | 5.6 | 139 | -0.01 |
| 61 | 4-Chlorophenyl-phenylether | 0.694 | 0.673 | 3.0 | 145 | 0.00 |
| 62 | 4-Nitroaniline | 0.241 | 0.257 | -6.6 | 141 | 0.00 |
| 63 | Azobenzene | 1.689 | 1.598 | 5.4 | 140 | 0.00 |
| 64 I | Phanthrene-d10 | 1.000 | 1.000 | 0.0 | 133 | 0.00 |
| 65 | 4,6-Dinitro-2-methylphenol | 0.054 | 0.068 | -25.9# | 172# | 0.00 |
| 66 c | n-Nitrosodiphenylamine | 0.536 | 0.564 | -5.2 | 140 | 0.00 |
| 67 | 4-Bromophenyl-phenylether | 0.212 | 0.224 | -5.7 | 144 | 0.00 |
| 68 | Hexachlorobenzene | 0.250 | 0.257 | -2.8 | 136 | 0.00 |
| 69 | Atrazine | 0.184 | 0.179 | 2.7 | 125 | 0.00 |
| 70 C | Pentachlorophenol | 0.137 | 0.152 | -10.9 | 148 | 0.00 |
| 71 | Phanthrene | 0.962 | 0.951 | 1.1 | 131 | 0.00 |
| 72 | Anthracene | 0.943 | 0.935 | 0.8 | 131 | 0.00 |
| 73 | Carbazole | 0.909 | 0.851 | 6.4 | 122 | 0.00 |
| 74 | Di-n-butylphthalate | 1.116 | 1.075 | 3.7 | 125 | 0.00 |
| 75 C | Fluoranthene | 1.180 | 1.071 | 9.2 | 119 | 0.00 |
| 76 I | Chrysene-d12 | 1.000 | 1.000 | 0.0 | 106 | 0.02 |
| 77 | Benzidine | 0.420 | 0.354 | 15.7 | 79 | 0.00 |
| 78 | Pyrene | 1.307 | 1.407 | -7.7 | 116 | 0.01 |
| 79 S | Terphenyl-d14 | 1.029 | 1.109 | -7.8 | 117 | 0.01 |
| 80 | Butylbenzylphthalate | 0.437 | 0.510 | -16.7 | 115 | 0.00 |
| 81 | Benzo(a)anthracene | 1.256 | 1.261 | -0.4 | 106 | 0.01 |
| 82 | 3,3'-Dichlorobenzidine | 0.403 | 0.412 | -2.2 | 102 | 0.00 |
| 83 | Chrysene | 1.185 | 1.178 | 0.6 | 105 | 0.01 |
| 84 | Bis(2-ethylhexyl)phthalate | 0.742 | 0.805 | -8.5 | 109 | 0.01 |
| 85 c | Di-n-octyl phthalate | 1.199 | 1.270 | -5.9 | 108 | 0.01 |
| 86 I | Perylene-d12 | 1.000 | 1.000 | 0.0 | 102 | 0.02 |
| 87 | Indeno(1,2,3-cd)pyrene | 1.296 | 1.332 | -2.8 | 102 | 0.00 |
| 88 | Benzo(b)fluoranthene | 1.149 | 1.160 | -1.0 | 102 | 0.01 |
| 89 | Benzo(k)fluoranthene | 1.121 | 1.120 | 0.1 | 100 | 0.01 |
| 90 C | Benzo(a)pyrene | 0.997 | 1.013 | -1.6 | 102 | 0.02 |
| 91 | Dibenzo(a,h)anthracene | 1.077 | 1.109 | -3.0 | 103 | 0.02 |
| 92 | Benzo(g,h,i)perylene | 1.080 | 1.121 | -3.8 | 104 | 0.03 |

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081324\
Data File : BP021489.D
Acq On : 13 Aug 2024 19:03
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
ICVBP081324

Quant Time: Aug 14 00:10:09 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 14 00:00:30 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

| Compound | AvgRF | CCRF | %Dev | Area% | Dev(min) |
|----------|-------|------|------|-------|----------|
|----------|-------|------|------|-------|----------|

(#) = Out of Range SPCC's out = 0 CCC's out = 1

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081324\
 Data File : BP021489.D
 Acq On : 13 Aug 2024 19:03
 Operator : RC/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 ICVBP081324

Quant Time: Aug 14 00:10:09 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 14 00:00:30 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

| | Compound | Amount | Calc. | %Dev | Area% | Dev(min) |
|------|-----------------------------|--------|--------|-------|-------|----------|
| 1 I | 1,4-Dichlorobenzene-d4 | 20.000 | 20.000 | 0.0 | 132 | 0.00 |
| 2 | 1,4-Dioxane | 40.000 | 37.948 | 5.1 | 125 | 0.00 |
| 3 | Pyridine | 40.000 | 40.445 | -1.1 | 133 | 0.00 |
| 4 | n-Nitrosodimethylamine | 40.000 | 40.233 | -0.6 | 134 | 0.00 |
| 5 S | 2-Fluorophenol | 80.000 | 80.810 | -1.0 | 134 | 0.00 |
| 6 | Aniline | 40.000 | 41.596 | -4.0 | 137 | 0.00 |
| 7 S | Phenol-d6 | 80.000 | 85.856 | -7.3 | 141 | 0.00 |
| 8 | 2-Chlorophenol | 40.000 | 42.300 | -5.7 | 140 | 0.00 |
| 9 | Benzaldehyde | 40.000 | 40.434 | -1.1 | 141 | 0.00 |
| 10 C | Phenol | 40.000 | 42.607 | -6.5 | 140 | 0.00 |
| 11 | bis(2-Chloroethyl)ether | 40.000 | 41.543 | -3.9 | 139 | 0.00 |
| 12 | 1,3-Dichlorobenzene | 40.000 | 39.674 | 0.8 | 134 | 0.00 |
| 13 C | 1,4-Dichlorobenzene | 40.000 | 39.564 | 1.1 | 133 | 0.00 |
| 14 | 1,2-Dichlorobenzene | 40.000 | 39.873 | 0.3 | 135 | 0.00 |
| 15 | Benzyl Alcohol | 40.000 | 44.472 | -11.2 | 147 | 0.00 |
| 16 | 2,2'-oxybis(1-Chloropropane | 40.000 | 41.875 | -4.7 | 142 | 0.00 |
| 17 | 2-Methylphenol | 40.000 | 43.687 | -9.2 | 144 | 0.00 |
| 18 | Hexachloroethane | 40.000 | 40.755 | -1.9 | 136 | 0.00 |
| 19 P | n-Nitroso-di-n-propylamine | 40.000 | 43.304 | -8.3 | 145 | 0.00 |
| 20 | 3+4-Methylphenols | 40.000 | 44.850 | -12.1 | 147 | 0.00 |
| 21 I | Naphthalene-d8 | 20.000 | 20.000 | 0.0 | 145 | 0.00 |
| 22 | Acetophenone | 40.000 | 39.936 | 0.2 | 145 | 0.00 |
| 23 S | Nitrobenzene-d5 | 80.000 | 89.361 | -11.7 | 153 | 0.00 |
| 24 | Nitrobenzene | 40.000 | 43.244 | -8.1 | 149 | 0.00 |
| 25 | Isophorone | 40.000 | 40.700 | -1.8 | 148 | 0.00 |
| 26 C | 2-Nitrophenol | 40.000 | 47.780 | -19.5 | 193 | 0.00 |
| 27 | 2,4-Dimethylphenol | 40.000 | 41.179 | -2.9 | 150 | 0.00 |
| 28 | bis(2-Chloroethoxy)methane | 40.000 | 40.438 | -1.1 | 149 | 0.00 |
| 29 C | 2,4-Dichlorophenol | 40.000 | 43.442 | -8.6 | 154 | 0.00 |
| 30 | 1,2,4-Trichlorobenzene | 40.000 | 39.146 | 2.1 | 143 | 0.00 |
| 31 | Naphthalene | 40.000 | 39.629 | 0.9 | 144 | 0.00 |
| 32 | Benzoic acid | 40.000 | 49.839 | -24.6 | 210 | 0.02 |
| 33 | 4-Chloroaniline | 40.000 | 40.617 | -1.5 | 146 | 0.00 |
| 34 C | Hexachlorobutadiene | 40.000 | 39.102 | 2.2 | 142 | 0.00 |
| 35 | Caprolactam | 40.000 | 41.786 | -4.5 | 148 | 0.00 |
| 36 C | 4-Chloro-3-methylphenol | 40.000 | 42.987 | -7.5 | 151 | 0.00 |
| 37 | 2-Methylnaphthalene | 40.000 | 40.412 | -1.0 | 148 | 0.00 |
| 38 | 1-Methylnaphthalene | 40.000 | 40.173 | -0.4 | 147 | 0.00 |
| 39 I | Acenaphthene-d10 | 20.000 | 20.000 | 0.0 | 150 | 0.00 |
| 40 | 1,2,4,5-Tetrachlorobenzene | 40.000 | 40.261 | -0.7 | 149 | 0.01 |
| 41 P | Hexachlorocyclopentadiene | 40.000 | 43.183 | -8.0 | 157 | 0.00 |
| 42 S | 2,4,6-Tribromophenol | 80.000 | 83.477 | -4.3 | 147 | 0.00 |
| 43 C | 2,4,6-Trichlorophenol | 40.000 | 44.249 | -10.6 | 158 | 0.00 |
| 44 | 2,4,5-Trichlorophenol | 40.000 | 44.396 | -11.0 | 157 | 0.00 |
| 45 S | 2-Fluorobiphenyl | 80.000 | 78.649 | 1.7 | 146 | 0.00 |
| 46 | 1,1'-Biphenyl | 40.000 | 39.542 | 1.1 | 147 | 0.00 |
| 47 | 2-Chloronaphthalene | 40.000 | 39.695 | 0.8 | 147 | 0.00 |

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081324\
 Data File : BP021489.D
 Acq On : 13 Aug 2024 19:03
 Operator : RC/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 ICVBP081324

Quant Time: Aug 14 00:10:09 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 14 00:00:30 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

| | Compound | Amount | Calc. | %Dev | Area% | Dev(min) |
|------|----------------------------|--------|--------|-------|-------|----------|
| 48 | 2-Nitroaniline | 40.000 | 43.044 | -7.6 | 172 | 0.00 |
| 49 | Acenaphthylene | 40.000 | 39.531 | 1.2 | 146 | 0.01 |
| 50 | Dimethylphthalate | 40.000 | 39.127 | 2.2 | 145 | 0.00 |
| 51 | 2,6-Dinitrotoluene | 40.000 | 41.424 | -3.6 | 162 | 0.00 |
| 52 C | Acenaphthene | 40.000 | 39.366 | 1.6 | 146 | 0.00 |
| 53 | 3-Nitroaniline | 40.000 | 40.483 | -1.2 | 156 | 0.00 |
| 54 P | 2,4-Dinitrophenol | 40.000 | 45.065 | -12.7 | 184 | 0.00 |
| 55 | Dibenzofuran | 40.000 | 38.543 | 3.6 | 143 | 0.00 |
| 56 P | 4-Nitrophenol | 40.000 | 39.582 | 1.0 | 152 | -0.01 |
| 57 | 2,4-Dinitrotoluene | 40.000 | 40.864 | -2.2 | 161 | 0.00 |
| 58 | Fluorene | 40.000 | 38.055 | 4.9 | 141 | 0.00 |
| 59 | 2,3,4,6-Tetrachlorophenol | 40.000 | 42.310 | -5.8 | 148 | 0.00 |
| 60 | Diethylphthalate | 40.000 | 37.752 | 5.6 | 139 | -0.01 |
| 61 | 4-Chlorophenyl-phenylether | 40.000 | 38.777 | 3.1 | 145 | 0.00 |
| 62 | 4-Nitroaniline | 40.000 | 38.254 | 4.4 | 141 | 0.00 |
| 63 | Azobenzene | 40.000 | 37.833 | 5.4 | 140 | 0.00 |
| 64 I | Phenanthrene-d10 | 20.000 | 20.000 | 0.0 | 133 | 0.00 |
| 65 | 4,6-Dinitro-2-methylphenol | 40.000 | 46.281 | -15.7 | 172 | 0.00 |
| 66 c | n-Nitrosodiphenylamine | 40.000 | 42.093 | -5.2 | 140 | 0.00 |
| 67 | 4-Bromophenyl-phenylether | 40.000 | 42.289 | -5.7 | 144 | 0.00 |
| 68 | Hexachlorobenzene | 40.000 | 41.161 | -2.9 | 136 | 0.00 |
| 69 | Atrazine | 40.000 | 38.851 | 2.9 | 125 | 0.00 |
| 70 C | Pentachlorophenol | 40.000 | 44.446 | -11.1 | 148 | 0.00 |
| 71 | Phenanthrene | 40.000 | 39.541 | 1.1 | 131 | 0.00 |
| 72 | Anthracene | 40.000 | 39.639 | 0.9 | 131 | 0.00 |
| 73 | Carbazole | 40.000 | 37.457 | 6.4 | 122 | 0.00 |
| 74 | Di-n-butylphthalate | 40.000 | 38.509 | 3.7 | 125 | 0.00 |
| 75 C | Fluoranthene | 40.000 | 36.288 | 9.3 | 119 | 0.00 |
| 76 I | Chrysene-d12 | 20.000 | 20.000 | 0.0 | 106 | 0.02 |
| 77 | Benzidine | 40.000 | 30.320 | 24.2 | 79 | 0.00 |
| 78 | Pyrene | 40.000 | 43.044 | -7.6 | 116 | 0.01 |
| 79 S | Terphenyl-d14 | 80.000 | 86.216 | -7.8 | 117 | 0.01 |
| 80 | Butylbenzylphthalate | 40.000 | 41.647 | -4.1 | 115 | 0.00 |
| 81 | Benzo(a)anthracene | 40.000 | 40.172 | -0.4 | 106 | 0.01 |
| 82 | 3,3'-Dichlorobenzidine | 40.000 | 40.936 | -2.3 | 102 | 0.00 |
| 83 | Chrysene | 40.000 | 39.759 | 0.6 | 105 | 0.01 |
| 84 | Bis(2-ethylhexyl)phthalate | 40.000 | 43.381 | -8.5 | 109 | 0.01 |
| 85 c | Di-n-octyl phthalate | 40.000 | 39.871 | 0.3 | 108 | 0.01 |
| 86 I | Perylene-d12 | 20.000 | 20.000 | 0.0 | 102 | 0.02 |
| 87 | Indeno(1,2,3-cd)pyrene | 40.000 | 41.105 | -2.8 | 102 | 0.00 |
| 88 | Benzo(b)fluoranthene | 40.000 | 40.381 | -1.0 | 102 | 0.01 |
| 89 | Benzo(k)fluoranthene | 40.000 | 39.984 | 0.0 | 100 | 0.01 |
| 90 C | Benzo(a)pyrene | 40.000 | 40.657 | -1.6 | 102 | 0.02 |
| 91 | Dibenzo(a,h)anthracene | 40.000 | 41.172 | -2.9 | 103 | 0.02 |
| 92 | Benzo(g,h,i)perylene | 40.000 | 41.532 | -3.8 | 104 | 0.03 |

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081324\
Data File : BP021489.D
Acq On : 13 Aug 2024 19:03
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
ICVBP081324

Quant Time: Aug 14 00:10:09 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 14 00:00:30 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

| Compound | Amount | Calc. | %Dev | Area% | Dev(min) |
|----------|--------|-------|------|-------|----------|
|----------|--------|-------|------|-------|----------|

(#) = Out of Range SPCC's out = 0 CCC's out = 0



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

7C

SEMOVOLATILE CONTINUING CALIBRATION CHECK

| | | | | | |
|-----------------|--------------------|-----------|------------------------|-------------------|-------------------|
| Lab Name: | <u>CHEMTECH</u> | | Contract: | <u>JAC005</u> | |
| Lab Code: | <u>CHEM</u> | Case No.: | <u>P3657</u> | SAS No.: | <u>P3657</u> |
| Instrument ID: | <u>BNA_P</u> | | Calibration Date/Time: | <u>08/20/2024</u> | <u>10:07</u> |
| Lab File ID: | <u>BP021566.D</u> | | Init. Calib. Date(s): | <u>08/13/2024</u> | <u>08/13/2024</u> |
| EPA Sample No.: | <u>SSTDCCCC040</u> | | Init. Calib. Time(s): | <u>12:32</u> | <u>17:36</u> |
| GC Column: | <u>ZB-GR</u> | ID: | <u>0.25</u> | (mm) | |

| COMPOUND | RRF | RRF040 | MIN RRF | %D | MAX%D |
|----------------------------|-------|--------|---------|-------|-------|
| Pyridine | 1.783 | 1.602 | | -10.2 | |
| 2-Fluorophenol | 1.347 | 1.353 | | 0.4 | |
| Benzaldehyde | 1.258 | 1.358 | | 7.9 | |
| Phenol-d6 | 1.965 | 1.932 | | -1.7 | |
| 2-Methylphenol | 1.286 | 1.272 | | -1.1 | |
| 3+4-Methylphenols | 1.734 | 1.797 | | 3.6 | |
| Nitrobenzene-d5 | 0.387 | 0.426 | | 10.1 | |
| Hexachloroethane | 0.599 | 0.597 | | -0.3 | |
| Nitrobenzene | 0.421 | 0.432 | | 2.6 | |
| Naphthalene | 1.045 | 1.030 | | -1.4 | |
| Hexachlorobutadiene | 0.215 | 0.208 | | -3.3 | 20.0 |
| 2-Methylnaphthalene | 0.713 | 0.693 | | -2.8 | |
| 2,4,6-Trichlorophenol | 0.326 | 0.378 | | 16.0 | 20.0 |
| 2-Fluorobiphenyl | 1.310 | 1.241 | | -5.3 | |
| 2,4,5-Trichlorophenol | 0.362 | 0.427 | | 18.0 | |
| Acenaphthylene | 1.619 | 1.589 | | -1.9 | |
| Acenaphthene | 1.063 | 1.082 | | 1.8 | 20.0 |
| Dibenzofuran | 1.635 | 1.602 | | -2.0 | |
| 2,4-Dinitrotoluene | 0.284 | 0.425 | | 49.6 | |
| Fluorene | 1.343 | 1.294 | | -3.6 | |
| 2,4,6-Tribromophenol | 0.223 | 0.267 | | 19.7 | |
| Hexachlorobenzene | 0.250 | 0.242 | | -3.2 | |
| Pentachlorophenol | 0.137 | 0.170 | | 24.1 | 20.0 |
| Phenanthrene | 0.962 | 0.951 | | -1.1 | |
| Anthracene | 0.943 | 0.947 | | 0.4 | |
| Carbazole | 0.909 | 0.938 | | 3.2 | |
| Di-n-butylphthalate | 1.116 | 1.121 | | 0.4 | |
| Fluoranthene | 1.180 | 1.227 | | 4.0 | 20.0 |
| Pyrene | 1.307 | 1.288 | | -1.5 | |
| Terphenyl-d14 | 1.029 | 0.954 | | -7.3 | |
| Benzo(a)anthracene | 1.256 | 1.256 | | 0.0 | |
| Chrysene | 1.185 | 1.188 | | 0.3 | |
| Bis(2-ethylhexyl)phthalate | 0.742 | 0.771 | | 3.9 | |
| Benzo(b)fluoranthene | 1.149 | 1.120 | | -2.5 | |
| Benzo(k)fluoranthene | 1.121 | 1.084 | | -3.3 | |
| Benzo(a)pyrene | 0.997 | 1.000 | | 0.3 | 20.0 |
| Indeno(1,2,3-cd)pyrene | 1.296 | 1.324 | | 2.2 | |
| Dibenzo(a,h)anthracene | 1.077 | 1.092 | | 1.4 | |
| Benzo(g,h,i)perylene | 1.080 | 1.120 | | 3.7 | |



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

7C

SEMOVOLATILE CONTINUING CALIBRATION CHECK

| | | | | | |
|-----------------|-------------------|-----------------|------------------------|-------------------|-------------------|
| Lab Name: | <u>CHEMTECH</u> | | Contract: | <u>JAC005</u> | |
| Lab Code: | <u>CHEM</u> | Case No.: | <u>P3657</u> | SAS No.: | <u>P3657</u> |
| Instrument ID: | <u>BNA_P</u> | | Calibration Date/Time: | <u>08/20/2024</u> | <u>10:07</u> |
| Lab File ID: | <u>BP021566.D</u> | | Init. Calib. Date(s): | <u>08/13/2024</u> | <u>08/13/2024</u> |
| EPA Sample No.: | <u>SSTDCCC040</u> | | Init. Calib. Time(s): | <u>12:32</u> | <u>17:36</u> |
| GC Column: | <u>ZB-GR</u> | ID: <u>0.25</u> | (mm) | | |

| COMPOUND | RRF | RRF040 | MIN RRF | %D | MAX%D |
|---------------------|-------|--------|---------|-------|-------|
| 1,4-Dioxane | 0.639 | 0.561 | | -12.2 | 20.0 |
| 1-Methylnaphthalene | 0.703 | 0.682 | | -3.0 | |

All other compounds must meet a minimum RRF of 0.010.

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP082024\
 Data File : BP021566.D
 Acq On : 20 Aug 2024 10:07
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040

Quant Time: Aug 20 11:20:14 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 14 00:00:30 2024
 Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/21/2024
 Supervised By :mohammad ahmed 08/22/2024

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|------------------------------------|--------|------|----------|--------|-------|----------|
| Internal Standards | | | | | | |
| 1) 1,4-Dichlorobenzene-d4 | 7.805 | 152 | 357434 | 20.000 | ng | 0.00 |
| 21) Naphthalene-d8 | 10.593 | 136 | 1454006 | 20.000 | ng | 0.00 |
| 39) Acenaphthene-d10 | 14.451 | 164 | 971251 | 20.000 | ng | 0.00 |
| 64) Phenanthrene-d10 | 17.263 | 188 | 2102943 | 20.000 | ng | 0.01 |
| 76) Chrysene-d12 | 21.727 | 240 | 2067038 | 20.000 | ng | 0.03 |
| 86) Perylene-d12 | 25.174 | 264 | 2510089 | 20.000 | ng | 0.07 |
| System Monitoring Compounds | | | | | | |
| 5) 2-Fluorophenol | 5.405 | 112 | 1934361 | 80.358 | ng | 0.01 |
| 7) Phenol-d6 | 6.975 | 99 | 2762029 | 78.656 | ng | 0.01 |
| 23) Nitrobenzene-d5 | 8.952 | 82 | 2476910 | 87.927 | ng | 0.00 |
| 42) 2,4,6-Tribromophenol | 15.963 | 330 | 1037187 | 95.950 | ng | 0.00 |
| 45) 2-Fluorobiphenyl | 13.063 | 172 | 4820142 | 75.740 | ng | 0.00 |
| 79) Terphenyl-d14 | 19.992 | 244 | 7890058 | 74.156 | ng | 0.02 |
| Target Compounds | | | | | | |
| | | | | Qvalue | | |
| 2) 1,4-Dioxane | 3.322 | 88 | 401142 | 35.145 | ng | 99 |
| 3) Pyridine | 3.717 | 79 | 1144919 | 35.932 | ng | 96 |
| 4) n-Nitrosodimethylamine | 3.622 | 42 | 351901 | 36.744 | ng | 93 |
| 6) Aniline | 7.134 | 93 | 1285899 | 36.630 | ng | 97 |
| 8) 2-Chlorophenol | 7.375 | 128 | 925647 | 41.611 | ng | 95 |
| 9) Benzaldehyde | 6.946 | 77 | 970805m | 43.178 | ng | |
| 10) Phenol | 6.999 | 94 | 1391597 | 38.270 | ng | 97 |
| 11) bis(2-Chloroethyl)ether | 7.228 | 93 | 1102315 | 35.826 | ng | 97 |
| 12) 1,3-Dichlorobenzene | 7.699 | 146 | 1028196 | 38.912 | ng | 96 |
| 13) 1,4-Dichlorobenzene | 7.840 | 146 | 1041774 | 39.044 | ng | 98 |
| 14) 1,2-Dichlorobenzene | 8.157 | 146 | 989860 | 38.497 | ng | 99 |
| 15) Benzyl Alcohol | 8.040 | 79 | 963573 | 38.244 | ng | 95 |
| 16) 2,2'-oxybis(1-Chloropr... | 8.328 | 45 | 988701 | 34.893 | ng | 98 |
| 17) 2-Methylphenol | 8.246 | 107 | 909170 | 39.549 | ng | 97 |
| 18) Hexachloroethane | 8.887 | 117 | 426491 | 39.861 | ng | 99 |
| 19) n-Nitroso-di-n-propyla... | 8.605 | 70 | 792638 | 32.677 | ng | 96 |
| 20) 3+4-Methylphenols | 8.569 | 107 | 1284831 | 41.450 | ng | 98 |
| 22) Acetophenone | 8.622 | 105 | 1697255 | 37.996 | ng | # 96 |
| 24) Nitrobenzene | 8.993 | 77 | 1256678 | 41.074 | ng | 93 |
| 25) Isophorone | 9.522 | 82 | 2333788 | 35.068 | ng | 98 |
| 26) 2-Nitrophenol | 9.704 | 139 | 480680 | 56.961 | ng | 95 |
| 27) 2,4-Dimethylphenol | 9.769 | 122 | 666408 | 40.576 | ng | 95 |
| 28) bis(2-Chloroethoxy)met... | 10.004 | 93 | 1468190 | 36.141 | ng | 100 |
| 29) 2,4-Dichlorophenol | 10.240 | 162 | 872962 | 44.986 | ng | 98 |
| 30) 1,2,4-Trichlorobenzene | 10.457 | 180 | 966653 | 38.855 | ng | 97 |
| 31) Naphthalene | 10.646 | 128 | 2995089 | 39.431 | ng | 100 |
| 32) Benzoic acid | 9.899 | 122 | 796276 | 64.703 | ng | 97 |
| 33) 4-Chloroaniline | 10.746 | 127 | 1156545 | 40.535 | ng | 96 |
| 34) Hexachlorobutadiene | 10.934 | 225 | 604655 | 38.600 | ng | 99 |
| 35) Caprolactam | 11.522 | 113 | 378013 | 46.800 | ng | 91 |
| 36) 4-Chloro-3-methylphenol | 11.875 | 107 | 1181215 | 43.882 | ng | 97 |
| 37) 2-Methylnaphthalene | 12.263 | 142 | 2014464 | 38.859 | ng | 99 |
| 38) 1-Methylnaphthalene | 12.481 | 142 | 1982500 | 38.775 | ng | 100 |
| 40) 1,2,4,5-Tetrachloroben... | 12.628 | 216 | 1085575 | 38.147 | ng | 100 |
| 41) Hexachlorocyclopentadiene | 12.616 | 237 | 456339 | 50.596 | ng | 98 |
| 43) 2,4,6-Trichlorophenol | 12.869 | 196 | 734554 | 46.468 | ng | 99 |

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP082024\
 Data File : BP021566.D
 Acq On : 20 Aug 2024 10:07
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040

Quant Time: Aug 20 11:20:14 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 14 00:00:30 2024
 Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/21/2024
 Supervised By :mohammad ahmed 08/22/2024

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|-------------------------------|--------|------|----------|--------|-------|----------|
| 44) 2,4,5-Trichlorophenol | 12.940 | 196 | 829555 | 47.232 | ng | 95 |
| 46) 1,1'-Biphenyl | 13.275 | 154 | 2616716 | 38.080 | ng | 98 |
| 47) 2-Chloronaphthalene | 13.316 | 162 | 2065003 | 38.599 | ng | 99 |
| 48) 2-Nitroaniline | 13.510 | 65 | 704708 | 47.106 | ng | # 86 |
| 49) Acenaphthylene | 14.169 | 152 | 3086915 | 39.251 | ng | 99 |
| 50) Dimethylphthalate | 13.904 | 163 | 2580187 | 39.669 | ng | 99 |
| 51) 2,6-Dinitrotoluene | 14.010 | 165 | 594069 | 45.498 | ng | 92 |
| 52) Acenaphthene | 14.522 | 154 | 2102181 | 40.738 | ng | 99 |
| 53) 3-Nitroaniline | 14.339 | 138 | 627496 | 48.589 | ng | # 95 |
| 54) 2,4-Dinitrophenol | 14.563 | 184 | 267685 | 65.093 | ng | 88 |
| 55) Dibenzofuran | 14.863 | 168 | 3112579 | 39.211 | ng | 99 |
| 56) 4-Nitrophenol | 14.657 | 139 | 556879 | 53.424 | ng | 88 |
| 57) 2,4-Dinitrotoluene | 14.816 | 165 | 825714 | 49.320 | ng | 88 |
| 58) Fluorene | 15.528 | 166 | 2513424 | 38.542 | ng | 100 |
| 59) 2,3,4,6-Tetrachlorophenol | 15.087 | 232 | 716552 | 48.067 | ng | 99 |
| 60) Diethylphthalate | 15.292 | 149 | 2656153 | 39.800 | ng | 100 |
| 61) 4-Chlorophenyl-phenyle... | 15.522 | 204 | 1250366 | 37.093 | ng | 99 |
| 62) 4-Nitroaniline | 15.534 | 138 | 654363 | 49.265 | ng | 91 |
| 63) Azobenzene | 15.816 | 77 | 2864106 | 34.913 | ng | 97 |
| 65) 4,6-Dinitro-2-methylph... | 15.598 | 198 | 403578 | 58.939 | ng | 95 |
| 66) n-Nitrosodiphenylamine | 15.734 | 169 | 2179381 | 38.672 | ng | 100 |
| 67) 4-Bromophenyl-phenylether | 16.434 | 248 | 835316 | 37.531 | ng | 99 |
| 68) Hexachlorobenzene | 16.557 | 284 | 1019155 | 38.748 | ng | 99 |
| 69) Atrazine | 16.710 | 200 | 638902 | 33.020 | ng | 99 |
| 70) Pentachlorophenol | 16.910 | 266 | 714044 | 49.739 | ng | 99 |
| 71) Phenanthrene | 17.310 | 178 | 3999697 | 39.548 | ng | 100 |
| 72) Anthracene | 17.404 | 178 | 3981334 | 40.140 | ng | 100 |
| 73) Carbazole | 17.675 | 167 | 3947153 | 41.294 | ng | 99 |
| 74) Di-n-butylphthalate | 18.280 | 149 | 4715211 | 40.166 | ng | 100 |
| 75) Fluoranthene | 19.386 | 202 | 5158752 | 41.572 | ng | 99 |
| 77) Benzidine | 19.580 | 184 | 2092255 | 43.787 | ng | 100 |
| 78) Pyrene | 19.763 | 202 | 5322876 | 39.398 | ng | 100 |
| 80) Butylbenzylphthalate | 20.716 | 149 | 2221372 | 43.672 | ng | 95 |
| 81) Benzo(a)anthracene | 21.710 | 228 | 5190787 | 40.003 | ng | 100 |
| 82) 3,3'-Dichlorobenzidine | 21.627 | 252 | 1922208 | 46.202 | ng | 100 |
| 83) Chrysene | 21.780 | 228 | 4912854 | 40.101 | ng | 100 |
| 84) Bis(2-ethylhexyl)phtha... | 21.663 | 149 | 3185624 | 41.537 | ng | 100 |
| 85) Di-n-octyl phthalate | 22.986 | 149 | 5659478 | 42.612 | ng | 97 |
| 87) Indeno(1,2,3-cd)pyrene | 29.103 | 276 | 6648231m | 40.860 | ng | |
| 88) Benzo(b)fluoranthene | 24.086 | 252 | 5622632 | 39.003 | ng | 99 |
| 89) Benzo(k)fluoranthene | 24.163 | 252 | 5443792 | 38.710 | ng | 99 |
| 90) Benzo(a)pyrene | 25.004 | 252 | 5022273 | 40.145 | ng | 99 |
| 91) Dibenzo(a,h)anthracene | 29.192 | 278 | 5480813 | 40.541 | ng | 100 |
| 92) Benzo(g,h,i)perylene | 30.309 | 276 | 5623792 | 41.504 | ng | 99 |

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP082024\
 Data File : BP021566.D
 Acq On : 20 Aug 2024 10:07
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1


Quant Time: Aug 20 11:20:14 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 14 00:00:30 2024
 Response via : Initial Calibration

Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 08/21/2024
 Supervised By :mohammad ahmed 08/22/2024







#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.805 min Scan# 8
Delta R.T. -0.006 min
Lab File: BP021566.D
Acq: 20 Aug 2024 10:07

Instrument : BNA_P
ClientSampleId : SSTDCCC040


Manual Integrations APPROVED




Reviewed By :Yogesh Patel 08/21/2024
Supervised By :mohammad ahmed 08/22/2024



#2
1,4-Dioxane
Concen: 35.145 ng
RT: 3.322 min Scan# 57
Delta R.T. 0.006 min
Lab File: BP021566.D
Acq: 20 Aug 2024 10:07

Tgt Ion: 88 Resp: 401142
Ion Ratio Lower Upper
88 100
58 61.6 48.7 73.1
43 19.8 15.4 23.2









#3

Pyridine
Concen: 35.932 ng
RT: 3.717 min Scan# 1
Delta R.T. 0.012 min
Lab File: BP021566.D
Acq: 20 Aug 2024 10:07

Instrument : BNA_P
ClientSampleId : SSTDCCC040

Manual Integrations APPROVED


Reviewed By :Yogesh Patel 08/21/2024
Supervised By :mohammad ahmed 08/22/2024






#4

n-Nitrosodimethylamine
Concen: 36.744 ng
RT: 3.622 min Scan# 108
Delta R.T. 0.006 min
Lab File: BP021566.D
Acq: 20 Aug 2024 10:07

Tgt Ion: 42 Resp: 351901
Ion Ratio Lower Upper
42 100
74 182.0 154.3 231.5
44 6.8 6.2 9.4





#5

2-Fluorophenol

Concen: 80.358 ng

RT: 5.405 min Scan# 4

Delta R.T. 0.012 min

Lab File: BP021566.D

Acq: 20 Aug 2024 10:07

Instrument :

BNA_P

ClientSampleId :

SSTDCCC040

Tgt Ion:112 Resp: 193436

Ion Ratio Lower Upper

112 100

64 54.5 49.0 73.6





63 27.8 24.6 36.8

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 08/21/2024

Supervised By :mohammad ahmed 08/22/2024



#6
Aniline
Concen: 36.630 ng
RT: 7.134 min Scan# 705
Delta R.T. -0.000 min
Lab File: BP021566.D
Acq: 20 Aug 2024 10:07


Tgt Ion: 93 Resp: 1285899




Ion Ratio Lower Upper

93 100

66 35.7 30.2 45.2

65 17.3 14.7 22.1









#7
Phenol-d6
Concen: 78.656 ng
RT: 6.975 min Scan# 6
Delta R.T. 0.012 min
Lab File: BP021566.D
Acq: 20 Aug 2024 10:07

Instrument : BNA_P
ClientSampleId : SSTDCCC040


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 08/21/2024
Supervised By :mohammad ahmed 08/22/2024



#8
2-Chlorophenol
Concen: 41.611 ng
RT: 7.375 min Scan# 746
Delta R.T. -0.000 min
Lab File: BP021566.D
Acq: 20 Aug 2024 10:07

Tgt Ion:128 Resp: 925647
Ion Ratio Lower Upper
128 100
130 31.3 11.6 51.6
64 49.4 35.3 75.3







#9
Benzaldehyde
Concen: 43.178 ng m
RT: 6.946 min Scan# 6
Delta R.T. -0.000 min
Lab File: BP021566.D
Acq: 20 Aug 2024 10:07

Instrument : BNA_P
ClientSampleId : SSTDCCC040


Manual Integrations
APPROVED







Reviewed By :Yogesh Patel 08/21/2024
Supervised By :mohammad ahmed 08/22/2024



#10
Phenol
Concen: 38.270 ng
RT: 6.999 min Scan# 682
Delta R.T. 0.012 min
Lab File: BP021566.D
Acq: 20 Aug 2024 10:07

Tgt Ion: 94 Resp: 1391597
Ion Ratio Lower Upper
94 100
65 25.4 6.5 46.5
66 32.7 14.7 54.7






#11
bis(2-Chloroethyl)ether
Concen: 35.826 ng
RT: 7.228 min Scan# 7
Delta R.T. -0.000 min
Lab File: BP021566.D
Acq: 20 Aug 2024 10:07

Instrument : BNA_P
ClientSampleId : SSTDCCC040


Manual Integrations APPROVED







Reviewed By :Yogesh Patel 08/21/2024
Supervised By :mohammad ahmed 08/22/2024



#12
1,3-Dichlorobenzene
Concen: 38.912 ng
RT: 7.699 min Scan# 801
Delta R.T. -0.000 min
Lab File: BP021566.D
Acq: 20 Aug 2024 10:07

Tgt Ion:146 Resp: 1028196
Ion Ratio Lower Upper
146 100
148 63.6 51.8 77.6
75 33.7 31.3 46.9





#13

1,4-Dichlorobenzene

Concen: 39.044 ng

RT: 7.840 min Scan# 8

Delta R.T. -0.000 min

Lab File: BP021566.D

Acq: 20 Aug 2024 10:07

Instrument :

BNA_P


ClientSampleId :

SSTDCCC040

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/21/2024

Supervised By :mohammad ahmed 08/22/2024



#14

1,2-Dichlorobenzene

Concen: 38.497 ng

RT: 8.157 min Scan# 879

Delta R.T. -0.000 min

Lab File: BP021566.D

Acq: 20 Aug 2024 10:07


Tgt Ion:146 Resp: 989860


Ion Ratio Lower Upper

146 100

148 64.0 51.5 77.3

111 50.4 41.1 61.7







#15

Benzyl Alcohol
Concen: 38.244 ng
RT: 8.040 min Scan# 8
Delta R.T. 0.006 min
Lab File: BP021566.D
Acq: 20 Aug 2024 10:07

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Manual Integrations APPROVED


Reviewed By :Yogesh Patel 08/21/2024
Supervised By :mohammad ahmed 08/22/2024






#16

2,2'-oxybis(1-Chloropropane)
Concen: 34.893 ng
RT: 8.328 min Scan# 908
Delta R.T. -0.006 min
Lab File: BP021566.D
Acq: 20 Aug 2024 10:07

Tgt Ion: 45 Resp: 988701
Ion Ratio Lower Upper
45 100
77 20.5 1.9 41.9
79 15.8 0.0 36.6





#17

2-Methylphenol

Concen: 39.549 ng

RT: 8.246 min Scan# 8

Delta R.T. 0.012 min

Lab File: BP021566.D

Acq: 20 Aug 2024 10:07

Instrument :

BNA_P




ClientSampleId :

SSTDCCC040

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/21/2024

Supervised By :mohammad ahmed 08/22/2024



#18

Hexachloroethane

Concen: 39.861 ng

RT: 8.887 min Scan# 1003

Delta R.T. -0.006 min

Lab File: BP021566.D

Acq: 20 Aug 2024 10:07


Tgt Ion:117 Resp: 426491




Ion Ratio Lower Upper

117 100

119 93.1 73.9 110.9

201 95.3 74.9 112.3





#19

n-Nitroso-di-n-propylamine
Concen: 32.677 ng
RT: 8.605 min Scan# 949
Delta R.T. -0.000 min
Lab File: BP021566.D
Acq: 20 Aug 2024 10:07

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040





Manual Integrations APPROVED

Reviewed By :Yogesh Patel 08/21/2024
Supervised By :mohammad ahmed 08/22/2024

Tgt Ion: 70 Resp: 792638

| | | Lower | Upper |
|-----|------|-------|-------|
| 70 | 100 | | |
| 42 | 43.6 | 33.1 | 49.7 |
| 101 | 10.7 | 7.4 | 11.2 |
| 130 | 18.9 | 13.3 | 19.9 |

Abundance




#20


3+4-Methylphenols
Concen: 41.450 ng
RT: 8.569 min Scan# 949
Delta R.T. 0.012 min
Lab File: BP021566.D
Acq: 20 Aug 2024 10:07

Tgt Ion:107 Resp: 1284831

| | | Lower | Upper |
|-----|------|-------|-------|
| 107 | 100 | | |
| 108 | 88.7 | 67.6 | 107.6 |
| 77 | 32.7 | 15.0 | 55.0 |
| 79 | 28.9 | 8.9 | 48.9 |

Abundance







#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.593 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP021566.D
Acq: 20 Aug 2024 10:07

Instrument : BNA_P
ClientSampleId : SSTDCCC040


Manual Integrations APPROVED




Reviewed By :Yogesh Patel 08/21/2024
Supervised By :mohammad ahmed 08/22/2024



#22
Acetophenone
Concen: 37.996 ng
RT: 8.622 min Scan# 958
Delta R.T. 0.006 min
Lab File: BP021566.D
Acq: 20 Aug 2024 10:07

Tgt Ion:105 Resp: 1697255
Ion Ratio Lower Upper
105 100
71 2.5 4.6 7.0#
51 18.5 16.3 24.5
120 20.8 16.2 24.2





#23

Nitrobenzene-d5

Concen: 87.927 ng

RT: 8.952 min Scan# 1

Delta R.T. -0.000 min

Lab File: BP021566.D

Acq: 20 Aug 2024 10:07

Instrument :

BNA_P





ClientSampleId :

SSTDCCC040

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/21/2024

Supervised By :mohammad ahmed 08/22/2024



#24

Nitrobenzene

Concen: 41.074 ng

RT: 8.993 min Scan# 1021

Delta R.T. -0.000 min

Lab File: BP021566.D

Acq: 20 Aug 2024 10:07


Tgt Ion: 77 Resp: 1256678







Ion Ratio Lower Upper

77 100


123 42.1 29.5 44.3

65 12.6 9.9 14.9









#25

Isophorone
Concen: 35.068 ngRT: 9.522 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP021566.D
Acq: 20 Aug 2024 10:07Instrument : BNA_P
ClientSampleId : SSTDCCC040**Manual Integrations
APPROVED**Reviewed By :Yogesh Patel 08/21/2024
Supervised By :mohammad ahmed 08/22/2024

#26

2-Nitrophenol
Concen: 56.961 ng
RT: 9.704 min Scan# 1142
Delta R.T. 0.005 min
Lab File: BP021566.D
Acq: 20 Aug 2024 10:07Tgt Ion:139 Resp: 480680
Ion Ratio Lower Upper
139 100
109 34.9 28.9 43.3
65 55.8 49.0 73.6



#27

2,4-Dimethylphenol

Concen: 40.576 ng

RT: 9.769 min Scan# 1

Delta R.T. 0.012 min

Lab File: BP021566.D

Acq: 20 Aug 2024 10:07

Instrument :

BNA_P

ClientSampleId :

SSTDCCC040

Tgt Ion:122 Resp: 666403

Ion Ratio Lower Upper

122 100




107 125.8 106.2 159.4

121 58.2 48.8 73.2

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/21/2024

Supervised By :mohammad ahmed 08/22/2024



#28

bis(2-Chloroethoxy)methane

Concen: 36.141 ng

RT: 10.004 min Scan# 1193

Delta R.T. -0.000 min

Lab File: BP021566.D

Acq: 20 Aug 2024 10:07


Tgt Ion: 93 Resp: 1468190




Ion Ratio Lower Upper

93 100

95 32.7 26.2 39.4

123 9.6 7.6 11.4





#29

2,4-Dichlorophenol

Concen: 44.986 ng

RT: 10.240 min Scan# 1

Delta R.T. -0.000 min

Lab File: BP021566.D

Acq: 20 Aug 2024 10:07





Instrument : BNA_P

ClientSampleId : SSTDCCC040

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/21/2024

Supervised By :mohammad ahmed 08/22/2024



#30

1,2,4-Trichlorobenzene

Concen: 38.855 ng

RT: 10.457 min Scan# 1270

Delta R.T. -0.000 min

Lab File: BP021566.D

Acq: 20 Aug 2024 10:07


Tgt Ion:180 Resp: 966653

Ion Ratio Lower Upper

180 100

182 94.7 78.3 117.5

145 30.2 24.0 36.0



#31

Naphthalene

Concen: 39.431 ng

RT: 10.646 min Scan# 1

Delta R.T. -0.000 min

Lab File: BP021566.D

Acq: 20 Aug 2024 10:07

Instrument :

BNA_P

ClientSampleId :

SSTDCCC040

Tgt Ion:128 Resp: 2995089

Ion Ratio Lower Upper

128 100

129 10.8 8.6 12.8

127 13.1 10.6 15.8

Manual Integrations


APPROVED

Reviewed By :Yogesh Patel 08/21/2024

Supervised By :mohammad ahmed 08/22/2024

Abundance

10.646



#32

Benzoic acid

Concen: 64.703 ng

RT: 9.899 min Scan# 1175

Delta R.T. 0.024 min

Lab File: BP021566.D

Acq: 20 Aug 2024 10:07

Tgt Ion:122 Resp: 796276

Ion Ratio Lower Upper




122 100

105 143.8 124.6 164.6

77 108.7 96.2 136.2

Abundance

9.899



#33

4-Chloroaniline

Concen: 40.535 ng

RT: 10.746 min Scan# 1

Delta R.T. 0.006 min

Lab File: BP021566.D




Acq: 20 Aug 2024 10:07

Instrument :

BNA_P

ClientSampleId :

SSTDCCC040



Tgt Ion:127 Resp: 115654




Ion Ratio Lower Upper

| | | | |
|-----|------|------|------|
| 127 | 100 | | |
| 129 | 31.7 | 24.5 | 36.7 |
| 65 | 31.9 | 28.9 | 43.3 |
| 92 | 22.9 | 19.3 | 28.9 |

Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel 08/21/2024

Supervised By :mohammad ahmed 08/22/2024




#34
Hexachlorobutadiene
Concen: 38.600 ng
RT: 10.934 min Scan# 1351
Delta R.T. -0.006 min
Lab File: BP021566.D
Acq: 20 Aug 2024 10:07

Tgt Ion:225 Resp: 604655

Ion Ratio Lower Upper

| | | | |
|-----|------|------|------|
| 225 | 100 | | |
| 223 | 63.0 | 49.6 | 74.4 |
| 227 | 62.9 | 50.1 | 75.1 |




Abundance Scan 1351 (10.934 min): BP021566.D\data.ms (-)

Sub

Abundance Scan 1351 (10.934 min): BP021566.D\data.ms (-)

m/z-->





#35
Caprolactam
Concen: 46.800 ng
RT: 11.522 min Scan# 14
Delta R.T. 0.023 min
Lab File: BP021566.D
Acq: 20 Aug 2024 10:07

Instrument :
BNA_P
ClientSampleId :
SSTDCCCC040

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/21/2024
Supervised By :mohammad ahmed 08/22/2024







#36
4-Chloro-3-methylphenol
Concen: 43.882 ng
RT: 11.875 min Scan# 1511
Delta R.T. 0.012 min
Lab File: BP021566.D
Acq: 20 Aug 2024 10:07

```

Tgt Ion:107 Resp: 1181215
Ion Ratio Lower Upper
107 100
144 21.3 16.6 24.8
142 66.0 51.0 76.4

```





#37

2-Methylnaphthalene

Concen: 38.859 ng

RT: 12.263 min Scan# 1

Delta R.T. -0.000 min

Lab File: BP021566.D

Acq: 20 Aug 2024 10:07

Instrument :

BNA_P

ClientSampleId :

SSTDCCC040

Tgt Ion:142 Resp: 201446

Ion Ratio Lower Upper

142 100

141 88.6 71.7 107.5




115 38.5 31.4 47.2

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 08/21/2024

Supervised By :mohammad ahmed 08/22/2024



#38

1-Methylnaphthalene

Concen: 38.775 ng

RT: 12.481 min Scan# 1614

Delta R.T. -0.000 min

Lab File: BP021566.D

Acq: 20 Aug 2024 10:07


Tgt Ion:142 Resp: 1982500


Ion Ratio Lower Upper

142 100


141 92.3 73.5 110.3

116 4.3 3.4 5.2






Abundance Scan 1949 (14.451 min): BP021566.D\data.ms (-)



Abundance Scan 1949 (14.451 min): BP021566.D\data.ms (-)



#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 14.451 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP021566.D

Acq: 20 Aug 2024 10:07

Instrument :

BNA_P


ClientSampleId :

SSTDCCC040


**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/21/2024


Supervised By :mohammad ahmed 08/22/2024



Abundance Scan 1639 (12.628 min): BP021566.D\data.ms (-)



Abundance Scan 1639 (12.628 min): BP021566.D\data.ms (-)



#40

1,2,4,5-Tetrachlorobenzene

Concen: 38.147 ng

RT: 12.628 min Scan# 1639

Delta R.T. -0.000 min

Lab File: BP021566.D

Acq: 20 Aug 2024 10:07

Tgt Ion:216 Resp: 1085575

Ion Ratio Lower Upper


216 100


214 78.1 62.5 93.7

179 20.3 16.4 24.6

108 20.5 16.7 25.1

Abundance







#41
Hexachlorocyclopentadiene
Concen: 50.596 ng
RT: 12.616 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP021566.D
Acq: 20 Aug 2024 10:07

Instrument : BNA_P
ClientSampleId : SSTDCCC040


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/21/2024
Supervised By :mohammad ahmed 08/22/2024



#42
2,4,6-Tribromophenol
Concen: 95.950 ng
RT: 15.963 min Scan# 2206
Delta R.T. -0.006 min
Lab File: BP021566.D
Acq: 20 Aug 2024 10:07

Tgt Ion:330 Resp: 1037187
Ion Ratio Lower Upper
330 100
332 96.9 77.4 116.0
141 29.9 24.5 36.7







#43
2,4,6-Trichlorophenol
Concen: 46.468 ng
RT: 12.869 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP021566.D
Acq: 20 Aug 2024 10:07

Instrument : BNA_P
ClientSampleId : SSTDCCC040


Manual Integrations
APPROVED




Reviewed By :Yogesh Patel 08/21/2024
Supervised By :mohammad ahmed 08/22/2024



#44
2,4,5-Trichlorophenol
Concen: 47.232 ng
RT: 12.940 min Scan# 1692
Delta R.T. 0.006 min
Lab File: BP021566.D
Acq: 20 Aug 2024 10:07

Tgt Ion:196 Resp: 829555
Ion Ratio Lower Upper
196 100
198 94.8 78.9 118.3
97 54.9 49.4 74.0
132 25.7 21.1 31.7





#45

2-Fluorobiphenyl

Concen: 75.740 ng

RT: 13.063 min Scan# 1

Delta R.T. -0.000 min

Lab File: BP021566.D

Acq: 20 Aug 2024 10:07

Instrument:

BNA_P




ClientSampleId :

SSTDCCC040

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/21/2024

Supervised By :mohammad ahmed 08/22/2024



#46

1,1'-Biphenyl

Concen: 38.080 ng

RT: 13.275 min Scan# 1749

Delta R.T. -0.006 min

Lab File: BP021566.D

Acq: 20 Aug 2024 10:07


Tgt Ion:154 Resp: 2616716

Ion Ratio Lower Upper

154 100

153 40.3 20.7 60.7




76 15.8 0.0 37.7



Abundance Scan 1749 (13.275 min): BP021566.D\data.ms (-)

Sub

m/z--> 50 100 150 200 250 300 350 400 450 500



#47

2-Chloronaphthalene

Concen: 38.599 ng

RT: 13.316 min Scan# 1

Delta R.T. -0.000 min

Lab File: BP021566.D

Acq: 20 Aug 2024 10:07

Instrument :

BNA_P





ClientSampleId :

SSTDCCC040

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/21/2024

Supervised By :mohammad ahmed 08/22/2024



#48

2-Nitroaniline

Concen: 47.106 ng

RT: 13.510 min Scan# 1789

Delta R.T. 0.006 min

Lab File: BP021566.D

Acq: 20 Aug 2024 10:07


Tgt Ion: 65 Resp: 704708


Ion Ratio Lower Upper

65 100

92 79.2 58.1 87.1

138 99.1 65.5 98.3#







#49
 Acenaphthylene
 Concen: 39.251 ng
 RT: 14.169 min Scan# 1
 Delta R.T. -0.000 min
 Lab File: BP021566.D
 Acq: 20 Aug 2024 10:07

Instrument : BNA_P
 ClientSampleId : SSTDCCC040


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/21/2024
 Supervised By :mohammad ahmed 08/22/2024



#50
 Dimethylphthalate
 Concen: 39.669 ng
 RT: 13.904 min Scan# 1856
 Delta R.T. -0.006 min
 Lab File: BP021566.D
 Acq: 20 Aug 2024 10:07

Tgt Ion:163 Resp: 2580187
 Ion Ratio Lower Upper
 163 100
 194 4.1 3.3 4.9
 164 10.0 8.3 12.5







#51
2,6-Dinitrotoluene
Concen: 45.498 ng
RT: 14.010 min Scan# 1
Delta R.T. -0.012 min
Lab File: BP021566.D
Acq: 20 Aug 2024 10:07

Instrument : BNA_P
ClientSampleId : SSTDCCC040


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 08/21/2024
Supervised By :mohammad ahmed 08/22/2024



#52
Acenaphthene
Concen: 40.738 ng
RT: 14.522 min Scan# 1961
Delta R.T. -0.000 min
Lab File: BP021566.D
Acq: 20 Aug 2024 10:07

Tgt Ion:154 Resp: 2102181
Ion Ratio Lower Upper
154 100
153 112.5 91.1 136.7
152 54.1 43.0 64.4







#53
3-Nitroaniline
Concen: 48.589 ng
RT: 14.339 min Scan# 1
Delta R.T. -0.012 min
Lab File: BP021566.D
Acq: 20 Aug 2024 10:07

Instrument : BNA_P
ClientSampleId : SSTDCCC040


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/21/2024
Supervised By :mohammad ahmed 08/22/2024



#54
2,4-Dinitrophenol
Concen: 65.093 ng
RT: 14.563 min Scan# 1968
Delta R.T. 0.012 min
Lab File: BP021566.D
Acq: 20 Aug 2024 10:07

Tgt Ion:184 Resp: 267685
Ion Ratio Lower Upper
184 100
63 65.1 64.9 97.3
154 61.3 52.1 78.1






#55
Dibenzofuran
Concen: 39.211 ng
RT: 14.863 min Scan# 2
Delta R.T. 0.006 min
Lab File: BP021566.D
Acq: 20 Aug 2024 10:07

Instrument : BNA_P
ClientSampleId : SSTDCCC040


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/21/2024
Supervised By :mohammad ahmed 08/22/2024



#56
4-Nitrophenol
Concen: 53.424 ng
RT: 14.657 min Scan# 1984
Delta R.T. 0.006 min
Lab File: BP021566.D
Acq: 20 Aug 2024 10:07

Tgt Ion:139 Resp: 556879
Ion Ratio Lower Upper
139 100
109 83.4 70.2 110.2
65 103.7 101.8 141.8







#57
2,4-Dinitrotoluene
Concen: 49.320 ng
RT: 14.816 min Scan# 2130
Delta R.T. 0.006 min
Lab File: BP021566.D
Acq: 20 Aug 2024 10:07

Instrument : BNA_P
ClientSampleId : SSTDCCC040


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/21/2024
Supervised By :mohammad ahmed 08/22/2024



#58
Fluorene
Concen: 38.542 ng
RT: 15.528 min Scan# 2132
Delta R.T. 0.012 min
Lab File: BP021566.D
Acq: 20 Aug 2024 10:07

Tgt Ion:166 Resp: 2513424
Ion Ratio Lower Upper
166 100
165 97.8 78.1 117.1
167 13.1 11.0 16.4







#59
2,3,4,6-Tetrachlorophenol
Concen: 48.067 ng
RT: 15.087 min Scan# 2
Delta R.T. -0.000 min
Lab File: BP021566.D
Acq: 20 Aug 2024 10:07

Instrument : BNA_P
ClientSampleId : SSTDCCC040


Manual Integrations APPROVED




Reviewed By :Yogesh Patel 08/21/2024
Supervised By :mohammad ahmed 08/22/2024



#60
Diethylphthalate
Concen: 39.800 ng
RT: 15.292 min Scan# 2092
Delta R.T. -0.012 min
Lab File: BP021566.D
Acq: 20 Aug 2024 10:07

Tgt Ion:149 Resp: 2656153
Ion Ratio Lower Upper
149 100
177 22.4 17.8 26.8
150 12.6 10.1 15.1









#61
4-Chlorophenyl-phenylether
Concen: 37.093 ng
RT: 15.522 min Scan# 2131
Delta R.T. -0.000 min
Lab File: BP021566.D
Acq: 20 Aug 2024 10:07

Instrument : BNA_P
ClientSampleId : SSTDCCC040


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 08/21/2024
Supervised By :mohammad ahmed 08/22/2024



#62
4-Nitroaniline
Concen: 49.265 ng
RT: 15.534 min Scan# 2133
Delta R.T. 0.006 min
Lab File: BP021566.D
Acq: 20 Aug 2024 10:07

Tgt Ion:138 Resp: 654363
Ion Ratio Lower Upper
138 100
92 58.4 46.4 86.4
108 85.2 73.2 113.2







#63
Azobenzene
Concen: 34.913 ng
RT: 15.816 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP021566.D
Acq: 20 Aug 2024 10:07

Instrument : BNA_P
ClientSampleId : SSTDCCC040


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/21/2024
Supervised By :mohammad ahmed 08/22/2024



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.263 min Scan# 2427
Delta R.T. 0.012 min
Lab File: BP021566.D
Acq: 20 Aug 2024 10:07

Tgt Ion:188 Resp: 2102943
Ion Ratio Lower Upper
188 100
94 12.3 10.2 15.4
80 13.5 11.6 17.4






#65
4,6-Dinitro-2-methylphenol
Concen: 58.939 ng
RT: 15.598 min Scan# 2142
Delta R.T. 0.011 min
Lab File: BP021566.D
Acq: 20 Aug 2024 10:07

Instrument : BNA_P
ClientSampleId : SSTDCCC040



Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/21/2024
Supervised By :mohammad ahmed 08/22/2024



#66
n-Nitrosodiphenylamine
Concen: 38.672 ng
RT: 15.734 min Scan# 2167
Delta R.T. -0.006 min
Lab File: BP021566.D
Acq: 20 Aug 2024 10:07

Tgt Ion:169 Resp: 2179381
Ion Ratio Lower Upper
169 100
168 66.6 53.4 80.2
167 36.2 28.7 43.1




#67
 4-Bromophenyl-phenylether
 Concen: 37.531 ng
 RT: 16.434 min Scan# 2
 Delta R.T. -0.006 min
 Lab File: BP021566.D
 Acq: 20 Aug 2024 10:07

Instrument : BNA_P
 ClientSampleId : SSTDCCC040

Tgt Ion:248 Resp: 835310
 Ion Ratio Lower Upper
 248 100
 250 99.6 78.6 117.8
 141 68.6 53.9 80.9




Manual Integrations APPROVED

Reviewed By :Yogesh Patel 08/21/2024
 Supervised By :mohammad ahmed 08/22/2024



#68
 Hexachlorobenzene
 Concen: 38.748 ng
 RT: 16.557 min Scan# 2307
 Delta R.T. 0.006 min
 Lab File: BP021566.D
 Acq: 20 Aug 2024 10:07

Tgt Ion:284 Resp: 1019155
 Ion Ratio Lower Upper
 284 100
 142 29.9 23.4 35.2
 249 30.2 24.1 36.1



#69

Atrazine

Concen: 33.020 ng

RT: 16.710 min Scan# 2

Delta R.T. -0.006 min

Lab File: BP021566.D

Acq: 20 Aug 2024 10:07

Instrument :

BNA_P

ClientSampleId :

SSTDCCC040

Tgt Ion:200 Resp: 63890

Ion Ratio Lower Upper

200 100

173 26.0 6.2 46.2

215 49.6 28.4 68.4

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 08/21/2024

Supervised By :mohammad ahmed 08/22/2024

Abundance

Scan 2334 (16.716 min): BP021485.D\data.ms (-)

200.1

m/z-->

50 100 150 200 250 300 350 400 450 500

Abundance

Scan 2333 (16.710 min): BP021566.D\data.ms (-)

200.1

m/z-->

50 100 150 200 250 300 350 400 450 500

Abundance

Scan 2333 (16.710 min): BP021566.D\data.ms (-)

200.1

m/z-->

50 100 150 200 250 300 350 400 450 500

Sub

58.1

132.0

287.7354.4422.2 526.1

m/z-->

50 100 150 200 250 300 350 400 450 500

Sub

58.1

131.9

269.3 384.0452.0519.8

m/z-->

50 100 150 200 250 300 350 400 450 500

Abundance

Scan 2364 (16.892 min): BP021485.D\data.ms (-)

265.9

m/z-->

50 100 150 200 250 300 350 400 450 500

Abundance

Scan 2367 (16.910 min): BP021566.D\data.ms (-)

265.9

m/z-->

50 100 150 200 250 300 350 400 450 500

Sub

58.1

164.9

332.9398.3 521.2

m/z-->

50 100 150 200 250 300 350 400 450 500

Abundance

Scan 2367 (16.910 min): BP021566.D\data.ms (-)

265.9

m/z-->

50 100 150 200 250 300 350 400 450 500

Sub

58.1

164.9

331.8398.3463.7531.4

m/z-->

Time-->

16.60 16.70 16.80

Abundance

16.710

Time-->


16.60 16.70 16.80

Abundance

16.910

Time-->

16.80 16.90 17.00



Ref

50 58.0 132.0 200.1 334.5 402.3467.6 536.1

m/z-->

50 100 150 200 250 300 350 400 450 500

Abundance Scan 2333 (16.710 min): BP021566.D\data.ms (-)

200.1

m/z-->

50 100 150 200 250 300 350 400 450 500

Raw

50 58.1 132.0 200.1 287.7 354.4 4422.2 526.1

m/z-->

50 100 150 200 250 300 350 400 450 500

Abundance Scan 2333 (16.710 min): BP021566.D\data.ms (-)

200.1

m/z-->

50 100 150 200 250 300 350 400 450 500

Sub

50 58.1 131.9 200.1 269.3 384.0 452.0 519.8

m/z-->

50 100 150 200 250 300 350 400 450 500

Abundance Scan 2364 (16.892 min): BP021485.D\data.ms (-)

265.9

m/z-->

50 100 150 200 250 300 350 400 450 500

Ref

50 95.0 164.9 200.1 333.6 398.9 465.1 530.8

m/z-->

50 100 150 200 250 300 350 400 450 500

Abundance Scan 2367 (16.910 min): BP021566.D\data.ms (-)

265.9

m/z-->

50 100 150 200 250 300 350 400 450 500

Raw

50 95.0 164.9 200.1 332.9 398.3 521.2

m/z-->

50 100 150 200 250 300 350 400 450 500

Abundance Scan 2367 (16.910 min): BP021566.D\data.ms (-)

265.9

m/z-->

50 100 150 200 250 300 350 400 450 500

Sub

50 95.0 164.9 200.1 331.8 398.3 463.7 531.4

m/z-->

50 100 150 200 250 300 350 400 450 500

Abundance

#70

Pentachlorophenol

Concen: 49.739 ng

RT: 16.910 min Scan# 2367

Delta R.T. 0.018 min

Lab File: BP021566.D

Acq: 20 Aug 2024 10:07

Tgt Ion:266 Resp: 714044

Ion Ratio Lower Upper

266 100

268 63.5 51.8 77.6

264 61.9 50.5 75.7

m/z-->


50 100 150 200 250 300 350 400 450 500

Abundance

16.910

Time-->

16.80 16.90 17.00



Ref

50 95.0 164.9 200.1 333.6 398.9 465.1 530.8

m/z-->

50 100 150 200 250 300 350 400 450 500

Abundance Scan 2367 (16.910 min): BP021566.D\data.ms (-)

265.9

m/z-->

50 100 150 200 250 300 350 400 450 500

Raw

50 95.0 164.9 200.1 332.9 398.3 521.2

m/z-->

50 100 150 200 250 300 350 400 450 500

Abundance Scan 2367 (16.910 min): BP021566.D\data.ms (-)

265.9

m/z-->




50 100 150 200 250 300 350 400 450 500

Sub

50 95.0 164.9 200.1 331.8 398.3 463.7 531.4

m/z-->

50 100 150 200 250 300 350 400 450 500



#71

Phenanthrene

Concen: 39.548 ng

RT: 17.310 min Scan# 2435

Delta R.T. 0.012 min

Lab File: BP021566.D

Acq: 20 Aug 2024 10:07

Instrument :

BNA_P





ClientSampleId :

SSTDCCC040

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/21/2024

Supervised By :mohammad ahmed 08/22/2024



#72

Anthracene

Concen: 40.140 ng

RT: 17.404 min Scan# 2451

Delta R.T. 0.006 min

Lab File: BP021566.D

Acq: 20 Aug 2024 10:07


Tgt Ion:178 Resp: 3981334


Ion Ratio Lower Upper

178 100

176 18.5 14.9 22.3

179 15.6 12.5 18.7







#73
Carbazole
Concen: 41.294 ng
RT: 17.675 min Scan# 2
Delta R.T. 0.006 min
Lab File: BP021566.D
Acq: 20 Aug 2024 10:07

Instrument : BNA_P
ClientSampleId : SSTDCCC040


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/21/2024
Supervised By :mohammad ahmed 08/22/2024




#74
Di-n-butylphthalate
Concen: 40.166 ng
RT: 18.280 min Scan# 2600
Delta R.T. 0.006 min
Lab File: BP021566.D
Acq: 20 Aug 2024 10:07

Tgt Ion:149 Resp: 4715211
Ion Ratio Lower Upper
149 100
150 9.4 7.5 11.3
104 5.8 4.9 7.3






Abundance Scan 2788 (19.386 min): BP021566.D\data.ms (-)



Abundance Scan 2788 (19.386 min): BP021566.D\data.ms (-)



#75

Fluoranthene

Concen: 41.572 ng

RT: 19.386 min Scan# 2

Delta R.T. 0.006 min

Lab File: BP021566.D

Acq: 20 Aug 2024 10:07

Instrument :

BNA_P

ClientSampleId :

SSTDCCC040

Tgt Ion:202 Resp: 515875:

Ion Ratio Lower Upper

202 100

101 14.3 0.0 34.0


203 17.2 0.0 37.5

Manual Integrations


APPROVED

Reviewed By :Yogesh Patel 08/21/2024


Supervised By :mohammad ahmed 08/22/2024



Abundance Scan 3186 (21.727 min): BP021566.D\data.ms (-)



Abundance Scan 3186 (21.727 min): BP021566.D\data.ms (-)



#76

Chrysene-d₁₂

Concen: 20.000 ng

RT: 21.727 min Scan# 3186

Delta R.T. 0.029 min

Lab File: BP021566.D

Acq: 20 Aug 2024 10:07


Tgt Ion:240 Resp: 2067038

Ion Ratio Lower Upper


240 100

120 11.9 9.9 14.9


236 25.5 20.6 30.8



Abundance Scan 19.386 min: BP021566.D\data.ms (-)



Abundance Scan 21.698 min: BP021485.D\data.ms (-)



Abundance Scan 21.727 min: BP021566.D\data.ms (-)

#76

Chrysene-d₁₂

Concen: 20.000 ng

RT: 21.727 min Scan# 3186

Delta R.T. 0.029 min

Lab File: BP021566.D

Acq: 20 Aug 2024 10:07


Tgt Ion:240 Resp: 2067038

Ion Ratio Lower Upper


240 100

120 11.9 9.9 14.9

236 25.5 20.6 30.8






Abundance Scan 21.698 min: BP021485.D\data.ms (-)



Abundance Scan 21.727 min: BP021566.D\data.ms (-)



Abundance Scan 21.727 min: BP021566.D\data.ms (-)







#77
Benzidine
Concen: 43.787 ng
RT: 19.580 min Scan# 21
Delta R.T. 0.005 min
Lab File: BP021566.D
Acq: 20 Aug 2024 10:07

Instrument : BNA_P
ClientSampleId : SSTDCCC040


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 08/21/2024
Supervised By :mohammad ahmed 08/22/2024



#78
Pyrene
Concen: 39.398 ng
RT: 19.763 min Scan# 2852
Delta R.T. 0.012 min
Lab File: BP021566.D
Acq: 20 Aug 2024 10:07

Tgt Ion:202 Resp: 5322876
Ion Ratio Lower Upper
202 100
200 20.7 16.7 25.1
203 17.6 14.1 21.1







#79
Terphenyl-d14
Concen: 74.156 ng
RT: 19.992 min Scan# 2
Delta R.T. 0.017 min
Lab File: BP021566.D
Acq: 20 Aug 2024 10:07

Instrument : BNA_P
ClientSampleId : SSTDCCC040


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/21/2024
Supervised By :mohammad ahmed 08/22/2024



#80
Butylbenzylphthalate
Concen: 43.672 ng
RT: 20.716 min Scan# 3014
Delta R.T. 0.006 min
Lab File: BP021566.D
Acq: 20 Aug 2024 10:07

Tgt Ion:149 Resp: 2221372
Ion Ratio Lower Upper
149 100
91 85.3 73.3 109.9
206 21.5 17.1 25.7



#81

Benzo(a)anthracene

Concen: 40.003 ng

RT: 21.710 min Scan# 3

Instrument :

Delta R.T. 0.029 min

BNA_P

Lab File: BP021566.D

ClientSampleId :

Acq: 20 Aug 2024 10:07

SSTDCCC040

#83

Chrysene

Concen: 40.101 ng

RT: 21.780 min Scan# 3

Delta R.T. 0.029 min

Lab File: BP021566.D

Acq: 20 Aug 2024 10:07

Instrument :

BNA_P

ClientSampleId :

SSTDCCC040

Tgt Ion:228 Resp: 4912854

Ion Ratio Lower Upper

228 100

226 29.8 23.9 35.9

229 20.0 15.8 23.6

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 08/21/2024

Supervised By :mohammad ahmed 08/22/2024

Abundance

3000000


2000000

1000000

0

Time-->

21.70 21.80 21.90





#85
Di-n-octyl phthalate
Concen: 42.612 ng
RT: 22.986 min Scan# 3400
Delta R.T. 0.023 min
Lab File: BP021566.D
Acq: 20 Aug 2024 10:07

Instrument : BNA_P
ClientSampleId : SSTDCCC040


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/21/2024
Supervised By :mohammad ahmed 08/22/2024



#86
Perylene-d₁₂
Concen: 20.000 ng
RT: 25.174 min Scan# 3772
Delta R.T. 0.070 min
Lab File: BP021566.D
Acq: 20 Aug 2024 10:07

Tgt Ion:264 Resp: 2510089
Ion Ratio Lower Upper
264 100
260 22.4 18.6 28.0
265 22.1 17.1 25.7



#87

Indeno(1,2,3-cd)pyrene

Concen: 40.860 ng m

RT: 29.103 min Scan# 4

Delta R.T. 0.112 min

Lab File: BP021566.D

Acq: 20 Aug 2024 10:07

Instrument :

BNA_P

ClientSampleId :

SSTDCCC040

Tgt Ion:276 Resp: 664823

Ion Ratio Lower Upper

276 100

138 16.2 12.8 19.2

277 22.7 16.5 24.7

Manual Integrations

APPROVED



Reviewed By :Yogesh Patel 08/21/2024

Supervised By :mohammad ahmed 08/22/2024

Abundance

Scan 4440 (29.103 min): BP021566.D\data.ms (-)

276.1



Time-->

#88

Benzo(b)fluoranthene

Concen: 39.003 ng

RT: 24.086 min Scan# 3587

Delta R.T. 0.059 min

Lab File: BP021566.D

Acq: 20 Aug 2024 10:07

Tgt Ion:252 Resp: 5622632

Ion Ratio Lower Upper

252 100



253 21.3 17.6 26.4

125 10.6 8.6 13.0


Abundance

Scan 3587 (24.086 min): BP021566.D\data.ms (-)

252.1



Time-->



#87

Indeno(1,2,3-cd)pyrene

Concen: 40.860 ng m

RT: 29.103 min Scan# 4

Delta R.T. 0.112 min

Lab File: BP021566.D


Acq: 20 Aug 2024 10:07

Instrument :

BNA_P

ClientSampleId :

SSTDCCC040



#88

Benzo(b)fluoranthene

Concen: 39.003 ng

RT: 24.086 min Scan# 3587

Delta R.T. 0.059 min

Lab File: BP021566.D

Acq: 20 Aug 2024 10:07

Tgt Ion:252 Resp: 5622632

Ion Ratio Lower Upper

252 100



253 21.3 17.6 26.4

125 10.6 8.6 13.0

Abundance

Scan 3587 (24.086 min): BP021566.D\data.ms (-)

252.1



Time-->

#89

Benzo(k)fluoranthene

Concen: 38.710 ng

RT: 24.163 min Scan# 3

Delta R.T. 0.053 min

Lab File: BP021566.D



Acq: 20 Aug 2024 10:07

Instrument :

BNA_P

ClientSampleId :

SSTDCCC040



Tgt Ion:252 Resp: 544379:

Ion Ratio Lower Upper

252 100





253 22.0 17.5 26.3

125 10.7 7.8 11.8

Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel 08/21/2024

Supervised By :mohammad ahmed 08/22/2024



#90

Benzo(a)pyrene

Concen: 40.145 ng

RT: 25.004 min Scan# 3743

Delta R.T. 0.053 min

Lab File: BP021566.D

Acq: 20 Aug 2024 10:07


Tgt Ion:252 Resp: 5022273


Ion Ratio Lower Upper

252 100

253 21.9 17.7 26.5

125 11.4 8.7 13.1






#91
Dibenzo(a,h)anthracene
Concen: 40.541 ng
RT: 29.192 min Scan# 4455
Delta R.T. 0.106 min
Lab File: BP021566.D
Acq: 20 Aug 2024 10:07

Instrument : BNA_P
ClientSampleId : SSTDCCC040

Tgt Ion:278 Resp: 548081:
Ion Ratio Lower Upper
278 100
139 15.3 12.2 18.4
279 23.7 19.1 28.7



Manual Integrations APPROVED

Reviewed By :Yogesh Patel 08/21/2024
Supervised By :mohammad ahmed 08/22/2024



#92
Benzo(g,h,i)perylene
Concen: 41.504 ng
RT: 30.309 min Scan# 4645
Delta R.T. 0.135 min
Lab File: BP021566.D
Acq: 20 Aug 2024 10:07

Tgt Ion:276 Resp: 5623792
Ion Ratio Lower Upper
276 100
277 23.9 19.0 28.6
138 21.3 16.0 24.0



Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP082024\
 Data File : BP021566.D
 Acq On : 20 Aug 2024 10:07
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 20 11:20:14 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 14 00:00:30 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

| | Compound | AvgRF | CCRF | %Dev | Area% | Dev(min) |
|------|-----------------------------|-------|-------|---------|-------|----------|
| 1 I | 1,4-Dichlorobenzene-d4 | 1.000 | 1.000 | 0.0 | 107 | 0.00 |
| 2 | 1,4-Dioxane | 0.639 | 0.561 | 12.2 | 93 | 0.00 |
| 3 | Pyridine | 1.783 | 1.602 | 10.2 | 95 | 0.01 |
| 4 | n-Nitrosodimethylamine | 0.536 | 0.492 | 8.2 | 99 | 0.00 |
| 5 S | 2-Fluorophenol | 1.347 | 1.353 | -0.4 | 107 | 0.01 |
| 6 | Aniline | 1.964 | 1.799 | 8.4 | 97 | 0.00 |
| 7 S | Phenol-d6 | 1.965 | 1.932 | 1.7 | 104 | 0.01 |
| 8 | 2-Chlorophenol | 1.245 | 1.295 | -4.0 | 111 | 0.00 |
| 9 | Benzaldehyde | 1.258 | 1.358 | -7.9 | 121 | 0.00 |
| 10 C | Phenol | 2.035 | 1.947 | 4.3 | 102 | 0.01 |
| 11 | bis(2-Chloroethyl)ether | 1.722 | 1.542 | 10.5 | 97 | 0.00 |
| 12 | 1,3-Dichlorobenzene | 1.479 | 1.438 | 2.8 | 106 | 0.00 |
| 13 C | 1,4-Dichlorobenzene | 1.493 | 1.457 | 2.4 | 106 | 0.00 |
| 14 | 1,2-Dichlorobenzene | 1.439 | 1.385 | 3.8 | 105 | 0.00 |
| 15 | Benzyl Alcohol | 1.410 | 1.348 | 4.4 | 102 | 0.00 |
| 16 | 2,2'-oxybis(1-Chloropropane | 1.585 | 1.383 | 12.7 | 96 | 0.00 |
| 17 | 2-Methylphenol | 1.286 | 1.272 | 1.1 | 105 | 0.01 |
| 18 | Hexachloroethane | 0.599 | 0.597 | 0.3 | 107 | 0.00 |
| 19 P | n-Nitroso-di-n-propylamine | 1.357 | 1.109 | 18.3 | 88 | 0.00 |
| 20 | 3+4-Methylphenols | 1.734 | 1.797 | -3.6 | 109 | 0.01 |
| 21 I | Naphthalene-d8 | 1.000 | 1.000 | 0.0 | 105 | 0.00 |
| 22 | Acetophenone | 0.614 | 0.584 | 4.9 | 101 | 0.00 |
| 23 S | Nitrobenzene-d5 | 0.387 | 0.426 | -10.1 | 109 | 0.00 |
| 24 | Nitrobenzene | 0.421 | 0.432 | -2.6 | 103 | 0.00 |
| 25 | Isophorone | 0.915 | 0.803 | 12.2 | 93 | 0.00 |
| 26 C | 2-Nitrophenol | 0.086 | 0.165 | -91.9# | 185# | 0.00 |
| 27 | 2,4-Dimethylphenol | 0.226 | 0.229 | -1.3 | 107 | 0.01 |
| 28 | bis(2-Chloroethoxy)methane | 0.559 | 0.505 | 9.7 | 97 | 0.00 |
| 29 C | 2,4-Dichlorophenol | 0.267 | 0.300 | -12.4 | 116 | 0.00 |
| 30 | 1,2,4-Trichlorobenzene | 0.342 | 0.332 | 2.9 | 103 | 0.00 |
| 31 | Naphthalene | 1.045 | 1.030 | 1.4 | 104 | 0.00 |
| 32 | Benzoic acid | 0.137 | 0.274 | -100.0# | 223# | 0.02 |
| 33 | 4-Chloroaniline | 0.392 | 0.398 | -1.5 | 106 | 0.00 |
| 34 C | Hexachlorobutadiene | 0.215 | 0.208 | 3.3 | 102 | 0.00 |
| 35 | Caprolactam | 0.111 | 0.130 | -17.1 | 120 | 0.02 |
| 36 C | 4-Chloro-3-methylphenol | 0.370 | 0.406 | -9.7 | 112 | 0.01 |
| 37 | 2-Methylnaphthalene | 0.713 | 0.693 | 2.8 | 104 | 0.00 |
| 38 | 1-Methylnaphthalene | 0.703 | 0.682 | 3.0 | 103 | 0.00 |
| 39 I | Acenaphthene-d10 | 1.000 | 1.000 | 0.0 | 109 | 0.00 |
| 40 | 1,2,4,5-Tetrachlorobenzene | 0.586 | 0.559 | 4.6 | 103 | 0.00 |
| 41 P | Hexachlorocyclopentadiene | 0.186 | 0.235 | -26.3# | 134 | 0.00 |
| 42 S | 2,4,6-Tribromophenol | 0.223 | 0.267 | -19.7 | 123 | 0.00 |
| 43 C | 2,4,6-Trichlorophenol | 0.326 | 0.378 | -16.0 | 120 | 0.00 |
| 44 | 2,4,5-Trichlorophenol | 0.362 | 0.427 | -18.0 | 122 | 0.00 |
| 45 S | 2-Fluorobiphenyl | 1.310 | 1.241 | 5.3 | 102 | 0.00 |
| 46 | 1,1'-Biphenyl | 1.415 | 1.347 | 4.8 | 103 | 0.00 |
| 47 | 2-Chloronaphthalene | 1.102 | 1.063 | 3.5 | 104 | 0.00 |

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP082024\
 Data File : BP021566.D
 Acq On : 20 Aug 2024 10:07
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 20 11:20:14 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 14 00:00:30 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

| | Compound | AvgRF | CCRF | %Dev | Area% | Dev(min) |
|------|----------------------------|-------|-------|---------|-------|----------|
| 48 | 2-Nitroaniline | 0.256 | 0.363 | -41.8# | 139 | 0.00 |
| 49 | Acenaphthylene | 1.619 | 1.589 | 1.9 | 105 | 0.00 |
| 50 | Dimethylphthalate | 1.339 | 1.328 | 0.8 | 107 | 0.00 |
| 51 | 2,6-Dinitrotoluene | 0.232 | 0.306 | -31.9# | 131 | -0.01 |
| 52 C | Acenaphthene | 1.063 | 1.082 | -1.8 | 110 | 0.00 |
| 53 | 3-Nitroaniline | 0.228 | 0.323 | -41.7# | 139 | -0.01 |
| 54 P | 2,4-Dinitrophenol | 0.069 | 0.138 | -100.0# | 229# | 0.01 |
| 55 | Dibenzofuran | 1.635 | 1.602 | 2.0 | 106 | 0.00 |
| 56 P | 4-Nitrophenol | 0.195 | 0.287 | -47.2# | 155# | 0.00 |
| 57 | 2,4-Dinitrotoluene | 0.284 | 0.425 | -49.6# | 145 | 0.00 |
| 58 | Fluorene | 1.343 | 1.294 | 3.6 | 104 | 0.01 |
| 59 | 2,3,4,6-Tetrachlorophenol | 0.307 | 0.369 | -20.2 | 123 | 0.00 |
| 60 | Diethylphthalate | 1.374 | 1.367 | 0.5 | 107 | -0.01 |
| 61 | 4-Chlorophenyl-phenylether | 0.694 | 0.644 | 7.2 | 101 | 0.00 |
| 62 | 4-Nitroaniline | 0.241 | 0.337 | -39.8# | 135 | 0.00 |
| 63 | Azobenzene | 1.689 | 1.474 | 12.7 | 94 | 0.00 |
| 64 I | Phenanthrene-d10 | 1.000 | 1.000 | 0.0 | 110 | 0.01 |
| 65 | 4,6-Dinitro-2-methylphenol | 0.054 | 0.096 | -77.8# | 203# | 0.01 |
| 66 c | n-Nitrosodiphenylamine | 0.536 | 0.518 | 3.4 | 106 | 0.00 |
| 67 | 4-Bromophenyl-phenylether | 0.212 | 0.199 | 6.1 | 106 | 0.00 |
| 68 | Hexachlorobenzene | 0.250 | 0.242 | 3.2 | 106 | 0.00 |
| 69 | Atrazine | 0.184 | 0.152 | 17.4 | 88 | 0.00 |
| 70 C | Pentachlorophenol | 0.137 | 0.170 | -24.1# | 137 | 0.02 |
| 71 | Phenanthrene | 0.962 | 0.951 | 1.1 | 109 | 0.01 |
| 72 | Anthracene | 0.943 | 0.947 | -0.4 | 110 | 0.00 |
| 73 | Carbazole | 0.909 | 0.938 | -3.2 | 111 | 0.00 |
| 74 | Di-n-butylphthalate | 1.116 | 1.121 | -0.4 | 108 | 0.00 |
| 75 C | Fluoranthene | 1.180 | 1.227 | -4.0 | 113 | 0.00 |
| 76 I | Chrysene-d12 | 1.000 | 1.000 | 0.0 | 114 | 0.03 |
| 77 | Benzidine | 0.420 | 0.506 | -20.5 | 121 | 0.00 |
| 78 | Pyrene | 1.307 | 1.288 | 1.5 | 113 | 0.01 |
| 79 S | Terphenyl-d14 | 1.029 | 0.954 | 7.3 | 108 | 0.02 |
| 80 | Butylbenzylphthalate | 0.437 | 0.537 | -22.9 | 130 | 0.00 |
| 81 | Benzo(a)anthracene | 1.256 | 1.256 | 0.0 | 112 | 0.03 |
| 82 | 3,3'-Dichlorobenzidine | 0.403 | 0.465 | -15.4 | 123 | 0.02 |
| 83 | Chrysene | 1.185 | 1.188 | -0.3 | 113 | 0.03 |
| 84 | Bis(2-ethylhexyl)phthalate | 0.742 | 0.771 | -3.9 | 112 | 0.02 |
| 85 c | Di-n-octyl phthalate | 1.199 | 1.369 | -14.2 | 125 | 0.02 |
| 86 I | Perylene-d12 | 1.000 | 1.000 | 0.0 | 120 | 0.07 |
| 87 | Indeno(1,2,3-cd)pyrene | 1.296 | 1.324 | -2.2 | 120 | 0.11 |
| 88 | Benzo(b)fluoranthene | 1.149 | 1.120 | 2.5 | 117 | 0.06 |
| 89 | Benzo(k)fluoranthene | 1.121 | 1.084 | 3.3 | 114 | 0.05 |
| 90 C | Benzo(a)pyrene | 0.997 | 1.000 | -0.3 | 119 | 0.05 |
| 91 | Dibenzo(a,h)anthracene | 1.077 | 1.092 | -1.4 | 119 | 0.11 |
| 92 | Benzo(g,h,i)perylene | 1.080 | 1.120 | -3.7 | 122 | 0.14 |

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP082024\
Data File : BP021566.D
Acq On : 20 Aug 2024 10:07
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Aug 20 11:20:14 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 14 00:00:30 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

| Compound | AvgRF | CCRF | %Dev | Area% | Dev(min) |
|----------|-------|------|------|-------|----------|
|----------|-------|------|------|-------|----------|

(#) = Out of Range SPCC's out = 0 CCC's out = 2

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP082024\
 Data File : BP021566.D
 Acq On : 20 Aug 2024 10:07
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 20 11:20:14 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 14 00:00:30 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

| | Compound | Amount | Calc. | %Dev | Area% | Dev(min) |
|------|-----------------------------|--------|--------|--------|-------|----------|
| 1 I | 1,4-Dichlorobenzene-d4 | 20.000 | 20.000 | 0.0 | 107 | 0.00 |
| 2 | 1,4-Dioxane | 40.000 | 35.145 | 12.1 | 93 | 0.00 |
| 3 | Pyridine | 40.000 | 35.932 | 10.2 | 95 | 0.01 |
| 4 | n-Nitrosodimethylamine | 40.000 | 36.744 | 8.1 | 99 | 0.00 |
| 5 S | 2-Fluorophenol | 80.000 | 80.358 | -0.4 | 107 | 0.01 |
| 6 | Aniline | 40.000 | 36.630 | 8.4 | 97 | 0.00 |
| 7 S | Phenol-d6 | 80.000 | 78.656 | 1.7 | 104 | 0.01 |
| 8 | 2-Chlorophenol | 40.000 | 41.611 | -4.0 | 111 | 0.00 |
| 9 | Benzaldehyde | 40.000 | 43.178 | -7.9 | 121 | 0.00 |
| 10 C | Phenol | 40.000 | 38.270 | 4.3 | 102 | 0.01 |
| 11 | bis(2-Chloroethyl)ether | 40.000 | 35.826 | 10.4 | 97 | 0.00 |
| 12 | 1,3-Dichlorobenzene | 40.000 | 38.912 | 2.7 | 106 | 0.00 |
| 13 C | 1,4-Dichlorobenzene | 40.000 | 39.044 | 2.4 | 106 | 0.00 |
| 14 | 1,2-Dichlorobenzene | 40.000 | 38.497 | 3.8 | 105 | 0.00 |
| 15 | Benzyl Alcohol | 40.000 | 38.244 | 4.4 | 102 | 0.00 |
| 16 | 2,2'-oxybis(1-Chloropropane | 40.000 | 34.893 | 12.8 | 96 | 0.00 |
| 17 | 2-Methylphenol | 40.000 | 39.549 | 1.1 | 105 | 0.01 |
| 18 | Hexachloroethane | 40.000 | 39.861 | 0.3 | 107 | 0.00 |
| 19 P | n-Nitroso-di-n-propylamine | 40.000 | 32.677 | 18.3 | 88 | 0.00 |
| 20 | 3+4-Methylphenols | 40.000 | 41.450 | -3.6 | 109 | 0.01 |
| 21 I | Naphthalene-d8 | 20.000 | 20.000 | 0.0 | 105 | 0.00 |
| 22 | Acetophenone | 40.000 | 37.996 | 5.0 | 101 | 0.00 |
| 23 S | Nitrobenzene-d5 | 80.000 | 87.927 | -9.9 | 109 | 0.00 |
| 24 | Nitrobenzene | 40.000 | 41.074 | -2.7 | 103 | 0.00 |
| 25 | Isophorone | 40.000 | 35.068 | 12.3 | 93 | 0.00 |
| 26 C | 2-Nitrophenol | 40.000 | 56.961 | -42.4# | 185 | 0.00 |
| 27 | 2,4-Dimethylphenol | 40.000 | 40.576 | -1.4 | 107 | 0.01 |
| 28 | bis(2-Chloroethoxy)methane | 40.000 | 36.141 | 9.6 | 97 | 0.00 |
| 29 C | 2,4-Dichlorophenol | 40.000 | 44.986 | -12.5 | 116 | 0.00 |
| 30 | 1,2,4-Trichlorobenzene | 40.000 | 38.855 | 2.9 | 103 | 0.00 |
| 31 | Naphthalene | 40.000 | 39.431 | 1.4 | 104 | 0.00 |
| 32 | Benzoic acid | 40.000 | 64.703 | -61.8# | 223 | 0.02 |
| 33 | 4-Chloroaniline | 40.000 | 40.535 | -1.3 | 106 | 0.00 |
| 34 C | Hexachlorobutadiene | 40.000 | 38.600 | 3.5 | 102 | 0.00 |
| 35 | Caprolactam | 40.000 | 46.800 | -17.0 | 120 | 0.02 |
| 36 C | 4-Chloro-3-methylphenol | 40.000 | 43.882 | -9.7 | 112 | 0.01 |
| 37 | 2-Methylnaphthalene | 40.000 | 38.859 | 2.9 | 104 | 0.00 |
| 38 | 1-Methylnaphthalene | 40.000 | 38.775 | 3.1 | 103 | 0.00 |
| 39 I | Acenaphthene-d10 | 20.000 | 20.000 | 0.0 | 109 | 0.00 |
| 40 | 1,2,4,5-Tetrachlorobenzene | 40.000 | 38.147 | 4.6 | 103 | 0.00 |
| 41 P | Hexachlorocyclopentadiene | 40.000 | 50.596 | -26.5# | 134 | 0.00 |
| 42 S | 2,4,6-Tribromophenol | 80.000 | 95.950 | -19.9 | 123 | 0.00 |
| 43 C | 2,4,6-Trichlorophenol | 40.000 | 46.468 | -16.2 | 120 | 0.00 |
| 44 | 2,4,5-Trichlorophenol | 40.000 | 47.232 | -18.1 | 122 | 0.00 |
| 45 S | 2-Fluorobiphenyl | 80.000 | 75.740 | 5.3 | 102 | 0.00 |
| 46 | 1,1'-Biphenyl | 40.000 | 38.080 | 4.8 | 103 | 0.00 |
| 47 | 2-Chloronaphthalene | 40.000 | 38.599 | 3.5 | 104 | 0.00 |

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP082024\
 Data File : BP021566.D
 Acq On : 20 Aug 2024 10:07
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 20 11:20:14 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 14 00:00:30 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

| | Compound | Amount | Calc. | %Dev | Area% | Dev(min) |
|------|----------------------------|--------|--------|--------|-------|----------|
| 48 | 2-Nitroaniline | 40.000 | 47.106 | -17.8 | 139 | 0.00 |
| 49 | Acenaphthylene | 40.000 | 39.251 | 1.9 | 105 | 0.00 |
| 50 | Dimethylphthalate | 40.000 | 39.669 | 0.8 | 107 | 0.00 |
| 51 | 2,6-Dinitrotoluene | 40.000 | 45.498 | -13.7 | 131 | -0.01 |
| 52 C | Acenaphthene | 40.000 | 40.738 | -1.8 | 110 | 0.00 |
| 53 | 3-Nitroaniline | 40.000 | 48.589 | -21.5 | 139 | -0.01 |
| 54 P | 2,4-Dinitrophenol | 40.000 | 65.093 | -62.7# | 229 | 0.01 |
| 55 | Dibenzofuran | 40.000 | 39.211 | 2.0 | 106 | 0.00 |
| 56 P | 4-Nitrophenol | 40.000 | 53.424 | -33.6# | 155 | 0.00 |
| 57 | 2,4-Dinitrotoluene | 40.000 | 49.320 | -23.3 | 145 | 0.00 |
| 58 | Fluorene | 40.000 | 38.542 | 3.6 | 104 | 0.01 |
| 59 | 2,3,4,6-Tetrachlorophenol | 40.000 | 48.067 | -20.2 | 123 | 0.00 |
| 60 | Diethylphthalate | 40.000 | 39.800 | 0.5 | 107 | -0.01 |
| 61 | 4-Chlorophenyl-phenylether | 40.000 | 37.093 | 7.3 | 101 | 0.00 |
| 62 | 4-Nitroaniline | 40.000 | 49.265 | -23.2 | 135 | 0.00 |
| 63 | Azobenzene | 40.000 | 34.913 | 12.7 | 94 | 0.00 |
| 64 I | Phenanthrene-d10 | 20.000 | 20.000 | 0.0 | 110 | 0.01 |
| 65 | 4,6-Dinitro-2-methylphenol | 40.000 | 58.939 | -47.3# | 203 | 0.01 |
| 66 c | n-Nitrosodiphenylamine | 40.000 | 38.672 | 3.3 | 106 | 0.00 |
| 67 | 4-Bromophenyl-phenylether | 40.000 | 37.531 | 6.2 | 106 | 0.00 |
| 68 | Hexachlorobenzene | 40.000 | 38.748 | 3.1 | 106 | 0.00 |
| 69 | Atrazine | 40.000 | 33.020 | 17.4 | 88 | 0.00 |
| 70 C | Pentachlorophenol | 40.000 | 49.739 | -24.3# | 137 | 0.02 |
| 71 | Phenanthrene | 40.000 | 39.548 | 1.1 | 109 | 0.01 |
| 72 | Anthracene | 40.000 | 40.140 | -0.4 | 110 | 0.00 |
| 73 | Carbazole | 40.000 | 41.294 | -3.2 | 111 | 0.00 |
| 74 | Di-n-butylphthalate | 40.000 | 40.166 | -0.4 | 108 | 0.00 |
| 75 C | Fluoranthene | 40.000 | 41.572 | -3.9 | 113 | 0.00 |
| 76 I | Chrysene-d12 | 20.000 | 20.000 | 0.0 | 114 | 0.03 |
| 77 | Benzidine | 40.000 | 43.787 | -9.5 | 121 | 0.00 |
| 78 | Pyrene | 40.000 | 39.398 | 1.5 | 113 | 0.01 |
| 79 S | Terphenyl-d14 | 80.000 | 74.156 | 7.3 | 108 | 0.02 |
| 80 | Butylbenzylphthalate | 40.000 | 43.672 | -9.2 | 130 | 0.00 |
| 81 | Benzo(a)anthracene | 40.000 | 40.003 | -0.0 | 112 | 0.03 |
| 82 | 3,3'-Dichlorobenzidine | 40.000 | 46.202 | -15.5 | 123 | 0.02 |
| 83 | Chrysene | 40.000 | 40.101 | -0.3 | 113 | 0.03 |
| 84 | Bis(2-ethylhexyl)phthalate | 40.000 | 41.537 | -3.8 | 112 | 0.02 |
| 85 c | Di-n-octyl phthalate | 40.000 | 42.612 | -6.5 | 125 | 0.02 |
| 86 I | Perylene-d12 | 20.000 | 20.000 | 0.0 | 120 | 0.07 |
| 87 | Indeno(1,2,3-cd)pyrene | 40.000 | 40.860 | -2.1 | 120 | 0.11 |
| 88 | Benzo(b)fluoranthene | 40.000 | 39.003 | 2.5 | 117 | 0.06 |
| 89 | Benzo(k)fluoranthene | 40.000 | 38.710 | 3.2 | 114 | 0.05 |
| 90 C | Benzo(a)pyrene | 40.000 | 40.145 | -0.4 | 119 | 0.05 |
| 91 | Dibenzo(a,h)anthracene | 40.000 | 40.541 | -1.4 | 119 | 0.11 |
| 92 | Benzo(g,h,i)perylene | 40.000 | 41.504 | -3.8 | 122 | 0.14 |

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP082024\
Data File : BP021566.D
Acq On : 20 Aug 2024 10:07
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Aug 20 11:20:14 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 14 00:00:30 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

| Compound | Amount | Calc. | %Dev | Area% | Dev(min) |
|----------|--------|-------|------|-------|----------|
|----------|--------|-------|------|-------|----------|

(#) = Out of Range SPCC's out = 0 CCC's out = 2



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

7C

SEMOVOLATILE CONTINUING CALIBRATION CHECK

| | | | | | |
|-----------------|--------------------|-----------|------------------------|-------------------|-------------------|
| Lab Name: | <u>CHEMTECH</u> | | Contract: | <u>JAC005</u> | |
| Lab Code: | <u>CHEM</u> | Case No.: | <u>P3657</u> | SAS No.: | <u>P3657</u> |
| Instrument ID: | <u>BNA_P</u> | | Calibration Date/Time: | <u>08/20/2024</u> | <u>16:40</u> |
| Lab File ID: | <u>BP021575.D</u> | | Init. Calib. Date(s): | <u>08/13/2024</u> | <u>08/13/2024</u> |
| EPA Sample No.: | <u>SSTDCCCC040</u> | | Init. Calib. Time(s): | <u>12:32</u> | <u>17:36</u> |
| GC Column: | <u>ZB-GR</u> | ID: | <u>0.25</u> | (mm) | |

| COMPOUND | RRF | RRF040 | MIN RRF | %D | MAX%D |
|----------------------------|-------|--------|---------|------|-------|
| Pyridine | 1.783 | 1.623 | | -9.0 | |
| 2-Fluorophenol | 1.347 | 1.372 | | 1.9 | |
| Benzaldehyde | 1.258 | 1.392 | | 10.7 | |
| Phenol-d6 | 1.965 | 1.965 | | 0.0 | |
| 2-Methylphenol | 1.286 | 1.311 | | 1.9 | |
| 3+4-Methylphenols | 1.734 | 1.820 | | 5.0 | |
| Nitrobenzene-d5 | 0.387 | 0.421 | | 8.8 | |
| Hexachloroethane | 0.599 | 0.610 | | 1.8 | |
| Nitrobenzene | 0.421 | 0.430 | | 2.1 | |
| Naphthalene | 1.045 | 1.030 | | -1.4 | |
| Hexachlorobutadiene | 0.215 | 0.206 | | -4.2 | 20.0 |
| 2-Methylnaphthalene | 0.713 | 0.699 | | -2.0 | |
| 2,4,6-Trichlorophenol | 0.326 | 0.377 | | 15.6 | 20.0 |
| 2-Fluorobiphenyl | 1.310 | 1.212 | | -7.5 | |
| 2,4,5-Trichlorophenol | 0.362 | 0.422 | | 16.6 | |
| Acenaphthylene | 1.619 | 1.596 | | -1.4 | |
| Acenaphthene | 1.063 | 1.097 | | 3.2 | 20.0 |
| Dibenzofuran | 1.635 | 1.623 | | -0.7 | |
| 2,4-Dinitrotoluene | 0.284 | 0.440 | | 54.9 | |
| Fluorene | 1.343 | 1.311 | | -2.4 | |
| 2,4,6-Tribromophenol | 0.223 | 0.265 | | 18.8 | |
| Hexachlorobenzene | 0.250 | 0.240 | | -4.0 | |
| Pentachlorophenol | 0.137 | 0.168 | | 22.6 | 20.0 |
| Phenanthrene | 0.962 | 0.949 | | -1.4 | |
| Anthracene | 0.943 | 0.925 | | -1.9 | |
| Carbazole | 0.909 | 0.943 | | 3.7 | |
| Di-n-butylphthalate | 1.116 | 1.161 | | 4.0 | |
| Fluoranthene | 1.180 | 1.230 | | 4.2 | 20.0 |
| Pyrene | 1.307 | 1.267 | | -3.1 | |
| Terphenyl-d14 | 1.029 | 0.958 | | -6.9 | |
| Benzo(a)anthracene | 1.256 | 1.254 | | -0.2 | |
| Chrysene | 1.185 | 1.172 | | -1.1 | |
| Bis(2-ethylhexyl)phthalate | 0.742 | 0.836 | | 12.7 | |
| Benzo(b)fluoranthene | 1.149 | 1.131 | | -1.6 | |
| Benzo(k)fluoranthene | 1.121 | 1.065 | | -5.0 | |
| Benzo(a)pyrene | 0.997 | 0.991 | | -0.6 | 20.0 |
| Indeno(1,2,3-cd)pyrene | 1.296 | 1.308 | | 0.9 | |
| Dibenzo(a,h)anthracene | 1.077 | 1.071 | | -0.6 | |
| Benzo(g,h,i)perylene | 1.080 | 1.083 | | 0.3 | |



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

7C

SEMOVOLATILE CONTINUING CALIBRATION CHECK

| | | | | | |
|-----------------|--------------------|-----------------|------------------------|-------------------|-------------------|
| Lab Name: | <u>CHEMTECH</u> | | Contract: | <u>JAC005</u> | |
| Lab Code: | <u>CHEM</u> | Case No.: | <u>P3657</u> | SAS No.: | <u>P3657</u> |
| Instrument ID: | <u>BNA_P</u> | | Calibration Date/Time: | <u>08/20/2024</u> | <u>16:40</u> |
| Lab File ID: | <u>BP021575.D</u> | | Init. Calib. Date(s): | <u>08/13/2024</u> | <u>08/13/2024</u> |
| EPA Sample No.: | <u>SSTDCCCC040</u> | | Init. Calib. Time(s): | <u>12:32</u> | <u>17:36</u> |
| GC Column: | <u>ZB-GR</u> | ID: <u>0.25</u> | (mm) | | |

| COMPOUND | RRF | RRF040 | MIN RRF | %D | MAX%D |
|---------------------|-------|--------|---------|-------|-------|
| 1,4-Dioxane | 0.639 | 0.570 | | -10.8 | 20.0 |
| 1-Methylnaphthalene | 0.703 | 0.693 | | -1.4 | |

All other compounds must meet a minimum RRF of 0.010.

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP082024\
 Data File : BP021575.D
 Acq On : 20 Aug 2024 16:40
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040

Quant Time: Aug 20 17:15:31 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 14 00:00:30 2024
 Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/21/2024
 Supervised By :mohammad ahmed 08/22/2024

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|------------------------------------|--------|------|----------|--------|-------|----------|
| Internal Standards | | | | | | |
| 1) 1,4-Dichlorobenzene-d4 | 7.805 | 152 | 372530 | 20.000 | ng | 0.00 |
| 21) Naphthalene-d8 | 10.593 | 136 | 1563751 | 20.000 | ng | 0.00 |
| 39) Acenaphthene-d10 | 14.463 | 164 | 1068846 | 20.000 | ng | 0.00 |
| 64) Phenanthrene-d10 | 17.275 | 188 | 2403929 | 20.000 | ng | 0.02 |
| 76) Chrysene-d12 | 21.739 | 240 | 2426350 | 20.000 | ng | 0.04 |
| 86) Perylene-d12 | 25.180 | 264 | 2940308 | 20.000 | ng | 0.08 |
| System Monitoring Compounds | | | | | | |
| 5) 2-Fluorophenol | 5.399 | 112 | 2044869 | 81.506 | ng | 0.00 |
| 7) Phenol-d6 | 6.969 | 99 | 2928471 | 80.016 | ng | 0.00 |
| 23) Nitrobenzene-d5 | 8.952 | 82 | 2633490 | 86.924 | ng | 0.00 |
| 42) 2,4,6-Tribromophenol | 15.981 | 330 | 1133914 | 95.320 | ng | 0.01 |
| 45) 2-Fluorobiphenyl | 13.063 | 172 | 5181271 | 73.981 | ng | 0.00 |
| 79) Terphenyl-d14 | 19.998 | 244 | 9297974 | 74.447 | ng | 0.02 |
| Target Compounds | | | | | | |
| | | | | Qvalue | | |
| 2) 1,4-Dioxane | 3.317 | 88 | 424462 | 35.681 | ng | 99 |
| 3) Pyridine | 3.717 | 79 | 1208987 | 36.406 | ng | 97 |
| 4) n-Nitrosodimethylamine | 3.622 | 42 | 371373 | 37.206 | ng | 95 |
| 6) Aniline | 7.134 | 93 | 1348749 | 36.863 | ng | 96 |
| 8) 2-Chlorophenol | 7.375 | 128 | 979003 | 42.226 | ng | 95 |
| 9) Benzaldehyde | 6.946 | 77 | 1037198 | 44.262 | ng | 93 |
| 10) Phenol | 6.999 | 94 | 1481572 | 39.093 | ng | 96 |
| 11) bis(2-Chloroethyl)ether | 7.228 | 93 | 1172312 | 36.557 | ng | 96 |
| 12) 1,3-Dichlorobenzene | 7.699 | 146 | 1084084 | 39.364 | ng | 97 |
| 13) 1,4-Dichlorobenzene | 7.840 | 146 | 1102920 | 39.661 | ng | 98 |
| 14) 1,2-Dichlorobenzene | 8.158 | 146 | 1043542 | 38.940 | ng | 99 |
| 15) Benzyl Alcohol | 8.040 | 79 | 1015548 | 38.673 | ng | 96 |
| 16) 2,2'-oxybis(1-Chloropr... | 8.334 | 45 | 1061205 | 35.934 | ng | 97 |
| 17) 2-Methylphenol | 8.240 | 107 | 976782 | 40.769 | ng | 97 |
| 18) Hexachloroethane | 8.887 | 117 | 454819 | 40.786 | ng | 98 |
| 19) n-Nitroso-di-n-propyla... | 8.605 | 70 | 850374 | 33.637 | ng | 95 |
| 20) 3+4-Methylphenols | 8.563 | 107 | 1356088 | 41.976 | ng | 99 |
| 22) Acetophenone | 8.616 | 105 | 1802202 | 37.514 | ng | 99 |
| 24) Nitrobenzene | 8.993 | 77 | 1345015 | 40.876 | ng | 93 |
| 25) Isophorone | 9.522 | 82 | 2538499 | 35.467 | ng | 98 |
| 26) 2-Nitrophenol | 9.704 | 139 | 508327 | 56.363 | ng | 92 |
| 27) 2,4-Dimethylphenol | 9.769 | 122 | 711129 | 40.260 | ng | 98 |
| 28) bis(2-Chloroethoxy)met... | 9.999 | 93 | 1593547 | 36.474 | ng | 99 |
| 29) 2,4-Dichlorophenol | 10.240 | 162 | 947435 | 45.397 | ng | 98 |
| 30) 1,2,4-Trichlorobenzene | 10.457 | 180 | 1037019 | 38.758 | ng | 98 |
| 31) Naphthalene | 10.646 | 128 | 3220205 | 39.419 | ng | 100 |
| 32) Benzoic acid | 9.905 | 122 | 875031 | 65.645 | ng | 92 |
| 33) 4-Chloroaniline | 10.746 | 127 | 1251607 | 40.788 | ng | 97 |
| 34) Hexachlorobutadiene | 10.940 | 225 | 645550 | 38.319 | ng | 98 |
| 35) Caprolactam | 11.534 | 113 | 416030 | 47.892 | ng | 88 |
| 36) 4-Chloro-3-methylphenol | 11.881 | 107 | 1284015 | 44.353 | ng | 96 |
| 37) 2-Methylnaphthalene | 12.263 | 142 | 2186594 | 39.219 | ng | 100 |
| 38) 1-Methylnaphthalene | 12.481 | 142 | 2166454 | 39.399 | ng | 99 |
| 40) 1,2,4,5-Tetrachloroben... | 12.634 | 216 | 1175676 | 37.541 | ng | 100 |
| 41) Hexachlorocyclopentadiene | 12.616 | 237 | 464856 | 46.834 | ng | 99 |
| 43) 2,4,6-Trichlorophenol | 12.875 | 196 | 805218 | 46.287 | ng | 99 |

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP082024\
 Data File : BP021575.D
 Acq On : 20 Aug 2024 16:40
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040

Quant Time: Aug 20 17:15:31 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 14 00:00:30 2024
 Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/21/2024
 Supervised By :mohammad ahmed 08/22/2024

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|-------------------------------|--------|------|----------|--------|-------|----------|
| 44) 2,4,5-Trichlorophenol | 12.946 | 196 | 901983 | 46.667 | ng | 97 |
| 46) 1,1'-Biphenyl | 13.281 | 154 | 2870734 | 37.962 | ng | 98 |
| 47) 2-Chloronaphthalene | 13.322 | 162 | 2258356 | 38.359 | ng | 99 |
| 48) 2-Nitroaniline | 13.516 | 65 | 775883 | 47.126 | ng | 88 |
| 49) Acenaphthylene | 14.181 | 152 | 3412612 | 39.430 | ng | 100 |
| 50) Dimethylphthalate | 13.916 | 163 | 2903067 | 40.557 | ng | 100 |
| 51) 2,6-Dinitrotoluene | 14.022 | 165 | 668456 | 46.431 | ng | 91 |
| 52) Acenaphthene | 14.534 | 154 | 2344983m | 41.294 | ng | |
| 53) 3-Nitroaniline | 14.351 | 138 | 704846 | 49.513 | ng | # 90 |
| 54) 2,4-Dinitrophenol | 14.569 | 184 | 301078 | 66.049 | ng | 90 |
| 55) Dibenzofuran | 14.869 | 168 | 3469479 | 39.716 | ng | 99 |
| 56) 4-Nitrophenol | 14.663 | 139 | 623596 | 54.273 | ng | 88 |
| 57) 2,4-Dinitrotoluene | 14.828 | 165 | 940265 | 50.859 | ng | # 85 |
| 58) Fluorene | 15.534 | 166 | 2801684 | 39.039 | ng | 100 |
| 59) 2,3,4,6-Tetrachlorophenol | 15.092 | 232 | 787234 | 47.986 | ng | 98 |
| 60) Diethylphthalate | 15.304 | 149 | 3034612 | 41.319 | ng | 100 |
| 61) 4-Chlorophenyl-phenyle... | 15.528 | 204 | 1408811 | 37.977 | ng | 99 |
| 62) 4-Nitroaniline | 15.539 | 138 | 744682 | 50.846 | ng | 92 |
| 63) Azobenzene | 15.822 | 77 | 3217680 | 35.641 | ng | 97 |
| 65) 4,6-Dinitro-2-methylph... | 15.604 | 198 | 455554 | 58.431 | ng | 99 |
| 66) n-Nitrosodiphenylamine | 15.739 | 169 | 2431907 | 37.750 | ng | 100 |
| 67) 4-Bromophenyl-phenylether | 16.439 | 248 | 941570 | 37.008 | ng | 99 |
| 68) Hexachlorobenzene | 16.563 | 284 | 1154799 | 38.408 | ng | 99 |
| 69) Atrazine | 16.722 | 200 | 704532 | 31.853 | ng | 100 |
| 70) Pentachlorophenol | 16.916 | 266 | 808216 | 49.250 | ng | 98 |
| 71) Phenanthrene | 17.322 | 178 | 4562430 | 39.464 | ng | 100 |
| 72) Anthracene | 17.416 | 178 | 4446404 | 39.216 | ng | 100 |
| 73) Carbazole | 17.686 | 167 | 4532118 | 41.477 | ng | 99 |
| 74) Di-n-butylphthalate | 18.298 | 149 | 5580792 | 41.587 | ng | 100 |
| 75) Fluoranthene | 19.404 | 202 | 5912433 | 41.680 | ng | 100 |
| 77) Benzidine | 19.592 | 184 | 2026298 | 35.722 | ng | 100 |
| 78) Pyrene | 19.780 | 202 | 6147221 | 38.762 | ng | 99 |
| 80) Butylbenzylphthalate | 20.722 | 149 | 2641829 | 44.200 | ng | 97 |
| 81) Benzo(a)anthracene | 21.716 | 228 | 6085084 | 39.950 | ng | 100 |
| 82) 3,3'-Dichlorobenzidine | 21.627 | 252 | 2216157 | 45.379 | ng | 100 |
| 83) Chrysene | 21.786 | 228 | 5688074 | 39.553 | ng | 99 |
| 84) Bis(2-ethylhexyl)phtha... | 21.663 | 149 | 4058627 | 45.083 | ng | 99 |
| 85) Di-n-octyl phthalate | 22.992 | 149 | 7075138 | 45.068 | ng | 98 |
| 87) Indeno(1,2,3-cd)pyrene | 29.109 | 276 | 7692660m | 40.361 | ng | |
| 88) Benzo(b)fluoranthene | 24.092 | 252 | 6649815 | 39.379 | ng | 99 |
| 89) Benzo(k)fluoranthene | 24.168 | 252 | 6261875 | 38.012 | ng | 100 |
| 90) Benzo(a)pyrene | 25.021 | 252 | 5826473 | 39.759 | ng | 99 |
| 91) Dibenzo(a,h)anthracene | 29.221 | 278 | 6298888 | 39.775 | ng | 100 |
| 92) Benzo(g,h,i)perylene | 30.303 | 276 | 6370124 | 40.133 | ng | 99 |

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP082024\
Data File : BP021575.D
Acq On : 20 Aug 2024 16:40
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1


Quant Time: Aug 20 17:15:31 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 14 00:00:30 2024
Response via : Initial Calibration

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/21/2024
Supervised By :mohammad ahmed 08/22/2024







#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.805 min Scan# 8
Delta R.T. -0.006 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40

Instrument : BNA_P
ClientSampleId : SSTDCCC040


Manual Integrations
APPROVED




Reviewed By :Yogesh Patel 08/21/2024
Supervised By :mohammad ahmed 08/22/2024



#2
1,4-Dioxane
Concen: 35.681 ng
RT: 3.317 min Scan# 56
Delta R.T. -0.000 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40

Tgt Ion: 88 Resp: 424462
Ion Ratio Lower Upper
88 100
58 61.5 48.7 73.1
43 20.1 15.4 23.2









#3
Pyridine
Concen: 36.406 ng
RT: 3.717 min Scan# 1
Delta R.T. 0.012 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40

Instrument : BNA_P
ClientSampleId : SSTDCCCC040


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/21/2024
Supervised By :mohammad ahmed 08/22/2024



#4
n-Nitrosodimethylamine
Concen: 37.206 ng
RT: 3.622 min Scan# 108
Delta R.T. 0.006 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40

Tgt Ion: 42 Resp: 371373
Ion Ratio Lower Upper
42 100
74 185.7 154.3 231.5
44 6.5 6.2 9.4







#5
2-Fluorophenol
Concen: 81.506 ng
RT: 5.399 min Scan# 4
Instrument : BNA_P
Delta R.T. 0.006 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40

ClientSampleId : SSTDCCC040


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/21/2024
Supervised By :mohammad ahmed 08/22/2024



#6
Aniline
Concen: 36.863 ng
RT: 7.134 min Scan# 705
Delta R.T. -0.000 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40

Tgt Ion: 93 Resp: 1348749
Ion Ratio Lower Upper
93 100
66 35.1 30.2 45.2
65 17.5 14.7 22.1







#7
 Phenol-d6
 Concen: 80.016 ng
 RT: 6.969 min Scan# 6
 Delta R.T. 0.006 min
 Lab File: BP021575.D
 Acq: 20 Aug 2024 16:40

Instrument : BNA_P
 ClientSampleId : SSTDCCC040


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/21/2024
 Supervised By :mohammad ahmed 08/22/2024



#8
 2-Chlorophenol
 Concen: 42.226 ng
 RT: 7.375 min Scan# 746
 Delta R.T. -0.000 min
 Lab File: BP021575.D
 Acq: 20 Aug 2024 16:40

Tgt Ion:128 Resp: 979003
 Ion Ratio Lower Upper
 128 100
 130 31.6 11.6 51.6
 64 49.4 35.3 75.3







#9
 Benzaldehyde
 Concen: 44.262 ng
 RT: 6.946 min Scan# 6
 Delta R.T. -0.000 min
 Lab File: BP021575.D
 Acq: 20 Aug 2024 16:40

Instrument : BNA_P
 ClientSampleId : SSTDCCC040


Manual Integrations
APPROVED







Reviewed By :Yogesh Patel 08/21/2024
 Supervised By :mohammad ahmed 08/22/2024



#10
 Phenol
 Concen: 39.093 ng
 RT: 6.999 min Scan# 682
 Delta R.T. 0.012 min
 Lab File: BP021575.D
 Acq: 20 Aug 2024 16:40

Tgt Ion: 94 Resp: 1481572
 Ion Ratio Lower Upper
 94 100
 65 24.1 6.5 46.5
 66 32.5 14.7 54.7






#11
bis(2-Chloroethyl)ether
Concen: 36.557 ng
RT: 7.228 min Scan# 7
Delta R.T. -0.000 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40

Instrument : BNA_P
ClientSampleId : SSTDCCC040


Manual Integrations APPROVED







Reviewed By :Yogesh Patel 08/21/2024
Supervised By :mohammad ahmed 08/22/2024



#12
1,3-Dichlorobenzene
Concen: 39.364 ng
RT: 7.699 min Scan# 801
Delta R.T. -0.000 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40

Tgt Ion:146 Resp: 1084084
Ion Ratio Lower Upper
146 100
148 63.8 51.8 77.6
75 34.8 31.3 46.9





#13

1,4-Dichlorobenzene

Concen: 39.661 ng

RT: 7.840 min Scan# 8

Delta R.T. -0.000 min

Lab File: BP021575.D

Acq: 20 Aug 2024 16:40

Instrument :

BNA_P


ClientSampleId :

SSTDCCC040

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 08/21/2024

Supervised By :mohammad ahmed 08/22/2024



#14

1,2-Dichlorobenzene

Concen: 38.940 ng

RT: 8.158 min Scan# 879

Delta R.T. -0.000 min

Lab File: BP021575.D

Acq: 20 Aug 2024 16:40


Tgt Ion:146 Resp: 1043542




Ion Ratio Lower Upper

146 100

148 65.0 51.5 77.3

111 50.9 41.1 61.7





#15

Benzyl Alcohol

Concen: 38.673 ng

RT: 8.040 min Scan# 8

Delta R.T. 0.006 min

Lab File: BP021575.D

Acq: 20 Aug 2024 16:40

Instrument :

BNA_P





ClientSampleId :

SSTDCCC040

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/21/2024

Supervised By :mohammad ahmed 08/22/2024



#16

2,2'-oxybis(1-Chloropropane)

Concen: 35.934 ng

RT: 8.334 min Scan# 909

Delta R.T. -0.000 min

Lab File: BP021575.D

Acq: 20 Aug 2024 16:40


Tgt Ion: 45 Resp: 1061205




Ion Ratio Lower Upper

45 100

77 20.0 1.9 41.9

79 15.8 0.0 36.6





#17

2-Methylphenol

Concen: 40.769 ng

RT: 8.240 min Scan# 8

Delta R.T. 0.006 min

Lab File: BP021575.D

Acq: 20 Aug 2024 16:40

Instrument :

BNA_P




ClientSampleId :

SSTDCCC040

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/21/2024

Supervised By :mohammad ahmed 08/22/2024



#18

Hexachloroethane

Concen: 40.786 ng

RT: 8.887 min Scan# 1003

Delta R.T. -0.006 min

Lab File: BP021575.D

Acq: 20 Aug 2024 16:40


Tgt Ion:117 Resp: 454819




Ion Ratio Lower Upper

117 100

119 94.3 73.9 110.9

201 95.3 74.9 112.3









#19

n-Nitroso-di-n-propylamine
Concen: 33.637 ng
RT: 8.605 min Scan# 948
Delta R.T. -0.000 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Manual Integrations APPROVED


Reviewed By :Yogesh Patel 08/21/2024
Supervised By :mohammad ahmed 08/22/2024




#20

3+4-Methylphenols
Concen: 41.976 ng
RT: 8.563 min Scan# 948
Delta R.T. 0.006 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40

Tgt Ion:107 Resp: 1356088
Ion Ratio Lower Upper
107 100
108 87.5 67.6 107.6
77 32.4 15.0 55.0
79 28.2 8.9 48.9







#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.593 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40

Instrument : BNA_P
ClientSampleId : SSTDCCC040


Manual Integrations APPROVED




Reviewed By :Yogesh Patel 08/21/2024
Supervised By :mohammad ahmed 08/22/2024



#22
Acetophenone
Concen: 37.514 ng
RT: 8.616 min Scan# 957
Delta R.T. -0.000 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40

Tgt Ion:105 Resp: 1802202
Ion Ratio Lower Upper
105 100
71 5.0 4.6 7.0
51 19.7 16.3 24.5
120 20.7 16.2 24.2





#23

Nitrobenzene-d5

Concen: 86.924 ng

RT: 8.952 min Scan# 1021

Delta R.T. -0.000 min

Lab File: BP021575.D

Acq: 20 Aug 2024 16:40





Instrument : BNA_P

ClientSampleId : SSTDCCC040

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/21/2024

Supervised By :mohammad ahmed 08/22/2024



#24

Nitrobenzene

Concen: 40.876 ng

RT: 8.993 min Scan# 1021

Delta R.T. -0.000 min

Lab File: BP021575.D

Acq: 20 Aug 2024 16:40


Tgt Ion: 77 Resp: 1345015




Ion Ratio Lower Upper

77 100

123 42.6 29.5 44.3

65 12.3 9.9 14.9









#25
Isophorone
Concen: 35.467 ng
RT: 9.522 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40

Instrument : BNA_P
ClientSampleId : SSTDCCC040


Manual Integrations
APPROVED







Reviewed By :Yogesh Patel 08/21/2024
Supervised By :mohammad ahmed 08/22/2024



#26
2-Nitrophenol
Concen: 56.363 ng
RT: 9.704 min Scan# 1142
Delta R.T. 0.005 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40

Tgt Ion:139 Resp: 508327
Ion Ratio Lower Upper
139 100
109 34.3 28.9 43.3
65 53.4 49.0 73.6





#27

2,4-Dimethylphenol

Concen: 40.260 ng

RT: 9.769 min Scan# 1

Delta R.T. 0.012 min

Lab File: BP021575.D

Acq: 20 Aug 2024 16:40

Instrument :

BNA_P


ClientSampleId :

SSTDCCC040

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/21/2024

Supervised By :mohammad ahmed 08/22/2024



#28

bis(2-Chloroethoxy)methane

Concen: 36.474 ng

RT: 9.999 min Scan# 1192

Delta R.T. -0.006 min

Lab File: BP021575.D

Acq: 20 Aug 2024 16:40


Tgt Ion: 93 Resp: 1593547


Ion Ratio Lower Upper

93 100

95 32.2 26.2 39.4

123 9.7 7.6 11.4







#29
2,4-Dichlorophenol
Concen: 45.397 ng
RT: 10.240 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40

Instrument : BNA_P
ClientSampleId : SSTDCCC040


Manual Integrations
APPROVED




Reviewed By :Yogesh Patel 08/21/2024
Supervised By :mohammad ahmed 08/22/2024



#30
1,2,4-Trichlorobenzene
Concen: 38.758 ng
RT: 10.457 min Scan# 1270
Delta R.T. -0.000 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40

Tgt Ion:180 Resp: 1037019
Ion Ratio Lower Upper
180 100
182 95.4 78.3 117.5
145 30.7 24.0 36.0





#31

Naphthalene

Concen: 39.419 ng

RT: 10.646 min Scan# 1

Delta R.T. -0.000 min

Lab File: BP021575.D

Acq: 20 Aug 2024 16:40

Instrument:

BNA_P

ClientSampleId :

SSTDCCC040

Tgt Ion:128 Resp: 322020

Ion Ratio Lower Upper

128 100

129 10.6 8.6 12.8





127 13.0 10.6 15.8

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 08/21/2024

Supervised By :mohammad ahmed 08/22/2024



#32

Benzoic acid

Concen: 65.645 ng

RT: 9.905 min Scan# 1176

Delta R.T. 0.030 min

Lab File: BP021575.D

Acq: 20 Aug 2024 16:40


Tgt Ion:122 Resp: 875031

Ion Ratio Lower Upper

122 100

105 137.7 124.6 164.6

77 103.6 96.2 136.2



#33

4-Chloroaniline

Concen: 40.788 ng

RT: 10.746 min Scan# 1




Delta R.T. 0.006 min

Lab File: BP021575.D

Acq: 20 Aug 2024 16:40

Instrument : BNA_P

ClientSampleId : SSTDCCC040



Tgt Ion:127 Resp: 125160





Ion Ratio Lower Upper

| | | | |
|-----|------|------|------|
| 127 | 100 | | |
| 129 | 31.8 | 24.5 | 36.7 |
| 65 | 33.2 | 28.9 | 43.3 |
| 92 | 23.4 | 19.3 | 28.9 |

Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel 08/21/2024

Supervised By :mohammad ahmed 08/22/2024



#34

Hexachlorobutadiene

Concen: 38.319 ng

RT: 10.940 min Scan# 1352

Delta R.T. -0.000 min


Lab File: BP021575.D

Acq: 20 Aug 2024 16:40

Tgt Ion:225 Resp: 645550

Ion Ratio Lower Upper

| | | | |
|-----|------|------|------|
| 225 | 100 | | |
| 223 | 62.5 | 49.6 | 74.4 |
| 227 | 64.6 | 50.1 | 75.1 |



#35

Caprolactam

Concen: 47.892 ng

RT: 11.534 min Scan# 1

Delta R.T. 0.035 min

Lab File: BP021575.D




Acq: 20 Aug 2024 16:40

Instrument :

BNA_P

ClientSampleId :

SSTDCCC040



Tgt Ion:113 Resp: 416030

Ion Ratio Lower Upper

113 100





55 133.6 130.6 170.6

56 102.6 94.6 134.6

Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel 08/21/2024

Supervised By :mohammad ahmed 08/22/2024



#36

4-Chloro-3-methylphenol

Concen: 44.353 ng

RT: 11.881 min Scan# 1512

Delta R.T. 0.018 min

Lab File: BP021575.D

Acq: 20 Aug 2024 16:40


Tgt Ion:107 Resp: 1284015


Ion Ratio Lower Upper

107 100

144 21.5 16.6 24.8

142 67.2 51.0 76.4





#37

2-Methylnaphthalene

Concen: 39.219 ng

RT: 12.263 min Scan# 1

Delta R.T. -0.000 min

Lab File: BP021575.D

Acq: 20 Aug 2024 16:40

Instrument :

BNA_P


ClientSampleId :

SSTDCCC040

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/21/2024

Supervised By :mohammad ahmed 08/22/2024



#38

1-Methylnaphthalene

Concen: 39.399 ng

RT: 12.481 min Scan# 1614

Delta R.T. -0.000 min

Lab File: BP021575.D

Acq: 20 Aug 2024 16:40


Tgt Ion:142 Resp: 2166454

Ion Ratio Lower Upper

142 100

141 92.7 73.5 110.3

116 4.5 3.4 5.2



#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 14.463 min Scan# 1

Delta R.T. 0.006 min

Lab File: BP021575.D




Acq: 20 Aug 2024 16:40

Instrument :

BNA_P

ClientSampleId :

SSTDCCC040



Tgt Ion:164 Resp: 1068840

Ion Ratio Lower Upper

164 100





162 100.1 81.5 122.3

160 43.4 35.2 52.8

Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel 08/21/2024

Supervised By :mohammad ahmed 08/22/2024



#40

1,2,4,5-Tetrachlorobenzene

Concen: 37.541 ng

RT: 12.634 min Scan# 1640

Delta R.T. 0.006 min

Lab File: BP021575.D

Acq: 20 Aug 2024 16:40

Tgt Ion:216 Resp: 1175676


Ion Ratio Lower Upper


216 100

214 78.2 62.5 93.7

179 20.6 16.4 24.6

108 20.7 16.7 25.1







#41
Hexachlorocyclopentadiene
Concen: 46.834 ng
RT: 12.616 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40

Instrument : BNA_P
ClientSampleId : SSTDCCC040


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/21/2024
Supervised By :mohammad ahmed 08/22/2024



#42
2,4,6-Tribromophenol
Concen: 95.320 ng
RT: 15.981 min Scan# 2209
Delta R.T. 0.012 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40

Tgt Ion:330 Resp: 1133914
Ion Ratio Lower Upper
330 100
332 97.2 77.4 116.0
141 31.0 24.5 36.7







#43
2,4,6-Trichlorophenol
Concen: 46.287 ng
RT: 12.875 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40

Instrument : BNA_P
ClientSampleId : SSTDCCC040


Manual Integrations
APPROVED




Reviewed By :Yogesh Patel 08/21/2024
Supervised By :mohammad ahmed 08/22/2024



#44
2,4,5-Trichlorophenol
Concen: 46.667 ng
RT: 12.946 min Scan# 1693
Delta R.T. 0.012 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40

Tgt Ion:196 Resp: 901983
Ion Ratio Lower Upper
196 100
198 98.0 78.9 118.3
97 56.1 49.4 74.0
132 26.1 21.1 31.7





#45

2-Fluorobiphenyl

Concen: 73.981 ng

RT: 13.063 min Scan# 1

Delta R.T. -0.000 min

Lab File: BP021575.D

Acq: 20 Aug 2024 16:40

Instrument:

BNA_P





ClientSampleId :

SSTDCCC040

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/21/2024

Supervised By :mohammad ahmed 08/22/2024



#46

1,1'-Biphenyl

Concen: 37.962 ng

RT: 13.281 min Scan# 1750

Delta R.T. -0.000 min

Lab File: BP021575.D

Acq: 20 Aug 2024 16:40


Tgt Ion:154 Resp: 2870734

Ion Ratio Lower Upper

154 100

153 40.5 20.7 60.7

76 15.7 0.0 37.7



#47

2-Chloronaphthalene

Concen: 38.359 ng

RT: 13.322 min Scan# 1

Delta R.T. 0.006 min

Lab File: BP021575.D

Acq: 20 Aug 2024 16:40

Instrument :

BNA_P

ClientSampleId :

SSTDCCC040

Tgt Ion:162 Resp: 2258350

Ion Ratio Lower Upper

162 100

127 39.9 31.9 47.9

164 32.4 26.8 40.2

Manual Integrations

APPROVED



Reviewed By :Yogesh Patel 08/21/2024

Supervised By :mohammad ahmed 08/22/2024

Abundance

Scan 1757 (13.322 min): BP021575.D\data.ms (-)

162.1



Abundance

Scan 1788 (13.504 min): BP021485.D\data.ms (-)

65.0 138.0

#48
2-Nitroaniline
Concen: 47.126 ng
RT: 13.516 min Scan# 1790
Delta R.T. 0.012 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40

Tgt Ion: 65 Resp: 775883

Ion Ratio Lower Upper

65 100





92 78.4 58.1 87.1

138 97.3 65.5 98.3

Abundance

Scan 1790 (13.516 min): BP021575.D\data.ms (-)

65.0 138.1



#49

Acenaphthylene

Concen: 39.430 ng

RT: 14.181 min Scan# 1

Delta R.T. 0.012 min

Lab File: BP021575.D

Acq: 20 Aug 2024 16:40

Instrument :

BNA_P

ClientSampleId :

SSTDCCC040

Tgt Ion:152 Resp: 341261

Ion Ratio Lower Upper

152 100

151 19.8 15.9 23.9

153 13.1 10.6 16.0

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 08/21/2024

Supervised By :mohammad ahmed 08/22/2024

Abundance


2000000 14.181

1500000

1000000

500000

0



#50

Dimethylphthalate

Concen: 40.557 ng

RT: 13.916 min Scan# 1858

Delta R.T. 0.006 min

Lab File: BP021575.D

Acq: 20 Aug 2024 16:40

Tgt Ion:163 Resp: 2903067

Ion Ratio Lower Upper

163 100

194 4.2 3.3 4.9

164 10.3 8.3 12.5

Abundance




2000000 13.916


1500000

1000000

500000

0







#51
2,6-Dinitrotoluene
Concen: 46.431 ng
RT: 14.022 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40

Instrument : BNA_P
ClientSampleId : SSTDCCC040


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 08/21/2024
Supervised By :mohammad ahmed 08/22/2024



#52
Acenaphthene
Concen: 41.294 ng
RT: 14.534 min Scan# 1963
Delta R.T. 0.012 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40

Tgt Ion:154 Resp: 2344983
Ion Ratio Lower Upper
154 100
153 112.8 91.1 136.7
152 53.0 43.0 64.4







#53
3-Nitroaniline
Concen: 49.513 ng
RT: 14.351 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40

Instrument : BNA_P
ClientSampleId : SSTDCCC040


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/21/2024
Supervised By :mohammad ahmed 08/22/2024



#54
2,4-Dinitrophenol
Concen: 66.049 ng
RT: 14.569 min Scan# 1969
Delta R.T. 0.018 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40

Tgt Ion:184 Resp: 301078
Ion Ratio Lower Upper
184 100
63 67.1 64.9 97.3
154 63.6 52.1 78.1







#55
Dibenzofuran
Concen: 39.716 ng
RT: 14.869 min Scan# 2
Delta R.T. 0.012 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40

Instrument : BNA_P
ClientSampleId : SSTDCCC040


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/21/2024
Supervised By :mohammad ahmed 08/22/2024



#56
4-Nitrophenol
Concen: 54.273 ng
RT: 14.663 min Scan# 1985
Delta R.T. 0.012 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40

Tgt Ion:139 Resp: 623596
Ion Ratio Lower Upper
139 100
109 82.5 70.2 110.2
65 104.4 101.8 141.8




#57
 2,4-Dinitrotoluene
 Concen: 50.859 ng
 RT: 14.828 min Scan# 2133
 Delta R.T. 0.018 min
 Lab File: BP021575.D
 Acq: 20 Aug 2024 16:40

Instrument : BNA_P
 ClientSampleId : SSTDCCC040

| Tgt | Ion:165 | Resp: | 940269 |
|-----|---------|-------|--------|
| Ion | Ratio | Lower | Upper |
| 165 | 100 | | |
| 63 | 42.2 | 42.5 | 63.7 |
| 89 | 81.5 | 77.1 | 115.7 |
| 182 | 2.7 | 2.0 | 3.0 |




Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/21/2024
 Supervised By :mohammad ahmed 08/22/2024



#58
 Fluorene
 Concen: 39.039 ng
 RT: 15.534 min Scan# 2133
 Delta R.T. 0.018 min
 Lab File: BP021575.D
 Acq: 20 Aug 2024 16:40

| Tgt | Ion:166 | Resp: | 2801684 |
|-----|---------|-------|---------|
| Ion | Ratio | Lower | Upper |
| 166 | 100 | | |
| 165 | 97.4 | 78.1 | 117.1 |
| 167 | 13.5 | 11.0 | 16.4 |




#59
 2,3,4,6-Tetrachlorophenol
 Concen: 47.986 ng
 RT: 15.092 min Scan# 2
 Delta R.T. 0.006 min
 Lab File: BP021575.D
 Acq: 20 Aug 2024 16:40

Instrument : BNA_P
 ClientSampleId : SSTDCCC040

Tgt Ion:232 Resp: 787234
 Ion Ratio Lower Upper
 232 100
 131 42.3 33.1 49.7
 130 2.3 1.7 2.5
 166 29.8 23.0 34.4




Manual Integrations APPROVED


Reviewed By :Yogesh Patel 08/21/2024
 Supervised By :mohammad ahmed 08/22/2024



#60
 Diethylphthalate
 Concen: 41.319 ng
 RT: 15.304 min Scan# 2094
 Delta R.T. -0.000 min
 Lab File: BP021575.D
 Acq: 20 Aug 2024 16:40

Tgt Ion:149 Resp: 3034612
 Ion Ratio Lower Upper
 149 100
 177 22.1 17.8 26.8
 150 12.7 10.1 15.1







#61
4-Chlorophenyl-phenylether
Concen: 37.977 ng
RT: 15.528 min Scan# 2132
Delta R.T. 0.006 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40

Instrument : BNA_P
ClientSampleId : SSTDCCC040


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/21/2024
Supervised By :mohammad ahmed 08/22/2024



#62
4-Nitroaniline
Concen: 50.846 ng
RT: 15.539 min Scan# 2134
Delta R.T. 0.011 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40

Tgt Ion:138 Resp: 744682
Ion Ratio Lower Upper
138 100
92 59.5 46.4 86.4
108 86.0 73.2 113.2







#63
Azobenzene
Concen: 35.641 ng
RT: 15.822 min Scan# 2
Delta R.T. -0.000 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40

Instrument : BNA_P
ClientSampleId : SSTDCCC040


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/21/2024
Supervised By :mohammad ahmed 08/22/2024



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.275 min Scan# 2429
Delta R.T. 0.023 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40

Tgt Ion:188 Resp: 2403929
Ion Ratio Lower Upper
188 100
94 12.2 10.2 15.4
80 13.1 11.6 17.4




#65
 4,6-Dinitro-2-methylphenol
 Concen: 58.431 ng
 RT: 15.604 min Scan# 2142
 Delta R.T. 0.017 min
 Lab File: BP021575.D
 Acq: 20 Aug 2024 16:40

Instrument : BNA_P
 ClientSampleId : SSTDCCC040

Tgt Ion:198 Resp: 455554
 Ion Ratio Lower Upper
 198 100
 51 40.2 21.5 61.5
 105 52.2 32.9 72.9






Manual Integrations APPROVED

Reviewed By :Yogesh Patel 08/21/2024
 Supervised By :mohammad ahmed 08/22/2024



#66
 n-Nitrosodiphenylamine
 Concen: 37.750 ng
 RT: 15.739 min Scan# 2168
 Delta R.T. -0.000 min
 Lab File: BP021575.D
 Acq: 20 Aug 2024 16:40

Tgt Ion:169 Resp: 2431907
 Ion Ratio Lower Upper
 169 100
 168 66.8 53.4 80.2
 167 36.3 28.7 43.1




#67
 4-Bromophenyl-phenylether
 Concen: 37.008 ng
 RT: 16.439 min Scan# 2
 Delta R.T. -0.000 min
 Lab File: BP021575.D
 Acq: 20 Aug 2024 16:40

Instrument : BNA_P
 ClientSampleId : SSTDCCC040

Tgt Ion:248 Resp: 941570
 Ion Ratio Lower Upper
 248 100
 250 97.9 78.6 117.8
 141 69.2 53.9 80.9



Manual Integrations APPROVED




Reviewed By :Yogesh Patel 08/21/2024
 Supervised By :mohammad ahmed 08/22/2024



#68
 Hexachlorobenzene
 Concen: 38.408 ng
 RT: 16.563 min Scan# 2308
 Delta R.T. 0.012 min
 Lab File: BP021575.D
 Acq: 20 Aug 2024 16:40

Tgt Ion:284 Resp: 1154799
 Ion Ratio Lower Upper
 284 100
 142 30.3 23.4 35.2
 249 30.6 24.1 36.1





#69

Atrazine

Concen: 31.853 ng

RT: 16.722 min Scan# 2

Delta R.T. 0.006 min

Lab File: BP021575.D

Acq: 20 Aug 2024 16:40

Instrument :

BNA_P





ClientSampleId :

SSTDCCC040

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/21/2024

Supervised By :mohammad ahmed 08/22/2024



#70

Pentachlorophenol

Concen: 49.250 ng

RT: 16.916 min Scan# 2368

Delta R.T. 0.024 min

Lab File: BP021575.D

Acq: 20 Aug 2024 16:40


Tgt Ion:266 Resp: 808216


Ion Ratio Lower Upper

266 100


268 65.9 51.8 77.6

264 65.3 50.5 75.7





Abundance Scan 2437 (17.322 min): BP021575.D\data.ms



#71

Phenanthrene

Concen: 39.464 ng

RT: 17.322 min Scan# 2437

Delta R.T. 0.023 min

Lab File: BP021575.D

Acq: 20 Aug 2024 16:40

Instrument :

BNA_P


ClientSampleId :

SSTDCCC040

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/21/2024

Supervised By :mohammad ahmed 08/22/2024



#72

Anthracene

Concen: 39.216 ng

RT: 17.416 min Scan# 2453

Delta R.T. 0.018 min

Lab File: BP021575.D

Acq: 20 Aug 2024 16:40


Tgt Ion:178 Resp: 4446404


Ion Ratio Lower Upper

178 100


176 18.5 14.9 22.3

179 15.3 12.5 18.7






Abundance Scan 2499 (17.686 min): BP021575.D\data.ms (-)



Abundance Scan 2499 (17.686 min): BP021575.D\data.ms (-)



#73

Carbazole

Concen: 41.477 ng

RT: 17.686 min Scan# 2

Delta R.T. 0.018 min

Lab File: BP021575.D

Acq: 20 Aug 2024 16:40

Instrument :

BNA_P

ClientSampleId :

SSTDCCC040

Tgt Ion:167 Resp: 453211:

Ion Ratio Lower Upper

167 100

166 21.7 17.0 25.4


139 13.0 9.9 14.9

Manual Integrations


APPROVED

Reviewed By :Yogesh Patel 08/21/2024


Supervised By :mohammad ahmed 08/22/2024



Abundance Scan 2603 (18.298 min): BP021575.D\data.ms (-)



Abundance Scan 2603 (18.298 min): BP021575.D\data.ms (-)



#74

Di-n-butylphthalate

Concen: 41.587 ng

RT: 18.298 min Scan# 2603

Delta R.T. 0.023 min

Lab File: BP021575.D

Acq: 20 Aug 2024 16:40


Tgt Ion:149 Resp: 5580792




Ion Ratio Lower Upper

149 100

150 9.3 7.5 11.3

104 5.9 4.9 7.3









#75
Fluoranthene
Concen: 41.680 ng
RT: 19.404 min Scan# 2
Delta R.T. 0.023 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40

Instrument : BNA_P
ClientSampleId : SSTDCCC040


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 08/21/2024
Supervised By :mohammad ahmed 08/22/2024



#76
Chrysene-d₁₂
Concen: 20.000 ng
RT: 21.739 min Scan# 3188
Delta R.T. 0.041 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40

Tgt Ion:240 Resp: 2426350
Ion Ratio Lower Upper
240 100
120 11.0 9.9 14.9
236 25.9 20.6 30.8





#77
 Benzidine
 Concen: 35.722 ng
 RT: 19.592 min Scan# 21
 Delta R.T. 0.017 min
 Lab File: BP021575.D
 Acq: 20 Aug 2024 16:40

Instrument : BNA_P
 ClientSampleId : SSTDCCC040

Tgt Ion:184 Resp: 202629:
 Ion Ratio Lower Upper
 184 100
 185 14.1 11.4 17.0
 183 11.7 9.3 13.9



Manual Integrations APPROVED

Reviewed By :Yogesh Patel 08/21/2024
 Supervised By :mohammad ahmed 08/22/2024



#78
 Pyrene
 Concen: 38.762 ng
 RT: 19.780 min Scan# 2855
 Delta R.T. 0.029 min
 Lab File: BP021575.D
 Acq: 20 Aug 2024 16:40

Tgt Ion:202 Resp: 6147221
 Ion Ratio Lower Upper
 202 100
 200 20.6 16.7 25.1
 203 17.4 14.1 21.1



#79

Terphenyl-d14

Concen: 74.447 ng

RT: 19.998 min Scan# 2

Delta R.T. 0.023 min

Lab File: BP021575.D

Acq: 20 Aug 2024 16:40

Instrument :

BNA_P

ClientSampleId :

SSTDCCC040

Tgt Ion:244 Resp: 929797

Ion Ratio Lower Upper

244 100

212 7.0 5.6 8.4

122 11.3 9.4 14.2

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 08/21/2024

Supervised By :mohammad ahmed 08/22/2024


Abundance

19.998

4000000

2000000

0



#80

Butylbenzylphthalate

Concen: 44.200 ng

RT: 20.722 min Scan# 3015

Delta R.T. 0.012 min

Lab File: BP021575.D

Acq: 20 Aug 2024 16:40

Tgt Ion:149 Resp: 2641829

Ion Ratio Lower Upper

149 100

91 88.4 73.3 109.9

206 21.8 17.1 25.7

Abundance

20.722




2000000


1500000

1000000

500000

0







#81
Benzo(a)anthracene
Concen: 39.950 ng
RT: 21.716 min Scan# 3
Delta R.T. 0.035 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40

Instrument : BNA_P
ClientSampleId : SSTDCCC040


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/21/2024
Supervised By :mohammad ahmed 08/22/2024



#82
3,3'-Dichlorobenzidine
Concen: 45.379 ng
RT: 21.627 min Scan# 3169
Delta R.T. 0.018 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40

Tgt Ion:252 Resp: 2216157
Ion Ratio Lower Upper
252 100
254 64.5 51.7 77.5
126 11.1 8.7 13.1



#83

Chrysene

Concen: 39.553 ng

RT: 21.786 min Scan# 3

Delta R.T. 0.035 min

Lab File: BP021575.D

Acq: 20 Aug 2024 16:40

Instrument :

BNA_P

ClientSampleId :

SSTDCCC040

Tgt Ion:228 Resp: 5688074

Ion Ratio Lower Upper

228 100

226 29.6 23.9 35.9

229 19.4 15.8 23.6

Manual Integrations


APPROVED

Reviewed By :Yogesh Patel 08/21/2024

Supervised By :mohammad ahmed 08/22/2024

Abundance

21.786



Time--> 21.70 21.80 21.90

Abundance

#84

Bis(2-ethylhexyl)phthalate

Concen: 45.083 ng

RT: 21.663 min Scan# 3175

Delta R.T. 0.018 min

Lab File: BP021575.D

Acq: 20 Aug 2024 16:40

Tgt Ion:149 Resp: 4058627

Ion Ratio Lower Upper


149 100

167 27.5 22.3 33.5

279 4.6 3.8 5.6

Abundance

21.663




Time--> 21.60 21.70

Abundance


21.751

Time--> 21.60 21.70



Abundance


21.751



Time--> 21.60 21.70

Abundance


21.663




Time--> 21.60 21.70

Abundance

21.751



Time--> 21.60 21.70





#85
Di-n-octyl phthalate
Concen: 45.068 ng
RT: 22.992 min Scan# 3401
Delta R.T. 0.029 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40

Instrument : BNA_P
ClientSampleId : SSTDCCC040


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 08/21/2024
Supervised By :mohammad ahmed 08/22/2024



#86
Perylene-d₁₂
Concen: 20.000 ng
RT: 25.180 min Scan# 3773
Delta R.T. 0.076 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40

Tgt Ion:264 Resp: 2940308
Ion Ratio Lower Upper
264 100
260 23.1 18.6 28.0
265 22.0 17.1 25.7



#87

Indeno(1,2,3-cd)pyrene

Concen: 40.361 ng m

RT: 29.109 min Scan# 4

Delta R.T. 0.118 min

Lab File: BP021575.D

Acq: 20 Aug 2024 16:40

Instrument :

BNA_P

ClientSampleId :

SSTDCCC040

Tgt Ion:276 Resp: 7692660

Ion Ratio Lower Upper

276 100

138 15.8 12.8 19.2

277 20.4 16.5 24.7

Manual Integrations

APPROVED


Reviewed By :Yogesh Patel 08/21/2024

Supervised By :mohammad ahmed 08/22/2024


Abundance

Scan 4441 (29.109 min): BP021575.D\data.ms (-)

276.1



| | Tgt Ion: | Ion Ratio | Lower | Upper |
|-----|----------|-----------|-------|-------|
| 276 | 100 | | | |
| 138 | 15.8 | 12.8 | 19.2 | |
| 277 | 20.4 | 16.5 | 24.7 | |



Time-->

Abundance

Scan 4441 (29.109 min): BP021575.D\data.ms (-)

276.1



m/z-->

Time-->

12

13

14

15

16

17

18

19

20

21

22

23

24

25

26

27

28

29

30

31

32

33

34

35

36

37

38

39

40

41

42

43

44

45

46

47

48

49

50

51

52

53

54

55

56

57

58

59

60

61

62

63

64

65

66

67

68

69

70

71

72

73

74

75

76

77

78

79

80

81

82

83

84

85

86

87

88

89

90

91

92

93

94

95

96

97

98

99

100

101

102

103

104

105

106

107

108

109

110

111

112

113

114

115

116

117

118

119

120

121

122

123

124

125

126

127

128

129

130

131

132

133

134

135

136

137

138

139

140

141

142

143

144

145

146

147

148

149

150

151

152

153

154

155

156

157

158

159

160

161

162

163

164

165

166

167

168

169

170

171

172

173

174

175

176

177

178

179

180

181

182

183

184

185

186

187

188

189

190

191

192

193

194

195

196

197

198

199

200

201

202

203

204

205

206

207

208

209

210

211

212

213

214

215

216

217

218

219

220

221

222

223

224

225

226

227

228

229

230

231

232

233

234

235

236

237

238

239

240

241

242

243

244

245

246

247

248

249

250

251

252

253

254

255

256

257

258

259

260

261

262

263

264

265

266

267

268

269

270

271

272

273

274

275

276

277

278

279

280

281

282

283

284

285

286

287

288

289

290

291

292

293

294

295

296

297

298

299

300

301

302

303

304

305

306

307

308

309

310

311

312

#89

Benzo(k)fluoranthene

Concen: 38.012 ng

RT: 24.168 min Scan# 3

Delta R.T. 0.059 min

Lab File: BP021575.D




Acq: 20 Aug 2024 16:40

Instrument :

BNA_P

ClientSampleId :

SSTDCCC040



Tgt Ion:252 Resp: 626187

Ion Ratio Lower Upper

252 100





253 22.0 17.5 26.3

125 9.7 7.8 11.8

Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel 08/21/2024

Supervised By :mohammad ahmed 08/22/2024



#90

Benzo(a)pyrene

Concen: 39.759 ng

RT: 25.021 min Scan# 3746

Delta R.T. 0.071 min

Lab File: BP021575.D

Acq: 20 Aug 2024 16:40


Tgt Ion:252 Resp: 5826473




Ion Ratio Lower Upper

252 100

253 22.2 17.7 26.5

125 12.0 8.7 13.1









#91
Dibenzo(a,h)anthracene
Concen: 39.775 ng
RT: 29.221 min Scan# 4460
Delta R.T. 0.135 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40

Instrument : BNA_P
ClientSampleId : SSTDCCC040


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 08/21/2024
Supervised By :mohammad ahmed 08/22/2024



#92
Benzo(g,h,i)perylene
Concen: 40.133 ng
RT: 30.303 min Scan# 4644
Delta R.T. 0.129 min
Lab File: BP021575.D
Acq: 20 Aug 2024 16:40

Tgt Ion:276 Resp: 6370124
Ion Ratio Lower Upper
276 100
277 23.9 19.0 28.6
138 19.0 16.0 24.0



Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP082024\
 Data File : BP021575.D
 Acq On : 20 Aug 2024 16:40
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 20 17:15:31 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 14 00:00:30 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

| | Compound | AvgRF | CCRF | %Dev | Area% | Dev(min) |
|------|-----------------------------|-------|-------|---------|-------|----------|
| 1 I | 1,4-Dichlorobenzene-d4 | 1.000 | 1.000 | 0.0 | 111 | 0.00 |
| 2 | 1,4-Dioxane | 0.639 | 0.570 | 10.8 | 99 | 0.00 |
| 3 | Pyridine | 1.783 | 1.623 | 9.0 | 100 | 0.01 |
| 4 | n-Nitrosodimethylamine | 0.536 | 0.498 | 7.1 | 104 | 0.00 |
| 5 S | 2-Fluorophenol | 1.347 | 1.372 | -1.9 | 113 | 0.00 |
| 6 | Aniline | 1.964 | 1.810 | 7.8 | 102 | 0.00 |
| 7 S | Phenol-d6 | 1.965 | 1.965 | 0.0 | 111 | 0.00 |
| 8 | 2-Chlorophenol | 1.245 | 1.314 | -5.5 | 117 | 0.00 |
| 9 | Benzaldehyde | 1.258 | 1.392 | -10.7 | 129 | 0.00 |
| 10 C | Phenol | 2.035 | 1.989 | 2.3 | 108 | 0.01 |
| 11 | bis(2-Chloroethyl)ether | 1.722 | 1.573 | 8.7 | 103 | 0.00 |
| 12 | 1,3-Dichlorobenzene | 1.479 | 1.455 | 1.6 | 112 | 0.00 |
| 13 C | 1,4-Dichlorobenzene | 1.493 | 1.480 | 0.9 | 112 | 0.00 |
| 14 | 1,2-Dichlorobenzene | 1.439 | 1.401 | 2.6 | 111 | 0.00 |
| 15 | Benzyl Alcohol | 1.410 | 1.363 | 3.3 | 107 | 0.00 |
| 16 | 2,2'-oxybis(1-Chloropropane | 1.585 | 1.424 | 10.2 | 103 | 0.00 |
| 17 | 2-Methylphenol | 1.286 | 1.311 | -1.9 | 113 | 0.00 |
| 18 | Hexachloroethane | 0.599 | 0.610 | -1.8 | 114 | 0.00 |
| 19 P | n-Nitroso-di-n-propylamine | 1.357 | 1.141 | 15.9 | 95 | 0.00 |
| 20 | 3+4-Methylphenols | 1.734 | 1.820 | -5.0 | 115 | 0.00 |
| 21 I | Naphthalene-d8 | 1.000 | 1.000 | 0.0 | 113 | 0.00 |
| 22 | Acetophenone | 0.614 | 0.576 | 6.2 | 107 | 0.00 |
| 23 S | Nitrobenzene-d5 | 0.387 | 0.421 | -8.8 | 116 | 0.00 |
| 24 | Nitrobenzene | 0.421 | 0.430 | -2.1 | 110 | 0.00 |
| 25 | Isophorone | 0.915 | 0.812 | 11.3 | 101 | 0.00 |
| 26 C | 2-Nitrophenol | 0.086 | 0.163 | -89.5# | 196# | 0.00 |
| 27 | 2,4-Dimethylphenol | 0.226 | 0.227 | -0.4 | 114 | 0.01 |
| 28 | bis(2-Chloroethoxy)methane | 0.559 | 0.510 | 8.8 | 105 | 0.00 |
| 29 C | 2,4-Dichlorophenol | 0.267 | 0.303 | -13.5 | 126 | 0.00 |
| 30 | 1,2,4-Trichlorobenzene | 0.342 | 0.332 | 2.9 | 111 | 0.00 |
| 31 | Naphthalene | 1.045 | 1.030 | 1.4 | 112 | 0.00 |
| 32 | Benzoic acid | 0.137 | 0.280 | -104.4# | 245# | 0.03 |
| 33 | 4-Chloroaniline | 0.392 | 0.400 | -2.0 | 115 | 0.00 |
| 34 C | Hexachlorobutadiene | 0.215 | 0.206 | 4.2 | 109 | 0.00 |
| 35 | Caprolactam | 0.111 | 0.133 | -19.8 | 132 | 0.04 |
| 36 C | 4-Chloro-3-methylphenol | 0.370 | 0.411 | -11.1 | 122 | 0.02 |
| 37 | 2-Methylnaphthalene | 0.713 | 0.699 | 2.0 | 112 | 0.00 |
| 38 | 1-Methylnaphthalene | 0.703 | 0.693 | 1.4 | 113 | 0.00 |
| 39 I | Acenaphthene-d10 | 1.000 | 1.000 | 0.0 | 120 | 0.00 |
| 40 | 1,2,4,5-Tetrachlorobenzene | 0.586 | 0.550 | 6.1 | 111 | 0.00 |
| 41 P | Hexachlorocyclopentadiene | 0.186 | 0.217 | -16.7 | 136 | 0.00 |
| 42 S | 2,4,6-Tribromophenol | 0.223 | 0.265 | -18.8 | 135 | 0.01 |
| 43 C | 2,4,6-Trichlorophenol | 0.326 | 0.377 | -15.6 | 132 | 0.00 |
| 44 | 2,4,5-Trichlorophenol | 0.362 | 0.422 | -16.6 | 132 | 0.01 |
| 45 S | 2-Fluorobiphenyl | 1.310 | 1.212 | 7.5 | 110 | 0.00 |
| 46 | 1,1'-Biphenyl | 1.415 | 1.343 | 5.1 | 113 | 0.00 |
| 47 | 2-Chloronaphthalene | 1.102 | 1.056 | 4.2 | 114 | 0.00 |

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP082024\
 Data File : BP021575.D
 Acq On : 20 Aug 2024 16:40
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 20 17:15:31 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 14 00:00:30 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

| | Compound | AvgRF | CCRF | %Dev | Area% | Dev(min) |
|------|----------------------------|-------|-------|---------|-------|----------|
| 48 | 2-Nitroaniline | 0.256 | 0.363 | -41.8# | 153# | 0.01 |
| 49 | Acenaphthylene | 1.619 | 1.596 | 1.4 | 117 | 0.01 |
| 50 | Dimethylphthalate | 1.339 | 1.358 | -1.4 | 121 | 0.00 |
| 51 | 2,6-Dinitrotoluene | 0.232 | 0.313 | -34.9# | 147 | 0.00 |
| 52 C | Acenaphthene | 1.063 | 1.097 | -3.2 | 123 | 0.01 |
| 53 | 3-Nitroaniline | 0.228 | 0.330 | -44.7# | 156# | 0.00 |
| 54 P | 2,4-Dinitrophenol | 0.069 | 0.141 | -104.3# | 258# | 0.02 |
| 55 | Dibenzofuran | 1.635 | 1.623 | 0.7 | 118 | 0.01 |
| 56 P | 4-Nitrophenol | 0.195 | 0.292 | -49.7# | 174# | 0.01 |
| 57 | 2,4-Dinitrotoluene | 0.284 | 0.440 | -54.9# | 165# | 0.02 |
| 58 | Fluorene | 1.343 | 1.311 | 2.4 | 116 | 0.02 |
| 59 | 2,3,4,6-Tetrachlorophenol | 0.307 | 0.368 | -19.9 | 135 | 0.00 |
| 60 | Diethylphthalate | 1.374 | 1.420 | -3.3 | 122 | 0.00 |
| 61 | 4-Chlorophenyl-phenylether | 0.694 | 0.659 | 5.0 | 114 | 0.00 |
| 62 | 4-Nitroaniline | 0.241 | 0.348 | -44.4# | 153# | 0.01 |
| 63 | Azobenzene | 1.689 | 1.505 | 10.9 | 105 | 0.00 |
| 64 I | Phenanthrene-d10 | 1.000 | 1.000 | 0.0 | 126 | 0.02 |
| 65 | 4,6-Dinitro-2-methylphenol | 0.054 | 0.095 | -75.9# | 229# | 0.02 |
| 66 c | n-Nitrosodiphenylamine | 0.536 | 0.506 | 5.6 | 119 | 0.00 |
| 67 | 4-Bromophenyl-phenylether | 0.212 | 0.196 | 7.5 | 119 | 0.00 |
| 68 | Hexachlorobenzene | 0.250 | 0.240 | 4.0 | 120 | 0.01 |
| 69 | Atrazine | 0.184 | 0.147 | 20.1 | 97 | 0.00 |
| 70 C | Pentachlorophenol | 0.137 | 0.168 | -22.6# | 155# | 0.02 |
| 71 | Phenanthrene | 0.962 | 0.949 | 1.4 | 124 | 0.02 |
| 72 | Anthracene | 0.943 | 0.925 | 1.9 | 123 | 0.02 |
| 73 | Carbazole | 0.909 | 0.943 | -3.7 | 128 | 0.02 |
| 74 | Di-n-butylphthalate | 1.116 | 1.161 | -4.0 | 127 | 0.02 |
| 75 C | Fluoranthene | 1.180 | 1.230 | -4.2 | 129 | 0.02 |
| 76 I | Chrysene-d12 | 1.000 | 1.000 | 0.0 | 133 | 0.04 |
| 77 | Benzidine | 0.420 | 0.418 | 0.5 | 117 | 0.02 |
| 78 | Pyrene | 1.307 | 1.267 | 3.1 | 131 | 0.03 |
| 79 S | Terphenyl-d14 | 1.029 | 0.958 | 6.9 | 127 | 0.02 |
| 80 | Butylbenzylphthalate | 0.437 | 0.544 | -24.5 | 154# | 0.01 |
| 81 | Benzo(a)anthracene | 1.256 | 1.254 | 0.2 | 132 | 0.04 |
| 82 | 3,3'-Dichlorobenzidine | 0.403 | 0.457 | -13.4 | 142 | 0.02 |
| 83 | Chrysene | 1.185 | 1.172 | 1.1 | 130 | 0.04 |
| 84 | Bis(2-ethylhexyl)phthalate | 0.742 | 0.836 | -12.7 | 143 | 0.02 |
| 85 c | Di-n-octyl phthalate | 1.199 | 1.458 | -21.6# | 156# | 0.03 |
| 86 I | Perylene-d12 | 1.000 | 1.000 | 0.0 | 141 | 0.08 |
| 87 | Indeno(1,2,3-cd)pyrene | 1.296 | 1.308 | -0.9 | 139 | 0.12 |
| 88 | Benzo(b)fluoranthene | 1.149 | 1.131 | 1.6 | 138 | 0.06 |
| 89 | Benzo(k)fluoranthene | 1.121 | 1.065 | 5.0 | 132 | 0.06 |
| 90 C | Benzo(a)pyrene | 0.997 | 0.991 | 0.6 | 138 | 0.07 |
| 91 | Dibenzo(a,h)anthracene | 1.077 | 1.071 | 0.6 | 137 | 0.14 |
| 92 | Benzo(g,h,i)perylene | 1.080 | 1.083 | -0.3 | 139 | 0.13 |

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP082024\
Data File : BP021575.D
Acq On : 20 Aug 2024 16:40
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Aug 20 17:15:31 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 14 00:00:30 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

| Compound | AvgRF | CCRF | %Dev | Area% | Dev(min) |
|----------|-------|------|------|-------|----------|
|----------|-------|------|------|-------|----------|

(#) = Out of Range SPCC's out = 0 CCC's out = 3

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP082024\
 Data File : BP021575.D
 Acq On : 20 Aug 2024 16:40
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 20 17:15:31 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 14 00:00:30 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

| | Compound | Amount | Calc. | %Dev | Area% | Dev(min) |
|------|-----------------------------|--------|--------|--------|-------|----------|
| 1 I | 1,4-Dichlorobenzene-d4 | 20.000 | 20.000 | 0.0 | 111 | 0.00 |
| 2 | 1,4-Dioxane | 40.000 | 35.681 | 10.8 | 99 | 0.00 |
| 3 | Pyridine | 40.000 | 36.406 | 9.0 | 100 | 0.01 |
| 4 | n-Nitrosodimethylamine | 40.000 | 37.206 | 7.0 | 104 | 0.00 |
| 5 S | 2-Fluorophenol | 80.000 | 81.506 | -1.9 | 113 | 0.00 |
| 6 | Aniline | 40.000 | 36.863 | 7.8 | 102 | 0.00 |
| 7 S | Phenol-d6 | 80.000 | 80.016 | -0.0 | 111 | 0.00 |
| 8 | 2-Chlorophenol | 40.000 | 42.226 | -5.6 | 117 | 0.00 |
| 9 | Benzaldehyde | 40.000 | 44.262 | -10.7 | 129 | 0.00 |
| 10 C | Phenol | 40.000 | 39.093 | 2.3 | 108 | 0.01 |
| 11 | bis(2-Chloroethyl)ether | 40.000 | 36.557 | 8.6 | 103 | 0.00 |
| 12 | 1,3-Dichlorobenzene | 40.000 | 39.364 | 1.6 | 112 | 0.00 |
| 13 C | 1,4-Dichlorobenzene | 40.000 | 39.661 | 0.8 | 112 | 0.00 |
| 14 | 1,2-Dichlorobenzene | 40.000 | 38.940 | 2.7 | 111 | 0.00 |
| 15 | Benzyl Alcohol | 40.000 | 38.673 | 3.3 | 107 | 0.00 |
| 16 | 2,2'-oxybis(1-Chloropropane | 40.000 | 35.934 | 10.2 | 103 | 0.00 |
| 17 | 2-Methylphenol | 40.000 | 40.769 | -1.9 | 113 | 0.00 |
| 18 | Hexachloroethane | 40.000 | 40.786 | -2.0 | 114 | 0.00 |
| 19 P | n-Nitroso-di-n-propylamine | 40.000 | 33.637 | 15.9 | 95 | 0.00 |
| 20 | 3+4-Methylphenols | 40.000 | 41.976 | -4.9 | 115 | 0.00 |
| 21 I | Naphthalene-d8 | 20.000 | 20.000 | 0.0 | 113 | 0.00 |
| 22 | Acetophenone | 40.000 | 37.514 | 6.2 | 107 | 0.00 |
| 23 S | Nitrobenzene-d5 | 80.000 | 86.924 | -8.7 | 116 | 0.00 |
| 24 | Nitrobenzene | 40.000 | 40.876 | -2.2 | 110 | 0.00 |
| 25 | Isophorone | 40.000 | 35.467 | 11.3 | 101 | 0.00 |
| 26 C | 2-Nitrophenol | 40.000 | 56.363 | -40.9# | 196 | 0.00 |
| 27 | 2,4-Dimethylphenol | 40.000 | 40.260 | -0.6 | 114 | 0.01 |
| 28 | bis(2-Chloroethoxy)methane | 40.000 | 36.474 | 8.8 | 105 | 0.00 |
| 29 C | 2,4-Dichlorophenol | 40.000 | 45.397 | -13.5 | 126 | 0.00 |
| 30 | 1,2,4-Trichlorobenzene | 40.000 | 38.758 | 3.1 | 111 | 0.00 |
| 31 | Naphthalene | 40.000 | 39.419 | 1.5 | 112 | 0.00 |
| 32 | Benzoic acid | 40.000 | 65.645 | -64.1# | 245 | 0.03 |
| 33 | 4-Chloroaniline | 40.000 | 40.788 | -2.0 | 115 | 0.00 |
| 34 C | Hexachlorobutadiene | 40.000 | 38.319 | 4.2 | 109 | 0.00 |
| 35 | Caprolactam | 40.000 | 47.892 | -19.7 | 132 | 0.04 |
| 36 C | 4-Chloro-3-methylphenol | 40.000 | 44.353 | -10.9 | 122 | 0.02 |
| 37 | 2-Methylnaphthalene | 40.000 | 39.219 | 2.0 | 112 | 0.00 |
| 38 | 1-Methylnaphthalene | 40.000 | 39.399 | 1.5 | 113 | 0.00 |
| 39 I | Acenaphthene-d10 | 20.000 | 20.000 | 0.0 | 120 | 0.00 |
| 40 | 1,2,4,5-Tetrachlorobenzene | 40.000 | 37.541 | 6.1 | 111 | 0.00 |
| 41 P | Hexachlorocyclopentadiene | 40.000 | 46.834 | -17.1 | 136 | 0.00 |
| 42 S | 2,4,6-Tribromophenol | 80.000 | 95.320 | -19.1 | 135 | 0.01 |
| 43 C | 2,4,6-Trichlorophenol | 40.000 | 46.287 | -15.7 | 132 | 0.00 |
| 44 | 2,4,5-Trichlorophenol | 40.000 | 46.667 | -16.7 | 132 | 0.01 |
| 45 S | 2-Fluorobiphenyl | 80.000 | 73.981 | 7.5 | 110 | 0.00 |
| 46 | 1,1'-Biphenyl | 40.000 | 37.962 | 5.1 | 113 | 0.00 |
| 47 | 2-Chloronaphthalene | 40.000 | 38.359 | 4.1 | 114 | 0.00 |

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP082024\
 Data File : BP021575.D
 Acq On : 20 Aug 2024 16:40
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 20 17:15:31 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 14 00:00:30 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

| | Compound | Amount | Calc. | %Dev | Area% | Dev(min) |
|------|----------------------------|--------|--------|--------|-------|----------|
| 48 | 2-Nitroaniline | 40.000 | 47.126 | -17.8 | 153 | 0.01 |
| 49 | Acenaphthylene | 40.000 | 39.430 | 1.4 | 117 | 0.01 |
| 50 | Dimethylphthalate | 40.000 | 40.557 | -1.4 | 121 | 0.00 |
| 51 | 2,6-Dinitrotoluene | 40.000 | 46.431 | -16.1 | 147 | 0.00 |
| 52 C | Acenaphthene | 40.000 | 41.294 | -3.2 | 123 | 0.01 |
| 53 | 3-Nitroaniline | 40.000 | 49.513 | -23.8 | 156 | 0.00 |
| 54 P | 2,4-Dinitrophenol | 40.000 | 66.049 | -65.1# | 258 | 0.02 |
| 55 | Dibenzofuran | 40.000 | 39.716 | 0.7 | 118 | 0.01 |
| 56 P | 4-Nitrophenol | 40.000 | 54.273 | -35.7# | 174 | 0.01 |
| 57 | 2,4-Dinitrotoluene | 40.000 | 50.859 | -27.1# | 165 | 0.02 |
| 58 | Fluorene | 40.000 | 39.039 | 2.4 | 116 | 0.02 |
| 59 | 2,3,4,6-Tetrachlorophenol | 40.000 | 47.986 | -20.0 | 135 | 0.00 |
| 60 | Diethylphthalate | 40.000 | 41.319 | -3.3 | 122 | 0.00 |
| 61 | 4-Chlorophenyl-phenylether | 40.000 | 37.977 | 5.1 | 114 | 0.00 |
| 62 | 4-Nitroaniline | 40.000 | 50.846 | -27.1# | 153 | 0.01 |
| 63 | Azobenzene | 40.000 | 35.641 | 10.9 | 105 | 0.00 |
| 64 I | Phenanthrene-d10 | 20.000 | 20.000 | 0.0 | 126 | 0.02 |
| 65 | 4,6-Dinitro-2-methylphenol | 40.000 | 58.431 | -46.1# | 229 | 0.02 |
| 66 c | n-Nitrosodiphenylamine | 40.000 | 37.750 | 5.6 | 119 | 0.00 |
| 67 | 4-Bromophenyl-phenylether | 40.000 | 37.008 | 7.5 | 119 | 0.00 |
| 68 | Hexachlorobenzene | 40.000 | 38.408 | 4.0 | 120 | 0.01 |
| 69 | Atrazine | 40.000 | 31.853 | 20.4 | 97 | 0.00 |
| 70 C | Pentachlorophenol | 40.000 | 49.250 | -23.1# | 155 | 0.02 |
| 71 | Phenanthrene | 40.000 | 39.464 | 1.3 | 124 | 0.02 |
| 72 | Anthracene | 40.000 | 39.216 | 2.0 | 123 | 0.02 |
| 73 | Carbazole | 40.000 | 41.477 | -3.7 | 128 | 0.02 |
| 74 | Di-n-butylphthalate | 40.000 | 41.587 | -4.0 | 127 | 0.02 |
| 75 C | Fluoranthene | 40.000 | 41.680 | -4.2 | 129 | 0.02 |
| 76 I | Chrysene-d12 | 20.000 | 20.000 | 0.0 | 133 | 0.04 |
| 77 | Benzidine | 40.000 | 35.722 | 10.7 | 117 | 0.02 |
| 78 | Pyrene | 40.000 | 38.762 | 3.1 | 131 | 0.03 |
| 79 S | Terphenyl-d14 | 80.000 | 74.447 | 6.9 | 127 | 0.02 |
| 80 | Butylbenzylphthalate | 40.000 | 44.200 | -10.5 | 154 | 0.01 |
| 81 | Benzo(a)anthracene | 40.000 | 39.950 | 0.1 | 132 | 0.04 |
| 82 | 3,3'-Dichlorobenzidine | 40.000 | 45.379 | -13.4 | 142 | 0.02 |
| 83 | Chrysene | 40.000 | 39.553 | 1.1 | 130 | 0.04 |
| 84 | Bis(2-ethylhexyl)phthalate | 40.000 | 45.083 | -12.7 | 143 | 0.02 |
| 85 c | Di-n-octyl phthalate | 40.000 | 45.068 | -12.7 | 156 | 0.03 |
| 86 I | Perylene-d12 | 20.000 | 20.000 | 0.0 | 141 | 0.08 |
| 87 | Indeno(1,2,3-cd)pyrene | 40.000 | 40.361 | -0.9 | 139 | 0.12 |
| 88 | Benzo(b)fluoranthene | 40.000 | 39.379 | 1.6 | 138 | 0.06 |
| 89 | Benzo(k)fluoranthene | 40.000 | 38.012 | 5.0 | 132 | 0.06 |
| 90 C | Benzo(a)pyrene | 40.000 | 39.759 | 0.6 | 138 | 0.07 |
| 91 | Dibenzo(a,h)anthracene | 40.000 | 39.775 | 0.6 | 137 | 0.14 |
| 92 | Benzo(g,h,i)perylene | 40.000 | 40.133 | -0.3 | 139 | 0.13 |

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP082024\
Data File : BP021575.D
Acq On : 20 Aug 2024 16:40
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Aug 20 17:15:31 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 14 00:00:30 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

| Compound | Amount | Calc. | %Dev | Area% | Dev(min) |
|----------|--------|-------|------|-------|----------|
|----------|--------|-------|------|-------|----------|

(#) = Out of Range SPCC's out = 0 CCC's out = 2



QC SAMPLE


DATA

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081324\
 Data File : BP021480.D
 Acq On : 13 Aug 2024 11:49
 Operator : RC/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Aug 14 00:00:30 2024




AutoFind: Scans 2509, 2510, 2511; Background Corrected with Scan 2499

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 51 | 198 | 10 | 80 | 30.1 | 212028 | PASS |
| 68 | 69 | 0.00 | 2 | 1.5 | 4256 | PASS |
| 69 | 198 | 0.00 | 100 | 40.6 | 286100 | PASS |
| 70 | 69 | 0.00 | 2 | 0.4 | 1197 | PASS |
| 127 | 198 | 10 | 80 | 43.9 | 309160 | PASS |
| 197 | 198 | 0.00 | 2 | 0.5 | 3666 | PASS |
| 198 | 198 | 100 | 100 | 100.0 | 704123 | PASS |
| 199 | 198 | 5 | 9 | 6.6 | 46731 | PASS |
| 275 | 198 | 10 | 60 | 28.5 | 200892 | PASS |
| 365 | 198 | 1 | 100 | 3.5 | 24968 | PASS |
| 441 | 198 | 0.01 | 100 | 15.0 | 105600 | PASS |
| 442 | 442 | 100 | 100 | 100.0 | 684335 | PASS |
| 443 | 442 | 15 | 24 | 19.3 | 131982 | PASS |

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081324\
 Data File : BP021480.D
 Acq On : 13 Aug 2024 11:49
 Operator : RC/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DFTPP

Quant Time: Aug 14 00:14:23 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 14 00:00:30 2024
 Response via : Initial Calibration



(70) Pentachlorophenol (C)

16.904min (+ 0.012) 6855729.98 ng

response 795611

| Ion | Exp% | Act% |
|--------|--------|--------|
| 265.70 | 100.00 | 100.00 |
| 268.00 | 64.70 | 64.10 |
| 264.00 | 63.10 | 63.36 |
| 0.00 | 0.00 | 0.00 |

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081324\
 Data File : BP021480.D
 Acq On : 13 Aug 2024 11:49
 Operator : RC/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DFTPP

Quant Time: Aug 14 00:14:23 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 14 00:00:30 2024
 Response via : Initial Calibration

Abundance

Ion 184.00 (183.70 to 184.70): BP021480.D\data.ms
 Ion 185.00 (184.70 to 185.70): BP021480.D\data.ms
 Ion 183.00 (182.70 to 183.70): BP021480.D\data.ms


19.58 Tailing = 1.05

\$ E

Time--> 18.60 18.70 18.80 18.90 19.00 19.10 19.20 19.30 19.40 19.50 19.60 19.70 19.80 19.90 20.00 20.10 20.20 20.30 20.40 20.50 20.60 20.70


Scan 2821 (19.580 min): BP021480.D\data.ms

184.1



Scan 2820 (19.575 min): BP021485.D\data.ms (-2809) (-)

184.1



TIC: BP021480.D\data.ms

(77) Benzidine

19.580min (+ 0.005) -20.00 ng

response 5830874

| Ion | Exp% | Act% |
|--------|--------|--------|
| 184.00 | 100.00 | 100.00 |
| 185.00 | 14.20 | 14.05 |
| 183.00 | 11.60 | 11.75 |
| 0.00 | 0.00 | 0.00 |

Instrument :
BNA_P
ClientSampleId :
DFTPP

DDT Breakdown


| Date | Instrument Name | DFTPP Data File |
|---------------|------------------|--------------------|
| 8/13/2024 | BNA_P | BP021480.D |
| Compound Name | Response | Retention Time |
| DDT | 2081689 | 20.88 |
| DDD | 17224 | 20.486 |
| DDE | 158 | 19.962 |
| SUM(DDD+DDE) | SUM(DDT+DDD+DDE) | % Breakdown Of DDT |
| 17382 | 2099071 | 0.83 |

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP082024\
 Data File : BP021565.D
 Acq On : 20 Aug 2024 09:25
 Operator : RC/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Aug 14 00:00:30 2024



AutoFind: Scans 2508, 2509, 2510; Background Corrected with Scan 2498

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 51 | 198 | 10 | 80 | 29.2 | 324123 | PASS |
| 68 | 69 | 0.00 | 2 | 1.6 | 6547 | PASS |
| 69 | 198 | 0.00 | 100 | 36.7 | 407478 | PASS |
| 70 | 69 | 0.00 | 2 | 0.5 | 2016 | PASS |
| 127 | 198 | 10 | 80 | 44.9 | 498814 | PASS |
| 197 | 198 | 0.00 | 2 | 0.5 | 5669 | PASS |
| 198 | 198 | 100 | 100 | 100.0 | 1109972 | PASS |
| 199 | 198 | 5 | 9 | 6.7 | 74231 | PASS |
| 275 | 198 | 10 | 60 | 28.9 | 320288 | PASS |
| 365 | 198 | 1 | 100 | 3.8 | 42187 | PASS |
| 441 | 198 | 0.01 | 100 | 14.4 | 160344 | PASS |
| 442 | 442 | 100 | 100 | 100.0 | 1057190 | PASS |
| 443 | 442 | 15 | 24 | 19.1 | 202151 | PASS |

Instrument :
BNA_P
ClientSampleId :
DFTPP

DDT Breakdown


| Date | Instrument Name | DFTPP Data File |
|---------------|------------------|--------------------|
| 8/20/2024 | BNA_P | BP021565.D |
| Compound Name | Response | Retention Time |
| DDT | 4160702 | 20.892 |
| DDD | 231762 | 20.422 |
| DDE | 2972 | 19.898 |
| SUM(DDD+DDE) | SUM(DDT+DDD+DDE) | % Breakdown Of DDT |
| 234734 | 4395436 | 5.34 |

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP082024\
 Data File : BP021565.D
 Acq On : 20 Aug 2024 09:25
 Operator : RC/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :
 DFTPP

Quant Time: Aug 20 11:19:26 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 14 00:00:30 2024
 Response via : Initial Calibration

Ion 265.70 (265.40 to 266.40): BP021565.D\data.ms
 Ion 268.00 (267.70 to 268.70): BP021565.D\data.ms
 Ion 264.00 (263.70 to 264.70): BP021565.D\data.ms



Scan 2366 (16.904 min): BP021565.D\data.ms



TIC: BP021565.D\data.ms

(70) Pentachlorophenol (C)
 16.904min (+ 0.012) 916265.56 ng



response 1726349

| Ion | Exp% | Act% |
|--------|--------|--------|
| 265.70 | 100.00 | 100.00 |
| 268.00 | 64.70 | 64.57 |
| 264.00 | 63.10 | 62.71 |
| 0.00 | 0.00 | 0.00 |

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP082024\
 Data File : BP021565.D
 Acq On : 20 Aug 2024 09:25
 Operator : RC/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DFTPP

Quant Time: Aug 20 11:19:26 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 14 00:00:30 2024
 Response via : Initial Calibration



(77) Benzidine

19.592min (+ 0.017) -20.00 ng

response 9393500


| Ion | Exp% | Act% |
|--------|--------|--------|
| 184.00 | 100.00 | 100.00 |
| 185.00 | 14.20 | 14.73 |
| 183.00 | 11.60 | 12.22 |
| 0.00 | 0.00 | 0.00 |

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP082024\
 Data File : BP021574.D
 Acq On : 20 Aug 2024 15:57
 Operator : RC/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Aug 14 00:00:30 2024



AutoFind: Scans 2508, 2509, 2510; Background Corrected with Scan 2496

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 51 | 198 | 10 | 80 | 28.5 | 318494 | PASS |
| 68 | 69 | 0.00 | 2 | 1.6 | 6404 | PASS |
| 69 | 198 | 0.00 | 100 | 36.5 | 407976 | PASS |
| 70 | 69 | 0.00 | 2 | 0.4 | 1654 | PASS |
| 127 | 198 | 10 | 80 | 44.3 | 495609 | PASS |
| 197 | 198 | 0.00 | 2 | 0.3 | 3742 | PASS |
| 198 | 198 | 100 | 100 | 100.0 | 1117661 | PASS |
| 199 | 198 | 5 | 9 | 6.8 | 76061 | PASS |
| 275 | 198 | 10 | 60 | 30.1 | 336655 | PASS |
| 365 | 198 | 1 | 100 | 4.0 | 45002 | PASS |
| 441 | 198 | 0.01 | 100 | 15.2 | 169704 | PASS |
| 442 | 442 | 100 | 100 | 100.0 | 1099215 | PASS |
| 443 | 442 | 15 | 24 | 19.1 | 210240 | PASS |

Instrument :
BNA_P
ClientSampleId :
DFTPP



DDT Breakdown

| Date | Instrument Name | DFTPP Data File |
|---------------|------------------|--------------------|
| 8/20/2024 | BNA_P | BP021574.D |
| Compound Name | Response | Retention Time |
| DDT | 4861860 | 20.892 |
| DDD | 175942 | 20.427 |
| DDE | 3862 | 19.898 |
| SUM(DDD+DDE) | SUM(DDT+DDD+DDE) | % Breakdown Of DDT |
| 179804 | 5041664 | 3.57 |

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP082024\
 Data File : BP021574.D
 Acq On : 20 Aug 2024 15:57
 Operator : RC/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DFTPP

Quant Time: Aug 20 16:56:21 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 14 00:00:30 2024
 Response via : Initial Calibration





(70) Pentachlorophenol (C)
 16.904min (+ 0.012) 1466831.19 ng
 response 1882505

| Ion | Exp% | Act% |
|--------|--------|--------|
| 265.70 | 100.00 | 100.00 |
| 268.00 | 64.70 | 65.26 |
| 264.00 | 63.10 | 62.86 |
| 0.00 | 0.00 | 0.00 |

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP082024\
 Data File : BP021574.D
 Acq On : 20 Aug 2024 15:57
 Operator : RC/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DFTPP

Quant Time: Aug 20 16:56:21 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 14 00:00:30 2024
 Response via : Initial Calibration



(77) Benzidine

19.592min (+ 0.017) -20.00 ng

response 10375641

| Ion | Exp% | Act% |
|--------|--------|--------|
| 184.00 | 100.00 | 100.00 |
| 185.00 | 14.20 | 14.60 |
| 183.00 | 11.60 | 12.36 |
| 0.00 | 0.00 | 0.00 |



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

| | | | | | |
|--------------------|---------------------------------------|--------|----|-----------------|---------------|
| Client: | JACOBS Engineering Group, Inc. | | | Date Collected: | |
| Project: | Former Schlumberger Site Princeton NJ | | | Date Received: | |
| Client Sample ID: | PB162822BL | | | SDG No.: | P3657 |
| Lab Sample ID: | PB162822BL | | | Matrix: | Water |
| Analytical Method: | SW8270 | | | % Solid: | 0 |
| Sample Wt/Vol: | 1000 | Units: | mL | Final Vol: | 1000 uL |
| Soil Aliquot Vol: | uL | | | Test: | SVOCMS Group6 |
| Extraction Type : | Decanted : N | | | Level : | LOW |
| Injection Volume : | GPC Factor : 1.0 | | | GPC Cleanup : | N PH : |
| Prep Method : | SW3510C | | | | |

| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed | Prep Batch ID |
|-------------------|-----------|----------------|----------------|---------------|
| BP021567.D | 1 | 08/19/24 09:40 | 08/20/24 10:49 | PB162822 |

| CAS Number | Parameter | Conc. | Qualifier | MDL | LOQ / CRQL | Units |
|------------|-----------|-------|-----------|-----|------------|-------|
|------------|-----------|-------|-----------|-----|------------|-------|

TARGETS

| | | | | | | |
|------------|----------------------------|------|---|------|------|------|
| 110-86-1 | Pyridine | 1.60 | U | 1.60 | 5.00 | ug/L |
| 100-52-7 | Benzaldehyde | 4.00 | U | 4.00 | 10.0 | ug/L |
| 95-48-7 | 2-Methylphenol | 1.10 | U | 1.10 | 5.00 | ug/L |
| 65794-96-9 | 3+4-Methylphenols | 1.20 | U | 1.20 | 10.0 | ug/L |
| 67-72-1 | Hexachloroethane | 1.00 | U | 1.00 | 5.00 | ug/L |
| 98-95-3 | Nitrobenzene | 1.30 | U | 1.30 | 5.00 | ug/L |
| 91-20-3 | Naphthalene | 1.00 | U | 1.00 | 5.00 | ug/L |
| 87-68-3 | Hexachlorobutadiene | 1.30 | U | 1.30 | 5.00 | ug/L |
| 91-57-6 | 2-Methylnaphthalene | 1.10 | U | 1.10 | 5.00 | ug/L |
| 88-06-2 | 2,4,6-Trichlorophenol | 0.89 | U | 0.89 | 5.00 | ug/L |
| 95-95-4 | 2,4,5-Trichlorophenol | 1.00 | U | 1.00 | 5.00 | ug/L |
| 208-96-8 | Acenaphthylene | 1.00 | U | 1.00 | 5.00 | ug/L |
| 83-32-9 | Acenaphthene | 0.81 | U | 0.81 | 5.00 | ug/L |
| 132-64-9 | Dibenzofuran | 0.93 | U | 0.93 | 5.00 | ug/L |
| 121-14-2 | 2,4-Dinitrotoluene | 1.50 | U | 1.50 | 5.00 | ug/L |
| 86-73-7 | Fluorene | 0.96 | U | 0.96 | 5.00 | ug/L |
| 118-74-1 | Hexachlorobenzene | 1.10 | U | 1.10 | 5.00 | ug/L |
| 87-86-5 | Pentachlorophenol | 1.90 | U | 1.90 | 10.0 | ug/L |
| 85-01-8 | Phenanthrene | 0.89 | U | 0.89 | 5.00 | ug/L |
| 120-12-7 | Anthracene | 1.10 | U | 1.10 | 5.00 | ug/L |
| 86-74-8 | Carbazole | 1.20 | U | 1.20 | 5.00 | ug/L |
| 84-74-2 | Di-n-butylphthalate | 1.50 | U | 1.50 | 5.00 | ug/L |
| 206-44-0 | Fluoranthene | 1.30 | U | 1.30 | 5.00 | ug/L |
| 129-00-0 | Pyrene | 1.10 | U | 1.10 | 5.00 | ug/L |
| 56-55-3 | Benzo(a)anthracene | 0.94 | U | 0.94 | 5.00 | ug/L |
| 218-01-9 | Chrysene | 0.86 | U | 0.86 | 5.00 | ug/L |
| 117-81-7 | Bis(2-ethylhexyl)phthalate | 1.90 | U | 1.90 | 5.00 | ug/L |
| 205-99-2 | Benzo(b)fluoranthene | 1.10 | U | 1.10 | 5.00 | ug/L |
| 207-08-9 | Benzo(k)fluoranthene | 1.20 | U | 1.20 | 5.00 | ug/L |



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

| | | | | | |
|--------------------|---------------------------------------|--------|----|-----------------|---------------|
| Client: | JACOBS Engineering Group, Inc. | | | Date Collected: | |
| Project: | Former Schlumberger Site Princeton NJ | | | Date Received: | |
| Client Sample ID: | PB162822BL | | | SDG No.: | P3657 |
| Lab Sample ID: | PB162822BL | | | Matrix: | Water |
| Analytical Method: | SW8270 | | | % Solid: | 0 |
| Sample Wt/Vol: | 1000 | Units: | mL | Final Vol: | 1000 uL |
| Soil Aliquot Vol: | uL | | | Test: | SVOCMS Group6 |
| Extraction Type : | Decanted : N | | | Level : | LOW |
| Injection Volume : | GPC Factor : 1.0 | | | GPC Cleanup : | N PH : |
| Prep Method : | SW3510C | | | | |

| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed | Prep Batch ID |
|-------------------|-----------|----------------|----------------|---------------|
| BP021567.D | 1 | 08/19/24 09:40 | 08/20/24 10:49 | PB162822 |

| CAS Number | Parameter | Conc. | Qualifier | MDL | LOQ / CRQL | Units |
|------------|------------------------|-------|-----------|------|------------|-------|
| 50-32-8 | Benzo(a)pyrene | 1.70 | U | 1.70 | 5.00 | ug/L |
| 193-39-5 | Indeno(1,2,3-cd)pyrene | 1.00 | U | 1.00 | 5.00 | ug/L |
| 53-70-3 | Dibenzo(a,h)anthracene | 1.20 | U | 1.20 | 5.00 | ug/L |
| 191-24-2 | Benzo(g,h,i)perylene | 1.20 | U | 1.20 | 5.00 | ug/L |
| 123-91-1 | 1,4-Dioxane | 1.30 | U | 1.30 | 5.00 | ug/L |
| 90-12-0 | 1-Methylnaphthalene | 0.86 | U | 0.86 | 5.00 | ug/L |

SURROGATES

| | | | | | |
|------------|----------------------|------|---------------------|------|----------|
| 367-12-4 | 2-Fluorophenol | 130 | 15 (10) - 110 (139) | 87% | SPK: 150 |
| 13127-88-3 | Phenol-d6 | 120 | 15 (10) - 110 (134) | 80% | SPK: 150 |
| 4165-60-0 | Nitrobenzene-d5 | 90.5 | 30 (49) - 130 (133) | 91% | SPK: 100 |
| 321-60-8 | 2-Fluorobiphenyl | 78.7 | 30 (52) - 130 (132) | 79% | SPK: 100 |
| 118-79-6 | 2,4,6-Tribromophenol | 158 | 15 (44) - 110 (137) | 105% | SPK: 150 |
| 1718-51-0 | Terphenyl-d14 | 82.4 | 30 (48) - 130 (125) | 82% | SPK: 100 |

INTERNAL STANDARDS

| | | | |
|------------|------------------------|---------|--------|
| 3855-82-1 | 1,4-Dichlorobenzene-d4 | 385000 | 7.805 |
| 1146-65-2 | Naphthalene-d8 | 1560000 | 10.593 |
| 15067-26-2 | Acenaphthene-d10 | 1020000 | 14.445 |
| 1517-22-2 | Phenanthrene-d10 | 2410000 | 17.257 |
| 1719-03-5 | Chrysene-d12 | 2580000 | 21.722 |
| 1520-96-3 | Perylene-d12 | 2890000 | 25.18 |

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP082024\
 Data File : BP021567.D
 Acq On : 20 Aug 2024 10:49
 Operator : RC/JU
 Sample : PB162822BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 PB162822BL

Quant Time: Aug 20 11:21:31 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 14 00:00:30 2024
 Response via : Initial Calibration


| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|------------------------------------|--------|------|----------|---------|-------|----------|
| Internal Standards | | | | | | |
| 1) 1,4-Dichlorobenzene-d4 | 7.805 | 152 | 385209 | 20.000 | ng | 0.00 |
| 21) Naphthalene-d8 | 10.593 | 136 | 1560541 | 20.000 | ng | 0.00 |
| 39) Acenaphthene-d10 | 14.445 | 164 | 1021934 | 20.000 | ng | -0.01 |
| 64) Phenanthrene-d10 | 17.257 | 188 | 2408801 | 20.000 | ng | 0.00 |
| 76) Chrysene-d12 | 21.722 | 240 | 2576176 | 20.000 | ng | 0.02 |
| 86) Perylene-d12 | 25.180 | 264 | 2892860 | 20.000 | ng | 0.08 |
| System Monitoring Compounds | | | | | | |
| 5) 2-Fluorophenol | 5.405 | 112 | 3380481 | 130.307 | ng | 0.01 |
| 7) Phenol-d6 | 6.969 | 99 | 4524489 | 119.556 | ng | 0.00 |
| 23) Nitrobenzene-d5 | 8.952 | 82 | 2737375 | 90.539 | ng | 0.00 |
| 42) 2,4,6-Tribromophenol | 15.963 | 330 | 1796946 | 157.991 | ng | 0.00 |
| 45) 2-Fluorobiphenyl | 13.063 | 172 | 5272853 | 78.745 | ng | 0.00 |
| 79) Terphenyl-d14 | 19.986 | 244 | 10921233 | 82.358 | ng | 0.01 |


| Target Compounds | Qvalue |
|--|--------|
| (#= qualifier out of range (m) = manual integration (+) = signals summed | |

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP082024\
Data File : BP021567.D
Acq On : 20 Aug 2024 10:49
Operator : RC/JU
Sample : PB162822BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
PB162822BL

Quant Time: Aug 20 11:21:31 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 14 00:00:30 2024
Response via : Initial Calibration







#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.805 min Scan# 8
Delta R.T. -0.006 min
Lab File: BP021567.D
Acq: 20 Aug 2024 10:49


Instrument : BNA_P
ClientSampleId : PB162822BL


Tgt Ion:152 Resp: 385209
Ion Ratio Lower Upper
152 100
150 153.6 122.4 183.6
115 68.3 54.7 82.1



#5
2-Fluorophenol
Concen: 130.307 ng
RT: 5.405 min Scan# 411
Delta R.T. 0.012 min
Lab File: BP021567.D
Acq: 20 Aug 2024 10:49

Tgt Ion:112 Resp: 3380481
Ion Ratio Lower Upper
112 100
64 52.8 49.0 73.6
63 27.0 24.6 36.8







#7
 Phenol-d6
 Concen: 119.556 ng
 RT: 6.969 min Scan# 6
 Delta R.T. 0.006 min
 Lab File: BP021567.D
 Acq: 20 Aug 2024 10:49


Instrument : BNA_P
 ClientSampleId : PB162822BL


Tgt Ion: 99 Resp: 4524489
 Ion Ratio Lower Upper
 99 100
 42 12.3 10.2 15.4
 71 31.1 25.8 38.8



#21
 Naphthalene-d8
 Concen: 20.000 ng
 RT: 10.593 min Scan# 1293
 Delta R.T. -0.006 min
 Lab File: BP021567.D
 Acq: 20 Aug 2024 10:49

Tgt Ion:136 Resp: 1560541
 Ion Ratio Lower Upper
 136 100
 137 10.6 8.8 13.2
 54 7.5 6.3 9.5
 68 7.1 6.8 10.2







#23
 Nitrobenzene-d5
 Concen: 90.539 ng
 RT: 8.952 min Scan# 1
 Delta R.T. -0.000 min
 Lab File: BP021567.D
 Acq: 20 Aug 2024 10:49


Instrument : BNA_P
 ClientSampleId : PB162822BL


Tgt Ion: 82 Resp: 2737375
 Ion Ratio Lower Upper
 82 100
 128 38.3 26.5 39.7
 54 38.9 30.3 45.5



#39
 Acenaphthene-d10
 Concen: 20.000 ng
 RT: 14.445 min Scan# 1948
 Delta R.T. -0.012 min
 Lab File: BP021567.D
 Acq: 20 Aug 2024 10:49

Tgt Ion: 164 Resp: 1021934
 Ion Ratio Lower Upper
 164 100
 162 98.6 81.5 122.3
 160 44.1 35.2 52.8







#42
2,4,6-Tribromophenol
Concen: 157.991 ng
RT: 15.963 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP021567.D
Acq: 20 Aug 2024 10:49


Instrument : BNA_P
ClientSampleId : PB162822BL


Tgt Ion:330 Resp: 1796946
Ion Ratio Lower Upper
330 100
332 96.9 77.4 116.0
141 30.0 24.5 36.7



#45
2-Fluorobiphenyl
Concen: 78.745 ng
RT: 13.063 min Scan# 1713
Delta R.T. -0.000 min
Lab File: BP021567.D
Acq: 20 Aug 2024 10:49



Tgt Ion:172 Resp: 5272853
Ion Ratio Lower Upper
172 100
171 35.2 28.2 42.4
170 23.2 18.4 27.6






#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.257 min Scan# 2
Instrument: BNA_P
Delta R.T. 0.006 min
Lab File: BP021567.D ClientSampleId :
Acq: 20 Aug 2024 10:49 PB162822BL

Tgt Ion:188 Resp: 2408801
Ion Ratio Lower Upper
188 100
94 11.6 10.2 15.4
80 12.6 11.6 17.4



#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.722 min Scan# 3185
Delta R.T. 0.024 min
Lab File: BP021567.D
Acq: 20 Aug 2024 10:49

Tgt Ion:240 Resp: 2576176
Ion Ratio Lower Upper
240 100
120 11.8 9.9 14.9
236 25.7 20.6 30.8



#79

Terphenyl-d14

Concen: 82.358 ng

RT: 19.986 min Scan# 2

Delta R.T. 0.011 min

Lab File: BP021567.D

Acq: 20 Aug 2024 10:49

Instrument :

BNA_P

ClientSampleId :

PB162822BL

Tgt Ion:244 Resp:10921233

Ion Ratio Lower Upper


244 100

212 7.0 5.6 8.4

122 11.1 9.4 14.2

Abundance

19.986



#86

Perylene-d12

Concen: 20.000 ng

RT: 25.180 min Scan# 3773

Delta R.T. 0.076 min

Lab File: BP021567.D

Acq: 20 Aug 2024 10:49

Tgt Ion:264 Resp: 2892860

Ion Ratio Lower Upper




264 100

260 23.7 18.6 28.0

265 21.1 17.1 25.7

Abundance

25.180





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

| | | | | | |
|--------------------|---------------------------------------|--------|----|-----------------|---------------|
| Client: | JACOBS Engineering Group, Inc. | | | Date Collected: | |
| Project: | Former Schlumberger Site Princeton NJ | | | Date Received: | |
| Client Sample ID: | PB162822BS | | | SDG No.: | P3657 |
| Lab Sample ID: | PB162822BS | | | Matrix: | Water |
| Analytical Method: | SW8270 | | | % Solid: | 0 |
| Sample Wt/Vol: | 1000 | Units: | mL | Final Vol: | 1000 uL |
| Soil Aliquot Vol: | uL | | | Test: | SVOCMS Group6 |
| Extraction Type : | Decanted : N | | | Level : | LOW |
| Injection Volume : | GPC Factor : 1.0 | | | GPC Cleanup : | N PH : |
| Prep Method : | SW3510C | | | | |

| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed | Prep Batch ID |
|-------------------|-----------|----------------|----------------|---------------|
| BP021568.D | 1 | 08/19/24 09:40 | 08/20/24 11:31 | PB162822 |

| CAS Number | Parameter | Conc. | Qualifier | MDL | LOQ / CRQL | Units |
|----------------|----------------------------|-------|-----------|------|------------|-------|
| TARGETS | | | | | | |
| 110-86-1 | Pyridine | 28.7 | | 1.60 | 5.00 | ug/L |
| 100-52-7 | Benzaldehyde | 16.8 | | 4.00 | 10.0 | ug/L |
| 95-48-7 | 2-Methylphenol | 49.7 | | 1.10 | 5.00 | ug/L |
| 65794-96-9 | 3+4-Methylphenols | 51.5 | | 1.20 | 10.0 | ug/L |
| 67-72-1 | Hexachloroethane | 45.4 | | 1.00 | 5.00 | ug/L |
| 98-95-3 | Nitrobenzene | 44.7 | | 1.30 | 5.00 | ug/L |
| 91-20-3 | Naphthalene | 43.9 | | 1.00 | 5.00 | ug/L |
| 87-68-3 | Hexachlorobutadiene | 41.5 | | 1.30 | 5.00 | ug/L |
| 91-57-6 | 2-Methylnaphthalene | 44.8 | | 1.10 | 5.00 | ug/L |
| 88-06-2 | 2,4,6-Trichlorophenol | 57.1 | | 0.89 | 5.00 | ug/L |
| 95-95-4 | 2,4,5-Trichlorophenol | 55.3 | | 1.00 | 5.00 | ug/L |
| 208-96-8 | Acenaphthylene | 48.9 | | 1.00 | 5.00 | ug/L |
| 83-32-9 | Acenaphthene | 51.6 | | 0.81 | 5.00 | ug/L |
| 132-64-9 | Dibenzofuran | 45.7 | | 0.93 | 5.00 | ug/L |
| 121-14-2 | 2,4-Dinitrotoluene | 58.1 | | 1.50 | 5.00 | ug/L |
| 86-73-7 | Fluorene | 43.7 | | 0.96 | 5.00 | ug/L |
| 118-74-1 | Hexachlorobenzene | 44.5 | | 1.10 | 5.00 | ug/L |
| 87-86-5 | Pentachlorophenol | 110 | E | 1.90 | 10.0 | ug/L |
| 85-01-8 | Phenanthrene | 46.6 | | 0.89 | 5.00 | ug/L |
| 120-12-7 | Anthracene | 47.6 | | 1.10 | 5.00 | ug/L |
| 86-74-8 | Carbazole | 47.8 | | 1.20 | 5.00 | ug/L |
| 84-74-2 | Di-n-butylphthalate | 48.4 | | 1.50 | 5.00 | ug/L |
| 206-44-0 | Fluoranthene | 46.1 | | 1.30 | 5.00 | ug/L |
| 129-00-0 | Pyrene | 48.5 | | 1.10 | 5.00 | ug/L |
| 56-55-3 | Benzo(a)anthracene | 47.9 | | 0.94 | 5.00 | ug/L |
| 218-01-9 | Chrysene | 48.0 | | 0.86 | 5.00 | ug/L |
| 117-81-7 | Bis(2-ethylhexyl)phthalate | 54.2 | | 1.90 | 5.00 | ug/L |
| 205-99-2 | Benzo(b)fluoranthene | 51.7 | | 1.10 | 5.00 | ug/L |
| 207-08-9 | Benzo(k)fluoranthene | 50.4 | | 1.20 | 5.00 | ug/L |



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

| | | | | | |
|--------------------|---------------------------------------|--------|----|-----------------|---------------|
| Client: | JACOBS Engineering Group, Inc. | | | Date Collected: | |
| Project: | Former Schlumberger Site Princeton NJ | | | Date Received: | |
| Client Sample ID: | PB162822BS | | | SDG No.: | P3657 |
| Lab Sample ID: | PB162822BS | | | Matrix: | Water |
| Analytical Method: | SW8270 | | | % Solid: | 0 |
| Sample Wt/Vol: | 1000 | Units: | mL | Final Vol: | 1000 uL |
| Soil Aliquot Vol: | uL | | | Test: | SVOCMS Group6 |
| Extraction Type : | Decanted : N | | | Level : | LOW |
| Injection Volume : | GPC Factor : 1.0 | | | GPC Cleanup : | N PH : |
| Prep Method : | SW3510C | | | | |

| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed | Prep Batch ID |
|-------------------|-----------|----------------|----------------|---------------|
| BP021568.D | 1 | 08/19/24 09:40 | 08/20/24 11:31 | PB162822 |

| CAS Number | Parameter | Conc. | Qualifier | MDL | LOQ / CRQL | Units |
|-------------------|------------------------|-------|-----------|---------------------|------------|----------|
| 50-32-8 | Benzo(a)pyrene | 54.3 | | 1.70 | 5.00 | ug/L |
| 193-39-5 | Indeno(1,2,3-cd)pyrene | 54.2 | | 1.00 | 5.00 | ug/L |
| 53-70-3 | Dibenzo(a,h)anthracene | 53.3 | | 1.20 | 5.00 | ug/L |
| 191-24-2 | Benzo(g,h,i)perylene | 51.4 | | 1.20 | 5.00 | ug/L |
| 123-91-1 | 1,4-Dioxane | 28.5 | | 1.30 | 5.00 | ug/L |
| 90-12-0 | 1-Methylnaphthalene | 42.7 | | 0.86 | 5.00 | ug/L |
| SURROGATES | | | | | | |
| 367-12-4 | 2-Fluorophenol | 139 | | 15 (10) - 110 (139) | 93% | SPK: 150 |
| 13127-88-3 | Phenol-d6 | 131 | | 15 (10) - 110 (134) | 87% | SPK: 150 |
| 4165-60-0 | Nitrobenzene-d5 | 93.7 | | 30 (49) - 130 (133) | 94% | SPK: 100 |
| 321-60-8 | 2-Fluorobiphenyl | 79.6 | | 30 (52) - 130 (132) | 80% | SPK: 100 |
| 118-79-6 | 2,4,6-Tribromophenol | 157 | | 15 (44) - 110 (137) | 105% | SPK: 150 |
| 1718-51-0 | Terphenyl-d14 | 85.2 | | 30 (48) - 130 (125) | 85% | SPK: 100 |

INTERNAL STANDARDS

| | | | |
|------------|------------------------|---------|--------|
| 3855-82-1 | 1,4-Dichlorobenzene-d4 | 495000 | 7.81 |
| 1146-65-2 | Naphthalene-d8 | 2070000 | 10.599 |
| 15067-26-2 | Acenaphthene-d10 | 1390000 | 14.451 |
| 1517-22-2 | Phenanthrene-d10 | 3040000 | 17.257 |
| 1719-03-5 | Chrysene-d12 | 2680000 | 21.727 |
| 1520-96-3 | Perylene-d12 | 2880000 | 25.168 |

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP082024\
 Data File : BP021568.D
 Acq On : 20 Aug 2024 11:31
 Operator : RC/JU
 Sample : PB162822BS
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 PB162822BS

Quant Time: Aug 20 12:11:11 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 14 00:00:30 2024
 Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/21/2024
 Supervised By :mohammad ahmed 08/22/2024

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|------------------------------------|--------|------|----------|---------|-------|----------|
| Internal Standards | | | | | | |
| 1) 1,4-Dichlorobenzene-d4 | 7.810 | 152 | 494651 | 20.000 | ng | 0.00 |
| 21) Naphthalene-d8 | 10.599 | 136 | 2073175 | 20.000 | ng | 0.00 |
| 39) Acenaphthene-d10 | 14.451 | 164 | 1389376 | 20.000 | ng | 0.00 |
| 64) Phenanthrene-d10 | 17.257 | 188 | 3038381 | 20.000 | ng | 0.00 |
| 76) Chrysene-d12 | 21.727 | 240 | 2684396 | 20.000 | ng | 0.03 |
| 86) Perylene-d12 | 25.168 | 264 | 2880701 | 20.000 | ng | 0.06 |
| System Monitoring Compounds | | | | | | |
| 5) 2-Fluorophenol | 5.411 | 112 | 4632548 | 139.062 | ng | 0.02 |
| 7) Phenol-d6 | 6.981 | 99 | 6354941 | 130.771 | ng | 0.02 |
| 23) Nitrobenzene-d5 | 8.957 | 82 | 3764035 | 93.712 | ng | 0.00 |
| 42) 2,4,6-Tribromophenol | 15.963 | 330 | 2430215 | 157.161 | ng | 0.00 |
| 45) 2-Fluorobiphenyl | 13.063 | 172 | 7244303 | 79.575 | ng | 0.00 |
| 79) Terphenyl-d14 | 19.980 | 244 | 11775634 | 85.222 | ng | 0.00 |
| Target Compounds | | | | | | |
| | | | | Qvalue | | |
| 2) 1,4-Dioxane | 3.322 | 88 | 450308 | 28.508 | ng | 99 |
| 3) Pyridine | 3.717 | 79 | 1265449m | 28.698 | ng | |
| 4) n-Nitrosodimethylamine | 3.622 | 42 | 562224 | 42.420 | ng | 90 |
| 6) Aniline | 7.134 | 93 | 1353041 | 27.851 | ng | 97 |
| 8) 2-Chlorophenol | 7.381 | 128 | 1635284 | 53.119 | ng | 94 |
| 9) Benzaldehyde | 6.946 | 77 | 523494m | 16.825 | ng | |
| 10) Phenol | 7.005 | 94 | 2456949 | 48.824 | ng | 97 |
| 11) bis(2-Chloroethyl)ether | 7.228 | 93 | 1787223 | 41.972 | ng | 96 |
| 12) 1,3-Dichlorobenzene | 7.705 | 146 | 1603283 | 43.844 | ng | 96 |
| 13) 1,4-Dichlorobenzene | 7.846 | 146 | 1626757 | 44.056 | ng | 98 |
| 14) 1,2-Dichlorobenzene | 8.157 | 146 | 1576656 | 44.308 | ng | 99 |
| 15) Benzyl Alcohol | 8.046 | 79 | 1665499 | 47.766 | ng | 96 |
| 16) 2,2'-oxybis(1-Chloropr... | 8.328 | 45 | 1679419 | 42.828 | ng | 97 |
| 17) 2-Methylphenol | 8.246 | 107 | 1581707 | 49.719 | ng | 97 |
| 18) Hexachloroethane | 8.893 | 117 | 671831 | 45.372 | ng | 97 |
| 19) n-Nitroso-di-n-propyla... | 8.610 | 70 | 1291565 | 38.475 | ng | 95 |
| 20) 3+4-Methylphenols | 8.575 | 107 | 2208573 | 51.486 | ng | 98 |
| 22) Acetophenone | 8.622 | 105 | 2535541 | 39.810 | ng | 98 |
| 24) Nitrobenzene | 8.999 | 77 | 1948611 | 44.668 | ng | 91 |
| 25) Isophorone | 9.528 | 82 | 3876788 | 40.856 | ng | 97 |
| 26) 2-Nitrophenol | 9.704 | 139 | 810333 | 63.198 | ng | 94 |
| 27) 2,4-Dimethylphenol | 9.769 | 122 | 1260392 | 53.822 | ng | 97 |
| 28) bis(2-Chloroethoxy)met... | 10.004 | 93 | 2441309 | 42.148 | ng | 99 |
| 29) 2,4-Dichlorophenol | 10.240 | 162 | 1550319 | 56.031 | ng | 97 |
| 30) 1,2,4-Trichlorobenzene | 10.457 | 180 | 1545646 | 43.572 | ng | 98 |
| 31) Naphthalene | 10.646 | 128 | 4750296 | 43.861 | ng | 100 |
| 32) Benzoic acid | 9.928 | 122 | 1255905 | 69.217 | ng | 91 |
| 33) 4-Chloroaniline | 10.746 | 127 | 1147903 | 28.216 | ng | 97 |
| 34) Hexachlorobutadiene | 10.940 | 225 | 927946 | 41.546 | ng | 99 |
| 35) Caprolactam | 11.534 | 113 | 591792m | 51.386 | ng | |
| 36) 4-Chloro-3-methylphenol | 11.875 | 107 | 2044802 | 53.276 | ng | 96 |
| 37) 2-Methylnaphthalene | 12.257 | 142 | 3311877 | 44.806 | ng | 100 |
| 38) 1-Methylnaphthalene | 12.481 | 142 | 3109322 | 42.651 | ng | 100 |
| 40) 1,2,4,5-Tetrachloroben... | 12.634 | 216 | 1660023 | 40.778 | ng | 99 |
| 41) Hexachlorocyclopentadiene | 12.610 | 237 | 1502725 | 116.470 | ng | 100 |
| 43) 2,4,6-Trichlorophenol | 12.869 | 196 | 1291184 | 57.099 | ng | 99 |

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP082024\
 Data File : BP021568.D
 Acq On : 20 Aug 2024 11:31
 Operator : RC/JU
 Sample : PB162822BS
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 PB162822BS

Quant Time: Aug 20 12:11:11 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 14 00:00:30 2024
 Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/21/2024
 Supervised By :mohammad ahmed 08/22/2024

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|-------------------------------|--------|------|-----------|---------|-------|----------|
| 44) 2,4,5-Trichlorophenol | 12.945 | 196 | 1389021 | 55.286 | ng | 98 |
| 46) 1,1'-Biphenyl | 13.275 | 154 | 3996688 | 40.658 | ng | 98 |
| 47) 2-Chloronaphthalene | 13.316 | 162 | 3361372 | 43.922 | ng | 99 |
| 48) 2-Nitroaniline | 13.510 | 65 | 1188075 | 54.598 | ng | # 85 |
| 49) Acenaphthylene | 14.169 | 152 | 5500221 | 48.890 | ng | 100 |
| 50) Dimethylphthalate | 13.904 | 163 | 4461019 | 47.945 | ng | 99 |
| 51) 2,6-Dinitrotoluene | 14.010 | 165 | 982980 | 52.003 | ng | 93 |
| 52) Acenaphthene | 14.516 | 154 | 3809412m | 51.606 | ng | |
| 53) 3-Nitroaniline | 14.345 | 138 | 796079 | 43.541 | ng | # 90 |
| 54) 2,4-Dinitrophenol | 14.557 | 184 | 1036408 | 123.071 | ng | # 87 |
| 55) Dibenzofuran | 14.857 | 168 | 5192990 | 45.731 | ng | 99 |
| 56) 4-Nitrophenol | 14.669 | 139 | 1830474 | 116.168 | ng | # 85 |
| 57) 2,4-Dinitrotoluene | 14.816 | 165 | 1414718 | 58.078 | ng | # 84 |
| 58) Fluorene | 15.516 | 166 | 4077699 | 43.711 | ng | 100 |
| 59) 2,3,4,6-Tetrachlorophenol | 15.087 | 232 | 1266843 | 59.406 | ng | 99 |
| 60) Diethylphthalate | 15.287 | 149 | 4548361 | 47.643 | ng | 99 |
| 61) 4-Chlorophenyl-phenyle... | 15.510 | 204 | 2084668 | 43.232 | ng | 98 |
| 62) 4-Nitroaniline | 15.528 | 138 | 1019125 | 53.376 | ng | 92 |
| 63) Azobenzene | 15.810 | 77 | 4900001 | 41.754 | ng | 95 |
| 65) 4,6-Dinitro-2-methylph... | 15.592 | 198 | 760962 | 70.607 | ng | 95 |
| 66) n-Nitrosodiphenylamine | 15.728 | 169 | 3761401 | 46.195 | ng | 99 |
| 67) 4-Bromophenyl-phenylether | 16.428 | 248 | 1446723 | 44.989 | ng | 99 |
| 68) Hexachlorobenzene | 16.545 | 284 | 1689715 | 44.464 | ng | 100 |
| 69) Atrazine | 16.710 | 200 | 1219298 | 43.616 | ng | 100 |
| 70) Pentachlorophenol | 16.898 | 266 | 2234935 | 107.752 | ng | 99 |
| 71) Phenanthrene | 17.304 | 178 | 6812697 | 46.623 | ng | 100 |
| 72) Anthracene | 17.392 | 178 | 6814116 | 47.550 | ng | 100 |
| 73) Carbazole | 17.669 | 167 | 6597878 | 47.774 | ng | 99 |
| 74) Di-n-butylphthalate | 18.275 | 149 | 8213695 | 48.427 | ng | 99 |
| 75) Fluoranthene | 19.386 | 202 | 8258942 | 46.065 | ng | 99 |
| 77) Benzidine | 19.575 | 184 | 1737317 | 27.868 | ng | 99 |
| 78) Pyrene | 19.763 | 202 | 8508675 | 48.494 | ng | 100 |
| 80) Butylbenzylphthalate | 20.716 | 149 | 3674312 | 54.638 | ng | 95 |
| 81) Benzo(a)anthracene | 21.704 | 228 | 8065910 | 47.864 | ng | 99 |
| 82) 3,3'-Dichlorobenzidine | 21.621 | 252 | 2385563 | 44.152 | ng | 99 |
| 83) Chrysene | 21.774 | 228 | 7638418 | 48.009 | ng | 99 |
| 84) Bis(2-ethylhexyl)phtha... | 21.657 | 149 | 5396475 | 54.182 | ng | 100 |
| 85) Di-n-octyl phthalate | 22.980 | 149 | 9325220 | 52.766 | ng | 97 |
| 87) Indeno(1,2,3-cd)pyrene | 29.092 | 276 | 10117374m | 54.182 | ng | |
| 88) Benzo(b)fluoranthene | 24.086 | 252 | 8554644 | 51.707 | ng | 100 |
| 89) Benzo(k)fluoranthene | 24.157 | 252 | 8136624 | 50.415 | ng | 99 |
| 90) Benzo(a)pyrene | 25.010 | 252 | 7802789 | 54.347 | ng | 99 |
| 91) Dibenzo(a,h)anthracene | 29.198 | 278 | 8263413 | 53.260 | ng | 99 |
| 92) Benzo(g,h,i)perylene | 30.303 | 276 | 7985298 | 51.350 | ng | 100 |

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP082024\
Data File : BP021568.D
Acq On : 20 Aug 2024 11:31
Operator : RC/JU
Sample : PB162822BS
Misc :
ALS Vial : 4 Sample Multiplier: 1


Quant Time: Aug 20 12:11:11 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 14 00:00:30 2024
Response via : Initial Calibration

Instrument :
BNA_P
ClientSampleId :
PB162822BS

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 08/21/2024
Supervised By :mohammad ahmed 08/22/2024








```
#1  
1,4-Dichlorobenzene-d4  
Concen: 20.000 ng  
RT: 7.810 min Scan# 1  
Delta R.T. -0.001 min  
Lab File: BP021568.D  
Acq: 20 Aug 2024 11:31
```

Instrument :
BNA_P
ClientSampleId :
PB162822BS

| Tgt | Ion:152 | Resp: | 49465 |
|-----|---------|-------|-------|
| Ion | Ratio | Lower | Upper |
| 152 | 100 | | |
| 150 | 155.9 | 122.4 | 183.6 |
| 115 | 66.5 | 54.7 | 82.1 |





**Manual Integrations
APPROVED**




Reviewed By :Yogesh Patel 08/21/2024
Supervised By :mohammad ahmed 08/22/2024



#2
1,4-Dioxane
Concen: 28.508 ng
RT: 3.322 min Scan# 57
Delta R.T. 0.006 min
Lab File: BP021568.D
Acq: 20 Aug 2024 11:31

| Tgt | Ion: | 88 | Resp: | 450308 |
|-----|-------|------|-------|--------|
| Ion | Ratio | | Lower | Upper |
| 88 | 100 | | | |
| 58 | 60.3 | 48.7 | 73.1 | |
| 43 | 19.4 | 15.4 | 23.2 | |









#3
Pyridine
 Concen: 28.698 ng m
 RT: 3.717 min Scan# 1
 Delta R.T. 0.012 min
 Lab File: BP021568.D
 Acq: 20 Aug 2024 11:31

Instrument : BNA_P
 ClientSampleId : PB162822BS


1 Manual Integrations
 2 APPROVED




3 Reviewed By :Yogesh Patel 08/21/2024
 4 Supervised By :mohammad ahmed 08/22/2024



#4
n-Nitrosodimethylamine
 Concen: 42.420 ng
 RT: 3.622 min Scan# 108
 Delta R.T. 0.006 min
 Lab File: BP021568.D
 Acq: 20 Aug 2024 11:31

Tgt Ion: 42 Resp: 562224
 Ion Ratio Lower Upper
 42 100
 74 178.2 154.3 231.5
 44 6.6 6.2 9.4





#5

2-Fluorophenol

Concen: 139.062 ng

RT: 5.411 min Scan# 4

Delta R.T. 0.018 min

Lab File: BP021568.D

Acq: 20 Aug 2024 11:31

Instrument : BNA_P

ClientSampleId : PB162822BS

Tgt Ion:112 Resp: 463254

Ion Ratio Lower Upper

112 100

64 53.1 49.0 73.6





63 27.3 24.6 36.8

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 08/21/2024

Supervised By :mohammad ahmed 08/22/2024



#6

Aniline

Concen: 27.851 ng

RT: 7.134 min Scan# 705

Delta R.T. -0.000 min

Lab File: BP021568.D

Acq: 20 Aug 2024 11:31


Tgt Ion: 93 Resp: 1353041


Ion Ratio Lower Upper

93 100

66 35.5 30.2 45.2

65 17.8 14.7 22.1







#7
Phenol-d6
Concen: 130.771 ng
RT: 6.981 min Scan# 6
Delta R.T. 0.018 min
Lab File: BP021568.D
Acq: 20 Aug 2024 11:31

Instrument : BNA_P
ClientSampleId : PB162822BS


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/21/2024
Supervised By :mohammad ahmed 08/22/2024



#8
2-Chlorophenol
Concen: 53.119 ng
RT: 7.381 min Scan# 747
Delta R.T. 0.006 min
Lab File: BP021568.D
Acq: 20 Aug 2024 11:31

Tgt Ion:128 Resp: 1635284
Ion Ratio Lower Upper
128 100
130 31.7 11.6 51.6
64 48.0 35.3 75.3







#9
Benzaldehyde
Concen: 16.825 ng m
RT: 6.946 min Scan# 6
Delta R.T. -0.000 min
Lab File: BP021568.D
Acq: 20 Aug 2024 11:31

Instrument : BNA_P
ClientSampleId : PB162822BS


Manual Integrations
APPROVED




Reviewed By :Yogesh Patel 08/21/2024
Supervised By :mohammad ahmed 08/22/2024



#10
Phenol
Concen: 48.824 ng
RT: 7.005 min Scan# 683
Delta R.T. 0.018 min
Lab File: BP021568.D
Acq: 20 Aug 2024 11:31

Tgt Ion: 94 Resp: 2456949
Ion Ratio Lower Upper
94 100
65 24.2 6.5 46.5
66 33.4 14.7 54.7








#11
bis(2-Chloroethyl)ether
Concen: 41.972 ng
RT: 7.228 min Scan# 7
Delta R.T. -0.000 min
Lab File: BP021568.D
Acq: 20 Aug 2024 11:31

Instrument : BNA_P
ClientSampleId : PB162822BS



Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 08/21/2024
Supervised By :mohammad ahmed 08/22/2024



#12
1,3-Dichlorobenzene
Concen: 43.844 ng
RT: 7.705 min Scan# 802
Delta R.T. 0.006 min
Lab File: BP021568.D
Acq: 20 Aug 2024 11:31

Tgt Ion:146 Resp: 1603283
Ion Ratio Lower Upper
146 100
148 64.0 51.8 77.6
75 33.9 31.3 46.9





#13

1,4-Dichlorobenzene

Concen: 44.056 ng

RT: 7.846 min Scan# 8

Delta R.T. 0.006 min

Lab File: BP021568.D

Acq: 20 Aug 2024 11:31


Instrument : BNA_P

ClientSampleId : PB162822BS

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/21/2024

Supervised By :mohammad ahmed 08/22/2024



#14

1,2-Dichlorobenzene

Concen: 44.308 ng

RT: 8.157 min Scan# 879

Delta R.T. -0.000 min

Lab File: BP021568.D

Acq: 20 Aug 2024 11:31


Tgt Ion:146 Resp: 1576656


Ion Ratio Lower Upper

146 100

148 64.4 51.5 77.3


111 49.4 41.1 61.7





#15
Benzyl Alcohol
Concen: 47.766 ng
RT: 8.046 min Scan# 8
Delta R.T. 0.012 min
Lab File: BP021568.D
Acq: 20 Aug 2024 11:31




Instrument : BNA_P
ClientSampleId : PB162822BS




Tgt Ion: 79 Resp: 1665499
Ion Ratio Lower Upper
79 100
108 78.3 57.9 86.9
77 63.6 51.8 77.8

Manual Integrations APPROVED



Reviewed By :Yogesh Patel 08/21/2024
Supervised By :mohammad ahmed 08/22/2024






#16
2,2'-oxybis(1-Chloropropane)
Concen: 42.828 ng
RT: 8.328 min Scan# 908
Delta R.T. -0.006 min
Lab File: BP021568.D
Acq: 20 Aug 2024 11:31



Tgt Ion: 45 Resp: 1679419
Ion Ratio Lower Upper
45 100
77 20.8 1.9 41.9
79 15.3 0.0 36.6





#17

2-Methylphenol

Concen: 49.719 ng

RT: 8.246 min Scan# 8

Delta R.T. 0.012 min

Lab File: BP021568.D

Acq: 20 Aug 2024 11:31

Instrument :

BNA_P




ClientSampleId :

PB162822BS

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/21/2024

Supervised By :mohammad ahmed 08/22/2024



#18

Hexachloroethane

Concen: 45.372 ng

RT: 8.893 min Scan# 1004

Delta R.T. -0.000 min

Lab File: BP021568.D

Acq: 20 Aug 2024 11:31


Tgt Ion:117 Resp: 671831







Ion Ratio Lower Upper

117 100

119 94.2 73.9 110.9

201 96.6 74.9 112.3






#19

n-Nitroso-di-n-propylamine
Concen: 38.475 ng
RT: 8.610 min Scan# 9
Delta R.T. 0.006 min
Lab File: BP021568.D
Acq: 20 Aug 2024 11:31

Instrument :
BNA_P
ClientSampleId :
PB162822BS

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 08/21/2024
Supervised By :mohammad ahmed 08/22/2024







#20

3+4-Methylphenols
Concen: 51.486 ng
RT: 8.575 min Scan# 950
Delta R.T. 0.018 min
Lab File: BP021568.D
Acq: 20 Aug 2024 11:31

Tgt Ion:107 Resp: 2208573
Ion Ratio Lower Upper

| | 100 | | |
|-----|------|------|-------|
| 107 | 100 | | |
| 108 | 89.5 | 67.6 | 107.6 |
| 77 | 32.3 | 15.0 | 55.0 |
| 79 | 29.2 | 8.9 | 48.9 |





#21

Naphthalene-d8

Concen: 20.000 ng

RT: 10.599 min Scan# 1

Delta R.T. -0.000 min

Lab File: BP021568.D

Acq: 20 Aug 2024 11:31

Instrument :

BNA_P




ClientSampleId :

PB162822BS

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/21/2024

Supervised By :mohammad ahmed 08/22/2024



#22

Acetophenone

Concen: 39.810 ng

RT: 8.622 min Scan# 958

Delta R.T. 0.006 min

Lab File: BP021568.D

Acq: 20 Aug 2024 11:31

Tgt Ion:105 Resp: 2535541


Ion Ratio Lower Upper


105 100

71 5.2 4.6 7.0

51 19.4 16.3 24.5


120 20.6 16.2 24.2





#23
Nitrobenzene-d5
Concen: 93.712 ng
RT: 8.957 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP021568.D
Acq: 20 Aug 2024 11:31




Instrument : BNA_P
ClientSampleId : PB162822BS




Tgt Ion: 82 Resp: 376403
Ion Ratio Lower Upper
82 100
128 38.2 26.5 39.7
54 39.0 30.3 45.5

Manual Integrations APPROVED



Reviewed By :Yogesh Patel 08/21/2024
Supervised By :mohammad ahmed 08/22/2024




#24
Nitrobenzene
Concen: 44.668 ng
RT: 8.999 min Scan# 1022
Delta R.T. 0.006 min
Lab File: BP021568.D
Acq: 20 Aug 2024 11:31





Tgt Ion: 77 Resp: 1948611
Ion Ratio Lower Upper
77 100
123 43.7 29.5 44.3
65 12.7 9.9 14.9





Abundance Scan 1112 (9.528 min): BP021568.D\data.ms




Abundance Scan 1141 (9.699 min): BP021485.D\data.ms (-1)



Abundance Scan 1142 (9.704 min): BP021568.D\data.ms



Abundance Scan 1142 (9.704 min): BP021568.D\data.ms (-1)



#25

Isophorone

Concen: 40.856 ng

RT: 9.528 min Scan# 1

Delta R.T. 0.006 min

Lab File: BP021568.D

Acq: 20 Aug 2024 11:31

Instrument :

BNA_P


ClientSampleId :

PB162822BS

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 08/21/2024

Supervised By :mohammad ahmed 08/22/2024



#26

2-Nitrophenol

Concen: 63.198 ng

RT: 9.704 min Scan# 1142

Delta R.T. 0.005 min

Lab File: BP021568.D

Acq: 20 Aug 2024 11:31


Tgt Ion:139 Resp: 810333




Ion Ratio Lower Upper

139 100

109 34.6 28.9 43.3

65 54.7 49.0 73.6





#27

2,4-Dimethylphenol

Concen: 53.822 ng

RT: 9.769 min Scan# 1

Delta R.T. 0.012 min

Lab File: BP021568.D

Acq: 20 Aug 2024 11:31

Instrument :

BNA_P




ClientSampleId :

PB162822BS

**Manual Integrations
APPROVED**



Reviewed By :Yogesh Patel 08/21/2024




Supervised By :mohammad ahmed 08/22/2024



#28
bis(2-Chloroethoxy)methane
Concen: 42.148 ng
RT: 10.004 min Scan# 1193
Delta R.T. -0.000 min
Lab File: BP021568.D
Acq: 20 Aug 2024 11:31

Tgt Ion: 93 Resp: 2441309
Ion Ratio Lower Upper
93 100
95 32.5 26.2 39.4
123 10.0 7.6 11.4









#29
2,4-Dichlorophenol
Concen: 56.031 ng
RT: 10.240 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP021568.D
Acq: 20 Aug 2024 11:31

Instrument : BNA_P
ClientSampleId : PB162822BS


Manual Integrations
APPROVED




Reviewed By :Yogesh Patel 08/21/2024
Supervised By :mohammad ahmed 08/22/2024



#30
1,2,4-Trichlorobenzene
Concen: 43.572 ng
RT: 10.457 min Scan# 1270
Delta R.T. -0.000 min
Lab File: BP021568.D
Acq: 20 Aug 2024 11:31

Tgt Ion:180 Resp: 1545646
Ion Ratio Lower Upper
180 100
182 95.8 78.3 117.5
145 30.2 24.0 36.0





#31

Naphthalene

Concen: 43.861 ng

RT: 10.646 min Scan# 1

Delta R.T. -0.000 min

Lab File: BP021568.D

Acq: 20 Aug 2024 11:31

Instrument :

BNA_P




ClientSampleId :

PB162822BS

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 08/21/2024


Supervised By :mohammad ahmed 08/22/2024



#32
 Benzoic acid
 Concen: 69.217 ng
 RT: 9.928 min Scan# 1180
 Delta R.T. 0.053 min
 Lab File: BP021568.D
 Acq: 20 Aug 2024 11:31

Tgt Ion:122 Resp: 1255905
 Ion Ratio Lower Upper
 122 100
 105 138.0 124.6 164.6
 77 102.2 96.2 136.2







#33
4-Chloroaniline
Concen: 28.216 ng
RT: 10.746 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP021568.D
Acq: 20 Aug 2024 11:31

Instrument :
BNA_P
ClientSampleId :
PB162822BS


**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/21/2024
Supervised By :mohammad ahmed 08/22/2024



#34
Hexachlorobutadiene
Concen: 41.546 ng
RT: 10.940 min Scan# 1352
Delta R.T. -0.000 min
Lab File: BP021568.D
Acq: 20 Aug 2024 11:31

| Tgt | Ion:225 | Resp: | 927946 |
|-----|---------|-------|--------|
| Ion | Ratio | Lower | Upper |
| 225 | 100 | | |
| 223 | 62.1 | 49.6 | 74.4 |
| 227 | 63.9 | 50.1 | 75.1 |



#35

Caprolactam

Concen: 51.386 ng/m

RT: 11.534 min Scan# 1447

Delta R.T. 0.035 min

Lab File: BP021568.D




Acq: 20 Aug 2024 11:31

Instrument :

BNA_P

ClientSampleId :

PB162822BS



Tgt Ion:113 Resp: 591792

Ion Ratio Lower Upper

113 100





55 138.6 130.6 170.6

56 105.8 94.6 134.6

Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel 08/21/2024

Supervised By :mohammad ahmed 08/22/2024



#36

4-Chloro-3-methylphenol

Concen: 53.276 ng

RT: 11.875 min Scan# 1511

Delta R.T. 0.012 min

Lab File: BP021568.D

Acq: 20 Aug 2024 11:31


Tgt Ion:107 Resp: 2044802

Ion Ratio Lower Upper

107 100

144 21.4 16.6 24.8

142 67.5 51.0 76.4



#37

2-Methylnaphthalene

Concen: 44.806 ng

RT: 12.257 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP021568.D


Acq: 20 Aug 2024 11:31

Instrument :


BNA_P

ClientSampleId :


PB162822BS



Abundance Scan 1576 (12.257 min): BP021568.D\data.ms (-)



Abundance Scan 1576 (12.257 min): BP021568.D\data.ms (-)



Tgt Ion:142 Resp: 331187

Ion Ratio Lower Upper

142 100

141 89.6 71.7 107.5



115 39.2 31.4 47.2

Manual Integrations


APPROVED

Reviewed By :Yogesh Patel 08/21/2024


Supervised By :mohammad ahmed 08/22/2024



Abundance Scan 1614 (12.481 min): BP021568.D\data.ms (-)



Abundance Scan 1614 (12.481 min): BP021568.D\data.ms (-)



#38

1-Methylnaphthalene

Concen: 42.651 ng

RT: 12.481 min Scan# 1614

Delta R.T. -0.000 min

Lab File: BP021568.D

Acq: 20 Aug 2024 11:31


Tgt Ion:142 Resp: 3109322

Ion Ratio Lower Upper

142 100

141 92.2 73.5 110.3

116 4.3 3.4 5.2



#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 14.451 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP021568.D




Acq: 20 Aug 2024 11:31

Instrument:

BNA_P

ClientSampleId :

PB162822BS



Tgt Ion:164 Resp: 1389370

Ion Ratio Lower Upper

164 100





162 99.6 81.5 122.3

160 43.9 35.2 52.8

Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel 08/21/2024

Supervised By :mohammad ahmed 08/22/2024



#40

1,2,4,5-Tetrachlorobenzene

Concen: 40.778 ng

RT: 12.634 min Scan# 1640

Delta R.T. 0.006 min

Lab File: BP021568.D

Acq: 20 Aug 2024 11:31

Tgt Ion:216 Resp: 1660023


Ion Ratio Lower Upper


216 100

214 78.6 62.5 93.7

179 20.6 16.4 24.6

108 21.8 16.7 25.1







#41
Hexachlorocyclopentadiene
Concen: 116.470 ng
RT: 12.610 min Scan# 1
Delta R.T. -0.012 min
Lab File: BP021568.D
Acq: 20 Aug 2024 11:31

Instrument : BNA_P
ClientSampleId : PB162822BS


Manual Integrations
APPROVED




Reviewed By :Yogesh Patel 08/21/2024
Supervised By :mohammad ahmed 08/22/2024



#42
2,4,6-Tribromophenol
Concen: 157.161 ng
RT: 15.963 min Scan# 2206
Delta R.T. -0.006 min
Lab File: BP021568.D
Acq: 20 Aug 2024 11:31

Tgt Ion:330 Resp: 2430215
Ion Ratio Lower Upper
330 100
332 96.1 77.4 116.0
141 29.4 24.5 36.7





#43

2,4,6-Trichlorophenol

Concen: 57.099 ng

RT: 12.869 min Scan# 1

Delta R.T. -0.000 min

Lab File: BP021568.D

Acq: 20 Aug 2024 11:31

Instrument :

BNA_P





ClientSampleId :

PB162822BS

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/21/2024

Supervised By :mohammad ahmed 08/22/2024



#44

2,4,5-Trichlorophenol

Concen: 55.286 ng


RT: 12.945 min Scan# 1693




Delta R.T. 0.012 min

Lab File: BP021568.D

Acq: 20 Aug 2024 11:31

| Tgt | Ion:196 | Resp: 1389021 | |
|-----|---------|---------------|-------|
| Ion | Ratio | Lower | Upper |
| 196 | 100 | | |
| 198 | 98.3 | 78.9 | 118.3 |
| 97 | 57.4 | 49.4 | 74.0 |
| 132 | 27.3 | 21.1 | 31.7 |





#45

2-Fluorobiphenyl

Concen: 79.575 ng

RT: 13.063 min Scan# 1

Delta R.T. -0.000 min

Lab File: BP021568.D

Acq: 20 Aug 2024 11:31

Instrument :

BNA_P

ClientSampleId :


PB162822BS

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/21/2024

Supervised By :mohammad ahmed 08/22/2024

Abundance



#46

1,1'-Biphenyl

Concen: 40.658 ng

RT: 13.275 min Scan# 1749

Delta R.T. -0.006 min

Lab File: BP021568.D

Acq: 20 Aug 2024 11:31

Tgt Ion:154 Resp: 3996688





Ion Ratio Lower Upper




154 100

153 40.5 20.7 60.7

76 15.0 0.0 37.7

Abundance





#47

2-Chloronaphthalene

Concen: 43.922 ng

RT: 13.316 min Scan# 1

Delta R.T. -0.000 min

Lab File: BP021568.D

Acq: 20 Aug 2024 11:31

Instrument:

BNA_P




ClientSampleId :

PB162822BS

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/21/2024

Supervised By :mohammad ahmed 08/22/2024



#48

2-Nitroaniline

Concen: 54.598 ng

RT: 13.510 min Scan# 1789

Delta R.T. 0.006 min

Lab File: BP021568.D

Acq: 20 Aug 2024 11:31


Tgt Ion: 65 Resp: 1188075




Ion Ratio Lower Upper

65 100

92 79.3 58.1 87.1

138 100.0 65.5 98.3#





#49

Acenaphthylene

Concen: 48.890 ng

RT: 14.169 min Scan# 1

Delta R.T. -0.000 min

Lab File: BP021568.D

Acq: 20 Aug 2024 11:31





Instrument : BNA_P

ClientSampleId : PB162822BS

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/21/2024

Supervised By :mohammad ahmed 08/22/2024



#50

Dimethylphthalate

Concen: 47.945 ng

RT: 13.904 min Scan# 1856

Delta R.T. -0.006 min

Lab File: BP021568.D

Acq: 20 Aug 2024 11:31


Tgt Ion:163 Resp: 4461019


Ion Ratio Lower Upper

163 100

194 4.1 3.3 4.9

164 10.0 8.3 12.5







#51
2,6-Dinitrotoluene
Concen: 52.003 ng
RT: 14.010 min Scan# 1
Delta R.T. -0.012 min
Lab File: BP021568.D
Acq: 20 Aug 2024 11:31

Instrument : BNA_P
ClientSampleId : PB162822BS


1 Manual Integrations
2 APPROVED




3 Reviewed By :Yogesh Patel 08/21/2024
4 Supervised By :mohammad ahmed 08/22/2024



#52
Acenaphthene
Concen: 51.606 ng
RT: 14.516 min Scan# 1960
Delta R.T. -0.006 min
Lab File: BP021568.D
Acq: 20 Aug 2024 11:31

Tgt Ion:154 Resp: 3809412
Ion Ratio Lower Upper
154 100
153 112.3 91.1 136.7
152 53.2 43.0 64.4









#53
3-Nitroaniline
Concen: 43.541 ng
RT: 14.345 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP021568.D
Acq: 20 Aug 2024 11:31

Instrument : BNA_P
ClientSampleId : PB162822BS


Manual Integrations APPROVED




Reviewed By :Yogesh Patel 08/21/2024
Supervised By :mohammad ahmed 08/22/2024



#54
2,4-Dinitrophenol
Concen: 123.071 ng
RT: 14.557 min Scan# 1967
Delta R.T. 0.006 min
Lab File: BP021568.D
Acq: 20 Aug 2024 11:31

Tgt Ion:184 Resp: 1036408
Ion Ratio Lower Upper
184 100
63 63.8 64.9 97.3#
154 61.2 52.1 78.1









#55
Dibenzofuran
Concen: 45.731 ng
RT: 14.857 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP021568.D
Acq: 20 Aug 2024 11:31

Instrument : BNA_P
ClientSampleId : PB162822BS


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/21/2024
Supervised By :mohammad ahmed 08/22/2024



#56
4-Nitrophenol
Concen: 116.168 ng
RT: 14.669 min Scan# 1986
Delta R.T. 0.018 min
Lab File: BP021568.D
Acq: 20 Aug 2024 11:31

Tgt Ion:139 Resp: 1830474
Ion Ratio Lower Upper
139 100
109 81.7 70.2 110.2
65 100.2 101.8 141.8#




#57
 2,4-Dinitrotoluene
 Concen: 58.078 ng
 RT: 14.816 min Scan# 2130
 Delta R.T. 0.006 min
 Lab File: BP021568.D
 Acq: 20 Aug 2024 11:31

Instrument : BNA_P
 ClientSampleId : PB162822BS

Tgt Ion:165 Resp: 141471:
 Ion Ratio Lower Upper
 165 100
 63 41.8 42.5 63.7
 89 80.1 77.1 115.7
 182 2.9 2.0 3.0








Manual Integrations APPROVED

Reviewed By :Yogesh Patel 08/21/2024
 Supervised By :mohammad ahmed 08/22/2024



#58
 Fluorene
 Concen: 43.711 ng
 RT: 15.516 min Scan# 2130
 Delta R.T. -0.000 min
 Lab File: BP021568.D
 Acq: 20 Aug 2024 11:31

Tgt Ion:166 Resp: 4077699
 Ion Ratio Lower Upper
 166 100
 165 97.8 78.1 117.1
 167 13.6 11.0 16.4




#59
 2,3,4,6-Tetrachlorophenol
 Concen: 59.406 ng
 RT: 15.087 min Scan# 2
 Delta R.T. -0.000 min
 Lab File: BP021568.D
 Acq: 20 Aug 2024 11:31

Instrument : BNA_P
 ClientSampleId : PB162822BS


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/21/2024
 Supervised By :mohammad ahmed 08/22/2024



Tgt Ion:232 Resp: 126684
 Ion Ratio Lower Upper



| | | | |
|-----|------|------|------|
| 232 | 100 | | |
| 131 | 42.1 | 33.1 | 49.7 |
| 130 | 2.2 | 1.7 | 2.5 |
| 166 | 29.4 | 23.0 | 34.4 |




#60
 Diethylphthalate
 Concen: 47.643 ng
 RT: 15.287 min Scan# 2091
 Delta R.T. -0.018 min
 Lab File: BP021568.D
 Acq: 20 Aug 2024 11:31


Tgt Ion:149 Resp: 4548361
 Ion Ratio Lower Upper

| | | | |
|-----|------|------|------|
| 149 | 100 | | |
| 177 | 23.2 | 17.8 | 26.8 |
| 150 | 12.7 | 10.1 | 15.1 |






Abundance Scan 2129 (15.510 min): BP021568.D\data.ms



Abundance Scan 2129 (15.510 min): BP021568.D\data.ms (-)



#61
4-Chlorophenyl-phenylether
Concen: 43.232 ng
RT: 15.510 min Scan# 2129
Delta R.T. -0.012 min
Lab File: BP021568.D
Acq: 20 Aug 2024 11:31

Tgt Ion:204 Resp: 2084663

Ion Ratio Lower Upper

204 100

206 33.4 26.4 39.6

141 55.9 46.1 69.1


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 08/21/2024


Supervised By :mohammad ahmed 08/22/2024

Abundance


Time--> 15.40 15.50 15.60



Abundance Scan 2132 (15.528 min): BP021568.D\data.ms



Abundance Scan 2132 (15.528 min): BP021568.D\data.ms (-)



#62
4-Nitroaniline
Concen: 53.376 ng
RT: 15.528 min Scan# 2132
Delta R.T. -0.000 min
Lab File: BP021568.D
Acq: 20 Aug 2024 11:31

Tgt Ion:138 Resp: 1019125




Ion Ratio Lower Upper

138 100

92 59.0 46.4 86.4

108 87.2 73.2 113.2

Time--> 15.40 15.528 15.60



#63

Azobenzene

Concen: 41.754 ng

RT: 15.810 min Scan# 2180

Delta R.T. -0.012 min

Lab File: BP021568.D

Acq: 20 Aug 2024 11:31

Instrument :

BNA_P




ClientSampleId :

PB162822BS

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/21/2024

Supervised By :mohammad ahmed 08/22/2024



#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 17.257 min Scan# 2426

Delta R.T. 0.006 min

Lab File: BP021568.D

Acq: 20 Aug 2024 11:31


Tgt Ion:188 Resp: 3038381

Ion Ratio Lower Upper

188 100

94 11.5 10.2 15.4

80 12.7 11.6 17.4




#65
 4,6-Dinitro-2-methylphenol
 Concen: 70.607 ng
 RT: 15.592 min Scan# 2165
 Delta R.T. 0.005 min
 Lab File: BP021568.D
 Acq: 20 Aug 2024 11:31

Instrument : BNA_P
 ClientSampleId : PB162822BS

Tgt Ion:198 Resp: 760963
 Ion Ratio Lower Upper
 198 100
 51 37.4 21.5 61.5
 105 49.9 32.9 72.9


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 08/21/2024
 Supervised By :mohammad ahmed 08/22/2024



#66
 n-Nitrosodiphenylamine
 Concen: 46.195 ng
 RT: 15.728 min Scan# 2166
 Delta R.T. -0.012 min
 Lab File: BP021568.D
 Acq: 20 Aug 2024 11:31

Tgt Ion:169 Resp: 3761401
 Ion Ratio Lower Upper
 169 100
 168 66.9 53.4 80.2
 167 36.6 28.7 43.1



#1

#2

#3

#4

#5

#6

#7

#8

#9

#10

#11

#12



#13

#14

#15

#16

#17




#67
 4-Bromophenyl-phenylether
 Concen: 44.989 ng
 RT: 16.428 min Scan# 2
 Delta R.T. -0.012 min
 Lab File: BP021568.D
 Acq: 20 Aug 2024 11:31

Instrument : BNA_P
 ClientSampleId : PB162822BS

Tgt Ion:248 Resp: 144672
 Ion Ratio Lower Upper
 248 100
 250 99.0 78.6 117.8
 141 65.7 53.9 80.9




Manual Integrations APPROVED




Reviewed By :Yogesh Patel 08/21/2024
 Supervised By :mohammad ahmed 08/22/2024



#68
 Hexachlorobenzene
 Concen: 44.464 ng
 RT: 16.545 min Scan# 2305
 Delta R.T. -0.006 min
 Lab File: BP021568.D
 Acq: 20 Aug 2024 11:31

Tgt Ion:284 Resp: 1689715
 Ion Ratio Lower Upper
 284 100
 142 29.6 23.4 35.2
 249 30.3 24.1 36.1





#69

Atrazine

Concen: 43.616 ng

RT: 16.710 min Scan# 2

Delta R.T. -0.006 min

Lab File: BP021568.D

Acq: 20 Aug 2024 11:31

Instrument :

BNA_P





ClientSampleId :

PB162822BS

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/21/2024

Supervised By :mohammad ahmed 08/22/2024



#70

Pentachlorophenol

Concen: 107.752 ng

RT: 16.898 min Scan# 2365

Delta R.T. 0.006 min

Lab File: BP021568.D

Acq: 20 Aug 2024 11:31


Tgt Ion:266 Resp: 2234935


Ion Ratio Lower Upper

266 100


268 64.1 51.8 77.6

264 63.7 50.5 75.7






Abundance Scan 2434 (17.304 min): BP021568.D\data.ms (-)



Abundance Scan 2434 (17.304 min): BP021568.D\data.ms (-)



#71

Phenanthrene

Concen: 46.623 ng

RT: 17.304 min Scan# 2449

Delta R.T. 0.006 min

Lab File: BP021568.D

Acq: 20 Aug 2024 11:31

Instrument :

BNA_P


ClientSampleId :

PB162822BS


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/21/2024


Supervised By :mohammad ahmed 08/22/2024



Abundance Scan 2449 (17.392 min): BP021568.D\data.ms (-)



Abundance Scan 2449 (17.392 min): BP021568.D\data.ms (-)



#72

Anthracene

Concen: 47.550 ng

RT: 17.392 min Scan# 2449

Delta R.T. -0.006 min

Lab File: BP021568.D

Acq: 20 Aug 2024 11:31



Tgt Ion:178 Resp: 6814116




Ion Ratio Lower Upper

178 100

176 18.9 14.9 22.3

179 15.7 12.5 18.7









#73
Carbazole
Concen: 47.774 ng
RT: 17.669 min Scan# 2
Delta R.T. -0.000 min
Lab File: BP021568.D
Acq: 20 Aug 2024 11:31

Instrument : BNA_P
ClientSampleId : PB162822BS


Manual Integrations APPROVED







Reviewed By :Yogesh Patel 08/21/2024
Supervised By :mohammad ahmed 08/22/2024



#74
Di-n-butylphthalate
Concen: 48.427 ng
RT: 18.275 min Scan# 2599
Delta R.T. -0.000 min
Lab File: BP021568.D
Acq: 20 Aug 2024 11:31

Tgt Ion:149 Resp: 8213695
Ion Ratio Lower Upper
149 100
150 9.3 7.5 11.3
104 5.8 4.9 7.3





#75

Fluoranthene

Concen: 46.065 ng

RT: 19.386 min Scan# 2

Delta R.T. 0.006 min

Lab File: BP021568.D

Acq: 20 Aug 2024 11:31

Instrument :

BNA_P


ClientSampleId :

PB162822BS

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 08/21/2024

Supervised By :mohammad ahmed 08/22/2024



#76

Chrysene-d12

Concen: 20.000 ng

RT: 21.727 min Scan# 3186

Delta R.T. 0.029 min

Lab File: BP021568.D

Acq: 20 Aug 2024 11:31

Tgt Ion:240 Resp: 2684396

Ion Ratio Lower Upper

240 100

120 10.9


236 25.7




9.9

14.9

20.6

30.8









#77
Benzidine
Concen: 27.868 ng
RT: 19.575 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP021568.D
Acq: 20 Aug 2024 11:31

Instrument : BNA_P
ClientSampleId : PB162822BS


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/21/2024
Supervised By :mohammad ahmed 08/22/2024



#78
Pyrene
Concen: 48.494 ng
RT: 19.763 min Scan# 2852
Delta R.T. 0.012 min
Lab File: BP021568.D
Acq: 20 Aug 2024 11:31

Tgt Ion:202 Resp: 8508675
Ion Ratio Lower Upper
202 100
200 21.1 16.7 25.1
203 17.8 14.1 21.1



#79

Terphenyl-d14

Concen: 85.222 ng

RT: 19.980 min Scan# 2

Delta R.T. 0.005 min

Lab File: BP021568.D

Acq: 20 Aug 2024 11:31

Instrument :

BNA_P

ClientSampleId :

PB162822BS

Tgt Ion:244 Resp:1177563

Ion Ratio Lower Upper

244 100

212 8.0 5.6 8.4

122 11.8 9.4 14.2

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 08/21/2024

Supervised By :mohammad ahmed 08/22/2024

Abundance

8000000 19.980

6000000

4000000

2000000

0

Time-->

Time-->

#80

Butylbenzylphthalate

Concen: 54.638 ng

RT: 20.716 min Scan# 3014

Delta R.T. 0.006 min

Lab File: BP021568.D

Acq: 20 Aug 2024 11:31

Tgt Ion:149 Resp: 3674312

Ion Ratio Lower Upper

149 100

91 85.9 73.3 109.9

206 21.4 17.1 25.7

Abundance

3000000 20.716







2000000

1000000

0

Time-->

20.70



#81

Benzo(a)anthracene

Concen: 47.864 ng

RT: 21.704 min Scan# 3

Delta R.T. 0.023 min

Lab File: BP021568.D

Acq: 20 Aug 2024 11:31

Instrument :

BNA_P

ClientSampleId :

PB162822BS

Tgt Ion:228 Resp: 8065910

Ion Ratio Lower Upper

228 100

226 27.6 21.4 32.2

229 20.2 15.8 23.6

Manual Integrations


APPROVED

Reviewed By :Yogesh Patel 08/21/2024

Supervised By :mohammad ahmed 08/22/2024

Abundance

21.704



#82

3,3'-Dichlorobenzidine

Concen: 44.152 ng

RT: 21.621 min Scan# 3168

Delta R.T. 0.012 min

Lab File: BP021568.D

Acq: 20 Aug 2024 11:31

Tgt Ion:252 Resp: 2385563

Ion Ratio Lower Upper




252 100

254 63.8 51.7 77.5

126 10.8 8.7 13.1

Abundance

21.621



#83

Chrysene

Concen: 48.009 ng

RT: 21.774 min Scan# 3

Delta R.T. 0.023 min

Lab File: BP021568.D

Acq: 20 Aug 2024 11:31

Instrument :

BNA_P

ClientSampleId :

PB162822BS

Tgt Ion:228 Resp: 763841

Ion Ratio Lower Upper

228 100

226 30.5 23.9 35.9

229 20.3 15.8 23.6

Manual Integrations


APPROVED

Reviewed By :Yogesh Patel 08/21/2024

Supervised By :mohammad ahmed 08/22/2024

Abundance

21.774



#84

Bis(2-ethylhexyl)phthalate

Concen: 54.182 ng

RT: 21.657 min Scan# 3174

Delta R.T. 0.012 min

Lab File: BP021568.D

Acq: 20 Aug 2024 11:31

Tgt Ion:149 Resp: 5396475

Ion Ratio Lower Upper




149 100


167 27.9 22.3 33.5

279 4.8 3.8 5.6

Abundance

21.657







#85
Di-n-octyl phthalate
Concen: 52.766 ng
RT: 22.980 min Scan# 3
Delta R.T. 0.017 min
Lab File: BP021568.D
Acq: 20 Aug 2024 11:31

Instrument : BNA_P
ClientSampleId : PB162822BS


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 08/21/2024
Supervised By :mohammad ahmed 08/22/2024









#86
Perylene-d12
Concen: 20.000 ng
RT: 25.168 min Scan# 3771
Delta R.T. 0.065 min
Lab File: BP021568.D
Acq: 20 Aug 2024 11:31

Tgt Ion:264 Resp: 2880701
Ion Ratio Lower Upper
264 100
260 23.7 18.6 28.0
265 21.7 17.1 25.7



#87


Indeno(1,2,3-cd)pyrene
 Concen: 54.182 ng/m
 RT: 29.092 min Scan# 4
 Delta R.T. 0.100 min
 Lab File: BP021568.D
 Acq: 20 Aug 2024 11:31



Instrument : BNA_P
 ClientSampleId : PB162822BS


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 08/21/2024
 Supervised By :mohammad ahmed 08/22/2024



#88
 Benzo(b)fluoranthene
 Concen: 51.707 ng
 RT: 24.086 min Scan# 3587
 Delta R.T. 0.059 min
 Lab File: BP021568.D
 Acq: 20 Aug 2024 11:31

Tgt Ion:252 Resp: 8554644
 Ion Ratio Lower Upper
 252 100
 253 22.0 17.6 26.4
 125 10.3 8.6 13.0



#89

Benzo(k)fluoranthene

Concen: 50.415 ng

RT: 24.157 min Scan# 3

Delta R.T. 0.047 min

Lab File: BP021568.D

Acq: 20 Aug 2024 11:31

Instrument :

BNA_P

ClientSampleId :

PB162822BS

Tgt Ion:252 Resp: 813662

Ion Ratio Lower Upper

252 100

253 22.0 17.5 26.3

125 10.4 7.8 11.8

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 08/21/2024


Supervised By :mohammad ahmed 08/22/2024

Abundance

323.1 392.1 458.1 548.1

318.2 406.8 480.3 546.1

Time-->



#90

Benzo(a)pyrene

Concen: 54.347 ng

RT: 25.010 min Scan# 3744

Delta R.T. 0.059 min

Lab File: BP021568.D

Acq: 20 Aug 2024 11:31

Tgt Ion:252 Resp: 7802789

Ion Ratio Lower Upper

252 100

253 21.8 17.7 26.5




125 11.6 8.7 13.1







Abundance

324.1 392.2 458.1 539.5

318.1 446.6 518.3

Time-->





#91

Dibenzo(a,h)anthracene

Concen: 53.260 ng

RT: 29.198 min Scan# 4

Delta R.T. 0.112 min

Lab File: BP021568.D

Acq: 20 Aug 2024 11:31

Instrument :

BNA_P


ClientSampleId :

PB162822BS

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 08/21/2024

Supervised By :mohammad ahmed 08/22/2024



#92

Benzo(g,h,i)perylene

Concen: 51.350 ng

RT: 30.303 min Scan# 4644

Delta R.T. 0.129 min


Lab File: BP021568.D

Acq: 20 Aug 2024 11:31

Tgt Ion:276 Resp: 7985298

Ion Ratio Lower Upper

| | 100 | | |
|-----|------|------|------|
| 276 | 100 | | |
| 277 | 23.7 | 19.0 | 28.6 |
| 138 | 19.8 | 16.0 | 24.0 |





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

| | | | | | |
|--------------------|---------------------------------------|--------------|-----|-----------------|---------------|
| Client: | JACOBS Engineering Group, Inc. | | | Date Collected: | |
| Project: | Former Schlumberger Site Princeton NJ | | | Date Received: | |
| Client Sample ID: | PB162822BSD | | | SDG No.: | P3657 |
| Lab Sample ID: | PB162822BSD | | | Matrix: | Water |
| Analytical Method: | SW8270 | | | % Solid: | 0 |
| Sample Wt/Vol: | 1000 | Units: | mL | Final Vol: | 1000 uL |
| Soil Aliquot Vol: | | | uL | Test: | SVOCMS Group6 |
| Extraction Type : | | Decanted : | N | Level : | LOW |
| Injection Volume : | | GPC Factor : | 1.0 | GPC Cleanup : | N PH : |
| Prep Method : | SW3510C | | | | |

| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed | Prep Batch ID |
|-------------------|-----------|----------------|----------------|---------------|
| BP021569.D | 1 | 08/19/24 09:40 | 08/20/24 12:13 | PB162822 |

| CAS Number | Parameter | Conc. | Qualifier | MDL | LOQ / CRQL | Units |
|----------------|----------------------------|-------|-----------|------|------------|-------|
| TARGETS | | | | | | |
| 110-86-1 | Pyridine | 28.8 | | 1.60 | 5.00 | ug/L |
| 100-52-7 | Benzaldehyde | 21.4 | | 4.00 | 10.0 | ug/L |
| 95-48-7 | 2-Methylphenol | 48.7 | | 1.10 | 5.00 | ug/L |
| 65794-96-9 | 3+4-Methylphenols | 50.1 | | 1.20 | 10.0 | ug/L |
| 67-72-1 | Hexachloroethane | 44.9 | | 1.00 | 5.00 | ug/L |
| 98-95-3 | Nitrobenzene | 44.7 | | 1.30 | 5.00 | ug/L |
| 91-20-3 | Naphthalene | 43.7 | | 1.00 | 5.00 | ug/L |
| 87-68-3 | Hexachlorobutadiene | 41.6 | | 1.30 | 5.00 | ug/L |
| 91-57-6 | 2-Methylnaphthalene | 44.2 | | 1.10 | 5.00 | ug/L |
| 88-06-2 | 2,4,6-Trichlorophenol | 57.4 | | 0.89 | 5.00 | ug/L |
| 95-95-4 | 2,4,5-Trichlorophenol | 55.2 | | 1.00 | 5.00 | ug/L |
| 208-96-8 | Acenaphthylene | 49.4 | | 1.00 | 5.00 | ug/L |
| 83-32-9 | Acenaphthene | 51.4 | | 0.81 | 5.00 | ug/L |
| 132-64-9 | Dibenzofuran | 46.3 | | 0.93 | 5.00 | ug/L |
| 121-14-2 | 2,4-Dinitrotoluene | 57.8 | | 1.50 | 5.00 | ug/L |
| 86-73-7 | Fluorene | 43.6 | | 0.96 | 5.00 | ug/L |
| 118-74-1 | Hexachlorobenzene | 45.0 | | 1.10 | 5.00 | ug/L |
| 87-86-5 | Pentachlorophenol | 110 | E | 1.90 | 10.0 | ug/L |
| 85-01-8 | Phenanthrene | 47.6 | | 0.89 | 5.00 | ug/L |
| 120-12-7 | Anthracene | 48.7 | | 1.10 | 5.00 | ug/L |
| 86-74-8 | Carbazole | 48.9 | | 1.20 | 5.00 | ug/L |
| 84-74-2 | Di-n-butylphthalate | 48.1 | | 1.50 | 5.00 | ug/L |
| 206-44-0 | Fluoranthene | 46.4 | | 1.30 | 5.00 | ug/L |
| 129-00-0 | Pyrene | 48.0 | | 1.10 | 5.00 | ug/L |
| 56-55-3 | Benzo(a)anthracene | 48.7 | | 0.94 | 5.00 | ug/L |
| 218-01-9 | Chrysene | 48.8 | | 0.86 | 5.00 | ug/L |
| 117-81-7 | Bis(2-ethylhexyl)phthalate | 53.5 | | 1.90 | 5.00 | ug/L |
| 205-99-2 | Benzo(b)fluoranthene | 49.6 | | 1.10 | 5.00 | ug/L |
| 207-08-9 | Benzo(k)fluoranthene | 51.0 | | 1.20 | 5.00 | ug/L |



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

| | | | | | |
|--------------------|---------------------------------------|--------|----|-----------------|---------------|
| Client: | JACOBS Engineering Group, Inc. | | | Date Collected: | |
| Project: | Former Schlumberger Site Princeton NJ | | | Date Received: | |
| Client Sample ID: | PB162822BSD | | | SDG No.: | P3657 |
| Lab Sample ID: | PB162822BSD | | | Matrix: | Water |
| Analytical Method: | SW8270 | | | % Solid: | 0 |
| Sample Wt/Vol: | 1000 | Units: | mL | Final Vol: | 1000 uL |
| Soil Aliquot Vol: | uL | | | Test: | SVOCMS Group6 |
| Extraction Type : | Decanted : N | | | Level : | LOW |
| Injection Volume : | GPC Factor : 1.0 | | | GPC Cleanup : | N PH : |
| Prep Method : | SW3510C | | | | |

| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed | Prep Batch ID |
|-------------------|-----------|----------------|----------------|---------------|
| BP021569.D | 1 | 08/19/24 09:40 | 08/20/24 12:13 | PB162822 |

| CAS Number | Parameter | Conc. | Qualifier | MDL | LOQ / CRQL | Units |
|-------------------|------------------------|-------|-----------|---------------------|------------|----------|
| 50-32-8 | Benzo(a)pyrene | 54.0 | | 1.70 | 5.00 | ug/L |
| 193-39-5 | Indeno(1,2,3-cd)pyrene | 53.9 | | 1.00 | 5.00 | ug/L |
| 53-70-3 | Dibenzo(a,h)anthracene | 53.3 | | 1.20 | 5.00 | ug/L |
| 191-24-2 | Benzo(g,h,i)perylene | 50.9 | | 1.20 | 5.00 | ug/L |
| 123-91-1 | 1,4-Dioxane | 28.8 | | 1.30 | 5.00 | ug/L |
| 90-12-0 | 1-Methylnaphthalene | 42.0 | | 0.86 | 5.00 | ug/L |
| SURROGATES | | | | | | |
| 367-12-4 | 2-Fluorophenol | 137 | | 15 (10) - 110 (139) | 91% | SPK: 150 |
| 13127-88-3 | Phenol-d6 | 128 | | 15 (10) - 110 (134) | 85% | SPK: 150 |
| 4165-60-0 | Nitrobenzene-d5 | 93.0 | | 30 (49) - 130 (133) | 93% | SPK: 100 |
| 321-60-8 | 2-Fluorobiphenyl | 80.2 | | 30 (52) - 130 (132) | 80% | SPK: 100 |
| 118-79-6 | 2,4,6-Tribromophenol | 154 | | 15 (44) - 110 (137) | 103% | SPK: 150 |
| 1718-51-0 | Terphenyl-d14 | 88.7 | | 30 (48) - 130 (125) | 89% | SPK: 100 |

INTERNAL STANDARDS

| | | | |
|------------|------------------------|---------|--------|
| 3855-82-1 | 1,4-Dichlorobenzene-d4 | 491000 | 7.805 |
| 1146-65-2 | Naphthalene-d8 | 2020000 | 10.593 |
| 15067-26-2 | Acenaphthene-d10 | 1320000 | 14.457 |
| 1517-22-2 | Phenanthrene-d10 | 2800000 | 17.269 |
| 1719-03-5 | Chrysene-d12 | 2520000 | 21.739 |
| 1520-96-3 | Perylene-d12 | 2790000 | 25.186 |

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP082024\
 Data File : BP021569.D
 Acq On : 20 Aug 2024 12:13
 Operator : RC/JU
 Sample : PB162822BSD
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 PB162822BSD

Quant Time: Aug 20 12:41:50 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 14 00:00:30 2024
 Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/21/2024
 Supervised By :mohammad ahmed 08/22/2024

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|------------------------------------|--------|------|----------|---------|-------|----------|
| Internal Standards | | | | | | |
| 1) 1,4-Dichlorobenzene-d4 | 7.805 | 152 | 491039 | 20.000 | ng | 0.00 |
| 21) Naphthalene-d8 | 10.593 | 136 | 2018197 | 20.000 | ng | 0.00 |
| 39) Acenaphthene-d10 | 14.457 | 164 | 1318739 | 20.000 | ng | 0.00 |
| 64) Phenanthrene-d10 | 17.269 | 188 | 2796133 | 20.000 | ng | 0.02 |
| 76) Chrysene-d12 | 21.739 | 240 | 2515502 | 20.000 | ng | 0.04 |
| 86) Perylene-d12 | 25.186 | 264 | 2791666 | 20.000 | ng | 0.08 |
| System Monitoring Compounds | | | | | | |
| 5) 2-Fluorophenol | 5.411 | 112 | 4528303 | 136.932 | ng | 0.02 |
| 7) Phenol-d6 | 6.975 | 99 | 6165876 | 127.814 | ng | 0.01 |
| 23) Nitrobenzene-d5 | 8.958 | 82 | 3635318 | 92.973 | ng | 0.00 |
| 42) 2,4,6-Tribromophenol | 15.975 | 330 | 2257963 | 153.843 | ng | 0.00 |
| 45) 2-Fluorobiphenyl | 13.063 | 172 | 6931075 | 80.212 | ng | 0.00 |
| 79) Terphenyl-d14 | 19.992 | 244 | 11484368 | 88.694 | ng | 0.02 |
| Target Compounds | | | | | | |
| | | | | Qvalue | | |
| 2) 1,4-Dioxane | 3.323 | 88 | 451297 | 28.781 | ng | 99 |
| 3) Pyridine | 3.717 | 79 | 1261547m | 28.820 | ng | |
| 4) n-Nitrosodimethylamine | 3.623 | 42 | 558739 | 42.467 | ng | 93 |
| 6) Aniline | 7.134 | 93 | 1298454 | 26.924 | ng | 96 |
| 8) 2-Chlorophenol | 7.375 | 128 | 1596706 | 52.247 | ng | 94 |
| 9) Benzaldehyde | 6.946 | 77 | 660523m | 21.385 | ng | |
| 10) Phenol | 7.005 | 94 | 2374061 | 47.524 | ng | 96 |
| 11) bis(2-Chloroethyl)ether | 7.228 | 93 | 1760124 | 41.640 | ng | 96 |
| 12) 1,3-Dichlorobenzene | 7.699 | 146 | 1573718 | 43.352 | ng | 97 |
| 13) 1,4-Dichlorobenzene | 7.840 | 146 | 1603764 | 43.753 | ng | 98 |
| 14) 1,2-Dichlorobenzene | 8.158 | 146 | 1559102 | 44.137 | ng | 100 |
| 15) Benzyl Alcohol | 8.040 | 79 | 1626253 | 46.983 | ng | 96 |
| 16) 2,2'-oxybis(1-Chloropr... | 8.322 | 45 | 1647523 | 42.324 | ng | 97 |
| 17) 2-Methylphenol | 8.240 | 107 | 1536652 | 48.658 | ng | 97 |
| 18) Hexachloroethane | 8.881 | 117 | 659730 | 44.883 | ng | 99 |
| 19) n-Nitroso-di-n-propyla... | 8.605 | 70 | 1249300 | 37.490 | ng | 95 |
| 20) 3+4-Methylphenols | 8.569 | 107 | 2134400 | 50.123 | ng | 98 |
| 22) Acetophenone | 8.622 | 105 | 2456033 | 39.612 | ng | # 98 |
| 24) Nitrobenzene | 8.999 | 77 | 1897954 | 44.692 | ng | 92 |
| 25) Isophorone | 9.522 | 82 | 3709853 | 40.161 | ng | 97 |
| 26) 2-Nitrophenol | 9.705 | 139 | 792325 | 63.370 | ng | 93 |
| 27) 2,4-Dimethylphenol | 9.763 | 122 | 1213284 | 53.222 | ng | 97 |
| 28) bis(2-Chloroethoxy)met... | 10.005 | 93 | 2367345 | 41.984 | ng | 99 |
| 29) 2,4-Dichlorophenol | 10.240 | 162 | 1480451 | 54.964 | ng | 97 |
| 30) 1,2,4-Trichlorobenzene | 10.458 | 180 | 1492819 | 43.230 | ng | 98 |
| 31) Naphthalene | 10.646 | 128 | 4603512 | 43.664 | ng | 100 |
| 32) Benzoic acid | 9.922 | 122 | 1217860 | 69.039 | ng | 93 |
| 33) 4-Chloroaniline | 10.746 | 127 | 1119273 | 28.262 | ng | 96 |
| 34) Hexachlorobutadiene | 10.940 | 225 | 903824 | 41.569 | ng | 98 |
| 35) Caprolactam | 11.540 | 113 | 566372m | 50.518 | ng | |
| 36) 4-Chloro-3-methylphenol | 11.881 | 107 | 1952285 | 52.252 | ng | 97 |
| 37) 2-Methylnaphthalene | 12.257 | 142 | 3183184 | 44.238 | ng | 99 |
| 38) 1-Methylnaphthalene | 12.481 | 142 | 2982545 | 42.027 | ng | 99 |
| 40) 1,2,4,5-Tetrachloroben... | 12.634 | 216 | 1566485 | 40.541 | ng | 99 |
| 41) Hexachlorocyclopentadiene | 12.616 | 237 | 1393597 | 113.798 | ng | 98 |
| 43) 2,4,6-Trichlorophenol | 12.869 | 196 | 1231858 | 57.393 | ng | 100 |

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP082024\
 Data File : BP021569.D
 Acq On : 20 Aug 2024 12:13
 Operator : RC/JU
 Sample : PB162822BSD
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 PB162822BSD

Quant Time: Aug 20 12:41:50 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 14 00:00:30 2024
 Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/21/2024
 Supervised By :mohammad ahmed 08/22/2024

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|-------------------------------|--------|------|----------|---------|-------|----------|
| 44) 2,4,5-Trichlorophenol | 12.952 | 196 | 1315468 | 55.163 | ng | 96 |
| 46) 1,1'-Biphenyl | 13.275 | 154 | 3776973 | 40.481 | ng | 98 |
| 47) 2-Chloronaphthalene | 13.316 | 162 | 3185536 | 43.854 | ng | 99 |
| 48) 2-Nitroaniline | 13.516 | 65 | 1126118 | 54.530 | ng | # 86 |
| 49) Acenaphthylene | 14.175 | 152 | 5275845 | 49.407 | ng | 100 |
| 50) Dimethylphthalate | 13.910 | 163 | 4233153 | 47.933 | ng | 100 |
| 51) 2,6-Dinitrotoluene | 14.016 | 165 | 940351 | 52.381 | ng | 91 |
| 52) Acenaphthene | 14.522 | 154 | 3601366m | 51.401 | ng | |
| 53) 3-Nitroaniline | 14.346 | 138 | 775492 | 44.582 | ng | # 90 |
| 54) 2,4-Dinitrophenol | 14.569 | 184 | 965462 | 121.666 | ng | # 85 |
| 55) Dibenzofuran | 14.863 | 168 | 4988509 | 46.284 | ng | 99 |
| 56) 4-Nitrophenol | 14.669 | 139 | 1714088 | 114.676 | ng | 87 |
| 57) 2,4-Dinitrotoluene | 14.816 | 165 | 1335111 | 57.774 | ng | 89 |
| 58) Fluorene | 15.528 | 166 | 3864218 | 43.641 | ng | 100 |
| 59) 2,3,4,6-Tetrachlorophenol | 15.087 | 232 | 1185204 | 58.555 | ng | 99 |
| 60) Diethylphthalate | 15.304 | 149 | 4259031 | 47.002 | ng | 99 |
| 61) 4-Chlorophenyl-phenyle... | 15.516 | 204 | 1966279 | 42.961 | ng | 98 |
| 62) 4-Nitroaniline | 15.534 | 138 | 951854 | 52.570 | ng | 91 |
| 63) Azobenzene | 15.822 | 77 | 4620647 | 41.483 | ng | 95 |
| 65) 4,6-Dinitro-2-methylph... | 15.598 | 198 | 702004 | 70.723 | ng | 96 |
| 66) n-Nitrosodiphenylamine | 15.740 | 169 | 3529833 | 47.107 | ng | 99 |
| 67) 4-Bromophenyl-phenylether | 16.434 | 248 | 1341601 | 45.335 | ng | 99 |
| 68) Hexachlorobenzene | 16.557 | 284 | 1574085 | 45.010 | ng | 99 |
| 69) Atrazine | 16.716 | 200 | 1144071 | 44.470 | ng | 99 |
| 70) Pentachlorophenol | 16.910 | 266 | 2107986 | 110.436 | ng | 99 |
| 71) Phenanthrene | 17.310 | 178 | 6401785 | 47.607 | ng | 100 |
| 72) Anthracene | 17.398 | 178 | 6424495 | 48.715 | ng | 99 |
| 73) Carbazole | 17.675 | 167 | 6219869 | 48.939 | ng | 99 |
| 74) Di-n-butylphthalate | 18.292 | 149 | 7501228 | 48.058 | ng | 100 |
| 75) Fluoranthene | 19.398 | 202 | 7658708 | 46.418 | ng | 99 |
| 77) Benzidine | 19.586 | 184 | 1654718 | 28.295 | ng | 100 |
| 78) Pyrene | 19.781 | 202 | 7888730 | 47.980 | ng | 100 |
| 80) Butylbenzylphthalate | 20.722 | 149 | 3397196 | 53.957 | ng | 95 |
| 81) Benzo(a)anthracene | 21.716 | 228 | 7693671 | 48.721 | ng | 99 |
| 82) 3,3'-Dichlorobenzidine | 21.633 | 252 | 2315165 | 45.726 | ng | 99 |
| 83) Chrysene | 21.792 | 228 | 7276214 | 48.803 | ng | 99 |
| 84) Bis(2-ethylhexyl)phtha... | 21.675 | 149 | 4995239 | 53.521 | ng | 100 |
| 85) Di-n-octyl phthalate | 22.992 | 149 | 8711748 | 52.619 | ng | 97 |
| 87) Indeno(1,2,3-cd)pyrene | 29.104 | 276 | 9758827 | 53.928 | ng | 97 |
| 88) Benzo(b)fluoranthene | 24.098 | 252 | 7945368 | 49.556 | ng | 100 |
| 89) Benzo(k)fluoranthene | 24.180 | 252 | 7973591m | 50.980 | ng | |
| 90) Benzo(a)pyrene | 25.033 | 252 | 7515255 | 54.014 | ng | 99 |
| 91) Dibenzo(a,h)anthracene | 29.192 | 278 | 8019357 | 53.335 | ng | 99 |
| 92) Benzo(g,h,i)perylene | 30.327 | 276 | 7676781 | 50.941 | ng | 98 |

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP082024\
Data File : BP021569.D
Acq On : 20 Aug 2024 12:13
Operator : RC/JU
Sample : PB162822BSD
Misc :
ALS Vial : 5 Sample Multiplier: 1


Quant Time: Aug 20 12:41:50 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 14 00:00:30 2024
Response via : Initial Calibration

Instrument :
BNA_P
ClientSampleId :
PB162822BSD

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/21/2024
Supervised By :mohammad ahmed 08/22/2024







#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.805 min Scan# 8
Delta R.T. -0.006 min
Lab File: BP021569.D
Acq: 20 Aug 2024 12:13

Instrument : BNA_P
ClientSampleId : PB162822BSD


Manual Integrations APPROVED




Reviewed By :Yogesh Patel 08/21/2024
Supervised By :mohammad ahmed 08/22/2024



#2
1,4-Dioxane
Concen: 28.781 ng
RT: 3.323 min Scan# 57
Delta R.T. 0.006 min
Lab File: BP021569.D
Acq: 20 Aug 2024 12:13

Tgt Ion: 88 Resp: 451297
Ion Ratio Lower Upper
88 100
58 61.9 48.7 73.1
43 19.9 15.4 23.2









#3
Pyridine
Concen: 28.820 ng m
RT: 3.717 min Scan# 1
Delta R.T. 0.012 min
Lab File: BP021569.D
Acq: 20 Aug 2024 12:13

Instrument : BNA_P
ClientSampleId : PB162822BSD


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/21/2024
Supervised By :mohammad ahmed 08/22/2024



#4
n-Nitrosodimethylamine
Concen: 42.467 ng
RT: 3.623 min Scan# 108
Delta R.T. 0.006 min
Lab File: BP021569.D
Acq: 20 Aug 2024 12:13

Tgt Ion: 42 Resp: 558739
Ion Ratio Lower Upper
42 100
74 182.4 154.3 231.5
44 6.7 6.2 9.4







#5
2-Fluorophenol
Concen: 136.932 ng
RT: 5.411 min Scan# 4
Instrument : BNA_P
Delta R.T. 0.018 min
Lab File: BP021569.D
Acq: 20 Aug 2024 12:13
ClientSampleId : PB162822BSD


Manual Integrations
APPROVED




Reviewed By :Yogesh Patel 08/21/2024
Supervised By :mohammad ahmed 08/22/2024



#6
Aniline
Concen: 26.924 ng
RT: 7.134 min Scan# 705
Delta R.T. 0.000 min
Lab File: BP021569.D
Acq: 20 Aug 2024 12:13

Tgt Ion: 93 Resp: 1298454
Ion Ratio Lower Upper
93 100
66 34.9 30.2 45.2
65 17.2 14.7 22.1





#7

Phenol-d6

Concen: 127.814 ng

RT: 6.975 min Scan# 6

Delta R.T. 0.012 min

Lab File: BP021569.D

Acq: 20 Aug 2024 12:13

Instrument :

BNA_P





ClientSampleId :

PB162822BSD

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/21/2024

Supervised By :mohammad ahmed 08/22/2024



#8

2-Chlorophenol

Concen: 52.247 ng

RT: 7.375 min Scan# 746

Delta R.T. 0.000 min

Lab File: BP021569.D

Acq: 20 Aug 2024 12:13


Tgt Ion:128 Resp: 1596706




Ion Ratio Lower Upper

128 100

130 31.9 11.6 51.6

64 49.1 35.3 75.3





#9

Benzaldehyde

Concen: 21.385 ng m

RT: 6.946 min Scan# 6

Delta R.T. -0.000 min

Lab File: BP021569.D

Acq: 20 Aug 2024 12:13

Instrument :

BNA_P




ClientSampleId :

PB162822BSD

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/21/2024

Supervised By :mohammad ahmed 08/22/2024



#10

Phenol

Concen: 47.524 ng


RT: 7.005 min Scan# 683

Delta R.T. 0.018 min


Lab File: BP021569.D




Acq: 20 Aug 2024 12:13

| Tgt Ion: | Ion Ratio | Lower | Upper |
|----------|-----------|-------|-------|
| 94 | 100 | | |
| 65 | 24.7 | 6.5 | 46.5 |
| 66 | 32.4 | 14.7 | 54.7 |



Abundance Scan 683 (7.005 min): BP021569.D\data.ms (-62)









#11
bis(2-Chloroethyl)ether
Concen: 41.640 ng
RT: 7.228 min Scan# 7
Delta R.T. 0.000 min
Lab File: BP021569.D
Acq: 20 Aug 2024 12:13

Instrument : BNA_P
ClientSampleId : PB162822BSD


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/21/2024
Supervised By :mohammad ahmed 08/22/2024



#12
1,3-Dichlorobenzene
Concen: 43.352 ng
RT: 7.699 min Scan# 801
Delta R.T. 0.000 min
Lab File: BP021569.D
Acq: 20 Aug 2024 12:13

Tgt Ion:146 Resp: 1573718
Ion Ratio Lower Upper
146 100
148 64.2 51.8 77.6
75 35.1 31.3 46.9






#13
1,4-Dichlorobenzene
Concen: 43.753 ng
RT: 7.840 min Scan# 81
Delta R.T. 0.000 min
Lab File: BP021569.D
Acq: 20 Aug 2024 12:13

Instrument : BNA_P
ClientSampleId : PB162822BSD


Manual Integrations
APPROVED




Reviewed By :Yogesh Patel 08/21/2024
Supervised By :mohammad ahmed 08/22/2024



#14
1,2-Dichlorobenzene
Concen: 44.137 ng
RT: 8.158 min Scan# 879
Delta R.T. 0.000 min
Lab File: BP021569.D
Acq: 20 Aug 2024 12:13

Tgt Ion:146 Resp: 1559102
Ion Ratio Lower Upper
146 100
148 64.1 51.5 77.3
111 51.0 41.1 61.7





#15

Benzyl Alcohol

Concen: 46.983 ng

RT: 8.040 min Scan# 8

Delta R.T. 0.006 min

Lab File: BP021569.D

Acq: 20 Aug 2024 12:13

Instrument :

BNA_P





ClientSampleId :

PB162822BSD

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/21/2024

Supervised By :mohammad ahmed 08/22/2024



#16

2,2'-oxybis(1-Chloropropane)

Concen: 42.324 ng

RT: 8.322 min Scan# 907

Delta R.T. -0.012 min

Lab File: BP021569.D

Acq: 20 Aug 2024 12:13


Tgt Ion: 45 Resp: 1647523




Ion Ratio Lower Upper

45 100

77 20.8 1.9 41.9

79 15.0 0.0 36.6





#17

2-Methylphenol

Concen: 48.658 ng

RT: 8.240 min Scan# 8

Delta R.T. 0.006 min

Lab File: BP021569.D

Acq: 20 Aug 2024 12:13

Instrument :

BNA_P




ClientSampleId :

PB162822BSD

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 08/21/2024

Supervised By :mohammad ahmed 08/22/2024



#18

Hexachloroethane

Concen: 44.883 ng

RT: 8.881 min Scan# 1002

Delta R.T. -0.012 min

Lab File: BP021569.D

Acq: 20 Aug 2024 12:13


Tgt Ion:117 Resp: 659730




Ion Ratio Lower Upper

117 100

119 94.0 73.9 110.9

201 92.6 74.9 112.3





#19

n-Nitroso-di-n-propylamine
Concen: 37.490 ng
RT: 8.605 min Scan# 9

Instrument :
BNA_P
ClientSampleId :
PB162822BSD





Tgt Ion: 70 Resp: 1249300

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 70 | 100 | | |
| 42 | 44.0 | 33.1 | 49.7 |
| 101 | 10.9 | 7.4 | 11.2 |
| 130 | 19.4 | 13.3 | 19.9 |

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 08/21/2024
Supervised By :mohammad ahmed 08/22/2024

Abundance




#20


3+4-Methylphenols
Concen: 50.123 ng
RT: 8.569 min Scan# 949
Delta R.T. 0.012 min
Lab File: BP021569.D
Acq: 20 Aug 2024 12:13

Tgt Ion:107 Resp: 2134400

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 107 | 100 | | |
| 108 | 88.5 | 67.6 | 107.6 |
| 77 | 32.3 | 15.0 | 55.0 |
| 79 | 28.6 | 8.9 | 48.9 |

Abundance






#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.593 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP021569.D
Acq: 20 Aug 2024 12:13

Instrument : BNA_P
ClientSampleId : PB162822BSD



Manual Integrations APPROVED


Reviewed By :Yogesh Patel 08/21/2024
Supervised By :mohammad ahmed 08/22/2024



#22
Acetophenone
Concen: 39.612 ng
RT: 8.622 min Scan# 958
Delta R.T. 0.006 min
Lab File: BP021569.D
Acq: 20 Aug 2024 12:13


Tgt Ion:105 Resp: 2456033
Ion Ratio Lower Upper
105 100
71 3.5 4.6 7.0#
51 19.2 16.3 24.5
120 20.6 16.2 24.2





#23
Nitrobenzene-d5
Concen: 92.973 ng
RT: 8.958 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP021569.D
Acq: 20 Aug 2024 12:13




Instrument : BNA_P
ClientSampleId : PB162822BSD




Tgt Ion: 82 Resp: 363531
Ion Ratio Lower Upper
82 100
128 38.3 26.5 39.7
54 38.6 30.3 45.5

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 08/21/2024
Supervised By :mohammad ahmed 08/22/2024




#24
Nitrobenzene
Concen: 44.692 ng
RT: 8.999 min Scan# 1022
Delta R.T. 0.006 min
Lab File: BP021569.D
Acq: 20 Aug 2024 12:13




Tgt Ion: 77 Resp: 1897954
Ion Ratio Lower Upper
77 100
123 43.6 29.5 44.3
65 12.4 9.9 14.9






Abundance Scan 1111 (9.522 min): BP021569.D\data.ms



Abundance Scan 1111 (9.522 min): BP021569.D\data.ms (-1)




#25
Isophorone
Concen: 40.161 ng
RT: 9.522 min Scan# 1111
Delta R.T. 0.000 min
Lab File: BP021569.D
Acq: 20 Aug 2024 12:13


Instrument : BNA_P
ClientSampleId : PB162822BSD

Manual Integrations APPROVED


Reviewed By :Yogesh Patel 08/21/2024
Supervised By :mohammad ahmed 08/22/2024



Abundance Scan 1142 (9.705 min): BP021569.D\data.ms










Abundance Scan 1142 (9.705 min): BP021569.D\data.ms (-1)



#26
2-Nitrophenol
Concen: 63.370 ng
RT: 9.705 min Scan# 1142
Delta R.T. 0.006 min
Lab File: BP021569.D
Acq: 20 Aug 2024 12:13

Tgt Ion:139 Resp: 792325
Ion Ratio Lower Upper
139 100
109 34.1 28.9 43.3
65 54.1 49.0 73.6





#27

2,4-Dimethylphenol

Concen: 53.222 ng

RT: 9.763 min Scan# 1

Delta R.T. 0.006 min

Lab File: BP021569.D

Acq: 20 Aug 2024 12:13

Instrument :

BNA_P


ClientSampleId :

PB162822BSD

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/21/2024

Supervised By :mohammad ahmed 08/22/2024



#28

bis(2-Chloroethoxy)methane

Concen: 41.984 ng

RT: 10.005 min Scan# 1193

Delta R.T. 0.000 min

Lab File: BP021569.D

Acq: 20 Aug 2024 12:13


Tgt Ion: 93 Resp: 2367345




Ion Ratio Lower Upper

93 100

95 32.4 26.2 39.4

123 10.1 7.6 11.4





#29

2,4-Dichlorophenol

Concen: 54.964 ng

RT: 10.240 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP021569.D

Acq: 20 Aug 2024 12:13




Instrument : BNA_P

ClientSampleId : PB162822BSD

**Manual Integrations
APPROVED**



Reviewed By :Yogesh Patel 08/21/2024


Supervised By :mohammad ahmed 08/22/2024



#30
1,2,4-Trichlorobenzene
Concen: 43.230 ng
RT: 10.458 min Scan# 1270
Delta R.T. 0.000 min
Lab File: BP021569.D
Acq: 20 Aug 2024 12:13

Tgt Ion:180 Resp: 1492819
Ion Ratio Lower Upper
180 100
182 96.1 78.3 117.5
145 30.8 24.0 36.0





#31

Naphthalene

Concen: 43.664 ng

RT: 10.646 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP021569.D

Acq: 20 Aug 2024 12:13

Instrument :

BNA_P



ClientSampleId :

PB162822BSD

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/21/2024

Supervised By :mohammad ahmed 08/22/2024



#32

Benzoic acid

Concen: 69.039 ng

RT: 9.922 min Scan# 1179

Delta R.T. 0.047 min

Lab File: BP021569.D

Acq: 20 Aug 2024 12:13


Tgt Ion:122 Resp: 1217860

Ion Ratio Lower Upper

122 100

105 140.9 124.6 164.6

77 103.1 96.2 136.2



#33

4-Chloroaniline

Concen: 28.262 ng

RT: 10.746 min Scan# 1

Delta R.T. 0.006 min

Lab File: BP021569.D

Acq: 20 Aug 2024 12:13

Instrument :

BNA_P

ClientSampleId :

PB162822BSD

Tgt Ion:127 Resp: 111927

Ion Ratio Lower Upper

127 100

129 31.9 24.5 36.7

65 32.1 28.9 43.3

92 22.8 19.3 28.9


Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel 08/21/2024

Supervised By :mohammad ahmed 08/22/2024

Abundance

10.746



#34

Hexachlorobutadiene

Concen: 41.569 ng

RT: 10.940 min Scan# 1352

Delta R.T. 0.000 min

Lab File: BP021569.D

Acq: 20 Aug 2024 12:13

Tgt Ion:225 Resp: 903824

Ion Ratio Lower Upper




225 100

223 61.6 49.6 74.4

227 64.8 50.1 75.1

Abundance

10.940



#35

Caprolactam

Concen: 50.518 ng m

RT: 11.540 min Scan# 14

Delta R.T. 0.041 min

Lab File: BP021569.D




Acq: 20 Aug 2024 12:13

Instrument :

BNA_P

ClientSampleId :

PB162822BSD



Tgt Ion:113 Resp: 56637:

Ion Ratio Lower Upper

113 100





55 135.5 130.6 170.6

56 102.6 94.6 134.6

Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel 08/21/2024

Supervised By :mohammad ahmed 08/22/2024

#36
4-Chloro-3-methylphenol
Concen: 52.252 ng
RT: 11.881 min Scan# 1512
Delta R.T. 0.018 min
Lab File: BP021569.D
Acq: 20 Aug 2024 12:13


Tgt Ion:107 Resp: 1952285

Ion Ratio Lower Upper

107 100

144 21.5 16.6 24.8

142 66.8 51.0 76.4



#37

2-Methylnaphthalene

Concen: 44.238 ng

RT: 12.257 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP021569.D


Acq: 20 Aug 2024 12:13

Instrument :


BNA_P

ClientSampleId :


PB162822BSD



Abundance Scan 1576 (12.257 min): BP021569.D\data.ms (-)



Abundance Scan 1576 (12.257 min): BP021569.D\data.ms (-)



Tgt Ion:142 Resp: 3183184

Ion Ratio Lower Upper

142 100



141 88.7 71.7 107.5

115 39.1 31.4 47.2


Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel 08/21/2024


Supervised By :mohammad ahmed 08/22/2024



Abundance Scan 1614 (12.481 min): BP021569.D\data.ms (-)



Abundance Scan 1614 (12.481 min): BP021569.D\data.ms (-)



#38

1-Methylnaphthalene

Concen: 42.027 ng

RT: 12.481 min Scan# 1614

Delta R.T. 0.000 min

Lab File: BP021569.D

Acq: 20 Aug 2024 12:13


Tgt Ion:142 Resp: 2982545

Ion Ratio Lower Upper

142 100

141 92.6 73.5 110.3

116 4.3 3.4 5.2



#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 14.457 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP021569.D

Acq: 20 Aug 2024 12:13

Instrument :

BNA_P

ClientSampleId :

PB162822BSD

Tgt Ion:164 Resp: 1318739

Ion Ratio Lower Upper

164 100

162 100.3 81.5 122.3

160 44.4 35.2 52.8

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 08/21/2024

Supervised By :mohammad ahmed 08/22/2024

Abundance

800000

600000

400000

200000

0

14.457

Time-->

14.40 14.50

#40

1,2,4,5-Tetrachlorobenzene

Concen: 40.541 ng

RT: 12.634 min Scan# 1640

Delta R.T. 0.006 min

Lab File: BP021569.D

Acq: 20 Aug 2024 12:13

Tgt Ion:216 Resp: 1566485

Ion Ratio Lower Upper

216 100

214 78.6 62.5 93.7

179 20.6 16.4 24.6

108 22.0 16.7 25.1

Abundance

1000000

800000

600000

400000







200000


0

12.634

Time-->

12.50 12.60 12.70







#41
Hexachlorocyclopentadiene
Concen: 113.798 ng
RT: 12.616 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP021569.D
Acq: 20 Aug 2024 12:13

Instrument : BNA_P
ClientSampleId : PB162822BSD


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 08/21/2024
Supervised By :mohammad ahmed 08/22/2024



#42
2,4,6-Tribromophenol
Concen: 153.843 ng
RT: 15.975 min Scan# 2208
Delta R.T. 0.006 min
Lab File: BP021569.D
Acq: 20 Aug 2024 12:13

Tgt Ion:330 Resp: 2257963
Ion Ratio Lower Upper
330 100
332 96.9 77.4 116.0
141 29.7 24.5 36.7





#43

2,4,6-Trichlorophenol

Concen: 57.393 ng

RT: 12.869 min Scan# 1680

Delta R.T. 0.000 min

Lab File: BP021569.D

Acq: 20 Aug 2024 12:13



Instrument : BNA_P

ClientSampleId : PB162822BSD

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/21/2024

Supervised By :mohammad ahmed 08/22/2024



#44

2,4,5-Trichlorophenol

Concen: 55.163 ng

RT: 12.952 min Scan# 1694

Delta R.T. 0.018 min

Lab File: BP021569.D

Acq: 20 Aug 2024 12:13

Tgt Ion:196 Resp: 1315468


Ion Ratio Lower Upper




196 100

198 95.6 78.9 118.3

97 55.1 49.4 74.0

132 26.6 21.1 31.7





#45

2-Fluorobiphenyl

Concen: 80.212 ng

RT: 13.063 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP021569.D

Acq: 20 Aug 2024 12:13




Instrument : BNA_P

ClientSampleId : PB162822BSD

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/21/2024

Supervised By :mohammad ahmed 08/22/2024



#46

1,1'-Biphenyl

Concen: 40.481 ng

RT: 13.275 min Scan# 1749

Delta R.T. -0.006 min

Lab File: BP021569.D

Acq: 20 Aug 2024 12:13

Tgt Ion:154 Resp: 3776973



Ion Ratio Lower Upper




154 100

153 40.1 20.7 60.7

76 15.0 0.0 37.7

1 2 3 4 5 6 7 8 9 10 11 12 13 14 15 16 17





#47

2-Chloronaphthalene

Concen: 43.854 ng

RT: 13.316 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP021569.D

Acq: 20 Aug 2024 12:13





Instrument : BNA_P

ClientSampleId : PB162822BSD

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/21/2024

Supervised By :mohammad ahmed 08/22/2024



#48

2-Nitroaniline

Concen: 54.530 ng

RT: 13.516 min Scan# 1790

Delta R.T. 0.012 min

Lab File: BP021569.D

Acq: 20 Aug 2024 12:13


Tgt Ion: 65 Resp: 1126118


Ion Ratio Lower Upper

65 100

92 78.6 58.1 87.1

138 100.2 65.5 98.3#





#49

Acenaphthylene

Concen: 49.407 ng

RT: 14.175 min Scan# 1

Delta R.T. 0.006 min

Lab File: BP021569.D

Acq: 20 Aug 2024 12:13

Instrument :

BNA_P



ClientSampleId :

PB162822BSD

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/21/2024

Supervised By :mohammad ahmed 08/22/2024



#50

Dimethylphthalate

Concen: 47.933 ng

RT: 13.910 min Scan# 1857

Delta R.T. 0.000 min

Lab File: BP021569.D

Acq: 20 Aug 2024 12:13


Tgt Ion:163 Resp: 4233153


Ion Ratio Lower Upper

163 100

194 4.1 3.3 4.9

164 10.2 8.3 12.5







#51
2,6-Dinitrotoluene
Concen: 52.381 ng
RT: 14.016 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP021569.D
Acq: 20 Aug 2024 12:13

Instrument : BNA_P
ClientSampleId : PB162822BSD


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 08/21/2024
Supervised By :mohammad ahmed 08/22/2024



#52
Acenaphthene
Concen: 51.401 ng
RT: 14.522 min Scan# 1961
Delta R.T. 0.000 min
Lab File: BP021569.D
Acq: 20 Aug 2024 12:13

Tgt Ion:154 Resp: 3601366
Ion Ratio Lower Upper
154 100
153 113.4 91.1 136.7
152 53.4 43.0 64.4







#53
3-Nitroaniline
Concen: 44.582 ng
RT: 14.346 min Scan# 1
Delta R.T. -0.005 min
Lab File: BP021569.D
Acq: 20 Aug 2024 12:13

Instrument : BNA_P
ClientSampleId : PB162822BSD


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/21/2024
Supervised By :mohammad ahmed 08/22/2024



#54
2,4-Dinitrophenol
Concen: 121.666 ng
RT: 14.569 min Scan# 1969
Delta R.T. 0.018 min
Lab File: BP021569.D
Acq: 20 Aug 2024 12:13

Tgt Ion:184 Resp: 965462
Ion Ratio Lower Upper
184 100
63 61.1 64.9 97.3#
154 60.2 52.1 78.1







#55
Dibenzofuran
Concen: 46.284 ng
RT: 14.863 min Scan# 2
Delta R.T. 0.006 min
Lab File: BP021569.D
Acq: 20 Aug 2024 12:13

Instrument : BNA_P
ClientSampleId : PB162822BSD


1 Manual Integrations
2 APPROVED

3 Reviewed By :Yogesh Patel 08/21/2024
4 Supervised By :mohammad ahmed 08/22/2024



#56
4-Nitrophenol
Concen: 114.676 ng
RT: 14.669 min Scan# 1986
Delta R.T. 0.018 min
Lab File: BP021569.D
Acq: 20 Aug 2024 12:13

Tgt Ion:139 Resp: 1714088
Ion Ratio Lower Upper
139 100
109 83.3 70.2 110.2
65 103.2 101.8 141.8




#57
 2,4-Dinitrotoluene
 Concen: 57.774 ng
 RT: 14.816 min Scan# 2130
 Delta R.T. 0.006 min
 Lab File: BP021569.D
 Acq: 20 Aug 2024 12:13

Instrument : BNA_P
 ClientSampleId : PB162822BSD

Tgt Ion:165 Resp: 133511:
 Ion Ratio Lower Upper
 165 100
 63 45.7 42.5 63.7
 89 84.4 77.1 115.7
 182 2.8 2.0 3.0




Manual Integrations APPROVED

Reviewed By :Yogesh Patel 08/21/2024
 Supervised By :mohammad ahmed 08/22/2024



#58
 Fluorene
 Concen: 43.641 ng
 RT: 15.528 min Scan# 2132
 Delta R.T. 0.012 min
 Lab File: BP021569.D
 Acq: 20 Aug 2024 12:13

Tgt Ion:166 Resp: 3864218
 Ion Ratio Lower Upper
 166 100
 165 97.7 78.1 117.1
 167 13.6 11.0 16.4





#59
 2,3,4,6-Tetrachlorophenol
 Concen: 58.555 ng
 RT: 15.087 min Scan# 2
 Delta R.T. 0.000 min
 Lab File: BP021569.D
 Acq: 20 Aug 2024 12:13

Instrument : BNA_P
 ClientSampleId : PB162822BSD

| Tgt | Ion | Ratio | Lower | Upper |
|-----|------|-------|-------|-------|
| 232 | 100 | | | |
| 131 | 41.0 | 33.1 | 49.7 | |
| 130 | 2.2 | 1.7 | 2.5 | |
| 166 | 29.6 | 23.0 | 34.4 | |


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 08/21/2024
 Supervised By :mohammad ahmed 08/22/2024





#60
 Diethylphthalate
 Concen: 47.002 ng
 RT: 15.304 min Scan# 2094
 Delta R.T. 0.000 min
 Lab File: BP021569.D
 Acq: 20 Aug 2024 12:13

| Tgt | Ion | Ratio | Lower | Upper |
|-----|------|-------|-------|-------|
| 149 | 100 | | | |
| 177 | 23.2 | 17.8 | 26.8 | |
| 150 | 12.4 | 10.1 | 15.1 | |




Abundance






Abundance Scan 2130 (15.516 min): BP021569.D\data.ms



Abundance Scan 2130 (15.516 min): BP021569.D\data.ms (-)




#61

4-Chlorophenyl-phenylether
Concen: 42.961 ng
RT: 15.516 min Scan# 2130
Delta R.T. -0.006 min
Lab File: BP021569.D
Acq: 20 Aug 2024 12:13


Instrument : BNA_P
ClientSampleId : PB162822BSD

Manual Integrations APPROVED


Reviewed By :Yogesh Patel 08/21/2024
Supervised By :mohammad ahmed 08/22/2024



Abundance Scan 2133 (15.534 min): BP021569.D\data.ms




Abundance Scan 2133 (15.534 min): BP021569.D\data.ms (-)



#62

4-Nitroaniline
Concen: 52.570 ng
RT: 15.534 min Scan# 2133
Delta R.T. 0.006 min
Lab File: BP021569.D
Acq: 20 Aug 2024 12:13

Tgt Ion:138 Resp: 951854
Ion Ratio Lower Upper
138 100
92 58.7 46.4 86.4
108 84.3 73.2 113.2





#63
Azobenzene
Concen: 41.483 ng
RT: 15.822 min Scan# 2182
Delta R.T. 0.000 min
Lab File: BP021569.D
Acq: 20 Aug 2024 12:13

Instrument : BNA_P
ClientSampleId : PB162822BSD


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/21/2024
Supervised By :mohammad ahmed 08/22/2024



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.269 min Scan# 2428
Delta R.T. 0.018 min
Lab File: BP021569.D
Acq: 20 Aug 2024 12:13

Tgt Ion:188 Resp: 2796133
Ion Ratio Lower Upper
188 100
94 11.6 10.2 15.4
80 12.7 11.6 17.4






#65
4,6-Dinitro-2-methylphenol
Concen: 70.723 ng
RT: 15.598 min Scan# 2142
Delta R.T. 0.011 min
Lab File: BP021569.D
Acq: 20 Aug 2024 12:13

Instrument : BNA_P
ClientSampleId : PB162822BSD


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 08/21/2024
Supervised By :mohammad ahmed 08/22/2024



#66
n-Nitrosodiphenylamine
Concen: 47.107 ng
RT: 15.740 min Scan# 2168
Delta R.T. 0.000 min
Lab File: BP021569.D
Acq: 20 Aug 2024 12:13

Tgt Ion:169 Resp: 3529833
Ion Ratio Lower Upper
169 100
168 66.3 53.4 80.2
167 36.5 28.7 43.1




#67
 4-Bromophenyl-phenylether
 Concen: 45.335 ng
 RT: 16.434 min Scan# 21
 Delta R.T. -0.006 min
 Lab File: BP021569.D
 Acq: 20 Aug 2024 12:13

Instrument : BNA_P
 ClientSampleId : PB162822BSD

Tgt Ion:248 Resp: 134160
 Ion Ratio Lower Upper
 248 100
 250 99.2 78.6 117.8
 141 66.3 53.9 80.9




Manual Integrations APPROVED




Reviewed By :Yogesh Patel 08/21/2024
 Supervised By :mohammad ahmed 08/22/2024



#68
 Hexachlorobenzene
 Concen: 45.010 ng
 RT: 16.557 min Scan# 2307
 Delta R.T. 0.006 min
 Lab File: BP021569.D
 Acq: 20 Aug 2024 12:13

Tgt Ion:284 Resp: 1574085
 Ion Ratio Lower Upper
 284 100
 142 28.3 23.4 35.2
 249 29.7 24.1 36.1





#69

Atrazine

Concen: 44.470 ng

RT: 16.716 min Scan# 2

Delta R.T. 0.000 min

Lab File: BP021569.D

Acq: 20 Aug 2024 12:13

Instrument :

BNA_P





ClientSampleId :

PB162822BSD

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/21/2024

Supervised By :mohammad ahmed 08/22/2024



#70

Pentachlorophenol

Concen: 110.436 ng

RT: 16.910 min Scan# 2367

Delta R.T. 0.018 min

Lab File: BP021569.D

Acq: 20 Aug 2024 12:13


Tgt Ion:266 Resp: 2107986

Ion Ratio Lower Upper

266 100

268 63.5 51.8 77.6

264 62.8 50.5 75.7



#71

Phenanthrene

Concen: 47.607 ng

RT: 17.310 min Scan# 2

Delta R.T. 0.012 min

Lab File: BP021569.D


Acq: 20 Aug 2024 12:13

Instrument :





BNA_P

ClientSampleId :

PB162822BSD



Abundance Scan 2435 (17.310 min): BP021569.D\data.ms



Tgt Ion:178 Resp: 640178

Ion Ratio Lower Upper

178 100

176 19.1 15.4 23.2


179 15.6 12.4 18.6

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 08/21/2024

Supervised By :mohammad ahmed 08/22/2024



#72

Anthracene

Concen: 48.715 ng

RT: 17.398 min Scan# 2450

Delta R.T. 0.000 min

Lab File: BP021569.D

Acq: 20 Aug 2024 12:13


Tgt Ion:178 Resp: 6424495


Ion Ratio Lower Upper

178 100

176 18.9 14.9 22.3

179 15.8 12.5 18.7







#73
Carbazole
Concen: 48.939 ng
RT: 17.675 min Scan# 2219869
Delta R.T. 0.006 min
Lab File: BP021569.D
Acq: 20 Aug 2024 12:13

Instrument : BNA_P
ClientSampleId : PB162822BSD


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 08/21/2024
Supervised By :mohammad ahmed 08/22/2024



#74
Di-n-butylphthalate
Concen: 48.058 ng
RT: 18.292 min Scan# 2602
Delta R.T. 0.018 min
Lab File: BP021569.D
Acq: 20 Aug 2024 12:13

Tgt Ion:149 Resp: 7501228
Ion Ratio Lower Upper
149 100
150 9.3 7.5 11.3
104 5.9 4.9 7.3



#75

Fluoranthene

Concen: 46.418 ng

RT: 19.398 min Scan# 2

Delta R.T. 0.018 min

Lab File: BP021569.D




Acq: 20 Aug 2024 12:13

Instrument :

BNA_P

ClientSampleId :

PB162822BSD



Tgt Ion:202 Resp: 765870

Ion Ratio Lower Upper

202 100





101 14.4 0.0 34.0

203 18.1 0.0 37.5

Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel 08/21/2024

Supervised By :mohammad ahmed 08/22/2024



#76

Chrysene-d₁₂

Concen: 20.000 ng

RT: 21.739 min Scan# 3188

Delta R.T. 0.041 min

Lab File: BP021569.D

Acq: 20 Aug 2024 12:13


Tgt Ion:240 Resp: 2515502




Ion Ratio Lower Upper

240 100

120 11.8 9.9 14.9

236 25.9 20.6 30.8









#77
Benzidine
Concen: 28.295 ng
RT: 19.586 min Scan# 21
Delta R.T. 0.011 min
Lab File: BP021569.D
Acq: 20 Aug 2024 12:13

Instrument : BNA_P
ClientSampleId : PB162822BSD


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 08/21/2024
Supervised By :mohammad ahmed 08/22/2024



#78
Pyrene
Concen: 47.980 ng
RT: 19.781 min Scan# 2855
Delta R.T. 0.029 min
Lab File: BP021569.D
Acq: 20 Aug 2024 12:13

Tgt Ion:202 Resp: 7888730
Ion Ratio Lower Upper
202 100
200 20.9 16.7 25.1
203 17.8 14.1 21.1



#79

Terphenyl-d14

Concen: 88.694 ng

RT: 19.992 min Scan# 2

Delta R.T. 0.017 min

Lab File: BP021569.D

Acq: 20 Aug 2024 12:13

Instrument :

BNA_P

ClientSampleId :

PB162822BSD

Tgt Ion:244 Resp:11484363

Ion Ratio Lower Upper

244 100

212 7.2 5.6 8.4

122 11.1 9.4 14.2

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 08/21/2024

Supervised By :mohammad ahmed 08/22/2024

Abundance

Scan 2891 (19.992 min): BP021569.D\data.ms (-)

244.3

m/z-->

Raw 50

50 100 150 200 250 300 350 400 450 500

Abundance

Scan 2891 (19.992 min): BP021569.D\data.ms (-)

244.3

m/z-->

Sub 50

50 100 150 200 250 300 350 400 450 500

Abundance

Scan 2891 (19.992 min): BP021569.D\data.ms (-)

244.3

m/z-->

Ref 50

50 100 150 200 250 300 350 400 450 500

Abundance

Scan 3013 (20.710 min): BP021485.D\data.ms (-)

149.0

m/z-->

Ref 50

50 100 150 200 250 300 350 400 450 500

Abundance

Scan 3015 (20.722 min): BP021569.D\data.ms (-)

149.1

m/z-->

Raw 50

50 100 150 200 250 300 350 400 450 500

Abundance

Scan 3015 (20.722 min): BP021569.D\data.ms (-)

149.0

m/z-->

Sub 50

50 100 150 200 250 300 350 400 450 500

Abundance

Scan 3015 (20.722 min): BP021569.D\data.ms (-)

149.0

m/z-->

#80

Butylbenzylphthalate

Concen: 53.957 ng

RT: 20.722 min Scan# 3015

Delta R.T. 0.012 min

Lab File: BP021569.D

Acq: 20 Aug 2024 12:13

Tgt Ion:149 Resp: 3397196

Ion Ratio Lower Upper

149 100

91 86.3 73.3 109.9

206 21.7 17.1 25.7

Abundance

Scan 3013 (20.710 min): BP021485.D\data.ms (-)

149.0

m/z-->

Raw 50

50 100 150 200 250 300 350 400 450 500

Abundance

Scan 3015 (20.722 min): BP021569.D\data.ms (-)

149.1

m/z-->

Sub 50

50 100 150 200 250 300 350 400 450 500

Abundance

Scan 3015 (20.722 min): BP021569.D\data.ms (-)

149.0

m/z-->

#80

Butylbenzylphthalate

Concen: 53.957 ng

RT: 20.722 min Scan# 3015

Delta R.T. 0.012 min

Lab File: BP021569.D

Acq: 20 Aug 2024 12:13

Tgt Ion:149 Resp: 3397196

Ion Ratio Lower Upper

149 100

91 86.3 73.3 109.9

206 21.7 17.1 25.7

Abundance

Scan 3013 (20.710 min): BP021485.D\data.ms (-)

149.0

m/z-->

Raw 50

50 100 150 200 250 300 350 400 450 500

Abundance

Scan 3015 (20.722 min): BP021569.D\data.ms (-)

149.1

m/z-->

Sub 50

50 100 150 200 250 300 350 400 450 500

Abundance

Scan 3015 (20.722 min): BP021569.D\data.ms (-)

149.0

m/z-->

#80

Butylbenzylphthalate

Concen: 53.957 ng

RT: 20.722 min Scan# 3015

Delta R.T. 0.012 min

Lab File: BP021569.D

Acq: 20 Aug 2024 12:13

Tgt Ion:149 Resp: 3397196

Ion Ratio Lower Upper

149 100

91 86.3 73.3 109.9

206 21.7 17.1 25.7

Abundance

Scan 3013 (20.710 min): BP021485.D\data.ms (-)

149.0

m/z-->

Raw 50

50 100 150 200 250 300 350 400 450 500

Abundance

Scan 3015 (20.722 min): BP021569.D\data.ms (-)

149.1

m/z-->

Sub 50

50 100 150 200 250 300 350 400 450 500

Abundance

Scan 3015 (20.722 min): BP021569.D\data.ms (-)

149.0

m/z-->

#80

Butylbenzylphthalate

Concen: 53.957 ng

RT: 20.722 min Scan# 3015

Delta R.T. 0.012 min

Lab File: BP021569.D

Acq: 20 Aug 2024 12:13

Tgt Ion:149 Resp: 3397196

Ion Ratio Lower Upper

149 100

91 86.3 73.3 109.9

206 21.7 17.1 25.7

Abundance

Scan 3013 (20.710 min): BP021485.D\data.ms (-)

149.0

m/z-->

Raw 50

50 100 150 200 250 300 350 400 450 500

Abundance

Scan 3015 (20.722 min): BP021569.D\data.ms (-)

149.1

m/z-->

Sub 50

50 100 150 200 250 300 350 400 450 500

Abundance

Scan 3015 (20.722 min): BP021569.D\data.ms (-)

149.0

m/z-->

#80

Butylbenzylphthalate

Concen: 53.957 ng

RT: 20.722 min Scan# 3015

Delta R.T. 0.012 min

Lab File: BP021569.D

Acq: 20 Aug 2024 12:13

Tgt Ion:149 Resp: 3397196

Ion Ratio Lower Upper

149 100

91 86.3 73.3 109.9

206 21.7 17.1 25.7

Abundance

Scan 3013 (20.710 min): BP021485.D\data.ms (-)

149.0

m/z-->

Raw 50

50 100 150 200 250 300 350 400 450 500

Abundance

Scan 3015 (20.722 min): BP021569.D\data.ms (-)

149.1

m/z-->

Sub 50

50 100 150 200 250 300 350 400 450 500

Abundance

Scan 3015 (20.722 min): BP021569.D\data.ms (-)

149.0

m/z-->

#80

Butylbenzylphthalate

Concen: 53.957 ng

RT: 20.722 min Scan# 3015

Delta R.T. 0.012 min

Lab File: BP021569.D

Acq: 20 Aug 2024 12:13

Tgt Ion:149 Resp: 3397196

Ion Ratio Lower Upper

149 100

91 86.3 73.3 109.9

206 21.7 17.1 25.7

Abundance

Scan 3013 (20.710 min): BP021485.D\data.ms (-)

149.0

m/z-->

Raw 50

50 100 150 200 250 300 350 400 450 500

Abundance

Scan 3015 (20.722 min): BP021569.D\data.ms (-)

149.1

m/z-->

Sub 50

50 100 150 200 250 300 350 400 450 500

Abundance

Scan 3015 (20.722 min): BP021569.D\data.ms (-)

149.0

m/z-->

#80

Butylbenzylphthalate

Concen: 53.957 ng

RT: 20.722 min Scan# 3015

Delta R.T. 0.012 min

Lab File: BP021569.D

Acq: 20 Aug 2024 12:13

Tgt Ion:149 Resp: 3397196

Ion Ratio Lower Upper

149 100

91 86.3 73.3 109.9

206 21.7 17.1 25.7

Abundance

Scan 3013 (20.710 min): BP021485.D\data.ms (-)

149.0

m/z-->

Raw 50

50 100 150 200 250 300 350 400 450 500

Abundance

Scan 3015 (20.722 min): BP021569.D\data.ms (-)

149.1

m/z-->

Sub 50

50 100 150 200 250 300 350 400 450 500

Abundance

Scan 3015 (20.722 min): BP021569.D\data.ms (-)

149.0

m/z-->

#80

Butylbenzylphthalate

Concen: 53.957 ng

RT: 20.722 min Scan# 3015

Delta R.T. 0.012 min

Lab File: BP021569.D

Acq: 20 Aug 2024 12:13

Tgt Ion:149 Resp: 3397196

Ion Ratio Lower Upper

149 100

91 86.3 73.3 109.9

206 21.7 17.1 25.7

Abundance

#81

Benzo(a)anthracene

Concen: 48.721 ng

RT: 21.716 min Scan# 3

Instrument :

Delta R.T. 0.035 min

BNA_P

Lab File: BP021569.D

ClientSampleId :

Acq: 20 Aug 2024 12:13

PB162822BSD

Tgt Ion:228 Resp: 769367

Ion Ratio Lower Upper

228 100


226 27.3 21.4 32.2

229 20.0 15.8 23.6

Manual Integrations**APPROVED**


Reviewed By :Yogesh Patel 08/21/2024

Supervised By :mohammad ahmed 08/22/2024






#82
3,3'-Dichlorobenzidine
Concen: 45.726 ng
RT: 21.633 min Scan# 3170
Delta R.T. 0.024 min
Lab File: BP021569.D
Acq: 20 Aug 2024 12:13

Tgt Ion:252 Resp: 2315165
Ion Ratio Lower Upper
252 100
254 63.9 51.7 77.5
126 11.7 8.7 13.1



BP021569.D 8270E-BP081324.M

P3657-SVOCMS Group6



#83

Chrysene

Concen: 48.803 ng

RT: 21.792 min Scan# 3190

Delta R.T. 0.041 min

Lab File: BP021569.D

Acq: 20 Aug 2024 12:13

Instrument :

BNA_P




ClientSampleId :

PB162822BSD

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 08/21/2024

Supervised By :mohammad ahmed 08/22/2024



#84

Bis(2-ethylhexyl)phthalate

Concen: 53.521 ng

RT: 21.675 min Scan# 3177

Delta R.T. 0.029 min

Lab File: BP021569.D

Acq: 20 Aug 2024 12:13


Tgt Ion:149 Resp: 4995239


Ion Ratio Lower Upper

149 100

167 27.6 22.3 33.5

279 4.8 3.8 5.6







#85
Di-n-octyl phthalate
Concen: 52.619 ng
RT: 22.992 min Scan# 3401
Delta R.T. 0.029 min
Lab File: BP021569.D
Acq: 20 Aug 2024 12:13

Instrument : BNA_P
ClientSampleId : PB162822BSD


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 08/21/2024
Supervised By :mohammad ahmed 08/22/2024



#86
Perylene-d12
Concen: 20.000 ng
RT: 25.186 min Scan# 3774
Delta R.T. 0.082 min
Lab File: BP021569.D
Acq: 20 Aug 2024 12:13

Tgt Ion:264 Resp: 2791666
Ion Ratio Lower Upper
264 100
260 22.9 18.6 28.0
265 21.8 17.1 25.7



#87

Indeno(1,2,3-cd)pyrene

Concen: 53.928 ng

RT: 29.104 min Scan# 4

Delta R.T. 0.112 min

Lab File: BP021569.D

Acq: 20 Aug 2024 12:13

Instrument :

BNA_P

ClientSampleId :

PB162822BSD

Tgt Ion:276 Resp: 975882

Ion Ratio Lower Upper

276 100

138 17.1 12.8 19.2

277 21.8 16.5 24.7

Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel 08/21/2024

Supervised By :mohammad ahmed 08/22/2024

Abundance

1500000

1000000

500000

0

Time-->

29.00

31.00

33.00

35.00

37.00

39.00

41.00

43.00

45.00

47.00

49.00

51.00

53.00

55.00

57.00

59.00

61.00

63.00

65.00

67.00

69.00

71.00

73.00

75.00

77.00

79.00

81.00

83.00

85.00

87.00

89.00

91.00

93.00

95.00

97.00

99.00

101.00

103.00

105.00

107.00

109.00

111.00

113.00

115.00

117.00

119.00

121.00

123.00

125.00

127.00

129.00

131.00

133.00

135.00

137.00

139.00

141.00

143.00

145.00

147.00

149.00

151.00

153.00

155.00

157.00

159.00

161.00

163.00

165.00

167.00

169.00

171.00

173.00

175.00

177.00

179.00

181.00

183.00

185.00

187.00

189.00

191.00

193.00

195.00

197.00

199.00

201.00

203.00

205.00

207.00

209.00

211.00

213.00

215.00

217.00

219.00

221.00

223.00

225.00

227.00

229.00

231.00

233.00

235.00

237.00

239.00

241.00

243.00

245.00

247.00

249.00

251.00

253.00

255.00

257.00

259.00

261.00

263.00

265.00

267.00

269.00

271.00

273.00

275.00

277.00

279.00

281.00

283.00

285.00

287.00

289.00

291.00

293.00

295.00

297.00

299.00

301.00

303.00

305.00

307.00

309.00

311.00

313.00

315.00

317.00

319.00

321.00

323.00

325.00

327.00

329.00

331.00

333.00

335.00

337.00

339.00

341.00

343.00

345.00

347.00

349.00

351.00

353.00

355.00

357.00

359.00

361.00

363.00

365.00

367.00

369.00

371.00

373.00

375.00

377.00

379.00

381.00

383.00

385.00

387.00

389.00

391.00

393.00

395.00

397.00

399.00

401.00

403.00

405.00

407.00

409.00

411.00

413.00

415.00

417.00

419.00

421.00

423.00

425.00

427.00

429.00

431.00

433.00

435.00

437.00

439.00

441.00

443.00

445.00

447.00

449.00

451.00

453.00

455.00

457.00

459.00

461.00

463.00

465.00

467.00

469.00

471.00

473.00

475.00

477.00

479.00

481.00

483.00

485.00

487.00

489.00

491.00

493.00

495.00

497.00

499.00

501.00

503.00

505.00

507.00

509.00

511.00

513.00

515.00

517.00

519.00

521.00

523.00

525.00

527.00

529.00

531.00

533.00

535.00

537.00

539.00

541.00

543.00

545.00

547.00

549.00

551.00

553.00

555.00

557.00

559.00

561.00

563.00

565.00

567.00

569.00

571.00

573.00

575.00

577.00

579.00

581.00

583.00

585.00

587.00

589.00

591.00

593.00

595.00

#89

Benzo(k)fluoranthene

Concen: 50.980 ng m

RT: 24.180 min Scan# 3

Delta R.T. 0.071 min

Lab File: BP021569.D




Acq: 20 Aug 2024 12:13

Instrument :

BNA_P

ClientSampleId :

PB162822BSD



Tgt Ion:252 Resp: 797359:

Ion Ratio Lower Upper

252 100





253 22.0 17.5 26.3

125 10.0 7.8 11.8

Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel 08/21/2024

Supervised By :mohammad ahmed 08/22/2024



#90

Benzo(a)pyrene

Concen: 54.014 ng

RT: 25.033 min Scan# 3748

Delta R.T. 0.082 min

Lab File: BP021569.D

Acq: 20 Aug 2024 12:13


Tgt Ion:252 Resp: 7515255




Ion Ratio Lower Upper

252 100

253 21.7 17.7 26.5

125 11.6 8.7 13.1









#91
Dibenzo(a,h)anthracene
Concen: 53.335 ng
RT: 29.192 min Scan# 4455
Delta R.T. 0.106 min
Lab File: BP021569.D
Acq: 20 Aug 2024 12:13

Instrument : BNA_P
ClientSampleId : PB162822BSD


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 08/21/2024
Supervised By :mohammad ahmed 08/22/2024



#92
Benzo(g,h,i)perylene
Concen: 50.941 ng
RT: 30.327 min Scan# 4648
Delta R.T. 0.153 min
Lab File: BP021569.D
Acq: 20 Aug 2024 12:13

Tgt Ion:276 Resp: 7676781
Ion Ratio Lower Upper
276 100
277 23.7 19.0 28.6
138 21.6 16.0 24.0



Manual Integration Report

| | | | |
|-----------|----------|------------|-------|
| Sequence: | BP081324 | Instrument | BNA_p |
|-----------|----------|------------|-------|

| Sample ID | File ID | Parameter | Review By | Review On | Supervised By | Supervised On | Reason |
|-------------|------------|------------------------------|-----------|----------------------|---------------|----------------------|-----------------------------|
| SSTDICC005 | BP021482.D | Indeno(1,2,3-cd)pyrene | yogesh | 8/14/2024 5:48:54 AM | mohammad | 8/14/2024 6:23:14 AM | Peak Integrated by Software |
| SSTDICC010 | BP021483.D | 2,2"-oxybis(1-Chloropropane) | yogesh | 8/14/2024 5:48:55 AM | mohammad | 8/14/2024 6:23:14 AM | Peak Integrated by Software |
| SSTDICC020 | BP021484.D | Indeno(1,2,3-cd)pyrene | yogesh | 8/14/2024 5:48:57 AM | mohammad | 8/14/2024 6:23:14 AM | Peak Integrated by Software |
| SSTDICCC040 | BP021485.D | Indeno(1,2,3-cd)pyrene | yogesh | 8/14/2024 5:48:59 AM | mohammad | 8/14/2024 6:23:14 AM | Peak Integrated by Software |
| SSTDICCC040 | BP021485.D | Pyridine | yogesh | 8/14/2024 5:48:59 AM | mohammad | 8/14/2024 6:23:14 AM | Peak Integrated by Software |
| SSTDICC050 | BP021486.D | Indeno(1,2,3-cd)pyrene | yogesh | 8/14/2024 5:49:00 AM | mohammad | 8/14/2024 6:23:14 AM | Peak Integrated by Software |
| SSTDICC050 | BP021486.D | Pyridine | yogesh | 8/14/2024 5:49:00 AM | mohammad | 8/14/2024 6:23:14 AM | Peak Integrated by Software |
| SSTDICC060 | BP021487.D | Indeno(1,2,3-cd)pyrene | yogesh | 8/14/2024 5:49:01 AM | mohammad | 8/14/2024 6:23:14 AM | Peak Integrated by Software |
| SSTDICC060 | BP021487.D | Pyridine | yogesh | 8/14/2024 5:49:01 AM | mohammad | 8/14/2024 6:23:14 AM | Peak Integrated by Software |
| SSTDICC080 | BP021488.D | Caprolactam | yogesh | 8/14/2024 5:49:03 AM | mohammad | 8/14/2024 6:23:14 AM | Peak Integrated by Software |
| SSTDICC080 | BP021488.D | Indeno(1,2,3-cd)pyrene | yogesh | 8/14/2024 5:49:03 AM | mohammad | 8/14/2024 6:23:14 AM | Peak Integrated by Software |
| SSTDICC080 | BP021488.D | Pyridine | yogesh | 8/14/2024 5:49:03 AM | mohammad | 8/14/2024 6:23:14 AM | Peak Integrated by Software |
| SSTDICV040 | BP021489.D | Indeno(1,2,3-cd)pyrene | yogesh | 8/14/2024 5:49:04 AM | mohammad | 8/14/2024 6:23:14 AM | Peak Integrated by Software |



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

| | | | |
|-----------|----------|------------|-------|
| Sequence: | BP081324 | Instrument | BNA_p |
|-----------|----------|------------|-------|

| Sample ID | File ID | Parameter | Review By | Review On | Supervised By | Supervised On | Reason |
|------------|------------|-----------|-----------|----------------------|---------------|----------------------|-----------------------------|
| SSTDICV040 | BP021489.D | Pyridine | yogesh | 8/14/2024 5:49:04 AM | mohammad | 8/14/2024 6:23:14 AM | Peak Integrated by Software |

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17

Manual Integration Report

| | | | |
|-----------|----------|------------|-------|
| Sequence: | BP082024 | Instrument | BNA_p |
|-----------|----------|------------|-------|

| Sample ID | File ID | Parameter | Review By | Review On | Supervised By | Supervised On | Reason |
|-------------|------------|------------------------|-----------|----------------------|---------------|----------------------|-----------------------------|
| SSTDCCC040 | BP021566.D | Benzaldehyde | yogesh | 8/21/2024 1:30:12 AM | mohammad | 8/22/2024 7:51:42 AM | Peak Integrated by Software |
| SSTDCCC040 | BP021566.D | Indeno(1,2,3-cd)pyrene | yogesh | 8/21/2024 1:30:12 AM | mohammad | 8/22/2024 7:51:42 AM | Peak Integrated by Software |
| PB162822BS | BP021568.D | Acenaphthene | yogesh | 8/21/2024 1:30:14 AM | mohammad | 8/22/2024 7:51:42 AM | Peak Integrated by Software |
| PB162822BS | BP021568.D | Benzaldehyde | yogesh | 8/21/2024 1:30:14 AM | mohammad | 8/22/2024 7:51:42 AM | Peak Integrated by Software |
| PB162822BS | BP021568.D | Caprolactam | yogesh | 8/21/2024 1:30:14 AM | mohammad | 8/22/2024 7:51:42 AM | Peak Integrated by Software |
| PB162822BS | BP021568.D | Indeno(1,2,3-cd)pyrene | yogesh | 8/21/2024 1:30:14 AM | mohammad | 8/22/2024 7:51:42 AM | Peak Integrated by Software |
| PB162822BS | BP021568.D | Pyridine | yogesh | 8/21/2024 1:30:14 AM | mohammad | 8/22/2024 7:51:42 AM | Peak Integrated by Software |
| PB162822BSD | BP021569.D | Acenaphthene | yogesh | 8/21/2024 1:30:15 AM | mohammad | 8/22/2024 7:51:42 AM | Peak Integrated by Software |
| PB162822BSD | BP021569.D | Benzaldehyde | yogesh | 8/21/2024 1:30:15 AM | mohammad | 8/22/2024 7:51:42 AM | Peak Integrated by Software |
| PB162822BSD | BP021569.D | Benzo(k)fluoranthene | yogesh | 8/21/2024 1:30:15 AM | mohammad | 8/22/2024 7:51:42 AM | Peak Integrated by Software |
| PB162822BSD | BP021569.D | Caprolactam | yogesh | 8/21/2024 1:30:15 AM | mohammad | 8/22/2024 7:51:42 AM | Peak Integrated by Software |
| PB162822BSD | BP021569.D | Pyridine | yogesh | 8/21/2024 1:30:15 AM | mohammad | 8/22/2024 7:51:42 AM | Peak Integrated by Software |
| SSTDCCC040 | BP021573.D | Benzaldehyde | yogesh | 8/21/2024 1:30:19 AM | mohammad | 8/22/2024 7:51:42 AM | Peak Integrated by Software |

 1
 2
 3
 4
 5
 6
 7
 8
 9
 10
 11
 12
 13
 14
 15
 16
 17

Manual Integration Report

| | | | |
|-----------|----------|------------|-------|
| Sequence: | BP082024 | Instrument | BNA_p |
|-----------|----------|------------|-------|

| Sample ID | File ID | Parameter | Review By | Review On | Supervised By | Supervised On | Reason |
|------------|------------|------------------------|-----------|-----------------------|---------------|----------------------|-----------------------------|
| SSTDCCC040 | BP021573.D | Indeno(1,2,3-cd)pyrene | yogesh | 8/21/2024 1:30:19 AM | mohammad | 8/22/2024 7:51:42 AM | Peak Integrated by Software |
| SSTDCCC040 | BP021575.D | Acenaphthene | yogesh | 8/21/2024 1:30:23 AM | mohammad | 8/22/2024 7:51:42 AM | Peak Integrated by Software |
| SSTDCCC040 | BP021575.D | Indeno(1,2,3-cd)pyrene | yogesh | 8/21/2024 1:30:23 AM | mohammad | 8/22/2024 7:51:42 AM | Peak Integrated by Software |
| SSTDCCC040 | BP021590.D | Acenaphthene | Jagrut | 8/21/2024 10:20:54 AM | mohammad | 8/22/2024 7:51:42 AM | Peak Integrated by Software |
| SSTDCCC040 | BP021590.D | Indeno(1,2,3-cd)pyrene | Jagrut | 8/21/2024 10:20:54 AM | mohammad | 8/22/2024 7:51:42 AM | Peak Integrated by Software |

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17

Instrument ID: **BNA_P**

Daily Analysis Runlog For Sequence/QCBatch ID # BP081324

| | | | |
|--|---|-------------------|----------------------|
| Review By | yogesh | Review On | 8/14/2024 5:49:15 AM |
| Supervise By | mohammad | Supervise On | 8/14/2024 6:23:14 AM |
| SubDirectory | BP081324 | HP Acquire Method | BNA_P |
| STD. NAME | STD REF.# | | |
| Tune/Reschk Initial Calibration Stds | SP6573 SP6595,SP6594,SP6593,SP6592,SP6591,SP6590,SP6589,SP6588 | | |
| CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard | SP6591 S12041,10ul/1000ul sample SP6559 | | |

| Sr# | SampleId | Data File Name | Date-Time | Operator | Status |
|-----|-------------|----------------|-------------------|----------|--------|
| 1 | DFTPP | BP021480.D | 13 Aug 2024 11:49 | RC/JU | Ok |
| 2 | SSTDICC2.5 | BP021481.D | 13 Aug 2024 12:32 | RC/JU | Ok |
| 3 | SSTDICC005 | BP021482.D | 13 Aug 2024 13:15 | RC/JU | Ok,M |
| 4 | SSTDICC010 | BP021483.D | 13 Aug 2024 13:59 | RC/JU | Ok,M |
| 5 | SSTDICC020 | BP021484.D | 13 Aug 2024 14:42 | RC/JU | Ok,M |
| 6 | SSTDICCC040 | BP021485.D | 13 Aug 2024 15:26 | RC/JU | Ok,M |
| 7 | SSTDICC050 | BP021486.D | 13 Aug 2024 16:09 | RC/JU | Ok,M |
| 8 | SSTDICC060 | BP021487.D | 13 Aug 2024 16:53 | RC/JU | Ok,M |
| 9 | SSTDICC080 | BP021488.D | 13 Aug 2024 17:36 | RC/JU | Ok,M |
| 10 | SSTDICCV040 | BP021489.D | 13 Aug 2024 19:03 | RC/JU | Ok,M |
| 11 | PB162612TB | BP021490.D | 13 Aug 2024 19:46 | RC/JU | Ok |

M : Manual Integration

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QCBatch ID # BP082024

| Review By | yogesh | Review On | 8/21/2024 1:30:32 AM |
|--|---|-------------------|----------------------|
| Supervise By | mohammad | Supervise On | 8/22/2024 7:51:42 AM |
| SubDirectory | BP082024 | HP Acquire Method | BNA_P |
| HP Processing Method | BP081324 | | |
| STD. NAME | STD REF.# | | |
| Tune/Reschk Initial Calibration Stds | SP6573 SP6595,SP6594,SP6593,SP6592,SP6591,SP6590,SP6589,SP6588 | | |
| CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard | SP6591 S12041,10ul/1000ul sample SP6559 | | |

| Sr# | SampleId | Data File Name | Date-Time | Operator | Status |
|-----|-------------|----------------|-------------------|----------|----------|
| 1 | DFTPP | BP021565.D | 20 Aug 2024 09:25 | RC/JU | Ok |
| 2 | SSTDCCC040 | BP021566.D | 20 Aug 2024 10:07 | RC/JU | Ok,M |
| 3 | PB162822BL | BP021567.D | 20 Aug 2024 10:49 | RC/JU | Ok |
| 4 | PB162822BS | BP021568.D | 20 Aug 2024 11:31 | RC/JU | Ok,M |
| 5 | PB162822BSD | BP021569.D | 20 Aug 2024 12:13 | RC/JU | Ok,M |
| 6 | P3650-01 | BP021570.D | 20 Aug 2024 12:59 | RC/JU | Ok |
| 7 | P3650-04 | BP021571.D | 20 Aug 2024 13:42 | RC/JU | Ok |
| 8 | P3519-03RE | BP021572.D | 20 Aug 2024 14:25 | RC/JU | Confirms |
| 9 | SSTDCCC040 | BP021573.D | 20 Aug 2024 15:14 | RC/JU | Ok,M |
| 10 | DFTPP | BP021574.D | 20 Aug 2024 15:57 | RC/JU | Ok |
| 11 | SSTDCCC040 | BP021575.D | 20 Aug 2024 16:40 | RC/JU | Ok,M |
| 12 | PB162631BL | BP021576.D | 20 Aug 2024 17:24 | RC/JU | Ok |
| 13 | P3643-10 | BP021577.D | 20 Aug 2024 18:07 | RC/JU | Ok |
| 14 | P3643-12 | BP021578.D | 20 Aug 2024 18:51 | RC/JU | Ok |
| 15 | P3643-14 | BP021579.D | 20 Aug 2024 19:34 | RC/JU | Ok |
| 16 | P3600-06 | BP021580.D | 20 Aug 2024 20:17 | RC/JU | Ok,M |
| 17 | P3600-06MS | BP021581.D | 20 Aug 2024 21:01 | RC/JU | Ok,M |
| 18 | P3600-06MSD | BP021582.D | 20 Aug 2024 21:44 | RC/JU | Ok,M |
| 19 | P3518-15RE | BP021583.D | 20 Aug 2024 22:27 | RC/JU | Confirms |
| 20 | P3518-15MS | BP021584.D | 20 Aug 2024 23:10 | RC/JU | Ok,M |
| 21 | P3518-15MSD | BP021585.D | 20 Aug 2024 23:53 | RC/JU | Ok,M |

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QCBatch ID # BP082024

| Review By | yogesh | Review On | 8/21/2024 1:30:32 AM |
|--|---|-------------------|----------------------|
| Supervise By | mohammad | Supervise On | 8/22/2024 7:51:42 AM |
| SubDirectory | BP082024 | HP Acquire Method | BNA_P |
| HP Processing Method | BP081324 | | |
| STD. NAME | STD REF.# | | |
| Tune/Reschk Initial Calibration Stds | SP6573 SP6595,SP6594,SP6593,SP6592,SP6591,SP6590,SP6589,SP6588 | | |
| CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard | SP6591 S12041,10ul/1000ul sample SP6559 | | |

| | | | | | |
|----|------------|------------|-------------------|-------|----------|
| 22 | P3518-23RE | BP021586.D | 21 Aug 2024 00:36 | RC/JU | Confirms |
| 23 | P3518-25RE | BP021587.D | 21 Aug 2024 01:19 | RC/JU | Confirms |
| 24 | P3518-29RE | BP021588.D | 21 Aug 2024 02:02 | RC/JU | Confirms |
| 25 | P3657-01 | BP021589.D | 21 Aug 2024 02:44 | RC/JU | Ok |
| 26 | SSTDCCC040 | BP021590.D | 21 Aug 2024 03:27 | RC/JU | Ok,M |

M : Manual Integration

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QCBatch ID # BP081324

| Review By | yogesh | Review On | 8/14/2024 5:49:15 AM | | |
|--------------------------|---|-------------------|----------------------|----------------------|----------|
| Supervise By | mohammad | Supervise On | 8/14/2024 6:23:14 AM | | |
| SubDirectory | BP081324 | HP Acquire Method | BNA_P | HP Processing Method | BP081324 |
| STD. NAME | STD REF.# | | | | |
| Tune/Reschk | SP6573 | | | | |
| Initial Calibration Stds | SP6595,SP6594,SP6593,SP6592,SP6591,SP6590,SP6589,SP6588 | | | | |
| CCC | SP6591 | | | | |
| Internal Standard/PEM | S12041,10ul/1000ul sample | | | | |
| ICV/I.BLK | SP6559 | | | | |
| Surrogate Standard | | | | | |
| MS/MSD Standard | | | | | |
| LCS Standard | | | | | |

| Sr# | SampleId | ClientID | Data File Name | Date-Time | Comment | Operator | Status |
|-----|-------------|-------------|----------------|-------------------|--|----------|--------|
| 1 | DFTPP | DFTPP | BP021480.D | 13 Aug 2024 11:49 | | RC/JU | Ok |
| 2 | SSTDICC2.5 | SSTDICC2.5 | BP021481.D | 13 Aug 2024 12:32 | | RC/JU | Ok |
| 3 | SSTDICC005 | SSTDICC005 | BP021482.D | 13 Aug 2024 13:15 | Compound #32,54,56,65,70,77,85 removed from 5ppm | RC/JU | Ok,M |
| 4 | SSTDICC010 | SSTDICC010 | BP021483.D | 13 Aug 2024 13:59 | | RC/JU | Ok,M |
| 5 | SSTDICC020 | SSTDICC020 | BP021484.D | 13 Aug 2024 14:42 | Compound #48,51,53,56,57,62,80,85 Kept on LR and Compound #26,32,54,65,77 Kept on QR | RC/JU | Ok,M |
| 6 | SSTDICCC040 | SSTDICCC040 | BP021485.D | 13 Aug 2024 15:26 | The Calibration is Good For 8270 DOD except com#32,77 and good for 8270E | RC/JU | Ok,M |
| 7 | SSTDICC050 | SSTDICC050 | BP021486.D | 13 Aug 2024 16:09 | | RC/JU | Ok,M |
| 8 | SSTDICC060 | SSTDICC060 | BP021487.D | 13 Aug 2024 16:53 | | RC/JU | Ok,M |
| 9 | SSTDICC080 | SSTDICC080 | BP021488.D | 13 Aug 2024 17:36 | Compound #9 removed from 80ppm | RC/JU | Ok,M |
| 10 | SSTDICCV040 | ICVBP081324 | BP021489.D | 13 Aug 2024 19:03 | Com#32 and 77 failed in ICV for 8270 DOD Criteria | RC/JU | Ok,M |
| 11 | PB162612TB | PB162612TB | BP021490.D | 13 Aug 2024 19:46 | Already analyzed with ok status | RC/JU | Ok |

M : Manual Integration

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QCBatch ID # BP082024

| Review By | yogesh | Review On | 8/21/2024 1:30:32 AM | | |
|--------------------------|---|-------------------|----------------------|----------------------|----------|
| Supervise By | mohammad | Supervise On | 8/22/2024 7:51:42 AM | | |
| SubDirectory | BP082024 | HP Acquire Method | BNA_P | HP Processing Method | BP081324 |
| STD. NAME | STD REF.# | | | | |
| Tune/Reschk | SP6573 | | | | |
| Initial Calibration Stds | SP6595,SP6594,SP6593,SP6592,SP6591,SP6590,SP6589,SP6588 | | | | |
| CCC | SP6591 | | | | |
| Internal Standard/PEM | S12041,10ul/1000ul sample | | | | |
| ICV/I.BLK | SP6559 | | | | |
| Surrogate Standard | | | | | |
| MS/MSD Standard | | | | | |
| LCS Standard | | | | | |

| Sr# | SampleId | ClientID | Data File Name | Date-Time | Comment | Operator | Status |
|-----|-------------|-----------------------|----------------|-------------------|----------------|----------|----------|
| 1 | DFTPP | DFTPP | BP021565.D | 20 Aug 2024 09:25 | | RC/JU | Ok |
| 2 | SSTDCCC040 | SSTDCCC040 | BP021566.D | 20 Aug 2024 10:07 | | RC/JU | Ok,M |
| 3 | PB162822BL | PB162822BL | BP021567.D | 20 Aug 2024 10:49 | | RC/JU | Ok |
| 4 | PB162822BS | PB162822BS | BP021568.D | 20 Aug 2024 11:31 | | RC/JU | Ok,M |
| 5 | PB162822BSD | PB162822BSD | BP021569.D | 20 Aug 2024 12:13 | | RC/JU | Ok,M |
| 6 | P3650-01 | RW7-SP100-20240814 | BP021570.D | 20 Aug 2024 12:59 | | RC/JU | Ok |
| 7 | P3650-04 | RW7-SP303-20240814 | BP021571.D | 20 Aug 2024 13:42 | | RC/JU | Ok |
| 8 | P3519-03RE | OU4-TS-17-080724RE | BP021572.D | 20 Aug 2024 14:25 | Surrogate Fail | RC/JU | Confirms |
| 9 | SSTDCCC040 | SSTDCCC040EC | BP021573.D | 20 Aug 2024 15:14 | | RC/JU | Ok,M |
| 10 | DFTPP | DFTPP | BP021574.D | 20 Aug 2024 15:57 | | RC/JU | Ok |
| 11 | SSTDCCC040 | SSTDCCC040 | BP021575.D | 20 Aug 2024 16:40 | | RC/JU | Ok,M |
| 12 | PB162631BL | PB162631BL | BP021576.D | 20 Aug 2024 17:24 | | RC/JU | Ok |
| 13 | P3643-10 | 924-K1-SD-0.5-1-08152 | BP021577.D | 20 Aug 2024 18:07 | | RC/JU | Ok |
| 14 | P3643-12 | 921-J-SD-0.5-1-081524 | BP021578.D | 20 Aug 2024 18:51 | | RC/JU | Ok |
| 15 | P3643-14 | 921-J-SD-0.5-1-081524 | BP021579.D | 20 Aug 2024 19:34 | | RC/JU | Ok |
| 16 | P3600-06 | VCPS-B3-081324-10-1 | BP021580.D | 20 Aug 2024 20:17 | | RC/JU | Ok,M |
| 17 | P3600-06MS | VCPS-B3-081324-10-1 | BP021581.D | 20 Aug 2024 21:01 | | RC/JU | Ok,M |
| 18 | P3600-06MSD | VCPS-B3-081324-10-1 | BP021582.D | 20 Aug 2024 21:44 | | RC/JU | Ok,M |

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QCBatch ID # BP082024

| Review By | yogesh | Review On | 8/21/2024 1:30:32 AM | | |
|--------------------------|---|-------------------|----------------------|----------------------|----------|
| Supervise By | mohammad | Supervise On | 8/22/2024 7:51:42 AM | | |
| SubDirectory | BP082024 | HP Acquire Method | BNA_P | HP Processing Method | BP081324 |
| STD. NAME | STD REF.# | | | | |
| Tune/Reschk | SP6573 | | | | |
| Initial Calibration Stds | SP6595,SP6594,SP6593,SP6592,SP6591,SP6590,SP6589,SP6588 | | | | |
| CCC | SP6591 | | | | |
| Internal Standard/PEM | S12041,10ul/1000ul sample | | | | |
| ICV/I.BLK | SP6559 | | | | |
| Surrogate Standard | | | | | |
| MS/MSD Standard | | | | | |
| LCS Standard | | | | | |

| | | | | | | | |
|----|-------------|--------------------|------------|-------------------|----------------|-------|----------|
| 19 | P3518-15RE | OU4-TS-08-080724RE | BP021583.D | 20 Aug 2024 22:27 | Surrogate Fail | RC/JU | Confirms |
| 20 | P3518-15MS | OU4-TS-08-080724MS | BP021584.D | 20 Aug 2024 23:10 | | RC/JU | Ok,M |
| 21 | P3518-15MSD | OU4-TS-08-080724MS | BP021585.D | 20 Aug 2024 23:53 | | RC/JU | Ok,M |
| 22 | P3518-23RE | OU4-TS-12-080724RE | BP021586.D | 21 Aug 2024 00:36 | Surrogate Fail | RC/JU | Confirms |
| 23 | P3518-25RE | OU4-TS-13-080724RE | BP021587.D | 21 Aug 2024 01:19 | Surrogate Fail | RC/JU | Confirms |
| 24 | P3518-29RE | OU4-TS-15-080724RE | BP021588.D | 21 Aug 2024 02:02 | Surrogate Fail | RC/JU | Confirms |
| 25 | P3657-01 | 917-J-WS-081624 | BP021589.D | 21 Aug 2024 02:44 | | RC/JU | Ok |
| 26 | SSTDCCC040 | SSTDCCC040EC | BP021590.D | 21 Aug 2024 03:27 | | RC/JU | Ok,M |

M : Manual Integration

| | | | |
|--------------------|--|-------------------------|------------|
| SOP ID: | M3510C,3580A-Extraction SVOC-20 | | |
| Clean Up SOP #: | N/A | Extraction Start Date : | 08/19/2024 |
| Matrix : | Water | Extraction Start Time : | 09:40 |
| Weigh By: | N/A | Extraction End Date : | 08/19/2024 |
| Balance check: | N/A | Extraction End Time : | 14:35 |
| Balance ID: | N/A | Concentration By: | RS |
| pH Strip Lot#: | E3574 | Hood ID: | 4,5,6,7 |
| Extraction Method: | <input checked="" type="checkbox"/> Separatory Funne <input type="checkbox"/> Continous Liquid/Liquid <input type="checkbox"/> Sonication <input type="checkbox"/> Waste Dilution <input type="checkbox"/> Soxhlet | | |

| Standard Name | MLS USED | Concentration ug/mL | STD REF. # FROM LOG |
|---------------|----------|---------------------|---------------------|
| Spike Sol 1 | 1.0ML | 50/100 PPM | SP6604 |
| Surrogate | 1.0ML | 100/150 PPM | SP6524 |
| N/A | N/A | N/A | N/A |
| N/A | N/A | N/A | N/A |
| N/A | N/A | N/A | N/A |

| Chemical Used | ML/SAMPLE USED | Lot Number |
|--------------------|----------------|------------|
| Methylene Chloride | N/A | E3787 |
| Baked Na2SO4 | N/A | EP2521 |
| H2SO4 1:1 | N/A | EP2524 |
| 10N NaOH | N/A | EP2523 |
| N/A | N/A | N/A |

Extraction Conformance/Non-Conformance Comments:

pH Adjusted to < 2 with 1:1 H2SO4 and > 11 with 10N NaOH, 1.5ML Vial Lot # 2210673.

KD Bath ID: WATER BATH-1 Envap ID: NE VAP-02
 KD Bath Temperature: 60 °C Envap Temperature: 40 °C

| Date / Time | Prepped Sample Relinquished By/Location | Received By/Location |
|-------------|---|----------------------|
| 08/19/24 | RP (Ept. 2nd) | PC/SVOC |
| 14:40 | Preparation Group | Analysis Group |

Analytical Method: M3510C,3580A-Extraction SVOC-20

Concentration Date: 08/19/2024

| Sample ID | Client Sample ID | Test | g / mL | PH | Surr/Spike By: | | Final Vol.(mL) | JarID | Comments | Prep Pos |
|--------------|--------------------|---------------|--------|----|----------------|------------|----------------|-------|----------|----------|
| | | | | | AddedBy | VerifiedBy | | | | |
| PB162822BL | SBLK822 | SVOCMS Group5 | 1000 | 6 | RUPESH | rajesh | 1 | | | SEP-10 |
| PB162822BS | SLCS822 | SVOCMS Group5 | 1000 | 6 | RUPESH | rajesh | 1 | | | 11 |
| PB162822BS D | SLCSD822 | SVOCMS Group5 | 1000 | 6 | RUPESH | rajesh | 1 | | | 12 |
| P3650-01 | RW7-SP100-20240814 | SVOCMS Group5 | 1000 | 6 | RUPESH | rajesh | 1 | B | | 13 |
| P3650-04 | RW7-SP303-20240814 | SVOCMS Group5 | 1000 | 6 | RUPESH | rajesh | 1 | B | | 14 |
| P3657-01 | 917-J-WS-081624 | SVOCMS Group6 | 960 | 6 | RUPESH | rajesh | 1 | E | | 15 |

* Extracts relinquished on the same date as received.



16287
102521
9:40 12/5/19

WORKLIST(Hardcopy Internal Chain)

WorkList Name : P3651

WorkList ID : 182767

Department : Extraction

Date : 08-19-2024 09:38:44

| Sample | Customer Sample | Matrix | Test | Preservative | Customer | Raw Sample Storage Location | Collect Date | Method |
|----------|--------------------|--------|---------------|--------------|----------|-----------------------------|--------------|--------|
| P3650-01 | RW7-SP100-20240814 | Water | SVOCMS Group5 | Cool 4 deg C | TETR06 | G11 | 08/14/2024 | 8270E |
| P3650-04 | RW7-SP303-20240814 | Water | SVOCMS Group5 | Cool 4 deg C | TETR06 | G11 | 08/14/2024 | 8270E |
| P3657-01 | 917-J-WS-081624 | Water | SVOCMS Group6 | Cool 4 deg C | JACO05 | G11 | 08/16/2024 | 8270E |

Date/Time 08/19/24 9:38
Raw Sample Received by: RS (Ext Lab)
Raw Sample Relinquished by: ZJW

Date/Time 08/19/24 10:00
Raw Sample Received by: ZJW
Raw Sample Relinquished by: RS (Ext Lab)

Prep Standard - Chemical Standard Summary

Order ID : P3657

Test : SVOCMS Group6

Prepbatch ID : PB162822,

Sequence ID/Qc Batch ID: BP082024,BP082024,BP082024,

Standard ID :

EP2521,EP2523,EP2524,SP6524,SP6558,SP6559,SP6573,SP6587,SP6588,SP6589,SP6590,SP6591,SP6592,SP6593,SP6594,SP6595,SP6604,

Chemical ID :

10ul/1000ul

sample,E3551,E3657,E3744,E3768,E3780,E3786,E3787,M5037,S10103,S10247,S10398,S10583,S10972,S10973,S10974,S10975,S10976,S10977,S10996,S10997,S10998,S10999,S11000,S11001,S11002,S11003,S11012,S11084,S11093,S11097,S11103,S11138,S11141,S11494,S11555,S11561,S11566,S11567,S11766,S11767,S11768,S11769,S11770,S12033,S12039,S12041,S12077,S12097,S12098,S12099,S12100,S12101,S12102,S12103,S12104,S12105,S12112,S12117,S12118,S12119,S12120,S12121,S12122,S12123,S12124,S12125,S12126,S12449,S12450,S12451,S12452,S12453,W3112,

Extractions STANDARD PREPARATION LOG

| <u>Recipe ID</u> | <u>NAME</u> | <u>NO.</u> | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u> | <u>PipetteID</u> | <u>Supervised By</u> |
|------------------|----------------------|------------------------|------------------|------------------------|--------------------|----------------------------------|------------------|--------------------------------|
| 3923 | Baked Sodium Sulfate | EP2521 | 08/10/2024 | 01/03/2025 | Rajesh Parikh | Extraction_SC ALE_2 (EX-SC-2) | None | RUPESHKUMAR SHAH 08/10/2024 |

FROM 4000.00000gram of E3551 = Final Quantity: 4000.000 gram

| <u>Recipe ID</u> | <u>NAME</u> | <u>NO.</u> | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u> | <u>PipetteID</u> | <u>Supervised By</u> |
|------------------|----------------------------|------------------------|------------------|------------------------|--------------------|----------------------------------|------------------|--------------------------------|
| 1874 | 10 N SODIUM HYDROXIDE SOLN | EP2523 | 08/14/2024 | 02/14/2025 | Rajesh Parikh | Extraction_SC ALE_2 (EX-SC-2) | None | RUPESHKUMAR SHAH 08/14/2024 |

FROM 1000.00000ml of W3112 + 400.00000gram of E3657 = Final Quantity: 1000.000 ml

Extractions STANDARD PREPARATION LOG

| <u>Recipe ID</u> | <u>NAME</u> | <u>NO.</u> | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u> | <u>PipetteID</u> | <u>Supervised By</u> |
|------------------|----------------|------------------------|------------------|------------------------|--------------------|----------------|------------------|--------------------------------|
| 314 | 1.1 H2SO4 SOLN | EP2524 | 08/14/2024 | 12/15/2024 | Rajesh Parikh | None | None | RUPESHKUMAR SHAH 08/14/2024 |

FROM 1000.00000ml of M5037 + 1000.00000ml of W3112 = Final Quantity: 2000.000 ml

| <u>Recipe ID</u> | <u>NAME</u> | <u>NO.</u> | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u> | <u>PipetteID</u> | <u>Supervised By</u> |
|------------------|--|------------------------|------------------|------------------------|--------------------|----------------|------------------|----------------------------|
| 19 | 8270/CLP Surrogate Solution, 100 PPM BN/150 PPM ACID | SP6524 | 05/31/2024 | 11/29/2024 | Jagrut Upadhyay | None | None | Yogesh Patel 06/13/2024 |

FROM 1930.00000ml of E3744 + 2.90000ml of S11003 + 3.00000ml of S10977 + 5.30000ml of S10996 + 5.30000ml of S10997 + 5.30000ml of S10998 + 5.30000ml of S10999 + 5.30000ml of S11000 + 5.30000ml of S11001 + 5.30000ml of S11002 + 5.40000ml of S10972 + 5.40000ml of S10973 + 5.40000ml of S10974 + 5.40000ml of S10975 + 5.40000ml of S10976 = Final Quantity: 2000.000 ml

SVOC STANDARD PREPARATION LOG

| <u>Recipe ID</u> | <u>NAME</u> | <u>NO.</u> | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u> | <u>PipetteID</u> | <u>Supervised By</u> |
|------------------|---|------------------------|------------------|------------------------|--------------------|----------------|------------------|------------------------------|
| 18 | Second Source Calibration Stock Standard, 100 PPM, (8270/625/CLP) | SP6558 | 07/09/2024 | 11/30/2024 | Jagrut Upadhyay | None | None | mohammad ahmed 07/11/2024 |
| <u>FROM</u> | 0.04000ml of S10977 + 0.08000ml of S11003 + 0.10000ml of S11766 + 0.20000ml of S11566 + 0.20000ml of S12097 + 0.20000ml of S12117 + 1.18000ml of E3768 = Final Quantity: 2.000 ml | | | | | | | |

| <u>Recipe ID</u> | <u>NAME</u> | <u>NO.</u> | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u> | <u>PipetteID</u> | <u>Supervised By</u> |
|------------------|---|------------------------|------------------|------------------------|--------------------|----------------|------------------|------------------------------|
| 416 | 40 ng BNA ICV, 40 PPM | SP6559 | 07/09/2024 | 11/30/2024 | Jagrut Upadhyay | None | None | mohammad ahmed 07/11/2024 |
| <u>FROM</u> | 0.01000ml of S12033 + 0.60000ml of E3768 + 0.40000ml of SP6558 = Final Quantity: 1.010 ml | | | | | | | |

SVOC STANDARD PREPARATION LOG

| <u>Recipe ID</u> | <u>NAME</u> | <u>NO.</u> | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u> | <u>PipetteID</u> | <u>Supervised By</u> |
|------------------|----------------------|------------------------|------------------|------------------------|--------------------|----------------|------------------|----------------------------|
| 3895 | 50 ug/ml DFTPP 8270E | SP6573 | 07/15/2024 | 01/08/2025 | Rahul Chavli | None | None | Yogesh Patel 07/17/2024 |

FROM 1.00000ml of S10247 + 19.00000ml of E3768 = Final Quantity: 20.000 ml

| <u>Recipe ID</u> | <u>NAME</u> | <u>NO.</u> | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u> | <u>PipetteID</u> | <u>Supervised By</u> |
|------------------|--------------------------------|------------------------|------------------|------------------------|--------------------|----------------|------------------|----------------------------|
| 3764 | 8270/625 Stock solution 100 ng | SP6587 | 08/08/2024 | 01/09/2025 | Jagrut Upadhyay | None | None | Yogesh Patel 08/17/2024 |

FROM 0.26700ml of S10103 + 0.40000ml of S11494 + 0.50000ml of S12112 + 1.00000ml of S11084 + 1.00000ml of S11093 + 1.00000ml of S11097 + 1.00000ml of S11103 + 1.00000ml of S12077 + 3.83300ml of E3786 = Final Quantity: 10.000 ml

SVOC STANDARD PREPARATION LOG

| <u>Recipe ID</u> | <u>NAME</u> | <u>NO.</u> | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u> | <u>PipetteID</u> | <u>Supervised By</u> |
|------------------|-----------------------|------------------------|------------------|------------------------|--------------------|----------------|------------------|----------------------------|
| 413 | 80 ng BNA ICC, 80 PPM | SP6588 | 08/08/2024 | 01/09/2025 | Jagrut Upadhyay | None | None | Yogesh Patel 08/17/2024 |

FROM 0.01000ml of S12039 + 0.20000ml of E3786 + 0.80000ml of SP6587 = Final Quantity: 1.010 ml

| <u>Recipe ID</u> | <u>NAME</u> | <u>NO.</u> | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u> | <u>PipetteID</u> | <u>Supervised By</u> |
|------------------|-----------------------|------------------------|------------------|------------------------|--------------------|----------------|------------------|----------------------------|
| 412 | 60 ng BNA ICC, 60 PPM | SP6589 | 08/08/2024 | 01/09/2025 | Jagrut Upadhyay | None | None | Yogesh Patel 08/17/2024 |

FROM 0.01000ml of S12039 + 0.40000ml of E3786 + 0.60000ml of SP6587 = Final Quantity: 1.010 ml

SVOC STANDARD PREPARATION LOG

| <u>Recipe ID</u> | <u>NAME</u> | <u>NO.</u> | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u> | <u>PipetteID</u> | <u>Supervised By</u> |
|------------------|-----------------------|------------------------|------------------|------------------------|--------------------|----------------|------------------|----------------------------|
| 411 | 50 ng BNA ICC, 50 PPM | SP6590 | 08/08/2024 | 01/09/2025 | Jagrut Upadhyay | None | None | Yogesh Patel 08/17/2024 |

FROM 0.01000ml of S12039 + 0.50000ml of E3786 + 0.50000ml of SP6587 = Final Quantity: 1.010 ml

| <u>Recipe ID</u> | <u>NAME</u> | <u>NO.</u> | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u> | <u>PipetteID</u> | <u>Supervised By</u> |
|------------------|-----------------------|------------------------|------------------|------------------------|--------------------|----------------|------------------|----------------------------|
| 410 | 40 ng BNA ICC, 40 PPM | SP6591 | 08/08/2024 | 01/09/2025 | Jagrut Upadhyay | None | None | Yogesh Patel 08/17/2024 |

FROM 0.01000ml of S12039 + 0.60000ml of E3786 + 0.40000ml of SP6587 = Final Quantity: 1.010 ml

SVOC STANDARD PREPARATION LOG

| <u>Recipe ID</u> | <u>NAME</u> | <u>NO.</u> | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u> | <u>PipetteID</u> | <u>Supervised By</u> |
|------------------|-----------------------|------------------------|------------------|------------------------|--------------------|----------------|------------------|----------------------------|
| 3678 | 20 ng BNA ICC, 20 PPM | SP6592 | 08/08/2024 | 01/09/2025 | Jagrut Upadhyay | None | None | Yogesh Patel 08/17/2024 |

FROM 0.01000ml of S12039 + 0.80000ml of E3786 + 0.20000ml of SP6587 = Final Quantity: 1.010 ml

| <u>Recipe ID</u> | <u>NAME</u> | <u>NO.</u> | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u> | <u>PipetteID</u> | <u>Supervised By</u> |
|------------------|-----------------------|------------------------|------------------|------------------------|--------------------|----------------|------------------|----------------------------|
| 408 | 10 ng BNA ICC, 10 PPM | SP6593 | 08/08/2024 | 01/09/2025 | Jagrut Upadhyay | None | None | Yogesh Patel 08/17/2024 |

FROM 0.01000ml of S12039 + 0.90000ml of E3786 + 0.10000ml of SP6587 = Final Quantity: 1.010 ml

SVOC STANDARD PREPARATION LOG

| <u>Recipe ID</u> | <u>NAME</u> | <u>NO.</u> | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u> | <u>PipetteID</u> | <u>Supervised By</u> |
|------------------|---------------------|------------------------|------------------|------------------------|--------------------|----------------|------------------|----------------------------|
| 407 | 5 ng BNA ICC, 5 PPM | SP6594 | 08/08/2024 | 01/09/2025 | Jagrut Upadhyay | None | None | Yogesh Patel 08/17/2024 |

FROM 0.01000ml of S12039 + 0.95000ml of E3786 + 0.05000ml of SP6587 = Final Quantity: 1.010 ml

| <u>Recipe ID</u> | <u>NAME</u> | <u>NO.</u> | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u> | <u>PipetteID</u> | <u>Supervised By</u> |
|------------------|-------------------------|------------------------|------------------|------------------------|--------------------|----------------|------------------|----------------------------|
| 175 | 2.5 ng BNA ICC, 2.5 PPM | SP6595 | 08/08/2024 | 01/09/2025 | Jagrut Upadhyay | None | None | Yogesh Patel 08/17/2024 |

FROM 0.01000ml of S12039 + 0.50000ml of E3786 + 0.50000ml of SP6594 = Final Quantity: 1.010 ml

SVOC STANDARD PREPARATION LOG

| <u>Recipe ID</u> | <u>NAME</u> | <u>NO.</u> | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u> | <u>PipetteID</u> | <u>Supervised By</u> |
|------------------|---|------------------------|------------------|------------------------|--------------------|----------------|------------------|----------------------------|
| 171 | 8270/625 Spike Solution, 50/100 PPM | SP6604 | 08/12/2024 | 10/26/2024 | Rahul Chavli | None | None | Yogesh Patel 08/21/2024 |
| FROM | 0.20000ml of S12126 + 0.40000ml of S10398 + 0.40000ml of S10583 + 0.40000ml of S11012 + 0.40000ml of S11138 + 0.40000ml of S11141 + 0.50000ml of S12105 + 0.50000ml of S12453 + 0.70000ml of S12097 + 0.80000ml of S11566 + 0.80000ml of S12118 + 1.20000ml of S11555 + 1.20000ml of S11561 + 1.20000ml of S11767 + 1.20000ml of S11769 + 1.20000ml of S12099 + 1.20000ml of S12101 + 1.20000ml of S12104 + 1.20000ml of S12119 + 1.20000ml of S12121 + 1.20000ml of S12449 + 1.20000ml of S12451 + 1.30000ml of S11567 + 1.30000ml of S11768 + 1.30000ml of S11770 + 1.30000ml of S12098 + 1.30000ml of S12100 + 1.30000ml of S12102 + 1.30000ml of S12103 + 1.30000ml of S12120 + 1.30000ml of S12122 + 1.30000ml of S12123 + 1.30000ml of S12124 + 1.30000ml of S12450 + 1.30000ml of S12452 + 1.40000ml of S12125 + 163.00000ml of E3780 = Final Quantity: 200.000 ml | | | | | | | |

CHEMICAL RECEIPT LOG BOOK

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|-----------------------------|---|--------------|------------------------|--------------------------------|------------------------------------|-----------------------|
| PCI Scientific Supply, Inc. | PC19631-100 / SODIUM SULFATE, ANHYDROUS, PEST GRADE, 1 | 313201 | 01/03/2025 | 01/03/2024 / Rajesh | 07/20/2023 / Rajesh | E3551 |
| PCI Scientific Supply, Inc. | PC19510-5 / Sodium Hydroxide Pellets 2.5 Kg, Pk of 4 | 23B1556310 | 12/31/2025 | 12/04/2023 / Rajesh | 12/01/2023 / Rajesh | E3657 |
| Seidler Chemical | BA-9254-03 / Acetone, Ultra Resi (cs/4x4L) | 23H14626005 | 11/29/2024 | 05/29/2024 / Rajesh | 05/23/2024 / Rajesh | E3744 |
| Seidler Chemical | BA-9644-A4 / Methylene Chloride,U-Resi, Cycle-Tainer (215L) | 24E2462004 | 01/08/2025 | 07/08/2024 / Rajesh | 06/21/2024 / Rajesh | E3768 |
| Seidler Chemical | BA-9254-03 / Acetone, Ultra Resi (cs/4x4L) | 23H1462005 | 02/10/2025 | 08/10/2024 / Rajesh | 07/24/2024 / Rajesh | E3780 |
| Seidler Chemical | BA-9644-A4 / Methylene Chloride,U-Resi, Cycle-Tainer (215L) | 24F1062004 | 02/01/2025 | 08/01/2024 / Rajesh | 07/16/2024 / Rajesh | E3786 |

CHEMICAL RECEIPT LOG BOOK

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|-------------------|---|------------|-----------------|-------------------------|-----------------------------|----------------|
| Seidler Chemical | BA-9644-A4 / Methylene Chloride,U-Resi, Cycle-Tainer (215L) | 24G0862022 | 02/13/2025 | 08/13/2024 / Rajesh | 08/07/2024 / Rajesh | E3787 |
| Seidler Chemical | BA-9673-33 / Sulfuric Acid, Instra-Analyzed (cs/6c2.5L) | 0000250349 | 12/15/2024 | 01/06/2022 / mohan | 09/18/2021 / mohan | M5037 |
| CPI International | Z-112090-04 / CLP Acid Surrogate Solution, 7500 mg/L, 1ml | 440246 | 02/08/2025 | 08/08/2024 / Jagrut | 12/09/2021 / Christian | S10103 |
| Restek | 31615 / SV Mixture, GC/MS Tuning Mixture, CH2Cl2, 1mL, | A0182667 | 01/15/2025 | 07/15/2024 / Rahul | 03/18/2022 / Christian | S10247 |
| Restek | 555871 / Custom Standard, 4-nitrophenol Std [CS 5238-4] | A0185300 | 10/26/2024 | 04/26/2024 / Rahul | 05/18/2022 / Christian | S10398 |
| Restek | 555868 / Custom Standard, Benzidine Std [CS 5328-1] | A0186373 | 02/12/2025 | 08/12/2024 / Rahul | 07/05/2022 / Christian | S10583 |

CHEMICAL RECEIPT LOG BOOK

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|----------|--|----------|-----------------|-------------------------|-----------------------------|----------------|
| Restek | 31087 / Acid Surrogate 10,000ug/ml,methanol,5ml/ ampul | A0188108 | 08/31/2030 | 05/31/2024 / Jagrut | 12/28/2022 / Christian | S10972 |
| Restek | 31087 / Acid Surrogate 10,000ug/ml,methanol,5ml/ ampul | A0188108 | 11/30/2024 | 05/31/2024 / Jagrut | 12/28/2022 / Christian | S10973 |
| Restek | 31087 / Acid Surrogate 10,000ug/ml,methanol,5ml/ ampul | A0188108 | 11/30/2024 | 05/31/2024 / Jagrut | 12/28/2022 / Christian | S10974 |
| Restek | 31087 / Acid Surrogate 10,000ug/ml,methanol,5ml/ ampul | A0188108 | 11/30/2024 | 05/31/2024 / Jagrut | 12/28/2022 / Christian | S10975 |
| Restek | 31087 / Acid Surrogate 10,000ug/ml,methanol,5ml/ ampul | A0188108 | 11/30/2024 | 05/31/2024 / Jagrut | 12/28/2022 / Christian | S10976 |
| Restek | 31087 / Acid Surrogate 10,000ug/ml,methanol,5ml/ ampul | A0188108 | 11/30/2024 | 05/31/2024 / Jagrut | 12/28/2022 / Christian | S10977 |

CHEMICAL RECEIPT LOG BOOK

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|----------|--|----------|-----------------|-------------------------|-----------------------------|----------------|
| Restek | 31086 / Base Neutral Surrogate 5000ug/ml,CH2Cl2,5ml | A0189418 | 08/31/2028 | 05/31/2024 / Jagrut | 12/28/2022 / Christian | S10996 |
| Restek | 31086 / Base Neutral Surrogate 5000ug/ml,CH2Cl2,5ml | A0189418 | 11/30/2024 | 05/31/2024 / Jagrut | 12/28/2022 / Christian | S10997 |
| Restek | 31086 / Base Neutral Surrogate 5000ug/ml,CH2Cl2,5ml | A0189418 | 11/30/2024 | 05/31/2024 / Jagrut | 12/28/2022 / Christian | S10998 |
| Restek | 31086 / Base Neutral Surrogate 5000ug/ml,CH2Cl2,5ml | A0189418 | 11/30/2024 | 05/31/2024 / Jagrut | 12/28/2022 / Christian | S10999 |
| Restek | 31086 / Base Neutral Surrogate 5000ug/ml,CH2Cl2,5ml | A0189418 | 11/30/2024 | 05/31/2024 / Jagrut | 12/28/2022 / Christian | S11000 |
| Restek | 31086 / Base Neutral Surrogate 5000ug/ml,CH2Cl2,5ml | A0189418 | 11/30/2024 | 05/31/2024 / Jagrut | 12/28/2022 / Christian | S11001 |

CHEMICAL RECEIPT LOG BOOK

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|-------------------|---|----------|-----------------|-------------------------|-----------------------------|----------------|
| Restek | 31086 / Base Neutral Surrogate 5000ug/ml,CH2Cl2,5ml | A0189418 | 11/30/2024 | 05/31/2024 / Jagrut | 12/28/2022 / Christian | S11002 |
| Restek | 31086 / Base Neutral Surrogate 5000ug/ml,CH2Cl2,5ml | A0189418 | 11/30/2024 | 05/31/2024 / Jagrut | 12/28/2022 / Christian | S11003 |
| Restek | 555872 / Custom Standard, pentachlorophenol Std [CS 5328-5] | A0193449 | 10/26/2024 | 04/26/2024 / Rahul | 01/13/2023 / Christian | S11012 |
| CPI International | Z-010074-07 / 3,3'-Dichlorobenzidine Solution, 1,000 mg/L, 1 ml, (Maximum Expiration: 180 days) | 406703 | 02/08/2025 | 08/08/2024 / Jagrut | 02/07/2023 / Christian | S11084 |
| CPI International | Z-110817-01 / Custom 8270 Mix, 4-55, 1000 mg/L, 1 ml, (Maximum Expiration: 90 Days) | 414125 | 02/08/2025 | 08/08/2024 / Jagrut | 02/07/2023 / Christian | S11093 |
| CPI International | z-110381-01 / 8270 Calibration Solution, 76-1, 500 & 1,000 mg/L, 1ml | 495831 | 02/08/2025 | 08/08/2024 / Jagrut | 02/07/2023 / Christian | S11097 |

CHEMICAL RECEIPT LOG BOOK

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|-------------------|--|----------|-----------------|-------------------------|-----------------------------|----------------|
| CPI International | Z-010442-07 / Benzaldehyde Solution, 1000 mg/L, 1.3 ml, (Maximum Expiration: 90 Days) | 495833 | 02/08/2025 | 08/08/2024 / Jagrut | 02/07/2023 / Christian | S11103 |
| Restek | 555870 / Custom Standard, 2,4-dinitrophenol Std [CS 5328-3] | A0194698 | 02/12/2025 | 08/12/2024 / Rahul | 02/20/2023 / Christian | S11138 |
| Restek | 555869 / Custom Standard, hexachlorocyclopentadiene Std [CS 5328-2] | A0194702 | 02/12/2025 | 08/12/2024 / Rahul | 02/20/2023 / Christian | S11141 |
| CPI International | Z-110094-02 / CLP Base/Neutral Surrogate Solution, 5000 mg/L, 1ml | 506889 | 02/08/2025 | 08/08/2024 / Jagrut | 08/11/2023 / Yogesh | S11494 |
| Restek | 555223 / Custom 8270 Plus Std #1 [2nd lot at \$100 per ampul if requested - contact ARM with Request] [CS 4978-1] | A0201940 | 02/12/2025 | 08/12/2024 / Rahul | 09/18/2023 / Kiran | S11555 |
| Restek | 555223 / Custom 8270 Plus Std #1 [2nd lot at \$100 per ampul if requested - contact ARM with Request] [CS 4978-1] | A0201940 | 02/12/2025 | 08/12/2024 / Rahul | 09/18/2023 / Kiran | S11561 |

CHEMICAL RECEIPT LOG BOOK

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|-------------|---|----------|-----------------|-------------------------|-----------------------------|----------------|
| Restek | 555223 / Custom 8270 Plus Std #1 [2nd lot at \$100 per ampul if requested - contact ARM with Request] | A0201940 | 12/05/2024 | 06/05/2024 / Rahul | 09/18/2023 / Kiran | S11566 |
| [CS 4978-1] | | | | | | |
| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
| Restek | 555223 / Custom 8270 Plus Std #1 [2nd lot at \$100 per ampul if requested - contact ARM with Request] | A0201940 | 02/12/2025 | 08/12/2024 / Rahul | 09/18/2023 / Kiran | S11567 |
| [CS 4978-1] | | | | | | |
| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
| Restek | 31853 / 1,4-Dioxane, 2000 ug/ml , Solvent: Methylene Chloride | A0196453 | 12/14/2024 | 06/14/2024 / Rahul | 11/21/2023 / Rahul | S11766 |
| | | | | | | |
| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
| Restek | 31853 / 1,4-Dioxane, 2000 ug/ml , Solvent: Methylene Chloride | A0196453 | 01/12/2025 | 07/12/2024 / Rahul | 11/21/2023 / Rahul | S11767 |
| | | | | | | |
| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
| Restek | 31853 / 1,4-Dioxane, 2000 ug/ml , Solvent: Methylene Chloride | A0196453 | 02/12/2025 | 08/12/2024 / Rahul | 11/21/2023 / Rahul | S11768 |
| | | | | | | |
| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
| Restek | 31853 / 1,4-Dioxane, 2000 ug/ml , Solvent: Methylene Chloride | A0196453 | 02/12/2025 | 08/12/2024 / Rahul | 11/21/2023 / Rahul | S11769 |

CHEMICAL RECEIPT LOG BOOK

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|-------------------|--|----------|-----------------|-------------------------|-----------------------------|----------------|
| Restek | 31853 / 1,4-Dioxane, 2000 ug/ml , Solvent: Methylene Chloride | A0196453 | 02/12/2025 | 08/12/2024 / Rahul | 11/21/2023 / Rahul | S11770 |
| Restek | 31206 / SV Mix, CLP method, Internal Std, 2000ug/mL, CH2Cl2, 1mL | A0201320 | 01/01/2025 | 07/01/2024 / Rahul | 12/21/2023 / Rahul | S12033 |
| Restek | 31206 / SV Mix, CLP method, Internal Std, 2000ug/mL, CH2Cl2, 1mL | A0201320 | 02/07/2025 | 08/07/2024 / anahy | 12/21/2023 / Rahul | S12039 |
| Restek | 31206 / SV Mix, CLP method, Internal Std, 2000ug/mL, CH2Cl2, 1mL | A0201320 | 02/13/2025 | 08/13/2024 / anahy | 12/21/2023 / Rahul | S12041 |
| CPI International | Z-110816-01 / Custom 8270 Mix, 4-79, 1000 mg/L, 1 mL, (Maximum Expiration: 180 Days) | 414127 | 02/08/2025 | 08/08/2024 / Jagrut | 01/31/2024 / Rahul | S12077 |
| Restek | 555224 / Custom 8270 Plus Std #2 [2nd lot at \$85 per ampul if requested - contact ARM with Request] | A0207706 | 01/09/2025 | 07/09/2024 / Jagrut | 02/05/2024 / Rahul | S12097 |

[CS 4978-2]

CHEMICAL RECEIPT LOG BOOK

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|----------|---|----------|-----------------|-------------------------|-----------------------------|----------------|
| Restek | 555224 / Custom 8270 Plus Std #2 [2nd lot at \$85 per ampul if requested - contact ARM with Request] | A0207706 | 02/12/2025 | 08/12/2024 / Rahul | 02/05/2024 / Rahul | S12098 |

[CS 4978-2]

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|----------|---|----------|-----------------|-------------------------|-----------------------------|----------------|
| Restek | 555224 / Custom 8270 Plus Std #2 [2nd lot at \$85 per ampul if requested - contact ARM with Request] | A0207706 | 02/12/2025 | 08/12/2024 / Rahul | 02/05/2024 / Rahul | S12099 |

[CS 4978-2]

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|----------|---|----------|-----------------|-------------------------|-----------------------------|----------------|
| Restek | 555224 / Custom 8270 Plus Std #2 [2nd lot at \$85 per ampul if requested - contact ARM with Request] | A0207706 | 02/12/2025 | 08/12/2024 / Rahul | 02/05/2024 / Rahul | S12100 |

[CS 4978-2]

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|----------|---|----------|-----------------|-------------------------|-----------------------------|----------------|
| Restek | 555224 / Custom 8270 Plus Std #2 [2nd lot at \$85 per ampul if requested - contact ARM with Request] | A0207706 | 02/12/2025 | 08/12/2024 / Rahul | 02/05/2024 / Rahul | S12101 |

[CS 4978-2]

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|----------|---|----------|-----------------|-------------------------|-----------------------------|----------------|
| Restek | 555224 / Custom 8270 Plus Std #2 [2nd lot at \$85 per ampul if requested - contact ARM with Request] | A0207706 | 02/12/2025 | 08/12/2024 / Rahul | 02/05/2024 / Rahul | S12102 |

[CS 4978-2]

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|----------|---|----------|-----------------|-------------------------|-----------------------------|----------------|
| Restek | 555224 / Custom 8270 Plus Std #2 [2nd lot at \$85 per ampul if requested - contact ARM with Request] | A0207706 | 02/12/2025 | 08/12/2024 / Rahul | 02/05/2024 / Rahul | S12103 |

[CS 4978-2]

CHEMICAL RECEIPT LOG BOOK

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|-------------------|--|----------|-----------------|-------------------------|-----------------------------|----------------|
| Restek | 555224 / Custom 8270 Plus Std #2 [2nd lot at \$85 per ampul if requested - contact ARM with Request] | A0207706 | 02/12/2025 | 08/12/2024 / Rahul | 02/05/2024 / Rahul | S12104 |
| [CS 4978-2] | | | | | | |
| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
| Restek | 555224 / Custom 8270 Plus Std #2 [2nd lot at \$85 per ampul if requested - contact ARM with Request] | A0207706 | 02/12/2025 | 08/12/2024 / Rahul | 02/05/2024 / Rahul | S12105 |
| [CS 4978-2] | | | | | | |
| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
| CPI International | z-010223-01 / 1,4-Dioxane Solution, 2,000mg/L, 1ml | 454157 | 01/09/2025 | 07/09/2024 / Jagrut | 03/08/2024 / Rahul | S12112 |
| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
| Restek | 31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH ₂ Cl ₂ [New Solvent 100% CH ₂ Cl ₂] | A0203726 | 12/05/2024 | 06/05/2024 / Rahul | 03/15/2024 / Rahul | S12117 |
| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
| Restek | 31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH ₂ Cl ₂ [New Solvent 100% CH ₂ Cl ₂] | A0203726 | 04/30/2025 | 01/10/2025 / Jagrut | 03/15/2024 / Rahul | S12118 |
| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
| Restek | 31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH ₂ Cl ₂ [New Solvent 100% CH ₂ Cl ₂] | A0203726 | 02/12/2025 | 08/12/2024 / Rahul | 03/15/2024 / Rahul | S12119 |

CHEMICAL RECEIPT LOG BOOK

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|----------|--|----------|-----------------|-------------------------|-----------------------------|----------------|
| Restek | 31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH2Cl2 [New Solvent 100% CH2Cl2] | A0203726 | 02/12/2025 | 08/12/2024 / Rahul | 03/15/2024 / Rahul | S12120 |
| Restek | 31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH2Cl2 [New Solvent 100% CH2Cl2] | A0203726 | 02/12/2025 | 08/12/2024 / Rahul | 03/15/2024 / Rahul | S12121 |
| Restek | 31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH2Cl2 [New Solvent 100% CH2Cl2] | A0203726 | 02/12/2025 | 08/12/2024 / Rahul | 03/15/2024 / Rahul | S12122 |
| Restek | 31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH2Cl2 [New Solvent 100% CH2Cl2] | A0203726 | 02/12/2025 | 08/12/2024 / Rahul | 03/15/2024 / Rahul | S12123 |
| Restek | 31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH2Cl2 [New Solvent 100% CH2Cl2] | A0203726 | 02/12/2025 | 08/12/2024 / Rahul | 03/15/2024 / Rahul | S12124 |
| Restek | 31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH2Cl2 [New Solvent 100% CH2Cl2] | A0203726 | 02/12/2025 | 08/12/2024 / Rahul | 03/15/2024 / Rahul | S12125 |

CHEMICAL RECEIPT LOG BOOK

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|----------|--|----------|-----------------|-------------------------|-----------------------------|----------------|
| Restek | 31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH ₂ Cl ₂ [New Solvent 100% CH ₂ Cl ₂] | A0203726 | 02/12/2025 | 08/12/2024 / Rahul | 03/15/2024 / Rahul | S12126 |

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|----------|--|----------|-----------------|-------------------------|-----------------------------|----------------|
| Restek | 555223 / Custom 8270 Plus Std #1 [2nd lot at \$100 per ampul if requested - contact ARM with Request] | A0214021 | 02/12/2025 | 08/12/2024 / Rahul | 07/23/2024 / RAHUL | S12449 |

[CS 4978-1]

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|----------|--|----------|-----------------|-------------------------|-----------------------------|----------------|
| Restek | 555223 / Custom 8270 Plus Std #1 [2nd lot at \$100 per ampul if requested - contact ARM with Request] | A0214021 | 02/12/2025 | 08/12/2024 / Rahul | 07/23/2024 / RAHUL | S12450 |

[CS 4978-1]

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|----------|--|----------|-----------------|-------------------------|-----------------------------|----------------|
| Restek | 555223 / Custom 8270 Plus Std #1 [2nd lot at \$100 per ampul if requested - contact ARM with Request] | A0214021 | 02/12/2025 | 08/12/2024 / Rahul | 07/23/2024 / RAHUL | S12451 |

[CS 4978-1]

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|----------|--|----------|-----------------|-------------------------|-----------------------------|----------------|
| Restek | 555223 / Custom 8270 Plus Std #1 [2nd lot at \$100 per ampul if requested - contact ARM with Request] | A0214021 | 02/12/2025 | 08/12/2024 / Rahul | 07/23/2024 / RAHUL | S12452 |

[CS 4978-1]

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|----------|--|----------|-----------------|-------------------------|-----------------------------|----------------|
| Restek | 555223 / Custom 8270 Plus Std #1 [2nd lot at \$100 per ampul if requested - contact ARM with Request] | A0214021 | 02/12/2025 | 08/12/2024 / Rahul | 07/23/2024 / RAHUL | S12453 |

[CS 4978-1]

CHEMICAL RECEIPT LOG BOOK

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|------------------|---------------------|---------------------|-----------------|-------------------------|-----------------------------|----------------|
| Seidler Chemical | DIW / DI Water | Daily Lab-Certified | 07/03/2029 | 07/03/2024 / Iwona | 07/03/2024 / Iwona | W3112 |



5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received: _____

Certificate of Analysis

Rev 0

Page 1 of 1

Catalog No.: Lot No.: Storage: Solvent: Exp. Date: Description:
Z-010074-07 406703 ≤ -10 °C Methylene Chloride 3/30/2025 3,3'-Dichlorobenzidine Solution, 1,000 mg/L, 1 mL

| Compound | CAS No. | Purity (%) | Compound Lot No. | Concentration, mg/L |
|------------------------|---------|------------|------------------|---------------------|
| 3,3'-dichlorobenzidine | 91-94-1 | 99.5 | 74.3.26P | 989 ± 7.53 |

Received on
02/07/23

by

C6

S11084

to

S11088

*Not a certified value

Certified By:

Jacob Mulloy
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values
listed are determined gravimetrically.



5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received: _____

Certificate of Analysis

Rev 0

Page 1 of 1

Catalog No.: Lot No.: Storage: Solvent: Exp. Date: Description:
Z-110817-01 414125 ≤ -10 °C Methylene Chloride 6/21/2025 Custom 8270 Mix, 4-55, 1000 mg/L, 1 mL

| Compound | CAS No. | Purity (%) | Compound Lot No. | Concentration, mg/L |
|----------------------------|---------|------------|------------------|---------------------|
| acetophenone | 98-86-2 | 99.2 | 85.8.1P | 998 ± 11.5 |
| benzoic acid | 65-85-0 | 100 | 123.7.1P | 1010 ± 5.88 |
| biphenyl | 92-52-4 | 99.9 | 366.29.1P | 999 ± 5.82 |
| 1,2,4,5-tetrachlorobenzene | 95-94-3 | 99.7 | 53.7.2P | 993 ± 5.79 |

Received on
02/07/23
by
CG
S 11089
to
S 11093

*Not a certified value

Manufactured by o2si smart solutions, Accredited to ISO 9001:2008 by NSF and ISO/IEC 17025:2005 (Certification No. 3031.01) and ISO Guide 34:2009 (Certification No. 3031.02) by A2LA

Certified By: _____
Shane Overcash
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values
listed are determined gravimetrically.



5580 Skylane Blvd
Santa Rosa, CA 95403

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Date Received: _____

Certificate of Analysis

Rev 0

Page 1 of 1

Catalog No.: Lot No.: Storage: Solvent: Exp. Date: Description:
Z-112090 440246 $\leq -10^{\circ}\text{C}$ Methylene Chloride 2/16/2026 CLP Acid Surrogate Solution, 7,500 mg/L, 1 mL
-04

| Compound | CAS No. | Purity (%) | Compound Lot No. | Concentration, mg/L |
|-------------------------------|------------|------------|------------------|---------------------|
| 2-chlorophenol-d ₄ | 93951-73-6 | 99.3 | 248.12.7P | 7487 \pm 17.2 |
| 2-fluorophenol | 367-12-4 | 99.8 | 10.7.3.3P | 7513 \pm 17.26 |
| phenol-d ₆ | 13127-88-3 | 99.9 | 949.120.8P | 7481 \pm 17.19 |
| 2,4,6-tribromophenol | 118-79-6 | 99.8 | 12.1.6P | 7469 \pm 17.17 |

Received on

02/25/21

by
CG

S9236
+0

S9240

*Not a certified value

Manufactured by o2si smart solutions, Accredited to ISO 9001:2008 by NSF and ISO/IEC 17025:2005 (Certification No. 3031.01) and ISO Guide 34:2009 (Certification No. 3031.02) by A2LA

Certified By:

Erica Castiglione
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetrically.



5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Received on
02/07/23 by C6

SH067 S11096
to
S11099
Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received: _____

Certificate of Analysis

Rev 0

Page 1 of 4

Catalog No.: Lot No.: Storage: Solvent: Exp. Date: Description:
Z-110381-01 495831 ≤ -10 °C Methylene Chloride 10/30/2027 Method 8270 Calibration Solution, 76-1, 500 & 1,000 mg/L, 1 mL

| Compound | CAS No. | Purity (%) | Compound Lot No. | Concentration, mg/L |
|-----------------------------------|----------|------------|------------------|---------------------|
| acenaphthene | 83-32-9 | 99.9 | 13.1.5P | 1003 ± 17.27 |
| acenaphthylene | 208-96-8 | 97.6 | 14.290.1P | 999.8 ± 17.22 |
| aniline | 62-53-3 | 99.9 | 64.7.1P | 995 ± 17.13 |
| anthracene | 120-12-7 | 99.5 | 15.7.1P | 1001 ± 17.24 |
| azobenzene | 103-33-3 | 98.1 | 252.7.2P | 999.1 ± 17.21 |
| benzo[a]anthracene | 56-55-3 | 100 | 16.7.3P | 1001 ± 17.24 |
| benzo[b]fluoranthene | 205-99-2 | 99.8 | 17.421.3P | 1001 ± 19.91 |
| benzo[k]fluoranthene | 207-08-9 | 98.9 | 18.421.4P | 1001 ± 17.92 |
| benzo[ghi]perylene | 191-24-2 | 93 | 19.286.4P | 999.6 ± 19.88 |
| benzo[a]pyrene | 50-32-8 | 97 | 20.286.2P | 999.1 ± 26.35 |
| benzyl alcohol | 100-51-6 | 99.9 | 65.18.1P | 1001 ± 17.24 |
| bis(2-chloroethoxy)methane | 111-91-1 | 99.1 | 31.3.15P | 999.7 ± 17.89 |
| bis(2-chloroethyl)ether | 111-44-4 | 99.8 | 32.7.1P | 1001 ± 17.23 |
| bis(2-chloro-1-methylethyl) ether | 108-60-1 | 99.5 | 34.3.13P | 999.5 ± 17.89 |
| bis(2-ethylhexyl)adipate | 103-23-1 | 99.5 | 874.7.1P | 999.5 ± 17.21 |
| bis(2-ethylhexyl)phthalate | 117-81-7 | 99.4 | 33.29.1P | 998.8 ± 19.86 |
| 4-bromophenyl phenyl ether | 101-55-3 | 99.4 | 35.7.1P | 999.1 ± 17.2 |
| butyl benzyl phthalate | 85-68-7 | 98.4 | 36.1.6P | 984.7 ± 19.58 |
| carbazole | 86-74-8 | 99.4 | 239.7.2P | 1000 ± 17.22 |

*Not a certified value

Manufactured by o2si smart solutions, Accredited to ISO 9001:2008 by NSF and ISO/IEC 17025:2005 (Certification No. 3031.01) and ISO Guide 34:2009 (Certification No. 3031.02) by A2LA

Certified By:

Briana Smith
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetrically.

Certificate of Analysis

Page 4 of 4

Catalog No.: Z-110381-01

Lot No.: 495831

Expiration Date: 10/30/2027

| Compound | CAS No. | Purity (%) | Compound Lot No. | Concentration, mg/L |
|------------------------|----------|------------|------------------|---------------------|
| 1,2,4-trichlorobenzene | 120-82-1 | 99.6 | 54.29.1P | 1000 ± 17.22 |
| 2,4,5-trichlorophenol | 95-95-4 | 96.5 | 121.7.1.1P | 1000 ± 17.22 |
| 2,4,6-trichlorophenol | 88-06-2 | 99.6 | 113.7.1P | 1002 ± 17.25 |

*Not a certified value

Manufactured by o2si smart solutions, Accredited to ISO 9001:2008 by NSF and ISO/IEC 17025:2005 (Certification No. 3031.01) and ISO Guide 34:2009 (Certification No. 3031.02) by A2LA

Certified By:

Briana Smith
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetrically.



5580 Skylane Blvd
Santa Rosa, CA 95403

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Date Received: _____

Certificate of Analysis

Rev 0

Page 1 of 1

Catalog No.: Lot No.: Z-010442-07 **Storage:** 495833 $\leq -10^{\circ}\text{C}$ **Solvent:** Methylene Chloride **Exp. Date:** 1/16/2028 **Description:** Benzaldehyde Solution, 1000 mg/L, 1.3 mL

| Compound | CAS No. | Purity (%) | Compound Lot No. | Concentration, mg/L |
|-----------------|----------------|-------------------|-------------------------|----------------------------|
| benzaldehyde | 100-52-7 | 98.3 | 442.421.1P | 996.8 \pm 11.49 |

Received on
02/07/23
by CG

S11101
to
S11103

*Not a certified value

Certified By:

Scott Hunter
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values
listed are determined gravimetrically.

RESTEK® CERTIFIED REFERENCE MATERIAL

110 Benner Circle
 Bellefonte, PA 16823-8812
 Tel: (800)356-1688
 Fax: (814)353-1309

www.restek.com



Certificate of Analysis



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Received on
 03/16/22
 by
 CG

S10242
 to
 S10247

| | | | |
|-------------------|--|----------|-----------------------|
| Catalog No. : | <u>31615</u> | Lot No.: | <u>A0182667</u> |
| Description : | GC/MS Tuning Mixture | | |
| | GC/MS Tuning Mixture 1,000 μ g/mL, Methylene Chloride, 1mL/ampul | | |
| Container Size : | <u>2 mL</u> | Pkg Amt: | <u>> 1 mL</u> |
| Expiration Date : | <u>March 31, 2025</u> | Storage: | <u>10°C or colder</u> |
| Handling: | <u>Contains carcinogen/reproductive toxin.</u> | | |
| Ship: | <u>Ambient</u> | | |

C E R T I F I E D V A L U E S

| Elution Order | Compound | Grav. Conc. (weight/volume) | Expanded Uncertainty (95% C.L.; K=2) | | |
|---------------|---|--------------------------------|---|------------------------|------------------------|
| 1 | Pentachlorophenol CAS # 87-86-5 Purity 99% | 1,003.6 μ g/mL | +/- 5.8897 μ g/mL | +/- 45.7132 μ g/mL | +/- 66.0037 μ g/mL |
| 2 | DFTPP (Decafluorotriphenylphosphine) CAS # 5074-71-5 Purity 95% | 1,006.6 μ g/mL | +/- 5.9074 μ g/mL | +/- 45.8508 μ g/mL | +/- 66.2023 μ g/mL |
| 3 | Benzidine CAS # 92-87-5 Purity 99% | 1,008.4 μ g/mL | +/- 5.9179 μ g/mL | +/- 45.9318 μ g/mL | +/- 66.3193 μ g/mL |
| 4 | 4,4'-DDT CAS # 50-29-3 Purity 99% | 1,007.6 μ g/mL | +/- 5.9132 μ g/mL | +/- 45.8954 μ g/mL | +/- 66.2667 μ g/mL |

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Column:

30m x 0.25mm x 0.25 μ m
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant pressure 10 psi.

Temp. Program:

75°C (hold 1 min.) to 330°C
@ 20°C/min. (hold 10 min.)

Inj. Temp:


250°C

Det. Temp:

330°C

Det. Type:

FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Morgan Craighead - Mix Technician

Date Mixed: 08-Mar-2022 Balance: B345965662

Marilina Cowan - Operations Tech I

Date Passed: 10-Mar-2022

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

RESTEK® CERTIFIED REFERENCE MATERIAL

110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com



Gravimetric Certificate

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555871

Lot No.: A0185300

Description : Custom 4-Nitrophenol Standard

Custom 4-Nitrophenol Standard 25,000 μ g/mL, Methanol, 1mL/ampul

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : May 31, 2025

Storage: 10°C or colder

Ship: Ambient

Received by

CG on

05/18/22

\$10393

+0

\$10402

C E R T I F I E D V A L U E S

| Component # | Compound | Grav. Conc. (weight/volume) | Expanded Uncertainty (95% C.L.; K=2) | | |
|-------------|---|--------------------------------|---|-------------------------|---------------------------------------|
| 1 | 4-Nitrophenol CAS # 100-02-7 Purity 99% | 25,060.0 μ g/mL | +/- 231.9100 μ g/mL | +/- 753.2622 μ g/mL | Gravimetric Unstressed Stressed |

Solvent: Methanol
CAS # 67-56-1
Purity 99%

Katelyn McGinn - Operations Tech I

Date Mixed: 16-May-2022 Balance: 1128342314

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/ μ ECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified combined stressed uncertainty value (includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{\text{combined stressed}} = k \sqrt{U_{\text{gravimetric}}^2 + U_{\text{homogeneity}}^2 + U_{\text{storage stability}}^2 + U_{\text{shipping stability}}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- It is important to note that the shipping stability uncertainty was obtained under temperature extremes for specific time intervals; therefore, the certified combined stressed uncertainty value should only be applied to the product if it was stored at non-standard temperature conditions up to and including 7 days. Contact Restek Technical Service at www.restek.com/Contact-Us for use recommendations if your shipment was in-transit for more than 7 days at non-standard temperature conditions.
- Apply the certified combined unstressed uncertainty value if the product was received under standard shipping conditions. Apply the certified combined stressed uncertainty value if the product was received under non-standard conditions as specified below.

| Label Conditions | Standard Conditions | Non-Standard Conditions |
|---|---------------------|-------------------------|
| 25°C Nominal (Room Temperature) | < 60°C | ≥ 60°C up to 7 days |
| 10°C or colder (Refrigerate) | < 40°C | ≥ 40°C up to 7 days |
| 0°C or colder (Freezer) -20°C or colder (Deep Freezer) | < 25°C | ≥ 25°C up to 7 days |

- Separate (not combined) uncertainty values for gravimetric uncertainty are also displayed on the certificate, if needed, separate homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty values are available by contacting Restek Technical Service at www.restek.com/Contact-Us.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampules are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampules. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.



CERTIFIED REFERENCE MATERIAL

110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com



Gravimetric Certificate



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555868

Lot No.: A0186373

Description : Custom Benzidine Standard

Custom Benzidine Standard 25,000 μ g/mL, Methanol, 1mL/ampul

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : June 30, 2025

Storage: 10°C or colder

Handling: Contains carcinogen/reproductive toxin.

Ship: Ambient

Received by

CG

on

07/05/22

S 10583

to

S 10592

C E R T I F I E D V A L U E S

| Component # | Compound | Grav. Conc. (weight/volume) | Expanded Uncertainty (95% C.L.; K=2) | | |
|-------------|--|--------------------------------|---|------------|-------------|
| 1 | Benzidine CAS # 92-87-5 Purity 99% | 25,200.0 μ g/mL | +/- 233.2055 | μ g/mL | Gravimetric |
| | (Lot 220511RSR) | | +/- 351.6606 | μ g/mL | Unstressed |
| | | | +/- 512.6054 | μ g/mL | Stressed |

Solvent: Methanol
CAS # 67-56-1
Purity 99%

Tom Suckar - Mix Technician

Date Mixed: 16-Jun-2022 Balance: 1122030677

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/ μ ECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified combined stressed uncertainty value (includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{\text{combined stressed}} = k \sqrt{U_{\text{gravimetric}}^2 + U_{\text{homogeneity}}^2 + U_{\text{storage stability}}^2 + U_{\text{shipping stability}}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- It is important to note that the shipping stability uncertainty was obtained under temperature extremes for specific time intervals; therefore, the certified combined stressed uncertainty value should only be applied to the product if it was stored at non-standard temperature conditions up to and including 7 days. Contact Restek Technical Service at www.restek.com/Contact-Us for use recommendations if your shipment was in-transit for more than 7 days at non-standard temperature conditions.
- Apply the certified combined unstressed uncertainty value if the product was received under standard shipping conditions. Apply the certified combined stressed uncertainty value if the product was received under non-standard conditions as specified below.

| Label Conditions | Standard Conditions | Non-Standard Conditions |
|---|---------------------|-------------------------|
| 25°C Nominal (Room Temperature) | < 60°C | ≥ 60°C up to 7 days |
| 10°C or colder (Refrigerate) | < 40°C | ≥ 40°C up to 7 days |
| 0°C or colder (Freezer) -20°C or colder (Deep Freezer) | < 25°C | ≥ 25°C up to 7 days |

- Separate (not combined) uncertainty values for gravimetric uncertainty are also displayed on the certificate, if needed, separate homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty values are available by contacting Restek Technical Service at www.restek.com/Contact-Us.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampules are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampules. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.

RESTEK® CERTIFIED REFERENCE MATERIAL

110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com



Certificate of Analysis



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31087

Lot No.: A0188108

Description : Acid Surrogate Mix (4/89 SOW)

Acid Surrogate 10,000 μ g/mL, Methanol, 5mL/ampul

Container Size : 5 mL

Pkg Amt: > 5 mL

Expiration Date : August 31, 2030

Storage: 10°C or colder

Ship: Ambient

Received by
CG on
12/28/22
S10951
to
S10980

C E R T I F I E D V A L U E S

| Elution Order | Compound | Grav. Conc. (weight/volume) | Expanded Uncertainty (95% C.L.; K=2) | | |
|---------------|--|--------------------------------|---|------------|-------------|
| 1 | 2-Fluorophenol CAS # 367-12-4 Purity 99% | 10,088.5 μ g/mL | +/- 58.6554 | μ g/mL | Gravimetric |
| | (Lot STBF3761V) | | +/- 294.4162 | μ g/mL | Unstressed |
| | | | +/- 357.2628 | μ g/mL | Stressed |
| 2 | Phenol-d6 CAS # 13127-88-3 Purity 99% | 10,043.3 μ g/mL | +/- 58.3923 | μ g/mL | Gravimetric |
| | (Lot PR-31262) | | +/- 293.0957 | μ g/mL | Unstressed |
| | | | +/- 355.6603 | μ g/mL | Stressed |
| 3 | 2,4,6-Tribromophenol CAS # 118-79-6 Purity 99% | 10,010.0 μ g/mL | +/- 58.1990 | μ g/mL | Gravimetric |
| | (Lot MKCJ7664) | | +/- 292.1253 | μ g/mL | Unstressed |
| | | | +/- 354.4829 | μ g/mL | Stressed |

Solvent: Methanol
CAS # 67-56-1
Purity 99%

Column:

30m x 0.25mm x 0.25 μ m
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant pressure 10 psi.

Temp. Program:

40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

Inj. Temp:


250°C

Det. Temp:

330°C

Det. Type:

FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.



Morgan Craighead - Mix Technician

Date Mixed: 02-Aug-2022 Balance: 1127510105

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 05-Aug-2022

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

RESTEK® CERTIFIED REFERENCE MATERIAL

110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com



Certificate of Analysis



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31087

Lot No.: A0188108

Description : Acid Surrogate Mix (4/89 SOW)

Acid Surrogate 10,000 μ g/mL, Methanol, 5mL/ampul

Container Size : 5 mL

Pkg Amt: > 5 mL

Expiration Date : August 31, 2030

Storage: 10°C or colder

Ship: Ambient

Received by
CG on
12/28/22
S10951
to
S10980

C E R T I F I E D V A L U E S

| Elution Order | Compound | Grav. Conc. (weight/volume) | Expanded Uncertainty (95% C.L.; K=2) | | |
|---------------|--|--------------------------------|---|------------|-------------|
| 1 | 2-Fluorophenol CAS # 367-12-4 Purity 99% | 10,088.5 μ g/mL | +/- 58.6554 | μ g/mL | Gravimetric |
| | (Lot STBF3761V) | | +/- 294.4162 | μ g/mL | Unstressed |
| | | | +/- 357.2628 | μ g/mL | Stressed |
| 2 | Phenol-d6 CAS # 13127-88-3 Purity 99% | 10,043.3 μ g/mL | +/- 58.3923 | μ g/mL | Gravimetric |
| | (Lot PR-31262) | | +/- 293.0957 | μ g/mL | Unstressed |
| | | | +/- 355.6603 | μ g/mL | Stressed |
| 3 | 2,4,6-Tribromophenol CAS # 118-79-6 Purity 99% | 10,010.0 μ g/mL | +/- 58.1990 | μ g/mL | Gravimetric |
| | (Lot MKCJ7664) | | +/- 292.1253 | μ g/mL | Unstressed |
| | | | +/- 354.4829 | μ g/mL | Stressed |

Solvent: Methanol
CAS # 67-56-1
Purity 99%

Column:

30m x 0.25mm x 0.25 μ m
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant pressure 10 psi.

Temp. Program:

40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

Inj. Temp:


250°C

Det. Temp:

330°C

Det. Type:

FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.



Morgan Craighead - Mix Technician

Date Mixed: 02-Aug-2022 Balance: 1127510105

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 05-Aug-2022

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

RESTEK® CERTIFIED REFERENCE MATERIAL

110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com



Certificate of Analysis



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31087

Lot No.: A0188108

Description : Acid Surrogate Mix (4/89 SOW)

Acid Surrogate 10,000 μ g/mL, Methanol, 5mL/ampul

Container Size : 5 mL

Pkg Amt: > 5 mL

Expiration Date : August 31, 2030

Storage: 10°C or colder

Ship: Ambient

Received by
CG on
12/28/22
S10951
to
S10980

C E R T I F I E D V A L U E S

| Elution Order | Compound | Grav. Conc. (weight/volume) | Expanded Uncertainty (95% C.L.; K=2) | | |
|---------------|--|--------------------------------|---|------------|-------------|
| 1 | 2-Fluorophenol CAS # 367-12-4 Purity 99% | 10,088.5 μ g/mL | +/- 58.6554 | μ g/mL | Gravimetric |
| | (Lot STBF3761V) | | +/- 294.4162 | μ g/mL | Unstressed |
| | | | +/- 357.2628 | μ g/mL | Stressed |
| 2 | Phenol-d6 CAS # 13127-88-3 Purity 99% | 10,043.3 μ g/mL | +/- 58.3923 | μ g/mL | Gravimetric |
| | (Lot PR-31262) | | +/- 293.0957 | μ g/mL | Unstressed |
| | | | +/- 355.6603 | μ g/mL | Stressed |
| 3 | 2,4,6-Tribromophenol CAS # 118-79-6 Purity 99% | 10,010.0 μ g/mL | +/- 58.1990 | μ g/mL | Gravimetric |
| | (Lot MKCJ7664) | | +/- 292.1253 | μ g/mL | Unstressed |
| | | | +/- 354.4829 | μ g/mL | Stressed |

Solvent: Methanol
CAS # 67-56-1
Purity 99%

Column:

30m x 0.25mm x 0.25 μ m
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant pressure 10 psi.

Temp. Program:

40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

Inj. Temp:


250°C

Det. Temp:

330°C

Det. Type:

FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.



Morgan Craighead - Mix Technician

Date Mixed: 02-Aug-2022 Balance: 1127510105

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 05-Aug-2022

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

RESTEK® CERTIFIED REFERENCE MATERIAL

110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com



Certificate of Analysis



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31087

Lot No.: A0188108

Description : Acid Surrogate Mix (4/89 SOW)

Acid Surrogate 10,000 μ g/mL, Methanol, 5mL/ampul

Container Size : 5 mL

Pkg Amt: > 5 mL

Expiration Date : August 31, 2030

Storage: 10°C or colder

Ship: Ambient

Received by
CG on
12/28/22
S10951
to
S10980

C E R T I F I E D V A L U E S

| Elution Order | Compound | Grav. Conc. (weight/volume) | Expanded Uncertainty (95% C.L.; K=2) | | |
|---------------|--|--------------------------------|---|------------|-------------|
| 1 | 2-Fluorophenol CAS # 367-12-4 Purity 99% | 10,088.5 μ g/mL | +/- 58.6554 | μ g/mL | Gravimetric |
| | (Lot STBF3761V) | | +/- 294.4162 | μ g/mL | Unstressed |
| | | | +/- 357.2628 | μ g/mL | Stressed |
| 2 | Phenol-d6 CAS # 13127-88-3 Purity 99% | 10,043.3 μ g/mL | +/- 58.3923 | μ g/mL | Gravimetric |
| | (Lot PR-31262) | | +/- 293.0957 | μ g/mL | Unstressed |
| | | | +/- 355.6603 | μ g/mL | Stressed |
| 3 | 2,4,6-Tribromophenol CAS # 118-79-6 Purity 99% | 10,010.0 μ g/mL | +/- 58.1990 | μ g/mL | Gravimetric |
| | (Lot MKCJ7664) | | +/- 292.1253 | μ g/mL | Unstressed |
| | | | +/- 354.4829 | μ g/mL | Stressed |

Solvent: Methanol
CAS # 67-56-1
Purity 99%

Column:

30m x 0.25mm x 0.25 μ m
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant pressure 10 psi.

Temp. Program:

40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

Inj. Temp:


250°C

Det. Temp:

330°C

Det. Type:

FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.



Morgan Craighead - Mix Technician

Date Mixed: 02-Aug-2022 Balance: 1127510105

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 05-Aug-2022

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

RESTEK® CERTIFIED REFERENCE MATERIAL

110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com



Certificate of Analysis



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31087

Lot No.: A0188108

Description : Acid Surrogate Mix (4/89 SOW)

Acid Surrogate 10,000 μ g/mL, Methanol, 5mL/ampul

Container Size : 5 mL

Pkg Amt: > 5 mL

Expiration Date : August 31, 2030

Storage: 10°C or colder

Ship: Ambient

Received by
CG on
12/28/22
S10951
to
S10980

C E R T I F I E D V A L U E S

| Elution Order | Compound | Grav. Conc. (weight/volume) | Expanded Uncertainty (95% C.L.; K=2) | | |
|---------------|--|--------------------------------|---|------------|-------------|
| 1 | 2-Fluorophenol CAS # 367-12-4 Purity 99% | 10,088.5 μ g/mL | +/- 58.6554 | μ g/mL | Gravimetric |
| | (Lot STBF3761V) | | +/- 294.4162 | μ g/mL | Unstressed |
| | | | +/- 357.2628 | μ g/mL | Stressed |
| 2 | Phenol-d6 CAS # 13127-88-3 Purity 99% | 10,043.3 μ g/mL | +/- 58.3923 | μ g/mL | Gravimetric |
| | (Lot PR-31262) | | +/- 293.0957 | μ g/mL | Unstressed |
| | | | +/- 355.6603 | μ g/mL | Stressed |
| 3 | 2,4,6-Tribromophenol CAS # 118-79-6 Purity 99% | 10,010.0 μ g/mL | +/- 58.1990 | μ g/mL | Gravimetric |
| | (Lot MKCJ7664) | | +/- 292.1253 | μ g/mL | Unstressed |
| | | | +/- 354.4829 | μ g/mL | Stressed |

Solvent: Methanol
CAS # 67-56-1
Purity 99%

Column:

30m x 0.25mm x 0.25 μ m
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant pressure 10 psi.

Temp. Program:

40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

Inj. Temp:


250°C

Det. Temp:

330°C

Det. Type:

FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.



Morgan Craighead - Mix Technician

Date Mixed: 02-Aug-2022 Balance: 1127510105

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 05-Aug-2022

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

RESTEK® CERTIFIED REFERENCE MATERIAL

110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com



Certificate of Analysis



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31087

Lot No.: A0188108

Description : Acid Surrogate Mix (4/89 SOW)

Acid Surrogate 10,000 μ g/mL, Methanol, 5mL/ampul

Container Size : 5 mL

Pkg Amt: > 5 mL

Expiration Date : August 31, 2030

Storage: 10°C or colder

Ship: Ambient

Received by
CG on
12/28/22
S10951
to
S10980

C E R T I F I E D V A L U E S

| Elution Order | Compound | Grav. Conc. (weight/volume) | Expanded Uncertainty (95% C.L.; K=2) | | |
|---------------|--|--------------------------------|---|------------|-------------|
| 1 | 2-Fluorophenol CAS # 367-12-4 Purity 99% | 10,088.5 μ g/mL | +/- 58.6554 | μ g/mL | Gravimetric |
| | (Lot STBF3761V) | | +/- 294.4162 | μ g/mL | Unstressed |
| | | | +/- 357.2628 | μ g/mL | Stressed |
| 2 | Phenol-d6 CAS # 13127-88-3 Purity 99% | 10,043.3 μ g/mL | +/- 58.3923 | μ g/mL | Gravimetric |
| | (Lot PR-31262) | | +/- 293.0957 | μ g/mL | Unstressed |
| | | | +/- 355.6603 | μ g/mL | Stressed |
| 3 | 2,4,6-Tribromophenol CAS # 118-79-6 Purity 99% | 10,010.0 μ g/mL | +/- 58.1990 | μ g/mL | Gravimetric |
| | (Lot MKCJ7664) | | +/- 292.1253 | μ g/mL | Unstressed |
| | | | +/- 354.4829 | μ g/mL | Stressed |

Solvent: Methanol
CAS # 67-56-1
Purity 99%

Column:

30m x 0.25mm x 0.25 μ m
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant pressure 10 psi.

Temp. Program:

40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

Inj. Temp:


250°C

Det. Temp:

330°C

Det. Type:

FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.



Morgan Craighead - Mix Technician

Date Mixed: 02-Aug-2022 Balance: 1127510105

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 05-Aug-2022

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL



Certificate of Analysis



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31086 Lot No.: A0189418
 Description : B/N Surrogate Mix (4/89 SOW)
Base Neutral Surrogate 5000 μ g/mL, Methylene Chloride, 5mL/ampul
 Container Size : 5 mL Pkg Amt: > 5 mL
 Expiration Date : August 31, 2028 Storage: 10°C or colder
 Handling: Sonicate prior to use. Ship: Ambient

Received by
CG on
12/28/22
Storage
to
Silo 10

C E R T I F I E D V A L U E S

| Elution Order | Compound | Grav. Conc. (weight/volume) | Expanded Uncertainty (95% C.L.; K=2) | | |
|---------------|--|--------------------------------|---|-------------------------|-------------------------|
| 1 | Nitrobenzene-d5 CAS # 4165-60-0 Purity 99% | 5,009.8 μ g/mL | +/- 29.1271 μ g/mL | +/- 225.6421 μ g/mL | +/- 250.3778 μ g/mL |
| 2 | 2-Fluorobiphenyl CAS # 321-60-8 Purity 99% | 5,026.6 μ g/mL | +/- 29.2250 μ g/mL | +/- 226.4003 μ g/mL | +/- 251.2191 μ g/mL |
| 3 | p-Terphenyl-d14 CAS # 1718-51-0 Purity 99% | 5,027.3 μ g/mL | +/- 29.2289 μ g/mL | +/- 226.4304 μ g/mL | +/- 251.2524 μ g/mL |

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Tech Tips:

Due to the limited solubility of p-terphenyl-d14 in methanol, we do not recommend that this mixture be diluted in methanol.

Column:
30m x 0.25mm x 0.25 μ m
Rtx-5 (cat.#10223)


Carrier Gas:
hydrogen-constant pressure 10 psi.

Temp. Program:
40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

Inj. Temp:
250°C

Det. Temp:
330°C

Det. Type:
FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

[Signature]
John Friedline - Operations Technician I

Date Mixed: 09-Sep-2022 Balance: 1128353505

[Signature]
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 13-Sep-2022

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL



Certificate of Analysis



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31086 Lot No.: A0189418
 Description : B/N Surrogate Mix (4/89 SOW)
Base Neutral Surrogate 5000 μ g/mL, Methylene Chloride, 5mL/ampul
 Container Size : 5 mL Pkg Amt: > 5 mL
 Expiration Date : August 31, 2028 Storage: 10°C or colder
 Handling: Sonicate prior to use. Ship: Ambient

Received by
CG on
12/28/22
Storage
to
Silo 10

C E R T I F I E D V A L U E S

| Elution Order | Compound | Grav. Conc. (weight/volume) | Expanded Uncertainty (95% C.L.; K=2) | | |
|---------------|--|--------------------------------|---|-------------------------|-------------------------|
| 1 | Nitrobenzene-d5 CAS # 4165-60-0 Purity 99% | 5,009.8 μ g/mL | +/- 29.1271 μ g/mL | +/- 225.6421 μ g/mL | +/- 250.3778 μ g/mL |
| 2 | 2-Fluorobiphenyl CAS # 321-60-8 Purity 99% | 5,026.6 μ g/mL | +/- 29.2250 μ g/mL | +/- 226.4003 μ g/mL | +/- 251.2191 μ g/mL |
| 3 | p-Terphenyl-d14 CAS # 1718-51-0 Purity 99% | 5,027.3 μ g/mL | +/- 29.2289 μ g/mL | +/- 226.4304 μ g/mL | +/- 251.2524 μ g/mL |

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Tech Tips:

Due to the limited solubility of p-terphenyl-d14 in methanol, we do not recommend that this mixture be diluted in methanol.

Column:
30m x 0.25mm x 0.25 μ m
Rtx-5 (cat.#10223)


Carrier Gas:
hydrogen-constant pressure 10 psi.

Temp. Program:
40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

Inj. Temp:
250°C

Det. Temp:
330°C

Det. Type:
FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

[Signature]
John Friedline - Operations Technician I

Date Mixed: 09-Sep-2022 Balance: 1128353505

[Signature]
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 13-Sep-2022

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL



Certificate of Analysis



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31086 Lot No.: A0189418
 Description : B/N Surrogate Mix (4/89 SOW)
Base Neutral Surrogate 5000 μ g/mL, Methylene Chloride, 5mL/ampul
 Container Size : 5 mL Pkg Amt: > 5 mL
 Expiration Date : August 31, 2028 Storage: 10°C or colder
 Handling: Sonicate prior to use. Ship: Ambient

Received by
CG on
12/28/22
Storage
to
Silo 10

C E R T I F I E D V A L U E S

| Elution Order | Compound | Grav. Conc. (weight/volume) | Expanded Uncertainty (95% C.L.; K=2) | | |
|---------------|--|--------------------------------|---|-------------------------|-------------------------|
| 1 | Nitrobenzene-d5 CAS # 4165-60-0 Purity 99% | 5,009.8 μ g/mL | +/- 29.1271 μ g/mL | +/- 225.6421 μ g/mL | +/- 250.3778 μ g/mL |
| 2 | 2-Fluorobiphenyl CAS # 321-60-8 Purity 99% | 5,026.6 μ g/mL | +/- 29.2250 μ g/mL | +/- 226.4003 μ g/mL | +/- 251.2191 μ g/mL |
| 3 | p-Terphenyl-d14 CAS # 1718-51-0 Purity 99% | 5,027.3 μ g/mL | +/- 29.2289 μ g/mL | +/- 226.4304 μ g/mL | +/- 251.2524 μ g/mL |

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Tech Tips:

Due to the limited solubility of p-terphenyl-d14 in methanol, we do not recommend that this mixture be diluted in methanol.

Column:
30m x 0.25mm x 0.25 μ m
Rtx-5 (cat.#10223)


Carrier Gas:
hydrogen-constant pressure 10 psi.

Temp. Program:
40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

Inj. Temp:
250°C

Det. Temp:
330°C

Det. Type:
FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

[Signature]
John Friedline - Operations Technician I

Date Mixed: 09-Sep-2022 Balance: 1128353505

[Signature]
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 13-Sep-2022

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL



Certificate of Analysis



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

| | | | |
|---|------------------------------|----------|----------------|
| Catalog No. : | 31086 | Lot No.: | A0189418 |
| Description : | B/N Surrogate Mix (4/89 SOW) | | |
| Base Neutral Surrogate 5000 μ g/mL, Methylene Chloride, 5mL/ampul | | | |
| Container Size : | 5 mL | Pkg Amt: | > 5 mL |
| Expiration Date : | August 31, 2028 | Storage: | 10°C or colder |
| Handling: | Sonicate prior to use. | | |
| | | Ship: | Ambient |

Received by
CG on
12/28/22
Storage
to
Silo 10

C E R T I F I E D V A L U E S

| Elution Order | Compound | Grav. Conc. (weight/volume) | Expanded Uncertainty (95% C.L.; K=2) | | |
|---------------|--|--------------------------------|---|-------------------------|---------------------------|
| 1 | Nitrobenzene-d5 CAS # 4165-60-0 Purity 99% | 5,009.8 μ g/mL | +/- 29.1271 μ g/mL | +/- 225.6421 μ g/mL | Gravimetric Unstressed |
| | (Lot PR-29940A) | | +/- 250.3778 μ g/mL | +/- 251.2191 μ g/mL | Stressed |
| 2 | 2-Fluorobiphenyl CAS # 321-60-8 Purity 99% | 5,026.6 μ g/mL | +/- 29.2250 μ g/mL | +/- 226.4003 μ g/mL | Gravimetric Unstressed |
| | (Lot 00021384) | | +/- 251.2191 μ g/mL | +/- 251.2524 μ g/mL | Stressed |
| 3 | p-Terphenyl-d14 CAS # 1718-51-0 Purity 99% | 5,027.3 μ g/mL | +/- 29.2289 μ g/mL | +/- 226.4304 μ g/mL | Gravimetric Unstressed |
| | (Lot PR-30504) | | +/- 251.2524 μ g/mL | +/- 251.2524 μ g/mL | Stressed |

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Tech Tips:

Due to the limited solubility of p-terphenyl-d14 in methanol, we do not recommend that this mixture be diluted in methanol.

Column:
30m x 0.25mm x 0.25 μ m
Rtx-5 (cat.#10223)


Carrier Gas:
hydrogen-constant pressure 10 psi.

Temp. Program:
40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

Inj. Temp:
250°C

Det. Temp:
330°C

Det. Type:
FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

[Signature]
John Friedline - Operations Technician I

Date Mixed: 09-Sep-2022 Balance: 1128353505

[Signature]
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 13-Sep-2022

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL



Certificate of Analysis



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

| | | | |
|---|------------------------------|----------|----------------|
| Catalog No. : | 31086 | Lot No.: | A0189418 |
| Description : | B/N Surrogate Mix (4/89 SOW) | | |
| Base Neutral Surrogate 5000 μ g/mL, Methylene Chloride, 5mL/ampul | | | |
| Container Size : | 5 mL | Pkg Amt: | > 5 mL |
| Expiration Date : | August 31, 2028 | Storage: | 10°C or colder |
| Handling: | Sonicate prior to use. | | |
| | | Ship: | Ambient |

Received by
CG on
12/28/22
Storage
to
Silo 10

C E R T I F I E D V A L U E S

| Elution Order | Compound | Grav. Conc. (weight/volume) | Expanded Uncertainty (95% C.L.; K=2) | | |
|---------------|--|--------------------------------|---|-------------------------|---------------------------|
| 1 | Nitrobenzene-d5 CAS # 4165-60-0 Purity 99% | 5,009.8 μ g/mL | +/- 29.1271 μ g/mL | +/- 225.6421 μ g/mL | Gravimetric Unstressed |
| | (Lot PR-29940A) | | +/- 250.3778 μ g/mL | +/- 251.2191 μ g/mL | Stressed |
| 2 | 2-Fluorobiphenyl CAS # 321-60-8 Purity 99% | 5,026.6 μ g/mL | +/- 29.2250 μ g/mL | +/- 226.4003 μ g/mL | Gravimetric Unstressed |
| | (Lot 00021384) | | +/- 251.2191 μ g/mL | +/- 251.2524 μ g/mL | Stressed |
| 3 | p-Terphenyl-d14 CAS # 1718-51-0 Purity 99% | 5,027.3 μ g/mL | +/- 29.2289 μ g/mL | +/- 226.4304 μ g/mL | Gravimetric Unstressed |
| | (Lot PR-30504) | | +/- 251.2524 μ g/mL | +/- 251.2524 μ g/mL | Stressed |

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Tech Tips:

Due to the limited solubility of p-terphenyl-d14 in methanol, we do not recommend that this mixture be diluted in methanol.

Column:
30m x 0.25mm x 0.25 μ m
Rtx-5 (cat.#10223)


Carrier Gas:
hydrogen-constant pressure 10 psi.

Temp. Program:
40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

Inj. Temp:
250°C

Det. Temp:
330°C

Det. Type:
FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

[Signature]
John Friedline - Operations Technician I

Date Mixed: 09-Sep-2022 Balance: 1128353505

[Signature]
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 13-Sep-2022

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL



Certificate of Analysis



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

| | | | |
|---|------------------------------|----------|----------------|
| Catalog No. : | 31086 | Lot No.: | A0189418 |
| Description : | B/N Surrogate Mix (4/89 SOW) | | |
| Base Neutral Surrogate 5000 μ g/mL, Methylene Chloride, 5mL/ampul | | | |
| Container Size : | 5 mL | Pkg Amt: | > 5 mL |
| Expiration Date : | August 31, 2028 | Storage: | 10°C or colder |
| Handling: | Sonicate prior to use. | | |
| | | Ship: | Ambient |

Received by
CG on
12/28/22
Storage
to
Silo 10


C E R T I F I E D V A L U E S

| Elution Order | Compound | Grav. Conc. (weight/volume) | Expanded Uncertainty (95% C.L.; K=2) | | |
|---------------|--|--------------------------------|---|-------------------------|---------------------------|
| 1 | Nitrobenzene-d5 CAS # 4165-60-0 Purity 99% | 5,009.8 μ g/mL | +/- 29.1271 μ g/mL | +/- 225.6421 μ g/mL | Gravimetric Unstressed |
| | (Lot PR-29940A) | | +/- 250.3778 μ g/mL | +/- 251.2191 μ g/mL | Stressed |
| 2 | 2-Fluorobiphenyl CAS # 321-60-8 Purity 99% | 5,026.6 μ g/mL | +/- 29.2250 μ g/mL | +/- 226.4003 μ g/mL | Gravimetric Unstressed |
| | (Lot 00021384) | | +/- 251.2191 μ g/mL | +/- 251.2524 μ g/mL | Stressed |
| 3 | p-Terphenyl-d14 CAS # 1718-51-0 Purity 99% | 5,027.3 μ g/mL | +/- 29.2289 μ g/mL | +/- 226.4304 μ g/mL | Gravimetric Unstressed |
| | (Lot PR-30504) | | +/- 251.2524 μ g/mL | +/- 251.2524 μ g/mL | Stressed |

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Tech Tips:

Due to the limited solubility of p-terphenyl-d14 in methanol, we do not recommend that this mixture be diluted in methanol.





110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL



Certificate of Analysis



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

| | | | |
|---|------------------------------|----------|----------------|
| Catalog No. : | 31086 | Lot No.: | A0189418 |
| Description : | B/N Surrogate Mix (4/89 SOW) | | |
| Base Neutral Surrogate 5000 μ g/mL, Methylene Chloride, 5mL/ampul | | | |
| Container Size : | 5 mL | Pkg Amt: | > 5 mL |
| Expiration Date : | August 31, 2028 | Storage: | 10°C or colder |
| Handling: | Sonicate prior to use. | | |
| | | Ship: | Ambient |

Received by
CG on
12/28/22
Storage
to
Silo 10

C E R T I F I E D V A L U E S

| Elution Order | Compound | Grav. Conc. (weight/volume) | Expanded Uncertainty (95% C.L.; K=2) | | |
|---------------|--|--------------------------------|---|-------------------------|---------------------------|
| 1 | Nitrobenzene-d5 CAS # 4165-60-0 Purity 99% | 5,009.8 μ g/mL | +/- 29.1271 μ g/mL | +/- 225.6421 μ g/mL | Gravimetric Unstressed |
| | (Lot PR-29940A) | | +/- 250.3778 μ g/mL | +/- 251.2191 μ g/mL | Stressed |
| 2 | 2-Fluorobiphenyl CAS # 321-60-8 Purity 99% | 5,026.6 μ g/mL | +/- 29.2250 μ g/mL | +/- 226.4003 μ g/mL | Gravimetric Unstressed |
| | (Lot 00021384) | | +/- 251.2191 μ g/mL | +/- 251.2524 μ g/mL | Stressed |
| 3 | p-Terphenyl-d14 CAS # 1718-51-0 Purity 99% | 5,027.3 μ g/mL | +/- 29.2289 μ g/mL | +/- 226.4304 μ g/mL | Gravimetric Unstressed |
| | (Lot PR-30504) | | +/- 251.2524 μ g/mL | +/- 251.2524 μ g/mL | Stressed |

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Tech Tips:

Due to the limited solubility of p-terphenyl-d14 in methanol, we do not recommend that this mixture be diluted in methanol.

Column:
30m x 0.25mm x 0.25 μ m
Rtx-5 (cat.#10223)


Carrier Gas:
hydrogen-constant pressure 10 psi.

Temp. Program:
40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

Inj. Temp:
250°C

Det. Temp:
330°C

Det. Type:
FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

[Signature]
John Friedline - Operations Technician I

Date Mixed: 09-Sep-2022 Balance: 1128353505

[Signature]
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 13-Sep-2022

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL



Certificate of Analysis



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31086 Lot No.: A0189418
 Description : B/N Surrogate Mix (4/89 SOW)
Base Neutral Surrogate 5000 μ g/mL, Methylene Chloride, 5mL/ampul
 Container Size : 5 mL Pkg Amt: > 5 mL
 Expiration Date : August 31, 2028 Storage: 10°C or colder
 Handling: Sonicate prior to use. Ship: Ambient

Received by
CG on
12/28/22
Storage
to
Silo 10

C E R T I F I E D V A L U E S

| Elution Order | Compound | Grav. Conc. (weight/volume) | Expanded Uncertainty (95% C.L.; K=2) | | |
|---------------|--|--------------------------------|---|-------------------------|-------------------------|
| 1 | Nitrobenzene-d5 CAS # 4165-60-0 Purity 99% | 5,009.8 μ g/mL | +/- 29.1271 μ g/mL | +/- 225.6421 μ g/mL | +/- 250.3778 μ g/mL |
| 2 | 2-Fluorobiphenyl CAS # 321-60-8 Purity 99% | 5,026.6 μ g/mL | +/- 29.2250 μ g/mL | +/- 226.4003 μ g/mL | +/- 251.2191 μ g/mL |
| 3 | p-Terphenyl-d14 CAS # 1718-51-0 Purity 99% | 5,027.3 μ g/mL | +/- 29.2289 μ g/mL | +/- 226.4304 μ g/mL | +/- 251.2524 μ g/mL |

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Tech Tips:

Due to the limited solubility of p-terphenyl-d14 in methanol, we do not recommend that this mixture be diluted in methanol.

Column:
30m x 0.25mm x 0.25 μ m
Rtx-5 (cat.#10223)


Carrier Gas:
hydrogen-constant pressure 10 psi.

Temp. Program:
40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

Inj. Temp:
250°C

Det. Temp:
330°C

Det. Type:
FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

[Signature]
John Friedline - Operations Technician I

Date Mixed: 09-Sep-2022 Balance: 1128353505

[Signature]
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 13-Sep-2022

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555872

Lot No.: A0193449

Description : Custom Pentachlorophenol Standard

Custom Pentachlorophenol Standard 25,000 μ g/mL, Methanol,
1mL/ampul

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : January 31, 2026

Storage: 10°C or colder

Ship: Ambient

Received on

01/3/23

by

C6

S11011

to

S11015

C E R T I F I E D V A L U E S

| Component # | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|-------------|-------------------|---------|----------|--------|--------------------------------|--|
| 1 | Pentachlorophenol | 87-86-5 | RP221012 | 99% | 25,050.0 μ g/mL | +/- 778.6378 |

Solvent: Methanol
CAS # 67-56-1
Purity 99%

Russ Bookhamer - Operations Technician I

Date Mixed: 11-Jan-2023

Balance: B442140311

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/ μ ECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{\text{combined uncertainty}} = k \sqrt{u_{\text{gravimetric}}^2 + u_{\text{homogeneity}}^2 + u_{\text{storage stability}}^2 + u_{\text{shipping stability}}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555870

Lot No.: A0194698

Description : Custom 2,4-Dinitrophenol Standard

Custom 2,4-Dinitrophenol Standard 25,000 μ g/mL, Methanol, 1mL/ampul

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : February 28, 2026


Storage: 10°C or colder

Ship: Ambient

C E R T I F I C A T E

| Component # | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) |
|-------------|-------------------|---------|-------------|--------|--------------------------------|
| 1 | 2,4-Dinitrophenol | 51-28-5 | DR221221RSR | 99% | 25,195.0 μ g/mL |

Solvent: Methanol
CAS # 67-56-1
Purity 99%


Russ Bookhamer - Operations Technician |

Date Mixed: 15-Feb-2023 Balance: B442140311

Manufactured under Restek
Registered Quality
Certificate #FM 8

Certified Reference Material Notes

Notes:

The date valid for unopened ampul stored in compliance with the recommended conditions. The identity, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Chemical identity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, LC/MS, RI, and/or melting point.

Products with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the compound in solution.

Isomeric compounds is reported as the sum of the isomers.

Values are rounded to the nearest whole number.

Uncertainty Value Notes:

Uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{\text{combined uncertainty}} = k \sqrt{u_{\text{gravimetric}}^2 + u_{\text{homogeneity}}^2 + u_{\text{storage stability}}^2 + u_{\text{shipping stability}}^2}$$

A coverage factor of 2, which gives a level of confidence of approximately 95%.

Sampled amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Notes:

This note is based upon gravimetric preparation using either a balance whose calibration has been verified daily or traceable weights, and/or dilutions with Class A glassware.

The unopened product, when stored in compliance with the recommended conditions, is guaranteed through the date displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.

If dissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555869

Lot No.: A0194702

Description : Custom Hexachlorocyclopentadiene Standard

Custom Hexachlorocyclopentadiene Standard 25,000 μ g/mL, Methanol,
1mL/ampul

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : February 28, 2026

Storage: 10°C or colder

Ship: Ambient

1 2 3 4 5 6 7 8 9 10 11 12 13 14 15 16 17

CERTIFICATE

| Component # | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) |
|-------------|---------------------------|---------|---------|--------|--------------------------------|
| 1 | Hexachlorocyclopentadiene | 77-47-4 | 0012019 | 99% | 25,008.0 μ g/mL |

Solvent: Methanol
CAS # 67-56-1
Purity 99%

Russ Bookhamer - Operations Technician I

Date Mixed: 15-Feb-2023 Balance: B442140311

Manufactured under Restek
Registered Quality
Certificate #FM 1

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17

Certified Reference Material Notes

Notes:

The date valid for unopened ampul stored in compliance with the recommended conditions. The purity, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, LC/MS, RI, and/or melting point.

Ampuls with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the compound in solution.

Isomeric compounds are reported as the sum of the isomers.

Values are rounded to the nearest whole number.

Uncertainty Value Notes:

Uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{\text{combined uncertainty}} = k \sqrt{u_{\text{gravimetric}}^2 + u_{\text{homogeneity}}^2 + u_{\text{storage stability}}^2 + u_{\text{shipping stability}}^2}$$

A coverage factor of 2, which gives a level of confidence of approximately 95%.

The stated amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure the maximum packaged amount can be sufficiently transferred.

Notes:

Preparation is based upon gravimetric preparation using either a balance whose calibration has been verified daily or traceable weights, and/or dilutions with Class A glassware.

:

The unopened product, when stored in compliance with the recommended conditions, is guaranteed through the stated expiration date displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with 1mL and 2mL ampuls packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom service. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, with complete instructions.

If dissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



PRODUCTOS
QUÍMICOS
MONTERREY, S.A. DE C.V.

MIRADOR 201, COL. MIRADOR
MONTERREY, N.L. MEXICO
CP 64070
TEL +52 81 13 52 57 57
www.pqm.com.mx

CERTIFICATE OF ANALYSIS

| | | | |
|------------------------|-----------------------------------|---------------|---------------------------------|
| PRODUCT : | SODIUM SULFATE CRYSTALS ANHYDROUS | | |
| QUALITY : | ACS (CODE RMB3375) | FORMULA : | Na ₂ SO ₄ |
| SPECIFICATION NUMBER : | 6399 | RELEASE DATE: | ABR/21/2023 |
| LOT NUMBER : | 313201 | | |

| TEST | SPECIFICATIONS | LOT VALUES |
|--|----------------|-------------|
| Assay (Na ₂ SO ₄) | Min. 99.0% | 99.7 % |
| pH of a 5% solution at 25°C | 5.2 - 9.2 | 6.1 |
| Insoluble matter | Max. 0.01% | 0.005 % |
| Loss on ignition | Max. 0.5% | 0.1 % |
| Chloride (Cl) | Max. 0.001% | <0.001 % |
| Nitrogen compounds (as N) | Max. 5 ppm | <5 ppm |
| Phosphate (PO ₄) | Max. 0.001% | <0.001 % |
| Heavy metals (as Pb) | Max. 5 ppm | <5 ppm |
| Iron (Fe) | Max. 0.001% | <0.001 % |
| Calcium (Ca) | Max. 0.01% | 0.002 % |
| Magnesium (Mg) | Max. 0.005% | 0.001 % |
| Potassium (K) | Max. 0.008% | 0.003 % |
| Extraction-concentration suitability | Passes test | Passes test |
| Appearance | Passes test | Passes test |
| Identification | Passes test | Passes test |
| Solubility and foreing matter | Passes test | Passes test |
| Retained on US Standard No. 10 sieve | Max. 1% | 0.1 % |
| Retained on US Standard No. 60 sieve | Min. 94% | 97.3 % |
| Through US Standard No. 60 sieve | Max. 5% | 2.5 % |
| Through US Standard No. 100 sieve | Max. 10% | 0.1 % |
| COMMENTS | | |

QC: PhC Irma Belmares

If you need further details, please call our factory or contact our local distributor.

Recd. by R3 on 7/29/23 [E 3551]

RC-02-01, Ed. 3



Certificate of Analysis

Sodium Hydroxide (Pellets)

Material: 0583
Grade: ACS GRADE
Batch Number: 23B1556310

Chemical Formula: NaOH
Molecular Weight: 40
CAS #: 1310-73-2
Appearance:
Pellets

Manufacture Date: 12/14/2022
Expiration Date: 12/31/2025
Storage: Room Temperature

| TEST | SPECIFICATION | ANALYSIS | DISPOSITION |
|--------------------|---------------|----------|-------------|
| Calcium | <= 0.005 % | <0.005 % | PASS |
| Chloride | <= 0.005 % | 0.002 % | PASS |
| Heavy Metals | <= 0.002 % | <0.002 % | PASS |
| Iron | <= 0.001 % | <0.001 % | PASS |
| Magnesium | <= 0.002 % | <0.002 % | PASS |
| Mercury | <= 0.1 ppm | <0.1 ppm | PASS |
| Nickel | <= 0.001 % | <0.001 % | PASS |
| Nitrogen Compounds | <= 0.001 % | <0.001 % | PASS |
| Phosphate | <= 0.001 % | <0.001 % | PASS |
| Potassium | <= 0.02 % | <0.02 % | PASS |
| Purity | >= 97.0 % | 99.2 % | PASS |
| Sodium Carbonate | <= 1.0 % | 0.5 % | PASS |
| Sulfate | <= 0.003 % | <0.003 % | PASS |

Internal ID #: 710

Signature

Additional Information

We certify that this batch conforms to the specifications listed.

Analysis may have been rounded to significant digits in specification limits.

This document has been electronically produced and is valid without a signature.

Product meets analytical specifications of the grades listed.

Leona Edwardson, Quality Control Sr. Manager - Solon
VWR Chemicals, LLC.
28600 Fountain Parkway, Solon OH 44139 USA

Acetone

BAKER RESI-ANALYZED® Reagent

For Organic Residue Analysis

avantor™



Material No.: 9254-03
Batch No.: 23H1462005
Manufactured Date: 2023-07-26
Expiration Date: 2026-07-25
Revision No.: 0

Certificate of Analysis

| Test | Specification | Result |
|---|---------------|-------------|
| Assay ((CH ₃) ₂ CO) (by GC, corrected for water) | ≥ 99.4 % | 99.7 % |
| Color (APHA) | ≤ 10 | 5 |
| Residue after Evaporation | ≤ 1.0 ppm | 0.3 ppm |
| Substances Reducing Permanganate | Passes Test | Passes Test |
| Titrable Acid (μeq/g) | ≤ 0.3 | 0.1 |
| Titrable Base (μeq/g) | ≤ 0.6 | < 0.1 |
| Water (H ₂ O) | ≤ 0.5 % | 0.3 % |
| FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL) | ≤ 5 | < 1 |
| ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL) | ≤ 10 | 1 |

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: USA

Packaging Site: Phillipsburg Mfg Ctr & DC

Received by AP on 5/23/24

E 3744

A handwritten signature in black ink, appearing to read "Ken Koehlein".

Ken Koehlein
Sr. Manager, Quality Assurance

Methylene Chloride
ULTRA RESI-ANALYZED
For Organic Residue Analysis
(dichloromethane)



Material No.: 9266-A4
Batch No.: 24E2462004
Manufactured Date: 2024-04-10
Expiration Date: 2025-07-10
Revision No.: 0

Certificate of Analysis

| Test | Specification | Result |
|--|---------------|----------|
| FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL) | ≤ 5 | 3 |
| ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL) | ≤ 10 | 3 |
| Assay (CH_2Cl_2) (by GC, exclusive of preservative, corrected for water) | ≥ 99.8 % | 100.0 % |
| Color (APHA) | ≤ 10 | 5 |
| Residue after Evaporation | ≤ 1.0 ppm | 0.1 ppm |
| Titrable Acid ($\mu\text{eq/g}$) | ≤ 0.3 | < 0.1 |
| Chloride (Cl) | ≤ 10 ppm | 5 ppm |
| Water (by KF, coulometric) | ≤ 0.02 % | < 0.01 % |

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC
Manufacturer source batch: MG24D10725

E 3768

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials, LLC

100 Matsonford Rd, Suite 200, Radnor, PA 19087, U.S.A. Phone 610.386.1700

Page 1 of 1

Acetone

BAKER RESI-ANALYZED® Reagent

For Organic Residue Analysis

avantor™



Material No.: 9254-03
Batch No.: 23H1462005
Manufactured Date: 2023-07-26
Expiration Date: 2026-07-25
Revision No.: 0

Certificate of Analysis

| Test | Specification | Result |
|---|---------------|-------------|
| Assay ((CH ₃) ₂ CO) (by GC, corrected for water) | ≥ 99.4 % | 99.7 % |
| Color (APHA) | ≤ 10 | 5 |
| Residue after Evaporation | ≤ 1.0 ppm | 0.3 ppm |
| Substances Reducing Permanganate | Passes Test | Passes Test |
| Titrable Acid (μeq/g) | ≤ 0.3 | 0.1 |
| Titrable Base (μeq/g) | ≤ 0.6 | < 0.1 |
| Water (H ₂ O) | ≤ 0.5 % | 0.3 % |
| FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL) | ≤ 5 | < 1 |
| ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL) | ≤ 10 | 1 |

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: USA

Packaging Site: Phillipsburg Mfg Ctr & DC

Rec'd by RP on 7/24/24

E 3480

A handwritten signature in black ink, appearing to read "Ken Koehlein".

Ken Koehlein
Sr. Manager, Quality Assurance

Methylene Chloride
ULTRA RESI-ANALYZED
For Organic Residue Analysis
(dichloromethane)



Material No.: 9266-A4
Batch No.: 24F1062004
Manufactured Date: 2024-04-15
Expiration Date: 2025-07-15
Revision No.: 0

Certificate of Analysis

| Test | Specification | Result |
|--|---------------|----------|
| FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL) | ≤ 5 | < 1 |
| ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL) | ≤ 10 | 7 |
| Assay (CH_2Cl_2) (by GC, exclusive of preservative, corrected for water) | ≥ 99.8 % | 100.0 % |
| Color (APHA) | ≤ 10 | 5 |
| Residue after Evaporation | ≤ 1.0 ppm | 0.1 ppm |
| Titrable Acid ($\mu\text{eq/g}$) | ≤ 0.3 | < 0.1 |
| Chloride (Cl) | ≤ 10 ppm | < 5 ppm |
| Water (by KF, coulometric) | ≤ 0.02 % | < 0.01 % |

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC
Manufacturer source batch: MG24D15750

E 3786

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials, LLC

100 Matsonford Rd, Suite 200, Radnor, PA 19087, U.S.A. Phone 610.386.1700

Page 1 of 1

Methylene Chloride
ULTRA RESI-ANALYZED
For Organic Residue Analysis
(dichloromethane)

avantor™



Material No.: 9266-A4
Batch No.: 24G0862022
Manufactured Date: 2024-06-05
Expiration Date: 2025-09-04
Revision No.: 0

Certificate of Analysis

| Test | Specification | Result |
|--|---------------|----------|
| FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL) | ≤ 5 | 3 |
| ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL) | ≤ 10 | 4 |
| Assay (CH ₂ Cl ₂) (by GC, exclusive of preservative, corrected for water) | ≥ 99.8 % | 100.0 % |
| Color (APHA) | ≤ 10 | 5 |
| Residue after Evaporation | ≤ 1.0 ppm | 0.3 ppm |
| Titrable Acid (μeq/g) | ≤ 0.3 | < 0.1 |
| Chloride (Cl) | ≤ 10 ppm | < 5 ppm |
| Water (by KF, coulometric) | ≤ 0.02 % | < 0.01 % |

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC
Manufacturer source batch: MG24F05012

E 3787

A handwritten signature of Jamie Croak.

Jamie Croak
Director Quality Operations, Bioscience Production

Sulfuric Acid
BAKER INSTRA-ANALYZED® Reagent
For Trace Metal Analysis
Low Selenium

M5037-38-3n-40
no



Material No.: 9673-33
Batch No.: 0000250349
Manufactured Date: 2019/12/17
Retest Date: 2024/12/15
Revision No: 1

Certificate of Analysis

| Test | Specification | Result |
|--|---------------|--------|
| ACS - Assay (H ₂ SO ₄) | 95.0 – 98.0 % | 96.5 |
| Appearance | Passes Test | PT |
| ACS - Color (APHA) | <= 10 | 5 |
| ACS - Residue after Ignition | <= 3 ppm | 1 |
| ACS - Substances Reducing Permanganate (as SO ₂) | <= 2 ppm | < 2 |
| Ammonium (NH ₄) | <= 1 ppm | < 1 |
| Chloride (Cl) | <= 0.1 ppm | < 0.1 |
| Nitrate (NO ₃) | <= 0.2 ppm | < 0.1 |
| Phosphate (PO ₄) | <= 0.5 ppm | < 0.1 |
| Trace Impurities - Aluminum (Al) | <= 30.0 ppb | 0.2 |
| Arsenic and Antimony (as As) | <= 4 ppb | < 2 |
| Trace Impurities - Barium (Ba) | <= 10.0 ppb | < 1.0 |
| Trace Impurities - Beryllium (Be) | <= 10.0 ppb | < 1.0 |
| Trace Impurities - Bismuth (Bi) | <= 10.0 ppb | < 1.0 |
| Trace Impurities - Boron (B) | <= 10.0 ppb | < 5.0 |
| Trace Impurities - Cadmium (Cd) | <= 2.0 ppb | < 0.3 |
| Trace Impurities - Calcium (Ca) | <= 50.0 ppb | 2.9 |
| Trace Impurities - Chromium (Cr) | <= 6.0 ppb | < 0.4 |
| Trace Impurities - Cobalt (Co) | <= 0.5 ppb | < 0.3 |
| Trace Impurities - Copper (Cu) | <= 1.0 ppb | < 0.1 |
| Trace Impurities - Gallium (Ga) | <= 10.0 ppb | < 1.0 |
| Trace Impurities - Germanium (Ge) | <= 10.0 ppb | < 10.0 |
| Trace Impurities - Gold (Au) | <= 10.0 ppb | < 0.2 |
| Heavy Metals (as Pb) | <= 500 ppb | < 100 |

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials, LLC


100 Matsonford Rd, Suite 200, Radnor, PA 19087. U.S.A. Phone: 610.386.1700

| Test | Specification | Result |
|------------------------------------|---------------|--------|
| Trace Impurities - Iron (Fe) | <= 50.0 ppb | 4.1 |
| Trace Impurities - Lead (Pb) | <= 0.5 ppb | < 0.5 |
| Trace Impurities - Lithium (Li) | <= 10.0 ppb | < 1.0 |
| Trace Impurities - Magnesium (Mg) | <= 7.0 ppb | 0.4 |
| Trace Impurities - Manganese (Mn) | <= 1.0 ppb | < 0.4 |
| Trace Impurities - Mercury (Hg) | <= 0.5 ppb | < 0.1 |
| Trace Impurities - Molybdenum (Mo) | <= 10.0 ppb | < 5.0 |
| Trace Impurities - Nickel (Ni) | <= 2.0 ppb | < 0.3 |
| Trace Impurities - Niobium (Nb) | <= 10.0 ppb | < 1.0 |
| Trace Impurities - Potassium (K) | <= 500.0 ppb | < 2.0 |
| Trace Impurities - Selenium (Se) | <= 50.0 ppb | 22.9 |
| Trace Impurities - Silicon (Si) | <= 100.0 ppb | < 10.0 |
| Trace Impurities - Silver (Ag) | <= 1.0 ppb | < 0.3 |
| Trace Impurities - Sodium (Na) | <= 500.0 ppb | 2.7 |
| Trace Impurities - Strontium (Sr) | <= 5.0 ppb | < 0.2 |
| Trace Impurities - Tantalum (Ta) | <= 10.0 ppb | < 5.0 |
| Trace Impurities - Thallium (Tl) | <= 20.0 ppb | < 5.0 |
| Trace Impurities - Tin (Sn) | <= 5.0 ppb | < 0.8 |
| Trace Impurities - Titanium (Ti) | <= 10.0 ppb | < 1.0 |
| Trace Impurities - Vanadium (V) | <= 10.0 ppb | < 1.0 |
| Trace Impurities - Zinc (Zn) | <= 5.0 ppb | 0.3 |
| Trace Impurities - Zirconium (Zr) | <= 10.0 ppb | < 1.0 |

For Laboratory, Research or Manufacturing Use

Country of Origin: US

Packaging Site: Phillipsburg Mfg Ctr & DC



Jamie Ethier
Vice President Global Quality

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials, LLC
100 Matsonford Rd, Suite 200, Radnor, PA 19087. U.S.A. Phone: 610.386.1700



5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received: _____

Certificate of Analysis

Rev 0

Page 1 of 1

Catalog No.: Lot No.: Storage: Solvent: Exp. Date: Description:
Z-110094-02 506889 ≤ -10 °C Methylene Chloride 7/25/2028 CLP Base/Neutral Surrogate Solution, 5,000 mg/L, 1 ml

| Compound | CAS No. | Purity (%) | Compound Lot No. | Concentration, mg/L |
|------------------------------------|-----------|------------|------------------|---------------------|
| 1,2-dichlorobenzene-d ₄ | 2199-69-1 | 99.7 | 247.29.3P | 5035 ± 28.02 |
| 2-fluorobiphenyl | 321-60-8 | 99.69 | 8.286.1.1P | 4999 ± 103.66 |
| nitrobenzene-d ₅ | 4165-60-0 | 99.67 | 7.9.3P | 4988 ± 27.32 |
| p-terphenyl-d ₁₄ | 1718-51-0 | 99.3 | 9.120.8P | 5005 ± 27.85 |

511494 } Y.P.
↓ } 08/11/2023
511498

*Not a certified value

Certified By: _____

Clint Tipton
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values
listed are determined gravimetrically.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis *gravimetric*



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No.: 555223

Description: Custom 8270 Plus Standard #1

Custom 8270 Plus Standard #1 1,000 μ g/mL, Methylene Chloride,
1mL/ampul

Container Size: 2 mL

Expiration Date: September 30, 2025

Handling: This product is photosensitive.

Lot No.: A0201940

Pkg Amt: > 1 mL

Storage: 10°C or colder

Ship: Ambient

511539

↓

511568

7.6
og119

C E R T I F I E D V A L U E S

| Component # | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|-------------|------------------------|-----------|------------|--------|--------------------------------|---|
| 1 | 3,3'-Dichlorobenzidine | 91-94-1 | S230321RSR | 99% | 1,001.0 μ g/mL | +/- 22.9799 |
| 2 | Atrazine | 1912-24-9 | 5FYWL | 99% | 1,010.0 μ g/mL | +/- 23.1865 |
| 3 | Benzidine | 92-87-5 | S221205RSR | 99% | 1,008.0 μ g/mL | +/- 23.1406 |
| 4 | epsilon-Caprolactam | 105-60-2 | I16X016 | 99% | 1,008.0 μ g/mL | +/- 23.1406 |

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Jennifer Pollino
Sam Moodler - Operations Tech I

Date Mixed: 13-Sep-2023 Balance: B345965662

REVIEWED

By Jennifer Pollino at 7:10 am, Sep 13, 2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/pECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{\text{combined uncertainty}} = k \sqrt{u_{\text{gravimetric}}^2 + u_{\text{homogeneity}}^2 + u_{\text{storage stability}}^2 + u_{\text{shipping stability}}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis *gravimetric*



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555223

Description : Custom 8270 Plus Standard #1

Custom 8270 Plus Standard #1 1,000 μ g/mL, Methylene Chloride,
1mL/ampul

Container Size : 2 mL

Expiration Date : September 30, 2025

Handling: This product is photosensitive.

Lot No.: A0201940

Pkg Amt: > 1 mL

Storage: 10°C or colder

Ship: Ambient

511539

↓
511568

7.6
} 09/11/19

C E R T I F I E D V A L U E S

| Component # | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|-------------|------------------------|-----------|------------|--------|--------------------------------|---|
| 1 | 3,3'-Dichlorobenzidine | 91-94-1 | S230321RSR | 99% | 1,001.0 μ g/mL | +/- 22.9799 |
| 2 | Atrazine | 1912-24-9 | 5FYWL | 99% | 1,010.0 μ g/mL | +/- 23.1865 |
| 3 | Benzidine | 92-87-5 | S221205RSR | 99% | 1,008.0 μ g/mL | +/- 23.1406 |
| 4 | epsilon-Caprolactam | 105-60-2 | I16X016 | 99% | 1,008.0 μ g/mL | +/- 23.1406 |

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Jennifer Pollino
Sam Moodler - Operations Tech I

Date Mixed: 13-Sep-2023 Balance: B345965662

REVIEWED

By Jennifer Pollino at 7:10 am, Sep 13, 2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/pECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{\text{combined uncertainty}} = k \sqrt{u_{\text{gravimetric}}^2 + u_{\text{homogeneity}}^2 + u_{\text{storage stability}}^2 + u_{\text{shipping stability}}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis *gravimetric*



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555223

Description : Custom 8270 Plus Standard #1

Custom 8270 Plus Standard #1 1,000 μ g/mL, Methylene Chloride,
1mL/ampul

Container Size : 2 mL

Expiration Date : September 30, 2025

Handling: This product is photosensitive.

Lot No.: A0201940

Pkg Amt: > 1 mL

Storage: 10°C or colder

Ship: Ambient

511539

↓

511568

7.6
og119

C E R T I F I E D V A L U E S

| Component # | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|-------------|------------------------|-----------|------------|--------|--------------------------------|---|
| 1 | 3,3'-Dichlorobenzidine | 91-94-1 | S230321RSR | 99% | 1,001.0 μ g/mL | +/- 22.9799 |
| 2 | Atrazine | 1912-24-9 | 5FYWL | 99% | 1,010.0 μ g/mL | +/- 23.1865 |
| 3 | Benzidine | 92-87-5 | S221205RSR | 99% | 1,008.0 μ g/mL | +/- 23.1406 |
| 4 | epsilon-Caprolactam | 105-60-2 | I16X016 | 99% | 1,008.0 μ g/mL | +/- 23.1406 |

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Jennifer Pollino
Sam Moodler - Operations Tech I

Date Mixed: 13-Sep-2023 Balance: B345965662

REVIEWED

By Jennifer Pollino at 7:10 am, Sep 13, 2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/pECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{\text{combined uncertainty}} = k \sqrt{u_{\text{gravimetric}}^2 + u_{\text{homogeneity}}^2 + u_{\text{storage stability}}^2 + u_{\text{shipping stability}}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis *gravimetric*



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No.: 555223

Description: Custom 8270 Plus Standard #1

Custom 8270 Plus Standard #1 1,000 μ g/mL, Methylene Chloride,
1mL/ampul

Container Size: 2 mL

Expiration Date: September 30, 2025

Handling: This product is photosensitive.

Lot No.: A0201940

Pkg Amt: > 1 mL

Storage: 10°C or colder

Ship: Ambient

511539

↓
511568

7.6
} 09/11
19

C E R T I F I E D V A L U E S

| Component # | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|-------------|------------------------|-----------|------------|--------|--------------------------------|---|
| 1 | 3,3'-Dichlorobenzidine | 91-94-1 | S230321RSR | 99% | 1,001.0 μ g/mL | +/- 22.9799 |
| 2 | Atrazine | 1912-24-9 | 5FYWL | 99% | 1,010.0 μ g/mL | +/- 23.1865 |
| 3 | Benzidine | 92-87-5 | S221205RSR | 99% | 1,008.0 μ g/mL | +/- 23.1406 |
| 4 | epsilon-Caprolactam | 105-60-2 | I16X016 | 99% | 1,008.0 μ g/mL | +/- 23.1406 |

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Jennifer Pollino
Sam Moodler - Operations Tech I

Date Mixed: 13-Sep-2023 Balance: B345965662

REVIEWED

By Jennifer Pollino at 7:10 am, Sep 13, 2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/pECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{\text{combined uncertainty}} = k \sqrt{u_{\text{gravimetric}}^2 + u_{\text{homogeneity}}^2 + u_{\text{storage stability}}^2 + u_{\text{shipping stability}}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309
www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31853

Lot No.: A0196453

Description : 1,4-dioxane

1,4-Dioxane 2,000 μ g/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : March 31, 2028

Storage: 0°C or colder

Ship: Ambient

SI1749
↓ { RC /
SI1794 } 11/30/23

C E R T I F I E D V A L U E S

| Elution Order | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|---------------|-------------|----------|----------|--------|-----------------------------|--|
| 1 | 1,4-Dioxane | 123-91-1 | SHBN3770 | 99% | 2,013.0 μ g/mL | +/- 25.0521 |

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride

CAS # 75-09-2

Purity 99%

Quality Confirmation Test

Column:

30m x 0.25mm x 0.25 μ m
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant flow 1.8 mL/min.

Temp. Program:

80°C (hold 0.1 min.) to 330°C
@ 9.6°C/min. (hold 2.86 min.)

Inj. Temp:

250°C

Det. Temp:

340°C

Det. Type:


FID

Split Vent:

100 mL/min.

Inj. Vol

1 μ L



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Sam Moodier
Sam Moodier - Operations Tech I

Date Mixed: 30-Mar-2023 Balance Serial #: B707717271

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 31-Mar-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/µECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{\text{combined uncertainty}} = k \sqrt{u_{\text{gravimetric}}^2 + u_{\text{homogeneity}}^2 + u_{\text{storage stability}}^2 + u_{\text{shipping stability}}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309
www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31853

Lot No.: A0196453

Description : 1,4-dioxane

1,4-Dioxane 2,000 μ g/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : March 31, 2028

Storage: 0°C or colder

Ship: Ambient

SI1749
↓ { RC /
SI1794 } 11/30/23

C E R T I F I E D V A L U E S

| Elution Order | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|---------------|-------------|----------|----------|--------|-----------------------------|--|
| 1 | 1,4-Dioxane | 123-91-1 | SHBN3770 | 99% | 2,013.0 μ g/mL | +/- 25.0521 |

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride

CAS # 75-09-2

Purity 99%

Quality Confirmation Test

Column:

30m x 0.25mm x 0.25 μ m
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant flow 1.8 mL/min.

Temp. Program:

80°C (hold 0.1 min.) to 330°C
@ 9.6°C/min. (hold 2.86 min.)

Inj. Temp:

250°C

Det. Temp:

340°C

Det. Type:


FID

Split Vent:

100 mL/min.

Inj. Vol

1 μ L



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Sam Moodier
Sam Moodier - Operations Tech I

Date Mixed: 30-Mar-2023 Balance Serial #: B707717271

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 31-Mar-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/µECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{\text{combined uncertainty}} = k \sqrt{u_{\text{gravimetric}}^2 + u_{\text{homogeneity}}^2 + u_{\text{storage stability}}^2 + u_{\text{shipping stability}}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309
www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31853

Lot No.: A0196453

Description : 1,4-dioxane

1,4-Dioxane 2,000 μ g/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : March 31, 2028

Storage: 0°C or colder

Ship: Ambient

SI1749
↓ { RC /
SI1794 } 11/30/23

C E R T I F I E D V A L U E S

| Elution Order | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|---------------|-------------|----------|----------|--------|-----------------------------|--|
| 1 | 1,4-Dioxane | 123-91-1 | SHBN3770 | 99% | 2,013.0 μ g/mL | +/- 25.0521 |

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride

CAS # 75-09-2

Purity 99%

Quality Confirmation Test

Column:

30m x 0.25mm x 0.25 μ m
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant flow 1.8 mL/min.

Temp. Program:

80°C (hold 0.1 min.) to 330°C
@ 9.6°C/min. (hold 2.86 min.)

Inj. Temp:

250°C

Det. Temp:

340°C

Det. Type:


FID

Split Vent:

100 mL/min.

Inj. Vol

1 μ L



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Sam Moodier
Sam Moodier - Operations Tech I

Date Mixed: 30-Mar-2023 Balance Serial #: B707717271

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 31-Mar-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/µECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{\text{combined uncertainty}} = k \sqrt{u_{\text{gravimetric}}^2 + u_{\text{homogeneity}}^2 + u_{\text{storage stability}}^2 + u_{\text{shipping stability}}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309
www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31853

Lot No.: A0196453

Description : 1,4-dioxane

1,4-Dioxane 2,000 μ g/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : March 31, 2028

Storage: 0°C or colder

Ship: Ambient

SI1749
↓ { RC /
SI1794 } 11/30/23

C E R T I F I E D V A L U E S

| Elution Order | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|---------------|-------------|----------|----------|--------|-----------------------------|--|
| 1 | 1,4-Dioxane | 123-91-1 | SHBN3770 | 99% | 2,013.0 μ g/mL | +/- 25.0521 |

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride

CAS # 75-09-2

Purity 99%

Quality Confirmation Test

Column:

30m x 0.25mm x 0.25 μ m
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant flow 1.8 mL/min.

Temp. Program:

80°C (hold 0.1 min.) to 330°C
@ 9.6°C/min. (hold 2.86 min.)

Inj. Temp:

250°C

Det. Temp:

340°C

Det. Type:


FID

Split Vent:

100 mL/min.

Inj. Vol

1 μ L



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Samuel Moodier
Sam Moodier - Operations Tech I

Date Mixed: 30-Mar-2023 Balance Serial #: B707717271

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 31-Mar-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/µECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{\text{combined uncertainty}} = k \sqrt{u_{\text{gravimetric}}^2 + u_{\text{homogeneity}}^2 + u_{\text{storage stability}}^2 + u_{\text{shipping stability}}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309
www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31853

Lot No.: A0196453

Description : 1,4-dioxane

1,4-Dioxane 2,000 μ g/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : March 31, 2028

Storage: 0°C or colder

Ship: Ambient

SI1749
↓ { RC /
SI1794 } 11/30/23

C E R T I F I E D V A L U E S

| Elution Order | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|---------------|-------------|----------|----------|--------|-----------------------------|--|
| 1 | 1,4-Dioxane | 123-91-1 | SHBN3770 | 99% | 2,013.0 μ g/mL | +/- 25.0521 |

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Quality Confirmation Test

Column:

30m x 0.25mm x 0.25 μ m
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant flow 1.8 mL/min.

Temp. Program:

80°C (hold 0.1 min.) to 330°C
@ 9.6°C/min. (hold 2.86 min.)

Inj. Temp:

250°C

Det. Temp:

340°C

Det. Type:


FID

Split Vent:

100 mL/min.

Inj. Vol

1 μ L



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Sam Moodier
Sam Moodier - Operations Tech I

Date Mixed: 30-Mar-2023 Balance Serial #: B707717271

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 31-Mar-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/µECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{\text{combined uncertainty}} = k \sqrt{u_{\text{gravimetric}}^2 + u_{\text{homogeneity}}^2 + u_{\text{storage stability}}^2 + u_{\text{shipping stability}}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL



ILAC-MRA
ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01



ILAC-MRA
ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

Certificate of Analysis *chromatographic plus*

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31206

Lot No.: A0201320

Description : SV Internal Standard Mix 2mg/ml

SV Internal Standard Mix 2mg/ml 2000 µg/ml, Methylene Chloride,
1mL/ampul

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : July 31, 2029

Storage: 10°C or colder

Handling: Sonication required. Mix is
photosensitive.

Ship: Ambient

S12013
↓
S12042 } 12/26/23
RC }

C E R T I F I E D V A L U E S

| Elution Order | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|---------------|------------------------|------------|----------|--------|--------------------------------|---|
| 1 | 1,4-Dichlorobenzene-d4 | 3855-82-1 | PR-30447 | 99% | 2,017.0 µg/mL | +/- 90.8469 |
| 2 | Naphthalene-d8 | 1146-65-2 | M-2180 | 99% | 2,011.3 µg/mL | +/- 90.5917 |
| 3 | Acenaphthene-d10 | 15067-26-2 | PR-33507 | 99% | 2,008.6 µg/mL | +/- 90.4685 |
| 4 | Phenanthrene-d10 | 1517-22-2 | PR-32303 | 99% | 2,019.4 µg/mL | +/- 90.9550 |
| 5 | Chrysene-d12 | 1719-03-5 | PR-32210 | 99% | 2,013.7 µg/mL | +/- 90.6968 |
| 6 | Perylene-d12 | 1520-96-3 | PR-33205 | 99% | 2,012.7 µg/mL | +/- 90.6517 |

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride

CAS # 75-09-2

Purity 99%

Quality Confirmation Test

Column:

30m x 0.25mm x 0.25 μ m
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant pressure 10 psi.

Temp. Program:

75°C (hold 1 min.) to 330°C
@ 20°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

330°C

Det. Type:


FID

Split Vent:

10 ml/min.

Inj. Vol


1 μ l



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Peter Robbins - Operations Technician I

Date Mixed: 23-Aug-2023 Balance Serial #: B345965662


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 25-Aug-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL



Certificate of Analysis *chromatographic plus*

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31206

Lot No.: A0201320

Description : SV Internal Standard Mix 2mg/ml

SV Internal Standard Mix 2mg/ml 2000 µg/ml, Methylene Chloride,
1mL/ampul

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : July 31, 2029

Storage: 10°C or colder

Handling: Sonication required. Mix is
photosensitive.

Ship: Ambient

S12013
↓
S12042 } 12/26/23
RC }

C E R T I F I E D V A L U E S

| Elution Order | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|---------------|------------------------|------------|----------|--------|--------------------------------|---|
| 1 | 1,4-Dichlorobenzene-d4 | 3855-82-1 | PR-30447 | 99% | 2,017.0 µg/mL | +/- 90.8469 |
| 2 | Naphthalene-d8 | 1146-65-2 | M-2180 | 99% | 2,011.3 µg/mL | +/- 90.5917 |
| 3 | Acenaphthene-d10 | 15067-26-2 | PR-33507 | 99% | 2,008.6 µg/mL | +/- 90.4685 |
| 4 | Phenanthrene-d10 | 1517-22-2 | PR-32303 | 99% | 2,019.4 µg/mL | +/- 90.9550 |
| 5 | Chrysene-d12 | 1719-03-5 | PR-32210 | 99% | 2,013.7 µg/mL | +/- 90.6968 |
| 6 | Perylene-d12 | 1520-96-3 | PR-33205 | 99% | 2,012.7 µg/mL | +/- 90.6517 |

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride

CAS # 75-09-2

Purity 99%

Quality Confirmation Test

Column:

30m x 0.25mm x 0.25 μ m
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant pressure 10 psi.

Temp. Program:

75°C (hold 1 min.) to 330°C
@ 20°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

330°C

Det. Type:


FID

Split Vent:

10 ml/min.

Inj. Vol


1 μ l



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Peter Robbins - Operations Technician I

Date Mixed: 23-Aug-2023 Balance Serial #: B345965662


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 25-Aug-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL



ILAC-MRA
ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01



ILAC-MRA
ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

Certificate of Analysis *chromatographic plus*

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31206

Lot No.: A0201320

Description : SV Internal Standard Mix 2mg/ml

SV Internal Standard Mix 2mg/ml 2000 µg/ml, Methylene Chloride,
1mL/ampul

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : July 31, 2029

Storage: 10°C or colder

Handling: Sonication required. Mix is
photosensitive.

Ship: Ambient

S12013
↓
S12042 } 12/26/23
RC }

C E R T I F I E D V A L U E S

| Elution Order | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|---------------|------------------------|------------|----------|--------|--------------------------------|---|
| 1 | 1,4-Dichlorobenzene-d4 | 3855-82-1 | PR-30447 | 99% | 2,017.0 µg/mL | +/- 90.8469 |
| 2 | Naphthalene-d8 | 1146-65-2 | M-2180 | 99% | 2,011.3 µg/mL | +/- 90.5917 |
| 3 | Acenaphthene-d10 | 15067-26-2 | PR-33507 | 99% | 2,008.6 µg/mL | +/- 90.4685 |
| 4 | Phenanthrene-d10 | 1517-22-2 | PR-32303 | 99% | 2,019.4 µg/mL | +/- 90.9550 |
| 5 | Chrysene-d12 | 1719-03-5 | PR-32210 | 99% | 2,013.7 µg/mL | +/- 90.6968 |
| 6 | Perylene-d12 | 1520-96-3 | PR-33205 | 99% | 2,012.7 µg/mL | +/- 90.6517 |

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride

CAS # 75-09-2

Purity 99%

Quality Confirmation Test

Column:

30m x 0.25mm x 0.25 μ m
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant pressure 10 psi.

Temp. Program:

75°C (hold 1 min.) to 330°C
@ 20°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

330°C

Det. Type:


FID

Split Vent:

10 ml/min.

Inj. Vol


1 μ l



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Peter Robbins - Operations Technician I

Date Mixed: 23-Aug-2023 Balance Serial #: B345965662


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 25-Aug-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received: _____

Certificate of Analysis

Rev 0

Page 1 of 1

| Catalog No.: Lot No.: | Storage: | Solvent: | Exp. Date: | Description: | |
|-----------------------|----------|--------------------|------------|---|---------------------|
| Z-110816-01 | 414127 | Methylene Chloride | 6/21/2025 | Custom 8270 Mix, 4-79, 1000 mg/L, 1 mL | |
| Compound | | CAS No. | Purity (%) | Compound Lot No. | Concentration, mg/L |
| atrazine | | 1912-24-9 | 99.5 | 337.7.3P | 997 ± 5.81 |
| benzidine | | 92-87-5 | 99.9 | 124.18.6.2P | 991.8 ± 5.77 |
| caprolactam | | 105-60-2 | 99.9 | 271.1.6P | 999 ± 5.82 |

512075 } RC
↓ } 02/01/24
512079 }

Manufactured by o2si smart solutions, Accredited to ISO 9001:2008 by NSF and ISO/IEC 17025:2005 (Certification No. 3031.01) and ISO Guide 34:2009 (Certification No. 3031.02) by A2LA

*Not a certified value

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetrically.

Certified By:

Shane Overcash
Chemist



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555224

Lot No.: A0207706

Description : Custom 8270 Plus Standard #2

Custom 8270 Plus Standard #2 1,000 μ g/mL, Methylene Chloride,
1mL/ampul

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : February 28, 2026

Storage: 10°C or colder

Ship: Ambient

S12082 }
↓ RC /
S12111 } 02/22/24


C E R T I F I E D V A L U E S

| Component # | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|-------------|----------------------------|----------|--------------|--------|--------------------------------|--|
| 1 | 1,2,4,5-Tetrachlorobenzene | 95-94-3 | MKCT9480 | 99% | 1,001.0 μ g/mL | +/- 29.424320 |
| 2 | Acetophenone | 98-86-2 | STBH8205 | 99% | 1,004.0 μ g/mL | +/- 29.512504 |
| 3 | Benzaldehyde | 100-52-7 | RD231129RSRA | 99% | 1,005.0 μ g/mL | +/- 29.541899 |
| 4 | Benzoic acid | 65-85-0 | MKCR2694 | 99% | 1,003.0 μ g/mL | +/- 29.483110 |
| 5 | Biphenyl | 92-52-4 | MKCL6515 | 99% | 1,006.0 μ g/mL | +/- 29.571294 |

Solvent: Methylene chloride

CAS # 75-09-2

Purity 99%


John Friedline - Operations Technician I

Date Mixed: 12-Feb-2024

Balance: B345965662

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555224

Lot No.: A0207706

Description : Custom 8270 Plus Standard #2

Custom 8270 Plus Standard #2 1,000 μ g/mL, Methylene Chloride,
1mL/ampul

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : February 28, 2026

Storage: 10°C or colder

Ship: Ambient

S12082 } RC/
↓ S12111 } 02/22/24


C E R T I F I E D V A L U E S

| Component # | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|-------------|----------------------------|----------|--------------|--------|--------------------------------|---|
| 1 | 1,2,4,5-Tetrachlorobenzene | 95-94-3 | MKCT9480 | 99% | 1,001.0 μ g/mL | +/- 29.424320 |
| 2 | Acetophenone | 98-86-2 | STBH8205 | 99% | 1,004.0 μ g/mL | +/- 29.512504 |
| 3 | Benzaldehyde | 100-52-7 | RD231129RSRA | 99% | 1,005.0 μ g/mL | +/- 29.541899 |
| 4 | Benzoic acid | 65-85-0 | MKCR2694 | 99% | 1,003.0 μ g/mL | +/- 29.483110 |
| 5 | Biphenyl | 92-52-4 | MKCL6515 | 99% | 1,006.0 μ g/mL | +/- 29.571294 |

Solvent: Methylene chloride

CAS # 75-09-2

Purity 99%


John Friedline - Operations Technician I

Date Mixed: 12-Feb-2024

Balance: B345965662

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555224

Lot No.: A0207706

Description : Custom 8270 Plus Standard #2

Custom 8270 Plus Standard #2 1,000 μ g/mL, Methylene Chloride,
1mL/ampul

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : February 28, 2026

Storage: 10°C or colder

Ship: Ambient

S12082 } RC/
↓ S12111 } 02/22/24


C E R T I F I E D V A L U E S

| Component # | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|-------------|----------------------------|----------|--------------|--------|--------------------------------|---|
| 1 | 1,2,4,5-Tetrachlorobenzene | 95-94-3 | MKCT9480 | 99% | 1,001.0 μ g/mL | +/- 29.424320 |
| 2 | Acetophenone | 98-86-2 | STBH8205 | 99% | 1,004.0 μ g/mL | +/- 29.512504 |
| 3 | Benzaldehyde | 100-52-7 | RD231129RSRA | 99% | 1,005.0 μ g/mL | +/- 29.541899 |
| 4 | Benzoic acid | 65-85-0 | MKCR2694 | 99% | 1,003.0 μ g/mL | +/- 29.483110 |
| 5 | Biphenyl | 92-52-4 | MKCL6515 | 99% | 1,006.0 μ g/mL | +/- 29.571294 |

Solvent: Methylene chloride

CAS # 75-09-2

Purity 99%


John Friedline - Operations Technician I

Date Mixed: 12-Feb-2024

Balance: B345965662

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL



Certificate of Analysis

gravimetric

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555224

Lot No.: A0207706

Description : Custom 8270 Plus Standard #2

Custom 8270 Plus Standard #2 1,000 μ g/mL, Methylene Chloride,
1mL/ampul

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : February 28, 2026

Storage: 10°C or colder

Ship: Ambient

S12082 }
↓ RC /
S12111 } 02/22/24

C E R T I F I E D V A L U E S

| Component # | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|-------------|----------------------------|----------|--------------|--------|--------------------------------|---|
| 1 | 1,2,4,5-Tetrachlorobenzene | 95-94-3 | MKCT9480 | 99% | 1,001.0 μ g/mL | +/- 29.424320 |
| 2 | Acetophenone | 98-86-2 | STBH8205 | 99% | 1,004.0 μ g/mL | +/- 29.512504 |
| 3 | Benzaldehyde | 100-52-7 | RD231129RSRA | 99% | 1,005.0 μ g/mL | +/- 29.541899 |
| 4 | Benzoic acid | 65-85-0 | MKCR2694 | 99% | 1,003.0 μ g/mL | +/- 29.483110 |
| 5 | Biphenyl | 92-52-4 | MKCL6515 | 99% | 1,006.0 μ g/mL | +/- 29.571294 |

Solvent: Methylene chloride

CAS # 75-09-2

Purity 99%

John Friedline - Operations Technician I

Date Mixed: 12-Feb-2024

Balance: B345965662

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL



Certificate of Analysis

gravimetric

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555224

Lot No.: A0207706

Description : Custom 8270 Plus Standard #2

Custom 8270 Plus Standard #2 1,000 μ g/mL, Methylene Chloride,
1mL/ampul

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : February 28, 2026

Storage: 10°C or colder

Ship: Ambient

S12082 } RC/
↓ S12111 } 02/22/24


C E R T I F I E D V A L U E S

| Component # | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|-------------|----------------------------|----------|--------------|--------|--------------------------------|---|
| 1 | 1,2,4,5-Tetrachlorobenzene | 95-94-3 | MKCT9480 | 99% | 1,001.0 μ g/mL | +/- 29.424320 |
| 2 | Acetophenone | 98-86-2 | STBH8205 | 99% | 1,004.0 μ g/mL | +/- 29.512504 |
| 3 | Benzaldehyde | 100-52-7 | RD231129RSRA | 99% | 1,005.0 μ g/mL | +/- 29.541899 |
| 4 | Benzoic acid | 65-85-0 | MKCR2694 | 99% | 1,003.0 μ g/mL | +/- 29.483110 |
| 5 | Biphenyl | 92-52-4 | MKCL6515 | 99% | 1,006.0 μ g/mL | +/- 29.571294 |

Solvent: Methylene chloride

CAS # 75-09-2

Purity 99%


John Friedline - Operations Technician I

Date Mixed: 12-Feb-2024

Balance: B345965662

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555224

Lot No.: A0207706

Description : Custom 8270 Plus Standard #2

Custom 8270 Plus Standard #2 1,000 μ g/mL, Methylene Chloride,
1mL/ampul

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : February 28, 2026

Storage: 10°C or colder

Ship: Ambient

S12082 } RC/
↓ S12111 } 02/22/24

C E R T I F I E D V A L U E S

| Component # | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|-------------|----------------------------|----------|--------------|--------|--------------------------------|---|
| 1 | 1,2,4,5-Tetrachlorobenzene | 95-94-3 | MKCT9480 | 99% | 1,001.0 μ g/mL | +/- 29.424320 |
| 2 | Acetophenone | 98-86-2 | STBH8205 | 99% | 1,004.0 μ g/mL | +/- 29.512504 |
| 3 | Benzaldehyde | 100-52-7 | RD231129RSRA | 99% | 1,005.0 μ g/mL | +/- 29.541899 |
| 4 | Benzoic acid | 65-85-0 | MKCR2694 | 99% | 1,003.0 μ g/mL | +/- 29.483110 |
| 5 | Biphenyl | 92-52-4 | MKCL6515 | 99% | 1,006.0 μ g/mL | +/- 29.571294 |

Solvent: Methylene chloride

CAS # 75-09-2

Purity 99%

John Friedline - Operations Technician I

Date Mixed: 12-Feb-2024

Balance: B345965662

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL



Certificate of Analysis

gravimetric

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555224

Lot No.: A0207706

Description : Custom 8270 Plus Standard #2

Custom 8270 Plus Standard #2 1,000 μ g/mL, Methylene Chloride,
1mL/ampul

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : February 28, 2026

Storage: 10°C or colder

Ship: Ambient

S12082 } RC/
↓ S12111 } 02/22/24

C E R T I F I E D V A L U E S

| Component # | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|-------------|----------------------------|----------|--------------|--------|--------------------------------|---|
| 1 | 1,2,4,5-Tetrachlorobenzene | 95-94-3 | MKCT9480 | 99% | 1,001.0 μ g/mL | +/- 29.424320 |
| 2 | Acetophenone | 98-86-2 | STBH8205 | 99% | 1,004.0 μ g/mL | +/- 29.512504 |
| 3 | Benzaldehyde | 100-52-7 | RD231129RSRA | 99% | 1,005.0 μ g/mL | +/- 29.541899 |
| 4 | Benzoic acid | 65-85-0 | MKCR2694 | 99% | 1,003.0 μ g/mL | +/- 29.483110 |
| 5 | Biphenyl | 92-52-4 | MKCL6515 | 99% | 1,006.0 μ g/mL | +/- 29.571294 |

Solvent: Methylene chloride

CAS # 75-09-2

Purity 99%

John Friedline - Operations Technician I

Date Mixed: 12-Feb-2024

Balance: B345965662

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555224

Lot No.: A0207706

Description : Custom 8270 Plus Standard #2

Custom 8270 Plus Standard #2 1,000 μ g/mL, Methylene Chloride,
1mL/ampul

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : February 28, 2026

Storage: 10°C or colder

Ship: Ambient

S12082 }
↓ RC /
S12111 } 02/22/24

C E R T I F I E D V A L U E S

| Component # | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|-------------|----------------------------|----------|--------------|--------|--------------------------------|--|
| 1 | 1,2,4,5-Tetrachlorobenzene | 95-94-3 | MKCT9480 | 99% | 1,001.0 μ g/mL | +/- 29.424320 |
| 2 | Acetophenone | 98-86-2 | STBH8205 | 99% | 1,004.0 μ g/mL | +/- 29.512504 |
| 3 | Benzaldehyde | 100-52-7 | RD231129RSRA | 99% | 1,005.0 μ g/mL | +/- 29.541899 |
| 4 | Benzoic acid | 65-85-0 | MKCR2694 | 99% | 1,003.0 μ g/mL | +/- 29.483110 |
| 5 | Biphenyl | 92-52-4 | MKCL6515 | 99% | 1,006.0 μ g/mL | +/- 29.571294 |

Solvent: Methylene chloride

CAS # 75-09-2

Purity 99%

John Friedline - Operations Technician I

Date Mixed: 12-Feb-2024

Balance: B345965662

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555224

Lot No.: A0207706

Description : Custom 8270 Plus Standard #2

Custom 8270 Plus Standard #2 1,000 μ g/mL, Methylene Chloride,
1mL/ampul

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : February 28, 2026

Storage: 10°C or colder

Ship: Ambient

S12082 } RC/
↓ S12111 } 02/22/24


C E R T I F I E D V A L U E S

| Component # | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|-------------|----------------------------|----------|--------------|--------|--------------------------------|---|
| 1 | 1,2,4,5-Tetrachlorobenzene | 95-94-3 | MKCT9480 | 99% | 1,001.0 μ g/mL | +/- 29.424320 |
| 2 | Acetophenone | 98-86-2 | STBH8205 | 99% | 1,004.0 μ g/mL | +/- 29.512504 |
| 3 | Benzaldehyde | 100-52-7 | RD231129RSRA | 99% | 1,005.0 μ g/mL | +/- 29.541899 |
| 4 | Benzoic acid | 65-85-0 | MKCR2694 | 99% | 1,003.0 μ g/mL | +/- 29.483110 |
| 5 | Biphenyl | 92-52-4 | MKCL6515 | 99% | 1,006.0 μ g/mL | +/- 29.571294 |

Solvent: Methylene chloride

CAS # 75-09-2

Purity 99%


John Friedline - Operations Technician I

Date Mixed: 12-Feb-2024

Balance: B345965662

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received: _____

Certificate of Analysis

Rev 0

Page 1 of 1

Catalog No.: Lot No.: Storage: Solvent: Exp. Date: Description:
Z-020223-01 454157 ≤ -10 °C P/T Methanol 6/10/2026 1,4-Dioxane Solution, 2000 mg/L,
1 mL

| Compound | CAS No. | Purity (%) | Compound Lot No. | Concentration, mg/L |
|-------------|----------|------------|------------------|---------------------|
| 1,4-dioxane | 123-91-1 | 100 | 223.1.3P | 1997 ± 57.08 |

512112 } RC/
↓
512116 } 03/08/24

*Not a certified value

Certified By:

Melissa Workoff
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values
listed are determined gravimetrically.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL



Certificate of Analysis

chromatographic plus

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31850

Lot No.: A0203726

Description : 8270 MegaMix®

8270 MegaMix® 500-1000 µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : April 30, 2025

Storage: 0°C or colder

Handling: Sonication required. Mix is photosensitive.

Ship: Ambient

512117 } RC/
↓ 03/18/24
512146

C E R T I F I E D V A L U E S

| Elution Order | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|---------------|------------------------------|----------|-------------|--------|--------------------------------|---|
| 1 | Pyridine | 110-86-1 | SHBP6240 | 99% | 1,001.6 µg/mL | +/- 36.4412 |
| 2 | N-Nitrosodimethylamine | 62-75-9 | 230209JLM | 99% | 1,005.9 µg/mL | +/- 36.5968 |
| 3 | Phenol | 108-95-2 | MKCK1120 | 99% | 1,003.3 µg/mL | +/- 36.5038 |
| 4 | Aniline | 62-53-3 | X22F726 | 99% | 1,005.8 µg/mL | +/- 36.5928 |
| 5 | Bis(2-chloroethyl)ether | 111-44-4 | SHBL6942 | 99% | 1,008.1 µg/mL | +/- 36.6776 |
| 6 | 2-Chlorophenol | 95-57-8 | STBJ3909 | 99% | 1,001.8 µg/mL | +/- 36.4492 |
| 7 | 1,3-Dichlorobenzene | 541-73-1 | BCCD5315 | 99% | 1,002.3 µg/mL | +/- 36.4654 |
| 8 | 1,4-Dichlorobenzene | 106-46-7 | MKBS7929V | 99% | 1,003.7 µg/mL | +/- 36.5159 |
| 9 | Benzyl alcohol | 100-51-6 | SHBK5469 | 99% | 1,008.7 µg/mL | +/- 36.6979 |
| 10 | 1,2-Dichlorobenzene | 95-50-1 | SHBN3835 | 99% | 1,000.3 µg/mL | +/- 36.3926 |
| 11 | 2-Methylphenol (o-cresol) | 95-48-7 | SHBN7598 | 99% | 1,003.5 µg/mL | +/- 36.5099 |
| 12 | 2,2'-oxybis(1-chloropropane) | 108-60-1 | 29-MAR-45-5 | 99% | 1,007.3 µg/mL | +/- 36.6493 |
| 13 | 3-Methylphenol (m-cresol) | 108-39-4 | STBJ0710 | 99% | 504.3 µg/mL | +/- 18.3500 |
| 14 | 4-Methylphenol (p-cresol) | 106-44-5 | SHBN3411 | 99% | 503.6 µg/mL | +/- 18.3237 |
| 15 | N-Nitroso-di-n-propylamine | 621-64-7 | N63MG | 99% | 1,008.3 µg/mL | +/- 36.6857 |
| 16 | Hexachloroethane | 67-72-1 | QTORH | 99% | 1,007.5 µg/mL | +/- 36.6554 |
| 17 | Nitrobenzene | 98-95-3 | 10224044 | 99% | 1,008.6 µg/mL | +/- 36.6938 |

| | | | | | | | | |
|----|---|-----------|-------------|-----|---------|-------|-----|---------|
| 18 | Isophorone | 78-59-1 | MKCC9506 | 99% | 1,005.9 | µg/mL | +/- | 36.5988 |
| 19 | 2-Nitrophenol | 88-75-5 | RP230710 | 99% | 1,003.2 | µg/mL | +/- | 36.4998 |
| 20 | 2,4-Dimethylphenol | 105-67-9 | XW5GK | 99% | 1,003.8 | µg/mL | +/- | 36.5200 |
| 21 | Bis(2-chloroethoxy)methane | 111-91-1 | 13670200 | 99% | 1,002.1 | µg/mL | +/- | 36.4573 |
| 22 | 2,4-Dichlorophenol | 120-83-2 | BCBZ6787 | 99% | 1,003.7 | µg/mL | +/- | 36.5180 |
| 23 | 1,2,4-Trichlorobenzene | 120-82-1 | SHBP5900 | 99% | 1,007.6 | µg/mL | +/- | 36.6574 |
| 24 | Naphthalene | 91-20-3 | STBL1057 | 99% | 1,008.3 | µg/mL | +/- | 36.6837 |
| 25 | 4-Chloroaniline | 106-47-8 | BCCJ3217 | 99% | 1,001.3 | µg/mL | +/- | 36.4290 |
| 26 | Hexachlorobutadiene | 87-68-3 | RP230823RSR | 98% | 1,008.3 | µg/mL | +/- | 36.6829 |
| 27 | 4-Chloro-3-methylphenol | 59-50-7 | BCCD4461 | 99% | 1,003.1 | µg/mL | +/- | 36.4937 |
| 28 | 2-Methylnaphthalene | 91-57-6 | STBK0259 | 96% | 1,001.9 | µg/mL | +/- | 36.4505 |
| 29 | 1-Methylnaphthalene | 90-12-0 | 5234.00-8 | 98% | 1,000.0 | µg/mL | +/- | 36.3838 |
| 30 | Hexachlorocyclopentadiene | 77-47-4 | 099063I14L | 98% | 1,008.5 | µg/mL | +/- | 36.6909 |
| 31 | 2,4,6-Trichlorophenol | 88-06-2 | STBJ5914 | 99% | 1,004.4 | µg/mL | +/- | 36.5442 |
| 32 | 2,4,5-Trichlorophenol | 95-95-4 | FHN01 | 98% | 1,001.9 | µg/mL | +/- | 36.4512 |
| 33 | 2-Chloronaphthalene | 91-58-7 | RPN7O | 99% | 1,001.1 | µg/mL | +/- | 36.4230 |
| 34 | 2-Nitroaniline | 88-74-4 | RP230531 | 99% | 1,002.9 | µg/mL | +/- | 36.4876 |
| 35 | 1,4-Dinitrobenzene | 100-25-4 | RP230816 | 99% | 1,005.7 | µg/mL | +/- | 36.5887 |
| 36 | Acenaphthylene | 208-96-8 | p06V | 98% | 1,009.5 | µg/mL | +/- | 36.7265 |
| 37 | 1,3-Dinitrobenzene | 99-65-0 | 1-DXX-24-1 | 99% | 1,004.4 | µg/mL | +/- | 36.5422 |
| 38 | Dimethylphthalate | 131-11-3 | 358221L17K | 99% | 1,005.9 | µg/mL | +/- | 36.5968 |
| 39 | 2,6-Dinitrotoluene | 606-20-2 | BCCG1833 | 99% | 1,003.2 | µg/mL | +/- | 36.4998 |
| 40 | 1,2-Dinitrobenzene | 528-29-0 | RP230428 | 99% | 1,002.2 | µg/mL | +/- | 36.4634 |
| 41 | Acenaphthene | 83-32-9 | MKCR7169 | 99% | 1,009.3 | µg/mL | +/- | 36.7221 |
| 42 | 3-Nitroaniline | 99-09-2 | RP230822RSR | 99% | 1,003.9 | µg/mL | +/- | 36.5240 |
| 43 | 2,4-Dinitrophenol | 51-28-5 | DR230417RSR | 99% | 1,002.0 | µg/mL | +/- | 36.4553 |
| 44 | Dibenzofuran | 132-64-9 | MKCD9952 | 99% | 1,006.7 | µg/mL | +/- | 36.6251 |
| 45 | 2,4-Dinitrotoluene | 121-14-2 | MKAA0690V | 99% | 1,003.8 | µg/mL | +/- | 36.5220 |
| 46 | 4-Nitrophenol | 100-02-7 | RP230627 | 99% | 1,002.3 | µg/mL | +/- | 36.4674 |
| 47 | 2,3,4,6-Tetrachlorophenol | 58-90-2 | PR-30126 | 99% | 1,008.7 | µg/mL | +/- | 36.6979 |
| 48 | 2,3,5,6-Tetrachlorophenol | 935-95-5 | RP230919 | 99% | 1,006.3 | µg/mL | +/- | 36.6130 |
| 49 | Fluorene | 86-73-7 | 10241100 | 99% | 1,008.3 | µg/mL | +/- | 36.6857 |
| 50 | 4-Chlorophenyl phenyl ether | 7005-72-3 | MKCT7248 | 99% | 1,003.8 | µg/mL | +/- | 36.5220 |
| 51 | Diethylphthalate | 84-66-2 | MKCD2547 | 99% | 1,008.6 | µg/mL | +/- | 36.6958 |
| 52 | 4-Nitroaniline | 100-01-6 | RP230111 | 99% | 1,001.1 | µg/mL | +/- | 36.4230 |
| 53 | 4,6-Dinitro-2-methylphenol (Dinitro-o-cresol) | 534-52-1 | 230718JLM | 99% | 1,002.0 | µg/mL | +/- | 36.4553 |

| | | | | | | | |
|----|----------------------------|----------|---------------|-----|---------|-------|-------------|
| 54 | Diphenylamine | 122-39-4 | MKCH1042 | 99% | 1,002.3 | µg/mL | +/- 36.4674 |
| 55 | Azobenzene | 103-33-3 | BCCK0887 | 99% | 1,005.8 | µg/mL | +/- 36.5928 |
| 56 | 4-Bromophenyl phenyl ether | 101-55-3 | STBH6361 | 99% | 1,003.0 | µg/mL | +/- 36.4917 |
| 57 | Hexachlorobenzene | 118-74-1 | 14821700 | 99% | 1,007.5 | µg/mL | +/- 36.6554 |
| 58 | Pentachlorophenol | 87-86-5 | RP230530RSR | 99% | 1,008.8 | µg/mL | +/- 36.7019 |
| 59 | Phenanthrene | 85-01-8 | MKCQ8876 | 99% | 1,008.4 | µg/mL | +/- 36.6877 |
| 60 | Anthracene | 120-12-7 | MKCR0570 | 99% | 1,009.0 | µg/mL | +/- 36.7100 |
| 61 | Carbazole | 86-74-8 | 14351100 | 99% | 1,000.9 | µg/mL | +/- 36.4149 |
| 62 | Di-n-butylphthalate | 84-74-2 | MKCN4337 | 99% | 1,007.6 | µg/mL | +/- 36.6595 |
| 63 | Fluoranthene | 206-44-0 | MKCQ4728 | 99% | 1,009.6 | µg/mL | +/- 36.7302 |
| 64 | Pyrene | 129-00-0 | BCCG8479 | 98% | 1,007.2 | µg/mL | +/- 36.6453 |
| 65 | Benzyl butyl phthalate | 85-68-7 | X12I018 | 99% | 1,002.1 | µg/mL | +/- 36.4573 |
| 66 | Bis(2-ethylhexyl)adipate | 103-23-1 | MKCM1988 | 99% | 1,005.2 | µg/mL | +/- 36.5705 |
| 67 | Benz(a)anthracene | 56-55-3 | I220012022BAA | 99% | 1,002.2 | µg/mL | +/- 36.4614 |
| 68 | Chrysene | 218-01-9 | RP230601 | 99% | 1,008.3 | µg/mL | +/- 36.6837 |
| 69 | Bis(2-ethylhexyl)phthalate | 117-81-7 | MKCQ3468 | 99% | 1,001.8 | µg/mL | +/- 36.4472 |
| 70 | Di-n-octyl phthalate | 117-84-0 | 14382700 | 99% | 1,006.0 | µg/mL | +/- 36.6008 |
| 71 | Benzo(b)fluoranthene | 205-99-2 | 012013B | 99% | 1,002.8 | µg/mL | +/- 36.4836 |
| 72 | Benzo(k)fluoranthene | 207-08-9 | 012022K | 99% | 1,003.0 | µg/mL | +/- 36.4917 |
| 73 | Benzo(a)pyrene | 50-32-8 | P54915-0703 | 99% | 1,002.3 | µg/mL | +/- 36.4674 |
| 74 | Indeno(1,2,3-cd)pyrene | 193-39-5 | 12-JKL-118-9 | 97% | 1,009.4 | µg/mL | +/- 36.7243 |
| 75 | Dibenz(a,h)anthracene | 53-70-3 | 2-ASA-59-1 | 99% | 1,007.6 | µg/mL | +/- 36.6595 |
| 76 | Benzo(g,h,i)perylene | 191-24-2 | RP231003RSR | 99% | 1,002.9 | µg/mL | +/- 36.4876 |

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride

CAS # 75-09-2

Purity 99%



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL



Certificate of Analysis

chromatographic plus

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31850

Lot No.: A0203726

Description : 8270 MegaMix®

8270 MegaMix® 500-1000 µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : April 30, 2025

Storage: 0°C or colder

Handling: Sonication required. Mix is photosensitive.

Ship: Ambient

512117 } RC/
↓ } 03/18/24
512146

C E R T I F I E D V A L U E S

| Elution Order | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|---------------|------------------------------|----------|-------------|--------|--------------------------------|---|
| 1 | Pyridine | 110-86-1 | SHBP6240 | 99% | 1,001.6 µg/mL | +/- 36.4412 |
| 2 | N-Nitrosodimethylamine | 62-75-9 | 230209JLM | 99% | 1,005.9 µg/mL | +/- 36.5968 |
| 3 | Phenol | 108-95-2 | MKCK1120 | 99% | 1,003.3 µg/mL | +/- 36.5038 |
| 4 | Aniline | 62-53-3 | X22F726 | 99% | 1,005.8 µg/mL | +/- 36.5928 |
| 5 | Bis(2-chloroethyl)ether | 111-44-4 | SHBL6942 | 99% | 1,008.1 µg/mL | +/- 36.6776 |
| 6 | 2-Chlorophenol | 95-57-8 | STBJ3909 | 99% | 1,001.8 µg/mL | +/- 36.4492 |
| 7 | 1,3-Dichlorobenzene | 541-73-1 | BCCD5315 | 99% | 1,002.3 µg/mL | +/- 36.4654 |
| 8 | 1,4-Dichlorobenzene | 106-46-7 | MKBS7929V | 99% | 1,003.7 µg/mL | +/- 36.5159 |
| 9 | Benzyl alcohol | 100-51-6 | SHBK5469 | 99% | 1,008.7 µg/mL | +/- 36.6979 |
| 10 | 1,2-Dichlorobenzene | 95-50-1 | SHBN3835 | 99% | 1,000.3 µg/mL | +/- 36.3926 |
| 11 | 2-Methylphenol (o-cresol) | 95-48-7 | SHBN7598 | 99% | 1,003.5 µg/mL | +/- 36.5099 |
| 12 | 2,2'-oxybis(1-chloropropane) | 108-60-1 | 29-MAR-45-5 | 99% | 1,007.3 µg/mL | +/- 36.6493 |
| 13 | 3-Methylphenol (m-cresol) | 108-39-4 | STBJ0710 | 99% | 504.3 µg/mL | +/- 18.3500 |
| 14 | 4-Methylphenol (p-cresol) | 106-44-5 | SHBN3411 | 99% | 503.6 µg/mL | +/- 18.3237 |
| 15 | N-Nitroso-di-n-propylamine | 621-64-7 | N63MG | 99% | 1,008.3 µg/mL | +/- 36.6857 |
| 16 | Hexachloroethane | 67-72-1 | QTORH | 99% | 1,007.5 µg/mL | +/- 36.6554 |
| 17 | Nitrobenzene | 98-95-3 | 10224044 | 99% | 1,008.6 µg/mL | +/- 36.6938 |

| | | | | | | | | |
|----|---|-----------|-------------|-----|---------|-------|-----|---------|
| 18 | Isophorone | 78-59-1 | MKCC9506 | 99% | 1,005.9 | µg/mL | +/- | 36.5988 |
| 19 | 2-Nitrophenol | 88-75-5 | RP230710 | 99% | 1,003.2 | µg/mL | +/- | 36.4998 |
| 20 | 2,4-Dimethylphenol | 105-67-9 | XW5GK | 99% | 1,003.8 | µg/mL | +/- | 36.5200 |
| 21 | Bis(2-chloroethoxy)methane | 111-91-1 | 13670200 | 99% | 1,002.1 | µg/mL | +/- | 36.4573 |
| 22 | 2,4-Dichlorophenol | 120-83-2 | BCBZ6787 | 99% | 1,003.7 | µg/mL | +/- | 36.5180 |
| 23 | 1,2,4-Trichlorobenzene | 120-82-1 | SHBP5900 | 99% | 1,007.6 | µg/mL | +/- | 36.6574 |
| 24 | Naphthalene | 91-20-3 | STBL1057 | 99% | 1,008.3 | µg/mL | +/- | 36.6837 |
| 25 | 4-Chloroaniline | 106-47-8 | BCCJ3217 | 99% | 1,001.3 | µg/mL | +/- | 36.4290 |
| 26 | Hexachlorobutadiene | 87-68-3 | RP230823RSR | 98% | 1,008.3 | µg/mL | +/- | 36.6829 |
| 27 | 4-Chloro-3-methylphenol | 59-50-7 | BCCD4461 | 99% | 1,003.1 | µg/mL | +/- | 36.4937 |
| 28 | 2-Methylnaphthalene | 91-57-6 | STBK0259 | 96% | 1,001.9 | µg/mL | +/- | 36.4505 |
| 29 | 1-Methylnaphthalene | 90-12-0 | 5234.00-8 | 98% | 1,000.0 | µg/mL | +/- | 36.3838 |
| 30 | Hexachlorocyclopentadiene | 77-47-4 | 099063I14L | 98% | 1,008.5 | µg/mL | +/- | 36.6909 |
| 31 | 2,4,6-Trichlorophenol | 88-06-2 | STBJ5914 | 99% | 1,004.4 | µg/mL | +/- | 36.5442 |
| 32 | 2,4,5-Trichlorophenol | 95-95-4 | FHN01 | 98% | 1,001.9 | µg/mL | +/- | 36.4512 |
| 33 | 2-Chloronaphthalene | 91-58-7 | RPN7O | 99% | 1,001.1 | µg/mL | +/- | 36.4230 |
| 34 | 2-Nitroaniline | 88-74-4 | RP230531 | 99% | 1,002.9 | µg/mL | +/- | 36.4876 |
| 35 | 1,4-Dinitrobenzene | 100-25-4 | RP230816 | 99% | 1,005.7 | µg/mL | +/- | 36.5887 |
| 36 | Acenaphthylene | 208-96-8 | p06V | 98% | 1,009.5 | µg/mL | +/- | 36.7265 |
| 37 | 1,3-Dinitrobenzene | 99-65-0 | 1-DXX-24-1 | 99% | 1,004.4 | µg/mL | +/- | 36.5422 |
| 38 | Dimethylphthalate | 131-11-3 | 358221L17K | 99% | 1,005.9 | µg/mL | +/- | 36.5968 |
| 39 | 2,6-Dinitrotoluene | 606-20-2 | BCCG1833 | 99% | 1,003.2 | µg/mL | +/- | 36.4998 |
| 40 | 1,2-Dinitrobenzene | 528-29-0 | RP230428 | 99% | 1,002.2 | µg/mL | +/- | 36.4634 |
| 41 | Acenaphthene | 83-32-9 | MKCR7169 | 99% | 1,009.3 | µg/mL | +/- | 36.7221 |
| 42 | 3-Nitroaniline | 99-09-2 | RP230822RSR | 99% | 1,003.9 | µg/mL | +/- | 36.5240 |
| 43 | 2,4-Dinitrophenol | 51-28-5 | DR230417RSR | 99% | 1,002.0 | µg/mL | +/- | 36.4553 |
| 44 | Dibenzofuran | 132-64-9 | MKCD9952 | 99% | 1,006.7 | µg/mL | +/- | 36.6251 |
| 45 | 2,4-Dinitrotoluene | 121-14-2 | MKAA0690V | 99% | 1,003.8 | µg/mL | +/- | 36.5220 |
| 46 | 4-Nitrophenol | 100-02-7 | RP230627 | 99% | 1,002.3 | µg/mL | +/- | 36.4674 |
| 47 | 2,3,4,6-Tetrachlorophenol | 58-90-2 | PR-30126 | 99% | 1,008.7 | µg/mL | +/- | 36.6979 |
| 48 | 2,3,5,6-Tetrachlorophenol | 935-95-5 | RP230919 | 99% | 1,006.3 | µg/mL | +/- | 36.6130 |
| 49 | Fluorene | 86-73-7 | 10241100 | 99% | 1,008.3 | µg/mL | +/- | 36.6857 |
| 50 | 4-Chlorophenyl phenyl ether | 7005-72-3 | MKCT7248 | 99% | 1,003.8 | µg/mL | +/- | 36.5220 |
| 51 | Diethylphthalate | 84-66-2 | MKCD2547 | 99% | 1,008.6 | µg/mL | +/- | 36.6958 |
| 52 | 4-Nitroaniline | 100-01-6 | RP230111 | 99% | 1,001.1 | µg/mL | +/- | 36.4230 |
| 53 | 4,6-Dinitro-2-methylphenol (Dinitro-o-cresol) | 534-52-1 | 230718JLM | 99% | 1,002.0 | µg/mL | +/- | 36.4553 |

| | | | | | | | |
|----|----------------------------|----------|---------------|-----|---------|-------|-------------|
| 54 | Diphenylamine | 122-39-4 | MKCH1042 | 99% | 1,002.3 | µg/mL | +/- 36.4674 |
| 55 | Azobenzene | 103-33-3 | BCCK0887 | 99% | 1,005.8 | µg/mL | +/- 36.5928 |
| 56 | 4-Bromophenyl phenyl ether | 101-55-3 | STBH6361 | 99% | 1,003.0 | µg/mL | +/- 36.4917 |
| 57 | Hexachlorobenzene | 118-74-1 | 14821700 | 99% | 1,007.5 | µg/mL | +/- 36.6554 |
| 58 | Pentachlorophenol | 87-86-5 | RP230530RSR | 99% | 1,008.8 | µg/mL | +/- 36.7019 |
| 59 | Phenanthrene | 85-01-8 | MKCQ8876 | 99% | 1,008.4 | µg/mL | +/- 36.6877 |
| 60 | Anthracene | 120-12-7 | MKCR0570 | 99% | 1,009.0 | µg/mL | +/- 36.7100 |
| 61 | Carbazole | 86-74-8 | 14351100 | 99% | 1,000.9 | µg/mL | +/- 36.4149 |
| 62 | Di-n-butylphthalate | 84-74-2 | MKCN4337 | 99% | 1,007.6 | µg/mL | +/- 36.6595 |
| 63 | Fluoranthene | 206-44-0 | MKCQ4728 | 99% | 1,009.6 | µg/mL | +/- 36.7302 |
| 64 | Pyrene | 129-00-0 | BCCG8479 | 98% | 1,007.2 | µg/mL | +/- 36.6453 |
| 65 | Benzyl butyl phthalate | 85-68-7 | X12I018 | 99% | 1,002.1 | µg/mL | +/- 36.4573 |
| 66 | Bis(2-ethylhexyl)adipate | 103-23-1 | MKCM1988 | 99% | 1,005.2 | µg/mL | +/- 36.5705 |
| 67 | Benz(a)anthracene | 56-55-3 | I220012022BAA | 99% | 1,002.2 | µg/mL | +/- 36.4614 |
| 68 | Chrysene | 218-01-9 | RP230601 | 99% | 1,008.3 | µg/mL | +/- 36.6837 |
| 69 | Bis(2-ethylhexyl)phthalate | 117-81-7 | MKCQ3468 | 99% | 1,001.8 | µg/mL | +/- 36.4472 |
| 70 | Di-n-octyl phthalate | 117-84-0 | 14382700 | 99% | 1,006.0 | µg/mL | +/- 36.6008 |
| 71 | Benzo(b)fluoranthene | 205-99-2 | 012013B | 99% | 1,002.8 | µg/mL | +/- 36.4836 |
| 72 | Benzo(k)fluoranthene | 207-08-9 | 012022K | 99% | 1,003.0 | µg/mL | +/- 36.4917 |
| 73 | Benzo(a)pyrene | 50-32-8 | P54915-0703 | 99% | 1,002.3 | µg/mL | +/- 36.4674 |
| 74 | Indeno(1,2,3-cd)pyrene | 193-39-5 | 12-JKL-118-9 | 97% | 1,009.4 | µg/mL | +/- 36.7243 |
| 75 | Dibenz(a,h)anthracene | 53-70-3 | 2-ASA-59-1 | 99% | 1,007.6 | µg/mL | +/- 36.6595 |
| 76 | Benzo(g,h,i)perylene | 191-24-2 | RP231003RSR | 99% | 1,002.9 | µg/mL | +/- 36.4876 |

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride

CAS # 75-09-2

Purity 99%



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL



Certificate of Analysis

chromatographic plus

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31850

Lot No.: A0203726

Description : 8270 MegaMix®

8270 MegaMix® 500-1000 µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : April 30, 2025

Storage: 0°C or colder

Handling: Sonication required. Mix is photosensitive.

Ship: Ambient

512117 } RC/
↓ 03/18/24
512146

C E R T I F I E D V A L U E S

| Elution Order | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|---------------|------------------------------|----------|-------------|--------|--------------------------------|---|
| 1 | Pyridine | 110-86-1 | SHBP6240 | 99% | 1,001.6 µg/mL | +/- 36.4412 |
| 2 | N-Nitrosodimethylamine | 62-75-9 | 230209JLM | 99% | 1,005.9 µg/mL | +/- 36.5968 |
| 3 | Phenol | 108-95-2 | MKCK1120 | 99% | 1,003.3 µg/mL | +/- 36.5038 |
| 4 | Aniline | 62-53-3 | X22F726 | 99% | 1,005.8 µg/mL | +/- 36.5928 |
| 5 | Bis(2-chloroethyl)ether | 111-44-4 | SHBL6942 | 99% | 1,008.1 µg/mL | +/- 36.6776 |
| 6 | 2-Chlorophenol | 95-57-8 | STBJ3909 | 99% | 1,001.8 µg/mL | +/- 36.4492 |
| 7 | 1,3-Dichlorobenzene | 541-73-1 | BCCD5315 | 99% | 1,002.3 µg/mL | +/- 36.4654 |
| 8 | 1,4-Dichlorobenzene | 106-46-7 | MKBS7929V | 99% | 1,003.7 µg/mL | +/- 36.5159 |
| 9 | Benzyl alcohol | 100-51-6 | SHBK5469 | 99% | 1,008.7 µg/mL | +/- 36.6979 |
| 10 | 1,2-Dichlorobenzene | 95-50-1 | SHBN3835 | 99% | 1,000.3 µg/mL | +/- 36.3926 |
| 11 | 2-Methylphenol (o-cresol) | 95-48-7 | SHBN7598 | 99% | 1,003.5 µg/mL | +/- 36.5099 |
| 12 | 2,2'-oxybis(1-chloropropane) | 108-60-1 | 29-MAR-45-5 | 99% | 1,007.3 µg/mL | +/- 36.6493 |
| 13 | 3-Methylphenol (m-cresol) | 108-39-4 | STBJ0710 | 99% | 504.3 µg/mL | +/- 18.3500 |
| 14 | 4-Methylphenol (p-cresol) | 106-44-5 | SHBN3411 | 99% | 503.6 µg/mL | +/- 18.3237 |
| 15 | N-Nitroso-di-n-propylamine | 621-64-7 | N63MG | 99% | 1,008.3 µg/mL | +/- 36.6857 |
| 16 | Hexachloroethane | 67-72-1 | QTORH | 99% | 1,007.5 µg/mL | +/- 36.6554 |
| 17 | Nitrobenzene | 98-95-3 | 10224044 | 99% | 1,008.6 µg/mL | +/- 36.6938 |

| | | | | | | | | |
|----|---|-----------|-------------|-----|---------|-------|-----|---------|
| 18 | Isophorone | 78-59-1 | MKCC9506 | 99% | 1,005.9 | µg/mL | +/- | 36.5988 |
| 19 | 2-Nitrophenol | 88-75-5 | RP230710 | 99% | 1,003.2 | µg/mL | +/- | 36.4998 |
| 20 | 2,4-Dimethylphenol | 105-67-9 | XW5GK | 99% | 1,003.8 | µg/mL | +/- | 36.5200 |
| 21 | Bis(2-chloroethoxy)methane | 111-91-1 | 13670200 | 99% | 1,002.1 | µg/mL | +/- | 36.4573 |
| 22 | 2,4-Dichlorophenol | 120-83-2 | BCBZ6787 | 99% | 1,003.7 | µg/mL | +/- | 36.5180 |
| 23 | 1,2,4-Trichlorobenzene | 120-82-1 | SHBP5900 | 99% | 1,007.6 | µg/mL | +/- | 36.6574 |
| 24 | Naphthalene | 91-20-3 | STBL1057 | 99% | 1,008.3 | µg/mL | +/- | 36.6837 |
| 25 | 4-Chloroaniline | 106-47-8 | BCCJ3217 | 99% | 1,001.3 | µg/mL | +/- | 36.4290 |
| 26 | Hexachlorobutadiene | 87-68-3 | RP230823RSR | 98% | 1,008.3 | µg/mL | +/- | 36.6829 |
| 27 | 4-Chloro-3-methylphenol | 59-50-7 | BCCD4461 | 99% | 1,003.1 | µg/mL | +/- | 36.4937 |
| 28 | 2-Methylnaphthalene | 91-57-6 | STBK0259 | 96% | 1,001.9 | µg/mL | +/- | 36.4505 |
| 29 | 1-Methylnaphthalene | 90-12-0 | 5234.00-8 | 98% | 1,000.0 | µg/mL | +/- | 36.3838 |
| 30 | Hexachlorocyclopentadiene | 77-47-4 | 099063I14L | 98% | 1,008.5 | µg/mL | +/- | 36.6909 |
| 31 | 2,4,6-Trichlorophenol | 88-06-2 | STBJ5914 | 99% | 1,004.4 | µg/mL | +/- | 36.5442 |
| 32 | 2,4,5-Trichlorophenol | 95-95-4 | FHN01 | 98% | 1,001.9 | µg/mL | +/- | 36.4512 |
| 33 | 2-Chloronaphthalene | 91-58-7 | RPN7O | 99% | 1,001.1 | µg/mL | +/- | 36.4230 |
| 34 | 2-Nitroaniline | 88-74-4 | RP230531 | 99% | 1,002.9 | µg/mL | +/- | 36.4876 |
| 35 | 1,4-Dinitrobenzene | 100-25-4 | RP230816 | 99% | 1,005.7 | µg/mL | +/- | 36.5887 |
| 36 | Acenaphthylene | 208-96-8 | p06V | 98% | 1,009.5 | µg/mL | +/- | 36.7265 |
| 37 | 1,3-Dinitrobenzene | 99-65-0 | 1-DXX-24-1 | 99% | 1,004.4 | µg/mL | +/- | 36.5422 |
| 38 | Dimethylphthalate | 131-11-3 | 358221L17K | 99% | 1,005.9 | µg/mL | +/- | 36.5968 |
| 39 | 2,6-Dinitrotoluene | 606-20-2 | BCCG1833 | 99% | 1,003.2 | µg/mL | +/- | 36.4998 |
| 40 | 1,2-Dinitrobenzene | 528-29-0 | RP230428 | 99% | 1,002.2 | µg/mL | +/- | 36.4634 |
| 41 | Acenaphthene | 83-32-9 | MKCR7169 | 99% | 1,009.3 | µg/mL | +/- | 36.7221 |
| 42 | 3-Nitroaniline | 99-09-2 | RP230822RSR | 99% | 1,003.9 | µg/mL | +/- | 36.5240 |
| 43 | 2,4-Dinitrophenol | 51-28-5 | DR230417RSR | 99% | 1,002.0 | µg/mL | +/- | 36.4553 |
| 44 | Dibenzofuran | 132-64-9 | MKCD9952 | 99% | 1,006.7 | µg/mL | +/- | 36.6251 |
| 45 | 2,4-Dinitrotoluene | 121-14-2 | MKAA0690V | 99% | 1,003.8 | µg/mL | +/- | 36.5220 |
| 46 | 4-Nitrophenol | 100-02-7 | RP230627 | 99% | 1,002.3 | µg/mL | +/- | 36.4674 |
| 47 | 2,3,4,6-Tetrachlorophenol | 58-90-2 | PR-30126 | 99% | 1,008.7 | µg/mL | +/- | 36.6979 |
| 48 | 2,3,5,6-Tetrachlorophenol | 935-95-5 | RP230919 | 99% | 1,006.3 | µg/mL | +/- | 36.6130 |
| 49 | Fluorene | 86-73-7 | 10241100 | 99% | 1,008.3 | µg/mL | +/- | 36.6857 |
| 50 | 4-Chlorophenyl phenyl ether | 7005-72-3 | MKCT7248 | 99% | 1,003.8 | µg/mL | +/- | 36.5220 |
| 51 | Diethylphthalate | 84-66-2 | MKCD2547 | 99% | 1,008.6 | µg/mL | +/- | 36.6958 |
| 52 | 4-Nitroaniline | 100-01-6 | RP230111 | 99% | 1,001.1 | µg/mL | +/- | 36.4230 |
| 53 | 4,6-Dinitro-2-methylphenol (Dinitro-o-cresol) | 534-52-1 | 230718JLM | 99% | 1,002.0 | µg/mL | +/- | 36.4553 |

| | | | | | | | |
|----|----------------------------|----------|---------------|-----|---------|-------|-------------|
| 54 | Diphenylamine | 122-39-4 | MKCH1042 | 99% | 1,002.3 | µg/mL | +/- 36.4674 |
| 55 | Azobenzene | 103-33-3 | BCCK0887 | 99% | 1,005.8 | µg/mL | +/- 36.5928 |
| 56 | 4-Bromophenyl phenyl ether | 101-55-3 | STBH6361 | 99% | 1,003.0 | µg/mL | +/- 36.4917 |
| 57 | Hexachlorobenzene | 118-74-1 | 14821700 | 99% | 1,007.5 | µg/mL | +/- 36.6554 |
| 58 | Pentachlorophenol | 87-86-5 | RP230530RSR | 99% | 1,008.8 | µg/mL | +/- 36.7019 |
| 59 | Phenanthrene | 85-01-8 | MKCQ8876 | 99% | 1,008.4 | µg/mL | +/- 36.6877 |
| 60 | Anthracene | 120-12-7 | MKCR0570 | 99% | 1,009.0 | µg/mL | +/- 36.7100 |
| 61 | Carbazole | 86-74-8 | 14351100 | 99% | 1,000.9 | µg/mL | +/- 36.4149 |
| 62 | Di-n-butylphthalate | 84-74-2 | MKCN4337 | 99% | 1,007.6 | µg/mL | +/- 36.6595 |
| 63 | Fluoranthene | 206-44-0 | MKCQ4728 | 99% | 1,009.6 | µg/mL | +/- 36.7302 |
| 64 | Pyrene | 129-00-0 | BCCG8479 | 98% | 1,007.2 | µg/mL | +/- 36.6453 |
| 65 | Benzyl butyl phthalate | 85-68-7 | X12I018 | 99% | 1,002.1 | µg/mL | +/- 36.4573 |
| 66 | Bis(2-ethylhexyl)adipate | 103-23-1 | MKCM1988 | 99% | 1,005.2 | µg/mL | +/- 36.5705 |
| 67 | Benz(a)anthracene | 56-55-3 | I220012022BAA | 99% | 1,002.2 | µg/mL | +/- 36.4614 |
| 68 | Chrysene | 218-01-9 | RP230601 | 99% | 1,008.3 | µg/mL | +/- 36.6837 |
| 69 | Bis(2-ethylhexyl)phthalate | 117-81-7 | MKCQ3468 | 99% | 1,001.8 | µg/mL | +/- 36.4472 |
| 70 | Di-n-octyl phthalate | 117-84-0 | 14382700 | 99% | 1,006.0 | µg/mL | +/- 36.6008 |
| 71 | Benzo(b)fluoranthene | 205-99-2 | 012013B | 99% | 1,002.8 | µg/mL | +/- 36.4836 |
| 72 | Benzo(k)fluoranthene | 207-08-9 | 012022K | 99% | 1,003.0 | µg/mL | +/- 36.4917 |
| 73 | Benzo(a)pyrene | 50-32-8 | P54915-0703 | 99% | 1,002.3 | µg/mL | +/- 36.4674 |
| 74 | Indeno(1,2,3-cd)pyrene | 193-39-5 | 12-JKL-118-9 | 97% | 1,009.4 | µg/mL | +/- 36.7243 |
| 75 | Dibenz(a,h)anthracene | 53-70-3 | 2-ASA-59-1 | 99% | 1,007.6 | µg/mL | +/- 36.6595 |
| 76 | Benzo(g,h,i)perylene | 191-24-2 | RP231003RSR | 99% | 1,002.9 | µg/mL | +/- 36.4876 |

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride

CAS # 75-09-2

Purity 99%



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL



Certificate of Analysis

chromatographic plus

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31850

Lot No.: A0203726

Description : 8270 MegaMix®

8270 MegaMix® 500-1000 µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : April 30, 2025

Storage: 0°C or colder

Handling: Sonication required. Mix is photosensitive.

Ship: Ambient

512117 } RC/
↓ 03/18/24
512146

C E R T I F I E D V A L U E S

| Elution Order | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|---------------|------------------------------|----------|-------------|--------|--------------------------------|---|
| 1 | Pyridine | 110-86-1 | SHBP6240 | 99% | 1,001.6 µg/mL | +/- 36.4412 |
| 2 | N-Nitrosodimethylamine | 62-75-9 | 230209JLM | 99% | 1,005.9 µg/mL | +/- 36.5968 |
| 3 | Phenol | 108-95-2 | MKCK1120 | 99% | 1,003.3 µg/mL | +/- 36.5038 |
| 4 | Aniline | 62-53-3 | X22F726 | 99% | 1,005.8 µg/mL | +/- 36.5928 |
| 5 | Bis(2-chloroethyl)ether | 111-44-4 | SHBL6942 | 99% | 1,008.1 µg/mL | +/- 36.6776 |
| 6 | 2-Chlorophenol | 95-57-8 | STBJ3909 | 99% | 1,001.8 µg/mL | +/- 36.4492 |
| 7 | 1,3-Dichlorobenzene | 541-73-1 | BCCD5315 | 99% | 1,002.3 µg/mL | +/- 36.4654 |
| 8 | 1,4-Dichlorobenzene | 106-46-7 | MKBS7929V | 99% | 1,003.7 µg/mL | +/- 36.5159 |
| 9 | Benzyl alcohol | 100-51-6 | SHBK5469 | 99% | 1,008.7 µg/mL | +/- 36.6979 |
| 10 | 1,2-Dichlorobenzene | 95-50-1 | SHBN3835 | 99% | 1,000.3 µg/mL | +/- 36.3926 |
| 11 | 2-Methylphenol (o-cresol) | 95-48-7 | SHBN7598 | 99% | 1,003.5 µg/mL | +/- 36.5099 |
| 12 | 2,2'-oxybis(1-chloropropane) | 108-60-1 | 29-MAR-45-5 | 99% | 1,007.3 µg/mL | +/- 36.6493 |
| 13 | 3-Methylphenol (m-cresol) | 108-39-4 | STBJ0710 | 99% | 504.3 µg/mL | +/- 18.3500 |
| 14 | 4-Methylphenol (p-cresol) | 106-44-5 | SHBN3411 | 99% | 503.6 µg/mL | +/- 18.3237 |
| 15 | N-Nitroso-di-n-propylamine | 621-64-7 | N63MG | 99% | 1,008.3 µg/mL | +/- 36.6857 |
| 16 | Hexachloroethane | 67-72-1 | QTORH | 99% | 1,007.5 µg/mL | +/- 36.6554 |
| 17 | Nitrobenzene | 98-95-3 | 10224044 | 99% | 1,008.6 µg/mL | +/- 36.6938 |

| | | | | | | | | |
|----|---|-----------|-------------|-----|---------|-------|-----|---------|
| 18 | Isophorone | 78-59-1 | MKCC9506 | 99% | 1,005.9 | µg/mL | +/- | 36.5988 |
| 19 | 2-Nitrophenol | 88-75-5 | RP230710 | 99% | 1,003.2 | µg/mL | +/- | 36.4998 |
| 20 | 2,4-Dimethylphenol | 105-67-9 | XW5GK | 99% | 1,003.8 | µg/mL | +/- | 36.5200 |
| 21 | Bis(2-chloroethoxy)methane | 111-91-1 | 13670200 | 99% | 1,002.1 | µg/mL | +/- | 36.4573 |
| 22 | 2,4-Dichlorophenol | 120-83-2 | BCBZ6787 | 99% | 1,003.7 | µg/mL | +/- | 36.5180 |
| 23 | 1,2,4-Trichlorobenzene | 120-82-1 | SHBP5900 | 99% | 1,007.6 | µg/mL | +/- | 36.6574 |
| 24 | Naphthalene | 91-20-3 | STBL1057 | 99% | 1,008.3 | µg/mL | +/- | 36.6837 |
| 25 | 4-Chloroaniline | 106-47-8 | BCCJ3217 | 99% | 1,001.3 | µg/mL | +/- | 36.4290 |
| 26 | Hexachlorobutadiene | 87-68-3 | RP230823RSR | 98% | 1,008.3 | µg/mL | +/- | 36.6829 |
| 27 | 4-Chloro-3-methylphenol | 59-50-7 | BCCD4461 | 99% | 1,003.1 | µg/mL | +/- | 36.4937 |
| 28 | 2-Methylnaphthalene | 91-57-6 | STBK0259 | 96% | 1,001.9 | µg/mL | +/- | 36.4505 |
| 29 | 1-Methylnaphthalene | 90-12-0 | 5234.00-8 | 98% | 1,000.0 | µg/mL | +/- | 36.3838 |
| 30 | Hexachlorocyclopentadiene | 77-47-4 | 099063I14L | 98% | 1,008.5 | µg/mL | +/- | 36.6909 |
| 31 | 2,4,6-Trichlorophenol | 88-06-2 | STBJ5914 | 99% | 1,004.4 | µg/mL | +/- | 36.5442 |
| 32 | 2,4,5-Trichlorophenol | 95-95-4 | FHN01 | 98% | 1,001.9 | µg/mL | +/- | 36.4512 |
| 33 | 2-Chloronaphthalene | 91-58-7 | RPN7O | 99% | 1,001.1 | µg/mL | +/- | 36.4230 |
| 34 | 2-Nitroaniline | 88-74-4 | RP230531 | 99% | 1,002.9 | µg/mL | +/- | 36.4876 |
| 35 | 1,4-Dinitrobenzene | 100-25-4 | RP230816 | 99% | 1,005.7 | µg/mL | +/- | 36.5887 |
| 36 | Acenaphthylene | 208-96-8 | p06V | 98% | 1,009.5 | µg/mL | +/- | 36.7265 |
| 37 | 1,3-Dinitrobenzene | 99-65-0 | 1-DXX-24-1 | 99% | 1,004.4 | µg/mL | +/- | 36.5422 |
| 38 | Dimethylphthalate | 131-11-3 | 358221L17K | 99% | 1,005.9 | µg/mL | +/- | 36.5968 |
| 39 | 2,6-Dinitrotoluene | 606-20-2 | BCCG1833 | 99% | 1,003.2 | µg/mL | +/- | 36.4998 |
| 40 | 1,2-Dinitrobenzene | 528-29-0 | RP230428 | 99% | 1,002.2 | µg/mL | +/- | 36.4634 |
| 41 | Acenaphthene | 83-32-9 | MKCR7169 | 99% | 1,009.3 | µg/mL | +/- | 36.7221 |
| 42 | 3-Nitroaniline | 99-09-2 | RP230822RSR | 99% | 1,003.9 | µg/mL | +/- | 36.5240 |
| 43 | 2,4-Dinitrophenol | 51-28-5 | DR230417RSR | 99% | 1,002.0 | µg/mL | +/- | 36.4553 |
| 44 | Dibenzofuran | 132-64-9 | MKCD9952 | 99% | 1,006.7 | µg/mL | +/- | 36.6251 |
| 45 | 2,4-Dinitrotoluene | 121-14-2 | MKAA0690V | 99% | 1,003.8 | µg/mL | +/- | 36.5220 |
| 46 | 4-Nitrophenol | 100-02-7 | RP230627 | 99% | 1,002.3 | µg/mL | +/- | 36.4674 |
| 47 | 2,3,4,6-Tetrachlorophenol | 58-90-2 | PR-30126 | 99% | 1,008.7 | µg/mL | +/- | 36.6979 |
| 48 | 2,3,5,6-Tetrachlorophenol | 935-95-5 | RP230919 | 99% | 1,006.3 | µg/mL | +/- | 36.6130 |
| 49 | Fluorene | 86-73-7 | 10241100 | 99% | 1,008.3 | µg/mL | +/- | 36.6857 |
| 50 | 4-Chlorophenyl phenyl ether | 7005-72-3 | MKCT7248 | 99% | 1,003.8 | µg/mL | +/- | 36.5220 |
| 51 | Diethylphthalate | 84-66-2 | MKCD2547 | 99% | 1,008.6 | µg/mL | +/- | 36.6958 |
| 52 | 4-Nitroaniline | 100-01-6 | RP230111 | 99% | 1,001.1 | µg/mL | +/- | 36.4230 |
| 53 | 4,6-Dinitro-2-methylphenol (Dinitro-o-cresol) | 534-52-1 | 230718JLM | 99% | 1,002.0 | µg/mL | +/- | 36.4553 |

| | | | | | | | |
|----|----------------------------|----------|---------------|-----|---------|-------|-------------|
| 54 | Diphenylamine | 122-39-4 | MKCH1042 | 99% | 1,002.3 | µg/mL | +/- 36.4674 |
| 55 | Azobenzene | 103-33-3 | BCCK0887 | 99% | 1,005.8 | µg/mL | +/- 36.5928 |
| 56 | 4-Bromophenyl phenyl ether | 101-55-3 | STBH6361 | 99% | 1,003.0 | µg/mL | +/- 36.4917 |
| 57 | Hexachlorobenzene | 118-74-1 | 14821700 | 99% | 1,007.5 | µg/mL | +/- 36.6554 |
| 58 | Pentachlorophenol | 87-86-5 | RP230530RSR | 99% | 1,008.8 | µg/mL | +/- 36.7019 |
| 59 | Phenanthrene | 85-01-8 | MKCQ8876 | 99% | 1,008.4 | µg/mL | +/- 36.6877 |
| 60 | Anthracene | 120-12-7 | MKCR0570 | 99% | 1,009.0 | µg/mL | +/- 36.7100 |
| 61 | Carbazole | 86-74-8 | 14351100 | 99% | 1,000.9 | µg/mL | +/- 36.4149 |
| 62 | Di-n-butylphthalate | 84-74-2 | MKCN4337 | 99% | 1,007.6 | µg/mL | +/- 36.6595 |
| 63 | Fluoranthene | 206-44-0 | MKCQ4728 | 99% | 1,009.6 | µg/mL | +/- 36.7302 |
| 64 | Pyrene | 129-00-0 | BCCG8479 | 98% | 1,007.2 | µg/mL | +/- 36.6453 |
| 65 | Benzyl butyl phthalate | 85-68-7 | X12I018 | 99% | 1,002.1 | µg/mL | +/- 36.4573 |
| 66 | Bis(2-ethylhexyl)adipate | 103-23-1 | MKCM1988 | 99% | 1,005.2 | µg/mL | +/- 36.5705 |
| 67 | Benz(a)anthracene | 56-55-3 | I220012022BAA | 99% | 1,002.2 | µg/mL | +/- 36.4614 |
| 68 | Chrysene | 218-01-9 | RP230601 | 99% | 1,008.3 | µg/mL | +/- 36.6837 |
| 69 | Bis(2-ethylhexyl)phthalate | 117-81-7 | MKCQ3468 | 99% | 1,001.8 | µg/mL | +/- 36.4472 |
| 70 | Di-n-octyl phthalate | 117-84-0 | 14382700 | 99% | 1,006.0 | µg/mL | +/- 36.6008 |
| 71 | Benzo(b)fluoranthene | 205-99-2 | 012013B | 99% | 1,002.8 | µg/mL | +/- 36.4836 |
| 72 | Benzo(k)fluoranthene | 207-08-9 | 012022K | 99% | 1,003.0 | µg/mL | +/- 36.4917 |
| 73 | Benzo(a)pyrene | 50-32-8 | P54915-0703 | 99% | 1,002.3 | µg/mL | +/- 36.4674 |
| 74 | Indeno(1,2,3-cd)pyrene | 193-39-5 | 12-JKL-118-9 | 97% | 1,009.4 | µg/mL | +/- 36.7243 |
| 75 | Dibenz(a,h)anthracene | 53-70-3 | 2-ASA-59-1 | 99% | 1,007.6 | µg/mL | +/- 36.6595 |
| 76 | Benzo(g,h,i)perylene | 191-24-2 | RP231003RSR | 99% | 1,002.9 | µg/mL | +/- 36.4876 |

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride

CAS # 75-09-2

Purity 99%



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL



Certificate of Analysis

chromatographic plus

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31850

Lot No.: A0203726

Description : 8270 MegaMix®

8270 MegaMix® 500-1000 µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : April 30, 2025

Storage: 0°C or colder

Handling: Sonication required. Mix is photosensitive.

Ship: Ambient

512117 } RC/
↓ } 03/18/24
512146

C E R T I F I E D V A L U E S

| Elution Order | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|---------------|------------------------------|----------|-------------|--------|--------------------------------|---|
| 1 | Pyridine | 110-86-1 | SHBP6240 | 99% | 1,001.6 µg/mL | +/- 36.4412 |
| 2 | N-Nitrosodimethylamine | 62-75-9 | 230209JLM | 99% | 1,005.9 µg/mL | +/- 36.5968 |
| 3 | Phenol | 108-95-2 | MKCK1120 | 99% | 1,003.3 µg/mL | +/- 36.5038 |
| 4 | Aniline | 62-53-3 | X22F726 | 99% | 1,005.8 µg/mL | +/- 36.5928 |
| 5 | Bis(2-chloroethyl)ether | 111-44-4 | SHBL6942 | 99% | 1,008.1 µg/mL | +/- 36.6776 |
| 6 | 2-Chlorophenol | 95-57-8 | STBJ3909 | 99% | 1,001.8 µg/mL | +/- 36.4492 |
| 7 | 1,3-Dichlorobenzene | 541-73-1 | BCCD5315 | 99% | 1,002.3 µg/mL | +/- 36.4654 |
| 8 | 1,4-Dichlorobenzene | 106-46-7 | MKBS7929V | 99% | 1,003.7 µg/mL | +/- 36.5159 |
| 9 | Benzyl alcohol | 100-51-6 | SHBK5469 | 99% | 1,008.7 µg/mL | +/- 36.6979 |
| 10 | 1,2-Dichlorobenzene | 95-50-1 | SHBN3835 | 99% | 1,000.3 µg/mL | +/- 36.3926 |
| 11 | 2-Methylphenol (o-cresol) | 95-48-7 | SHBN7598 | 99% | 1,003.5 µg/mL | +/- 36.5099 |
| 12 | 2,2'-oxybis(1-chloropropane) | 108-60-1 | 29-MAR-45-5 | 99% | 1,007.3 µg/mL | +/- 36.6493 |
| 13 | 3-Methylphenol (m-cresol) | 108-39-4 | STBJ0710 | 99% | 504.3 µg/mL | +/- 18.3500 |
| 14 | 4-Methylphenol (p-cresol) | 106-44-5 | SHBN3411 | 99% | 503.6 µg/mL | +/- 18.3237 |
| 15 | N-Nitroso-di-n-propylamine | 621-64-7 | N63MG | 99% | 1,008.3 µg/mL | +/- 36.6857 |
| 16 | Hexachloroethane | 67-72-1 | QTORH | 99% | 1,007.5 µg/mL | +/- 36.6554 |
| 17 | Nitrobenzene | 98-95-3 | 10224044 | 99% | 1,008.6 µg/mL | +/- 36.6938 |

| | | | | | | | | |
|----|---|-----------|-------------|-----|---------|-------|-----|---------|
| 18 | Isophorone | 78-59-1 | MKCC9506 | 99% | 1,005.9 | µg/mL | +/- | 36.5988 |
| 19 | 2-Nitrophenol | 88-75-5 | RP230710 | 99% | 1,003.2 | µg/mL | +/- | 36.4998 |
| 20 | 2,4-Dimethylphenol | 105-67-9 | XW5GK | 99% | 1,003.8 | µg/mL | +/- | 36.5200 |
| 21 | Bis(2-chloroethoxy)methane | 111-91-1 | 13670200 | 99% | 1,002.1 | µg/mL | +/- | 36.4573 |
| 22 | 2,4-Dichlorophenol | 120-83-2 | BCBZ6787 | 99% | 1,003.7 | µg/mL | +/- | 36.5180 |
| 23 | 1,2,4-Trichlorobenzene | 120-82-1 | SHBP5900 | 99% | 1,007.6 | µg/mL | +/- | 36.6574 |
| 24 | Naphthalene | 91-20-3 | STBL1057 | 99% | 1,008.3 | µg/mL | +/- | 36.6837 |
| 25 | 4-Chloroaniline | 106-47-8 | BCCJ3217 | 99% | 1,001.3 | µg/mL | +/- | 36.4290 |
| 26 | Hexachlorobutadiene | 87-68-3 | RP230823RSR | 98% | 1,008.3 | µg/mL | +/- | 36.6829 |
| 27 | 4-Chloro-3-methylphenol | 59-50-7 | BCCD4461 | 99% | 1,003.1 | µg/mL | +/- | 36.4937 |
| 28 | 2-Methylnaphthalene | 91-57-6 | STBK0259 | 96% | 1,001.9 | µg/mL | +/- | 36.4505 |
| 29 | 1-Methylnaphthalene | 90-12-0 | 5234.00-8 | 98% | 1,000.0 | µg/mL | +/- | 36.3838 |
| 30 | Hexachlorocyclopentadiene | 77-47-4 | 099063I14L | 98% | 1,008.5 | µg/mL | +/- | 36.6909 |
| 31 | 2,4,6-Trichlorophenol | 88-06-2 | STBJ5914 | 99% | 1,004.4 | µg/mL | +/- | 36.5442 |
| 32 | 2,4,5-Trichlorophenol | 95-95-4 | FHN01 | 98% | 1,001.9 | µg/mL | +/- | 36.4512 |
| 33 | 2-Chloronaphthalene | 91-58-7 | RPN7O | 99% | 1,001.1 | µg/mL | +/- | 36.4230 |
| 34 | 2-Nitroaniline | 88-74-4 | RP230531 | 99% | 1,002.9 | µg/mL | +/- | 36.4876 |
| 35 | 1,4-Dinitrobenzene | 100-25-4 | RP230816 | 99% | 1,005.7 | µg/mL | +/- | 36.5887 |
| 36 | Acenaphthylene | 208-96-8 | p06V | 98% | 1,009.5 | µg/mL | +/- | 36.7265 |
| 37 | 1,3-Dinitrobenzene | 99-65-0 | 1-DXX-24-1 | 99% | 1,004.4 | µg/mL | +/- | 36.5422 |
| 38 | Dimethylphthalate | 131-11-3 | 358221L17K | 99% | 1,005.9 | µg/mL | +/- | 36.5968 |
| 39 | 2,6-Dinitrotoluene | 606-20-2 | BCCG1833 | 99% | 1,003.2 | µg/mL | +/- | 36.4998 |
| 40 | 1,2-Dinitrobenzene | 528-29-0 | RP230428 | 99% | 1,002.2 | µg/mL | +/- | 36.4634 |
| 41 | Acenaphthene | 83-32-9 | MKCR7169 | 99% | 1,009.3 | µg/mL | +/- | 36.7221 |
| 42 | 3-Nitroaniline | 99-09-2 | RP230822RSR | 99% | 1,003.9 | µg/mL | +/- | 36.5240 |
| 43 | 2,4-Dinitrophenol | 51-28-5 | DR230417RSR | 99% | 1,002.0 | µg/mL | +/- | 36.4553 |
| 44 | Dibenzofuran | 132-64-9 | MKCD9952 | 99% | 1,006.7 | µg/mL | +/- | 36.6251 |
| 45 | 2,4-Dinitrotoluene | 121-14-2 | MKAA0690V | 99% | 1,003.8 | µg/mL | +/- | 36.5220 |
| 46 | 4-Nitrophenol | 100-02-7 | RP230627 | 99% | 1,002.3 | µg/mL | +/- | 36.4674 |
| 47 | 2,3,4,6-Tetrachlorophenol | 58-90-2 | PR-30126 | 99% | 1,008.7 | µg/mL | +/- | 36.6979 |
| 48 | 2,3,5,6-Tetrachlorophenol | 935-95-5 | RP230919 | 99% | 1,006.3 | µg/mL | +/- | 36.6130 |
| 49 | Fluorene | 86-73-7 | 10241100 | 99% | 1,008.3 | µg/mL | +/- | 36.6857 |
| 50 | 4-Chlorophenyl phenyl ether | 7005-72-3 | MKCT7248 | 99% | 1,003.8 | µg/mL | +/- | 36.5220 |
| 51 | Diethylphthalate | 84-66-2 | MKCD2547 | 99% | 1,008.6 | µg/mL | +/- | 36.6958 |
| 52 | 4-Nitroaniline | 100-01-6 | RP230111 | 99% | 1,001.1 | µg/mL | +/- | 36.4230 |
| 53 | 4,6-Dinitro-2-methylphenol (Dinitro-o-cresol) | 534-52-1 | 230718JLM | 99% | 1,002.0 | µg/mL | +/- | 36.4553 |

| | | | | | | | |
|----|----------------------------|----------|---------------|-----|---------|-------|-------------|
| 54 | Diphenylamine | 122-39-4 | MKCH1042 | 99% | 1,002.3 | µg/mL | +/- 36.4674 |
| 55 | Azobenzene | 103-33-3 | BCCK0887 | 99% | 1,005.8 | µg/mL | +/- 36.5928 |
| 56 | 4-Bromophenyl phenyl ether | 101-55-3 | STBH6361 | 99% | 1,003.0 | µg/mL | +/- 36.4917 |
| 57 | Hexachlorobenzene | 118-74-1 | 14821700 | 99% | 1,007.5 | µg/mL | +/- 36.6554 |
| 58 | Pentachlorophenol | 87-86-5 | RP230530RSR | 99% | 1,008.8 | µg/mL | +/- 36.7019 |
| 59 | Phenanthrene | 85-01-8 | MKCQ8876 | 99% | 1,008.4 | µg/mL | +/- 36.6877 |
| 60 | Anthracene | 120-12-7 | MKCR0570 | 99% | 1,009.0 | µg/mL | +/- 36.7100 |
| 61 | Carbazole | 86-74-8 | 14351100 | 99% | 1,000.9 | µg/mL | +/- 36.4149 |
| 62 | Di-n-butylphthalate | 84-74-2 | MKCN4337 | 99% | 1,007.6 | µg/mL | +/- 36.6595 |
| 63 | Fluoranthene | 206-44-0 | MKCQ4728 | 99% | 1,009.6 | µg/mL | +/- 36.7302 |
| 64 | Pyrene | 129-00-0 | BCCG8479 | 98% | 1,007.2 | µg/mL | +/- 36.6453 |
| 65 | Benzyl butyl phthalate | 85-68-7 | X12I018 | 99% | 1,002.1 | µg/mL | +/- 36.4573 |
| 66 | Bis(2-ethylhexyl)adipate | 103-23-1 | MKCM1988 | 99% | 1,005.2 | µg/mL | +/- 36.5705 |
| 67 | Benz(a)anthracene | 56-55-3 | I220012022BAA | 99% | 1,002.2 | µg/mL | +/- 36.4614 |
| 68 | Chrysene | 218-01-9 | RP230601 | 99% | 1,008.3 | µg/mL | +/- 36.6837 |
| 69 | Bis(2-ethylhexyl)phthalate | 117-81-7 | MKCQ3468 | 99% | 1,001.8 | µg/mL | +/- 36.4472 |
| 70 | Di-n-octyl phthalate | 117-84-0 | 14382700 | 99% | 1,006.0 | µg/mL | +/- 36.6008 |
| 71 | Benzo(b)fluoranthene | 205-99-2 | 012013B | 99% | 1,002.8 | µg/mL | +/- 36.4836 |
| 72 | Benzo(k)fluoranthene | 207-08-9 | 012022K | 99% | 1,003.0 | µg/mL | +/- 36.4917 |
| 73 | Benzo(a)pyrene | 50-32-8 | P54915-0703 | 99% | 1,002.3 | µg/mL | +/- 36.4674 |
| 74 | Indeno(1,2,3-cd)pyrene | 193-39-5 | 12-JKL-118-9 | 97% | 1,009.4 | µg/mL | +/- 36.7243 |
| 75 | Dibenz(a,h)anthracene | 53-70-3 | 2-ASA-59-1 | 99% | 1,007.6 | µg/mL | +/- 36.6595 |
| 76 | Benzo(g,h,i)perylene | 191-24-2 | RP231003RSR | 99% | 1,002.9 | µg/mL | +/- 36.4876 |

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride

CAS # 75-09-2

Purity 99%



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL



Certificate of Analysis

chromatographic plus

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31850

Lot No.: A0203726

Description : 8270 MegaMix®

8270 MegaMix® 500-1000 µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : April 30, 2025

Storage: 0°C or colder

Handling: Sonication required. Mix is photosensitive.

Ship: Ambient

512117 } RC/
↓ 03/18/24
512146

C E R T I F I E D V A L U E S

| Elution Order | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|---------------|------------------------------|----------|-------------|--------|--------------------------------|---|
| 1 | Pyridine | 110-86-1 | SHBP6240 | 99% | 1,001.6 µg/mL | +/- 36.4412 |
| 2 | N-Nitrosodimethylamine | 62-75-9 | 230209JLM | 99% | 1,005.9 µg/mL | +/- 36.5968 |
| 3 | Phenol | 108-95-2 | MKCK1120 | 99% | 1,003.3 µg/mL | +/- 36.5038 |
| 4 | Aniline | 62-53-3 | X22F726 | 99% | 1,005.8 µg/mL | +/- 36.5928 |
| 5 | Bis(2-chloroethyl)ether | 111-44-4 | SHBL6942 | 99% | 1,008.1 µg/mL | +/- 36.6776 |
| 6 | 2-Chlorophenol | 95-57-8 | STBJ3909 | 99% | 1,001.8 µg/mL | +/- 36.4492 |
| 7 | 1,3-Dichlorobenzene | 541-73-1 | BCCD5315 | 99% | 1,002.3 µg/mL | +/- 36.4654 |
| 8 | 1,4-Dichlorobenzene | 106-46-7 | MKBS7929V | 99% | 1,003.7 µg/mL | +/- 36.5159 |
| 9 | Benzyl alcohol | 100-51-6 | SHBK5469 | 99% | 1,008.7 µg/mL | +/- 36.6979 |
| 10 | 1,2-Dichlorobenzene | 95-50-1 | SHBN3835 | 99% | 1,000.3 µg/mL | +/- 36.3926 |
| 11 | 2-Methylphenol (o-cresol) | 95-48-7 | SHBN7598 | 99% | 1,003.5 µg/mL | +/- 36.5099 |
| 12 | 2,2'-oxybis(1-chloropropane) | 108-60-1 | 29-MAR-45-5 | 99% | 1,007.3 µg/mL | +/- 36.6493 |
| 13 | 3-Methylphenol (m-cresol) | 108-39-4 | STBJ0710 | 99% | 504.3 µg/mL | +/- 18.3500 |
| 14 | 4-Methylphenol (p-cresol) | 106-44-5 | SHBN3411 | 99% | 503.6 µg/mL | +/- 18.3237 |
| 15 | N-Nitroso-di-n-propylamine | 621-64-7 | N63MG | 99% | 1,008.3 µg/mL | +/- 36.6857 |
| 16 | Hexachloroethane | 67-72-1 | QTORH | 99% | 1,007.5 µg/mL | +/- 36.6554 |
| 17 | Nitrobenzene | 98-95-3 | 10224044 | 99% | 1,008.6 µg/mL | +/- 36.6938 |

| | | | | | | | | |
|----|---|-----------|-------------|-----|---------|-------|-----|---------|
| 18 | Isophorone | 78-59-1 | MKCC9506 | 99% | 1,005.9 | µg/mL | +/- | 36.5988 |
| 19 | 2-Nitrophenol | 88-75-5 | RP230710 | 99% | 1,003.2 | µg/mL | +/- | 36.4998 |
| 20 | 2,4-Dimethylphenol | 105-67-9 | XW5GK | 99% | 1,003.8 | µg/mL | +/- | 36.5200 |
| 21 | Bis(2-chloroethoxy)methane | 111-91-1 | 13670200 | 99% | 1,002.1 | µg/mL | +/- | 36.4573 |
| 22 | 2,4-Dichlorophenol | 120-83-2 | BCBZ6787 | 99% | 1,003.7 | µg/mL | +/- | 36.5180 |
| 23 | 1,2,4-Trichlorobenzene | 120-82-1 | SHBP5900 | 99% | 1,007.6 | µg/mL | +/- | 36.6574 |
| 24 | Naphthalene | 91-20-3 | STBL1057 | 99% | 1,008.3 | µg/mL | +/- | 36.6837 |
| 25 | 4-Chloroaniline | 106-47-8 | BCCJ3217 | 99% | 1,001.3 | µg/mL | +/- | 36.4290 |
| 26 | Hexachlorobutadiene | 87-68-3 | RP230823RSR | 98% | 1,008.3 | µg/mL | +/- | 36.6829 |
| 27 | 4-Chloro-3-methylphenol | 59-50-7 | BCCD4461 | 99% | 1,003.1 | µg/mL | +/- | 36.4937 |
| 28 | 2-Methylnaphthalene | 91-57-6 | STBK0259 | 96% | 1,001.9 | µg/mL | +/- | 36.4505 |
| 29 | 1-Methylnaphthalene | 90-12-0 | 5234.00-8 | 98% | 1,000.0 | µg/mL | +/- | 36.3838 |
| 30 | Hexachlorocyclopentadiene | 77-47-4 | 099063I14L | 98% | 1,008.5 | µg/mL | +/- | 36.6909 |
| 31 | 2,4,6-Trichlorophenol | 88-06-2 | STBJ5914 | 99% | 1,004.4 | µg/mL | +/- | 36.5442 |
| 32 | 2,4,5-Trichlorophenol | 95-95-4 | FHN01 | 98% | 1,001.9 | µg/mL | +/- | 36.4512 |
| 33 | 2-Chloronaphthalene | 91-58-7 | RPN7O | 99% | 1,001.1 | µg/mL | +/- | 36.4230 |
| 34 | 2-Nitroaniline | 88-74-4 | RP230531 | 99% | 1,002.9 | µg/mL | +/- | 36.4876 |
| 35 | 1,4-Dinitrobenzene | 100-25-4 | RP230816 | 99% | 1,005.7 | µg/mL | +/- | 36.5887 |
| 36 | Acenaphthylene | 208-96-8 | p06V | 98% | 1,009.5 | µg/mL | +/- | 36.7265 |
| 37 | 1,3-Dinitrobenzene | 99-65-0 | 1-DXX-24-1 | 99% | 1,004.4 | µg/mL | +/- | 36.5422 |
| 38 | Dimethylphthalate | 131-11-3 | 358221L17K | 99% | 1,005.9 | µg/mL | +/- | 36.5968 |
| 39 | 2,6-Dinitrotoluene | 606-20-2 | BCCG1833 | 99% | 1,003.2 | µg/mL | +/- | 36.4998 |
| 40 | 1,2-Dinitrobenzene | 528-29-0 | RP230428 | 99% | 1,002.2 | µg/mL | +/- | 36.4634 |
| 41 | Acenaphthene | 83-32-9 | MKCR7169 | 99% | 1,009.3 | µg/mL | +/- | 36.7221 |
| 42 | 3-Nitroaniline | 99-09-2 | RP230822RSR | 99% | 1,003.9 | µg/mL | +/- | 36.5240 |
| 43 | 2,4-Dinitrophenol | 51-28-5 | DR230417RSR | 99% | 1,002.0 | µg/mL | +/- | 36.4553 |
| 44 | Dibenzofuran | 132-64-9 | MKCD9952 | 99% | 1,006.7 | µg/mL | +/- | 36.6251 |
| 45 | 2,4-Dinitrotoluene | 121-14-2 | MKAA0690V | 99% | 1,003.8 | µg/mL | +/- | 36.5220 |
| 46 | 4-Nitrophenol | 100-02-7 | RP230627 | 99% | 1,002.3 | µg/mL | +/- | 36.4674 |
| 47 | 2,3,4,6-Tetrachlorophenol | 58-90-2 | PR-30126 | 99% | 1,008.7 | µg/mL | +/- | 36.6979 |
| 48 | 2,3,5,6-Tetrachlorophenol | 935-95-5 | RP230919 | 99% | 1,006.3 | µg/mL | +/- | 36.6130 |
| 49 | Fluorene | 86-73-7 | 10241100 | 99% | 1,008.3 | µg/mL | +/- | 36.6857 |
| 50 | 4-Chlorophenyl phenyl ether | 7005-72-3 | MKCT7248 | 99% | 1,003.8 | µg/mL | +/- | 36.5220 |
| 51 | Diethylphthalate | 84-66-2 | MKCD2547 | 99% | 1,008.6 | µg/mL | +/- | 36.6958 |
| 52 | 4-Nitroaniline | 100-01-6 | RP230111 | 99% | 1,001.1 | µg/mL | +/- | 36.4230 |
| 53 | 4,6-Dinitro-2-methylphenol (Dinitro-o-cresol) | 534-52-1 | 230718JLM | 99% | 1,002.0 | µg/mL | +/- | 36.4553 |

| | | | | | | | |
|----|----------------------------|----------|---------------|-----|---------|-------|-------------|
| 54 | Diphenylamine | 122-39-4 | MKCH1042 | 99% | 1,002.3 | µg/mL | +/- 36.4674 |
| 55 | Azobenzene | 103-33-3 | BCCK0887 | 99% | 1,005.8 | µg/mL | +/- 36.5928 |
| 56 | 4-Bromophenyl phenyl ether | 101-55-3 | STBH6361 | 99% | 1,003.0 | µg/mL | +/- 36.4917 |
| 57 | Hexachlorobenzene | 118-74-1 | 14821700 | 99% | 1,007.5 | µg/mL | +/- 36.6554 |
| 58 | Pentachlorophenol | 87-86-5 | RP230530RSR | 99% | 1,008.8 | µg/mL | +/- 36.7019 |
| 59 | Phenanthrene | 85-01-8 | MKCQ8876 | 99% | 1,008.4 | µg/mL | +/- 36.6877 |
| 60 | Anthracene | 120-12-7 | MKCR0570 | 99% | 1,009.0 | µg/mL | +/- 36.7100 |
| 61 | Carbazole | 86-74-8 | 14351100 | 99% | 1,000.9 | µg/mL | +/- 36.4149 |
| 62 | Di-n-butylphthalate | 84-74-2 | MKCN4337 | 99% | 1,007.6 | µg/mL | +/- 36.6595 |
| 63 | Fluoranthene | 206-44-0 | MKCQ4728 | 99% | 1,009.6 | µg/mL | +/- 36.7302 |
| 64 | Pyrene | 129-00-0 | BCCG8479 | 98% | 1,007.2 | µg/mL | +/- 36.6453 |
| 65 | Benzyl butyl phthalate | 85-68-7 | X12I018 | 99% | 1,002.1 | µg/mL | +/- 36.4573 |
| 66 | Bis(2-ethylhexyl)adipate | 103-23-1 | MKCM1988 | 99% | 1,005.2 | µg/mL | +/- 36.5705 |
| 67 | Benz(a)anthracene | 56-55-3 | I220012022BAA | 99% | 1,002.2 | µg/mL | +/- 36.4614 |
| 68 | Chrysene | 218-01-9 | RP230601 | 99% | 1,008.3 | µg/mL | +/- 36.6837 |
| 69 | Bis(2-ethylhexyl)phthalate | 117-81-7 | MKCQ3468 | 99% | 1,001.8 | µg/mL | +/- 36.4472 |
| 70 | Di-n-octyl phthalate | 117-84-0 | 14382700 | 99% | 1,006.0 | µg/mL | +/- 36.6008 |
| 71 | Benzo(b)fluoranthene | 205-99-2 | 012013B | 99% | 1,002.8 | µg/mL | +/- 36.4836 |
| 72 | Benzo(k)fluoranthene | 207-08-9 | 012022K | 99% | 1,003.0 | µg/mL | +/- 36.4917 |
| 73 | Benzo(a)pyrene | 50-32-8 | P54915-0703 | 99% | 1,002.3 | µg/mL | +/- 36.4674 |
| 74 | Indeno(1,2,3-cd)pyrene | 193-39-5 | 12-JKL-118-9 | 97% | 1,009.4 | µg/mL | +/- 36.7243 |
| 75 | Dibenz(a,h)anthracene | 53-70-3 | 2-ASA-59-1 | 99% | 1,007.6 | µg/mL | +/- 36.6595 |
| 76 | Benzo(g,h,i)perylene | 191-24-2 | RP231003RSR | 99% | 1,002.9 | µg/mL | +/- 36.4876 |

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride

CAS # 75-09-2

Purity 99%



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL



Certificate of Analysis

chromatographic plus

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31850

Lot No.: A0203726

Description : 8270 MegaMix®

8270 MegaMix® 500-1000 µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : April 30, 2025

Storage: 0°C or colder

Handling: Sonication required. Mix is photosensitive.

Ship: Ambient

512117 } RC/
↓ 03/18/24
512146

C E R T I F I E D V A L U E S

| Elution Order | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|---------------|------------------------------|----------|-------------|--------|--------------------------------|---|
| 1 | Pyridine | 110-86-1 | SHBP6240 | 99% | 1,001.6 µg/mL | +/- 36.4412 |
| 2 | N-Nitrosodimethylamine | 62-75-9 | 230209JLM | 99% | 1,005.9 µg/mL | +/- 36.5968 |
| 3 | Phenol | 108-95-2 | MKCK1120 | 99% | 1,003.3 µg/mL | +/- 36.5038 |
| 4 | Aniline | 62-53-3 | X22F726 | 99% | 1,005.8 µg/mL | +/- 36.5928 |
| 5 | Bis(2-chloroethyl)ether | 111-44-4 | SHBL6942 | 99% | 1,008.1 µg/mL | +/- 36.6776 |
| 6 | 2-Chlorophenol | 95-57-8 | STBJ3909 | 99% | 1,001.8 µg/mL | +/- 36.4492 |
| 7 | 1,3-Dichlorobenzene | 541-73-1 | BCCD5315 | 99% | 1,002.3 µg/mL | +/- 36.4654 |
| 8 | 1,4-Dichlorobenzene | 106-46-7 | MKBS7929V | 99% | 1,003.7 µg/mL | +/- 36.5159 |
| 9 | Benzyl alcohol | 100-51-6 | SHBK5469 | 99% | 1,008.7 µg/mL | +/- 36.6979 |
| 10 | 1,2-Dichlorobenzene | 95-50-1 | SHBN3835 | 99% | 1,000.3 µg/mL | +/- 36.3926 |
| 11 | 2-Methylphenol (o-cresol) | 95-48-7 | SHBN7598 | 99% | 1,003.5 µg/mL | +/- 36.5099 |
| 12 | 2,2'-oxybis(1-chloropropane) | 108-60-1 | 29-MAR-45-5 | 99% | 1,007.3 µg/mL | +/- 36.6493 |
| 13 | 3-Methylphenol (m-cresol) | 108-39-4 | STBJ0710 | 99% | 504.3 µg/mL | +/- 18.3500 |
| 14 | 4-Methylphenol (p-cresol) | 106-44-5 | SHBN3411 | 99% | 503.6 µg/mL | +/- 18.3237 |
| 15 | N-Nitroso-di-n-propylamine | 621-64-7 | N63MG | 99% | 1,008.3 µg/mL | +/- 36.6857 |
| 16 | Hexachloroethane | 67-72-1 | QTORH | 99% | 1,007.5 µg/mL | +/- 36.6554 |
| 17 | Nitrobenzene | 98-95-3 | 10224044 | 99% | 1,008.6 µg/mL | +/- 36.6938 |

| | | | | | | | | |
|----|---|-----------|-------------|-----|---------|-------|-----|---------|
| 18 | Isophorone | 78-59-1 | MKCC9506 | 99% | 1,005.9 | µg/mL | +/- | 36.5988 |
| 19 | 2-Nitrophenol | 88-75-5 | RP230710 | 99% | 1,003.2 | µg/mL | +/- | 36.4998 |
| 20 | 2,4-Dimethylphenol | 105-67-9 | XW5GK | 99% | 1,003.8 | µg/mL | +/- | 36.5200 |
| 21 | Bis(2-chloroethoxy)methane | 111-91-1 | 13670200 | 99% | 1,002.1 | µg/mL | +/- | 36.4573 |
| 22 | 2,4-Dichlorophenol | 120-83-2 | BCBZ6787 | 99% | 1,003.7 | µg/mL | +/- | 36.5180 |
| 23 | 1,2,4-Trichlorobenzene | 120-82-1 | SHBP5900 | 99% | 1,007.6 | µg/mL | +/- | 36.6574 |
| 24 | Naphthalene | 91-20-3 | STBL1057 | 99% | 1,008.3 | µg/mL | +/- | 36.6837 |
| 25 | 4-Chloroaniline | 106-47-8 | BCCJ3217 | 99% | 1,001.3 | µg/mL | +/- | 36.4290 |
| 26 | Hexachlorobutadiene | 87-68-3 | RP230823RSR | 98% | 1,008.3 | µg/mL | +/- | 36.6829 |
| 27 | 4-Chloro-3-methylphenol | 59-50-7 | BCCD4461 | 99% | 1,003.1 | µg/mL | +/- | 36.4937 |
| 28 | 2-Methylnaphthalene | 91-57-6 | STBK0259 | 96% | 1,001.9 | µg/mL | +/- | 36.4505 |
| 29 | 1-Methylnaphthalene | 90-12-0 | 5234.00-8 | 98% | 1,000.0 | µg/mL | +/- | 36.3838 |
| 30 | Hexachlorocyclopentadiene | 77-47-4 | 099063I14L | 98% | 1,008.5 | µg/mL | +/- | 36.6909 |
| 31 | 2,4,6-Trichlorophenol | 88-06-2 | STBJ5914 | 99% | 1,004.4 | µg/mL | +/- | 36.5442 |
| 32 | 2,4,5-Trichlorophenol | 95-95-4 | FHN01 | 98% | 1,001.9 | µg/mL | +/- | 36.4512 |
| 33 | 2-Chloronaphthalene | 91-58-7 | RPN7O | 99% | 1,001.1 | µg/mL | +/- | 36.4230 |
| 34 | 2-Nitroaniline | 88-74-4 | RP230531 | 99% | 1,002.9 | µg/mL | +/- | 36.4876 |
| 35 | 1,4-Dinitrobenzene | 100-25-4 | RP230816 | 99% | 1,005.7 | µg/mL | +/- | 36.5887 |
| 36 | Acenaphthylene | 208-96-8 | p06V | 98% | 1,009.5 | µg/mL | +/- | 36.7265 |
| 37 | 1,3-Dinitrobenzene | 99-65-0 | 1-DXX-24-1 | 99% | 1,004.4 | µg/mL | +/- | 36.5422 |
| 38 | Dimethylphthalate | 131-11-3 | 358221L17K | 99% | 1,005.9 | µg/mL | +/- | 36.5968 |
| 39 | 2,6-Dinitrotoluene | 606-20-2 | BCCG1833 | 99% | 1,003.2 | µg/mL | +/- | 36.4998 |
| 40 | 1,2-Dinitrobenzene | 528-29-0 | RP230428 | 99% | 1,002.2 | µg/mL | +/- | 36.4634 |
| 41 | Acenaphthene | 83-32-9 | MKCR7169 | 99% | 1,009.3 | µg/mL | +/- | 36.7221 |
| 42 | 3-Nitroaniline | 99-09-2 | RP230822RSR | 99% | 1,003.9 | µg/mL | +/- | 36.5240 |
| 43 | 2,4-Dinitrophenol | 51-28-5 | DR230417RSR | 99% | 1,002.0 | µg/mL | +/- | 36.4553 |
| 44 | Dibenzofuran | 132-64-9 | MKCD9952 | 99% | 1,006.7 | µg/mL | +/- | 36.6251 |
| 45 | 2,4-Dinitrotoluene | 121-14-2 | MKAA0690V | 99% | 1,003.8 | µg/mL | +/- | 36.5220 |
| 46 | 4-Nitrophenol | 100-02-7 | RP230627 | 99% | 1,002.3 | µg/mL | +/- | 36.4674 |
| 47 | 2,3,4,6-Tetrachlorophenol | 58-90-2 | PR-30126 | 99% | 1,008.7 | µg/mL | +/- | 36.6979 |
| 48 | 2,3,5,6-Tetrachlorophenol | 935-95-5 | RP230919 | 99% | 1,006.3 | µg/mL | +/- | 36.6130 |
| 49 | Fluorene | 86-73-7 | 10241100 | 99% | 1,008.3 | µg/mL | +/- | 36.6857 |
| 50 | 4-Chlorophenyl phenyl ether | 7005-72-3 | MKCT7248 | 99% | 1,003.8 | µg/mL | +/- | 36.5220 |
| 51 | Diethylphthalate | 84-66-2 | MKCD2547 | 99% | 1,008.6 | µg/mL | +/- | 36.6958 |
| 52 | 4-Nitroaniline | 100-01-6 | RP230111 | 99% | 1,001.1 | µg/mL | +/- | 36.4230 |
| 53 | 4,6-Dinitro-2-methylphenol (Dinitro-o-cresol) | 534-52-1 | 230718JLM | 99% | 1,002.0 | µg/mL | +/- | 36.4553 |

| | | | | | | | |
|----|----------------------------|----------|---------------|-----|---------|-------|-------------|
| 54 | Diphenylamine | 122-39-4 | MKCH1042 | 99% | 1,002.3 | µg/mL | +/- 36.4674 |
| 55 | Azobenzene | 103-33-3 | BCCK0887 | 99% | 1,005.8 | µg/mL | +/- 36.5928 |
| 56 | 4-Bromophenyl phenyl ether | 101-55-3 | STBH6361 | 99% | 1,003.0 | µg/mL | +/- 36.4917 |
| 57 | Hexachlorobenzene | 118-74-1 | 14821700 | 99% | 1,007.5 | µg/mL | +/- 36.6554 |
| 58 | Pentachlorophenol | 87-86-5 | RP230530RSR | 99% | 1,008.8 | µg/mL | +/- 36.7019 |
| 59 | Phenanthrene | 85-01-8 | MKCQ8876 | 99% | 1,008.4 | µg/mL | +/- 36.6877 |
| 60 | Anthracene | 120-12-7 | MKCR0570 | 99% | 1,009.0 | µg/mL | +/- 36.7100 |
| 61 | Carbazole | 86-74-8 | 14351100 | 99% | 1,000.9 | µg/mL | +/- 36.4149 |
| 62 | Di-n-butylphthalate | 84-74-2 | MKCN4337 | 99% | 1,007.6 | µg/mL | +/- 36.6595 |
| 63 | Fluoranthene | 206-44-0 | MKCQ4728 | 99% | 1,009.6 | µg/mL | +/- 36.7302 |
| 64 | Pyrene | 129-00-0 | BCCG8479 | 98% | 1,007.2 | µg/mL | +/- 36.6453 |
| 65 | Benzyl butyl phthalate | 85-68-7 | X12I018 | 99% | 1,002.1 | µg/mL | +/- 36.4573 |
| 66 | Bis(2-ethylhexyl)adipate | 103-23-1 | MKCM1988 | 99% | 1,005.2 | µg/mL | +/- 36.5705 |
| 67 | Benz(a)anthracene | 56-55-3 | I220012022BAA | 99% | 1,002.2 | µg/mL | +/- 36.4614 |
| 68 | Chrysene | 218-01-9 | RP230601 | 99% | 1,008.3 | µg/mL | +/- 36.6837 |
| 69 | Bis(2-ethylhexyl)phthalate | 117-81-7 | MKCQ3468 | 99% | 1,001.8 | µg/mL | +/- 36.4472 |
| 70 | Di-n-octyl phthalate | 117-84-0 | 14382700 | 99% | 1,006.0 | µg/mL | +/- 36.6008 |
| 71 | Benzo(b)fluoranthene | 205-99-2 | 012013B | 99% | 1,002.8 | µg/mL | +/- 36.4836 |
| 72 | Benzo(k)fluoranthene | 207-08-9 | 012022K | 99% | 1,003.0 | µg/mL | +/- 36.4917 |
| 73 | Benzo(a)pyrene | 50-32-8 | P54915-0703 | 99% | 1,002.3 | µg/mL | +/- 36.4674 |
| 74 | Indeno(1,2,3-cd)pyrene | 193-39-5 | 12-JKL-118-9 | 97% | 1,009.4 | µg/mL | +/- 36.7243 |
| 75 | Dibenz(a,h)anthracene | 53-70-3 | 2-ASA-59-1 | 99% | 1,007.6 | µg/mL | +/- 36.6595 |
| 76 | Benzo(g,h,i)perylene | 191-24-2 | RP231003RSR | 99% | 1,002.9 | µg/mL | +/- 36.4876 |

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride

CAS # 75-09-2

Purity 99%



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL



Certificate of Analysis

chromatographic plus

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31850
Description : 8270 MegaMix®
 8270 MegaMix® 500-1000 µg/mL, Methylene Chloride, 1mL/ampul
Container Size : 2 mL
Expiration Date : April 30, 2025
Handling: Sonication required. Mix is photosensitive.

Lot No.: A0203726
Pkg Amt: > 1 mL
Storage: 0°C or colder
Ship: Ambient

512117 } RC/
 ↓ 03/18/24
 512146

C E R T I F I E D V A L U E S

| Elution Order | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|---------------|------------------------------|----------|-------------|--------|--------------------------------|---|
| 1 | Pyridine | 110-86-1 | SHBP6240 | 99% | 1,001.6 µg/mL | +/- 36.4412 |
| 2 | N-Nitrosodimethylamine | 62-75-9 | 230209JLM | 99% | 1,005.9 µg/mL | +/- 36.5968 |
| 3 | Phenol | 108-95-2 | MKCK1120 | 99% | 1,003.3 µg/mL | +/- 36.5038 |
| 4 | Aniline | 62-53-3 | X22F726 | 99% | 1,005.8 µg/mL | +/- 36.5928 |
| 5 | Bis(2-chloroethyl)ether | 111-44-4 | SHBL6942 | 99% | 1,008.1 µg/mL | +/- 36.6776 |
| 6 | 2-Chlorophenol | 95-57-8 | STBJ3909 | 99% | 1,001.8 µg/mL | +/- 36.4492 |
| 7 | 1,3-Dichlorobenzene | 541-73-1 | BCCD5315 | 99% | 1,002.3 µg/mL | +/- 36.4654 |
| 8 | 1,4-Dichlorobenzene | 106-46-7 | MKBS7929V | 99% | 1,003.7 µg/mL | +/- 36.5159 |
| 9 | Benzyl alcohol | 100-51-6 | SHBK5469 | 99% | 1,008.7 µg/mL | +/- 36.6979 |
| 10 | 1,2-Dichlorobenzene | 95-50-1 | SHBN3835 | 99% | 1,000.3 µg/mL | +/- 36.3926 |
| 11 | 2-Methylphenol (o-cresol) | 95-48-7 | SHBN7598 | 99% | 1,003.5 µg/mL | +/- 36.5099 |
| 12 | 2,2'-oxybis(1-chloropropane) | 108-60-1 | 29-MAR-45-5 | 99% | 1,007.3 µg/mL | +/- 36.6493 |
| 13 | 3-Methylphenol (m-cresol) | 108-39-4 | STBJ0710 | 99% | 504.3 µg/mL | +/- 18.3500 |
| 14 | 4-Methylphenol (p-cresol) | 106-44-5 | SHBN3411 | 99% | 503.6 µg/mL | +/- 18.3237 |
| 15 | N-Nitroso-di-n-propylamine | 621-64-7 | N63MG | 99% | 1,008.3 µg/mL | +/- 36.6857 |
| 16 | Hexachloroethane | 67-72-1 | QTORH | 99% | 1,007.5 µg/mL | +/- 36.6554 |
| 17 | Nitrobenzene | 98-95-3 | 10224044 | 99% | 1,008.6 µg/mL | +/- 36.6938 |

| | | | | | | | | |
|----|---|-----------|-------------|-----|---------|-------|-----|---------|
| 18 | Isophorone | 78-59-1 | MKCC9506 | 99% | 1,005.9 | µg/mL | +/- | 36.5988 |
| 19 | 2-Nitrophenol | 88-75-5 | RP230710 | 99% | 1,003.2 | µg/mL | +/- | 36.4998 |
| 20 | 2,4-Dimethylphenol | 105-67-9 | XW5GK | 99% | 1,003.8 | µg/mL | +/- | 36.5200 |
| 21 | Bis(2-chloroethoxy)methane | 111-91-1 | 13670200 | 99% | 1,002.1 | µg/mL | +/- | 36.4573 |
| 22 | 2,4-Dichlorophenol | 120-83-2 | BCBZ6787 | 99% | 1,003.7 | µg/mL | +/- | 36.5180 |
| 23 | 1,2,4-Trichlorobenzene | 120-82-1 | SHBP5900 | 99% | 1,007.6 | µg/mL | +/- | 36.6574 |
| 24 | Naphthalene | 91-20-3 | STBL1057 | 99% | 1,008.3 | µg/mL | +/- | 36.6837 |
| 25 | 4-Chloroaniline | 106-47-8 | BCCJ3217 | 99% | 1,001.3 | µg/mL | +/- | 36.4290 |
| 26 | Hexachlorobutadiene | 87-68-3 | RP230823RSR | 98% | 1,008.3 | µg/mL | +/- | 36.6829 |
| 27 | 4-Chloro-3-methylphenol | 59-50-7 | BCCD4461 | 99% | 1,003.1 | µg/mL | +/- | 36.4937 |
| 28 | 2-Methylnaphthalene | 91-57-6 | STBK0259 | 96% | 1,001.9 | µg/mL | +/- | 36.4505 |
| 29 | 1-Methylnaphthalene | 90-12-0 | 5234.00-8 | 98% | 1,000.0 | µg/mL | +/- | 36.3838 |
| 30 | Hexachlorocyclopentadiene | 77-47-4 | 099063I14L | 98% | 1,008.5 | µg/mL | +/- | 36.6909 |
| 31 | 2,4,6-Trichlorophenol | 88-06-2 | STBJ5914 | 99% | 1,004.4 | µg/mL | +/- | 36.5442 |
| 32 | 2,4,5-Trichlorophenol | 95-95-4 | FHN01 | 98% | 1,001.9 | µg/mL | +/- | 36.4512 |
| 33 | 2-Chloronaphthalene | 91-58-7 | RPN7O | 99% | 1,001.1 | µg/mL | +/- | 36.4230 |
| 34 | 2-Nitroaniline | 88-74-4 | RP230531 | 99% | 1,002.9 | µg/mL | +/- | 36.4876 |
| 35 | 1,4-Dinitrobenzene | 100-25-4 | RP230816 | 99% | 1,005.7 | µg/mL | +/- | 36.5887 |
| 36 | Acenaphthylene | 208-96-8 | p06V | 98% | 1,009.5 | µg/mL | +/- | 36.7265 |
| 37 | 1,3-Dinitrobenzene | 99-65-0 | 1-DXX-24-1 | 99% | 1,004.4 | µg/mL | +/- | 36.5422 |
| 38 | Dimethylphthalate | 131-11-3 | 358221L17K | 99% | 1,005.9 | µg/mL | +/- | 36.5968 |
| 39 | 2,6-Dinitrotoluene | 606-20-2 | BCCG1833 | 99% | 1,003.2 | µg/mL | +/- | 36.4998 |
| 40 | 1,2-Dinitrobenzene | 528-29-0 | RP230428 | 99% | 1,002.2 | µg/mL | +/- | 36.4634 |
| 41 | Acenaphthene | 83-32-9 | MKCR7169 | 99% | 1,009.3 | µg/mL | +/- | 36.7221 |
| 42 | 3-Nitroaniline | 99-09-2 | RP230822RSR | 99% | 1,003.9 | µg/mL | +/- | 36.5240 |
| 43 | 2,4-Dinitrophenol | 51-28-5 | DR230417RSR | 99% | 1,002.0 | µg/mL | +/- | 36.4553 |
| 44 | Dibenzofuran | 132-64-9 | MKCD9952 | 99% | 1,006.7 | µg/mL | +/- | 36.6251 |
| 45 | 2,4-Dinitrotoluene | 121-14-2 | MKAA0690V | 99% | 1,003.8 | µg/mL | +/- | 36.5220 |
| 46 | 4-Nitrophenol | 100-02-7 | RP230627 | 99% | 1,002.3 | µg/mL | +/- | 36.4674 |
| 47 | 2,3,4,6-Tetrachlorophenol | 58-90-2 | PR-30126 | 99% | 1,008.7 | µg/mL | +/- | 36.6979 |
| 48 | 2,3,5,6-Tetrachlorophenol | 935-95-5 | RP230919 | 99% | 1,006.3 | µg/mL | +/- | 36.6130 |
| 49 | Fluorene | 86-73-7 | 10241100 | 99% | 1,008.3 | µg/mL | +/- | 36.6857 |
| 50 | 4-Chlorophenyl phenyl ether | 7005-72-3 | MKCT7248 | 99% | 1,003.8 | µg/mL | +/- | 36.5220 |
| 51 | Diethylphthalate | 84-66-2 | MKCD2547 | 99% | 1,008.6 | µg/mL | +/- | 36.6958 |
| 52 | 4-Nitroaniline | 100-01-6 | RP230111 | 99% | 1,001.1 | µg/mL | +/- | 36.4230 |
| 53 | 4,6-Dinitro-2-methylphenol (Dinitro-o-cresol) | 534-52-1 | 230718JLM | 99% | 1,002.0 | µg/mL | +/- | 36.4553 |

| | | | | | | | |
|----|----------------------------|----------|---------------|-----|---------|-------|-------------|
| 54 | Diphenylamine | 122-39-4 | MKCH1042 | 99% | 1,002.3 | µg/mL | +/- 36.4674 |
| 55 | Azobenzene | 103-33-3 | BCCK0887 | 99% | 1,005.8 | µg/mL | +/- 36.5928 |
| 56 | 4-Bromophenyl phenyl ether | 101-55-3 | STBH6361 | 99% | 1,003.0 | µg/mL | +/- 36.4917 |
| 57 | Hexachlorobenzene | 118-74-1 | 14821700 | 99% | 1,007.5 | µg/mL | +/- 36.6554 |
| 58 | Pentachlorophenol | 87-86-5 | RP230530RSR | 99% | 1,008.8 | µg/mL | +/- 36.7019 |
| 59 | Phenanthrene | 85-01-8 | MKCQ8876 | 99% | 1,008.4 | µg/mL | +/- 36.6877 |
| 60 | Anthracene | 120-12-7 | MKCR0570 | 99% | 1,009.0 | µg/mL | +/- 36.7100 |
| 61 | Carbazole | 86-74-8 | 14351100 | 99% | 1,000.9 | µg/mL | +/- 36.4149 |
| 62 | Di-n-butylphthalate | 84-74-2 | MKCN4337 | 99% | 1,007.6 | µg/mL | +/- 36.6595 |
| 63 | Fluoranthene | 206-44-0 | MKCQ4728 | 99% | 1,009.6 | µg/mL | +/- 36.7302 |
| 64 | Pyrene | 129-00-0 | BCCG8479 | 98% | 1,007.2 | µg/mL | +/- 36.6453 |
| 65 | Benzyl butyl phthalate | 85-68-7 | X12I018 | 99% | 1,002.1 | µg/mL | +/- 36.4573 |
| 66 | Bis(2-ethylhexyl)adipate | 103-23-1 | MKCM1988 | 99% | 1,005.2 | µg/mL | +/- 36.5705 |
| 67 | Benz(a)anthracene | 56-55-3 | I220012022BAA | 99% | 1,002.2 | µg/mL | +/- 36.4614 |
| 68 | Chrysene | 218-01-9 | RP230601 | 99% | 1,008.3 | µg/mL | +/- 36.6837 |
| 69 | Bis(2-ethylhexyl)phthalate | 117-81-7 | MKCQ3468 | 99% | 1,001.8 | µg/mL | +/- 36.4472 |
| 70 | Di-n-octyl phthalate | 117-84-0 | 14382700 | 99% | 1,006.0 | µg/mL | +/- 36.6008 |
| 71 | Benzo(b)fluoranthene | 205-99-2 | 012013B | 99% | 1,002.8 | µg/mL | +/- 36.4836 |
| 72 | Benzo(k)fluoranthene | 207-08-9 | 012022K | 99% | 1,003.0 | µg/mL | +/- 36.4917 |
| 73 | Benzo(a)pyrene | 50-32-8 | P54915-0703 | 99% | 1,002.3 | µg/mL | +/- 36.4674 |
| 74 | Indeno(1,2,3-cd)pyrene | 193-39-5 | 12-JKL-118-9 | 97% | 1,009.4 | µg/mL | +/- 36.7243 |
| 75 | Dibenz(a,h)anthracene | 53-70-3 | 2-ASA-59-1 | 99% | 1,007.6 | µg/mL | +/- 36.6595 |
| 76 | Benzo(g,h,i)perylene | 191-24-2 | RP231003RSR | 99% | 1,002.9 | µg/mL | +/- 36.4876 |

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride

CAS # 75-09-2

Purity 99%



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL



Certificate of Analysis

chromatographic plus

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31850

Lot No.: A0203726

Description : 8270 MegaMix®

8270 MegaMix® 500-1000 µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : April 30, 2025

Storage: 0°C or colder

Handling: Sonication required. Mix is photosensitive.

Ship: Ambient

512117 } RC/
↓ 03/18/24
512146

C E R T I F I E D V A L U E S

| Elution Order | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|---------------|------------------------------|----------|-------------|--------|--------------------------------|---|
| 1 | Pyridine | 110-86-1 | SHBP6240 | 99% | 1,001.6 µg/mL | +/- 36.4412 |
| 2 | N-Nitrosodimethylamine | 62-75-9 | 230209JLM | 99% | 1,005.9 µg/mL | +/- 36.5968 |
| 3 | Phenol | 108-95-2 | MKCK1120 | 99% | 1,003.3 µg/mL | +/- 36.5038 |
| 4 | Aniline | 62-53-3 | X22F726 | 99% | 1,005.8 µg/mL | +/- 36.5928 |
| 5 | Bis(2-chloroethyl)ether | 111-44-4 | SHBL6942 | 99% | 1,008.1 µg/mL | +/- 36.6776 |
| 6 | 2-Chlorophenol | 95-57-8 | STBJ3909 | 99% | 1,001.8 µg/mL | +/- 36.4492 |
| 7 | 1,3-Dichlorobenzene | 541-73-1 | BCCD5315 | 99% | 1,002.3 µg/mL | +/- 36.4654 |
| 8 | 1,4-Dichlorobenzene | 106-46-7 | MKBS7929V | 99% | 1,003.7 µg/mL | +/- 36.5159 |
| 9 | Benzyl alcohol | 100-51-6 | SHBK5469 | 99% | 1,008.7 µg/mL | +/- 36.6979 |
| 10 | 1,2-Dichlorobenzene | 95-50-1 | SHBN3835 | 99% | 1,000.3 µg/mL | +/- 36.3926 |
| 11 | 2-Methylphenol (o-cresol) | 95-48-7 | SHBN7598 | 99% | 1,003.5 µg/mL | +/- 36.5099 |
| 12 | 2,2'-oxybis(1-chloropropane) | 108-60-1 | 29-MAR-45-5 | 99% | 1,007.3 µg/mL | +/- 36.6493 |
| 13 | 3-Methylphenol (m-cresol) | 108-39-4 | STBJ0710 | 99% | 504.3 µg/mL | +/- 18.3500 |
| 14 | 4-Methylphenol (p-cresol) | 106-44-5 | SHBN3411 | 99% | 503.6 µg/mL | +/- 18.3237 |
| 15 | N-Nitroso-di-n-propylamine | 621-64-7 | N63MG | 99% | 1,008.3 µg/mL | +/- 36.6857 |
| 16 | Hexachloroethane | 67-72-1 | QTORH | 99% | 1,007.5 µg/mL | +/- 36.6554 |
| 17 | Nitrobenzene | 98-95-3 | 10224044 | 99% | 1,008.6 µg/mL | +/- 36.6938 |

| | | | | | | | | |
|----|---|-----------|-------------|-----|---------|-------|-----|---------|
| 18 | Isophorone | 78-59-1 | MKCC9506 | 99% | 1,005.9 | µg/mL | +/- | 36.5988 |
| 19 | 2-Nitrophenol | 88-75-5 | RP230710 | 99% | 1,003.2 | µg/mL | +/- | 36.4998 |
| 20 | 2,4-Dimethylphenol | 105-67-9 | XW5GK | 99% | 1,003.8 | µg/mL | +/- | 36.5200 |
| 21 | Bis(2-chloroethoxy)methane | 111-91-1 | 13670200 | 99% | 1,002.1 | µg/mL | +/- | 36.4573 |
| 22 | 2,4-Dichlorophenol | 120-83-2 | BCBZ6787 | 99% | 1,003.7 | µg/mL | +/- | 36.5180 |
| 23 | 1,2,4-Trichlorobenzene | 120-82-1 | SHBP5900 | 99% | 1,007.6 | µg/mL | +/- | 36.6574 |
| 24 | Naphthalene | 91-20-3 | STBL1057 | 99% | 1,008.3 | µg/mL | +/- | 36.6837 |
| 25 | 4-Chloroaniline | 106-47-8 | BCCJ3217 | 99% | 1,001.3 | µg/mL | +/- | 36.4290 |
| 26 | Hexachlorobutadiene | 87-68-3 | RP230823RSR | 98% | 1,008.3 | µg/mL | +/- | 36.6829 |
| 27 | 4-Chloro-3-methylphenol | 59-50-7 | BCCD4461 | 99% | 1,003.1 | µg/mL | +/- | 36.4937 |
| 28 | 2-Methylnaphthalene | 91-57-6 | STBK0259 | 96% | 1,001.9 | µg/mL | +/- | 36.4505 |
| 29 | 1-Methylnaphthalene | 90-12-0 | 5234.00-8 | 98% | 1,000.0 | µg/mL | +/- | 36.3838 |
| 30 | Hexachlorocyclopentadiene | 77-47-4 | 099063I14L | 98% | 1,008.5 | µg/mL | +/- | 36.6909 |
| 31 | 2,4,6-Trichlorophenol | 88-06-2 | STBJ5914 | 99% | 1,004.4 | µg/mL | +/- | 36.5442 |
| 32 | 2,4,5-Trichlorophenol | 95-95-4 | FHN01 | 98% | 1,001.9 | µg/mL | +/- | 36.4512 |
| 33 | 2-Chloronaphthalene | 91-58-7 | RPN7O | 99% | 1,001.1 | µg/mL | +/- | 36.4230 |
| 34 | 2-Nitroaniline | 88-74-4 | RP230531 | 99% | 1,002.9 | µg/mL | +/- | 36.4876 |
| 35 | 1,4-Dinitrobenzene | 100-25-4 | RP230816 | 99% | 1,005.7 | µg/mL | +/- | 36.5887 |
| 36 | Acenaphthylene | 208-96-8 | p06V | 98% | 1,009.5 | µg/mL | +/- | 36.7265 |
| 37 | 1,3-Dinitrobenzene | 99-65-0 | 1-DXX-24-1 | 99% | 1,004.4 | µg/mL | +/- | 36.5422 |
| 38 | Dimethylphthalate | 131-11-3 | 358221L17K | 99% | 1,005.9 | µg/mL | +/- | 36.5968 |
| 39 | 2,6-Dinitrotoluene | 606-20-2 | BCCG1833 | 99% | 1,003.2 | µg/mL | +/- | 36.4998 |
| 40 | 1,2-Dinitrobenzene | 528-29-0 | RP230428 | 99% | 1,002.2 | µg/mL | +/- | 36.4634 |
| 41 | Acenaphthene | 83-32-9 | MKCR7169 | 99% | 1,009.3 | µg/mL | +/- | 36.7221 |
| 42 | 3-Nitroaniline | 99-09-2 | RP230822RSR | 99% | 1,003.9 | µg/mL | +/- | 36.5240 |
| 43 | 2,4-Dinitrophenol | 51-28-5 | DR230417RSR | 99% | 1,002.0 | µg/mL | +/- | 36.4553 |
| 44 | Dibenzofuran | 132-64-9 | MKCD9952 | 99% | 1,006.7 | µg/mL | +/- | 36.6251 |
| 45 | 2,4-Dinitrotoluene | 121-14-2 | MKAA0690V | 99% | 1,003.8 | µg/mL | +/- | 36.5220 |
| 46 | 4-Nitrophenol | 100-02-7 | RP230627 | 99% | 1,002.3 | µg/mL | +/- | 36.4674 |
| 47 | 2,3,4,6-Tetrachlorophenol | 58-90-2 | PR-30126 | 99% | 1,008.7 | µg/mL | +/- | 36.6979 |
| 48 | 2,3,5,6-Tetrachlorophenol | 935-95-5 | RP230919 | 99% | 1,006.3 | µg/mL | +/- | 36.6130 |
| 49 | Fluorene | 86-73-7 | 10241100 | 99% | 1,008.3 | µg/mL | +/- | 36.6857 |
| 50 | 4-Chlorophenyl phenyl ether | 7005-72-3 | MKCT7248 | 99% | 1,003.8 | µg/mL | +/- | 36.5220 |
| 51 | Diethylphthalate | 84-66-2 | MKCD2547 | 99% | 1,008.6 | µg/mL | +/- | 36.6958 |
| 52 | 4-Nitroaniline | 100-01-6 | RP230111 | 99% | 1,001.1 | µg/mL | +/- | 36.4230 |
| 53 | 4,6-Dinitro-2-methylphenol (Dinitro-o-cresol) | 534-52-1 | 230718JLM | 99% | 1,002.0 | µg/mL | +/- | 36.4553 |

| | | | | | | | |
|----|----------------------------|----------|---------------|-----|---------|-------|-------------|
| 54 | Diphenylamine | 122-39-4 | MKCH1042 | 99% | 1,002.3 | µg/mL | +/- 36.4674 |
| 55 | Azobenzene | 103-33-3 | BCCK0887 | 99% | 1,005.8 | µg/mL | +/- 36.5928 |
| 56 | 4-Bromophenyl phenyl ether | 101-55-3 | STBH6361 | 99% | 1,003.0 | µg/mL | +/- 36.4917 |
| 57 | Hexachlorobenzene | 118-74-1 | 14821700 | 99% | 1,007.5 | µg/mL | +/- 36.6554 |
| 58 | Pentachlorophenol | 87-86-5 | RP230530RSR | 99% | 1,008.8 | µg/mL | +/- 36.7019 |
| 59 | Phenanthrene | 85-01-8 | MKCQ8876 | 99% | 1,008.4 | µg/mL | +/- 36.6877 |
| 60 | Anthracene | 120-12-7 | MKCR0570 | 99% | 1,009.0 | µg/mL | +/- 36.7100 |
| 61 | Carbazole | 86-74-8 | 14351100 | 99% | 1,000.9 | µg/mL | +/- 36.4149 |
| 62 | Di-n-butylphthalate | 84-74-2 | MKCN4337 | 99% | 1,007.6 | µg/mL | +/- 36.6595 |
| 63 | Fluoranthene | 206-44-0 | MKCQ4728 | 99% | 1,009.6 | µg/mL | +/- 36.7302 |
| 64 | Pyrene | 129-00-0 | BCCG8479 | 98% | 1,007.2 | µg/mL | +/- 36.6453 |
| 65 | Benzyl butyl phthalate | 85-68-7 | X12I018 | 99% | 1,002.1 | µg/mL | +/- 36.4573 |
| 66 | Bis(2-ethylhexyl)adipate | 103-23-1 | MKCM1988 | 99% | 1,005.2 | µg/mL | +/- 36.5705 |
| 67 | Benz(a)anthracene | 56-55-3 | I220012022BAA | 99% | 1,002.2 | µg/mL | +/- 36.4614 |
| 68 | Chrysene | 218-01-9 | RP230601 | 99% | 1,008.3 | µg/mL | +/- 36.6837 |
| 69 | Bis(2-ethylhexyl)phthalate | 117-81-7 | MKCQ3468 | 99% | 1,001.8 | µg/mL | +/- 36.4472 |
| 70 | Di-n-octyl phthalate | 117-84-0 | 14382700 | 99% | 1,006.0 | µg/mL | +/- 36.6008 |
| 71 | Benzo(b)fluoranthene | 205-99-2 | 012013B | 99% | 1,002.8 | µg/mL | +/- 36.4836 |
| 72 | Benzo(k)fluoranthene | 207-08-9 | 012022K | 99% | 1,003.0 | µg/mL | +/- 36.4917 |
| 73 | Benzo(a)pyrene | 50-32-8 | P54915-0703 | 99% | 1,002.3 | µg/mL | +/- 36.4674 |
| 74 | Indeno(1,2,3-cd)pyrene | 193-39-5 | 12-JKL-118-9 | 97% | 1,009.4 | µg/mL | +/- 36.7243 |
| 75 | Dibenz(a,h)anthracene | 53-70-3 | 2-ASA-59-1 | 99% | 1,007.6 | µg/mL | +/- 36.6595 |
| 76 | Benzo(g,h,i)perylene | 191-24-2 | RP231003RSR | 99% | 1,002.9 | µg/mL | +/- 36.4876 |

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride

CAS # 75-09-2

Purity 99%



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL



Certificate of Analysis

chromatographic plus

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31850

Lot No.: A0203726

Description : 8270 MegaMix®

8270 MegaMix® 500-1000 µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : April 30, 2025

Storage: 0°C or colder

Handling: Sonication required. Mix is photosensitive.

Ship: Ambient

512117 } RC/
↓ 03/18/24
512146

C E R T I F I E D V A L U E S

| Elution Order | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|---------------|------------------------------|----------|-------------|--------|--------------------------------|---|
| 1 | Pyridine | 110-86-1 | SHBP6240 | 99% | 1,001.6 µg/mL | +/- 36.4412 |
| 2 | N-Nitrosodimethylamine | 62-75-9 | 230209JLM | 99% | 1,005.9 µg/mL | +/- 36.5968 |
| 3 | Phenol | 108-95-2 | MKCK1120 | 99% | 1,003.3 µg/mL | +/- 36.5038 |
| 4 | Aniline | 62-53-3 | X22F726 | 99% | 1,005.8 µg/mL | +/- 36.5928 |
| 5 | Bis(2-chloroethyl)ether | 111-44-4 | SHBL6942 | 99% | 1,008.1 µg/mL | +/- 36.6776 |
| 6 | 2-Chlorophenol | 95-57-8 | STBJ3909 | 99% | 1,001.8 µg/mL | +/- 36.4492 |
| 7 | 1,3-Dichlorobenzene | 541-73-1 | BCCD5315 | 99% | 1,002.3 µg/mL | +/- 36.4654 |
| 8 | 1,4-Dichlorobenzene | 106-46-7 | MKBS7929V | 99% | 1,003.7 µg/mL | +/- 36.5159 |
| 9 | Benzyl alcohol | 100-51-6 | SHBK5469 | 99% | 1,008.7 µg/mL | +/- 36.6979 |
| 10 | 1,2-Dichlorobenzene | 95-50-1 | SHBN3835 | 99% | 1,000.3 µg/mL | +/- 36.3926 |
| 11 | 2-Methylphenol (o-cresol) | 95-48-7 | SHBN7598 | 99% | 1,003.5 µg/mL | +/- 36.5099 |
| 12 | 2,2'-oxybis(1-chloropropane) | 108-60-1 | 29-MAR-45-5 | 99% | 1,007.3 µg/mL | +/- 36.6493 |
| 13 | 3-Methylphenol (m-cresol) | 108-39-4 | STBJ0710 | 99% | 504.3 µg/mL | +/- 18.3500 |
| 14 | 4-Methylphenol (p-cresol) | 106-44-5 | SHBN3411 | 99% | 503.6 µg/mL | +/- 18.3237 |
| 15 | N-Nitroso-di-n-propylamine | 621-64-7 | N63MG | 99% | 1,008.3 µg/mL | +/- 36.6857 |
| 16 | Hexachloroethane | 67-72-1 | QTORH | 99% | 1,007.5 µg/mL | +/- 36.6554 |
| 17 | Nitrobenzene | 98-95-3 | 10224044 | 99% | 1,008.6 µg/mL | +/- 36.6938 |

| | | | | | | | | |
|----|---|-----------|-------------|-----|---------|-------|-----|---------|
| 18 | Isophorone | 78-59-1 | MKCC9506 | 99% | 1,005.9 | µg/mL | +/- | 36.5988 |
| 19 | 2-Nitrophenol | 88-75-5 | RP230710 | 99% | 1,003.2 | µg/mL | +/- | 36.4998 |
| 20 | 2,4-Dimethylphenol | 105-67-9 | XW5GK | 99% | 1,003.8 | µg/mL | +/- | 36.5200 |
| 21 | Bis(2-chloroethoxy)methane | 111-91-1 | 13670200 | 99% | 1,002.1 | µg/mL | +/- | 36.4573 |
| 22 | 2,4-Dichlorophenol | 120-83-2 | BCBZ6787 | 99% | 1,003.7 | µg/mL | +/- | 36.5180 |
| 23 | 1,2,4-Trichlorobenzene | 120-82-1 | SHBP5900 | 99% | 1,007.6 | µg/mL | +/- | 36.6574 |
| 24 | Naphthalene | 91-20-3 | STBL1057 | 99% | 1,008.3 | µg/mL | +/- | 36.6837 |
| 25 | 4-Chloroaniline | 106-47-8 | BCCJ3217 | 99% | 1,001.3 | µg/mL | +/- | 36.4290 |
| 26 | Hexachlorobutadiene | 87-68-3 | RP230823RSR | 98% | 1,008.3 | µg/mL | +/- | 36.6829 |
| 27 | 4-Chloro-3-methylphenol | 59-50-7 | BCCD4461 | 99% | 1,003.1 | µg/mL | +/- | 36.4937 |
| 28 | 2-Methylnaphthalene | 91-57-6 | STBK0259 | 96% | 1,001.9 | µg/mL | +/- | 36.4505 |
| 29 | 1-Methylnaphthalene | 90-12-0 | 5234.00-8 | 98% | 1,000.0 | µg/mL | +/- | 36.3838 |
| 30 | Hexachlorocyclopentadiene | 77-47-4 | 099063I14L | 98% | 1,008.5 | µg/mL | +/- | 36.6909 |
| 31 | 2,4,6-Trichlorophenol | 88-06-2 | STBJ5914 | 99% | 1,004.4 | µg/mL | +/- | 36.5442 |
| 32 | 2,4,5-Trichlorophenol | 95-95-4 | FHN01 | 98% | 1,001.9 | µg/mL | +/- | 36.4512 |
| 33 | 2-Chloronaphthalene | 91-58-7 | RPN7O | 99% | 1,001.1 | µg/mL | +/- | 36.4230 |
| 34 | 2-Nitroaniline | 88-74-4 | RP230531 | 99% | 1,002.9 | µg/mL | +/- | 36.4876 |
| 35 | 1,4-Dinitrobenzene | 100-25-4 | RP230816 | 99% | 1,005.7 | µg/mL | +/- | 36.5887 |
| 36 | Acenaphthylene | 208-96-8 | p06V | 98% | 1,009.5 | µg/mL | +/- | 36.7265 |
| 37 | 1,3-Dinitrobenzene | 99-65-0 | 1-DXX-24-1 | 99% | 1,004.4 | µg/mL | +/- | 36.5422 |
| 38 | Dimethylphthalate | 131-11-3 | 358221L17K | 99% | 1,005.9 | µg/mL | +/- | 36.5968 |
| 39 | 2,6-Dinitrotoluene | 606-20-2 | BCCG1833 | 99% | 1,003.2 | µg/mL | +/- | 36.4998 |
| 40 | 1,2-Dinitrobenzene | 528-29-0 | RP230428 | 99% | 1,002.2 | µg/mL | +/- | 36.4634 |
| 41 | Acenaphthene | 83-32-9 | MKCR7169 | 99% | 1,009.3 | µg/mL | +/- | 36.7221 |
| 42 | 3-Nitroaniline | 99-09-2 | RP230822RSR | 99% | 1,003.9 | µg/mL | +/- | 36.5240 |
| 43 | 2,4-Dinitrophenol | 51-28-5 | DR230417RSR | 99% | 1,002.0 | µg/mL | +/- | 36.4553 |
| 44 | Dibenzofuran | 132-64-9 | MKCD9952 | 99% | 1,006.7 | µg/mL | +/- | 36.6251 |
| 45 | 2,4-Dinitrotoluene | 121-14-2 | MKAA0690V | 99% | 1,003.8 | µg/mL | +/- | 36.5220 |
| 46 | 4-Nitrophenol | 100-02-7 | RP230627 | 99% | 1,002.3 | µg/mL | +/- | 36.4674 |
| 47 | 2,3,4,6-Tetrachlorophenol | 58-90-2 | PR-30126 | 99% | 1,008.7 | µg/mL | +/- | 36.6979 |
| 48 | 2,3,5,6-Tetrachlorophenol | 935-95-5 | RP230919 | 99% | 1,006.3 | µg/mL | +/- | 36.6130 |
| 49 | Fluorene | 86-73-7 | 10241100 | 99% | 1,008.3 | µg/mL | +/- | 36.6857 |
| 50 | 4-Chlorophenyl phenyl ether | 7005-72-3 | MKCT7248 | 99% | 1,003.8 | µg/mL | +/- | 36.5220 |
| 51 | Diethylphthalate | 84-66-2 | MKCD2547 | 99% | 1,008.6 | µg/mL | +/- | 36.6958 |
| 52 | 4-Nitroaniline | 100-01-6 | RP230111 | 99% | 1,001.1 | µg/mL | +/- | 36.4230 |
| 53 | 4,6-Dinitro-2-methylphenol (Dinitro-o-cresol) | 534-52-1 | 230718JLM | 99% | 1,002.0 | µg/mL | +/- | 36.4553 |

| | | | | | | | |
|----|----------------------------|----------|---------------|-----|---------|-------|-------------|
| 54 | Diphenylamine | 122-39-4 | MKCH1042 | 99% | 1,002.3 | µg/mL | +/- 36.4674 |
| 55 | Azobenzene | 103-33-3 | BCCK0887 | 99% | 1,005.8 | µg/mL | +/- 36.5928 |
| 56 | 4-Bromophenyl phenyl ether | 101-55-3 | STBH6361 | 99% | 1,003.0 | µg/mL | +/- 36.4917 |
| 57 | Hexachlorobenzene | 118-74-1 | 14821700 | 99% | 1,007.5 | µg/mL | +/- 36.6554 |
| 58 | Pentachlorophenol | 87-86-5 | RP230530RSR | 99% | 1,008.8 | µg/mL | +/- 36.7019 |
| 59 | Phenanthrene | 85-01-8 | MKCQ8876 | 99% | 1,008.4 | µg/mL | +/- 36.6877 |
| 60 | Anthracene | 120-12-7 | MKCR0570 | 99% | 1,009.0 | µg/mL | +/- 36.7100 |
| 61 | Carbazole | 86-74-8 | 14351100 | 99% | 1,000.9 | µg/mL | +/- 36.4149 |
| 62 | Di-n-butylphthalate | 84-74-2 | MKCN4337 | 99% | 1,007.6 | µg/mL | +/- 36.6595 |
| 63 | Fluoranthene | 206-44-0 | MKCQ4728 | 99% | 1,009.6 | µg/mL | +/- 36.7302 |
| 64 | Pyrene | 129-00-0 | BCCG8479 | 98% | 1,007.2 | µg/mL | +/- 36.6453 |
| 65 | Benzyl butyl phthalate | 85-68-7 | X12I018 | 99% | 1,002.1 | µg/mL | +/- 36.4573 |
| 66 | Bis(2-ethylhexyl)adipate | 103-23-1 | MKCM1988 | 99% | 1,005.2 | µg/mL | +/- 36.5705 |
| 67 | Benz(a)anthracene | 56-55-3 | I220012022BAA | 99% | 1,002.2 | µg/mL | +/- 36.4614 |
| 68 | Chrysene | 218-01-9 | RP230601 | 99% | 1,008.3 | µg/mL | +/- 36.6837 |
| 69 | Bis(2-ethylhexyl)phthalate | 117-81-7 | MKCQ3468 | 99% | 1,001.8 | µg/mL | +/- 36.4472 |
| 70 | Di-n-octyl phthalate | 117-84-0 | 14382700 | 99% | 1,006.0 | µg/mL | +/- 36.6008 |
| 71 | Benzo(b)fluoranthene | 205-99-2 | 012013B | 99% | 1,002.8 | µg/mL | +/- 36.4836 |
| 72 | Benzo(k)fluoranthene | 207-08-9 | 012022K | 99% | 1,003.0 | µg/mL | +/- 36.4917 |
| 73 | Benzo(a)pyrene | 50-32-8 | P54915-0703 | 99% | 1,002.3 | µg/mL | +/- 36.4674 |
| 74 | Indeno(1,2,3-cd)pyrene | 193-39-5 | 12-JKL-118-9 | 97% | 1,009.4 | µg/mL | +/- 36.7243 |
| 75 | Dibenz(a,h)anthracene | 53-70-3 | 2-ASA-59-1 | 99% | 1,007.6 | µg/mL | +/- 36.6595 |
| 76 | Benzo(g,h,i)perylene | 191-24-2 | RP231003RSR | 99% | 1,002.9 | µg/mL | +/- 36.4876 |

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride

CAS # 75-09-2

Purity 99%



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL



Certificate of Analysis

gravimetric

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555223 **Lot No.:** A0214021

Description : Custom 8270 Plus Standard #1

Custom 8270 Plus Standard #1 1,000 μ g/mL, Methylene Chloride,
1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2026 **Storage:** 10°C or colder

Handling: This product is photosensitive. **Ship:** Ambient

C E R T I F I E D V A L U E S

| Component # | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|-------------|------------------------|-----------|------------|--------|--------------------------------|--|
| 1 | 3,3'-Dichlorobenzidine | 91-94-1 | S240326RSR | 99% | 1,004.0 μ g/mL | +/- 23.0487 |
| 2 | Atrazine | 1912-24-9 | 5FYWL | 99% | 1,005.0 μ g/mL | +/- 23.0717 |
| 3 | Benzidine | 92-87-5 | S240430RSR | 99% | 1,006.0 μ g/mL | +/- 23.0947 |
| 4 | epsilon-Caprolactam | 105-60-2 | Y16H012 | 99% | 1,000.0 μ g/mL | +/- 22.9569 |

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

S12449 } RC/
↓ } 7/24/24
S12508 }

Rebecca Gingerich - Operations Tech II

Date Mixed: 18-Jul-2024

Balance: 1128353505

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{\text{combined uncertainty}} = k \sqrt{u_{\text{gravimetric}}^2 + u_{\text{homogeneity}}^2 + u_{\text{storage stability}}^2 + u_{\text{shipping stability}}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL



Certificate of Analysis

gravimetric

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555223 **Lot No.:** A0214021

Description : Custom 8270 Plus Standard #1

Custom 8270 Plus Standard #1 1,000 μ g/mL, Methylene Chloride,
1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2026 **Storage:** 10°C or colder

Handling: This product is photosensitive. **Ship:** Ambient

C E R T I F I E D V A L U E S

| Component # | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|-------------|------------------------|-----------|------------|--------|--------------------------------|--|
| 1 | 3,3'-Dichlorobenzidine | 91-94-1 | S240326RSR | 99% | 1,004.0 μ g/mL | +/- 23.0487 |
| 2 | Atrazine | 1912-24-9 | 5FYWL | 99% | 1,005.0 μ g/mL | +/- 23.0717 |
| 3 | Benzidine | 92-87-5 | S240430RSR | 99% | 1,006.0 μ g/mL | +/- 23.0947 |
| 4 | epsilon-Caprolactam | 105-60-2 | Y16H012 | 99% | 1,000.0 μ g/mL | +/- 22.9569 |

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

S12449 } RC/
↓ } 7/24/24
S12508 }

Rebecca Gingerich - Operations Tech II

Date Mixed: 18-Jul-2024

Balance: 1128353505

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{\text{combined uncertainty}} = k \sqrt{u_{\text{gravimetric}}^2 + u_{\text{homogeneity}}^2 + u_{\text{storage stability}}^2 + u_{\text{shipping stability}}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL



Certificate of Analysis

gravimetric

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555223

Lot No.: A0214021

Description : Custom 8270 Plus Standard #1

Custom 8270 Plus Standard #1 1,000 μ g/mL, Methylene Chloride,
1mL/ampul

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : July 31, 2026

Storage: 10°C or colder

Handling: This product is photosensitive.

Ship: Ambient

C E R T I F I E D V A L U E S

| Component # | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|-------------|------------------------|-----------|------------|--------|--------------------------------|--|
| 1 | 3,3'-Dichlorobenzidine | 91-94-1 | S240326RSR | 99% | 1,004.0 μ g/mL | +/- 23.0487 |
| 2 | Atrazine | 1912-24-9 | 5FYWL | 99% | 1,005.0 μ g/mL | +/- 23.0717 |
| 3 | Benzidine | 92-87-5 | S240430RSR | 99% | 1,006.0 μ g/mL | +/- 23.0947 |
| 4 | epsilon-Caprolactam | 105-60-2 | Y16H012 | 99% | 1,000.0 μ g/mL | +/- 22.9569 |

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

S12449 } RC/
↓ } 7/24/24
S12508 }

Rebecca Gingerich - Operations Tech II

Date Mixed: 18-Jul-2024

Balance: 1128353505

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{\text{combined uncertainty}} = k \sqrt{u_{\text{gravimetric}}^2 + u_{\text{homogeneity}}^2 + u_{\text{storage stability}}^2 + u_{\text{shipping stability}}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL



Certificate of Analysis

gravimetric

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555223 **Lot No.:** A0214021

Description : Custom 8270 Plus Standard #1

Custom 8270 Plus Standard #1 1,000 μ g/mL, Methylene Chloride,
1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2026 **Storage:** 10°C or colder

Handling: This product is photosensitive. **Ship:** Ambient

C E R T I F I E D V A L U E S

| Component # | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|-------------|------------------------|-----------|------------|--------|--------------------------------|--|
| 1 | 3,3'-Dichlorobenzidine | 91-94-1 | S240326RSR | 99% | 1,004.0 μ g/mL | +/- 23.0487 |
| 2 | Atrazine | 1912-24-9 | 5FYWL | 99% | 1,005.0 μ g/mL | +/- 23.0717 |
| 3 | Benzidine | 92-87-5 | S240430RSR | 99% | 1,006.0 μ g/mL | +/- 23.0947 |
| 4 | epsilon-Caprolactam | 105-60-2 | Y16H012 | 99% | 1,000.0 μ g/mL | +/- 22.9569 |

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

S12449 } RC/
↓ } 7/24/24
S12508 }

Rebecca Gingerich - Operations Tech II

Date Mixed: 18-Jul-2024

Balance: 1128353505

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{\text{combined uncertainty}} = k \sqrt{u_{\text{gravimetric}}^2 + u_{\text{homogeneity}}^2 + u_{\text{storage stability}}^2 + u_{\text{shipping stability}}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL



Certificate of Analysis

gravimetric

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555223

Lot No.: A0214021

Description : Custom 8270 Plus Standard #1

Custom 8270 Plus Standard #1 1,000 μ g/mL, Methylene Chloride,
1mL/ampul

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : July 31, 2026

Storage: 10°C or colder

Handling: This product is photosensitive.

Ship: Ambient

C E R T I F I E D V A L U E S

| Componen t # | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|-----------------|------------------------|-----------|------------|--------|--------------------------------|--|
| 1 | 3,3'-Dichlorobenzidine | 91-94-1 | S240326RSR | 99% | 1,004.0 μ g/mL | +/- 23.0487 |
| 2 | Atrazine | 1912-24-9 | 5FYWL | 99% | 1,005.0 μ g/mL | +/- 23.0717 |
| 3 | Benzidine | 92-87-5 | S240430RSR | 99% | 1,006.0 μ g/mL | +/- 23.0947 |
| 4 | epsilon-Caprolactam | 105-60-2 | Y16H012 | 99% | 1,000.0 μ g/mL | +/- 22.9569 |

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

S12449 } RC/
↓ } 7/24/24
S12508 }

Rebecca Gingerich - Operations Tech II

Date Mixed: 18-Jul-2024

Balance: 1128353505

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{\text{combined uncertainty}} = k \sqrt{u_{\text{gravimetric}}^2 + u_{\text{homogeneity}}^2 + u_{\text{storage stability}}^2 + u_{\text{shipping stability}}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



SHIPPING DOCUMENTS

CLIENT INFORMATION

CLIENT PROJECT INFORMATION

CLIENT BILLING INFORMATION

REPORT TO BE SENT TO:

COMPANY: Jacobs

ADDRESS: 412 Mt. Kumble Ave Suite H100

CITY: Morristown STATE: NJ ZIP: 07960

ATTENTION: John Yantane

PHONE: (201) 444-1719 FAX:

PROJECT NAME: STC PTC

PROJECT NO.: D3779422 LOCATION: Princeton Junction

PROJECT MANAGER: Mary Murphy

e-mail: Mary.Murphy@Jacobs.com

PHONE: (201) 936-0586 FAX:

BILL TO: Mary Murphy

PO#:

ADDRESS:

CITY STATE ZIP:

ATTENTION: PHONE:

ANALYSIS

DATA TURNAROUND INFORMATION

FAX (RUSH) Standard TAT DAYS*

HARDCOPY (DATA PACKAGE): DAYS*

EDD: DAYS*

*TO BE APPROVED BY CHEMTECH

STANDARD HARDCOPY TURNAROUND TIME IS 10 BUSINESS DAYS

DATA DELIVERABLE INFORMATION

- Level 1 (Results Only) Level 4 (QC + Full Raw Data)
 Level 2 (Results + QC) NJ Reduced US EPA CLP
 Level 3 (Results + QC) NYS ASP A NYS ASP B
 + Raw Data Other _____
 EDD FORMAT

VOCs 8/26/01
 SPX - 8/26/01
 PATH - 8/26/01
 Rebar - 8/26/01
 241-A 602081-A9
 CV-CVII 7/14/01
 SW-Jacob's A

| CHEMTECH SAMPLE ID | PROJECT SAMPLE IDENTIFICATION | SAMPLE MATRIX | SAMPLE TYPE | | SAMPLE COLLECTION | | # OF BOTTLES | PRESERVATIVES | | | | | | | | | COMMENTS | | |
|--------------------------|----------------------------------|------------------|----------------|------|----------------------|------|--------------|----------------------------|---|---|---|---|---|---|---|---|-------------------------|---------|--|
| | | | COMP | GRAB | DATE | TIME | | A/E E B/E E | | | | | | | | | ← Specify Preservatives | | |
| | | | | | | | | 1 | 2 | 3 | 4 | 5 | 6 | 7 | 8 | 9 | A-HCl | D-NaOH | |
| 1. | 717-T-WS-081624 | WS | X | | 8-16-24 | 0930 | 8 | 2 | 4 | 1 | 1 | | | | | | B-HN03 | E-ICE | |
| 2. | TB-01-081624 | DI | X | | 8-16-24 | 1055 | 1 | 1 | | | | | | | | | C-H2SO4 | F-OTHER | |
| 3. | | | | | | | | | | | | | | | | | | | |
| 4. | | | | | | | | | | | | | | | | | | | |
| 5. | | | | | | | | | | | | | | | | | | | |
| 6. | | | | | | | | | | | | | | | | | | | |
| 7. | | | | | | | | | | | | | | | | | | | |
| 8. | | | | | | | | | | | | | | | | | | | |
| 9. | | | | | | | | | | | | | | | | | | | |
| 10. | | | | | | | | | | | | | | | | | | | |

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER:
1. *Murphy*

DATE/TIME: 8-16-24 1245

RECEIVED BY:
1. *J. Murphy*Conditions of bottles or coolers at receipt: COMPLIANT NON COMPLIANT COOLER TEMP 2-4 °C
 Comments: See attached table for this required analytes listRELINQUISHED BY SAMPLER:
2.

DATE/TIME:

RECEIVED BY:
2.RELINQUISHED BY SAMPLER:
3.

DATE/TIME:

RECEIVED BY:
3.Page 1 of 1 CLIENT: Hand Delivered Other
 CHEMTECH: Picked Up Field SamplingShipment Complete
 YES NO

Table 3. Surface Water Target Analytes, Methods, Action Levels, and Control Limits*Site Sampling Plan for Ecological Evaluation**Princeton Technology Center, West Windsor Township, New Jersey*

| Method | Analyte | CAS Number | Units | Higher of PQL and Ground Water Quality Criterion ^a | Fresh Surface Water Chronic NJDEP Ecological Criterion ^b |
|------------------|----------------------------------|------------|-------|---|---|
| ECO-SVOCs | | | | | |
| SW8270E | 1,4-Dioxane | 123-91-1 | µg/L | 0.4 | -- |
| SW8270E | 1-Methylnaphthalene | 90-12-0 | µg/L | -- | |
| SW8270E | 2,4,5-Trichlorophenol | 95-95-4 | µg/L | 700 | |
| SW8270E | 2,4,6-Trichlorophenol | 88-06-2 | µg/L | 20 | |
| SW8270E | 2,4-Dinitrotoluene | 121-14-2 | µg/L | 10 | |
| SW8270E | 2-Methylnaphthalene | 91-57-6 | µg/L | 30 | |
| SW8270E | 2-Methylphenol | 95-48-7 | µg/L | -- | |
| SW8270E | 3 & 4-Methylphenol (m,p-Cresols) | 65794-96-9 | µg/L | -- | |
| SW8270E | Acenaphthene | 83-32-9 | µg/L | 400 | |
| SW8270E | Acenaphthylene | 208-96-8 | µg/L | -- | |
| SW8270E | Anthracene | 120-12-7 | µg/L | 2000 | |
| SW8270E | Benzaldehyde | 100-52-7 | µg/L | -- | |
| SW8270E | Benzo(a)anthracene | 56-55-3 | µg/L | -- | |
| SW8270E | Benzo(a)pyrene | 50-32-8 | µg/L | 0.1 | |
| SW8270E | Benzo(b)fluoranthene | 205-99-2 | µg/L | 0.5 | |
| SW8270E | Benzo(g,h,i)perylene | 191-24-2 | µg/L | -- | |
| SW8270E | Benzo(k)fluoranthene | 207-08-9 | µg/L | 0.5 | |
| SW8270E | Bis (2-ethylhexyl) phthalate | 117-81-7 | µg/L | -- | |
| SW8270E | Carbazole | 86-74-8 | µg/L | -- | |
| SW8270E | Chrysene | 218-01-9 | µg/L | 5 | |
| SW8270E | Dibenzo(a,h)anthracene | 53-70-3 | µg/L | 0.3 | |
| SW8270E | Dibenzofuran | 132-64-9 | µg/L | -- | |
| SW8270E | Di-N-Butylphthalate | 84-74-2 | µg/L | -- | |
| SW8270E | Fluoranthene | 206-44-0 | µg/L | 300 | |
| SW8270E | Fluorene | 86-73-7 | µg/L | 300 | |
| SW8270E | Hexachlorobenzene | 118-74-1 | µg/L | 0.02 | |
| SW8270E | Hexachlorobutadiene | 87-68-3 | µg/L | 1 | |
| SW8270E | Hexachloroethane | 67-72-1 | µg/L | 7 | |
| SW8270E | Indeno(1,2,3-Cd)Pyrene | 193-39-5 | µg/L | 0.2 | |
| SW8270E | Naphthalene | 91-20-3 | µg/L | 300 | |
| SW8270E | Nitrobenzene | 98-95-3 | µg/L | 6 | |
| SW8270E | Pentachlorophenol | 87-86-5 | µg/L | 0.3 | |
| SW8270E | Phenanthrene | 85-01-8 | µg/L | -- | |
| SW8270E | Pyrene | 129-00-0 | µg/L | 200 | |
| SW8270E | Pyridine | 110-86-1 | µg/L | -- | |
| ECO-VOCs | | | | | |
| SW8260D | 1,1,1-Trichloroethane | 71-55-6 | µg/L | 30 | 76 |

Table 3. Surface Water Target Analytes, Methods, Action Levels, and Control Limits

*Site Sampling Plan for Ecological Evaluation**Princeton Technology Center, West Windsor Township, New Jersey*

| Method | Analyte | CAS Number | Units | Higher of PQL and Ground Water Quality Criterion ^a | Fresh Surface Water Chronic NJDEP Ecological Criterion ^b |
|-----------------|----------------------------|-------------|-------|---|---|
| SW8260D | 1,1,2-Trichloroethane | 79-00-5 | µg/L | 3 | 500 |
| SW8260D | 1,1-Dichloroethane | 75-34-3 | µg/L | 50 | -- |
| SW8260D | 1,1-Dichloroethene | 75-35-4 | µg/L | 1 | 65 |
| SW8260D | 1,2-Dichlorobenzene | 95-50-1 | µg/L | 600 | 14 |
| SW8260D | 1,2-Dichloroethane | 107-06-2 | µg/L | 2 | 910 |
| SW8260D | 1,2-Dichloroethene (Total) | 540-59-0 | µg/L | | |
| SW8260D | 1,4-Dichlorobenzene | 106-46-7 | µg/L | 75 | 9.4 |
| SW8260D | 2-Butanone | 78-93-3 | µg/L | 300 | -- |
| SW8260D | Acetone | 67-64-1 | µg/L | 6000 | -- |
| SW8260D | Benzene | 71-43-2 | µg/L | 1 | 114 |
| SW8260D | Bromodichloromethane | 75-27-4 | µg/L | 1 | -- |
| SW8260D | Bromomethane | 74-83-9 | µg/L | 10 | -- |
| SW8260D | Carbon disulfide | 75-15-0 | µg/L | 700 | -- |
| SW8260D | Carbon tetrachloride | 56-23-5 | µg/L | 1 | 240 |
| SW8260D | Chlorobenzene | 108-90-7 | µg/L | 50 | 47 |
| SW8260D | Chloroethane | 75-00-3 | µg/L | -- | -- |
| SW8260D | Chloroform | 67-66-3 | µg/L | 70 | 140 |
| SW8260D | Chloromethane | 74-87-3 | µg/L | -- | -- |
| SW8260D | cis-1,2-Dichloroethene | 156-59-2 | µg/L | 70 | -- |
| SW8260D | Cyclohexane | 110-82-7 | µg/L | -- | -- |
| SW8260D | Dibromochloromethane | 124-48-1 | µg/L | 1 | -- |
| SW8260D | Dichlorodifluoromethane | 75-71-8 | µg/L | 1000 | -- |
| SW8260D | Ethylbenzene | 100-41-4 | µg/L | 700 | 14 |
| SW8260D | Freon TF | 76-13-1 | µg/L | 20000 | -- |
| SW8260D | Isopropylbenzene | 98-82-8 | µg/L | 700 | -- |
| SW8260D | m&p-Xylene | 179601-23-1 | µg/L | 1000 | 27 |
| SW8260D | Methylcyclohexane | 108-87-2 | µg/L | -- | -- |
| SW8260D | Methylene Chloride | 75-09-2 | µg/L | 3 | 940 |
| SW8260D | MTBE | 1634-04-4 | µg/L | 70 | 51000 |
| SW8260D | o-Xylene | 95-47-6 | µg/L | 1000 | 27 |
| SW8260D | Tetrachloroethene | 127-18-4 | µg/L | 1 | 45 |
| SW8260D | Toluene | 108-88-3 | µg/L | 600 | 253 |
| SW8260D | trans-1,2-Dichloroethene | 156-60-5 | µg/L | 100 | 970 |
| SW8260D | Trichloroethene | 79-01-6 | µg/L | 1 | 47 |
| SW8260D | Vinyl chloride | 75-01-4 | µg/L | 1 | 930 |
| SW8260D | Xylenes, Total | 1330-20-7 | µg/L | | |
| ECO-PAHs | | | | | |
| SW8270E SIM | 1,4-Dioxane | 123-91-1 | µg/L | 0.4 | -- |

Table 3. Surface Water Target Analytes, Methods, Action Levels, and Control Limits

*Site Sampling Plan for Ecological Evaluation**Princeton Technology Center, West Windsor Township, New Jersey*

| Method | Analyte | CAS Number | Units | Higher of PQL and Ground Water Quality Criterion ^a | Fresh Surface Water Chronic NJDEP Ecological Criterion ^b |
|--------------------------|------------------------|------------|-------|---|---|
| SW8270E SIM | 2-Methylnaphthalene | 91-57-6 | µg/L | 30 | 330 |
| SW8270E SIM | Acenaphthene | 83-32-9 | µg/L | 400 | 38 |
| SW8270E SIM | Acenaphthylene | 208-96-8 | µg/L | -- | 4840 |
| SW8270E SIM | Anthracene | 120-12-7 | µg/L | 2000 | 0.035 |
| SW8270E SIM | Benzo(a)anthracene | 56-55-3 | µg/L | 0.1 | 0.025 |
| SW8270E SIM | Benzo(a)pyrene | 50-32-8 | µg/L | 0.1 | 0.014 |
| SW8270E SIM | Benzo(b)fluoranthene | 205-99-2 | µg/L | 0.2 | 9.07 |
| SW8270E SIM | Benzo(g,h,i)perylene | 191-24-2 | µg/L | -- | 7.64 |
| SW8270E SIM | Benzo(k)fluoranthene | 207-08-9 | µg/L | 0.5 | -- |
| SW8270E SIM | Chrysene | 218-01-9 | µg/L | 5 | -- |
| SW8270E SIM | Dibenz(a,h)anthracene | 53-70-3 | µg/L | 0.3 | -- |
| SW8270E SIM | Fluoranthene | 206-44-0 | µg/L | 300 | 1.9 |
| SW8270E SIM | Fluorene | 86-73-7 | µg/L | 300 | 19 |
| SW8270E SIM | Indeno[1,2,3-cd]pyrene | 193-39-5 | µg/L | 0.2 | 4.31 |
| SW8270E SIM | Naphthalene | 91-20-3 | µg/L | 300 | 13 |
| SW8270E SIM | Phenanthrene | 85-01-8 | µg/L | -- | 3.6 |
| SW8270E SIM | Pyrene | 129-00-0 | µg/L | 200 | 0.3 |
| <i>ECO-Metals</i> | | | | | |
| SW3060A/7196A | Hexavalent Chromium | 18540-29-9 | µg/L | -- | 10 |
| SW7470A | Mercury | 7439-97-6 | µg/L | 2 | 0.77 |
| SW6020B | Aluminum | 7429-90-5 | µg/L | -- | -- |
| SW6020B | Antimony | 7440-36-0 | µg/L | 6 | 80 |
| SW6020B | Arsenic | 7440-38-2 | µg/L | 3 | 150 |
| SW6020B | Barium | 7440-39-3 | µg/L | 6000 | 220 |
| SW6020B | Beryllium | 7440-41-7 | µg/L | 1 | 3.6 |
| SW6020B | Cadmium | 7440-43-9 | µg/L | 4 | -- |
| SW6020B | Calcium | 7440-70-2 | µg/L | -- | -- |
| SW6020B | Chromium | 7440-47-3 | µg/L | -- | 42 |
| SW6020B | Cobalt | 7440-48-4 | µg/L | 100 | 24 |
| SW6020B | Copper | 7440-50-8 | µg/L | 1300 | -- |
| SW6020B | Iron | 7439-89-6 | µg/L | -- | -- |
| SW6020B | Lead | 7439-92-1 | µg/L | 5 | 5.4 |
| SW6020B | Magnesium | 7439-95-4 | µg/L | -- | -- |
| SW6020B | Manganese | 7439-96-5 | µg/L | -- | -- |
| SW6020B | Nickel | 7440-02-0 | µg/L | 100 | -- |
| SW6020B | Potassium | 7440-09-7 | µg/L | -- | -- |
| SW6020B | Selenium | 7782-49-2 | µg/L | 40 | 5 |
| EPA 200.7 | Silica | 7631-86-9 | µg/L | -- | -- |

Table 3. Surface Water Target Analytes, Methods, Action Levels, and Control Limits*Site Sampling Plan for Ecological Evaluation**Princeton Technology Center, West Windsor Township, New Jersey*

| Method | Analyte | CAS Number | Units | Higher of PQL and Ground Water Quality Criterion ^a | Fresh Surface Water Chronic NJDEP Ecological Criterion ^b |
|---------|----------|------------|-------|---|---|
| SW6020B | Silver | 7440-22-4 | µg/L | 40 | 0.12 |
| SW6020B | Sodium | 7440-23-5 | µg/L | -- | -- |
| SW6020B | Thallium | 7440-28-0 | µg/L | -- | 10 |
| SW6020B | Vanadium | 7440-62-2 | µg/L | -- | 12 |
| SW6020B | Zinc | 7440-66-6 | µg/L | 2000 | -- |

Notes:

^a New Jersey Department of Environmental Protection (NJDEP) Ground Water Quality Standards - Class IIA by Constituent. May 2021. New Jersey Administration Code 7:9C-1.4: Remediation Standards.

^b NJDEP Ground Water Quality Standards - Class IIA by Constituent. May 2021. New Jersey Administration Code 7:9C-1.4: Remediation Standards. NJDEP Ecological Surface Water SSLs. March 2009.

Bold = MDL and RL exceed screening criteria.

-- = not available (no standard)

µg/L = microgram(s) per liter

CAS = Chemical Abstracts Service

Freon TF = 1,1,2-Trichloro-1,2,2-trifluoroethane

MDL = method detection limit

MTBE = methyl tert butyl ether

NJDEP = New Jersey Department of Environmental Protection

PAH = polycyclic aromatic hydrocarbon

PQL = Practical Quantitation Level as defined in N.J.A.C. 7:9C-1.4

RL = reporting limit

SIM = selected ion method

SVOC = semivolatile organic compound

VOC = volatile organic compound

Laboratory Certification

| Certified By | License No. |
|----------------------|------------------|
| CAS EPA CLP Contract | 68HERH20D0011 |
| Connecticut | PH-0830 |
| DOD ELAP (L-A-B) | L2219 |
| Maine | 2024021 |
| Maryland | 296 |
| New Hampshire | 255423 |
| New Jersey | 20012 |
| New York | 11376 |
| Pennsylvania | 68-00548 |
| Soil Permit | 525-24-234-08441 |
| Texas | T104704488 |

LOGIN REPORT/SAMPLE TRANSFER

| | | | | | |
|-------------------|-------------------------|--------|--------------------|----------------------------|-----------------------|
| Order ID : | P3657 | JACO05 | Order Date : | 8/16/2024 2:45:00 PM | Project Mgr : |
| Client Name : | JACOBS Engineering Grou | | Project Name : | Former Schlumberger Site I | |
| Client Contact : | Mary I. Murphy | | Receive DateTime : | 8/16/2024 12:00:00 AM | Report Type : Level 4 |
| Invoice Name : | JACOBS Engineering Grou | | Purchase Order : | 12:45 | EDD Type : CH2MHILL |
| Invoice Contact : | Mary I. Murphy | | | | Hard Copy Date : |
| | | | | | Date Signoff : |

| LAB ID | CLIENT ID | MATRIX | SAMPLE DATE | SAMPLE TIME | TEST | TEST GROUP | METHOD | FAX DATE | DUE DATES |
|----------|------------------------|--------|-------------|-------------|--------------|------------|----------|--------------|-----------|
| P3657-01 | 717-J-WS-081624 917 | Water | 08/16/2024 | 09:30 | VOCMS Group6 | | 8260-Low | 10 Bus. Days | |
| P3657-02 | TB-01-081624 | Water | 08/16/2024 | 10:55 | VOCMS Group6 | | 8260-Low | 10 Bus. Days | |

Relinquished By : JM
 Date / Time : 08-16-24 1524

Received By : Sam
 Date / Time : 8/16/24 15:21 Aug 4

Storage Area : VOA Refrigerator Room