

Surrogate Summary

SW-846

SDG No.: P3657

Client: JACOBS Engineering Group, Inc.

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
P3657-01	917-J-WS-081624	2-Fluorophenol	150	52.1	35		15 (10)	110 (139)
		Phenol-d6	150	32.9	22		15 (10)	110 (134)
		Nitrobenzene-d5	100	82.0	82		30 (49)	130 (133)
		2-Fluorobiphenyl	100	76.3	76		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	179	119	*	15 (44)	110 (137)
		Terphenyl-d14	100	96.0	96		30 (48)	130 (125)
PB162822BL	PB162822BL	2-Fluorophenol	150	130	87		15 (10)	110 (139)
		Phenol-d6	150	120	80		15 (10)	110 (134)
		Nitrobenzene-d5	100	90.5	91		30 (49)	130 (133)
		2-Fluorobiphenyl	100	78.7	79		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	158	105		15 (44)	110 (137)
		Terphenyl-d14	100	82.4	82		30 (48)	130 (125)
PB162822BS	PB162822BS	2-Fluorophenol	150	139	93		15 (10)	110 (139)
		Phenol-d6	150	131	87		15 (10)	110 (134)
		Nitrobenzene-d5	100	93.7	94		30 (49)	130 (133)
		2-Fluorobiphenyl	100	79.6	80		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	157	105		15 (44)	110 (137)
		Terphenyl-d14	100	85.2	85		30 (48)	130 (125)
PB162822BSD	PB162822BSD	2-Fluorophenol	150	137	91		15 (10)	110 (139)
		Phenol-d6	150	128	85		15 (10)	110 (134)
		Nitrobenzene-d5	100	93.0	93		30 (49)	130 (133)
		2-Fluorobiphenyl	100	80.2	80		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	154	103		15 (44)	110 (137)
		Terphenyl-d14	100	88.7	89		30 (48)	130 (125)