

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P3657 SAS No.: P3657 SDG NO.: P3657

EPA Sample No.: SSTDCCC040 Date Analyzed: 08/20/2024

Lab File ID: BP021575.D Time Analyzed: 16:40

Instrument ID: BNA_P GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	372530	7.805	1563750	10.59	1068850	14.46
UPPER LIMIT	745060	8.305	3127500	11.093	2137700	14.963
LOWER LIMIT	186265	7.305	781875	10.093	534425	13.963
EPA SAMPLE NO.						
01 917-J-WS-081624	352980	7.81	1458860	10.60	956855	14.46

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.