

Cover Page

Order ID : P3657

Project ID : Former Schlumberger Site Princeton NJ

Client : JACOBS Engineering Group, Inc.

Lab Sample Number

P3657-01
P3657-02

Client Sample Number

917-J-WS-081624
TB-01-081624

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : _____

Date: 8/29/2024

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012

CASE NARRATIVE

JACOBS Engineering Group, Inc.

Project Name: Former Schlumberger Site Princeton NJ

Project # N/A

Chemtech Project # P3657

Test Name: VOCMS Group6

A. Number of Samples and Date of Receipt:

2 Water samples were received on 08/16/2024.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: Hexavalent Chromium, Mercury, Metals Group4, SVOCMS Group3, SVOCMS Group6 and VOCMS Group6. This data package contains results for VOCMS Group6.

C. Analytical Techniques:

The analysis performed on instrument MSVOA_N were done using GC column RXI-624SIL MS 30m 0.25mm 1.4 um. Cat#13868. The analysis of VOCMS Group6 was based on method 8260D.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria.

The Internal Standards Areas met the acceptable requirements.

The Retention Times were acceptable for all samples.

The RPD met criteria .

The Blank Spike met requirements for all samples .

The Blank Spike Duplicate met requirements for all samples .

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements .

The Continuous Calibration met the requirements .

The Tuning criteria met requirements.

E. Additional Comments:

Samples for MS/MSD for VOC analysis were not provided with this set of samples. The Blank Spike Duplicate is reported with the data.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <15% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 15% for the Initial Calibration curve for SW-846 analysis.

F. Manual Integration Comments:



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following “ Results Qualifiers” are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. “10 U”. This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as “12 B”.
E	Indicates the analyte ‘s concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a “P”.
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

ALLIANCE 284 Sheffield Street, Mountainside New Jersey 07092

NEW JERSEY LAB ID#: 20012: NEW YORK LAB ID#: 11376

GC/MS VOA CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NUMBER: P3657

MATRIX: Water

METHOD: 8260D

		NA	NO	YES
1.	Chromatograms Labeled/Compounds Identified. (Field samples and Method Blanks)			✓
2.	GC/MS Tuning Specifications BFB Meet Criteria (NOTE THAT THERE ARE DIFFERENT CRITERIA FOR NY ASP CLP, CLP AND NJ)			✓
3.	GC/MS Tuning Frequency - Performed every 24 hours for 600 series and 12 hours for 8000 Series.			✓
4.	GC/MS Calibration - Initial Calibration performed before sample analysis and continuing calibration performed within 24 hours of sample analysis for 600 series and 12 hours for 8000 series.			✓
5.	GC/MS Calibration Requirements. The Initial Calibration met the requirements . The Continuous Calibration met the requirements .			✓
6.	Blank Contamination - If yes, list compounds and concentrations in each blank:		✓	
7.	Surrogate Recoveries Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable ranges.			✓
8.	Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range. The Blank Spike met requirements for all samples . The Blank Spike Duplicate met requirements for all samples .			✓
9.	Internal Standard Area/Retention Time Shift Meet Criteria Comments:			✓
10.	Analysis Holding Time Met If not met, list number of days exceeded for each sample:			✓

ALLIANCE 284 Sheffield Street, Mountainside New Jersey 07092

NEW JERSEY LAB ID#: 20012: NEW YORK LAB ID#: 11376

GC/MS VOA CONFORMANCE/NON-CONFORMANCE SUMMARY (CONTINUED)

NA NO YES

ADDITIONAL COMMENTS:

Samples for MS/MSD for VOC analysis were not provided with this set of samples. The Blank Spike Duplicate is reported with the data.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <15% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 15% for the Initial Calibration curve for SW-846 analysis.

QA REVIEW

Date

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: P3657

Completed

For thorough review, the report must have the following:

GENERAL:

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Custody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

ANALYTICAL:

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

1st Level QA Review Signature: SOHIL JODHANI

Date: 08/29/2024

2nd Level QA Review Signature: _____

Date: _____

LAB CHRONICLE

OrderID:	P3657	OrderDate:	8/16/2024 2:45:00 PM
Client:	JACOBS Engineering Group, Inc.	Project:	Former Schlumberger Site Princeton NJ
Contact:	Mary I. Murphy	Location:	G11,VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
P3657-01	917-J-WS-081624	Water	VOCMS Group6	8260-Low	08/16/24		08/19/24	08/16/24
P3657-02	TB-01-081624	Water	VOCMS Group6	8260-Low	08/16/24		08/19/24	08/16/24



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: P3657
Client: JACOBS Engineering Group, Inc.

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID:	917-J-WS-081624							
P3657-01	917-J-WS-081624	Water	Acetone	7.30		1.40	5.00	ug/L
			Total Voc :		7.30			
			Total Concentration:		7.30			



QC SUMMARY

Surrogate Summary

SDG No.: **P3657**

Client: **JACOBS Engineering Group, Inc.**

Analytical Method: **SW8260-Low**

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
P3657-01	917-J-WS-081624	1,2-Dichloroethane-d4	50	53.3	107	70 (74)	130 (125)
		Dibromofluoromethane	50	48.9	98	70 (75)	130 (124)
		Toluene-d8	50	48.9	98	70 (86)	130 (113)
		4-Bromofluorobenzene	50	51.3	103	70 (77)	130 (121)
P3657-02	TB-01-081624	1,2-Dichloroethane-d4	50	53.1	106	70 (74)	130 (125)
		Dibromofluoromethane	50	48.6	97	70 (75)	130 (124)
		Toluene-d8	50	46.7	93	70 (86)	130 (113)
		4-Bromofluorobenzene	50	49.7	99	70 (77)	130 (121)
VN0819WBL01	VN0819WBL01	1,2-Dichloroethane-d4	50	51.4	103	70 (74)	130 (125)
		Dibromofluoromethane	50	48.4	97	70 (75)	130 (124)
		Toluene-d8	50	48.4	97	70 (86)	130 (113)
		4-Bromofluorobenzene	50	51.6	103	70 (77)	130 (121)
VN0819WBS01	VN0819WBS01	1,2-Dichloroethane-d4	50	56.0	112	70 (74)	130 (125)
		Dibromofluoromethane	50	54.2	108	70 (75)	130 (124)
		Toluene-d8	50	53.4	107	70 (86)	130 (113)
		4-Bromofluorobenzene	50	52.4	105	70 (77)	130 (121)
VN0819WBSD0	VN0819WBSD01	1,2-Dichloroethane-d4	50	57.5	115	70 (74)	130 (125)
		Dibromofluoromethane	50	54.9	110	70 (75)	130 (124)
		Toluene-d8	50	52.6	105	70 (86)	130 (113)
		4-Bromofluorobenzene	50	52.2	104	70 (77)	130 (121)

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P3657

Client: JACOBS Engineering Group, Inc.

Analytical Method: SW8260-Low

Datafile : VN083370.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN0819WBS01	Dichlorodifluoromethane	20	17.6	ug/L	88			40 (69)	160 (116)	
	Chloromethane	20	18.0	ug/L	90			40 (65)	160 (116)	
	Vinyl chloride	20	18.2	ug/L	91			70 (65)	130 (117)	
	Bromomethane	20	17.4	ug/L	87			40 (58)	160 (125)	
	1,1,2-Trichlorotrifluoroethane	20	19.7	ug/L	99			70 (80)	130 (112)	
	Acetone	100	94.6	ug/L	95			40 (60)	160 (125)	
	Carbon disulfide	20	14.8	ug/L	74			40 (64)	160 (112)	
	Methyl tert-butyl Ether	20	19.3	ug/L	97			70 (78)	130 (114)	
	Methylene Chloride	20	17.9	ug/L	90			70 (72)	130 (114)	
	trans-1,2-Dichloroethene	20	17.8	ug/L	89			70 (75)	130 (108)	
	Cyclohexane	20	17.3	ug/L	86			70 (75)	130 (110)	
	2-Butanone	100	96.8	ug/L	97			40 (65)	160 (122)	
	Carbon Tetrachloride	20	19.3	ug/L	97			70 (77)	130 (113)	
	cis-1,2-Dichloroethene	20	19.0	ug/L	95			70 (77)	130 (110)	
	Chloroform	20	20.9	ug/L	104			70 (79)	130 (113)	
	1,1,1-Trichloroethane	20	20.4	ug/L	102			70 (80)	130 (108)	
	Methylcyclohexane	20	16.6	ug/L	83			70 (72)	130 (115)	
	Benzene	20	19.3	ug/L	97			70 (82)	130 (109)	
	1,2-Dichloroethane	20	20.2	ug/L	101			70 (80)	130 (115)	
	Trichloroethene	20	19.2	ug/L	96			70 (77)	130 (113)	
	Bromodichloromethane	20	19.9	ug/L	100			70 (83)	130 (110)	
	Toluene	20	19.1	ug/L	96			70 (82)	130 (110)	
	1,1,2-Trichloroethane	20	20.5	ug/L	103			70 (83)	130 (112)	
	Dibromochloromethane	20	20.9	ug/L	104			70 (82)	130 (110)	
	Tetrachloroethene	20	19.1	ug/L	96			70 (67)	130 (123)	
	Chlorobenzene	20	19.2	ug/L	96			70 (82)	130 (109)	
	Ethyl Benzene	20	18.7	ug/L	94			70 (83)	130 (109)	
	m/p-Xylenes	40	37.1	ug/L	93			70 (82)	130 (110)	
	o-Xylene	20	18.2	ug/L	91			70 (83)	130 (109)	
	Isopropylbenzene	20	18.8	ug/L	94			70 (83)	130 (112)	
	1,4-Dichlorobenzene	20	18.6	ug/L	93			70 (82)	130 (107)	
	1,2-Dichlorobenzene	20	19.2	ug/L	96			70 (82)	130 (109)	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P3657

Client: JACOBS Engineering Group, Inc.

Analytical Method: SW8260-Low

Datafile : VN083371.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN0819WBSD01	Dichlorodifluoromethane	20	17.7	ug/L	89	1		40 (69)	160 (116)	20 (20)
	Chloromethane	20	18.6	ug/L	93	3		40 (65)	160 (116)	20 (20)
	Vinyl chloride	20	19.2	ug/L	96	5		70 (65)	130 (117)	20 (20)
	Bromomethane	20	18.4	ug/L	92	6		40 (58)	160 (125)	20 (20)
	1,1,2-Trichlorotrifluoroethane	20	19.7	ug/L	99	0		70 (80)	130 (112)	20 (20)
	Acetone	100	100	ug/L	100	5		40 (60)	160 (125)	20 (20)
	Carbon disulfide	20	15.1	ug/L	76	3		40 (64)	160 (112)	20 (20)
	Methyl tert-butyl Ether	20	20.0	ug/L	100	3		70 (78)	130 (114)	20 (20)
	Methylene Chloride	20	19.1	ug/L	96	6		70 (72)	130 (114)	20 (20)
	trans-1,2-Dichloroethene	20	18.0	ug/L	90	1		70 (75)	130 (108)	20 (20)
	Cyclohexane	20	17.7	ug/L	89	3		70 (75)	130 (110)	20 (20)
	2-Butanone	100	100	ug/L	100	3		40 (65)	160 (122)	20 (20)
	Carbon Tetrachloride	20	19.4	ug/L	97	0		70 (77)	130 (113)	20 (20)
	cis-1,2-Dichloroethene	20	19.7	ug/L	99	4		70 (77)	130 (110)	20 (20)
	Chloroform	20	21.2	ug/L	106	2		70 (79)	130 (113)	20 (20)
	1,1,1-Trichloroethane	20	20.7	ug/L	104	2		70 (80)	130 (108)	20 (20)
	Methylcyclohexane	20	16.8	ug/L	84	1		70 (72)	130 (115)	20 (20)
	Benzene	20	19.4	ug/L	97	0		70 (82)	130 (109)	20 (20)
	1,2-Dichloroethane	20	21.3	ug/L	106	5		70 (80)	130 (115)	20 (20)
	Trichloroethene	20	19.3	ug/L	97	1		70 (77)	130 (113)	20 (20)
	Bromodichloromethane	20	20.3	ug/L	102	2		70 (83)	130 (110)	20 (20)
	Toluene	20	19.5	ug/L	98	2		70 (82)	130 (110)	20 (20)
	1,1,2-Trichloroethane	20	21.0	ug/L	105	2		70 (83)	130 (112)	20 (20)
	Dibromochloromethane	20	21.2	ug/L	106	2		70 (82)	130 (110)	20 (20)
	Tetrachloroethene	20	18.3	ug/L	92	4		70 (67)	130 (123)	20 (20)
	Chlorobenzene	20	19.3	ug/L	97	1		70 (82)	130 (109)	20 (20)
	Ethyl Benzene	20	18.5	ug/L	93	1		70 (83)	130 (109)	20 (20)
	m/p-Xylenes	40	37.0	ug/L	93	0		70 (82)	130 (110)	20 (20)
	o-Xylene	20	18.6	ug/L	93	2		70 (83)	130 (109)	20 (20)
	Isopropylbenzene	20	18.7	ug/L	94	0		70 (83)	130 (112)	20 (20)
	1,4-Dichlorobenzene	20	18.8	ug/L	94	1		70 (82)	130 (107)	20 (20)
	1,2-Dichlorobenzene	20	18.8	ug/L	94	2		70 (82)	130 (109)	20 (20)

VOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

VN0819WBL01

Lab Name: CHEMTECH

Contract: JACO05

Lab Code: CHEM Case No.: P3657

SAS No.: P3657 SDG NO.: P3657

Lab File ID: VN083369.D

Lab Sample ID: VN0819WBL01

Date Analyzed: 08/19/2024

Time Analyzed: 13:37

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: (Y/N) N

Instrument ID: MSVOA_N

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
VN0819WBS01	VN0819WBS01	VN083370.D	08/19/2024
VN0819WBSD01	VN0819WBSD01	VN083371.D	08/19/2024
TB-01-081624	P3657-02	VN083378.D	08/19/2024
917-J-WS-081624	P3657-01	VN083379.D	08/19/2024

COMMENTS: _____



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: JACO05
Lab Code: CHEM Case No.: P3657 SAS No.: P3657 SDG NO.: P3657
Lab File ID: VN083134.D BFB Injection Date: 08/07/2024
Instrument ID: MSVOA_N BFB Injection Time: 10:09
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	19.7
75	30.0 - 60.0% of mass 95	52.6
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	7.5
173	Less than 2.0% of mass 174	0.2 (0.3) 1
174	50.0 - 100.0% of mass 95	66.5
175	5.0 - 9.0% of mass 174	4.9 (7.4) 1
176	95.0 - 101.0% of mass 174	64.4 (96.9) 1
177	5.0 - 9.0% of mass 176	4.1 (6.3) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDICC100	VSTDICC100	VN083135.D	08/07/2024	10:33
VSTDICCC050	VSTDICCC050	VN083136.D	08/07/2024	10:58
VSTDICC020	VSTDICC020	VN083137.D	08/07/2024	11:22
VSTDICC010	VSTDICC010	VN083138.D	08/07/2024	11:46
VSTDICC005	VSTDICC005	VN083139.D	08/07/2024	12:10
VSTDICC001	VSTDICC001	VN083140.D	08/07/2024	12:34



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: JACO05
Lab Code: CHEM Case No.: P3657 SAS No.: P3657 SDG NO.: P3657
Lab File ID: VN083366.D BFB Injection Date: 08/19/2024
Instrument ID: MSVOA_N BFB Injection Time: 10:59
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	19.3
75	30.0 - 60.0% of mass 95	53.4
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.5
173	Less than 2.0% of mass 174	0.6 (0.9) 1
174	50.0 - 100.0% of mass 95	66.8
175	5.0 - 9.0% of mass 174	5.3 (8) 1
176	95.0 - 101.0% of mass 174	66.2 (99.1) 1
177	5.0 - 9.0% of mass 176	4.4 (6.7) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VN083367.D	08/19/2024	11:43
VN0819WBL01	VN0819WBL01	VN083369.D	08/19/2024	13:37
VN0819WBS01	VN0819WBS01	VN083370.D	08/19/2024	14:14
VN0819WBSD01	VN0819WBSD01	VN083371.D	08/19/2024	14:39
TB-01-081624	P3657-02	VN083378.D	08/19/2024	17:29
917-J-WS-081624	P3657-01	VN083379.D	08/19/2024	17:53

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: JACO05
Lab Code: CHEM Case No.: P3657 SAS No.: P3657 SDG NO.: P3657
Lab File ID: VN083367.D Date Analyzed: 08/19/2024
Instrument ID: MSVOA_N Time Analyzed: 11:43
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	172122	8.22	293866	9.10	257989	11.87
UPPER LIMIT	344244	8.724	587732	9.6	515978	12.365
LOWER LIMIT	86061	7.724	146933	8.6	128995	11.365
EPA SAMPLE NO.						
917-J-WS-081624	139398	8.22	273365	9.11	283963	11.87
TB-01-081624	133915	8.22	268412	9.11	271721	11.87
VN0819WBL01	144938	8.22	283183	9.10	290527	11.87
VN0819WBS01	162452	8.22	284677	9.10	251348	11.87
VN0819WBSD01	154957	8.22	272633	9.11	241038	11.87

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: JACO05
 Lab Code: CHEM Case No.: P3657 SAS No.: P3657 SDG NO.: P3657
 Lab File ID: VN083367.D Date Analyzed: 08/19/2024
 Instrument ID: MSVOA_N Time Analyzed: 11:43
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	128940	13.794				
UPPER LIMIT	257880	14.294				
LOWER LIMIT	64470	13.294				
EPA SAMPLE NO.						
917-J-WS-081624	126838	13.79				
TB-01-081624	120282	13.79				
VN0819WBL01	134032	13.79				
VN0819WBS01	116728	13.79				
VN0819WBSD01	112767	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.



SAMPLE DATA

Report of Analysis

Client:	JACOBS Engineering Group, Inc.	Date Collected:	08/16/24
Project:	Former Schlumberger Site Princeton NJ	Date Received:	08/16/24
Client Sample ID:	917-J-WS-081624	SDG No.:	P3657
Lab Sample ID:	P3657-01	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group6
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN083379.D	1		08/19/24 17:53	VN081924

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.21	U	0.21	1.00	ug/L
74-87-3	Chloromethane	0.35	U	0.35	1.00	ug/L
75-01-4	Vinyl Chloride	0.34	U	0.34	1.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	1.00	ug/L
67-64-1	Acetone	7.30		1.40	5.00	ug/L
75-15-0	Carbon Disulfide	0.32	U	0.32	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	1.00	ug/L
75-09-2	Methylene Chloride	0.32	U	0.32	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.25	U	0.25	1.00	ug/L
110-82-7	Cyclohexane	1.60	U	1.60	5.00	ug/L
78-93-3	2-Butanone	1.30	U	1.30	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.25	U	0.25	1.00	ug/L
67-66-3	Chloroform	0.26	U	0.26	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.19	U	0.19	1.00	ug/L
108-87-2	Methylcyclohexane	0.19	U	0.19	1.00	ug/L
71-43-2	Benzene	0.16	U	0.16	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.24	U	0.24	1.00	ug/L
79-01-6	Trichloroethene	0.32	U	0.32	1.00	ug/L
75-27-4	Bromodichloromethane	0.24	U	0.24	1.00	ug/L
108-88-3	Toluene	0.18	U	0.18	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	1.00	ug/L
124-48-1	Dibromochloromethane	0.18	U	0.18	1.00	ug/L
127-18-4	Tetrachloroethene	0.25	U	0.25	1.00	ug/L
108-90-7	Chlorobenzene	0.13	U	0.13	1.00	ug/L
100-41-4	Ethyl Benzene	0.16	U	0.16	1.00	ug/L
179601-23-1	m/p-Xylenes	0.31	U	0.31	2.00	ug/L
1330-20-7	Total Xylenes	0.45	U	0.45	3.00	ug/L
95-47-6	o-Xylene	0.14	U	0.14	1.00	ug/L

Report of Analysis

Client:	JACOBS Engineering Group, Inc.	Date Collected:	08/16/24
Project:	Former Schlumberger Site Princeton NJ	Date Received:	08/16/24
Client Sample ID:	917-J-WS-081624	SDG No.:	P3657
Lab Sample ID:	P3657-01	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5	Units:	mL
Soil Aliquot Vol:			uL
GC Column:	RXI-624	ID :	0.25
Prep Method :		Level :	LOW

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN083379.D	1		08/19/24 17:53	VN081924

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
98-82-8	Isopropylbenzene	0.13	U	0.13	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.27	U	0.27	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.19	U	0.19	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	53.3		70 (74) - 130 (125)	107%	SPK: 50
1868-53-7	Dibromofluoromethane	48.9		70 (75) - 130 (124)	98%	SPK: 50
2037-26-5	Toluene-d8	48.9		70 (86) - 130 (113)	98%	SPK: 50
460-00-4	4-Bromofluorobenzene	51.3		70 (77) - 130 (121)	103%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	139000	8.224			
540-36-3	1,4-Difluorobenzene	273000	9.106			
3114-55-4	Chlorobenzene-d5	284000	11.865			
3855-82-1	1,4-Dichlorobenzene-d4	127000	13.794			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081924\
 Data File : VN083379.D
 Acq On : 19 Aug 2024 17:53
 Operator : JC\MD
 Sample : P3657-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 917-J-WS-081624

Quant Time: Aug 20 04:49:26 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	8.224	168	139398	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	273365	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	283963	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	126838	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	105845	53.345	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	106.680%
35) Dibromofluoromethane	8.171	113	83464	48.916	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.840%
50) Toluene-d8	10.565	98	311490	48.939	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.880%
62) 4-Bromofluorobenzene	12.853	95	127359	51.326	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	102.660%
Target Compounds						
16) Acetone	4.447	43	5705	7.348	ug/l	95


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081924\
Data File : VN083379.D
Acq On : 19 Aug 2024 17:53
Operator : JC\MD
Sample : P3657-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
917-J-WS-081624

Quant Time: Aug 20 04:49:26 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 1066
Delta R.T. -0.000 min
Lab File: VN083379.D
Acq: 19 Aug 2024 17:53




Instrument :
MSVOA_N
ClientSampleId :
917-J-WS-081624


Tgt Ion:168 Resp: 139398
Ion Ratio Lower Upper
168 100
99 63.5 48.2 72.4



#16
Acetone
Concen: 7.348 ug/l
RT: 4.447 min Scan# 424
Delta R.T. 0.018 min
Lab File: VN083379.D
Acq: 19 Aug 2024 17:53

Tgt Ion: 43 Resp: 5705
Ion Ratio Lower Upper
43 100
58 24.6 21.8 32.6





#33

1,2-Dichloroethane-d4

Concen: 53.345 ug/l

RT: 8.582 min Scan# 1127

Delta R.T. 0.000 min

Lab File: VN083379.D

Acq: 19 Aug 2024 17:53

Instrument :

MSVOA_N

ClientSampleId :





917-J-WS-081624

Tgt Ion: 65 Resp: 105845

Ion Ratio Lower Upper

65 100

67 49.2 0.0 104.4



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.106 min Scan# 1216

Delta R.T. 0.006 min

Lab File: VN083379.D

Acq: 19 Aug 2024 17:53




Tgt Ion: 114 Resp: 273365


Ion Ratio Lower Upper

114 100

63 22.7 0.0 44.6

88 17.5 0.0 31.4





#35

Dibromofluoromethane

Concen: 48.916 ug/l

RT: 8.171 min Scan# 1056

Delta R.T. 0.006 min

Lab File: VN083379.D

Acq: 19 Aug 2024 17:53

Instrument :

MSVOA_N

ClientSampleId :

917-J-WS-081624





Tgt Ion:113 Resp: 83464

Ion Ratio Lower Upper

113 100

111 105.3 82.4 123.6

192 17.2 14.9 22.3



#50

Toluene-d8

Concen: 48.939 ug/l

RT: 10.565 min Scan# 1464

Delta R.T. -0.000 min

Lab File: VN083379.D




Acq: 19 Aug 2024 17:53


Tgt Ion: 98 Resp: 311490

Ion Ratio Lower Upper

98 100

100 64.6 51.5 77.3









#62
4-Bromofluorobenzene
Concen: 51.326 ug/l
RT: 12.853 min Scan# 1853
Delta R.T. 0.006 min
Lab File: VN083379.D
Acq: 19 Aug 2024 17:53




Instrument :
MSVOA_N
ClientSampleId :
917-J-WS-081624


Tgt Ion: 95 Resp: 127359
Ion Ratio Lower Upper
95 100
174 69.6 0.0 159.2
176 67.3 0.0 147.6



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 1685
Delta R.T. -0.000 min
Lab File: VN083379.D
Acq: 19 Aug 2024 17:53

Tgt Ion: 117 Resp: 283963
Ion Ratio Lower Upper
117 100
82 59.5 47.5 71.3
119 31.9 25.6 38.4





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.794 min Scan# 20

Delta R.T. -0.000 min

Lab File: VN083379.D


Acq: 19 Aug 2024 17:53

Instrument :

MSVOA_N

ClientSampleId :

917-J-WS-081624





Tgt Ion:152 Resp: 126838

Ion Ratio Lower Upper

152 100

115 63.6 30.6 91.6

150 157.0 0.0 348.6



Report of Analysis

Client:	JACOBS Engineering Group, Inc.	Date Collected:	08/16/24
Project:	Former Schlumberger Site Princeton NJ	Date Received:	08/16/24
Client Sample ID:	TB-01-081624	SDG No.:	P3657
Lab Sample ID:	P3657-02	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group6
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN083378.D	1		08/19/24 17:29	VN081924

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.21	U	0.21	1.00	ug/L
74-87-3	Chloromethane	0.35	U	0.35	1.00	ug/L
75-01-4	Vinyl Chloride	0.34	U	0.34	1.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	1.00	ug/L
67-64-1	Acetone	1.40	U	1.40	5.00	ug/L
75-15-0	Carbon Disulfide	0.32	U	0.32	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	1.00	ug/L
75-09-2	Methylene Chloride	0.32	U	0.32	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.25	U	0.25	1.00	ug/L
110-82-7	Cyclohexane	1.60	U	1.60	5.00	ug/L
78-93-3	2-Butanone	1.30	U	1.30	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.25	U	0.25	1.00	ug/L
67-66-3	Chloroform	0.26	U	0.26	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.19	U	0.19	1.00	ug/L
108-87-2	Methylcyclohexane	0.19	U	0.19	1.00	ug/L
71-43-2	Benzene	0.16	U	0.16	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.24	U	0.24	1.00	ug/L
79-01-6	Trichloroethene	0.32	U	0.32	1.00	ug/L
75-27-4	Bromodichloromethane	0.24	U	0.24	1.00	ug/L
108-88-3	Toluene	0.18	U	0.18	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	1.00	ug/L
124-48-1	Dibromochloromethane	0.18	U	0.18	1.00	ug/L
127-18-4	Tetrachloroethene	0.25	U	0.25	1.00	ug/L
108-90-7	Chlorobenzene	0.13	U	0.13	1.00	ug/L
100-41-4	Ethyl Benzene	0.16	U	0.16	1.00	ug/L
179601-23-1	m/p-Xylenes	0.31	U	0.31	2.00	ug/L
1330-20-7	Total Xylenes	0.45	U	0.45	3.00	ug/L
95-47-6	o-Xylene	0.14	U	0.14	1.00	ug/L

Report of Analysis

Client:	JACOBS Engineering Group, Inc.	Date Collected:	08/16/24
Project:	Former Schlumberger Site Princeton NJ	Date Received:	08/16/24
Client Sample ID:	TB-01-081624	SDG No.:	P3657
Lab Sample ID:	P3657-02	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5	Units:	mL
Soil Aliquot Vol:			uL
GC Column:	RXI-624	ID :	0.25
Prep Method :		Level :	LOW

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN083378.D	1		08/19/24 17:29	VN081924

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
98-82-8	Isopropylbenzene	0.13	U	0.13	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.27	U	0.27	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.19	U	0.19	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	53.1		70 (74) - 130 (125)	106%	SPK: 50
1868-53-7	Dibromofluoromethane	48.6		70 (75) - 130 (124)	97%	SPK: 50
2037-26-5	Toluene-d8	46.7		70 (86) - 130 (113)	93%	SPK: 50
460-00-4	4-Bromofluorobenzene	49.7		70 (77) - 130 (121)	99%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	134000	8.224			
540-36-3	1,4-Difluorobenzene	268000	9.106			
3114-55-4	Chlorobenzene-d5	272000	11.865			
3855-82-1	1,4-Dichlorobenzene-d4	120000	13.794			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of OC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081924\
 Data File : VN083378.D
 Acq On : 19 Aug 2024 17:29
 Operator : JC\MD
 Sample : P3657-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB-01-081624

Quant Time: Aug 20 04:49:05 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	133915	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	268412	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	271721	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	120282	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	101252	53.119	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	106.240%
35) Dibromofluoromethane	8.171	113	81404	48.589	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.180%
50) Toluene-d8	10.565	98	292111	46.741	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	93.480%
62) 4-Bromofluorobenzene	12.847	95	121199	49.745	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	99.480%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081924\
Data File : VN083378.D
Acq On : 19 Aug 2024 17:29
Operator : JC\MD
Sample : P3657-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB-01-081624

Quant Time: Aug 20 04:49:05 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 8.224 min Scan# 1066

Delta R.T. -0.000 min

Lab File: VN083378.D

Acq: 19 Aug 2024 17:29

Instrument :

MSVOA_N

ClientSampleId :





TB-01-081624

Tgt Ion:168 Resp: 133915

Ion Ratio Lower Upper

168 100

99 64.2 48.2 72.4



#33

1,2-Dichloroethane-d4

Concen: 53.119 ug/l

RT: 8.577 min Scan# 1126

Delta R.T. -0.005 min

Lab File: VN083378.D




Acq: 19 Aug 2024 17:29


Tgt Ion: 65 Resp: 101252

Ion Ratio Lower Upper

65 100

67 49.3 0.0 104.4





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.106 min Scan# 11

Delta R.T. 0.006 min

Lab File: VN083378.D

Acq: 19 Aug 2024 17:29

Instrument :

MSVOA_N

ClientSampleId :

TB-01-081624





Tgt Ion:114 Resp: 268412

Ion Ratio Lower Upper

114 100

63 21.1 0.0 44.6

88 16.5 0.0 31.4



#35

Dibromofluoromethane

Concen: 48.589 ug/l

RT: 8.171 min Scan# 1057

Delta R.T. 0.006 min

Lab File: VN083378.D

Acq: 19 Aug 2024 17:29




Tgt Ion:113 Resp: 81404


Ion Ratio Lower Upper

113 100

111 101.8 82.4 123.6

192 16.9 14.9 22.3









#50
Toluene-d8
Concen: 46.741 ug/l
RT: 10.565 min Scan# 1464
Delta R.T. -0.000 min
Lab File: VN083378.D
Acq: 19 Aug 2024 17:29




Instrument :
MSVOA_N
ClientSampleId :
TB-01-081624


Tgt Ion: 98 Resp: 292111
Ion Ratio Lower Upper
98 100
100 66.3 51.5 77.3



#62
4-Bromofluorobenzene
Concen: 49.745 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. 0.000 min
Lab File: VN083378.D
Acq: 19 Aug 2024 17:29

Tgt Ion: 95 Resp: 121199
Ion Ratio Lower Upper
95 100
174 69.6 0.0 159.2
176 66.9 0.0 147.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 11
Delta R.T. 0.000 min
Lab File: VN083378.D
Acq: 19 Aug 2024 17:29




Instrument :
MSVOA_N
ClientSampleId :
TB-01-081624

Tgt Ion:117 Resp: 271721
Ion Ratio Lower Upper
117 100
82 60.5 47.5 71.3
119 32.0 25.6 38.4



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.794 min Scan# 2013
Delta R.T. 0.000 min
Lab File: VN083378.D
Acq: 19 Aug 2024 17:29

Tgt Ion:152 Resp: 120282
Ion Ratio Lower Upper
152 100
115 64.0 30.6 91.6
150 157.3 0.0 348.6





CALIBRATION SUMMARY



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: JAC005
 Lab Code: CHEM Case No.: P3657 SAS No.: P3657 SDG No.: P3657
 Instrument ID: MSVOA_N Calibration Date(s): 08/07/2024 08/07/2024
 Heated Purge: (Y/N) N Calibration Time(s): 10:33 12:34
 GC Column: RXI-624 ID: 0.25 (mm)

LAB FILE ID:		RRF100 = VN083135.D		RRF050 = VN083136.D		RRF020 = VN083137.D			
		RRF010 = VN083138.D		RRF005 = VN083139.D		RRF001 = VN083140.D			
COMPOUND	RRF100	RRF050	RRF020	RRF010	RRF005	RRF001	RRF	% RSD	
Dichlorodifluoromethane	0.548	0.571	0.566	0.579	0.597	0.542	0.567	3.6	
Chloromethane	0.547	0.591	0.584	0.589	0.595	0.576	0.581	3.1	
Vinyl Chloride	0.562	0.598	0.588	0.612	0.586	0.607	0.592	3	
Bromomethane	0.313	0.357	0.370	0.385	0.413		0.368	10.1	
1,1,2-Trichlorotrifluoroethane	0.524	0.548	0.548	0.550	0.558	0.510	0.539	3.4	
Acetone	0.302	0.286	0.260	0.284	0.274	0.265	0.278	5.5	
Carbon Disulfide	1.466	1.560	1.596	1.599	1.630	1.879	1.622	8.5	
Methyl tert-butyl Ether	1.927	2.028	2.028	2.015	2.053	1.953	2.001	2.5	
Methylene Chloride	0.569	0.599	0.619	0.614	0.640	0.805	0.641	13	
trans-1,2-Dichloroethene	0.533	0.574	0.577	0.586	0.568	0.599	0.573	3.9	
Cyclohexane	0.933	0.982	1.010	1.110	1.238		1.055	11.5	
2-Butanone	0.418	0.427	0.424	0.430	0.414	0.453	0.428	3.2	
Carbon Tetrachloride	0.524	0.546	0.548	0.541	0.539	0.493	0.532	3.9	
cis-1,2-Dichloroethene	0.658	0.689	0.691	0.701	0.687	0.721	0.691	3	
Chloroform	1.087	1.137	1.110	1.132	1.122	1.102	1.115	1.7	
1,1,1-Trichloroethane	1.027	1.075	1.066	1.076	1.081	1.007	1.055	2.9	
Methylcyclohexane	0.565	0.597	0.587	0.583	0.575	0.572	0.580	2	
Benzene	1.386	1.447	1.431	1.429	1.403	1.343	1.406	2.7	
1,2-Dichloroethane	0.494	0.520	0.514	0.524	0.524	0.498	0.512	2.6	
Trichloroethene	0.331	0.344	0.343	0.341	0.340	0.310	0.335	3.9	
Bromodichloromethane	0.521	0.539	0.540	0.533	0.541	0.544	0.537	1.6	
Toluene	0.892	0.926	0.905	0.900	0.885	0.824	0.889	3.9	
1,1,2-Trichloroethane	0.322	0.334	0.331	0.327	0.314	0.283	0.318	6	
Dibromochloromethane	0.400	0.413	0.397	0.388	0.383	0.329	0.385	7.7	
Tetrachloroethene	0.319	0.343	0.338	0.343	0.324	0.320	0.331	3.5	
Chlorobenzene	1.070	1.142	1.101	1.153	1.094	1.069	1.105	3.2	
Ethyl Benzene	1.967	2.075	2.061	2.070	2.032	1.957	2.027	2.6	
m/p-Xylenes	0.747	0.787	0.771	0.785	0.759	0.706	0.759	4	
o-Xylene	0.735	0.773	0.759	0.773	0.750	0.703	0.749	3.6	
Isopropylbenzene	3.822	4.249	4.133	4.386	4.249	4.254	4.182	4.6	

* Compounds with required minimum RRF and maximum %RSD values.
 All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: JAC005
 Lab Code: CHEM Case No.: P3657 SAS No.: P3657 SDG No.: P3657
 Instrument ID: MSVOA_N Calibration Date(s): 08/07/2024 08/07/2024
 Heated Purge: (Y/N) N Calibration Time(s): 10:33 12:34
 GC Column: RXI-624 ID: 0.25 (mm)

LAB FILE ID:		RRF100 = VN083135.D		RRF050 = VN083136.D		RRF020 = VN083137.D			
		RRF010 = VN083138.D		RRF005 = VN083139.D		RRF001 = VN083140.D			
COMPOUND	RRF100	RRF050	RRF020	RRF010	RRF005	RRF001	RRF	% RSD	
1,4-Dichlorobenzene	1.655	1.783	1.728	1.751	1.795	1.862	1.762	4	
1,2-Dichlorobenzene	1.571	1.704	1.664	1.739	1.731	1.741	1.692	3.9	
1,2-Dichloroethane-d4	0.717	0.745	0.601	0.710	0.786		0.712	9.7	
Dibromofluoromethane	0.311	0.332	0.264	0.307	0.346		0.312	10	
Toluene-d8	1.201	1.261	0.985	1.133	1.240		1.164	9.6	
4-Bromofluorobenzene	0.473	0.492	0.390	0.432	0.483		0.454	9.4	

* Compounds with required minimum RRF and maximum %RSD values.
 All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.

Method Path : Z:\voasrv\HPCHEM1\MSVOA_N\methods\
 Method File : 82N080724W.M
 Title : SW846 8260
 Last Update : Thu Aug 08 06:30:41 2024
 Response Via : Initial Calibration

Calibration Files

1 =VN083140.D 5 =VN083139.D 10 =VN083138.D 20 =VN083137.D 50 =VN083136.D 100 =VN083135.D

	Compound	1	5	10	20	50	100	Avg	%RSD
1) I	Pentafluorobenzene	-----ISTD-----							
2) T	Dichlorodifluo...	0.542	0.597	0.579	0.566	0.571	0.548	0.567	3.55
3) P	Chloromethane	0.576	0.595	0.589	0.584	0.591	0.547	0.581	3.06
4) C	Vinyl Chloride	0.607	0.586	0.612	0.588	0.598	0.562	0.592	3.02#
5) T	Bromomethane		0.413	0.385	0.370	0.357	0.313	0.368	10.07
6) T	Chloroethane	0.436	0.381	0.355	0.356	0.356	0.340	0.371	9.33
7) T	Trichlorofluor...	0.901	1.003	1.013	1.005	0.995	0.955	0.979	4.42
8) T	Diethyl Ether	0.338	0.380	0.385	0.367	0.366	0.348	0.364	4.97
9) T	1,1,2-Trichlor...	0.510	0.558	0.550	0.548	0.548	0.524	0.539	3.41
10) T	Methyl Iodide		0.688	0.710	0.703	0.740	0.707	0.710	2.72
11) T	Tert butyl alc...		0.155	0.155	0.146	0.147	0.136	0.148	5.22
12) CM	1,1-Dichloroet...	0.575	0.581	0.564	0.538	0.544	0.524	0.554	4.08#
13) T	Acrolein		0.095	0.101	0.083	0.102	0.100	0.096	7.99
14) T	Allyl chloride	1.210	1.099	1.083	1.063	0.933	0.896	1.047	11.00
15) T	Acrylonitrile	0.306	0.311	0.306	0.304	0.306	0.293	0.304	1.95
16) T	Acetone	0.265	0.274	0.284	0.260	0.286	0.302	0.278	5.51
17) T	Carbon Disulfide	1.879	1.630	1.599	1.596	1.560	1.466	1.622	8.52
18) T	Methyl Acetate	0.985	0.844	0.842	0.773	0.776	0.758	0.830	10.20
19) T	Methyl tert-bu...	1.953	2.053	2.015	2.028	2.028	1.927	2.001	2.46
20) T	Methylene Chlo...	0.805	0.640	0.614	0.619	0.599	0.569	0.641	13.03
21) T	trans-1,2-Dich...	0.599	0.568	0.586	0.577	0.574	0.533	0.573	3.93
22) T	Diisopropyl ether	1.945	1.966	2.009	1.976	2.008	1.909	1.969	1.94
23) T	Vinyl Acetate	1.930	2.006	2.040	2.046	2.096	1.990	2.018	2.80
24) P	1,1-Dichloroet...	1.035	1.093	1.096	1.093	1.084	1.038	1.073	2.65
25) T	2-Butanone	0.453	0.414	0.430	0.424	0.427	0.418	0.428	3.22
26) T	2,2-Dichloropr...	0.925	0.984	1.032	1.019	1.023	0.997	0.997	3.95
27) T	cis-1,2-Dichlo...	0.721	0.687	0.701	0.691	0.689	0.658	0.691	2.99
28) T	Bromochloromet...	0.465	0.489	0.391	0.420	0.439	0.427	0.439	7.91
29) T	Tetrahydrofuran	0.271	0.280	0.286	0.272	0.279	0.270	0.276	2.24
30) C	Chloroform	1.102	1.122	1.132	1.110	1.137	1.087	1.115	1.71#
31) T	Cyclohexane		1.238	1.110	1.010	0.982	0.933	1.055	11.50
32) T	1,1,1-Trichlor...	1.007	1.081	1.076	1.066	1.075	1.027	1.055	2.89
33) S	1,2-Dichloroet...		0.786	0.710	0.601	0.745	0.717	0.712	9.68
34) I	1,4-Difluorobenzene	-----ISTD-----							
35) S	Dibromofluorom...		0.346	0.307	0.264	0.332	0.311	0.312	9.99
36) T	1,1-Dichloropr...	0.488	0.463	0.477	0.472	0.475	0.457	0.472	2.33
37) T	Ethyl Acetate	0.665	0.515	0.512	0.502	0.499	0.479	0.529	12.83
38) T	Carbon Tetrach...	0.493	0.539	0.541	0.548	0.546	0.524	0.532	3.93
39) T	Methylcyclohexane	0.572	0.575	0.583	0.587	0.597	0.565	0.580	1.96
40) TM	Benzene	1.343	1.403	1.429	1.431	1.447	1.386	1.406	2.68
41) T	Methacrylonitrile	0.377	0.290	0.277	0.300	0.283	0.276	0.301	12.78
42) TM	1,2-Dichloroet...	0.498	0.524	0.524	0.514	0.520	0.494	0.512	2.63
43) T	Isopropyl Acetate	2.156	1.289	1.075	0.956	0.921	0.869	1.211	40.18
44) TM	Trichloroethene	0.310	0.340	0.341	0.343	0.344	0.331	0.335	3.94
45) C	1,2-Dichloropr...	0.316	0.326	0.341	0.344	0.346	0.331	0.334	3.50#
46) T	Dibromomethane	0.218	0.243	0.240	0.245	0.251	0.238	0.239	4.73
47) T	Bromodichlorom...	0.544	0.541	0.533	0.540	0.539	0.521	0.537	1.55
48) T	Methyl methacr...	0.506	0.423	0.432	0.421	0.425	0.415	0.437	7.88
49) T	1,4-Dioxane	0.008	0.008	0.008	0.008	0.008	0.008	0.008	2.64
50) S	Toluene-d8		1.240	1.133	0.985	1.261	1.201	1.164	9.56
51) T	4-Methyl-2-Pen...	0.457	0.499	0.516	0.516	0.517	0.495	0.500	4.62
52) CM	Toluene	0.824	0.885	0.900	0.905	0.926	0.892	0.889	3.91#
53) T	t-1,3-Dichloro...	0.469	0.546	0.552	0.573	0.597	0.569	0.551	7.99
54) T	cis-1,3-Dichlo...	0.546	0.588	0.595	0.597	0.606	0.586	0.586	3.54
55) T	1,1,2-Trichlor...	0.283	0.314	0.327	0.331	0.334	0.322	0.318	5.95
56) T	Ethyl methacry...	0.553	0.594	0.612	0.617	0.624	0.601	0.600	4.24

Method Path : Z:\voasrv\HPCHEM1\MSVOA_N\methods\
 Method File : 82N080724W.M

57)	T	1,3-Dichloropr...	0.552	0.545	0.575	0.581	0.592	0.561	0.568	3.17
58)	T	2-Chloroethyl ...	0.243	0.250	0.262	0.242	0.261	0.264	0.254	3.91
59)	T	2-Hexanone	0.367	0.385	0.389	0.395	0.397	0.388	0.387	2.70
60)	T	Dibromochlorom...	0.329	0.383	0.388	0.397	0.413	0.400	0.385	7.68
61)	T	1,2-Dibromoethane	0.317	0.332	0.342	0.336	0.347	0.334	0.335	3.13
62)	S	4-Bromofluorob...	0.483	0.432	0.390	0.492	0.473	0.454		9.36
63)	I	Chlorobenzene-d5	-----ISTD-----							
64)	T	Tetrachloroethene	0.320	0.324	0.343	0.338	0.343	0.319	0.331	3.46
65)	PM	Chlorobenzene	1.069	1.094	1.153	1.101	1.142	1.070	1.105	3.23
66)	T	1,1,1,2-Tetrac...	0.370	0.401	0.403	0.385	0.399	0.380	0.390	3.37
67)	C	Ethyl Benzene	1.957	2.032	2.070	2.061	2.075	1.967	2.027	2.59#
68)	T	m/p-Xylenes	0.706	0.759	0.785	0.771	0.787	0.747	0.759	3.99
69)	T	o-Xylene	0.703	0.750	0.773	0.759	0.773	0.735	0.749	3.60
70)	T	Styrene	1.100	1.261	1.300	1.288	1.336	1.261	1.258	6.51
71)	P	Bromoform	0.267	0.275	0.292	0.309	0.318	0.310	0.295	7.01
72)	I	1,4-Dichlorobenzen...	-----ISTD-----							
73)	T	Isopropylbenzene	4.254	4.249	4.386	4.133	4.249	3.822	4.182	4.63
74)	T	N-amyl acetate	2.341	2.051	2.108	1.945	2.003	1.826	2.046	8.50
75)	P	1,1,2,2-Tetrac...	1.234	1.237	1.210	1.163	1.167	1.085	1.183	4.85
76)	T	1,2,3-Trichlor...	1.233	1.083	1.128	1.094	1.063	0.984	1.097	7.47
77)	T	Bromobenzene	0.900	0.975	0.946	0.927	0.959	0.868	0.929	4.27
78)	T	n-propylbenzene	4.800	4.849	4.998	4.753	4.968	4.526	4.816	3.55
79)	T	2-Chlorotoluene	3.146	3.083	3.184	3.027	3.076	2.802	3.053	4.42
80)	T	1,3,5-Trimethy...	3.479	3.563	3.641	3.490	3.589	3.245	3.501	3.98
81)	T	trans-1,4-Dich...	0.530	0.468	0.524	0.493	0.508	0.504		4.97
82)	T	4-Chlorotoluene	3.261	3.081	3.116	3.014	3.104	2.799	3.062	4.99
83)	T	tert-Butylbenzene	3.103	3.199	3.192	3.098	3.162	2.850	3.101	4.19
84)	T	1,2,4-Trimethy...	3.416	3.592	3.679	3.526	3.654	3.305	3.529	4.10
85)	T	sec-Butylbenzene	4.235	4.265	4.325	4.202	4.384	3.976	4.231	3.33
86)	T	p-Isopropyltol...	3.355	3.569	3.510	3.491	3.682	3.354	3.493	3.62
87)	T	1,3-Dichlorobe...	1.870	1.708	1.757	1.739	1.771	1.643	1.748	4.29
88)	T	1,4-Dichlorobe...	1.862	1.795	1.751	1.728	1.783	1.655	1.762	3.96
89)	T	n-Butylbenzene	2.851	3.017	3.045	3.023	3.216	3.011	3.027	3.84
90)	T	Hexachloroethane	0.620	0.696	0.708	0.686	0.689	0.650	0.675	4.92
91)	T	1,2-Dichlorobe...	1.741	1.731	1.739	1.664	1.704	1.571	1.692	3.90
92)	T	1,2-Dibromo-3-...	0.310	0.307	0.289	0.277	0.281	0.258	0.287	6.84
93)	T	1,2,4-Trichlor...	1.014	0.893	0.938	0.943	0.984	0.914	0.948	4.72
94)	T	Hexachlorobuta...	0.480	0.445	0.429	0.397	0.409	0.368	0.421	9.32
95)	T	Naphthalene	3.429	3.230	3.435	3.365	3.481	3.202	3.357	3.44
96)	T	1,2,3-Trichlor...	0.972	0.920	0.974	0.920	0.963	0.877	0.938	4.09

(#) = Out of Range

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080724\
 Data File : VN083135.D
 Acq On : 07 Aug 2024 10:33
 Operator : JC\MD
 Sample : VSTDICC100
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC100

Manual Integrations
 APPROVED

Reviewed By :John Carlone 08/08/2024
 Supervised By :Mahesh Dadoda 08/09/2024

Quant Time: Aug 08 06:14:38 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 05:50:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	217317	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	373033	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	329850	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	162834	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	311531	100.713	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 201.420%#			
35) Dibromofluoromethane	8.165	113	231805	99.556	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 199.120%#			
50) Toluene-d8	10.565	98	896022	103.163	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 206.320%#			
62) 4-Bromofluorobenzene	12.847	95	352947	104.234	ug/l	0.00
Spiked Amount 50.000	Range 64 - 133		Recovery = 208.460%#			
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	2.124	85	238119	96.624	ug/l	99
3) Chloromethane	2.359	50	237633	94.185	ug/l	98
4) Vinyl Chloride	2.512	62	244329	94.904	ug/l	100
5) Bromomethane	2.924	94	135920	85.085	ug/l	96
6) Chloroethane	3.100	64	147611	91.645	ug/l	99
7) Trichlorofluoromethane	3.489	101	414913	97.558	ug/l	96
8) Diethyl Ether	3.959	74	151291	95.597	ug/l	83
9) 1,1,2-Trichlorotrifluo...	4.365	101	227550	97.044	ug/l	98
10) Methyl Iodide	4.583	142	307476	99.686	ug/l	97
11) Tert butyl alcohol	5.524	59	296420	461.112	ug/l	98
12) 1,1-Dichloroethene	4.336	96	227537	94.453	ug/l	90
13) Acrolein	4.177	56	218130	520.640	ug/l	98
14) Allyl chloride	5.018	41	389580	85.577	ug/l	93
15) Acrylonitrile	5.718	53	636894	481.734	ug/l	99
16) Acetone	4.424	43	656348	542.264	ug/l	95
17) Carbon Disulfide	4.712	76	637255	90.406	ug/l	99
18) Methyl Acetate	5.024	43	329457	91.357	ug/l	92
19) Methyl tert-butyl Ether	5.794	73	837562	96.324	ug/l	99
20) Methylene Chloride	5.277	84	247283	88.771	ug/l	88
21) trans-1,2-Dichloroethene	5.783	96	231530	92.991	ug/l	95
22) Diisopropyl ether	6.671	45	829706	96.959	ug/l	96
23) Vinyl Acetate	6.600	43	4324915	493.177	ug/l	95
24) 1,1-Dichloroethane	6.571	63	451253	96.746	ug/l	99
25) 2-Butanone	7.482	43	908992	489.098	ug/l	91
26) 2,2-Dichloropropane	7.488	77	433489	100.075	ug/l	98
27) cis-1,2-Dichloroethene	7.488	96	286100	95.225	ug/l	93
28) Bromochloromethane	7.812	49	185533	97.333	ug/l	85
29) Tetrahydrofuran	7.841	42	587324	488.865	ug/l	88
30) Chloroform	7.965	83	472257	97.456	ug/l	99
31) Cyclohexane	8.259	56	405666	88.496	ug/l	98
32) 1,1,1-Trichloroethane	8.171	97	446446	97.332	ug/l	95
36) 1,1-Dichloropropene	8.371	75	340980	96.808	ug/l	98
37) Ethyl Acetate	7.559	43	357308	90.602	ug/l #	93
38) Carbon Tetrachloride	8.359	117	391148	98.597	ug/l	96
39) Methylcyclohexane	9.600	83	421743	97.486	ug/l	97
40) Benzene	8.606	78	1034059	98.550	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080724\
 Data File : VN083135.D
 Acq On : 07 Aug 2024 10:33
 Operator : JC\MD
 Sample : VSTDICC100
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC100

Manual Integrations
 APPROVED

Reviewed By :John Carlone 08/08/2024
 Supervised By :Mahesh Dadoda 08/09/2024

Quant Time: Aug 08 06:14:38 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 05:50:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.777	41	205844	91.799	ug/l	95
42) 1,2-Dichloroethane	8.671	62	368412	96.385	ug/l	97
43) Isopropyl Acetate	8.688	43	648582	99.962	ug/l #	93
44) Trichloroethene	9.353	130	246689	98.774	ug/l	93
45) 1,2-Dichloropropane	9.624	63	247262	99.275	ug/l	98
46) Dibromomethane	9.712	93	177294	99.449	ug/l	98
47) Bromodichloromethane	9.888	83	389020	97.184	ug/l	100
48) Methyl methacrylate	9.682	41	309367	94.886	ug/l	91
49) 1,4-Dioxane	9.694	88	117342	1994.879	ug/l	94
51) 4-Methyl-2-Pentanone	10.447	43	1846346	495.091	ug/l	92
52) Toluene	10.629	92	665274	100.347	ug/l	97
53) t-1,3-Dichloropropene	10.835	75	424825	103.317	ug/l	98
54) cis-1,3-Dichloropropene	10.312	75	436928	99.903	ug/l	90
55) 1,1,2-Trichloroethane	11.018	97	240237	101.114	ug/l	96
56) Ethyl methacrylate	10.876	69	448632	100.164	ug/l #	84
57) 1,3-Dichloropropane	11.165	76	418505	98.824	ug/l	98
58) 2-Chloroethyl Vinyl ether	10.159	63	984478	519.987	ug/l	96
59) 2-Hexanone	11.194	43	1447307	501.642	ug/l	92
60) Dibromochloromethane	11.359	129	298498	103.914	ug/l	98
61) 1,2-Dibromoethane	11.470	107	249072	99.794	ug/l	96
64) Tetrachloroethene	11.106	164	210334	96.284	ug/l	96
65) Chlorobenzene	11.894	112	705635	96.805	ug/l	100
66) 1,1,1,2-Tetrachloroethane	11.959	131	250772	97.546	ug/l	99
67) Ethyl Benzene	11.965	91	1297632	97.041	ug/l	99
68) m/p-Xylenes	12.070	106	984984	196.641	ug/l	99
69) o-Xylene	12.400	106	484673	98.107	ug/l	97
70) Styrene	12.412	104	831903	100.261	ug/l	99
71) Bromoform	12.576	173	204380	104.942	ug/l #	99
73) Isopropylbenzene	12.694	105	1244845	91.399	ug/l	100
74) N-amyl acetate	12.494	43	594797	89.282	ug/l	94
75) 1,1,2,2-Tetrachloroethane	12.941	83	353346	91.727	ug/l	99
76) 1,2,3-Trichloropropane	12.994	75	320302m	89.478	ug/l	
77) Bromobenzene	12.982	156	282676	93.431	ug/l	96
78) n-propylbenzene	13.035	91	1473868	93.979	ug/l	99
79) 2-Chlorotoluene	13.123	91	912618	91.785	ug/l	98
80) 1,3,5-Trimethylbenzene	13.170	105	1056912	92.690	ug/l	98
81) trans-1,4-Dichloro-2-b...	12.735	75	165326	100.628	ug/l	89
82) 4-Chlorotoluene	13.223	91	911384	91.381	ug/l	98
83) tert-Butylbenzene	13.441	119	928195	91.918	ug/l	99
84) 1,2,4-Trimethylbenzene	13.482	105	1076449	93.675	ug/l	99
85) sec-Butylbenzene	13.617	105	1294900	93.978	ug/l	100
86) p-Isopropyltoluene	13.729	119	1092439	96.022	ug/l	99
87) 1,3-Dichlorobenzene	13.735	146	535193	94.009	ug/l	99
88) 1,4-Dichlorobenzene	13.811	146	538924	93.900	ug/l	99
89) n-Butylbenzene	14.059	91	980482	99.461	ug/l	98
90) Hexachloroethane	14.335	117	211710	96.320	ug/l	95
91) 1,2-Dichlorobenzene	14.106	146	511584	92.863	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	14.723	75	83930	89.788	ug/l	98
93) 1,2,4-Trichlorobenzene	15.394	180	297699	96.456	ug/l	99
94) Hexachlorobutadiene	15.500	225	119783	87.268	ug/l	98
95) Naphthalene	15.641	128	1042864	95.391	ug/l	99
96) 1,2,3-Trichlorobenzene	15.841	180	285769	93.575	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080724\
Data File : VN083135.D
Acq On : 07 Aug 2024 10:33
Operator : JC\MD
Sample : VSTDICC100
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Manual Integrations
APPROVED

Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024

Quant Time: Aug 08 06:14:38 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 05:50:26 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080724\
 Data File : VN083135.D
 Acq On : 07 Aug 2024 10:33
 Operator : JC\MD
 Sample : VSTDICC100
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :

MSVOA_N

Client Sample Id :

VSTDICC100

Manual Integrations

APPROVED

Reviewed By : John Carlone 08/08/2024

Supervised By : Mahesh Dadoda 08/09/2024


Quant Time: Aug 08 06:14:38 2024


Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M

Quant Title : SW846 8260

QLast Update : Thu Aug 08 05:50:26 2024


Response via : Initial Calibration





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 1066
Delta R.T. -0.000 min
Lab File: VN083135.D
Acq: 07 Aug 2024 10:33




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100



Tgt Ion:168 Resp: 21731
Ion Ratio Lower Upper
168 100
99 68.1 48.2 72.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#2
Dichlorodifluoromethane
Concen: 96.624 ug/l
RT: 2.124 min Scan# 29
Delta R.T. -0.000 min
Lab File: VN083135.D
Acq: 07 Aug 2024 10:33

Tgt Ion: 85 Resp: 238119
Ion Ratio Lower Upper
85 100
87 32.1 15.7 47.0









#3
Chloromethane
Concen: 94.185 ug/l
RT: 2.359 min Scan# 69
Delta R.T. -0.000 min
Lab File: VN083135.D
Acq: 07 Aug 2024 10:33

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Tgt Ion: 50 Resp: 23763
Ion Ratio Lower Upper
50 100
52 32.4 25.0 37.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#4
Vinyl Chloride
Concen: 94.904 ug/l
RT: 2.512 min Scan# 95
Delta R.T. -0.000 min
Lab File: VN083135.D
Acq: 07 Aug 2024 10:33


Tgt Ion: 62 Resp: 244329
Ion Ratio Lower Upper
62 100
64 31.4 25.0 37.6





#5
Bromomethane
Concen: 85.085 ug/l
RT: 2.924 min Scan# 1
Delta R.T. -0.024 min
Lab File: VN083135.D
Acq: 07 Aug 2024 10:33




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100



Tgt Ion: 94 Resp: 135920
Ion Ratio Lower Upper
94 100
96 94.0 78.0 117.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#6
Chloroethane
Concen: 91.645 ug/l
RT: 3.100 min Scan# 195
Delta R.T. -0.012 min
Lab File: VN083135.D
Acq: 07 Aug 2024 10:33


Tgt Ion: 64 Resp: 147611
Ion Ratio Lower Upper
64 100
66 32.9 26.6 40.0





#7
Trichlorofluoromethane
Concen: 97.558 ug/l
RT: 3.489 min Scan# 262
Delta R.T. -0.006 min
Lab File: VN083135.D
Acq: 07 Aug 2024 10:33




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100



Tgt Ion:101 Resp: 41491
Ion Ratio Lower Upper
101 100
103 63.0 52.7 79.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#8
Diethyl Ether
Concen: 95.597 ug/l
RT: 3.959 min Scan# 341
Delta R.T. -0.000 min
Lab File: VN083135.D
Acq: 07 Aug 2024 10:33

Tgt Ion: 74 Resp: 151291
Ion Ratio Lower Upper
74 100
45 93.4 55.5 166.3









#9
1,1,2-Trichlorotrifluoroethane
Concen: 97.044 ug/l
RT: 4.365 min Scan# 410
Delta R.T. -0.000 min
Lab File: VN083135.D
Acq: 07 Aug 2024 10:33

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Tgt Ion:101 Resp: 227550
Ion Ratio Lower Upper
101 100
85 45.6 38.3 57.5
151 75.2 59.3 88.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#10
Methyl Iodide
Concen: 99.686 ug/l
RT: 4.583 min Scan# 447
Delta R.T. -0.006 min
Lab File: VN083135.D
Acq: 07 Aug 2024 10:33

Tgt Ion:142 Resp: 307476
Ion Ratio Lower Upper
142 100
127 48.4 37.5 56.3
141 15.0 13.1 19.7





#11

Tert butyl alcohol

Concen: 461.112 ug/l

RT: 5.524 min Scan# 608

Delta R.T. -0.006 min

Lab File: VN083135.D

Acq: 07 Aug 2024 10:33

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC100

Tgt Ion: 59 Resp: 296420

Ion Ratio Lower Upper

59 100





57 10.5 8.9 13.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#12

1,1-Dichloroethene

Concen: 94.453 ug/l

RT: 4.336 min Scan# 405

Delta R.T. -0.006 min

Lab File: VN083135.D

Acq: 07 Aug 2024 10:33




Tgt Ion: 96 Resp: 227537


Ion Ratio Lower Upper

96 100

61 169.4 149.7 224.5

98 66.5 50.1 75.1





#13

Acrolein

Concen: 520.640 ug/l

RT: 4.177 min Scan# 31

Delta R.T. -0.006 min

Lab File: VN083135.D

Acq: 07 Aug 2024 10:33

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100

Tgt Ion: 56 Resp: 218130

Ion Ratio Lower Upper

56 100





55 68.5 56.4 84.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#14

Allyl chloride

Concen: 85.577 ug/l

RT: 5.018 min Scan# 521

Delta R.T. -0.000 min

Lab File: VN083135.D

Acq: 07 Aug 2024 10:33




Tgt Ion: 41 Resp: 389580


Ion Ratio Lower Upper

41 100

39 78.4 67.8 101.6

76 36.6 25.8 38.8





#15

Acrylonitrile

Concen: 481.734 ug/l

RT: 5.718 min Scan# 64

Delta R.T. -0.000 min

Lab File: VN083135.D

Acq: 07 Aug 2024 10:33

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100

Tgt Ion: 53 Resp: 636894

Ion Ratio Lower Upper

53 100

52 82.5 66.6 100.0





51 36.4 30.1 45.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#16

Acetone

Concen: 542.264 ug/l

RT: 4.424 min Scan# 420

Delta R.T. -0.006 min

Lab File: VN083135.D




Acq: 07 Aug 2024 10:33


Tgt Ion: 43 Resp: 656348

Ion Ratio Lower Upper

43 100

58 29.7 21.8 32.6









#17
Carbon Disulfide
Concen: 90.406 ug/l
RT: 4.712 min Scan# 469
Delta R.T. -0.000 min
Lab File: VN083135.D
Acq: 07 Aug 2024 10:33

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Tgt Ion: 76 Resp: 63725
Ion Ratio Lower Upper
76 100
78 9.2 7.2 10.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#18
Methyl Acetate
Concen: 91.357 ug/l
RT: 5.024 min Scan# 522
Delta R.T. -0.000 min
Lab File: VN083135.D
Acq: 07 Aug 2024 10:33

Tgt Ion: 43 Resp: 329457
Ion Ratio Lower Upper
43 100
74 24.0 16.3 24.5





#19

Methyl tert-butyl Ether

Concen: 96.324 ug/l

RT: 5.794 min Scan# 653

Delta R.T. -0.000 min

Lab File: VN083135.D

Acq: 07 Aug 2024 10:33

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100

Tgt Ion: 73 Resp: 837562

Ion Ratio Lower Upper

73 100





57 21.9 17.9 26.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#20

Methylene Chloride

Concen: 88.771 ug/l

RT: 5.277 min Scan# 565

Delta R.T. -0.000 min

Lab File: VN083135.D

Acq: 07 Aug 2024 10:33

Tgt Ion: 84 Resp: 247283




Ion Ratio Lower Upper


84 100

49 127.4 119.6 179.4

51 40.0 34.8 52.2


86 63.6 52.9 79.3





#21
trans-1,2-Dichloroethene
Concen: 92.991 ug/l
RT: 5.783 min Scan# 61
Delta R.T. -0.012 min
Lab File: VN083135.D
Acq: 07 Aug 2024 10:33




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100



Tgt Ion: 96 Resp: 231530
Ion Ratio Lower Upper
96 100
61 145.4 123.4 185.2
98 63.2 49.3 73.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#22
Diisopropyl ether
Concen: 96.959 ug/l
RT: 6.671 min Scan# 802
Delta R.T. -0.000 min
Lab File: VN083135.D
Acq: 07 Aug 2024 10:33


Tgt Ion: 45 Resp: 829706
Ion Ratio Lower Upper
45 100
43 55.8 44.0 66.0
87 29.5 19.7 29.5
59 12.1 8.5 12.7





#23
Vinyl Acetate
Concen: 493.177 ug/l
RT: 6.600 min Scan# 791
Delta R.T. -0.006 min
Lab File: VN083135.D
Acq: 07 Aug 2024 10:33




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100



Tgt Ion: 43 Resp: 4324915
Ion Ratio Lower Upper
43 100
86 10.9 7.3 10.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#24
1,1-Dichloroethane
Concen: 96.746 ug/l
RT: 6.571 min Scan# 785
Delta R.T. -0.000 min
Lab File: VN083135.D
Acq: 07 Aug 2024 10:33


Tgt Ion: 63 Resp: 451253
Ion Ratio Lower Upper
63 100
98 6.8 3.3 9.9
100 4.5 2.0 6.0





#25
2-Butanone
Concen: 489.098 ug/l
RT: 7.482 min Scan# 940
Delta R.T. -0.000 min
Lab File: VN083135.D
Acq: 07 Aug 2024 10:33




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100



Tgt Ion: 43 Resp: 908992
Ion Ratio Lower Upper
43 100
72 27.2 18.2 27.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#26
2,2-Dichloropropane
Concen: 100.075 ug/l
RT: 7.488 min Scan# 941
Delta R.T. -0.000 min
Lab File: VN083135.D
Acq: 07 Aug 2024 10:33

Tgt Ion: 77 Resp: 433489
Ion Ratio Lower Upper
77 100
97 21.7 10.3 30.9





#27

cis-1,2-Dichloroethene

Concen: 95.225 ug/l

RT: 7.488 min Scan# 941

Delta R.T. -0.000 min

Lab File: VN083135.D


Acq: 07 Aug 2024 10:33

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100






Tgt Ion:	96	Resp:	286100
Ion	Ratio	Lower	Upper
96	100		
61	151.6	0.0	329.6
98	64.3	0.0	128.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#28

Bromochloromethane


Concen: 97.333 ug/l

RT: 7.812 min Scan# 996



Delta R.T. -0.000 min


Lab File: VN083135.D

Acq: 07 Aug 2024 10:33



Tgt Ion:	49	Resp:	185533
Ion	Ratio	Lower	Upper
49	100		
129	1.9	0.0	3.8
130	74.9	50.5	75.7





#29

Tetrahydrofuran

Concen: 488.865 ug/l

RT: 7.841 min Scan# 1000

Delta R.T. 0.006 min

Lab File: VN083135.D

Acq: 07 Aug 2024 10:33

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC100

Tgt Ion: 42 Resp: 587324

Ion Ratio Lower Upper

42 100

72 44.6 30.0 45.0





71 41.9 28.3 42.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#30

Chloroform

Concen: 97.456 ug/l

RT: 7.965 min Scan# 1022

Delta R.T. -0.000 min

Lab File: VN083135.D




Acq: 07 Aug 2024 10:33


Tgt Ion: 83 Resp: 472257

Ion Ratio Lower Upper

83 100


85 64.2 50.9 76.3





#31
Cyclohexane
Concen: 88.496 ug/l
RT: 8.259 min Scan# 1072
Delta R.T. -0.000 min
Lab File: VN083135.D
Acq: 07 Aug 2024 10:33




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100



Tgt Ion: 56 Resp: 405660
Ion Ratio Lower Upper
56 100
69 29.7 24.5 36.7
84 84.5 66.0 99.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#32
1,1,1-Trichloroethane
Concen: 97.332 ug/l
RT: 8.171 min Scan# 1057
Delta R.T. -0.000 min
Lab File: VN083135.D
Acq: 07 Aug 2024 10:33

Tgt Ion: 97 Resp: 446446
Ion Ratio Lower Upper
97 100
99 64.7 52.0 78.0
61 45.7 42.1 63.1





#33

1,2-Dichloroethane-d4

Concen: 100.713 ug/l

RT: 8.577 min Scan# 1126

Delta R.T. -0.006 min

Lab File: VN083135.D

Acq: 07 Aug 2024 10:33

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100

Tgt Ion: 65 Resp: 31153

Ion Ratio Lower Upper

65 100





67 50.6 0.0 104.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.100 min Scan# 1215

Delta R.T. -0.000 min

Lab File: VN083135.D

Acq: 07 Aug 2024 10:33




Tgt Ion:114 Resp: 373033


Ion Ratio Lower Upper

114 100

63 22.5 0.0 44.6

88 17.1 0.0 31.4





#35

Dibromofluoromethane

Concen: 99.556 ug/l

RT: 8.165 min Scan# 1056

Delta R.T. -0.000 min

Lab File: VN083135.D

Acq: 07 Aug 2024 10:33

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC100

Tgt Ion:113 Resp: 23180

Ion Ratio Lower Upper

113 100

111 103.7 82.4 123.6





192 17.7 14.9 22.3

Manual Integrations

APPROVED

Reviewed By :John Carbone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#36

1,1-Dichloropropene

Concen: 96.808 ug/l

RT: 8.371 min Scan# 1091

Delta R.T. -0.000 min

Lab File: VN083135.D

Acq: 07 Aug 2024 10:33




Tgt Ion: 75 Resp: 340980


Ion Ratio Lower Upper

75 100

110 34.1 16.3 48.9

77 30.6 24.6 37.0





#37

Ethyl Acetate

Concen: 90.602 ug/l

RT: 7.559 min Scan# 91

Delta R.T. -0.000 min

Lab File: VN083135.D

Acq: 07 Aug 2024 10:33

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC100

Tgt Ion: 43 Resp: 357308

Ion Ratio Lower Upper

43 100

61 14.4 9.3 13.9





70 11.4 7.0 10.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#38

Carbon Tetrachloride

Concen: 98.597 ug/l

RT: 8.359 min Scan# 1089

Delta R.T. -0.000 min

Lab File: VN083135.D

Acq: 07 Aug 2024 10:33




Tgt Ion:117 Resp: 391148


Ion Ratio Lower Upper

117 100

119 97.5 74.9 112.3

121 31.9 24.3 36.5





#39

Methylcyclohexane

Concen: 97.486 ug/l

RT: 9.600 min Scan# 1130

Delta R.T. -0.000 min

Lab File: VN083135.D

Acq: 07 Aug 2024 10:33

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100

Tgt Ion: 83 Resp: 42174

Ion Ratio Lower Upper

83 100

55 80.6 66.4 99.6





98 49.2 40.7 61.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#40

Benzene

Concen: 98.550 ug/l

RT: 8.606 min Scan# 1131

Delta R.T. -0.000 min

Lab File: VN083135.D




Acq: 07 Aug 2024 10:33


Tgt Ion: 78 Resp: 1034059

Ion Ratio Lower Upper

78 100


77 23.1 19.0 28.4





#41
Methacrylonitrile
Concen: 91.799 ug/l
RT: 7.777 min Scan# 990
Delta R.T. -0.000 min
Lab File: VN083135.D
Acq: 07 Aug 2024 10:33




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100



Tgt Ion: 41 Resp: 205844
Ion Ratio Lower Upper
41 100
39 55.7 45.5 68.3
67 70.9 50.5 75.7
52 30.1 24.0 36.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#42
1,2-Dichloroethane
Concen: 96.385 ug/l
RT: 8.671 min Scan# 1142
Delta R.T. -0.000 min
Lab File: VN083135.D
Acq: 07 Aug 2024 10:33

Tgt Ion: 62 Resp: 368412
Ion Ratio Lower Upper
62 100
98 8.8 0.0 15.8





#43

Isopropyl Acetate

Concen: 99.962 ug/l

RT: 8.688 min Scan# 1145

Delta R.T. 0.000 min

Lab File: VN083135.D

Acq: 07 Aug 2024 10:33

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100

Tgt Ion: 43 Resp: 648582

Ion Ratio Lower Upper

43 100

61 25.2 17.8 26.6





87 13.0 8.2 12.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#44

Trichloroethene

Concen: 98.774 ug/l

RT: 9.353 min Scan# 1258

Delta R.T. -0.000 min

Lab File: VN083135.D




Acq: 07 Aug 2024 10:33


Tgt Ion:130 Resp: 246689

Ion Ratio Lower Upper

130 100

95 105.7 0.0 197.8





#45

1,2-Dichloropropane

Concen: 99.275 ug/l

RT: 9.624 min Scan# 1304

Delta R.T. -0.000 min

Lab File: VN083135.D

Acq: 07 Aug 2024 10:33

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100

Tgt Ion: 63 Resp: 247262

Ion Ratio Lower Upper

63 100





65 31.6 24.4 36.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#46

Dibromomethane

Concen: 99.449 ug/l

RT: 9.712 min Scan# 1319

Delta R.T. -0.000 min

Lab File: VN083135.D

Acq: 07 Aug 2024 10:33




Tgt Ion: 93 Resp: 177294


Ion Ratio Lower Upper

93 100

95 83.4 65.8 98.6

174 92.1 71.7 107.5





#47

Bromodichloromethane

Concen: 97.184 ug/l

RT: 9.888 min Scan# 1349

Delta R.T. -0.000 min

Lab File: VN083135.D

Acq: 07 Aug 2024 10:33

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC100

Tgt Ion: 83 Resp: 389020

Ion Ratio Lower Upper

83 100

85 63.9 51.1 76.7





127 8.3 7.2 10.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#48

Methyl methacrylate

Concen: 94.886 ug/l

RT: 9.682 min Scan# 1314

Delta R.T. -0.000 min

Lab File: VN083135.D

Acq: 07 Aug 2024 10:33




Tgt Ion: 41 Resp: 309367


Ion Ratio Lower Upper

41 100

69 81.0 56.3 84.5

39 58.7 50.3 75.5





#49

1,4-Dioxane

Concen: 1994.879 ug/l

RT: 9.694 min Scan# 11734

Delta R.T. -0.000 min

Lab File: VN083135.D

Acq: 07 Aug 2024 10:33

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100

Tgt Ion: 88 Resp: 11734

Ion Ratio Lower Upper

88 100

43 28.2 27.8 41.8





58 71.8 59.4 89.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#50

Toluene-d8

Concen: 103.163 ug/l

RT: 10.565 min Scan# 1464

Delta R.T. -0.000 min

Lab File: VN083135.D




Acq: 07 Aug 2024 10:33


Tgt Ion: 98 Resp: 896022

Ion Ratio Lower Upper

98 100


100 65.8 51.5 77.3





#51
4-Methyl-2-Pentanone
Concen: 495.091 ug/l
RT: 10.447 min Scan# 1444
Delta R.T. -0.000 min
Lab File: VN083135.D
Acq: 07 Aug 2024 10:33




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100



Tgt Ion: 43 Resp: 1846340
Ion Ratio Lower Upper
43 100
58 39.7 28.2 42.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#52
Toluene
Concen: 100.347 ug/l
RT: 10.629 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VN083135.D
Acq: 07 Aug 2024 10:33


Tgt Ion: 92 Resp: 665274
Ion Ratio Lower Upper
92 100
91 169.4 139.4 209.0





#53
t-1,3-Dichloropropene
Concen: 103.317 ug/l
RT: 10.835 min Scan# 1510
Delta R.T. -0.000 min
Lab File: VN083135.D
Acq: 07 Aug 2024 10:33




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100



Tgt Ion: 75 Resp: 424829
Ion Ratio Lower Upper
75 100
77 30.9 25.7 38.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#54
cis-1,3-Dichloropropene
Concen: 99.903 ug/l
RT: 10.312 min Scan# 1421
Delta R.T. -0.000 min
Lab File: VN083135.D
Acq: 07 Aug 2024 10:33


Tgt Ion: 75 Resp: 436928
Ion Ratio Lower Upper
75 100
77 32.7 24.3 36.5
39 53.2 50.5 75.7





#55
1,1,2-Trichloroethane
Concen: 101.114 ug/l
RT: 11.018 min Scan# 1541
Delta R.T. -0.000 min
Lab File: VN083135.D
Acq: 07 Aug 2024 10:33




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100




Tgt Ion: 97 Resp: 24023
Ion Ratio Lower Upper
97 100
83 85.3 71.4 107.0
85 54.8 45.8 68.6
99 61.8 52.0 78.0

Manual Integrations
APPROVED



Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024




#56
Ethyl methacrylate
Concen: 100.164 ug/l
RT: 10.876 min Scan# 1517
Delta R.T. -0.000 min
Lab File: VN083135.D
Acq: 07 Aug 2024 10:33



Tgt Ion: 69 Resp: 448632
Ion Ratio Lower Upper
69 100
41 66.4 63.4 95.2
39 34.5 37.4 56.0#





#57

1,3-Dichloropropane

Concen: 98.824 ug/l

RT: 11.165 min Scan# 1566

Delta R.T. -0.000 min

Lab File: VN083135.D

Acq: 07 Aug 2024 10:33

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC100

Tgt Ion: 76 Resp: 41850

Ion Ratio Lower Upper

76 100





78 32.9 25.6 38.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#58

2-Chloroethyl Vinyl ether

Concen: 519.987 ug/l

RT: 10.159 min Scan# 1395

Delta R.T. -0.000 min

Lab File: VN083135.D




Acq: 07 Aug 2024 10:33


Tgt Ion: 63 Resp: 984478

Ion Ratio Lower Upper

63 100


106 28.8 21.4 32.0





#59
2-Hexanone
Concen: 501.642 ug/l
RT: 11.194 min Scan# 1571
Delta R.T. -0.000 min
Lab File: VN083135.D
Acq: 07 Aug 2024 10:33




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100



Tgt Ion: 43 Resp: 1447307
Ion Ratio Lower Upper
43 100
58 54.7 24.6 73.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#60
Dibromochloromethane
Concen: 103.914 ug/l
RT: 11.359 min Scan# 1599
Delta R.T. -0.000 min
Lab File: VN083135.D
Acq: 07 Aug 2024 10:33


Tgt Ion:129 Resp: 298498
Ion Ratio Lower Upper
129 100
127 76.3 39.2 117.6





#61
1,2-Dibromoethane
Concen: 99.794 ug/l
RT: 11.470 min Scan# 1618
Delta R.T. -0.000 min
Lab File: VN083135.D
Acq: 07 Aug 2024 10:33




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100



Tgt Ion:107 Resp: 249072
Ion Ratio Lower Upper
107 100
109 95.5 73.7 110.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#62
4-Bromofluorobenzene
Concen: 104.234 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. -0.000 min
Lab File: VN083135.D
Acq: 07 Aug 2024 10:33


Tgt Ion: 95 Resp: 352947
Ion Ratio Lower Upper
95 100
174 69.9 0.0 159.2
176 67.6 0.0 147.6





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 1685
Delta R.T. -0.000 min
Lab File: VN083135.D
Acq: 07 Aug 2024 10:33




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100



Tgt Ion:117 Resp: 329850
Ion Ratio Lower Upper
117 100
82 59.0 47.5 71.3
119 31.4 25.6 38.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#64
Tetrachloroethene
Concen: 96.284 ug/l
RT: 11.106 min Scan# 1556
Delta R.T. -0.000 min
Lab File: VN083135.D
Acq: 07 Aug 2024 10:33


Tgt Ion:164 Resp: 210334
Ion Ratio Lower Upper
164 100
166 128.2 101.4 152.0
129 102.1 75.7 113.5
131 96.1 73.8 110.6





#65
Chlorobenzene
Concen: 96.805 ug/l
RT: 11.894 min Scan# 1690
Delta R.T. -0.000 min
Lab File: VN083135.D
Acq: 07 Aug 2024 10:33




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100




Tgt Ion:112 Resp: 705639
Ion Ratio Lower Upper
112 100
114 32.5 26.2 39.2

Manual Integrations
APPROVED



Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024




#66
1,1,1,2-Tetrachloroethane
Concen: 97.546 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. -0.006 min
Lab File: VN083135.D
Acq: 07 Aug 2024 10:33



Tgt Ion:131 Resp: 250772
Ion Ratio Lower Upper
131 100
133 93.9 47.3 142.0
119 63.6 32.5 97.4





#67

Ethyl Benzene

Concen: 97.041 ug/l

RT: 11.965 min Scan# 1702

Delta R.T. -0.000 min

Lab File: VN083135.D

Acq: 07 Aug 2024 10:33

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100

Tgt Ion: 91 Resp: 129763

Ion Ratio Lower Upper

91 100





106 30.5 24.0 36.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#68

m/p-Xylenes

Concen: 196.641 ug/l

RT: 12.070 min Scan# 1720

Delta R.T. -0.000 min

Lab File: VN083135.D




Acq: 07 Aug 2024 10:33


Tgt Ion:106 Resp: 984984

Ion Ratio Lower Upper

106 100

91 208.8 166.1 249.1





#69

o-Xylene

Concen: 98.107 ug/l

RT: 12.400 min Scan# 1776

Delta R.T. -0.000 min

Lab File: VN083135.D

Acq: 07 Aug 2024 10:33

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC100

Tgt Ion:106 Resp: 484673

Ion Ratio Lower Upper

106 100





91 218.1 111.3 333.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#70

Styrene

Concen: 100.261 ug/l

RT: 12.412 min Scan# 1778

Delta R.T. -0.000 min

Lab File: VN083135.D

Acq: 07 Aug 2024 10:33




Tgt Ion:104 Resp: 831903


Ion Ratio Lower Upper

104 100

78 53.5 41.6 62.4


103 54.4 44.0 66.0





#71
Bromoform
Concen: 104.942 ug/l
RT: 12.576 min Scan# 1806
Delta R.T. -0.000 min
Lab File: VN083135.D
Acq: 07 Aug 2024 10:33




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100




Tgt Ion:173 Resp: 204380
Ion Ratio Lower Upper
173 100
175 47.6 24.0 72.0
254 0.0 0.0 0.0

Manual Integrations
APPROVED



Reviewed By :John Carbone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.794 min Scan# 2013
Delta R.T. -0.000 min
Lab File: VN083135.D
Acq: 07 Aug 2024 10:33




Tgt Ion:152 Resp: 162834
Ion Ratio Lower Upper
152 100
115 63.9 30.6 91.6
150 189.9 0.0 348.6





#73
Isopropylbenzene
Concen: 91.399 ug/l
RT: 12.694 min Scan# 1826
Delta R.T. -0.000 min
Lab File: VN083135.D
Acq: 07 Aug 2024 10:33




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100



Tgt Ion:105 Resp: 124484
Ion Ratio Lower Upper
105 100
120 25.9 12.9 38.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#74
N-ethyl acetate
Concen: 89.282 ug/l
RT: 12.494 min Scan# 1792
Delta R.T. -0.000 min
Lab File: VN083135.D
Acq: 07 Aug 2024 10:33


Tgt Ion: 43 Resp: 594797
Ion Ratio Lower Upper
43 100
70 42.7 29.8 44.6
55 23.4 18.7 28.1
61 24.7 17.4 26.0





#75
1,1,2,2-Tetrachloroethane
Concen: 91.727 ug/l
RT: 12.941 min Scan# 1868
Delta R.T. 0.006 min
Lab File: VN083135.D
Acq: 07 Aug 2024 10:33




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100



Tgt Ion: 83 Resp: 353340
Ion Ratio Lower Upper
83 100
131 10.3 5.0 14.9
85 64.2 31.9 95.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#76
1,2,3-Trichloropropane
Concen: 89.478 ug/l m
RT: 12.994 min Scan# 1877
Delta R.T. -0.000 min
Lab File: VN083135.D
Acq: 07 Aug 2024 10:33


Tgt Ion: 75 Resp: 320302
Ion Ratio Lower Upper
75 100
77 203.7 110.9 332.6





#77
Bromobenzene
Concen: 93.431 ug/l
RT: 12.982 min Scan# 1875
Delta R.T. -0.000 min
Lab File: VN083135.D
Acq: 07 Aug 2024 10:33




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100



Tgt Ion:156 Resp: 282670
Ion Ratio Lower Upper
156 100
77 230.9 111.3 333.8
158 96.4 48.3 144.9




Manual Integrations
APPROVED


Reviewed By :John Carbone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#78
n-propylbenzene
Concen: 93.979 ug/l
RT: 13.035 min Scan# 1884
Delta R.T. -0.000 min
Lab File: VN083135.D
Acq: 07 Aug 2024 10:33

Tgt Ion: 91 Resp: 1473868
Ion Ratio Lower Upper
91 100
120 22.3 10.9 32.9





#79

2-Chlorotoluene

Concen: 91.785 ug/l

RT: 13.123 min Scan# 1899

Delta R.T. -0.000 min

Lab File: VN083135.D

Acq: 07 Aug 2024 10:33

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100

Tgt Ion: 91 Resp: 912618

Ion Ratio Lower Upper

91 100





126 31.1 16.2 48.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#80

1,3,5-Trimethylbenzene

Concen: 92.690 ug/l

RT: 13.170 min Scan# 1907

Delta R.T. -0.000 min

Lab File: VN083135.D




Acq: 07 Aug 2024 10:33


Tgt Ion:105 Resp: 1056912

Ion Ratio Lower Upper

105 100


120 47.4 24.3 72.8





#81
trans-1,4-Dichloro-2-butene
Concen: 100.628 ug/l
RT: 12.735 min Scan# 1833
Delta R.T. -0.000 min
Lab File: VN083135.D
Acq: 07 Aug 2024 10:33




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100



Tgt Ion: 75 Resp: 165320
Ion Ratio Lower Upper
75 100
53 96.5 88.4 132.6
89 41.9 36.3 54.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#82
4-Chlorotoluene
Concen: 91.381 ug/l
RT: 13.223 min Scan# 1916
Delta R.T. -0.000 min
Lab File: VN083135.D
Acq: 07 Aug 2024 10:33


Tgt Ion: 91 Resp: 911384
Ion Ratio Lower Upper
91 100
126 31.4 16.2 48.6





#83
tert-Butylbenzene
Concen: 91.918 ug/l
RT: 13.441 min Scan# 1953
Delta R.T. -0.000 min
Lab File: VN083135.D
Acq: 07 Aug 2024 10:33




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100



Tgt Ion:119 Resp: 92819
Ion Ratio Lower Upper
119 100
91 69.3 35.0 105.1
134 26.1 13.3 39.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#84
1,2,4-Trimethylbenzene
Concen: 93.675 ug/l
RT: 13.482 min Scan# 1960
Delta R.T. -0.000 min
Lab File: VN083135.D
Acq: 07 Aug 2024 10:33


Tgt Ion:105 Resp: 1076449
Ion Ratio Lower Upper
105 100
120 44.4 21.9 65.8





#85
sec-Butylbenzene
Concen: 93.978 ug/l
RT: 13.617 min Scan# 1983
Delta R.T. -0.000 min
Lab File: VN083135.D
Acq: 07 Aug 2024 10:33




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100



Tgt Ion:105 Resp: 1294900
Ion Ratio Lower Upper
105 100
134 19.4 9.8 29.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#86
p-Isopropyltoluene
Concen: 96.022 ug/l
RT: 13.729 min Scan# 2002
Delta R.T. -0.000 min
Lab File: VN083135.D
Acq: 07 Aug 2024 10:33

Tgt Ion:119 Resp: 1092439
Ion Ratio Lower Upper
119 100
134 25.5 13.0 39.0
91 25.5 12.3 36.9





#87

1,3-Dichlorobenzene

Concen: 94.009 ug/l

RT: 13.735 min Scan# 2003

Delta R.T. -0.000 min

Lab File: VN083135.D


Acq: 07 Aug 2024 10:33

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC100



Tgt Ion:146 Resp: 53519

Ion Ratio Lower Upper

146 100

111 43.5 20.6 61.9




148 63.6 31.8 95.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#88

1,4-Dichlorobenzene


Concen: 93.900 ug/l

RT: 13.811 min Scan# 2016

Delta R.T. -0.000 min

Lab File: VN083135.D

Acq: 07 Aug 2024 10:33





Tgt Ion:146 Resp: 538924


Ion Ratio Lower Upper

146 100

111 41.5 20.1 60.3

148 63.6 32.2 96.6





#89

n-Butylbenzene

Concen: 99.461 ug/l

RT: 14.059 min Scan# 2058

Delta R.T. -0.000 min

Lab File: VN083135.D

Acq: 07 Aug 2024 10:33

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC100

Tgt Ion:	91	Resp:	98048
Ion Ratio	Lower	Upper	
91	100		
92	53.8	26.1	78.3
134	25.2	12.1	36.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#90

Hexachloroethane

Concen: 96.320 ug/l




RT: 14.335 min Scan# 2105


Delta R.T. -0.000 min

Lab File: VN083135.D

Acq: 07 Aug 2024 10:33


Tgt Ion:	117	Resp:	211710
Ion Ratio	Lower	Upper	
117	100		
201	67.7	35.8	107.3





#91
1,2-Dichlorobenzene
Concen: 92.863 ug/l
RT: 14.106 min Scan# 2066
Delta R.T. -0.000 min
Lab File: VN083135.D
Acq: 07 Aug 2024 10:33




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100



Tgt Ion:146 Resp: 511584
Ion Ratio Lower Upper
146 100
111 45.4 21.4 64.2
148 63.9 32.1 96.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#92
1,2-Dibromo-3-Chloropropane
Concen: 89.788 ug/l
RT: 14.723 min Scan# 2171
Delta R.T. 0.006 min
Lab File: VN083135.D
Acq: 07 Aug 2024 10:33

Tgt Ion: 75 Resp: 83930
Ion Ratio Lower Upper
75 100
155 75.2 36.6 109.8
157 95.1 46.9 140.6





#93

1,2,4-Trichlorobenzene

Concen: 96.456 ug/l

RT: 15.394 min Scan# 2285

Delta R.T. -0.000 min

Lab File: VN083135.D

Acq: 07 Aug 2024 10:33

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC100

Tgt Ion:180 Resp: 297699

Ion Ratio Lower Upper

180 100

182 94.1 47.3 141.8





145 33.6 15.7 47.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#94

Hexachlorobutadiene

Concen: 87.268 ug/l

RT: 15.500 min Scan# 2303

Delta R.T. -0.000 min

Lab File: VN083135.D

Acq: 07 Aug 2024 10:33




Tgt Ion:225 Resp: 119783


Ion Ratio Lower Upper

225 100

223 63.0 31.9 95.7


227 63.3 32.5 97.5





#95
Naphthalene
Concen: 95.391 ug/l
RT: 15.641 min Scan# 2327
Delta R.T. -0.000 min
Lab File: VN083135.D
Acq: 07 Aug 2024 10:33




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100



Tgt Ion:128 Resp: 1042864
Ion Ratio Lower Upper
128 100
127 12.7 10.6 16.0
129 10.9 8.6 12.8




Manual Integrations
APPROVED

Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#96
1,2,3-Trichlorobenzene
Concen: 93.575 ug/l
RT: 15.841 min Scan# 2361
Delta R.T. -0.000 min
Lab File: VN083135.D
Acq: 07 Aug 2024 10:33

Tgt Ion:180 Resp: 285769
Ion Ratio Lower Upper
180 100
182 95.1 48.9 146.8
145 36.0 16.8 50.4



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080724\
 Data File : VN083136.D
 Acq On : 07 Aug 2024 10:58
 Operator : JC\MD
 Sample : VSTDICCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICCC050

Manual Integrations
 APPROVED

Reviewed By : John Carlone 08/08/2024
 Supervised By : Mahesh Dadoda 08/09/2024

Quant Time: Aug 08 06:15:41 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 05:50:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	204914	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	348813	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	303118	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	143466	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	152684	52.348	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 104.700%			
35) Dibromofluoromethane	8.165	113	115863	53.216	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 106.440%			
50) Toluene-d8	10.565	98	439976	54.174	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 108.340%			
62) 4-Bromofluorobenzene	12.847	95	171524	54.173	ug/l	0.00
Spiked Amount 50.000	Range 64 - 133		Recovery = 108.340%			
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	2.124	85	116913	50.312	ug/l	97
3) Chloromethane	2.359	50	121130	50.915	ug/l	93
4) Vinyl Chloride	2.512	62	122548	50.482	ug/l	98
5) Bromomethane	2.948	94	73113	48.538	ug/l	95
6) Chloroethane	3.112	64	72966	48.043	ug/l	98
7) Trichlorofluoromethane	3.495	101	203848	50.832	ug/l	98
8) Diethyl Ether	3.959	74	75087	50.318	ug/l	84
9) 1,1,2-Trichlorotrifluo...	4.365	101	112201	50.747	ug/l	98
10) Methyl Iodide	4.589	142	151731	52.170	ug/l	96
11) Tert butyl alcohol	5.530	59	150267	247.904	ug/l	100
12) 1,1-Dichloroethene	4.342	96	111564	49.114	ug/l	87
13) Acrolein	4.183	56	104276	263.954	ug/l	98
14) Allyl chloride	5.018	41	191195	44.541	ug/l	93
15) Acrylonitrile	5.718	53	313090	251.149	ug/l	98
16) Acetone	4.430	43	293091	256.804	ug/l	97
17) Carbon Disulfide	4.712	76	319654	48.094	ug/l	98
18) Methyl Acetate	5.024	43	159022	46.765	ug/l	91
19) Methyl tert-butyl Ether	5.794	73	415489	50.676	ug/l	97
20) Methylene Chloride	5.277	84	122780	46.744	ug/l	85
21) trans-1,2-Dichloroethene	5.794	96	117685	50.127	ug/l	85
22) Diisopropyl ether	6.671	45	411409	50.987	ug/l	96
23) Vinyl Acetate	6.606	43	2147591m	259.716	ug/l	
24) 1,1-Dichloroethane	6.571	63	222071	50.493	ug/l	99
25) 2-Butanone	7.483	43	437392	249.591	ug/l	91
26) 2,2-Dichloropropane	7.488	77	209679	51.336	ug/l	97
27) cis-1,2-Dichloroethene	7.488	96	141110	49.809	ug/l	92
28) Bromochloromethane	7.812	49	89942	50.041	ug/l #	83
29) Tetrahydrofuran	7.835	42	285853	252.334	ug/l	88
30) Chloroform	7.965	83	232933	50.978	ug/l	99
31) Cyclohexane	8.259	56	201273	46.565	ug/l	96
32) 1,1,1-Trichloroethane	8.171	97	220296	50.935	ug/l	95
36) 1,1-Dichloropropene	8.371	75	165826	50.349	ug/l	99
37) Ethyl Acetate	7.559	43	174105	47.213	ug/l #	92
38) Carbon Tetrachloride	8.359	117	190357	51.315	ug/l	97
39) Methylcyclohexane	9.600	83	208177	51.461	ug/l	93
40) Benzene	8.606	78	504662	51.436	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080724\
 Data File : VN083136.D
 Acq On : 07 Aug 2024 10:58
 Operator : JC\MD
 Sample : VSTDICCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICCC050

Manual Integrations
 APPROVED

Reviewed By :John Carlone 08/08/2024
 Supervised By :Mahesh Dadoda 08/09/2024

Quant Time: Aug 08 06:15:41 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 05:50:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.777	41	98869	47.154	ug/l	94
42) 1,2-Dichloroethane	8.671	62	181440	50.765	ug/l	98
43) Isopropyl Acetate	8.688	43	321088	50.219	ug/l #	93
44) Trichloroethene	9.353	130	119998	51.383	ug/l	94
45) 1,2-Dichloropropane	9.624	63	120518	51.747	ug/l	97
46) Dibromomethane	9.712	93	87411	52.436	ug/l	99
47) Bromodichloromethane	9.888	83	188069	50.245	ug/l	99
48) Methyl methacrylate	9.682	41	148264	48.631	ug/l	91
49) 1,4-Dioxane	9.694	88	57033	1036.917	ug/l	95
51) 4-Methyl-2-Pentanone	10.447	43	901412	258.493	ug/l	92
52) Toluene	10.629	92	323148	52.127	ug/l	98
53) t-1,3-Dichloropropene	10.835	75	208411	54.205	ug/l	100
54) cis-1,3-Dichloropropene	10.312	75	211396	51.692	ug/l	91
55) 1,1,2-Trichloroethane	11.018	97	116384	52.387	ug/l	99
56) Ethyl methacrylate	10.876	69	217635	51.964	ug/l #	85
57) 1,3-Dichloropropane	11.165	76	206442	52.133	ug/l	100
58) 2-Chloroethyl Vinyl ether	10.159	63	454871	256.939	ug/l	95
59) 2-Hexanone	11.194	43	691708	256.396	ug/l	92
60) Dibromochloromethane	11.359	129	144208	53.688	ug/l	99
61) 1,2-Dibromoethane	11.471	107	121153	51.912	ug/l	98
64) Tetrachloroethene	11.106	164	103955	51.784	ug/l	96
65) Chlorobenzene	11.894	112	346172	51.679	ug/l	96
66) 1,1,1,2-Tetrachloroethane	11.965	131	121057	51.242	ug/l	100
67) Ethyl Benzene	11.965	91	628906	51.179	ug/l	98
68) m/p-Xylenes	12.070	106	477261	103.683	ug/l	99
69) o-Xylene	12.400	106	234432	51.638	ug/l	100
70) Styrene	12.412	104	405002	53.115	ug/l	98
71) Bromoform	12.576	173	96377	53.850	ug/l #	99
73) Isopropylbenzene	12.694	105	609529	50.795	ug/l	100
74) N-amyl acetate	12.494	43	287319	48.950	ug/l	94
75) 1,1,2,2-Tetrachloroethane	12.935	83	167489	49.349	ug/l	99
76) 1,2,3-Trichloropropane	12.994	75	152531m	48.363	ug/l	
77) Bromobenzene	12.982	156	137514	51.587	ug/l	97
78) n-propylbenzene	13.035	91	712768	51.584	ug/l	99
79) 2-Chlorotoluene	13.123	91	441262	50.371	ug/l	98
80) 1,3,5-Trimethylbenzene	13.170	105	514918	51.254	ug/l	98
81) trans-1,4-Dichloro-2-b...	12.735	75	70776	48.894	ug/l	89
82) 4-Chlorotoluene	13.223	91	445369	50.684	ug/l	99
83) tert-Butylbenzene	13.441	119	453597	50.983	ug/l	98
84) 1,2,4-Trimethylbenzene	13.482	105	524226	51.778	ug/l	100
85) sec-Butylbenzene	13.617	105	628959	51.809	ug/l	100
86) p-Isopropyltoluene	13.729	119	528236	52.699	ug/l	98
87) 1,3-Dichlorobenzene	13.735	146	254058	50.651	ug/l	99
88) 1,4-Dichlorobenzene	13.812	146	255814	50.589	ug/l	99
89) n-Butylbenzene	14.059	91	461372	53.120	ug/l	97
90) Hexachloroethane	14.335	117	98836	51.037	ug/l	98
91) 1,2-Dichlorobenzene	14.106	146	244462	50.365	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	14.717	75	40369	49.017	ug/l	98
93) 1,2,4-Trichlorobenzene	15.394	180	141229	51.937	ug/l	99
94) Hexachlorobutadiene	15.500	225	58664	48.509	ug/l	98
95) Naphthalene	15.641	128	499372	51.844	ug/l	99
96) 1,2,3-Trichlorobenzene	15.841	180	138125	51.335	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080724\
Data File : VN083136.D
Acq On : 07 Aug 2024 10:58
Operator : JC\MD
Sample : VSTDICCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Manual Integrations
APPROVED

Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024

Quant Time: Aug 08 06:15:41 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 05:50:26 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080724\
Data File : VN083136.D
Acq On : 07 Aug 2024 10:58
Operator : JC\MD
Sample : VSTDICCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1


Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED

Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024


Quant Time: Aug 08 06:15:41 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 05:50:26 2024
Response via : Initial Calibration





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 1066
Delta R.T. -0.000 min
Lab File: VN083136.D
Acq: 07 Aug 2024 10:58




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050



Tgt Ion:168 Resp: 204914
Ion Ratio Lower Upper
168 100
99 65.4 48.2 72.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#2
Dichlorodifluoromethane
Concen: 50.312 ug/l
RT: 2.124 min Scan# 29
Delta R.T. 0.000 min
Lab File: VN083136.D
Acq: 07 Aug 2024 10:58

Tgt Ion: 85 Resp: 116913
Ion Ratio Lower Upper
85 100
87 33.0 15.7 47.0









#3
Chloromethane
Concen: 50.915 ug/l
RT: 2.359 min Scan# 69
Delta R.T. 0.000 min
Lab File: VN083136.D
Acq: 07 Aug 2024 10:58

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Tgt Ion: 50 Resp: 121130
Ion Ratio Lower Upper
50 100
52 35.1 25.0 37.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#4
Vinyl Chloride
Concen: 50.482 ug/l
RT: 2.512 min Scan# 95
Delta R.T. 0.000 min
Lab File: VN083136.D
Acq: 07 Aug 2024 10:58

Tgt Ion: 62 Resp: 122548
Ion Ratio Lower Upper
62 100
64 32.3 25.0 37.6





#5

Bromomethane

Concen: 48.538 ug/l

RT: 2.948 min Scan# 169

Delta R.T. 0.000 min

Lab File: VN083136.D

Acq: 07 Aug 2024 10:58

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 94 Resp: 7311

Ion Ratio Lower Upper

94 100





96 92.2 78.0 117.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#6

Chloroethane

Concen: 48.043 ug/l

RT: 3.112 min Scan# 197

Delta R.T. 0.000 min

Lab File: VN083136.D




Acq: 07 Aug 2024 10:58


Tgt Ion: 64 Resp: 72966

Ion Ratio Lower Upper

64 100


66 34.3 26.6 40.0





#7
Trichlorofluoromethane
Concen: 50.832 ug/l
RT: 3.495 min Scan# 262
Delta R.T. -0.000 min
Lab File: VN083136.D
Acq: 07 Aug 2024 10:58




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050



Tgt Ion:101 Resp: 20384
Ion Ratio Lower Upper
101 100
103 64.3 52.7 79.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#8
Diethyl Ether
Concen: 50.318 ug/l
RT: 3.959 min Scan# 341
Delta R.T. 0.000 min
Lab File: VN083136.D
Acq: 07 Aug 2024 10:58

Tgt Ion: 74 Resp: 75087
Ion Ratio Lower Upper
74 100
45 94.0 55.5 166.3





#9

1,1,2-Trichlorotrifluoroethane

Concen: 50.747 ug/l

RT: 4.365 min Scan# 410

Delta R.T. 0.000 min

Lab File: VN083136.D

Acq: 07 Aug 2024 10:58

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion:101 Resp: 11220

Ion Ratio Lower Upper

101 100

85 45.9 38.3 57.5





151 75.1 59.3 88.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#10

Methyl Iodide

Concen: 52.170 ug/l

RT: 4.589 min Scan# 448

Delta R.T. 0.000 min

Lab File: VN083136.D

Acq: 07 Aug 2024 10:58




Tgt Ion:142 Resp: 151731


Ion Ratio Lower Upper

142 100

127 44.2 37.5 56.3

141 14.3 13.1 19.7





#11

Tert butyl alcohol

Concen: 247.904 ug/l

RT: 5.530 min Scan# 608

Delta R.T. 0.000 min

Lab File: VN083136.D

Acq: 07 Aug 2024 10:58

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 59 Resp: 15026

Ion Ratio Lower Upper

59 100





57 11.0 8.9 13.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#12

1,1-Dichloroethene

Concen: 49.114 ug/l

RT: 4.342 min Scan# 406

Delta R.T. 0.000 min

Lab File: VN083136.D

Acq: 07 Aug 2024 10:58




Tgt Ion: 96 Resp: 111564


Ion Ratio Lower Upper

96 100

61 162.0 149.7 224.5

98 61.2 50.1 75.1





#13

Acrolein

Concen: 263.954 ug/l

RT: 4.183 min Scan# 379

Delta R.T. 0.000 min

Lab File: VN083136.D

Acq: 07 Aug 2024 10:58

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 56 Resp: 104270

Ion Ratio Lower Upper

56 100





55 68.8 56.4 84.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#14

Allyl chloride

Concen: 44.541 ug/l

RT: 5.018 min Scan# 521

Delta R.T. 0.000 min

Lab File: VN083136.D

Acq: 07 Aug 2024 10:58




Tgt Ion: 41 Resp: 191195


Ion Ratio Lower Upper

41 100

39 78.9 67.8 101.6


76 38.1 25.8 38.8





#15
Acrylonitrile
Concen: 251.149 ug/l
RT: 5.718 min Scan# 640
Delta R.T. 0.000 min
Lab File: VN083136.D
Acq: 07 Aug 2024 10:58




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050



Tgt Ion: 53 Resp: 313090
Ion Ratio Lower Upper
53 100
52 81.2 66.6 100.0
51 36.5 30.1 45.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#16
Acetone
Concen: 256.804 ug/l
RT: 4.430 min Scan# 421
Delta R.T. 0.000 min
Lab File: VN083136.D
Acq: 07 Aug 2024 10:58

Tgt Ion: 43 Resp: 293091
Ion Ratio Lower Upper
43 100
58 29.0 21.8 32.6





#17

Carbon Disulfide

Concen: 48.094 ug/l

RT: 4.712 min Scan# 469

Delta R.T. 0.000 min

Lab File: VN083136.D

Acq: 07 Aug 2024 10:58

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 76 Resp: 319654

Ion Ratio Lower Upper

76 100





78 9.6 7.2 10.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#18

Methyl Acetate

Concen: 46.765 ug/l

RT: 5.024 min Scan# 522

Delta R.T. 0.000 min

Lab File: VN083136.D




Acq: 07 Aug 2024 10:58


Tgt Ion: 43 Resp: 159022

Ion Ratio Lower Upper

43 100

74 24.3 16.3 24.5





#19

Methyl tert-butyl Ether

Concen: 50.676 ug/l

RT: 5.794 min Scan# 653

Delta R.T. 0.000 min

Lab File: VN083136.D

Acq: 07 Aug 2024 10:58

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 73 Resp: 415489

Ion Ratio Lower Upper

73 100





57 20.7 17.9 26.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#20

Methylene Chloride

Concen: 46.744 ug/l

RT: 5.277 min Scan# 565

Delta R.T. 0.000 min

Lab File: VN083136.D

Acq: 07 Aug 2024 10:58

Tgt Ion: 84 Resp: 122780




Ion Ratio Lower Upper


84 100

49 119.8 119.6 179.4

51 39.8 34.8 52.2


86 63.4 52.9 79.3





#21
trans-1,2-Dichloroethene
Concen: 50.127 ug/l
RT: 5.794 min Scan# 61
Delta R.T. 0.000 min
Lab File: VN083136.D
Acq: 07 Aug 2024 10:58




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050



Tgt Ion: 96 Resp: 11768
Ion Ratio Lower Upper
96 100
61 128.4 123.4 185.2
98 58.7 49.3 73.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#22
Diisopropyl ether
Concen: 50.987 ug/l
RT: 6.671 min Scan# 802
Delta R.T. 0.000 min
Lab File: VN083136.D
Acq: 07 Aug 2024 10:58

Tgt Ion: 45 Resp: 411409
Ion Ratio Lower Upper
45 100
43 52.8 44.0 66.0
87 27.7 19.7 29.5
59 11.5 8.5 12.7









#23
Vinyl Acetate
Concen: 259.716 ug/l m
RT: 6.606 min Scan# 791
Delta R.T. 0.000 min
Lab File: VN083136.D
Acq: 07 Aug 2024 10:58

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Tgt Ion: 43 Resp: 214759
Ion Ratio Lower Upper
43 100
86 11.0 7.3 10.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#24
1,1-Dichloroethane
Concen: 50.493 ug/l
RT: 6.571 min Scan# 785
Delta R.T. 0.000 min
Lab File: VN083136.D
Acq: 07 Aug 2024 10:58

Tgt Ion: 63 Resp: 222071
Ion Ratio Lower Upper
63 100
98 6.9 3.3 9.9
100 4.6 2.0 6.0





#25

2-Butanone

Concen: 249.591 ug/l

RT: 7.483 min Scan# 940

Delta R.T. 0.000 min

Lab File: VN083136.D

Acq: 07 Aug 2024 10:58

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 43 Resp: 437392

Ion Ratio Lower Upper

43 100





72 27.1 18.2 27.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#26

2,2-Dichloropropane

Concen: 51.336 ug/l

RT: 7.488 min Scan# 941

Delta R.T. 0.000 min

Lab File: VN083136.D




Acq: 07 Aug 2024 10:58


Tgt Ion: 77 Resp: 209679

Ion Ratio Lower Upper

77 100

97 21.9 10.3 30.9





#27

cis-1,2-Dichloroethene

Concen: 49.809 ug/l

RT: 7.488 min Scan# 941

Delta R.T. 0.000 min

Lab File: VN083136.D

Acq: 07 Aug 2024 10:58

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 96 Resp: 141110

Ion Ratio Lower Upper

96 100

61 150.4 0.0 329.6





98 64.5 0.0 128.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#28

Bromochloromethane

Concen: 50.041 ug/l

RT: 7.812 min Scan# 996

Delta R.T. 0.000 min

Lab File: VN083136.D

Acq: 07 Aug 2024 10:58




Tgt Ion: 49 Resp: 89942


Ion Ratio Lower Upper

49 100

129 1.8 0.0 3.8

130 76.7 50.5 75.7#





#29

Tetrahydrofuran

Concen: 252.334 ug/l

RT: 7.835 min Scan# 1000

Delta R.T. 0.000 min

Lab File: VN083136.D

Acq: 07 Aug 2024 10:58

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 42 Resp: 28585

Ion Ratio Lower Upper

42 100

72 44.6 30.0 45.0





71 42.4 28.3 42.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#30

Chloroform

Concen: 50.978 ug/l

RT: 7.965 min Scan# 1022

Delta R.T. 0.000 min

Lab File: VN083136.D




Acq: 07 Aug 2024 10:58


Tgt Ion: 83 Resp: 232933

Ion Ratio Lower Upper

83 100

85 64.5 50.9 76.3





#31

Cyclohexane

Concen: 46.565 ug/l

RT: 8.259 min Scan# 1072

Delta R.T. 0.000 min

Lab File: VN083136.D

Acq: 07 Aug 2024 10:58

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 56 Resp: 20127

Ion Ratio Lower Upper

56 100

69 32.7 24.5 36.7





84 86.4 66.0 99.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#32

1,1,1-Trichloroethane

Concen: 50.935 ug/l

RT: 8.171 min Scan# 1057

Delta R.T. 0.000 min

Lab File: VN083136.D

Acq: 07 Aug 2024 10:58




Tgt Ion: 97 Resp: 220296


Ion Ratio Lower Upper

97 100

99 62.5 52.0 78.0

61 46.6 42.1 63.1





#33

1,2-Dichloroethane-d4

Concen: 52.348 ug/l

RT: 8.582 min Scan# 1127

Delta R.T. 0.000 min

Lab File: VN083136.D

Acq: 07 Aug 2024 10:58

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 65 Resp: 152684

Ion Ratio Lower Upper

65 100





67 50.5 0.0 104.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.100 min Scan# 1215

Delta R.T. 0.000 min

Lab File: VN083136.D

Acq: 07 Aug 2024 10:58




Tgt Ion:114 Resp: 348813


Ion Ratio Lower Upper

114 100

63 22.2 0.0 44.6

88 17.3 0.0 31.4





#35

Dibromofluoromethane

Concen: 53.216 ug/l

RT: 8.165 min Scan# 1056

Delta R.T. 0.000 min

Lab File: VN083136.D

Acq: 07 Aug 2024 10:58

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion:113 Resp: 11586

Ion Ratio Lower Upper

113 100

111 101.9 82.4 123.6





192 17.5 14.9 22.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#36

1,1-Dichloropropene

Concen: 50.349 ug/l

RT: 8.371 min Scan# 1091

Delta R.T. 0.000 min

Lab File: VN083136.D

Acq: 07 Aug 2024 10:58




Tgt Ion: 75 Resp: 165826


Ion Ratio Lower Upper

75 100

110 34.2 16.3 48.9

77 30.8 24.6 37.0





#37

Ethyl Acetate

Concen: 47.213 ug/l

RT: 7.559 min Scan# 91

Delta R.T. 0.000 min

Lab File: VN083136.D

Acq: 07 Aug 2024 10:58

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 43 Resp: 17410

Ion Ratio Lower Upper

43 100

61 14.4 9.3 13.9





70 12.1 7.0 10.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#38

Carbon Tetrachloride

Concen: 51.315 ug/l

RT: 8.359 min Scan# 1089

Delta R.T. 0.000 min

Lab File: VN083136.D

Acq: 07 Aug 2024 10:58




Tgt Ion:117 Resp: 190357


Ion Ratio Lower Upper

117 100

119 98.0 74.9 112.3

121 30.3 24.3 36.5





#39

Methylcyclohexane

Concen: 51.461 ug/l

RT: 9.600 min Scan# 1130

Delta R.T. 0.000 min

Lab File: VN083136.D

Acq: 07 Aug 2024 10:58

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050





Tgt Ion:	83	Resp:	20817
Ion Ratio	Lower	Upper	
83	100		
55	76.0	66.4	99.6
98	46.3	40.7	61.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#40

Benzene

Concen: 51.436 ug/l




RT: 8.606 min Scan# 1131


Delta R.T. 0.000 min

Lab File: VN083136.D

Acq: 07 Aug 2024 10:58

Tgt Ion:	78	Resp:	504662
Ion Ratio	Lower	Upper	
78	100		
77	24.8	19.0	28.4









#41
Methacrylonitrile
Concen: 47.154 ug/l
RT: 7.777 min Scan# 990
Delta R.T. 0.000 min
Lab File: VN083136.D
Acq: 07 Aug 2024 10:58

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Tgt Ion: 41 Resp: 98869
Ion Ratio Lower Upper
41 100
39 55.4 45.5 68.3
67 72.0 50.5 75.7
52 30.0 24.0 36.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#42
1,2-Dichloroethane
Concen: 50.765 ug/l
RT: 8.671 min Scan# 1142
Delta R.T. 0.000 min
Lab File: VN083136.D
Acq: 07 Aug 2024 10:58

Tgt Ion: 62 Resp: 181440
Ion Ratio Lower Upper
62 100
98 8.7 0.0 15.8





#43

Isopropyl Acetate

Concen: 50.219 ug/l

RT: 8.688 min Scan# 1145

Delta R.T. 0.000 min

Lab File: VN083136.D


Acq: 07 Aug 2024 10:58

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050






Tgt Ion:	43	Resp:	32108
Ion	Ratio	Lower	Upper
43	100		
61	25.3	17.8	26.6
87	13.0	8.2	12.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#44

Trichloroethene

Concen: 51.383 ug/l




RT: 9.353 min Scan# 1258


Delta R.T. 0.000 min

Lab File: VN083136.D

Acq: 07 Aug 2024 10:58

Tgt Ion:	130	Resp:	119998
Ion	Ratio	Lower	Upper
130	100		
95	104.7	0.0	197.8





#45

1,2-Dichloropropane

Concen: 51.747 ug/l

RT: 9.624 min Scan# 1304

Delta R.T. 0.000 min

Lab File: VN083136.D

Acq: 07 Aug 2024 10:58

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 63 Resp: 120518

Ion Ratio Lower Upper

63 100





65 32.1 24.4 36.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#46

Dibromomethane

Concen: 52.436 ug/l

RT: 9.712 min Scan# 1319

Delta R.T. 0.000 min

Lab File: VN083136.D

Acq: 07 Aug 2024 10:58




Tgt Ion: 93 Resp: 87411


Ion Ratio Lower Upper

93 100

95 81.7 65.8 98.6

174 90.1 71.7 107.5





#47

Bromodichloromethane

Concen: 50.245 ug/l

RT: 9.888 min Scan# 1349

Delta R.T. 0.000 min

Lab File: VN083136.D

Acq: 07 Aug 2024 10:58

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 83 Resp: 188069

Ion Ratio Lower Upper

83 100

85 64.3 51.1 76.7





127 8.5 7.2 10.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#48

Methyl methacrylate

Concen: 48.631 ug/l

RT: 9.682 min Scan# 1314

Delta R.T. 0.000 min

Lab File: VN083136.D

Acq: 07 Aug 2024 10:58




Tgt Ion: 41 Resp: 148264


Ion Ratio Lower Upper

41 100

69 82.5 56.3 84.5

39 61.0 50.3 75.5





#49

1,4-Dioxane

Concen: 1036.917 ug/l

RT: 9.694 min Scan# 1316

Delta R.T. 0.000 min

Lab File: VN083136.D

Acq: 07 Aug 2024 10:58

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 88 Resp: 5703

Ion Ratio Lower Upper

88 100

43 29.1 27.8 41.8





58 72.4 59.4 89.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#50

Toluene-d8

Concen: 54.174 ug/l

RT: 10.565 min Scan# 1464

Delta R.T. 0.000 min

Lab File: VN083136.D




Acq: 07 Aug 2024 10:58


Tgt Ion: 98 Resp: 439976

Ion Ratio Lower Upper

98 100


100 65.5 51.5 77.3





#51
4-Methyl-2-Pentanone
Concen: 258.493 ug/l
RT: 10.447 min Scan# 1444
Delta R.T. 0.000 min
Lab File: VN083136.D
Acq: 07 Aug 2024 10:58




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050



Tgt Ion: 43 Resp: 901412
Ion Ratio Lower Upper
43 100
58 39.8 28.2 42.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#52
Toluene
Concen: 52.127 ug/l
RT: 10.629 min Scan# 1475
Delta R.T. 0.000 min
Lab File: VN083136.D
Acq: 07 Aug 2024 10:58


Tgt Ion: 92 Resp: 323148
Ion Ratio Lower Upper
92 100
91 172.1 139.4 209.0





#53
t-1,3-Dichloropropene
Concen: 54.205 ug/l
RT: 10.835 min Scan# 1510
Delta R.T. 0.000 min
Lab File: VN083136.D
Acq: 07 Aug 2024 10:58




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050



Tgt Ion: 75 Resp: 20841
Ion Ratio Lower Upper
75 100
77 31.9 25.7 38.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#54
cis-1,3-Dichloropropene
Concen: 51.692 ug/l
RT: 10.312 min Scan# 1421
Delta R.T. 0.000 min
Lab File: VN083136.D
Acq: 07 Aug 2024 10:58


Tgt Ion: 75 Resp: 211396
Ion Ratio Lower Upper
75 100
77 32.8 24.3 36.5
39 54.0 50.5 75.7





#55
1,1,2-Trichloroethane
Concen: 52.387 ug/l
RT: 11.018 min Scan# 1541
Delta R.T. 0.000 min
Lab File: VN083136.D
Acq: 07 Aug 2024 10:58




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050



Tgt Ion: 97 Resp: 116384
Ion Ratio Lower Upper
97 100
83 89.4 71.4 107.0
85 55.4 45.8 68.6
99 64.4 52.0 78.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#56
Ethyl methacrylate
Concen: 51.964 ug/l
RT: 10.876 min Scan# 1517
Delta R.T. 0.000 min
Lab File: VN083136.D
Acq: 07 Aug 2024 10:58

Tgt Ion: 69 Resp: 217635
Ion Ratio Lower Upper
69 100
41 66.9 63.4 95.2
39 35.2 37.4 56.0#





#57

1,3-Dichloropropane

Concen: 52.133 ug/l

RT: 11.165 min Scan# 1566

Delta R.T. 0.000 min

Lab File: VN083136.D

Acq: 07 Aug 2024 10:58

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 76 Resp: 20644

Ion Ratio Lower Upper

76 100





78 32.2 25.6 38.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#58

2-Chloroethyl Vinyl ether

Concen: 256.939 ug/l

RT: 10.159 min Scan# 1395

Delta R.T. 0.000 min

Lab File: VN083136.D




Acq: 07 Aug 2024 10:58


Tgt Ion: 63 Resp: 454871

Ion Ratio Lower Upper

63 100


106 29.0 21.4 32.0





#59
2-Hexanone
Concen: 256.396 ug/l
RT: 11.194 min Scan# 1571
Delta R.T. 0.000 min
Lab File: VN083136.D
Acq: 07 Aug 2024 10:58




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050



Tgt Ion: 43 Resp: 691708
Ion Ratio Lower Upper
43 100
58 54.9 24.6 73.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#60
Dibromochloromethane
Concen: 53.688 ug/l
RT: 11.359 min Scan# 1599
Delta R.T. 0.000 min
Lab File: VN083136.D
Acq: 07 Aug 2024 10:58

Tgt Ion:129 Resp: 144208
Ion Ratio Lower Upper
129 100
127 77.4 39.2 117.6









#61
1,2-Dibromoethane
Concen: 51.912 ug/l
RT: 11.471 min Scan# 1618
Delta R.T. 0.000 min
Lab File: VN083136.D
Acq: 07 Aug 2024 10:58

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Tgt Ion:107 Resp: 12115
Ion Ratio Lower Upper
107 100
109 94.4 73.7 110.5




Manual Integrations
APPROVED


Reviewed By :John Carbone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#62
4-Bromofluorobenzene
Concen: 54.173 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. 0.000 min
Lab File: VN083136.D
Acq: 07 Aug 2024 10:58


Tgt Ion: 95 Resp: 171524
Ion Ratio Lower Upper
95 100
174 70.0 0.0 159.2
176 66.9 0.0 147.6





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 1685
Delta R.T. 0.000 min
Lab File: VN083136.D
Acq: 07 Aug 2024 10:58




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050



Tgt Ion:117 Resp: 303118
Ion Ratio Lower Upper
117 100
82 57.9 47.5 71.3
119 32.7 25.6 38.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#64
Tetrachloroethene
Concen: 51.784 ug/l
RT: 11.106 min Scan# 1556
Delta R.T. 0.000 min
Lab File: VN083136.D
Acq: 07 Aug 2024 10:58


Tgt Ion:164 Resp: 103955
Ion Ratio Lower Upper
164 100
166 124.8 101.4 152.0
129 103.3 75.7 113.5
131 95.3 73.8 110.6





#65
Chlorobenzene
Concen: 51.679 ug/l
RT: 11.894 min Scan# 1690
Delta R.T. 0.000 min
Lab File: VN083136.D
Acq: 07 Aug 2024 10:58




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050



Tgt Ion:112 Resp: 346172
Ion Ratio Lower Upper
112 100
114 30.6 26.2 39.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#66
1,1,1,2-Tetrachloroethane
Concen: 51.242 ug/l
RT: 11.965 min Scan# 1702
Delta R.T. 0.000 min
Lab File: VN083136.D
Acq: 07 Aug 2024 10:58

Tgt Ion:131 Resp: 121057
Ion Ratio Lower Upper
131 100
133 94.9 47.3 142.0
119 64.7 32.5 97.4





#67

Ethyl Benzene

Concen: 51.179 ug/l

RT: 11.965 min Scan# 1702

Delta R.T. 0.000 min

Lab File: VN083136.D

Acq: 07 Aug 2024 10:58

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 91 Resp: 628900

Ion Ratio Lower Upper

91 100





106 31.3 24.0 36.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#68

m/p-Xylenes

Concen: 103.683 ug/l

RT: 12.070 min Scan# 1720

Delta R.T. 0.000 min

Lab File: VN083136.D




Acq: 07 Aug 2024 10:58


Tgt Ion:106 Resp: 477261

Ion Ratio Lower Upper

106 100

91 209.9 166.1 249.1





#69

o-Xylene

Concen: 51.638 ug/l

RT: 12.400 min Scan# 1776

Delta R.T. 0.000 min

Lab File: VN083136.D

Acq: 07 Aug 2024 10:58

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion:106 Resp: 234432

Ion Ratio Lower Upper

106 100





91 222.4 111.3 333.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#70

Styrene

Concen: 53.115 ug/l

RT: 12.412 min Scan# 1778

Delta R.T. 0.000 min

Lab File: VN083136.D

Acq: 07 Aug 2024 10:58




Tgt Ion:104 Resp: 405002


Ion Ratio Lower Upper

104 100

78 53.4 41.6 62.4


103 53.7 44.0 66.0





#71
Bromoform
Concen: 53.850 ug/l
RT: 12.576 min Scan# 1806
Delta R.T. 0.000 min
Lab File: VN083136.D
Acq: 07 Aug 2024 10:58




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050




Tgt Ion:173 Resp: 96372
Ion Ratio Lower Upper
173 100
175 48.9 24.0 72.0
254 0.0 0.0 0.0

Manual Integrations
APPROVED



Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.794 min Scan# 1213
Delta R.T. 0.000 min
Lab File: VN083136.D
Acq: 07 Aug 2024 10:58




Tgt Ion:152 Resp: 143466
Ion Ratio Lower Upper
152 100
115 63.9 30.6 91.6
150 177.7 0.0 348.6





#73
Isopropylbenzene
Concen: 50.795 ug/l
RT: 12.694 min Scan# 1826
Delta R.T. 0.000 min
Lab File: VN083136.D
Acq: 07 Aug 2024 10:58




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050



Tgt Ion:105 Resp: 609529
Ion Ratio Lower Upper
105 100
120 25.8 12.9 38.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#74
N-ethyl acetate
Concen: 48.950 ug/l
RT: 12.494 min Scan# 1792
Delta R.T. 0.000 min
Lab File: VN083136.D
Acq: 07 Aug 2024 10:58


Tgt Ion: 43 Resp: 287319
Ion Ratio Lower Upper
43 100
70 42.7 29.8 44.6
55 23.8 18.7 28.1
61 24.6 17.4 26.0





#75
1,1,2,2-Tetrachloroethane
Concen: 49.349 ug/l
RT: 12.935 min Scan# 1867
Delta R.T. 0.000 min
Lab File: VN083136.D
Acq: 07 Aug 2024 10:58




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050



Tgt Ion: 83 Resp: 167489
Ion Ratio Lower Upper
83 100
131 10.4 5.0 14.9
85 64.4 31.9 95.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#76
1,2,3-Trichloropropane
Concen: 48.363 ug/l m
RT: 12.994 min Scan# 1877
Delta R.T. 0.000 min
Lab File: VN083136.D
Acq: 07 Aug 2024 10:58


Tgt Ion: 75 Resp: 152531
Ion Ratio Lower Upper
75 100
77 207.4 110.9 332.6





#77
Bromobenzene
Concen: 51.587 ug/l
RT: 12.982 min Scan# 1875
Delta R.T. 0.000 min
Lab File: VN083136.D
Acq: 07 Aug 2024 10:58




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050



Tgt Ion:156 Resp: 137514
Ion Ratio Lower Upper
156 100
77 230.1 111.3 333.8
158 97.4 48.3 144.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#78
n-propylbenzene
Concen: 51.584 ug/l
RT: 13.035 min Scan# 1884
Delta R.T. 0.000 min
Lab File: VN083136.D
Acq: 07 Aug 2024 10:58

Tgt Ion: 91 Resp: 712768
Ion Ratio Lower Upper
91 100
120 22.3 10.9 32.9





#79

2-Chlorotoluene

Concen: 50.371 ug/l

RT: 13.123 min Scan# 1899

Delta R.T. 0.000 min

Lab File: VN083136.D

Acq: 07 Aug 2024 10:58

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 91 Resp: 441262

Ion Ratio Lower Upper

91 100





126 31.1 16.2 48.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#80

1,3,5-Trimethylbenzene

Concen: 51.254 ug/l

RT: 13.170 min Scan# 1907

Delta R.T. 0.000 min

Lab File: VN083136.D




Acq: 07 Aug 2024 10:58


Tgt Ion:105 Resp: 514918

Ion Ratio Lower Upper

105 100


120 47.5 24.3 72.8





#81
trans-1,4-Dichloro-2-butene
Concen: 48.894 ug/l
RT: 12.735 min Scan# 1833
Delta R.T. 0.000 min
Lab File: VN083136.D
Acq: 07 Aug 2024 10:58




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050



Tgt Ion: 75 Resp: 70770
Ion Ratio Lower Upper
75 100
53 94.2 88.4 132.6
89 44.3 36.3 54.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#82
4-Chlorotoluene
Concen: 50.684 ug/l
RT: 13.223 min Scan# 1916
Delta R.T. 0.000 min
Lab File: VN083136.D
Acq: 07 Aug 2024 10:58


Tgt Ion: 91 Resp: 445369
Ion Ratio Lower Upper
91 100
126 31.6 16.2 48.6





#83
tert-Butylbenzene
Concen: 50.983 ug/l
RT: 13.441 min Scan# 1953
Delta R.T. 0.000 min
Lab File: VN083136.D
Acq: 07 Aug 2024 10:58




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050



Tgt Ion:119 Resp: 453597
Ion Ratio Lower Upper
119 100
91 68.9 35.0 105.1
134 24.1 13.3 39.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#84
1,2,4-Trimethylbenzene
Concen: 51.778 ug/l
RT: 13.482 min Scan# 1960
Delta R.T. 0.000 min
Lab File: VN083136.D
Acq: 07 Aug 2024 10:58


Tgt Ion:105 Resp: 524226
Ion Ratio Lower Upper
105 100
120 44.1 21.9 65.8





#85
sec-Butylbenzene
Concen: 51.809 ug/l
RT: 13.617 min Scan# 1983
Delta R.T. 0.000 min
Lab File: VN083136.D
Acq: 07 Aug 2024 10:58




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050



Tgt Ion:105 Resp: 628959
Ion Ratio Lower Upper
105 100
134 19.6 9.8 29.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#86
p-Isopropyltoluene
Concen: 52.699 ug/l
RT: 13.729 min Scan# 2002
Delta R.T. 0.000 min
Lab File: VN083136.D
Acq: 07 Aug 2024 10:58

Tgt Ion:119 Resp: 528236
Ion Ratio Lower Upper
119 100
134 24.8 13.0 39.0
91 25.4 12.3 36.9





#87

1,3-Dichlorobenzene

Concen: 50.651 ug/l

RT: 13.735 min Scan# 2003

Delta R.T. 0.000 min

Lab File: VN083136.D

Acq: 07 Aug 2024 10:58

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion:146 Resp: 254058

Ion Ratio Lower Upper

146 100

111 43.1 20.6 61.9





148 63.5 31.8 95.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#88

1,4-Dichlorobenzene

Concen: 50.589 ug/l

RT: 13.812 min Scan# 2016

Delta R.T. 0.000 min

Lab File: VN083136.D

Acq: 07 Aug 2024 10:58




Tgt Ion:146 Resp: 255814


Ion Ratio Lower Upper

146 100

111 42.0 20.1 60.3

148 64.0 32.2 96.6









#89
n-Butylbenzene
Concen: 53.120 ug/l
RT: 14.059 min Scan# 2058
Delta R.T. 0.000 min
Lab File: VN083136.D
Acq: 07 Aug 2024 10:58

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Tgt Ion: 91 Resp: 461372
Ion Ratio Lower Upper
91 100
92 54.4 26.1 78.3
134 25.1 12.1 36.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#90
Hexachloroethane
Concen: 51.037 ug/l
RT: 14.335 min Scan# 2105
Delta R.T. 0.000 min
Lab File: VN083136.D
Acq: 07 Aug 2024 10:58


Tgt Ion:117 Resp: 98836
Ion Ratio Lower Upper
117 100
201 69.9 35.8 107.3





#91
1,2-Dichlorobenzene
Concen: 50.365 ug/l
RT: 14.106 min Scan# 2066
Delta R.T. 0.000 min
Lab File: VN083136.D
Acq: 07 Aug 2024 10:58




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050



Tgt Ion:146 Resp: 244462
Ion Ratio Lower Upper
146 100
111 45.3 21.4 64.2
148 63.6 32.1 96.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#92
1,2-Dibromo-3-Chloropropane
Concen: 49.017 ug/l
RT: 14.717 min Scan# 2170
Delta R.T. 0.000 min
Lab File: VN083136.D
Acq: 07 Aug 2024 10:58

Tgt Ion: 75 Resp: 40369
Ion Ratio Lower Upper
75 100
155 76.6 36.6 109.8
157 93.4 46.9 140.6





#93

1,2,4-Trichlorobenzene

Concen: 51.937 ug/l

RT: 15.394 min Scan# 2285

Delta R.T. 0.000 min

Lab File: VN083136.D

Acq: 07 Aug 2024 10:58

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion:180 Resp: 141229

Ion Ratio Lower Upper

180 100

182 94.1 47.3 141.8





145 33.3 15.7 47.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#94

Hexachlorobutadiene

Concen: 48.509 ug/l

RT: 15.500 min Scan# 2303

Delta R.T. 0.000 min

Lab File: VN083136.D

Acq: 07 Aug 2024 10:58




Tgt Ion:225 Resp: 58664


Ion Ratio Lower Upper

225 100

223 63.9 31.9 95.7


227 62.3 32.5 97.5





#95
Naphthalene
Concen: 51.844 ug/l
RT: 15.641 min Scan# 2327
Delta R.T. 0.000 min
Lab File: VN083136.D
Acq: 07 Aug 2024 10:58




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050



Tgt Ion:128 Resp: 499372
Ion Ratio Lower Upper
128 100
127 12.9 10.6 16.0
129 10.7 8.6 12.8




Manual Integrations
APPROVED

Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#96
1,2,3-Trichlorobenzene
Concen: 51.335 ug/l
RT: 15.841 min Scan# 2361
Delta R.T. 0.000 min
Lab File: VN083136.D
Acq: 07 Aug 2024 10:58

Tgt Ion:180 Resp: 138125
Ion Ratio Lower Upper
180 100
182 95.2 48.9 146.8
145 35.8 16.8 50.4



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080724\
 Data File : VN083137.D
 Acq On : 07 Aug 2024 11:22
 Operator : JC\MD
 Sample : VSTDICC020
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC020

Manual Integrations
 APPROVED

Reviewed By : John Carlone 08/08/2024
 Supervised By : Mahesh Dadoda 08/09/2024

Quant Time: Aug 08 06:16:37 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 05:50:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	212913	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	362741	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	313255	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	149244	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	51159	16.881	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	33.760%#		
35) Dibromofluoromethane	8.171	113	38354	16.940	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	33.880%#		
50) Toluene-d8	10.565	98	142952	16.926	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	33.860%#		
62) 4-Bromofluorobenzene	12.847	95	56566	17.179	ug/l	0.00
Spiked Amount 50.000	Range 64 - 133		Recovery =	34.360%#		
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	2.124	85	48191	19.959	ug/l	95
3) Chloromethane	2.359	50	49777	20.137	ug/l	99
4) Vinyl Chloride	2.512	62	50108	19.866	ug/l	97
5) Bromomethane	2.942	94	31525	20.143	ug/l	89
6) Chloroethane	3.112	64	30331	19.221	ug/l	100
7) Trichlorofluoromethane	3.489	101	85550	20.531	ug/l	99
8) Diethyl Ether	3.959	74	31256	20.158	ug/l	85
9) 1,1,2-Trichlorotrifluo...	4.365	101	46683	20.321	ug/l	99
10) Methyl Iodide	4.589	142	59865	19.810	ug/l	96
11) Tert butyl alcohol	5.530	59	62317	98.946	ug/l	97
12) 1,1-Dichloroethene	4.342	96	45786	19.399	ug/l	89
13) Acrolein	4.183	56	35516	86.524	ug/l	97
14) Allyl chloride	5.018	41	90544	20.301	ug/l	87
15) Acrylonitrile	5.718	53	129316	99.835	ug/l	99
16) Acetone	4.430	43	110676	93.330	ug/l	95
17) Carbon Disulfide	4.712	76	135927	19.683	ug/l	97
18) Methyl Acetate	5.030	43	65871	18.644	ug/l #	91
19) Methyl tert-butyl Ether	5.800	73	172740	20.277	ug/l	97
20) Methylene Chloride	5.271	84	52716	19.316	ug/l	84
21) trans-1,2-Dichloroethene	5.789	96	49121	20.137	ug/l	94
22) Diisopropyl ether	6.671	45	168267	20.070	ug/l #	95
23) Vinyl Acetate	6.600	43	871215m	101.401	ug/l	
24) 1,1-Dichloroethane	6.571	63	93066	20.366	ug/l	97
25) 2-Butanone	7.488	43	180458	99.107	ug/l	93
26) 2,2-Dichloropropane	7.488	77	86767	20.445	ug/l	99
27) cis-1,2-Dichloroethene	7.488	96	58852	19.993	ug/l	92
28) Bromochloromethane	7.812	49	35768	19.153	ug/l	86
29) Tetrahydrofuran	7.841	42	115821	98.399	ug/l #	86
30) Chloroform	7.965	83	94560	19.917	ug/l	98
31) Cyclohexane	8.253	56	85978	19.144	ug/l	96
32) 1,1,1-Trichloroethane	8.165	97	90748	20.194	ug/l	91
36) 1,1-Dichloropropene	8.377	75	68425	19.978	ug/l	98
37) Ethyl Acetate	7.565	43	72806	18.985	ug/l #	93
38) Carbon Tetrachloride	8.359	117	79487	20.605	ug/l	98
39) Methylcyclohexane	9.600	83	85173	20.246	ug/l	95
40) Benzene	8.606	78	207581	20.345	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080724\
 Data File : VN083137.D
 Acq On : 07 Aug 2024 11:22
 Operator : JC\MD
 Sample : VSTDICC020
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC020

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024

Quant Time: Aug 08 06:16:37 2024

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M

Quant Title : SW846 8260

QLast Update : Thu Aug 08 05:50:26 2024

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.777	41	43539	19.968	ug/l	97
42) 1,2-Dichloroethane	8.671	62	74567	20.062	ug/l	97
43) Isopropyl Acetate	8.688	43	138715	19.459	ug/l #	95
44) Trichloroethene	9.353	130	49712	20.469	ug/l	98
45) 1,2-Dichloropropane	9.624	63	49878	20.594	ug/l	100
46) Dibromomethane	9.712	93	35560	20.512	ug/l	98
47) Bromodichloromethane	9.888	83	78421	20.147	ug/l	99
48) Methyl methacrylate	9.682	41	61044	19.254	ug/l	92
49) 1,4-Dioxane	9.694	88	23083	403.558	ug/l	98
51) 4-Methyl-2-Pentanone	10.447	43	374332	103.224	ug/l	93
52) Toluene	10.629	92	131247	20.358	ug/l	99
53) t-1,3-Dichloropropene	10.835	75	83139	20.793	ug/l	97
54) cis-1,3-Dichloropropene	10.312	75	86551	20.351	ug/l	92
55) 1,1,2-Trichloroethane	11.018	97	48081	20.811	ug/l	94
56) Ethyl methacrylate	10.877	69	89503	20.550	ug/l	88
57) 1,3-Dichloropropane	11.165	76	84272	20.464	ug/l	99
58) 2-Chloroethyl Vinyl ether	10.159	63	175345	95.242	ug/l	96
59) 2-Hexanone	11.194	43	286222	102.020	ug/l	92
60) Dibromochloromethane	11.359	129	57571	20.610	ug/l	98
61) 1,2-Dibromoethane	11.471	107	48753	20.088	ug/l	98
64) Tetrachloroethene	11.106	164	42332	20.405	ug/l	94
65) Chlorobenzene	11.894	112	137984	19.933	ug/l	100
66) 1,1,1,2-Tetrachloroethane	11.959	131	48247	19.762	ug/l	99
67) Ethyl Benzene	11.965	91	258237	20.335	ug/l	98
68) m/p-Xylenes	12.071	106	193256	40.625	ug/l	99
69) o-Xylene	12.400	106	95153	20.281	ug/l	99
70) Styrene	12.412	104	161342	20.475	ug/l	98
71) Bromoform	12.576	173	38761	20.957	ug/l #	100
73) Isopropylbenzene	12.694	105	246711	19.764	ug/l	100
74) N-amyl acetate	12.494	43	116100	19.014	ug/l	94
75) 1,1,2,2-Tetrachloroethane	12.935	83	69449	19.670	ug/l	99
76) 1,2,3-Trichloropropane	12.994	75	65295m	19.901	ug/l	
77) Bromobenzene	12.982	156	55327	19.952	ug/l	94
78) n-propylbenzene	13.035	91	283745	19.740	ug/l	99
79) 2-Chlorotoluene	13.123	91	180682	19.827	ug/l	97
80) 1,3,5-Trimethylbenzene	13.170	105	208367	19.938	ug/l	98
81) trans-1,4-Dichloro-2-b...	12.741	75	31260	20.759	ug/l #	83
82) 4-Chlorotoluene	13.223	91	179932	19.684	ug/l	98
83) tert-Butylbenzene	13.441	119	184965	19.985	ug/l	99
84) 1,2,4-Trimethylbenzene	13.482	105	210474	19.984	ug/l	100
85) sec-Butylbenzene	13.618	105	250829	19.862	ug/l	100
86) p-Isopropyltoluene	13.729	119	208404	19.986	ug/l	99
87) 1,3-Dichlorobenzene	13.735	146	103819	19.897	ug/l	98
88) 1,4-Dichlorobenzene	13.812	146	103132	19.605	ug/l	97
89) n-Butylbenzene	14.053	91	180483	19.975	ug/l	97
90) Hexachloroethane	14.335	117	40982	20.343	ug/l	94
91) 1,2-Dichlorobenzene	14.106	146	99334	19.673	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	14.723	75	16552	19.320	ug/l	98
93) 1,2,4-Trichlorobenzene	15.394	180	56299	19.902	ug/l	98
94) Hexachlorobutadiene	15.500	225	23725	18.859	ug/l	99
95) Naphthalene	15.641	128	200875	20.047	ug/l	99
96) 1,2,3-Trichlorobenzene	15.841	180	54944	19.630	ug/l	96

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080724\
Data File : VN083137.D
Acq On : 07 Aug 2024 11:22
Operator : JC\MD
Sample : VSTDICC020
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Manual Integrations
APPROVED

Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024

Quant Time: Aug 08 06:16:37 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 05:50:26 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080724\
Data File : VN083137.D
Acq On : 07 Aug 2024 11:22
Operator : JC\MD
Sample : VSTDIC020
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Quant Time: Aug 08 06:16:37 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 05:50:26 2024
Response via : Initial Calibration

Instrument :

MSVOA_N

Client Sample Id :


VSTDIC020


Manual Integrations

APPROVED

Reviewed By : John Carlone 08/08/2024


Supervised By : Mahesh Dadoda 08/09/2024





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 1066
Delta R.T. -0.000 min
Lab File: VN083137.D
Acq: 07 Aug 2024 11:22




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020



Tgt Ion:168 Resp: 21291
Ion Ratio Lower Upper
168 100
99 67.5 48.2 72.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#2
Dichlorodifluoromethane
Concen: 19.959 ug/l
RT: 2.124 min Scan# 29
Delta R.T. 0.000 min
Lab File: VN083137.D
Acq: 07 Aug 2024 11:22


Tgt Ion: 85 Resp: 48191
Ion Ratio Lower Upper
85 100
87 34.0 15.7 47.0





#3
Chloromethane
Concen: 20.137 ug/l
RT: 2.359 min Scan# 69
Delta R.T. 0.000 min
Lab File: VN083137.D
Acq: 07 Aug 2024 11:22




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020



Tgt Ion: 50 Resp: 4977
Ion Ratio Lower Upper
50 100
52 31.9 25.0 37.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#4
Vinyl Chloride
Concen: 19.866 ug/l
RT: 2.512 min Scan# 95
Delta R.T. 0.000 min
Lab File: VN083137.D
Acq: 07 Aug 2024 11:22


Tgt Ion: 62 Resp: 50108
Ion Ratio Lower Upper
62 100
64 33.1 25.0 37.6





#5
Bromomethane
Concen: 20.143 ug/l
RT: 2.942 min Scan# 168
Delta R.T. -0.006 min
Lab File: VN083137.D
Acq: 07 Aug 2024 11:22




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020



Tgt Ion: 94 Resp: 3152
Ion Ratio Lower Upper
94 100
96 87.1 78.0 117.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#6
Chloroethane
Concen: 19.221 ug/l
RT: 3.112 min Scan# 197
Delta R.T. 0.000 min
Lab File: VN083137.D
Acq: 07 Aug 2024 11:22

Tgt Ion: 64 Resp: 30331
Ion Ratio Lower Upper
64 100
66 33.0 26.6 40.0









#7
Trichlorofluoromethane
Concen: 20.531 ug/l
RT: 3.489 min Scan# 262
Delta R.T. -0.006 min
Lab File: VN083137.D
Acq: 07 Aug 2024 11:22

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Tgt Ion:101 Resp: 85550
Ion Ratio Lower Upper
101 100
103 65.0 52.7 79.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#8
Diethyl Ether
Concen: 20.158 ug/l
RT: 3.959 min Scan# 341
Delta R.T. 0.000 min
Lab File: VN083137.D
Acq: 07 Aug 2024 11:22

Tgt Ion: 74 Resp: 31256
Ion Ratio Lower Upper
74 100
45 94.8 55.5 166.3









#9
1,1,2-Trichlorotrifluoroethane
Concen: 20.321 ug/l
RT: 4.365 min Scan# 410
Delta R.T. 0.000 min
Lab File: VN083137.D
Acq: 07 Aug 2024 11:22

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Tgt Ion:101 Resp: 4668
Ion Ratio Lower Upper
101 100
85 46.5 38.3 57.5
151 74.6 59.3 88.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#10
Methyl Iodide
Concen: 19.810 ug/l
RT: 4.589 min Scan# 448
Delta R.T. 0.000 min
Lab File: VN083137.D
Acq: 07 Aug 2024 11:22

Tgt Ion:142 Resp: 59865
Ion Ratio Lower Upper
142 100
127 50.3 37.5 56.3
141 15.4 13.1 19.7









#11
Tert butyl alcohol
Concen: 98.946 ug/l
RT: 5.530 min Scan# 608
Delta R.T. 0.000 min
Lab File: VN083137.D
Acq: 07 Aug 2024 11:22

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Tgt Ion: 59 Resp: 6231
Ion Ratio Lower Upper
59 100
57 12.2 8.9 13.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#12
1,1-Dichloroethene
Concen: 19.399 ug/l
RT: 4.342 min Scan# 406
Delta R.T. 0.000 min
Lab File: VN083137.D
Acq: 07 Aug 2024 11:22

Tgt Ion: 96 Resp: 45786
Ion Ratio Lower Upper
96 100
61 167.8 149.7 224.5
98 64.7 50.1 75.1





#13

Acrolein

Concen: 86.524 ug/l

RT: 4.183 min Scan# 379

Delta R.T. 0.000 min

Lab File: VN083137.D

Acq: 07 Aug 2024 11:22

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC020

Tgt Ion: 56 Resp: 35510

Ion Ratio Lower Upper

56 100





55 72.6 56.4 84.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#14

Allyl chloride

Concen: 20.301 ug/l

RT: 5.018 min Scan# 521

Delta R.T. 0.000 min

Lab File: VN083137.D

Acq: 07 Aug 2024 11:22




Tgt Ion: 41 Resp: 90544


Ion Ratio Lower Upper

41 100

39 69.0 67.8 101.6

76 32.7 25.8 38.8





#15

Acrylonitrile

Concen: 99.835 ug/l

RT: 5.718 min Scan# 640

Delta R.T. 0.000 min

Lab File: VN083137.D

Acq: 07 Aug 2024 11:22

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC020

Tgt Ion: 53 Resp: 129310

Ion Ratio Lower Upper

53 100

52 82.1 66.6 100.0





51 37.7 30.1 45.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#16

Acetone

Concen: 93.330 ug/l

RT: 4.430 min Scan# 421

Delta R.T. 0.000 min

Lab File: VN083137.D




Acq: 07 Aug 2024 11:22


Tgt Ion: 43 Resp: 110676

Ion Ratio Lower Upper

43 100


58 29.8 21.8 32.6





#17
Carbon Disulfide
Concen: 19.683 ug/l
RT: 4.712 min Scan# 469
Delta R.T. 0.000 min
Lab File: VN083137.D
Acq: 07 Aug 2024 11:22




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020



Tgt Ion: 76 Resp: 135927
Ion Ratio Lower Upper
76 100
78 9.9 7.2 10.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#18
Methyl Acetate
Concen: 18.644 ug/l
RT: 5.030 min Scan# 523
Delta R.T. 0.006 min
Lab File: VN083137.D
Acq: 07 Aug 2024 11:22

Tgt Ion: 43 Resp: 65871
Ion Ratio Lower Upper
43 100
74 24.5 16.3 24.5









#19
Methyl tert-butyl Ether
Concen: 20.277 ug/l
RT: 5.800 min Scan# 653
Delta R.T. 0.006 min
Lab File: VN083137.D
Acq: 07 Aug 2024 11:22

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020

Tgt Ion: 73 Resp: 172740
Ion Ratio Lower Upper
73 100
57 21.0 17.9 26.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#20
Methylene Chloride
Concen: 19.316 ug/l
RT: 5.271 min Scan# 564
Delta R.T. -0.006 min
Lab File: VN083137.D
Acq: 07 Aug 2024 11:22

Tgt Ion: 84 Resp: 52716
Ion Ratio Lower Upper
84 100
49 120.5 119.6 179.4
51 36.5 34.8 52.2
86 62.2 52.9 79.3









#21
trans-1,2-Dichloroethene
Concen: 20.137 ug/l
RT: 5.789 min Scan# 61
Delta R.T. -0.006 min
Lab File: VN083137.D
Acq: 07 Aug 2024 11:22

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Tgt Ion: 96 Resp: 4912
Ion Ratio Lower Upper
96 100
61 143.3 123.4 185.2
98 62.4 49.3 73.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#22
Diisopropyl ether
Concen: 20.070 ug/l
RT: 6.671 min Scan# 802
Delta R.T. 0.000 min
Lab File: VN083137.D
Acq: 07 Aug 2024 11:22

Tgt Ion: 45 Resp: 168267
Ion Ratio Lower Upper
45 100
43 57.6 44.0 66.0
87 29.6 19.7 29.5#
59 12.2 8.5 12.7









#23
Vinyl Acetate
Concen: 101.401 ug/l m
RT: 6.600 min Scan# 791
Delta R.T. -0.006 min
Lab File: VN083137.D
Acq: 07 Aug 2024 11:22

Instrument :
MSVOA_N
Client Sample Id :
VSTDICC020

Tgt Ion: 43 Resp: 871215
Ion Ratio Lower Upper
43 100
86 10.4 7.3 10.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#24
1,1-Dichloroethane
Concen: 20.366 ug/l
RT: 6.571 min Scan# 785
Delta R.T. 0.000 min
Lab File: VN083137.D
Acq: 07 Aug 2024 11:22

Tgt Ion: 63 Resp: 93066
Ion Ratio Lower Upper
63 100
98 7.6 3.3 9.9
100 5.0 2.0 6.0





#25

2-Butanone

Concen: 99.107 ug/l

RT: 7.488 min Scan# 941

Delta R.T. 0.006 min

Lab File: VN083137.D

Acq: 07 Aug 2024 11:22

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC020

Tgt Ion: 43 Resp: 180458

Ion Ratio Lower Upper

43 100





72 25.9 18.2 27.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#26

2,2-Dichloropropane

Concen: 20.445 ug/l

RT: 7.488 min Scan# 941

Delta R.T. 0.000 min

Lab File: VN083137.D




Acq: 07 Aug 2024 11:22


Tgt Ion: 77 Resp: 86767

Ion Ratio Lower Upper

77 100

97 21.3 10.3 30.9









#27
cis-1,2-Dichloroethene
Concen: 19.993 ug/l
RT: 7.488 min Scan# 941
Delta R.T. 0.000 min
Lab File: VN083137.D
Acq: 07 Aug 2024 11:22

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Tgt Ion: 96 Resp: 5885
Ion Ratio Lower Upper
96 100
61 149.6 0.0 329.6
98 63.7 0.0 128.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#28
Bromochloromethane
Concen: 19.153 ug/l
RT: 7.812 min Scan# 996
Delta R.T. 0.000 min
Lab File: VN083137.D
Acq: 07 Aug 2024 11:22

Tgt Ion: 49 Resp: 35768
Ion Ratio Lower Upper
49 100
129 1.6 0.0 3.8
130 74.4 50.5 75.7





#29

Tetrahydrofuran

Concen: 98.399 ug/l

RT: 7.841 min Scan# 1000

Delta R.T. 0.006 min

Lab File: VN083137.D

Acq: 07 Aug 2024 11:22

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC020

Tgt Ion: 42 Resp: 11582

Ion Ratio Lower Upper

42 100

72 45.8 30.0 45.0





71 44.0 28.3 42.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#30

Chloroform

Concen: 19.917 ug/l

RT: 7.965 min Scan# 1022

Delta R.T. 0.000 min

Lab File: VN083137.D




Acq: 07 Aug 2024 11:22


Tgt Ion: 83 Resp: 94560

Ion Ratio Lower Upper

83 100

85 64.9 50.9 76.3





#31

Cyclohexane

Concen: 19.144 ug/l

RT: 8.253 min Scan# 1071

Delta R.T. -0.006 min

Lab File: VN083137.D

Acq: 07 Aug 2024 11:22

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020

Tgt Ion: 56 Resp: 85978

Ion Ratio Lower Upper

56 100

69 30.4 24.5 36.7





84 87.9 66.0 99.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#32

1,1,1-Trichloroethane

Concen: 20.194 ug/l

RT: 8.165 min Scan# 1056

Delta R.T. -0.006 min

Lab File: VN083137.D

Acq: 07 Aug 2024 11:22




Tgt Ion: 97 Resp: 90748


Ion Ratio Lower Upper

97 100

99 58.9 52.0 78.0

61 45.3 42.1 63.1





#33

1,2-Dichloroethane-d4

Concen: 16.881 ug/l

RT: 8.583 min Scan# 1127

Delta R.T. 0.000 min

Lab File: VN083137.D

Acq: 07 Aug 2024 11:22

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020

Tgt Ion: 65 Resp: 5115

Ion Ratio Lower Upper

65 100





67 51.1 0.0 104.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.100 min Scan# 1215

Delta R.T. 0.000 min

Lab File: VN083137.D

Acq: 07 Aug 2024 11:22




Tgt Ion:114 Resp: 362741


Ion Ratio Lower Upper

114 100

63 22.0 0.0 44.6

88 17.6 0.0 31.4





#35

Dibromofluoromethane

Concen: 16.940 ug/l

RT: 8.171 min Scan# 1056

Delta R.T. 0.006 min

Lab File: VN083137.D

Acq: 07 Aug 2024 11:22

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC020





Tgt Ion:	113	Resp:	38354
Ion	Ratio	Lower	Upper
113	100		
111	101.3	82.4	123.6
192	17.8	14.9	22.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#36

1,1-Dichloropropene

Concen: 19.978 ug/l




RT: 8.377 min Scan# 1092


Delta R.T. 0.006 min

Lab File: VN083137.D

Acq: 07 Aug 2024 11:22

Tgt Ion:	75	Resp:	68425
Ion	Ratio	Lower	Upper
75	100		
110	34.3	16.3	48.9
77	31.3	24.6	37.0





#37

Ethyl Acetate

Concen: 18.985 ug/l

RT: 7.565 min Scan# 91

Delta R.T. 0.006 min

Lab File: VN083137.D

Acq: 07 Aug 2024 11:22

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020

Tgt Ion: 43 Resp: 72800

Ion Ratio Lower Upper

43 100

61 14.3 9.3 13.9





70 11.2 7.0 10.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#38

Carbon Tetrachloride

Concen: 20.605 ug/l

RT: 8.359 min Scan# 1089

Delta R.T. 0.000 min

Lab File: VN083137.D

Acq: 07 Aug 2024 11:22




Tgt Ion:117 Resp: 79487


Ion Ratio Lower Upper

117 100

119 95.1 74.9 112.3

121 28.3 24.3 36.5





#39

Methylcyclohexane

Concen: 20.246 ug/l

RT: 9.600 min Scan# 1130

Delta R.T. 0.000 min

Lab File: VN083137.D

Acq: 07 Aug 2024 11:22

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC020





Tgt Ion:	83	Resp:	8517
Ion	Ratio	Lower	Upper
83	100		
55	79.7	66.4	99.6
98	46.1	40.7	61.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#40

Benzene

Concen: 20.345 ug/l




RT: 8.606 min Scan# 1131


Delta R.T. 0.000 min

Lab File: VN083137.D

Acq: 07 Aug 2024 11:22


Tgt Ion:	78	Resp:	207581
Ion	Ratio	Lower	Upper
78	100		
77	23.1	19.0	28.4





#41
Methacrylonitrile
Concen: 19.968 ug/l
RT: 7.777 min Scan# 990
Delta R.T. 0.000 min
Lab File: VN083137.D
Acq: 07 Aug 2024 11:22




Instrument :
MSVOA_N
Client Sample Id :
VSTDICC020



Tgt Ion: 41 Resp: 43539
Ion Ratio Lower Upper
41 100
39 54.6 45.5 68.3
67 65.2 50.5 75.7
52 28.6 24.0 36.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#42
1,2-Dichloroethane
Concen: 20.062 ug/l
RT: 8.671 min Scan# 1142
Delta R.T. 0.000 min
Lab File: VN083137.D
Acq: 07 Aug 2024 11:22

Tgt Ion: 62 Resp: 74567
Ion Ratio Lower Upper
62 100
98 9.0 0.0 15.8





#43

Isopropyl Acetate

Concen: 19.459 ug/l

RT: 8.688 min Scan# 1145

Delta R.T. 0.000 min

Lab File: VN083137.D


Acq: 07 Aug 2024 11:22

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020






Tgt Ion:	43	Resp:	13871
Ion Ratio	100	Lower	Upper
43	100		
61	24.5	17.8	26.6
87	12.7	8.2	12.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#44

Trichloroethene

Concen: 20.469 ug/l




RT: 9.353 min Scan# 1258


Delta R.T. 0.000 min

Lab File: VN083137.D

Acq: 07 Aug 2024 11:22

Tgt Ion:	130	Resp:	49712
Ion Ratio	100	Lower	Upper
130	100		
95	101.0	0.0	197.8





#45

1,2-Dichloropropane

Concen: 20.594 ug/l

RT: 9.624 min Scan# 1304

Delta R.T. 0.000 min

Lab File: VN083137.D

Acq: 07 Aug 2024 11:22

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC020

Tgt Ion: 63 Resp: 49878

Ion Ratio Lower Upper

63 100





65 30.4 24.4 36.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#46

Dibromomethane

Concen: 20.512 ug/l

RT: 9.712 min Scan# 1319

Delta R.T. 0.000 min

Lab File: VN083137.D

Acq: 07 Aug 2024 11:22




Tgt Ion: 93 Resp: 35560


Ion Ratio Lower Upper

93 100

95 81.8 65.8 98.6

174 92.9 71.7 107.5





#47

Bromodichloromethane

Concen: 20.147 ug/l

RT: 9.888 min Scan# 1349

Delta R.T. 0.000 min

Lab File: VN083137.D

Acq: 07 Aug 2024 11:22

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020





Tgt Ion:	83	Resp:	7842
Ion Ratio	Lower	Upper	
83	100		
85	63.7	51.1	76.7
127	8.1	7.2	10.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#48

Methyl methacrylate

Concen: 19.254 ug/l




RT: 9.682 min Scan# 1314


Delta R.T. 0.000 min

Lab File: VN083137.D

Acq: 07 Aug 2024 11:22

Tgt Ion:	41	Resp:	61044
Ion Ratio	Lower	Upper	
41	100		
69	82.1	56.3	84.5
39	62.5	50.3	75.5





#49

1,4-Dioxane

Concen: 403.558 ug/l

RT: 9.694 min Scan# 1316

Delta R.T. 0.000 min

Lab File: VN083137.D

Acq: 07 Aug 2024 11:22

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020

Tgt Ion: 88 Resp: 2308

Ion Ratio Lower Upper

88 100

43 32.7 27.8 41.8





58 75.6 59.4 89.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#50

Toluene-d8

Concen: 16.926 ug/l

RT: 10.565 min Scan# 1464

Delta R.T. 0.000 min

Lab File: VN083137.D




Acq: 07 Aug 2024 11:22


Tgt Ion: 98 Resp: 142952

Ion Ratio Lower Upper

98 100


100 65.8 51.5 77.3





#51
4-Methyl-2-Pentanone
Concen: 103.224 ug/l
RT: 10.447 min Scan# 1444
Delta R.T. 0.000 min
Lab File: VN083137.D
Acq: 07 Aug 2024 11:22




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020



Tgt Ion: 43 Resp: 37433
Ion Ratio Lower Upper
43 100
58 39.4 28.2 42.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#52
Toluene
Concen: 20.358 ug/l
RT: 10.629 min Scan# 1475
Delta R.T. 0.000 min
Lab File: VN083137.D
Acq: 07 Aug 2024 11:22


Tgt Ion: 92 Resp: 131247
Ion Ratio Lower Upper
92 100
91 173.0 139.4 209.0





#53
t-1,3-Dichloropropene
Concen: 20.793 ug/l
RT: 10.835 min Scan# 1510
Delta R.T. 0.000 min
Lab File: VN083137.D
Acq: 07 Aug 2024 11:22




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020



Tgt Ion: 75 Resp: 83139
Ion Ratio Lower Upper
75 100
77 30.4 25.7 38.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#54
cis-1,3-Dichloropropene
Concen: 20.351 ug/l
RT: 10.312 min Scan# 1421
Delta R.T. 0.000 min
Lab File: VN083137.D
Acq: 07 Aug 2024 11:22


Tgt Ion: 75 Resp: 86551
Ion Ratio Lower Upper
75 100
77 31.7 24.3 36.5
39 54.5 50.5 75.7





#55
1,1,2-Trichloroethane
Concen: 20.811 ug/l
RT: 11.018 min Scan# 1541
Delta R.T. 0.000 min
Lab File: VN083137.D
Acq: 07 Aug 2024 11:22




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020



Tgt Ion: 97 Resp: 4808
Ion Ratio Lower Upper
97 100
83 81.1 71.4 107.0
85 56.0 45.8 68.6
99 59.2 52.0 78.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#56
Ethyl methacrylate
Concen: 20.550 ug/l
RT: 10.877 min Scan# 1517
Delta R.T. 0.000 min
Lab File: VN083137.D
Acq: 07 Aug 2024 11:22

Tgt Ion: 69 Resp: 89503
Ion Ratio Lower Upper
69 100
41 67.5 63.4 95.2
39 41.2 37.4 56.0





#57

1,3-Dichloropropane

Concen: 20.464 ug/l

RT: 11.165 min Scan# 1566

Delta R.T. 0.000 min

Lab File: VN083137.D

Acq: 07 Aug 2024 11:22

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020

Tgt Ion: 76 Resp: 8427

Ion Ratio Lower Upper

76 100





78 32.5 25.6 38.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#58

2-Chloroethyl Vinyl ether

Concen: 95.242 ug/l

RT: 10.159 min Scan# 1395

Delta R.T. 0.000 min

Lab File: VN083137.D




Acq: 07 Aug 2024 11:22


Tgt Ion: 63 Resp: 175345

Ion Ratio Lower Upper

63 100

106 28.9 21.4 32.0





#59

2-Hexanone

Concen: 102.020 ug/l

RT: 11.194 min Scan# 1571

Delta R.T. 0.000 min

Lab File: VN083137.D

Acq: 07 Aug 2024 11:22

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020

Tgt Ion: 43 Resp: 28622

Ion Ratio Lower Upper

43 100





58 54.7 24.6 73.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#60

Dibromochloromethane

Concen: 20.610 ug/l

RT: 11.359 min Scan# 1599

Delta R.T. 0.000 min

Lab File: VN083137.D




Acq: 07 Aug 2024 11:22


Tgt Ion:129 Resp: 57571

Ion Ratio Lower Upper

129 100


127 79.9 39.2 117.6





#61
1,2-Dibromoethane
Concen: 20.088 ug/l
RT: 11.471 min Scan# 1618
Delta R.T. 0.000 min
Lab File: VN083137.D
Acq: 07 Aug 2024 11:22




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020



Tgt Ion:107 Resp: 4875
Ion Ratio Lower Upper
107 100
109 94.4 73.7 110.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#62
4-Bromofluorobenzene
Concen: 17.179 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. 0.000 min
Lab File: VN083137.D
Acq: 07 Aug 2024 11:22


Tgt Ion: 95 Resp: 56566
Ion Ratio Lower Upper
95 100
174 69.8 0.0 159.2
176 66.2 0.0 147.6





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 1685
Delta R.T. 0.000 min
Lab File: VN083137.D
Acq: 07 Aug 2024 11:22




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020



Tgt Ion:117 Resp: 31325
Ion Ratio Lower Upper
117 100
82 61.1 47.5 71.3
119 33.4 25.6 38.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#64
Tetrachloroethene
Concen: 20.405 ug/l
RT: 11.106 min Scan# 1556
Delta R.T. 0.000 min
Lab File: VN083137.D
Acq: 07 Aug 2024 11:22


Tgt Ion:164 Resp: 42332
Ion Ratio Lower Upper
164 100
166 129.4 101.4 152.0
129 103.2 75.7 113.5
131 98.7 73.8 110.6





#65
Chlorobenzene
Concen: 19.933 ug/l
RT: 11.894 min Scan# 11894
Delta R.T. 0.000 min
Lab File: VN083137.D
Acq: 07 Aug 2024 11:22




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020



Tgt Ion:112 Resp: 137984
Ion Ratio Lower Upper
112 100
114 32.7 26.2 39.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#66
1,1,1,2-Tetrachloroethane
Concen: 19.762 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. -0.006 min
Lab File: VN083137.D
Acq: 07 Aug 2024 11:22

Tgt Ion:131 Resp: 48247
Ion Ratio Lower Upper
131 100
133 96.1 47.3 142.0
119 64.5 32.5 97.4





#67

Ethyl Benzene

Concen: 20.335 ug/l

RT: 11.965 min Scan# 1702

Delta R.T. 0.000 min

Lab File: VN083137.D

Acq: 07 Aug 2024 11:22

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020

Tgt Ion: 91 Resp: 25823

Ion Ratio Lower Upper

91 100





106 31.2 24.0 36.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#68

m/p-Xylenes

Concen: 40.625 ug/l

RT: 12.071 min Scan# 1720

Delta R.T. 0.000 min

Lab File: VN083137.D




Acq: 07 Aug 2024 11:22


Tgt Ion:106 Resp: 193256

Ion Ratio Lower Upper

106 100

91 209.9 166.1 249.1





#69

o-Xylene

Concen: 20.281 ug/l

RT: 12.400 min Scan# 1776

Delta R.T. 0.000 min

Lab File: VN083137.D

Acq: 07 Aug 2024 11:22

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020

Tgt Ion:106 Resp: 9515

Ion Ratio Lower Upper

106 100





91 221.3 111.3 333.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#70

Styrene

Concen: 20.475 ug/l

RT: 12.412 min Scan# 1778

Delta R.T. 0.000 min

Lab File: VN083137.D

Acq: 07 Aug 2024 11:22




Tgt Ion:104 Resp: 161342


Ion Ratio Lower Upper

104 100

78 53.7 41.6 62.4


103 53.8 44.0 66.0





#71
Bromoform
Concen: 20.957 ug/l
RT: 12.576 min Scan# 1806
Delta R.T. 0.000 min
Lab File: VN083137.D
Acq: 07 Aug 2024 11:22




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020




Tgt Ion:173 Resp: 3876
Ion Ratio Lower Upper
173 100
175 47.9 24.0 72.0
254 0.0 0.0 0.0

Manual Integrations
APPROVED



Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.794 min Scan# 2013
Delta R.T. 0.000 min
Lab File: VN083137.D
Acq: 07 Aug 2024 11:22




Tgt Ion:152 Resp: 149244
Ion Ratio Lower Upper
152 100
115 63.1 30.6 91.6
150 162.7 0.0 348.6





#73
Isopropylbenzene
Concen: 19.764 ug/l
RT: 12.694 min Scan# 1826
Delta R.T. 0.000 min
Lab File: VN083137.D
Acq: 07 Aug 2024 11:22




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020



Tgt Ion: 105 Resp: 24671
Ion Ratio Lower Upper
105 100
120 25.8 12.9 38.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#74
N-ethyl acetate
Concen: 19.014 ug/l
RT: 12.494 min Scan# 1792
Delta R.T. 0.000 min
Lab File: VN083137.D
Acq: 07 Aug 2024 11:22


Tgt Ion: 43 Resp: 116100
Ion Ratio Lower Upper
43 100
70 42.8 29.8 44.6
55 24.0 18.7 28.1
61 24.8 17.4 26.0





#75
1,1,2-Tetrachloroethane
Concen: 19.670 ug/l
RT: 12.935 min Scan# 1867
Delta R.T. 0.000 min
Lab File: VN083137.D
Acq: 07 Aug 2024 11:22




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020



Tgt Ion: 83 Resp: 69449
Ion Ratio Lower Upper
83 100
131 10.6 5.0 14.9
85 64.6 31.9 95.7




Manual Integrations
APPROVED


Reviewed By :John Carbone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#76
1,2,3-Trichloropropane
Concen: 19.901 ug/l m
RT: 12.994 min Scan# 1877
Delta R.T. 0.000 min
Lab File: VN083137.D
Acq: 07 Aug 2024 11:22

Tgt Ion: 75 Resp: 65295
Ion Ratio Lower Upper
75 100
77 199.0 110.9 332.6









#77
Bromobenzene
Concen: 19.952 ug/l
RT: 12.982 min Scan# 1875
Delta R.T. 0.000 min
Lab File: VN083137.D
Acq: 07 Aug 2024 11:22

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020

Tgt Ion:156 Resp: 5532
Ion Ratio Lower Upper
156 100
77 234.8 111.3 333.8
158 95.8 48.3 144.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#78
n-propylbenzene
Concen: 19.740 ug/l
RT: 13.035 min Scan# 1884
Delta R.T. 0.000 min
Lab File: VN083137.D
Acq: 07 Aug 2024 11:22


Tgt Ion: 91 Resp: 283745
Ion Ratio Lower Upper
91 100
120 22.4 10.9 32.9





#79
2-Chlorotoluene
Concen: 19.827 ug/l
RT: 13.123 min Scan# 1899
Delta R.T. 0.000 min
Lab File: VN083137.D
Acq: 07 Aug 2024 11:22




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020




Tgt Ion: 91 Resp: 180682
Ion Ratio Lower Upper
91 100
126 30.7 16.2 48.6

Manual Integrations
APPROVED



Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024




#80
1,3,5-Trimethylbenzene
Concen: 19.938 ug/l
RT: 13.170 min Scan# 1907
Delta R.T. 0.000 min
Lab File: VN083137.D
Acq: 07 Aug 2024 11:22




Tgt Ion:105 Resp: 208367
Ion Ratio Lower Upper
105 100
120 47.2 24.3 72.8





#81
trans-1,4-Dichloro-2-butene
Concen: 20.759 ug/l
RT: 12.741 min Scan# 1834
Delta R.T. 0.006 min
Lab File: VN083137.D
Acq: 07 Aug 2024 11:22




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020



Tgt Ion: 75 Resp: 31260
Ion Ratio Lower Upper
75 100
53 88.0 88.4 132.6
89 41.4 36.3 54.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#82
4-Chlorotoluene
Concen: 19.684 ug/l
RT: 13.223 min Scan# 1916
Delta R.T. 0.000 min
Lab File: VN083137.D
Acq: 07 Aug 2024 11:22


Tgt Ion: 91 Resp: 179932
Ion Ratio Lower Upper
91 100
126 31.3 16.2 48.6





#83
tert-Butylbenzene
Concen: 19.985 ug/l
RT: 13.441 min Scan# 1953
Delta R.T. 0.000 min
Lab File: VN083137.D
Acq: 07 Aug 2024 11:22




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020



Tgt Ion:119 Resp: 18496
Ion Ratio Lower Upper
119 100
91 70.2 35.0 105.1
134 25.2 13.3 39.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#84
1,2,4-Trimethylbenzene
Concen: 19.984 ug/l
RT: 13.482 min Scan# 1960
Delta R.T. 0.000 min
Lab File: VN083137.D
Acq: 07 Aug 2024 11:22

Tgt Ion:105 Resp: 210474
Ion Ratio Lower Upper
105 100
120 44.0 21.9 65.8





#85

sec-Butylbenzene

Concen: 19.862 ug/l

RT: 13.618 min Scan# 1983

Delta R.T. 0.000 min

Lab File: VN083137.D

Acq: 07 Aug 2024 11:22

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020

Tgt Ion:105 Resp: 250829

Ion Ratio Lower Upper

105 100





134 19.3 9.8 29.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#86

p-Isopropyltoluene

Concen: 19.986 ug/l

RT: 13.729 min Scan# 2002

Delta R.T. 0.000 min

Lab File: VN083137.D

Acq: 07 Aug 2024 11:22




Tgt Ion:119 Resp: 208404


Ion Ratio Lower Upper

119 100

134 25.7 13.0 39.0

91 25.8 12.3 36.9





#87

1,3-Dichlorobenzene

Concen: 19.897 ug/l

RT: 13.735 min Scan# 2003

Delta R.T. 0.000 min

Lab File: VN083137.D

Acq: 07 Aug 2024 11:22

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020

Tgt Ion:146 Resp: 103819

Ion Ratio Lower Upper

146 100

111 43.7 20.6 61.9





148 62.7 31.8 95.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#88

1,4-Dichlorobenzene

Concen: 19.605 ug/l

RT: 13.812 min Scan# 2016

Delta R.T. 0.000 min

Lab File: VN083137.D

Acq: 07 Aug 2024 11:22




Tgt Ion:146 Resp: 103132


Ion Ratio Lower Upper

146 100

111 43.2 20.1 60.3

148 63.5 32.2 96.6





#89

n-Butylbenzene

Concen: 19.975 ug/l

RT: 14.053 min Scan# 2058

Delta R.T. -0.006 min

Lab File: VN083137.D

Acq: 07 Aug 2024 11:22

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020

Tgt Ion: 91 Resp: 18048

Ion Ratio Lower Upper

91 100

92 54.3 26.1 78.3





134 25.4 12.1 36.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#90

Hexachloroethane

Concen: 20.343 ug/l

RT: 14.335 min Scan# 2105

Delta R.T. 0.000 min

Lab File: VN083137.D




Acq: 07 Aug 2024 11:22


Tgt Ion:117 Resp: 40982

Ion Ratio Lower Upper

117 100


201 66.5 35.8 107.3





#91
1,2-Dichlorobenzene
Concen: 19.673 ug/l
RT: 14.106 min Scan# 2066
Delta R.T. 0.000 min
Lab File: VN083137.D
Acq: 07 Aug 2024 11:22




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020




Tgt Ion:146 Resp: 99334
Ion Ratio Lower Upper
146 100
111 45.3 21.4 64.2
148 63.5 32.1 96.5

Manual Integrations
APPROVED



Reviewed By :John Carbone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024




#92
1,2-Dibromo-3-Chloropropane
Concen: 19.320 ug/l
RT: 14.723 min Scan# 2171
Delta R.T. 0.006 min
Lab File: VN083137.D
Acq: 07 Aug 2024 11:22



Tgt Ion: 75 Resp: 16552
Ion Ratio Lower Upper
75 100
155 75.5 36.6 109.8
157 95.6 46.9 140.6





#93

1,2,4-Trichlorobenzene

Concen: 19.902 ug/l

RT: 15.394 min Scan# 2285

Delta R.T. 0.000 min

Lab File: VN083137.D

Acq: 07 Aug 2024 11:22

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020

Tgt Ion:180 Resp: 56299

Ion Ratio Lower Upper

180 100

182 94.3 47.3 141.8





145 34.5 15.7 47.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#94

Hexachlorobutadiene

Concen: 18.859 ug/l

RT: 15.500 min Scan# 2303

Delta R.T. 0.000 min

Lab File: VN083137.D

Acq: 07 Aug 2024 11:22




Tgt Ion:225 Resp: 23725


Ion Ratio Lower Upper

225 100

223 64.3 31.9 95.7


227 65.7 32.5 97.5





#95
Naphthalene
Concen: 20.047 ug/l
RT: 15.641 min Scan# 2327
Delta R.T. 0.000 min
Lab File: VN083137.D
Acq: 07 Aug 2024 11:22




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020




Tgt Ion:128 Resp: 200879
Ion Ratio Lower Upper
128 100
127 12.8 10.6 16.0
129 10.8 8.6 12.8

Manual Integrations
APPROVED



Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#96
1,2,3-Trichlorobenzene
Concen: 19.630 ug/l
RT: 15.841 min Scan# 2361
Delta R.T. 0.000 min
Lab File: VN083137.D
Acq: 07 Aug 2024 11:22



Tgt Ion:180 Resp: 54944
Ion Ratio Lower Upper
180 100
182 93.5 48.9 146.8
145 35.9 16.8 50.4



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080724\
 Data File : VN083138.D
 Acq On : 07 Aug 2024 11:46
 Operator : JC\MD
 Sample : VSTDIC010
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC010

Manual Integrations
 APPROVED

Reviewed By : John Carlone 08/08/2024
 Supervised By : Mahesh Dadoda 08/09/2024

Quant Time: Aug 08 06:17:39 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 05:50:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	191754	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	332105	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	279409	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	125725	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	27223	9.974	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	19.940%#		
35) Dibromofluoromethane	8.165	113	20382	9.832	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	19.660%#		
50) Toluene-d8	10.565	98	75261	9.733	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	19.460%#		
62) 4-Bromofluorobenzene	12.853	95	28691	9.517	ug/l	0.00
Spiked Amount 50.000	Range 64 - 133		Recovery =	19.040%#		
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	2.124	85	22206	10.212	ug/l	99
3) Chloromethane	2.359	50	22596	10.150	ug/l	92
4) Vinyl Chloride	2.512	62	23461	10.328	ug/l	99
5) Bromomethane	2.959	94	14783	10.488	ug/l	95
6) Chloroethane	3.118	64	13616	9.581	ug/l	95
7) Trichlorofluoromethane	3.500	101	38868	10.357	ug/l	97
8) Diethyl Ether	3.965	74	14761	10.571	ug/l	86
9) 1,1,2-Trichlorotrifluo...	4.365	101	21081	10.189	ug/l	100
10) Methyl Iodide	4.589	142	27229	10.005	ug/l	96
11) Tert butyl alcohol	5.530	59	29665	52.299	ug/l	97
12) 1,1-Dichloroethene	4.336	96	21633	10.177	ug/l	83
13) Acrolein	4.177	56	19397	52.469	ug/l	92
14) Allyl chloride	5.024	41	41546	10.343	ug/l	89
15) Acrylonitrile	5.724	53	58618	50.248	ug/l	99
16) Acetone	4.430	43	54397	50.933	ug/l	92
17) Carbon Disulfide	4.712	76	61323	9.860	ug/l	97
18) Methyl Acetate	5.036	43	32282	10.145	ug/l	95
19) Methyl tert-butyl Ether	5.794	73	77278	10.072	ug/l	94
20) Methylene Chloride	5.277	84	23553	9.582	ug/l #	83
21) trans-1,2-Dichloroethene	5.788	96	22473	10.229	ug/l	93
22) Diisopropyl ether	6.671	45	77046	10.204	ug/l	91
23) Vinyl Acetate	6.606	43	391151	50.550	ug/l #	94
24) 1,1-Dichloroethane	6.571	63	42035	10.213	ug/l	100
25) 2-Butanone	7.488	43	82451	50.278	ug/l #	90
26) 2,2-Dichloropropane	7.488	77	39561	10.351	ug/l	99
27) cis-1,2-Dichloroethene	7.488	96	26902	10.148	ug/l	93
28) Bromochloromethane	7.818	49	14997	8.916	ug/l	85
29) Tetrahydrofuran	7.841	42	54815	51.708	ug/l	88
30) Chloroform	7.971	83	43421	10.155	ug/l	98
31) Cyclohexane	8.259	56	42564	10.523	ug/l	92
32) 1,1,1-Trichloroethane	8.171	97	41247	10.191	ug/l	92
36) 1,1-Dichloropropene	8.371	75	31681	10.103	ug/l	99
37) Ethyl Acetate	7.565	43	34005	9.685	ug/l #	96
38) Carbon Tetrachloride	8.365	117	35950	10.179	ug/l	93
39) Methylcyclohexane	9.600	83	38738	10.058	ug/l	93
40) Benzene	8.606	78	94885	10.157	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080724\
 Data File : VN083138.D
 Acq On : 07 Aug 2024 11:46
 Operator : JC\MD
 Sample : VSTDIC010
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC010

Manual Integrations
 APPROVED

Reviewed By : John Carlone 08/08/2024
 Supervised By : Mahesh Dadoda 08/09/2024

Quant Time: Aug 08 06:17:39 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 05:50:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.777	41	18392	9.213	ug/l	93
42) 1,2-Dichloroethane	8.671	62	34826	10.234	ug/l	97
43) Isopropyl Acetate	8.694	43	71420	10.170	ug/l	98
44) Trichloroethene	9.353	130	22669	10.195	ug/l	100
45) 1,2-Dichloropropane	9.624	63	22642	10.211	ug/l	98
46) Dibromomethane	9.706	93	15937	10.041	ug/l	97
47) Bromodichloromethane	9.888	83	35389	9.930	ug/l	97
48) Methyl methacrylate	9.682	41	28711	9.891	ug/l	93
49) 1,4-Dioxane	9.694	88	10545	201.364	ug/l	95
51) 4-Methyl-2-Pentanone	10.447	43	171277	51.587	ug/l	94
52) Toluene	10.629	92	59784	10.129	ug/l	97
53) t-1,3-Dichloropropene	10.835	75	36670	10.017	ug/l	100
54) cis-1,3-Dichloropropene	10.312	75	39503	10.145	ug/l	91
55) 1,1,2-Trichloroethane	11.018	97	21732	10.274	ug/l	94
56) Ethyl methacrylate	10.871	69	40676	10.201	ug/l #	86
57) 1,3-Dichloropropane	11.165	76	38200	10.132	ug/l	98
58) 2-Chloroethyl Vinyl ether	10.159	63	87140	51.698	ug/l	96
59) 2-Hexanone	11.194	43	129231	50.312	ug/l	93
60) Dibromochloromethane	11.359	129	25798	10.088	ug/l	99
61) 1,2-Dibromoethane	11.470	107	22707	10.219	ug/l	99
64) Tetrachloroethene	11.106	164	19174	10.362	ug/l	92
65) Chlorobenzene	11.894	112	64445	10.437	ug/l	96
66) 1,1,1,2-Tetrachloroethane	11.965	131	22498	10.331	ug/l	99
67) Ethyl Benzene	11.965	91	115664	10.211	ug/l	99
68) m/p-Xylenes	12.070	106	87775	20.687	ug/l	100
69) o-Xylene	12.400	106	43216	10.327	ug/l	97
70) Styrene	12.412	104	72641	10.335	ug/l	97
71) Bromoform	12.582	173	16320	9.893	ug/l #	95
73) Isopropylbenzene	12.694	105	110277	10.487	ug/l	99
74) N-amyl acetate	12.494	43	52998	10.303	ug/l	96
75) 1,1,2,2-Tetrachloroethane	12.941	83	30438	10.234	ug/l	98
76) 1,2,3-Trichloropropane	12.994	75	28359m	10.261	ug/l	
77) Bromobenzene	12.982	156	23781	10.180	ug/l	93
78) n-propylbenzene	13.035	91	125678	10.379	ug/l	99
79) 2-Chlorotoluene	13.123	91	80065	10.429	ug/l	96
80) 1,3,5-Trimethylbenzene	13.176	105	91559	10.400	ug/l	98
81) trans-1,4-Dichloro-2-b...	12.735	75	11760	9.271	ug/l	95
82) 4-Chlorotoluene	13.223	91	78348	10.174	ug/l	97
83) tert-Butylbenzene	13.441	119	80263	10.294	ug/l	98
84) 1,2,4-Trimethylbenzene	13.482	105	92502	10.426	ug/l	97
85) sec-Butylbenzene	13.617	105	108740	10.221	ug/l	98
86) p-Isopropyltoluene	13.729	119	88256	10.047	ug/l	97
87) 1,3-Dichlorobenzene	13.735	146	44168	10.048	ug/l	97
88) 1,4-Dichlorobenzene	13.811	146	44036	9.937	ug/l	96
89) n-Butylbenzene	14.059	91	76567	10.060	ug/l	99
90) Hexachloroethane	14.335	117	17798	10.487	ug/l	93
91) 1,2-Dichlorobenzene	14.106	146	43734	10.282	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	14.723	75	7258	10.056	ug/l	98
93) 1,2,4-Trichlorobenzene	15.394	180	23580	9.895	ug/l	98
94) Hexachlorobutadiene	15.500	225	10788	10.179	ug/l	99
95) Naphthalene	15.641	128	86373	10.232	ug/l	99
96) 1,2,3-Trichlorobenzene	15.841	180	24492	10.387	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080724\
Data File : VN083138.D
Acq On : 07 Aug 2024 11:46
Operator : JC\MD
Sample : VSTDICC010
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010

Manual Integrations
APPROVED

Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024

Quant Time: Aug 08 06:17:39 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 05:50:26 2024
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010

Manual Integrations APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 1066
Delta R.T. -0.000 min
Lab File: VN083138.D
Acq: 07 Aug 2024 11:46




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010



Tgt Ion:168 Resp: 191754
Ion Ratio Lower Upper
168 100
99 67.6 48.2 72.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#2
Dichlorodifluoromethane
Concen: 10.212 ug/l
RT: 2.124 min Scan# 29
Delta R.T. -0.000 min
Lab File: VN083138.D
Acq: 07 Aug 2024 11:46

Tgt Ion: 85 Resp: 22206
Ion Ratio Lower Upper
85 100
87 31.6 15.7 47.0





#3

Chloromethane

Concen: 10.150 ug/l

RT: 2.359 min Scan# 69

Delta R.T. -0.000 min

Lab File: VN083138.D

Acq: 07 Aug 2024 11:46

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC010

Tgt Ion: 50 Resp: 22590

Ion Ratio Lower Upper

50 100





52 35.8 25.0 37.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#4

Vinyl Chloride

Concen: 10.328 ug/l

RT: 2.512 min Scan# 95

Delta R.T. -0.000 min

Lab File: VN083138.D




Acq: 07 Aug 2024 11:46


Tgt Ion: 62 Resp: 23461

Ion Ratio Lower Upper

62 100

64 30.8 25.0 37.6





#5

Bromomethane

Concen: 10.488 ug/l

RT: 2.959 min Scan# 11

Delta R.T. 0.012 min

Lab File: VN083138.D

Acq: 07 Aug 2024 11:46

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC010

Tgt Ion: 94 Resp: 1478

Ion Ratio Lower Upper

94 100





96 102.2 78.0 117.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#6

Chloroethane

Concen: 9.581 ug/l

RT: 3.118 min Scan# 198

Delta R.T. 0.006 min

Lab File: VN083138.D




Acq: 07 Aug 2024 11:46


Tgt Ion: 64 Resp: 13616

Ion Ratio Lower Upper

64 100

66 30.6 26.6 40.0









#7
Trichlorofluoromethane
Concen: 10.357 ug/l
RT: 3.500 min Scan# 20
Delta R.T. 0.005 min
Lab File: VN083138.D
Acq: 07 Aug 2024 11:46

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010

Tgt Ion:101 Resp: 3886
Ion Ratio Lower Upper
101 100
103 63.5 52.7 79.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#8
Diethyl Ether
Concen: 10.571 ug/l
RT: 3.965 min Scan# 342
Delta R.T. 0.006 min
Lab File: VN083138.D
Acq: 07 Aug 2024 11:46

Tgt Ion: 74 Resp: 14761
Ion Ratio Lower Upper
74 100
45 95.9 55.5 166.3









#9
1,1,2-Trichlorotrifluoroethane
Concen: 10.189 ug/l
RT: 4.365 min Scan# 410
Delta R.T. -0.000 min
Lab File: VN083138.D
Acq: 07 Aug 2024 11:46

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010

Tgt Ion	Ratio	Lower	Upper
101	100		
85	48.0	38.3	57.5
151	74.1	59.3	88.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#10
Methyl Iodide
Concen: 10.005 ug/l
RT: 4.589 min Scan# 448
Delta R.T. -0.000 min
Lab File: VN083138.D
Acq: 07 Aug 2024 11:46

Tgt Ion	Ratio	Lower	Upper
142	100		
127	45.0	37.5	56.3
141	13.9	13.1	19.7





#11

Tert butyl alcohol

Concen: 52.299 ug/l

RT: 5.530 min Scan# 608

Delta R.T. -0.000 min

Lab File: VN083138.D

Acq: 07 Aug 2024 11:46

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC010

Tgt Ion: 59 Resp: 2966

Ion Ratio Lower Upper

59 100





57 12.1 8.9 13.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#12

1,1-Dichloroethene

Concen: 10.177 ug/l

RT: 4.336 min Scan# 405

Delta R.T. -0.006 min

Lab File: VN083138.D

Acq: 07 Aug 2024 11:46




Tgt Ion: 96 Resp: 21633


Ion Ratio Lower Upper

96 100

61 156.5 149.7 224.5

98 59.7 50.1 75.1





#13

Acrolein

Concen: 52.469 ug/l

RT: 4.177 min Scan# 31

Delta R.T. -0.006 min

Lab File: VN083138.D

Acq: 07 Aug 2024 11:46

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC010

Tgt Ion: 56 Resp: 1939

Ion Ratio Lower Upper

56 100





55 77.0 56.4 84.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#14

Allyl chloride

Concen: 10.343 ug/l

RT: 5.024 min Scan# 522

Delta R.T. 0.006 min

Lab File: VN083138.D

Acq: 07 Aug 2024 11:46




Tgt Ion: 41 Resp: 41546


Ion Ratio Lower Upper

41 100

39 72.0 67.8 101.6

76 33.2 25.8 38.8





#15

Acrylonitrile

Concen: 50.248 ug/l

RT: 5.724 min Scan# 641

Delta R.T. 0.006 min

Lab File: VN083138.D

Acq: 07 Aug 2024 11:46

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC010

Tgt Ion: 53 Resp: 58618

Ion Ratio Lower Upper

53 100

52 83.1 66.6 100.0





51 36.6 30.1 45.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#16

Acetone

Concen: 50.933 ug/l

RT: 4.430 min Scan# 421

Delta R.T. -0.000 min

Lab File: VN083138.D




Acq: 07 Aug 2024 11:46


Tgt Ion: 43 Resp: 54397

Ion Ratio Lower Upper

43 100


58 31.4 21.8 32.6





#17
Carbon Disulfide
Concen: 9.860 ug/l
RT: 4.712 min Scan# 469
Delta R.T. -0.000 min
Lab File: VN083138.D
Acq: 07 Aug 2024 11:46




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010



Tgt Ion: 76 Resp: 6132
Ion Ratio Lower Upper
76 100
78 9.9 7.2 10.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#18
Methyl Acetate
Concen: 10.145 ug/l
RT: 5.036 min Scan# 524
Delta R.T. 0.012 min
Lab File: VN083138.D
Acq: 07 Aug 2024 11:46

Tgt Ion: 43 Resp: 32282
Ion Ratio Lower Upper
43 100
74 22.8 16.3 24.5









#19
Methyl tert-butyl Ether
Concen: 10.072 ug/l
RT: 5.794 min Scan# 653
Delta R.T. -0.000 min
Lab File: VN083138.D
Acq: 07 Aug 2024 11:46

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010

Tgt Ion: 73 Resp: 77278
Ion Ratio Lower Upper
73 100
57 19.3 17.9 26.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#20
Methylene Chloride
Concen: 9.582 ug/l
RT: 5.277 min Scan# 565
Delta R.T. -0.000 min
Lab File: VN083138.D
Acq: 07 Aug 2024 11:46


Tgt Ion: 84 Resp: 23553
Ion Ratio Lower Upper
84 100
49 123.1 119.6 179.4
51 29.0 34.8 52.2#
86 63.2 52.9 79.3





#21
trans-1,2-Dichloroethene
Concen: 10.229 ug/l
RT: 5.788 min Scan# 61
Delta R.T. -0.006 min
Lab File: VN083138.D
Acq: 07 Aug 2024 11:46




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010



Tgt Ion: 96 Resp: 2247
Ion Ratio Lower Upper
96 100
61 144.0 123.4 185.2
98 58.3 49.3 73.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#22
Diisopropyl ether
Concen: 10.204 ug/l
RT: 6.671 min Scan# 802
Delta R.T. -0.000 min
Lab File: VN083138.D
Acq: 07 Aug 2024 11:46


Tgt Ion: 45 Resp: 77046
Ion Ratio Lower Upper
45 100
43 64.3 44.0 66.0
87 26.1 19.7 29.5
59 11.0 8.5 12.7





#23
Vinyl Acetate
Concen: 50.550 ug/l
RT: 6.606 min Scan# 791
Delta R.T. -0.000 min
Lab File: VN083138.D
Acq: 07 Aug 2024 11:46




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010



Tgt Ion: 43 Resp: 39115
Ion Ratio Lower Upper
43 100
86 11.2 7.3 10.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#24
1,1-Dichloroethane
Concen: 10.213 ug/l
RT: 6.571 min Scan# 785
Delta R.T. -0.000 min
Lab File: VN083138.D
Acq: 07 Aug 2024 11:46

Tgt Ion: 63 Resp: 42035
Ion Ratio Lower Upper
63 100
98 6.5 3.3 9.9
100 3.8 2.0 6.0





#25

2-Butanone

Concen: 50.278 ug/l

RT: 7.488 min Scan# 941

Delta R.T. 0.006 min

Lab File: VN083138.D

Acq: 07 Aug 2024 11:46

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC010

Tgt Ion: 43 Resp: 8245

Ion Ratio Lower Upper

43 100





72 27.4 18.2 27.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#26

2,2-Dichloropropane

Concen: 10.351 ug/l

RT: 7.488 min Scan# 941

Delta R.T. -0.000 min

Lab File: VN083138.D




Acq: 07 Aug 2024 11:46


Tgt Ion: 77 Resp: 39561

Ion Ratio Lower Upper

77 100


97 20.2 10.3 30.9





#27
cis-1,2-Dichloroethene
Concen: 10.148 ug/l
RT: 7.488 min Scan# 941
Delta R.T. -0.000 min
Lab File: VN083138.D
Acq: 07 Aug 2024 11:46




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010



Tgt Ion: 96 Resp: 2690
Ion Ratio Lower Upper
96 100
61 152.1 0.0 329.6
98 64.4 0.0 128.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#28
Bromochloromethane
Concen: 8.916 ug/l
RT: 7.818 min Scan# 997
Delta R.T. 0.006 min
Lab File: VN083138.D
Acq: 07 Aug 2024 11:46

Tgt Ion: 49 Resp: 14997
Ion Ratio Lower Upper
49 100
129 0.9 0.0 3.8
130 75.4 50.5 75.7





#29

Tetrahydrofuran

Concen: 51.708 ug/l

RT: 7.841 min Scan# 1001

Delta R.T. 0.006 min

Lab File: VN083138.D

Acq: 07 Aug 2024 11:46

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC010





Tgt Ion:	42	Resp:	5481
Ion	Ratio	Lower	Upper
42	100		
72	44.9	30.0	45.0
71	41.8	28.3	42.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#30

Chloroform

Concen: 10.155 ug/l




RT: 7.971 min Scan# 1023


Delta R.T. 0.006 min

Lab File: VN083138.D

Acq: 07 Aug 2024 11:46

Tgt Ion:	83	Resp:	43421
Ion	Ratio	Lower	Upper
83	100		
85	65.1	50.9	76.3





#31

Cyclohexane

Concen: 10.523 ug/l

RT: 8.259 min Scan# 1072

Delta R.T. -0.000 min

Lab File: VN083138.D

Acq: 07 Aug 2024 11:46

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC010

Tgt Ion: 56 Resp: 42564

Ion Ratio Lower Upper

56 100

69 34.0 24.5 36.7





84 90.8 66.0 99.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#32

1,1,1-Trichloroethane

Concen: 10.191 ug/l

RT: 8.171 min Scan# 1057

Delta R.T. -0.000 min

Lab File: VN083138.D

Acq: 07 Aug 2024 11:46




Tgt Ion: 97 Resp: 41247


Ion Ratio Lower Upper

97 100

99 59.5 52.0 78.0

61 46.2 42.1 63.1





#33

1,2-Dichloroethane-d4

Concen: 9.974 ug/l

RT: 8.582 min Scan# 1127

Delta R.T. -0.000 min

Lab File: VN083138.D

Acq: 07 Aug 2024 11:46

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC010

Tgt Ion: 65 Resp: 2722

Ion Ratio Lower Upper

65 100





67 50.3 0.0 104.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.100 min Scan# 1215

Delta R.T. -0.000 min

Lab File: VN083138.D

Acq: 07 Aug 2024 11:46




Tgt Ion:114 Resp: 332105


Ion Ratio Lower Upper

114 100

63 21.7 0.0 44.6

88 16.8 0.0 31.4





#35

Dibromofluoromethane

Concen: 9.832 ug/l

RT: 8.165 min Scan# 1056

Delta R.T. -0.000 min

Lab File: VN083138.D

Acq: 07 Aug 2024 11:46

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC010





Tgt Ion	Ratio	Lower	Upper
113	100		
111	105.9	82.4	123.6
192	16.7	14.9	22.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#36

1,1-Dichloropropene

Concen: 10.103 ug/l




RT: 8.371 min Scan# 1091


Delta R.T. -0.000 min

Lab File: VN083138.D

Acq: 07 Aug 2024 11:46

Tgt Ion	Ratio	Lower	Upper
75	100		
110	33.9	16.3	48.9
77	31.1	24.6	37.0





#37

Ethyl Acetate

Concen: 9.685 ug/l

RT: 7.565 min Scan# 91

Delta R.T. 0.006 min

Lab File: VN083138.D

Acq: 07 Aug 2024 11:46

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC010

Tgt Ion: 43 Resp: 3400

Ion Ratio Lower Upper

43 100

61 12.0 9.3 13.9





70 11.4 7.0 10.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#38

Carbon Tetrachloride

Concen: 10.179 ug/l

RT: 8.365 min Scan# 1090

Delta R.T. 0.006 min

Lab File: VN083138.D

Acq: 07 Aug 2024 11:46




Tgt Ion:117 Resp: 35950


Ion Ratio Lower Upper

117 100

119 87.5 74.9 112.3

121 25.6 24.3 36.5





#39

Methylcyclohexane

Concen: 10.058 ug/l

RT: 9.600 min Scan# 1130

Delta R.T. -0.000 min

Lab File: VN083138.D

Acq: 07 Aug 2024 11:46

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC010





Tgt Ion:	83	Resp:	38738
Ion	Ratio	Lower	Upper
83	100		
55	77.1	66.4	99.6
98	45.9	40.7	61.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#40

Benzene

Concen: 10.157 ug/l




RT: 8.606 min Scan# 1131


Delta R.T. -0.000 min

Lab File: VN083138.D

Acq: 07 Aug 2024 11:46


Tgt Ion:	78	Resp:	94885
Ion	Ratio	Lower	Upper
78	100		
77	22.2	19.0	28.4





#41
Methacrylonitrile
Concen: 9.213 ug/l
RT: 7.777 min Scan# 990
Delta R.T. -0.000 min
Lab File: VN083138.D
Acq: 07 Aug 2024 11:46




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC010




Tgt Ion: 41 Resp: 18392
Ion Ratio Lower Upper
41 100
39 62.6 45.5 68.3
67 71.3 50.5 75.7
52 30.1 24.0 36.0

Manual Integrations
APPROVED



Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024




#42
1,2-Dichloroethane
Concen: 10.234 ug/l
RT: 8.671 min Scan# 1142
Delta R.T. -0.000 min
Lab File: VN083138.D
Acq: 07 Aug 2024 11:46



Tgt Ion: 62 Resp: 34826
Ion Ratio Lower Upper
62 100
98 9.1 0.0 15.8





#43

Isopropyl Acetate

Concen: 10.170 ug/l

RT: 8.694 min Scan# 1146

Delta R.T. 0.006 min

Lab File: VN083138.D


Acq: 07 Aug 2024 11:46

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC010






Tgt Ion:	43	Resp:	71420
Ion	Ratio	Lower	Upper
43	100		
61	20.7	17.8	26.6
87	10.4	8.2	12.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#44

Trichloroethene

Concen: 10.195 ug/l




RT: 9.353 min Scan# 1258


Delta R.T. -0.000 min

Lab File: VN083138.D

Acq: 07 Aug 2024 11:46

Tgt Ion:	130	Resp:	22669
Ion	Ratio	Lower	Upper
130	100		
95	98.5	0.0	197.8





#45

1,2-Dichloropropane

Concen: 10.211 ug/l

RT: 9.624 min Scan# 11

Delta R.T. -0.000 min

Lab File: VN083138.D

Acq: 07 Aug 2024 11:46

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC010

Tgt Ion: 63 Resp: 2264

Ion Ratio Lower Upper

63 100





65 29.4 24.4 36.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#46

Dibromomethane

Concen: 10.041 ug/l

RT: 9.706 min Scan# 1318

Delta R.T. -0.006 min

Lab File: VN083138.D

Acq: 07 Aug 2024 11:46




Tgt Ion: 93 Resp: 15937


Ion Ratio Lower Upper

93 100

95 86.3 65.8 98.6

174 91.4 71.7 107.5





#47

Bromodichloromethane

Concen: 9.930 ug/l

RT: 9.888 min Scan# 1349

Delta R.T. -0.000 min

Lab File: VN083138.D

Acq: 07 Aug 2024 11:46

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC010





Tgt Ion:	83	Resp:	3538
Ion	Ratio	Lower	Upper
83	100		
85	61.1	51.1	76.7
127	8.8	7.2	10.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#48

Methyl methacrylate

Concen: 9.891 ug/l




RT: 9.682 min Scan# 1314


Delta R.T. -0.000 min

Lab File: VN083138.D

Acq: 07 Aug 2024 11:46

Tgt Ion:	41	Resp:	28711
Ion	Ratio	Lower	Upper
41	100		
69	81.0	56.3	84.5
39	63.6	50.3	75.5





#49

1,4-Dioxane

Concen: 201.364 ug/l

RT: 9.694 min Scan# 11

Delta R.T. -0.000 min

Lab File: VN083138.D

Acq: 07 Aug 2024 11:46

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC010

Tgt Ion: 88 Resp: 10549

Ion Ratio Lower Upper

88 100

43 39.2 27.8 41.8





58 77.8 59.4 89.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#50

Toluene-d8

Concen: 9.733 ug/l

RT: 10.565 min Scan# 1464

Delta R.T. -0.000 min

Lab File: VN083138.D




Acq: 07 Aug 2024 11:46


Tgt Ion: 98 Resp: 75261

Ion Ratio Lower Upper

98 100


100 66.0 51.5 77.3





#51
4-Methyl-2-Pentanone
Concen: 51.587 ug/l
RT: 10.447 min Scan# 1444
Delta R.T. -0.000 min
Lab File: VN083138.D
Acq: 07 Aug 2024 11:46




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010



Tgt Ion: 43 Resp: 17127
Ion Ratio Lower Upper
43 100
58 38.8 28.2 42.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#52
Toluene
Concen: 10.129 ug/l
RT: 10.629 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VN083138.D
Acq: 07 Aug 2024 11:46


Tgt Ion: 92 Resp: 59784
Ion Ratio Lower Upper
92 100
91 170.0 139.4 209.0





#53
t-1,3-Dichloropropene
Concen: 10.017 ug/l
RT: 10.835 min Scan# 1510
Delta R.T. -0.000 min
Lab File: VN083138.D
Acq: 07 Aug 2024 11:46




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC010



Tgt Ion: 75 Resp: 36670
Ion Ratio Lower Upper
75 100
77 32.2 25.7 38.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#54
cis-1,3-Dichloropropene
Concen: 10.145 ug/l
RT: 10.312 min Scan# 1421
Delta R.T. -0.000 min
Lab File: VN083138.D
Acq: 07 Aug 2024 11:46

Tgt Ion: 75 Resp: 39503
Ion Ratio Lower Upper
75 100
77 31.1 24.3 36.5
39 53.8 50.5 75.7





#55

1,1,2-Trichloroethane

Concen: 10.274 ug/l

RT: 11.018 min Scan# 1541

Delta R.T. -0.000 min

Lab File: VN083138.D

Acq: 07 Aug 2024 11:46

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC010

Tgt Ion: 97 Resp: 2173

Ion Ratio Lower Upper

97 100

83 84.0 71.4 107.0

85 54.3 45.8 68.6





99 58.9 52.0 78.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#56

Ethyl methacrylate

Concen: 10.201 ug/l

RT: 10.871 min Scan# 1516

Delta R.T. -0.006 min

Lab File: VN083138.D

Acq: 07 Aug 2024 11:46




Tgt Ion: 69 Resp: 40676


Ion Ratio Lower Upper

69 100

41 67.9 63.4 95.2


39 36.8 37.4 56.0#





#57
1,3-Dichloropropane
Concen: 10.132 ug/l
RT: 11.165 min Scan# 1566
Delta R.T. -0.000 min
Lab File: VN083138.D
Acq: 07 Aug 2024 11:46




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC010



Tgt Ion: 76 Resp: 38200
Ion Ratio Lower Upper
76 100
78 33.1 25.6 38.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#58
2-Chloroethyl Vinyl ether
Concen: 51.698 ug/l
RT: 10.159 min Scan# 1395
Delta R.T. -0.000 min
Lab File: VN083138.D
Acq: 07 Aug 2024 11:46

Tgt Ion: 63 Resp: 87140
Ion Ratio Lower Upper
63 100
106 28.9 21.4 32.0





#59

2-Hexanone

Concen: 50.312 ug/l

RT: 11.194 min Scan# 1571

Delta R.T. -0.000 min

Lab File: VN083138.D

Acq: 07 Aug 2024 11:46

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC010

Tgt Ion: 43 Resp: 12923

Ion Ratio Lower Upper

43 100





58 54.0 24.6 73.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#60

Dibromochloromethane

Concen: 10.088 ug/l

RT: 11.359 min Scan# 1599

Delta R.T. -0.000 min

Lab File: VN083138.D




Acq: 07 Aug 2024 11:46


Tgt Ion:129 Resp: 25798

Ion Ratio Lower Upper

129 100


127 77.5 39.2 117.6





#61
1,2-Dibromoethane
Concen: 10.219 ug/l
RT: 11.470 min Scan# 1618
Delta R.T. -0.000 min
Lab File: VN083138.D
Acq: 07 Aug 2024 11:46




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC010



Tgt Ion:107 Resp: 2270
Ion Ratio Lower Upper
107 100
109 92.7 73.7 110.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#62
4-Bromofluorobenzene
Concen: 9.517 ug/l
RT: 12.853 min Scan# 1853
Delta R.T. 0.006 min
Lab File: VN083138.D
Acq: 07 Aug 2024 11:46

Tgt Ion: 95 Resp: 28691
Ion Ratio Lower Upper
95 100
174 69.5 0.0 159.2
176 66.2 0.0 147.6





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.865 min Scan# 1685

Delta R.T. -0.000 min

Lab File: VN083138.D

Acq: 07 Aug 2024 11:46

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC010

Tgt Ion:117 Resp: 279409

Ion Ratio Lower Upper

117 100

82 60.2 47.5 71.3





119 31.9 25.6 38.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#64

Tetrachloroethene

Concen: 10.362 ug/l

RT: 11.106 min Scan# 1556

Delta R.T. -0.000 min

Lab File: VN083138.D

Acq: 07 Aug 2024 11:46

Tgt Ion:164 Resp: 19174




Ion Ratio Lower Upper


164 100

166 135.5 101.4 152.0

129 102.0 75.7 113.5


131 101.9 73.8 110.6





#65
Chlorobenzene
Concen: 10.437 ug/l
RT: 11.894 min Scan# 1690
Delta R.T. -0.000 min
Lab File: VN083138.D
Acq: 07 Aug 2024 11:46




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC010



Tgt Ion:112 Resp: 64445
Ion Ratio Lower Upper
112 100
114 30.4 26.2 39.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#66
1,1,1,2-Tetrachloroethane
Concen: 10.331 ug/l
RT: 11.965 min Scan# 1702
Delta R.T. -0.000 min
Lab File: VN083138.D
Acq: 07 Aug 2024 11:46

Tgt Ion:131 Resp: 22498
Ion Ratio Lower Upper
131 100
133 95.7 47.3 142.0
119 66.5 32.5 97.4





#67

Ethyl Benzene

Concen: 10.211 ug/l

RT: 11.965 min Scan# 1702

Delta R.T. -0.000 min

Lab File: VN083138.D

Acq: 07 Aug 2024 11:46

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC010

Tgt Ion: 91 Resp: 115664

Ion Ratio Lower Upper

91 100





106 30.7 24.0 36.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#68

m/p-Xylenes

Concen: 20.687 ug/l

RT: 12.070 min Scan# 1720

Delta R.T. -0.000 min

Lab File: VN083138.D




Acq: 07 Aug 2024 11:46


Tgt Ion:106 Resp: 87775

Ion Ratio Lower Upper

106 100


91 206.9 166.1 249.1





#69
o-Xylene
Concen: 10.327 ug/l
RT: 12.400 min Scan# 1776
Delta R.T. -0.000 min
Lab File: VN083138.D
Acq: 07 Aug 2024 11:46




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC010



Tgt Ion:106 Resp: 43210
Ion Ratio Lower Upper
106 100
91 217.0 111.3 333.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#70
Styrene
Concen: 10.335 ug/l
RT: 12.412 min Scan# 1778
Delta R.T. -0.000 min
Lab File: VN083138.D
Acq: 07 Aug 2024 11:46


Tgt Ion:104 Resp: 72641
Ion Ratio Lower Upper
104 100
78 53.1 41.6 62.4
103 52.2 44.0 66.0





#71
Bromoform
Concen: 9.893 ug/l
RT: 12.582 min Scan# 1807
Delta R.T. 0.006 min
Lab File: VN083138.D
Acq: 07 Aug 2024 11:46




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC010



Tgt Ion:173 Resp: 16320
Ion Ratio Lower Upper
173 100
175 51.7 24.0 72.0
254 0.0 0.0 0.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.794 min Scan# 2013
Delta R.T. -0.000 min
Lab File: VN083138.D
Acq: 07 Aug 2024 11:46


Tgt Ion:152 Resp: 125725
Ion Ratio Lower Upper
152 100
115 62.4 30.6 91.6
150 158.1 0.0 348.6





#73
Isopropylbenzene
Concen: 10.487 ug/l
RT: 12.694 min Scan# 1826
Delta R.T. -0.000 min
Lab File: VN083138.D
Acq: 07 Aug 2024 11:46




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC010



Tgt Ion: 105 Resp: 11027
Ion Ratio Lower Upper
105 100
120 26.0 12.9 38.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#74
N-ethyl acetate
Concen: 10.303 ug/l
RT: 12.494 min Scan# 1792
Delta R.T. -0.000 min
Lab File: VN083138.D
Acq: 07 Aug 2024 11:46


Tgt Ion: 43 Resp: 52998
Ion Ratio Lower Upper
43 100
70 39.8 29.8 44.6
55 24.3 18.7 28.1
61 24.0 17.4 26.0





#75
1,1,2,2-Tetrachloroethane
Concen: 10.234 ug/l
RT: 12.941 min Scan# 1867
Delta R.T. 0.006 min
Lab File: VN083138.D
Acq: 07 Aug 2024 11:46




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC010



Tgt Ion: 83 Resp: 30438
Ion Ratio Lower Upper
83 100
131 9.8 5.0 14.9
85 65.6 31.9 95.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#76
1,2,3-Trichloropropane
Concen: 10.261 ug/l m
RT: 12.994 min Scan# 1877
Delta R.T. -0.000 min
Lab File: VN083138.D
Acq: 07 Aug 2024 11:46

Tgt Ion: 75 Resp: 28359
Ion Ratio Lower Upper
75 100
77 197.9 110.9 332.6









#77
Bromobenzene
Concen: 10.180 ug/l
RT: 12.982 min Scan# 1875
Delta R.T. -0.000 min
Lab File: VN083138.D
Acq: 07 Aug 2024 11:46

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC010

Tgt Ion:156 Resp: 2378
Ion Ratio Lower Upper
156 100
77 236.0 111.3 333.8
158 98.9 48.3 144.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#78
n-propylbenzene
Concen: 10.379 ug/l
RT: 13.035 min Scan# 1884
Delta R.T. -0.000 min
Lab File: VN083138.D
Acq: 07 Aug 2024 11:46

Tgt Ion: 91 Resp: 125678
Ion Ratio Lower Upper
91 100
120 22.2 10.9 32.9









#79
2-Chlorotoluene
Concen: 10.429 ug/l
RT: 13.123 min Scan# 1899
Delta R.T. -0.000 min
Lab File: VN083138.D
Acq: 07 Aug 2024 11:46

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010

Tgt Ion: 91 Resp: 8006
Ion Ratio Lower Upper
91 100
126 30.4 16.2 48.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#80
1,3,5-Trimethylbenzene
Concen: 10.400 ug/l
RT: 13.176 min Scan# 1908
Delta R.T. 0.006 min
Lab File: VN083138.D
Acq: 07 Aug 2024 11:46


Tgt Ion:105 Resp: 91559
Ion Ratio Lower Upper
105 100
120 47.3 24.3 72.8





#81
trans-1,4-Dichloro-2-butene
Concen: 9.271 ug/l
RT: 12.735 min Scan# 1833
Delta R.T. -0.000 min
Lab File: VN083138.D
Acq: 07 Aug 2024 11:46




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010



Tgt Ion: 75 Resp: 11760
Ion Ratio Lower Upper
75 100
53 116.7 88.4 132.6
89 47.3 36.3 54.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#82
4-Chlorotoluene
Concen: 10.174 ug/l
RT: 13.223 min Scan# 1916
Delta R.T. -0.000 min
Lab File: VN083138.D
Acq: 07 Aug 2024 11:46


Tgt Ion: 91 Resp: 78348
Ion Ratio Lower Upper
91 100
126 31.0 16.2 48.6





#83
tert-Butylbenzene
Concen: 10.294 ug/l
RT: 13.441 min Scan# 1953
Delta R.T. -0.000 min
Lab File: VN083138.D
Acq: 07 Aug 2024 11:46




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010



Tgt Ion:119 Resp: 8026
Ion Ratio Lower Upper
119 100
91 71.7 35.0 105.1
134 25.3 13.3 39.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#84
1,2,4-Trimethylbenzene
Concen: 10.426 ug/l
RT: 13.482 min Scan# 1960
Delta R.T. -0.000 min
Lab File: VN083138.D
Acq: 07 Aug 2024 11:46


Tgt Ion:105 Resp: 92502
Ion Ratio Lower Upper
105 100
120 42.1 21.9 65.8





#85
sec-Butylbenzene
Concen: 10.221 ug/l
RT: 13.617 min Scan# 1983
Delta R.T. -0.000 min
Lab File: VN083138.D
Acq: 07 Aug 2024 11:46




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC010



Tgt Ion:105 Resp: 108740
Ion Ratio Lower Upper
105 100
134 18.8 9.8 29.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#86
p-Isopropyltoluene
Concen: 10.047 ug/l
RT: 13.729 min Scan# 2002
Delta R.T. -0.000 min
Lab File: VN083138.D
Acq: 07 Aug 2024 11:46

Tgt Ion:119 Resp: 88256
Ion Ratio Lower Upper
119 100
134 26.2 13.0 39.0
91 27.0 12.3 36.9





#87

1,3-Dichlorobenzene

Concen: 10.048 ug/l

RT: 13.735 min Scan# 2003

Delta R.T. -0.000 min

Lab File: VN083138.D

Acq: 07 Aug 2024 11:46

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC010

Tgt Ion:146 Resp: 4416

Ion Ratio Lower Upper

146 100

111 44.1 20.6 61.9





148 65.1 31.8 95.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#88

1,4-Dichlorobenzene

Concen: 9.937 ug/l

RT: 13.811 min Scan# 2016

Delta R.T. -0.000 min

Lab File: VN083138.D

Acq: 07 Aug 2024 11:46




Tgt Ion:146 Resp: 44036


Ion Ratio Lower Upper

146 100

111 45.6 20.1 60.3

148 64.0 32.2 96.6





#89

n-Butylbenzene

Concen: 10.060 ug/l

RT: 14.059 min Scan# 2058

Delta R.T. -0.000 min

Lab File: VN083138.D

Acq: 07 Aug 2024 11:46

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC010





Tgt Ion:	91	Resp:	7656
Ion	Ratio	Lower	Upper
91	100		
92	53.1	26.1	78.3
134	24.7	12.1	36.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#90

Hexachloroethane

Concen: 10.487 ug/l




RT: 14.335 min Scan# 2105


Delta R.T. -0.000 min

Lab File: VN083138.D

Acq: 07 Aug 2024 11:46

Tgt Ion:	117	Resp:	17798
Ion	Ratio	Lower	Upper
117	100		
201	65.3	35.8	107.3





#91

1,2-Dichlorobenzene

Concen: 10.282 ug/l

RT: 14.106 min Scan# 2066

Delta R.T. -0.000 min

Lab File: VN083138.D

Acq: 07 Aug 2024 11:46

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC010





Tgt Ion:	146	Resp:	43734
Ion	Ratio	Lower	Upper
146	100		
111	43.8	21.4	64.2
148	62.3	32.1	96.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#92

1,2-Dibromo-3-Chloropropane

Concen: 10.056 ug/l




RT: 14.723 min Scan# 2171


Delta R.T. 0.006 min

Lab File: VN083138.D

Acq: 07 Aug 2024 11:46

Tgt Ion:	75	Resp:	7258
Ion	Ratio	Lower	Upper
75	100		
155	75.6	36.6	109.8
157	95.2	46.9	140.6





#93

1,2,4-Trichlorobenzene

Concen: 9.895 ug/l

RT: 15.394 min Scan# 2285

Delta R.T. -0.000 min

Lab File: VN083138.D

Acq: 07 Aug 2024 11:46

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC010

Tgt Ion:180 Resp: 23580

Ion Ratio Lower Upper

180 100

182 96.2 47.3 141.8





145 33.2 15.7 47.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#94

Hexachlorobutadiene

Concen: 10.179 ug/l

RT: 15.500 min Scan# 2303

Delta R.T. -0.000 min

Lab File: VN083138.D

Acq: 07 Aug 2024 11:46




Tgt Ion:225 Resp: 10788


Ion Ratio Lower Upper

225 100

223 64.4 31.9 95.7

227 64.4 32.5 97.5









#95
Naphthalene
Concen: 10.232 ug/l
RT: 15.641 min Scan# 2327
Delta R.T. -0.000 min
Lab File: VN083138.D
Acq: 07 Aug 2024 11:46

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC010

Tgt Ion	Ratio	Lower	Upper
128	100		
127	12.6	10.6	16.0
129	10.8	8.6	12.8




Manual Integrations
APPROVED

Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#96
1,2,3-Trichlorobenzene
Concen: 10.387 ug/l
RT: 15.841 min Scan# 2361
Delta R.T. -0.000 min
Lab File: VN083138.D
Acq: 07 Aug 2024 11:46

Tgt Ion	Ratio	Lower	Upper
180	100		
182	95.3	48.9	146.8
145	34.8	16.8	50.4



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080724\
 Data File : VN083139.D
 Acq On : 07 Aug 2024 12:10
 Operator : JC\MD
 Sample : VSTDICC005
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC005

Manual Integrations
 APPROVED

Reviewed By :John Carlone 08/08/2024
 Supervised By :Mahesh Dadoda 08/09/2024

Quant Time: Aug 08 06:18:41 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 05:50:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	201859	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	347562	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	293513	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	133537	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	15867	5.522	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	11.040%#		
35) Dibromofluoromethane	8.159	113	12039	5.549	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	11.100%#		
50) Toluene-d8	10.565	98	43104	5.326	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	10.660%#		
62) 4-Bromofluorobenzene	12.847	95	16776	5.317	ug/l	0.00
Spiked Amount 50.000	Range 64 - 133		Recovery =	10.640%#		
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	2.124	85	12045	5.262	ug/l	97
3) Chloromethane	2.359	50	12017	5.128	ug/l	97
4) Vinyl Chloride	2.512	62	11837	4.950	ug/l	96
5) Bromomethane	2.942	94	8328	5.612	ug/l	88
6) Chloroethane	3.106	64	7688	5.139	ug/l	90
7) Trichlorofluoromethane	3.489	101	20242	5.124	ug/l	93
8) Diethyl Ether	3.959	74	7674	5.220	ug/l	91
9) 1,1,2-Trichlorotrifluo...	4.377	101	11261	5.170	ug/l	99
10) Methyl Iodide	4.595	142	13878	4.844	ug/l #	95
11) Tert butyl alcohol	5.524	59	15685	26.268	ug/l	99
12) 1,1-Dichloroethene	4.342	96	11718	5.237	ug/l	90
13) Acrolein	4.177	56	9615	24.707	ug/l	89
14) Allyl chloride	5.018	41	22183	5.246	ug/l	87
15) Acrylonitrile	5.724	53	31357	25.534	ug/l	96
16) Acetone	4.436	43	27636	24.581	ug/l	97
17) Carbon Disulfide	4.712	76	32908	5.026	ug/l	96
18) Methyl Acetate	5.030	43	17029	5.084	ug/l	98
19) Methyl tert-butyl Ether	5.789	73	41439	5.131	ug/l	93
20) Methylene Chloride	5.271	84	12913	4.991	ug/l #	84
21) trans-1,2-Dichloroethene	5.783	96	11465	4.957	ug/l	88
22) Diisopropyl ether	6.677	45	39688	4.993	ug/l #	97
23) Vinyl Acetate	6.600	43	202471m	24.856	ug/l	
24) 1,1-Dichloroethane	6.565	63	22058	5.091	ug/l	98
25) 2-Butanone	7.488	43	41750	24.185	ug/l	91
26) 2,2-Dichloropropane	7.494	77	19864	4.937	ug/l	95
27) cis-1,2-Dichloroethene	7.483	96	13862	4.967	ug/l	94
28) Bromochloromethane	7.812	49	9874	5.577	ug/l #	85
29) Tetrahydrofuran	7.841	42	28260	25.324	ug/l	89
30) Chloroform	7.965	83	22650	5.032	ug/l	99
31) Cyclohexane	8.265	56	24999	5.871	ug/l	96
32) 1,1,1-Trichloroethane	8.165	97	21825	5.123	ug/l #	49
36) 1,1-Dichloropropene	8.377	75	16102	4.907	ug/l	99
37) Ethyl Acetate	7.559	43	17909	4.874	ug/l #	96
38) Carbon Tetrachloride	8.365	117	18728	5.067	ug/l	98
39) Methylcyclohexane	9.600	83	19981	4.957	ug/l	91
40) Benzene	8.606	78	48765	4.988	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080724\
 Data File : VN083139.D
 Acq On : 07 Aug 2024 12:10
 Operator : JC\MD
 Sample : VSTDICC005
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC005

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024

Quant Time: Aug 08 06:18:41 2024

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M

Quant Title : SW846 8260

QLast Update : Thu Aug 08 05:50:26 2024

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.777	41	10086	4.828	ug/l	95
42) 1,2-Dichloroethane	8.671	62	18219	5.116	ug/l	97
43) Isopropyl Acetate	8.694	43	44817	5.438	ug/l #	92
44) Trichloroethene	9.353	130	11830	5.084	ug/l	94
45) 1,2-Dichloropropane	9.624	63	11320	4.878	ug/l	90
46) Dibromomethane	9.712	93	8436	5.079	ug/l	98
47) Bromodichloromethane	9.888	83	18799	5.040	ug/l	98
48) Methyl methacrylate	9.682	41	14704	4.840	ug/l	94
49) 1,4-Dioxane	9.694	88	5239	95.593	ug/l #	77
51) 4-Methyl-2-Pentanone	10.447	43	86679	24.946	ug/l	93
52) Toluene	10.629	92	30769	4.981	ug/l	99
53) t-1,3-Dichloropropene	10.835	75	18964	4.950	ug/l	100
54) cis-1,3-Dichloropropene	10.312	75	20431	5.014	ug/l	93
55) 1,1,2-Trichloroethane	11.018	97	10913	4.930	ug/l	98
56) Ethyl methacrylate	10.876	69	20661	4.951	ug/l	93
57) 1,3-Dichloropropane	11.165	76	18934	4.799	ug/l	94
58) 2-Chloroethyl Vinyl ether	10.159	63	43528	24.676	ug/l	96
59) 2-Hexanone	11.194	43	66854	24.870	ug/l	95
60) Dibromochloromethane	11.359	129	13307	4.972	ug/l	97
61) 1,2-Dibromoethane	11.471	107	11523	4.955	ug/l	96
64) Tetrachloroethene	11.106	164	9517	4.896	ug/l	96
65) Chlorobenzene	11.894	112	32114	4.951	ug/l	96
66) 1,1,1,2-Tetrachloroethane	11.959	131	11758	5.140	ug/l	94
67) Ethyl Benzene	11.965	91	59648	5.013	ug/l	97
68) m/p-Xylenes	12.070	106	44572	10.000	ug/l	96
69) o-Xylene	12.400	106	22003	5.005	ug/l	99
70) Styrene	12.412	104	37023	5.014	ug/l	98
71) Bromoform	12.582	173	8085	4.665	ug/l #	89
73) Isopropylbenzene	12.694	105	56741	5.080	ug/l	100
74) N-amyl acetate	12.494	43	27393	5.014	ug/l	95
75) 1,1,2,2-Tetrachloroethane	12.935	83	16514	5.227	ug/l	97
76) 1,2,3-Trichloropropane	12.994	75	14463m	4.927	ug/l	
77) Bromobenzene	12.982	156	13024	5.249	ug/l	94
78) n-propylbenzene	13.035	91	64751	5.035	ug/l	99
79) 2-Chlorotoluene	13.123	91	41176	5.050	ug/l	99
80) 1,3,5-Trimethylbenzene	13.170	105	47581	5.088	ug/l	97
81) trans-1,4-Dichloro-2-b...	12.735	75	7079	5.254	ug/l	92
82) 4-Chlorotoluene	13.223	91	41137	5.030	ug/l	97
83) tert-Butylbenzene	13.435	119	42723	5.159	ug/l	100
84) 1,2,4-Trimethylbenzene	13.482	105	47965	5.090	ug/l	98
85) sec-Butylbenzene	13.617	105	56947	5.040	ug/l	98
86) p-Isopropyltoluene	13.729	119	47654	5.108	ug/l	98
87) 1,3-Dichlorobenzene	13.735	146	22814	4.887	ug/l	98
88) 1,4-Dichlorobenzene	13.812	146	23971	5.093	ug/l	94
89) n-Butylbenzene	14.053	91	40282	4.983	ug/l	99
90) Hexachloroethane	14.335	117	9298	5.158	ug/l	93
91) 1,2-Dichlorobenzene	14.106	146	23116	5.117	ug/l	96
92) 1,2-Dibromo-3-Chloropr...	14.717	75	4097	5.345	ug/l	94
93) 1,2,4-Trichlorobenzene	15.394	180	11922	4.710	ug/l	94
94) Hexachlorobutadiene	15.500	225	5946	5.282	ug/l	93
95) Naphthalene	15.641	128	43131	4.811	ug/l	99
96) 1,2,3-Trichlorobenzene	15.841	180	12285	4.905	ug/l	94

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080724\
Data File : VN083139.D
Acq On : 07 Aug 2024 12:10
Operator : JC\MD
Sample : VSTDICC005
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED

Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024

Quant Time: Aug 08 06:18:41 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 05:50:26 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080724\
Data File : VN083139.D
Acq On : 07 Aug 2024 12:10
Operator : JC\MD
Sample : VSTDICC005
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1


Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED

Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024


Quant Time: Aug 08 06:18:41 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 05:50:26 2024
Response via : Initial Calibration





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 1066
Delta R.T. -0.000 min
Lab File: VN083139.D
Acq: 07 Aug 2024 12:10




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005



Tgt Ion:168 Resp: 201859
Ion Ratio Lower Upper
168 100
99 62.6 48.2 72.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#2
Dichlorodifluoromethane
Concen: 5.262 ug/l
RT: 2.124 min Scan# 29
Delta R.T. 0.000 min
Lab File: VN083139.D
Acq: 07 Aug 2024 12:10

Tgt Ion: 85 Resp: 12045
Ion Ratio Lower Upper
85 100
87 32.8 15.7 47.0









#3
Chloromethane
Concen: 5.128 ug/l
RT: 2.359 min Scan# 69
Delta R.T. 0.000 min
Lab File: VN083139.D
Acq: 07 Aug 2024 12:10

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Tgt Ion: 50 Resp: 1201
Ion Ratio Lower Upper
50 100
52 32.7 25.0 37.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#4
Vinyl Chloride
Concen: 4.950 ug/l
RT: 2.512 min Scan# 95
Delta R.T. 0.000 min
Lab File: VN083139.D
Acq: 07 Aug 2024 12:10

Tgt Ion: 62 Resp: 11837
Ion Ratio Lower Upper
62 100
64 33.6 25.0 37.6





#5

Bromomethane

Concen: 5.612 ug/l

RT: 2.942 min Scan# 1

Delta R.T. -0.006 min

Lab File: VN083139.D


Acq: 07 Aug 2024 12:10

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005



Tgt Ion: 94 Resp: 8328

Ion Ratio Lower Upper

94 100




96 85.3 78.0 117.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#6

Chloroethane

Concen: 5.139 ug/l

RT: 3.106 min Scan# 196

Delta R.T. -0.006 min

Lab File: VN083139.D




Acq: 07 Aug 2024 12:10


Tgt Ion: 64 Resp: 7688

Ion Ratio Lower Upper

64 100


66 27.9 26.6 40.0





#7
Trichlorofluoromethane
Concen: 5.124 ug/l
RT: 3.489 min Scan# 262
Delta R.T. -0.006 min
Lab File: VN083139.D
Acq: 07 Aug 2024 12:10




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005



Tgt Ion:101 Resp: 2024
Ion Ratio Lower Upper
101 100
103 60.3 52.7 79.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#8
Diethyl Ether
Concen: 5.220 ug/l
RT: 3.959 min Scan# 341
Delta R.T. 0.000 min
Lab File: VN083139.D
Acq: 07 Aug 2024 12:10

Tgt Ion: 74 Resp: 7674
Ion Ratio Lower Upper
74 100
45 101.8 55.5 166.3





#9

1,1,2-Trichlorotrifluoroethane

Concen: 5.170 ug/l

RT: 4.377 min Scan# 410

Delta R.T. 0.012 min

Lab File: VN083139.D

Acq: 07 Aug 2024 12:10

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005

Tgt Ion:101 Resp: 1126

Ion Ratio Lower Upper

101 100

85 47.2 38.3 57.5





151 74.4 59.3 88.9

Manual Integrations

APPROVED

Reviewed By :John Carbone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#10

Methyl Iodide

Concen: 4.844 ug/l

RT: 4.595 min Scan# 449

Delta R.T. 0.006 min

Lab File: VN083139.D

Acq: 07 Aug 2024 12:10




Tgt Ion:142 Resp: 13878


Ion Ratio Lower Upper

142 100

127 45.4 37.5 56.3

141 10.5 13.1 19.7#





#11

Tert butyl alcohol

Concen: 26.268 ug/l

RT: 5.524 min Scan# 608

Delta R.T. -0.006 min

Lab File: VN083139.D

Acq: 07 Aug 2024 12:10

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC005

Tgt Ion: 59 Resp: 1568

Ion Ratio Lower Upper

59 100





57 11.4 8.9 13.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#12

1,1-Dichloroethene

Concen: 5.237 ug/l

RT: 4.342 min Scan# 406

Delta R.T. 0.000 min

Lab File: VN083139.D

Acq: 07 Aug 2024 12:10




Tgt Ion: 96 Resp: 11718


Ion Ratio Lower Upper

96 100

61 168.3 149.7 224.5

98 60.4 50.1 75.1





#13

Acrolein

Concen: 24.707 ug/l

RT: 4.177 min Scan# 31

Delta R.T. -0.006 min

Lab File: VN083139.D

Acq: 07 Aug 2024 12:10

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC005

Tgt Ion: 56 Resp: 961

Ion Ratio Lower Upper

56 100





55 80.0 56.4 84.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#14

Allyl chloride

Concen: 5.246 ug/l

RT: 5.018 min Scan# 521

Delta R.T. 0.000 min

Lab File: VN083139.D

Acq: 07 Aug 2024 12:10




Tgt Ion: 41 Resp: 22183


Ion Ratio Lower Upper

41 100

39 69.3 67.8 101.6


76 31.2 25.8 38.8





#15
Acrylonitrile
Concen: 25.534 ug/l
RT: 5.724 min Scan# 641
Delta R.T. 0.006 min
Lab File: VN083139.D
Acq: 07 Aug 2024 12:10




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005



Tgt Ion: 53 Resp: 3135
Ion Ratio Lower Upper
53 100
52 79.9 66.6 100.0
51 35.3 30.1 45.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#16
Acetone
Concen: 24.581 ug/l
RT: 4.436 min Scan# 422
Delta R.T. 0.006 min
Lab File: VN083139.D
Acq: 07 Aug 2024 12:10


Tgt Ion: 43 Resp: 27636
Ion Ratio Lower Upper
43 100
58 28.6 21.8 32.6





#17
Carbon Disulfide
Concen: 5.026 ug/l
RT: 4.712 min Scan# 469
Delta R.T. 0.000 min
Lab File: VN083139.D
Acq: 07 Aug 2024 12:10




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005



Tgt Ion: 76 Resp: 32908
Ion Ratio Lower Upper
76 100
78 7.4 7.2 10.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#18
Methyl Acetate
Concen: 5.084 ug/l
RT: 5.030 min Scan# 523
Delta R.T. 0.006 min
Lab File: VN083139.D
Acq: 07 Aug 2024 12:10

Tgt Ion: 43 Resp: 17029
Ion Ratio Lower Upper
43 100
74 21.4 16.3 24.5





#19

Methyl tert-butyl Ether

Concen: 5.131 ug/l

RT: 5.789 min Scan# 61

Delta R.T. -0.006 min

Lab File: VN083139.D

Acq: 07 Aug 2024 12:10

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005

Tgt Ion: 73 Resp: 41439

Ion Ratio Lower Upper

73 100





57 19.2 17.9 26.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#20

Methylene Chloride

Concen: 4.991 ug/l

RT: 5.271 min Scan# 564

Delta R.T. -0.006 min

Lab File: VN083139.D

Acq: 07 Aug 2024 12:10

Tgt Ion: 84 Resp: 12913




Ion Ratio Lower Upper


84 100

49 117.8 119.6 179.4#

51 42.0 34.8 52.2


86 61.5 52.9 79.3





#21
trans-1,2-Dichloroethene
Concen: 4.957 ug/l
RT: 5.783 min Scan# 61
Delta R.T. -0.012 min
Lab File: VN083139.D
Acq: 07 Aug 2024 12:10




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005



Tgt Ion: 96 Resp: 1146
Ion Ratio Lower Upper
96 100
61 133.0 123.4 185.2
98 63.1 49.3 73.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#22
Diisopropyl ether
Concen: 4.993 ug/l
RT: 6.677 min Scan# 803
Delta R.T. 0.006 min
Lab File: VN083139.D
Acq: 07 Aug 2024 12:10


Tgt Ion: 45 Resp: 39688
Ion Ratio Lower Upper
45 100
43 54.9 44.0 66.0
87 29.0 19.7 29.5
59 13.1 8.5 12.7#





#23
Vinyl Acetate
Concen: 24.856 ug/l m
RT: 6.600 min Scan# 791
Delta R.T. -0.006 min
Lab File: VN083139.D
Acq: 07 Aug 2024 12:10




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005



Tgt Ion: 43 Resp: 20247
Ion Ratio Lower Upper
43 100
86 10.2 7.3 10.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#24
1,1-Dichloroethane
Concen: 5.091 ug/l
RT: 6.565 min Scan# 784
Delta R.T. -0.006 min
Lab File: VN083139.D
Acq: 07 Aug 2024 12:10


Tgt Ion: 63 Resp: 22058
Ion Ratio Lower Upper
63 100
98 7.3 3.3 9.9
100 5.0 2.0 6.0





#25
2-Butanone
Concen: 24.185 ug/l
RT: 7.488 min Scan# 941
Delta R.T. 0.006 min
Lab File: VN083139.D
Acq: 07 Aug 2024 12:10




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005



Tgt Ion: 43 Resp: 41750
Ion Ratio Lower Upper
43 100
72 27.1 18.2 27.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#26
2,2-Dichloropropane
Concen: 4.937 ug/l
RT: 7.494 min Scan# 942
Delta R.T. 0.006 min
Lab File: VN083139.D
Acq: 07 Aug 2024 12:10

Tgt Ion: 77 Resp: 19864
Ion Ratio Lower Upper
77 100
97 22.7 10.3 30.9





#27

cis-1,2-Dichloroethene

Concen: 4.967 ug/l

RT: 7.483 min Scan# 941

Delta R.T. -0.006 min

Lab File: VN083139.D

Acq: 07 Aug 2024 12:10

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC005

Tgt Ion: 96 Resp: 1386

Ion Ratio Lower Upper

96 100

61 154.7 0.0 329.6





98 62.6 0.0 128.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#28

Bromochloromethane

Concen: 5.577 ug/l

RT: 7.812 min Scan# 996

Delta R.T. 0.000 min

Lab File: VN083139.D

Acq: 07 Aug 2024 12:10




Tgt Ion: 49 Resp: 9874


Ion Ratio Lower Upper

49 100

129 0.0 0.0 3.8

130 75.2 50.5 75.7





#29

Tetrahydrofuran

Concen: 25.324 ug/l

RT: 7.841 min Scan# 1000

Delta R.T. 0.006 min

Lab File: VN083139.D

Acq: 07 Aug 2024 12:10

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005





Tgt Ion:	42	Resp:	28260
Ion Ratio	Lower	Upper	
42	100		
72	44.9	30.0	45.0
71	41.3	28.3	42.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#30

Chloroform

Concen: 5.032 ug/l




RT: 7.965 min Scan# 1022


Delta R.T. 0.000 min

Lab File: VN083139.D

Acq: 07 Aug 2024 12:10


Tgt Ion:	83	Resp:	22650
Ion Ratio	Lower	Upper	
83	100		
85	64.3	50.9	76.3





#31
Cyclohexane
Concen: 5.871 ug/l
RT: 8.265 min Scan# 1072
Delta R.T. 0.006 min
Lab File: VN083139.D
Acq: 07 Aug 2024 12:10




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005



Tgt Ion: 56 Resp: 24999
Ion Ratio Lower Upper
56 100
69 33.5 24.5 36.7
84 85.8 66.0 99.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#32
1,1,1-Trichloroethane
Concen: 5.123 ug/l
RT: 8.165 min Scan# 1056
Delta R.T. -0.006 min
Lab File: VN083139.D
Acq: 07 Aug 2024 12:10

Tgt Ion: 97 Resp: 21825
Ion Ratio Lower Upper
97 100
99 0.0 52.0 78.0#
61 44.1 42.1 63.1





#33

1,2-Dichloroethane-d4

Concen: 5.522 ug/l

RT: 8.582 min Scan# 1127

Delta R.T. 0.000 min

Lab File: VN083139.D

Acq: 07 Aug 2024 12:10

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005

Tgt Ion: 65 Resp: 1586

Ion Ratio Lower Upper

65 100





67 51.8 0.0 104.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.100 min Scan# 1215

Delta R.T. 0.000 min

Lab File: VN083139.D

Acq: 07 Aug 2024 12:10




Tgt Ion:114 Resp: 347562


Ion Ratio Lower Upper

114 100

63 22.1 0.0 44.6

88 16.8 0.0 31.4





#35

Dibromofluoromethane

Concen: 5.549 ug/l

RT: 8.159 min Scan# 1055

Delta R.T. -0.006 min

Lab File: VN083139.D

Acq: 07 Aug 2024 12:10

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005

Tgt Ion: 113 Resp: 12039

Ion Ratio Lower Upper

113 100

111 101.4 82.4 123.6





192 16.2 14.9 22.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#36

1,1-Dichloropropene

Concen: 4.907 ug/l

RT: 8.377 min Scan# 1092

Delta R.T. 0.006 min

Lab File: VN083139.D

Acq: 07 Aug 2024 12:10




Tgt Ion: 75 Resp: 16102


Ion Ratio Lower Upper

75 100

110 34.1 16.3 48.9

77 30.8 24.6 37.0





#37

Ethyl Acetate

Concen: 4.874 ug/l

RT: 7.559 min Scan# 91

Delta R.T. 0.000 min

Lab File: VN083139.D

Acq: 07 Aug 2024 12:10

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005





Tgt Ion:	43	Resp:	1790
Ion	Ratio	Lower	Upper
43	100		
61	12.1	9.3	13.9
70	11.6	7.0	10.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#38

Carbon Tetrachloride

Concen: 5.067 ug/l




RT: 8.365 min Scan# 1090


Delta R.T. 0.006 min

Lab File: VN083139.D

Acq: 07 Aug 2024 12:10

Tgt Ion:	117	Resp:	18728
Ion	Ratio	Lower	Upper
117	100		
119	94.8	74.9	112.3
121	27.0	24.3	36.5





#39

Methylcyclohexane

Concen: 4.957 ug/l

RT: 9.600 min Scan# 1130

Delta R.T. 0.000 min

Lab File: VN083139.D

Acq: 07 Aug 2024 12:10

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005





Tgt Ion:	83	Resp:	1998
Ion	Ratio	Lower	Upper
83	100		
55	74.1	66.4	99.6
98	55.9	40.7	61.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#40

Benzene

Concen: 4.988 ug/l




RT: 8.606 min Scan# 1131


Delta R.T. 0.000 min

Lab File: VN083139.D

Acq: 07 Aug 2024 12:10


Tgt Ion:	78	Resp:	48765
Ion	Ratio	Lower	Upper
78	100		
77	22.8	19.0	28.4





#41
Methacrylonitrile
Concen: 4.828 ug/l
RT: 7.777 min Scan# 990
Delta R.T. 0.000 min
Lab File: VN083139.D
Acq: 07 Aug 2024 12:10




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005



Tgt Ion: 41 Resp: 1008
Ion Ratio Lower Upper
41 100
39 60.2 45.5 68.3
67 68.2 50.5 75.7
52 27.9 24.0 36.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#42
1,2-Dichloroethane
Concen: 5.116 ug/l
RT: 8.671 min Scan# 1142
Delta R.T. 0.000 min
Lab File: VN083139.D
Acq: 07 Aug 2024 12:10

Tgt Ion: 62 Resp: 18219
Ion Ratio Lower Upper
62 100
98 9.0 0.0 15.8





#43

Isopropyl Acetate

Concen: 5.438 ug/l

RT: 8.694 min Scan# 1146

Delta R.T. 0.006 min

Lab File: VN083139.D

Acq: 07 Aug 2024 12:10

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC005





Tgt Ion:	43	Resp:	4481
Ion	Ratio	Lower	Upper
43	100		
61	17.5	17.8	26.6
87	9.2	8.2	12.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#44

Trichloroethene

Concen: 5.084 ug/l




RT: 9.353 min Scan# 1258


Delta R.T. 0.000 min

Lab File: VN083139.D

Acq: 07 Aug 2024 12:10

Tgt Ion:	130	Resp:	11830
Ion	Ratio	Lower	Upper
130	100		
95	93.4	0.0	197.8





#45

1,2-Dichloropropane

Concen: 4.878 ug/l

RT: 9.624 min Scan# 11320

Delta R.T. 0.000 min

Lab File: VN083139.D

Acq: 07 Aug 2024 12:10

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005

Tgt Ion: 63 Resp: 11320

Ion Ratio Lower Upper

63 100





65 35.7 24.4 36.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#46

Dibromomethane

Concen: 5.079 ug/l

RT: 9.712 min Scan# 1319

Delta R.T. 0.000 min

Lab File: VN083139.D

Acq: 07 Aug 2024 12:10




Tgt Ion: 93 Resp: 8436


Ion Ratio Lower Upper

93 100

95 82.9 65.8 98.6

174 92.2 71.7 107.5





#47

Bromodichloromethane

Concen: 5.040 ug/l

RT: 9.888 min Scan# 1349

Delta R.T. 0.000 min

Lab File: VN083139.D

Acq: 07 Aug 2024 12:10

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005





Tgt Ion:	83	Resp:	1879
Ion	Ratio	Lower	Upper
83	100		
85	62.3	51.1	76.7
127	10.3	7.2	10.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#48

Methyl methacrylate

Concen: 4.840 ug/l




RT: 9.682 min Scan# 1314


Delta R.T. 0.000 min

Lab File: VN083139.D

Acq: 07 Aug 2024 12:10

Tgt Ion:	41	Resp:	14704
Ion	Ratio	Lower	Upper
41	100		
69	79.2	56.3	84.5
39	63.9	50.3	75.5





#49

1,4-Dioxane

Concen: 95.593 ug/l

RT: 9.694 min Scan# 1316

Delta R.T. 0.000 min

Lab File: VN083139.D

Acq: 07 Aug 2024 12:10

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005

Tgt Ion: 88 Resp: 5239

Ion Ratio Lower Upper

88 100

43 0.0 27.8 41.8





58 79.3 59.4 89.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#50

Toluene-d8

Concen: 5.326 ug/l

RT: 10.565 min Scan# 1464

Delta R.T. 0.000 min

Lab File: VN083139.D




Acq: 07 Aug 2024 12:10


Tgt Ion: 98 Resp: 43104

Ion Ratio Lower Upper

98 100


100 65.2 51.5 77.3





#51
4-Methyl-2-Pentanone
Concen: 24.946 ug/l
RT: 10.447 min Scan# 1444
Delta R.T. 0.000 min
Lab File: VN083139.D
Acq: 07 Aug 2024 12:10




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005



Tgt Ion: 43 Resp: 86679
Ion Ratio Lower Upper
43 100
58 39.3 28.2 42.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#52
Toluene
Concen: 4.981 ug/l
RT: 10.629 min Scan# 1475
Delta R.T. 0.000 min
Lab File: VN083139.D
Acq: 07 Aug 2024 12:10


Tgt Ion: 92 Resp: 30769
Ion Ratio Lower Upper
92 100
91 172.3 139.4 209.0





#53
t-1,3-Dichloropropene
Concen: 4.950 ug/l
RT: 10.835 min Scan# 1510
Delta R.T. 0.000 min
Lab File: VN083139.D
Acq: 07 Aug 2024 12:10




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005



Tgt Ion: 75 Resp: 18964
Ion Ratio Lower Upper
75 100
77 32.0 25.7 38.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#54
cis-1,3-Dichloropropene
Concen: 5.014 ug/l
RT: 10.312 min Scan# 1421
Delta R.T. 0.000 min
Lab File: VN083139.D
Acq: 07 Aug 2024 12:10


Tgt Ion: 75 Resp: 20431
Ion Ratio Lower Upper
75 100
77 33.8 24.3 36.5
39 57.5 50.5 75.7





#55
1,1,2-Trichloroethane
Concen: 4.930 ug/l
RT: 11.018 min Scan# 1541
Delta R.T. 0.000 min
Lab File: VN083139.D
Acq: 07 Aug 2024 12:10




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005



Tgt Ion	Ratio	Lower	Upper
97	100		
83	91.7	71.4	107.0
85	54.2	45.8	68.6
99	65.3	52.0	78.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#56
Ethyl methacrylate
Concen: 4.951 ug/l
RT: 10.876 min Scan# 1517
Delta R.T. 0.000 min
Lab File: VN083139.D
Acq: 07 Aug 2024 12:10

Tgt Ion	Ratio	Lower	Upper
69	100		
41	74.3	63.4	95.2
39	39.8	37.4	56.0





#57

1,3-Dichloropropane

Concen: 4.799 ug/l

RT: 11.165 min Scan# 1566

Delta R.T. 0.000 min

Lab File: VN083139.D

Acq: 07 Aug 2024 12:10

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC005

Tgt Ion: 76 Resp: 18934

Ion Ratio Lower Upper

76 100





78 35.3 25.6 38.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#58

2-Chloroethyl Vinyl ether

Concen: 24.676 ug/l

RT: 10.159 min Scan# 1395

Delta R.T. 0.000 min

Lab File: VN083139.D




Acq: 07 Aug 2024 12:10


Tgt Ion: 63 Resp: 43528

Ion Ratio Lower Upper

63 100


106 28.8 21.4 32.0





#59
2-Hexanone
Concen: 24.870 ug/l
RT: 11.194 min Scan# 1571
Delta R.T. 0.000 min
Lab File: VN083139.D
Acq: 07 Aug 2024 12:10




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005



Tgt Ion: 43 Resp: 66854
Ion Ratio Lower Upper
43 100
58 52.9 24.6 73.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#60
Dibromochloromethane
Concen: 4.972 ug/l
RT: 11.359 min Scan# 1599
Delta R.T. 0.000 min
Lab File: VN083139.D
Acq: 07 Aug 2024 12:10


Tgt Ion:129 Resp: 13307
Ion Ratio Lower Upper
129 100
127 75.6 39.2 117.6





#61
1,2-Dibromoethane
Concen: 4.955 ug/l
RT: 11.471 min Scan# 1618
Delta R.T. 0.000 min
Lab File: VN083139.D
Acq: 07 Aug 2024 12:10




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005



Tgt Ion:107 Resp: 1152
Ion Ratio Lower Upper
107 100
109 95.9 73.7 110.5




Manual Integrations
APPROVED


Reviewed By :John Carbone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#62
4-Bromofluorobenzene
Concen: 5.317 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. 0.000 min
Lab File: VN083139.D
Acq: 07 Aug 2024 12:10


Tgt Ion: 95 Resp: 16776
Ion Ratio Lower Upper
95 100
174 68.1 0.0 159.2
176 67.0 0.0 147.6





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 1685
Delta R.T. 0.000 min
Lab File: VN083139.D
Acq: 07 Aug 2024 12:10




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005



Tgt Ion:117 Resp: 29351
Ion Ratio Lower Upper
117 100
82 62.1 47.5 71.3
119 32.0 25.6 38.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#64
Tetrachloroethene
Concen: 4.896 ug/l
RT: 11.106 min Scan# 1556
Delta R.T. 0.000 min
Lab File: VN083139.D
Acq: 07 Aug 2024 12:10

Tgt Ion:164 Resp: 9517
Ion Ratio Lower Upper
164 100
166 129.2 101.4 152.0
129 93.1 75.7 113.5
131 100.8 73.8 110.6





#65

Chlorobenzene

Concen: 4.951 ug/l

RT: 11.894 min Scan# 1

Delta R.T. 0.000 min

Lab File: VN083139.D

Acq: 07 Aug 2024 12:10

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005

Tgt Ion:112 Resp: 32114

Ion Ratio Lower Upper

112 100





114 34.9 26.2 39.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#66

1,1,1,2-Tetrachloroethane

Concen: 5.140 ug/l

RT: 11.959 min Scan# 1701

Delta R.T. -0.006 min

Lab File: VN083139.D

Acq: 07 Aug 2024 12:10




Tgt Ion:131 Resp: 11758


Ion Ratio Lower Upper

131 100

133 101.6 47.3 142.0

119 62.2 32.5 97.4





#67

Ethyl Benzene

Concen: 5.013 ug/l

RT: 11.965 min Scan# 1702

Delta R.T. 0.000 min

Lab File: VN083139.D

Acq: 07 Aug 2024 12:10

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005

Tgt Ion: 91 Resp: 59648

Ion Ratio Lower Upper

91 100





106 28.6 24.0 36.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#68

m/p-Xylenes

Concen: 10.000 ug/l

RT: 12.070 min Scan# 1720

Delta R.T. 0.000 min

Lab File: VN083139.D




Acq: 07 Aug 2024 12:10


Tgt Ion:106 Resp: 44572

Ion Ratio Lower Upper

106 100

91 214.3 166.1 249.1





#69

o-Xylene

Concen: 5.005 ug/l

RT: 12.400 min Scan# 1776

Delta R.T. 0.000 min

Lab File: VN083139.D

Acq: 07 Aug 2024 12:10

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005

Tgt Ion:106 Resp: 22000

Ion Ratio Lower Upper

106 100





91 221.7 111.3 333.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#70

Styrene

Concen: 5.014 ug/l

RT: 12.412 min Scan# 1778

Delta R.T. 0.000 min

Lab File: VN083139.D

Acq: 07 Aug 2024 12:10




Tgt Ion:104 Resp: 37023


Ion Ratio Lower Upper

104 100

78 54.3 41.6 62.4


103 55.4 44.0 66.0





#71
Bromoform
Concen: 4.665 ug/l
RT: 12.582 min Scan# 1806
Delta R.T. 0.006 min
Lab File: VN083139.D
Acq: 07 Aug 2024 12:10




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005



Tgt Ion:173 Resp: 808
Ion Ratio Lower Upper
173 100
175 55.6 24.0 72.0
254 0.0 0.0 0.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.794 min Scan# 2013
Delta R.T. 0.000 min
Lab File: VN083139.D
Acq: 07 Aug 2024 12:10


Tgt Ion:152 Resp: 133537
Ion Ratio Lower Upper
152 100
115 64.0 30.6 91.6
150 158.4 0.0 348.6





#73
Isopropylbenzene
Concen: 5.080 ug/l
RT: 12.694 min Scan# 1826
Delta R.T. 0.000 min
Lab File: VN083139.D
Acq: 07 Aug 2024 12:10




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005



Tgt Ion: 105 Resp: 5674
Ion Ratio Lower Upper
105 100
120 25.7 12.9 38.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#74
N-ethyl acetate
Concen: 5.014 ug/l
RT: 12.494 min Scan# 1792
Delta R.T. 0.000 min
Lab File: VN083139.D
Acq: 07 Aug 2024 12:10


Tgt Ion: 43 Resp: 27393
Ion Ratio Lower Upper
43 100
70 41.4 29.8 44.6
55 24.3 18.7 28.1
61 23.4 17.4 26.0





#75
1,1,2,2-Tetrachloroethane
Concen: 5.227 ug/l
RT: 12.935 min Scan# 1867
Delta R.T. 0.000 min
Lab File: VN083139.D
Acq: 07 Aug 2024 12:10




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005



Tgt Ion: 83 Resp: 16514
Ion Ratio Lower Upper
83 100
131 8.8 5.0 14.9
85 66.4 31.9 95.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#76
1,2,3-Trichloropropane
Concen: 4.927 ug/l m
RT: 12.994 min Scan# 1877
Delta R.T. 0.000 min
Lab File: VN083139.D
Acq: 07 Aug 2024 12:10


Tgt Ion: 75 Resp: 14463
Ion Ratio Lower Upper
75 100
77 212.4 110.9 332.6





#77
Bromobenzene
Concen: 5.249 ug/l
RT: 12.982 min Scan# 1875
Delta R.T. 0.000 min
Lab File: VN083139.D
Acq: 07 Aug 2024 12:10




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005



Tgt Ion:156 Resp: 13024
Ion Ratio Lower Upper
156 100
77 235.8 111.3 333.8
158 95.0 48.3 144.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#78
n-propylbenzene
Concen: 5.035 ug/l
RT: 13.035 min Scan# 1884
Delta R.T. 0.000 min
Lab File: VN083139.D
Acq: 07 Aug 2024 12:10

Tgt Ion: 91 Resp: 64751
Ion Ratio Lower Upper
91 100
120 22.4 10.9 32.9





#79

2-Chlorotoluene

Concen: 5.050 ug/l

RT: 13.123 min Scan# 1899

Delta R.T. 0.000 min

Lab File: VN083139.D

Acq: 07 Aug 2024 12:10

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005

Tgt Ion: 91 Resp: 41170

Ion Ratio Lower Upper

91 100





126 31.6 16.2 48.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#80

1,3,5-Trimethylbenzene

Concen: 5.088 ug/l

RT: 13.170 min Scan# 1907

Delta R.T. 0.000 min

Lab File: VN083139.D




Acq: 07 Aug 2024 12:10


Tgt Ion:105 Resp: 47581

Ion Ratio Lower Upper

105 100


120 46.5 24.3 72.8





#81
trans-1,4-Dichloro-2-butene
Concen: 5.254 ug/l
RT: 12.735 min Scan# 1833
Delta R.T. 0.000 min
Lab File: VN083139.D
Acq: 07 Aug 2024 12:10




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005



Tgt Ion: 75 Resp: 7079
Ion Ratio Lower Upper
75 100
53 100.6 88.4 132.6
89 41.6 36.3 54.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#82
4-Chlorotoluene
Concen: 5.030 ug/l
RT: 13.223 min Scan# 1916
Delta R.T. 0.000 min
Lab File: VN083139.D
Acq: 07 Aug 2024 12:10

Tgt Ion: 91 Resp: 41137
Ion Ratio Lower Upper
91 100
126 30.6 16.2 48.6









#83
tert-Butylbenzene
Concen: 5.159 ug/l
RT: 13.435 min Scan# 1953
Delta R.T. -0.006 min
Lab File: VN083139.D
Acq: 07 Aug 2024 12:10

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005

Tgt Ion	Ratio	Lower	Upper
119	100		
91	70.0	35.0	105.1
134	26.6	13.3	39.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#84
1,2,4-Trimethylbenzene
Concen: 5.090 ug/l
RT: 13.482 min Scan# 1960
Delta R.T. 0.000 min
Lab File: VN083139.D
Acq: 07 Aug 2024 12:10


Tgt Ion	Ratio	Lower	Upper
105	100		
120	42.3	21.9	65.8





#85
sec-Butylbenzene
Concen: 5.040 ug/l
RT: 13.617 min Scan# 1983
Delta R.T. 0.000 min
Lab File: VN083139.D
Acq: 07 Aug 2024 12:10




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005



Tgt Ion:105 Resp: 5694
Ion Ratio Lower Upper
105 100
134 18.5 9.8 29.3




Manual Integrations
APPROVED


Reviewed By :John Carbone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#86
p-Isopropyltoluene
Concen: 5.108 ug/l
RT: 13.729 min Scan# 2002
Delta R.T. 0.000 min
Lab File: VN083139.D
Acq: 07 Aug 2024 12:10

Tgt Ion:119 Resp: 47654
Ion Ratio Lower Upper
119 100
134 24.9 13.0 39.0
91 26.0 12.3 36.9





#87

1,3-Dichlorobenzene

Concen: 4.887 ug/l

RT: 13.735 min Scan# 2003

Delta R.T. 0.000 min

Lab File: VN083139.D

Acq: 07 Aug 2024 12:10

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC005





Tgt Ion:	146	Resp:	22814
Ion Ratio	Lower	Upper	
146	100		
111	43.0	20.6	61.9
148	64.7	31.8	95.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#88

1,4-Dichlorobenzene

Concen: 5.093 ug/l




RT: 13.812 min Scan# 2016


Delta R.T. 0.000 min

Lab File: VN083139.D

Acq: 07 Aug 2024 12:10

Tgt Ion:	146	Resp:	23971
Ion Ratio	Lower	Upper	
146	100		
111	47.6	20.1	60.3
148	65.9	32.2	96.6





#89

n-Butylbenzene

Concen: 4.983 ug/l

RT: 14.053 min Scan# 2058

Delta R.T. -0.006 min

Lab File: VN083139.D

Acq: 07 Aug 2024 12:10

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005





Tgt Ion:	91	Resp:	4028
Ion	Ratio	Lower	Upper
91	100		
92	52.3	26.1	78.3
134	25.3	12.1	36.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#90

Hexachloroethane

Concen: 5.158 ug/l




RT: 14.335 min Scan# 2105


Delta R.T. 0.000 min

Lab File: VN083139.D

Acq: 07 Aug 2024 12:10


Tgt Ion:	117	Resp:	9298
Ion	Ratio	Lower	Upper
117	100		
201	65.4	35.8	107.3





#91
1,2-Dichlorobenzene
Concen: 5.117 ug/l
RT: 14.106 min Scan# 2066
Delta R.T. 0.000 min
Lab File: VN083139.D
Acq: 07 Aug 2024 12:10




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005



Tgt Ion:146 Resp: 23110
Ion Ratio Lower Upper
146 100
111 45.5 21.4 64.2
148 61.5 32.1 96.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#92
1,2-Dibromo-3-Chloropropane
Concen: 5.345 ug/l
RT: 14.717 min Scan# 2170
Delta R.T. 0.000 min
Lab File: VN083139.D
Acq: 07 Aug 2024 12:10


Tgt Ion: 75 Resp: 4097
Ion Ratio Lower Upper
75 100
155 68.1 36.6 109.8
157 88.3 46.9 140.6





#93
1,2,4-Trichlorobenzene
Concen: 4.710 ug/l
RT: 15.394 min Scan# 2285
Delta R.T. 0.000 min
Lab File: VN083139.D
Acq: 07 Aug 2024 12:10




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005



Tgt Ion:180 Resp: 1192
Ion Ratio Lower Upper
180 100
182 98.9 47.3 141.8
145 38.1 15.7 47.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#94
Hexachlorobutadiene
Concen: 5.282 ug/l
RT: 15.500 min Scan# 2303
Delta R.T. 0.000 min
Lab File: VN083139.D
Acq: 07 Aug 2024 12:10


Tgt Ion:225 Resp: 5946
Ion Ratio Lower Upper
225 100
223 57.3 31.9 95.7
227 60.5 32.5 97.5





#95
Naphthalene
Concen: 4.811 ug/l
RT: 15.641 min Scan# 2327
Delta R.T. 0.000 min
Lab File: VN083139.D
Acq: 07 Aug 2024 12:10




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005



Tgt Ion:128 Resp: 4313
Ion Ratio Lower Upper
128 100
127 13.0 10.6 16.0
129 10.8 8.6 12.8




Manual Integrations
APPROVED

Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#96
1,2,3-Trichlorobenzene
Concen: 4.905 ug/l
RT: 15.841 min Scan# 2361
Delta R.T. 0.000 min
Lab File: VN083139.D
Acq: 07 Aug 2024 12:10

Tgt Ion:180 Resp: 12285
Ion Ratio Lower Upper
180 100
182 91.5 48.9 146.8
145 35.5 16.8 50.4



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080724\
 Data File : VN083140.D
 Acq On : 07 Aug 2024 12:34
 Operator : JC\MD
 Sample : VSTDICC001
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC001

Manual Integrations
 APPROVED

Reviewed By : John Carlone 08/08/2024
 Supervised By : Mahesh Dadoda 08/09/2024

Quant Time: Aug 08 06:19:42 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 05:50:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	202290	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	347394	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	292647	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	122922	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	0.000	65	0d	0.000	ug/l	
Spiked Amount	50.000	Range	74 - 125	Recovery	=	0.000%#
35) Dibromofluoromethane	0.000	113	0d	0.000	ug/l	
Spiked Amount	50.000	Range	75 - 124	Recovery	=	0.000%#
50) Toluene-d8	0.000	98	0d	0.000	ug/l	
Spiked Amount	50.000	Range	86 - 113	Recovery	=	0.000%#
62) 4-Bromofluorobenzene	0.000	95	0d	0.000	ug/l	
Spiked Amount	50.000	Range	64 - 133	Recovery	=	0.000%#
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	2.124	85	2193	0.956	ug/l	84
3) Chloromethane	2.365	50	2331	0.993	ug/l	96
4) Vinyl Chloride	2.512	62	2457	1.025	ug/l	94
6) Chloroethane	3.124	64	1763	1.176	ug/l #	54
7) Trichlorofluoromethane	3.495	101	3645	0.921	ug/l	88
8) Diethyl Ether	3.965	74	1368	0.929	ug/l	95
9) 1,1,2-Trichlorotrifluo...	4.371	101	2064	0.946	ug/l #	88
12) 1,1-Dichloroethene	4.342	96	2328	1.038	ug/l #	72
14) Allyl chloride	5.018	41	4894	1.155	ug/l #	75
15) Acrylonitrile	5.724	53	6198	5.036	ug/l	94
16) Acetone	4.442	43	5369	4.765	ug/l	95
17) Carbon Disulfide	4.718	76	7603	1.159	ug/l	99
18) Methyl Acetate	5.036	43	3987	1.188	ug/l #	89
19) Methyl tert-butyl Ether	5.795	73	7900	0.976	ug/l	96
20) Methylene Chloride	5.283	84	3255	1.255	ug/l #	63
21) trans-1,2-Dichloroethene	5.783	96	2425	1.046	ug/l #	75
22) Diisopropyl ether	6.677	45	7871	0.988	ug/l	94
23) Vinyl Acetate	6.606	43	39044	4.783	ug/l #	94
24) 1,1-Dichloroethane	6.571	63	4189	0.965	ug/l #	83
25) 2-Butanone	7.495	43	9164	5.297	ug/l	93
26) 2,2-Dichloropropane	7.489	77	3741	0.928	ug/l	95
27) cis-1,2-Dichloroethene	7.495	96	2919	1.044	ug/l	90
28) Bromochloromethane	7.812	49	1883	1.061	ug/l #	92
29) Tetrahydrofuran	7.842	42	5490	4.909	ug/l #	87
30) Chloroform	7.971	83	4457	0.988	ug/l	95
32) 1,1,1-Trichloroethane	8.171	97	4076	0.955	ug/l #	49
36) 1,1-Dichloropropene	8.377	75	3393	1.034	ug/l	97
37) Ethyl Acetate	7.565	43	4617	1.257	ug/l #	90
38) Carbon Tetrachloride	8.359	117	3422	0.926	ug/l #	84
39) Methylcyclohexane	9.600	83	3974	0.986	ug/l	87
40) Benzene	8.606	78	9334	0.955	ug/l	99
41) Methacrylonitrile	7.783	41	2618	1.254	ug/l #	85
42) 1,2-Dichloroethane	8.677	62	3457	0.971	ug/l	82
43) Isopropyl Acetate	8.689	43	14981	0.753	ug/l #	80
44) Trichloroethene	9.359	130	2151	0.925	ug/l	98
45) 1,2-Dichloropropane	9.618	63	2194	0.946	ug/l	93

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080724\
 Data File : VN083140.D
 Acq On : 07 Aug 2024 12:34
 Operator : JC\MD
 Sample : VSTDICC001
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC001

Manual Integrations
 APPROVED

Reviewed By : John Carlone 08/08/2024
 Supervised By : Mahesh Dadoda 08/09/2024

Quant Time: Aug 08 06:19:42 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 05:50:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) Dibromomethane	9.700	93	1513	0.911	ug/l	95
47) Bromodichloromethane	9.894	83	3783	1.015	ug/l #	83
48) Methyl methacrylate	9.683	41	3518	1.159	ug/l	96
49) 1,4-Dioxane	9.700	88	1089	19.880	ug/l #	77
51) 4-Methyl-2-Pentanone	10.447	43	15872	4.570	ug/l	90
52) Toluene	10.630	92	5723	0.927	ug/l	97
53) t-1,3-Dichloropropene	10.835	75	3260	0.851	ug/l	100
54) cis-1,3-Dichloropropene	10.312	75	3797	0.932	ug/l #	87
55) 1,1,2-Trichloroethane	11.018	97	1963	0.887	ug/l	96
56) Ethyl methacrylate	10.871	69	3843	0.921	ug/l #	90
57) 1,3-Dichloropropane	11.165	76	3837	0.973	ug/l	99
58) 2-Chloroethyl Vinyl ether	10.159	63	8453	4.794	ug/l	93
59) 2-Hexanone	11.200	43	12761	4.749	ug/l	100
60) Dibromochloromethane	11.353	129	2283	0.853	ug/l	99
61) 1,2-Dibromoethane	11.465	107	2200	0.947	ug/l	98
64) Tetrachloroethene	11.106	164	1872	0.966	ug/l	93
65) Chlorobenzene	11.888	112	6259	0.968	ug/l	93
66) 1,1,1,2-Tetrachloroethane	11.965	131	2168	0.951	ug/l #	63
67) Ethyl Benzene	11.965	91	11455	0.966	ug/l	98
68) m/p-Xylenes	12.065	106	8266	1.860	ug/l	94
69) o-Xylene	12.400	106	4113	0.938	ug/l	98
70) Styrene	12.412	104	6441	0.875	ug/l	90
71) Bromoform	12.577	173	1561	0.903	ug/l #	88
73) Isopropylbenzene	12.694	105	10459	1.017	ug/l	99
74) N-amyl acetate	12.494	43	5755	1.144	ug/l #	92
75) 1,1,2,2-Tetrachloroethane	12.941	83	3034	1.043	ug/l	99
76) 1,2,3-Trichloropropane	12.988	75	3031m	1.122	ug/l	
77) Bromobenzene	12.977	156	2212	0.969	ug/l	86
78) n-propylbenzene	13.029	91	11800	0.997	ug/l	99
79) 2-Chlorotoluene	13.124	91	7735	1.031	ug/l	90
80) 1,3,5-Trimethylbenzene	13.176	105	8552	0.994	ug/l	95
82) 4-Chlorotoluene	13.224	91	8018	1.065	ug/l	93
83) tert-Butylbenzene	13.435	119	7628	1.001	ug/l	98
84) 1,2,4-Trimethylbenzene	13.482	105	8397	0.968	ug/l	100
85) sec-Butylbenzene	13.618	105	10411	1.001	ug/l	96
86) p-Isopropyltoluene	13.735	119	8247	0.960	ug/l	99
87) 1,3-Dichlorobenzene	13.735	146	4598	1.070	ug/l	97
88) 1,4-Dichlorobenzene	13.812	146	4578m	1.057	ug/l	
89) n-Butylbenzene	14.053	91	7008	0.942	ug/l	100
90) Hexachloroethane	14.335	117	1524	0.918	ug/l	93
91) 1,2-Dichlorobenzene	14.106	146	4279	1.029	ug/l	95
92) 1,2-Dibromo-3-Chloropr...	14.718	75	763	1.081	ug/l	78
93) 1,2,4-Trichlorobenzene	15.394	180	2493	1.070	ug/l	97
94) Hexachlorobutadiene	15.512	225	1181	1.140	ug/l	92
95) Naphthalene	15.641	128	8430	1.021	ug/l #	96
96) 1,2,3-Trichlorobenzene	15.835	180	2389	1.036	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080724\
Data File : VN083140.D
Acq On : 07 Aug 2024 12:34
Operator : JC\MD
Sample : VSTDIC001
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1


Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001

Manual Integrations

Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024


Quant Time: Aug 08 06:19:42 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 05:50:26 2024
Response via : Initial Calibration





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 1066
Delta R.T. -0.000 min
Lab File: VN083140.D
Acq: 07 Aug 2024 12:34




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001



Tgt Ion:168 Resp: 202290
Ion Ratio Lower Upper
168 100
99 64.1 48.2 72.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#2
Dichlorodifluoromethane
Concen: 0.956 ug/l
RT: 2.124 min Scan# 29
Delta R.T. 0.000 min
Lab File: VN083140.D
Acq: 07 Aug 2024 12:34


Tgt Ion: 85 Resp: 2193
Ion Ratio Lower Upper
85 100
87 40.2 15.7 47.0





#3
Chloromethane
Concen: 0.993 ug/l
RT: 2.365 min Scan# 70
Delta R.T. 0.006 min
Lab File: VN083140.D
Acq: 07 Aug 2024 12:34




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001



Tgt Ion: 50 Resp: 233
Ion Ratio Lower Upper
50 100
52 33.3 25.0 37.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#4
Vinyl Chloride
Concen: 1.025 ug/l
RT: 2.512 min Scan# 95
Delta R.T. 0.000 min
Lab File: VN083140.D
Acq: 07 Aug 2024 12:34


Tgt Ion: 62 Resp: 2457
Ion Ratio Lower Upper
62 100
64 34.7 25.0 37.6





#6
Chloroethane
Concen: 1.176 ug/l
RT: 3.124 min Scan# 199
Delta R.T. 0.012 min
Lab File: VN083140.D
Acq: 07 Aug 2024 12:34




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001



Tgt Ion: 64 Resp: 176
Ion Ratio Lower Upper
64 100
66 7.1 26.6 40.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#7
Trichlorofluoromethane
Concen: 0.921 ug/l
RT: 3.495 min Scan# 262
Delta R.T. -0.000 min
Lab File: VN083140.D
Acq: 07 Aug 2024 12:34

Tgt Ion:101 Resp: 3645
Ion Ratio Lower Upper
101 100
103 75.0 52.7 79.1









#8
Diethyl Ether
Concen: 0.929 ug/l
RT: 3.965 min Scan# 341
Delta R.T. 0.006 min
Lab File: VN083140.D
Acq: 07 Aug 2024 12:34

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001

Tgt Ion: 74 Resp: 1368
Ion Ratio Lower Upper
74 100
45 105.1 55.5 166.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#9
1,1,2-Trichlorotrifluoroethane
Concen: 0.946 ug/l
RT: 4.371 min Scan# 411
Delta R.T. 0.006 min
Lab File: VN083140.D
Acq: 07 Aug 2024 12:34

Tgt Ion:101 Resp: 2064
Ion Ratio Lower Upper
101 100
85 37.3 38.3 57.5#
151 82.8 59.3 88.9





#12

1,1-Dichloroethene

Concen: 1.038 ug/l

RT: 4.342 min Scan# 406

Delta R.T. 0.000 min

Lab File: VN083140.D

Acq: 07 Aug 2024 12:34

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001

Tgt Ion: 96 Resp: 2328

Ion Ratio Lower Upper

96 100

61 135.3 149.7 224.5





98 57.8 50.1 75.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#14

Allyl chloride

Concen: 1.155 ug/l

RT: 5.018 min Scan# 521

Delta R.T. 0.000 min

Lab File: VN083140.D

Acq: 07 Aug 2024 12:34




Tgt Ion: 41 Resp: 4894


Ion Ratio Lower Upper

41 100

39 65.9 67.8 101.6#

76 11.9 25.8 38.8#





#15

Acrylonitrile

Concen: 5.036 ug/l

RT: 5.724 min Scan# 641

Delta R.T. 0.006 min

Lab File: VN083140.D

Acq: 07 Aug 2024 12:34

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC001





Tgt Ion: 53	Resp: 6198
Ion Ratio	Lower Upper
53	100
52	77.5 66.6 100.0
51	34.3 30.1 45.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#16

Acetone

Concen: 4.765 ug/l




RT: 4.442 min Scan# 423


Delta R.T. 0.012 min

Lab File: VN083140.D

Acq: 07 Aug 2024 12:34

Tgt Ion: 43	Resp: 5369
Ion Ratio	Lower Upper
43	100
58	29.9 21.8 32.6





#17

Carbon Disulfide

Concen: 1.159 ug/l

RT: 4.718 min Scan# 41

Delta R.T. 0.006 min

Lab File: VN083140.D

Acq: 07 Aug 2024 12:34

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC001

Tgt Ion: 76 Resp: 760

Ion Ratio Lower Upper

76 100





78 9.3 7.2 10.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#18

Methyl Acetate

Concen: 1.188 ug/l

RT: 5.036 min Scan# 524

Delta R.T. 0.012 min

Lab File: VN083140.D




Acq: 07 Aug 2024 12:34


Tgt Ion: 43 Resp: 3987

Ion Ratio Lower Upper

43 100

74 15.5 16.3 24.5#









#19
Methyl tert-butyl Ether
Concen: 0.976 ug/l
RT: 5.795 min Scan# 61
Delta R.T. 0.000 min
Lab File: VN083140.D
Acq: 07 Aug 2024 12:34

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001

Tgt Ion: 73 Resp: 7900
Ion Ratio Lower Upper
73 100
57 20.5 17.9 26.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#20
Methylene Chloride
Concen: 1.255 ug/l
RT: 5.283 min Scan# 566
Delta R.T. 0.006 min
Lab File: VN083140.D
Acq: 07 Aug 2024 12:34


Tgt Ion: 84 Resp: 3255
Ion Ratio Lower Upper
84 100
49 90.3 119.6 179.4#
51 22.8 34.8 52.2#
86 52.9 52.9 79.3#





#21
trans-1,2-Dichloroethene
Concen: 1.046 ug/l
RT: 5.783 min Scan# 61
Delta R.T. -0.012 min
Lab File: VN083140.D
Acq: 07 Aug 2024 12:34




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001



Tgt Ion: 96 Resp: 242
Ion Ratio Lower Upper
96 100
61 109.9 123.4 185.2
98 60.5 49.3 73.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#22
Diisopropyl ether
Concen: 0.988 ug/l
RT: 6.677 min Scan# 803
Delta R.T. 0.006 min
Lab File: VN083140.D
Acq: 07 Aug 2024 12:34

Tgt Ion: 45 Resp: 7871
Ion Ratio Lower Upper
45 100
43 61.5 44.0 66.0
87 24.9 19.7 29.5
59 10.0 8.5 12.7





#23

Vinyl Acetate

Concen: 4.783 ug/l

RT: 6.606 min Scan# 791

Delta R.T. 0.000 min

Lab File: VN083140.D

Acq: 07 Aug 2024 12:34

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001

Tgt Ion: 43 Resp: 3904

Ion Ratio Lower Upper

43 100





86 11.4 7.3 10.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#24

1,1-Dichloroethane

Concen: 0.965 ug/l

RT: 6.571 min Scan# 785

Delta R.T. 0.000 min

Lab File: VN083140.D

Acq: 07 Aug 2024 12:34




Tgt Ion: 63 Resp: 4189


Ion Ratio Lower Upper

63 100

98 0.0 3.3 9.9#


100 0.0 2.0 6.0#





#25
2-Butanone
Concen: 5.297 ug/l
RT: 7.495 min Scan# 942
Delta R.T. 0.012 min
Lab File: VN083140.D
Acq: 07 Aug 2024 12:34




Instrument :
MSVOA_N
Client Sample Id :
VSTDICC001



Tgt Ion: 43 Resp: 9164
Ion Ratio Lower Upper
43 100
72 25.9 18.2 27.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#26
2,2-Dichloropropane
Concen: 0.928 ug/l
RT: 7.489 min Scan# 941
Delta R.T. 0.000 min
Lab File: VN083140.D
Acq: 07 Aug 2024 12:34

Tgt Ion: 77 Resp: 3741
Ion Ratio Lower Upper
77 100
97 22.9 10.3 30.9





#27

cis-1,2-Dichloroethene

Concen: 1.044 ug/l

RT: 7.495 min Scan# 941

Delta R.T. 0.006 min

Lab File: VN083140.D

Acq: 07 Aug 2024 12:34

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001





Tgt Ion:	96	Resp:	2919
Ion	Ratio	Lower	Upper
96	100		
61	146.7	0.0	329.6
98	65.4	0.0	128.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#28

Bromochloromethane

Concen: 1.061 ug/l




RT: 7.812 min Scan# 996


Delta R.T. 0.000 min

Lab File: VN083140.D

Acq: 07 Aug 2024 12:34

Tgt Ion:	49	Resp:	1883
Ion	Ratio	Lower	Upper
49	100		
129	0.0	0.0	3.8
130	69.5	50.5	75.7





#29

Tetrahydrofuran

Concen: 4.909 ug/l

RT: 7.842 min Scan# 1000

Delta R.T. 0.006 min

Lab File: VN083140.D

Acq: 07 Aug 2024 12:34

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001





Tgt Ion:	42	Resp:	5490
Ion Ratio	Lower	Upper	
42	100		
72	46.5	30.0	45.0
71	41.2	28.3	42.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#30

Chloroform

Concen: 0.988 ug/l




RT: 7.971 min Scan# 1023


Delta R.T. 0.006 min

Lab File: VN083140.D

Acq: 07 Aug 2024 12:34

Tgt Ion:	83	Resp:	4457
Ion Ratio	Lower	Upper	
83	100		
85	59.8	50.9	76.3





#32

1,1,1-Trichloroethane

Concen: 0.955 ug/l

RT: 8.171 min Scan# 1057

Delta R.T. 0.000 min

Lab File: VN083140.D

Acq: 07 Aug 2024 12:34

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001

Tgt Ion: 97 Resp: 4070

Ion Ratio Lower Upper

97 100

99 0.0 52.0 78.0





61 44.6 42.1 63.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#33

1,2-Dichloroethane-d4

Concen: 0.000 ug/l

Expected RT: 8.58 min

Lab File: VN083140.D


Acq: 07 Aug 2024 12:34


Tgt Ion: 65

Sig Exp Ratio

65 100

67 52.2





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.100 min Scan# 11

Delta R.T. 0.000 min

Lab File: VN083140.D

Acq: 07 Aug 2024 12:34

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001

Tgt Ion:114 Resp: 347394

Ion Ratio Lower Upper

114 100

63 22.6 0.0 44.6





88 16.9 0.0 31.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#35

Dibromofluoromethane

Concen: 0.000 ug/l

Expected RT: 8.16 min

Lab File: VN083140.D

Acq: 07 Aug 2024 12:34


Tgt Ion: 113


Sig Exp Ratio

113 100

111 103.0

192 18.6





#36

1,1-Dichloropropene

Concen: 1.034 ug/l

RT: 8.377 min Scan# 1091

Delta R.T. 0.006 min

Lab File: VN083140.D

Acq: 07 Aug 2024 12:34

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001





Tgt Ion:	75	Resp:	339
Ion	Ratio	Lower	Upper
75	100		
110	32.0	16.3	48.9
77	33.3	24.6	37.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#37

Ethyl Acetate

Concen: 1.257 ug/l




RT: 7.565 min Scan# 954


Delta R.T. 0.006 min

Lab File: VN083140.D

Acq: 07 Aug 2024 12:34


Tgt Ion:	43	Resp:	4617
Ion	Ratio	Lower	Upper
43	100		
61	9.2	9.3	13.9#
70	3.8	7.0	10.6#





#38
Carbon Tetrachloride
Concen: 0.926 ug/l
RT: 8.359 min Scan# 1089
Delta R.T. 0.000 min
Lab File: VN083140.D
Acq: 07 Aug 2024 12:34




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001




Tgt Ion: 117 Resp: 342
Ion Ratio Lower Upper
117 100
119 79.9 74.9 112.3
121 42.0 24.3 36.5

Manual Integrations
APPROVED



Reviewed By : John Carlone 08/08/2024
Supervised By : Mahesh Dadoda 08/09/2024




#39
Methylcyclohexane
Concen: 0.986 ug/l
RT: 9.600 min Scan# 1300
Delta R.T. 0.000 min
Lab File: VN083140.D
Acq: 07 Aug 2024 12:34



Tgt Ion: 83 Resp: 3974
Ion Ratio Lower Upper
83 100
55 96.3 66.4 99.6
98 43.7 40.7 61.1





#40

Benzene

Concen: 0.955 ug/l

RT: 8.606 min Scan# 1131

Delta R.T. 0.000 min

Lab File: VN083140.D

Acq: 07 Aug 2024 12:34

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001

Tgt Ion: 78 Resp: 9334

Ion Ratio Lower Upper

78 100





77 24.2 19.0 28.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#41

Methacrylonitrile

Concen: 1.254 ug/l

RT: 7.783 min Scan# 991

Delta R.T. 0.006 min

Lab File: VN083140.D

Acq: 07 Aug 2024 12:34

Tgt Ion: 41 Resp: 2618




Ion Ratio Lower Upper


41 100

39 52.4 45.5 68.3

67 54.2 50.5 75.7

52 8.4 24.0 36.0#





#42

1,2-Dichloroethane

Concen: 0.971 ug/l

RT: 8.677 min Scan# 1142

Delta R.T. 0.006 min

Lab File: VN083140.D

Acq: 07 Aug 2024 12:34

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001





Tgt Ion	Ratio	Lower	Upper
62	100		
98	1.6	0.0	15.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#43

Isopropyl Acetate

Concen: 0.753 ug/l




RT: 8.689 min Scan# 1145


Delta R.T. 0.001 min

Lab File: VN083140.D

Acq: 07 Aug 2024 12:34


Tgt Ion	Ratio	Lower	Upper
43	100		
61	11.5	17.8	26.6#
87	4.4	8.2	12.2#





#44
Trichloroethene
Concen: 0.925 ug/l
RT: 9.359 min Scan# 11
Delta R.T. 0.006 min
Lab File: VN083140.D
Acq: 07 Aug 2024 12:34




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001




Tgt Ion:130 Resp: 215
Ion Ratio Lower Upper
130 100
95 96.7 0.0 197.8

Manual Integrations
APPROVED



Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024




#45
1,2-Dichloropropane
Concen: 0.946 ug/l
RT: 9.618 min Scan# 1303
Delta R.T. -0.006 min
Lab File: VN083140.D
Acq: 07 Aug 2024 12:34



Tgt Ion: 63 Resp: 2194
Ion Ratio Lower Upper
63 100
65 34.2 24.4 36.6





#46

Dibromomethane

Concen: 0.911 ug/l

RT: 9.700 min Scan# 1319

Delta R.T. -0.012 min

Lab File: VN083140.D

Acq: 07 Aug 2024 12:34

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001





Tgt Ion:	93	Resp:	151
Ion	Ratio	Lower	Upper
93	100		
95	89.4	65.8	98.6
174	87.3	71.7	107.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#47

Bromodichloromethane

Concen: 1.015 ug/l




RT: 9.894 min Scan# 1350


Delta R.T. 0.006 min

Lab File: VN083140.D

Acq: 07 Aug 2024 12:34

Tgt Ion:	83	Resp:	3783
Ion	Ratio	Lower	Upper
83	100		
85	51.4	51.1	76.7
127	0.0	7.2	10.8





#48

Methyl methacrylate

Concen: 1.159 ug/l

RT: 9.683 min Scan# 1314

Delta R.T. 0.000 min

Lab File: VN083140.D

Acq: 07 Aug 2024 12:34

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001





Tgt Ion:	41	Resp:	3518
Ion	Ratio	Lower	Upper
41	100		
69	70.6	56.3	84.5
39	57.3	50.3	75.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#49

1,4-Dioxane

Concen: 19.880 ug/l




RT: 9.700 min Scan# 1317


Delta R.T. 0.006 min

Lab File: VN083140.D

Acq: 07 Aug 2024 12:34

Tgt Ion:	88	Resp:	1089
Ion	Ratio	Lower	Upper
88	100		
43	0.0	27.8	41.8#
58	79.4	59.4	89.0





#50
Toluene-d8
Concen: 0.000 ug/l
Expected RT: 10.56 min



Lab File: VN083140.D
Acq: 07 Aug 2024 12:34

Tgt Ion: 98
Sig Exp Ratio
98 100
100 64.4

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#51
4-Methyl-2-Pentanone
Concen: 4.570 ug/l
RT: 10.447 min Scan# 1444
Delta R.T. 0.000 min
Lab File: VN083140.D
Acq: 07 Aug 2024 12:34


Tgt Ion: 43 Resp: 15872
Ion Ratio Lower Upper
43 100
58 41.1 28.2 42.4





#52
Toluene
Concen: 0.927 ug/l
RT: 10.630 min Scan# 1475
Delta R.T. 0.000 min
Lab File: VN083140.D
Acq: 07 Aug 2024 12:34




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001




Tgt Ion: 92 Resp: 572.0
Ion Ratio Lower Upper
92 100
91 178.1 139.4 209.0

Manual Integrations
APPROVED



Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024




#53
t-1,3-Dichloropropene
Concen: 0.851 ug/l
RT: 10.835 min Scan# 1510
Delta R.T. 0.000 min
Lab File: VN083140.D
Acq: 07 Aug 2024 12:34




Tgt Ion: 75 Resp: 3260
Ion Ratio Lower Upper
75 100
77 32.0 25.7 38.5





#54
cis-1,3-Dichloropropene
Concen: 0.932 ug/l
RT: 10.312 min Scan# 1421
Delta R.T. 0.000 min
Lab File: VN083140.D
Acq: 07 Aug 2024 12:34




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001



Tgt Ion: 75 Resp: 379
Ion Ratio Lower Upper
75 100
77 32.5 24.3 36.5
39 49.0 50.5 75.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#55
1,1,2-Trichloroethane
Concen: 0.887 ug/l
RT: 11.018 min Scan# 1541
Delta R.T. 0.000 min
Lab File: VN083140.D
Acq: 07 Aug 2024 12:34


Tgt Ion: 97 Resp: 1963
Ion Ratio Lower Upper
97 100
83 88.2 71.4 107.0
85 47.2 45.8 68.6
99 65.4 52.0 78.0





#56
Ethyl methacrylate
Concen: 0.921 ug/l
RT: 10.871 min Scan# 1516
Delta R.T. -0.006 min
Lab File: VN083140.D
Acq: 07 Aug 2024 12:34




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001



Tgt Ion: 69 Resp: 384
Ion Ratio Lower Upper
69 100
41 83.6 63.4 95.2
39 33.7 37.4 56.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#57
1,3-Dichloropropane
Concen: 0.973 ug/l
RT: 11.165 min Scan# 1566
Delta R.T. 0.000 min
Lab File: VN083140.D
Acq: 07 Aug 2024 12:34

Tgt Ion: 76 Resp: 3837
Ion Ratio Lower Upper
76 100
78 31.3 25.6 38.4









#58
2-Chloroethyl Vinyl ether
Concen: 4.794 ug/l
RT: 10.159 min Scan# 1395
Delta R.T. 0.000 min
Lab File: VN083140.D
Acq: 07 Aug 2024 12:34

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001

Tgt Ion: 63 Resp: 845
Ion Ratio Lower Upper
63 100
106 30.3 21.4 32.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#59
2-Hexanone
Concen: 4.749 ug/l
RT: 11.200 min Scan# 1572
Delta R.T. 0.006 min
Lab File: VN083140.D
Acq: 07 Aug 2024 12:34


Tgt Ion: 43 Resp: 12761
Ion Ratio Lower Upper
43 100
58 49.3 24.6 73.8





#60
Dibromochloromethane
Concen: 0.853 ug/l
RT: 11.353 min Scan# 1598
Delta R.T. -0.006 min
Lab File: VN083140.D
Acq: 07 Aug 2024 12:34




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001




Tgt Ion:129 Resp: 2283
Ion Ratio Lower Upper
129 100
127 77.5 39.2 117.6

Manual Integrations
APPROVED



Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024




#61
1,2-Dibromoethane
Concen: 0.947 ug/l
RT: 11.465 min Scan# 1617
Delta R.T. -0.006 min
Lab File: VN083140.D
Acq: 07 Aug 2024 12:34



Tgt Ion:107 Resp: 2200
Ion Ratio Lower Upper
107 100
109 90.4 73.7 110.5





#62
4-Bromofluorobenzene
Concen: 0.000 ug/l
Expected RT: 12.85 min



Lab File: VN083140.D
Acq: 07 Aug 2024 12:34

Tgt Ion: 95
Sig Exp Ratio
95 100
174 79.6
176 73.8

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 1685
Delta R.T. 0.000 min
Lab File: VN083140.D
Acq: 07 Aug 2024 12:34


Tgt Ion:117 Resp: 292647
Ion Ratio Lower Upper
117 100
82 60.6 47.5 71.3
119 32.4 25.6 38.4





#64
Tetrachloroethene
Concen: 0.966 ug/l
RT: 11.106 min Scan# 1556
Delta R.T. 0.000 min
Lab File: VN083140.D
Acq: 07 Aug 2024 12:34




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001



Tgt Ion:164 Resp: 1872
Ion Ratio Lower Upper
164 100
166 127.0 101.4 152.0
129 112.2 75.7 113.5
131 96.5 73.8 110.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#65
Chlorobenzene
Concen: 0.968 ug/l
RT: 11.888 min Scan# 1689
Delta R.T. -0.006 min
Lab File: VN083140.D
Acq: 07 Aug 2024 12:34


Tgt Ion:112 Resp: 6259
Ion Ratio Lower Upper
112 100
114 29.0 26.2 39.2





#66
1,1,1,2-Tetrachloroethane
Concen: 0.951 ug/l
RT: 11.965 min Scan# 1702
Delta R.T. 0.000 min
Lab File: VN083140.D
Acq: 07 Aug 2024 12:34




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001



Tgt Ion: 131 Resp: 2168
Ion Ratio Lower Upper
131 100
133 100.6 47.3 142.0
119 0.0 32.5 97.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#67
Ethyl Benzene
Concen: 0.966 ug/l
RT: 11.965 min Scan# 1702
Delta R.T. 0.000 min
Lab File: VN083140.D
Acq: 07 Aug 2024 12:34


Tgt Ion: 91 Resp: 11455
Ion Ratio Lower Upper
91 100
106 28.7 24.0 36.0





#68
m/p-Xylenes
Concen: 1.860 ug/l
RT: 12.065 min Scan# 1719
Delta R.T. -0.006 min
Lab File: VN083140.D
Acq: 07 Aug 2024 12:34




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001



Tgt Ion:106 Resp: 826
Ion Ratio Lower Upper
106 100
91 217.0 166.1 249.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#69
o-Xylene
Concen: 0.938 ug/l
RT: 12.400 min Scan# 1776
Delta R.T. 0.000 min
Lab File: VN083140.D
Acq: 07 Aug 2024 12:34


Tgt Ion:106 Resp: 4113
Ion Ratio Lower Upper
106 100
91 226.2 111.3 333.9





#70
Styrene
Concen: 0.875 ug/l
RT: 12.412 min Scan# 1778
Delta R.T. 0.000 min
Lab File: VN083140.D
Acq: 07 Aug 2024 12:34




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001



Tgt Ion:104 Resp: 644
Ion Ratio Lower Upper
104 100
78 60.2 41.6 62.4
103 60.7 44.0 66.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#71
Bromoform
Concen: 0.903 ug/l
RT: 12.577 min Scan# 1806
Delta R.T. 0.000 min
Lab File: VN083140.D
Acq: 07 Aug 2024 12:34


Tgt Ion:173 Resp: 1561
Ion Ratio Lower Upper
173 100
175 40.2 24.0 72.0
254 0.0 0.0 0.0





#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.794 min Scan# 2013
Delta R.T. 0.000 min
Lab File: VN083140.D
Acq: 07 Aug 2024 12:34




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001



Tgt Ion:152 Resp: 12292
Ion Ratio Lower Upper
152 100
115 64.2 30.6 91.6
150 157.5 0.0 348.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#73
Isopropylbenzene
Concen: 1.017 ug/l
RT: 12.694 min Scan# 1826
Delta R.T. 0.000 min
Lab File: VN083140.D
Acq: 07 Aug 2024 12:34


Tgt Ion:105 Resp: 10459
Ion Ratio Lower Upper
105 100
120 25.1 12.9 38.6





#74
N-amy1 acetate
Concen: 1.144 ug/l
RT: 12.494 min Scan# 1792
Delta R.T. 0.000 min
Lab File: VN083140.D
Acq: 07 Aug 2024 12:34




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001



Tgt Ion: 43 Resp: 575
Ion Ratio Lower Upper
43 100
70 33.3 29.8 44.6
55 28.9 18.7 28.1
61 19.1 17.4 26.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#75
1,1,2,2-Tetrachloroethane
Concen: 1.043 ug/l
RT: 12.941 min Scan# 1868
Delta R.T. 0.006 min
Lab File: VN083140.D
Acq: 07 Aug 2024 12:34

Tgt Ion: 83 Resp: 3034
Ion Ratio Lower Upper
83 100
131 8.5 5.0 14.9
85 64.0 31.9 95.7









#76
1,2,3-Trichloropropane
Concen: 1.122 ug/l m
RT: 12.988 min Scan# 1876
Delta R.T. -0.006 min
Lab File: VN083140.D
Acq: 07 Aug 2024 12:34

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001

Tgt Ion: 75 Resp: 303
Ion Ratio Lower Upper
75 100
77 184.1 110.9 332.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#77
Bromobenzene
Concen: 0.969 ug/l
RT: 12.977 min Scan# 1874
Delta R.T. -0.006 min
Lab File: VN083140.D
Acq: 07 Aug 2024 12:34

Tgt Ion:156 Resp: 2212
Ion Ratio Lower Upper
156 100
77 252.2 111.3 333.8
158 101.5 48.3 144.9









#78
n-propylbenzene
Concen: 0.997 ug/l
RT: 13.029 min Scan# 1883
Delta R.T. -0.006 min
Lab File: VN083140.D
Acq: 07 Aug 2024 12:34

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001

Tgt Ion: 91 Resp: 11800
Ion Ratio Lower Upper
91 100
120 21.7 10.9 32.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#79
2-Chlorotoluene
Concen: 1.031 ug/l
RT: 13.124 min Scan# 1899
Delta R.T. 0.000 min
Lab File: VN083140.D
Acq: 07 Aug 2024 12:34


Tgt Ion: 91 Resp: 7735
Ion Ratio Lower Upper
91 100
126 26.9 16.2 48.6





#80
1,3,5-Trimethylbenzene
Concen: 0.994 ug/l
RT: 13.176 min Scan# 1907
Delta R.T. 0.006 min
Lab File: VN083140.D
Acq: 07 Aug 2024 12:34




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001



Tgt Ion:105 Resp: 8552
Ion Ratio Lower Upper
105 100
120 51.7 24.3 72.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#82
4-Chlorotoluene
Concen: 1.065 ug/l
RT: 13.224 min Scan# 1916
Delta R.T. 0.000 min
Lab File: VN083140.D
Acq: 07 Aug 2024 12:34

Tgt Ion: 91 Resp: 8018
Ion Ratio Lower Upper
91 100
126 28.6 16.2 48.6





#83

tert-Butylbenzene

Concen: 1.001 ug/l

RT: 13.435 min Scan# 1953

Delta R.T. -0.006 min

Lab File: VN083140.D


Acq: 07 Aug 2024 12:34

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001



Tgt Ion:119 Resp: 7628

Ion Ratio Lower Upper

119 100

91 69.2 35.0 105.1




134 23.5 13.3 39.8

Manual Integrations

APPROVED

Reviewed By :John Carbone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#84

1,2,4-Trimethylbenzene

Concen: 0.968 ug/l

RT: 13.482 min Scan# 1960

Delta R.T. 0.000 min

Lab File: VN083140.D




Acq: 07 Aug 2024 12:34


Tgt Ion:105 Resp: 8397

Ion Ratio Lower Upper

105 100


120 44.1 21.9 65.8





#85
sec-Butylbenzene
Concen: 1.001 ug/l
RT: 13.618 min Scan# 1983
Delta R.T. 0.000 min
Lab File: VN083140.D
Acq: 07 Aug 2024 12:34




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001



Tgt Ion:105 Resp: 1041
Ion Ratio Lower Upper
105 100
134 17.7 9.8 29.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#86
p-Isopropyltoluene
Concen: 0.960 ug/l
RT: 13.735 min Scan# 2003
Delta R.T. 0.006 min
Lab File: VN083140.D
Acq: 07 Aug 2024 12:34

Tgt Ion:119 Resp: 8247
Ion Ratio Lower Upper
119 100
134 26.4 13.0 39.0
91 24.2 12.3 36.9





#87

1,3-Dichlorobenzene

Concen: 1.070 ug/l

RT: 13.735 min Scan# 2003

Delta R.T. 0.000 min

Lab File: VN083140.D

Acq: 07 Aug 2024 12:34

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC001





Tgt Ion:	146	Resp:	4598
Ion	Ratio	Lower	Upper
146	100		
111	41.5	20.6	61.9
148	59.5	31.8	95.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#88

1,4-Dichlorobenzene

Concen: 1.057 ug/l m




RT: 13.812 min Scan# 2016


Delta R.T. 0.000 min

Lab File: VN083140.D

Acq: 07 Aug 2024 12:34

Tgt Ion:	146	Resp:	4578
Ion	Ratio	Lower	Upper
146	100		
111	41.7	20.1	60.3
148	59.7	32.2	96.6





#89

n-Butylbenzene

Concen: 0.942 ug/l

RT: 14.053 min Scan# 2058

Delta R.T. -0.006 min

Lab File: VN083140.D

Acq: 07 Aug 2024 12:34

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001





Tgt Ion:	91	Resp:	7008
Ion	Ratio	Lower	Upper
91	100		
92	52.0	26.1	78.3
134	24.1	12.1	36.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#90

Hexachloroethane

Concen: 0.918 ug/l




RT: 14.335 min Scan# 2105


Delta R.T. 0.000 min

Lab File: VN083140.D

Acq: 07 Aug 2024 12:34

Tgt Ion:	117	Resp:	1524
Ion	Ratio	Lower	Upper
117	100		
201	65.7	35.8	107.3









#91
1,2-Dichlorobenzene
Concen: 1.029 ug/l
RT: 14.106 min Scan# 2066
Delta R.T. 0.000 min
Lab File: VN083140.D
Acq: 07 Aug 2024 12:34

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001

Tgt Ion:146 Resp: 4279
Ion Ratio Lower Upper
146 100
111 40.7 21.4 64.2
148 59.6 32.1 96.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#92
1,2-Dibromo-3-Chloropropane
Concen: 1.081 ug/l
RT: 14.718 min Scan# 2170
Delta R.T. 0.000 min
Lab File: VN083140.D
Acq: 07 Aug 2024 12:34

Tgt Ion: 75 Resp: 763
Ion Ratio Lower Upper
75 100
155 59.5 36.6 109.8
157 67.9 46.9 140.6





#93

1,2,4-Trichlorobenzene

Concen: 1.070 ug/l

RT: 15.394 min Scan# 21

Delta R.T. 0.000 min

Lab File: VN083140.D

Acq: 07 Aug 2024 12:34

Instrument :

MSVOA_N

ClientSampleId :





VSTDICC001

Tgt Ion:	180	Resp:	249
Ion Ratio	Lower	Upper	
180	100		
182	95.5	47.3	141.8
145	25.6	15.7	47.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#94

Hexachlorobutadiene

Concen: 1.140 ug/l




RT: 15.512 min Scan# 2305


Delta R.T. 0.012 min

Lab File: VN083140.D

Acq: 07 Aug 2024 12:34


Tgt Ion:	225	Resp:	1181
Ion Ratio	Lower	Upper	
225	100		
223	57.0	31.9	95.7
227	59.4	32.5	97.5





#95
Naphthalene
Concen: 1.021 ug/l
RT: 15.641 min Scan# 2327
Delta R.T. 0.000 min
Lab File: VN083140.D
Acq: 07 Aug 2024 12:34




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001



Tgt Ion:128 Resp: 8430
Ion Ratio Lower Upper
128 100
127 10.4 10.6 16.0
129 10.9 8.6 12.8




Manual Integrations
APPROVED

Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#96
1,2,3-Trichlorobenzene
Concen: 1.036 ug/l
RT: 15.835 min Scan# 2360
Delta R.T. -0.006 min
Lab File: VN083140.D
Acq: 07 Aug 2024 12:34

Tgt Ion:180 Resp: 2389
Ion Ratio Lower Upper
180 100
182 93.8 48.9 146.8
145 36.6 16.8 50.4



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080724\
 Data File : VN083142.D
 Acq On : 07 Aug 2024 14:49
 Operator : JC\MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ICVVN080724

Quant Time: Aug 08 06:35:30 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : John Carlone 08/08/2024
 Supervised By : Mahesh Dadoda 08/09/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	223686	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	380000	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	330951	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	157193	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	161051	50.583	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 101.160%			
35) Dibromofluoromethane	8.165	113	122268	51.549	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 103.100%			
50) Toluene-d8	10.565	98	461627	52.175	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 104.340%			
62) 4-Bromofluorobenzene	12.847	95	181344	52.574	ug/l	0.00
Spiked Amount 50.000	Range 64 - 133		Recovery = 105.140%			
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	2.124	85	119481	47.102	ug/l	99
3) Chloromethane	2.360	50	120081	46.238	ug/l	97
4) Vinyl Chloride	2.512	62	121730	45.937	ug/l	99
5) Bromomethane	2.948	94	70678	42.984	ug/l	99
6) Chloroethane	3.112	64	72491	43.725	ug/l	96
7) Trichlorofluoromethane	3.495	101	206291	47.124	ug/l	98
8) Diethyl Ether	3.959	74	75355	46.259	ug/l	85
9) 1,1,2-Trichlorotrifluo...	4.365	101	113860	47.176	ug/l	99
10) Methyl Iodide	4.589	142	150327	47.349	ug/l	97
11) Tert butyl alcohol	5.524	59	137718	208.135	ug/l	99
12) 1,1-Dichloroethene	4.336	96	112976	45.562	ug/l	93
13) Acrolein	4.177	56	93826	217.571	ug/l	97
14) Allyl chloride	5.018	41	192929	41.173	ug/l	94
15) Acrylonitrile	5.712	53	296594	217.950	ug/l	99
16) Acetone	4.424	43	262677	210.840	ug/l	97
17) Carbon Disulfide	4.712	76	316980	43.689	ug/l	100
18) Methyl Acetate	5.024	43	157374	42.397	ug/l #	91
19) Methyl tert-butyl Ether	5.795	73	418691	46.781	ug/l	97
20) Methylene Chloride	5.277	84	125142	43.645	ug/l #	82
21) trans-1,2-Dichloroethene	5.783	96	117045	45.671	ug/l	92
22) Diisopropyl ether	6.671	45	412931	46.881	ug/l	97
23) Vinyl Acetate	6.600	43	2125343	235.414	ug/l	95
24) 1,1-Dichloroethane	6.565	63	224136	46.685	ug/l	99
25) 2-Butanone	7.483	43	405134	211.782	ug/l	91
26) 2,2-Dichloropropane	7.489	77	213355	47.853	ug/l	97
27) cis-1,2-Dichloroethene	7.483	96	142606	46.113	ug/l	93
28) Bromochloromethane	7.812	49	90226	45.986	ug/l	86
29) Tetrahydrofuran	7.836	42	274710	222.147	ug/l	89
30) Chloroform	7.965	83	236359	47.387	ug/l	97
31) Cyclohexane	8.253	56	205124	43.474	ug/l	96
32) 1,1,1-Trichloroethane	8.165	97	222538	47.135	ug/l	96
36) 1,1-Dichloropropene	8.371	75	171262	47.732	ug/l	99
37) Ethyl Acetate	7.559	43	173246	43.124	ug/l #	96
38) Carbon Tetrachloride	8.365	117	193592	47.904	ug/l	100
39) Methylcyclohexane	9.600	83	210745	47.821	ug/l	98
40) Benzene	8.606	78	509159	47.635	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080724\
 Data File : VN083142.D
 Acq On : 07 Aug 2024 14:49
 Operator : JC\MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ICVVN080724

Quant Time: Aug 08 06:35:30 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : John Carlone 08/08/2024
 Supervised By : Mahesh Dadoda 08/09/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.777	41	98251	43.013	ug/l	95
42) 1,2-Dichloroethane	8.671	62	184210	47.310	ug/l	98
43) Isopropyl Acetate	8.689	43	319540	45.577	ug/l #	95
44) Trichloroethene	9.353	130	122105	47.994	ug/l	99
45) 1,2-Dichloropropane	9.624	63	122313	48.208	ug/l	99
46) Dibromomethane	9.712	93	85581	47.124	ug/l	99
47) Bromodichloromethane	9.888	83	191842	47.047	ug/l	98
48) Methyl methacrylate	9.683	41	148161	44.609	ug/l	92
49) 1,4-Dioxane	9.694	88	53042	885.211	ug/l	96
51) 4-Methyl-2-Pentanone	10.441	43	877246	230.917	ug/l	94
52) Toluene	10.630	92	327682	48.520	ug/l	98
53) t-1,3-Dichloropropene	10.835	75	208495	49.776	ug/l	99
54) cis-1,3-Dichloropropene	10.312	75	216020	48.487	ug/l	93
55) 1,1,2-Trichloroethane	11.018	97	117023	48.351	ug/l	99
56) Ethyl methacrylate	10.871	69	216248	47.395	ug/l	88
57) 1,3-Dichloropropane	11.165	76	204441	47.391	ug/l	98
58) 2-Chloroethyl Vinyl ether	10.159	63	472357	244.918	ug/l	96
59) 2-Hexanone	11.194	43	661033	224.916	ug/l	92
60) Dibromochloromethane	11.359	129	144886	49.513	ug/l	99
61) 1,2-Dibromoethane	11.471	107	119834	47.133	ug/l	98
64) Tetrachloroethene	11.106	164	105591	48.175	ug/l	97
65) Chlorobenzene	11.888	112	347560	47.522	ug/l	95
66) 1,1,1,2-Tetrachloroethane	11.959	131	120826	46.843	ug/l	98
67) Ethyl Benzene	11.965	91	640827	47.764	ug/l	99
68) m/p-Xylenes	12.071	106	484653	96.434	ug/l	98
69) o-Xylene	12.400	106	231292	46.662	ug/l	97
70) Styrene	12.412	104	403657	48.487	ug/l	98
71) Bromoform	12.577	173	93397	47.797	ug/l #	99
73) Isopropylbenzene	12.694	105	616544	46.893	ug/l	99
74) N-amyl acetate	12.494	43	281352	43.748	ug/l	94
75) 1,1,2,2-Tetrachloroethane	12.941	83	159962	43.016	ug/l	99
76) 1,2,3-Trichloropropane	12.994	75	145309m	42.119	ug/l	
77) Bromobenzene	12.982	156	136249	46.649	ug/l	97
78) n-propylbenzene	13.035	91	721825	47.678	ug/l	99
79) 2-Chlorotoluene	13.124	91	445353	46.398	ug/l	98
80) 1,3,5-Trimethylbenzene	13.171	105	524769	47.673	ug/l	98
81) trans-1,4-Dichloro-2-b...	12.735	75	74654	47.070	ug/l	84
82) 4-Chlorotoluene	13.224	91	447810	46.512	ug/l	98
83) tert-Butylbenzene	13.435	119	450104	46.173	ug/l	99
84) 1,2,4-Trimethylbenzene	13.482	105	515759	46.493	ug/l	99
85) sec-Butylbenzene	13.618	105	625404	47.018	ug/l	100
86) p-Isopropyltoluene	13.729	119	522411	47.566	ug/l	98
87) 1,3-Dichlorobenzene	13.735	146	252579	45.959	ug/l	98
88) 1,4-Dichlorobenzene	13.812	146	252088	45.499	ug/l	98
89) n-Butylbenzene	14.059	91	461440	48.489	ug/l	98
90) Hexachloroethane	14.335	117	99176	46.741	ug/l	95
91) 1,2-Dichlorobenzene	14.106	146	240673	45.255	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	14.723	75	37677	41.753	ug/l	99
93) 1,2,4-Trichlorobenzene	15.394	180	137214	46.054	ug/l	96
94) Hexachlorobutadiene	15.500	225	57394	43.315	ug/l	98
95) Naphthalene	15.641	128	471674	44.692	ug/l	99
96) 1,2,3-Trichlorobenzene	15.841	180	133315	45.221	ug/l	95

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080724\
Data File : VN083142.D
Acq On : 07 Aug 2024 14:49
Operator : JC\MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ICVVN080724

Manual Integrations
APPROVED

Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024

Quant Time: Aug 08 06:35:30 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080724\
Data File : VN083142.D
Acq On : 07 Aug 2024 14:49
Operator : JC\MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1


Instrument :
MSVOA_N
ClientSampleId :
ICVVN080724

Manual Integrations APPROVED

Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024


Quant Time: Aug 08 06:35:30 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 1066
Delta R.T. -0.000 min
Lab File: VN083142.D
Acq: 07 Aug 2024 14:49




Instrument :
MSVOA_N
ClientSampleId :
ICVVN080724



Tgt Ion:168 Resp: 223680
Ion Ratio Lower Upper
168 100
99 67.4 48.2 72.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#2
Dichlorodifluoromethane
Concen: 47.102 ug/l
RT: 2.124 min Scan# 29
Delta R.T. 0.000 min
Lab File: VN083142.D
Acq: 07 Aug 2024 14:49


Tgt Ion: 85 Resp: 119481
Ion Ratio Lower Upper
85 100
87 31.6 15.7 47.0





#3
Chloromethane
Concen: 46.238 ug/l
RT: 2.360 min Scan# 69
Delta R.T. 0.000 min
Lab File: VN083142.D
Acq: 07 Aug 2024 14:49




Instrument :
MSVOA_N
ClientSampleId :
ICVVN080724



Tgt Ion: 50 Resp: 12008
Ion Ratio Lower Upper
50 100
52 32.9 25.0 37.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#4
Vinyl Chloride
Concen: 45.937 ug/l
RT: 2.512 min Scan# 95
Delta R.T. 0.000 min
Lab File: VN083142.D
Acq: 07 Aug 2024 14:49


Tgt Ion: 62 Resp: 121730
Ion Ratio Lower Upper
62 100
64 32.0 25.0 37.6





#5
Bromomethane
Concen: 42.984 ug/l
RT: 2.948 min Scan# 169
Delta R.T. 0.000 min
Lab File: VN083142.D
Acq: 07 Aug 2024 14:49




Instrument :
MSVOA_N
ClientSampleId :
ICVVN080724



Tgt Ion: 94 Resp: 70678
Ion Ratio Lower Upper
94 100
96 96.9 78.0 117.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#6
Chloroethane
Concen: 43.725 ug/l
RT: 3.112 min Scan# 197
Delta R.T. 0.000 min
Lab File: VN083142.D
Acq: 07 Aug 2024 14:49


Tgt Ion: 64 Resp: 72491
Ion Ratio Lower Upper
64 100
66 31.3 26.6 40.0





#7
Trichlorofluoromethane
Concen: 47.124 ug/l
RT: 3.495 min Scan# 262
Delta R.T. -0.000 min
Lab File: VN083142.D
Acq: 07 Aug 2024 14:49




Instrument :
MSVOA_N
ClientSampleId :
ICVVN080724



Tgt Ion:101 Resp: 20629
Ion Ratio Lower Upper
101 100
103 67.1 52.7 79.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#8
Diethyl Ether
Concen: 46.259 ug/l
RT: 3.959 min Scan# 341
Delta R.T. 0.000 min
Lab File: VN083142.D
Acq: 07 Aug 2024 14:49


Tgt Ion: 74 Resp: 75355
Ion Ratio Lower Upper
74 100
45 94.7 55.5 166.3





#9
1,1,2-Trichlorotrifluoroethane
Concen: 47.176 ug/l
RT: 4.365 min Scan# 410
Delta R.T. 0.000 min
Lab File: VN083142.D
Acq: 07 Aug 2024 14:49




Instrument :
MSVOA_N
ClientSampleId :
ICVVN080724



Tgt Ion:101 Resp: 113860
Ion Ratio Lower Upper
101 100
85 47.0 38.3 57.5
151 75.1 59.3 88.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#10
Methyl Iodide
Concen: 47.349 ug/l
RT: 4.589 min Scan# 448
Delta R.T. 0.000 min
Lab File: VN083142.D
Acq: 07 Aug 2024 14:49

Tgt Ion:142 Resp: 150327
Ion Ratio Lower Upper
142 100
127 49.1 37.5 56.3
141 15.1 13.1 19.7





#11

Tert butyl alcohol

Concen: 208.135 ug/l

RT: 5.524 min Scan# 607

Delta R.T. -0.006 min

Lab File: VN083142.D


Acq: 07 Aug 2024 14:49

Instrument :

MSVOA_N

ClientSampleId :

ICVVN080724



Tgt Ion: 59 Resp: 137718

Ion Ratio Lower Upper

59 100




57 10.9 8.9 13.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#12

1,1-Dichloroethene

Concen: 45.562 ug/l

RT: 4.336 min Scan# 405

Delta R.T. -0.006 min

Lab File: VN083142.D

Acq: 07 Aug 2024 14:49




Tgt Ion: 96 Resp: 112976


Ion Ratio Lower Upper

96 100

61 173.9 149.7 224.5

98 63.8 50.1 75.1





#13

Acrolein

Concen: 217.571 ug/l

RT: 4.177 min Scan# 31

Delta R.T. -0.006 min

Lab File: VN083142.D

Acq: 07 Aug 2024 14:49

Instrument :

MSVOA_N

ClientSampleId :

ICVVN080724

Tgt Ion: 56 Resp: 93820

Ion Ratio Lower Upper

56 100





55 73.1 56.4 84.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#14

Allyl chloride

Concen: 41.173 ug/l

RT: 5.018 min Scan# 521

Delta R.T. 0.000 min

Lab File: VN083142.D

Acq: 07 Aug 2024 14:49




Tgt Ion: 41 Resp: 192929


Ion Ratio Lower Upper

41 100

39 79.9 67.8 101.6

76 36.7 25.8 38.8





#15

Acrylonitrile

Concen: 217.950 ug/l

RT: 5.712 min Scan# 61

Delta R.T. -0.006 min

Lab File: VN083142.D


Acq: 07 Aug 2024 14:49

Instrument :

MSVOA_N

ClientSampleId :

ICVVN080724






Tgt Ion:	53	Resp:	296594
Ion	Ratio	Lower	Upper
53	100		
52	82.9	66.6	100.0
51	36.5	30.1	45.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#16

Acetone

Concen: 210.840 ug/l




RT: 4.424 min Scan# 420


Delta R.T. -0.006 min

Lab File: VN083142.D

Acq: 07 Aug 2024 14:49


Tgt Ion:	43	Resp:	262677
Ion	Ratio	Lower	Upper
43	100		
58	28.9	21.8	32.6





#17
Carbon Disulfide
Concen: 43.689 ug/l
RT: 4.712 min Scan# 469
Delta R.T. 0.000 min
Lab File: VN083142.D
Acq: 07 Aug 2024 14:49




Instrument :
MSVOA_N
ClientSampleId :
ICVVN080724



Tgt Ion: 76 Resp: 316980
Ion Ratio Lower Upper
76 100
78 9.1 7.2 10.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#18
Methyl Acetate
Concen: 42.397 ug/l
RT: 5.024 min Scan# 522
Delta R.T. 0.000 min
Lab File: VN083142.D
Acq: 07 Aug 2024 14:49

Tgt Ion: 43 Resp: 157374
Ion Ratio Lower Upper
43 100
74 24.6 16.3 24.5#





#19

Methyl tert-butyl Ether

Concen: 46.781 ug/l

RT: 5.795 min Scan# 61

Delta R.T. 0.000 min

Lab File: VN083142.D

Acq: 07 Aug 2024 14:49

Instrument :

MSVOA_N

ClientSampleId :

ICVVN080724

Tgt Ion: 73 Resp: 41869

Ion Ratio Lower Upper

73 100





57 21.1 17.9 26.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#20

Methylene Chloride

Concen: 43.645 ug/l

RT: 5.277 min Scan# 565

Delta R.T. 0.000 min

Lab File: VN083142.D

Acq: 07 Aug 2024 14:49

Tgt Ion: 84 Resp: 125142




Ion Ratio Lower Upper


84 100

49 115.8 119.6 179.4#

51 36.5 34.8 52.2


86 63.6 52.9 79.3





#21
trans-1,2-Dichloroethene
Concen: 45.671 ug/l
RT: 5.783 min Scan# 61
Delta R.T. -0.012 min
Lab File: VN083142.D
Acq: 07 Aug 2024 14:49




Instrument :
MSVOA_N
ClientSampleId :
ICVVN080724



Tgt Ion: 96 Resp: 117049
Ion Ratio Lower Upper
96 100
61 141.1 123.4 185.2
98 62.9 49.3 73.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#22
Diisopropyl ether
Concen: 46.881 ug/l
RT: 6.671 min Scan# 802
Delta R.T. 0.000 min
Lab File: VN083142.D
Acq: 07 Aug 2024 14:49


Tgt Ion: 45 Resp: 412931
Ion Ratio Lower Upper
45 100
43 55.6 44.0 66.0
87 28.6 19.7 29.5
59 12.4 8.5 12.7





#23
Vinyl Acetate
Concen: 235.414 ug/l
RT: 6.600 min Scan# 791
Delta R.T. -0.006 min
Lab File: VN083142.D
Acq: 07 Aug 2024 14:49




Instrument :
MSVOA_N
Client Sample Id :
ICVVN080724



Tgt Ion: 43 Resp: 212534
Ion Ratio Lower Upper
43 100
86 10.9 7.3 10.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#24
1,1-Dichloroethane
Concen: 46.685 ug/l
RT: 6.565 min Scan# 784
Delta R.T. -0.006 min
Lab File: VN083142.D
Acq: 07 Aug 2024 14:49


Tgt Ion: 63 Resp: 224136
Ion Ratio Lower Upper
63 100
98 6.5 3.3 9.9
100 4.9 2.0 6.0





#25
2-Butanone
Concen: 211.782 ug/l
RT: 7.483 min Scan# 940
Delta R.T. 0.000 min
Lab File: VN083142.D
Acq: 07 Aug 2024 14:49




Instrument :
MSVOA_N
ClientSampleId :
ICVVN080724



Tgt Ion: 43 Resp: 405134
Ion Ratio Lower Upper
43 100
72 26.9 18.2 27.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#26
2,2-Dichloropropane
Concen: 47.853 ug/l
RT: 7.489 min Scan# 941
Delta R.T. 0.000 min
Lab File: VN083142.D
Acq: 07 Aug 2024 14:49

Tgt Ion: 77 Resp: 213355
Ion Ratio Lower Upper
77 100
97 21.9 10.3 30.9





#27

cis-1,2-Dichloroethene

Concen: 46.113 ug/l

RT: 7.483 min Scan# 941

Delta R.T. -0.006 min

Lab File: VN083142.D

Acq: 07 Aug 2024 14:49

Instrument :

MSVOA_N

ClientSampleId :

ICVVN080724

Tgt Ion: 96 Resp: 142600

Ion Ratio Lower Upper

96 100

61 151.9 0.0 329.6





98 65.2 0.0 128.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#28

Bromochloromethane

Concen: 45.986 ug/l

RT: 7.812 min Scan# 996

Delta R.T. 0.000 min

Lab File: VN083142.D

Acq: 07 Aug 2024 14:49




Tgt Ion: 49 Resp: 90226


Ion Ratio Lower Upper

49 100

129 2.0 0.0 3.8

130 73.9 50.5 75.7





#29

Tetrahydrofuran

Concen: 222.147 ug/l

RT: 7.836 min Scan# 1000

Delta R.T. 0.000 min

Lab File: VN083142.D

Acq: 07 Aug 2024 14:49

Instrument :

MSVOA_N

ClientSampleId :

ICVVN080724

Tgt Ion: 42 Resp: 274710

Ion Ratio Lower Upper

42 100

72 44.2 30.0 45.0





71 41.5 28.3 42.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#30

Chloroform

Concen: 47.387 ug/l

RT: 7.965 min Scan# 1022

Delta R.T. 0.000 min

Lab File: VN083142.D




Acq: 07 Aug 2024 14:49


Tgt Ion: 83 Resp: 236359

Ion Ratio Lower Upper

83 100

85 66.1 50.9 76.3





#31

Cyclohexane

Concen: 43.474 ug/l

RT: 8.253 min Scan# 1071

Delta R.T. -0.006 min

Lab File: VN083142.D

Acq: 07 Aug 2024 14:49

Instrument :

MSVOA_N

ClientSampleId :

ICVVN080724

Tgt Ion: 56 Resp: 205124

Ion Ratio Lower Upper

56 100

69 32.3 24.5 36.7





84 86.7 66.0 99.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#32

1,1,1-Trichloroethane

Concen: 47.135 ug/l

RT: 8.165 min Scan# 1056

Delta R.T. -0.006 min

Lab File: VN083142.D

Acq: 07 Aug 2024 14:49




Tgt Ion: 97 Resp: 222538


Ion Ratio Lower Upper

97 100

99 64.0 52.0 78.0

61 47.5 42.1 63.1





#33

1,2-Dichloroethane-d4

Concen: 50.583 ug/l

RT: 8.577 min Scan# 1126

Delta R.T. -0.006 min

Lab File: VN083142.D

Acq: 07 Aug 2024 14:49

Instrument :

MSVOA_N

ClientSampleId :

ICVVN080724

Tgt Ion: 65 Resp: 16105

Ion Ratio Lower Upper

65 100





67 51.3 0.0 104.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.100 min Scan# 1215

Delta R.T. 0.000 min

Lab File: VN083142.D

Acq: 07 Aug 2024 14:49




Tgt Ion:114 Resp: 380000


Ion Ratio Lower Upper

114 100

63 22.6 0.0 44.6

88 17.4 0.0 31.4





#35

Dibromofluoromethane

Concen: 51.549 ug/l

RT: 8.165 min Scan# 1056

Delta R.T. 0.000 min

Lab File: VN083142.D

Acq: 07 Aug 2024 14:49

Instrument :

MSVOA_N

ClientSampleId :

ICVVN080724

Tgt Ion:113 Resp: 12226

Ion Ratio Lower Upper

113 100

111 102.3 82.4 123.6





192 17.6 14.9 22.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#36

1,1-Dichloropropene

Concen: 47.732 ug/l

RT: 8.371 min Scan# 1091

Delta R.T. 0.000 min

Lab File: VN083142.D

Acq: 07 Aug 2024 14:49




Tgt Ion: 75 Resp: 171262


Ion Ratio Lower Upper

75 100

110 33.2 16.3 48.9

77 30.5 24.6 37.0





#37

Ethyl Acetate

Concen: 43.124 ug/l

RT: 7.559 min Scan# 91

Delta R.T. 0.000 min

Lab File: VN083142.D


Acq: 07 Aug 2024 14:49

Instrument :

MSVOA_N

ClientSampleId :

ICVVN080724






Tgt Ion:	43	Resp:	173240
Ion	Ratio	Lower	Upper
43	100		
61	12.9	9.3	13.9
70	11.0	7.0	10.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#38

Carbon Tetrachloride


Concen: 47.904 ug/l

RT: 8.365 min Scan# 1090



Delta R.T. 0.006 min


Lab File: VN083142.D

Acq: 07 Aug 2024 14:49



Tgt Ion:	117	Resp:	193592
Ion	Ratio	Lower	Upper
117	100		
119	94.0	74.9	112.3
121	30.3	24.3	36.5





#39

Methylcyclohexane

Concen: 47.821 ug/l

RT: 9.600 min Scan# 1130

Delta R.T. 0.000 min

Lab File: VN083142.D

Acq: 07 Aug 2024 14:49

Instrument :

MSVOA_N

ClientSampleId :

ICVVN080724





Tgt Ion:	83	Resp:	21074
Ion	Ratio	Lower	Upper
83	100		
55	81.3	66.4	99.6
98	49.5	40.7	61.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#40

Benzene

Concen: 47.635 ug/l




RT: 8.606 min Scan# 1131


Delta R.T. 0.000 min

Lab File: VN083142.D

Acq: 07 Aug 2024 14:49


Tgt Ion:	78	Resp:	509159
Ion	Ratio	Lower	Upper
78	100		
77	23.4	19.0	28.4





#41
Methacrylonitrile
Concen: 43.013 ug/l
RT: 7.777 min Scan# 990
Delta R.T. 0.000 min
Lab File: VN083142.D
Acq: 07 Aug 2024 14:49




Instrument :
MSVOA_N
ClientSampleId :
ICVVN080724




Tgt Ion: 41 Resp: 9825
Ion Ratio Lower Upper
41 100
39 57.7 45.5 68.3
67 70.8 50.5 75.7
52 31.0 24.0 36.0

Manual Integrations
APPROVED



Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024




#42
1,2-Dichloroethane
Concen: 47.310 ug/l
RT: 8.671 min Scan# 1142
Delta R.T. 0.000 min
Lab File: VN083142.D
Acq: 07 Aug 2024 14:49



Tgt Ion: 62 Resp: 184210
Ion Ratio Lower Upper
62 100
98 8.6 0.0 15.8





#43

Isopropyl Acetate

Concen: 45.577 ug/l

RT: 8.689 min Scan# 1145

Delta R.T. 0.001 min

Lab File: VN083142.D


Acq: 07 Aug 2024 14:49

Instrument :

MSVOA_N

ClientSampleId :

ICVVN080724






Tgt Ion:	43	Resp:	319540
Ion Ratio	100	Lower	Upper
43	100		
61	24.3	17.8	26.6
87	13.0	8.2	12.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#44

Trichloroethene

Concen: 47.994 ug/l




RT: 9.353 min Scan# 1258


Delta R.T. 0.000 min

Lab File: VN083142.D

Acq: 07 Aug 2024 14:49

Tgt Ion:	130	Resp:	122105
Ion Ratio	100	Lower	Upper
130	100		
95	100.4	0.0	197.8





#45

1,2-Dichloropropane

Concen: 48.208 ug/l

RT: 9.624 min Scan# 1304

Delta R.T. 0.000 min

Lab File: VN083142.D

Acq: 07 Aug 2024 14:49

Instrument :

MSVOA_N

ClientSampleId :

ICVVN080724

Tgt Ion: 63 Resp: 12231

Ion Ratio Lower Upper

63 100





65 31.2 24.4 36.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#46

Dibromomethane

Concen: 47.124 ug/l

RT: 9.712 min Scan# 1319

Delta R.T. 0.000 min

Lab File: VN083142.D

Acq: 07 Aug 2024 14:49




Tgt Ion: 93 Resp: 85581


Ion Ratio Lower Upper

93 100

95 83.0 65.8 98.6

174 91.4 71.7 107.5





#47

Bromodichloromethane

Concen: 47.047 ug/l

RT: 9.888 min Scan# 1349

Delta R.T. 0.000 min

Lab File: VN083142.D

Acq: 07 Aug 2024 14:49

Instrument :

MSVOA_N

ClientSampleId :

ICVVN080724





Tgt Ion:	83	Resp:	19184
Ion	Ratio	Lower	Upper
83	100		
85	65.5	51.1	76.7
127	8.5	7.2	10.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#48

Methyl methacrylate

Concen: 44.609 ug/l




RT: 9.683 min Scan# 1314


Delta R.T. 0.000 min

Lab File: VN083142.D

Acq: 07 Aug 2024 14:49

Tgt Ion:	41	Resp:	148161
Ion	Ratio	Lower	Upper
41	100		
69	81.0	56.3	84.5
39	60.7	50.3	75.5





#49

1,4-Dioxane

Concen: 885.211 ug/l

RT: 9.694 min Scan# 11

Delta R.T. 0.000 min

Lab File: VN083142.D

Acq: 07 Aug 2024 14:49

Instrument :

MSVOA_N

ClientSampleId :

ICVVN080724

Tgt Ion: 88 Resp: 5304

Ion Ratio Lower Upper

88 100

43 32.0 27.8 41.8





58 71.4 59.4 89.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#50

Toluene-d8

Concen: 52.175 ug/l

RT: 10.565 min Scan# 1464

Delta R.T. 0.000 min

Lab File: VN083142.D




Acq: 07 Aug 2024 14:49


Tgt Ion: 98 Resp: 461627

Ion Ratio Lower Upper

98 100


100 65.5 51.5 77.3





#51
4-Methyl-2-Pentanone
Concen: 230.917 ug/l
RT: 10.441 min Scan# 1444
Delta R.T. -0.006 min
Lab File: VN083142.D
Acq: 07 Aug 2024 14:49




Instrument :
MSVOA_N
ClientSampleId :
ICVVN080724



Tgt Ion: 43 Resp: 877240
Ion Ratio Lower Upper
43 100
58 39.0 28.2 42.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#52
Toluene
Concen: 48.520 ug/l
RT: 10.630 min Scan# 1475
Delta R.T. 0.000 min
Lab File: VN083142.D
Acq: 07 Aug 2024 14:49


Tgt Ion: 92 Resp: 327682
Ion Ratio Lower Upper
92 100
91 170.9 139.4 209.0





#53
t-1,3-Dichloropropene
Concen: 49.776 ug/l
RT: 10.835 min Scan# 1510
Delta R.T. 0.000 min
Lab File: VN083142.D
Acq: 07 Aug 2024 14:49




Instrument :
MSVOA_N
ClientSampleId :
ICVVN080724



Tgt Ion: 75 Resp: 20849
Ion Ratio Lower Upper
75 100
77 32.8 25.7 38.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#54
cis-1,3-Dichloropropene
Concen: 48.487 ug/l
RT: 10.312 min Scan# 1421
Delta R.T. 0.000 min
Lab File: VN083142.D
Acq: 07 Aug 2024 14:49


Tgt Ion: 75 Resp: 216020
Ion Ratio Lower Upper
75 100
77 31.6 24.3 36.5
39 56.0 50.5 75.7





#55
1,1,2-Trichloroethane
Concen: 48.351 ug/l
RT: 11.018 min Scan# 1541
Delta R.T. 0.000 min
Lab File: VN083142.D
Acq: 07 Aug 2024 14:49




Instrument :
MSVOA_N
ClientSampleId :
ICVVN080724



Tgt Ion: 97 Resp: 117023
Ion Ratio Lower Upper
97 100
83 88.2 71.4 107.0
85 57.0 45.8 68.6
99 62.8 52.0 78.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#56
Ethyl methacrylate
Concen: 47.395 ug/l
RT: 10.871 min Scan# 1516
Delta R.T. -0.006 min
Lab File: VN083142.D
Acq: 07 Aug 2024 14:49

Tgt Ion: 69 Resp: 216248
Ion Ratio Lower Upper
69 100
41 67.3 63.4 95.2
39 40.2 37.4 56.0





#57

1,3-Dichloropropane

Concen: 47.391 ug/l

RT: 11.165 min Scan# 1566

Delta R.T. 0.000 min

Lab File: VN083142.D

Acq: 07 Aug 2024 14:49

Instrument :

MSVOA_N

ClientSampleId :

ICVVN080724

Tgt Ion: 76 Resp: 20444

Ion Ratio Lower Upper

76 100





78 32.9 25.6 38.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#58

2-Chloroethyl Vinyl ether

Concen: 244.918 ug/l

RT: 10.159 min Scan# 1395

Delta R.T. 0.000 min

Lab File: VN083142.D




Acq: 07 Aug 2024 14:49


Tgt Ion: 63 Resp: 472357

Ion Ratio Lower Upper

63 100

106 28.6 21.4 32.0





#59

2-Hexanone

Concen: 224.916 ug/l

RT: 11.194 min Scan# 1571

Delta R.T. 0.000 min

Lab File: VN083142.D


Acq: 07 Aug 2024 14:49

Instrument :

MSVOA_N

ClientSampleId :

ICVVN080724



Tgt Ion: 43 Resp: 661033

Ion Ratio Lower Upper

43 100




58 54.6 24.6 73.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#60

Dibromochloromethane

Concen: 49.513 ug/l

RT: 11.359 min Scan# 1599

Delta R.T. 0.000 min

Lab File: VN083142.D




Acq: 07 Aug 2024 14:49


Tgt Ion:129 Resp: 144886

Ion Ratio Lower Upper

129 100


127 77.3 39.2 117.6





#61
1,2-Dibromoethane
Concen: 47.133 ug/l
RT: 11.471 min Scan# 119834
Delta R.T. 0.000 min
Lab File: VN083142.D
Acq: 07 Aug 2024 14:49




Instrument :
MSVOA_N
ClientSampleId :
ICVVN080724



Tgt Ion:107 Resp: 119834
Ion Ratio Lower Upper
107 100
109 94.5 73.7 110.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#62
4-Bromofluorobenzene
Concen: 52.574 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. 0.000 min
Lab File: VN083142.D
Acq: 07 Aug 2024 14:49


Tgt Ion: 95 Resp: 181344
Ion Ratio Lower Upper
95 100
174 68.0 0.0 159.2
176 65.5 0.0 147.6





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 1685
Delta R.T. 0.000 min
Lab File: VN083142.D
Acq: 07 Aug 2024 14:49




Instrument :
MSVOA_N
ClientSampleId :
ICVVN080724



Tgt Ion:117 Resp: 33095
Ion Ratio Lower Upper
117 100
82 59.4 47.5 71.3
119 31.9 25.6 38.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#64
Tetrachloroethene
Concen: 48.175 ug/l
RT: 11.106 min Scan# 1556
Delta R.T. 0.000 min
Lab File: VN083142.D
Acq: 07 Aug 2024 14:49


Tgt Ion:164 Resp: 105591
Ion Ratio Lower Upper
164 100
166 123.0 101.4 152.0
129 99.7 75.7 113.5
131 93.5 73.8 110.6





#65
Chlorobenzene
Concen: 47.522 ug/l
RT: 11.888 min Scan# 1689
Delta R.T. -0.006 min
Lab File: VN083142.D
Acq: 07 Aug 2024 14:49




Instrument :
MSVOA_N
ClientSampleId :
ICVVN080724




Tgt Ion:112 Resp: 347560
Ion Ratio Lower Upper
112 100
114 29.9 26.2 39.2

Manual Integrations
APPROVED



Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024




#66
1,1,1,2-Tetrachloroethane
Concen: 46.843 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. -0.006 min
Lab File: VN083142.D
Acq: 07 Aug 2024 14:49



Tgt Ion:131 Resp: 120826
Ion Ratio Lower Upper
131 100
133 97.1 47.3 142.0
119 64.8 32.5 97.4





#67

Ethyl Benzene

Concen: 47.764 ug/l

RT: 11.965 min Scan# 1702

Delta R.T. 0.000 min

Lab File: VN083142.D

Acq: 07 Aug 2024 14:49

Instrument :

MSVOA_N

ClientSampleId :

ICVVN080724

Tgt Ion: 91 Resp: 64082

Ion Ratio Lower Upper

91 100





106 30.4 24.0 36.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#68

m/p-Xylenes

Concen: 96.434 ug/l

RT: 12.071 min Scan# 1720

Delta R.T. 0.000 min

Lab File: VN083142.D




Acq: 07 Aug 2024 14:49


Tgt Ion:106 Resp: 484653

Ion Ratio Lower Upper

106 100


91 210.2 166.1 249.1





#69
o-Xylene
Concen: 46.662 ug/l
RT: 12.400 min Scan# 1776
Delta R.T. 0.000 min
Lab File: VN083142.D
Acq: 07 Aug 2024 14:49




Instrument :
MSVOA_N
ClientSampleId :
ICVVN080724



Tgt Ion:106 Resp: 231292
Ion Ratio Lower Upper
106 100
91 227.0 111.3 333.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#70
Styrene
Concen: 48.487 ug/l
RT: 12.412 min Scan# 1778
Delta R.T. 0.000 min
Lab File: VN083142.D
Acq: 07 Aug 2024 14:49


Tgt Ion:104 Resp: 403657
Ion Ratio Lower Upper
104 100
78 53.9 41.6 62.4
103 54.6 44.0 66.0





#71
Bromoform
Concen: 47.797 ug/l
RT: 12.577 min Scan# 1806
Delta R.T. 0.000 min
Lab File: VN083142.D
Acq: 07 Aug 2024 14:49




Instrument :
MSVOA_N
ClientSampleId :
ICVVN080724



Tgt Ion:173 Resp: 93397
Ion Ratio Lower Upper
173 100
175 47.4 24.0 72.0
254 0.0 0.0 0.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.794 min Scan# 2013
Delta R.T. 0.000 min
Lab File: VN083142.D
Acq: 07 Aug 2024 14:49


Tgt Ion:152 Resp: 157193
Ion Ratio Lower Upper
152 100
115 63.3 30.6 91.6
150 172.2 0.0 348.6





#73
Isopropylbenzene
Concen: 46.893 ug/l
RT: 12.694 min Scan# 1826
Delta R.T. 0.000 min
Lab File: VN083142.D
Acq: 07 Aug 2024 14:49




Instrument :
MSVOA_N
ClientSampleId :
ICVVN080724



Tgt Ion:105 Resp: 616544
Ion Ratio Lower Upper
105 100
120 26.0 12.9 38.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#74
N-ethyl acetate
Concen: 43.748 ug/l
RT: 12.494 min Scan# 1792
Delta R.T. 0.000 min
Lab File: VN083142.D
Acq: 07 Aug 2024 14:49


Tgt Ion: 43 Resp: 281352
Ion Ratio Lower Upper
43 100
70 43.1 29.8 44.6
55 23.3 18.7 28.1
61 24.5 17.4 26.0





#75
1,1,2,2-Tetrachloroethane
Concen: 43.016 ug/l
RT: 12.941 min Scan# 1868
Delta R.T. 0.006 min
Lab File: VN083142.D
Acq: 07 Aug 2024 14:49




Instrument :
MSVOA_N
ClientSampleId :
ICVVN080724



Tgt Ion: 83 Resp: 159962
Ion Ratio Lower Upper
83 100
131 10.6 5.0 14.9
85 64.5 31.9 95.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#76
1,2,3-Trichloropropane
Concen: 42.119 ug/l m
RT: 12.994 min Scan# 1877
Delta R.T. 0.000 min
Lab File: VN083142.D
Acq: 07 Aug 2024 14:49


Tgt Ion: 75 Resp: 145309
Ion Ratio Lower Upper
75 100
77 214.8 110.9 332.6





#77
Bromobenzene
Concen: 46.649 ug/l
RT: 12.982 min Scan# 1875
Delta R.T. 0.000 min
Lab File: VN083142.D
Acq: 07 Aug 2024 14:49




Instrument :
MSVOA_N
ClientSampleId :
ICVVN080724



Tgt Ion:156 Resp: 136249
Ion Ratio Lower Upper
156 100
77 229.1 111.3 333.8
158 98.0 48.3 144.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#78
n-propylbenzene
Concen: 47.678 ug/l
RT: 13.035 min Scan# 1884
Delta R.T. 0.000 min
Lab File: VN083142.D
Acq: 07 Aug 2024 14:49


Tgt Ion: 91 Resp: 721825
Ion Ratio Lower Upper
91 100
120 22.2 10.9 32.9





#79
2-Chlorotoluene
Concen: 46.398 ug/l
RT: 13.124 min Scan# 1899
Delta R.T. 0.000 min
Lab File: VN083142.D
Acq: 07 Aug 2024 14:49




Instrument :
MSVOA_N
ClientSampleId :
ICVVN080724



Tgt Ion: 91 Resp: 44535
Ion Ratio Lower Upper
91 100
126 31.1 16.2 48.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#80
1,3,5-Trimethylbenzene
Concen: 47.673 ug/l
RT: 13.171 min Scan# 1907
Delta R.T. 0.000 min
Lab File: VN083142.D
Acq: 07 Aug 2024 14:49


Tgt Ion:105 Resp: 524769
Ion Ratio Lower Upper
105 100
120 46.9 24.3 72.8





#81
trans-1,4-Dichloro-2-butene
Concen: 47.070 ug/l
RT: 12.735 min Scan# 1833
Delta R.T. 0.000 min
Lab File: VN083142.D
Acq: 07 Aug 2024 14:49




Instrument :
MSVOA_N
ClientSampleId :
ICVVN080724



Tgt Ion: 75 Resp: 74654
Ion Ratio Lower Upper
75 100
53 89.6 88.4 132.6
89 41.1 36.3 54.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#82
4-Chlorotoluene
Concen: 46.512 ug/l
RT: 13.224 min Scan# 1916
Delta R.T. 0.000 min
Lab File: VN083142.D
Acq: 07 Aug 2024 14:49


Tgt Ion: 91 Resp: 447810
Ion Ratio Lower Upper
91 100
126 31.2 16.2 48.6





#83
tert-Butylbenzene
Concen: 46.173 ug/l
RT: 13.435 min Scan# 1953
Delta R.T. -0.006 min
Lab File: VN083142.D
Acq: 07 Aug 2024 14:49




Instrument :
MSVOA_N
ClientSampleId :
ICVVN080724



Tgt Ion:119 Resp: 450104
Ion Ratio Lower Upper
119 100
91 70.4 35.0 105.1
134 24.4 13.3 39.8




Manual Integrations
APPROVED


Reviewed By :John Carbone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#84
1,2,4-Trimethylbenzene
Concen: 46.493 ug/l
RT: 13.482 min Scan# 1960
Delta R.T. 0.000 min
Lab File: VN083142.D
Acq: 07 Aug 2024 14:49


Tgt Ion:105 Resp: 515759
Ion Ratio Lower Upper
105 100
120 44.4 21.9 65.8





#85
sec-Butylbenzene
Concen: 47.018 ug/l
RT: 13.618 min Scan# 1983
Delta R.T. 0.000 min
Lab File: VN083142.D
Acq: 07 Aug 2024 14:49




Instrument :
MSVOA_N
ClientSampleId :
ICVVN080724



Tgt Ion:105 Resp: 625404
Ion Ratio Lower Upper
105 100
134 19.5 9.8 29.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#86
p-Isopropyltoluene
Concen: 47.566 ug/l
RT: 13.729 min Scan# 2002
Delta R.T. 0.000 min
Lab File: VN083142.D
Acq: 07 Aug 2024 14:49

Tgt Ion:119 Resp: 522411
Ion Ratio Lower Upper
119 100
134 25.6 13.0 39.0
91 25.9 12.3 36.9





#87

1,3-Dichlorobenzene

Concen: 45.959 ug/l

RT: 13.735 min Scan# 2003

Delta R.T. 0.000 min

Lab File: VN083142.D

Acq: 07 Aug 2024 14:49

Instrument :

MSVOA_N

ClientSampleId :

ICVVN080724

Tgt Ion:146 Resp: 252579

Ion Ratio Lower Upper

146 100

111 43.4 20.6 61.9





148 62.8 31.8 95.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#88

1,4-Dichlorobenzene

Concen: 45.499 ug/l

RT: 13.812 min Scan# 2016

Delta R.T. 0.000 min

Lab File: VN083142.D

Acq: 07 Aug 2024 14:49




Tgt Ion:146 Resp: 252088


Ion Ratio Lower Upper

146 100

111 42.6 20.1 60.3

148 64.5 32.2 96.6





#89

n-Butylbenzene

Concen: 48.489 ug/l

RT: 14.059 min Scan# 2058

Delta R.T. 0.000 min

Lab File: VN083142.D

Acq: 07 Aug 2024 14:49

Instrument :

MSVOA_N

ClientSampleId :

ICVVN080724

Tgt Ion: 91 Resp: 461440

Ion Ratio Lower Upper

91 100

92 53.9 26.1 78.3





134 24.5 12.1 36.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#90

Hexachloroethane

Concen: 46.741 ug/l

RT: 14.335 min Scan# 2105

Delta R.T. 0.000 min

Lab File: VN083142.D




Acq: 07 Aug 2024 14:49


Tgt Ion:117 Resp: 99176

Ion Ratio Lower Upper

117 100


201 67.3 35.8 107.3





#91
1,2-Dichlorobenzene
Concen: 45.255 ug/l
RT: 14.106 min Scan# 2066
Delta R.T. 0.000 min
Lab File: VN083142.D
Acq: 07 Aug 2024 14:49




Instrument :
MSVOA_N
ClientSampleId :
ICVVN080724




Tgt Ion:146 Resp: 240677
Ion Ratio Lower Upper
146 100
111 45.2 21.4 64.2
148 64.4 32.1 96.5

Manual Integrations
APPROVED



Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024




#92
1,2-Dibromo-3-Chloropropane
Concen: 41.753 ug/l
RT: 14.723 min Scan# 2171
Delta R.T. 0.006 min
Lab File: VN083142.D
Acq: 07 Aug 2024 14:49



Tgt Ion: 75 Resp: 37677
Ion Ratio Lower Upper
75 100
155 72.8 36.6 109.8
157 95.3 46.9 140.6





#93

1,2,4-Trichlorobenzene

Concen: 46.054 ug/l

RT: 15.394 min Scan# 2285

Delta R.T. 0.000 min

Lab File: VN083142.D

Acq: 07 Aug 2024 14:49

Instrument :

MSVOA_N

ClientSampleId :

ICVVN080724

Tgt Ion:180 Resp: 137214

Ion Ratio Lower Upper

180 100

182 97.2 47.3 141.8





145 34.6 15.7 47.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#94

Hexachlorobutadiene

Concen: 43.315 ug/l

RT: 15.500 min Scan# 2303

Delta R.T. 0.000 min

Lab File: VN083142.D

Acq: 07 Aug 2024 14:49




Tgt Ion:225 Resp: 57394


Ion Ratio Lower Upper

225 100

223 63.3 31.9 95.7


227 62.6 32.5 97.5





#95
Naphthalene
Concen: 44.692 ug/l
RT: 15.641 min Scan# 2327
Delta R.T. 0.000 min
Lab File: VN083142.D
Acq: 07 Aug 2024 14:49




Instrument :
MSVOA_N
ClientSampleId :
ICVVN080724



Tgt Ion:128 Resp: 471674
Ion Ratio Lower Upper
128 100
127 12.4 10.6 16.0
129 10.8 8.6 12.8




Manual Integrations
APPROVED

Reviewed By :John Carlone 08/08/2024
Supervised By :Mahesh Dadoda 08/09/2024



#96
1,2,3-Trichlorobenzene
Concen: 45.221 ug/l
RT: 15.841 min Scan# 2361
Delta R.T. 0.000 min
Lab File: VN083142.D
Acq: 07 Aug 2024 14:49

Tgt Ion:180 Resp: 133315
Ion Ratio Lower Upper
180 100
182 93.7 48.9 146.8
145 36.6 16.8 50.4



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080724\
 Data File : VN083142.D
 Acq On : 07 Aug 2024 14:49
 Operator : JC\MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ICVVN080724

Quant Time: Aug 08 06:35:30 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	109	0.00
2 T	Dichlorodifluoromethane	0.567	0.534	5.8	102	0.00
3 P	Chloromethane	0.581	0.537	7.6	99	0.00
4 C	Vinyl Chloride	0.592	0.544	8.1#	99	0.00
5 T	Bromomethane	0.368	0.316	14.1	97	0.00
6 T	Chloroethane	0.371	0.324	12.7	99	0.00
7 T	Trichlorofluoromethane	0.979	0.922	5.8	101	0.00
8 T	Diethyl Ether	0.364	0.337	7.4	100	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.539	0.509	5.6	101	0.00
10 T	Methyl Iodide	0.710	0.672	5.4	99	0.00
11 T	Tert butyl alcohol	0.148	0.123	16.9	92	0.00
12 CM	1,1-Dichloroethene	0.554	0.505	8.8#	101	0.00
13 T	Acrolein	0.096	0.084	12.5	90	0.00
14 T	Allyl chloride	1.047	0.862	17.7	101	0.00
15 T	Acrylonitrile	0.304	0.265	12.8	95	0.00
16 T	Acetone	0.278	0.235	15.5	90	0.00
17 T	Carbon Disulfide	1.622	1.417	12.6	99	0.00
18 T	Methyl Acetate	0.830	0.704	15.2	99	0.00
19 T	Methyl tert-butyl Ether	2.001	1.872	6.4	101	0.00
20 T	Methylene Chloride	0.641	0.559	12.8	102	0.00
21 T	trans-1,2-Dichloroethene	0.573	0.523	8.7	99	-0.01
22 T	Diisopropyl ether	1.969	1.846	6.2	100	0.00
23 T	Vinyl Acetate	2.018	1.900	5.8	99	0.00
24 P	1,1-Dichloroethane	1.073	1.002	6.6	101	0.00
25 T	2-Butanone	0.428	0.362	15.4	93	0.00
26 T	2,2-Dichloropropane	0.997	0.954	4.3	102	0.00
27 T	cis-1,2-Dichloroethene	0.691	0.638	7.7	101	0.00
28 T	Bromochloromethane	0.439	0.403	8.2	100	0.00
29 T	Tetrahydrofuran	0.276	0.246	10.9	96	0.00
30 C	Chloroform	1.115	1.057	5.2#	101	0.00
31 T	Cyclohexane	1.055	0.917	13.1	102	0.00
32 T	1,1,1-Trichloroethane	1.055	0.995	5.7	101	0.00
33 S	1,2-Dichloroethane-d4	0.712	0.720	-1.1	105	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	109	0.00
35 S	Dibromofluoromethane	0.312	0.322	-3.2	106	0.00
36 T	1,1-Dichloropropene	0.472	0.451	4.4	103	0.00
37 T	Ethyl Acetate	0.529	0.456	13.8	100	0.00
38 T	Carbon Tetrachloride	0.532	0.509	4.3	102	0.00
39 T	Methylcyclohexane	0.580	0.555	4.3	101	0.00
40 TM	Benzene	1.406	1.340	4.7	101	0.00
41 T	Methacrylonitrile	0.301	0.259	14.0	99	0.00
42 TM	1,2-Dichloroethane	0.512	0.485	5.3	102	0.00
43 T	Isopropyl Acetate	1.211	0.841	30.6#	100	0.00
44 TM	Trichloroethene	0.335	0.321	4.2	102	0.00
45 C	1,2-Dichloropropane	0.334	0.322	3.6#	101	0.00
46 T	Dibromomethane	0.239	0.225	5.9	98	0.00
47 T	Bromodichloromethane	0.537	0.505	6.0	102	0.00
48 T	Methyl methacrylate	0.437	0.390	10.8	100	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080724\
 Data File : VN083142.D
 Acq On : 07 Aug 2024 14:49
 Operator : JC\MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ICSVN080724

Quant Time: Aug 08 06:35:30 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.008	0.007	12.5	93	0.00
50 S	Toluene-d8	1.164	1.215	-4.4	105	0.00
51 T	4-Methyl-2-Pentanone	0.500	0.462	7.6	97	0.00
52 CM	Toluene	0.889	0.862	3.0#	101	0.00
53 T	t-1,3-Dichloropropene	0.551	0.549	0.4	100	0.00
54 T	cis-1,3-Dichloropropene	0.586	0.568	3.1	102	0.00
55 T	1,1,2-Trichloroethane	0.318	0.308	3.1	101	0.00
56 T	Ethyl methacrylate	0.600	0.569	5.2	99	0.00
57 T	1,3-Dichloropropane	0.568	0.538	5.3	99	0.00
58 T	2-Chloroethyl Vinyl ether	0.254	0.249	2.0	104	0.00
59 T	2-Hexanone	0.387	0.348	10.1	96	0.00
60 T	Dibromochloromethane	0.385	0.381	1.0	100	0.00
61 T	1,2-Dibromoethane	0.335	0.315	6.0	99	0.00
62 S	4-Bromofluorobenzene	0.454	0.477	-5.1	106	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	109	0.00
64 T	Tetrachloroethene	0.331	0.319	3.6	102	0.00
65 PM	Chlorobenzene	1.105	1.050	5.0	100	0.00
66 T	1,1,1,2-Tetrachloroethane	0.390	0.365	6.4	100	0.00
67 C	Ethyl Benzene	2.027	1.936	4.5#	102	0.00
68 T	m/p-Xylenes	0.759	0.732	3.6	102	0.00
69 T	o-Xylene	0.749	0.699	6.7	99	0.00
70 T	Styrene	1.258	1.220	3.0	100	0.00
71 P	Bromoform	0.295	0.282	4.4	97	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	110	0.00
73 T	Isopropylbenzene	4.182	3.922	6.2	101	0.00
74 T	N-amyl acetate	2.046	1.790	12.5	98	0.00
75 P	1,1,2,2-Tetrachloroethane	1.183	1.018	13.9	96	0.00
76 T	1,2,3-Trichloropropane	1.097	0.924	15.8	95	0.00
77 T	Bromobenzene	0.929	0.867	6.7	99	0.00
78 T	n-propylbenzene	4.816	4.592	4.7	101	0.00
79 T	2-Chlorotoluene	3.053	2.833	7.2	101	0.00
80 T	1,3,5-Trimethylbenzene	3.501	3.338	4.7	102	0.00
81 T	trans-1,4-Dichloro-2-butene	0.504	0.475	5.8	105	0.00
82 T	4-Chlorotoluene	3.062	2.849	7.0	101	0.00
83 T	tert-Butylbenzene	3.101	2.863	7.7	99	0.00
84 T	1,2,4-Trimethylbenzene	3.529	3.281	7.0	98	0.00
85 T	sec-Butylbenzene	4.231	3.979	6.0	99	0.00
86 T	p-Isopropyltoluene	3.493	3.323	4.9	99	0.00
87 T	1,3-Dichlorobenzene	1.748	1.607	8.1	99	0.00
88 T	1,4-Dichlorobenzene	1.762	1.604	9.0	99	0.00
89 T	n-Butylbenzene	3.027	2.935	3.0	100	0.00
90 T	Hexachloroethane	0.675	0.631	6.5	100	0.00
91 T	1,2-Dichlorobenzene	1.692	1.531	9.5	98	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.287	0.240	16.4	93	0.00
93 T	1,2,4-Trichlorobenzene	0.948	0.873	7.9	97	0.00
94 T	Hexachlorobutadiene	0.421	0.365	13.3	98	0.00
95 T	Naphthalene	3.357	3.001	10.6	94	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080724\
Data File : VN083142.D
Acq On : 07 Aug 2024 14:49
Operator : JC\MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ICVVN080724

Quant Time: Aug 08 06:35:30 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.938	0.848	9.6	97	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080724\
 Data File : VN083142.D
 Acq On : 07 Aug 2024 14:49
 Operator : JC\MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ICVVN080724

Quant Time: Aug 08 06:35:30 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	109	0.00
2 T	Dichlorodifluoromethane	50.000	47.102	5.8	102	0.00
3 P	Chloromethane	50.000	46.238	7.5	99	0.00
4 C	Vinyl Chloride	50.000	45.937	8.1#	99	0.00
5 T	Bromomethane	50.000	42.984	14.0	97	0.00
6 T	Chloroethane	50.000	43.725	12.5	99	0.00
7 T	Trichlorofluoromethane	50.000	47.124	5.8	101	0.00
8 T	Diethyl Ether	50.000	46.259	7.5	100	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	47.176	5.6	101	0.00
10 T	Methyl Iodide	50.000	47.349	5.3	99	0.00
11 T	Tert butyl alcohol	250.000	208.135	16.7	92	0.00
12 CM	1,1-Dichloroethene	50.000	45.562	8.9#	101	0.00
13 T	Acrolein	250.000	217.571	13.0	90	0.00
14 T	Allyl chloride	50.000	41.173	17.7	101	0.00
15 T	Acrylonitrile	250.000	217.950	12.8	95	0.00
16 T	Acetone	250.000	210.840	15.7	90	0.00
17 T	Carbon Disulfide	50.000	43.689	12.6	99	0.00
18 T	Methyl Acetate	50.000	42.397	15.2	99	0.00
19 T	Methyl tert-butyl Ether	50.000	46.781	6.4	101	0.00
20 T	Methylene Chloride	50.000	43.645	12.7	102	0.00
21 T	trans-1,2-Dichloroethene	50.000	45.671	8.7	99	-0.01
22 T	Diisopropyl ether	50.000	46.881	6.2	100	0.00
23 T	Vinyl Acetate	250.000	235.414	5.8	99	0.00
24 P	1,1-Dichloroethane	50.000	46.685	6.6	101	0.00
25 T	2-Butanone	250.000	211.782	15.3	93	0.00
26 T	2,2-Dichloropropane	50.000	47.853	4.3	102	0.00
27 T	cis-1,2-Dichloroethene	50.000	46.113	7.8	101	0.00
28 T	Bromochloromethane	50.000	45.986	8.0	100	0.00
29 T	Tetrahydrofuran	250.000	222.147	11.1	96	0.00
30 C	Chloroform	50.000	47.387	5.2#	101	0.00
31 T	Cyclohexane	50.000	43.474	13.1	102	0.00
32 T	1,1,1-Trichloroethane	50.000	47.135	5.7	101	0.00
33 S	1,2-Dichloroethane-d4	50.000	50.583	-1.2	105	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	109	0.00
35 S	Dibromofluoromethane	50.000	51.549	-3.1	106	0.00
36 T	1,1-Dichloropropene	50.000	47.732	4.5	103	0.00
37 T	Ethyl Acetate	50.000	43.124	13.8	100	0.00
38 T	Carbon Tetrachloride	50.000	47.904	4.2	102	0.00
39 T	Methylcyclohexane	50.000	47.821	4.4	101	0.00
40 TM	Benzene	50.000	47.635	4.7	101	0.00
41 T	Methacrylonitrile	50.000	43.013	14.0	99	0.00
42 TM	1,2-Dichloroethane	50.000	47.310	5.4	102	0.00
43 T	Isopropyl Acetate	50.000	45.577	8.8	100	0.00
44 TM	Trichloroethene	50.000	47.994	4.0	102	0.00
45 C	1,2-Dichloropropane	50.000	48.208	3.6#	101	0.00
46 T	Dibromomethane	50.000	47.124	5.8	98	0.00
47 T	Bromodichloromethane	50.000	47.047	5.9	102	0.00
48 T	Methyl methacrylate	50.000	44.609	10.8	100	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080724\
 Data File : VN083142.D
 Acq On : 07 Aug 2024 14:49
 Operator : JC\MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ICSVN080724

Quant Time: Aug 08 06:35:30 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	885.211	11.5	93	0.00
50 S	Toluene-d8	50.000	52.175	-4.3	105	0.00
51 T	4-Methyl-2-Pentanone	250.000	230.917	7.6	97	0.00
52 CM	Toluene	50.000	48.520	3.0#	101	0.00
53 T	t-1,3-Dichloropropene	50.000	49.776	0.4	100	0.00
54 T	cis-1,3-Dichloropropene	50.000	48.487	3.0	102	0.00
55 T	1,1,2-Trichloroethane	50.000	48.351	3.3	101	0.00
56 T	Ethyl methacrylate	50.000	47.395	5.2	99	0.00
57 T	1,3-Dichloropropane	50.000	47.391	5.2	99	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	244.918	2.0	104	0.00
59 T	2-Hexanone	250.000	224.916	10.0	96	0.00
60 T	Dibromochloromethane	50.000	49.513	1.0	100	0.00
61 T	1,2-Dibromoethane	50.000	47.133	5.7	99	0.00
62 S	4-Bromofluorobenzene	50.000	52.574	-5.1	106	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	109	0.00
64 T	Tetrachloroethene	50.000	48.175	3.7	102	0.00
65 PM	Chlorobenzene	50.000	47.522	5.0	100	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	46.843	6.3	100	0.00
67 C	Ethyl Benzene	50.000	47.764	4.5#	102	0.00
68 T	m/p-Xylenes	100.000	96.434	3.6	102	0.00
69 T	o-Xylene	50.000	46.662	6.7	99	0.00
70 T	Styrene	50.000	48.487	3.0	100	0.00
71 P	Bromoform	50.000	47.797	4.4	97	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	110	0.00
73 T	Isopropylbenzene	50.000	46.893	6.2	101	0.00
74 T	N-amyl acetate	50.000	43.748	12.5	98	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	43.016	14.0	96	0.00
76 T	1,2,3-Trichloropropane	50.000	42.119	15.8	95	0.00
77 T	Bromobenzene	50.000	46.649	6.7	99	0.00
78 T	n-propylbenzene	50.000	47.678	4.6	101	0.00
79 T	2-Chlorotoluene	50.000	46.398	7.2	101	0.00
80 T	1,3,5-Trimethylbenzene	50.000	47.673	4.7	102	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	47.070	5.9	105	0.00
82 T	4-Chlorotoluene	50.000	46.512	7.0	101	0.00
83 T	tert-Butylbenzene	50.000	46.173	7.7	99	0.00
84 T	1,2,4-Trimethylbenzene	50.000	46.493	7.0	98	0.00
85 T	sec-Butylbenzene	50.000	47.018	6.0	99	0.00
86 T	p-Isopropyltoluene	50.000	47.566	4.9	99	0.00
87 T	1,3-Dichlorobenzene	50.000	45.959	8.1	99	0.00
88 T	1,4-Dichlorobenzene	50.000	45.499	9.0	99	0.00
89 T	n-Butylbenzene	50.000	48.489	3.0	100	0.00
90 T	Hexachloroethane	50.000	46.741	6.5	100	0.00
91 T	1,2-Dichlorobenzene	50.000	45.255	9.5	98	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	41.753	16.5	93	0.00
93 T	1,2,4-Trichlorobenzene	50.000	46.054	7.9	97	0.00
94 T	Hexachlorobutadiene	50.000	43.315	13.4	98	0.00
95 T	Naphthalene	50.000	44.692	10.6	94	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080724\
Data File : VN083142.D
Acq On : 07 Aug 2024 14:49
Operator : JC\MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ICVVN080724

Quant Time: Aug 08 06:35:30 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	45.221	9.6	97	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: JAC005

Lab Code: CHEM Case No.: P3657 SAS No.: P3657 SDG No.: P3657

Instrument ID: MSVOA_N Calibration Date/Time: 08/19/2024 11:43

Lab File ID: VN083367.D Init. Calib. Date(s): 08/07/2024 08/07/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 10:33 12:34

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.567	0.538		-5.11	20
Chloromethane	0.581	0.549	0.1	-5.51	20
Vinyl Chloride	0.592	0.595		0.51	20
Bromomethane	0.368	0.338		-8.15	20
1,1,2-Trichlorotrifluoroethane	0.539	0.570		5.75	20
Acetone	0.278	0.287		3.24	20
Carbon Disulfide	1.622	1.298		-19.98	20
Methyl tert-butyl Ether	2.001	2.094		4.65	20
Methylene Chloride	0.641	0.623		-2.81	20
trans-1,2-Dichloroethene	0.573	0.561		-2.09	20
Cyclohexane	1.055	0.926		-12.23	20
2-Butanone	0.428	0.438		2.34	20
Carbon Tetrachloride	0.532	0.570		7.14	20
cis-1,2-Dichloroethene	0.691	0.711		2.89	20
Chloroform	1.115	1.244		11.57	20
1,1,1-Trichloroethane	1.055	1.137		7.77	20
Methylcyclohexane	0.580	0.564		-2.76	20
Benzene	1.406	1.497		6.47	20
1,2-Dichloroethane	0.512	0.561		9.57	20
Trichloroethene	0.335	0.340		1.49	20
Bromodichloromethane	0.537	0.593		10.43	20
Toluene	0.889	0.961		8.1	20
1,1,2-Trichloroethane	0.318	0.365		14.78	20
Dibromochloromethane	0.385	0.450		16.88	20
Tetrachloroethene	0.331	0.347		4.83	20
Chlorobenzene	1.105	1.182	0.3	6.97	20
Ethyl Benzene	2.027	2.145		5.82	20
m/p-Xylenes	0.759	0.813		7.11	20
o-Xylene	0.749	0.796		6.28	20
Isopropylbenzene	4.182	4.153		-0.69	20
1,4-Dichlorobenzene	1.762	1.767		0.28	20
1,2-Dichlorobenzene	1.692	1.714		1.3	20
1,2-Dichloroethane-d4	0.712	0.763		7.16	20
Dibromofluoromethane	0.312	0.328		5.13	20
Toluene-d8	1.164	1.205		3.52	20
4-Bromofluorobenzene	0.454	0.477		5.07	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081924\
 Data File : VN083367.D
 Acq On : 19 Aug 2024 11:43
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 08/20/2024
 Supervised By :Mahesh Dadoda 08/20/2024

Quant Time: Aug 20 04:42:41 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	172122	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	293866	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	257989	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	128940	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	131386	53.628	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 107.260%			
35) Dibromofluoromethane	8.165	113	96420	52.566	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 105.140%			
50) Toluene-d8	10.565	98	353972	51.734	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 103.460%			
62) 4-Bromofluorobenzene	12.847	95	140275	52.587	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 105.180%			
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	2.124	85	92622	47.453	ug/l	98
3) Chloromethane	2.359	50	94447	47.263	ug/l	98
4) Vinyl Chloride	2.512	62	102460	50.248	ug/l	99
5) Bromomethane	2.948	94	58190	45.991	ug/l	92
6) Chloroethane	3.112	64	66873	52.420	ug/l	97
7) Trichlorofluoromethane	3.495	101	177971	52.834	ug/l	92
8) Diethyl Ether	3.959	74	63276	50.481	ug/l	86
9) 1,1,2-Trichlorotrifluo...	4.371	101	98140	52.844	ug/l	100
10) Methyl Iodide	4.583	142	107002	43.800	ug/l	96
11) Tert butyl alcohol	5.524	59	109673	215.405	ug/l	99
12) 1,1-Dichloroethene	4.342	96	92250	48.349	ug/l	91
13) Acrolein	4.177	56	34170	102.973	ug/l	100
14) Allyl chloride	5.024	41	172114	47.735	ug/l	89
15) Acrylonitrile	5.712	53	270822	258.631	ug/l	98
16) Acetone	4.424	43	247229	257.889	ug/l	100
17) Carbon Disulfide	4.712	76	223381	40.012	ug/l	99
18) Methyl Acetate	5.024	43	162749	56.980	ug/l	92
19) Methyl tert-butyl Ether	5.794	73	360356	52.325	ug/l	98
20) Methylene Chloride	5.277	84	107197	48.586	ug/l	88
21) trans-1,2-Dichloroethene	5.789	96	96613	48.992	ug/l	91
22) Diisopropyl ether	6.671	45	362023	53.415	ug/l	95
23) Vinyl Acetate	6.600	43	1826085m	262.861	ug/l	
24) 1,1-Dichloroethane	6.571	63	199796	54.083	ug/l	99
25) 2-Butanone	7.483	43	377107	256.187	ug/l	92
26) 2,2-Dichloropropane	7.488	77	190226	55.447	ug/l	100
27) cis-1,2-Dichloroethene	7.483	96	122308	51.398	ug/l	95
28) Bromochloromethane	7.812	49	77898	51.597	ug/l	87
29) Tetrahydrofuran	7.841	42	242146	254.476	ug/l	89
30) Chloroform	7.965	83	214076	55.777	ug/l	97
31) Cyclohexane	8.259	56	159458	43.919	ug/l	94
32) 1,1,1-Trichloroethane	8.171	97	195719	53.874	ug/l	96
36) 1,1-Dichloropropene	8.371	75	141011	50.820	ug/l	99
37) Ethyl Acetate	7.559	43	148708	47.866	ug/l #	96
38) Carbon Tetrachloride	8.359	117	167502	53.597	ug/l	97
39) Methylcyclohexane	9.600	83	165751	48.635	ug/l	99
40) Benzene	8.606	78	440008	53.231	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081924\
 Data File : VN083367.D
 Acq On : 19 Aug 2024 11:43
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 08/20/2024
 Supervised By :Mahesh Dadoda 08/20/2024

Quant Time: Aug 20 04:42:41 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.783	41	85769	48.555	ug/l	93
42) 1,2-Dichloroethane	8.671	62	164962	54.784	ug/l	98
43) Isopropyl Acetate	8.688	43	279244	51.960	ug/l #	94
44) Trichloroethene	9.353	130	99861	50.756	ug/l	84
45) 1,2-Dichloropropane	9.624	63	108858	55.480	ug/l	97
46) Dibromomethane	9.706	93	78829	56.129	ug/l	99
47) Bromodichloromethane	9.888	83	174328	55.282	ug/l	99
48) Methyl methacrylate	9.682	41	128812	50.151	ug/l	91
49) 1,4-Dioxane	9.694	88	46308	999.349	ug/l	96
51) 4-Methyl-2-Pentanone	10.447	43	785486	267.367	ug/l	92
52) Toluene	10.629	92	282304	54.053	ug/l	96
53) t-1,3-Dichloropropene	10.835	75	181838	56.137	ug/l	96
54) cis-1,3-Dichloropropene	10.312	75	184159	53.451	ug/l	90
55) 1,1,2-Trichloroethane	11.018	97	107387	57.375	ug/l	97
56) Ethyl methacrylate	10.876	69	185494	52.571	ug/l	88
57) 1,3-Dichloropropane	11.165	76	186473	55.896	ug/l	97
58) 2-Chloroethyl Vinyl ether	10.159	63	371615	249.160	ug/l	96
59) 2-Hexanone	11.194	43	589752	259.478	ug/l	92
60) Dibromochloromethane	11.359	129	132356	58.489	ug/l	97
61) 1,2-Dibromoethane	11.471	107	108241	55.051	ug/l	99
64) Tetrachloroethene	11.100	164	89501	52.383	ug/l	97
65) Chlorobenzene	11.894	112	304948	53.488	ug/l	99
66) 1,1,1,2-Tetrachloroethane	11.959	131	109957	54.685	ug/l	99
67) Ethyl Benzene	11.965	91	553315	52.905	ug/l	99
68) m/p-Xylenes	12.070	106	419315	107.029	ug/l	98
69) o-Xylene	12.400	106	205471	53.176	ug/l	99
70) Styrene	12.412	104	357557	55.096	ug/l	98
71) Bromoform	12.576	173	85444	56.093	ug/l #	98
73) Isopropylbenzene	12.694	105	535547	49.657	ug/l	100
74) N-amyl acetate	12.494	43	243002	46.064	ug/l	94
75) 1,1,2,2-Tetrachloroethane	12.935	83	151940	49.811	ug/l	98
76) 1,2,3-Trichloropropane	12.994	75	140765m	49.742	ug/l	
77) Bromobenzene	12.982	156	122155	50.988	ug/l	97
78) n-propylbenzene	13.035	91	637985	51.374	ug/l	100
79) 2-Chlorotoluene	13.123	91	396349	50.341	ug/l	98
80) 1,3,5-Trimethylbenzene	13.170	105	463266	51.308	ug/l	97
81) trans-1,4-Dichloro-2-b...	12.735	75	58672	45.099	ug/l	97
82) 4-Chlorotoluene	13.223	91	399817	50.626	ug/l	97
83) tert-Butylbenzene	13.435	119	401516	50.214	ug/l	100
84) 1,2,4-Trimethylbenzene	13.482	105	463259	50.911	ug/l	100
85) sec-Butylbenzene	13.617	105	556906	51.042	ug/l	99
86) p-Isopropyltoluene	13.729	119	467596	51.904	ug/l	98
87) 1,3-Dichlorobenzene	13.735	146	231442	51.340	ug/l	99
88) 1,4-Dichlorobenzene	13.812	146	227888	50.144	ug/l	98
89) n-Butylbenzene	14.059	91	410788	52.624	ug/l	99
90) Hexachloroethane	14.335	117	87386	50.208	ug/l	95
91) 1,2-Dichlorobenzene	14.106	146	220974	50.655	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	14.717	75	34831	47.057	ug/l	99
93) 1,2,4-Trichlorobenzene	15.394	180	123914	50.703	ug/l	99
94) Hexachlorobutadiene	15.500	225	50083	46.079	ug/l	98
95) Naphthalene	15.641	128	411840	47.573	ug/l	99
96) 1,2,3-Trichlorobenzene	15.841	180	120989	50.032	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081924\
Data File : VN083367.D
Acq On : 19 Aug 2024 11:43
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024

Quant Time: Aug 20 04:42:41 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed



```
Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081924\
Data File : VN083367.D
Acq On    : 19 Aug 2024  11:43
Operator  : JC\MD
Sample    : VSTDCCC050
Misc      : 5.0mL/MSVOA_N/WATER
ALS Vial  : 1    Sample Multiplier: 1
```


Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Manual Integrations

Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024


Quant Time: Aug 20 04:42:41 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 1066
Delta R.T. -0.000 min
Lab File: VN083367.D
Acq: 19 Aug 2024 11:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion:168 Resp: 17212
Ion Ratio Lower Upper
168 100
99 65.1 48.2 72.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#2
Dichlorodifluoromethane
Concen: 47.453 ug/l
RT: 2.124 min Scan# 29
Delta R.T. -0.000 min
Lab File: VN083367.D
Acq: 19 Aug 2024 11:43


Tgt Ion: 85 Resp: 92622
Ion Ratio Lower Upper
85 100
87 30.4 15.7 47.0





#3
Chloromethane
Concen: 47.263 ug/l
RT: 2.359 min Scan# 69
Delta R.T. -0.000 min
Lab File: VN083367.D
Acq: 19 Aug 2024 11:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 50 Resp: 9444
Ion Ratio Lower Upper
50 100
52 32.2 25.0 37.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#4
Vinyl Chloride
Concen: 50.248 ug/l
RT: 2.512 min Scan# 95
Delta R.T. -0.000 min
Lab File: VN083367.D
Acq: 19 Aug 2024 11:43


Tgt Ion: 62 Resp: 102460
Ion Ratio Lower Upper
62 100
64 32.0 25.0 37.6





#5
Bromomethane
Concen: 45.991 ug/l
RT: 2.948 min Scan# 169
Delta R.T. -0.000 min
Lab File: VN083367.D
Acq: 19 Aug 2024 11:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 94 Resp: 58190
Ion Ratio Lower Upper
94 100
96 90.0 78.0 117.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#6
Chloroethane
Concen: 52.420 ug/l
RT: 3.112 min Scan# 197
Delta R.T. -0.000 min
Lab File: VN083367.D
Acq: 19 Aug 2024 11:43


Tgt Ion: 64 Resp: 66873
Ion Ratio Lower Upper
64 100
66 31.7 26.6 40.0





#7
Trichlorofluoromethane
Concen: 52.834 ug/l
RT: 3.495 min Scan# 262
Delta R.T. -0.000 min
Lab File: VN083367.D
Acq: 19 Aug 2024 11:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion:101 Resp: 17797
Ion Ratio Lower Upper
101 100
103 59.8 52.7 79.1




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#8
Diethyl Ether
Concen: 50.481 ug/l
RT: 3.959 min Scan# 341
Delta R.T. 0.000 min
Lab File: VN083367.D
Acq: 19 Aug 2024 11:43

Tgt Ion: 74 Resp: 63276
Ion Ratio Lower Upper
74 100
45 96.4 55.5 166.3









#9
1,1,2-Trichlorotrifluoroethane
Concen: 52.844 ug/l
RT: 4.371 min Scan# 411
Delta R.T. 0.006 min
Lab File: VN083367.D
Acq: 19 Aug 2024 11:43

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tgt Ion:101 Resp: 98140
Ion Ratio Lower Upper
101 100
85 47.7 38.3 57.5
151 73.7 59.3 88.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#10
Methyl Iodide
Concen: 43.800 ug/l
RT: 4.583 min Scan# 447
Delta R.T. -0.006 min
Lab File: VN083367.D
Acq: 19 Aug 2024 11:43

Tgt Ion:142 Resp: 107002
Ion Ratio Lower Upper
142 100
127 49.5 37.5 56.3
141 14.5 13.1 19.7





#11

Tert butyl alcohol

Concen: 215.405 ug/l

RT: 5.524 min Scan# 608

Delta R.T. -0.006 min

Lab File: VN083367.D

Acq: 19 Aug 2024 11:43

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion: 59 Resp: 10967

Ion Ratio Lower Upper

59 100





57 10.7 8.9 13.3

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 08/20/2024

Supervised By :Mahesh Dadoda 08/20/2024



#12

1,1-Dichloroethene

Concen: 48.349 ug/l

RT: 4.342 min Scan# 406

Delta R.T. -0.000 min

Lab File: VN083367.D

Acq: 19 Aug 2024 11:43




Tgt Ion: 96 Resp: 92250


Ion Ratio Lower Upper

96 100

61 170.6 149.7 224.5


98 63.6 50.1 75.1





#13
Acrolein
Concen: 102.973 ug/l
RT: 4.177 min Scan# 31
Delta R.T. -0.006 min
Lab File: VN083367.D
Acq: 19 Aug 2024 11:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050




Tgt Ion: 56 Resp: 34170
Ion Ratio Lower Upper
56 100
55 70.4 56.4 84.6

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024




#14
Allyl chloride
Concen: 47.735 ug/l
RT: 5.024 min Scan# 522
Delta R.T. 0.006 min
Lab File: VN083367.D
Acq: 19 Aug 2024 11:43




Tgt Ion: 41 Resp: 172114
Ion Ratio Lower Upper
41 100
39 71.3 67.8 101.6
76 33.9 25.8 38.8





#15
Acrylonitrile
Concen: 258.631 ug/l
RT: 5.712 min Scan# 61
Delta R.T. -0.006 min
Lab File: VN083367.D
Acq: 19 Aug 2024 11:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 53 Resp: 270822
Ion Ratio Lower Upper
53 100
52 82.1 66.6 100.0
51 36.3 30.1 45.1




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#16
Acetone
Concen: 257.889 ug/l
RT: 4.424 min Scan# 420
Delta R.T. -0.006 min
Lab File: VN083367.D
Acq: 19 Aug 2024 11:43


Tgt Ion: 43 Resp: 247229
Ion Ratio Lower Upper
43 100
58 27.0 21.8 32.6





#17
Carbon Disulfide
Concen: 40.012 ug/l
RT: 4.712 min Scan# 469
Delta R.T. -0.000 min
Lab File: VN083367.D
Acq: 19 Aug 2024 11:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 76 Resp: 22338
Ion Ratio Lower Upper
76 100
78 8.7 7.2 10.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#18
Methyl Acetate
Concen: 56.980 ug/l
RT: 5.024 min Scan# 522
Delta R.T. -0.000 min
Lab File: VN083367.D
Acq: 19 Aug 2024 11:43

Tgt Ion: 43 Resp: 162749
Ion Ratio Lower Upper
43 100
74 24.1 16.3 24.5





#19

Methyl tert-butyl Ether

Concen: 52.325 ug/l

RT: 5.794 min Scan# 653

Delta R.T. -0.000 min

Lab File: VN083367.D

Acq: 19 Aug 2024 11:43

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion: 73 Resp: 360350

Ion Ratio Lower Upper

73 100





57 21.6 17.9 26.9

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 08/20/2024

Supervised By :Mahesh Dadoda 08/20/2024



#20

Methylene Chloride

Concen: 48.586 ug/l

RT: 5.277 min Scan# 565

Delta R.T. -0.000 min

Lab File: VN083367.D

Acq: 19 Aug 2024 11:43

Tgt Ion: 84 Resp: 107197




Ion Ratio Lower Upper


84 100

49 126.5 119.6 179.4

51 38.2 34.8 52.2


86 66.2 52.9 79.3





#21
trans-1,2-Dichloroethene
Concen: 48.992 ug/l
RT: 5.789 min Scan# 61
Delta R.T. -0.006 min
Lab File: VN083367.D
Acq: 19 Aug 2024 11:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 96 Resp: 9661
Ion Ratio Lower Upper
96 100
61 143.0 123.4 185.2
98 68.3 49.3 73.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#22
Diisopropyl ether
Concen: 53.415 ug/l
RT: 6.671 min Scan# 802
Delta R.T. -0.000 min
Lab File: VN083367.D
Acq: 19 Aug 2024 11:43

Tgt Ion: 45 Resp: 362023
Ion Ratio Lower Upper
45 100
43 57.7 44.0 66.0
87 27.9 19.7 29.5
59 12.2 8.5 12.7





#23

Vinyl Acetate

Concen: 262.861 ug/l m

RT: 6.600 min Scan# 791

Delta R.T. -0.006 min

Lab File: VN083367.D

Acq: 19 Aug 2024 11:43

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion: 43 Resp: 182608

Ion Ratio Lower Upper

43 100





86 10.2 7.3 10.9

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 08/20/2024

Supervised By :Mahesh Dadoda 08/20/2024



#24

1,1-Dichloroethane

Concen: 54.083 ug/l

RT: 6.571 min Scan# 785

Delta R.T. -0.000 min

Lab File: VN083367.D

Acq: 19 Aug 2024 11:43




Tgt Ion: 63 Resp: 199796


Ion Ratio Lower Upper

63 100

98 6.6 3.3 9.9


100 4.5 2.0 6.0





#25
2-Butanone
Concen: 256.187 ug/l
RT: 7.483 min Scan# 940
Delta R.T. -0.000 min
Lab File: VN083367.D
Acq: 19 Aug 2024 11:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050




Tgt Ion: 43 Resp: 377107
Ion Ratio Lower Upper
43 100
72 26.8 18.2 27.2

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024




#26
2,2-Dichloropropane
Concen: 55.447 ug/l
RT: 7.488 min Scan# 941
Delta R.T. 0.000 min
Lab File: VN083367.D
Acq: 19 Aug 2024 11:43




Tgt Ion: 77 Resp: 190226
Ion Ratio Lower Upper
77 100
97 20.5 10.3 30.9





#27
cis-1,2-Dichloroethene
Concen: 51.398 ug/l
RT: 7.483 min Scan# 941
Delta R.T. -0.006 min
Lab File: VN083367.D
Acq: 19 Aug 2024 11:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 96 Resp: 122308
Ion Ratio Lower Upper
96 100
61 155.2 0.0 329.6
98 64.6 0.0 128.6




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#28
Bromochloromethane
Concen: 51.597 ug/l
RT: 7.812 min Scan# 996
Delta R.T. -0.000 min
Lab File: VN083367.D
Acq: 19 Aug 2024 11:43

Tgt Ion: 49 Resp: 77898
Ion Ratio Lower Upper
49 100
129 1.8 0.0 3.8
130 73.1 50.5 75.7





#29

Tetrahydrofuran

Concen: 254.476 ug/l

RT: 7.841 min Scan# 1000

Delta R.T. 0.006 min

Lab File: VN083367.D

Acq: 19 Aug 2024 11:43

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion: 42 Resp: 242140

Ion Ratio Lower Upper

42 100

72 44.5 30.0 45.0





71 41.5 28.3 42.5

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 08/20/2024

Supervised By :Mahesh Dadoda 08/20/2024



#30

Chloroform

Concen: 55.777 ug/l

RT: 7.965 min Scan# 1022

Delta R.T. -0.000 min

Lab File: VN083367.D




Acq: 19 Aug 2024 11:43


Tgt Ion: 83 Resp: 214076

Ion Ratio Lower Upper

83 100

85 66.3 50.9 76.3





#31

Cyclohexane

Concen: 43.919 ug/l

RT: 8.259 min Scan# 1072

Delta R.T. -0.000 min

Lab File: VN083367.D


Acq: 19 Aug 2024 11:43

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050



Tgt Ion: 56 Resp: 159458

Ion Ratio Lower Upper

56 100

69 31.3 24.5 36.7




84 89.8 66.0 99.0

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 08/20/2024

Supervised By :Mahesh Dadoda 08/20/2024



#32

1,1,1-Trichloroethane

Concen: 53.874 ug/l

RT: 8.171 min Scan# 1057

Delta R.T. -0.000 min

Lab File: VN083367.D

Acq: 19 Aug 2024 11:43




Tgt Ion: 97 Resp: 195719


Ion Ratio Lower Upper

97 100

99 66.4 52.0 78.0

61 48.0 42.1 63.1





#33

1,2-Dichloroethane-d4

Concen: 53.628 ug/l

RT: 8.577 min Scan# 1126

Delta R.T. -0.005 min

Lab File: VN083367.D

Acq: 19 Aug 2024 11:43

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion: 65 Resp: 131386

Ion Ratio Lower Upper

65 100





67 48.8 0.0 104.4

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 08/20/2024

Supervised By :Mahesh Dadoda 08/20/2024



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.100 min Scan# 1215

Delta R.T. -0.000 min

Lab File: VN083367.D

Acq: 19 Aug 2024 11:43




Tgt Ion:114 Resp: 293866


Ion Ratio Lower Upper

114 100

63 22.2 0.0 44.6


88 16.8 0.0 31.4





#35
Dibromofluoromethane
Concen: 52.566 ug/l
RT: 8.165 min Scan# 1056
Delta R.T. -0.000 min
Lab File: VN083367.D
Acq: 19 Aug 2024 11:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 113 Resp: 96420
Ion Ratio Lower Upper
113 100
111 103.2 82.4 123.6
192 17.6 14.9 22.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#36
1,1-Dichloropropene
Concen: 50.820 ug/l
RT: 8.371 min Scan# 1091
Delta R.T. -0.000 min
Lab File: VN083367.D
Acq: 19 Aug 2024 11:43

Tgt Ion: 75 Resp: 141011
Ion Ratio Lower Upper
75 100
110 33.4 16.3 48.9
77 31.6 24.6 37.0





#37

Ethyl Acetate

Concen: 47.866 ug/l

RT: 7.559 min Scan# 91

Delta R.T. 0.000 min

Lab File: VN083367.D

Acq: 19 Aug 2024 11:43

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion: 43 Resp: 148708

Ion Ratio Lower Upper

43 100

61 12.8 9.3 13.9





70 10.8 7.0 10.6

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 08/20/2024

Supervised By :Mahesh Dadoda 08/20/2024



#38

Carbon Tetrachloride

Concen: 53.597 ug/l

RT: 8.359 min Scan# 1089

Delta R.T. 0.000 min

Lab File: VN083367.D

Acq: 19 Aug 2024 11:43




Tgt Ion:117 Resp: 167502


Ion Ratio Lower Upper

117 100

119 97.0 74.9 112.3

121 29.8 24.3 36.5





#39

Methylcyclohexane

Concen: 48.635 ug/l

RT: 9.600 min Scan# 11300

Delta R.T. -0.000 min

Lab File: VN083367.D

Acq: 19 Aug 2024 11:43

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion: 83 Resp: 16575

Ion Ratio Lower Upper

83 100

55 81.6 66.4 99.6





98 51.0 40.7 61.1

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 08/20/2024

Supervised By :Mahesh Dadoda 08/20/2024



#40

Benzene

Concen: 53.231 ug/l

RT: 8.606 min Scan# 1131

Delta R.T. -0.000 min

Lab File: VN083367.D




Acq: 19 Aug 2024 11:43


Tgt Ion: 78 Resp: 440008

Ion Ratio Lower Upper

78 100


77 23.4 19.0 28.4





#41
Methacrylonitrile
Concen: 48.555 ug/l
RT: 7.783 min Scan# 991
Delta R.T. 0.006 min
Lab File: VN083367.D
Acq: 19 Aug 2024 11:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 41 Resp: 85769
Ion Ratio Lower Upper
41 100
39 60.2 45.5 68.3
67 72.3 50.5 75.7
52 31.8 24.0 36.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#42
1,2-Dichloroethane
Concen: 54.784 ug/l
RT: 8.671 min Scan# 1142
Delta R.T. -0.000 min
Lab File: VN083367.D
Acq: 19 Aug 2024 11:43


Tgt Ion: 62 Resp: 164962
Ion Ratio Lower Upper
62 100
98 8.7 0.0 15.8





#43
Isopropyl Acetate
Concen: 51.960 ug/l
RT: 8.688 min Scan# 1145
Delta R.T. 0.000 min
Lab File: VN083367.D
Acq: 19 Aug 2024 11:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 43 Resp: 279244
Ion Ratio Lower Upper
43 100
61 24.7 17.8 26.6
87 12.7 8.2 12.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#44
Trichloroethene
Concen: 50.756 ug/l
RT: 9.353 min Scan# 1258
Delta R.T. 0.000 min
Lab File: VN083367.D
Acq: 19 Aug 2024 11:43

Tgt Ion:130 Resp: 99861
Ion Ratio Lower Upper
130 100
95 114.6 0.0 197.8





#45

1,2-Dichloropropane

Concen: 55.480 ug/l

RT: 9.624 min Scan# 11

Delta R.T. -0.000 min

Lab File: VN083367.D

Acq: 19 Aug 2024 11:43

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion: 63 Resp: 108858

Ion Ratio Lower Upper

63 100





65 32.4 24.4 36.6

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 08/20/2024

Supervised By :Mahesh Dadoda 08/20/2024



#46

Dibromomethane

Concen: 56.129 ug/l

RT: 9.706 min Scan# 1318

Delta R.T. -0.006 min

Lab File: VN083367.D

Acq: 19 Aug 2024 11:43




Tgt Ion: 93 Resp: 78829


Ion Ratio Lower Upper

93 100

95 82.9 65.8 98.6

174 89.0 71.7 107.5





#47

Bromodichloromethane

Concen: 55.282 ug/l

RT: 9.888 min Scan# 1349

Delta R.T. -0.000 min

Lab File: VN083367.D

Acq: 19 Aug 2024 11:43

Instrument :

MSVOA_N

ClientSampleId :





VSTDCCC050

Tgt Ion: 83	Resp: 174328
Ion Ratio	Lower Upper
83 100	
85 62.9	51.1 76.7
127 8.5	7.2 10.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#48

Methyl methacrylate

Concen: 50.151 ug/l




RT: 9.682 min Scan# 1314


Delta R.T. -0.000 min

Lab File: VN083367.D

Acq: 19 Aug 2024 11:43

Tgt Ion: 41	Resp: 128812
Ion Ratio	Lower Upper
41 100	
69 80.9	56.3 84.5
39 58.8	50.3 75.5





#49

1,4-Dioxane

Concen: 999.349 ug/l

RT: 9.694 min Scan# 1316

Delta R.T. -0.000 min

Lab File: VN083367.D

Acq: 19 Aug 2024 11:43

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion: 88 Resp: 46308

Ion Ratio Lower Upper

88 100

43 30.7 27.8 41.8





58 72.1 59.4 89.0

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 08/20/2024

Supervised By :Mahesh Dadoda 08/20/2024



#50

Toluene-d8

Concen: 51.734 ug/l

RT: 10.565 min Scan# 1464

Delta R.T. -0.000 min

Lab File: VN083367.D




Acq: 19 Aug 2024 11:43


Tgt Ion: 98 Resp: 353972

Ion Ratio Lower Upper

98 100


100 65.3 51.5 77.3





#51
4-Methyl-2-Pentanone
Concen: 267.367 ug/l
RT: 10.447 min Scan# 1444
Delta R.T. -0.000 min
Lab File: VN083367.D
Acq: 19 Aug 2024 11:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 43 Resp: 785480
Ion Ratio Lower Upper
43 100
58 39.8 28.2 42.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#52
Toluene
Concen: 54.053 ug/l
RT: 10.629 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VN083367.D
Acq: 19 Aug 2024 11:43


Tgt Ion: 92 Resp: 282304
Ion Ratio Lower Upper
92 100
91 168.9 139.4 209.0





#53
t-1,3-Dichloropropene
Concen: 56.137 ug/l
RT: 10.835 min Scan# 1510
Delta R.T. -0.000 min
Lab File: VN083367.D
Acq: 19 Aug 2024 11:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 75 Resp: 181838
Ion Ratio Lower Upper
75 100
77 30.1 25.7 38.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#54
cis-1,3-Dichloropropene
Concen: 53.451 ug/l
RT: 10.312 min Scan# 1421
Delta R.T. -0.000 min
Lab File: VN083367.D
Acq: 19 Aug 2024 11:43


Tgt Ion: 75 Resp: 184159
Ion Ratio Lower Upper
75 100
77 33.3 24.3 36.5
39 53.8 50.5 75.7





#55
1,1,2-Trichloroethane
Concen: 57.375 ug/l
RT: 11.018 min Scan# 1541
Delta R.T. 0.000 min
Lab File: VN083367.D
Acq: 19 Aug 2024 11:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050




Tgt Ion: 97 Resp: 10738
Ion Ratio Lower Upper
97 100
83 85.3 71.4 107.0
85 55.4 45.8 68.6
99 64.5 52.0 78.0

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024




#56
Ethyl methacrylate
Concen: 52.571 ug/l
RT: 10.876 min Scan# 1517
Delta R.T. 0.000 min
Lab File: VN083367.D
Acq: 19 Aug 2024 11:43




Tgt Ion: 69 Resp: 185494
Ion Ratio Lower Upper
69 100
41 67.5 63.4 95.2
39 41.2 37.4 56.0





#57
1,3-Dichloropropane
Concen: 55.896 ug/l
RT: 11.165 min Scan# 1566
Delta R.T. 0.000 min
Lab File: VN083367.D
Acq: 19 Aug 2024 11:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 76 Resp: 18647
Ion Ratio Lower Upper
76 100
78 33.5 25.6 38.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#58
2-Chloroethyl Vinyl ether
Concen: 249.160 ug/l
RT: 10.159 min Scan# 1395
Delta R.T. -0.000 min
Lab File: VN083367.D
Acq: 19 Aug 2024 11:43


Tgt Ion: 63 Resp: 371615
Ion Ratio Lower Upper
63 100
106 28.7 21.4 32.0





#59
2-Hexanone
Concen: 259.478 ug/l
RT: 11.194 min Scan# 1571
Delta R.T. -0.000 min
Lab File: VN083367.D
Acq: 19 Aug 2024 11:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 43 Resp: 589752
Ion Ratio Lower Upper
43 100
58 55.0 24.6 73.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#60
Dibromochloromethane
Concen: 58.489 ug/l
RT: 11.359 min Scan# 1599
Delta R.T. -0.000 min
Lab File: VN083367.D
Acq: 19 Aug 2024 11:43


Tgt Ion:129 Resp: 132356
Ion Ratio Lower Upper
129 100
127 76.1 39.2 117.6





#61
1,2-Dibromoethane
Concen: 55.051 ug/l
RT: 11.471 min Scan# 1618
Delta R.T. -0.000 min
Lab File: VN083367.D
Acq: 19 Aug 2024 11:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050




Tgt Ion: 107 Resp: 10824
Ion Ratio Lower Upper
107 100
109 92.9 73.7 110.5

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024




#62
4-Bromofluorobenzene
Concen: 52.587 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. -0.000 min
Lab File: VN083367.D
Acq: 19 Aug 2024 11:43




Tgt Ion: 95 Resp: 140275
Ion Ratio Lower Upper
95 100
174 71.3 0.0 159.2
176 67.9 0.0 147.6





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 1685
Delta R.T. -0.000 min
Lab File: VN083367.D
Acq: 19 Aug 2024 11:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion:117 Resp: 257989
Ion Ratio Lower Upper
117 100
82 59.9 47.5 71.3
119 32.5 25.6 38.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#64
Tetrachloroethene
Concen: 52.383 ug/l
RT: 11.100 min Scan# 1555
Delta R.T. -0.006 min
Lab File: VN083367.D
Acq: 19 Aug 2024 11:43


Tgt Ion:164 Resp: 89501
Ion Ratio Lower Upper
164 100
166 122.5 101.4 152.0
129 97.8 75.7 113.5
131 94.4 73.8 110.6





#65
Chlorobenzene
Concen: 53.488 ug/l
RT: 11.894 min Scan# 11
Delta R.T. -0.000 min
Lab File: VN083367.D
Acq: 19 Aug 2024 11:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050




Tgt Ion:112 Resp: 304948
Ion Ratio Lower Upper
112 100
114 33.0 26.2 39.2

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024




#66
1,1,1,2-Tetrachloroethane
Concen: 54.685 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. -0.006 min
Lab File: VN083367.D
Acq: 19 Aug 2024 11:43



Tgt Ion:131 Resp: 109957
Ion Ratio Lower Upper
131 100
133 95.1 47.3 142.0
119 64.0 32.5 97.4





#67

Ethyl Benzene

Concen: 52.905 ug/l

RT: 11.965 min Scan# 1702

Delta R.T. -0.000 min

Lab File: VN083367.D

Acq: 19 Aug 2024 11:43

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion: 91 Resp: 553315

Ion Ratio Lower Upper

91 100





106 30.3 24.0 36.0

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 08/20/2024

Supervised By :Mahesh Dadoda 08/20/2024



#68

m/p-Xylenes

Concen: 107.029 ug/l

RT: 12.070 min Scan# 1720

Delta R.T. -0.000 min

Lab File: VN083367.D




Acq: 19 Aug 2024 11:43


Tgt Ion:106 Resp: 419315

Ion Ratio Lower Upper

106 100

91 211.3 166.1 249.1









#69
o-Xylene
Concen: 53.176 ug/l
RT: 12.400 min Scan# 1776
Delta R.T. -0.000 min
Lab File: VN083367.D
Acq: 19 Aug 2024 11:43

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tgt Ion:106 Resp: 20547
Ion Ratio Lower Upper
106 100
91 220.2 111.3 333.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#70
Styrene
Concen: 55.096 ug/l
RT: 12.412 min Scan# 1778
Delta R.T. -0.000 min
Lab File: VN083367.D
Acq: 19 Aug 2024 11:43


Tgt Ion:104 Resp: 357557
Ion Ratio Lower Upper
104 100
78 52.8 41.6 62.4
103 53.7 44.0 66.0





#71
Bromoform
Concen: 56.093 ug/l
RT: 12.576 min Scan# 1806
Delta R.T. -0.000 min
Lab File: VN083367.D
Acq: 19 Aug 2024 11:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050




Tgt Ion:173 Resp: 85444
Ion Ratio Lower Upper
173 100
175 49.4 24.0 72.0
254 0.0 0.0 0.0

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.794 min Scan# 2013
Delta R.T. -0.000 min
Lab File: VN083367.D
Acq: 19 Aug 2024 11:43




Tgt Ion:152 Resp: 128940
Ion Ratio Lower Upper
152 100
115 63.6 30.6 91.6
150 174.0 0.0 348.6





#73
Isopropylbenzene
Concen: 49.657 ug/l
RT: 12.694 min Scan# 1826
Delta R.T. -0.000 min
Lab File: VN083367.D
Acq: 19 Aug 2024 11:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion:105 Resp: 53554
Ion Ratio Lower Upper
105 100
120 25.9 12.9 38.6




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#74
N-ethyl acetate
Concen: 46.064 ug/l
RT: 12.494 min Scan# 1792
Delta R.T. -0.000 min
Lab File: VN083367.D
Acq: 19 Aug 2024 11:43

Tgt Ion: 43 Resp: 243002
Ion Ratio Lower Upper
43 100
70 42.4 29.8 44.6
55 24.3 18.7 28.1
61 24.0 17.4 26.0









#75
1,1,2,2-Tetrachloroethane
Concen: 49.811 ug/l
RT: 12.935 min Scan# 1867
Delta R.T. -0.000 min
Lab File: VN083367.D
Acq: 19 Aug 2024 11:43

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tgt Ion: 83 Resp: 151940
Ion Ratio Lower Upper
83 100
131 10.6 5.0 14.9
85 65.5 31.9 95.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#76
1,2,3-Trichloropropane
Concen: 49.742 ug/l m
RT: 12.994 min Scan# 1877
Delta R.T. -0.000 min
Lab File: VN083367.D
Acq: 19 Aug 2024 11:43


Tgt Ion: 75 Resp: 140765
Ion Ratio Lower Upper
75 100
77 197.9 110.9 332.6





#77
Bromobenzene
Concen: 50.988 ug/l
RT: 12.982 min Scan# 1875
Delta R.T. -0.000 min
Lab File: VN083367.D
Acq: 19 Aug 2024 11:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion:156 Resp: 122158
Ion Ratio Lower Upper
156 100
77 228.0 111.3 333.8
158 97.7 48.3 144.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#78
n-propylbenzene
Concen: 51.374 ug/l
RT: 13.035 min Scan# 1884
Delta R.T. 0.000 min
Lab File: VN083367.D
Acq: 19 Aug 2024 11:43


Tgt Ion: 91 Resp: 637985
Ion Ratio Lower Upper
91 100
120 22.0 10.9 32.9





#79
2-Chlorotoluene
Concen: 50.341 ug/l
RT: 13.123 min Scan# 1899
Delta R.T. -0.000 min
Lab File: VN083367.D
Acq: 19 Aug 2024 11:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050




Tgt Ion: 91 Resp: 396349
Ion Ratio Lower Upper
91 100
126 31.1 16.2 48.6

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024




#80
1,3,5-Trimethylbenzene
Concen: 51.308 ug/l
RT: 13.170 min Scan# 1907
Delta R.T. 0.000 min
Lab File: VN083367.D
Acq: 19 Aug 2024 11:43




Tgt Ion:105 Resp: 463266
Ion Ratio Lower Upper
105 100
120 46.6 24.3 72.8





#81
trans-1,4-Dichloro-2-butene
Concen: 45.099 ug/l
RT: 12.735 min Scan# 1833
Delta R.T. 0.000 min
Lab File: VN083367.D
Acq: 19 Aug 2024 11:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 75 Resp: 58672
Ion Ratio Lower Upper
75 100
53 113.5 88.4 132.6
89 46.7 36.3 54.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#82
4-Chlorotoluene
Concen: 50.626 ug/l
RT: 13.223 min Scan# 1916
Delta R.T. -0.000 min
Lab File: VN083367.D
Acq: 19 Aug 2024 11:43

Tgt Ion: 91 Resp: 399817
Ion Ratio Lower Upper
91 100
126 30.9 16.2 48.6









#83
tert-Butylbenzene
Concen: 50.214 ug/l
RT: 13.435 min Scan# 1953
Delta R.T. -0.006 min
Lab File: VN083367.D
Acq: 19 Aug 2024 11:43

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tgt Ion:119 Resp: 401510
Ion Ratio Lower Upper
119 100
91 70.1 35.0 105.1
134 26.0 13.3 39.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#84
1,2,4-Trimethylbenzene
Concen: 50.911 ug/l
RT: 13.482 min Scan# 1960
Delta R.T. -0.000 min
Lab File: VN083367.D
Acq: 19 Aug 2024 11:43


Tgt Ion:105 Resp: 463259
Ion Ratio Lower Upper
105 100
120 43.9 21.9 65.8





#85
sec-Butylbenzene
Concen: 51.042 ug/l
RT: 13.617 min Scan# 1983
Delta R.T. -0.000 min
Lab File: VN083367.D
Acq: 19 Aug 2024 11:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion:105 Resp: 556900
Ion Ratio Lower Upper
105 100
134 19.1 9.8 29.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#86
p-Isopropyltoluene
Concen: 51.904 ug/l
RT: 13.729 min Scan# 2002
Delta R.T. -0.000 min
Lab File: VN083367.D
Acq: 19 Aug 2024 11:43

Tgt Ion:119 Resp: 467596
Ion Ratio Lower Upper
119 100
134 25.3 13.0 39.0
91 25.7 12.3 36.9





#87

1,3-Dichlorobenzene

Concen: 51.340 ug/l

RT: 13.735 min Scan# 2003

Delta R.T. -0.000 min

Lab File: VN083367.D

Acq: 19 Aug 2024 11:43

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion:146 Resp: 23144

Ion Ratio Lower Upper

146 100

111 42.9 20.6 61.9





148 63.8 31.8 95.3

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 08/20/2024

Supervised By :Mahesh Dadoda 08/20/2024



#88

1,4-Dichlorobenzene

Concen: 50.144 ug/l

RT: 13.812 min Scan# 2016

Delta R.T. -0.000 min

Lab File: VN083367.D

Acq: 19 Aug 2024 11:43




Tgt Ion:146 Resp: 227888


Ion Ratio Lower Upper

146 100

111 42.8 20.1 60.3

148 64.0 32.2 96.6





#89

n-Butylbenzene

Concen: 52.624 ug/l

RT: 14.059 min Scan# 2058

Delta R.T. -0.000 min

Lab File: VN083367.D

Acq: 19 Aug 2024 11:43

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion: 91 Resp: 41078

Ion Ratio Lower Upper

91 100

92 53.3 26.1 78.3





134 24.5 12.1 36.3

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 08/20/2024

Supervised By :Mahesh Dadoda 08/20/2024



#90

Hexachloroethane

Concen: 50.208 ug/l

RT: 14.335 min Scan# 2105

Delta R.T. -0.000 min

Lab File: VN083367.D




Acq: 19 Aug 2024 11:43


Tgt Ion:117 Resp: 87386

Ion Ratio Lower Upper

117 100


201 67.0 35.8 107.3





#91
1,2-Dichlorobenzene
Concen: 50.655 ug/l
RT: 14.106 min Scan# 2066
Delta R.T. 0.000 min
Lab File: VN083367.D
Acq: 19 Aug 2024 11:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion:146 Resp: 220974
Ion Ratio Lower Upper
146 100
111 45.2 21.4 64.2
148 63.3 32.1 96.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#92
1,2-Dibromo-3-Chloropropane
Concen: 47.057 ug/l
RT: 14.717 min Scan# 2170
Delta R.T. -0.000 min
Lab File: VN083367.D
Acq: 19 Aug 2024 11:43


Tgt Ion: 75 Resp: 34831
Ion Ratio Lower Upper
75 100
155 73.0 36.6 109.8
157 91.4 46.9 140.6





#93
1,2,4-Trichlorobenzene
Concen: 50.703 ug/l
RT: 15.394 min Scan# 2285
Delta R.T. -0.000 min
Lab File: VN083367.D
Acq: 19 Aug 2024 11:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050




Tgt Ion:180 Resp: 123914
Ion Ratio Lower Upper
180 100
182 95.0 47.3 141.8
145 33.5 15.7 47.1

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024




#94
Hexachlorobutadiene
Concen: 46.079 ug/l
RT: 15.500 min Scan# 2303
Delta R.T. -0.000 min
Lab File: VN083367.D
Acq: 19 Aug 2024 11:43




Tgt Ion:225 Resp: 50083
Ion Ratio Lower Upper
225 100
223 63.0 31.9 95.7
227 63.3 32.5 97.5





#95
Naphthalene
Concen: 47.573 ug/l
RT: 15.641 min Scan# 2327
Delta R.T. -0.000 min
Lab File: VN083367.D
Acq: 19 Aug 2024 11:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion:128 Resp: 411840
Ion Ratio Lower Upper
128 100
127 13.0 10.6 16.0
129 10.9 8.6 12.8




Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#96
1,2,3-Trichlorobenzene
Concen: 50.032 ug/l
RT: 15.841 min Scan# 2361
Delta R.T. 0.000 min
Lab File: VN083367.D
Acq: 19 Aug 2024 11:43

Tgt Ion:180 Resp: 120989
Ion Ratio Lower Upper
180 100
182 95.4 48.9 146.8
145 36.0 16.8 50.4



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081924\
 Data File : VN083367.D
 Acq On : 19 Aug 2024 11:43
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 20 04:42:41 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	84	0.00
2 T	Dichlorodifluoromethane	0.567	0.538	5.1	79	0.00
3 P	Chloromethane	0.581	0.549	5.5	78	0.00
4 C	Vinyl Chloride	0.592	0.595	-0.5#	84	0.00
5 T	Bromomethane	0.368	0.338	8.2	80	0.00
6 T	Chloroethane	0.371	0.389	-4.9	92	0.00
7 T	Trichlorofluoromethane	0.979	1.034	-5.6	87	0.00
8 T	Diethyl Ether	0.364	0.368	-1.1	84	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.539	0.570	-5.8	87	0.00
10 T	Methyl Iodide	0.710	0.622	12.4	71	0.00
11 T	Tert butyl alcohol	0.148	0.127	14.2	73	0.00
12 CM	1,1-Dichloroethene	0.554	0.536	3.2#	83	0.00
13 T	Acrolein	0.096	0.040	58.3#	33#	0.00
14 T	Allyl chloride	1.047	1.000	4.5	90	0.00
15 T	Acrylonitrile	0.304	0.315	-3.6	86	0.00
16 T	Acetone	0.278	0.287	-3.2	84	0.00
17 T	Carbon Disulfide	1.622	1.298	20.0	70	0.00
18 T	Methyl Acetate	0.830	0.946	-14.0	102	0.00
19 T	Methyl tert-butyl Ether	2.001	2.094	-4.6	87	0.00
20 T	Methylene Chloride	0.641	0.623	2.8	87	0.00
21 T	trans-1,2-Dichloroethene	0.573	0.561	2.1	82	0.00
22 T	Diisopropyl ether	1.969	2.103	-6.8	88	0.00
23 T	Vinyl Acetate	2.018	2.122	-5.2	85	0.00
24 P	1,1-Dichloroethane	1.073	1.161	-8.2	90	0.00
25 T	2-Butanone	0.428	0.438	-2.3	86	0.00
26 T	2,2-Dichloropropane	0.997	1.105	-10.8	91	0.00
27 T	cis-1,2-Dichloroethene	0.691	0.711	-2.9	87	0.00
28 T	Bromochloromethane	0.439	0.453	-3.2	87	0.00
29 T	Tetrahydrofuran	0.276	0.281	-1.8	85	0.00
30 C	Chloroform	1.115	1.244	-11.6#	92	0.00
31 T	Cyclohexane	1.055	0.926	12.2	79	0.00
32 T	1,1,1-Trichloroethane	1.055	1.137	-7.8	89	0.00
33 S	1,2-Dichloroethane-d4	0.712	0.763	-7.2	86	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	84	0.00
35 S	Dibromofluoromethane	0.312	0.328	-5.1	83	0.00
36 T	1,1-Dichloropropene	0.472	0.480	-1.7	85	0.00
37 T	Ethyl Acetate	0.529	0.506	4.3	85	0.00
38 T	Carbon Tetrachloride	0.532	0.570	-7.1	88	0.00
39 T	Methylcyclohexane	0.580	0.564	2.8	80	0.00
40 TM	Benzene	1.406	1.497	-6.5	87	0.00
41 T	Methacrylonitrile	0.301	0.292	3.0	87	0.00
42 TM	1,2-Dichloroethane	0.512	0.561	-9.6	91	0.00
43 T	Isopropyl Acetate	1.211	0.950	21.6	87	0.00
44 TM	Trichloroethene	0.335	0.340	-1.5	83	0.00
45 C	1,2-Dichloropropane	0.334	0.370	-10.8#	90	0.00
46 T	Dibromomethane	0.239	0.268	-12.1	90	0.00
47 T	Bromodichloromethane	0.537	0.593	-10.4	93	0.00
48 T	Methyl methacrylate	0.437	0.438	-0.2	87	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081924\
 Data File : VN083367.D
 Acq On : 19 Aug 2024 11:43
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 20 04:42:41 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.008	0.008	0.0	81	0.00
50 S	Toluene-d8	1.164	1.205	-3.5	80	0.00
51 T	4-Methyl-2-Pentanone	0.500	0.535	-7.0	87	0.00
52 CM	Toluene	0.889	0.961	-8.1#	87	0.00
53 T	t-1,3-Dichloropropene	0.551	0.619	-12.3	87	0.00
54 T	cis-1,3-Dichloropropene	0.586	0.627	-7.0	87	0.00
55 T	1,1,2-Trichloroethane	0.318	0.365	-14.8	92	0.00
56 T	Ethyl methacrylate	0.600	0.631	-5.2	85	0.00
57 T	1,3-Dichloropropane	0.568	0.635	-11.8	90	0.00
58 T	2-Chloroethyl Vinyl ether	0.254	0.253	0.4	82	0.00
59 T	2-Hexanone	0.387	0.401	-3.6	85	0.00
60 T	Dibromochloromethane	0.385	0.450	-16.9	92	0.00
61 T	1,2-Dibromoethane	0.335	0.368	-9.9	89	0.00
62 S	4-Bromofluorobenzene	0.454	0.477	-5.1	82	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	85	0.00
64 T	Tetrachloroethene	0.331	0.347	-4.8	86	0.00
65 PM	Chlorobenzene	1.105	1.182	-7.0	88	0.00
66 T	1,1,1,2-Tetrachloroethane	0.390	0.426	-9.2	91	0.00
67 C	Ethyl Benzene	2.027	2.145	-5.8#	88	0.00
68 T	m/p-Xylenes	0.759	0.813	-7.1	88	0.00
69 T	o-Xylene	0.749	0.796	-6.3	88	0.00
70 T	Styrene	1.258	1.386	-10.2	88	0.00
71 P	Bromoform	0.295	0.331	-12.2	89	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	90	0.00
73 T	Isopropylbenzene	4.182	4.153	0.7	88	0.00
74 T	N-amyl acetate	2.046	1.885	7.9	85	0.00
75 P	1,1,2,2-Tetrachloroethane	1.183	1.178	0.4	91	0.00
76 T	1,2,3-Trichloropropane	1.097	1.092	0.5	92	0.00
77 T	Bromobenzene	0.929	0.947	-1.9	89	0.00
78 T	n-propylbenzene	4.816	4.948	-2.7	90	0.00
79 T	2-Chlorotoluene	3.053	3.074	-0.7	90	0.00
80 T	1,3,5-Trimethylbenzene	3.501	3.593	-2.6	90	0.00
81 T	trans-1,4-Dichloro-2-butene	0.504	0.455	9.7	83	0.00
82 T	4-Chlorotoluene	3.062	3.101	-1.3	90	0.00
83 T	tert-Butylbenzene	3.101	3.114	-0.4	89	0.00
84 T	1,2,4-Trimethylbenzene	3.529	3.593	-1.8	88	0.00
85 T	sec-Butylbenzene	4.231	4.319	-2.1	89	0.00
86 T	p-Isopropyltoluene	3.493	3.626	-3.8	89	0.00
87 T	1,3-Dichlorobenzene	1.748	1.795	-2.7	91	0.00
88 T	1,4-Dichlorobenzene	1.762	1.767	-0.3	89	0.00
89 T	n-Butylbenzene	3.027	3.186	-5.3	89	0.00
90 T	Hexachloroethane	0.675	0.678	-0.4	88	0.00
91 T	1,2-Dichlorobenzene	1.692	1.714	-1.3	90	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.287	0.270	5.9	86	0.00
93 T	1,2,4-Trichlorobenzene	0.948	0.961	-1.4	88	0.00
94 T	Hexachlorobutadiene	0.421	0.388	7.8	85	0.00
95 T	Naphthalene	3.357	3.194	4.9	82	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081924\
Data File : VN083367.D
Acq On : 19 Aug 2024 11:43
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Aug 20 04:42:41 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.938	0.938	0.0	88	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081924\
Data File : VN083367.D
Acq On : 19 Aug 2024 11:43
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Aug 20 04:42:41 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	84	0.00
2 T	Dichlorodifluoromethane	50.000	47.453	5.1	79	0.00
3 P	Chloromethane	50.000	47.263	5.5	78	0.00
4 C	Vinyl Chloride	50.000	50.248	-0.5#	84	0.00
5 T	Bromomethane	50.000	45.991	8.0	80	0.00
6 T	Chloroethane	50.000	52.420	-4.8	92	0.00
7 T	Trichlorofluoromethane	50.000	52.834	-5.7	87	0.00
8 T	Diethyl Ether	50.000	50.481	-1.0	84	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	52.844	-5.7	87	0.00
10 T	Methyl Iodide	50.000	43.800	12.4	71	0.00
11 T	Tert butyl alcohol	250.000	215.405	13.8	73	0.00
12 CM	1,1-Dichloroethene	50.000	48.349	3.3#	83	0.00
13 T	Acrolein	250.000	102.973	58.8#	33	0.00
14 T	Allyl chloride	50.000	47.735	4.5	90	0.00
15 T	Acrylonitrile	250.000	258.631	-3.5	86	0.00
16 T	Acetone	250.000	257.889	-3.2	84	0.00
17 T	Carbon Disulfide	50.000	40.012	20.0	70	0.00
18 T	Methyl Acetate	50.000	56.980	-14.0	102	0.00
19 T	Methyl tert-butyl Ether	50.000	52.325	-4.7	87	0.00
20 T	Methylene Chloride	50.000	48.586	2.8	87	0.00
21 T	trans-1,2-Dichloroethene	50.000	48.992	2.0	82	0.00
22 T	Diisopropyl ether	50.000	53.415	-6.8	88	0.00
23 T	Vinyl Acetate	250.000	262.861	-5.1	85	0.00
24 P	1,1-Dichloroethane	50.000	54.083	-8.2	90	0.00
25 T	2-Butanone	250.000	256.187	-2.5	86	0.00
26 T	2,2-Dichloropropane	50.000	55.447	-10.9	91	0.00
27 T	cis-1,2-Dichloroethene	50.000	51.398	-2.8	87	0.00
28 T	Bromochloromethane	50.000	51.597	-3.2	87	0.00
29 T	Tetrahydrofuran	250.000	254.476	-1.8	85	0.00
30 C	Chloroform	50.000	55.777	-11.6#	92	0.00
31 T	Cyclohexane	50.000	43.919	12.2	79	0.00
32 T	1,1,1-Trichloroethane	50.000	53.874	-7.7	89	0.00
33 S	1,2-Dichloroethane-d4	50.000	53.628	-7.3	86	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	84	0.00
35 S	Dibromofluoromethane	50.000	52.566	-5.1	83	0.00
36 T	1,1-Dichloropropene	50.000	50.820	-1.6	85	0.00
37 T	Ethyl Acetate	50.000	47.866	4.3	85	0.00
38 T	Carbon Tetrachloride	50.000	53.597	-7.2	88	0.00
39 T	Methylcyclohexane	50.000	48.635	2.7	80	0.00
40 TM	Benzene	50.000	53.231	-6.5	87	0.00
41 T	Methacrylonitrile	50.000	48.555	2.9	87	0.00
42 TM	1,2-Dichloroethane	50.000	54.784	-9.6	91	0.00
43 T	Isopropyl Acetate	50.000	51.960	-3.9	87	0.00
44 TM	Trichloroethene	50.000	50.756	-1.5	83	0.00
45 C	1,2-Dichloropropane	50.000	55.480	-11.0#	90	0.00
46 T	Dibromomethane	50.000	56.129	-12.3	90	0.00
47 T	Bromodichloromethane	50.000	55.282	-10.6	93	0.00
48 T	Methyl methacrylate	50.000	50.151	-0.3	87	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081924\
 Data File : VN083367.D
 Acq On : 19 Aug 2024 11:43
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 20 04:42:41 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	999.349	0.1	81	0.00
50 S	Toluene-d8	50.000	51.734	-3.5	80	0.00
51 T	4-Methyl-2-Pentanone	250.000	267.367	-6.9	87	0.00
52 CM	Toluene	50.000	54.053	-8.1#	87	0.00
53 T	t-1,3-Dichloropropene	50.000	56.137	-12.3	87	0.00
54 T	cis-1,3-Dichloropropene	50.000	53.451	-6.9	87	0.00
55 T	1,1,2-Trichloroethane	50.000	57.375	-14.8	92	0.00
56 T	Ethyl methacrylate	50.000	52.571	-5.1	85	0.00
57 T	1,3-Dichloropropane	50.000	55.896	-11.8	90	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	249.160	0.3	82	0.00
59 T	2-Hexanone	250.000	259.478	-3.8	85	0.00
60 T	Dibromochloromethane	50.000	58.489	-17.0	92	0.00
61 T	1,2-Dibromoethane	50.000	55.051	-10.1	89	0.00
62 S	4-Bromofluorobenzene	50.000	52.587	-5.2	82	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	85	0.00
64 T	Tetrachloroethene	50.000	52.383	-4.8	86	0.00
65 PM	Chlorobenzene	50.000	53.488	-7.0	88	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	54.685	-9.4	91	0.00
67 C	Ethyl Benzene	50.000	52.905	-5.8#	88	0.00
68 T	m/p-Xylenes	100.000	107.029	-7.0	88	0.00
69 T	o-Xylene	50.000	53.176	-6.4	88	0.00
70 T	Styrene	50.000	55.096	-10.2	88	0.00
71 P	Bromoform	50.000	56.093	-12.2	89	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	90	0.00
73 T	Isopropylbenzene	50.000	49.657	0.7	88	0.00
74 T	N-amyl acetate	50.000	46.064	7.9	85	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	49.811	0.4	91	0.00
76 T	1,2,3-Trichloropropane	50.000	49.742	0.5	92	0.00
77 T	Bromobenzene	50.000	50.988	-2.0	89	0.00
78 T	n-propylbenzene	50.000	51.374	-2.7	90	0.00
79 T	2-Chlorotoluene	50.000	50.341	-0.7	90	0.00
80 T	1,3,5-Trimethylbenzene	50.000	51.308	-2.6	90	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	45.099	9.8	83	0.00
82 T	4-Chlorotoluene	50.000	50.626	-1.3	90	0.00
83 T	tert-Butylbenzene	50.000	50.214	-0.4	89	0.00
84 T	1,2,4-Trimethylbenzene	50.000	50.911	-1.8	88	0.00
85 T	sec-Butylbenzene	50.000	51.042	-2.1	89	0.00
86 T	p-Isopropyltoluene	50.000	51.904	-3.8	89	0.00
87 T	1,3-Dichlorobenzene	50.000	51.340	-2.7	91	0.00
88 T	1,4-Dichlorobenzene	50.000	50.144	-0.3	89	0.00
89 T	n-Butylbenzene	50.000	52.624	-5.2	89	0.00
90 T	Hexachloroethane	50.000	50.208	-0.4	88	0.00
91 T	1,2-Dichlorobenzene	50.000	50.655	-1.3	90	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	47.057	5.9	86	0.00
93 T	1,2,4-Trichlorobenzene	50.000	50.703	-1.4	88	0.00
94 T	Hexachlorobutadiene	50.000	46.079	7.8	85	0.00
95 T	Naphthalene	50.000	47.573	4.9	82	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081924\
Data File : VN083367.D
Acq On : 19 Aug 2024 11:43
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Aug 20 04:42:41 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	50.032	-0.1	88	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6





QC SAMPLE DATA

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080724\
Data File : VN083134.D
Acq On : 07 Aug 2024 10:09
Operator : JC\MD
Sample : BFB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Title : SW846 8260
Last Update : Thu Aug 08 06:30:41 2024



AutoFind: Scans 1851, 1852, 1853; Background Corrected with Scan 1843



Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	19.7	9690	PASS
75	95	30	60	52.6	25907	PASS
95	95	100	100	100.0	49259	PASS
96	95	5	9	7.5	3697	PASS
173	174	0.00	2	0.3	99	PASS
174	95	50	100	66.5	32749	PASS
175	174	5	9	7.4	2414	PASS
176	174	95	101	96.9	31720	PASS
177	176	5	9	6.3	1999	PASS

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081924\
Data File : VN083366.D
Acq On : 19 Aug 2024 10:59
Operator : JC\MD
Sample : BFB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Title : SW846 8260
Last Update : Thu Aug 08 06:30:41 2024



AutoFind: Scans 1851, 1852, 1853; Background Corrected with Scan 1843

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	19.3	8998	PASS
75	95	30	60	53.4	24840	PASS
95	95	100	100	100.0	46504	PASS
96	95	5	9	6.5	3046	PASS
173	174	0.00	2	0.9	278	PASS
174	95	50	100	66.8	31069	PASS
175	174	5	9	8.0	2479	PASS
176	174	95	101	99.1	30787	PASS
177	176	5	9	6.7	2054	PASS

Report of Analysis

Client:	JACOBS Engineering Group, Inc.	Date Collected:	
Project:	Former Schlumberger Site Princeton NJ	Date Received:	
Client Sample ID:	VN0819WBL01	SDG No.:	P3657
Lab Sample ID:	VN0819WBL01	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group6
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN083369.D	1		08/19/24 13:37	VN081924

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.21	U	0.21	1.00	ug/L
74-87-3	Chloromethane	0.35	U	0.35	1.00	ug/L
75-01-4	Vinyl Chloride	0.34	U	0.34	1.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	1.00	ug/L
67-64-1	Acetone	1.40	U	1.40	5.00	ug/L
75-15-0	Carbon Disulfide	0.32	U	0.32	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	1.00	ug/L
75-09-2	Methylene Chloride	0.32	U	0.32	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.25	U	0.25	1.00	ug/L
110-82-7	Cyclohexane	1.60	U	1.60	5.00	ug/L
78-93-3	2-Butanone	1.30	U	1.30	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.25	U	0.25	1.00	ug/L
67-66-3	Chloroform	0.26	U	0.26	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.19	U	0.19	1.00	ug/L
108-87-2	Methylcyclohexane	0.19	U	0.19	1.00	ug/L
71-43-2	Benzene	0.16	U	0.16	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.24	U	0.24	1.00	ug/L
79-01-6	Trichloroethene	0.32	U	0.32	1.00	ug/L
75-27-4	Bromodichloromethane	0.24	U	0.24	1.00	ug/L
108-88-3	Toluene	0.18	U	0.18	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	1.00	ug/L
124-48-1	Dibromochloromethane	0.18	U	0.18	1.00	ug/L
127-18-4	Tetrachloroethene	0.25	U	0.25	1.00	ug/L
108-90-7	Chlorobenzene	0.13	U	0.13	1.00	ug/L
100-41-4	Ethyl Benzene	0.16	U	0.16	1.00	ug/L
179601-23-1	m/p-Xylenes	0.31	U	0.31	2.00	ug/L
1330-20-7	Total Xylenes	0.45	U	0.45	3.00	ug/L
95-47-6	o-Xylene	0.14	U	0.14	1.00	ug/L

Report of Analysis

Client:	JACOBS Engineering Group, Inc.	Date Collected:	
Project:	Former Schlumberger Site Princeton NJ	Date Received:	
Client Sample ID:	VN0819WBL01	SDG No.:	P3657
Lab Sample ID:	VN0819WBL01	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5	Units:	mL
Soil Aliquot Vol:			uL
GC Column:	RXI-624	ID :	0.25
Prep Method :		Level :	LOW

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN083369.D	1		08/19/24 13:37	VN081924

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
98-82-8	Isopropylbenzene	0.13	U	0.13	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.27	U	0.27	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.19	U	0.19	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	51.4		70 (74) - 130 (125)	103%	SPK: 50
1868-53-7	Dibromofluoromethane	48.4		70 (75) - 130 (124)	97%	SPK: 50
2037-26-5	Toluene-d8	48.4		70 (86) - 130 (113)	97%	SPK: 50
460-00-4	4-Bromofluorobenzene	51.6		70 (77) - 130 (121)	103%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	145000	8.224			
540-36-3	1,4-Difluorobenzene	283000	9.1			
3114-55-4	Chlorobenzene-d5	291000	11.865			
3855-82-1	1,4-Dichlorobenzene-d4	134000	13.794			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of OC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081924\
 Data File : VN083369.D
 Acq On : 19 Aug 2024 13:37
 Operator : JC\MD
 Sample : VN0819WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0819WBL01

Quant Time: Aug 20 04:43:59 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	144938	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	283183	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	290527	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	134032	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	105966	51.364	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	102.720%
35) Dibromofluoromethane	8.171	113	85532	48.390	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.780%
50) Toluene-d8	10.565	98	319405	48.443	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.880%
62) 4-Bromofluorobenzene	12.847	95	132664	51.610	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	103.220%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081924\
Data File : VN083369.D
Acq On : 19 Aug 2024 13:37
Operator : JC\MD
Sample : VN0819WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0819WBL01

Quant Time: Aug 20 04:43:59 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 1066
Delta R.T. 0.000 min
Lab File: VN083369.D
Acq: 19 Aug 2024 13:37




Instrument :
MSVOA_N
ClientSampleId :
VN0819WBL01


Tgt Ion:168 Resp: 144938
Ion Ratio Lower Upper
168 100
99 68.6 48.2 72.4



#33
1,2-Dichloroethane-d4
Concen: 51.364 ug/l
RT: 8.583 min Scan# 1127
Delta R.T. 0.001 min
Lab File: VN083369.D
Acq: 19 Aug 2024 13:37

Tgt Ion: 65 Resp: 105966
Ion Ratio Lower Upper
65 100
67 49.6 0.0 104.4





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.100 min Scan# 1

Delta R.T. 0.000 min

Lab File: VN083369.D

Acq: 19 Aug 2024 13:37

Instrument :

MSVOA_N

ClientSampleId :

VN0819WBL01





Tgt Ion:114 Resp: 283183

Ion Ratio Lower Upper

114 100

63 22.1 0.0 44.6

88 17.5 0.0 31.4



#35

Dibromofluoromethane

Concen: 48.390 ug/l

RT: 8.171 min Scan# 1057

Delta R.T. 0.006 min

Lab File: VN083369.D

Acq: 19 Aug 2024 13:37




Tgt Ion:113 Resp: 85532


Ion Ratio Lower Upper

113 100

111 103.4 82.4 123.6

192 16.9 14.9 22.3









#50
Toluene-d8
Concen: 48.443 ug/l
RT: 10.565 min Scan# 1464
Delta R.T. 0.000 min
Lab File: VN083369.D
Acq: 19 Aug 2024 13:37




Instrument :
MSVOA_N
ClientSampleId :
VN0819WBL01


Tgt Ion: 98 Resp: 319405
Ion Ratio Lower Upper
98 100
100 64.1 51.5 77.3



#62
4-Bromofluorobenzene
Concen: 51.610 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. 0.000 min
Lab File: VN083369.D
Acq: 19 Aug 2024 13:37

Tgt Ion: 95 Resp: 132664
Ion Ratio Lower Upper
95 100
174 70.4 0.0 159.2
176 67.5 0.0 147.6





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.865 min Scan# 1

Delta R.T. 0.000 min

Lab File: VN083369.D

Acq: 19 Aug 2024 13:37

Instrument :

MSVOA_N

ClientSampleId :

VN0819WBL01





Tgt Ion:117 Resp: 290527

Ion Ratio Lower Upper

117 100

82 59.2 47.5 71.3

119 31.8 25.6 38.4



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.794 min Scan# 2013

Delta R.T. 0.000 min

Lab File: VN083369.D

Acq: 19 Aug 2024 13:37




Tgt Ion:152 Resp: 134032

Ion Ratio Lower Upper

152 100

115 63.6 30.6 91.6

150 159.0 0.0 348.6



Report of Analysis

Client:	JACOBS Engineering Group, Inc.	Date Collected:	
Project:	Former Schlumberger Site Princeton NJ	Date Received:	
Client Sample ID:	VN0819WBS01	SDG No.:	P3657
Lab Sample ID:	VN0819WBS01	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group6
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN083370.D	1		08/19/24 14:14	VN081924

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	17.6		0.21	1.00	ug/L
74-87-3	Chloromethane	18.0		0.35	1.00	ug/L
75-01-4	Vinyl Chloride	18.2		0.34	1.00	ug/L
74-83-9	Bromomethane	17.4		1.40	5.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	19.7		0.25	1.00	ug/L
67-64-1	Acetone	94.6		1.40	5.00	ug/L
75-15-0	Carbon Disulfide	14.8		0.32	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	19.3		0.16	1.00	ug/L
75-09-2	Methylene Chloride	17.9		0.32	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	17.8		0.25	1.00	ug/L
110-82-7	Cyclohexane	17.3		1.60	5.00	ug/L
78-93-3	2-Butanone	96.8		1.30	5.00	ug/L
56-23-5	Carbon Tetrachloride	19.3		0.25	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	19.0		0.25	1.00	ug/L
67-66-3	Chloroform	20.9		0.26	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	20.4		0.19	1.00	ug/L
108-87-2	Methylcyclohexane	16.6		0.19	1.00	ug/L
71-43-2	Benzene	19.3		0.16	1.00	ug/L
107-06-2	1,2-Dichloroethane	20.2		0.24	1.00	ug/L
79-01-6	Trichloroethene	19.2		0.32	1.00	ug/L
75-27-4	Bromodichloromethane	19.9		0.24	1.00	ug/L
108-88-3	Toluene	19.1		0.18	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	20.5		0.21	1.00	ug/L
124-48-1	Dibromochloromethane	20.9		0.18	1.00	ug/L
127-18-4	Tetrachloroethene	19.1		0.25	1.00	ug/L
108-90-7	Chlorobenzene	19.2		0.13	1.00	ug/L
100-41-4	Ethyl Benzene	18.7		0.16	1.00	ug/L
179601-23-1	m/p-Xylenes	37.1		0.31	2.00	ug/L
1330-20-7	Total Xylenes	55.3		0.45	3.00	ug/L
95-47-6	o-Xylene	18.2		0.14	1.00	ug/L

Report of Analysis

Client:	JACOBS Engineering Group, Inc.	Date Collected:	
Project:	Former Schlumberger Site Princeton NJ	Date Received:	
Client Sample ID:	VN0819WBS01	SDG No.:	P3657
Lab Sample ID:	VN0819WBS01	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group6
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN083370.D	1		08/19/24 14:14	VN081924

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
98-82-8	Isopropylbenzene	18.8		0.13	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	18.6		0.27	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	19.2		0.19	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	56.0		70 (74) - 130 (125)	112%	SPK: 50
1868-53-7	Dibromofluoromethane	54.2		70 (75) - 130 (124)	108%	SPK: 50
2037-26-5	Toluene-d8	53.3		70 (86) - 130 (113)	107%	SPK: 50
460-00-4	4-Bromofluorobenzene	52.4		70 (77) - 130 (121)	105%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	162000	8.224			
540-36-3	1,4-Difluorobenzene	285000	9.1			
3114-55-4	Chlorobenzene-d5	251000	11.865			
3855-82-1	1,4-Dichlorobenzene-d4	117000	13.794			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081924\
 Data File : VN083370.D
 Acq On : 19 Aug 2024 14:14
 Operator : JC\MD
 Sample : VN0819WBS01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0819WBS01

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 08/20/2024
 Supervised By :Mahesh Dadoda 08/20/2024

Quant Time: Aug 20 04:44:22 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	162452	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	284677	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	251348	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	116728	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	129591	56.044	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 112.080%			
35) Dibromofluoromethane	8.171	113	96237	54.160	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 108.320%			
50) Toluene-d8	10.570	98	353604	53.348	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 106.700%			
62) 4-Bromofluorobenzene	12.847	95	135412	52.403	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 104.800%			
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	2.124	85	32436	17.607	ug/l	98
3) Chloromethane	2.359	50	33987	18.020	ug/l	94
4) Vinyl Chloride	2.512	62	35092	18.234	ug/l	95
5) Bromomethane	2.959	94	20728	17.358	ug/l	95
6) Chloroethane	3.118	64	24629	20.455	ug/l	99
7) Trichlorofluoromethane	3.500	101	61473	19.336	ug/l	95
8) Diethyl Ether	3.965	74	22007	18.602	ug/l	87
9) 1,1,2-Trichlorotrifluo...	4.371	101	34503	19.684	ug/l	98
10) Methyl Iodide	4.589	142	34826	15.104	ug/l	93
11) Tert butyl alcohol	5.530	59	42745	88.951	ug/l	100
12) 1,1-Dichloroethene	4.341	96	32275	17.922	ug/l	93
13) Acrolein	4.177	56	11293	36.058	ug/l	96
14) Allyl chloride	5.024	41	59807	17.574	ug/l	89
15) Acrylonitrile	5.724	53	97877	99.035	ug/l	99
16) Acetone	4.430	43	85624	94.633	ug/l	97
17) Carbon Disulfide	4.718	76	78191	14.839	ug/l	99
18) Methyl Acetate	5.024	43	59002	21.887	ug/l	94
19) Methyl tert-butyl Ether	5.800	73	125513	19.310	ug/l	95
20) Methylene Chloride	5.277	84	37348	17.935	ug/l	92
21) trans-1,2-Dichloroethene	5.788	96	33098	17.783	ug/l	94
22) Diisopropyl ether	6.671	45	125566	19.629	ug/l	97
23) Vinyl Acetate	6.606	43	627341m	95.680	ug/l	
24) 1,1-Dichloroethane	6.565	63	70035	20.086	ug/l	95
25) 2-Butanone	7.482	43	134423	96.756	ug/l	94
26) 2,2-Dichloropropane	7.488	77	61621	19.030	ug/l	99
27) cis-1,2-Dichloroethene	7.494	96	42567	18.953	ug/l	94
28) Bromochloromethane	7.812	49	27807	19.515	ug/l	90
29) Tetrahydrofuran	7.841	42	87293	97.199	ug/l #	87
30) Chloroform	7.965	83	75639	20.881	ug/l	98
31) Cyclohexane	8.253	56	59209	17.279	ug/l	94
32) 1,1,1-Trichloroethane	8.171	97	69922	20.393	ug/l	96
36) 1,1-Dichloropropene	8.371	75	47501	17.672	ug/l	99
37) Ethyl Acetate	7.559	43	54801	18.209	ug/l	96
38) Carbon Tetrachloride	8.359	117	58337	19.269	ug/l	99
39) Methylcyclohexane	9.600	83	54862	16.617	ug/l	100
40) Benzene	8.606	78	154182	19.255	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081924\
 Data File : VN083370.D
 Acq On : 19 Aug 2024 14:14
 Operator : JC\MD
 Sample : VN0819WBS01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0819WBS01

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 08/20/2024
 Supervised By :Mahesh Dadoda 08/20/2024

Quant Time: Aug 20 04:44:22 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.777	41	29307	17.126	ug/l	92
42) 1,2-Dichloroethane	8.671	62	58855	20.177	ug/l	99
43) Isopropyl Acetate	8.688	43	100017	17.725	ug/l #	95
44) Trichloroethene	9.353	130	36588	19.197	ug/l	94
45) 1,2-Dichloropropane	9.623	63	38361	20.182	ug/l	100
46) Dibromomethane	9.706	93	27162	19.965	ug/l	98
47) Bromodichloromethane	9.888	83	60730	19.880	ug/l	99
48) Methyl methacrylate	9.682	41	44290	17.800	ug/l	91
49) 1,4-Dioxane	9.700	88	17487	389.559	ug/l	98
51) 4-Methyl-2-Pentanone	10.447	43	285593	100.349	ug/l	93
52) Toluene	10.629	92	96572	19.088	ug/l	98
53) t-1,3-Dichloropropene	10.835	75	59863	19.077	ug/l	95
54) cis-1,3-Dichloropropene	10.312	75	63119	18.911	ug/l	92
55) 1,1,2-Trichloroethane	11.018	97	37184	20.508	ug/l	97
56) Ethyl methacrylate	10.876	69	62670	18.335	ug/l #	88
57) 1,3-Dichloropropane	11.165	76	64783	20.046	ug/l	99
58) 2-Chloroethyl Vinyl ether	10.159	63	136029	94.148	ug/l	96
59) 2-Hexanone	11.200	43	213145	96.806	ug/l	93
60) Dibromochloromethane	11.365	129	45859	20.920	ug/l	96
61) 1,2-Dibromoethane	11.470	107	37704	19.795	ug/l	96
64) Tetrachloroethene	11.106	164	31771	19.086	ug/l	91
65) Chlorobenzene	11.894	112	106774	19.223	ug/l	98
66) 1,1,1,2-Tetrachloroethane	11.959	131	38674	19.742	ug/l	99
67) Ethyl Benzene	11.965	91	190504	18.696	ug/l	97
68) m/p-Xylenes	12.070	106	141650	37.111	ug/l	98
69) o-Xylene	12.400	106	68572	18.215	ug/l	99
70) Styrene	12.412	104	118040	18.669	ug/l	98
71) Bromoform	12.582	173	28210	19.009	ug/l #	99
73) Isopropylbenzene	12.694	105	183779	18.823	ug/l	99
74) N-amyl acetate	12.494	43	79945	16.740	ug/l	93
75) 1,1,2,2-Tetrachloroethane	12.941	83	54585	19.767	ug/l	100
76) 1,2,3-Trichloropropane	12.994	75	51001m	19.908	ug/l	
77) Bromobenzene	12.982	156	41748	19.249	ug/l	96
78) n-propylbenzene	13.035	91	215301	19.151	ug/l	100
79) 2-Chlorotoluene	13.123	91	132186	18.546	ug/l	99
80) 1,3,5-Trimethylbenzene	13.176	105	155831	19.064	ug/l	98
81) trans-1,4-Dichloro-2-b...	12.735	75	20629	17.516	ug/l	88
82) 4-Chlorotoluene	13.223	91	133028	18.607	ug/l	98
83) tert-Butylbenzene	13.441	119	134819	18.624	ug/l	98
84) 1,2,4-Trimethylbenzene	13.482	105	156307	18.975	ug/l	99
85) sec-Butylbenzene	13.617	105	184059	18.634	ug/l	100
86) p-Isopropyltoluene	13.729	119	152949	18.754	ug/l	97
87) 1,3-Dichlorobenzene	13.735	146	77562	19.005	ug/l	98
88) 1,4-Dichlorobenzene	13.811	146	76574	18.612	ug/l	97
89) n-Butylbenzene	14.059	91	130544	18.473	ug/l	100
90) Hexachloroethane	14.335	117	29820	18.926	ug/l	90
91) 1,2-Dichlorobenzene	14.106	146	75859	19.209	ug/l	97
92) 1,2-Dibromo-3-Chloropr...	14.717	75	12055	17.990	ug/l	96
93) 1,2,4-Trichlorobenzene	15.394	180	39877	18.024	ug/l	97
94) Hexachlorobutadiene	15.505	225	16689	16.961	ug/l	98
95) Naphthalene	15.641	128	138298	17.647	ug/l	98
96) 1,2,3-Trichlorobenzene	15.841	180	39972	18.259	ug/l	96

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081924\
Data File : VN083370.D
Acq On : 19 Aug 2024 14:14
Operator : JC\MD
Sample : VN0819WBS01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0819WBS01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024

Quant Time: Aug 20 04:44:22 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081924\
 Data File : VN083370.D
 Acq On : 19 Aug 2024 14:14
 Operator : JC\MD
 Sample : VN0819WBS01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1


Instrument :
 MSVOA_N
 ClientSampleId :
 VN0819WBS01

Manual Integrations APPROVED

Reviewed By : Semsettin Yesilyurt 08/20/2024
 Supervised By : Mahesh Dadoda 08/20/2024


Quant Time: Aug 20 04:44:22 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 1066
Delta R.T. -0.000 min
Lab File: VN083370.D
Acq: 19 Aug 2024 14:14




Instrument :
MSVOA_N
ClientSampleId :
VN0819WBS01



Tgt Ion:168 Resp: 162452
Ion Ratio Lower Upper
168 100
99 64.6 48.2 72.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#2
Dichlorodifluoromethane
Concen: 17.607 ug/l
RT: 2.124 min Scan# 29
Delta R.T. -0.000 min
Lab File: VN083370.D
Acq: 19 Aug 2024 14:14


Tgt Ion: 85 Resp: 32436
Ion Ratio Lower Upper
85 100
87 30.4 15.7 47.0





#3
Chloromethane
Concen: 18.020 ug/l
RT: 2.359 min Scan# 69
Delta R.T. -0.000 min
Lab File: VN083370.D
Acq: 19 Aug 2024 14:14




Instrument :
MSVOA_N
ClientSampleId :
VN0819WBS01



Tgt Ion: 50 Resp: 3398
Ion Ratio Lower Upper
50 100
52 34.6 25.0 37.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#4
Vinyl Chloride
Concen: 18.234 ug/l
RT: 2.512 min Scan# 95
Delta R.T. -0.000 min
Lab File: VN083370.D
Acq: 19 Aug 2024 14:14


Tgt Ion: 62 Resp: 35092
Ion Ratio Lower Upper
62 100
64 34.1 25.0 37.6





#5
Bromomethane
Concen: 17.358 ug/l
RT: 2.959 min Scan# 11
Delta R.T. 0.012 min
Lab File: VN083370.D
Acq: 19 Aug 2024 14:14



Instrument :
MSVOA_N
ClientSampleId :
VN0819WBS01



Tgt Ion: 94 Resp: 20728
Ion Ratio Lower Upper
94 100
96 92.9 78.0 117.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#6
Chloroethane
Concen: 20.455 ug/l
RT: 3.118 min Scan# 198
Delta R.T. 0.006 min
Lab File: VN083370.D
Acq: 19 Aug 2024 14:14


Tgt Ion: 64 Resp: 24629
Ion Ratio Lower Upper
64 100
66 33.9 26.6 40.0





#7
Trichlorofluoromethane
Concen: 19.336 ug/l
RT: 3.500 min Scan# 20
Delta R.T. 0.005 min
Lab File: VN083370.D
Acq: 19 Aug 2024 14:14




Instrument :
MSVOA_N
ClientSampleId :
VN0819WBS01



Tgt Ion:101 Resp: 6147
Ion Ratio Lower Upper
101 100
103 62.3 52.7 79.1




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#8
Diethyl Ether
Concen: 18.602 ug/l
RT: 3.965 min Scan# 342
Delta R.T. 0.006 min
Lab File: VN083370.D
Acq: 19 Aug 2024 14:14

Tgt Ion: 74 Resp: 22007
Ion Ratio Lower Upper
74 100
45 97.0 55.5 166.3









#9
1,1,2-Trichlorotrifluoroethane
Concen: 19.684 ug/l
RT: 4.371 min Scan# 411
Delta R.T. 0.006 min
Lab File: VN083370.D
Acq: 19 Aug 2024 14:14

Instrument :
MSVOA_N
ClientSampleId :
VN0819WBS01

Tgt Ion	Ratio	Lower	Upper
101	100		
85	46.8	38.3	57.5
151	72.5	59.3	88.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#10
Methyl Iodide
Concen: 15.104 ug/l
RT: 4.589 min Scan# 448
Delta R.T. -0.000 min
Lab File: VN083370.D
Acq: 19 Aug 2024 14:14

Tgt Ion	Ratio	Lower	Upper
142	100		
127	52.9	37.5	56.3
141	16.7	13.1	19.7





#11

Tert butyl alcohol

Concen: 88.951 ug/l

RT: 5.530 min Scan# 608

Delta R.T. -0.000 min

Lab File: VN083370.D

Acq: 19 Aug 2024 14:14

Instrument :

MSVOA_N

ClientSampleId :

VN0819WBS01

Tgt Ion: 59 Resp: 42745

Ion Ratio Lower Upper

59 100





57 11.1 8.9 13.3

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 08/20/2024

Supervised By :Mahesh Dadoda 08/20/2024



#12

1,1-Dichloroethene

Concen: 17.922 ug/l

RT: 4.341 min Scan# 406

Delta R.T. -0.000 min

Lab File: VN083370.D

Acq: 19 Aug 2024 14:14




Tgt Ion: 96 Resp: 32275


Ion Ratio Lower Upper

96 100

61 174.4 149.7 224.5

98 60.2 50.1 75.1





#13

Acrolein

Concen: 36.058 ug/l

RT: 4.177 min Scan# 31

Delta R.T. -0.006 min

Lab File: VN083370.D

Acq: 19 Aug 2024 14:14

Instrument :

MSVOA_N

ClientSampleId :

VN0819WBS01

Tgt Ion: 56 Resp: 1129

Ion Ratio Lower Upper

56 100





55 74.0 56.4 84.6

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 08/20/2024

Supervised By :Mahesh Dadoda 08/20/2024



#14

Allyl chloride

Concen: 17.574 ug/l

RT: 5.024 min Scan# 522

Delta R.T. 0.006 min

Lab File: VN083370.D

Acq: 19 Aug 2024 14:14




Tgt Ion: 41 Resp: 59807


Ion Ratio Lower Upper

41 100

39 71.2 67.8 101.6


76 32.5 25.8 38.8





#15
Acrylonitrile
Concen: 99.035 ug/l
RT: 5.724 min Scan# 641
Delta R.T. 0.006 min
Lab File: VN083370.D
Acq: 19 Aug 2024 14:14




Instrument :
MSVOA_N
ClientSampleId :
VN0819WBS01



Tgt Ion: 53 Resp: 9787
Ion Ratio Lower Upper
53 100
52 82.6 66.6 100.0
51 36.9 30.1 45.1




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#16
Acetone
Concen: 94.633 ug/l
RT: 4.430 min Scan# 421
Delta R.T. -0.000 min
Lab File: VN083370.D
Acq: 19 Aug 2024 14:14


Tgt Ion: 43 Resp: 85624
Ion Ratio Lower Upper
43 100
58 28.9 21.8 32.6





#17
Carbon Disulfide
Concen: 14.839 ug/l
RT: 4.718 min Scan# 41
Delta R.T. 0.006 min
Lab File: VN083370.D
Acq: 19 Aug 2024 14:14




Instrument :
MSVOA_N
ClientSampleId :
VN0819WBS01



Tgt Ion: 76 Resp: 7819
Ion Ratio Lower Upper
76 100
78 9.5 7.2 10.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#18
Methyl Acetate
Concen: 21.887 ug/l
RT: 5.024 min Scan# 522
Delta R.T. -0.000 min
Lab File: VN083370.D
Acq: 19 Aug 2024 14:14

Tgt Ion: 43 Resp: 59002
Ion Ratio Lower Upper
43 100
74 23.4 16.3 24.5









#19
Methyl tert-butyl Ether
Concen: 19.310 ug/l
RT: 5.800 min Scan# 61
Delta R.T. 0.006 min
Lab File: VN083370.D
Acq: 19 Aug 2024 14:14

Instrument :
MSVOA_N
ClientSampleId :
VN0819WBS01

Tgt Ion: 73 Resp: 12551
Ion Ratio Lower Upper
73 100
57 20.2 17.9 26.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#20
Methylene Chloride
Concen: 17.935 ug/l
RT: 5.277 min Scan# 565
Delta R.T. -0.000 min
Lab File: VN083370.D
Acq: 19 Aug 2024 14:14


Tgt Ion: 84 Resp: 37348
Ion Ratio Lower Upper
84 100
49 134.3 119.6 179.4
51 40.1 34.8 52.2
86 66.5 52.9 79.3





#21
trans-1,2-Dichloroethene
Concen: 17.783 ug/l
RT: 5.788 min Scan# 61
Delta R.T. -0.006 min
Lab File: VN083370.D
Acq: 19 Aug 2024 14:14




Instrument :
MSVOA_N
ClientSampleId :
VN0819WBS01



Tgt Ion: 96 Resp: 33098
Ion Ratio Lower Upper
96 100
61 148.2 123.4 185.2
98 68.4 49.3 73.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#22
Diisopropyl ether
Concen: 19.629 ug/l
RT: 6.671 min Scan# 802
Delta R.T. -0.000 min
Lab File: VN083370.D
Acq: 19 Aug 2024 14:14


Tgt Ion: 45 Resp: 125566
Ion Ratio Lower Upper
45 100
43 58.1 44.0 66.0
87 25.5 19.7 29.5
59 11.4 8.5 12.7





#23
 Vinyl Acetate
 Concen: 95.680 ug/l m
 RT: 6.606 min Scan# 791
 Delta R.T. -0.000 min
 Lab File: VN083370.D
 Acq: 19 Aug 2024 14:14




Instrument :
 MSVOA_N
 ClientSampleId :
 VN0819WBS01



Tgt Ion: 43 Resp: 62734
 Ion Ratio Lower Upper
 43 100
 86 10.7 7.3 10.9




Manual Integrations
 APPROVED


Reviewed By :Semsettin Yesilyurt 08/20/2024
 Supervised By :Mahesh Dadoda 08/20/2024



#24
 1,1-Dichloroethane
 Concen: 20.086 ug/l
 RT: 6.565 min Scan# 784
 Delta R.T. -0.006 min
 Lab File: VN083370.D
 Acq: 19 Aug 2024 14:14


Tgt Ion: 63 Resp: 70035
 Ion Ratio Lower Upper
 63 100
 98 8.2 3.3 9.9
 100 5.6 2.0 6.0





#25
2-Butanone
Concen: 96.756 ug/l
RT: 7.482 min Scan# 940
Delta R.T. -0.000 min
Lab File: VN083370.D
Acq: 19 Aug 2024 14:14




Instrument :
MSVOA_N
ClientSampleId :
VN0819WBS01



Tgt Ion: 43 Resp: 13442
Ion Ratio Lower Upper
43 100
72 25.7 18.2 27.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#26
2,2-Dichloropropane
Concen: 19.030 ug/l
RT: 7.488 min Scan# 941
Delta R.T. -0.000 min
Lab File: VN083370.D
Acq: 19 Aug 2024 14:14


Tgt Ion: 77 Resp: 61621
Ion Ratio Lower Upper
77 100
97 21.0 10.3 30.9





#27
cis-1,2-Dichloroethene
Concen: 18.953 ug/l
RT: 7.494 min Scan# 941
Delta R.T. 0.006 min
Lab File: VN083370.D
Acq: 19 Aug 2024 14:14




Instrument :
MSVOA_N
ClientSampleId :
VN0819WBS01



Tgt Ion: 96 Resp: 4256
Ion Ratio Lower Upper
96 100
61 154.8 0.0 329.6
98 63.7 0.0 128.6




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#28
Bromochloromethane
Concen: 19.515 ug/l
RT: 7.812 min Scan# 996
Delta R.T. -0.000 min
Lab File: VN083370.D
Acq: 19 Aug 2024 14:14

Tgt Ion: 49 Resp: 27807
Ion Ratio Lower Upper
49 100
129 1.3 0.0 3.8
130 71.4 50.5 75.7





#29

Tetrahydrofuran

Concen: 97.199 ug/l

RT: 7.841 min Scan# 1000

Delta R.T. 0.006 min

Lab File: VN083370.D

Acq: 19 Aug 2024 14:14

Instrument :

MSVOA_N

ClientSampleId :





VN0819WBS01

Tgt Ion:	42	Resp:	8729
Ion Ratio	Lower	Upper	
42	100		
72	45.6	30.0	45.0
71	42.9	28.3	42.5

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 08/20/2024
 Supervised By :Mahesh Dadoda 08/20/2024



#30

Chloroform

Concen: 20.881 ug/l




RT: 7.965 min Scan# 1022


Delta R.T. -0.000 min

Lab File: VN083370.D

Acq: 19 Aug 2024 14:14


Tgt Ion:	83	Resp:	75639
Ion Ratio	Lower	Upper	
83	100		
85	62.3	50.9	76.3





#31
Cyclohexane
Concen: 17.279 ug/l
RT: 8.253 min Scan# 1071
Delta R.T. -0.006 min
Lab File: VN083370.D
Acq: 19 Aug 2024 14:14




Instrument :
MSVOA_N
ClientSampleId :
VN0819WBS01



Tgt Ion: 56 Resp: 59209
Ion Ratio Lower Upper
56 100
69 31.5 24.5 36.7
84 89.9 66.0 99.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#32
1,1,1-Trichloroethane
Concen: 20.393 ug/l
RT: 8.171 min Scan# 1057
Delta R.T. -0.000 min
Lab File: VN083370.D
Acq: 19 Aug 2024 14:14

Tgt Ion: 97 Resp: 69922
Ion Ratio Lower Upper
97 100
99 65.6 52.0 78.0
61 46.6 42.1 63.1





#33

1,2-Dichloroethane-d4

Concen: 56.044 ug/l

RT: 8.582 min Scan# 1127

Delta R.T. 0.000 min

Lab File: VN083370.D

Acq: 19 Aug 2024 14:14

Instrument :

MSVOA_N

ClientSampleId :

VN0819WBS01

Tgt Ion: 65 Resp: 12959

Ion Ratio Lower Upper

65 100





67 52.5 0.0 104.4

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 08/20/2024

Supervised By :Mahesh Dadoda 08/20/2024



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.100 min Scan# 1215

Delta R.T. -0.000 min

Lab File: VN083370.D

Acq: 19 Aug 2024 14:14




Tgt Ion:114 Resp: 284677


Ion Ratio Lower Upper

114 100

63 22.3 0.0 44.6

88 17.3 0.0 31.4





#35

Dibromofluoromethane

Concen: 54.160 ug/l

RT: 8.171 min Scan# 1056

Delta R.T. 0.006 min

Lab File: VN083370.D

Acq: 19 Aug 2024 14:14

Instrument :

MSVOA_N

ClientSampleId :





VN0819WBS01

Tgt Ion:	113	Resp:	9623
Ion Ratio	Lower	Upper	
113	100		
111	102.7	82.4	123.6
192	17.3	14.9	22.3

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#36

1,1-Dichloropropene

Concen: 17.672 ug/l




RT: 8.371 min Scan# 1091


Delta R.T. -0.000 min

Lab File: VN083370.D

Acq: 19 Aug 2024 14:14

Tgt Ion:	75	Resp:	47501
Ion Ratio	Lower	Upper	
75	100		
110	33.1	16.3	48.9
77	30.9	24.6	37.0





#37

Ethyl Acetate

Concen: 18.209 ug/l

RT: 7.559 min Scan# 91

Delta R.T. -0.000 min

Lab File: VN083370.D

Acq: 19 Aug 2024 14:14

Instrument :

MSVOA_N





ClientSampleId :

VN0819WBS01

Tgt Ion:	43	Resp:	5480
Ion	Ratio	Lower	Upper
43	100		
61	13.3	9.3	13.9
70	10.3	7.0	10.6

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#38

Carbon Tetrachloride

Concen: 19.269 ug/l




RT: 8.359 min Scan# 1089


Delta R.T. -0.000 min

Lab File: VN083370.D

Acq: 19 Aug 2024 14:14

Tgt Ion:	117	Resp:	58337
Ion	Ratio	Lower	Upper
117	100		
119	94.9	74.9	112.3
121	30.7	24.3	36.5





#39

Methylcyclohexane

Concen: 16.617 ug/l

RT: 9.600 min Scan# 11300

Delta R.T. -0.000 min

Lab File: VN083370.D

Acq: 19 Aug 2024 14:14

Instrument :

MSVOA_N

ClientSampleId :





VN0819WBS01

Tgt Ion:	83	Resp:	5486
Ion	Ratio	Lower	Upper
83	100		
55	83.3	66.4	99.6
98	50.7	40.7	61.1

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#40

Benzene

Concen: 19.255 ug/l




RT: 8.606 min Scan# 1131


Delta R.T. -0.000 min

Lab File: VN083370.D

Acq: 19 Aug 2024 14:14


Tgt Ion:	78	Resp:	154182
Ion	Ratio	Lower	Upper
78	100		
77	22.8	19.0	28.4





#41
Methacrylonitrile
Concen: 17.126 ug/l
RT: 7.777 min Scan# 990
Delta R.T. -0.000 min
Lab File: VN083370.D
Acq: 19 Aug 2024 14:14




Instrument :
MSVOA_N
ClientSampleId :
VN0819WBS01



Tgt Ion: 41 Resp: 2930
Ion Ratio Lower Upper
41 100
39 59.0 45.5 68.3
67 74.5 50.5 75.7
52 33.0 24.0 36.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#42
1,2-Dichloroethane
Concen: 20.177 ug/l
RT: 8.671 min Scan# 1142
Delta R.T. -0.000 min
Lab File: VN083370.D
Acq: 19 Aug 2024 14:14


Tgt Ion: 62 Resp: 58855
Ion Ratio Lower Upper
62 100
98 8.4 0.0 15.8





#43
Isopropyl Acetate
Concen: 17.725 ug/l
RT: 8.688 min Scan# 1145
Delta R.T. 0.000 min
Lab File: VN083370.D
Acq: 19 Aug 2024 14:14




Instrument :
MSVOA_N
ClientSampleId :
VN0819WBS01



Tgt Ion: 43 Resp: 10001
Ion Ratio Lower Upper
43 100
61 24.7 17.8 26.6
87 12.3 8.2 12.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#44
Trichloroethene
Concen: 19.197 ug/l
RT: 9.353 min Scan# 1258
Delta R.T. -0.000 min
Lab File: VN083370.D
Acq: 19 Aug 2024 14:14

Tgt Ion:130 Resp: 36588
Ion Ratio Lower Upper
130 100
95 105.3 0.0 197.8





#45

1,2-Dichloropropane

Concen: 20.182 ug/l

RT: 9.623 min Scan# 11

Delta R.T. -0.000 min

Lab File: VN083370.D

Acq: 19 Aug 2024 14:14

Instrument :

MSVOA_N

ClientSampleId :

VN0819WBS01

Tgt Ion: 63 Resp: 3836

Ion Ratio Lower Upper

63 100





65 30.4 24.4 36.6

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 08/20/2024

Supervised By :Mahesh Dadoda 08/20/2024



#46

Dibromomethane

Concen: 19.965 ug/l

RT: 9.706 min Scan# 1318

Delta R.T. -0.006 min

Lab File: VN083370.D

Acq: 19 Aug 2024 14:14




Tgt Ion: 93 Resp: 27162


Ion Ratio Lower Upper

93 100

95 85.1 65.8 98.6

174 88.7 71.7 107.5





#47

Bromodichloromethane

Concen: 19.880 ug/l

RT: 9.888 min Scan# 1349

Delta R.T. -0.000 min

Lab File: VN083370.D

Acq: 19 Aug 2024 14:14

Instrument :

MSVOA_N

ClientSampleId :





VN0819WBS01

Tgt Ion:	83	Resp:	60730
Ion	Ratio	Lower	Upper
83	100		
85	64.7	51.1	76.7
127	8.7	7.2	10.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#48

Methyl methacrylate

Concen: 17.800 ug/l




RT: 9.682 min Scan# 1314


Delta R.T. -0.000 min

Lab File: VN083370.D

Acq: 19 Aug 2024 14:14

Tgt Ion:	41	Resp:	44290
Ion	Ratio	Lower	Upper
41	100		
69	81.5	56.3	84.5
39	60.2	50.3	75.5





#49

1,4-Dioxane

Concen: 389.559 ug/l

RT: 9.700 min Scan# 11

Delta R.T. 0.006 min

Lab File: VN083370.D

Acq: 19 Aug 2024 14:14

Instrument :

MSVOA_N

ClientSampleId :

VN0819WBS01

Tgt Ion: 88 Resp: 1748

Ion Ratio Lower Upper

88 100

43 31.9 27.8 41.8





58 73.1 59.4 89.0

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 08/20/2024

Supervised By :Mahesh Dadoda 08/20/2024



#50

Toluene-d8

Concen: 53.348 ug/l

RT: 10.570 min Scan# 1465

Delta R.T. 0.005 min

Lab File: VN083370.D




Acq: 19 Aug 2024 14:14


Tgt Ion: 98 Resp: 353604

Ion Ratio Lower Upper

98 100


100 64.7 51.5 77.3





#51
4-Methyl-2-Pentanone
Concen: 100.349 ug/l
RT: 10.447 min Scan# 1444
Delta R.T. -0.000 min
Lab File: VN083370.D
Acq: 19 Aug 2024 14:14




Instrument :
MSVOA_N
ClientSampleId :
VN0819WBS01



Tgt Ion: 43 Resp: 28559
Ion Ratio Lower Upper
43 100
58 39.5 28.2 42.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#52
Toluene
Concen: 19.088 ug/l
RT: 10.629 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VN083370.D
Acq: 19 Aug 2024 14:14


Tgt Ion: 92 Resp: 96572
Ion Ratio Lower Upper
92 100
91 172.1 139.4 209.0





#53
t-1,3-Dichloropropene
Concen: 19.077 ug/l
RT: 10.835 min Scan# 1510
Delta R.T. -0.000 min
Lab File: VN083370.D
Acq: 19 Aug 2024 14:14




Instrument :
MSVOA_N
ClientSampleId :
VN0819WBS01



Tgt Ion: 75 Resp: 5986
Ion Ratio Lower Upper
75 100
77 35.1 25.7 38.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#54
cis-1,3-Dichloropropene
Concen: 18.911 ug/l
RT: 10.312 min Scan# 1421
Delta R.T. -0.000 min
Lab File: VN083370.D
Acq: 19 Aug 2024 14:14


Tgt Ion: 75 Resp: 63119
Ion Ratio Lower Upper
75 100
77 29.1 24.3 36.5
39 55.2 50.5 75.7





#55
1,1,2-Trichloroethane
Concen: 20.508 ug/l
RT: 11.018 min Scan# 1541
Delta R.T. -0.000 min
Lab File: VN083370.D
Acq: 19 Aug 2024 14:14




Instrument :
MSVOA_N
ClientSampleId :
VN0819WBS01



Tgt Ion: 97 Resp: 37184
Ion Ratio Lower Upper
97 100
83 84.9 71.4 107.0
85 58.5 45.8 68.6
99 66.1 52.0 78.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#56
Ethyl methacrylate
Concen: 18.335 ug/l
RT: 10.876 min Scan# 1517
Delta R.T. -0.000 min
Lab File: VN083370.D
Acq: 19 Aug 2024 14:14

Tgt Ion: 69 Resp: 62670
Ion Ratio Lower Upper
69 100
41 69.6 63.4 95.2
39 36.8 37.4 56.0#





#57

1,3-Dichloropropane

Concen: 20.046 ug/l

RT: 11.165 min Scan# 1566

Delta R.T. -0.000 min

Lab File: VN083370.D

Acq: 19 Aug 2024 14:14

Instrument :

MSVOA_N

ClientSampleId :

VN0819WBS01

Tgt Ion: 76 Resp: 6478

Ion Ratio Lower Upper

76 100





78 32.7 25.6 38.4

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 08/20/2024

Supervised By :Mahesh Dadoda 08/20/2024



#58

2-Chloroethyl Vinyl ether

Concen: 94.148 ug/l

RT: 10.159 min Scan# 1395

Delta R.T. -0.000 min

Lab File: VN083370.D




Acq: 19 Aug 2024 14:14


Tgt Ion: 63 Resp: 136029

Ion Ratio Lower Upper

63 100


106 28.7 21.4 32.0





#59
2-Hexanone
Concen: 96.806 ug/l
RT: 11.200 min Scan# 11
Delta R.T. 0.006 min
Lab File: VN083370.D
Acq: 19 Aug 2024 14:14




Instrument :
MSVOA_N
ClientSampleId :
VN0819WBS01



Tgt Ion: 43 Resp: 21314
Ion Ratio Lower Upper
43 100
58 53.8 24.6 73.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#60
Dibromochloromethane
Concen: 20.920 ug/l
RT: 11.365 min Scan# 1600
Delta R.T. 0.006 min
Lab File: VN083370.D
Acq: 19 Aug 2024 14:14


Tgt Ion:129 Resp: 45859
Ion Ratio Lower Upper
129 100
127 74.7 39.2 117.6





#61
1,2-Dibromoethane
Concen: 19.795 ug/l
RT: 11.470 min Scan# 1618
Delta R.T. -0.000 min
Lab File: VN083370.D
Acq: 19 Aug 2024 14:14




Instrument :
MSVOA_N
ClientSampleId :
VN0819WBS01




Tgt Ion:107 Resp: 37704
Ion Ratio Lower Upper
107 100
109 95.8 73.7 110.5

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024




#62
4-Bromofluorobenzene
Concen: 52.403 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. -0.000 min
Lab File: VN083370.D
Acq: 19 Aug 2024 14:14




Tgt Ion: 95 Resp: 135412
Ion Ratio Lower Upper
95 100
174 69.2 0.0 159.2
176 67.5 0.0 147.6





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 1685
Delta R.T. -0.000 min
Lab File: VN083370.D
Acq: 19 Aug 2024 14:14




Instrument :
MSVOA_N
ClientSampleId :
VN0819WBS01



Tgt Ion:117 Resp: 251348
Ion Ratio Lower Upper
117 100
82 59.1 47.5 71.3
119 30.7 25.6 38.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#64
Tetrachloroethene
Concen: 19.086 ug/l
RT: 11.106 min Scan# 1556
Delta R.T. -0.000 min
Lab File: VN083370.D
Acq: 19 Aug 2024 14:14


Tgt Ion:164 Resp: 31771
Ion Ratio Lower Upper
164 100
166 129.4 101.4 152.0
129 110.6 75.7 113.5
131 103.5 73.8 110.6





#65
Chlorobenzene
Concen: 19.223 ug/l
RT: 11.894 min Scan# 110
Delta R.T. -0.000 min
Lab File: VN083370.D
Acq: 19 Aug 2024 14:14




Instrument :
MSVOA_N
ClientSampleId :
VN0819WBS01




Tgt Ion:112 Resp: 106774
Ion Ratio Lower Upper
112 100
114 31.8 26.2 39.2

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024




#66
1,1,1,2-Tetrachloroethane
Concen: 19.742 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. -0.006 min
Lab File: VN083370.D
Acq: 19 Aug 2024 14:14




Tgt Ion:131 Resp: 38674
Ion Ratio Lower Upper
131 100
133 93.2 47.3 142.0
119 63.9 32.5 97.4





#67
Ethyl Benzene
Concen: 18.696 ug/l
RT: 11.965 min Scan# 1702
Delta R.T. -0.000 min
Lab File: VN083370.D
Acq: 19 Aug 2024 14:14




Instrument :
MSVOA_N
ClientSampleId :
VN0819WBS01



Tgt Ion: 91 Resp: 190504
Ion Ratio Lower Upper
91 100
106 31.8 24.0 36.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#68
m/p-Xylenes
Concen: 37.111 ug/l
RT: 12.070 min Scan# 1720
Delta R.T. -0.000 min
Lab File: VN083370.D
Acq: 19 Aug 2024 14:14

Tgt Ion:106 Resp: 141650
Ion Ratio Lower Upper
106 100
91 211.0 166.1 249.1









#69
o-Xylene
Concen: 18.215 ug/l
RT: 12.400 min Scan# 1776
Delta R.T. -0.000 min
Lab File: VN083370.D
Acq: 19 Aug 2024 14:14

Instrument :
MSVOA_N
ClientSampleId :
VN0819WBS01

Tgt Ion:106 Resp: 68572
Ion Ratio Lower Upper
106 100
91 220.4 111.3 333.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#70
Styrene
Concen: 18.669 ug/l
RT: 12.412 min Scan# 1778
Delta R.T. -0.000 min
Lab File: VN083370.D
Acq: 19 Aug 2024 14:14


Tgt Ion:104 Resp: 118040
Ion Ratio Lower Upper
104 100
78 54.1 41.6 62.4
103 53.6 44.0 66.0





#71
Bromoform
Concen: 19.009 ug/l
RT: 12.582 min Scan# 1807
Delta R.T. 0.006 min
Lab File: VN083370.D
Acq: 19 Aug 2024 14:14




Instrument :
MSVOA_N
ClientSampleId :
VN0819WBS01



Tgt Ion:173 Resp: 28210
Ion Ratio Lower Upper
173 100
175 48.7 24.0 72.0
254 0.0 0.0 0.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.794 min Scan# 2013
Delta R.T. -0.000 min
Lab File: VN083370.D
Acq: 19 Aug 2024 14:14


Tgt Ion:152 Resp: 116728
Ion Ratio Lower Upper
152 100
115 66.6 30.6 91.6
150 169.1 0.0 348.6





#73
Isopropylbenzene
Concen: 18.823 ug/l
RT: 12.694 min Scan# 1826
Delta R.T. -0.000 min
Lab File: VN083370.D
Acq: 19 Aug 2024 14:14




Instrument :
MSVOA_N
ClientSampleId :
VN0819WBS01



Tgt Ion: 105 Resp: 183779
Ion Ratio Lower Upper
105 100
120 25.3 12.9 38.6




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#74
N-ethyl acetate
Concen: 16.740 ug/l
RT: 12.494 min Scan# 1792
Delta R.T. -0.000 min
Lab File: VN083370.D
Acq: 19 Aug 2024 14:14


Tgt Ion: 43 Resp: 79945
Ion Ratio Lower Upper
43 100
70 42.9 29.8 44.6
55 24.7 18.7 28.1
61 24.3 17.4 26.0





#75
1,1,2,2-Tetrachloroethane
Concen: 19.767 ug/l
RT: 12.941 min Scan# 1868
Delta R.T. 0.006 min
Lab File: VN083370.D
Acq: 19 Aug 2024 14:14




Instrument :
MSVOA_N
ClientSampleId :
VN0819WBS01



Tgt Ion: 83 Resp: 5458
Ion Ratio Lower Upper
83 100
131 9.8 5.0 14.9
85 63.5 31.9 95.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#76
1,2,3-Trichloropropane
Concen: 19.908 ug/l m
RT: 12.994 min Scan# 1877
Delta R.T. -0.000 min
Lab File: VN083370.D
Acq: 19 Aug 2024 14:14


Tgt Ion: 75 Resp: 51001
Ion Ratio Lower Upper
75 100
77 188.8 110.9 332.6





#77
Bromobenzene
Concen: 19.249 ug/l
RT: 12.982 min Scan# 1875
Delta R.T. -0.000 min
Lab File: VN083370.D
Acq: 19 Aug 2024 14:14




Instrument :
MSVOA_N
ClientSampleId :
VN0819WBS01



Tgt Ion:156 Resp: 41748
Ion Ratio Lower Upper
156 100
77 230.6 111.3 333.8
158 97.6 48.3 144.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#78
n-propylbenzene
Concen: 19.151 ug/l
RT: 13.035 min Scan# 1884
Delta R.T. -0.000 min
Lab File: VN083370.D
Acq: 19 Aug 2024 14:14


Tgt Ion: 91 Resp: 215301
Ion Ratio Lower Upper
91 100
120 21.7 10.9 32.9





#79
2-Chlorotoluene
Concen: 18.546 ug/l
RT: 13.123 min Scan# 1899
Delta R.T. -0.000 min
Lab File: VN083370.D
Acq: 19 Aug 2024 14:14




Instrument :
MSVOA_N
ClientSampleId :
VN0819WBS01




Tgt Ion: 91 Resp: 132180
Ion Ratio Lower Upper
91 100
126 31.6 16.2 48.6

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024




#80
1,3,5-Trimethylbenzene
Concen: 19.064 ug/l
RT: 13.176 min Scan# 1908
Delta R.T. 0.006 min
Lab File: VN083370.D
Acq: 19 Aug 2024 14:14




Tgt Ion:105 Resp: 155831
Ion Ratio Lower Upper
105 100
120 46.9 24.3 72.8





#81
trans-1,4-Dichloro-2-butene
Concen: 17.516 ug/l
RT: 12.735 min Scan# 1833
Delta R.T. -0.000 min
Lab File: VN083370.D
Acq: 19 Aug 2024 14:14




Instrument :
MSVOA_N
ClientSampleId :
VN0819WBS01



Tgt Ion: 75 Resp: 20629
Ion Ratio Lower Upper
75 100
53 96.5 88.4 132.6
89 40.6 36.3 54.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#82
4-Chlorotoluene
Concen: 18.607 ug/l
RT: 13.223 min Scan# 1916
Delta R.T. -0.000 min
Lab File: VN083370.D
Acq: 19 Aug 2024 14:14


Tgt Ion: 91 Resp: 133028
Ion Ratio Lower Upper
91 100
126 31.1 16.2 48.6





#83
tert-Butylbenzene
Concen: 18.624 ug/l
RT: 13.441 min Scan# 1953
Delta R.T. -0.000 min
Lab File: VN083370.D
Acq: 19 Aug 2024 14:14




Instrument :
MSVOA_N
ClientSampleId :
VN0819WBS01



Tgt Ion:119 Resp: 134819
Ion Ratio Lower Upper
119 100
91 72.4 35.0 105.1
134 26.4 13.3 39.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#84
1,2,4-Trimethylbenzene
Concen: 18.975 ug/l
RT: 13.482 min Scan# 1960
Delta R.T. -0.000 min
Lab File: VN083370.D
Acq: 19 Aug 2024 14:14


Tgt Ion:105 Resp: 156307
Ion Ratio Lower Upper
105 100
120 43.5 21.9 65.8





#85
 sec-Butylbenzene
 Concen: 18.634 ug/l
 RT: 13.617 min Scan# 1983
 Delta R.T. -0.000 min
 Lab File: VN083370.D
 Acq: 19 Aug 2024 14:14




Instrument :
 MSVOA_N
 ClientSampleId :
 VN0819WBS01




Tgt Ion:105 Resp: 184059
 Ion Ratio Lower Upper
 105 100
 134 19.3 9.8 29.3

Manual Integrations
 APPROVED



Reviewed By :Semsettin Yesilyurt 08/20/2024
 Supervised By :Mahesh Dadoda 08/20/2024




#86
 p-Isopropyltoluene
 Concen: 18.754 ug/l
 RT: 13.729 min Scan# 2002
 Delta R.T. -0.000 min
 Lab File: VN083370.D
 Acq: 19 Aug 2024 14:14



Tgt Ion:119 Resp: 152949
 Ion Ratio Lower Upper
 119 100
 134 25.0 13.0 39.0
 91 26.4 12.3 36.9





#87

1,3-Dichlorobenzene

Concen: 19.005 ug/l

RT: 13.735 min Scan# 2003

Delta R.T. -0.000 min

Lab File: VN083370.D

Acq: 19 Aug 2024 14:14

Instrument :

MSVOA_N

ClientSampleId :





VN0819WBS01

Tgt Ion:146	Resp:	77562
Ion Ratio	Lower	Upper
146	100	
111	42.7	20.6 61.9
148	61.5	31.8 95.3

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 08/20/2024
 Supervised By :Mahesh Dadoda 08/20/2024



#88

1,4-Dichlorobenzene

Concen: 18.612 ug/l




RT: 13.811 min Scan# 2016


Delta R.T. -0.000 min

Lab File: VN083370.D

Acq: 19 Aug 2024 14:14

Tgt Ion:146	Resp:	76574
Ion Ratio	Lower	Upper
146	100	
111	44.6	20.1 60.3
148	64.4	32.2 96.6





#89

n-Butylbenzene

Concen: 18.473 ug/l

RT: 14.059 min Scan# 2058

Delta R.T. -0.000 min

Lab File: VN083370.D

Acq: 19 Aug 2024 14:14

Instrument :

MSVOA_N

ClientSampleId :





VN0819WBS01

Tgt Ion:	91	Resp:	130544
Ion	Ratio	Lower	Upper
91	100		
92	52.4	26.1	78.3
134	24.4	12.1	36.3

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#90

Hexachloroethane

Concen: 18.926 ug/l




RT: 14.335 min Scan# 2105


Delta R.T. -0.000 min

Lab File: VN083370.D

Acq: 19 Aug 2024 14:14

Tgt Ion:	117	Resp:	29820
Ion	Ratio	Lower	Upper
117	100		
201	62.9	35.8	107.3









#91
1,2-Dichlorobenzene
Concen: 19.209 ug/l
RT: 14.106 min Scan# 2066
Delta R.T. -0.000 min
Lab File: VN083370.D
Acq: 19 Aug 2024 14:14

Instrument :
MSVOA_N
ClientSampleId :
VN0819WBS01

Tgt Ion:	146	Resp:	7585
Ion Ratio	Lower	Upper	
146	100		
111	44.8	21.4	64.2
148	61.5	32.1	96.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#92
1,2-Dibromo-3-Chloropropane
Concen: 17.990 ug/l
RT: 14.717 min Scan# 2170
Delta R.T. -0.000 min
Lab File: VN083370.D
Acq: 19 Aug 2024 14:14

Tgt Ion:	75	Resp:	12055
Ion Ratio	Lower	Upper	
75	100		
155	73.5	36.6	109.8
157	99.6	46.9	140.6





#93

1,2,4-Trichlorobenzene

Concen: 18.024 ug/l

RT: 15.394 min Scan# 21

Delta R.T. -0.000 min

Lab File: VN083370.D

Acq: 19 Aug 2024 14:14

Instrument :

MSVOA_N

ClientSampleId :





VN0819WBS01

Tgt Ion	Ratio	Lower	Upper
180	100		
182	97.4	47.3	141.8
145	34.0	15.7	47.1

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 08/20/2024
 Supervised By :Mahesh Dadoda 08/20/2024



#94

Hexachlorobutadiene

Concen: 16.961 ug/l




RT: 15.505 min Scan# 2304


Delta R.T. 0.006 min

Lab File: VN083370.D

Acq: 19 Aug 2024 14:14


Tgt Ion	Ratio	Lower	Upper
225	100		
223	61.1	31.9	95.7
227	65.3	32.5	97.5





#95
Naphthalene
Concen: 17.647 ug/l
RT: 15.641 min Scan# 2327
Delta R.T. -0.000 min
Lab File: VN083370.D
Acq: 19 Aug 2024 14:14




Instrument :
MSVOA_N
ClientSampleId :
VN0819WBS01



Tgt Ion:128 Resp: 138298
Ion Ratio Lower Upper
128 100
127 12.5 10.6 16.0
129 10.2 8.6 12.8




Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#96
1,2,3-Trichlorobenzene
Concen: 18.259 ug/l
RT: 15.841 min Scan# 2361
Delta R.T. -0.000 min
Lab File: VN083370.D
Acq: 19 Aug 2024 14:14

Tgt Ion:180 Resp: 39972
Ion Ratio Lower Upper
180 100
182 93.6 48.9 146.8
145 35.2 16.8 50.4



Report of Analysis

Client:	JACOBS Engineering Group, Inc.	Date Collected:	
Project:	Former Schlumberger Site Princeton NJ	Date Received:	
Client Sample ID:	VN0819WBSD01	SDG No.:	P3657
Lab Sample ID:	VN0819WBSD01	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5	Units:	mL
Soil Aliquot Vol:			uL
GC Column:	RXI-624	ID :	0.25
Prep Method :		Level :	LOW

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN083371.D	1		08/19/24 14:39	VN081924

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	17.7		0.21	1.00	ug/L
74-87-3	Chloromethane	18.6		0.35	1.00	ug/L
75-01-4	Vinyl Chloride	19.2		0.34	1.00	ug/L
74-83-9	Bromomethane	18.4		1.40	5.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	19.7		0.25	1.00	ug/L
67-64-1	Acetone	100		1.40	5.00	ug/L
75-15-0	Carbon Disulfide	15.1		0.32	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	20.0		0.16	1.00	ug/L
75-09-2	Methylene Chloride	19.1		0.32	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	18.0		0.25	1.00	ug/L
110-82-7	Cyclohexane	17.7		1.60	5.00	ug/L
78-93-3	2-Butanone	100		1.30	5.00	ug/L
56-23-5	Carbon Tetrachloride	19.4		0.25	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	19.7		0.25	1.00	ug/L
67-66-3	Chloroform	21.2		0.26	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	20.7		0.19	1.00	ug/L
108-87-2	Methylcyclohexane	16.8		0.19	1.00	ug/L
71-43-2	Benzene	19.4		0.16	1.00	ug/L
107-06-2	1,2-Dichloroethane	21.3		0.24	1.00	ug/L
79-01-6	Trichloroethene	19.3		0.32	1.00	ug/L
75-27-4	Bromodichloromethane	20.3		0.24	1.00	ug/L
108-88-3	Toluene	19.5		0.18	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	21.0		0.21	1.00	ug/L
124-48-1	Dibromochloromethane	21.2		0.18	1.00	ug/L
127-18-4	Tetrachloroethene	18.3		0.25	1.00	ug/L
108-90-7	Chlorobenzene	19.3		0.13	1.00	ug/L
100-41-4	Ethyl Benzene	18.5		0.16	1.00	ug/L
179601-23-1	m/p-Xylenes	37.0		0.31	2.00	ug/L
1330-20-7	Total Xylenes	55.6		0.45	3.00	ug/L
95-47-6	o-Xylene	18.6		0.14	1.00	ug/L

Report of Analysis

Client:	JACOBS Engineering Group, Inc.	Date Collected:	
Project:	Former Schlumberger Site Princeton NJ	Date Received:	
Client Sample ID:	VN0819WBSD01	SDG No.:	P3657
Lab Sample ID:	VN0819WBSD01	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5	Units:	mL
Soil Aliquot Vol:			uL
GC Column:	RXI-624	ID :	0.25
Prep Method :		Level :	LOW

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN083371.D	1		08/19/24 14:39	VN081924

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
98-82-8	Isopropylbenzene	18.7		0.13	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	18.8		0.27	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	18.8		0.19	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	57.5		70 (74) - 130 (125)	115%	SPK: 50
1868-53-7	Dibromofluoromethane	54.9		70 (75) - 130 (124)	110%	SPK: 50
2037-26-5	Toluene-d8	52.6		70 (86) - 130 (113)	105%	SPK: 50
460-00-4	4-Bromofluorobenzene	52.2		70 (77) - 130 (121)	104%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	155000	8.224			
540-36-3	1,4-Difluorobenzene	273000	9.106			
3114-55-4	Chlorobenzene-d5	241000	11.865			
3855-82-1	1,4-Dichlorobenzene-d4	113000	13.794			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of OC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081924\
 Data File : VN083371.D
 Acq On : 19 Aug 2024 14:39
 Operator : JC\MD
 Sample : VN0819WBSD01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0819WBSD01

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 08/20/2024
 Supervised By :Mahesh Dadoda 08/20/2024

Quant Time: Aug 20 04:45:12 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	154957	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	272633	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	241038	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	112767	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	126736	57.460	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 114.920%			
35) Dibromofluoromethane	8.171	113	93410	54.892	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 109.780%			
50) Toluene-d8	10.570	98	333714	52.571	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 105.140%			
62) 4-Bromofluorobenzene	12.847	95	129072	52.156	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 104.320%			
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	2.124	85	31033	17.660	ug/l	95
3) Chloromethane	2.365	50	33546	18.646	ug/l	98
4) Vinyl Chloride	2.512	62	35169	19.158	ug/l	97
5) Bromomethane	2.959	94	20925	18.370	ug/l	92
6) Chloroethane	3.118	64	22619	19.695	ug/l	97
7) Trichlorofluoromethane	3.494	101	59780	19.713	ug/l	97
8) Diethyl Ether	3.965	74	21379	18.945	ug/l	89
9) 1,1,2-Trichlorotrifluo...	4.371	101	32964	19.716	ug/l	99
10) Methyl Iodide	4.589	142	33848	15.390	ug/l #	91
11) Tert butyl alcohol	5.530	59	44685	97.486	ug/l	98
12) 1,1-Dichloroethene	4.336	96	31722	18.467	ug/l	96
13) Acrolein	4.177	56	11069	37.052	ug/l	98
14) Allyl chloride	5.024	41	51369	15.825	ug/l	94
15) Acrylonitrile	5.724	53	99002	105.019	ug/l	99
16) Acetone	4.430	43	86575	100.312	ug/l	98
17) Carbon Disulfide	4.712	76	75758	15.073	ug/l	99
18) Methyl Acetate	5.030	43	60517	23.534	ug/l #	91
19) Methyl tert-butyl Ether	5.800	73	124104	20.016	ug/l	97
20) Methylene Chloride	5.277	84	37862	19.062	ug/l	87
21) trans-1,2-Dichloroethene	5.788	96	31985	18.016	ug/l	90
22) Diisopropyl ether	6.671	45	122681	20.106	ug/l	94
23) Vinyl Acetate	6.606	43	622023m	99.457	ug/l	
24) 1,1-Dichloroethane	6.571	63	67244	20.219	ug/l	98
25) 2-Butanone	7.482	43	137060	103.426	ug/l	94
26) 2,2-Dichloropropane	7.488	77	58753	19.022	ug/l	99
27) cis-1,2-Dichloroethene	7.488	96	42300	19.745	ug/l	92
28) Bromochloromethane	7.812	49	26799	19.717	ug/l	88
29) Tetrahydrofuran	7.841	42	90298	105.408	ug/l	90
30) Chloroform	7.971	83	73400	21.243	ug/l	98
31) Cyclohexane	8.253	56	57737	17.664	ug/l	96
32) 1,1,1-Trichloroethane	8.171	97	67760	20.718	ug/l	91
36) 1,1-Dichloropropene	8.371	75	47170	18.324	ug/l	98
37) Ethyl Acetate	7.565	43	54716	18.984	ug/l #	93
38) Carbon Tetrachloride	8.365	117	56360	19.439	ug/l	99
39) Methylcyclohexane	9.600	83	52965	16.751	ug/l	94
40) Benzene	8.606	78	148742	19.396	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081924\
 Data File : VN083371.D
 Acq On : 19 Aug 2024 14:39
 Operator : JC\MD
 Sample : VN0819WBSD01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0819WBSD01

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 08/20/2024
 Supervised By :Mahesh Dadoda 08/20/2024

Quant Time: Aug 20 04:45:12 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.782	41	30986	18.908	ug/l	95
42) 1,2-Dichloroethane	8.671	62	59445	21.279	ug/l	100
43) Isopropyl Acetate	8.694	43	99574	18.500	ug/l #	94
44) Trichloroethene	9.353	130	35214	19.292	ug/l	97
45) 1,2-Dichloropropane	9.623	63	37836	20.785	ug/l	97
46) Dibromomethane	9.712	93	27310	20.960	ug/l	99
47) Bromodichloromethane	9.888	83	59278	20.262	ug/l	99
48) Methyl methacrylate	9.682	41	45701	19.179	ug/l	93
49) 1,4-Dioxane	9.700	88	17913	416.678	ug/l	95
51) 4-Methyl-2-Pentanone	10.447	43	290215	106.478	ug/l	95
52) Toluene	10.629	92	94661	19.536	ug/l	99
53) t-1,3-Dichloropropene	10.835	75	58673	19.524	ug/l	99
54) cis-1,3-Dichloropropene	10.312	75	60119	18.808	ug/l	94
55) 1,1,2-Trichloroethane	11.018	97	36472	21.004	ug/l	98
56) Ethyl methacrylate	10.876	69	62114	18.975	ug/l	88
57) 1,3-Dichloropropane	11.165	76	65216	21.071	ug/l	99
58) 2-Chloroethyl Vinyl ether	10.159	63	134906	97.496	ug/l	96
59) 2-Hexanone	11.194	43	218635	103.686	ug/l	94
60) Dibromochloromethane	11.359	129	44597	21.243	ug/l	96
61) 1,2-Dibromoethane	11.470	107	36376	19.942	ug/l	95
64) Tetrachloroethene	11.100	164	29259	18.329	ug/l	94
65) Chlorobenzene	11.894	112	103038	19.344	ug/l	98
66) 1,1,1,2-Tetrachloroethane	11.959	131	37036	19.715	ug/l	99
67) Ethyl Benzene	11.965	91	180437	18.466	ug/l	99
68) m/p-Xylenes	12.070	106	135497	37.017	ug/l	97
69) o-Xylene	12.400	106	67097	18.586	ug/l	99
70) Styrene	12.412	104	113234	18.675	ug/l	98
71) Bromoform	12.582	173	27797	19.532	ug/l #	98
73) Isopropylbenzene	12.694	105	176424	18.705	ug/l	100
74) N-amyl acetate	12.494	43	80658	17.483	ug/l	93
75) 1,1,2,2-Tetrachloroethane	12.941	83	54020	20.250	ug/l	100
76) 1,2,3-Trichloropropane	12.994	75	49809m	20.125	ug/l	
77) Bromobenzene	12.982	156	40195	19.184	ug/l	94
78) n-propylbenzene	13.035	91	204577	18.836	ug/l	100
79) 2-Chlorotoluene	13.123	91	128030	18.593	ug/l	98
80) 1,3,5-Trimethylbenzene	13.176	105	147819	18.719	ug/l	98
81) trans-1,4-Dichloro-2-b...	12.735	75	17776	15.623	ug/l	92
82) 4-Chlorotoluene	13.223	91	129077	18.688	ug/l	98
83) tert-Butylbenzene	13.441	119	128002	18.304	ug/l	98
84) 1,2,4-Trimethylbenzene	13.482	105	153066	19.234	ug/l	98
85) sec-Butylbenzene	13.617	105	176890	18.538	ug/l	99
86) p-Isopropyltoluene	13.729	119	144342	18.320	ug/l	97
87) 1,3-Dichlorobenzene	13.735	146	73246	18.578	ug/l	99
88) 1,4-Dichlorobenzene	13.811	146	74550	18.756	ug/l	97
89) n-Butylbenzene	14.059	91	121664	17.821	ug/l	98
90) Hexachloroethane	14.335	117	27902	18.330	ug/l	92
91) 1,2-Dichlorobenzene	14.106	146	71760	18.809	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.717	75	12508	19.322	ug/l	98
93) 1,2,4-Trichlorobenzene	15.394	180	38477	18.002	ug/l	98
94) Hexachlorobutadiene	15.500	225	16274	17.120	ug/l	98
95) Naphthalene	15.641	128	137390	18.147	ug/l	99
96) 1,2,3-Trichlorobenzene	15.841	180	38265	18.093	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081924\
Data File : VN083371.D
Acq On : 19 Aug 2024 14:39
Operator : JC\MD
Sample : VN0819WBSD01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0819WBSD01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024

Quant Time: Aug 20 04:45:12 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081924\
Data File : VN083371.D
Acq On : 19 Aug 2024 14:39
Operator : JC\MD
Sample : VN0819WBSD01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1


Instrument :
MSVOA_N
ClientSampleId :
VN0819WBSD01

Manual Integrations

Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024


Quant Time: Aug 20 04:45:12 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 1066
Delta R.T. -0.000 min
Lab File: VN083371.D
Acq: 19 Aug 2024 14:39




Instrument :
MSVOA_N
ClientSampleId :
VN0819WBSD01



Tgt Ion:168 Resp: 15495
Ion Ratio Lower Upper
168 100
99 70.3 48.2 72.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#2
Dichlorodifluoromethane
Concen: 17.660 ug/l
RT: 2.124 min Scan# 29
Delta R.T. -0.000 min
Lab File: VN083371.D
Acq: 19 Aug 2024 14:39


Tgt Ion: 85 Resp: 31033
Ion Ratio Lower Upper
85 100
87 28.6 15.7 47.0





#3
Chloromethane
Concen: 18.646 ug/l
RT: 2.365 min Scan# 70
Delta R.T. 0.006 min
Lab File: VN083371.D
Acq: 19 Aug 2024 14:39




Instrument :
MSVOA_N
ClientSampleId :
VN0819WBSD01



Tgt Ion: 50 Resp: 33540
Ion Ratio Lower Upper
50 100
52 30.3 25.0 37.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#4
Vinyl Chloride
Concen: 19.158 ug/l
RT: 2.512 min Scan# 95
Delta R.T. -0.000 min
Lab File: VN083371.D
Acq: 19 Aug 2024 14:39


Tgt Ion: 62 Resp: 35169
Ion Ratio Lower Upper
62 100
64 29.7 25.0 37.6





#5
Bromomethane
Concen: 18.370 ug/l
RT: 2.959 min Scan# 11
Delta R.T. 0.012 min
Lab File: VN083371.D
Acq: 19 Aug 2024 14:39



Instrument :
MSVOA_N
ClientSampleId :
VN0819WBSD01



Tgt Ion: 94 Resp: 2092
Ion Ratio Lower Upper
94 100
96 89.2 78.0 117.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#6
Chloroethane
Concen: 19.695 ug/l
RT: 3.118 min Scan# 198
Delta R.T. 0.006 min
Lab File: VN083371.D
Acq: 19 Aug 2024 14:39


Tgt Ion: 64 Resp: 22619
Ion Ratio Lower Upper
64 100
66 35.1 26.6 40.0





#7
Trichlorofluoromethane
Concen: 19.713 ug/l
RT: 3.494 min Scan# 20
Delta R.T. -0.001 min
Lab File: VN083371.D
Acq: 19 Aug 2024 14:39




Instrument :
MSVOA_N
ClientSampleId :
VN0819WBSD01



Tgt Ion:101 Resp: 59780
Ion Ratio Lower Upper
101 100
103 68.0 52.7 79.1




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#8
Diethyl Ether
Concen: 18.945 ug/l
RT: 3.965 min Scan# 342
Delta R.T. 0.006 min
Lab File: VN083371.D
Acq: 19 Aug 2024 14:39

Tgt Ion: 74 Resp: 21379
Ion Ratio Lower Upper
74 100
45 99.7 55.5 166.3









#9
1,1,2-Trichlorotrifluoroethane
Concen: 19.716 ug/l
RT: 4.371 min Scan# 411
Delta R.T. 0.006 min
Lab File: VN083371.D
Acq: 19 Aug 2024 14:39

Instrument :
MSVOA_N
ClientSampleId :
VN0819WBSD01

Tgt Ion:101 Resp: 32964
Ion Ratio Lower Upper
101 100
85 47.8 38.3 57.5
151 72.5 59.3 88.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#10
Methyl Iodide
Concen: 15.390 ug/l
RT: 4.589 min Scan# 448
Delta R.T. -0.000 min
Lab File: VN083371.D
Acq: 19 Aug 2024 14:39

Tgt Ion:142 Resp: 33848
Ion Ratio Lower Upper
142 100
127 53.4 37.5 56.3
141 13.0 13.1 19.7#





#11

Tert butyl alcohol

Concen: 97.486 ug/l

RT: 5.530 min Scan# 608

Delta R.T. -0.000 min

Lab File: VN083371.D

Acq: 19 Aug 2024 14:39

Instrument :

MSVOA_N

ClientSampleId :

VN0819WBSD01

Tgt Ion: 59 Resp: 4468

Ion Ratio Lower Upper

59 100





57 10.2 8.9 13.3

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 08/20/2024

Supervised By :Mahesh Dadoda 08/20/2024



#12

1,1-Dichloroethene

Concen: 18.467 ug/l

RT: 4.336 min Scan# 405

Delta R.T. -0.006 min

Lab File: VN083371.D

Acq: 19 Aug 2024 14:39




Tgt Ion: 96 Resp: 31722


Ion Ratio Lower Upper

96 100

61 179.2 149.7 224.5

98 63.0 50.1 75.1





#13

Acrolein

Concen: 37.052 ug/l

RT: 4.177 min Scan# 31

Delta R.T. -0.006 min

Lab File: VN083371.D

Acq: 19 Aug 2024 14:39

Instrument :

MSVOA_N

ClientSampleId :

VN0819WBSD01

Tgt Ion: 56 Resp: 11069

Ion Ratio Lower Upper

56 100





55 68.6 56.4 84.6

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 08/20/2024

Supervised By :Mahesh Dadoda 08/20/2024



#14

Allyl chloride

Concen: 15.825 ug/l

RT: 5.024 min Scan# 522

Delta R.T. 0.006 min

Lab File: VN083371.D

Acq: 19 Aug 2024 14:39




Tgt Ion: 41 Resp: 51369


Ion Ratio Lower Upper

41 100

39 79.5 67.8 101.6


76 37.0 25.8 38.8





#15
Acrylonitrile
Concen: 105.019 ug/l
RT: 5.724 min Scan# 641
Delta R.T. 0.006 min
Lab File: VN083371.D
Acq: 19 Aug 2024 14:39




Instrument :
MSVOA_N
ClientSampleId :
VN0819WBSD01



Tgt Ion: 53 Resp: 9900
Ion Ratio Lower Upper
53 100
52 82.7 66.6 100.0
51 37.1 30.1 45.1




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#16
Acetone
Concen: 100.312 ug/l
RT: 4.430 min Scan# 421
Delta R.T. -0.000 min
Lab File: VN083371.D
Acq: 19 Aug 2024 14:39


Tgt Ion: 43 Resp: 86575
Ion Ratio Lower Upper
43 100
58 28.1 21.8 32.6





#17
Carbon Disulfide
Concen: 15.073 ug/l
RT: 4.712 min Scan# 469
Delta R.T. -0.000 min
Lab File: VN083371.D
Acq: 19 Aug 2024 14:39




Instrument :
MSVOA_N
ClientSampleId :
VN0819WBSD01



Tgt Ion: 76 Resp: 75758
Ion Ratio Lower Upper
76 100
78 8.6 7.2 10.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#18
Methyl Acetate
Concen: 23.534 ug/l
RT: 5.030 min Scan# 523
Delta R.T. 0.006 min
Lab File: VN083371.D
Acq: 19 Aug 2024 14:39

Tgt Ion: 43 Resp: 60517
Ion Ratio Lower Upper
43 100
74 24.6 16.3 24.5#









#19
Methyl tert-butyl Ether
Concen: 20.016 ug/l
RT: 5.800 min Scan# 614
Delta R.T. 0.006 min
Lab File: VN083371.D
Acq: 19 Aug 2024 14:39

Instrument :
MSVOA_N
ClientSampleId :
VN0819WBSD01

Tgt Ion: 73 Resp: 124104
Ion Ratio Lower Upper
73 100
57 21.1 17.9 26.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#20
Methylene Chloride
Concen: 19.062 ug/l
RT: 5.277 min Scan# 565
Delta R.T. -0.000 min
Lab File: VN083371.D
Acq: 19 Aug 2024 14:39


Tgt Ion: 84 Resp: 37862
Ion Ratio Lower Upper
84 100
49 126.3 119.6 179.4
51 36.8 34.8 52.2
86 64.4 52.9 79.3





#21
trans-1,2-Dichloroethene
Concen: 18.016 ug/l
RT: 5.788 min Scan# 61
Delta R.T. -0.006 min
Lab File: VN083371.D
Acq: 19 Aug 2024 14:39




Instrument :
MSVOA_N
ClientSampleId :
VN0819WBSD01



Tgt Ion: 96 Resp: 3198
Ion Ratio Lower Upper
96 100
61 141.3 123.4 185.2
98 69.3 49.3 73.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#22
Diisopropyl ether
Concen: 20.106 ug/l
RT: 6.671 min Scan# 802
Delta R.T. -0.000 min
Lab File: VN083371.D
Acq: 19 Aug 2024 14:39


Tgt Ion: 45 Resp: 122681
Ion Ratio Lower Upper
45 100
43 59.2 44.0 66.0
87 28.2 19.7 29.5
59 10.2 8.5 12.7





#23
 Vinyl Acetate
 Concen: 99.457 ug/l m
 RT: 6.606 min Scan# 791
 Delta R.T. -0.000 min
 Lab File: VN083371.D
 Acq: 19 Aug 2024 14:39




Instrument :
 MSVOA_N
 ClientSampleId :
 VN0819WBSD01



Tgt Ion: 43 Resp: 62202
 Ion Ratio Lower Upper
 43 100
 86 11.1 7.3 10.9




Manual Integrations
 APPROVED


Reviewed By :Semsettin Yesilyurt 08/20/2024
 Supervised By :Mahesh Dadoda 08/20/2024



#24
 1,1-Dichloroethane
 Concen: 20.219 ug/l
 RT: 6.571 min Scan# 785
 Delta R.T. -0.000 min
 Lab File: VN083371.D
 Acq: 19 Aug 2024 14:39


Tgt Ion: 63 Resp: 67244
 Ion Ratio Lower Upper
 63 100
 98 5.9 3.3 9.9
 100 4.4 2.0 6.0





#25
2-Butanone
Concen: 103.426 ug/l
RT: 7.482 min Scan# 940
Delta R.T. -0.000 min
Lab File: VN083371.D
Acq: 19 Aug 2024 14:39




Instrument :
MSVOA_N
ClientSampleId :
VN0819WBSD01



Tgt Ion: 43 Resp: 137060
Ion Ratio Lower Upper
43 100
72 25.5 18.2 27.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#26
2,2-Dichloropropane
Concen: 19.022 ug/l
RT: 7.488 min Scan# 941
Delta R.T. -0.000 min
Lab File: VN083371.D
Acq: 19 Aug 2024 14:39


Tgt Ion: 77 Resp: 58753
Ion Ratio Lower Upper
77 100
97 21.1 10.3 30.9





#27
cis-1,2-Dichloroethene
Concen: 19.745 ug/l
RT: 7.488 min Scan# 941
Delta R.T. -0.000 min
Lab File: VN083371.D
Acq: 19 Aug 2024 14:39




Instrument :
MSVOA_N
ClientSampleId :
VN0819WBSD01



Tgt Ion: 96 Resp: 42300
Ion Ratio Lower Upper
96 100
61 149.3 0.0 329.6
98 64.4 0.0 128.6




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#28
Bromochloromethane
Concen: 19.717 ug/l
RT: 7.812 min Scan# 996
Delta R.T. -0.000 min
Lab File: VN083371.D
Acq: 19 Aug 2024 14:39

Tgt Ion: 49 Resp: 26799
Ion Ratio Lower Upper
49 100
129 1.4 0.0 3.8
130 72.4 50.5 75.7





#29

Tetrahydrofuran

Concen: 105.408 ug/l

RT: 7.841 min Scan# 1001

Delta R.T. 0.006 min

Lab File: VN083371.D

Acq: 19 Aug 2024 14:39

Instrument :

MSVOA_N

ClientSampleId :





VN0819WBSD01

Tgt Ion:	42	Resp:	90298
Ion	Ratio	Lower	Upper
42	100		
72	44.7	30.0	45.0
71	40.4	28.3	42.5

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#30

Chloroform

Concen: 21.243 ug/l




RT: 7.971 min Scan# 1023


Delta R.T. 0.006 min

Lab File: VN083371.D

Acq: 19 Aug 2024 14:39


Tgt Ion:	83	Resp:	73400
Ion	Ratio	Lower	Upper
83	100		
85	65.5	50.9	76.3





#31
Cyclohexane
Concen: 17.664 ug/l
RT: 8.253 min Scan# 1072
Delta R.T. -0.006 min
Lab File: VN083371.D
Acq: 19 Aug 2024 14:39




Instrument :
MSVOA_N
ClientSampleId :
VN0819WBSD01



Tgt Ion: 56 Resp: 5773
Ion Ratio Lower Upper
56 100
69 33.0 24.5 36.7
84 86.4 66.0 99.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#32
1,1,1-Trichloroethane
Concen: 20.718 ug/l
RT: 8.171 min Scan# 1057
Delta R.T. -0.000 min
Lab File: VN083371.D
Acq: 19 Aug 2024 14:39

Tgt Ion: 97 Resp: 67760
Ion Ratio Lower Upper
97 100
99 58.2 52.0 78.0
61 46.0 42.1 63.1





#33

1,2-Dichloroethane-d4

Concen: 57.460 ug/l

RT: 8.582 min Scan# 1127

Delta R.T. 0.000 min

Lab File: VN083371.D

Acq: 19 Aug 2024 14:39

Instrument :

MSVOA_N

ClientSampleId :

VN0819WBSD01

Tgt Ion: 65 Resp: 126730

Ion Ratio Lower Upper

65 100





67 50.5 0.0 104.4

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 08/20/2024

Supervised By :Mahesh Dadoda 08/20/2024



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.106 min Scan# 1216

Delta R.T. 0.006 min

Lab File: VN083371.D

Acq: 19 Aug 2024 14:39




Tgt Ion:114 Resp: 272633


Ion Ratio Lower Upper

114 100

63 22.7 0.0 44.6

88 17.6 0.0 31.4





#35

Dibromofluoromethane

Concen: 54.892 ug/l

RT: 8.171 min Scan# 1056

Delta R.T. 0.006 min

Lab File: VN083371.D

Acq: 19 Aug 2024 14:39

Instrument :

MSVOA_N

ClientSampleId :

VN0819WBSD01

Tgt Ion:113 Resp: 93410

Ion Ratio Lower Upper

113 100

111 102.1 82.4 123.6





192 17.0 14.9 22.3

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 08/20/2024

Supervised By :Mahesh Dadoda 08/20/2024



#36

1,1-Dichloropropene

Concen: 18.324 ug/l

RT: 8.371 min Scan# 1091

Delta R.T. -0.000 min

Lab File: VN083371.D

Acq: 19 Aug 2024 14:39




Tgt Ion: 75 Resp: 47170


Ion Ratio Lower Upper

75 100

110 34.6 16.3 48.9

77 31.0 24.6 37.0





#37

Ethyl Acetate

Concen: 18.984 ug/l

RT: 7.565 min Scan# 91

Delta R.T. 0.006 min

Lab File: VN083371.D

Acq: 19 Aug 2024 14:39

Instrument :

MSVOA_N

ClientSampleId :





VN0819WBSD01

Tgt Ion:	43	Resp:	54710
Ion	Ratio	Lower	Upper
43	100		
61	14.5	9.3	13.9
70	10.9	7.0	10.6

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#38

Carbon Tetrachloride

Concen: 19.439 ug/l




RT: 8.365 min Scan# 1090


Delta R.T. 0.006 min

Lab File: VN083371.D

Acq: 19 Aug 2024 14:39

Tgt Ion:	117	Resp:	56360
Ion	Ratio	Lower	Upper
117	100		
119	94.0	74.9	112.3
121	32.3	24.3	36.5





#39

Methylcyclohexane

Concen: 16.751 ug/l

RT: 9.600 min Scan# 1130

Delta R.T. -0.000 min

Lab File: VN083371.D

Acq: 19 Aug 2024 14:39

Instrument :

MSVOA_N

ClientSampleId :





VN0819WBSD01

Tgt Ion:	83	Resp:	5296
Ion Ratio	Lower	Upper	
83	100		
55	77.0	66.4	99.6
98	48.3	40.7	61.1

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#40

Benzene

Concen: 19.396 ug/l




RT: 8.606 min Scan# 1131


Delta R.T. -0.000 min

Lab File: VN083371.D

Acq: 19 Aug 2024 14:39


Tgt Ion:	78	Resp:	148742
Ion Ratio	Lower	Upper	
78	100		
77	23.5	19.0	28.4





#41
Methacrylonitrile
Concen: 18.908 ug/l
RT: 7.782 min Scan# 991
Delta R.T. 0.006 min
Lab File: VN083371.D
Acq: 19 Aug 2024 14:39




Instrument :
MSVOA_N
ClientSampleId :
VN0819WBSD01



Tgt Ion: 41 Resp: 30980
Ion Ratio Lower Upper
41 100
39 57.2 45.5 68.3
67 70.8 50.5 75.7
52 30.7 24.0 36.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#42
1,2-Dichloroethane
Concen: 21.279 ug/l
RT: 8.671 min Scan# 1142
Delta R.T. -0.000 min
Lab File: VN083371.D
Acq: 19 Aug 2024 14:39

Tgt Ion: 62 Resp: 59445
Ion Ratio Lower Upper
62 100
98 7.9 0.0 15.8





#43

Isopropyl Acetate

Concen: 18.500 ug/l

RT: 8.694 min Scan# 1146

Delta R.T. 0.006 min

Lab File: VN083371.D

Acq: 19 Aug 2024 14:39

Instrument :

MSVOA_N

ClientSampleId :





VN0819WBSD01

Tgt Ion:	43	Resp:	99574
Ion	Ratio	Lower	Upper
43	100		
61	24.7	17.8	26.6
87	12.7	8.2	12.2

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#44

Trichloroethene

Concen: 19.292 ug/l




RT: 9.353 min Scan# 1258


Delta R.T. -0.000 min

Lab File: VN083371.D

Acq: 19 Aug 2024 14:39

Tgt Ion:	130	Resp:	35214
Ion	Ratio	Lower	Upper
130	100		
95	101.7	0.0	197.8





#45

1,2-Dichloropropane

Concen: 20.785 ug/l

RT: 9.623 min Scan# 1304

Delta R.T. -0.000 min

Lab File: VN083371.D

Acq: 19 Aug 2024 14:39

Instrument :

MSVOA_N

ClientSampleId :

VN0819WBSD01

Tgt Ion: 63 Resp: 3783

Ion Ratio Lower Upper

63 100





65 31.9 24.4 36.6

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 08/20/2024

Supervised By :Mahesh Dadoda 08/20/2024



#46

Dibromomethane

Concen: 20.960 ug/l

RT: 9.712 min Scan# 1319

Delta R.T. -0.000 min

Lab File: VN083371.D

Acq: 19 Aug 2024 14:39




Tgt Ion: 93 Resp: 27310


Ion Ratio Lower Upper

93 100

95 82.6 65.8 98.6

174 89.0 71.7 107.5





#47

Bromodichloromethane

Concen: 20.262 ug/l

RT: 9.888 min Scan# 1349

Delta R.T. -0.000 min

Lab File: VN083371.D

Acq: 19 Aug 2024 14:39

Instrument :

MSVOA_N

ClientSampleId :





VN0819WBSD01

Tgt Ion: 83	Resp: 59278
Ion Ratio	Lower Upper
83 100	
85 65.1	51.1 76.7
127 9.0	7.2 10.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#48

Methyl methacrylate

Concen: 19.179 ug/l




RT: 9.682 min Scan# 1314


Delta R.T. -0.000 min

Lab File: VN083371.D

Acq: 19 Aug 2024 14:39

Tgt Ion: 41	Resp: 45701
Ion Ratio	Lower Upper
41 100	
69 79.1	56.3 84.5
39 60.6	50.3 75.5





#49

1,4-Dioxane

Concen: 416.678 ug/l

RT: 9.700 min Scan# 11

Delta R.T. 0.006 min

Lab File: VN083371.D

Acq: 19 Aug 2024 14:39

Instrument :

MSVOA_N

ClientSampleId :

VN0819WBSD01

Tgt Ion: 88 Resp: 1791

Ion Ratio Lower Upper

88 100

43 32.5 27.8 41.8





58 70.1 59.4 89.0

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 08/20/2024

Supervised By :Mahesh Dadoda 08/20/2024



#50

Toluene-d8

Concen: 52.571 ug/l

RT: 10.570 min Scan# 1465

Delta R.T. 0.005 min

Lab File: VN083371.D




Acq: 19 Aug 2024 14:39


Tgt Ion: 98 Resp: 333714

Ion Ratio Lower Upper

98 100


100 66.1 51.5 77.3





#51
4-Methyl-2-Pentanone
Concen: 106.478 ug/l
RT: 10.447 min Scan# 1444
Delta R.T. -0.000 min
Lab File: VN083371.D
Acq: 19 Aug 2024 14:39




Instrument :
MSVOA_N
ClientSampleId :
VN0819WBSD01



Tgt Ion: 43 Resp: 290215
Ion Ratio Lower Upper
43 100
58 38.5 28.2 42.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#52
Toluene
Concen: 19.536 ug/l
RT: 10.629 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VN083371.D
Acq: 19 Aug 2024 14:39


Tgt Ion: 92 Resp: 94661
Ion Ratio Lower Upper
92 100
91 173.5 139.4 209.0





#53
t-1,3-Dichloropropene
Concen: 19.524 ug/l
RT: 10.835 min Scan# 1510
Delta R.T. -0.000 min
Lab File: VN083371.D
Acq: 19 Aug 2024 14:39




Instrument :
MSVOA_N
ClientSampleId :
VN0819WBSD01



Tgt Ion: 75 Resp: 5867
Ion Ratio Lower Upper
75 100
77 31.5 25.7 38.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#54
cis-1,3-Dichloropropene
Concen: 18.808 ug/l
RT: 10.312 min Scan# 1421
Delta R.T. -0.000 min
Lab File: VN083371.D
Acq: 19 Aug 2024 14:39


Tgt Ion: 75 Resp: 60119
Ion Ratio Lower Upper
75 100
77 33.2 24.3 36.5
39 58.3 50.5 75.7





#55
1,1,2-Trichloroethane
Concen: 21.004 ug/l
RT: 11.018 min Scan# 1541
Delta R.T. -0.000 min
Lab File: VN083371.D
Acq: 19 Aug 2024 14:39




Instrument :
MSVOA_N
ClientSampleId :
VN0819WBSD01



Tgt Ion: 97 Resp: 36472
Ion Ratio Lower Upper
97 100
83 89.0 71.4 107.0
85 53.5 45.8 68.6
99 65.6 52.0 78.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#56
Ethyl methacrylate
Concen: 18.975 ug/l
RT: 10.876 min Scan# 1517
Delta R.T. -0.000 min
Lab File: VN083371.D
Acq: 19 Aug 2024 14:39

Tgt Ion: 69 Resp: 62114
Ion Ratio Lower Upper
69 100
41 69.7 63.4 95.2
39 38.0 37.4 56.0





#57

1,3-Dichloropropane

Concen: 21.071 ug/l

RT: 11.165 min Scan# 1566

Delta R.T. -0.000 min

Lab File: VN083371.D

Acq: 19 Aug 2024 14:39

Instrument :

MSVOA_N

ClientSampleId :

VN0819WBSD01

Tgt Ion: 76 Resp: 65210

Ion Ratio Lower Upper

76 100





78 32.4 25.6 38.4

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 08/20/2024

Supervised By :Mahesh Dadoda 08/20/2024



#58

2-Chloroethyl Vinyl ether

Concen: 97.496 ug/l

RT: 10.159 min Scan# 1395

Delta R.T. -0.000 min

Lab File: VN083371.D




Acq: 19 Aug 2024 14:39


Tgt Ion: 63 Resp: 134906

Ion Ratio Lower Upper

63 100


106 28.7 21.4 32.0





#59
2-Hexanone
Concen: 103.686 ug/l
RT: 11.194 min Scan# 1571
Delta R.T. -0.000 min
Lab File: VN083371.D
Acq: 19 Aug 2024 14:39




Instrument :
MSVOA_N
ClientSampleId :
VN0819WBSD01



Tgt Ion: 43 Resp: 21863
Ion Ratio Lower Upper
43 100
58 53.5 24.6 73.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#60
Dibromochloromethane
Concen: 21.243 ug/l
RT: 11.359 min Scan# 1599
Delta R.T. -0.000 min
Lab File: VN083371.D
Acq: 19 Aug 2024 14:39


Tgt Ion:129 Resp: 44597
Ion Ratio Lower Upper
129 100
127 75.3 39.2 117.6





#61
1,2-Dibromoethane
Concen: 19.942 ug/l
RT: 11.470 min Scan# 1618
Delta R.T. -0.000 min
Lab File: VN083371.D
Acq: 19 Aug 2024 14:39




Instrument :
MSVOA_N
ClientSampleId :
VN0819WBSD01



Tgt Ion:107 Resp: 36370
Ion Ratio Lower Upper
107 100
109 97.1 73.7 110.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#62
4-Bromofluorobenzene
Concen: 52.156 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. -0.000 min
Lab File: VN083371.D
Acq: 19 Aug 2024 14:39


Tgt Ion: 95 Resp: 129072
Ion Ratio Lower Upper
95 100
174 69.7 0.0 159.2
176 66.6 0.0 147.6





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 1685
Delta R.T. -0.000 min
Lab File: VN083371.D
Acq: 19 Aug 2024 14:39




Instrument :
MSVOA_N
ClientSampleId :
VN0819WBSD01



Tgt Ion:117 Resp: 241038
Ion Ratio Lower Upper
117 100
82 58.9 47.5 71.3
119 31.8 25.6 38.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#64
Tetrachloroethene
Concen: 18.329 ug/l
RT: 11.100 min Scan# 1555
Delta R.T. -0.006 min
Lab File: VN083371.D
Acq: 19 Aug 2024 14:39


Tgt Ion:164 Resp: 29259
Ion Ratio Lower Upper
164 100
166 130.8 101.4 152.0
129 101.6 75.7 113.5
131 100.1 73.8 110.6





#65
Chlorobenzene
Concen: 19.344 ug/l
RT: 11.894 min Scan# 1100
Delta R.T. -0.000 min
Lab File: VN083371.D
Acq: 19 Aug 2024 14:39




Instrument :
MSVOA_N
ClientSampleId :
VN0819WBSD01



Tgt Ion:112 Resp: 103038
Ion Ratio Lower Upper
112 100
114 31.8 26.2 39.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#66
1,1,1,2-Tetrachloroethane
Concen: 19.715 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. -0.006 min
Lab File: VN083371.D
Acq: 19 Aug 2024 14:39


Tgt Ion:131 Resp: 37036
Ion Ratio Lower Upper
131 100
133 95.8 47.3 142.0
119 65.2 32.5 97.4





#67
Ethyl Benzene
Concen: 18.466 ug/l
RT: 11.965 min Scan# 1702
Delta R.T. -0.000 min
Lab File: VN083371.D
Acq: 19 Aug 2024 14:39




Instrument :
MSVOA_N
ClientSampleId :
VN0819WBSD01



Tgt Ion: 91 Resp: 180437
Ion Ratio Lower Upper
91 100
106 30.4 24.0 36.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#68
m/p-Xylenes
Concen: 37.017 ug/l
RT: 12.070 min Scan# 1720
Delta R.T. -0.000 min
Lab File: VN083371.D
Acq: 19 Aug 2024 14:39

Tgt Ion:106 Resp: 135497
Ion Ratio Lower Upper
106 100
91 212.9 166.1 249.1





#69

o-Xylene

Concen: 18.586 ug/l

RT: 12.400 min Scan# 1776

Delta R.T. -0.000 min

Lab File: VN083371.D

Acq: 19 Aug 2024 14:39

Instrument :

MSVOA_N

ClientSampleId :

VN0819WBSD01

Tgt Ion:106 Resp: 6709

Ion Ratio Lower Upper

106 100





91 223.9 111.3 333.9

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 08/20/2024

Supervised By :Mahesh Dadoda 08/20/2024



#70

Styrene

Concen: 18.675 ug/l

RT: 12.412 min Scan# 1778

Delta R.T. -0.000 min

Lab File: VN083371.D

Acq: 19 Aug 2024 14:39




Tgt Ion:104 Resp: 113234


Ion Ratio Lower Upper

104 100

78 54.2 41.6 62.4


103 54.5 44.0 66.0





#71
Bromoform
Concen: 19.532 ug/l
RT: 12.582 min Scan# 1807
Delta R.T. 0.006 min
Lab File: VN083371.D
Acq: 19 Aug 2024 14:39




Instrument :
MSVOA_N
ClientSampleId :
VN0819WBSD01




Tgt Ion:173 Resp: 2779
Ion Ratio Lower Upper
173 100
175 49.6 24.0 72.0
254 0.0 0.0 0.0

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.794 min Scan# 2013
Delta R.T. -0.000 min
Lab File: VN083371.D
Acq: 19 Aug 2024 14:39




Tgt Ion:152 Resp: 112767
Ion Ratio Lower Upper
152 100
115 63.2 30.6 91.6
150 164.8 0.0 348.6





#73
Isopropylbenzene
Concen: 18.705 ug/l
RT: 12.694 min Scan# 1826
Delta R.T. -0.000 min
Lab File: VN083371.D
Acq: 19 Aug 2024 14:39




Instrument :
MSVOA_N
ClientSampleId :
VN0819WBSD01



Tgt Ion:105 Resp: 176424
Ion Ratio Lower Upper
105 100
120 25.7 12.9 38.6




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#74
N-ethyl acetate
Concen: 17.483 ug/l
RT: 12.494 min Scan# 1792
Delta R.T. -0.000 min
Lab File: VN083371.D
Acq: 19 Aug 2024 14:39

Tgt Ion: 43 Resp: 80658
Ion Ratio Lower Upper
43 100
70 43.3 29.8 44.6
55 25.0 18.7 28.1
61 24.9 17.4 26.0









#75
1,1,2,2-Tetrachloroethane
Concen: 20.250 ug/l
RT: 12.941 min Scan# 1868
Delta R.T. 0.006 min
Lab File: VN083371.D
Acq: 19 Aug 2024 14:39

Instrument :
MSVOA_N
ClientSampleId :
VN0819WBSD01

Tgt Ion:	83	Resp:	54020
Ion	Ratio	Lower	Upper
83	100		
131	10.0	5.0	14.9
85	64.0	31.9	95.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#76
1,2,3-Trichloropropane
Concen: 20.125 ug/l m
RT: 12.994 min Scan# 1877
Delta R.T. -0.000 min
Lab File: VN083371.D
Acq: 19 Aug 2024 14:39


Tgt Ion:	75	Resp:	49809
Ion	Ratio	Lower	Upper
75	100		
77	189.1	110.9	332.6





#77
Bromobenzene
Concen: 19.184 ug/l
RT: 12.982 min Scan# 1875
Delta R.T. -0.000 min
Lab File: VN083371.D
Acq: 19 Aug 2024 14:39




Instrument :
MSVOA_N
ClientSampleId :
VN0819WBSD01



Tgt Ion:156 Resp: 4019
Ion Ratio Lower Upper
156 100
77 234.3 111.3 333.8
158 99.4 48.3 144.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#78
n-propylbenzene
Concen: 18.836 ug/l
RT: 13.035 min Scan# 1884
Delta R.T. -0.000 min
Lab File: VN083371.D
Acq: 19 Aug 2024 14:39


Tgt Ion: 91 Resp: 204577
Ion Ratio Lower Upper
91 100
120 21.7 10.9 32.9





#79
2-Chlorotoluene
Concen: 18.593 ug/l
RT: 13.123 min Scan# 1899
Delta R.T. -0.000 min
Lab File: VN083371.D
Acq: 19 Aug 2024 14:39




Instrument :
MSVOA_N
ClientSampleId :
VN0819WBSD01




Tgt Ion: 91 Resp: 128030
Ion Ratio Lower Upper
91 100
126 31.5 16.2 48.6

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024




#80
1,3,5-Trimethylbenzene
Concen: 18.719 ug/l
RT: 13.176 min Scan# 1908
Delta R.T. 0.006 min
Lab File: VN083371.D
Acq: 19 Aug 2024 14:39




Tgt Ion:105 Resp: 147819
Ion Ratio Lower Upper
105 100
120 47.1 24.3 72.8





#81
trans-1,4-Dichloro-2-butene
Concen: 15.623 ug/l
RT: 12.735 min Scan# 1833
Delta R.T. -0.000 min
Lab File: VN083371.D
Acq: 19 Aug 2024 14:39




Instrument :
MSVOA_N
ClientSampleId :
VN0819WBSD01



Tgt Ion: 75 Resp: 17770
Ion Ratio Lower Upper
75 100
53 121.8 88.4 132.6
89 46.5 36.3 54.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#82
4-Chlorotoluene
Concen: 18.688 ug/l
RT: 13.223 min Scan# 1916
Delta R.T. -0.000 min
Lab File: VN083371.D
Acq: 19 Aug 2024 14:39


Tgt Ion: 91 Resp: 129077
Ion Ratio Lower Upper
91 100
126 31.0 16.2 48.6





#83
tert-Butylbenzene
Concen: 18.304 ug/l
RT: 13.441 min Scan# 1953
Delta R.T. -0.000 min
Lab File: VN083371.D
Acq: 19 Aug 2024 14:39




Instrument :
MSVOA_N
ClientSampleId :
VN0819WBSD01



Tgt Ion:119 Resp: 128002
Ion Ratio Lower Upper
119 100
91 69.4 35.0 105.1
134 24.7 13.3 39.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#84
1,2,4-Trimethylbenzene
Concen: 19.234 ug/l
RT: 13.482 min Scan# 1960
Delta R.T. -0.000 min
Lab File: VN083371.D
Acq: 19 Aug 2024 14:39


Tgt Ion:105 Resp: 153066
Ion Ratio Lower Upper
105 100
120 42.4 21.9 65.8





#85
 sec-Butylbenzene
 Concen: 18.538 ug/l
 RT: 13.617 min Scan# 1983
 Delta R.T. -0.000 min
 Lab File: VN083371.D
 Acq: 19 Aug 2024 14:39




Instrument :
 MSVOA_N
 ClientSampleId :
 VN0819WBSD01




Tgt Ion:105 Resp: 176890
 Ion Ratio Lower Upper
 105 100
 134 18.8 9.8 29.3

Manual Integrations
 APPROVED



Reviewed By :Semsettin Yesilyurt 08/20/2024
 Supervised By :Mahesh Dadoda 08/20/2024




#86
 p-Isopropyltoluene
 Concen: 18.320 ug/l
 RT: 13.729 min Scan# 2002
 Delta R.T. -0.000 min
 Lab File: VN083371.D
 Acq: 19 Aug 2024 14:39



Tgt Ion:119 Resp: 144342
 Ion Ratio Lower Upper
 119 100
 134 25.2 13.0 39.0
 91 26.4 12.3 36.9





#87

1,3-Dichlorobenzene

Concen: 18.578 ug/l

RT: 13.735 min Scan# 2003

Delta R.T. -0.000 min

Lab File: VN083371.D


Acq: 19 Aug 2024 14:39

Instrument :

MSVOA_N

ClientSampleId :

VN0819WBSD01



Tgt Ion:146 Resp: 73240

Ion Ratio Lower Upper

146 100

111 42.5 20.6 61.9




148 63.0 31.8 95.3

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 08/20/2024

Supervised By :Mahesh Dadoda 08/20/2024



#88

1,4-Dichlorobenzene


Concen: 18.756 ug/l

RT: 13.811 min Scan# 2016

Delta R.T. -0.000 min

Lab File: VN083371.D

Acq: 19 Aug 2024 14:39





Tgt Ion:146 Resp: 74550


Ion Ratio Lower Upper

146 100

111 42.3 20.1 60.3

148 62.5 32.2 96.6





#89

n-Butylbenzene

Concen: 17.821 ug/l

RT: 14.059 min Scan# 2058

Delta R.T. -0.000 min

Lab File: VN083371.D

Acq: 19 Aug 2024 14:39

Instrument :

MSVOA_N

ClientSampleId :

VN0819WBSD01





Tgt Ion:	91	Resp:	121664
Ion	Ratio	Lower	Upper
91	100		
92	53.3	26.1	78.3
134	25.0	12.1	36.3

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 08/20/2024

Supervised By :Mahesh Dadoda 08/20/2024



#90

Hexachloroethane

Concen: 18.330 ug/l




RT: 14.335 min Scan# 2105


Delta R.T. -0.000 min

Lab File: VN083371.D

Acq: 19 Aug 2024 14:39


Tgt Ion:	117	Resp:	27902
Ion	Ratio	Lower	Upper
117	100		
201	65.2	35.8	107.3





#91
1,2-Dichlorobenzene
Concen: 18.809 ug/l
RT: 14.106 min Scan# 2066
Delta R.T. -0.000 min
Lab File: VN083371.D
Acq: 19 Aug 2024 14:39




Instrument :
MSVOA_N
ClientSampleId :
VN0819WBSD01



Tgt Ion:146 Resp: 71760
Ion Ratio Lower Upper
146 100
111 44.8 21.4 64.2
148 64.4 32.1 96.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#92
1,2-Dibromo-3-Chloropropane
Concen: 19.322 ug/l
RT: 14.717 min Scan# 2170
Delta R.T. -0.000 min
Lab File: VN083371.D
Acq: 19 Aug 2024 14:39


Tgt Ion: 75 Resp: 12508
Ion Ratio Lower Upper
75 100
155 70.8 36.6 109.8
157 94.5 46.9 140.6





#93
1,2,4-Trichlorobenzene
Concen: 18.002 ug/l
RT: 15.394 min Scan# 2285
Delta R.T. -0.000 min
Lab File: VN083371.D
Acq: 19 Aug 2024 14:39




Instrument :
MSVOA_N
ClientSampleId :
VN0819WBSD01




Tgt Ion:180 Resp: 3847
Ion Ratio Lower Upper
180 100
182 93.8 47.3 141.8
145 33.5 15.7 47.1

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024




#94
Hexachlorobutadiene
Concen: 17.120 ug/l
RT: 15.500 min Scan# 2303
Delta R.T. -0.000 min
Lab File: VN083371.D
Acq: 19 Aug 2024 14:39




Tgt Ion:225 Resp: 16274
Ion Ratio Lower Upper
225 100
223 62.7 31.9 95.7
227 62.5 32.5 97.5





#95
Naphthalene
Concen: 18.147 ug/l
RT: 15.641 min Scan# 2327
Delta R.T. -0.000 min
Lab File: VN083371.D
Acq: 19 Aug 2024 14:39




Instrument :
MSVOA_N
ClientSampleId :
VN0819WBSD01




Tgt Ion:128 Resp: 137390
Ion Ratio Lower Upper
128 100
127 12.7 10.6 16.0
129 11.0 8.6 12.8

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 08/20/2024
Supervised By :Mahesh Dadoda 08/20/2024



#96
1,2,3-Trichlorobenzene
Concen: 18.093 ug/l
RT: 15.841 min Scan# 2361
Delta R.T. -0.000 min
Lab File: VN083371.D
Acq: 19 Aug 2024 14:39



Tgt Ion:180 Resp: 38265
Ion Ratio Lower Upper
180 100
182 95.5 48.9 146.8
145 36.7 16.8 50.4



Manual Integration Report

Sequence:

vn080724

Instrument

MSVOA_n

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDICC100	VN083135.D	1,2,3-Trichloropropane	JOHN	8/8/2024 9:35:12 AM	MMDadoda	8/9/2024 6:15:44 AM	Peak Integrated by Software incorrectly
VSTDICCC050	VN083136.D	1,2,3-Trichloropropane	JOHN	8/8/2024 9:35:17 AM	MMDadoda	8/9/2024 6:15:45 AM	Peak Integrated by Software incorrectly
VSTDICCC050	VN083136.D	Vinyl Acetate	JOHN	8/8/2024 9:35:17 AM	MMDadoda	8/9/2024 6:15:45 AM	Peak Integrated by Software incorrectly
VSTDICC020	VN083137.D	1,2,3-Trichloropropane	JOHN	8/8/2024 9:35:21 AM	MMDadoda	8/9/2024 6:15:46 AM	Peak Integrated by Software incorrectly
VSTDICC020	VN083137.D	Vinyl Acetate	JOHN	8/8/2024 9:35:21 AM	MMDadoda	8/9/2024 6:15:46 AM	Peak Integrated by Software incorrectly
VSTDICC010	VN083138.D	1,2,3-Trichloropropane	JOHN	8/8/2024 9:35:26 AM	MMDadoda	8/9/2024 6:15:48 AM	Peak Integrated by Software incorrectly
VSTDICC005	VN083139.D	1,2,3-Trichloropropane	JOHN	8/8/2024 9:35:30 AM	MMDadoda	8/9/2024 6:15:49 AM	Peak Integrated by Software incorrectly
VSTDICC005	VN083139.D	Vinyl Acetate	JOHN	8/8/2024 9:35:30 AM	MMDadoda	8/9/2024 6:15:49 AM	Peak Integrated by Software incorrectly
VSTDICC001	VN083140.D	1,2,3-Trichloropropane	JOHN	8/8/2024 9:35:34 AM	MMDadoda	8/9/2024 6:15:51 AM	Peak Integrated by Software incorrectly
VSTDICC001	VN083140.D	1,4-Dichlorobenzene	JOHN	8/8/2024 9:35:34 AM	MMDadoda	8/9/2024 6:15:51 AM	Peak Integrated by Software incorrectly
VSTDICV050	VN083142.D	1,2,3-Trichloropropane	JOHN	8/8/2024 9:35:41 AM	MMDadoda	8/9/2024 6:15:52 AM	Peak Integrated by Software incorrectly

Manual Integration Report

Sequence:	vn081924	Instrument	MSVOA_n

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDCCC050	VN083367.D	1,2,3-Trichloropropane	SAM	8/20/2024 1:15:47 PM	MMDadoda	8/20/2024 4:36:00 PM	Peak Integrated by Software incorrectly
VSTDCCC050	VN083367.D	Vinyl Acetate	SAM	8/20/2024 1:15:47 PM	MMDadoda	8/20/2024 4:36:00 PM	Peak Integrated by Software incorrectly
VN0819WBS01	VN083370.D	1,2,3-Trichloropropane	SAM	8/20/2024 1:15:52 PM	MMDadoda	8/20/2024 4:36:01 PM	Peak Integrated by Software incorrectly
VN0819WBS01	VN083370.D	Vinyl Acetate	SAM	8/20/2024 1:15:52 PM	MMDadoda	8/20/2024 4:36:01 PM	Peak Integrated by Software incorrectly
VN0819WBSD0 1	VN083371.D	1,2,3-Trichloropropane	SAM	8/20/2024 1:16:01 PM	MMDadoda	8/20/2024 4:36:02 PM	Peak Integrated by Software incorrectly
VN0819WBSD0 1	VN083371.D	Vinyl Acetate	SAM	8/20/2024 1:16:01 PM	MMDadoda	8/20/2024 4:36:02 PM	Peak Integrated by Software incorrectly
VSTDCCC050	VN083389.D	1,2,3-Trichloropropane	SAM	8/20/2024 1:42:21 PM	MMDadoda	8/20/2024 4:36:04 PM	Peak Integrated by Software incorrectly
VSTDCCC050	VN083389.D	Vinyl Acetate	SAM	8/20/2024 1:42:21 PM	MMDadoda	8/20/2024 4:36:04 PM	Peak Integrated by Software incorrectly

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN080724

Review By	John Carlone	Review On	8/8/2024 9:36:00 AM
Supervise By	Maresh Dadoda	Supervise On	8/9/2024 6:16:03 AM
SubDirectory	VN080724	HP Acquire Method	HP Processing Method 82N080724W.M
STD. NAME	STD REF.#		
Tune/Reschk	VP129516		
Initial Calibration Stds	VP129526,VP129527,VP129528,VP129529,VP129530,VP129531		
CCC			
Internal Standard/PEM			
ICV/I.BLK	VP129532		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VN083134.D	07 Aug 2024 10:09	JC\MD	Ok
2	VSTDICC100	VN083135.D	07 Aug 2024 10:33	JC\MD	Ok,M
3	VSTDICCC050	VN083136.D	07 Aug 2024 10:58	JC\MD	Ok,M
4	VSTDICC020	VN083137.D	07 Aug 2024 11:22	JC\MD	Ok,M
5	VSTDICC010	VN083138.D	07 Aug 2024 11:46	JC\MD	Ok,M
6	VSTDICC005	VN083139.D	07 Aug 2024 12:10	JC\MD	Ok,M
7	VSTDICC001	VN083140.D	07 Aug 2024 12:34	JC\MD	Ok,M
8	IBLK	VN083141.D	07 Aug 2024 13:22	JC\MD	Ok
9	VSTDICV050	VN083142.D	07 Aug 2024 14:49	JC\MD	Ok,M

M : Manual Integration

Instrument ID: **MSVOA_N**

Daily Analysis Runlog For Sequence/QCBatch ID # VN081924

Review By	Semsettin Yesilyurt	Review On	8/20/2024 1:16:49 PM		
Supervise By	Maresh Dadoda	Supervise On	8/20/2024 4:36:05 PM		
SubDirectory	VN081924	HP Acquire Method	MSVOA_N	HP Processing Method	82n080724w.m
STD. NAME		STD REF.#			
Tune/Reschk Initial Calibration Stds		VP129836			
CCC		VP129832,VP129834			
Internal Standard/PEM		VP128298			
ICV/I.BLK					
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VN083366.D	19 Aug 2024 10:59	JC\MD	Ok
2	VSTDCCC050	VN083367.D	19 Aug 2024 11:43	JC\MD	Ok,M
3	VN0819MBL01	VN083368.D	19 Aug 2024 12:21	JC\MD	Ok
4	VN0819WBL01	VN083369.D	19 Aug 2024 13:37	JC\MD	Ok
5	VN0819WBS01	VN083370.D	19 Aug 2024 14:14	JC\MD	Ok,M
6	VN0819WBSD01	VN083371.D	19 Aug 2024 14:39	JC\MD	Ok,M
7	VIBLK	VN083372.D	19 Aug 2024 15:03	JC\MD	Ok
8	P3629-01	VN083373.D	19 Aug 2024 15:27	JC\MD	Ok
9	P3629-02	VN083374.D	19 Aug 2024 15:51	JC\MD	Ok
10	P3629-06	VN083375.D	19 Aug 2024 16:16	JC\MD	Ok
11	P3629-07	VN083376.D	19 Aug 2024 16:40	JC\MD	ReRun
12	P3643-16	VN083377.D	19 Aug 2024 17:04	JC\MD	Ok
13	P3657-02	VN083378.D	19 Aug 2024 17:29	JC\MD	Ok
14	P3657-01	VN083379.D	19 Aug 2024 17:53	JC\MD	Ok
15	P3645-01	VN083380.D	19 Aug 2024 18:17	JC\MD	Ok
16	P3645-02	VN083381.D	19 Aug 2024 18:42	JC\MD	Ok
17	P3646-01	VN083382.D	19 Aug 2024 19:06	JC\MD	Ok
18	P3646-02	VN083383.D	19 Aug 2024 19:30	JC\MD	Ok
19	P3662-01	VN083384.D	19 Aug 2024 19:55	JC\MD	Ok
20	P3662-02	VN083385.D	19 Aug 2024 20:19	JC\MD	Ok
21	P3660-09	VN083386.D	19 Aug 2024 20:43	JC\MD	Ok

Instrument ID: **MSVOA_N**

Daily Analysis Runlog For Sequence/QCBatch ID # VN081924

Review By	Semsettin Yesilyurt	Review On	8/20/2024 1:16:49 PM		
Supervise By	Mahesh Dadoda	Supervise On	8/20/2024 4:36:05 PM		
SubDirectory	VN081924	HP Acquire Method	MSVOA_N	HP Processing Method	82n080724w.m
STD. NAME		STD REF.#			
Tune/Reschk Initial Calibration Stds		VP129836			
CCC		VP129832,VP129834			
Internal Standard/PEM		VP128298			
ICV/I.BLK					
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

22	P3663-13	VN083387.D	19 Aug 2024 21:07	JC\MD	Ok
23	P3663-14	VN083388.D	19 Aug 2024 21:31	JC\MD	Ok
24	VSTDCCC050	VN083389.D	19 Aug 2024 21:56	JC\MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN080724

Review By	John Carlone	Review On	8/8/2024 9:36:00 AM
Supervise By	Maresh Dadoda	Supervise On	8/9/2024 6:16:03 AM
SubDirectory	VN080724	HP Acquire Method	HP Processing Method 82N080724W.M
STD. NAME	STD REF.#		
Tune/Reschk	VP129516		
Initial Calibration Stds	VP129526,VP129527,VP129528,VP129529,VP129530,VP129531		
CCC			
Internal Standard/PEM			
ICV/I.BLK	VP129532		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN083134.D	07 Aug 2024 10:09		JC\MD	Ok
2	VSTDICC100	VSTDICC100	VN083135.D	07 Aug 2024 10:33	Good for DOD	JC\MD	Ok,M
3	VSTDICCC050	VSTDICCC050	VN083136.D	07 Aug 2024 10:58	QR- 43	JC\MD	Ok,M
4	VSTDICC020	VSTDICC020	VN083137.D	07 Aug 2024 11:22		JC\MD	Ok,M
5	VSTDICC010	VSTDICC010	VN083138.D	07 Aug 2024 11:46		JC\MD	Ok,M
6	VSTDICC005	VSTDICC005	VN083139.D	07 Aug 2024 12:10		JC\MD	Ok,M
7	VSTDICC001	VSTDICC001	VN083140.D	07 Aug 2024 12:34		JC\MD	Ok,M
8	IBLK	IBLK	VN083141.D	07 Aug 2024 13:22		JC\MD	Ok
9	VSTDICV050	ICVVN080724	VN083142.D	07 Aug 2024 14:49		JC\MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN081924

Review By	Semsettin Yesilyurt	Review On	8/20/2024 1:16:49 PM				
Supervise By	Maresh Dadoda	Supervise On	8/20/2024 4:36:05 PM				
SubDirectory	VN081924	HP Acquire Method	MSVOA_N	HP Processing Method	82n080724w.m		
STD. NAME	STD REF.#						
Tune/Reschk Initial Calibration Stds	VP129836						
CCC	VP129832,VP129834						
Internal Standard/PEM	VP128298						
ICV/I.BLK							
Surrogate Standard							
MS/MSD Standard							
LCS Standard							

Sr#	Sampleld	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN083366.D	19 Aug 2024 10:59		JC\MD	Ok
2	VSTDCCC050	VSTDCCC050	VN083367.D	19 Aug 2024 11:43		JC\MD	Ok,M
3	VN0819MBL01	VN0819MBL01	VN083368.D	19 Aug 2024 12:21		JC\MD	Ok
4	VN0819WBL01	VN0819WBL01	VN083369.D	19 Aug 2024 13:37		JC\MD	Ok
5	VN0819WBS01	VN0819WBS01	VN083370.D	19 Aug 2024 14:14		JC\MD	Ok,M
6	VN0819WBSD01	VN0819WBSD01	VN083371.D	19 Aug 2024 14:39		JC\MD	Ok,M
7	VIBLK	VIBLK	VN083372.D	19 Aug 2024 15:03		JC\MD	Ok
8	P3629-01	50405	VN083373.D	19 Aug 2024 15:27	vial B pH<2 see non conf.	JC\MD	Ok
9	P3629-02	50398	VN083374.D	19 Aug 2024 15:51	vial B pH<2 see non conf.	JC\MD	Ok
10	P3629-06	SB-4	VN083375.D	19 Aug 2024 16:16	vial A pH<2	JC\MD	Ok
11	P3629-07	SB-5	VN083376.D	19 Aug 2024 16:40	vial A pH<2 Surrogate Fail	JC\MD	ReRun
12	P3643-16	TB-04-081524	VN083377.D	19 Aug 2024 17:04	vial A pH<2 TB, hit of com#16,no sample to run confirmation	JC\MD	Ok
13	P3657-02	TB-01-081624	VN083378.D	19 Aug 2024 17:29	vial A pH<2 TB	JC\MD	Ok
14	P3657-01	917-J-WS-081624	VN083379.D	19 Aug 2024 17:53	vial A pH<2	JC\MD	Ok
15	P3645-01	914-J-WS-081524	VN083380.D	19 Aug 2024 18:17	vial A pH<2	JC\MD	Ok
16	P3645-02	916-J-WS-081524	VN083381.D	19 Aug 2024 18:42	vial A pH<2	JC\MD	Ok
17	P3646-01	914-J-WPO-0.25-081524	VN083382.D	19 Aug 2024 19:06	vial A pH<2	JC\MD	Ok

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN081924

Review By	Semsettin Yesilyurt	Review On	8/20/2024 1:16:49 PM		
Supervise By	Mahesh Dadoda	Supervise On	8/20/2024 4:36:05 PM		
SubDirectory	VN081924	HP Acquire Method	MSVOA_N	HP Processing Method	82n080724w.m
STD. NAME		STD REF.#			
Tune/Reschk Initial Calibration Stds		VP129836			
CCC		VP129832,VP129834			
Internal Standard/PEM		VP128298			
ICV/I.BLK					
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

18	P3646-02	916-J-WPO-0.25-08152	VN083383.D	19 Aug 2024 19:30	vial A pH<2	JC\MD	Ok
19	P3662-01	917-J-WPO-0.25-08162	VN083384.D	19 Aug 2024 19:55	vial A pH<2	JC\MD	Ok
20	P3662-02	TB-02-081624	VN083385.D	19 Aug 2024 20:19	vial A pH<2 TB	JC\MD	Ok
21	P3660-09	TB-03-081624	VN083386.D	19 Aug 2024 20:43	vial A pH<2 TB	JC\MD	Ok
22	P3663-13	TB-04-081624	VN083387.D	19 Aug 2024 21:07	vial A pH<2 TB	JC\MD	Ok
23	P3663-14	TB-05-081624	VN083388.D	19 Aug 2024 21:31	vial A pH<2 TB	JC\MD	Ok
24	VSTDCCC050	VSTDCCC050EC	VN083389.D	19 Aug 2024 21:56		JC\MD	Ok,M

M : Manual Integration

Prep Standard - Chemical Standard Summary

Order ID : P3657

Test : VOCMS Group6

Prepbatch ID :

Sequence ID/Qc Batch ID: vn081924,

Standard ID :

VP126666,VP128290,VP128298,VP128762,VP128766,VP129228,VP129517,VP129832,VP129834,VP129836,

Chemical ID :

V12794,V12798,V13390,V13444,V13462,V13463,V13708,V13800,V13801,V13952,V13953,V14016,V14017,V14103,V14104,V14141,V14143,V14147,V14148,V14169,V14170,V14202,V14207,V14219,V14288,V 14411,V14412,V14413,V14414,W3112,

VOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
617	8260 Surrogate, 400PPM	VP126666	03/19/2024	09/19/2024	Semsettin Yesilyurt	None	None	Maresh Dadoda
03/28/2024								

FROM 0.80000ml of V13708 + 49.20000ml of V14141 = Final Quantity: 50.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
218	BFB, 25PPM	VP128290	06/10/2024	11/23/2024	Semsettin Yesilyurt	None	None	Maresh Dadoda
06/12/2024								

FROM 0.25000ml of V13390 + 24.75000ml of V14148 = Final Quantity: 25.000 ml

VOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
247	8260 Internal Standard, 250PPM	VP128298	06/10/2024	11/23/2024	Semsettin Yesilyurt	None	None	Maresh Dadoda
								06/12/2024

FROM 0.10000ml of V14288 + 9.90000ml of V14148 = Final Quantity: 10.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1810	8260 Working Std(2-CVE)-800ppm	VP128762	07/01/2024	12/11/2024	Semsettin Yesilyurt	None	None	Maresh Dadoda
								07/02/2024

FROM 0.50000ml of V12798 + 1.50000ml of V12794 + 23.00000ml of V14147 = Final Quantity: 25.000 ml

VOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
719	8260 Working STD (BCM)-First source, 400PPM	VP128766	07/01/2024	12/11/2024	Semsettin Yesilyurt	None	None	Maresh Dadoda
								07/02/2024

FROM 1.50000ml of V13462 + 1.50000ml of V13463 + 12.00000ml of V14147 = Final Quantity: 15.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
51	8260 Working STD (Acrolein) -first source, 800PPM	VP129228	07/25/2024	08/24/2024	Semsettin Yesilyurt	None	None	Maresh Dadoda
								07/30/2024

FROM 1.00000ml of V14411 + 1.00000ml of V14412 + 1.00000ml of V14413 + 1.00000ml of V14414 + 21.00000ml of V14143 =
Final Quantity: 25.000 ml

VOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
257	8260 Calibration Working STD Mix-First source, 160PPM	VP129517	08/05/2024	09/14/2024	Semsettin Yesilyurt	None	None	Maresh Dadoda 08/08/2024

FROM 0.40000ml of V13444 + 1.00000ml of V13800 + 1.00000ml of V13801 + 1.00000ml of V13952 + 1.00000ml of V13953 + 1.00000ml of V14016 + 1.00000ml of V14017 + 1.00000ml of V14103 + 1.00000ml of V14104 + 1.00000ml of V14169 + 1.00000ml of V14170 + 1.00000ml of V14219 + 1.50000ml of V14202 + 1.50000ml of V14207 + 10.60000ml of V14143 =
Final Quantity: 25.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
620	50 PPB CCC, 8260-Water	VP129832	08/19/2024	08/20/2024	Amit Patel	None	None	Maresh Dadoda 08/21/2024

FROM 39.94450ml of W3112 + 0.00500ml of VP126666 + 0.00500ml of VP128766 + 0.00800ml of VP128298 + 0.01250ml of VP128762 + 0.01250ml of VP129228 + 0.01250ml of VP129517 = Final Quantity: 40.000 ml



VOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
620	50 PPB CCC, 8260-Water	VP129834	08/19/2024	08/20/2024	Amit Patel	None	None	Mahesh Dadoda 08/21/2024
<u>FROM</u> 39.94450ml of W3112 + 0.00500ml of VP126666 + 0.00500ml of VP128766 + 0.00800ml of VP128298 + 0.01250ml of VP128762 + 0.01250ml of VP129228 + 0.01250ml of VP129517 = Final Quantity: 40.000 ml								

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
589	BFB TUNE CHECK	VP129836	08/19/2024	08/20/2024	Amit Patel	None	None	Mahesh Dadoda 08/21/2024
<u>FROM</u>	39.98400ml of W3112 + 0.01600ml of VP128290 = Final Quantity: 40.000 ml							

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	95318 / 2-Chloroethyl Vinyl Ether (Min = 5)	121321	12/13/2024	06/25/2024 / SAM	03/30/2022 / SAM	V12794

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	95318 / 2-Chloroethyl Vinyl Ether (Min = 5)	121321	12/13/2024	06/25/2024 / SAM	03/30/2022 / SAM	V12798

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30067 / BFB tuneing solution	A0191805	12/08/2024	12/08/2023 / SAM	01/13/2023 / SAM	V13390

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30470 / VOA Stock Solution, tert-butanol std, 1mL, P&TM	A0181905	12/14/2024	06/14/2024 / SAM	01/23/2023 / SAM	V13444

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30225 / VOA Mix, bromochloromethane, 2000ug/mL, P&TM, 1mL/ampul	A0193071	01/01/2025	07/01/2024 / SAM	01/27/2023 / SAM	V13462

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30225 / VOA Mix, bromochloromethane, 2000ug/mL, P&TM, 1mL/ampul	A0193071	01/01/2025	07/01/2024 / SAM	01/27/2023 / SAM	V13463

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555582 / Custom Mixture, 8260 A/B Surrogate Mix [CS 5179-2]	A0196865	09/19/2024	03/19/2024 / SAM	04/12/2023 / SAM	V13708

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30042 / VOA Mix,500 series method 502.2 Calibration Std #1 gases, 2000ug/ml, PTM, 1ml	A0194279	01/30/2025	07/30/2024 / SAM	05/31/2023 / SAM	V13800

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30042 / VOA Mix,500 series method 502.2 Calibration Std #1 gases, 2000ug/ml, PTM, 1ml	A0194279	12/28/2024	06/28/2024 / SAM	05/31/2023 / SAM	V13801

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30489 / VOA Mix, 8260B Acetates Mix, P&TM, 1mL	A0196115	09/30/2024	06/14/2024 / SAM	09/25/2023 / SAM	V13952

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30489 / VOA Mix, 8260B Acetates Mix, P&TM, 1mL	A0196115	09/30/2024	06/14/2024 / SAM	09/25/2023 / SAM	V13953

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	95319 / Revised Additions Mix (Min = 5)	032922	12/14/2024	06/14/2024 / SAM	11/22/2023 / SAM	V14016

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	95319 / Revised Additions Mix (Min = 5)	032922	12/14/2024	06/14/2024 / SAM	11/22/2023 / SAM	V14017

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555408 / Custom Standard, Vinyl Acetate Standard w/ Grav [CS 5066-6] TWO SEPARATE LOTS	A0205179	12/14/2024	06/14/2024 / SAM	12/22/2023 / SAM	V14103

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555408 / Custom Standard, Vinyl Acetate Standard w/ Grav [CS 5066-6] TWO SEPARATE LOTS	A0205179	12/14/2024	06/14/2024 / SAM	12/22/2023 / SAM	V14104

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA9077-02 / Methanol, Purge/Trap (cs=6x1L)	22L0562016	09/19/2024	03/19/2024 / SAM	02/06/2024 / SAM	V14141

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA9077-02 / Methanol, Purge/Trap (cs=6x1L)	22L0562016	01/22/2025	07/22/2024 / SAM	02/06/2024 / SAM	V14143

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA9077-02 / Methanol, Purge/Trap (cs=6x1L)	22L0562016	12/11/2024	06/11/2024 / pedro	02/06/2024 / SAM	V14147

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA9077-02 / Methanol, Purge/Trap (cs=6x1L)	22L0562016	11/23/2024	05/23/2024 / pedro	02/06/2024 / SAM	V14148

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	95317 / Universal VOA Mega Mix (Min order = 5)	021624	12/14/2024	06/14/2024 / SAM	02/20/2024 / SAM	V14169

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	95317 / Universal VOA Mega Mix (Min order = 5)	021624	12/14/2024	06/14/2024 / SAM	02/20/2024 / SAM	V14170

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30006 / VOA Mix, CLP method Calibration Std #1 ketones 5000uq/ml, PTM, 1ml	A0200785	12/25/2024	06/25/2024 / SAM	02/28/2024 / SAM	V14202

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30006 / VOA Mix, CLP method Calibration Std #1 ketones 5000uq/ml, PTM, 1ml	A0200785	12/25/2024	06/25/2024 / SAM	02/28/2024 / SAM	V14207

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30006 / VOA Mix, CLP method Calibration Std #1 ketones 5000uq/ml, PTM, 1ml	A0200785	12/25/2024	06/25/2024 / SAM	02/28/2024 / SAM	V14219

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555581 / Custom Standard, 8260 Internal Std [CS 5179-1]	A0210184	06/10/2025	06/10/2024 / SAM	04/15/2024 / SAM	V14288

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	91980 / Acrolin Std (Min = 5)	072424	08/24/2024	07/25/2024 / SAM	07/25/2024 / SAM	V14411

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	91980 / Acrolin Std (Min = 5)	072424	08/24/2024	07/25/2024 / SAM	07/25/2024 / SAM	V14412

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	91980 / Acrolin Std (Min = 5)	072424	08/24/2024	07/25/2024 / SAM	07/25/2024 / SAM	V14413

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	91980 / Acrolin Std (Min = 5)	072424	08/24/2024	07/25/2024 / SAM	07/25/2024 / SAM	V14414

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	DIW / DI Water	Daily Lab-Certified	07/03/2029	07/03/2024 / lwona	07/03/2024 / lwona	W3112



CERTIFIED WEIGHT REPORT

Part Number: **95317**
Lot Number: **021624**
Description: **Universal VOA Megamix**
69 components
Expiration Date: 021627
Recommended Storage: Freezer (0 °C)
Nominal Concentration (µg/mL): 2000
NIST Test ID#: 8UTB

Solvent(s): **Methanol**
Lot#: **EG359-USQ12**

Weight(s) shown below were combined and diluted to (mL): **100.0** **0.021** Balance Uncertainty: **5E-05** Flask Uncertainty: **0.021**


		021624
Formulated By:	Prashant Chauhan	DATE
		021624
Reviewed By:	Pedro L. Rentes	DATE

Compound	(RM#)	Lot	Di.	Initial	Initial	Nominal	Purity	Purity	Uncertainty	Target	Actual	Actual	Expanded	SDS Information		
														(Solvent Safety Info. On Attached pg.)		
														CAS#	OSHA PEL (TWA)	LD50
Part Number	Number	Factor	Vol. (mL)	Conc.(µg/mL)	Conc. (µg/mL)	(%)	Uncertainty	Pipette (mL)	Weight(g)	Weight(g)	Conc (µg/mL)	(+/-) (µg/mL)				
Acetonitrile	(0324)	021644	NA	NA	NA	2000	99.99	0.2	NA	0.20007	0.20020	2001.3	8.1	75-05-8	40 ppm (70mg/m3/8H)	or-rat 2400mg/kg
Allyl chloride (3-Chloropropene)	(0325)	102396	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20221	2001.4	8.2	107-05-1	1 ppm (3mg/m3/8H)	or-rat 700mg/kg
Carbon disulfide	(0060)	MKCR8581	NA	NA	NA	2000	99.99	0.2	NA	0.20007	0.20023	2001.6	8.1	75-15-0	4 ppm (12mg/m3) (skin)	or-rat 1200mg/kg
cis-1,4-Dichloro-2-butene	(1196)	14718EF	NA	NA	NA	2000	95	0.2	NA	0.21058	0.21069	2001.1	8.5	1476-11-5	N/A	N/A
trans-1,4-Dichloro-2-butene	(0486)	MKBP6041V	NA	NA	NA	2000	96.5	0.2	NA	0.20731	0.20748	2001.7	8.4	110-57-6	N/A	N/A
Diethyl ether	(0153)	IK18CA5000K	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20040	2001.5	8.1	60-29-7	N/A	N/A
Ethyl methacrylate	(0381)	06126PK	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20230	2002.3	8.2	97-63-2	N/A	or-rat 14800mg/kg
Iodomethane	(0489)	5HBF8718V	NA	NA	NA	2000	99.5	0.2	NA	0.20106	0.20121	2001.5	8.2	74-88-4	5 ppm (28mg/m3/8H) (skin)	or-rat 70mg/kg
2-Methyl-1-propanol	(0445)	15241EB	NA	NA	NA	2000	99.5	0.2	NA	0.20106	0.20120	2001.4	8.1	78-83-1	50 ppm (150mg/m3/8H)	or-rat 3480mg/kg
Methacrylonitrile	(0442)	00427ET	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20221	2001.4	8.2	126-98-7	1 ppm (3mg/m3/8H) (skin)	or-rat 1200mg/kg
Methyl acrylate	(1075)	SHBK0679	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20040	2001.5	8.1	96-33-3	10 ppm (35mg/m3/8H) (skin)	or-rat 277mg/kg
Methyl methacrylate	(0404)	MKBW5137V	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20041	2001.6	8.1	80-62-6	100 ppm (410mg/m3/8H)	or-rat 7872mg/kg
Nitrobenzene	(0228)	01213TV	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20220	2001.3	8.2	98-95-3	1 ppm (5mg/m3/8H) (skin)	or-rat 700mg/kg
2-Nitropropane	(0461)	14002JK	NA	NA	NA	2000	97.3	0.2	NA	0.20560	0.20577	2001.6	8.3	79-46-9	10 ppm (35mg/m3/8H)	or-rat 720mg/kg
Pentachloroethane	(0450)	HGA01	NA	NA	NA	2000	98	0.2	NA	0.20413	0.20430	2001.6	8.3	76-01-7	N/A	N/A
1,1,2-Trichlorotrifluoroethane	(0474)	18930	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20225	2001.8	8.2	76-13-1	1000 ppm (7800mg/m3/8H)	or-rat 43g/kg
Bromodichloromethane	35171	101623	0.05	5.00	40001.7	2000	NA	NA	0.017	NA	NA	1999.6	22.9	75-27-4	N/A	or-rat 910mg/kg
Dibromochloromethane	35171	101623	0.05	5.00	40002.1	2000	NA	NA	0.017	NA	NA	1999.6	23.0	124-48-1	N/A	or-rat 840mg/kg
cis-1,2-Dichloroethane	35171	101623	0.05	5.00	40003.1	2000	NA	NA	0.017	NA	NA	1999.7	22.9	156-59-2	N/A	N/A
trans-1,2-Dichloroethane	35171	101623	0.05	5.00	40002.4	2000	NA	NA	0.017	NA	NA	1999.6	23.0	156-60-5	N/A	or-rat 1235mg/kg
Methylene chloride	35171	101623	0.05	5.00	40002.8	2000	NA	NA	0.017	NA	NA	1999.6	22.9	75-09-2	500 ppm	or-rat 820mg/kg
1,1-Dichloroethene	32251	102023	0.10	10.00	20001.6	2000	NA	NA	0.042	NA	NA	1999.7	20.4	75-35-4	1 ppm (4mg/m3/8H)	or-rat 200mg/kg
Bromoform	95321	020724	0.10	10.00	20003.2	2000	NA	NA	0.042	NA	NA	1999.8	20.5	75-25-2	0.5 ppm (5mg/m3) (skin)	or-rat 930mg/kg
Carbon tetrachloride	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.8	20.4	56-23-5	2 ppm (12.6mg/m3/8H)	or-rat 2350mg/kg
Chloroform	95321	020724	0.10	10.00	20024.0	2000	NA	NA	0.042	NA	NA	2001.9	20.5	67-66-3	50 ppm (240mg/m3) (CL)	or-rat 900mg/kg
Dibromomethane	95321	020724	0.10	10.00	20002.9	2000	NA	NA	0.042	NA	NA	1999.8	20.5	74-95-3	N/A	or-rat 105mg/kg
1,1-Dichloroethane	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.8	20.5	75-34-3	100 ppm	or-rat 725mg/kg
2,2-Dichloropropane	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.8	20.4	594-20-7	N/A	N/A
Tetrachloroethane	95321	020724	0.10	10.00	20201.1	2000	NA	NA	0.042	NA	NA	2019.6	20.8	127-18-4	25 ppm (170mg/m3/8H) (final)	or-rat 2629mg/kg
1,1,1-Trichloroethane	95321	020724	0.10	10.00	20003.0	2000	NA	NA	0.042	NA	NA	1999.8	20.5	71-55-6	350 ppm (1900mg/m3/8H)	or-rat 10300mg/kg
1,2-Dibromo-3-chloropropane	35161	112322	0.05	5.00	40016.5	2000	NA	NA	0.017	NA	NA	2000.3	22.9	96-12-6	0.001 ppm	or-rat 170mg/kg
1,2-Dibromomethane	35161	112322	0.05	5.00	40024.8	2000	NA	NA	0.017	NA	NA	2000.7	22.9	106-93-4	20 ppm (8H)	or-rat 108mg/kg
1,2-Dichloroethane	35161	112322	0.05	5.00	40018.0	2000	NA	NA	0.017	NA	NA	2000.4	22.9	107-08-2	50 ppm (8H)	or-rat 870mg/kg
1,2-Dichloropropane	35161	112322	0.05	5.00	40051.0	2000	NA	NA	0.017	NA	NA	2000.2	22.9	78-57-5	75 ppm (350mg/m3/8H)	or-rat 1947mg/kg
1,3-Dichloropropane	35161	112322	0.05	5.00	40005.9	2000	NA	NA	0.017	NA	NA	1999.8	22.9	142-28-9	N/A	un-rat 3600mg/kg
1,1-Dichloropropene	35161	112322	0.05	5.00	40012.1	2000	NA	NA	0.017	NA	NA	2000.1	29.7	563-56-6	N/A	N/A
cis-1,3-Dichloropropene	35161	112322	0.05	5.00	40010.0	2000	NA	NA	0.017	NA	NA	2000.0	23.0	10061-01-5	N/A	N/A
trans-1,3-Dichloropropene	35161	112322	0.05	5.00	40017.8	2000	NA	NA	0.017	NA	NA	2000.4	23.0	10061-02-6	N/A	N/A
Hexachloro-1,3-butadiene	35161	112322	0.05	5.00	40021.9	2000	NA	NA	0.017	NA	NA	2000.6	29.7	87-68-3	0.02 ppm (0.24mg/m3/8H)	or-rat 82mg/kg
1,1,1,2-Tetrachloroethane	35161	112322	0.05	5.00	40011.9	2000	NA	NA	0.017	NA	NA	2000.1	22.9	630-20-6	N/A	or-rat 670mg/kg
1,1,2,2-Tetrachloroethane	35161	112322	0.05	5.00	40007.5	2000	NA	NA	0.017	NA	NA	1999.9	22.9	79-34-5	5 ppm (35mg/m3/8H) (skin)	or-rat 800mg/kg
1,1,2-Trichloroethane	35161	112322	0.05	5.00	40006.6	2000	NA	NA	0.017	NA	NA	1999.8	23.0	79-00-5	10 ppm (46mg/m3/8H) (skin)	or-rat 836mg/kg
2,3-Trichloropropane	35161	112322	0.05	5.00	40007.5	2000	NA	NA	0.017	NA	NA	2000.9	22.9	79-01-6	50 ppm (270mg/m3/8H)	or-rat 2400mg/kg
Benzene	35162	050823	0.05	5.00	40005.0	2000	NA	NA	0.017	NA	NA	1999.9	22.9	96-18-4	10 ppm (50mg/m3/8H)	or-rat 149.6mg/kg
Bromobenzene	35162	050823	0.05	5.00	40006.9	2000	NA	NA	0.017	NA	NA	1999.8	22.9	104-51-8	1 ppm	or-rat 4694mg/kg
n-Butyl benzene	35162	050823	0.05	5.00	40003.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	108-98-1	N/A	or-rat 2699mg/kg
Ethyl benzene	35162	050823	0.05	5.00	40004.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	104-51-8	N/A	N/A
Isopropyl toluene	35162	050823	0.05	5.00	40005.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	100-41-4	100 ppm (435mg/m3/8H)	or-rat >2000mg/kg
Naphthalene	35162	050823	0.05	5.00	40006.2	2000	NA	NA	0.017	NA	NA	1999.8	22.9	99-87-6	N/A	or-rat 4750mg/kg
Styrene	35162	050823	0.05	5.00	40004.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	91-20-3	10 ppm (50mg/m3/8H)	or-rat 490mg/kg
Toluene	35162	050823	0.05	5.00	40006.2	2000	NA	NA	0.017	NA	NA	1999.7	22.9	100-42-5	100 ppm	or-rat 5000mg/kg
2,3-Trichlorobenzene	35162	050823	0.05	5.00	40003.1	2000	NA	NA	0.017	NA	NA	1999.8	22.9	106-98-3	200 ppm	or-rat 5000mg/kg
2,4-Trichlorobenzene	35162	050823	0.05	5.00	40006.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	87-61-6	N/A	or-rat 1390mg/kg
2,4,6-Trimethylbenzene	35162	050823	0.05	5.00	40001.6	2000	NA	NA	0.017	NA	NA	1999.8	22.9	120-82-1	5 ppm (CL) (40mg/m3)	or-rat 750mg/kg
3,5-Trimethylbenzene	35162	050823	0.05	5.00	40006.7	2000	NA	NA	0.017	NA	NA	1999.8	22.9	95-63-6	N/A	or-rat 5g/kg
Xylene	35162	050823	0.05	5.00	40006.7	2000	NA	NA	0.017	NA	NA	1999.8	22.9	106-67-8	N/A	or-rat 5000mg/kg
tert-Butyl benzene	35163	101923	0.05	5.00	40001.2	2000	NA	NA	0.017	NA	NA	1999.6	22.9	98-06-6	100 ppm (435mg/m3/8H)	or-rat 5g/kg
sec-Butyl benzene	35163	101923	0.05	5.00	40002.4	2000	NA	NA	0.017	NA	NA	1999.6	22.9	135-98-8	N/A	N/A
Chlorobenzene	35163	101923	0.05	5.00	40003.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	108-90-7	75 ppm (350mg/m3/8H)	or-rat 2240mg/kg
Chlorotoluene	35163	101923	0.05	5.00	40000.3	2000	NA	NA	0.017	NA	NA	1999.5	22.9	108-90-7	75 ppm (350mg/m3/8H)	or-rat 2290mg/kg
Chlorotoluene	35163	101923	0.05	5.00	40003.3	2000	NA	NA	0.017	NA	NA	1999.5	22.9	95-49-8	50 ppm (250mg/m3/8H)	or-rat 3900mg/kg
2-Dichlorobenzene	35163	101923	0.05	5.00	40003.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	106-43-4	N/A	or-rat 2100mg/kg
3-Dichlorobenzene	35163	101923	0.05	5.00	40001.7	2000	NA	NA	0.017	NA	NA	1999.7	22.9	95-50-1	50 ppm (300mg/m3) (CL)	or-rat 500mg/kg
4-Dichlorobenzene	35163	101923	0.05	5.00	40001.8	2000	NA	NA	0.017	NA	NA	1999.6	22.9	106-46-7	75 ppm (450mg/m3/8H)	or-rat 500mg/kg
propylbenzene	35163	101923	0.05	5.00	40000.8	2000	NA	NA	0.017	NA	NA	1999.5	22.9	9		

Run 16, "P95317 L021624 [2000µg/mL in MeOH]"

Run Length: 60.00 min, 35998 points at 10 points/second.
Created: Sat, Feb 17, 2024 at 8:56:46 AM.
Sampled: Sequence "021624-GC5M1", Method "GC5-M1".
Analyzed using Method "GC5-M1".

Comments
GC5-M1 Analysis by Candice Warren
Column ID SPB-Vocol 105 meter X 0.53mm X 3.0µm film thickness
Flow rates: Total flow=290mL/min., Helium (carrier)=10mL/min.,
Helium(make-up)=10mL/min., Hydrogen(make-up)=40mL/min., Air(make-up)=230mL/min.
Oven Profile: Temp. 1=35°C (Time 1=10 min.), Temp 2=200°C (Time 2=8.75 min.),
Rate = 4°C/min., Total run time=60 min. Injector temp.=200°C, FID Temp.=200°C.
FID Signal = Edaq Channel 1
Standard injection = 0.5µL, Range=3



Peak #	Name	FID RT (min.)
1	Ether	9.97
2	1,1,2-Trichloro-1,2,2-trifluoroethane	10.33
3	1,1-Dichloroethane	11.10
4	Acetonitrile	12.00
5	Iodomethane	12.31
6	Allyl chloride	12.56
7	Carbon disulfide/Methylene chloride	13.04
8	trans-1,2-Dichloroethane	14.07
9	1,1-Dichloroethane	15.74
10	2,2-Dichloropropane	17.24
11	cis-1,2-Dichloroethane	18.00
12	Methacrylonitrile/methyl acrylate/Chloroform	18.46
13	Isobutane/1,1,1-Trichloroethane	19.01
14	1,1-Dichloropropane	20.46
15	Carbon tetrachloride	20.79
16	Benzene/1,2-Dichloroethane	21.49
17	Trichloroethane	23.58
18	1,2-Dichloropropane	24.38
19	Methyl methacrylate	24.52
20	Bromochloromethane	25.13
21	Dibromomethane/2-Hitropropane	25.46
22	cis-1,3-Dichloropropane	27.02
23	Toluene	28.05
24	Ethyl methacrylate/trans-1,3-Dichloropropane	28.73
25	1,1,2-Trichloroethane	29.34
26	Tetrachloroethene/1,3-Dichloropropane	30.24
27	Dibromochloromethane	31.16
28	1,2-Dibromomethane	31.84
29	Chlorobenzene	33.26
30	Ethylbenzene/1,1,1,2-Tetrachloroethane	33.40
31	m-Xylene/p-Xylene	33.66
32	o-Xylene	35.22
33	Styrene	35.39
34	Isopropylbenzene/Bromoforn	36.18
35	cis-1,4-Dichloro-3-butene	36.80
36	1,1,2,2-Tetrachloroethane	37.23
37	1,2,3-Trichloropropane	37.77
38	n-Propylbenzene	37.93
39	trans-1,4-Dichloro-3-butene	38.05
40	Bromobenzene	38.14
41	1,3,5-Trimethylbenzene	38.80
42	2-Chlorotoluene	38.82
43	4-Chlorotoluene	38.77
44	tert-Butylbenzene	39.74
45	1,2,4-Trimethylbenzene	39.91
46	Pentachloroethane	40.17
47	sec-Butylbenzene	40.52
48	p-Isopropyltoluene	41.02
49	1,3-Dichlorobenzene	41.42
50	1,4-Dichlorobenzene	41.83
51	n-Butylbenzene	42.53
52	1,2-Dichlorobenzene	43.18
53	1,2-Dibromo-3-chloropropane	46.12
54	Nitrobenzene	46.46
55	1,2,4-Trichlorobenzene	49.26
56	Hexachlorobutadiene	49.72
57	Naphthalene	50.26
58	1,2,3-Trichlorobenzene	51.16

Safety Data Sheet (SDS)

GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY ANALYTICAL STANDARD DISSOLVED IN METHANOL

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	1-800-535-5053
Address	44 Rossotto Dr. Hamden CT, 06514	Emergency Telephone International	1-352-323-3500
		Date Prepared/Revised	January 1, 2023

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

H225	Highly Flammable Liquid and Vapor	H301, 311, 331	Toxic if swallowed, skin contact, inhaled
H370	Cause damage to organs	H351	Suspected of causing cancer
P271	Use in ventilated area	P280	Use gloves, eye protection/face shield
P302,332	If on skin, wash with soap and water	P305,351,338	If in eyes, remove contacts, rinse with water



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))			% (optional)
Methanol	METHYL ALCOHOL	CAS#: 67-56-1	> 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance. Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Flammability	Flammable in the presence of a source of ignition when the temperature is above the flash point. Keep away from heat/sparks/open flame/hot surface. No smoking.
Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist.
Storage Conditions	Use ventilation. Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge. Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Methanol	67-56-1 TWA 200 ppm
Skin notation	TWA 200 ppm
Potential for skin absorption, ingestion and inhalation.	
Personal protective equipment	Respiratory protection. Handle with gloves. Gloves must be inspected prior to use. Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.	

Section IX - Physical/Chemical Characteristics

Boiling Point	65°C	Specific Gravity (H ₂ O = 1)	0.79
Vapor Pressure (mm Hg)	96	Melting Point	-98°C
Vapor Density (AIR = 1)	1.11	Evaporation rate (Butyl Acetate = 1)	4.6
Solubility in Water	COMPLETE		
Appearance and Odor	CLEAR, COLORLESS LIQUID WITH CHARACTERISTIC PUNGENT ODOR.		

Section X. STABILITY AND REACTIVITY

Chemical stability	Stable under recommended storage conditions.
Possibility of hazardous reactions	Vapours may form explosive mixture with air.
Conditions to avoid	Heat, flames, sparks, extreme temperature and sunlight.
Materials to avoid	Acid chlorides, Acid anhydrides, Oxidizing agents, Alkali metals, Reducing agents, Acids
Hazardous decomposition products formed under fire conditions.	- Carbon oxides

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - rat - 5,628 mg/kg
LC50 Inhalation - rat - 4 h - 64000 ppm
LD50 Dermal - rabbit - 15,800 mg/kg
Toxic if absorbed through skin. Causes skin irritation.
Eye damage/eye irritation
Toxic if inhaled. Causes respiratory tract irritation.
Toxic if swallowed.

Section XII. ECOLOGICAL INFORMATION FOR REPORTABLE QUANTITY OF 5000 lbs.

LC50 15,400 mg/l - 96 h
EC50 24,500.00 mg/l - 48 h
EC100 10,000.00 mg/l - 24 h

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)	IATA
UN number: 1230 Class: 3 Packing group: II	UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol	Proper shipping name: Methanol

Section XV. REGULATORY INFORMATION

OSHA Hazards Flammable liquid, Target Organ Effect, Toxic by inhalation., Toxic by ingestion, Toxic by skin absorption, Irritant
SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC. DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.



CERTIFIED WEIGHT REPORT

Part Number: **95317**
Lot Number: **021624**
Description: **Universal VOA Megamix**
69 components
Expiration Date: 021627
Recommended Storage: Freezer (0 °C)
Nominal Concentration (µg/mL): 2000
NIST Test ID#: 8UTB

Solvent(s): **Methanol**
Lot#: **EG359-USQ12**

Weight(s) shown below were combined and diluted to (mL): **100.0** **0.021** Balance Uncertainty: **5E-05** Flask Uncertainty: **0.021**


<i>Prashant Chauhan</i>		021624
Formulated By:	Prashant Chauhan	DATE
<i>Pedro L. Rentes</i>		021624
Reviewed By:	Pedro L. Rentes	DATE

Compound	(RM#)	Lot Number	Di. Factor	Initial Vol. (mL)	Initial Conc. (µg/mL)	Nominal Conc. (µg/mL)	Purity (%)	Purity Uncertainty	Uncertainty Pipette (mL)	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
														CAS#	OSHA PEL (TWA)	LD50
Acetonitrile	(0324)	021644	NA	NA	NA	2000	99.99	0.2	NA	0.20007	0.20020	2001.3	8.1	75-05-8	40 ppm (70mg/m3/8H)	or-rat 2400mg/kg
Allyl chloride (3-Chloropropene)	(0325)	102396	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20221	2001.4	8.2	107-05-1	1 ppm (3mg/m3/8H)	or-rat 700mg/kg
Carbon disulfide	(0060)	MKCR8581	NA	NA	NA	2000	99.99	0.2	NA	0.20007	0.20023	2001.6	8.1	75-15-0	4 ppm (12mg/m3) (skin)	or-rat 1200mg/kg
cis-1,4-Dichloro-2-butene	(1196)	14718EF	NA	NA	NA	2000	95	0.2	NA	0.21058	0.21069	2001.1	8.5	1476-11-5	N/A	N/A
trans-1,4-Dichloro-2-butene	(0486)	MKBP6041V	NA	NA	NA	2000	96.5	0.2	NA	0.20731	0.20748	2001.7	8.4	110-57-6	N/A	N/A
Diethyl ether	(0153)	IK18CA5000K	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20040	2001.5	8.1	60-29-7	N/A	N/A
Ethyl methacrylate	(0381)	06126PK	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20230	2002.3	8.2	97-63-2	N/A	or-rat 14800mg/kg
Iodomethane	(0489)	5HBF8718V	NA	NA	NA	2000	99.5	0.2	NA	0.20106	0.20121	2001.5	8.2	74-88-4	5 ppm (28mg/m3/8H) (skin)	or-rat 70mg/kg
2-Methyl-1-propanol	(0445)	15241EB	NA	NA	NA	2000	99.5	0.2	NA	0.20106	0.20120	2001.4	8.1	78-83-1	50 ppm (150mg/m3/8H)	or-rat 3480mg/kg
Methacrylonitrile	(0442)	00427ET	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20221	2001.4	8.2	126-98-7	1 ppm (3mg/m3/8H) (skin)	or-rat 1200mg/kg
Methyl acrylate	(1075)	SHBK0679	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20040	2001.5	8.1	96-33-3	10 ppm (35mg/m3/8H) (skin)	or-rat 277mg/kg
Methyl methacrylate	(0404)	MKBW5137V	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20041	2001.6	8.1	80-62-6	100 ppm (410mg/m3/8H)	or-rat 7872mg/kg
Nitrobenzene	(0228)	01213TV	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20220	2001.3	8.2	98-95-3	1 ppm (5mg/m3/8H) (skin)	or-rat 780mg/kg
2-Nitropropane	(0461)	14002JX	NA	NA	NA	2000	97.3	0.2	NA	0.20560	0.20577	2001.6	8.3	79-46-9	10 ppm (35mg/m3/8H)	or-rat 720mg/kg
Pentachloroethane	(0450)	HGA01	NA	NA	NA	2000	98	0.2	NA	0.20413	0.20430	2001.6	8.3	76-01-7	N/A	N/A
1,1,2-Trichlorotrifluoroethane	(0474)	18930	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20225	2001.8	8.2	76-13-1	1000 ppm (7800mg/m3/8H)	or-rat 43g/kg
Bromodichloromethane	35171	101623	0.05	5.00	40001.7	2000	NA	NA	0.017	NA	NA	1999.6	22.9	75-27-4	N/A	or-rat 910mg/kg
Dibromochloromethane	35171	101623	0.05	5.00	40002.1	2000	NA	NA	0.017	NA	NA	1999.6	23.0	124-48-1	N/A	or-rat 840mg/kg
cis-1,2-Dichloroethane	35171	101623	0.05	5.00	40003.1	2000	NA	NA	0.017	NA	NA	1999.7	22.9	156-59-2	N/A	N/A
trans-1,2-Dichloroethane	35171	101623	0.05	5.00	40002.4	2000	NA	NA	0.017	NA	NA	1999.6	23.0	156-60-5	N/A	or-rat 1235mg/kg
Methylene chloride	35171	101623	0.05	5.00	40002.8	2000	NA	NA	0.017	NA	NA	1999.6	22.9	75-09-2	500 ppm	or-rat 820mg/kg
1,1-Dichloroethene	32251	102023	0.10	10.00	20001.6	2000	NA	NA	0.042	NA	NA	1999.7	20.4	75-35-4	1 ppm (4mg/m3/8H)	or-rat 200mg/kg
Bromoform	95321	020724	0.10	10.00	20003.2	2000	NA	NA	0.042	NA	NA	1999.8	20.5	75-25-2	0.5 ppm (5mg/m3) (skin)	or-rat 930mg/kg
Carbon tetrachloride	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.8	20.4	56-23-5	2 ppm (12.6mg/m3/8H)	or-rat 2350mg/kg
Chloroform	95321	020724	0.10	10.00	20024.0	2000	NA	NA	0.042	NA	NA	2001.9	20.5	67-66-3	50 ppm (240mg/m3) (CL)	or-rat 900mg/kg
Dibromomethane	95321	020724	0.10	10.00	20002.9	2000	NA	NA	0.042	NA	NA	1999.8	20.5	74-85-3	N/A	or-rat 105mg/kg
1,1-Dichloroethane	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.8	20.5	75-34-3	100 ppm	or-rat 725mg/kg
2,2-Dichloropropane	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.8	20.4	594-20-7	N/A	N/A
Tetrachloroethane	95321	020724	0.10	10.00	20201.1	2000	NA	NA	0.042	NA	NA	2019.6	20.8	127-18-4	25 ppm (170mg/m3/8H) (final)	or-rat 2629mg/kg
1,1,1-Trichloroethane	95321	020724	0.10	10.00	20003.0	2000	NA	NA	0.042	NA	NA	1999.8	20.5	71-55-6	350 ppm (1900mg/m3/8H)	or-rat 10300mg/kg
1,2-Dibromo-3-chloropropane	35161	112322	0.05	5.00	40016.5	2000	NA	NA	0.017	NA	NA	2000.3	22.9	96-12-6	0.001 ppm	or-rat 170mg/kg
1,2-Dibromomethane	35161	112322	0.05	5.00	40024.8	2000	NA	NA	0.017	NA	NA	2000.7	22.9	106-93-4	20 ppm (8H)	or-rat 108mg/kg
1,2-Dichloroethane	35161	112322	0.05	5.00	40018.0	2000	NA	NA	0.017	NA	NA	2000.4	22.9	107-08-2	50 ppm (8H)	or-rat 870mg/kg
1,2-Dichloropropane	35161	112322	0.05	5.00	40051.0	2000	NA	NA	0.017	NA	NA	2002.0	22.9	78-57-5	75 ppm (350mg/m3/8H)	or-rat 1947mg/kg
1,3-Dichloropropane	35161	112322	0.05	5.00	40005.9	2000	NA	NA	0.017	NA	NA	1999.8	22.9	142-28-9	N/A	un-rat 3600mg/kg
1,1-Dichloropropene	35161	112322	0.05	5.00	40012.1	2000	NA	NA	0.017	NA	NA	2000.1	29.7	563-56-6	N/A	N/A
cis-1,3-Dichloropropene	35161	112322	0.05	5.00	40010.0	2000	NA	NA	0.017	NA	NA	2000.0	23.0	10061-01-5	N/A	N/A
trans-1,3-Dichloropropene	35161	112322	0.05	5.00	40017.8	2000	NA	NA	0.017	NA	NA	2000.4	23.0	10061-02-6	N/A	N/A
Hexachloro-1,3-butadiene	35161	112322	0.05	5.00	40021.9	2000	NA	NA	0.017	NA	NA	2000.6	29.7	87-68-3	0.02 ppm (0.24mg/m3/8H)	or-rat 82mg/kg
1,1,1,2-Tetrachloroethane	35161	112322	0.05	5.00	40011.9	2000	NA	NA	0.017	NA	NA	2000.1	22.9	630-20-6	N/A	or-rat 670mg/kg
1,1,2,2-Tetrachloroethane	35161	112322	0.05	5.00	40007.5	2000	NA	NA	0.017	NA	NA	1999.9	22.9	79-34-5	5 ppm (35mg/m3/8H) (skin)	or-rat 800mg/kg
1,1,2-Trichloroethane	35161	112322	0.05	5.00	40006.6	2000	NA	NA	0.017	NA	NA	1999.8	23.0	79-00-5	10 ppm (46mg/m3/8H) (skin)	or-rat 836mg/kg
2,3-Trichloropropane	35161	112322	0.05	5.00	40007.5	2000	NA	NA	0.017	NA	NA	2000.9	22.9	79-01-6	50 ppm (270mg/m3/8H)	or-rat 2400mg/kg
Benzene	35162	050823	0.05	5.00	40005.0	2000	NA	NA	0.017	NA	NA	1999.9	22.9	96-18-4	10 ppm (50mg/m3/8H)	or-rat 149.6mg/kg
Bromobenzene	35162	050823	0.05	5.00	40006.9	2000	NA	NA	0.017	NA	NA	1999.8	22.9	104-51-8	1 ppm	or-rat 4694mg/kg
n-Butyl benzene	35162	050823	0.05	5.00	40003.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	108-88-1	N/A	or-rat 2699mg/kg
Ethyl benzene	35162	050823	0.05	5.00	40004.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	104-51-8	N/A	N/A
Isopropyl toluene	35162	050823	0.05	5.00	40005.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	100-41-4	100 ppm (435mg/m3/8H)	or-rat >2000mg/kg
Naphthalene	35162	050823	0.05	5.00	40006.2	2000	NA	NA	0.017	NA	NA	1999.8	22.9	99-87-6	N/A	or-rat 4750mg/kg
Styrene	35162	050823	0.05	5.00	40004.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	91-20-3	10 ppm (50mg/m3/8H)	or-rat 490mg/kg
Toluene	35162	050823	0.05	5.00	40006.2	2000	NA	NA	0.017	NA	NA	1999.7	22.9	100-42-5	100 ppm	or-rat 5000mg/kg
2,3-Trichlorobenzene	35162	050823	0.05	5.00	40003.1	2000	NA	NA	0.017	NA	NA	1999.8	22.9	106-88-3	200 ppm	or-rat 5000mg/kg
2,4-Trichlorobenzene	35162	050823	0.05	5.00	40006.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	87-61-6	N/A	or-rat 1390mg/kg
2,4-Trimethylbenzene	35162	050823	0.05	5.00	40001.6	2000	NA	NA	0.017	NA	NA	1999.8	22.9	120-82-1	5 ppm (CL) (40mg/m3)	or-rat 750mg/kg
3,5-Trimethylbenzene	35162	050823	0.05	5.00	40006.7	2000	NA	NA	0.017	NA	NA	1999.6	23.0	95-63-6	N/A	or-rat 5g/kg
Xylene	35162	050823	0.05	5.00	40006.7	2000	NA	NA	0.017	NA	NA	1999.6	22.9	106-57-8	N/A	or-rat 5000mg/kg
tert-Butyl benzene	35163	101923	0.05	5.00	40001.2	2000	NA	NA	0.017	NA	NA	1999.6	22.9	98-06-6	100 ppm (435mg/m3/8H)	or-rat 5g/kg
sec-Butyl benzene	35163	101923	0.05	5.00	40002.4	2000	NA	NA	0.017	NA	NA	1999.6	22.9	135-98-8	N/A	N/A
Chlorobenzene	35163	101923	0.05	5.00	40003.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	95-49-8	75 ppm (350mg/m3/8H)	or-rat 2240mg/kg
Chlorotoluene	35163	101923	0.05	5.00	40003.3	2000	NA	NA	0.017	NA	NA	1999.5	22.9	108-90-7	50 ppm (250mg/m3/8H)	or-rat 2290mg/kg
Chlorotoluene	35163	101923	0.05	5.00	40003.3	2000	NA	NA	0.017	NA	NA	1999.7	22.9	106-43-4	N/A	or-rat 3900mg/kg
2-Dichlorobenzene	35163	101923	0.05	5.00	40003.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	95-50-1	50 ppm (300mg/m3) (CL)	or-rat 500mg/kg
3-Dichlorobenzene	35163	101923	0.05	5.00	40001.7	2000	NA	NA	0.017	NA	NA	1999.6	23.0	541-73-1	N/A	or-rat 500mg/kg
4-Dichlorobenzene	35163	101923	0.05	5.00	40001.8	2000	NA	NA	0.017	NA	NA	1999.6	22.9	106-46-7	75 ppm (450mg/m3/8H)	or-rat 900mg/kg
propylbenzene	35163	101923	0.05	5.00	40000.8	2000	NA	NA	0.017	NA	NA	1999.5	22.9	96-82-8	50 ppm (245mg/m3/8H)	or-rat 500mg/kg
propylbenzene	35163	101923	0.05	5.00												

Run 16, "P95317 L021624 [2000µg/mL in MeOH]"

Run Length: 60.00 min, 35998 points at 10 points/second.
Created: Sat, Feb 17, 2024 at 8:56:46 AM.
Sampled: Sequence "021624-GC5M1", Method "GC5-M1".
Analyzed using Method "GC5-M1".

Comments
GC5-M1 Analysis by Candice Warren
Column ID SPB-Vocol 105 meter X 0.53mm X 3.0µm film thickness
Flow rates: Total flow=290mL/min., Helium (carrier)=10mL/min.,
Helium(make-up)=10mL/min., Hydrogen(make-up)=40mL/min., Air(make-up)=230mL/min.
Oven Profile: Temp. 1=35°C (Time 1=10 min.), Temp 2=200°C (Time 2=8.75 min.),
Rate = 4°C/min., Total run time=60 min. Injector temp.=200°C, FID Temp.=200°C.
FID Signal = Edaq Channel 1
Standard injection = 0.5µL, Range=3



Peak #	Name	FID RT (min.)
1	Ether	9.97
2	1,1,2-Trichloro-1,2,2-trifluoroethane	10.33
3	1,1-Dichloroethane	11.10
4	Acetonitrile	12.00
5	Iodomethane	12.31
6	Allyl chloride	12.56
7	Carbon disulfide/Methylene chloride	13.04
8	trans-1,2-Dichloroethane	14.07
9	1,1-Dichloroethane	15.74
10	2,2-Dichloropropane	17.24
11	cis-1,2-Dichloroethane	18.00
12	Methacrylonitrile/methyl acrylate/Chloroform	18.46
13	Isobutane/1,1,1-Trichloroethane	19.01
14	1,1-Dichloropropane	20.46
15	Carbon tetrachloride	20.79
16	Benzene/1,2-Dichloroethane	21.49
17	Trichloroethane	23.58
18	1,2-Dichloropropane	24.38
19	Methyl methacrylate	24.52
20	Bromochloromethane	25.13
21	Dibromomethane/2-Hitropropane	25.46
22	cis-1,3-Dichloropropane	27.02
23	Toluene	28.05
24	Ethyl methacrylate/trans-1,3-Dichloropropane	28.73
25	1,1,2-Trichloroethane	29.34
26	Tetrachloroethene/1,3-Dichloropropane	30.24
27	Dibromochloromethane	31.16
28	1,2-Dibromomethane	31.84
29	Chlorobenzene	33.26
30	Ethylbenzene/1,1,1,2-Tetrachloroethane	33.40
31	m-Xylene/p-Xylene	33.66
32	o-Xylene	35.22
33	Styrene	35.39
34	Isopropylbenzene/Bromoforn	36.18
35	cis-1,4-Dichloro-3-butene	36.80
36	1,1,2,2-Tetrachloroethane	37.23
37	1,2,3-Trichloropropane	37.77
38	n-Propylbenzene	37.93
39	trans-1,4-Dichloro-3-butene	38.05
40	Bromobenzene	38.14
41	1,3,5-Trimethylbenzene	38.80
42	2-Chlorotoluene	38.82
43	4-Chlorotoluene	38.77
44	tert-Butylbenzene	39.74
45	1,2,4-Trimethylbenzene	39.91
46	Pentachloroethane	40.17
47	sec-Butylbenzene	40.52
48	p-Isopropyltoluene	41.02
49	1,3-Dichlorobenzene	41.42
50	1,4-Dichlorobenzene	41.83
51	n-Butylbenzene	42.53
52	1,2-Dichlorobenzene	43.18
53	1,2-Dibromo-3-chloropropane	46.12
54	Nitrobenzene	46.46
55	1,2,4-Trichlorobenzene	49.26
56	Hexachlorobutadiene	49.72
57	Naphthalene	50.26
58	1,2,3-Trichlorobenzene	51.16

Safety Data Sheet (SDS)

GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY ANALYTICAL STANDARD DISSOLVED IN METHANOL

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	1-800-535-5053
Address	44 Rossotto Dr. Hamden CT, 06514	Emergency Telephone International	1-352-323-3500
		Date Prepared/Revised	January 1, 2023

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

H225	Highly Flammable Liquid and Vapor	H301, 311, 331	Toxic if swallowed, skin contact, inhaled
H370	Cause damage to organs	H351	Suspected of causing cancer
P271	Use in ventilated area	P280	Use gloves, eye protection/face shield
P302,332	If on skin, wash with soap and water	P305,351,338	If in eyes, remove contacts, rinse with water



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))			% (optional)
Methanol	METHYL ALCOHOL	CAS#: 67-56-1	> 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance. Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Flammability	Flammable in the presence of a source of ignition when the temperature is above the flash point. Keep away from heat/sparks/open flame/hot surface. No smoking.
Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist.
Storage Conditions	Use ventilation. Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge. Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Methanol	67-56-1 TWA 200 ppm
Skin notation	TWA 200 ppm
Potential for skin absorption, ingestion and inhalation.	
Personal protective equipment	Respiratory protection. Handle with gloves. Gloves must be inspected prior to use. Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.	

Section IX - Physical/Chemical Characteristics

Boiling Point	65°C	Specific Gravity (H ₂ O = 1)	0.79
Vapor Pressure (mm Hg)	96	Melting Point	-98°C
Vapor Density (AIR = 1)	1.11	Evaporation rate (Butyl Acetate = 1)	4.6
Solubility in Water	COMPLETE		
Appearance and Odor	CLEAR, COLORLESS LIQUID WITH CHARACTERISTIC PUNGENT ODOR.		

Section X. STABILITY AND REACTIVITY

Chemical stability	Stable under recommended storage conditions.
Possibility of hazardous reactions	Vapours may form explosive mixture with air.
Conditions to avoid	Heat, flames, sparks, extreme temperature and sunlight.
Materials to avoid	Acid chlorides, Acid anhydrides, Oxidizing agents, Alkali metals, Reducing agents, Acids
Hazardous decomposition products formed under fire conditions.	- Carbon oxides

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - rat - 5,628 mg/kg
LC50 Inhalation - rat - 4 h - 64000 ppm
LD50 Dermal - rabbit - 15,800 mg/kg
Toxic if absorbed through skin. Causes skin irritation.
Eye damage/eye irritation
Toxic if inhaled. Causes respiratory tract irritation.
Toxic if swallowed.

Section XII. ECOLOGICAL INFORMATION FOR REPORTABLE QUANTITY OF 5000 lbs.

LC50 15,400 mg/l - 96 h
EC50 24,500.00 mg/l - 48 h
EC100 10,000.00 mg/l - 24 h

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)	IATA
UN number: 1230 Class: 3 Packing group: II	UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol	Proper shipping name: Methanol

Section XV. REGULATORY INFORMATION

OSHA Hazards Flammable liquid, Target Organ Effect, Toxic by inhalation., Toxic by ingestion, Toxic by skin absorption, Irritant
SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC. DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.



CERTIFIED WEIGHT REPORT


Part Number: 95319
Lot Number: 032922
Description: Revised Additions Mix
11 components
Expiration Date: 032925
Recommended Storage: Refrigerate (4 °C)
Nominal Concentration (µg/mL): Varied
NIST Test ID#: 6UTB

Solvent(s): Lot#
Methanol EC592-US

		032922
Formulated By:	Prashant Chauhan	DATE
		032922
Reviewed By:	Pedro L. Rentas	DATE

Weight(s) shown below were combined and diluted to (mL): 100.0 0.012 Balance Uncertainty
0.012 Flask Uncertainty

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. Acrylonitrile	7	4718CK	10000	99	0.2	1.01015	1.01030	10001.5	40.5	107-13-1	N/A	ori-rat 78 mg/kg
2. 1-Chlorobutane	1072	MKCM5711	2000	99.99	0.2	0.20003	0.20020	2001.7	8.1	109-69-3	N/A	ori-rat 2670mg/kg
3. Cyclohexane	1023	28930	2000	99	0.2	0.20203	0.20215	2001.2	8.2	110-82-7	300 ppm (1050mg/m3/8H)	ori-rat 12705mg/kg
4. Di-isopropyl ether (DIPE)	987	00412MX	2000	99	0.2	0.20203	0.20215	2001.2	8.2	108-20-3	500 ppm (2100mg/m3/8H)	ori-rat 8470mg/kg
5. 1,4-Dioxane	373	03853KE	40000	99	0.2	4.04060	4.04100	40004.0	161.9	123-91-1	25 ppm (90mg/m3/8H)(skin)	ori-mus 5700mg/kg
6. Hexachloroethane	199	12604HBV	2000	99	0.2	0.20203	0.20213	2001.0	8.2	67-72-1	1 ppm (10mg/m3/8H)(skin)	ori-gpg 4970mg/kg
7. Methylcyclohexane	1627	08046KN	2000	99	0.2	0.20203	0.20215	2001.2	8.2	108-87-2	N/A	N/A
8. Methyl tert-butyl ether (MTBE)	209	02197JJ	2000	99.8	0.2	0.20041	0.20055	2001.4	8.1	1634-04-4	N/A	ori-rat 4g/kg
9. Propionitrile	349	1395468	20000	99	0.2	2.02030	2.02045	20001.5	81.0	107-12-0	N/A	ori-rat 39mg/kg
10. Tetrahydrofuran	380	SHBH8330	10000	99.9	0.2	1.00105	1.00120	10001.5	40.1	109-99-9	20 ppm (590mg/m3/8H)	ori-rat 1650mg/kg
11. 1,2,3,4-Tetramethylbenzene	491	AP01	2000	93	0.2	0.21506	0.21520	2001.3	8.7	488-23-3	N/A	ori-rat 6408mg/kg



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Method GC6MSD-1: Column: Vocol (60m X 0.25mm ID X 1.5µm film thickness). Temp. 1 = 35°C (10min.), Temp. 2 = 200°C (8.75 min.), Rate = 4°C/min., Injector Temp. = 200°C, Detector Temp. = 220°C. Solvent Delay: 8 minutes. Analysis performed by Candice Warren.

Name	MSD RT (min.)
Methyl tert-butyl ether (MTBE)	13.56
Acrylonitrile	13.79
Di-isopropyl ether	15.44
Propionitrile	18.53
Tetrahydrofuran	20.17
Cyclohexane	20.58
1-Chlorobutane	20.83
Methylcyclohexane	24.84
1,4-Dioxane	26.84
Hexachloroethane	48.44
1,2,3,4-Tetramethylbenzene	51.62

Safety Data Sheet (SDS)

GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY ANALYTICAL STANDARD DISSOLVED IN METHANOL

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	1-800-535-5053
Address	44 Rosotto Dr. Hamden CT, 06514	Emergency Telephone International	1-352-323-3500
		Date Prepared/Revised	January 1, 2023

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

H225	Highly Flammable Liquid and Vapor	H301, 311, 331	Toxic if swallowed, skin contact, inhaled
H370	Cause damage to organs	H351	Suspected of causing cancer
P271	Use in ventilated area	P280	Use gloves, eye protection/face shield
P302,332	If on skin, wash with soap and water	P305,351,338	If in eyes, remove contacts, rinse with water



Signal Word: DANGER

Section III - Composition

Components:	CAS#:	LD50 Oral - Rat	OSHA PEL	% (optional)
Methanol	67-56-1	2,769 mg/kg	200 ppm	> 99

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance. Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Flammability	Flammable in the presence of a source of ignition when the temperature is above the flash point. Keep away from heat/sparks/open flame/hot surface. No smoking.
Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist.
	Use ventilation. Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge.
Storage Conditions	Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Methanol	67-56-1 TWA 200 ppm
Skin notation	TWA 200 ppm
Potential for skin absorption, ingestion and inhalation.	
Personal protective equipment	Respiratory protection. Handle with gloves. Gloves must be inspected prior to use. Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.	

Section IX - Physical/Chemical Characteristics

Boiling Point	65°C	Specific Gravity (H ₂ O = 1)	0.79
Vapor Pressure (mm Hg)	96	Melting Point	-98°C
Vapor Density (AIR = 1)	1.11	Evaporation rate (Butyl Acetate = 1)	4.6
Solubility in Water	COMPLETE		
Appearance and Odor	CLEAR, COLORLESS LIQUID WITH CHARACTERISTIC PUNGENT ODOR.		

Section X. STABILITY AND REACTIVITY

Chemical stability	Stable under recommended storage conditions.
Possibility of hazardous reactions	Vapours may form explosive mixture with air.
Conditions to avoid	Heat, flames, sparks, extreme temperature and sunlight.
Materials to avoid	Acid chlorides, Acid anhydrides, Oxidizing agents, Alkali metals, Reducing agents, Acids
Hazardous decomposition products formed under fire conditions.	- Carbon oxides

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - rat - 5,628 mg/kg
LC50 Inhalation - rat - 4 h - 64000 ppm
LD50 Dermal - rabbit - 15,800 mg/kg
Toxic if absorbed through skin. Causes skin irritation.
Eye damage/eye irritation
Toxic if inhaled. Causes respiratory tract irritation.
Toxic if swallowed.

Section XII. ECOLOGICAL INFORMATION FOR REPORTABLE QUANTITY OF 5000 lbs.

LC50 15,400 mg/l - 96 h
EC50 24,500.00 mg/l - 48 h
EC100 10,000.00 mg/l - 24 h

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)	IATA
UN number: 1230 Class: 3 Packing group: II	UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol	Proper shipping name: Methanol

Section XV. REGULATORY INFORMATION

OSHA Hazards Flammable liquid, Target Organ Effect, Toxic by inhalation., Toxic by ingestion, Toxic by skin absorption, Irritant
SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.



CERTIFIED WEIGHT REPORT


Part Number: **95319**
Lot Number: **032922**
Description: **Revised Additions Mix**
11 components

Solvent(s): **Lot#**
Methanol EC592-US

Expiration Date: **032925**Recommended Storage: **Refrigerate (4 °C)**Nominal Concentration (µg/mL): **Varied**NIST Test ID#: **6UTB****5E-05 Balance Uncertainty**Weight(s) shown below were combined and diluted to (mL): **100.0 0.012 Flask Uncertainty**

Prashant Chauhan **032922**
Formulated By: **Prashant Chauhan** DATE
Pedro L. Rentas **032922**
Reviewed By: **Pedro L. Rentas** DATE

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. Acrylonitrile	7	4718CK	10000	99	0.2	1.01015	1.01030	10001.5	40.5	107-13-1	N/A	ori-rat 78 mg/kg
2. 1-Chlorobutane	1072	MKCM5711	2000	99.99	0.2	0.20003	0.20020	2001.7	8.1	109-69-3	N/A	ori-rat 2670mg/kg
3. Cyclohexane	1023	28930	2000	99	0.2	0.20203	0.20215	2001.2	8.2	110-82-7	300 ppm (1050mg/m3/8H)	ori-rat 12705mg/kg
4. Di-isopropyl ether (DIPE)	987	00412MX	2000	99	0.2	0.20203	0.20215	2001.2	8.2	108-20-3	500 ppm (2100mg/m3/8H)	ori-rat 8470mg/kg
5. 1,4-Dioxane	373	03853KE	40000	99	0.2	4.04060	4.04100	40004.0	161.9	123-91-1	25 ppm (90mg/m3/8H)(skin)	ori-mus 5700mg/kg
6. Hexachloroethane	199	12604HBV	2000	99	0.2	0.20203	0.20213	2001.0	8.2	67-72-1	1 ppm (10mg/m3/8H)(skin)	ori-gpg 4970mg/kg
7. Methylcyclohexane	1627	08046KN	2000	99	0.2	0.20203	0.20215	2001.2	8.2	108-87-2	N/A	N/A
8. Methyl tert-butyl ether (MTBE)	209	02197JJ	2000	99.8	0.2	0.20041	0.20055	2001.4	8.1	1634-04-4	N/A	ori-rat 4g/kg
9. Propionitrile	349	1395468	20000	99	0.2	2.02030	2.02045	20001.5	81.0	107-12-0	N/A	ori-rat 39mg/kg
10. Tetrahydrofuran	380	SHBH8330	10000	99.9	0.2	1.00105	1.00120	10001.5	40.1	109-99-9	20 ppm (590mg/m3/8H)	ori-rat 1650mg/kg
11. 1,2,3,4-Tetramethylbenzene	491	AP01	2000	93	0.2	0.21506	0.21520	2001.3	8.7	488-23-3	N/A	ori-rat 6408mg/kg



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Name	MSD RT (min.)
Methyl tert-butyl ether (MTBE)	13.56
Acrylonitrile	13.79
Di-isopropyl ether	15.44
Propionitrile	18.53
Tetrahydrofuran	20.17
Cyclohexane	20.58
1-Chlorobutane	20.83
Methylcyclohexane	24.84
1,4-Dioxane	26.84
Hexachloroethane	48.44
1,2,3,4-Tetramethylbenzene	51.62

Safety Data Sheet (SDS)

GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY ANALYTICAL STANDARD DISSOLVED IN METHANOL

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	1-800-535-5053
Address	44 Rosotto Dr. Hamden CT, 06514	Emergency Telephone International	1-352-323-3500
		Date Prepared/Revised	January 1, 2023

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

H225	Highly Flammable Liquid and Vapor	H301, 311, 331	Toxic if swallowed, skin contact, inhaled
H370	Cause damage to organs	H351	Suspected of causing cancer
P271	Use in ventilated area	P280	Use gloves, eye protection/face shield
P302,332	If on skin, wash with soap and water	P305,351,338	If in eyes, remove contacts, rinse with water



Signal Word: DANGER

Section III - Composition

Components:	CAS#:	LD50 Oral - Rat	OSHA PEL	% (optional)
Methanol	67-56-1	2,769 mg/kg	200 ppm	> 99

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance. Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Flammability	Flammable in the presence of a source of ignition when the temperature is above the flash point. Keep away from heat/sparks/open flame/hot surface. No smoking.
Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist.
	Use ventilation. Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge.
Storage Conditions	Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Methanol	67-56-1 TWA 200 ppm
Skin notation	TWA 200 ppm
Potential for skin absorption, ingestion and inhalation.	
Personal protective equipment	Respiratory protection Handle with gloves. Gloves must be inspected prior to use. Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.	

Section IX - Physical/Chemical Characteristics

Boiling Point	65°C	Specific Gravity (H ₂ O = 1)	0.79
Vapor Pressure (mm Hg)	96	Melting Point	-98°C
Vapor Density (AIR = 1)	1.11	Evaporation rate (Butyl Acetate = 1)	4.6
Solubility in Water	COMPLETE		
Appearance and Odor	CLEAR, COLORLESS LIQUID WITH CHARACTERISTIC PUNGENT ODOR.		

Section X. STABILITY AND REACTIVITY

Chemical stability	Stable under recommended storage conditions.
Possibility of hazardous reactions	Vapours may form explosive mixture with air.
Conditions to avoid	Heat, flames, sparks, extreme temperature and sunlight.
Materials to avoid	Acid chlorides, Acid anhydrides, Oxidizing agents, Alkali metals, Reducing agents, Acids
Hazardous decomposition products formed under fire conditions.	- Carbon oxides

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - rat - 5,628 mg/kg
LC50 Inhalation - rat - 4 h - 64000 ppm
LD50 Dermal - rabbit - 15,800 mg/kg
Toxic if absorbed through skin. Causes skin irritation.
Eye damage/eye irritation
Toxic if inhaled. Causes respiratory tract irritation.
Toxic if swallowed.

Section XII. ECOLOGICAL INFORMATION FOR REPORTABLE QUANTITY OF 5000 lbs.

LC50 15,400 mg/l - 96 h
EC50 24,500.00 mg/l - 48 h
EC100 10,000.00 mg/l - 24 h

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)	IATA
UN number: 1230 Class: 3 Packing group: II	UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol	Proper shipping name: Methanol

Section XV. REGULATORY INFORMATION

OSHA Hazards Flammable liquid, Target Organ Effect, Toxic by inhalation., Toxic by ingestion, Toxic by skin absorption, Irritant
SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.



CERTIFIED WEIGHT REPORT

Part Number: 91980
Lot Number: 072423
Description: Acrolein

Solvent(s):
Water
Lot# 102422Q

5 Smu



Gabriel Holland		072423
Formulated By:	Gabriel Holland	DATE
Pedro L. Rentas		072423
Reviewed By:	Pedro L. Rentas	DATE

Expiration Date: 082423
Recommended Storage: Refrigerate (4 °C)
Nominal Concentration (µg/mL): 5000
NIST Test ID#: 6UTB
Weight(s) shown below were combined and diluted to (mL): 10.0

5E-05 Balance Uncertainty
0.001 Flask Uncertainty

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. Acrolein	5	103755R09M	5000	97.1	0.5	0.05160	0.05170	5009.2	52.5	107-02-8	0.1 ppm	ori-rat 46mg/kg

Method: GC6MSD-1. Detector: Mass Selective Detector (Scan mode). Column: Vocol (60m X 0.25mm ID X 1.5µm film thickness). Oven Profile: Temp. 1 = 35°C (Time 1 = 10min.), Temp. 2=200°C (Time 2 = 8.75 min.)
Rate = 4°C/min., Injector Temp. = 200°C, Detector Temp. = 220°C. Analyst: Pedro Rentas. NOTE: Due to the instability of acrolein in solution, all solutions of acrolein, and any dilutions thereof, should be used immediately
Long term storage is not recommended. Please contact our technical department if further information is required.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Safety Data Sheet (SDS) GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY ANALYTICAL STANDARD DISSOLVED IN WATER

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	1-800-535-5053
Address	44 Rossotto Dr.	Emergency Telephone International	1-352-323-3500
	Hamden CT, 06514	Date Prepared/Revised	May 1, 2022

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

P271	Use in ventilated area	H315	Causes skin and eye irritation.
P302,332	If on skin, wash with soap and water	P280	Use gloves, eye protection/face shield
		P305,351,338	If in eyes, remove contacts, rinse with water



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))
Water

CAS#: 7732-18-5

% (optional)
> 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance. Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.
Hazardous Decomposition products	Carbon oxides

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist.
Storage Conditions	Use ventilation. Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge. Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Water	CAS#: 7732-18-5	TWA: 500 ppm
Personal protective equipment Respiratory protection Handle with gloves. Gloves must be inspected prior to use. Eye protection.		
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.		

Section IX - PHYSICAL/CHEMICAL CHARACTERISTICS

Boiling Point	100°C	Specific Gravity (H ₂ O = 1)	1
Vapor Pressure (mm Hg)		Melting Point	

Vapor Density (AIR = 1)		NA	Evaporation rate (Butyl Acetate = 1)	0°C
Solubility in Water		Completely miscible	NA	
Appearance and Odor		CLEAR, COLORLESS LIQUID WITH SLIGHT CHEMICAL ODOR.		

Section X. STABILITY AND REACTIVITY

Chemical stability	Stable under recommended storage conditions.
Possibility of hazardous reactions	NA
Conditions to avoid	NA
Materials to avoid	NA
Hazardous decomposition products - No data available	

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - Rat	NA
LC50 Inhalation - Rat	NA
LD50 Dermal - Guinea pig	NA
Causes skin irritation.	
Eye irritation	

Section XII. ECOLOGICAL INFORMATION

LC50	NA
EC50	NA

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)	IATA
Not dangerous goods	Not dangerous goods
Proper shipping name: Water	Proper shipping name: Water

Section XV. REGULATORY INFORMATION

SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Material Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.



CERTIFIED WEIGHT REPORT

Part Number: 91980
Lot Number: 072423
Description: Acrolein

Solvent(s):
Water
Lot# 102422Q

Expiration Date: 082423

Recommended Storage: Refrigerate (4 °C)

Nominal Concentration (µg/mL): 5000


NIST Test ID#: 6UTB

Weight(s) shown below were combined and diluted to (mL): 10.0
5E-05 Balance Uncertainty
0.001 Flask Uncertainty

<i>Gabriel Holland</i>		072423
Formulated By:	Gabriel Holland	DATE
<i>Pedro L. Rentas</i>		072423
Reviewed By:	Pedro L. Rentas	DATE

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. Acrolein	5	103755R09M	5000	97.1	0.5	0.05160	0.05170	5009.2	52.5	107-02-8	0.1 ppm	ori-rat 46mg/kg

Method: GC6MSD-1. Detector: Mass Selective Detector (Scan mode). Column: Vocol (60m X 0.25mm ID X 1.5µm film thickness). Oven Profile: Temp. 1 = 35°C (Time 1 = 10min.), Temp. 2=200°C (Time 2 = 8.75 min.)
Rate = 4°C/min., Injector Temp. = 200°C, Detector Temp. = 220°C. Analyst: Pedro Rentas. NOTE: Due to the instability of acrolein in solution, all solutions of acrolein, and any dilutions thereof, should be used immediately
Long term storage is not recommended. Please contact our technical department if further information is required.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Safety Data Sheet (SDS) GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY ANALYTICAL STANDARD DISSOLVED IN WATER

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	1-800-535-5053
Address	44 Rossotto Dr.	Emergency Telephone International	1-352-323-3500
	Hamden CT, 06514	Date Prepared/Revised	May 1, 2022

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

P271	Use in ventilated area	H315	Causes skin and eye irritation.
P302,332	If on skin, wash with soap and water	P280	Use gloves, eye protection/face shield
		P305,351,338	If in eyes, remove contacts, rinse with water



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))
Water

CAS#: 7732-18-5

% (optional)
> 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance. Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.
Hazardous Decomposition products	Carbon oxides

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist.
Storage Conditions	Use ventilation. Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge. Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Water	CAS#: 7732-18-5	TWA: 500 ppm
Personal protective equipment Respiratory protection Handle with gloves. Gloves must be inspected prior to use. Eye protection.		
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.		

Section IX - PHYSICAL/CHEMICAL CHARACTERISTICS

Boiling Point	100°C	Specific Gravity (H ₂ O = 1)	1
Vapor Pressure (mm Hg)		Melting Point	

		NA		0°C
Vapor Density (AIR = 1)		NA	Evaporation rate (Butyl Acetate = 1)	NA
Solubility in Water	Completely miscible			
Appearance and Odor	CLEAR, COLORLESS LIQUID WITH SLIGHT CHEMICAL ODOR.			

Section X. STABILITY AND REACTIVITY

Chemical stability	Stable under recommended storage conditions.
Possibility of hazardous reactions	NA
Conditions to avoid	NA
Materials to avoid	NA
Hazardous decomposition products - No data available	

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - Rat	NA
LC50 Inhalation - Rat	NA
LD50 Dermal - Guinea pig	NA
Causes skin irritation.	
Eye irritation	

Section XII. ECOLOGICAL INFORMATION

LC50	NA
EC50	NA

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)	IATA
Not dangerous goods	Not dangerous goods
Proper shipping name: Water	Proper shipping name: Water

Section XV. REGULATORY INFORMATION

SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Material Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.



CERTIFIED WEIGHT REPORT

Part Number: 91980
Lot Number: 072423
Description: Acrolein

Solvent(s):
Water
Lot# 102422Q

Expiration Date: 082423

Recommended Storage: Refrigerate (4 °C)

Nominal Concentration (µg/mL): 5000



NIST Test ID#: 6UTB

Weight(s) shown below were combined and diluted to (mL): 10.0
5E-05 Balance Uncertainty
0.001 Flask Uncertainty

<i>Gabriel Holland</i>		072423
Formulated By:	Gabriel Holland	DATE
<i>Pedro L. Rentas</i>		072423
Reviewed By:	Pedro L. Rentas	DATE

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. Acrolein	5	103755R09M	5000	97.1	0.5	0.05160	0.05170	5009.2	52.5	107-02-8	0.1 ppm	ori-rat 46mg/kg

Method: GC6MSD-1. Detector: Mass Selective Detector (Scan mode). Column: Vocol (60m X 0.25mm ID X 1.5µm film thickness). Oven Profile: Temp. 1 = 35°C (Time 1 = 10min.), Temp. 2=200°C (Time 2 = 8.75 min.)
Rate = 4°C/min., Injector Temp. = 200°C, Detector Temp. = 220°C. Analyst: Pedro Rentas. NOTE: Due to the instability of acrolein in solution, all solutions of acrolein, and any dilutions thereof, should be used immediately
Long term storage is not recommended. Please contact our technical department if further information is required.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Safety Data Sheet (SDS) GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY ANALYTICAL STANDARD DISSOLVED IN WATER

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	1-800-535-5053
Address	44 Rossotto Dr.	Emergency Telephone International	1-352-323-3500
	Hamden CT, 06514	Date Prepared/Revised	May 1, 2022

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

P271	Use in ventilated area	H315	Causes skin and eye irritation.
P302,332	If on skin, wash with soap and water	P280	Use gloves, eye protection/face shield
		P305,351,338	If in eyes, remove contacts, rinse with water



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))
Water

CAS#: 7732-18-5

% (optional)
> 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance. Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.
Hazardous Decomposition products	Carbon oxides

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist.
Storage Conditions	Use ventilation. Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge. Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Water	CAS#: 7732-18-5	TWA: 500 ppm
Personal protective equipment Respiratory protection Handle with gloves. Gloves must be inspected prior to use. Eye protection.		
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.		

Section IX - PHYSICAL/CHEMICAL CHARACTERISTICS

Boiling Point	100°C	Specific Gravity (H ₂ O = 1)	1
Vapor Pressure (mm Hg)		Melting Point	

		NA		0°C
Vapor Density (AIR = 1)		NA	Evaporation rate (Butyl Acetate = 1)	NA
Solubility in Water	Completely miscible			
Appearance and Odor	CLEAR, COLORLESS LIQUID WITH SLIGHT CHEMICAL ODOR.			

Section X. STABILITY AND REACTIVITY

Chemical stability	Stable under recommended storage conditions.
Possibility of hazardous reactions	NA
Conditions to avoid	NA
Materials to avoid	NA
Hazardous decomposition products - No data available	

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - Rat	NA
LC50 Inhalation - Rat	NA
LD50 Dermal - Guinea pig	NA
Causes skin irritation.	
Eye irritation	

Section XII. ECOLOGICAL INFORMATION

LC50	NA
EC50	NA

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)	IATA
Not dangerous goods	Not dangerous goods
Proper shipping name: Water	Proper shipping name: Water

Section XV. REGULATORY INFORMATION

SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Material Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.



CERTIFIED WEIGHT REPORT

Part Number: 91980
Lot Number: 072423
Description: Acrolein

Solvent(s):
Water
Lot# 102422Q

Expiration Date: 082423

Recommended Storage: Refrigerate (4 °C)

Nominal Concentration (µg/mL): 5000

NIST Test ID#: 6UTB

Weight(s) shown below were combined and diluted to (mL): 10.0


5E-05 Balance Uncertainty

0.001 Flask Uncertainty

<i>Gabriel Holland</i>		072423
Formulated By:	Gabriel Holland	DATE
<i>Pedro L. Rentas</i>		072423
Reviewed By:	Pedro L. Rentas	DATE

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. Acrolein	5	103755R09M	5000	97.1	0.5	0.05160	0.05170	5009.2	52.5	107-02-8	0.1 ppm	ori-rat 46mg/kg

Method: GC6MSD-1. Detector: Mass Selective Detector (Scan mode). Column: Vocol (60m X 0.25mm ID X 1.5µm film thickness). Oven Profile: Temp. 1 = 35°C (Time 1 = 10min.), Temp. 2=200°C (Time 2 = 8.75 min.) Rate = 4°C/min., Injector Temp. = 200°C, Detector Temp. = 220°C. Analyst: Pedro Rentas. NOTE: Due to the instability of acrolein in solution, all solutions of acrolein, and any dilutions thereof, should be used immediately. Long term storage is not recommended. Please contact our technical department if further information is required.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Safety Data Sheet (SDS) GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY ANALYTICAL STANDARD DISSOLVED IN WATER

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	1-800-535-5053
Address	44 Rossotto Dr.	Emergency Telephone International	1-352-323-3500
	Hamden CT, 06514	Date Prepared/Revised	May 1, 2022

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

P271	Use in ventilated area	H315	Causes skin and eye irritation.
P302,332	If on skin, wash with soap and water	P280	Use gloves, eye protection/face shield
		P305,351,338	If in eyes, remove contacts, rinse with water



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))
Water

CAS#: 7732-18-5

% (optional)
> 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance. Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.
Hazardous Decomposition products	Carbon oxides

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist.
Storage Conditions	Use ventilation. Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge. Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Water	CAS#: 7732-18-5	TWA: 500 ppm
Personal protective equipment	Respiratory protection	Handle with gloves. Gloves must be inspected prior to use.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.		Eye protection.

Section IX - PHYSICAL/CHEMICAL CHARACTERISTICS

Boiling Point	100°C	Specific Gravity (H ₂ O = 1)	1
Vapor Pressure (mm Hg)		Melting Point	

		NA		0°C
Vapor Density (AIR = 1)		NA	Evaporation rate (Butyl Acetate = 1)	NA
Solubility in Water	Completely miscible			
Appearance and Odor	CLEAR, COLORLESS LIQUID WITH SLIGHT CHEMICAL ODOR.			

Section X. STABILITY AND REACTIVITY

Chemical stability	Stable under recommended storage conditions.
Possibility of hazardous reactions	NA
Conditions to avoid	NA
Materials to avoid	NA
Hazardous decomposition products - No data available	

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - Rat	NA
LC50 Inhalation - Rat	NA
LD50 Dermal - Guinea pig	NA
Causes skin irritation.	
Eye irritation	

Section XII. ECOLOGICAL INFORMATION

LC50	NA
EC50	NA

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION



DOT (US)	IATA
Not dangerous goods	Not dangerous goods
Proper shipping name: Water	Proper shipping name: Water

Section XV. REGULATORY INFORMATION

SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Material Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.



CERTIFIED WEIGHT REPORT

Part Number: **95318**
Lot Number: **121321**
Description: **2-Chloroethyl vinyl ether**




Solvent(s): **Methanol**
Lot#: **EA899-US**

		121321
Formulated By:	Benson Chan	DATE
		121321
Reviewed By:	Pedro L. Rentas	DATE



Expiration Date: **121324**
Recommended Storage: **Refrigerate (4 °C)**
Nominal Concentration (µg/mL): **10000**
NIST Test ID#: **6UTB**
Weight(s) shown below were combined and diluted to (mL): **30.0**
5E-05 Balance Uncertainty
0.0003 Flask Uncertainty

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight (g)	Actual Weight (g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. 2-Chloroethyl vinyl ether	74	MKCD0033	10000	99	0.2	0.30320	0.30411	10030.2	40.7	110-75-8	N/A	ori-rat 250mg/kg

Method: GC6MSD-1.M. Detector: MSD. Column: (60m X 0.25mm X 1.5 µm). Oven Profile: Temp 1 = 35°C (Time 1=10min.), Temp 2 = 200°C (Time 2=8.75 min.), Rate = 4°C/min., Injector B Temp = 200°C, Detector B Temp. = 220°C. Analyst: Candice Warren.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).



CERTIFIED WEIGHT REPORT

Part Number: **95318**
Lot Number: **121321**
Description: **2-Chloroethyl vinyl ether**

Solvent(s): **Methanol**
Lot#: **EA899-US**




Expiration Date: **121324**
Recommended Storage: **Refrigerate (4 °C)**
Nominal Concentration (µg/mL): **10000**
NIST Test ID#: **6UTB**

Weight(s) shown below were combined and diluted to (mL): **30.0**
5E-05 Balance Uncertainty
0.0003 Flask Uncertainty

		121321
Formulated By:	Benson Chan	DATE
		121321
Reviewed By:	Pedro L. Rentas	DATE

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight (g)	Actual Weight (g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. 2-Chloroethyl vinyl ether	74	MKCD0033	10000	99	0.2	0.30320	0.30411	10030.2	40.7	110-75-8	N/A	ori-rat 250mg/kg

Method: GC6MSD-1.M. Detector: MSD. Column: (60m X 0.25mm X 1.5 µm). Oven Profile: Temp 1 = 35°C (Time 1=10min.), Temp 2 = 200°C (Time 2=8.75 min.), Rate = 4°C/min., Injector B Temp = 200°C, Detector B Temp. = 220°C. Analyst: Candice Warren.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30470 Lot No.: A0181905
Description : tert-Butanol Standard
tert-Butanol Std 50,000µg/mL, P&T Methanol, 1mL/ampul
Container Size : 2 mL Pkg Amt: > 1 mL
Expiration Date : February 28, 2025 Storage: 0°C or colder
Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)			
1	tert-Butanol (TBA) CAS # 75-65-0 Purity 99% (Lot SHBM7694)	50,126.0 µg/mL	+/-	293.4988	µg/mL	Gravimetric
			+/-	1,073.7654	µg/mL	Unstressed
			+/-	1,104.9494	µg/mL	Stressed

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:


200°C

Det. Temp:

250°C

Det. Type:

FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

John Friedline - Operations Technician I

Date Mixed: 16-Feb-2022

Balance: B442140311

Marlene Cowan - Operations Tech I

Date Passed: 21-Feb-2022

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified combined stressed uncertainty value (includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ stressed} = k \sqrt{U_{gravimetric}^2 + U_{homogeneity}^2 + U_{storage\ stability}^2 + U_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- It is important to note that the shipping stability uncertainty was obtained under temperature extremes for specific time intervals; therefore, the certified combined stressed uncertainty value should only be applied to the product if it was stored at non-standard temperature conditions up to and including 7 days. Contact Restek Technical Service at www.restek.com/Contact-Us for use recommendations if your shipment was in-transit for more than 7 days at non-standard temperature conditions.
- Apply the certified combined unstressed uncertainty value if the product was received under standard shipping conditions. Apply the certified combined stressed uncertainty value if the product was received under non-standard conditions as specified below.

Label Conditions	Standard Conditions	Non-Standard Conditions
25°C Nominal (Room Temperature)	< 60°C	≥ 60°C up to 7 days
10°C or colder (Refrigerate)	< 40°C	≥ 40°C up to 7 days
0°C or colder (Freezer) -20°C or colder (Deep Freezer)	< 25°C	≥ 25°C up to 7 days

- Separate (not combined) uncertainty values for gravimetric uncertainty are also displayed on the certificate, if needed, separate homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty values are available by contacting Restek Technical Service at www.restek.com/Contact-Us.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampules are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampules. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30067 **Lot No.:** A0191805

Description : 4-Bromofluorobenzene Standard

4-Bromofluorobenzene Standard 2,500µg/mL, P&T Methanol,
1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : November 30, 2027 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1-Bromo-4-fluorobenzene (BFB)	460-00-4	184975	99%	2,483.9 µg/mL	+/- 139.5488

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:


40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Alicia Leathers - Operation Technician I

Date Mixed: 17-Nov-2022

Balance Serial # B251644995


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 21-Nov-2022

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30225 **Lot No.:** A0193071

Description : Bromochloromethane Standard
Bromochloromethane 2000µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : December 31, 2027 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Bromochloromethane	74-97-5	00008541	99%	2,018.0 µg/mL	+/- 113.3890

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Tom Suckar - Mix Technician

Date Mixed: 29-Dec-2022 Balance Serial # B707717271


Christie Mills - Operations Tech II - ARM QC

Date Passed: 03-Jan-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30225 **Lot No.:** A0193071

Description : Bromochloromethane Standard
Bromochloromethane 2000µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : December 31, 2027 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Bromochloromethane	74-97-5	00008541	99%	2,018.0 µg/mL	+/- 113.3890

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Tom Suckar - Mix Technician

Date Mixed: 29-Dec-2022 Balance Serial # B707717271


Christie Mills - Operations Tech II - ARM QC

Date Passed: 03-Jan-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30042 **Lot No.:** A0194279

Description : 502.2 Calibration Mix #1
502.2 Calibration Mix #1 2,000µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : October 31, 2029 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Dichlorodifluoromethane (CFC-12)	75-71-8	00012554	99%	2,001.5 µg/mL	+/- 112.7231
2	Chloromethane (methyl chloride)	74-87-3	SHBK6571	99%	2,001.2 µg/mL	+/- 112.5863
3	Vinyl chloride	75-01-4	00015559	99%	2,001.4 µg/mL	+/- 112.6561
4	Bromomethane (methyl bromide)	74-83-9	101604	99%	2,006.4 µg/mL	+/- 112.8262
5	Chloroethane (ethyl chloride)	75-00-3	107-401039114-1	99%	2,001.9 µg/mL	+/- 112.5897
6	Trichlorofluoromethane (CFC-11)	75-69-4	MKCL8411	99%	2,000.8 µg/mL	+/- 112.6473

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:

60m x 0.25mm x 1.4µm
Rtx-502.2 (cat.#10916)

Carrier Gas:

helium-constant flow 2.0 mL/min.

Temp. Program:

40°C (hold 6 min.) to 100°C
@ 6°C/min.

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


MSD

Split Vent:

Split ratio 10:1

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Tom Suckar - Mix Technician

Date Mixed: 03-Feb-2023

Balance Serial # B707717271


Christie Mills - Operations Tech II - ARM QC

Date Passed: 07-Feb-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/ μ ECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30042 **Lot No.:** A0194279

Description : 502.2 Calibration Mix #1
502.2 Calibration Mix #1 2,000µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : October 31, 2029 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Dichlorodifluoromethane (CFC-12)	75-71-8	00012554	99%	2,001.5 µg/mL	+/- 112.7231
2	Chloromethane (methyl chloride)	74-87-3	SHBK6571	99%	2,001.2 µg/mL	+/- 112.5863
3	Vinyl chloride	75-01-4	00015559	99%	2,001.4 µg/mL	+/- 112.6561
4	Bromomethane (methyl bromide)	74-83-9	101604	99%	2,006.4 µg/mL	+/- 112.8262
5	Chloroethane (ethyl chloride)	75-00-3	107-401039114-1	99%	2,001.9 µg/mL	+/- 112.5897
6	Trichlorofluoromethane (CFC-11)	75-69-4	MKCL8411	99%	2,000.8 µg/mL	+/- 112.6473

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:

60m x 0.25mm x 1.4µm
Rtx-502.2 (cat.#10916)

Carrier Gas:

helium-constant flow 2.0 mL/min.

Temp. Program:

40°C (hold 6 min.) to 100°C
@ 6°C/min.

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


MSD

Split Vent:

Split ratio 10:1

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Tom Suckar - Mix Technician

Date Mixed: 03-Feb-2023

Balance Serial # B707717271


Christie Mills - Operations Tech II - ARM QC

Date Passed: 07-Feb-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/ μ ECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30489 **Lot No.:** A0196115

Description : 8260B Acetates Mix

8260B Acetates Mix 2,000 µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : September 30, 2024 **Storage:** -20°C or colder

Handling: This product is photosensitive. **Ship:** On Ice

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Methyl acetate	79-20-9	SHBP3100	99%	2,013.7 µg/mL	+/- 69.6015
2	Vinyl acetate	108-05-4	RD220630	99%	2,020.0 µg/mL	+/- 69.8205
3	Ethyl acetate	141-78-6	SHBP9289	99%	2,019.3 µg/mL	+/- 69.7974
4	Isopropyl acetate	108-21-4	BCCG7069	99%	2,014.0 µg/mL	+/- 69.6131
5	Propyl acetate	109-60-4	TFFKL	99%	2,014.7 µg/mL	+/- 69.6361
6	Butyl acetate	123-86-4	SHBP6314	99%	2,014.0 µg/mL	+/- 69.6131
7	Amyl acetate	628-63-7	41325/1	97%	2,016.3 µg/mL	+/- 69.6928

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Tech Tips:

Vinyl acetate is a volatile organic ester included in the target lists of several US EPA and other methods. Under acidic conditions, esters react with alcohols to form new esters (transesterification). Methanol-based mixes containing halogenated compounds are slightly acidic, so it is important to minimize exposure of vinyl acetate to mixes of halogenated compounds in methanol. For this

reason, we offer vinyl acetate in individual solution, and suggest that it be introduced into the working level calibration solution immediately before use. This will minimize problems and ensure more consistent results.

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Bethany Lowery - Operations Tech I

Date Mixed: 21-Mar-2023

Balance Serial # B251644995


John Lidgett - AD Chemist

Date Passed: 29-Mar-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30489 **Lot No.:** A0196115

Description : 8260B Acetates Mix

8260B Acetates Mix 2,000 µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : September 30, 2024 **Storage:** -20°C or colder

Handling: This product is photosensitive. **Ship:** On Ice

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Methyl acetate	79-20-9	SHBP3100	99%	2,013.7 µg/mL	+/- 69.6015
2	Vinyl acetate	108-05-4	RD220630	99%	2,020.0 µg/mL	+/- 69.8205
3	Ethyl acetate	141-78-6	SHBP9289	99%	2,019.3 µg/mL	+/- 69.7974
4	Isopropyl acetate	108-21-4	BCCG7069	99%	2,014.0 µg/mL	+/- 69.6131
5	Propyl acetate	109-60-4	TFFKL	99%	2,014.7 µg/mL	+/- 69.6361
6	Butyl acetate	123-86-4	SHBP6314	99%	2,014.0 µg/mL	+/- 69.6131
7	Amyl acetate	628-63-7	41325/1	97%	2,016.3 µg/mL	+/- 69.6928

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Tech Tips:

Vinyl acetate is a volatile organic ester included in the target lists of several US EPA and other methods. Under acidic conditions, esters react with alcohols to form new esters (transesterification). Methanol-based mixes containing halogenated compounds are slightly acidic, so it is important to minimize exposure of vinyl acetate to mixes of halogenated compounds in methanol. For this

reason, we offer vinyl acetate in individual solution, and suggest that it be introduced into the working level calibration solution immediately before use. This will minimize problems and ensure more consistent results.

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Bethany Lowery - Operations Tech I

Date Mixed: 21-Mar-2023

Balance Serial # B251644995


John Lidgett - AD Chemist

Date Passed: 29-Mar-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555582 **Lot No.:** A0196865

Description : Custom 8260A/B Surrogate Mix
Custom 8260A/B Surrogate Mix 25,000µg/mL, P&T Methanol,
1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2026 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,2-Dichloroethane-d4	17060-07-0	PR-32845	99%	25,036.0 µg/mL	+/- 1,417.9179
2	1-Bromo-4-fluorobenzene (BFB)	460-00-4	184975	99%	25,132.0 µg/mL	+/- 1,423.3549
3	Dibromofluoromethane	1868-53-7	022013	99%	25,040.0 µg/mL	+/- 1,418.1445
4	Toluene-d8	2037-26-5	PR-33397	99%	25,028.0 µg/mL	+/- 1,417.4648

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Russ Bookhamer - Operations Technician I

Date Mixed: 11-Apr-2023

Balance: 1127510105

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30006 **Lot No.:** A0200785

Description : VOA Calibration Mix #1

VOA Calibration Mix #1 5,000µg/mL, P&T Methanol/Water(90:10), 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : November 30, 2026 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Acetone	67-64-1	SHBP8774	99%	5,018.5 µg/mL	+/- 173.4162
2	2-Butanone (MEK)	78-93-3	SHBL5543	99%	5,016.0 µg/mL	+/- 173.3298
3	4-Methyl-2-pentanone (MIBK)	108-10-1	SHBP4724	99%	5,010.7 µg/mL	+/- 173.1455
4	2-Hexanone	591-78-6	MKCQ6663	99%	5,015.0 µg/mL	+/- 173.2952

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol/Water (90:10)
CAS # 67-56-1/7732-18-5
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:


40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Laith Clemente - Operations Technician I

Date Mixed: 09-Aug-2023

Balance Serial # B707717271


Marlina Cowan - Operations Tech II ARM QC

Date Passed: 16-Aug-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30006 **Lot No.:** A0200785

Description : VOA Calibration Mix #1

VOA Calibration Mix #1 5,000µg/mL, P&T Methanol/Water(90:10), 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : November 30, 2026 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Acetone	67-64-1	SHBP8774	99%	5,018.5 µg/mL	+/- 173.4162
2	2-Butanone (MEK)	78-93-3	SHBL5543	99%	5,016.0 µg/mL	+/- 173.3298
3	4-Methyl-2-pentanone (MIBK)	108-10-1	SHBP4724	99%	5,010.7 µg/mL	+/- 173.1455
4	2-Hexanone	591-78-6	MKCQ6663	99%	5,015.0 µg/mL	+/- 173.2952

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol/Water (90:10)
CAS # 67-56-1/7732-18-5
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:


40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Laith Clemente - Operations Technician I

Date Mixed: 09-Aug-2023

Balance Serial # B707717271


Marlina Cowan - Operations Tech II ARM QC

Date Passed: 16-Aug-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30006 **Lot No.:** A0200785

Description : VOA Calibration Mix #1

VOA Calibration Mix #1 5,000µg/mL, P&T Methanol/Water(90:10), 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : November 30, 2026 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Acetone	67-64-1	SHBP8774	99%	5,018.5 µg/mL	+/- 173.4162
2	2-Butanone (MEK)	78-93-3	SHBL5543	99%	5,016.0 µg/mL	+/- 173.3298
3	4-Methyl-2-pentanone (MIBK)	108-10-1	SHBP4724	99%	5,010.7 µg/mL	+/- 173.1455
4	2-Hexanone	591-78-6	MKCQ6663	99%	5,015.0 µg/mL	+/- 173.2952

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol/Water (90:10)

CAS # 67-56-1/7732-18-5

Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:


40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Laith Clemente - Operations Technician I

Date Mixed: 09-Aug-2023

Balance Serial # B707717271


Marlina Cowan - Operations Tech II ARM QC

Date Passed: 16-Aug-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis
chromatographic



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555408-SL Lot No.: A0205179
Description : Custom Vinyl Acetate Standard
Custom Vinyl Acetate Standard 8,000µg/mL, P&T Methanol, 1mL/ampul
Container Size : 2 mL Pkg Amt: > 1 mL
Expiration Date : June 30, 2025 Storage: -20°C or colder
Handling: This product is photosensitive. Ship: On Ice

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Vinyl acetate	108-05-4	RP231030CTH	98%	8,075.2 µg/mL	+/- 279.1159

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Tech Tips:

Vinyl acetate is a volatile organic ester included in the target lists of several US EPA and other methods. Under acidic conditions, esters react with alcohols to form new esters (transesterification). Methanol-based mixes containing halogenated compounds are slightly acidic, so it is important to minimize exposure of vinyl acetate to mixes of halogenated compounds in methanol. For this reason, we offer vinyl acetate in individual solution, and suggest that it be introduced into the working level calibration solution immediately before use. This will minimize problems and ensure more consistent results.

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Daniel Wasson - Operations Tech I

Date Mixed: 06-Dec-2023

Balance Serial # 1127510105

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 11-Dec-2023

APPROVED
(Signature Required for Log to File)

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

12
CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555408-SL **Lot No.:** A0205179
Description : Custom Vinyl Acetate Standard
Custom Vinyl Acetate Standard 8,000µg/mL, P&T Methanol, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : June 30, 2025 **Storage:** -20°C or colder
Handling: This product is photosensitive. **Ship:** On Ice

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Vinyl acetate	108-05-4	RP231030CTH	98%	8,075.2 µg/mL	+/- 279.1159

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Tech Tips:

Vinyl acetate is a volatile organic ester included in the target lists of several US EPA and other methods. Under acidic conditions, esters react with alcohols to form new esters (transesterification). Methanol-based mixes containing halogenated compounds are slightly acidic, so it is important to minimize exposure of vinyl acetate to mixes of halogenated compounds in methanol. For this reason, we offer vinyl acetate in individual solution, and suggest that it be introduced into the working level calibration solution immediately before use. This will minimize problems and ensure more consistent results.

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Daniel Wasson - Operations Tech I

Date Mixed: 06-Dec-2023

Balance Serial # 1127510105


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 11-Dec-2023

APPROVED
By: Jennifer Pollino (Date: 11-Dec-2023)

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/ μ ECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555581 **Lot No.:** A0210184

Description : Custom 8260 Internal Standard Mix

Custom 8260 Internal Standard Mix 25,000µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL


Expiration Date : April 30, 2027 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,4-Dichlorobenzene-d4	3855-82-1	PR-30447	99%	25,212.0 µg/mL	+/- 1,427.8857
2	1,4-Difluorobenzene	540-36-3	MKCS8657	99%	25,220.0 µg/mL	+/- 1,428.3388
3	Chlorobenzene-d5	3114-55-4	PR-31132	99%	25,116.0 µg/mL	+/- 1,422.4487
4	Pentafluorobenzene	363-72-4	MKCR9383	99%	25,180.0 µg/mL	+/- 1,426.0734

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%


John Friedline - Operations Technician I

Date Mixed: 11-Apr-2024

Balance: 1127510105

REVISED
By Restek Review of L&P, Apr 16, 2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

Methanol
ULTRA RESI-ANALYZED
For Purge and Trap Analysis



Material No.: 9077-02
Batch No.: 22L0562016
Manufactured Date: 2022-10-26
Expiration Date: 2025-10-25
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay (CH ₃ OH) (by GC, corrected for water)	≥ 99.9 %	100.0 %
Residue after Evaporation	≤ 1.0 ppm	0.2 ppm
Titration Acid (µeq/g)	≤ 0.3	0.2
Titration Base (µeq/g)	≤ 0.10	0.03
Water (by KF, coulometric)	≤ 0.08 %	< 0.01 %
Volatile Organic Trace Analysis - Below EPA 8260B CRQL	Conforms	Conforms

For Laboratory, Research, or Manufacturing Use
Performance Tested for Use in EPA Methods
500 Series for Drinking Water
600 Series for Wastewater
846 for Solid Waste

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

Jamie Ethier
Vice President Global Quality

Methanol
ULTRA RESI-ANALYZED
For Purge and Trap Analysis



Material No.: 9077-02
Batch No.: 22L0562016
Manufactured Date: 2022-10-26
Expiration Date: 2025-10-25
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay (CH ₃ OH) (by GC, corrected for water)	≥ 99.9 %	100.0 %
Residue after Evaporation	≤ 1.0 ppm	0.2 ppm
Titration Acid (μeq/g)	≤ 0.3	0.2
Titration Base (μeq/g)	≤ 0.10	0.03
Water (by KF, coulometric)	≤ 0.08 %	< 0.01 %
Volatile Organic Trace Analysis - Below EPA 8260B CRQL	Conforms	Conforms

For Laboratory, Research, or Manufacturing Use
Performance Tested for Use in EPA Methods
500 Series for Drinking Water
600 Series for Wastewater
846 for Solid Waste

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

Jamie Ethier
Vice President Global Quality

Methanol
ULTRA RESI-ANALYZED
For Purge and Trap Analysis



Material No.: 9077-02
Batch No.: 22L0562016
Manufactured Date: 2022-10-26
Expiration Date: 2025-10-25
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay (CH ₃ OH) (by GC, corrected for water)	≥ 99.9 %	100.0 %
Residue after Evaporation	≤ 1.0 ppm	0.2 ppm
Titration Acid (μeq/g)	≤ 0.3	0.2
Titration Base (μeq/g)	≤ 0.10	0.03
Water (by KF, coulometric)	≤ 0.08 %	< 0.01 %
Volatile Organic Trace Analysis - Below EPA 8260B CRQL	Conforms	Conforms

For Laboratory, Research, or Manufacturing Use
Performance Tested for Use in EPA Methods
500 Series for Drinking Water
600 Series for Wastewater
846 for Solid Waste

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

Jamie Ethier
Vice President Global Quality

Methanol
ULTRA RESI-ANALYZED
For Purge and Trap Analysis



Material No.: 9077-02
Batch No.: 22L0562016
Manufactured Date: 2022-10-26
Expiration Date: 2025-10-25
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay (CH ₃ OH) (by GC, corrected for water)	≥ 99.9 %	100.0 %
Residue after Evaporation	≤ 1.0 ppm	0.2 ppm
Titration Acid (μeq/g)	≤ 0.3	0.2
Titration Base (μeq/g)	≤ 0.10	0.03
Water (by KF, coulometric)	≤ 0.08 %	< 0.01 %
Volatile Organic Trace Analysis - Below EPA 8260B CRQL	Conforms	Conforms

For Laboratory, Research, or Manufacturing Use
Performance Tested for Use in EPA Methods
500 Series for Drinking Water
600 Series for Wastewater
846 for Solid Waste

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

Jamie Ethier
Vice President Global Quality



SHIPPING DOCUMENTS



CHAIN OF CUSTODY RECORD

284 Sheffield Street, Mountainside, NJ 07092
(908) 789-8900 • Fax (908) 789-8922
www.chemtech.net

CHEMTECH PROJECT NO.

P3657

QUOTE NO.

COC Number 2041316

CLIENT INFORMATION

REPORT TO BE SENT TO:

COMPANY: Jacobs
ADDRESS: 412 Mt Kable Ave Suite H100
CITY: Morrisstown STATE: NJ ZIP: 07960
ATTENTION: John Yufante
PHONE: (281) 414-1719 FAX:

CLIENT PROJECT INFORMATION

PROJECT NAME: STC PTC
PROJECT NO.: D3779922 LOCATION: Princeton Junction
PROJECT MANAGER: Mary Murphy
e-mail: Mary.Murphy@Jacobs.com
PHONE: (201) 936-0586 FAX:

CLIENT BILLING INFORMATION

BILL TO: Mary Murphy PO#:
ADDRESS:
CITY: STATE: ZIP:
ATTENTION: PHONE:

ANALYSIS

DATA TURNAROUND INFORMATION

FAX (RUSH) Standard TAT DAYS*
HARDCOPY (DATA PACKAGE): DAYS*
EDD: DAYS*

*TO BE APPROVED BY CHEMTECH

STANDARD HARDCOPY TURNAROUND TIME IS 10 BUSINESS DAYS

DATA DELIVERABLE INFORMATION

☐ Level 1 (Results Only) ☐ Level 4 (QC + Full Raw Data)
☐ Level 2 (Results + QC) ☐ NJ Reduced ☐ US EPA CLP
☒ Level 3 (Results + QC) ☐ NYS ASP A ☐ NYS ASP B
+ Raw Data ☐ Other
☐ EDD FORMAT

VOCs 8260D
SVOCs 8260E
PAHs 8270E
Metals 6020B, Hg
Cr 6010D, Pb 6010A
Cu 6010D, Zn 6010A

PRESERVATIVES

COMMENTS

CHEMTECH SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE		SAMPLE COLLECTION		# OF BOTTLES	PRESERVATIVES									COMMENTS
			COMP	GRAB	DATE	TIME		A/E	F	B/E	E						
1.	717-J-WS-081624	WS		X	8-16-24	0930	8	2	4	1	1						
2.	TB-01-081624	DI		X	8-16-24	1055	1	1									TB is unpreserved!
3.																	
4.																	
5.																	
6.																	
7.																	
8.																	
9.																	
10.																	

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER: 1. <u>[Signature]</u>	DATE/TIME: <u>8-16-24 1245</u>	RECEIVED BY: 1. <u>[Signature]</u>	Conditions of bottles or coolers at receipt: ☒ COMPLIANT ☐ NON COMPLIANT ☒ COOLER TEMP <u>2.9</u> °C
RELINQUISHED BY SAMPLER: 2.	DATE/TIME:	RECEIVED BY: 2.	Comments: <u>See attached table for his required analytes list</u>
RELINQUISHED BY SAMPLER: 3.	DATE/TIME:	RECEIVED BY: 3.	

Page 1 of 1

CLIENT: ☐ Hand Delivered ☐ Other
CHEMTECH: ☒ Picked Up ☐ Field Sampling

Shipment Complete
☒ YES ☐ NO

Table 3. Surface Water Target Analytes, Methods, Action Levels, and Control Limits
Site Sampling Plan for Ecological Evaluation
Princeton Technology Center, West Windsor Township, New Jersey

Method	Analyte	CAS Number	Units	Higher of PQL and Ground Water Quality Criterion ^a	Fresh Surface Water Chronic NJDEP Ecological Criterion ^b
ECO-SVOCs					
SW8270E	1,4-Dioxane	123-91-1	µg/L	0.4	--
SW8270E	1-Methylnaphthalene	90-12-0	µg/L	--	
SW8270E	2,4,5-Trichlorophenol	95-95-4	µg/L	700	
SW8270E	2,4,6-Trichlorophenol	88-06-2	µg/L	20	
SW8270E	2,4-Dinitrotoluene	121-14-2	µg/L	10	
SW8270E	2-Methylnaphthalene	91-57-6	µg/L	30	
SW8270E	2-Methylphenol	95-48-7	µg/L	--	
SW8270E	3 & 4-Methylphenol (m,p-Cresols)	65794-96-9	µg/L	--	
SW8270E	Acenaphthene	83-32-9	µg/L	400	
SW8270E	Acenaphthylene	208-96-8	µg/L	--	
SW8270E	Anthracene	120-12-7	µg/L	2000	
SW8270E	Benzaldehyde	100-52-7	µg/L	--	
SW8270E	Benzo(a)anthracene	56-55-3	µg/L	--	
SW8270E	Benzo(a)pyrene	50-32-8	µg/L	0.1	
SW8270E	Benzo(b)fluoranthene	205-99-2	µg/L	0.5	
SW8270E	Benzo(g,h,i)perylene	191-24-2	µg/L	--	
SW8270E	Benzo(k)fluoranthene	207-08-9	µg/L	0.5	
SW8270E	Bis (2-ethylhexyl) phthalate	117-81-7	µg/L	--	
SW8270E	Carbazole	86-74-8	µg/L	--	
SW8270E	Chrysene	218-01-9	µg/L	5	
SW8270E	Dibenzo(a,h)anthracene	53-70-3	µg/L	0.3	
SW8270E	Dibenzofuran	132-64-9	µg/L	--	
SW8270E	Di-N-Butylphthalate	84-74-2	µg/L	--	
SW8270E	Fluoranthene	206-44-0	µg/L	300	
SW8270E	Fluorene	86-73-7	µg/L	300	
SW8270E	Hexachlorobenzene	118-74-1	µg/L	0.02	
SW8270E	Hexachlorobutadiene	87-68-3	µg/L	1	
SW8270E	Hexachloroethane	67-72-1	µg/L	7	
SW8270E	Indeno(1,2,3-Cd)Pyrene	193-39-5	µg/L	0.2	
SW8270E	Naphthalene	91-20-3	µg/L	300	
SW8270E	Nitrobenzene	98-95-3	µg/L	6	
SW8270E	Pentachlorophenol	87-86-5	µg/L	0.3	
SW8270E	Phenanthrene	85-01-8	µg/L	--	
SW8270E	Pyrene	129-00-0	µg/L	200	
SW8270E	Pyridine	110-86-1	µg/L	--	
ECO-VOCs					
SW8260D	1,1,1-Trichloroethane	71-55-6	µg/L	30	76

Table 3. Surface Water Target Analytes, Methods, Action Levels, and Control Limits
Site Sampling Plan for Ecological Evaluation
Princeton Technology Center, West Windsor Township, New Jersey

Method	Analyte	CAS Number	Units	Higher of PQL and Ground Water Quality Criterion ^a	Fresh Surface Water Chronic NDEP Ecological Criterion ^b
SW8260D	1,1,2-Trichloroethane	79-00-5	µg/L	3	500
SW8260D	1,1-Dichloroethane	75-34-3	µg/L	50	--
SW8260D	1,1-Dichloroethene	75-35-4	µg/L	1	65
SW8260D	1,2-Dichlorobenzene	95-50-1	µg/L	600	14
SW8260D	1,2-Dichloroethane	107-06-2	µg/L	2	910
SW8260D	1,2-Dichloroethene (Total)	540-59-0	µg/L		
SW8260D	1,4-Dichlorobenzene	106-46-7	µg/L	75	9.4
SW8260D	2-Butanone	78-93-3	µg/L	300	--
SW8260D	Acetone	67-64-1	µg/L	6000	--
SW8260D	Benzene	71-43-2	µg/L	1	114
SW8260D	Bromodichloromethane	75-27-4	µg/L	1	--
SW8260D	Bromomethane	74-83-9	µg/L	10	--
SW8260D	Carbon disulfide	75-15-0	µg/L	700	--
SW8260D	Carbon tetrachloride	56-23-5	µg/L	1	240
SW8260D	Chlorobenzene	108-90-7	µg/L	50	47
SW8260D	Chloroethane	75-00-3	µg/L	--	--
SW8260D	Chloroform	67-66-3	µg/L	70	140
SW8260D	Chloromethane	74-87-3	µg/L	--	--
SW8260D	cis-1,2-Dichloroethene	156-59-2	µg/L	70	--
SW8260D	Cyclohexane	110-82-7	µg/L	--	--
SW8260D	Dibromochloromethane	124-48-1	µg/L	1	--
SW8260D	Dichlorodifluoromethane	75-71-8	µg/L	1000	--
SW8260D	Ethylbenzene	100-41-4	µg/L	700	14
SW8260D	Freon TF	76-13-1	µg/L	20000	--
SW8260D	Isopropylbenzene	98-82-8	µg/L	700	--
SW8260D	m&p-Xylene	179601-23-1	µg/L	1000	27
SW8260D	Methylcyclohexane	108-87-2	µg/L	--	--
SW8260D	Methylene Chloride	75-09-2	µg/L	3	940
SW8260D	MTBE	1634-04-4	µg/L	70	51000
SW8260D	o-Xylene	95-47-6	µg/L	1000	27
SW8260D	Tetrachloroethene	127-18-4	µg/L	1	45
SW8260D	Toluene	108-88-3	µg/L	600	253
SW8260D	trans-1,2-Dichloroethene	156-60-5	µg/L	100	970
SW8260D	Trichloroethene	79-01-6	µg/L	1	47
SW8260D	Vinyl chloride	75-01-4	µg/L	1	930
SW8260D	Xylenes, Total	1330-20-7	µg/L		
ECO-PAHs					
SW8270E SIM	1,4-Dioxane	123-91-1	µg/L	0.4	--

Table 3. Surface Water Target Analytes, Methods, Action Levels, and Control Limits

Site Sampling Plan for Ecological Evaluation

Princeton Technology Center, West Windsor Township, New Jersey

Method	Analyte	CAS Number	Units	Higher of PQL and Ground Water Quality Criterion ^a	Fresh Surface Water Chronic NJDEP Ecological Criterion ^b
SW8270E SIM	2-Methylnaphthalene	91-57-6	µg/L	30	330
SW8270E SIM	Acenaphthene	83-32-9	µg/L	400	38
SW8270E SIM	Acenaphthylene	208-96-8	µg/L	--	4840
SW8270E SIM	Anthracene	120-12-7	µg/L	2000	0.035
SW8270E SIM	Benzo(a)anthracene	56-55-3	µg/L	0.1	0.025
SW8270E SIM	Benzo(a)pyrene	50-32-8	µg/L	0.1	0.014
SW8270E SIM	Benzo(b)fluoranthene	205-99-2	µg/L	0.2	9.07
SW8270E SIM	Benzo(g,h,i)perylene	191-24-2	µg/L	--	7.64
SW8270E SIM	Benzo(k)fluoranthene	207-08-9	µg/L	0.5	--
SW8270E SIM	Chrysene	218-01-9	µg/L	5	--
SW8270E SIM	Dibenz(a,h)anthracene	53-70-3	µg/L	0.3	--
SW8270E SIM	Fluoranthene	206-44-0	µg/L	300	1.9
SW8270E SIM	Fluorene	86-73-7	µg/L	300	19
SW8270E SIM	Indeno[1,2,3-cd]pyrene	193-39-5	µg/L	0.2	4.31
SW8270E SIM	Naphthalene	91-20-3	µg/L	300	13
SW8270E SIM	Phenanthrene	85-01-8	µg/L	--	3.6
SW8270E SIM	Pyrene	129-00-0	µg/L	200	0.3
ECO-Metals					
SW3060A/7196A	Hexavalent Chromium	18540-29-9	µg/L	--	10
SW7470A	Mercury	7439-97-6	µg/L	2	0.77
SW6020B	Aluminum	7429-90-5	µg/L	--	--
SW6020B	Antimony	7440-36-0	µg/L	6	80
SW6020B	Arsenic	7440-38-2	µg/L	3	150
SW6020B	Barium	7440-39-3	µg/L	6000	220
SW6020B	Beryllium	7440-41-7	µg/L	1	3.6
SW6020B	Cadmium	7440-43-9	µg/L	4	--
SW6020B	Calcium	7440-70-2	µg/L	--	--
SW6020B	Chromium	7440-47-3	µg/L	--	42
SW6020B	Cobalt	7440-48-4	µg/L	100	24
SW6020B	Copper	7440-50-8	µg/L	1300	--
SW6020B	Iron	7439-89-6	µg/L	--	--
SW6020B	Lead	7439-92-1	µg/L	5	5.4
SW6020B	Magnesium	7439-95-4	µg/L	--	--
SW6020B	Manganese	7439-96-5	µg/L	--	--
SW6020B	Nickel	7440-02-0	µg/L	100	--
SW6020B	Potassium	7440-09-7	µg/L	--	--
SW6020B	Selenium	7782-49-2	µg/L	40	5
EPA 200.7	Silica	7631-86-9	µg/L	--	--

Table 3. Surface Water Target Analytes, Methods, Action Levels, and Control Limits
 Site Sampling Plan for Ecological Evaluation
 Princeton Technology Center, West Windsor Township, New Jersey

Method	Analyte	CAS Number	Units	Higher of PQL and Ground Water Quality Criterion ^a	Fresh Surface Water Chronic NUDEP Ecological Criterion ^b
SW6020B	Silver	7440-22-4	µg/L	40	0.12
SW6020B	Sodium	7440-23-5	µg/L	--	--
SW6020B	Thallium	7440-28-0	µg/L	--	10
SW6020B	Vanadium	7440-62-2	µg/L	--	12
SW6020B	Zinc	7440-66-6	µg/L	2000	--

Notes:

^a New Jersey Department of Environmental Protection (NJDEP) Ground Water Quality Standards - Class IIA by Constituent. May 2021. New Jersey Administration Code 7:9C-1.4; Remediation Standards.

^b NJDEP Ground Water Quality Standards - Class IIA by Constituent. May 2021. New Jersey Administration Code 7:9C-1.4; Remediation Standards. NJDEP Ecological Surface Water SSLs. March 2009.

Bold = MDL and RL exceed screening criteria.

-- = not available (no standard)

µg/L = microgram(s) per liter

CAS = Chemical Abstracts Service

Freon TF = 1,1,2-Trichloro-1,2,2-trifluoroethane

MDL = method detection limit

MTBE = methyl tert butyl ether

NJDEP = New Jersey Department of Environmental Protection

PAH = polycyclic aromatic hydrocarbon

PQL = Practical Quantitation Level as defined in N.J.A.C. 7:9C-1.4

RL = reporting limit

SIM = selected ion method

SVOC = semivolatile organic compound

VOC = volatile organic compound



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Laboratory Certification

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (L-A-B)	L2219
Maine	2024021
Maryland	296
New Hampshire	255423
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	T104704488

LOGIN REPORT/SAMPLE TRANSFER

Order ID : P3657	JACO05	Order Date : 8/16/2024 2:45:00 PM	Project Mgr :
Client Name : JACOBS Engineering Grou		Project Name : Former Schlumberger Site I	Report Type : Level 4
Client Contact : Mary I. Murphy		Receive DateTime : 8/16/2024 12:00:00 AM	EDD Type : CH2MHILL
Invoice Name : JACOBS Engineering Grou		Purchase Order : 12:45	Hard Copy Date :
Invoice Contact : Mary I. Murphy			Date Signoff :

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
P3657-01	~~717~~ -J-WS-O81624 917	Water	08/16/2024	09:30					
					VOCMS Group6		8260-Low	10 Bus. Days	
P3657-02	TB-01-081624	Water	08/16/2024	10:55					
					VOCMS Group6		8260-Low	10 Bus. Days	

Relinquished By : MM
Date / Time : 08-16-24 1524

Received By : Sam
Date / Time : 8/16/24 15:27 ag # 4

Storage Area : VOA Refridgerator Room