

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: JACO05
Lab Code: CHEM Case No.: P3657 SAS No.: P3657 SDG NO.: P3657
Lab File ID: VN083367.D Date Analyzed: 08/19/2024
Instrument ID: MSVOA_N Time Analyzed: 11:43
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	172122	8.22	293866	9.10	257989	11.87
UPPER LIMIT	344244	8.724	587732	9.6	515978	12.365
LOWER LIMIT	86061	7.724	146933	8.6	128995	11.365
EPA SAMPLE NO.						
917-J-WS-081624	139398	8.22	273365	9.11	283963	11.87
TB-01-081624	133915	8.22	268412	9.11	271721	11.87
VN0819WBL01	144938	8.22	283183	9.10	290527	11.87
VN0819WBS01	162452	8.22	284677	9.10	251348	11.87
VN0819WBSD01	154957	8.22	272633	9.11	241038	11.87

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.