

Surrogate Summary

SDG No.: **P3657**

Client: **JACOBS Engineering Group, Inc.**

Analytical Method: **SW8260-Low**

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
P3657-01	917-J-WS-081624	1,2-Dichloroethane-d4	50	53.3	107	70 (74)	130 (125)
		Dibromofluoromethane	50	48.9	98	70 (75)	130 (124)
		Toluene-d8	50	48.9	98	70 (86)	130 (113)
		4-Bromofluorobenzene	50	51.3	103	70 (77)	130 (121)
P3657-02	TB-01-081624	1,2-Dichloroethane-d4	50	53.1	106	70 (74)	130 (125)
		Dibromofluoromethane	50	48.6	97	70 (75)	130 (124)
		Toluene-d8	50	46.7	93	70 (86)	130 (113)
		4-Bromofluorobenzene	50	49.7	99	70 (77)	130 (121)
VN0819WBL01	VN0819WBL01	1,2-Dichloroethane-d4	50	51.4	103	70 (74)	130 (125)
		Dibromofluoromethane	50	48.4	97	70 (75)	130 (124)
		Toluene-d8	50	48.4	97	70 (86)	130 (113)
		4-Bromofluorobenzene	50	51.6	103	70 (77)	130 (121)
VN0819WBS01	VN0819WBS01	1,2-Dichloroethane-d4	50	56.0	112	70 (74)	130 (125)
		Dibromofluoromethane	50	54.2	108	70 (75)	130 (124)
		Toluene-d8	50	53.4	107	70 (86)	130 (113)
		4-Bromofluorobenzene	50	52.4	105	70 (77)	130 (121)
VN0819WBSD01	VN0819WBSD01	1,2-Dichloroethane-d4	50	57.5	115	70 (74)	130 (125)
		Dibromofluoromethane	50	54.9	110	70 (75)	130 (124)
		Toluene-d8	50	52.6	105	70 (86)	130 (113)
		4-Bromofluorobenzene	50	52.2	104	70 (77)	130 (121)

