

ALLIANCE 284 Sheffield Street, Mountainside New Jersey 07092

NEW JERSEY LAB ID#: 20012: NEW YORK LAB ID#: 11376

GENERAL CHEMISTRY CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NUMBER: P3671

MATRIX: Solid

METHOD: 7196A, Lloyd Kahn

	NA	NO	YES
1. Blank Contamination - If yes, list compounds and concentrations in each blank:		✓	
2. Matrix Spike Duplicate Recoveries Met Criteria		✓	
If not met, list those compounds and their recoveries which fall outside the acceptable range.			
The Blank Spike met requirements for all samples. The Matrix Spike (S-929-K1-SO-0-0.5-082024MS) analysis met criteria for all samples except for TOC due to sample matrix interference. The Matrix Spike Duplicate (S-929-K1-SO-0-0.5-082024MSD) analysis met criteria for all samples except for TOC due to sample matrix interference.			
3. Sample Duplicate Analysis Met QC Criteria			✓
If not met, list those compounds and their recoveries which fall outside the acceptable range.			
4. Digestion Holding Time Met			✓
If not met, list number of days exceeded for each sample:			

ADDITIONAL COMMENTS: This Data Package has been Revised due to Client ID Change for some sample as per Client Request.

Lab has used least representable sample weight for the sample P3671-01, P3671-02, P3671-03, P3671-04, P3671-05, P3671-06, P3671-07, P3671-08, P3671-09, P3671-10, P3671-11, P3671-13 and P3671-14 for TOC analysis. Therefore Lab has reported the TOC result with "OR" qualifier.

QA REVIEW

Date