

Initial and Continuing Calibration Verification

Client: JACOBS Engineering Group, Inc.

SDG No.: P3671

Project: Former Schlumberger Site Princeton NJ

RunNo.: LB132146

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Hexavalent Chromium	mg/L	0.504	0.5	101	90-110	08/23/2024
Sample ID: CCV1 Hexavalent Chromium	mg/L	0.500	0.5	100	90-110	08/23/2024
Sample ID: CCV2 Hexavalent Chromium	mg/L	0.501	0.5	100	90-110	08/23/2024
Sample ID: CCV3 Hexavalent Chromium	mg/L	0.501	0.5	100	90-110	08/23/2024
Sample ID: CCV4 Hexavalent Chromium	mg/L	0.504	0.5	101	90-110	08/23/2024

Initial and Continuing Calibration Verification

Client: JACOBS Engineering Group, Inc.

SDG No.: P3671

Project: Former Schlumberger Site Princeton NJ

RunNo.: LB132204

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: TOC	ICV1 mg/L	1010	1000	101	90-110	08/07/2024
Sample ID: TOC	CCV1 mg/L	987	1000	99	90-110	08/29/2024
Sample ID: TOC	CCV2 mg/L	989	1000	99	90-110	08/29/2024
Sample ID: TOC	CCV3 mg/L	1010	1000	101	90-110	08/29/2024
Sample ID: TOC	CCV4 mg/L	986	1000	99	90-110	08/30/2024
Sample ID: TOC	CCV5 mg/L	1020	1000	102	90-110	08/30/2024