

Initial and Continuing Calibration Blank Summary

Client: JACOBS Engineering Group, Inc.

SDG No.: P3671

Project: Former Schlumberger Site Princeton NJ

RunNo.: LB132146

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.0027	0.01	08/23/2024
Sample ID: CCB1 Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.0027	0.01	08/23/2024
Sample ID: CCB2 Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.0027	0.01	08/23/2024
Sample ID: CCB3 Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.0027	0.01	08/23/2024
Sample ID: CCB4 Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.0027	0.01	08/23/2024

Initial and Continuing Calibration Blank Summary

Client: JACOBS Engineering Group, Inc.

SDG No.: P3671

Project: Former Schlumberger Site Princeton NJ

RunNo.: LB132204

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 TOC	mg/L	< 125.0000	125.0000	U	32.1	250	08/07/2024
Sample ID: CCB1 TOC	mg/L	< 125.0000	125.0000	U	32.1	250	08/29/2024
Sample ID: CCB2 TOC	mg/L	< 125.0000	125.0000	U	32.1	250	08/29/2024
Sample ID: CCB3 TOC	mg/L	< 125.0000	125.0000	U	32.1	250	08/29/2024
Sample ID: CCB4 TOC	mg/L	< 125.0000	125.0000	U	32.1	250	08/30/2024
Sample ID: CCB5 TOC	mg/L	< 125.0000	125.0000	U	32.1	250	08/30/2024