

Laboratory Control Sample Summary

Client: JACOBS Engineering Group, Inc.

SDG No.: P3671

Project: Former Schlumberger Site Princeton NJ

Run No.: LB132204

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB132204BSS							
TOC	mg/Kg	1000	976		98	1	90-110	08/29/2024

Laboratory Control Sample Summary

Client: JACOBS Engineering Group, Inc.

SDG No.: P3671

Project: Former Schlumberger Site Princeton NJ

Run No.: LB132204

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB132204BSS2							
TOC	mg/Kg	1000	977		98	1	90-110	08/30/2024

Laboratory Control Sample Summary

Client: JACOBS Engineering Group, Inc.

SDG No.: P3671

Project: Former Schlumberger Site Princeton NJ

Run No.: LB132146

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID PB162926BS								
Hexavalent Chromium	mg/Kg	20	20.2		101	1	84-110	08/23/2024