

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: JAC005

Lab Code: CHEM Case No.: P3671 SAS No.: P3671 SDG No.: P3671

Instrument ID: BNA_N Calibration Date/Time: 08/23/2024 09:48

Lab File ID: BN033582.D Init. Calib. Date(s): 08/19/2024 08/19/2024

EPA Sample No.: SSTDCCC0.4 Init. Calib. Time(s): 16:16 19:53

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.572	0.537		-6.1	20.0
Fluoranthene-d10	0.961	0.996		3.6	20.0
2-Fluorophenol	1.271	1.116		-12.2	20.0
Phenol-d6	1.512	1.361		-10.0	20.0
Nitrobenzene-d5	0.332	0.305		-8.1	20.0
Naphthalene	1.069	0.979		-8.4	20.0
2-Methylnaphthalene	0.677	0.621		-8.3	20.0
2-Fluorobiphenyl	1.634	1.456		-10.9	20.0
Acenaphthylene	1.754	1.541		-12.1	20.0
Acenaphthene	1.234	1.077		-12.7	20.0
Fluorene	1.555	1.395		-10.3	20.0
2,4,6-Tribromophenol	0.215	0.215		0.0	20.0
Phenanthrene	1.113	1.005		-9.7	20.0
Anthracene	0.985	0.894		-9.2	20.0
Fluoranthene	1.230	1.243		1.1	20.0
Pyrene	1.785	1.433		-19.7	20.0
Terphenyl-d14	0.909	0.768		-15.5	20.0
Benzo (a) anthracene	1.446	1.377		-4.8	20.0
Chrysene	1.437	1.334		-7.2	20.0
Benzo (b) fluoranthene	1.494	1.327		-11.2	20.0
Benzo (k) fluoranthene	1.470	1.255		-14.6	20.0
Benzo (a) pyrene	1.236	1.116		-9.7	20.0
Indeno (1,2,3-cd) pyrene	1.661	1.484		-10.7	20.0
Dibenzo (a,h) anthracene	1.328	1.183		-10.9	20.0
Benzo (g,h,i) perylene	1.420	1.247		-12.2	20.0
1,4-Dioxane	0.460	0.443		-3.7	20.0

All other compounds must meet a minimum RRF of 0.010.