



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN082224\
Data File : BN033562.D
Acq On : 22 Aug 2024 21:24
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN081924.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Fri Aug 23 00:22:18 2024






AutoFind: Scans 2451, 2452, 2453; Background Corrected with Scan 2444

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	37.9	183614	PASS
68	69	0.00	2	1.8	3432	PASS
69	198	0.00	100	40.0	194197	PASS
70	69	0.00	2	0.5	1014	PASS
127	198	10	80	49.7	241152	PASS
197	198	0.00	2	0.3	1497	PASS
198	198	100	100	100.0	485077	PASS
199	198	5	9	6.8	33064	PASS
275	198	10	60	27.0	131016	PASS
365	198	1	100	3.6	17521	PASS
441	198	0.01	100	10.6	51227	PASS
442	442	50	100	100.0	353387	PASS
443	442	15	24	18.4	65013	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN082224\
Data File : BN033562.D
Acq On : 22 Aug 2024 21:24
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Aug 23 07:44:10 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Aug 23 07:42:58 2024
Response via : Initial Calibration



TIC: BN033562.D\data.ms

(70) Pentachlorophenol (C)

16.587min (-0.006) 266711.29 ng




response 630018

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	63.86
264.00	61.60	65.33
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN082224\
Data File : BN033562.D
Acq On : 22 Aug 2024 21:24
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Aug 23 07:44:10 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Aug 23 07:42:58 2024
Response via : Initial Calibration



TIC: BN033562.D\data.ms

(77) Benzidine

19.204min (-0.006) 5344.12 ng

response 3912941

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	13.57
183.00	13.20	12.02
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
8/22/2024	BNA_N	<u>BN033562.D</u>
Compound Name	Response	Retention Time
DDT	1657614	20.445
DDD	210956	20.004
DDE	1311	19.492
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
212267	1869881	11.35

Instrument :
BNA_N
ClientSampleId :
DFTPP