

NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NUMBER: bn082224

SequenceID :	<u>bn082224</u>	NA	NO	YES
1. Chromatograms Labeled/Compounds Identified. (Field samples and Method Blanks)		_____	_____	_____✓_____
2. GC/MS Tuning Specifications. DFTPP Meet Criteria Criteria (NOTE THAT THERE ARE DIFFERENT CRITERIA FOR NY ASP CLP, CLP AND NJ)		_____	_____	_____✓_____
3. GC/MS Tuning Frequency - Performed every 24 hours for 600 series and 12 hours for 8000 series		_____	_____	_____✓_____
4. GC/MS Calibration - Initial Calibration performed within 30 days before sample analysis and continuing calibration performed within 24 hours of sample analysis for 600 series and 12 hours for 8000 series		_____	_____	_____✓_____
5. GC/MS Calibration Met:		_____	_____	_____✓_____
a. Initial calibration Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range.		_____	_____	_____✓_____
b. Continuous Calibration(CCC) Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range. Compound #20 is biased Low side in the SSTDCCC(BN033563.D) but it is not required for associated samples.		_____	_____✓_____	_____
6. Blank Contamination - If yes, list compounds and concentrations in each blank:		_____	_____✓_____	_____
a. B/N Fraction				
d. Acid Fraction				

7. Surrogate Recoveries Meet Criteria

If not met, list those compounds and their recoveries which fall outside the acceptable ranges.

a. B/N Fraction

d. Acid Fraction

____ ✓

8. Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria

If not met, list those compounds and their recoveries which fall outside the acceptable range.

a. B/N Fraction

d. Acid Fraction

✓ _____

9. Internal Standard Area/Retention Time Shift Meet Criteria

Comments:

____ ✓

10. Extraction Holding Time Met

If not met, list number of days exceeded for each sample:

____ ✓

11. Analysis Holding Time Met

If not met, list number of days exceeded for each sample:

____ ✓

ADDITIONAL COMMENTS:

Recovery and RPD of certain compounds were affected in P3660-06MS/ 07MSD, due to matrix interference. No corrective action is required.

Hitesh

Analyst

08/23/2024

Date

NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NUMBER: bn082224

SequenceID : bn082224 NA NO YES