

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081924\
 Data File : BN033480.D
 Acq On : 19 Aug 2024 16:52
 Operator : MA/JU
 Sample : SSTDICC0.2
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.2

Quant Time: Aug 19 23:22:19 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN081924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 19 23:20:26 2024
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.560	152	7894	0.400	ng	0.00
7) Naphthalene-d8	10.314	136	21153	0.400	ng	0.00
13) Acenaphthene-d10	14.189	164	11095	0.400	ng	0.00
19) Phenanthrene-d10	16.942	188	23305	0.400	ng	0.00
29) Chrysene-d12	21.148	240	15592	0.400	ng	0.00
35) Perylene-d12	23.323	264	15116	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.191	112	5055	0.228	ng	0.00
5) Phenol-d6	6.743	99	6221	0.214	ng	0.00
8) Nitrobenzene-d5	8.692	82	3428	0.214	ng	0.00
11) 2-Methylnaphthalene-d10	11.915	152	5985	0.188	ng	0.00
14) 2,4,6-Tribromophenol	15.688	330	1123	0.198	ng	0.00
15) 2-Fluorobiphenyl	12.810	172	8881	0.198	ng	0.00
27) Fluoranthene-d10	18.980	212	10713	0.175	ng	0.00
31) Terphenyl-d14	19.598	244	6989	0.233	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.190	88	2026	0.238	ng	95
3) n-Nitrosodimethylamine	3.479	42	2169	0.197	ng	# 99
6) bis(2-Chloroethyl)ether	6.996	93	4514	0.200	ng	99
9) Naphthalene	10.368	128	11298	0.197	ng	99
10) Hexachlorobutadiene	10.667	225	2251	0.204	ng	# 100
12) 2-Methylnaphthalene	11.990	142	7058	0.184	ng	99
16) Acenaphthylene	13.911	152	8927	0.175	ng	100
17) Acenaphthene	14.253	154	6529	0.186	ng	100
18) Fluorene	15.247	166	8225	0.179	ng	99
20) 4,6-Dinitro-2-methylph...	15.322	198	589	0.203	ng	# 58
21) 4-Bromophenyl-phenylether	16.147	248	2696	0.194	ng	97
22) Hexachlorobenzene	16.247	284	3027	0.195	ng	100
23) Atrazine	16.421	200	2139	0.193	ng	98
24) Pentachlorophenol	16.594	266	1137	0.179	ng	97
25) Phenanthrene	16.979	178	12438	0.187	ng	100
26) Anthracene	17.078	178	10626	0.180	ng	100
28) Fluoranthene	19.012	202	13434	0.166	ng	99
30) Pyrene	19.375	202	13742	0.219	ng	99
32) Benzo(a)anthracene	21.131	228	10672	0.184	ng	99
33) Chrysene	21.184	228	10851	0.188	ng	98
34) Bis(2-ethylhexyl)phtha...	21.104	149	7125	0.257	ng	99
36) Indeno(1,2,3-cd)pyrene	25.481	276	11942	0.191	ng	99
37) Benzo(b)fluoranthene	22.677	252	10821	0.192	ng	# 90
38) Benzo(k)fluoranthene	22.677	252	10821	0.189	ng	92
39) Benzo(a)pyrene	23.227	252	8779	0.184	ng	# 87
40) Dibenzo(a,h)anthracene	25.496	278	9553	0.193	ng	95
41) Benzo(g,h,i)perylene	26.136	276	10368	0.190	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081924\
Data File : BN033480.D
Acq On : 19 Aug 2024 16:52
Operator : MA/JU
Sample : SSTDICC0.2
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Quant Time: Aug 19 23:22:19 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN081924.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 19 23:20:26 2024
Response via : Initial Calibration








#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.560 min Scan# 61
 Delta R.T. 0.001 min
 Lab File: BN033480.D
 Acq: 19 Aug 2024 16:52

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.2







Tgt Ion:152 Resp: 7894
 Ion Ratio Lower Upper
 152 100
 150 150.9 122.2 183.2
 115 58.7 47.2 70.8



#2
 1,4-Dioxane
 Concen: 0.238 ng
 RT: 3.190 min Scan# 29
 Delta R.T. 0.000 min
 Lab File: BN033480.D
 Acq: 19 Aug 2024 16:52

Tgt Ion: 88 Resp: 2026
 Ion Ratio Lower Upper
 88 100
 43 34.3 25.0 37.4
 58 73.6 62.5 93.7









#3
n-Nitrosodimethylamine
Concen: 0.197 ng
RT: 3.479 min Scan# 69
Delta R.T. 0.000 min
Lab File: BN033480.D
Acq: 19 Aug 2024 16:52




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion: 42 Resp: 2169
Ion Ratio Lower Upper
42 100
74 124.3 100.2 150.2
44 10.2 5.3 7.9#



#4
2-Fluorophenol
Concen: 0.228 ng
RT: 5.191 min Scan# 306
Delta R.T. 0.000 min
Lab File: BN033480.D
Acq: 19 Aug 2024 16:52

Tgt Ion: 112 Resp: 5055
Ion Ratio Lower Upper
112 100
64 57.6 47.1 70.7
63 30.5 24.9 37.3









#5
Phenol-d6
Concen: 0.214 ng
RT: 6.743 min Scan# 51
Delta R.T. 0.000 min
Lab File: BN033480.D
Acq: 19 Aug 2024 16:52




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion: 99 Resp: 6221
Ion Ratio Lower Upper
99 100
42 21.1 16.6 24.8
71 32.5 26.2 39.4



#6
bis(2-Chloroethyl)ether
Concen: 0.200 ng
RT: 6.996 min Scan# 556
Delta R.T. 0.000 min
Lab File: BN033480.D
Acq: 19 Aug 2024 16:52

Tgt Ion: 93 Resp: 4514
Ion Ratio Lower Upper
93 100
63 79.2 63.0 94.4
95 32.9 26.0 39.0









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.314 min Scan# 91
Delta R.T. 0.000 min
Lab File: BN033480.D
Acq: 19 Aug 2024 16:52




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:	136	Resp:	21153
Ion Ratio	Lower	Upper	
136	100		
137	11.4	9.0	13.6
54	10.1	8.3	12.5
68	7.7	5.9	8.9



#8
Nitrobenzene-d5
Concen: 0.214 ng
RT: 8.692 min Scan# 774
Delta R.T. 0.000 min
Lab File: BN033480.D
Acq: 19 Aug 2024 16:52

Tgt Ion:	82	Resp:	3428
Ion Ratio	Lower	Upper	
82	100		
128	45.9	36.0	54.0
54	53.2	42.0	63.0









#9
Naphthalene
Concen: 0.197 ng
RT: 10.368 min Scan# 911
Delta R.T. 0.000 min
Lab File: BN033480.D
Acq: 19 Aug 2024 16:52




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:128 Resp: 11298
Ion Ratio Lower Upper
128 100
129 11.9 9.1 13.7
127 13.5 10.7 16.1



#10
Hexachlorobutadiene
Concen: 0.204 ng
RT: 10.667 min Scan# 959
Delta R.T. 0.000 min
Lab File: BN033480.D
Acq: 19 Aug 2024 16:52

Tgt Ion:225 Resp: 2251
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.7 51.2 76.8









#11
2-Methylnaphthalene-d10
Concen: 0.188 ng
RT: 11.915 min Scan# 1145
Delta R.T. 0.000 min
Lab File: BN033480.D
Acq: 19 Aug 2024 16:52




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:152 Resp: 5985
Ion Ratio Lower Upper
152 100
151 20.7 16.6 25.0



#12
2-Methylnaphthalene
Concen: 0.184 ng
RT: 11.990 min Scan# 1165
Delta R.T. 0.000 min
Lab File: BN033480.D
Acq: 19 Aug 2024 16:52

Tgt Ion:142 Resp: 7058
Ion Ratio Lower Upper
142 100
141 89.3 71.7 107.5
115 38.3 29.4 44.2





#13

Acenaphthene-d10

Concen: 0.400 ng

RT: 14.189 min Scan# 1476

Delta R.T. 0.000 min

Lab File: BN033480.D

Acq: 19 Aug 2024 16:52

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2





Tgt Ion:164 Resp: 11095

Ion Ratio Lower Upper

164 100

162 104.3 83.5 125.3

160 51.3 40.2 60.4



#14

2,4,6-Tribromophenol

Concen: 0.198 ng

RT: 15.688 min Scan# 1614

Delta R.T. 0.000 min

Lab File: BN033480.D

Acq: 19 Aug 2024 16:52




Tgt Ion:330 Resp: 1123


Ion Ratio Lower Upper

330 100

332 96.5 77.5 116.3

141 42.3 33.9 50.9









#15
2-Fluorobiphenyl
Concen: 0.198 ng
RT: 12.810 min Scan# 1347
Delta R.T. 0.000 min
Lab File: BN033480.D
Acq: 19 Aug 2024 16:52




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:172	Resp:	8881
Ion	Ratio	Lower Upper
172	100	
171	35.7	27.7 41.5
170	24.0	18.3 27.5



#16
Acenaphthylene
Concen: 0.175 ng
RT: 13.911 min Scan# 1450
Delta R.T. 0.000 min
Lab File: BN033480.D
Acq: 19 Aug 2024 16:52

Tgt Ion:152	Resp:	8927
Ion	Ratio	Lower Upper
152	100	
151	19.5	15.7 23.5
153	12.7	10.3 15.5









#17
Acenaphthene
Concen: 0.186 ng
RT: 14.253 min Scan# 1482
Delta R.T. 0.000 min
Lab File: BN033480.D
Acq: 19 Aug 2024 16:52




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:154 Resp: 6529
Ion Ratio Lower Upper
154 100
153 111.0 89.0 133.6
152 55.9 45.2 67.8



#18
Fluorene
Concen: 0.179 ng
RT: 15.247 min Scan# 1575
Delta R.T. 0.000 min
Lab File: BN033480.D
Acq: 19 Aug 2024 16:52

Tgt Ion:166 Resp: 8225
Ion Ratio Lower Upper
166 100
165 96.7 78.2 117.4
167 13.5 10.6 16.0





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 16.942 min Scan# 1

Delta R.T. 0.000 min

Lab File: BN033480.D

Acq: 19 Aug 2024 16:52

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2





Tgt Ion:188 Resp: 23305

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 9.7 7.8 11.8



#20

4,6-Dinitro-2-methylphenol

Concen: 0.203 ng

RT: 15.322 min Scan# 1582

Delta R.T. 0.000 min

Lab File: BN033480.D

Acq: 19 Aug 2024 16:52




Tgt Ion:198 Resp: 589


Ion Ratio Lower Upper

198 100

51 132.2 65.1 97.7#

105 71.3 44.8 67.2#









#21
4-Bromophenyl-phenylether
Concen: 0.194 ng
RT: 16.147 min Scan# 1651
Delta R.T. 0.000 min
Lab File: BN033480.D
Acq: 19 Aug 2024 16:52




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:248 Resp: 2696
Ion Ratio Lower Upper
248 100
250 97.7 79.2 118.8
141 51.0 37.9 56.9



#22
Hexachlorobenzene
Concen: 0.195 ng
RT: 16.247 min Scan# 1659
Delta R.T. 0.000 min
Lab File: BN033480.D
Acq: 19 Aug 2024 16:52

Tgt Ion:284 Resp: 3027
Ion Ratio Lower Upper
284 100
142 40.0 31.8 47.6
249 32.7 26.0 39.0









#23
Atrazine
Concen: 0.193 ng
RT: 16.421 min Scan# 1673
Delta R.T. 0.000 min
Lab File: BN033480.D
Acq: 19 Aug 2024 16:52




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:200 Resp: 2139
Ion Ratio Lower Upper
200 100
173 33.7 25.3 37.9
215 46.7 36.6 54.8



#24
Pentachlorophenol
Concen: 0.179 ng
RT: 16.594 min Scan# 1687
Delta R.T. 0.000 min
Lab File: BN033480.D
Acq: 19 Aug 2024 16:52

Tgt Ion:266 Resp: 1137
Ion Ratio Lower Upper
266 100
264 63.3 51.9 77.9
268 61.2 51.0 76.4





#25

Phenanthrene

Concen: 0.187 ng

RT: 16.979 min Scan# 1718

Delta R.T. 0.000 min

Lab File: BN033480.D

Acq: 19 Aug 2024 16:52

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2





Tgt Ion:178 Resp: 12438

Ion Ratio Lower Upper

178 100

176 19.3 15.3 22.9

179 15.5 12.3 18.5



#26

Anthracene

Concen: 0.180 ng

RT: 17.078 min Scan# 1726

Delta R.T. 0.013 min

Lab File: BN033480.D

Acq: 19 Aug 2024 16:52




Tgt Ion:178 Resp: 10626


Ion Ratio Lower Upper

178 100

176 18.5 15.0 22.6

179 15.5 12.4 18.6









#27
Fluoranthene-d10
Concen: 0.175 ng
RT: 18.980 min Scan# 1944
Delta R.T. -0.005 min
Lab File: BN033480.D
Acq: 19 Aug 2024 16:52




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:212 Resp: 10713
Ion Ratio Lower Upper
212 100
106 15.0 12.3 18.5
104 8.7 7.0 10.4



#28
Fluoranthene
Concen: 0.166 ng
RT: 19.012 min Scan# 1950
Delta R.T. 0.000 min
Lab File: BN033480.D
Acq: 19 Aug 2024 16:52

Tgt Ion:202 Resp: 13434
Ion Ratio Lower Upper
202 100
101 12.4 9.5 14.3
203 17.1 13.8 20.6





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.148 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN033480.D

Acq: 19 Aug 2024 16:52

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2





Tgt Ion:240 Resp: 15592

Ion Ratio Lower Upper

240 100

120 15.1 12.4 18.6

236 28.1 23.0 34.6



#30

Pyrene

Concen: 0.219 ng

RT: 19.375 min Scan# 2028

Delta R.T. 0.000 min

Lab File: BN033480.D

Acq: 19 Aug 2024 16:52




Tgt Ion:202 Resp: 13742


Ion Ratio Lower Upper

202 100

200 21.2 16.6 24.8

203 18.0 14.2 21.4





#31

Terphenyl-d14

Concen: 0.233 ng

RT: 19.598 min Scan# 2076

Delta R.T. 0.000 min

Lab File: BN033480.D

Acq: 19 Aug 2024 16:52

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2





Tgt Ion:244 Resp: 6989

Ion Ratio Lower Upper

244 100

212 10.1 7.8 11.6

122 17.3 13.3 19.9



#32

Benzo(a)anthracene

Concen: 0.184 ng

RT: 21.131 min Scan# 2393

Delta R.T. 0.000 min

Lab File: BN033480.D

Acq: 19 Aug 2024 16:52




Tgt Ion:228 Resp: 10672


Ion Ratio Lower Upper

228 100

226 28.1 21.8 32.6

229 19.5 15.8 23.6









#33
 Chrysene
 Concen: 0.188 ng
 RT: 21.184 min Scan# 2118
 Delta R.T. 0.000 min
 Lab File: BN033480.D
 Acq: 19 Aug 2024 16:52




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.2


Tgt Ion:	228	Resp:	10851
Ion Ratio	Lower	Upper	
228	100		
226	30.9	23.8	35.8
229	19.9	15.6	23.4



#34
 Bis(2-ethylhexyl)phthalate
 Concen: 0.257 ng
 RT: 21.104 min Scan# 2390
 Delta R.T. 0.009 min
 Lab File: BN033480.D
 Acq: 19 Aug 2024 16:52

Tgt Ion:	149	Resp:	7125
Ion Ratio	Lower	Upper	
149	100		
167	26.5	21.5	32.3
279	2.9	2.2	3.2





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.323 min Scan# 21

Delta R.T. 0.003 min

Lab File: BN033480.D

Acq: 19 Aug 2024 16:52

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2





Tgt Ion:264 Resp: 15116

Ion Ratio Lower Upper

264 100

260 26.3 20.8 31.2

265 68.3 52.2 78.2



#36

Indeno(1,2,3-cd)pyrene

Concen: 0.191 ng

RT: 25.481 min Scan# 3586

Delta R.T. -0.003 min

Lab File: BN033480.D

Acq: 19 Aug 2024 16:52




Tgt Ion:276 Resp: 11942


Ion Ratio Lower Upper

276 100

138 30.9 24.4 36.6

277 24.9 19.8 29.6









#37
Benzo(b)fluoranthene
Concen: 0.192 ng
RT: 22.677 min Scan# 2627
Delta R.T. 0.000 min
Lab File: BN033480.D
Acq: 19 Aug 2024 16:52




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:252 Resp: 10821
Ion Ratio Lower Upper
252 100
253 28.5 19.8 29.8
125 23.5 13.9 20.9#



#38
Benzo(k)fluoranthene
Concen: 0.189 ng
RT: 22.677 min Scan# 2627
Delta R.T. -0.044 min
Lab File: BN033480.D
Acq: 19 Aug 2024 16:52

Tgt Ion:252 Resp: 10821
Ion Ratio Lower Upper
252 100
253 28.5 19.8 29.8
125 23.5 15.8 23.8









#39
Benzo(a)pyrene
Concen: 0.184 ng
RT: 23.227 min Scan# 2815
Delta R.T. 0.000 min
Lab File: BN033480.D
Acq: 19 Aug 2024 16:52




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:252 Resp: 8779
Ion Ratio Lower Upper
252 100
253 31.8 21.5 32.3
125 29.9 17.0 25.4#



#40
Dibenzo(a,h)anthracene
Concen: 0.193 ng
RT: 25.496 min Scan# 3591
Delta R.T. 0.000 min
Lab File: BN033480.D
Acq: 19 Aug 2024 16:52

Tgt Ion:278 Resp: 9553
Ion Ratio Lower Upper
278 100
139 26.1 19.1 28.7
279 28.8 21.0 31.4





#41
 Benzo(g,h,i)perylene
 Concen: 0.190 ng
 RT: 26.136 min Scan# 3811
 Delta R.T. -0.003 min
 Lab File: BN033480.D
 Acq: 19 Aug 2024 16:52

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.2

Tgt Ion:	276	Resp:	10368
Ion Ratio	Lower	Upper	
276	100		
277	26.2	19.7	29.5
138	29.0	21.8	32.6

