

NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

**GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY**

CHEMTECH PROJECT NUMBER: BP082224

SequenceID : BP082224

	NA	NO	YES
1. Chromatograms Labeled/Compounds Identified. (Field samples and Method Blanks)	_____	_____	_____✓
2. GC/MS Tuning Specifications. DFTPP Meet Criteria Criteria (NOTE THAT THERE ARE DIFFERENT CRITERIA FOR NY ASP CLP, CLP AND NJ)	_____	_____	_____✓
3. GC/MS Tuning Frequency - Performed every 24 hours for 600 series and 12 hours for 8000 series	_____	_____	_____✓
4. GC/MS Calibration - Initial Calibration performed within 30 days before sample analysis and continuing calibration performed within 24 hours of sample analysis for 600 series and 12 hours for 8000 series	_____	_____	_____✓
5. GC/MS Calibration Met:	_____	_____	_____✓
a. Initial calibration Meet Criteria	_____	_____	_____✓
If not met, list those compounds and their recoveries which fall outside the acceptable range.	_____	_____	_____
b. Continuous Calibration(CCC) Meet Criteria	_____	_____✓	_____
If not met, list those compounds and their recoveries which fall outside the acceptable range.	_____	_____	_____
Compound #26,32,42,53,54,56,57,59,62,65,70 Failing high side in the SSTDCCC(BP021607.D) but not present in any samples so CCC is ok to use.	_____	_____	_____
6. Blank Contamination - If yes, list compounds and concentrations in each blank:	_____	_____✓	_____
a. B/N Fraction	_____	_____	_____
d. Acid Fraction	_____	_____	_____

7. Surrogate Recoveries Meet Criteria

If not met, list those compounds and their recoveries which fall outside the acceptable ranges.

\_\_\_\_ ✓ \_\_\_\_

a. B/N Fraction

d. Acid Fraction

8. Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria

If not met, list those compounds and their recoveries which fall outside the acceptable range.

\_\_\_\_ ✓ \_\_\_\_

a. B/N Fraction The Recovery and RPD Fail for Some Compounds in P3689-02MS/MSD due to Matrix interference

d. Acid Fraction

9. Internal Standard Area/Retention Time Shift Meet Criteria

Comments:

\_\_\_\_ \_\_\_\_ ✓

10. Extraction Holding Time Met

If not met, list number of days exceeded for each sample:

\_\_\_\_ \_\_\_\_ ✓

11. Analysis Holding Time Met

If not met, list number of days exceeded for each sample:

\_\_\_\_ ✓ \_\_\_\_

ADDITIONAL COMMENTS:

2,4,6-Tribromophenol was marginally high in PB162379BL. Recovery of certain compounds were biased high and Benzaldehyde being a poor recovery compound, was low in PB162379BS. The data will be hard copies.

Jagrut

Analyst

08/22/2024

Date

NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

**GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY**

CHEMTECH PROJECT NUMBER: BP082224

SequenceID : BP082224 NA NO YES