

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP082224\
 Data File : BP021621.D
 Acq On : 22 Aug 2024 20:06
 Operator : RC/JU
 Sample : P3671-11
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 S-903-J-SO-0-0.5-081924

Quant Time: Aug 23 00:54:20 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 22 11:12:31 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.805	152	360701	20.000	ng	0.00
21) Naphthalene-d8	10.587	136	1546563	20.000	ng	0.00
39) Acenaphthene-d10	14.440	164	1075680	20.000	ng	0.00
64) Phenanthrene-d10	17.245	188	2395933	20.000	ng	0.00
76) Chrysene-d12	21.716	240	2027364	20.000	ng	0.00
86) Perylene-d12	25.151	264	2358797	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.399	112	1128878	46.471	ng	0.00
7) Phenol-d6	6.963	99	1630099	46.001	ng	0.00
23) Nitrobenzene-d5	8.946	82	915399	30.551	ng	0.00
42) 2,4,6-Tribromophenol	15.957	330	516823	43.170	ng	0.00
45) 2-Fluorobiphenyl	13.057	172	1242392	17.627	ng	0.00
79) Terphenyl-d14	19.969	244	2284318	21.890	ng	0.00
Target Compounds						Qvalue
26) 2-Nitrophenol	9.681	139	107	Below Cal	#	36
32) Benzoic acid	9.810	122	3237	3.717	ng	# 77
48) 2-Nitroaniline	13.510	65	129	5.154	ng	# 53
51) 2,6-Dinitrotoluene	14.069	165	160	3.996	ng	90
53) 3-Nitroaniline	14.310	138	143	3.978	ng	# 48
54) 2,4-Dinitrophenol	14.522	184	115	2.474	ng	# 1
56) 4-Nitrophenol	14.510	139	1110	5.166	ng	93
57) 2,4-Dinitrotoluene	14.792	165	413	5.042	ng	# 68
62) 4-Nitroaniline	15.587	138	76	2.940	ng	87
65) 4,6-Dinitro-2-methylph...	15.575	198	143	4.632	ng	# 1
71) Phenanthrene	17.292	178	3650463	31.681	ng	99
72) Anthracene	17.386	178	472712	4.183	ng	99
73) Carbazole	17.663	167	311582	2.861	ng	98
75) Fluoranthene	19.380	202	7628233	53.956	ng	99
77) Benzidine	19.598	184	454	4.580	ng	# 1
78) Pyrene	19.757	202	5859772	44.221	ng	99
80) Butylbenzylphthalate	20.710	149	27904	4.116	ng	100
81) Benzo(a)anthracene	21.698	228	2667422	20.959	ng	98
83) Chrysene	21.763	228	3058021	25.449	ng	98
85) Di-n-octyl phthalate	22.980	149	11719	4.916	ng	# 78
87) Indeno(1,2,3-cd)pyrene	29.051	276	2010108	13.147	ng	# 90
88) Benzo(b)fluoranthene	24.068	252	4117274	30.392	ng	98
89) Benzo(k)fluoranthene	24.068	252	4117274	31.155	ng	97
90) Benzo(a)pyrene	24.986	252	2660944	22.634	ng	98
91) Dibenzo(a,h)anthracene	29.115	278	490549	3.861	ng	93
92) Benzo(g,h,i)perylene	30.250	276	1875557	14.729	ng	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP082224\
Data File : BP021621.D
Acq On : 22 Aug 2024 20:06
Operator : RC/JU
Sample : P3671-11
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
S-903-J-SO-0-0.5-081924

Quant Time: Aug 23 00:54:20 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Aug 22 11:12:31 2024
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.805 min Scan# 81
Delta R.T. -0.000 min
Lab File: BP021621.D
Acq: 22 Aug 2024 20:06




Instrument :
BNA_P
ClientSampleId :
S-903-J-SO-0-0.5-081924


Tgt Ion:152 Resp: 360701
Ion Ratio Lower Upper
152 100
150 158.7 122.4 183.6
115 69.6 54.7 82.1



#5
2-Fluorophenol
Concen: 46.471 ng
RT: 5.399 min Scan# 410
Delta R.T. -0.000 min
Lab File: BP021621.D
Acq: 22 Aug 2024 20:06

Tgt Ion:112 Resp: 1128878
Ion Ratio Lower Upper
112 100
64 53.8 49.0 73.6
63 28.3 24.6 36.8









#7
Phenol-d6
Concen: 46.001 ng
RT: 6.963 min Scan# 61
Delta R.T. -0.006 min
Lab File: BP021621.D
Acq: 22 Aug 2024 20:06




Instrument :
BNA_P
ClientSampleId :
S-903-J-SO-0-0.5-081924


Tgt Ion: 99 Resp: 1630099
Ion Ratio Lower Upper
99 100
42 12.6 10.2 15.4
71 30.2 25.8 38.8



#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.587 min Scan# 1292
Delta R.T. -0.000 min
Lab File: BP021621.D
Acq: 22 Aug 2024 20:06

Tgt Ion:136 Resp: 1546563
Ion Ratio Lower Upper
136 100
137 10.8 8.8 13.2
54 7.7 6.3 9.5
68 7.0 6.8 10.2





#23

Nitrobenzene-d5

Concen: 30.551 ng

RT: 8.946 min Scan# 1160

Delta R.T. -0.000 min

Lab File: BP021621.D

Acq: 22 Aug 2024 20:06

Instrument :

BNA_P

ClientSampleId :

S-903-J-SO-0-0.5-081924





Tgt Ion: 82 Resp: 915399

Ion Ratio Lower Upper

82 100

128 37.2 26.5 39.7

54 39.1 30.3 45.5



#32

Benzoic acid

Concen: 3.717 ng

RT: 9.810 min Scan# 1160

Delta R.T. -0.082 min

Lab File: BP021621.D

Acq: 22 Aug 2024 20:06




Tgt Ion: 122 Resp: 3237


Ion Ratio Lower Upper

122 100

105 101.6 124.6 164.6#

77 106.7 96.2 136.2





#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 14.440 min Scan# 1947

Delta R.T. -0.006 min

Lab File: BP021621.D

Acq: 22 Aug 2024 20:06

Instrument :

BNA_P

ClientSampleId :

S-903-J-SO-0-0.5-081924





Tgt Ion:164 Resp: 1075680

Ion Ratio Lower Upper

164 100

162 103.1 81.5 122.3

160 44.8 35.2 52.8



#42

2,4,6-Tribromophenol

Concen: 43.170 ng

RT: 15.957 min Scan# 2205

Delta R.T. -0.000 min

Lab File: BP021621.D

Acq: 22 Aug 2024 20:06




Tgt Ion:330 Resp: 516823


Ion Ratio Lower Upper

330 100

332 97.8 77.4 116.0

141 32.1 24.5 36.7





#45

2-Fluorobiphenyl

Concen: 17.627 ng

RT: 13.057 min Scan# 1

Delta R.T. 0.006 min

Lab File: BP021621.D

Acq: 22 Aug 2024 20:06

Instrument :

BNA_P

ClientSampleId :

S-903-J-SO-0-0.5-081924





Tgt Ion:172 Resp: 1242392

Ion Ratio Lower Upper

172 100

171 34.7 28.2 42.4

170 22.7 18.4 27.6



#48

2-Nitroaniline

Concen: 5.154 ng

RT: 13.510 min Scan# 1789

Delta R.T. 0.006 min

Lab File: BP021621.D

Acq: 22 Aug 2024 20:06




Tgt Ion: 65 Resp: 129


Ion Ratio Lower Upper

65 100

92 78.4 58.1 87.1

138 8.4 65.5 98.3#









#51
2,6-Dinitrotoluene
Concen: 3.996 ng
RT: 14.069 min Scan# 1884
Delta R.T. 0.065 min
Lab File: BP021621.D
Acq: 22 Aug 2024 20:06




Instrument :
BNA_P
ClientSampleId :
S-903-J-SO-0-0.5-081924


Tgt Ion:165	Resp:	160
Ion Ratio	Lower	Upper
165	100	
63	59.7	51.5 77.3
89	60.8	57.8 86.8



#53
3-Nitroaniline
Concen: 3.978 ng
RT: 14.310 min Scan# 1925
Delta R.T. -0.029 min
Lab File: BP021621.D
Acq: 22 Aug 2024 20:06

Tgt Ion:138	Resp:	143
Ion Ratio	Lower	Upper
138	100	
108	65.7	16.2 24.4#
92	94.1	122.2 183.4#





#54

2,4-Dinitrophenol

Concen: 2.474 ng

RT: 14.522 min Scan# 1961

Delta R.T. -0.024 min

Lab File: BP021621.D

Acq: 22 Aug 2024 20:06

Instrument :

BNA_P

ClientSampleId :

S-903-J-SO-0-0.5-081924





Tgt Ion:184 Resp: 115

Ion Ratio Lower Upper

184 100

63 1765.3 64.9 97.3#

154 24033.7 52.1 78.1#



#56

4-Nitrophenol

Concen: 5.166 ng

RT: 14.510 min Scan# 1959

Delta R.T. -0.141 min

Lab File: BP021621.D

Acq: 22 Aug 2024 20:06




Tgt Ion:139 Resp: 1110


Ion Ratio Lower Upper

139 100

109 93.4 70.2 110.2

65 133.5 101.8 141.8









#57
2,4-Dinitrotoluene
Concen: 5.042 ng
RT: 14.792 min Scan# 2010
Delta R.T. -0.012 min
Lab File: BP021621.D
Acq: 22 Aug 2024 20:06




Instrument :
BNA_P
ClientSampleId :
S-903-J-SO-0-0.5-081924


Tgt Ion:165 Resp: 413
Ion Ratio Lower Upper
165 100
63 94.7 42.5 63.7#
89 79.3 77.1 115.7
182 21.7 2.0 3.0#



#62
4-Nitroaniline
Concen: 2.940 ng
RT: 15.587 min Scan# 2142
Delta R.T. 0.076 min
Lab File: BP021621.D
Acq: 22 Aug 2024 20:06

Tgt Ion:138 Resp: 76
Ion Ratio Lower Upper
138 100
92 72.5 46.4 86.4
108 76.6 73.2 113.2





#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 17.245 min Scan# 2425

Delta R.T. -0.006 min

Lab File: BP021621.D

Acq: 22 Aug 2024 20:06

Instrument :

BNA_P

ClientSampleId :

S-903-J-SO-0-0.5-081924





Tgt Ion:188 Resp: 2395933

Ion Ratio Lower Upper

188 100

94 11.4 10.2 15.4

80 12.5 11.6 17.4



#65

4,6-Dinitro-2-methylphenol

Concen: 4.632 ng

RT: 15.575 min Scan# 2140

Delta R.T. -0.012 min

Lab File: BP021621.D

Acq: 22 Aug 2024 20:06




Tgt Ion:198 Resp: 143


Ion Ratio Lower Upper

198 100

51 273.1 21.5 61.5#

105 376.5 32.9 72.9#





#71

Phenanthrene

Concen: 31.681 ng

RT: 17.292 min Scan# 2433

Delta R.T. -0.006 min

Lab File: BP021621.D

Acq: 22 Aug 2024 20:06

Instrument :

BNA_P

ClientSampleId :

S-903-J-SO-0-0.5-081924





Tgt Ion:178 Resp: 3650463

Ion Ratio Lower Upper

178 100

176 18.9 15.4 23.2

179 15.2 12.4 18.6



#72

Anthracene

Concen: 4.183 ng

RT: 17.386 min Scan# 2448

Delta R.T. -0.000 min

Lab File: BP021621.D

Acq: 22 Aug 2024 20:06




Tgt Ion:178 Resp: 472712


Ion Ratio Lower Upper

178 100

176 18.7 14.9 22.3

179 16.1 12.5 18.7









#73
 Carbazole
 Concen: 2.861 ng
 RT: 17.663 min Scan# 2496
 Delta R.T. -0.000 min
 Lab File: BP021621.D
 Acq: 22 Aug 2024 20:06




Instrument :
 BNA_P
 ClientSampleId :
 S-903-J-SO-0-0.5-081924


Tgt Ion:167 Resp: 311582
 Ion Ratio Lower Upper
 167 100
 166 21.7 17.0 25.4
 139 13.7 9.9 14.9



#75
 Fluoranthene
 Concen: 53.956 ng
 RT: 19.380 min Scan# 2787
 Delta R.T. -0.000 min
 Lab File: BP021621.D
 Acq: 22 Aug 2024 20:06

Tgt Ion:202 Resp: 7628233
 Ion Ratio Lower Upper
 202 100
 101 13.6 0.0 34.0
 203 18.1 0.0 37.5





#76

Chrysene-d12

Concen: 20.000 ng

RT: 21.716 min Scan# 3181

Delta R.T. -0.000 min

Lab File: BP021621.D

Acq: 22 Aug 2024 20:06

Instrument :

BNA_P

ClientSampleId :

S-903-J-SO-0-0.5-081924





Tgt Ion:240 Resp: 2027364

Ion Ratio Lower Upper

240 100

120 12.2 9.9 14.9

236 25.8 20.6 30.8



#77

Benzidine

Concen: 4.580 ng

RT: 19.598 min Scan# 2824

Delta R.T. 0.023 min

Lab File: BP021621.D

Acq: 22 Aug 2024 20:06




Tgt Ion:184 Resp: 454


Ion Ratio Lower Upper

184 100

185 93.1 11.4 17.0#

183 44.4 9.3 13.9#





#78

Pyrene

Concen: 44.221 ng

RT: 19.757 min Scan# 2851

Delta R.T. -0.000 min

Lab File: BP021621.D

Acq: 22 Aug 2024 20:06

Instrument :

BNA_P

ClientSampleId :

S-903-J-SO-0-0.5-081924





Tgt Ion:202 Resp: 5859772

Ion Ratio Lower Upper

202 100

200 21.1 16.7 25.1

203 18.0 14.1 21.1



#79

Terphenyl-d14

Concen: 21.890 ng

RT: 19.969 min Scan# 2887

Delta R.T. -0.006 min

Lab File: BP021621.D

Acq: 22 Aug 2024 20:06




Tgt Ion:244 Resp: 2284318


Ion Ratio Lower Upper

244 100

212 6.8 5.6 8.4

122 11.1 9.4 14.2





#80

Butylbenzylphthalate

Concen: 4.116 ng

RT: 20.710 min Scan# 3013

Delta R.T. -0.000 min

Lab File: BP021621.D

Acq: 22 Aug 2024 20:06

Instrument :

BNA_P

ClientSampleId :

S-903-J-SO-0-0.5-081924





Tgt Ion:149 Resp: 27904

Ion Ratio Lower Upper

149 100

91 91.5 73.3 109.9

206 22.0 17.1 25.7



#81

Benzo(a)anthracene

Concen: 20.959 ng

RT: 21.698 min Scan# 3181

Delta R.T. -0.000 min

Lab File: BP021621.D

Acq: 22 Aug 2024 20:06




Tgt Ion:228 Resp: 2667422


Ion Ratio Lower Upper

228 100

226 28.3 21.4 32.2

229 20.4 15.8 23.6





#83

Chrysene

Concen: 25.449 ng

RT: 21.763 min Scan# 3190

Delta R.T. -0.006 min

Lab File: BP021621.D

Acq: 22 Aug 2024 20:06

Instrument :

BNA_P

ClientSampleId :

S-903-J-SO-0-0.5-081924





Tgt Ion:228 Resp: 3058021

Ion Ratio Lower Upper

228 100

226 29.7 23.9 35.9

229 21.8 15.8 23.6



#85

Di-n-octyl phthalate

Concen: 4.916 ng

RT: 22.980 min Scan# 3399

Delta R.T. 0.018 min

Lab File: BP021621.D

Acq: 22 Aug 2024 20:06




Tgt Ion:149 Resp: 11719


Ion Ratio Lower Upper

149 100

167 0.0 1.3 1.9#

43 0.0 7.3 10.9#





#86

Perylene-d12

Concen: 20.000 ng

RT: 25.151 min Scan# 31

Delta R.T. -0.000 min

Lab File: BP021621.D

Acq: 22 Aug 2024 20:06

Instrument :

BNA_P

ClientSampleId :

S-903-J-SO-0-0.5-081924





Tgt Ion:264 Resp: 2358797

Ion Ratio Lower Upper

264 100

260 23.1 18.6 28.0

265 22.3 17.1 25.7



#87

Indeno(1,2,3-cd)pyrene

Concen: 13.147 ng

RT: 29.051 min Scan# 4431

Delta R.T. -0.006 min

Lab File: BP021621.D

Acq: 22 Aug 2024 20:06




Tgt Ion:276 Resp: 2010108


Ion Ratio Lower Upper

276 100

138 20.7 12.8 19.2#

277 25.0 16.5 24.7#





#88

Benzo(b)fluoranthene

Concen: 30.392 ng

RT: 24.068 min Scan# 3584

Delta R.T. 0.012 min

Lab File: BP021621.D

Acq: 22 Aug 2024 20:06

Instrument :

BNA_P

ClientSampleId :

S-903-J-SO-0-0.5-081924





Tgt Ion:252 Resp: 4117274

Ion Ratio Lower Upper

252 100

253 22.9 17.6 26.4

125 11.7 8.6 13.0



#89

Benzo(k)fluoranthene

Concen: 31.155 ng

RT: 24.068 min Scan# 3584

Delta R.T. -0.065 min

Lab File: BP021621.D

Acq: 22 Aug 2024 20:06




Tgt Ion:252 Resp: 4117274


Ion Ratio Lower Upper

252 100

253 22.9 17.5 26.3

125 11.7 7.8 11.8









#90
Benzo(a)pyrene
Concen: 22.634 ng
RT: 24.986 min Scan# 31
Delta R.T. 0.006 min
Lab File: BP021621.D
Acq: 22 Aug 2024 20:06




Instrument :
BNA_P
ClientSampleId :
S-903-J-SO-0-0.5-081924


Tgt Ion:252 Resp: 2660944
Ion Ratio Lower Upper
252 100
253 22.6 17.7 26.5
125 12.1 8.7 13.1



#91
Dibenzo(a,h)anthracene
Concen: 3.861 ng
RT: 29.115 min Scan# 4442
Delta R.T. -0.018 min
Lab File: BP021621.D
Acq: 22 Aug 2024 20:06

Tgt Ion:278 Resp: 490549
Ion Ratio Lower Upper
278 100
139 17.6 12.2 18.4
279 27.8 19.1 28.7





#92

Benzo(g,h,i)perylene

Concen: 14.729 ng

RT: 30.250 min Scan# 4

Delta R.T. -0.012 min

Lab File: BP021621.D


Acq: 22 Aug 2024 20:06

Instrument :

BNA_P

ClientSampleId :

S-903-J-SO-0-0.5-081924



Tgt Ion:276 Resp: 1875557

Ion Ratio Lower Upper

276 100

277 23.8 19.0 28.6

138 20.3 16.0 24.0

