

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: JAC005

Lab Code: CHEM

SAS No.: P3671

SDG NO.: P3671

Lab File ID: BF139112.D

DFTPP Injection Date: 08/22/2024

Instrument ID: BNA_F

DFTPP Injection Time: 09:55

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	42.6
68	Less than 2.0% of mass 69	0.8 (1.9) 1
69	Mass 69 relative abundance	43.1
70	Less than 2.0% of mass 69	0.2 (0.6) 1
127	10.0 - 80.0% of mass 198	47.5
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.8
275	10.0 - 60.0% of mass 198	28.1
365	Greater than 1% of mass 198	3.6
441	Present, but less than mass 443	13.2
442	Greater than 50% of mass 198	85.7
443	15.0 - 24.0% of mass 442	16.5 (19.3) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BF139113.D	08/22/2024	10:25
S-903-J-SO-0.5-1.0-081924	P3671-12	BF139135.D	08/22/2024	21:20