

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF082024

Review By	Jagrut	Review On	8/21/2024 10:35:25 AM
Supervise By	mohammad	Supervise On	8/22/2024 7:52:09 AM
SubDirectory	BF082024	HP Acquire Method	BNA_F
		HP Processing Method	bf082024

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6595,SP6594,SP6593,SP6592,SP6591,SP6590,SP6589,SP6588
CCC	SP6591
Internal Standard/PEM	S12041,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF139077.D	20 Aug 2024 16:39		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BF139078.D	20 Aug 2024 17:10		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BF139079.D	20 Aug 2024 17:40	Compound #9,32,54,56,65 removed from 5ppm	RC/JU	Ok,M
4	SSTDICC010	SSTDICC010	BF139080.D	20 Aug 2024 18:10	Com#26 and 48 Fail for %RE Criteria in the Calibration.	RC/JU	Ok
5	SSTDICC020	SSTDICC020	BF139081.D	20 Aug 2024 18:40	Compound #26,32,40,48,51,53,57,62,92 Kept on LR and Compound #54,65 Kept on QR	RC/JU	Ok
6	SSTDICCC040	SSTDICCC040	BF139082.D	20 Aug 2024 19:10	The Calibration is Good For 8270 E NON-DOD except com#26 and 48 and good for 625.1 Method for all compounds.	RC/JU	Ok,M
7	SSTDICC050	SSTDICC050	BF139083.D	20 Aug 2024 19:41		RC/JU	Ok
8	SSTDICC060	SSTDICC060	BF139084.D	20 Aug 2024 20:11		RC/JU	Ok,M
9	SSTDICC080	SSTDICC080	BF139085.D	20 Aug 2024 20:41	Compound #9 removed from 80ppm	RC/JU	Ok,M
10	SSTDICV040	ICVBF082024	BF139086.D	20 Aug 2024 21:41		RC/JU	Ok
11	PB162687BL	PB162687BL	BF139087.D	20 Aug 2024 22:10		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF082224

Review By	yogesh	Review On	8/23/2024 6:48:24 AM
Supervise By	mohammad	Supervise On	8/23/2024 6:56:56 AM
SubDirectory	BF082224	HP Acquire Method	BNA_F
		HP Processing Method	bf082024

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6595,SP6594,SP6593,SP6592,SP6591,SP6590,SP6589,SP6588
CCC	SP6591
Internal Standard/PEM	S12041,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF139112.D	22 Aug 2024 09:55		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF139113.D	22 Aug 2024 10:25		RC/JU	Ok
3	PB162782BL	PB162782BL	BF139114.D	22 Aug 2024 10:54	Analyzed for contamination check	RC/JU	Not Ok
4	PB162687BS	PB162687BS	BF139115.D	22 Aug 2024 11:23		RC/JU	Ok,M
5	PB162687BSD	PB162687BSD	BF139116.D	22 Aug 2024 11:53		RC/JU	Ok,M
6	P3641-04	S-908-K1-0.5-1.0-0815	BF139117.D	22 Aug 2024 12:22		RC/JU	Ok,M
7	P3663-03	914-J-SD-0-0.5-081624	BF139118.D	22 Aug 2024 12:52		RC/JU	Ok,M
8	P3641-02	S-907-K1-0.5-1.0-0815	BF139119.D	22 Aug 2024 13:22		RC/JU	Ok,M
9	P3640-01	SP-1-COMP	BF139120.D	22 Aug 2024 13:51	Com#26 and 48 Fail for %RE Criteria in the Calibration.	RC/JU	Not Ok
10	P3641-06	S-909-K1-0.5-1.0-0815	BF139121.D	22 Aug 2024 14:20	SIM Analysis not required	RC/JU	Ok,M
11	P3641-08	S-910-K1-0-0.5-081524	BF139122.D	22 Aug 2024 14:50	FD,Sim Analysis Not Required	RC/JU	Ok,M
12	P3641-09	S-910-K1-0.5-1.0-0815	BF139123.D	22 Aug 2024 15:21	Sim Analysis not required	RC/JU	Ok,M
13	P3641-10	S-910-K1-0.5-1.0-0815	BF139124.D	22 Aug 2024 15:51	Sim Analysis not required	RC/JU	Ok,M
14	P3641-01	S-907-K1-0-0.5-081524	BF139125.D	22 Aug 2024 16:21		RC/JU	Ok,M
15	P3663-09	915-J-SD-0-0.5-081624	BF139126.D	22 Aug 2024 16:51	SIM Analysis not required	RC/JU	Ok,M
16	P3663-11	919-J-SD-0-0.5-1-0816	BF139127.D	22 Aug 2024 17:21	SIM Analysis not required	RC/JU	Ok,M
17	P3663-12	919-J-SD-0.5-1-081624	BF139128.D	22 Aug 2024 17:51		RC/JU	Ok,M

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF082224

Review By	yogesh	Review On	8/23/2024 6:48:24 AM		
Supervise By	mohammad	Supervise On	8/23/2024 6:56:56 AM		
SubDirectory	BF082224	HP Acquire Method	BNA_F	HP Processing Method	bf082024
STD. NAME	STD REF.#				
Tune/Reschk	SP6573				
Initial Calibration Stds	SP6595,SP6594,SP6593,SP6592,SP6591,SP6590,SP6589,SP6588				
CCC	SP6591				
Internal Standard/PEM	S12041,10ul/1000ul sample				
ICV/I.BLK	SP6559				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

18	P3641-03	S-908-K1-0-0.5-081524	BF139129.D	22 Aug 2024 18:21	Need 5X Dilution, SIM Analysis not required	RC/JU	Dilution
19	P3660-03DL	S-902-J-0-0.5-081624D	BF139130.D	22 Aug 2024 18:50		RC/JU	Ok,M
20	P3660-04DL	S-902-J-0.5-1-081624D	BF139131.D	22 Aug 2024 19:20		RC/JU	Ok,M
21	P3663-10DL	915-J-SD-0.5-1-081624	BF139132.D	22 Aug 2024 19:50		RC/JU	Ok,M
22	P3641-05	S-909-K1-0-0.5-081524	BF139133.D	22 Aug 2024 20:20	SIM Analysis not required	RC/JU	Ok,M
23	P3641-07	S-910-K1-0-0.5-081524	BF139134.D	22 Aug 2024 20:50	Sim Analysis not required	RC/JU	Ok,M
24	P3671-12	S-903-J-SO-0.5-1.0-08	BF139135.D	22 Aug 2024 21:20		RC/JU	Ok,M

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF082624

Review By	Jagrut	Review On	8/27/2024 9:04:58 AM
Supervise By	mohammad	Supervise On	8/28/2024 1:33:27 AM
SubDirectory	BF082624	HP Acquire Method	BNA_F
		HP Processing Method	BF082024

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6595,SP6594,SP6593,SP6592,SP6591,SP6590,SP6589,SP6588
CCC	SP6591
Internal Standard/PEM	S12041,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF139158.D	26 Aug 2024 09:14		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF139159.D	26 Aug 2024 09:43		RC/JU	Ok,M
3	PB162917BL	PB162917BL	BF139160.D	26 Aug 2024 10:13		RC/JU	Ok
4	P3671-01DL	S-904-J-SO-0-0.5-0819	BF139161.D	26 Aug 2024 10:51		RC/JU	Ok,M
5	P3641-03DL	S-908-K1-0-0.5-081524	BF139162.D	26 Aug 2024 11:21		RC/JU	Ok,M

M : Manual Integration

Instrument ID: BNA_M

Daily Analysis Runlog For Sequence/QC Batch ID # BM081024

Review By	Jagrut	Review On	8/12/2024 5:03:24 PM
Supervise By	mohammad	Supervise On	8/13/2024 5:17:30 AM
SubDirectory	BM081024	HP Acquire Method	BNA_M
		HP Processing Method	bm081024

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6550,SP6551,SP6552,SP6553,SP6554,SP6555,SP6556,SP6557
CCC	SP6553
Internal Standard/PEM	S12039,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BM047139.D	10 Aug 2024 12:26		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BM047140.D	10 Aug 2024 13:12		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BM047141.D	10 Aug 2024 13:52	Compound #54 removed from 5ppm	RC/JU	Ok,M
4	SSTDICC010	SSTDICC010	BM047142.D	10 Aug 2024 14:31	Com#77 failing in calibration for both 8270DOD and 8270E	RC/JU	Ok,M
5	SSTDICC020	SSTDICC020	BM047143.D	10 Aug 2024 15:11	Compound #9 Kept on QR	RC/JU	Ok,M
6	SSTDICCC040	SSTDICCC040	BM047144.D	10 Aug 2024 15:51	The Calibration is Good For 8270 DOD and also good for 625.1 Method except com#77	RC/JU	Ok,M
7	SSTDICC050	SSTDICC050	BM047145.D	10 Aug 2024 16:31		RC/JU	Ok,M
8	SSTDICC060	SSTDICC060	BM047146.D	10 Aug 2024 17:11		RC/JU	Ok,M
9	SSTDICC080	SSTDICC080	BM047147.D	10 Aug 2024 17:51	Compound #9 removed from 80ppm	RC/JU	Ok,M
10	SSTDICV040	ICVBM081024	BM047148.D	10 Aug 2024 19:10	Com#77 failing in ICV for Both 8270 DOD and 8270 E and also in 625.1.	RC/JU	Ok,M
11	PB162577TB	PB162577TB	BM047149.D	10 Aug 2024 19:50		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_M

Daily Analysis Runlog For Sequence/QC Batch ID # BM082224

Review By	yogesh	Review On	8/23/2024 6:46:01 AM
Supervise By	mohammad	Supervise On	8/23/2024 6:56:43 AM
SubDirectory	BM082224	HP Acquire Method	BNA_M
		HP Processing Method	bm081024

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6550,SP6551,SP6552,SP6553,SP6554,SP6555,SP6556,SP6557
CCC	SP6553
Internal Standard/PEM	S12041,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BM047307.D	22 Aug 2024 08:35		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BM047308.D	22 Aug 2024 09:14		RC/JU	Ok
3	PB162876BL	PB162876BL	BM047309.D	22 Aug 2024 09:54		RC/JU	Ok
4	PB162876BS	PB162876BS	BM047310.D	22 Aug 2024 10:33		RC/JU	Ok
5	PB162574BL	PB162574BL	BM047311.D	22 Aug 2024 11:13		RC/JU	Ok
6	PB162574BS	PB162574BS	BM047312.D	22 Aug 2024 11:52	Surrogate Fail and Com#9 Recovery is zero	RC/JU	Not Ok
7	PB162757BL	PB162757BL	BM047313.D	22 Aug 2024 12:32		RC/JU	Ok
8	P3671-14	S-912-J-SO-0.5-1-0819	BM047314.D	22 Aug 2024 13:11		RC/JU	Ok
9	P3671-09	S-928-K1-SO-0.5-081	BM047315.D	22 Aug 2024 13:51		RC/JU	Ok
10	P3671-13	S-912-J-SO-0.5-0819	BM047316.D	22 Aug 2024 14:31		RC/JU	Ok
11	P3671-10	S-928-K1-SO-0.5-1-081	BM047317.D	22 Aug 2024 15:10		RC/JU	Ok
12	P3671-10MS	S-928-K1-SO-0.5-1-081	BM047318.D	22 Aug 2024 15:50		RC/JU	Ok
13	P3671-10MSD	S-928-K1-SO-0.5-1-081	BM047319.D	22 Aug 2024 16:30		RC/JU	Ok
14	P3671-03	S-911-K1-SO-0.5-081	BM047320.D	22 Aug 2024 17:10		RC/JU	Ok
15	P3671-04	S-911-K1-SO-0.5-1-081	BM047321.D	22 Aug 2024 17:50		RC/JU	Ok
16	P3671-08	S-905-J-SO-0.5-1.0-081	BM047322.D	22 Aug 2024 18:29	SIM Analysis not required	RC/JU	Ok,M
17	P3685-01	Stock Pile-1	BM047323.D	22 Aug 2024 19:09		RC/JU	Ok

Instrument ID: BNA_M

Daily Analysis Runlog For Sequence/QC Batch ID # BM082224

Review By	yogesh	Review On	8/23/2024 6:46:01 AM		
Supervise By	mohammad	Supervise On	8/23/2024 6:56:43 AM		
SubDirectory	BM082224	HP Acquire Method	BNA_M	HP Processing Method	bm081024

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6550,SP6551,SP6552,SP6553,SP6554,SP6555,SP6556,SP6557
CCC	SP6553
Internal Standard/PEM	S12041,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

18	P3685-03	Stock Pile-2	BM047324.D	22 Aug 2024 19:49		RC/JU	Ok
----	----------	--------------	------------	-------------------	--	-------	----

M : Manual Integration

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP081324

Review By	yogesh	Review On	8/14/2024 5:49:15 AM
Supervise By	mohammad	Supervise On	8/14/2024 6:23:14 AM
SubDirectory	BP081324	HP Acquire Method	BNA_P
		HP Processing Method	BP081324
STD. NAME	STD REF.#		
Tune/Reschk	SP6573		
Initial Calibration Stds	SP6595,SP6594,SP6593,SP6592,SP6591,SP6590,SP6589,SP6588		
CCC	SP6591		
Internal Standard/PEM	S12041,10ul/1000ul sample		
ICV/I.BLK	SP6559		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BP021480.D	13 Aug 2024 11:49		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BP021481.D	13 Aug 2024 12:32		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BP021482.D	13 Aug 2024 13:15	Compound #32,54,56,65,70,77,85 removed from 5ppm	RC/JU	Ok,M
4	SSTDICC010	SSTDICC010	BP021483.D	13 Aug 2024 13:59		RC/JU	Ok,M
5	SSTDICC020	SSTDICC020	BP021484.D	13 Aug 2024 14:42	Compound #48,51,53,56,57,62,80,85 Kept on LR and Compound #26,32,54,65,77 Kept on QR	RC/JU	Ok,M
6	SSTDICCC040	SSTDICCC040	BP021485.D	13 Aug 2024 15:26	The Calibration is Good For 8270 DOD except com#32,77 and good for 8270E	RC/JU	Ok,M
7	SSTDICC050	SSTDICC050	BP021486.D	13 Aug 2024 16:09		RC/JU	Ok,M
8	SSTDICC060	SSTDICC060	BP021487.D	13 Aug 2024 16:53		RC/JU	Ok,M
9	SSTDICC080	SSTDICC080	BP021488.D	13 Aug 2024 17:36	Compound #9 removed from 80ppm	RC/JU	Ok,M
10	SSTDICV040	ICVBP081324	BP021489.D	13 Aug 2024 19:03	Com#32 and 77 failed in ICV for 8270 DOD Criteria	RC/JU	Ok,M
11	PB162612TB	PB162612TB	BP021490.D	13 Aug 2024 19:46	Already analyzed with ok status	RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP082224

Review By	yogesh	Review On	8/23/2024 6:45:58 AM
Supervise By	mohammad	Supervise On	8/23/2024 6:56:49 AM
SubDirectory	BP082224	HP Acquire Method	BNA_P
		HP Processing Method	BP081324

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6595,SP6594,SP6593,SP6592,SP6591,SP6590,SP6589,SP6588
CCC	SP6591
Internal Standard/PEM	S12041,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BP021606.D	22 Aug 2024 08:43		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BP021607.D	22 Aug 2024 10:09		RC/JU	Ok,M
3	PB162876BL	PB162876BL	BP021608.D	22 Aug 2024 10:51	Analyzed for Contamination check	RC/JU	Not Ok
4	PB162379BL	PB162379BL	BP021609.D	22 Aug 2024 11:33		RC/JU	Ok
5	PB162379BS	PB162379BS	BP021610.D	22 Aug 2024 12:15		RC/JU	Ok,M
6	P3671-06	S-905-KJ-SO-0-0.5-081	BP021611.D	22 Aug 2024 12:57		RC/JU	Ok,M
7	P3671-07	S-905-KJ-SO-0-0.5-081	BP021612.D	22 Aug 2024 13:39		RC/JU	Ok
8	P3671-05	S-911-K1-SO-0-0.5-081	BP021613.D	22 Aug 2024 14:21		RC/JU	Ok,M
9	P3671-01	S-904-J-SO-0-0.5-0819	BP021614.D	22 Aug 2024 15:04	Need 5X Dilution, SIM Analysis not required	RC/JU	Dilution
10	P3689-02	WC-1	BP021615.D	22 Aug 2024 15:48		RC/JU	Ok
11	P3689-02MS	WC-1MS	BP021616.D	22 Aug 2024 16:31		RC/JU	Ok,M
12	P3689-02MSD	WC-1MSD	BP021617.D	22 Aug 2024 17:14		RC/JU	Ok,M
13	P3518-21	OU4-TS-11-080724	BP021618.D	22 Aug 2024 17:58	END CCC Out of Tune Time,Surrogate Fail. Rerun with END CCC	RC/JU	ReRun
14	P3518-17	OU4-TS-09-080724	BP021619.D	22 Aug 2024 18:41	END CCC Out of Tune Time	RC/JU	Not Ok
15	P3671-02	S-904-J-SO-0.5-1.0-081	BP021620.D	22 Aug 2024 19:23	SIM Analysis not required	RC/JU	Ok,M
16	P3671-11	S-903-J-SO-0-0.5-0819	BP021621.D	22 Aug 2024 20:06	SIM Analysis not required	RC/JU	Ok,M

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP082224

Review By	yogesh	Review On	8/23/2024 6:45:58 AM		
Supervise By	mohammad	Supervise On	8/23/2024 6:56:49 AM		
SubDirectory	BP082224	HP Acquire Method	BNA_P	HP Processing Method	BP081324

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6595,SP6594,SP6593,SP6592,SP6591,SP6590,SP6589,SP6588
CCC	SP6591
Internal Standard/PEM	S12041,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

17	SSTDCCC040	SSTDCCC040EC	BP021622.D	22 Aug 2024 20:49	END CCC Out of Tune Time	RC/JU	Not Ok
----	------------	--------------	------------	-------------------	--------------------------	-------	--------

M : Manual Integration