

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P3671 SAS No.: P3671 SDG NO.: P3671

EPA Sample No.: SSTDCCC040 Date Analyzed: 08/26/2024

Lab File ID: BF139159.D Time Analyzed: 09:43

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	95428	6.898	349140	8.18	213475	9.94
UPPER LIMIT	190856	7.398	698280	8.681	426950	10.439
LOWER LIMIT	47714	6.398	174570	7.681	106738	9.439
EPA SAMPLE NO.						
01 S-904-J-SO-0-0.5-081924DL	92486	6.90	347656	8.18	198607	9.93

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.