

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P3671 SAS No.: P3671 SDG NO.: P3671

EPA Sample No.: SSTDCCC040 Date Analyzed: 08/26/2024

Lab File ID: BF139159.D Time Analyzed: 09:43

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	377412	11.422	203247	14.063	185607	15.539
UPPER LIMIT	754824	11.922	406494	14.563	371214	16.039
LOWER LIMIT	188706	10.922	101624	13.563	92803.5	15.039
EPA SAMPLE NO.						
01 S-904-J-SO-0-0.5-081924DL	323380	11.42	172392	14.06	177359	15.54

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.