

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: JACO05

Lab Code: CHEM Case No.: P3671 SAS No.: P3671 SDG No.: P3671

Instrument ID: MSVOA_Y Calibration Date/Time: 08/22/2024 01:25

Lab File ID: VY019191.D Init. Calib. Date(s): 08/15/2024 08/15/2024

Heated Purge: (Y/N) Y Init. Calib. Time(s): 17:24 19:50

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	Amount	Conc.	%Dev	MAX%D
Bromomethane	50.000	46.18	7.650002	20
Methylene Chloride	50.000	49.18	1.636002	20
Cyclohexane	50.000	48.57	2.860001	20