

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: JACO05

Lab Code: CHEM Case No.: P3671 SAS No.: P3671 SDG No.: P3671

Instrument ID: MSVOA\_Y Calibration Date/Time: 08/22/2024 11:22

Lab File ID: VY019211.D Init. Calib. Date(s): 08/15/2024 08/15/2024

Heated Purge: (Y/N) Y Init. Calib. Time(s): 17:24 19:50

GC Column: RXI-624 ID: 0.25 (mm)

| COMPOUND           | Amount | Conc. | %Dev      | MAX%D |
|--------------------|--------|-------|-----------|-------|
| Bromomethane       | 50.000 | 60.60 | -21.192   | 20    |
| Methylene Chloride | 50.000 | 51.60 | -3.204002 | 20    |
| Cyclohexane        | 50.000 | 52.65 | -5.302002 | 20    |