

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082124\
 Data File : VY019191.D
 Acq On : 22 Aug 2024 01:25
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050

Quant Time: Aug 22 03:46:16 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081524S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 16 05:10:30 2024
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 08/23/2024
 Supervised By :Semsettin Yesilyurt 08/23/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	132896	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	8.609	114	216735	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	191813	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.352	152	94816	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.055	65	67673	50.799	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery =	101.600%		
35) Dibromofluoromethane	7.628	113	63231	44.619	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery =	89.240%		
50) Toluene-d8	10.103	98	248885	48.803	ug/l	0.00
Spiked Amount 50.000	Range 58 - 134		Recovery =	97.600%		
62) 4-Bromofluorobenzene	12.408	95	83799	48.201	ug/l	0.00
Spiked Amount 50.000	Range 29 - 146		Recovery =	96.400%		
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.867	85	45882	47.057	ug/l	98
3) Chloromethane	2.068	50	74098	42.162	ug/l	99
4) Vinyl Chloride	2.202	62	89620	41.283	ug/l	99
5) Bromomethane	2.592	94	61804	46.175	ug/l	98
6) Chloroethane	2.732	64	57530	40.920	ug/l	98
7) Trichlorofluoromethane	3.049	101	134601	41.011	ug/l	96
8) Diethyl Ether	3.446	74	32960	46.525	ug/l	93
9) 1,1,2-Trichlorotrifluo...	3.812	101	67135	46.011	ug/l	98
10) Methyl Iodide	4.000	142	75012	52.449	ug/l	92
11) Tert butyl alcohol	4.830	59	29943m	266.517	ug/l	
12) 1,1-Dichloroethene	3.787	96	63351	47.789	ug/l	98
13) Acrolein	3.647	56	18601	180.429	ug/l	97
14) Allyl chloride	4.378	41	90744	51.486	ug/l	91
15) Acrylonitrile	5.043	53	71987	248.340	ug/l	97
16) Acetone	3.860	43	56076	226.976	ug/l	88
17) Carbon Disulfide	4.098	76	178810	47.823	ug/l	98
18) Methyl Acetate	4.378	43	37698	49.718	ug/l	96
19) Methyl tert-butyl Ether	5.116	73	160993	50.093	ug/l	97
20) Methylene Chloride	4.610	84	75718	49.182	ug/l	94
21) trans-1,2-Dichloroethene	5.104	96	72499	48.952	ug/l	98
22) Diisopropyl ether	6.012	45	209397	51.898	ug/l	94
23) Vinyl Acetate	5.951	43	773220	242.055	ug/l	97
24) 1,1-Dichloroethane	5.909	63	121850	49.383	ug/l	96
25) 2-Butanone	6.878	43	92314	232.016	ug/l #	87
26) 2,2-Dichloropropane	6.878	77	91885	45.986	ug/l	98
27) cis-1,2-Dichloroethene	6.884	96	84179	48.298	ug/l	98
28) Bromochloromethane	7.238	49	53495	51.683	ug/l	95
29) Tetrahydrofuran	7.256	42	63725	249.154	ug/l	91
30) Chloroform	7.415	83	136587	49.677	ug/l	97
31) Cyclohexane	7.689	56	92788	48.570	ug/l	96
32) 1,1,1-Trichloroethane	7.610	97	117635	48.684	ug/l	99
36) 1,1-Dichloropropene	7.829	75	93910	48.258	ug/l	98
37) Ethyl Acetate	6.976	43	43643	50.620	ug/l	96
38) Carbon Tetrachloride	7.811	117	113247	50.216	ug/l	96
39) Methylcyclohexane	9.103	83	114769	45.958	ug/l	96
40) Benzene	8.073	78	290534	48.680	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082124\
 Data File : VY019191.D
 Acq On : 22 Aug 2024 01:25
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 08/23/2024
 Supervised By :Semsettin Yesilyurt 08/23/2024

Quant Time: Aug 22 03:46:16 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081524S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 16 05:10:30 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.232	41	26746m	56.836	ug/l	
42) 1,2-Dichloroethane	8.152	62	82407	51.131	ug/l	95
43) Isopropyl Acetate	8.189	43	87961	50.447	ug/l	91
44) Trichloroethene	8.859	130	77302	46.419	ug/l	100
45) 1,2-Dichloropropane	9.140	63	62986	47.367	ug/l	97
46) Dibromomethane	9.231	93	40741	47.238	ug/l	99
47) Bromodichloromethane	9.420	83	105112	49.676	ug/l	98
48) Methyl methacrylate	9.219	41	40039	51.525	ug/l	86
49) 1,4-Dioxane	9.219	88	9451	904.612	ug/l #	96
51) 4-Methyl-2-Pentanone	9.999	43	237401	256.524	ug/l	91
52) Toluene	10.170	92	194854	49.737	ug/l	96
53) t-1,3-Dichloropropene	10.396	75	87688	49.957	ug/l	96
54) cis-1,3-Dichloropropene	9.853	75	100705	47.706	ug/l #	89
55) 1,1,2-Trichloroethane	10.572	97	53184	47.901	ug/l	97
56) Ethyl methacrylate	10.438	69	71231	49.015	ug/l #	79
57) 1,3-Dichloropropane	10.719	76	88558	48.929	ug/l	99
58) 2-Chloroethyl Vinyl ether	9.707	63	155687	225.822	ug/l	98
59) 2-Hexanone	10.761	43	159841	251.400	ug/l	89
60) Dibromochloromethane	10.914	129	73762	48.777	ug/l	99
61) 1,2-Dibromoethane	11.018	107	50662	47.210	ug/l	98
64) Tetrachloroethene	10.646	164	80693	48.129	ug/l	97
65) Chlorobenzene	11.444	112	209783	47.497	ug/l	98
66) 1,1,1,2-Tetrachloroethane	11.517	131	74802	48.159	ug/l	98
67) Ethyl Benzene	11.517	91	372382	48.576	ug/l	99
68) m/p-Xylenes	11.627	106	292684	96.221	ug/l	99
69) o-Xylene	11.956	106	136761	48.551	ug/l	99
70) Styrene	11.969	104	236434	49.221	ug/l	98
71) Bromoform	12.133	173	42507	47.874	ug/l #	98
73) Isopropylbenzene	12.255	105	356580	49.067	ug/l	99
74) N-amyl acetate	12.072	43	80732	50.946	ug/l	90
75) 1,1,2,2-Tetrachloroethane	12.511	83	57457	44.410	ug/l	97
76) 1,2,3-Trichloropropane	12.560	75	38803m	41.498	ug/l	
77) Bromobenzene	12.536	156	84190	48.239	ug/l	97
78) n-propylbenzene	12.597	91	411766	47.398	ug/l	99
79) 2-Chlorotoluene	12.682	91	237105	48.681	ug/l	99
80) 1,3,5-Trimethylbenzene	12.737	105	264001	44.593	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.310	75	17328	45.767	ug/l	86
82) 4-Chlorotoluene	12.779	91	226899	44.777	ug/l	99
83) tert-Butylbenzene	12.999	119	243855	44.603	ug/l	100
84) 1,2,4-Trimethylbenzene	13.048	105	264462	44.821	ug/l	100
85) sec-Butylbenzene	13.176	105	354356	44.477	ug/l	99
86) p-Isopropyltoluene	13.292	119	306302	46.805	ug/l	98
87) 1,3-Dichlorobenzene	13.292	146	160910	46.750	ug/l	100
88) 1,4-Dichlorobenzene	13.371	146	158953	46.459	ug/l	98
89) n-Butylbenzene	13.621	91	272677	45.702	ug/l	99
90) Hexachloroethane	13.883	117	60985	47.279	ug/l	99
91) 1,2-Dichlorobenzene	13.663	146	142597	47.106	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.279	75	8962	47.087	ug/l	93
93) 1,2,4-Trichlorobenzene	14.925	180	75477	44.213	ug/l	96
94) Hexachlorobutadiene	15.029	225	42627	47.939	ug/l	99
95) Naphthalene	15.145	128	141024	46.287	ug/l	99
96) 1,2,3-Trichlorobenzene	15.334	180	63779	44.296	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082124\
Data File : VY019191.D
Acq On : 22 Aug 2024 01:25
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 08/23/2024
Supervised By :Semsettin Yesilyurt 08/23/2024

Quant Time: Aug 22 03:46:16 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081524S.M
Quant Title : SW846 8260
QLast Update : Fri Aug 16 05:10:30 2024
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.707 min Scan# 91
Delta R.T. 0.012 min
Lab File: VY019191.D
Acq: 22 Aug 2024 01:25




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion:168 Resp: 132890
Ion Ratio Lower Upper
168 100
99 49.2 39.1 58.7




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 08/23/2024
Supervised By :Semsettin Yesilyurt 08/23/2024



#2
Dichlorodifluoromethane
Concen: 47.057 ug/l
RT: 1.867 min Scan# 24
Delta R.T. 0.006 min
Lab File: VY019191.D
Acq: 22 Aug 2024 01:25


Tgt Ion: 85 Resp: 45882
Ion Ratio Lower Upper
85 100
87 30.7 16.0 48.0





#3
Chloromethane
Concen: 42.162 ug/l
RT: 2.068 min Scan# 51
Delta R.T. 0.006 min
Lab File: VY019191.D
Acq: 22 Aug 2024 01:25




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 50 Resp: 74098
Ion Ratio Lower Upper
50 100
52 32.8 25.6 38.4




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 08/23/2024
Supervised By :Semsettin Yesilyurt 08/23/2024



#4
Vinyl Chloride
Concen: 41.283 ug/l
RT: 2.202 min Scan# 79
Delta R.T. 0.006 min
Lab File: VY019191.D
Acq: 22 Aug 2024 01:25


Tgt Ion: 62 Resp: 89620
Ion Ratio Lower Upper
62 100
64 33.4 26.1 39.1





#5
Bromomethane
Concen: 46.175 ug/l
RT: 2.592 min Scan# 142
Delta R.T. 0.006 min
Lab File: VY019191.D
Acq: 22 Aug 2024 01:25




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 94 Resp: 61804
Ion Ratio Lower Upper
94 100
96 94.5 74.1 111.1

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 08/23/2024
Supervised By :Semsettin Yesilyurt 08/23/2024




#6
Chloroethane
Concen: 40.920 ug/l
RT: 2.732 min Scan# 166
Delta R.T. 0.006 min
Lab File: VY019191.D
Acq: 22 Aug 2024 01:25




Tgt Ion: 64 Resp: 57530
Ion Ratio Lower Upper
64 100
66 31.6 26.1 39.1





#7
Trichlorofluoromethane
Concen: 41.011 ug/l
RT: 3.049 min Scan# 217
Delta R.T. 0.006 min
Lab File: VY019191.D
Acq: 22 Aug 2024 01:25




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion:101 Resp: 13460
Ion Ratio Lower Upper
101 100
103 64.7 49.3 73.9




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 08/23/2024
Supervised By :Semsettin Yesilyurt 08/23/2024



#8
Diethyl Ether
Concen: 46.525 ug/l
RT: 3.446 min Scan# 283
Delta R.T. 0.006 min
Lab File: VY019191.D
Acq: 22 Aug 2024 01:25

Tgt Ion: 74 Resp: 32960
Ion Ratio Lower Upper
74 100
45 89.1 41.3 124.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 46.011 ug/l
RT: 3.812 min Scan# 341
Delta R.T. 0.012 min
Lab File: VY019191.D
Acq: 22 Aug 2024 01:25

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion:101 Resp: 6713
Ion Ratio Lower Upper
101 100
85 45.4 34.5 51.7
151 78.6 61.7 92.5




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 08/23/2024
Supervised By :Semsettin Yesilyurt 08/23/2024



#10
Methyl Iodide
Concen: 52.449 ug/l
RT: 4.000 min Scan# 374
Delta R.T. 0.012 min
Lab File: VY019191.D
Acq: 22 Aug 2024 01:25


Tgt Ion:142 Resp: 75012
Ion Ratio Lower Upper
142 100
127 48.0 32.8 49.2
141 14.4 11.4 17.2





#11
Tert butyl alcohol
Concen: 266.517 ug/l m
RT: 4.830 min Scan# 512
Delta R.T. -0.012 min
Lab File: VY019191.D
Acq: 22 Aug 2024 01:25




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 59 Resp: 2994
Ion Ratio Lower Upper
59 100
57 8.7 3.0 4.6

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 08/23/2024
Supervised By :Semsettin Yesilyurt 08/23/2024




#12
1,1-Dichloroethene
Concen: 47.789 ug/l
RT: 3.787 min Scan# 339
Delta R.T. 0.012 min
Lab File: VY019191.D
Acq: 22 Aug 2024 01:25



Tgt Ion: 96 Resp: 63351
Ion Ratio Lower Upper
96 100
61 145.5 114.6 171.8
98 64.0 50.8 76.2





#13

Acrolein

Concen: 180.429 ug/l

RT: 3.647 min Scan# 314

Delta R.T. 0.012 min

Lab File: VY019191.D

Acq: 22 Aug 2024 01:25

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 56 Resp: 1860

Ion Ratio Lower Upper

56 100





55 72.3 55.7 83.5

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 08/23/2024

Supervised By :Semsettin Yesilyurt 08/23/2024



#14

Allyl chloride

Concen: 51.486 ug/l

RT: 4.378 min Scan# 436

Delta R.T. 0.012 min

Lab File: VY019191.D

Acq: 22 Aug 2024 01:25




Tgt Ion: 41 Resp: 90744


Ion Ratio Lower Upper

41 100

39 79.6 58.3 87.5

76 41.4 39.0 58.6









#15
Acrylonitrile
Concen: 248.340 ug/l
RT: 5.043 min Scan# 54
Delta R.T. 0.000 min
Lab File: VY019191.D
Acq: 22 Aug 2024 01:25

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion: 53 Resp: 7198
Ion Ratio Lower Upper
53 100
52 82.7 64.7 97.1
51 39.0 28.4 42.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 08/23/2024
Supervised By :Semsettin Yesilyurt 08/23/2024



#16
Acetone
Concen: 226.976 ug/l
RT: 3.860 min Scan# 351
Delta R.T. 0.006 min
Lab File: VY019191.D
Acq: 22 Aug 2024 01:25

Tgt Ion: 43 Resp: 56076
Ion Ratio Lower Upper
43 100
58 32.0 31.7 47.5





#17

Carbon Disulfide

Concen: 47.823 ug/l

RT: 4.098 min Scan# 390

Delta R.T. 0.006 min

Lab File: VY019191.D

Acq: 22 Aug 2024 01:25

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 76 Resp: 178810

Ion Ratio Lower Upper

76 100





78 9.1 6.8 10.2

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 08/23/2024

Supervised By :Semsettin Yesilyurt 08/23/2024



#18

Methyl Acetate

Concen: 49.718 ug/l

RT: 4.378 min Scan# 436

Delta R.T. 0.006 min

Lab File: VY019191.D




Acq: 22 Aug 2024 01:25


Tgt Ion: 43 Resp: 37698

Ion Ratio Lower Upper

43 100

74 26.7 23.3 34.9









#19
Methyl tert-butyl Ether
Concen: 50.093 ug/l
RT: 5.116 min Scan# 511
Delta R.T. 0.018 min
Lab File: VY019191.D
Acq: 22 Aug 2024 01:25

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion: 73 Resp: 16099
Ion Ratio Lower Upper
73 100
57 20.2 17.3 25.9




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 08/23/2024
Supervised By :Semsettin Yesilyurt 08/23/2024



#20
Methylene Chloride
Concen: 49.182 ug/l
RT: 4.610 min Scan# 474
Delta R.T. 0.006 min
Lab File: VY019191.D
Acq: 22 Aug 2024 01:25


Tgt Ion: 84 Resp: 75718
Ion Ratio Lower Upper
84 100
49 114.4 84.4 126.6
51 34.7 26.0 39.0
86 60.4 49.5 74.3





#21
trans-1,2-Dichloroethene
Concen: 48.952 ug/l
RT: 5.104 min Scan# 51
Delta R.T. 0.006 min
Lab File: VY019191.D
Acq: 22 Aug 2024 01:25




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 96 Resp: 72499
Ion Ratio Lower Upper
96 100
61 123.0 97.5 146.3
98 61.0 50.5 75.7




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 08/23/2024
Supervised By :Semsettin Yesilyurt 08/23/2024



#22
Diisopropyl ether
Concen: 51.898 ug/l
RT: 6.012 min Scan# 704
Delta R.T. 0.012 min
Lab File: VY019191.D
Acq: 22 Aug 2024 01:25


Tgt Ion: 45 Resp: 209397
Ion Ratio Lower Upper
45 100
43 58.6 42.2 63.2
87 30.2 25.0 37.6
59 11.6 10.5 15.7





#23
 Vinyl Acetate
 Concen: 242.055 ug/l
 RT: 5.951 min Scan# 694
 Delta R.T. 0.006 min
 Lab File: VY019191.D
 Acq: 22 Aug 2024 01:25




Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050




Tgt Ion: 43 Resp: 773220
 Ion Ratio Lower Upper
 43 100
 86 11.9 10.6 16.0

Manual Integrations
 APPROVED



Reviewed By :Mahesh Dadoda 08/23/2024
 Supervised By :Semsettin Yesilyurt 08/23/2024




#24
 1,1-Dichloroethane
 Concen: 49.383 ug/l
 RT: 5.909 min Scan# 687
 Delta R.T. 0.006 min
 Lab File: VY019191.D
 Acq: 22 Aug 2024 01:25



Tgt Ion: 63 Resp: 121850
 Ion Ratio Lower Upper
 63 100
 98 6.2 4.0 12.2
 100 4.5 2.6 8.0









#25
2-Butanone
Concen: 232.016 ug/l
RT: 6.878 min Scan# 846
Delta R.T. 0.006 min
Lab File: VY019191.D
Acq: 22 Aug 2024 01:25

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion: 43 Resp: 92314
Ion Ratio Lower Upper
43 100
72 26.7 27.3 40.9




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 08/23/2024
Supervised By :Semsettin Yesilyurt 08/23/2024



#26
2,2-Dichloropropane
Concen: 45.986 ug/l
RT: 6.878 min Scan# 846
Delta R.T. 0.012 min
Lab File: VY019191.D
Acq: 22 Aug 2024 01:25


Tgt Ion: 77 Resp: 91885
Ion Ratio Lower Upper
77 100
97 23.3 12.0 36.1





#27
 cis-1,2-Dichloroethene
 Concen: 48.298 ug/l
 RT: 6.884 min Scan# 845
 Delta R.T. 0.012 min
 Lab File: VY019191.D
 Acq: 22 Aug 2024 01:25




Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050



Tgt Ion: 96 Resp: 84179
 Ion Ratio Lower Upper
 96 100
 61 131.1 0.0 256.2
 98 64.9 0.0 130.0




Manual Integrations
 APPROVED


Reviewed By :Mahesh Dadoda 08/23/2024
 Supervised By :Semsettin Yesilyurt 08/23/2024



#28
 Bromochloromethane
 Concen: 51.683 ug/l
 RT: 7.238 min Scan# 905
 Delta R.T. 0.012 min
 Lab File: VY019191.D
 Acq: 22 Aug 2024 01:25

Tgt Ion: 49 Resp: 53495
 Ion Ratio Lower Upper
 49 100
 129 2.0 0.0 4.6
 130 86.4 72.7 109.1





#29

Tetrahydrofuran

Concen: 249.154 ug/l

RT: 7.256 min Scan# 906

Delta R.T. 0.012 min

Lab File: VY019191.D

Acq: 22 Aug 2024 01:25

Instrument :

MSVOA_Y

ClientSampleId :





VSTDCCC050

Tgt Ion:	42	Resp:	63725
Ion	Ratio	Lower	Upper
42	100		
72	48.1	43.0	64.4
71	44.4	40.8	61.2

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 08/23/2024
Supervised By :Semsettin Yesilyurt 08/23/2024



#30

Chloroform

Concen: 49.677 ug/l




RT: 7.415 min Scan# 934


Delta R.T. 0.006 min

Lab File: VY019191.D

Acq: 22 Aug 2024 01:25


Tgt Ion:	83	Resp:	136587
Ion	Ratio	Lower	Upper
83	100		
85	62.8	52.2	78.4





#31
Cyclohexane
Concen: 48.570 ug/l
RT: 7.689 min Scan# 91
Delta R.T. -0.000 min
Lab File: VY019191.D
Acq: 22 Aug 2024 01:25




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 56 Resp: 9278
Ion Ratio Lower Upper
56 100
69 38.0 28.7 43.1
84 92.7 77.7 116.5

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 08/23/2024
Supervised By :Semsettin Yesilyurt 08/23/2024




#32
1,1,1-Trichloroethane
Concen: 48.684 ug/l
RT: 7.610 min Scan# 966
Delta R.T. 0.006 min
Lab File: VY019191.D
Acq: 22 Aug 2024 01:25



Tgt Ion: 97 Resp: 117635
Ion Ratio Lower Upper
97 100
99 63.2 50.3 75.5
61 42.0 32.6 48.8





#33

1,2-Dichloroethane-d4

Concen: 50.799 ug/l

RT: 8.055 min Scan# 1130

Delta R.T. 0.006 min

Lab File: VY019191.D

Acq: 22 Aug 2024 01:25

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 65 Resp: 6767

Ion Ratio Lower Upper

65 100





67 51.1 0.0 109.6

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 08/23/2024

Supervised By :Semsettin Yesilyurt 08/23/2024



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.609 min Scan# 1130

Delta R.T. 0.006 min

Lab File: VY019191.D

Acq: 22 Aug 2024 01:25




Tgt Ion:114 Resp: 216735


Ion Ratio Lower Upper

114 100

63 17.8 0.0 35.0

88 15.1 0.0 27.2









#35
Dibromofluoromethane
Concen: 44.619 ug/l
RT: 7.628 min Scan# 90
Delta R.T. 0.006 min
Lab File: VY019191.D
Acq: 22 Aug 2024 01:25

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion:113 Resp: 6323
Ion Ratio Lower Upper
113 100
111 103.7 82.2 123.4
192 19.9 15.9 23.9




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 08/23/2024
Supervised By :Semsettin Yesilyurt 08/23/2024



#36
1,1-Dichloropropene
Concen: 48.258 ug/l
RT: 7.829 min Scan# 1002
Delta R.T. 0.006 min
Lab File: VY019191.D
Acq: 22 Aug 2024 01:25

Tgt Ion: 75 Resp: 93910
Ion Ratio Lower Upper
75 100
110 36.9 19.0 57.0
77 30.7 25.3 37.9









#37
Ethyl Acetate
Concen: 50.620 ug/l
RT: 6.976 min Scan# 861
Delta R.T. 0.006 min
Lab File: VY019191.D
Acq: 22 Aug 2024 01:25

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion: 43 Resp: 4364
Ion Ratio Lower Upper
43 100
61 14.4 12.6 19.0
70 11.2 7.8 11.8




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 08/23/2024
Supervised By :Semsettin Yesilyurt 08/23/2024



#38
Carbon Tetrachloride
Concen: 50.216 ug/l
RT: 7.811 min Scan# 999
Delta R.T. 0.006 min
Lab File: VY019191.D
Acq: 22 Aug 2024 01:25

Tgt Ion:117 Resp: 113247
Ion Ratio Lower Upper
117 100
119 98.4 75.8 113.6
121 32.0 24.0 36.0





#39

Methylcyclohexane

Concen: 45.958 ug/l

RT: 9.103 min Scan# 11

Delta R.T. 0.006 min

Lab File: VY019191.D

Acq: 22 Aug 2024 01:25

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 83 Resp: 114769

Ion Ratio Lower Upper

83 100

55 68.6 51.5 77.3





98 48.5 40.5 60.7

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 08/23/2024

Supervised By :Semsettin Yesilyurt 08/23/2024



#40

Benzene

Concen: 48.680 ug/l

RT: 8.073 min Scan# 1042

Delta R.T. 0.006 min

Lab File: VY019191.D




Acq: 22 Aug 2024 01:25


Tgt Ion: 78 Resp: 290534

Ion Ratio Lower Upper

78 100


77 23.8 19.5 29.3





#41
Methacrylonitrile
Concen: 56.836 ug/l m
RT: 7.232 min Scan# 904
Delta R.T. 0.000 min
Lab File: VY019191.D
Acq: 22 Aug 2024 01:25




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 41 Resp: 26740
Ion Ratio Lower Upper
41 100
39 61.6 48.0 72.0
67 0.0 75.1 112.7
52 0.0 29.8 44.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 08/23/2024
Supervised By :Semsettin Yesilyurt 08/23/2024



#42
1,2-Dichloroethane
Concen: 51.131 ug/l
RT: 8.152 min Scan# 1055
Delta R.T. 0.006 min
Lab File: VY019191.D
Acq: 22 Aug 2024 01:25

Tgt Ion: 62 Resp: 82407
Ion Ratio Lower Upper
62 100
98 9.2 0.0 22.0









#43
Isopropyl Acetate
Concen: 50.447 ug/l
RT: 8.189 min Scan# 1060
Delta R.T. 0.006 min
Lab File: VY019191.D
Acq: 22 Aug 2024 01:25

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion: 43 Resp: 8796
Ion Ratio Lower Upper
43 100
61 29.1 27.9 41.9
87 13.6 13.1 19.7




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 08/23/2024
Supervised By :Semsettin Yesilyurt 08/23/2024



#44
Trichloroethene
Concen: 46.419 ug/l
RT: 8.859 min Scan# 1171
Delta R.T. 0.006 min
Lab File: VY019191.D
Acq: 22 Aug 2024 01:25

Tgt Ion:130 Resp: 77302
Ion Ratio Lower Upper
130 100
95 93.6 0.0 188.0





#45

1,2-Dichloropropane

Concen: 47.367 ug/l

RT: 9.140 min Scan# 11

Delta R.T. 0.012 min

Lab File: VY019191.D

Acq: 22 Aug 2024 01:25

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 63 Resp: 6298

Ion Ratio Lower Upper

63 100





65 30.3 25.4 38.2

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 08/23/2024

Supervised By :Semsettin Yesilyurt 08/23/2024



#46

Dibromomethane

Concen: 47.238 ug/l

RT: 9.231 min Scan# 1232

Delta R.T. 0.012 min

Lab File: VY019191.D

Acq: 22 Aug 2024 01:25




Tgt Ion: 93 Resp: 40741


Ion Ratio Lower Upper

93 100

95 84.0 66.2 99.4

174 100.3 80.9 121.3









#47
Bromodichloromethane
Concen: 49.676 ug/l
RT: 9.420 min Scan# 1262
Delta R.T. 0.006 min
Lab File: VY019191.D
Acq: 22 Aug 2024 01:25

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion: 83 Resp: 10511
Ion Ratio Lower Upper
83 100
85 64.0 50.1 75.1
127 9.3 7.4 11.0




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 08/23/2024
Supervised By :Semsettin Yesilyurt 08/23/2024



#48
Methyl methacrylate
Concen: 51.525 ug/l
RT: 9.219 min Scan# 1230
Delta R.T. 0.012 min
Lab File: VY019191.D
Acq: 22 Aug 2024 01:25

Tgt Ion: 41 Resp: 40039
Ion Ratio Lower Upper
41 100
69 91.1 84.1 126.1
39 64.7 43.6 65.4





#49

1,4-Dioxane

Concen: 904.612 ug/l

RT: 9.219 min Scan# 1229

Delta R.T. 0.006 min

Lab File: VY019191.D

Acq: 22 Aug 2024 01:25

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 88 Resp: 945

Ion Ratio Lower Upper

88 100

43 32.0 0.0 0.0





58 64.5 49.3 73.9

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 08/23/2024

Supervised By :Semsettin Yesilyurt 08/23/2024



#50

Toluene-d8

Concen: 48.803 ug/l

RT: 10.103 min Scan# 1375

Delta R.T. 0.000 min

Lab File: VY019191.D




Acq: 22 Aug 2024 01:25


Tgt Ion: 98 Resp: 248885

Ion Ratio Lower Upper

98 100


100 67.1 52.0 78.0





#51
4-Methyl-2-Pentanone
Concen: 256.524 ug/l
RT: 9.999 min Scan# 1357
Delta R.T. 0.006 min
Lab File: VY019191.D
Acq: 22 Aug 2024 01:25




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 43 Resp: 23740
Ion Ratio Lower Upper
43 100
58 39.0 35.7 53.5




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 08/23/2024
Supervised By :Semsettin Yesilyurt 08/23/2024



#52
Toluene
Concen: 49.737 ug/l
RT: 10.170 min Scan# 1386
Delta R.T. 0.006 min
Lab File: VY019191.D
Acq: 22 Aug 2024 01:25


Tgt Ion: 92 Resp: 194854
Ion Ratio Lower Upper
92 100
91 168.7 139.0 208.6





#53
 t-1,3-Dichloropropene
 Concen: 49.957 ug/l
 RT: 10.396 min Scan# 1423
 Delta R.T. 0.012 min
 Lab File: VY019191.D
 Acq: 22 Aug 2024 01:25




Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050



Tgt Ion: 75 Resp: 87688
 Ion Ratio Lower Upper
 75 100
 77 32.1 24.1 36.1




Manual Integrations
 APPROVED


Reviewed By :Mahesh Dadoda 08/23/2024
 Supervised By :Semsettin Yesilyurt 08/23/2024



#54
 cis-1,3-Dichloropropene
 Concen: 47.706 ug/l
 RT: 9.853 min Scan# 1334
 Delta R.T. 0.006 min
 Lab File: VY019191.D
 Acq: 22 Aug 2024 01:25


Tgt Ion: 75 Resp: 100705
 Ion Ratio Lower Upper
 75 100
 77 32.2 25.4 38.0
 39 48.5 29.8 44.6#





#55
1,1,2-Trichloroethane
Concen: 47.901 ug/l
RT: 10.572 min Scan# 1451
Delta R.T. 0.006 min
Lab File: VY019191.D
Acq: 22 Aug 2024 01:25




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion:	97	Resp:	53184
Ion	Ratio	Lower	Upper
97	100		
83	83.7	67.5	101.3
85	54.0	42.9	64.3
99	64.1	46.6	70.0




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 08/23/2024
Supervised By :Semsettin Yesilyurt 08/23/2024



#56
Ethyl methacrylate
Concen: 49.015 ug/l
RT: 10.438 min Scan# 1430
Delta R.T. 0.006 min
Lab File: VY019191.D
Acq: 22 Aug 2024 01:25


Tgt Ion:	69	Resp:	71231
Ion	Ratio	Lower	Upper
69	100		
41	67.1	42.3	63.5#
39	40.4	22.9	34.3#





#57
1,3-Dichloropropane
Concen: 48.929 ug/l
RT: 10.719 min Scan# 1475
Delta R.T. 0.006 min
Lab File: VY019191.D
Acq: 22 Aug 2024 01:25




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 76 Resp: 88558
Ion Ratio Lower Upper
76 100
78 32.5 26.2 39.4

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 08/23/2024
Supervised By :Semsettin Yesilyurt 08/23/2024




#58
2-Chloroethyl Vinyl ether
Concen: 225.822 ug/l
RT: 9.707 min Scan# 1310
Delta R.T. 0.006 min
Lab File: VY019191.D
Acq: 22 Aug 2024 01:25



Tgt Ion: 63 Resp: 155687
Ion Ratio Lower Upper
63 100
106 30.8 23.7 35.5









#59
2-Hexanone
Concen: 251.400 ug/l
RT: 10.761 min Scan# 1483
Delta R.T. 0.006 min
Lab File: VY019191.D
Acq: 22 Aug 2024 01:25

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion: 43 Resp: 15984
Ion Ratio Lower Upper
43 100
58 54.6 31.4 94.0




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 08/23/2024
Supervised By :Semsettin Yesilyurt 08/23/2024



#60
Dibromochloromethane
Concen: 48.777 ug/l
RT: 10.914 min Scan# 1508
Delta R.T. 0.006 min
Lab File: VY019191.D
Acq: 22 Aug 2024 01:25

Tgt Ion:129 Resp: 73762
Ion Ratio Lower Upper
129 100
127 76.7 39.0 117.0









#61
1,2-Dibromoethane
Concen: 47.210 ug/l
RT: 11.018 min Scan# 1107
Delta R.T. 0.006 min
Lab File: VY019191.D
Acq: 22 Aug 2024 01:25

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion:107 Resp: 50662
Ion Ratio Lower Upper
107 100
109 96.4 75.9 113.9




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 08/23/2024
Supervised By :Semsettin Yesilyurt 08/23/2024



#62
4-Bromofluorobenzene
Concen: 48.201 ug/l
RT: 12.408 min Scan# 1753
Delta R.T. 0.006 min
Lab File: VY019191.D
Acq: 22 Aug 2024 01:25


Tgt Ion: 95 Resp: 83799
Ion Ratio Lower Upper
95 100
174 87.8 0.0 175.6
176 87.6 0.0 171.4





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.420 min Scan# 11
Delta R.T. 0.006 min
Lab File: VY019191.D
Acq: 22 Aug 2024 01:25




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion:117 Resp: 19181
Ion Ratio Lower Upper
117 100
82 51.5 42.4 63.6
119 32.6 25.9 38.9

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 08/23/2024
Supervised By :Semsettin Yesilyurt 08/23/2024




#64
Tetrachloroethene
Concen: 48.129 ug/l
RT: 10.646 min Scan# 1464
Delta R.T. 0.006 min
Lab File: VY019191.D
Acq: 22 Aug 2024 01:25




Tgt Ion:164 Resp: 80693
Ion Ratio Lower Upper
164 100
166 126.2 102.6 154.0
129 96.1 72.0 108.0
131 89.9 70.0 105.0





#65
Chlorobenzene
Concen: 47.497 ug/l
RT: 11.444 min Scan# 11
Delta R.T. 0.006 min
Lab File: VY019191.D
Acq: 22 Aug 2024 01:25




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion:112 Resp: 20978
Ion Ratio Lower Upper
112 100
114 31.9 26.4 39.6

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 08/23/2024
Supervised By :Semsettin Yesilyurt 08/23/2024




#66
1,1,1,2-Tetrachloroethane
Concen: 48.159 ug/l
RT: 11.517 min Scan# 1607
Delta R.T. 0.006 min
Lab File: VY019191.D
Acq: 22 Aug 2024 01:25




Tgt Ion:131 Resp: 74802
Ion Ratio Lower Upper
131 100
133 94.9 47.8 143.3
119 64.1 34.1 102.2





#67
Ethyl Benzene
Concen: 48.576 ug/l
RT: 11.517 min Scan# 1607
Delta R.T. 0.000 min
Lab File: VY019191.D
Acq: 22 Aug 2024 01:25




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 91 Resp: 37238
Ion Ratio Lower Upper
91 100
106 31.3 24.5 36.7




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 08/23/2024
Supervised By :Semsettin Yesilyurt 08/23/2024



#68
m/p-Xylenes
Concen: 96.221 ug/l
RT: 11.627 min Scan# 1625
Delta R.T. 0.000 min
Lab File: VY019191.D
Acq: 22 Aug 2024 01:25

Tgt Ion:106 Resp: 292684
Ion Ratio Lower Upper
106 100
91 196.5 156.0 234.0









#69
o-Xylene
Concen: 48.551 ug/l
RT: 11.956 min Scan# 1678
Delta R.T. 0.006 min
Lab File: VY019191.D
Acq: 22 Aug 2024 01:25

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion:106 Resp: 13676
Ion Ratio Lower Upper
106 100
91 208.1 103.6 310.8




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 08/23/2024
Supervised By :Semsettin Yesilyurt 08/23/2024



#70
Styrene
Concen: 49.221 ug/l
RT: 11.969 min Scan# 1681
Delta R.T. 0.000 min
Lab File: VY019191.D
Acq: 22 Aug 2024 01:25

Tgt Ion:104 Resp: 236434
Ion Ratio Lower Upper
104 100
78 47.4 37.0 55.4
103 53.8 44.2 66.2





#71

Bromoform

Concen: 47.874 ug/l

RT: 12.133 min Scan# 1707

Delta R.T. 0.006 min

Lab File: VY019191.D


Acq: 22 Aug 2024 01:25

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050



Tgt Ion:173 Resp: 4250

Ion Ratio Lower Upper

173 100

175 49.1 25.1 75.3




254 0.0 0.0 0.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 08/23/2024

Supervised By :Semsettin Yesilyurt 08/23/2024



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.352 min Scan# 1908

Delta R.T. 0.006 min

Lab File: VY019191.D

Acq: 22 Aug 2024 01:25




Tgt Ion:152 Resp: 94816


Ion Ratio Lower Upper

152 100

115 58.0 28.2 84.7

150 173.7 0.0 345.6





#73

Isopropylbenzene

Concen: 49.067 ug/l

RT: 12.255 min Scan# 1728

Delta R.T. 0.000 min

Lab File: VY019191.D

Acq: 22 Aug 2024 01:25

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion:105 Resp: 356580

Ion Ratio Lower Upper

105 100





120 27.1 13.4 40.1

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 08/23/2024

Supervised By :Semsettin Yesilyurt 08/23/2024



#74

N-ethyl acetate

Concen: 50.946 ug/l

RT: 12.072 min Scan# 1698

Delta R.T. 0.000 min

Lab File: VY019191.D

Acq: 22 Aug 2024 01:25

Tgt Ion: 43 Resp: 80732




Ion Ratio Lower Upper


43 100

70 43.2 41.4 62.2

55 25.6 23.3 34.9

61 23.3 22.6 33.8





#75

1,1,2,2-Tetrachloroethane

Concen: 44.410 ug/l

RT: 12.511 min Scan# 1770

Delta R.T. 0.006 min

Lab File: VY019191.D

Acq: 22 Aug 2024 01:25

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 83 Resp: 5745

Ion Ratio Lower Upper

83 100

131 10.9 5.0 14.9





85 66.0 32.0 96.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 08/23/2024

Supervised By :Semsettin Yesilyurt 08/23/2024



#76

1,2,3-Trichloropropane

Concen: 41.498 ug/l m

RT: 12.560 min Scan# 1778

Delta R.T. 0.006 min

Lab File: VY019191.D




Acq: 22 Aug 2024 01:25


Tgt Ion: 75 Resp: 38803

Ion Ratio Lower Upper

75 100

77 0.0 0.0 0.0





#77

Bromobenzene

Concen: 48.239 ug/l

RT: 12.536 min Scan# 1773

Delta R.T. 0.006 min

Lab File: VY019191.D

Acq: 22 Aug 2024 01:25

Instrument :

MSVOA_Y

ClientSampleId :





VSTDCCC050

Tgt Ion:	156	Resp:	84190
Ion Ratio	Lower	Upper	
156	100		
77	167.3	87.1	261.3
158	96.5	48.6	145.9

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 08/23/2024
Supervised By :Semsettin Yesilyurt 08/23/2024



#78

n-propylbenzene

Concen: 47.398 ug/l




RT: 12.597 min Scan# 1784


Delta R.T. 0.000 min

Lab File: VY019191.D

Acq: 22 Aug 2024 01:25

Tgt Ion:	91	Resp:	411766
Ion Ratio	Lower	Upper	
91	100		
120	23.6	11.5	34.4









#79
2-Chlorotoluene
Concen: 48.681 ug/l
RT: 12.682 min Scan# 1798
Delta R.T. 0.006 min
Lab File: VY019191.D
Acq: 22 Aug 2024 01:25

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion: 91 Resp: 23710
Ion Ratio Lower Upper
91 100
126 36.3 17.8 53.5




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 08/23/2024
Supervised By :Semsettin Yesilyurt 08/23/2024



#80
1,3,5-Trimethylbenzene
Concen: 44.593 ug/l
RT: 12.737 min Scan# 1807
Delta R.T. 0.000 min
Lab File: VY019191.D
Acq: 22 Aug 2024 01:25

Tgt Ion:105 Resp: 264001
Ion Ratio Lower Upper
105 100
120 49.2 25.1 75.2









#81
trans-1,4-Dichloro-2-butene
Concen: 45.767 ug/l
RT: 12.310 min Scan# 1736
Delta R.T. 0.006 min
Lab File: VY019191.D
Acq: 22 Aug 2024 01:25

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion: 75 Resp: 17328
Ion Ratio Lower Upper
75 100
53 98.7 66.8 100.2
89 49.8 35.8 53.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 08/23/2024
Supervised By :Semsettin Yesilyurt 08/23/2024



#82
4-Chlorotoluene
Concen: 44.777 ug/l
RT: 12.779 min Scan# 1814
Delta R.T. 0.000 min
Lab File: VY019191.D
Acq: 22 Aug 2024 01:25

Tgt Ion: 91 Resp: 226899
Ion Ratio Lower Upper
91 100
126 35.8 17.6 52.9





#83

tert-Butylbenzene

Concen: 44.603 ug/l

RT: 12.999 min Scan# 1849

Delta R.T. 0.006 min

Lab File: VY019191.D

Acq: 22 Aug 2024 01:25

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion:119 Resp: 24385

Ion Ratio Lower Upper

119 100

91 62.8 31.3 93.8





134 26.4 13.4 40.1

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 08/23/2024

Supervised By :Semsettin Yesilyurt 08/23/2024



#84

1,2,4-Trimethylbenzene

Concen: 44.821 ug/l

RT: 13.048 min Scan# 1858

Delta R.T. 0.006 min

Lab File: VY019191.D




Acq: 22 Aug 2024 01:25


Tgt Ion:105 Resp: 264462

Ion Ratio Lower Upper

105 100


120 46.2 23.2 69.6





#85
 sec-Butylbenzene
 Concen: 44.477 ug/l
 RT: 13.176 min Scan# 1879
 Delta R.T. 0.000 min
 Lab File: VY019191.D
 Acq: 22 Aug 2024 01:25




Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050




Tgt Ion:105 Resp: 354350
 Ion Ratio Lower Upper
 105 100
 134 20.7 10.1 30.1

Manual Integrations
 APPROVED



Reviewed By :Mahesh Dadoda 08/23/2024
 Supervised By :Semsettin Yesilyurt 08/23/2024




#86
 p-Isopropyltoluene
 Concen: 46.805 ug/l
 RT: 13.292 min Scan# 1898
 Delta R.T. 0.000 min
 Lab File: VY019191.D
 Acq: 22 Aug 2024 01:25




Tgt Ion:119 Resp: 306302
 Ion Ratio Lower Upper
 119 100
 134 27.0 13.1 39.3
 91 22.8 10.7 32.1





#87
1,3-Dichlorobenzene
Concen: 46.750 ug/l
RT: 13.292 min Scan# 1898
Delta R.T. 0.006 min
Lab File: VY019191.D
Acq: 22 Aug 2024 01:25




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion:146 Resp: 160910
Ion Ratio Lower Upper
146 100
111 38.1 19.0 57.0
148 63.9 31.7 95.1




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 08/23/2024
Supervised By :Semsettin Yesilyurt 08/23/2024



#88
1,4-Dichlorobenzene
Concen: 46.459 ug/l
RT: 13.371 min Scan# 1911
Delta R.T. 0.006 min
Lab File: VY019191.D
Acq: 22 Aug 2024 01:25

Tgt Ion:146 Resp: 158953
Ion Ratio Lower Upper
146 100
111 38.3 18.4 55.4
148 63.6 32.4 97.2









#89
n-Butylbenzene
Concen: 45.702 ug/l
RT: 13.621 min Scan# 1952
Delta R.T. 0.006 min
Lab File: VY019191.D
Acq: 22 Aug 2024 01:25

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion: 91 Resp: 27267
Ion Ratio Lower Upper
91 100
92 53.1 26.6 79.7
134 26.7 12.9 38.7




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 08/23/2024
Supervised By :Semsettin Yesilyurt 08/23/2024



#90
Hexachloroethane
Concen: 47.279 ug/l
RT: 13.883 min Scan# 1995
Delta R.T. 0.006 min
Lab File: VY019191.D
Acq: 22 Aug 2024 01:25

Tgt Ion:117 Resp: 60985
Ion Ratio Lower Upper
117 100
201 73.8 37.4 112.2









#91
1,2-Dichlorobenzene
Concen: 47.106 ug/l
RT: 13.663 min Scan# 1958
Delta R.T. 0.006 min
Lab File: VY019191.D
Acq: 22 Aug 2024 01:25

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion:146 Resp: 14259
Ion Ratio Lower Upper
146 100
111 39.9 19.9 59.7
148 62.4 31.8 95.4




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 08/23/2024
Supervised By :Semsettin Yesilyurt 08/23/2024



#92
1,2-Dibromo-3-Chloropropane
Concen: 47.087 ug/l
RT: 14.279 min Scan# 2060
Delta R.T. 0.006 min
Lab File: VY019191.D
Acq: 22 Aug 2024 01:25

Tgt Ion: 75 Resp: 8962
Ion Ratio Lower Upper
75 100
155 98.9 48.6 145.8
157 113.7 62.7 188.2





#93

1,2,4-Trichlorobenzene

Concen: 44.213 ug/l

RT: 14.925 min Scan# 2166

Delta R.T. 0.006 min

Lab File: VY019191.D

Acq: 22 Aug 2024 01:25

Instrument :

MSVOA_Y

ClientSampleId :





VSTDCCC050

Tgt Ion:180	Resp:	7547
Ion Ratio	Lower	Upper
180	100	
182	92.7	47.9 143.8
145	30.9	14.2 42.6

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 08/23/2024
Supervised By :Semsettin Yesilyurt 08/23/2024



#94

Hexachlorobutadiene

Concen: 47.939 ug/l




RT: 15.029 min Scan# 2183


Delta R.T. 0.006 min

Lab File: VY019191.D

Acq: 22 Aug 2024 01:25

Tgt Ion:225	Resp:	42627
Ion Ratio	Lower	Upper
225	100	
223	61.4	30.4 91.2
227	65.2	31.9 95.9





#95

Naphthalene

Concen: 46.287 ug/l

RT: 15.145 min Scan# 2202

Delta R.T. 0.000 min

Lab File: VY019191.D

Acq: 22 Aug 2024 01:25

Instrument :

MSVOA_Y

ClientSampleId :





VSTDCCC050

Tgt Ion:	128	Resp:	141024
Ion Ratio	Lower	Upper	
128	100		
127	12.9	10.6	16.0
129	11.0	8.6	13.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 08/23/2024
Supervised By :Semsettin Yesilyurt 08/23/2024



#96

1,2,3-Trichlorobenzene

Concen: 44.296 ug/l

RT: 15.334 min Scan# 2233

Delta R.T. 0.006 min

Lab File: VY019191.D

Acq: 22 Aug 2024 01:25

Tgt Ion:	180	Resp:	63779
Ion Ratio	Lower	Upper	
180	100		
182	95.1	47.6	142.9
145	32.0	15.4	46.1

