

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: JACO05
Lab Code: CHEM Case No.: P3671 SAS No.: P3671 SDG NO.: P3671
Lab File ID: VY019191.D Date Analyzed: 08/22/2024
Instrument ID: MSVOA_Y Time Analyzed: 01:25
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	94816	13.352				
UPPER LIMIT	189632	13.852				
LOWER LIMIT	47408	12.852				
EPA SAMPLE NO.						
S-904-J-SO-0.5-1.0-081924	47587	13.35				
S-911-K1-SO-0-0.5-081924	13012 *	13.35				
S-911-K1-SO-0.5-1-081924	31892 *	13.35				
S-911-K1-SO-0-0.5-081924-FD	38679 *	13.35				
S-905-KJ-SO-0-0.5-081924	50718	13.35				
S-905-J-SO-0.5-1.0-081924	65365	13.35				
S-928-K1-SO-0-0.5-081924	44883 *	13.35				
S-928-K1-SO-0.5-1-081924	302 *	13.35				
S-903-J-SO-0-0.5-081924	64388	13.35				
S-903-J-SO-0.5-1.0-081924	51958	13.35				
S-912-J-SO-0-0.5-081924	27197 *	13.35				
VY0821SBL02	72810	13.35				
VY0821SBS02	75884	13.35				
VY0821SBSD02	84746	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.